

เอกสารสอบเทียบเครื่องมือที่ใช้ในการวิเคราะห์



NSG-TISI-TIS 17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 63-400281-1 **Page : 1 of 2**

Submitted by : Special Lab Envi and Consultant Co., Ltd.
47/91 Moo 3, Tambol Tha-it, Pakkred, Nonthaburi 11120

Equipment : Digital Thermometer with TC probe
Temperature Indicator
Manufacturer : Thermo Scientific
Range : -250 °C to 1372 °C
Resolution : 0.1 °C
Model : TEMP 10K
ID No. : 4008958
SL-38

Environment :
Ambient Temperature : (23 ± 2) °C
Relative Humidity : (50 ± 15) %
Line Voltage : (220 ± 22) VAC

Date of Received : 30 May 2020
Date of Calibration : 04 June 2020
Date of Issue : 04 June 2020
Calibrated by : Bunjerd Masri

Calibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4003 by compared with PRT in the liquid bath at the constant controlled temperature.

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Platinum Resistance Thermometer (PRT)
ID No. **Cert. No.** **Due Date**
400001 TT-0016-20 04 Mar 2022
400016 TT-0058-19 07 May 2021
National Institute of Metrology Thailand (NIMT)
National Institute of Metrology Thailand (NIMT)

2. Standard Digital Thermometer
ID No. **Cert. No.** **Due Date**
400003 19E134 06 Jun 2021
400004 19E134 06 Jun 2021
National Institute of Metrology Thailand (NIMT)
National Institute of Metrology Thailand (NIMT)

Approved by :
(Surachai Promthong)
Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%.

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.

Certificate of Calibration

Certificate No. : 63-400281-1 **Page : 2 of 2**

Result of Calibration : Without Adjustment
UUC Condition As-Received : Good

Function : Temperature measurement with Thermocouple probe Type K

Model : Type K
Diameter : 2 mm.
Length : 1500 mm.
Sheath Material : Teflon
ID No. : SL-39

Immersion Depth (mm.)	Standard Reading (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
130	4.0024	4.6	-0.6	0.18
130	104.0005	104.2	-0.2	0.45
130	150.0033	150.0	0.0	0.58
130	180.0009	179.8	0.2	0.65

Model : AD-1218-230
Diameter : 3.5 mm.
Length : 230 mm.
Sheath Material : Stainless

ID No. : SL-40

Immersion Depth (mm.)	Standard Reading (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
124	250.0027	250.3	-0.3	1.2
124	380.0030	379.0	1.0	1.5

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

Signature

Signature



Certificate of Calibration

Certificate No. : 62-400577-1 Page : 1 of 2

Submitted by : Special Lab Envi and Consultant Co., Ltd.
47/91 Moo 3, Tambol Tha-it, Pakkred, Nonthaburi 11120Equipment : Digital Thermometer with TC probe
Temperature IndicatorManufacturer : Thermo Scientific Model : TEMP 10K
Range : -250 °C to 1372 °C Resolution : 0.1 °C
Serial No. : 4008958 ID No. : SL-38Environment : Ambient Temperature : (23 ± 2) °C
Relative Humidity : (50 ± 15) %
Line Voltage : (220 ± 22) VAC

Date of Calibration : 07 December 2019

Date of Issue : 09 December 2019

Calibrated by : Chortip Samchusri

Calibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4003
by compared with PRT in the liquid bath at the constant controlled temperature.

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Platinum Resistance Thermometer (PRT)
ID No. Cert.No. Due Date Traceability
400016 TT-0058-19 07 May 2021 National Institute of Metrology Thailand (NIMT)2. Standard Digital Thermometer
ID No. Cert.No. Due Date Traceability
400003 19E134 06 Jun 2021 National Institute of Metrology Thailand (NIMT)
400004 19E134 06 Jun 2021 National Institute of Metrology Thailand (NIMT)Approved by :
(Surachai Promthong)
Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.

Certificate of Calibration

Certificate No. : 62-400577-1 Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

Model : Type K Sheath Material : Teflon
Diameter : 2 mm. Length : 1500 mm.
Serial No. : N/A ID No. : SL-39

Immersion Depth (mm.)	Standard Reading (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
130	150.0005	150.1	-0.1	0.58

Temperature measurement

Model : AD-1218-230 Sheath Material : Stainless
Diameter : 3.5 mm. Length : 230 mm.
Serial No. : N/A ID No. : SL-40

Immersion Depth (mm.)	Standard Reading (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
124	380.0026	379.2	0.8	1.6

Remark

UUC : Unit Under Calibration

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2 ,
providing a level of confidence of approximately 95%

- 000 -



Certificate of Calibration

Certificate No. :

63-200136-1

Page : 1 of 2

Submitted by :

Special Lab Envi and Consultant Co., Ltd.

47/91 Moo 3, Tambol Tha-It, Pakkred, Nonthaburi 11120

Equipment :

Electronic Balance

Manufacturer : AND Model : GR-200

Serial No. : 14245322

Capacity : 210 g Resolution : 0.0001 g

Environment :

On site calibration was carried out at the Laboratory.

Special Lab Envi and Consultant Co., Ltd.

Ambient Temperature : (27.0 to 27.4) °C

Relative Humidity : (51.1 to 53.0) %

Air Pressure : 1011.0 mbar

Date of Received :

04 May 2020

Date of Calibration :

04 May 2020

Date of Issue :

15 May 2020

Calibrated by :

Akaradath Thippichai

Calibration Method :

In-house method CAL-M2001 based on UKAS Publication ref : LAB 14

Edition 5, July 2015

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Weights

ID No.

Cert. No.

Due Date

Traceability

E261-E2624

C02192873

14 Nov 2020

National Institute of Metrology (Thailand), (NIMT)

Approved by :

(Surachai Promthong)

Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.

Certificate of Calibration

Certificate No. : 63-200136-1

Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Departure of indication from nominal value

Nominal Value (g)	Correction (g)	Uncertainty ± (g)
0.001	0.0001	0.00011
0.01	0.0001	0.00011
0.1	0.0001	0.00011
0.5	-0.0001	0.00011
2	0.0000	0.00011
5	-0.0001	0.00012
10	0.0000	0.00012
50	0.0000	0.00014
100	0.0000	0.00020
200	-0.0001	0.00038

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2.11$, providing a level of confidence of approximately 95%

Eccentric error

Load test : 50 g

A

B

C

D

E

-0.0005


0.0001

0.0004

-0.0002

0.0000

g



Repeatability

Load test : 200 g

Sidev.

: 0.00005

g

-o0o-

Certificate of Calibration

Certificate No. : 63-400218-3

Page : 1 of 2

Submitted by : Special Lab Envi and Consultant Co., Ltd.

47/91 Moo 3 Thambol Tha-it, Pakkret, Nonthaburi 11120

Equipment : Air Chamber (Incubator)

 Manufacturer : Lovibond
 Range : N/A °C
 Model : FKU 1800
 Resolution : 0.1 °C
 Serial No. : 0914643-01
 ID No. : N/A

Environment : On site calibration was carried out at the Laboratory,

Special Lab Envi and Consultant Co., Ltd.

Ambient Temperature : (27.4 to 28.0) °C

Relative Humidity : (45 to 55) %

Line Voltage : (226.0 to 226.5) V

Date of Received : 04 May 2020

Date of Calibration : 04 May 2020

Date of Issue : 04 May 2020

Calibrated by : Bunjerd Masri

Calibration Method : CAL-M4004, TLAS G-20

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Digital Thermometer with Thermocouple probe

ID No.	Cert. No.	Due Date	Traceability
400022 & 400028	63-400107-1	29 Aug 2020	National Institute of Metrology Thailand (NIMT)

Approved by :



(Bunjerd Masri)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.

Certificate of Calibration

Certificate No. : 63-400218-3


Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (± °C)
			1	2	3	4	5	6	7	8	9	
20.0	20.0	20.0	20.0	19.8	20.0	20.0	20.0	19.9	20.2	20.2	20.1	0.77

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Uniformity (°C)	Measured Stability (°C)	Overall Variation (°C)
20.0	20.0	20.0	0.5	0.5	1.3

Remark: The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

-o0o-





MIRACLE INTERNATIONAL TECHNOLOGY CO.,LTD

214 Bangwack Rd. Bangpai Bangkok 10160
Tel.: 0-2865-4647-8 Fax: 0-2865-4649 <http://www.mit.in.th>



CALIBRATION CERTIFICATE

Certificate No. : AD2006-146-0001

Date Issued : 15-Jun-20

Customer : SPECIAL LAB ENVI AND CONSULTANT CO.,LTD.

47/91 Moo 3, Tha-It, Pak Kret, Nonthaburi 11120

Equipment : Hot Air Oven

Manufacturer : Memmert

Model : UN30

Serial No. : B120.0284

ID No./Tag No. : -

Date Received : 12-Jun-20

Date Calibrated : 13-Jun-20

Calibrated by : Mr. Surat Aumarb

Calibration Method or Calibration Procedure Used

Standard method : CP-05 TLAS G-20.

This certificate is traceable to national standards, which realize the units of measurement according to the International System of Units (SI).

Result of Calibration

The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor $k = 2$, providing a level confidence approximately 95 percent.

This certificate may not be reproduced other than in full except with the prior written approval of the Technical Manager, Miracle International Technology Company Limited.

Approved by :

(Mr. Tassanai Suksukon)
Technical Manager

Page 1 of 2



Certificate No. : AD2006-146-0001

Environment : Ambient Temperature : (25 ± 2)°C
Relative Humidity : (50 ± 15)%RH

Calibration Temperature (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Stability ¹ (°C)	Measured Uniformity ² (°C)	Overall Variation ³ (°C)
104	104.0	104.0	0.18	0.42	0.92
150	150.0	150.0	0.35	0.45	1.11
180	180.0	180.0	0.44	0.47	0.88

Without adjustment

Calibration Temperature (°C)	STD No. 1 (°C)	STD No. 2 (°C)	STD No. 3 (°C)	STD No. 4 (°C)	STD No. 5 (°C)	STD No. 6 (°C)	STD No. 7 (°C)	STD No. 8 (°C)	STD No. 9 (°C)	Uncertainty ⁴ (°C)
104	104.32	104.12	103.80	104.33	103.98	103.93	104.01	104.42	104.13	0.95
150	149.93	149.62	149.49	149.80	149.63	149.41	149.48	149.91	149.71	1.0
180	179.45	179.35	179.45	179.18	179.42	179.44	179.32	179.32	179.35	1.1

Note : Probe No. 9 is Reference Probe

Setting Air Fresh No. 0



Measurement Standards Used & Traceability :

The International System of Units (SI) through

MIT Certificate No. L2002-197 for Digital Thermometer with Probe (Agilent) Module 2 (08) TC Serial No. MY44000197, Due 26-Sep-20

- Notes :
1. The temperature stability is the one-half of greatest maximum difference of measured temperatures at any one probe.
 2. The temperature uniformity is the maximum difference of measured temperatures between of any probes and the measured temperature at the reference location which are observed at same time.
 3. Overall variation is the difference of maximum and minimum measured temperatures throughout observation time.
 4. The uncertainty of measurement is included temperature stability.
 5. The temperature uniformity, stability, overall variation and indicating temperature is applicable to all air or gas filled temperature controlled enclosures at atmospheric pressure.

End of Certificate

Page 2 of 2



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG, BANGKOK 10250
TEL. 0-2717-3000-24 FAX. 0-2719-9484



Certificate of Calibration

Certificate No. : 20T1897
Page : 1 of 2

Equipment : pH Meter With Sensor
Manufacturer: Eutech
Model : pH 700
Serial No.: 2858459
ID No.: SL-33
Condition As-Received: Used Item
Received Date: 25 August 2020
Calibration Date: 27 August 2020
Reference: 2008-0964WN
Ambient Temperature: (25 ± 3) °C
Relative Humidity: (50 ± 20) %
Submitted by: Special Lab Envi And Consultant Co.,Ltd
47/91 Moo 3 Tambon Tha-it, Pakkret Nonthaburi 11120

This certificate may not be reproduced other than in full,
except with the prior written approval of the head of
Corporate Services 3: Equipment Calibration and Testing Services.

Procedure used: Calibration were conducted using in-house calibration procedure CP-T01 according to comparison with
Platinum Resistance Thermometer (PRT) into liquid bath temperature controller.
The temperature scale used was based on ITS-90.

Condition of this result of calibration

1.Reference standards instruments :

Instrument	Model	Serial No.	Certificate No.	Due Date
1) Digital Thermometer	1529	A66176	1911397	01 Nov 2020
2) Platinum Resistance Temperature	162 P	3683	1911397	01 Nov 2020

2.The certificate is valid only to the item calibrated on date and place of calibration.
3.This Certification is traceable to the International System of Unit maintained at:-
-National Institute of Metrology Thailand (NIMT)

Calibrated by : Theerapong Ameen
Issue Date : 01 September 2020

Approved Signatory :

[] Phalinee Prabpaipal
[x] Chatchawan Khunpluek
[] Wanlop Larpkum

B 0241421



Cert. No.: 20T1897
Page.: 2 of 2

Result of Calibration:-

Function: Temperature measurement
Without Adjustment

This equipment was connected with Temperature Sensor ID No. SL-33/1
Dimension of probe : Diameter 3.5 mm., Length 115 mm. Sheath material : Stainless Steel

Immersion Depth (mm.)	Standard Temperature (°C)	UUC* Reading (°C)	Error (°C)	Uncertainty of Measurement (± °C)
100	25.0097	25.0	-0.0097	0.12

UUC* : Unit Under Calibration

The reported uncertainty of measurement was based on standard uncertainty multiplied
by a coverage factor k = 2, providing a level of confidence of approximately 95%.

-o0o-

a 1012543



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SIJANLUANG, SIJANLUANG BANGKOK 10250

TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert.No.: 20CH1254
Page.: 1 of 3

Certificate of Calibration

Equipment : pH Meter
Manufacturer : Eutech
Model : pH 700
Serial No. : 2858459
ID No. : SL-33
Condition As-Received: Used Item
Received Date : 25 August 2020
Calibration Date : 26 August 2020
Reference : 2008-0964WN-1
Submitted by : Special Lab Envi And Consultant Co.,Ltd
47/91 Moo 3, Tambon Tha-it,
Pakkret, Nonthaburi 11120
Ambient Temperature : (25 ± 2.5) °C
Relative Humidity : (50 ± 15) %
Calibration Procedure : In - house method :
- CP-CH5 : based on direct measurement by
using standard voltage calibrator and
certified reference material (CRM)
Calibrated by : Walalak Sirithean

Approved by : 
Approved Signatory

() Ponthippa Tameyakul
() Malee Butkruea
() Saithip Meangmai

Issue Date : 28 August 2020

The Uncertainties are for a confidence probability of approximately 95 %

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

A 0018380



Cert. No.: 20CH1254
Page.: 2 of 3

Condition of this calibration result

1. Reference Standard Instrument : -
Instrument : Document Process Calibrator
Model : 753
Serial No. : 46530031
ID No. : 130RC098
Cert. No. : 19E3994
Due Date : 10 Oct 2020
This certification is traceable to the International System of Unit maintained at:-
- Traceable to National Institute of Metrology (Thailand), NIMT

2. Certified Reference Materials : The measurement results are traceable to SI through Merck Ltd.,
Deutsche Akkreditierungsstelle, Accredited No.D-RM-15185-01-00

Buffer Solution	Manufacturer	Lot No.	Exp. date
pH 4.007	Merck	HC99078000	31 May 2022
pH 6.866	Merck	HC99138402	31 May 2022
pH 9.183	Merck	HC99627703	31 May 2021

3. This certificate is valid only to the item calibrated on date and place of calibration.

Calibration Results

Function : pH Measurement

Performing three buffers standard curve by using buffer nominal pH (4,7,9)

Unit Under Calibration	Standard pH Buffer Solution	Actual pH Reading	Actual mV Reading (mV)	Uncertainty of pH measurement (±)	Coverage factor k
pH Electrode S/N.: 2863304	4.007 6.866 9.183	4.01 6.86 9.18	173.3 5.1 -129.8	0.011 0.010 0.045	2.00 2.00 2.00



a 1017092

Certificate of Calibration

Page : 1 of 2

Certificate No. : 63-400218-4

Submitted by : Special Lab Envi and Consultant Co., Ltd.

Equipment : 47/91 Moo 3 Thambol Tha-it Pakkret Nonthaburi 11120

Manufacturer : Frozen

Range : N/A °C Model : CC-280C

Resolution : 0.1 °C

Serial No. : 2081307016 ID No. : N/A

Environment : On site calibration was carried out at the Laboratory.

Special Lab Envi and Consultant Co., Ltd.

Ambient Temperature : (30.0 to 31.3) °C

Relative Humidity : (50 to 55) %

Line Voltage : (226.0 to 226.5) V

Date of Received : 04 May 2020

Date of Calibration : 04 May 2020

Date of Issue : 04 May 2020

Calibrated by : Bunjerd Masri

Calibration Method : CAL-M4004, TLAS G-20

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units
 Standard Digital Thermometer with Thermocouple probe

ID No.

Cert. No.

Due Date

Traceability

400022 & 400023

63-400104-1

29 Aug 2020

National Institute of Metrology Thailand (NIMT)

Approved by :

(Bunjerd Masri)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.

CAL-F0031-03

Cert.No.: 20CH1254

Page.: 3 of 3

Calibration Results

Function : mV Measurement

Performing standard curve by Fluke at pH (4.7,10)

Unit Under Calibration	Nominal Value	Standard Voltage Input		Actual Reading		Uncertainty of Measurement (±mV)	Coverage factor k
		pH	mV	mV	pH		
pH Meter S/N.: 2858459	0.00	414.12	414	414	0.02	0.58	2.00
	1.00	354.96	355	355	1.02	0.58	2.00
	2.00	295.80	296	296	2.02	0.58	2.00
	3.00	236.64	237	237	3.01	0.58	2.00
	4.00	177.48	177.4	177.4	4.01	0.058	2.00
	5.00	118.32	118.2	118.2	5.01	0.11	2.52
	6.00	59.16	59.1	59.1	6.00	0.058	2.00
	6.86	8.28	8.2	8.2	6.86	0.058	2.00
	7.00	0.00	0.0	0.0	7.00	0.058	2.00
	8.00	-59.16	-59.2	-59.2	8.00	0.058	2.00
	9.00	-118.32	-118.3	-118.3	9.01	0.058	2.00
	9.18	-128.97	-129.0	-129.0	9.19	0.058	2.00
	10.00	-177.48	-177.5	-177.5	10.01	0.058	2.00
	11.00	-236.64	-237	-237	11.01	0.58	2.00
	12.00	-295.80	-296	-296	12.02	0.58	2.00
	13.00	-354.96	-355	-355	13.02	0.58	2.00
	14.00	-414.12	-414	-414	14.02	0.58	2.00

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

-o-o-

a 1013829

Certificate of Calibration

Certificate No. : 63-400218-4

Page : 2 of 2

Result of Calibration : Without Adjustment


UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)

Inside of Chamber

W = 1.02 m
D = 0.47 m
H = 1.48 m
Capacity = 0.71 m³



Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (± °C)
			1	2	3	4	5	6	7	8	9	
4.0	4.0	4.0	3.9	4.3	3.7	3.6	4.2	4.6	3.6	3.5	3.9	0.63

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Uniformity (°C)	Measured Stability (°C)	Overall Variation (°C)
4.0	4.0	4.0	0.8	0.3	1.5

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

-o0o-

Certificate No. : CAL-20-293

Page : 1 of 3

CERTIFICATE OF CALIBRATION

Equipment	:	Spectrophotometer
Manufacturer	:	Thermo Scientific
Model	:	Genesys 20
Serial No.	:	3SGT041007
ID No.	:	SL-34
Customer	:	Special Lab Envi And Consultant Co.,Ltd.
	:	47/91 Moo 3, Tambol Tait , Amphur Pakrad,
	:	Nonthaburi, 11120.
Location	:	Becthai Laboratory
Date of Receipt	:	15 May 2020
Date of Calibration	:	15 May 2020
Date of Issue	:	15 May 2020
Ambient Temperature	:	(25±10) °C
Relative Humidity	:	(60±20) %
Condition As-Received	:	Used Item

Calibrated by

(Ms. Alisa Lamor)

Calibration Engineer

Approved by

(Ms. Jintana Sangthajaroenlap)

Calibration Manager

The reported expended uncertainty of measurement was based on a combined standard uncertainty multiplied by a coverage factor k=2.00, providing a level of confidence of approximately 95%.

This certificate may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Indicated values are valid for the state of the Spectrophotometer at the time of calibration only.

ISSUE: 5 REV:4

FM-CAL-33/2

15/05/61



BECTHAI BANGKOK EQUIPMENT & CHEMICAL CO., LTD.
CALIBRATION LABORATORY

300 Phaholyothin Road, Phayathai, Bangkok 10400, Thailand Tel: +66 2615-2929 Fax: +66 2615-2350-1
E-mail: bkk@becthai.com Website: www.becthai.com



NSC-TISI-TIS 17925
CALIBRATION 0131

Certificate No. : CAL-20-293

Page : 2 of 3

CALIBRATION REPORT

Conditions of this result of calibration

1. Reference Standard Material :

Material	Model	Serial No.	Cert.No.	Due date
Holmium Glass Filter	RM-HG	12705	81255	16 Jan 22
Neutral Density Filter	RM-1N2N3N	8323	81257	16 Jan 22

2. Traceability : This certification is traceable to the International System of Unit maintained at:

The Siam Scientific Ltd. Accredited Calibration Laboratory No. 0659.

3. Method of calibration :

The calibration procedure was carried out according to the Guide to CPM-CAL-02 based on ASTM E275-08 (2013) and-
ASTM E925-09 (2014).

4. Result of calibration :

(✓) without adjustment () after adjustment

5. Equipment Specifications:

Spectral Bandwidth :	8	nm
Data Interval :	1	nm
Scan Speed :	N/A	nm/min

ISSUE: 5 REV:4

FM-CAL-33/2

15/05/61

- End of Report -

ISSUE: 5 REV:4

FM-CAL-33/2

15/05/61

Certificate No. : CAL-20-293

Page : 3 of 3

CALIBRATION REPORT

Wavelength Calibration

Certified Values of Reference Material (nm)	Nominal Value (nm)	UUC*Reading (nm)	Error (nm)	Uncertainty of Measurement (± nm)
418.40	418	418	-0.40	0.59
537.00	537	537	0.00	0.59
638.00	638	639	1.00	0.59

Photometric Calibration for Visible

Wavelength (nm)	Certified Values of Reference Material (A)	UUC* Reading (A)	Error (A)	Uncertainty of Measurement (± A)
420.0	Zero	0.000	0.0000	0.0028
	0.5717	0.575	0.0033	0.0033
	0.7341	0.736	0.0019	0.0036
	1.0726	1.077	0.0044	0.0032
440.0	Zero	0.000	0.0000	0.0028
	0.5611	0.562	0.0009	0.0032
	0.7168	0.716	-0.0008	0.0036
	1.0473	1.046	-0.0013	0.0032
465.0	Zero	0.000	0.0000	0.0028
	0.5114	0.515	0.0036	0.0032
	0.6610	0.664	0.0030	0.0035
	0.9651	0.968	0.0029	0.0032
546.1 (546.0)	Zero	0.000	0.0000	0.0028
	0.5233	0.522	-0.0013	0.0034
	0.6693	0.667	-0.0023	0.0032
	0.9796	0.977	-0.0026	0.0031
590.0	Zero	0.000	0.0000	0.0028
	0.5553	0.557	0.0017	0.0033
	0.6987	0.699	0.0003	0.0032
	1.0236	1.023	-0.0006	0.0030
635.0	Zero	0.000	0.0000	0.0028
	0.5411	0.541	-0.0001	0.0033
	0.6673	0.666	-0.0013	0.0032
	0.9771	0.976	-0.0011	0.0031

Remark : Each individual filter is measured against the empty filter holder (blank) used to zero the Spectrophotometer.

Note:

UUC* : Unit Under Calibration



Certificate of Calibration

Certificate No. : 63-210407-1

Page : 1 of 2

Submitted by :

Special Lab Envi and Consultant Co.,Ltd.

47/91 Moo 3, Tambol Tha-I, Pakkret, Nonthaburi 11120

Equipment :

Weight

Manufacturer : LS

Material : Stainless Steel

Weight size : 1 g

ID No. : 60-210017-1

Assumed density of weight : 7950 kg / m³Assumed Air density : 1.2 kg / m³

Environment :

Ambient Temperature : (20 ± 2) °C

Relative Humidity : (50 ± 10) %

Air Pressure : 1004.5 mbar

Date of Received :

22 August 2020

Date of Calibration :

27 August 2020

Date of Issue :

27 August 2020

Calibrated by :

Chanakan Pongsuwan

Calibration Method : In-house method CAL-M2101 based on OIML R 111-1 : 2004(E)

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Weights

ID No.

Cert. No.

Traceability

E2413-E2425 MM-0060-19 27 Mar 2022 National Institute of Metrology (Thailand), (NIMT)

Approved by :

(Surachai Promthong)

Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 63-210407-1

Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

No.	Nominal Value	Id.Mark	Conventional mass Value	Measuring Uncertainty
1	1 g	none	1 g -0.027 mg	± 0.023 mg

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

-o0o-

APD





Certificate of Calibration

Certificate No. : 63-210407-2

Page : 1 of 2

Submitted by :

Special Lab Envi and Consultant Co.,Ltd.
47/91 Moo 3, Tambol Tha-Ih, Pakkret, Nonthaburi 11120

Equipment :

Weight
Manufacturer : LS
Weight size : 100 g
Material : Stainless Steel

ID No. : 60-210017-2

Assumed density of weight : 7950 kg / m³
Assumed Air density : 1.2 kg / m³

Environment :

Ambient Temperature : (20 ± 2) °C
Relative Humidity : (50 ± 10) %
Air Pressure : 1004.2 mbar

Date of Received :

22 August 2020

Date of Calibration :

27 August 2020

Date of Issue :

27 August 2020

Calibrated by :

Charakan Pongsuwan

Calibration Method : In-house method CAL-M2101 based on OIML R 111-1 : 2004(E)

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Weights

ID No.

E2413-E2425

Cert.No.

MM-0060-19

Traceability

Due Date
27 Mar 2022
National Institute of Metrology (Thailand), (NIMT)

Approved by :

(Surachai Promthong)
Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 63-210407-2

Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

No.	Nominal Value	Id.Mark	Conventional mass Value	Measuring Uncertainty
1	100 g	none	100 g -0.20 mg	± 0.11 mg

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

-o-o-





Certificate of Calibration

Certificate No. : 63-210407-3 Page : 1 of 2

Submitted by :

Special Lab Envi and Consultant Co.,Ltd.

47/91 Moo 3, Tambol Tha-Irt, Pakkret, Nonthaburi 11120

Equipment :

Weight

Manufacturer : LS

Weight size : 200 g

Material : Stainless Steel

ID No. : 61-210565-1

Assumed density of weight : 7950 kg / m³Assumed Air density : 1.2 kg / m³

Environment :

Ambient Temperature : (20 ± 2) °C

Relative Humidity : (50 ± 10) %

Air Pressure : 1003.6 mbar

Date of Received :

22 August 2020

Date of Calibration :

27 August 2020

Date of Issue :

27 August 2020

Calibrated by :

Chanakan Pongsuwan

Calibration Method : In-house method CAL-M2101 based on OIML R 111-1 : 2004(E)

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Weights

ID No.

E2413-E2425

Cert. No.

MM-0060-19

Due Date

27 Mar 2022

Traceability

National Institute of Metrology (Thailand), (NIMT)

Approved by :

(Surachai Promthong)

Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.



Certificate of Calibration

Certificate No. : 63-210407-3

Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

No.	Nominal Value	Id.Mark	Conventional mass Value	Measuring Uncertainty
1	200 g	none	200 g -0.05 mg	± 0.17 mg

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

-o0o-

PR

