

ภาคผนวก จ

เอกสารสอบเทียบเครื่องมือ

ภาคผนวก จ-1

เอกสารสอบเทียบเครื่องมือ

บริษัท เอแอลเอส แล็บอราทอรี กรุ๊ป (ประเทศไทย) จำกัด



รายการเครื่องมือที่ใช้ในการวิเคราะห์ / ทดสอบ

Sample Name	Parameter	Equipment Name	ID No.	Calibrated Date	Next Cal	Freq. Calibrate (Months)
Ambient	Particulate Matter (PM-10)	High Volume	BKK_FS1062	-	-	On site Calibration
Ambient	Particulate Matter (PM-10)	High Volume	BKK_FS0387	-	-	On site Calibration
Ambient	Particulate Matter (PM-10)	High Volume	BKK_FS0388	-	-	On site Calibration
Ambient	Particulate Matter (PM-10)	High Volume	BKK_FS0385	-	-	On site Calibration
Ambient	Particulate Matter (PM-10)	Digital Balance	BKK_EN0004	10-Mar-21	10-Mar-22	12
Ambient	Total Suspended Particulate	High Volume	BKK_FS0359	-	-	On site Calibration
Ambient	Total Suspended Particulate	High Volume	BKK_FS0358	-	-	On site Calibration
Ambient	Total Suspended Particulate	High Volume	BKK_FS0368	-	-	On site Calibration
Ambient	Total Suspended Particulate	High Volume	BKK_FS0367	-	-	On site Calibration
Ambient	Total Suspended Particulate	Digital Balance	BKK_EN0004	10-Mar-21	10-Mar-22	12
Ambient	Nitrogen Dioxide	NO ₂ Analyzer	BKK_FS1092	4-Jan-22	4-Jul-22	6
Ambient	Nitrogen Dioxide	NO ₂ Analyzer	BKK_FS1088	4-Jan-22	4-Jul-22	6
Ambient	Nitrogen Dioxide	NO ₂ Analyzer	BKK_FS1072	4-Jan-22	4-Jul-22	6
Ambient	Nitrogen Dioxide	NO ₂ Analyzer	BKK_FS0782	4-Jan-22	4-Jul-22	6
Ambient	Sulfur Dioxide	SO ₂ Analyzer	BKK_FS1091	4-Jan-22	4-Jul-22	6
Ambient	Sulfur Dioxide	SO ₂ Analyzer	BKK_FS1087	4-Jan-22	4-Jul-22	6
Ambient	Sulfur Dioxide	SO ₂ Analyzer	BKK_FS1071	4-Jan-22	4-Jul-22	6
Ambient	Sulfur Dioxide	SO ₂ Analyzer	BKK_FS0781	4-Jan-22	4-Jul-22	6
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	BKK_FS0920	8-Oct-21	8-Apr-23	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	BKK_FS0917	1-Nov-21	2-May-23	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	BKK_FS0918	30-Aug-21	28-Feb-23	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	BKK_FS0919	30-Aug-21	28-Feb-23	18
Ambient	Lead	High Volume	BKK_FS0359	-	-	On site Calibration
Ambient	Lead	High Volume	BKK_FS0358	-	-	On site Calibration
Ambient	Lead	High Volume	BKK_FS0368	-	-	On site Calibration
Ambient	Lead	High Volume	BKK_FS0367	-	-	On site Calibration
Ambient	Nickel	High Volume	BKK_FS0359	-	-	On site Calibration
Ambient	Nickel	High Volume	BKK_FS0358	-	-	On site Calibration
Ambient	Nickel	High Volume	BKK_FS0368	-	-	On site Calibration
Ambient	Nickel	High Volume	BKK_FS0367	-	-	On site Calibration
Ambient	Nickel	High Volume	BKK_FS1056	-	-	On site Calibration
Ambient	Arsenic	High Volume	BKK_FS0359	-	-	On site Calibration
Ambient	Arsenic	High Volume	BKK_FS0358	-	-	On site Calibration
Ambient	Arsenic	High Volume	BKK_FS0368	-	-	On site Calibration
Ambient	Arsenic	High Volume	BKK_FS0367	-	-	On site Calibration

รายการเครื่องมือที่ใช้ในการวิเคราะห์ / ทดสอบ

Sample Name	Parameter	Equipment Name	ID No.	Calibrated Date	Next Cal	Freq. Calibrate (Months)
Ambient	Arsenic	High Volume	BKK_FS1056	-	-	On site Calibration
Ambient	Cadmium	High Volume	BKK_FS0359	-	-	On site Calibration
Ambient	Cadmium	High Volume	BKK_FS0358	-	-	On site Calibration
Ambient	Cadmium	High Volume	BKK_FS0368	-	-	On site Calibration
Ambient	Cadmium	High Volume	BKK_FS0367	-	-	On site Calibration
Ambient	Beryllium	High Volume	BKK_FS0359	-	-	On site Calibration
Ambient	Beryllium	High Volume	BKK_FS0358	-	-	On site Calibration
Ambient	Beryllium	High Volume	BKK_FS0368	-	-	On site Calibration
Ambient	Beryllium	High Volume	BKK_FS0367	-	-	On site Calibration
Ambient	Mercury (Particle Phase)	High Volume	BKK_FS0359	-	-	On site Calibration
Ambient	Mercury (Particle Phase)	High Volume	BKK_FS0358	-	-	On site Calibration
Ambient	Mercury (Particle Phase)	High Volume	BKK_FS0368	-	-	On site Calibration
Ambient	Mercury (Particle Phase)	High Volume	BKK_FS0367	-	-	On site Calibration
Noise	Leq 24 hrs	Sound Calibrator	BKK_FS0630	24-Mar-21	24-Mar-22	12
Noise	Leq 24 hrs	Sound Level Meter	BKK_FS0116	14-Dec-21	14-Dec-22	12
Noise	Leq 24 hrs	Sound Level Meter	BKK_FS0926	7-Jul-21	7-Jul-22	12
Noise	Noise Annoyance	Sound Calibrator	BKK_FS0630	24-Mar-21	24-Mar-22	12
Noise	Noise Annoyance	Sound Level Meter	BKK_FS0116	14-Dec-21	14-Dec-22	12
Noise	Noise Annoyance	Sound Level Meter	BKK_FS0926	7-Jul-21	7-Jul-22	12
Noise	Noise Annoyance	Sound Calibrator	BKK_FS0630	24-Mar-21	24-Mar-22	12
Noise	Noise Annoyance	Sound Level Meter	BKK_FS0115	14-Dec-21	14-Dec-22	12
Noise	Leq 5 min	Sound Calibrator	BKK_FS0631	26-Oct-21	26-Oct-22	12
Noise	Leq 5 min	Sound Level Meter	BKK_FS0875	2-Nov-21	2-Nov-22	12
Noise	Leq 8 hrs	Sound Calibrator	BKK_FS0631	26-Oct-21	26-Oct-22	12
Noise	Leq 8 hrs	Sound Level Meter	BKK_FS0968	12-Jan-22	12-Jan-23	12
Noise	Leq 8 hrs	Sound Level Meter	BKK_FS0927	7-Sep-21	7-Sep-22	12
Noise	Leq 8 hrs	Sound Level Meter	BKK_FS0930	12-Jan-22	12-Jan-23	12
Noise	Leq 8 hrs	Sound Level Meter	BKK_FS0969	12-Jan-22	12-Jan-23	12
Noise	Leq 8 hrs	Sound Level Meter	BKK_FS0876	2-Nov-21	2-Nov-22	12
Noise	Leq 8 hrs	Sound Level Meter	BKK_FS0922	12-Jan-22	12-Jan-23	12
Noise	Leq 8 hrs	Sound Level Meter	BKK_FS0924	12-Oct-21	12-Oct-22	12
Noise	Noise Dose, TWA	Dose Badge Reader	BKK_FS0627	7-Jul-21	7-Jul-22	12
Heat	Heat Stress	Heat Stress Monitor	BKK_FS0669	7-Jun-21	7-Jun-22	12
Heat	Heat Stress	Heat Stress Monitor	BKK_FS0670	11-Jun-21	11-Jun-22	12
Heat	Heat Stress	Heat Stress Monitor	BKK_FS0661	13-Dec-21	13-Dec-22	12



รายการเครื่องมือที่ใช้ในการวิเคราะห์ / ทดสอบ


Sample Name	Parameter	Equipment Name	ID No.	Calibrated Date	Next Cal	Freq. Calibrate (Months)
Heat	Heat Stress	Heat Stress Monitor	BKK_FS0671	30-Sep-21	30-Sep-22	12
Heat	Heat Stress	Heat Stress Monitor	BKK_FS0663	19-May-21	19-May-22	12
Heat	Heat Stress	Heat Stress Monitor	BKK_FS0667	14-Feb-22	14-Feb-23	12
Workplace	Ammonia	Field Rotameter	BKK_FS1016	4-Jan-22	4-Apr-22	3
Workplace	Ammonia	Spectrophotometer	BKK_EN0018	15-Oct-21	15-Oct-22	12
Workplace	Sodium hydroxide as NaOH	Field Rotameter	BKK_FS1016	4-Jan-22	4-Apr-22	3
Workplace	Benzene	Field Rotameter	BKK_FS1014	4-Jan-22	4-Apr-22	3
Workplace	Benzene	GC-MSD	BKK_EN0119	1-Oct-21	1-Apr-23	18
Workplace	Hydrochloric Acid	Field Rotameter	BKK_FS1015	4-Jan-22	4-Apr-22	3
Workplace	Hydrochloric Acid	Ion Chromatography	BKK_EN0069	12-Jan-22	12-Jan-23	12
Workplace	Methanol	Field Rotameter	BKK_FS1014	4-Jan-22	4-Apr-22	3
Workplace	Methanol	GC-FID	BKK_EN0126	21-Oct-21	21-Apr-23	18
Workplace	Total Dust	Field Rotameter	BKK_FS1016	4-Jan-22	4-Apr-22	3
Workplace	Total Dust	Digital Balance	BKK_EN0004	10-Mar-21	10-Mar-22	12
Workplace	Respirable Dust	Field Rotameter	BKK_FS1016	4-Jan-22	4-Apr-22	3
Workplace	Respirable Dust	Digital Balance	BKK_EN0004	10-Mar-21	10-Mar-22	12
Workplace	Toluene	Field Rotameter	BKK_FS1014	4-Jan-22	4-Apr-22	3
Workplace	Toluene	GC-MSD	BKK_EN0119	1-Oct-21	1-Apr-23	18
Workplace	Xylene	Field Rotameter	BKK_FS1014	4-Jan-22	4-Apr-22	3
Workplace	Xylene	GC-MSD	BKK_EN0119	1-Oct-21	1-Apr-23	18



High Volume Air Sampler Calibration Worksheet

Project Site :	Bangpoo Environmental Complex Co., Ltd.	Barometric Pressure (mm Hg) :	758
Calibrate Location :	ภายในพื้นที่โครงการ (Project area)	Temperature (°C) :	31
Calibrate Date :	15-Feb-22	High Volume ID :	BKK_FS1062
CalibrationSheet No.:	C-150222-BKK_FS1062	High Volume Model :	TE-5009X
Calibrator ID:	BKK_FS0624	High Volume S/N :	5686
Calibrator Model :	TE-5028A	Calibrator Slope :	1.64942
Calibrator S/N :	2584	Calibrator Intercept :	-0.02902

Test No.	Delta H ₂ O (inch)	Q _{ad} (m ³ /min)	I : Chart (CFM)	Linear Regression
1	2.2	0.9182	32	Slope : 35.3512 Intercept : -0.6096 Correlation Coefficient : 0.9985
2	2.8	1.0321	36	
3	3.6	1.1664	40	
4	4.2	1.2576	44	
5	5.0	1.3695	48	



Calibrated by : Winyou B.
(Mr. Winyou Boontanai)
Field Scientist(1)


Approved by : Wichan Chuehant
(Mr. Wichan Choonharat)
Manager



High Volume Air Sampler Calibration Worksheet

Project Site :	Bangpoo Environmental Complex Co., Ltd.	Barometric Pressure (mm Hg) :	758
Calibrate Location :	ภายในพื้นที่โครงการ (Community area 2)	Temperature (°C) :	31
Calibrate Date :	15-Feb-22	High Volume ID :	BKK_FS0387
CalibrationSheet No.:	C-150222-BKK_FS0387	High Volume Model :	GI051
Calibrator ID:	BKK_FS0624	High Volume S/N :	1626
Calibrator Model :	TE-5028A	Calibrator Slope :	1.64942
Calibrator S/N :	2584	Calibrator Intercept :	-0.02902

Test No.	Delta H ₂ O (inch)	Q _{ad} (m ³ /min)	I : Chart (CFM)	Linear Regression
1	3.2	1.1014	32	Slope : 26.3666 Intercept : 2.9429 Correlation Coefficient : 0.9994
2	4.2	1.2576	36	
3	5.2	1.3960	40	
4	6.6	1.5691	44	
5	7.8	1.7032	48	



Calibrated by : Winyou B.
(Mr. Winyou Boontanai)
Field Scientist(1)


Approved by : Wichan Chuehant
(Mr. Wichan Choonharat)
Manager



High Volume Air Sampler Calibration Worksheet

Project Site :	Bangpoo Environmental Complex Co., Ltd.	Barometric Pressure (mm Hg) :	758
Calibrate Location :	พื้นที่โรงงาน (Community area 3)	Temperature (°C) :	31
Calibrate Date :	15-Feb-22	High Volume ID :	BKK_FS0388
CalibrationSheet No.:	C-150222-BKK_FS0388	High Volume Model :	TE-5009X
Calibrator ID:	BKK_FS0624	High Volume S/N :	5328
Calibrator Model :	TE-5028A	Calibrator Slope :	1.64942
Calibrator S/N :	2584	Calibrator Intercept :	-0.02902

Test No.	Delta H ₂ O (inch)	Q _{ad} (m ³ /min)	I : Chart (CFM)	Linear Regression
1	2.2	0.9182	32	Slope : 30.3504 Intercept : 4.4852 Correlation Coefficient : 0.9982
2	2.8	1.0321	36	
3	3.6	1.1664	40	
4	4.4	1.2865	44	
5	5.6	1.4476	48	



Winyou B.

Calibrated by

(Mr. Winyou Boontanai)
Field Scientist(1)

Approved by :

Wichan Choonharat


(Mr. Wichan Choonharat)
Manager



High Volume Air Sampler Calibration Worksheet

Project Site :	Bangpoo Environmental Complex Co., Ltd.	Barometric Pressure (mm Hg) :	758
Calibrate Location :	พื้นที่โรงงาน 2 (Community area 4)	Temperature (°C) :	31
Calibrate Date :	15-Feb-22	High Volume ID :	BKK_FS0385
CalibrationSheet No.:	C-150222-BKK_FS0385	High Volume Model :	TE-5009X
Calibrator ID:	BKK_FS0624	High Volume S/N :	4789
Calibrator Model :	TE-5028A	Calibrator Slope :	1.64942
Calibrator S/N :	2584	Calibrator Intercept :	-0.02902

Test No.	Delta H ₂ O (inch)	Q _{ad} (m ³ /min)	I : Chart (CFM)	Linear Regression
1	2.0	0.8768	32	Slope : 35.9970 Intercept : 0.2846 Correlation Coefficient : 0.9986
2	2.6	0.9956	36	
3	3.2	1.1014	40	
4	4.0	1.2280	44	
5	4.6	1.3147	48	



Winyou B.

Calibrated by

(Mr. Winyou Boontanai)
Field Scientist(1)

Approved by :

Wichan Choonharat

(Mr. Wichan Choonharat)
Manager



Sartorius (Thailand) Co., Ltd.

129 Rama 9 Road, Huaykwang, Bangkok 10310
Tel: +66 2643 8361-6 Fax: +66 2643-8367 e-mail: service.thailand@sartorius.com

ISO 17025:2017
CALIBRATION 0215

SARTORIUS

REVIEW BY Smit 7
APPROVED BY LLAL
NEXT CAL DATE 10/3/23

Certificate

of Calibration

Model Number: **XP105DU**
Description: **Semi-micro Balance**
Serial Number: **1123091884**
Manufacturer: **Mettler Toledo**

Certificate No.: 21BCI0084
Issued Date: Friday, March 12, 2021
Reference No.: 501611
Page No.: 1 of 3

Customer Name: **ALS Laboratory Group (Thailand) Co., Ltd.**
104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250.

Calibrated Place: **Lab Room**

Calibrated By: **Mr. Chonchai Inthana**
Calibration Date: **Wednesday, March 10, 2021**

Calibration Procedure No.: This calibration was conducted by
Using in-house calibration procedure number (WI-003)
Based on UKAS LAB 14

Metrological data:
Capacity: **31/120 g** Readability: **0.01/0.1 mg** Temperature: **22.0 °C** ± **5.0 °C**
Humidity: **57.0 % RH** ± **10.0 % RH**
Pressure: **±**

Reasons for calibration
☐ New Installation ☐ Service / Repaired ☒ Re-calibration/ Maintenance ☐ Good Operate ☐ Fair

Measurement Method **UKAS Publication Ref : Lab 14**

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM). The calibration certificate documents the traceability to National Standards, which realise the unit of measurement according to the International Standard System of Units (SI).

Traceability:

Model Number	Description	Traceability	Certificate No.	Due Date
YCS011-522-00	Sartorius weight set 1mg - 200g E2/YCS011-522-00	Sartorius	119934 D-K-19398-01-00	10-Sep-2021
MHB-3825D	Humidity/Barometer/Temp. Lutron MHB-3825D	SPC-RT	C19203076	1-Sep-2021

This certificate relate and apply this equipment only.
This certificate may not be reproduced other than in full except with the prior written approval of the Verification Operation Division Sartorius (Thailand) Co., Ltd.

Mr. Chonchai Inthana (Technical Manager)

S T A M P



Sartorius (Thailand) Co., Ltd.

129 Rama 9 Road, Huaykwang, Bangkok 10310
Tel: +66 2643 8361-6 Fax: +66 2643-8367 e-mail: service.thailand@sartorius.com

Certificate

of Calibration

Model Number: **XP105DU**
Description: **Semi-micro Balance**
Serial Number: **1123091884**
Manufacturer: **Mettler Toledo**

Certificate No.: 21BCI0084
Issued Date: Friday, March 12, 2021
Reference No.: 501611
Page No.: 2 of 3

Calibration Results : Without Adjustment

Repeatability
The repeatability is the ability of a weighing instrument to display nearly identical readings under constant test conditions when the same load within a measurement range is placed repeatedly on the weighing pan in the same manner. The standard deviation is used to express repeatability quantitatively.

Nominal Value : (Low Load)	20 g
20.00001	20.00001
20.00002	20.00002
20.00003	20.00003
20.00004	20.00004
20.00005	20.00005
20.00006	20.00006
20.00007	20.00007
20.00008	20.00008
20.00009	20.00009
20.00010	20.00010
20.00011	20.00011
20.00012	20.00012
20.00013	20.00013
20.00014	20.00014
20.00015	20.00015
20.00016	20.00016
20.00017	20.00017
20.00018	20.00018
20.00019	20.00019
20.00020	20.00020
20.00021	20.00021
20.00022	20.00022
20.00023	20.00023
20.00024	20.00024
20.00025	20.00025
20.00026	20.00026
20.00027	20.00027
20.00028	20.00028
20.00029	20.00029
20.00030	20.00030
20.00031	20.00031
20.00032	20.00032
20.00033	20.00033
20.00034	20.00034
20.00035	20.00035
20.00036	20.00036
20.00037	20.00037
20.00038	20.00038
20.00039	20.00039
20.00040	20.00040
20.00041	20.00041
20.00042	20.00042
20.00043	20.00043
20.00044	20.00044
20.00045	20.00045
20.00046	20.00046
20.00047	20.00047
20.00048	20.00048
20.00049	20.00049
20.00050	20.00050
20.00051	20.00051
20.00052	20.00052
20.00053	20.00053
20.00054	20.00054
20.00055	20.00055
20.00056	20.00056
20.00057	20.00057
20.00058	20.00058
20.00059	20.00059
20.00060	20.00060
20.00061	20.00061
20.00062	20.00062
20.00063	20.00063
20.00064	20.00064
20.00065	20.00065
20.00066	20.00066
20.00067	20.00067
20.00068	20.00068
20.00069	20.00069
20.00070	20.00070
20.00071	20.00071
20.00072	20.00072
20.00073	20.00073
20.00074	20.00074
20.00075	20.00075
20.00076	20.00076
20.00077	20.00077
20.00078	20.00078
20.00079	20.00079
20.00080	20.00080
20.00081	20.00081
20.00082	20.00082
20.00083	20.00083
20.00084	20.00084
20.00085	20.00085
20.00086	20.00086
20.00087	20.00087
20.00088	20.00088
20.00089	20.00089
20.00090	20.00090
20.00091	20.00091
20.00092	20.00092
20.00093	20.00093
20.00094	20.00094
20.00095	20.00095
20.00096	20.00096
20.00097	20.00097
20.00098	20.00098
20.00099	20.00099
20.00100	20.00100

Eccentricity (Off-center loading error)
The off-center loading error is yielded by the difference between the resultant of the load, i.e. 1/3 or 1/4 of maximum capacity, placed in the middle of the weighing pan and between each of four additional measurement points (positions defined according to OIML R76).

Nominal value:	20 g
Tolerance	N/A

Difference

1	2	3	4	5	6
-0.00001	-0.00001	-0.00001	0.00002	0.00002	-

Linearity

The linearity, also called linearity error, describes the deviation of the characteristic curve of a weighing instrument from the linear slope.

Nominal Value (g)	Conventional Mass Value (g)	Displayed Value (g)	Deviation (g)	Uncertainty (g)
0.1	0.10000	0.10001	0.00001	0.000021
0.5	0.50000	0.50001	0.00001	0.000022
1	1.00000	1.00000	0.00000	0.000023
2	2.00001	2.00000	-0.00001	0.000024
5	5.00002	5.00002	0.00000	0.000027
10	10.00003	10.00002	-0.00001	0.000032
15	15.00004	15.00005	0.00001	0.000049
20	20.00001	20.00002	0.00001	0.000099
25	25.00003	25.00005	0.00002	0.000099
30	30.00004	30.00006	0.00002	0.000089

Certificate

of Calibration

Model Number : **XP105DU**
 Description : **Semi-micro Balance**
 Serial Number : **1123091884**
 Manufacturer : **Mettler Toledo**

Certificate No. : **21BC10084**
 Issued Date : **Friday, March 12, 2021**
 Reference No. : **501611**
 Page No. : **3 of 3**

Calibration Results : Without Adjustment

Repeatability		Eccentricity (Off-center loading error)	
The reproducibility is the ability of a weighing instrument to display nearly identical readouts under constant test conditions when the same load within a measurement series is placed repeatedly on the weighing pan in the same manner. The standard deviation is used to express reproducibility quantitatively.		The off-center loading error is yielded by the difference between the readout of the load, i.e. 1/8 or 1/4 of maximum capacity, placed in the middle of the weighing pan and between each of four additional measurement points (positions defined according to OIML R113).	
Nominal Value : (Low Load)	g	Nominal value :	g
Tolerance	N/A	Tolerance	N/A
Standard Deviation		Difference	
100 g		1	
N/A g		2	
100 g		3	
N/A g		4	
100 g		5	
N/A g		6	
Standard Deviation		0.00003	

Linearity		Linearity	
The linearity, also called linearity error, describes the deviation of the characteristic curve of a weighing instrument from the linear slope.		The linearity, also called linearity error, describes the deviation of the characteristic curve of a weighing instrument from the linear slope.	
Nominal Value	Conventional Mass Value	Displayed Value	Deviation
Tolerance	N/A	g	g
50	50.0001	50.0000	-0.0001
55	55.0001	55.0001	0.0000
60	60.0001	60.0001	0.0000
65	65.0001	65.0001	0.0000
70	70.0001	70.0000	-0.0001
80	80.0001	80.0001	0.0000
90	90.0001	90.0000	-0.0001
100	100.0001	100.0000	-0.0001
110	110.0001	110.0000	-0.0001
120	120.0001	120.0000	-0.0001
Uncertainty		g	
50		0.00015	
55		0.00015	
60		0.00015	
65		0.00015	
70		0.00015	
80		0.00016	
90		0.00017	
100		0.00017	
110		0.00026	
120		0.00026	

End of Report



High Volume Air Sampler Calibration Worksheet

Project Site : Bangpoo Environmental Complex Co., Ltd. Barometric Pressure (mm Hg) : 758

Calibrate Location : ภายในพื้นที่โครงการ (Project area) Temperature (°C) : 31

Calibrate Date : 15-Feb-22 High Volume ID : BKK_FS0359


Calibration Sheet No. : C-150222-BKK_FS0359 High Volume Model : TE-5009X

Calibrator ID : BKK_FS0624 High Volume S/N : 5194

Calibrator Model : TE-5028A Calibrator Slope : 1.64942

Calibrator S/N : 2584 Calibrator Intercept : -0.02902

Test No.	Delta H ₂ O (inch)	Q _{std} (m ³ /min)	I: Chart (CFM)	Linear Regression
1	2.8	1.0321	40	Slope : 33.1269
2	3.8	1.1976	46	Intercept : 5.9631
3	4.8	1.3424	50	Correlation Coefficient : 0.9989
4	6.0	1.4974	56	
5	7.2	1.6376	60	



Calibrated by

Winyou B.

Approved by :

Wichan Choonharat

(Mr. Winyou Boontana) Field Scientist(1)

(Mr. Wichan Choonharat) Manager



High Volume Air Sampler Calibration Worksheet



High Volume Air Sampler Calibration Worksheet

Project Site : Bangpoo Environmental Complex Co., Ltd. Barometric Pressure (mm Hg) : 758

Calibrate Location : พื้นที่ภายในอาคาร (Community area 2) Temperature (°C) : 31

Calibrate Date : 15-Feb-22 High Volume ID : BKK_FS0358


Calibration Sheet No.: C-150222-BKK_FS0358 High Volume Model : TE-5009X

Calibrator ID: BKK_FS0624 High Volume S/N : 5193

Calibrator Model : TE-5028A Calibrator Slope : 1.64942

Calibrator S/N : 2584 Calibrator Intercept : -0.02902

Test No.	Delta H ₂ O (inch)	Q _{std} (m ³ /min)	I : Chart (CFM)	Linear Regression
1	2.8	1.0321	40	Slope : 30.1445 Intercept : 8.9311 Correlation Coefficient : 0.9993
2	4.0	1.2280	46	
3	5.0	1.3695	50	
4	6.4	1.5456	56	
5	7.8	1.7032	60	



Winyou B.

Calibrated by

(Mr. Winyou Boontanai)
Field Scientist(1)

Wichan Chuanat

Approved by :

(Mr. Wichan Choonharat)
Manager

Project Site : Bangpoo Environmental Complex Co., Ltd. Barometric Pressure (mm Hg) : 758

Calibrate Location : พื้นที่ภายในอาคาร (Community area 3) Temperature (°C) : 31

Calibrate Date : 15-Feb-22 High Volume ID : BKK_FS0368


Calibration Sheet No.: C-150222-BKK_FS0368 High Volume Model : TE-5009X

Calibrator ID: BKK_FS0624 High Volume S/N : 4165

Calibrator Model : TE-5028A Calibrator Slope : 1.64942

Calibrator S/N : 2584 Calibrator Intercept : -0.02902

Test No.	Delta H ₂ O (inch)	Q _{std} (m ³ /min)	I : Chart (CFM)	Linear Regression
1	3.4	1.1344	40	Slope : 39.3324 Intercept : -4.8701 Correlation Coefficient : 0.9970
2	4.4	1.2865	46	
3	5.4	1.4221	50	
4	6.4	1.5456	56	
5	7.2	1.6376	60	



Winyou B.

Calibrated by

(Mr. Winyou Boontanai)
Field Scientist(1)

Wichan Chuanat

Approved by :


(Mr. Wichan Choonharat)
Manager



High Volume Air Sampler Calibration Worksheet

Project Site :	Bangpoo Environmental Complex Co., Ltd.	Barometric Pressure (mm Hg) :	758
Calibrate Location :	ชุมชนสิ่งแวดล้อม 2 (Community area 4)	Temperature (°C) :	31
Calibrate Date :	15-Feb-22	High Volume ID :	BKK_FS0367
Calibration Sheet No.:	C-150222-BKK_FS0367	High Volume Model :	TE-5009X
Calibrator ID:	BKK_FS0624	High Volume S/N :	4162
Calibrator Model :	TE-5028A	Calibrator Slope :	1.64942
Calibrator S/N :	2584	Calibrator Intercept :	-0.02902

Test No.	Delta H ₂ O (inch)	Q _{std} (m ³ /min)	I : Chart (CFM)	Linear Regression
1	3.2	1.1014	40	Slope : 42.3923 Intercept : -6.5203 Correlation Coefficient : 0.9992
2	4.0	1.2280	46	
3	4.8	1.3424	50	
4	5.8	1.4727	56	
5	6.6	1.5691	60	



Winyou B.

Calibrated by

(Mr. Winyou Boontanai)
Field Scientist(1)

Wichan Chuanat

Approved by :


(Mr. Wichan Choonharat)
Manager



MULTIPOINT CALIBRATION REPORT

Calibration Date	4-Jan-22	Equipment Name	NOx Analyzer
Manufacturer	HORIBA	Model	APNA-370
Serial No.	XLTWBRSJ	Equipment ID	BKK_FS1092
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947	Cylinder No.	LL36633
Std. Gas Concentration (PPM)	51.33	Certified By	Airgas Inc.
Cylinder Pressure (psi)	1200	Expired Date	18-Mar-22
Certified Date	18-Mar-14		

CALIBRATION RESULTS							
Point	Ideal	Actual NO	Error NO	%Error NO	Actual NOx	Error NOx	%Error NOx
ZERO	0.00	0.10	0.10	0.10	0.10	0.10	0.10
1	100.00	99.70	-0.30	-0.30	101.00	1.00	1.00
2	200.00	198.10	-1.90	-0.95	201.00	1.00	0.50
3	300.00	299.10	-0.90	-0.30	301.40	1.40	0.47
4	400.00	398.20	-1.80	-0.45	402.80	2.80	0.70
AVERAGE (%)				-0.38			0.55



Calibrated By

(Mr. Jirawut Sakarn)
Field Environmental Scientist (3)

Approved By


(Mr. Sarayuth Jitiranont)
Assistant General Manager



MULTIPOINT CALIBRATION REPORT

Calibration Date	4-Jan-22	Equipment Name	NOx Analyzer
Manufacturer	HORIBA	Model	APNA-370
Serial No.	PX13CWA0	Equipment ID	BKK_FS1088
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	51.33	Cylinder No.	LL36633
Cylinder Pressure (psi)	1200	Certified By	Airgas Inc.
Certified Date	18-Mar-14	Expired Date	18-Mar-22

CALIBRATION RESULTS							
Point	Ideal	Actual NO	Error NO	%Error NO	Actual NOx	Error NOx	%Error NOx
ZERO	0.00	0.10	0.10	0.10	0.10	0.10	0.10
1	100.00	99.00	-1.00	-1.00	100.30	0.30	0.30
2	200.00	198.30	-1.70	-0.85	198.10	-1.90	-0.95
3	300.00	298.40	-1.60	-0.53	301.70	1.70	0.57
4	400.00	396.70	-3.30	-0.83	398.30	-1.70	-0.42
AVERAGE (%)		-0.62				-0.08	



Calibrated By

[Signature]

(Mr.Jirawut Sakarn)
Field Environmental Scientist (3)

Approved By

[Signature]


(Mr.Sarayuth Jitiranont)
Assistant General Manager



MULTIPOINT CALIBRATION REPORT

Calibration Date	4-Jan-22	Equipment Name	NOx Analyzer
Manufacturer	HORIBA	Model	APNA-370
Serial No.	PHD13MC7	Equipment ID	BKK_FS1072
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	51.33	Cylinder No.	LL36633
Cylinder Pressure (psi)	1200	Certified By	Airgas Inc.
Certified Date	18-Mar-14	Expired Date	18-Mar-22

CALIBRATION RESULTS							
Point	Ideal	Actual NO	Error NO	%Error NO	Actual NOx	Error NOx	%Error NOx
ZERO	0.00	0.10	0.10	0.10	0.10	0.10	0.10
1	100.00	98.80	-1.20	-1.20	100.50	0.50	0.50
2	200.00	201.40	1.40	0.70	200.70	0.70	0.35
3	300.00	298.30	-1.70	-0.57	302.10	2.10	0.70
4	400.00	396.90	-3.10	-0.78	401.30	1.30	0.33
AVERAGE (%)		-0.35				0.40	



Calibrated By

[Signature]

(Mr.Jirawut Sakarn)
Field Environmental Scientist (3)

Approved By

[Signature]


(Mr.Sarayuth Jitiranont)
Assistant General Manager



MULTIPOINT CALIBRATION REPORT

Calibration Date	4-Jan-22	Equipment Name	NOx Analyzer
Manufacturer	HORIBA	Model	APNA-370
Serial No.	WPYQJMW	Equipment ID	BKK_FS0782
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	51.33	Cylinder No.	LL36633
Cylinder Pressure (psi)	1200	Certified By	Airgas Inc.
Certified Date	18-Mar-14	Expired Date	18-Mar-22

Point	CALIBRATION RESULTS						
	Ideal	Actual NO	Error NO	%Error NO	Actual NOx	Error NOx	%Error NOx
ZERO	0.00	0.10	0.10	0.10	0.10	0.10	0.10
1	100.00	99.60	-0.40	-0.40	100.60	0.60	0.60
2	200.00	199.70	-0.30	-0.15	201.30	1.30	0.65
3	300.00	298.50	-1.50	-0.50	298.30	-1.70	-0.57
4	400.00	398.70	-1.30	-0.33	399.00	-1.00	-0.25
AVERAGE (%)				-0.26			0.11



Calibrated By

(Mr. Jirawut Sakarn)
Field Environmental Scientist (3)

Approved By


(Mr. Sarayuth Jitranont)
Assistant General Manager



MULTIPOINT CALIBRATION REPORT

Calibration Date	4-Jan-22	Equipment Name	SO2 Analyzer
Manufacturer	HORIBA	Model	APSA-370
Serial No.	6BVW9P1K	Equipment ID	BKK_FS1091
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	50.87	Cylinder No.	LL36633
Cylinder Pressure (psi)	1200	Certified By	Airgas Inc.
Certified Date	18-Mar-14	Expired Date	18-Mar-22

Point	CALIBRATION RESULTS			
	Ideal	Actual	Error	%Error
ZERO	0.00	0.10	0.10	0.10
1	100.00	98.50	-1.50	-1.50
2	200.00	198.30	-1.70	-0.85
3	300.00	297.90	-2.10	-0.70
4	400.00	398.50	-1.50	-0.38
AVERAGE (%)				-0.67



Calibrated By

(Mr. Jirawut Sakarn)
Field Environmental Scientist (3)

Approved By


(Mr. Sarayuth Jitranont)
Assistant General Manager



MULTIPOINT CALIBRATION REPORT

Calibration Date	4-Jan-22	Equipment Name	SO2 Analyzer
Manufacturer	HORIBA	Model	APSA-370
Serial No.	XHV1S9F	Equipment ID	BKK_FS1087
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	50.87	Cylinder No.	LL36633
Cylinder Pressure (psi)	1200	Certified By	Airgas Inc.
Certified Date	18-Mar-14	Expired Date	18-Mar-22

Point	CALIBRATION RESULTS			
	Ideal	Actual	Error	%Error
ZERO	0.00	0.10	0.10	0.10
1	100.00	99.10	-0.90	-0.90
2	200.00	198.30	-1.70	-0.85
3	300.00	298.70	-1.30	-0.43
4	400.00	398.30	-1.70	-0.42
AVERAGE (%)				-0.50



Calibrated By

(Mr. Jirawut Sakarn)
Field Environmental Scientist (3)

Approved By


(Mr. Sarayuth Jitranont)
Assistant General Manager



MULTIPOINT CALIBRATION REPORT

Calibration Date	4-Jan-22	Equipment Name	SO2 Analyzer
Manufacturer	HORIBA	Model	APSA-370
Serial No.	ROH40S60	Equipment ID	BKK_FS1071
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	50.87	Cylinder No.	LL36633
Cylinder Pressure (psi)	1200	Certified By	Airgas Inc.
Certified Date	18-Mar-14	Expired Date	18-Mar-22

Point	CALIBRATION RESULTS			
	Ideal	Actual	Error	%Error
ZERO	0.00	0.10	0.10	0.10
1	100.00	101.30	1.30	1.30
2	200.00	197.30	-2.70	-1.35
3	300.00	298.50	-1.50	-0.50
4	400.00	398.00	-2.00	-0.50
AVERAGE (%)				-0.19



Calibrated By

(Mr. Jirawut Sakarn)
Field Environmental Scientist (3)

Approved By


(Mr. Sarayuth Jitranont)
Assistant General Manager



MULTIPOINT CALIBRATION REPORT

Calibration Date	4-Jan-22	Equipment Name	SO2 Analyzer
Manufacturer	HORIBA	Model	APSA-370
Serial No.	Y53SNSFB	Equipment ID	BKK_FS0781
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	50.87	Cylinder No.	LL36633
Cylinder Pressure (psi)	1200	Certified By	Airgas Inc.
Certified Date	18-Mar-14	Expired Date	18-Mar-22

CALIBRATION RESULTS				
Point	Ideal	Actual	Error	%Error
ZERO	0.00	0.10	0.10	0.10
1	100.00	99.70	-0.30	-0.30
2	200.00	199.20	-0.80	-0.40
3	300.00	298.50	-1.50	-0.50
4	400.00	397.40	-2.60	-0.65
AVERAGE (%)				-0.35



Calibrated By

(Mr. Jirawut Sakam)
Field Environmental Scientist (3)

Approved By

(Mr. Sarayuth Jitranont)
Assistant General Manager



63/14-15,67/35-36, Soi Petchkasem 7,7/1, Petchkasem Rd,
Wathapra, Bangkokkhai,Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranatee.com

CERTIFICATE OF CALIBRATION

Certificate No: WS-02102021

Page 1 of 2 pages

Measurement Item : Cup anemometer with data logger.

Manufacturer : Data logger: Novallux.

Model/Type : Cup anemometer: Novallux.

Serial Number : Data logger: A5380

ID No : Cup anemometer: WS-02P

Customer : Data logger: BKK_FS0920

Test Conditions : Cup anemometer: -

Calibration Procedure : ALS laboratory group (Thailand) co., ltd.

Traceability : 104 Phatthanaheh 40, Phatthanaheh Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand.

Test Conditions : Wind tunnel cross test section area : 900 cm²

Test Conditions : Anemometer frontal area : 100 cm²

Test Conditions : Diameter of mounting pipe : mm

Test Conditions : Blockage ratio of test object : 0.111 [-]

Test Conditions : Air temperature : 23.3 ±0.8 °C

Test Conditions : Air pressure : 1007.5 ±0.4 hPa

Test Conditions : Relative air humidity : 45.9 ±3.6 %RH

Calibration Procedure : Calibration was carried out base on:

Calibration Procedure : ISO 61400-12-1 ED.1: 2005-Power Performance Measurements of Electricity Producing Wind Turbines;

Calibration Procedure : MEASNET Anemometer Calibration Procedure - Version 2: 2009;

Traceability : This calibration documents the traceable to national standard, which realize the unit of measurements according to the international system of units (SI) through National Institute of Metrology Thailand (NIMT).

Measurement Date : Oct 08, 2021.

Issued Date : Oct 11, 2021.

Calibrated by : Mr. Screwit Thechad

Calibrated by : Miss Orathai Wivattitaya

Approved Signatory: Mr. Parinya Booncharoen

Approved Signatory: Technical Support and Calibration Manager

THIS CERTIFICATE MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.

Continuation of Certificate of Calibration Number

Certificate No: WS-02102021
Page 2 of 2 Pages

Result of calibration: ☒ Without adjustment ☐ With adjustment
Calibration in the range of 1 – 16 m/s at a calibration interval of 1 m/s.
The results of calibration and associated measurement uncertainties are reported in the table below.

V ₅₀ Reading m/s	V ₅₀ Reading m/s	Error (m/s)	Uncertainty (%)
2.040	2.0	0.0	2.7
4.090	4.0	-0.1	1.3
5.97	6.0	0.0	0.99
8.02	8.0	0.0	0.98
10.01	10.1	0.1	1.2
12.01	12.1	0.1	0.65
14.00	14.2	0.2	0.69
15.99	16.2	0.2	1.1
14.98	15.2	0.2	0.72
13.03	13.1	0.1	0.94
11.02	11.0	0.0	1.1
9.01	9.0	0.0	0.88
7.01	7.0	0.0	1.1
5.083	5.1	0.0	0.98
2.984	3.0	0.0	1.7
1.007	0.8	-0.2	4.6

UUC*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

Appendix 1: Instrumentations

NO	Sensor	Manufacturer	Model/Type	Calibration Date	Certificate Report Number	Range
1	Pitot static	TESSTO INC.	D6352145	Aug 07, 2021	MW-0034-21	5 – 30 m/s
2	Precision Differential Pressure Meter	Zoglab	DFM2500	Aug 07, 2021	MW-0034-21	5 – 30 m/s
3	Air velocity transducer (hot wire)	TSI INC.	84455-12	Aug 08, 2021	MW-0035-21	0 – 5 m/s
4	Temperature	Zoglab	DSR-THP	March 30, 2021	GL-027-64	-30 – 70 °C
5	Relative humidity	Zoglab	DSR-THP	March 30, 2021	RH-G3032021	0 – 100 %RH
6	Atmospheric pressure	Zoglab	DSR-THP	March 30, 2021	BP-G1032021	600 – 1100 hPa
7	Wind tunnel	ESSOM	MP3300	-	-	0 – 50 Hz

End of certificate of calibration



Performed by
☒ Mr. Sawit Thachalad
☐ Miss Orathai Wiatvitaya



Mr. Parnya Booncharoen
Technical Support
and Calibration Manager

Mr. Parnya Booncharoen
Technical Support
and Calibration Manager

CERTIFICATE OF CALIBRATION

Certificate No: WD-02102021
Page 1 of 2 pages

Measurement Item

: Wind direction sensor with data logger.

Manufacturer

: Data logger: Novolynx.
: Wind direction sensor: Novolynx.

Model/Type

: Data logger: 200-WS-26LB
: Wind direction sensor: WS-02F

Serial Number

: Data logger: A5380
: Wind direction sensor: -

ID No

: Data logger: BKH_FSO920
: Wind direction sensor: -

Customer

: ALS laboratory group (Thailand) Co.,Ltd.
104 Phatthanakan 40, Phatthanakan Rd,Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10260
Thailand.

Environmental Condition:

The measurement was carried out in an ambient temperature of (23±3)°C, and relative humidity of (40±10)%.

Measurement Method:

The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and line laser is used for axis control. The measurement were taken at 45° intervals in clockwise and counterclockwise directions.

Note: The UUC was warned up for 1 hour prior to the calibration being performed

Traceability:

The measurement results are traceable to the international system of units (SI) through Certificate No: CC563-07-0045, Certificate No: KWS64/0026.

Measurement Date : Oct 08, 2021.

Issued Date : Oct 11, 2021.

Continuation of Certificate of Calibration Number

Certificate No: WD-02102021
Pages 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment.
Calibration in the range of 0 - 360 ° at a calibration interval of 45°.
The results of calibration and associated measurement uncertainties are reported in table below.

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC* Reading (°)	Error (°)	Uncertainty ±(°)
1	Clockwise	0/360	360	359	-1	3.0
2		45	45	42	-3	3.0
3		90	90	87	-3	3.0
4		135	135	132	-3	3.0
5		180	180	179	-1	3.0
6		225	225	226	1	3.0
7		270	270	272	2	3.0
8		315	315	318	3	3.0
9	Counter Clockwise	0/360	360	359	-1	3.0
10		45	45	42	-3	3.0
11		90	90	87	-3	3.0
12		135	135	132	-3	3.0
13		180	180	179	-1	3.0
14		225	225	226	1	3.0
15		270	270	272	2	3.0
16		315	315	318	3	3.0

UUC*: Unit Under Calibration The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

End of Certificate of Calibration



CERTIFICATE OF CALIBRATION

Certificate No.: WD-01112021
Page 1 of 2 pages

Measurement Item : Wind direction sensor with data logger.

Manufacturer : Data logger: Novallux.
: Wind direction sensor: Novallux.

Model/Type : Data logger: 200-WS-25LB
: Wind direction sensor: WS-02P

Serial Number : Data logger: A5377
: Wind direction sensor: -

ID No : Data logger: BKK_PS0017
: Wind direction sensor: -

Customer : A.L.S laboratory group (Thailand) Co.,Ltd.
104 Phathanakan 40, Phathanakan Rd,Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250
Thailand.

Environmental Condition:

The measurement was carried out in an ambient temperature of (23±3) °C, and relative humidity of (40±10) %.

Measurement Method:

The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and line laser is used for axis control. The measurement were taken at 45° intervals in clockwise and counterclockwise directions.

Note: The UUC was warmed up for 1 hour prior to the calibration being performed

Traceability:

The measurement results are traceable to the international system of units (SI) through Certificate No.: Q21086014, Certificate No.: KWS64/0025.

Measurement Date : Nov 01, 2021.
Issued Date : Nov 01, 2021.

Performed by
☒ Mr. Sorawit Thachalad
☐ Miss Orathai Wiwatwitaya

Approved Signatory:

Mr. Parinya Booncharoen.
Calibration Department Manager



Result of calibration: ☐ Without adjustment ☒ With adjustment.
Calibration in the range of 0 - 360 ° at a calibration interval of 45°.
The results of calibration and associated measurement uncertainties are reported in table below.

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC* Reading (°)	Error (°)	Uncertainty ±(°)
1	Clockwise	0/360	360	359	-1	3.0
2		45	45	41	-4	3.0
3		90	90	87	-3	3.0
4		135	135	135	0	3.0
5		180	180	183	3	3.0
6		225	225	229	4	3.0
7		270	270	274	4	3.0
8		315	315	320	5	3.0
9	Counter Clockwise	0/360	360	359	-1	3.0
10		45	45	41	-4	3.0
11		90	90	87	-3	3.0
12		135	135	135	0	3.0
13		180	180	183	3	3.0
14		225	225	229	4	3.0
15		270	270	274	4	3.0
16		315	315	320	5	3.0

UUC*: Unit Under Calibration The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

End of Certificate of Calibration



CERTIFICATE OF CALIBRATION

Certificate No: WS-01112021
Page 1 of 2 pages

Measurement Item : Cup anemometer with data logger.
Manufacturer : Data logger: Novelmx
: Cup anemometer: Novelmx
Model/Type : Data logger: 200-WS-25LB
: Cup anemometer: WS-02F
Serial Number : Data logger: A5377
: Cup anemometer: -
ID No : Data logger: BKH_FS0917
: Cup anemometer: -
Customer : ALS laboratory group (Thailand) co. ltd.
: 104 Phathanakan 40, Phathanakan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand.
Test Conditions : Wind tunnel cross test section area 900 cm²
: Anemometer frontal area 100 cm²
: Diameter of mounting pipe - mm
: Blockage ratio of test object 0.111 [-]
Test Conditions : Air temperature 25.4 ±0.8 °C
: Air pressure 1015.1 ±0.4 hPa
: Relative air humidity 47.9 ±3.5 %RH

Calibration Procedure

Calibration was carried out base on:
ISO 61400-12-1 6D.1: 2005-Power Performance Measurements of Electricity Producing Wind Turbines;
MEXASNET Anemometer Calibration Procedure - Version 2: 2009;

Traceability

This calibration documents the traceable to national standard, which realize the unit of measurements according to the international system of units (SI) through National Institute of Metrology Thailand (NIMT).

Measurement Date

: Nov 01, 2021.
: Nov 01, 2021.

Calibrated by

☒ Mr. Soravit Thachalead
☐ Miss Orathai Witwitwattayee

Approved Signatory:



(Signature)

Mr. Panyee Boonchoren
Calibration Department Manager

Continuation of Certificate of Calibration Number

Certificate No: WS-01112021
Page 2 of 2 Pages

Result of calibration: ☒ Without adjustment ☐ With adjustment
Calibration in the range of 1 - 16 m/s at a calibration interval of 1 m/s.

The results of calibration and associated measurement uncertainties are reported in the table below.

V _{ISO} Reading m/s	V _{ISO} Reading m/s	Error (m/s)	Uncertainty (%)
1.994	1.9	-0.1	2.5
4.002	4.0	0.0	1.2
6.00	6.0	0.0	0.95
8.02	8.0	0.0	0.73
10.01	10.1	0.1	0.63
12.00	12.1	0.1	0.74
13.99	14.1	0.1	0.76
16.01	16.3	0.3	0.80
16.01	16.3	0.3	0.64
13.00	13.1	0.1	0.45
11.01	11.1	0.1	0.57
9.02	9.0	0.0	0.64
7.02	7.0	0.0	0.98
4.992	5.0	0.0	1.2
2.980	2.9	-0.1	1.5
0.996	0.9	-0.1	4.5

UUC*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

Appendix 1: Instrumentations

NO	Sensor	Manufacturer	Model/Type	Calibration Date	Certificate Report Number	Range
1	Pilot static	TGSTO INC.	DG352145	Aug 07, 2021	MW-0034-21	5 - 30 m/s
2	Precision Differential Pressure Meter	Zoglab	DPH2600	Aug 07, 2021	MW-0034-21	5 - 30 m/s
3	Air velocity transducer (hot wire)	TSI INC.	8456-12	Aug 08, 2021	MW-0035-21	0 - 5 m/s
4	Temperature	Zoglab	D8H-Thp	March 30, 2021	CL-027-44	-30 - 70°C
5	Relative humidity	Zoglab	D8H-Thp	March 30, 2021	PH-03032021	0 - 100 %RH
6	Atmospheric pressure	Zoglab	D8H-Thp	March 30, 2021	BP-01032021	500 - 1100 hPa
7	Wind tunnel	CSOIM	MP3300	-	-	0 - 50 Hz

End of certificate of calibration



CERTIFICATE OF CALIBRATION

Certificate No: WS-08082021
Page 1 of 2 pages

Measurement Item : Cup anemometer with data logger.

Manufacturer : Data logger: Novelynx

: Cup anemometer: Novelynx

Model/Type : Data logger: 200-WS-29LB

: Cup anemometer: WS-02P.

Serial Number : Data logger: A5378

: Cup anemometer: -

ID No : Data logger: BKK_PSO918

: Cup anemometer: -

Customer : ALS laboratory group (Thailand) co., Ltd.

: 104 Phattanasak 40, Phattanasak Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand.

Test Conditions : Wind tunnel cross test section area 900 cm²
: Anemometer frontal area 100 cm²
: Diameter of mounting pipe - mm
: Blockage ratio of test object 0.111 %

Test Conditions : Air temperature 22.6 ±0.8 °C
: Air pressure 1009.6 ±0.4 hPa
: Relative air humidity 50.4 ±3.5 %RH

Calibration Procedure

Calibration was carried out base on:
ISO 61400-12-1 ED.1: 2005-Power Performance Measurements of Electricity Producing Wind Turbines;
MEASNET Anemometer Calibration Procedure - Version 2: 2009;

Traceability

This calibration documents the traceable to national standard, which realize the unit of measurements according to the international system of units (SI) through National Institute of Metrology Thailand (NIMT).

Measurement Date

: Aug 30, 2021.

Issued Date

: Aug 31, 2021.

Calibrated by

☒ Mr. Sorawit Thachalad
☐ Miss Orattai Witsavitaya

Approved Signatory:



[Signature]

Mr. Parinya Booncharoen
Technical Support
and Calibration Manager

Continuation of Certificate of Calibration Number

Certificate No: WS-08082021
Page 2 of 2 Pages

Result of calibration: ☒ Without adjustment ☐ With adjustment

Calibration in the range of 1 – 16 m/s at a calibration interval of 1 m/s.

The results of calibration and associated measurement uncertainties are reported in the table below.

V_{rip} Reading m/s	V_{user} Reading m/s	Error (m/s)	Uncertainty (%)
2.021	1.8	-0.2	2.9
4.074	3.9	-0.2	1.6
6.98	6.0	0.0	0.99
8.03	8.0	0.0	0.84
10.03	10.2	0.2	0.66
11.99	12.3	0.3	0.96
13.98	14.4	0.4	0.47
16.00	16.6	0.6	0.48
18.02	18.6	0.6	0.69
19.98	19.4	0.4	0.67
20.99	21.2	0.2	0.69
8.97	9.0	0.0	0.97
7.01	7.0	0.0	0.90
5.085	5.0	-0.1	0.96
2.970	3.0	0.0	1.7
1.019	0.7	-0.3	5.4

UUC*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor $k=2$ providing a level of confidence of approximately 95%.

Appendix 1: Instrumentations

NO	Sensor	Manufacturer	Model/Type	Calibration Date	Certificate Report Number	Range
1	Phot static	TESST INC.	CS352145	Aug 07, 2021	KW-0034-21	5 - 30 m/s
2	Precision Differential Pressure Meter	Zoglab	DPM6000	Aug 07, 2021	KW-0034-21	5 - 30 m/s
3	Air velocity transducer (not wire)	TSI INC.	8455-12	Aug 08, 2021	KW-0035-21	0 - 5 m/s
4	Temperature	Zoglab	DS9-T4P	March 30, 2021	CL-027-64	-30 - 70°C
5	Relative humidity	Zoglab	DS9-T4P	March 30, 2021	RH-003032021	0 - 100 %RH
6	Atmospheric pressure	Zoglab	DS9-T4P	March 30, 2021	BP-01032021	500 - 1100 hPa
7	Wind tunnel	ESSCOM	MP330D	-	-	0 - 60 Hz

End of certificate of calibration



Performed by

☒ Mr. Sorawit Thachalad

☐ Miss Orathai Wiwatwittaya



Approved Signatory:  Mr. Parinya Booncharoen,
Technical Support
and Calibration Manager

CERTIFICATE OF CALIBRATION

Certificate No.: WD-07082021

Page 1 of 2 pages

Measurement Item	: Wind direction sensor with data logger.
------------------	---

Manufacturer : Data logger: Novalynx.

: Wind direction sensor; Novalynx.

Model/Type : Data logger: 200-WS-25LB,

: Wind direction sensor: WS-02F.

Serial Number : Data logger: A5378.

: Wind direction sensor; -

ID No : Data logger: BKK PSD918.

: Wind direction sensor: .

Customer : ALS laboratory group (Thailand) Co.Ltd.

104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250
Thailand.

Environmental Condition:

The measurement was carried out in an ambient temperature of $(23 \pm 3)^{\circ}\text{C}$. and relative humidity of $(40 \pm 10)\%$.

Measurement Methods:

The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and line laser is used for axis control. The measurement were taken at 45° intervals in clockwise and counterclockwise directions.

Note: The UUC was warmed up for 1 hour prior to the calibration being performed.

Traceability:

The measurement results are traceable to the international system of units (SI) through Certificate No: CC563-07-0045, Certificate No: KWS3/0044.

Measurement Date : Aug 30, 2021.

Issued Date : Aug 31, 2021.

THIS CERTIFICATE REPORT MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.

Continuation of Certificate of Calibration Number

Certificate No: WD-07082021
Pages 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment.
Calibration in the range of 0 - 360 ° at a calibration interval of 45°.

The results of calibration and associated measurement uncertainties are reported in table below.

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC* Reading (°)	Error (°)	Uncertainty ±(°)
1	Clockwise	0/360	360	369	-1	3.0
2		45	45	43	-2	3.0
3		90	90	87	-3	3.0
4		135	135	132	-3	3.0
5		180	180	180	0	3.0
6		225	225	228	3	3.0
7		270	270	274	4	3.0
8		315	315	319	4	3.0
9	Counter Clockwise	0/360	360	359	-1	3.0
10		45	45	43	-2	3.0
11		90	90	87	-3	3.0
12		135	135	132	-3	3.0
13		180	180	180	0	3.0
14		225	225	228	3	3.0
15		270	270	274	4	3.0
16		315	315	319	4	3.0

UUC*: Unit Under Calibration The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

End of Certificate of Calibration



CERTIFICATE OF CALIBRATION

Certificate No: WS-09082021
Page 1 of 2 pages

Measurement Item

: Cup anemometer with data logger.

Manufacturer

: Data logger: Novolynx.

: Cup anemometer: Novolynx.

Model/Type

: Data logger: 200-WS-28LB.

: Cup anemometer: WS-02F.

Serial Number

: Data logger: A6379.

: Cup anemometer: -.

ID No

: Data logger: BKK_FSO919.

: Cup anemometer: -.

Customer

: ALS laboratory group (Thailand) co., ltd.

: 104 Phatthanahan 40, Phatthanahan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand.

Test Conditions

: Wind tunnel cross test section area

900 cm²

: Anemometer frontal area

100 cm²

: Diameter of mounting pipe

- mm

: Blockage ratio of test object

0.111 [-]

Test Conditions

: Air temperature

22.7 ±0.8 °C

: Air pressure

1008.3 ±0.4 hPa

: Relative air humidity

45.8 ±3.6 %RH

Calibration Procedure

: Calibration was carried out base on:

IEC 61400-12-1 ED.1: 2005-Power Performance Measurements of Electricity Producing Wind Turbines;
MEASNET Anemometer Calibration Procedure - Version 2: 2009;

Traceability

This calibration documents the traceable to national standard, Which realize the unit of measurements according to the international system of units (SI) through National Institute of Metrology Thailand (NIMT).

Measurement Date

: Aug 30, 2021.

Issued Date

: Aug 31, 2021.

Calibrated by

☒ Mr. Soravit Thachalad

☐ Miss Orathai Wivattwittaya

Approved Signatory:

Mr. Panyee Booncharoen
Technical Support
and Calibration Manager



Continuation of Certificate of Calibration Number

Certificate No: WS-09082021
Page 2 of 2 Pages

Result of calibration: ☒ Without adjustment ☐ With adjustment
Calibration in the range of 1 – 16 m/s at a calibration interval of 1 m/s.

The results of calibration and associated measurement uncertainties are reported in the table below.

V _{ind} Reading m/s	V _{unc} Reading m/s	Error (m/s)	Uncertainty (%)
2.060	2.0	-0.1	2.7
4.064	4.1	0.0	1.3
5.99	6.0	0.0	0.98
8.00	8.0	0.0	0.74
10.01	10.2	0.2	0.66
11.98	12.3	0.3	0.48
14.03	14.4	0.4	0.47
15.98	16.6	0.6	0.39
16.00	16.4	0.4	0.51
12.98	13.4	0.4	0.55
10.99	11.2	0.2	0.53
8.97	9.0	0.0	0.65
6.97	7.0	0.0	0.84
5.062	5.1	0.0	0.93
2.972	3.1	0.1	1.6
1.021	0.9	-0.1	5.3

UUC*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

Appendix 1: Instrumentations

NO	Sensor	Manufacturer	Model/Type	Calibration Date	Certificate Report Number	Range
1	Pilot static	TGSI INC.	04352145	Aug 07, 2021	MW-0034-21	5 – 30 m/s
2	Precision Differential Pressure Meter	Zoglab	DPV2600	Aug 07, 2021	MW-0034-21	5 – 30 m/s
3	Air velocity transducer (hot wire)	TSI INC.	8456-12	Aug 06, 2021	MW-0035-21	0 – 6 m/s
4	Temperature	Zoglab	D98-T1P	March 30, 2021	CL-027-64	-50 – 70°C
5	Relative humidity	Zoglab	D98-T1P	March 30, 2021	PH-Q0302021	0 – 100 %RH
6	Atmospheric pressure	Zoglab	D98-T1P	March 30, 2021	BP-Q1032021	500 – 1100 hPa
7	Wind tunnel	CSOIM	MP3300	-	-	0 – 60 Hz

End of certificate of calibration



CERTIFICATE OF CALIBRATION

Certificate No: WD-08082021
Page 1 of 2 pages

Measurement Item : Wind direction sensor with data logger.

Manufacturer : Data logger: Novolynx

: Wind direction sensor: Novolynx

Model/Type : Data logger: 200-WS-25LB

: Wind direction sensor: WS-02F

Serial Number : Data logger: A5379

: Wind direction sensor: -

ID No : Data logger: BKK_FSD019

: Wind direction sensor: -

Customer : ALS laboratory group (Thailand) Co., Ltd.

104 Phatthaneken 40, Phatthaneken Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand.

Environmental Conditions

The measurement was carried out in an ambient temperature of (23±3)°C, and relative humidity of (40±10)%.

Measurement Method:

The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and line laser is used for axis control. The measurement were taken at 45° intervals in clockwise and counterclockwise directions.

Note: The UUC was warmed up for 1 hour prior to the calibration being performed

Traceability:

The measurement results are traceable to the international system of units (SI) through Certificate No: C05663-07-0045, Certificate No: KWS63/0044.

Measurement Date : Aug 30, 2021.

Issued Date : Aug 31, 2021.

Performed by

☒ Mr. Sawit Thachalad

☐ Miss Orathai Wiwatwitaya

Approved Signatory:.....

Mr. Parinya Booncharoen,

Technical Support

and Calibration Manager



Continuation of Certificate of Calibration Number

Certificate No: WD-08082021
Pages 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment.
Calibration in the range of 0 – 360 ° at a calibration interval of 45°.
The results of calibration and associated measurement uncertainties are reported in table below.

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC* Reading (°)	Error (°)	Uncertainty ±(°)
1	Clockwise	0/360	360	369	-1	3.0
2		45	45	41	-4	3.0
3		90	90	86	-4	3.0
4		135	135	133	-2	3.0
5		180	180	178	-2	3.0
6		225	225	227	2	3.0
7		270	270	273	3	3.0
8		315	315	318	3	3.0
9	Counter Clockwise	0/360	360	369	-1	3.0
10		45	45	41	-4	3.0
11		90	90	86	-4	3.0
12		135	135	133	-2	3.0
13		180	180	178	-2	3.0
14		225	225	227	2	3.0
15		270	270	273	3	3.0
16		315	315	318	3	3.0

UUC*: Unit Under Calibration The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

End of Certificate of Calibration



ELECTRICAL AND ELECTRONICS INSTITUTE
FOUNDATION FOR INDUSTRIAL DEVELOPMENT
975 Moo 4, Bangpoo Industrial Estate, Soi 8, Sukhumvit Road km 37,
Phraek Sa, Mueang Samut Prakan, Samut Prakan 10280
Tel: +66 2709 4860-8 Fax: +66 2324 0917-8



Certificate No.: 0146SV21
Operation No.: CP2021030033

Certificate of Calibration

Equipment: Sound Calibrator
Manufacturer: RION
Model/Type: NC-74
Serial No.: 34178117
ID No.: BKK_FS0630
Customer: ALS Laboratory Group (Thailand) Co.,Ltd.
Address: 104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Phatthanakan
Khet Suan Luang, Bangkok 10250 Thailand
Received Date: 19 March 2021
Calibrated Date: 24 March 2021
Issued Date: 25 March 2021
Calibrated by: Ms. Juntaporn Kunhakom



Approved by:
(Mr. Sittichai Swaksuriyawong)
Group Manager

The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor $k = 2.00$, providing a level of confidence of approximately 95%. This certificate may not be reproduced other than in full except with the prior written approval of the Electrical and Electronics Institute, Foundation for Industrial Development.

Certificate No.: 0146SV21

Calibration Report

Equipment: Sound Calibrator
Manufacturer: RION
Model/Type: NC-74
Serial No.: 34178117
ID No.: BKK_FS0630
Ambient Temperature: (23 ± 2) °C
Relative Humidity: (50 ± 15) %
Pressure: (101.3 ± 1.5) kPa

Method of Calibration :-
IEC 60942:2017

Condition of this result of calibration

1. Reference standards instrument :-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Standard microphone	4180	2661000	AA-1013-20	12 May 2021
2) Waveform Generator	33511B	MY52302264	0100RF20	17 June 2021
3) Audio Analyzing DMM	2015-P	000136E	E1U203927	16 November 2021
4) Pressure humidity and Temperature Transmitter	PTU301	F0640002	CL1-P200051	31 May 2021
			0305TE20	29 June 2021

2. This result of calibration was found accurate as shown on date and place of calibration only.

3. This certification is traceable to the international system of unit maintained at :-

- Reference standards instrument for Acoustic function
- National Institute of Metrology (Thailand)
Reference standards instrument for Electrical function
- Electrical and Electronics Institute; ONSC Accredited Calibration No.0119

Result of Calibration:-

1. Function : Sound pressure level

Normal Frequency (Hz)	Specified Sound Pressure level (dB)	Measured value (dB)	Deviated value ^[1] (dB)	Acceptance limit ^[3] (dB)
1000	94	94.10	0.10	±0.25

2. Function : Frequency

Normal Sound Pressure level (dB)	Specified Frequency (Hz)	Measured value (Hz)	Deviated value ^[2] (%)	Acceptance limit ^[3] (%)
94	1000	1001.7	0.2	±0.7

Certificate No.: 0146SV21

Calibration Report

3. Function : Total distortion + noise

Normal Sound Pressure level (dB)	Normal Frequency (Hz)	Measured value ^[4] (%)	Acceptance limit ^[5] (%)
94	1000	3.0	2.5

Uncertainty of measurement

Function	Uncertainty	Maximum-permitted uncertainty of measurement
Sound pressure level	0.10 dB	0.15 dB
Frequency	0.10 %	0.20 %
Total distortion + noise	0.40 %	0.50 %

Note: [1] The deviated value is the absolute value of the difference between the measured value and the corresponding specified sound pressure level.

[2] The deviated value is the absolute value of the difference in percent between the measured value and the corresponding specified frequency.

[3] The acceptance limit is for the deviated value.

[4] The measured value is the total distortion + noise, measured over the frequency range from 20 Hz to 20 kHz.

[5] The acceptance limit is for the Measured value.

Remarks: 1. Using the 1/2-inch microphone adaptor NC-74-002.

2. Acceptance limit was IEC 60942:2017 Class 1.

-- End of Report --



451-451/1 Sirinthorn Rd, Bangbunmru, Bangkok 10700 THAILAND.
Tel.0-2435-8800 Fax.0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.com

Cert. No. : ACL21172
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00858526 / 175176 / 85721
ID No.: BKK_FS0116

Condition As Found : GOOD
Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAEANG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location :
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %
Received Date : 09 DECEMBER 2021
Calibration Date : 14-15 DECEMBER 2021
Date of Issue : 16 DECEMBER 2021

Calibrated by : Nathakorn Pisutpaisan

Approved by :
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Cert. No. : ACL21172
Job No. : VC65AC0033
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.
For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	8846A	1997025	EEL.BP. 06/0264	05-Feb-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KAI	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.
3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
- 3.2 Thailand Institute of Scientific and Technological Research (TISTR).

T. Petchurai

Continuation of Calibration Certificate

Cert. No. : ACL21172
Job No. : VC65AC0033
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

Continuation of Calibration Certificate

Cert. No. : ACL21172
Job No. : VC65AC0033
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
15.1

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	13.4
C - weight	19.7
Flat	25.1

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight
125	0.2	0.2	0.2
1000	0.1	0.0	0.0
8000	1.9	2.0	2.0
			Acceptance Limits
			± 1.5
			± 1.0
			±5.0

Continuation of Calibration Certificate

Cert. No. : ACL21172
Job No. : VC65AC0033
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight Acceptance Limits
63	0.0	-0.2	-0.1 ±2.0
125	0.0	0.0	-0.1 ±1.5
250	0.0	0.0	-0.1 ±1.5
500	0.0	0.0	-0.1 ±1.5
1000	0.0	0.0	0.0 ±1.0
2000	0.0	0.0	0.0 ±2.0
4000	0.1	0.0	0.0 ±3.0
8000	0.1	0.0	0.0 ±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	93.9	94.0	0.1	± 0.3

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.1	0.1	± 1.1
135.0	135.1	0.1	± 1.1
134.0	134.1	0.1	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.1	0.1	± 1.1
114.0	114.1	0.1	± 1.1
109.0	109.1	0.1	± 1.1
104.0	104.1	0.1	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	29.9	-0.1	± 1.1
29.0	28.9	-0.1	± 1.1
28.0	27.9	-0.1	± 1.1
27.0	26.9	-0.1	± 1.1
26.0	25.9	-0.1	± 1.1
25.0	24.9	-0.1	± 1.1

Continuation of Calibration Certificate

Cert. No. : ACL21172
Job No. : VC65AC0033
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	116.9	-0.1	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
	0.25	1	99.0	98.8	-0.2	1.5 ; -5.0
SEL	2	8	108.0	107.9	-0.1	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, Lepeak (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.4	0.0	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

Continuation of Calibration Certificate

Cert. No. : ACL21172
Job No. : VC65AC0033
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	89.7	0.2	±1.5
Negative one-half cycle	89.7	0.2	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00584983 / 175177 / 85722
ID No.: BKK_FS0926

Condition As Found : GOOD
Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAEANG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location :
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 06 JULY 2021
Calibration Date : 07-08 JULY 2021
Date of Issue : 13 JULY 2021

Calibrated by : Nathakorn Pisutpaisan

Approved by : 
(Thanakul Petchurai)

REVIEW BY	
APPROVED BY	
NEXT CAL DATE	7/9/21

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	33461A	MY53220116	EEL.BP. 04/0264	10-Feb-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KAI	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.
3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
- 3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL21065
Job No. : VC64AC0051
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

Continuation of Calibration Certificate

Cert. No. : ACL21065
Job No. : VC64AC0051
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
13.4

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	11.6
C - weight	18.4
Flat	23.8

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight Acceptance Limits
125	0.3	0.3	± 1.5
1000	-0.1	-0.1	± 1.0
8000	-0.9	-0.8	±5.0

Continuation of Calibration Certificate

Cert. No. : ACL21065
Job No. : VC64AC0051
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight Acceptance Limits
63	0.0	0.0	±2.0
125	0.0	0.0	±1.5
250	0.0	0.0	±1.5
500	0.0	0.1	±1.5
1000	0.0	0.0	±1.0
2000	0.0	0.1	±2.0
4000	0.0	0.0	±3.0
8000	0.0	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

Continuation of Calibration Certificate

Cert. No. : ACL21065
Job No. : VC64AC0051
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.0	0.0	± 1.1
26.0	26.0	0.0	± 1.1
25.0	25.0	0.0	± 1.1

T. P. P. P.

T. P. P. P.

Continuation of Calibration Certificate

Cert. No. : ACL21065
Job No. : YC64AC0051
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, T _b (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.1	0.1	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.0	-0.4	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

Continuation of Calibration Certificate

Cert. No. : ACL21065
Job No. : YC64AC0051
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle	0.0	±1.5
89.6	89.6		

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation providing a level of confidence of approximately 95 %

End of Calibration Certificate



451-451/1 Sirinthorn Rd, Bangplud, Bangkok 10700 THAILAND.
Tel.0-2435-8800 Fax.0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.com

Cert. No. : ACL21171
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00858525 / 158776 / 58777
ID No.: BKK_FS0115

Condition As Found : GOOD
Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

REVIEW BY	Nathakorn P.
APPROVED BY	[Signature]
NEXT CAL DATE	14/12/22

Location :
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %
Received Date : 09 DECEMBER 2021
Calibration Date : 14-15 DECEMBER 2021
Date of Issue : 16 DECEMBER 2021

Calibrated by : Nathakorn Pisutpaisan

Approved by : [Signature]
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Cert. No. : ACL21171
Job No. : VC65AC0033
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.
For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	8846A	1997025	EEL.BP. 06/0264	05-Feb-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KAI	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.
3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
- 3.2 Thailand Institute of Scientific and Technological Research (TISTR).

[Signature]

Continuation of Calibration Certificate

Cert. No. : ACL21171
Job No. : VC65AC0033
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

Continuation of Calibration Certificate

Cert. No. : ACL21171
Job No. : VC65AC0033
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
13.4

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	11.6
C - weight	17.8
Flat	23.2

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight
125	0.4	0.4	0.5
1000	-0.1	-0.1	-0.1
8000	-2.8	-2.7	-2.7
			Acceptance Limits
			± 1.5
			± 1.0
			±5.0

Continuation of Calibration Certificate

Cert. No. : ACL21171
Job No. : VC65AC0033
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)	
	Flat	A-weight
63	0.0	0.0
125	0.0	0.0
250	0.0	0.0
500	0.0	0.0
1000	0.0	0.0
2000	0.0	0.0
4000	0.0	0.0
8000	0.0	0.1

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

Continuation of Calibration Certificate

Cert. No. : ACL21171
Job No. : VC65AC0033
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	29.9	-0.1	± 1.1
29.0	28.9	-0.1	± 1.1
28.0	27.9	-0.1	± 1.1
27.0	26.9	-0.1	± 1.1
26.0	25.9	-0.1	± 1.1
25.0	24.9	-0.1	± 1.1

Continuation of Calibration Certificate

Cert. No. : ACL21171
Job No. : VC65AC0033
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.1	0.1	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
SEL	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.3	-0.1	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.2	-0.2	±2.0
Negative half cycle	135.4	135.2	-0.2	±2.0

11. Overload indication

Measured value (dB)	Acceptance Limits (dB)	
	Positive one-half cycle	Negative one-half cycle
89.6	89.5	-0.1

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate



Calibration Certificate

Equipment : SOUND CALIBRATOR
Manufacturer : RION
Model : NC-74
Serial No.: 34178118
ID No.: BKK_FS0631

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAEANG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location :
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 08 OCTOBER 2021
Calibration Date : 26 OCTOBER 2021
Date of Issue : 26 OCTOBER 2021

Calibrated by :

Approved by :

T. Petchurai
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Calibration Procedure : CP-AC-03

Calibration Method :

This equipment was calibrated by based on IEC-60942-2003 Standard.
The sound pressure level, frequency and total distortion of the sound calibrator was measured using the reference microphone.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	8846A	1997025	EEL.BP. 06/0264	05-Feb-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KAI	34560495	AA-3003-21	16-Feb-22
Audio Analyzer	AVR-3360A	V744B6069	EF-0010-21	10-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.
3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
- 3.2 Thailand Institute of Scientific and Technological Research (TISTR).

T. Petchurai

Continuation of Calibration Certificate

Cert. No. : ACC21021
Job No. : VC65AC0003
Pages : 3 of 3

Result of calibration :

1. Sound pressure level

Specified sound pressure level (dB)	Measured value (dB)	Deviated value (dB)	Uncertainty (dB)	Tolerance limit (dB)
94	94.15	0.15	0.14	0.40

2. Frequency

Specified Frequency (Hz)	Measured value (Hz)	Deviated value (%)	Uncertainty (%)	Tolerance limit (%)
1000	1001.6	0.2	0.1	1.0

3. Total distortion

Measured value (%)	Uncertainty (%)	Tolerance limit (%)
1.76	0.10	3.0

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$, or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate



451-451/1 Sirinthorn Rd, Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel.0-2435-8800 Fax.0-2433-1679 e-mail:cal-center@sithiphorn.com http://www.sithiphorn.com

Cert. No. : ACL21146
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00572566 / 170403 / 72904
ID No.: BKK_FS0875
Condition As Found : GOOD
Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location :
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %
Received Date : 01 NOVEMBER 2021
Calibration Date : 02-04 NOVEMBER 2021
Date of Issue : 05 NOVEMBER 2021

Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

REVIEW BY *Nathakorn P.*
APPROVED BY *[Signature]*
NEXT CAL DATE 2/11/22

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Continuation of Calibration Certificate

Calibration Procedure : CP-AC-01

Cert. No. : ACL21146
Job No. : VC65AC0011
Pages : 2 of 8

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	8846A	1997025	EEL.BP. 06/0264	05-Feb-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KAI	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
- 3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL21146
Job No. : VC65AC0011
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

Continuation of Calibration Certificate

Continuation of Calibration Certificate

Cert. No. : ACL21146
Job No. : VC65AC0011
Pages : 4 of 8

Cert. No. : ACL21146
Job No. : VC65AC0011
Pages : 5 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.9G)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.6

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	9.9
C - weight	16.3
Flat	21.9

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight
125	0.5	0.5	0.5
1000	0.0	0.0	0.0
8000	-1.8	-1.8	-1.7

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight
63	0.0	-0.1	-0.1
125	0.0	0.0	-0.1
250	0.0	0.0	-0.1
500	0.0	0.0	-0.1
1000	0.0	0.0	0.0
2000	0.0	0.0	0.0
4000	0.0	0.0	0.0
8000	0.0	0.1	0.1

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

Continuation of Calibration Certificate

Cert. No. : ACL21146
Job No. : VC65AC0011
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.1	0.1	± 1.1
27.0	27.0	0.0	± 1.1
26.0	26.0	0.0	± 1.1
25.0	25.0	0.0	± 1.1

Continuation of Calibration Certificate

Cert. No. : ACL21146
Job No. : VC65AC0011
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
SEL	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, Lepeak (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.0	-0.4	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

Continuation of Calibration Certificate

Cert. No. : ACL21146
Job No. : VC65AC0011
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle	0.1	±1.5
89.6	89.7		

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

451-451/1 Sirinthorn Rd., Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:center@sithiporn.com http://www.sithiporn.com

Cert. No. : ACL22035
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00296511 / 179112 / 87520
ID No.: BKK_FS0968

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location :

Ambient Temperature : (23.0 ± 3) °C

Pressure : (101.3 ± 3) kPa

Relative Humidity : (50.0 ± 20) %

Received Date : 05 JANUARY 2022

Calibration Date : 12-14 JANUARY 2022

Date of Issue : 17 JANUARY 2022

Calibrated by :

Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

REVIEW BY	<i>Nathakorn P.</i>
APPROVED BY	<i>T. Petchurai</i>
NEXT CAL. DATE	12/1/23

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Continuation of Calibration Certificate

Cert. No. : ACL22035
Job No. : VC65AC0041
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	34461A	MY60024273	1-15180725251-1	15-Sep-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KAI	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
- 3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL22035
Job No. : VC65AC0041
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

Continuation of Calibration Certificate

Cert. No. : ACL22035
Job No. : VC65AC0041
Pages : 4 of 8**Result of calibration :****1. Absolute sensitivity**

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.6

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	12.6
C - weight	19.0
Flat	24.1

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight Acceptance Limits
125	0.4	0.4	0.4 ± 1.5
1000	0.0	0.0	0.0 ± 1.0
8000	-1.2	-1.1	-1.1 ±5.0

Continuation of Calibration Certificate

Cert. No. : ACL22035
Job No. : VC65AC0041
Pages : 5 of 8**4. Electrical signal tests of frequency weightings**

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	-0.1	-0.1	±2.0
125	0.0	0.0	0.0	±1.5
250	0.0	0.0	0.0	±1.5
500	0.0	0.0	0.0	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.1	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

Continuation of Calibration Certificate

Cert. No. : ACL22035
Job No. : VC65AC0041
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	33.9	-0.1	± 1.1
30.0	29.9	-0.1	± 1.1
29.0	28.9	-0.1	± 1.1
28.0	28.0	0.0	± 1.1
27.0	26.9	-0.1	± 1.1
26.0	26.0	0.0	± 1.1
25.0	24.9	-0.1	± 1.1

Continuation of Calibration Certificate

Cert. No. : ACL22035
Job No. : VC65AC0041
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighing	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.1	0.1	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
SEL	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.1	0.1	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, Lcpeak (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	135.4	-1.0	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.2	-0.2	±2.0
Negative half cycle	135.4	135.2	-0.2	±2.0

Continuation of Calibration Certificate

Cert. No. : ACL22035
Job No. : VC6SAC0041
Pages : 8 of 8

11. Overload indication

Measured value (dB)	Deviated Value (dB)		Acceptance Limits (dB)
	Positive one-half cycle	Negative one-half cycle	
89.7	89.6	-0.1	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate



451-451/1 Sirinthorn Rd, Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel.0-2435-8800 Fax.0-2433-1679 e-mail:cal-center@sithiphorn.com http://www.sithiphorn.com

Cert. No. : ACL21094
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00672737 / 170325 / 73077
ID No.: BKK_FS0927

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAEANG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 26 AUGUST 2021
Calibration Date : 07-08 SEPTEMBER 2021
Date of Issue : 08 SEPTEMBER 2021

Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

REVIEW BY *Nathakorn P.*
APPROVED BY *Nich-Chan*
NEXT CAL. DATE *4/8/22*

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Continuation of Calibration Certificate

Cert. No. : ACL21094
Job No. : VC64AC0062
Pages : 2 of 8Cert. No. : ACL21094
Job No. : VC64AC0062
Pages : 3 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	8846A	1997025	EEL.BP. 06/0264	05-Feb-22
Digital Multimeter	33461A	MY53220116	EEL.BP. 04/0264	10-Feb-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KA1	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

Continuation of Calibration Certificate

Cert. No. : ACL21094
Job No. : VC64AC0062
Pages : 4 of 8**Result of calibration :****1. Absolute sensitivity**

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise**2.1 Normal test**

Measured Value (dB)
17.5

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	12.6
C - weight	18.6
Flat	24.5

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight Acceptance Limits
125	0.3	0.4	0.4 ± 1.5
1000	0.0	0.0	0.0 ± 1.0
8000	-0.3	-0.2	-0.2 ±5.0

Continuation of Calibration Certificate

Cert. No. : ACL21094
Job No. : VC64AC0062
Pages : 5 of 8**4. Electrical signal tests of frequency weightings**

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight Acceptance Limits
63	-0.1	-0.1	-0.1 ±2.0
125	-0.1	0.0	-0.1 ±1.5
250	0.0	0.0	-0.1 ±1.5
500	0.0	0.0	-0.1 ±1.5
1000	0.0	0.0	0.0 ±1.0
2000	0.0	0.0	0.0 ±2.0
4000	0.0	0.0	0.0 ±3.0
8000	0.0	0.1	0.1 ±5.0

5. Frequency and time weightings at 1 kHz**5.1 Frequency weightings at 1 kHz**

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

Continuation of Calibration Certificate

Cert. No. : ACL21094

Job No. : VC64AC0062

Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	132.9	-0.1	± 1.1
132.0	131.9	-0.1	± 1.1
131.0	130.9	-0.1	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.0	0.0	± 1.1
26.0	26.0	0.0	± 1.1
25.0	25.0	0.0	± 1.1

Continuation of Calibration Certificate

Cert. No. : ACL21094

Job No. : VC64AC0062

Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	± 1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.0	0.0	± 1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	± 1.0
	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
SEL	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	± 1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, Lopeak (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.2	-0.2	± 3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	± 2.0
Negative half cycle	135.4	135.1	-0.3	± 2.0

Continuation of Calibration Certificate

Cert. No. : ACL21094
Job No. : VC64AC0062
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle	0.0	±1.5
89.6	89.6		

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate



451-451/1 Sirinthorn Rd., Bangbunru, Bangkok 10700 THAILAND.
Tel: 0-2435-8800 Fax: 0-2433-1679 e-mail: cal-center@sithiphorn.com http://www.sithiphorn.com

Cert. No. : ACL22043
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00873053 / 171587 / 73329
ID No.: BKK_FS0930

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location :
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 05 JANUARY 2022
Calibration Date : 12-14 JANUARY 2022
Date of Issue : 17 JANUARY 2022

Calibrated by : Nathakorn Pisurpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

REVIEW BY	<i>Nathakorn P.</i>
APPROVED BY	<i>T. Petchurai</i>
NEXT CAL. DATE	12/1/23

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Continuation of Calibration Certificate

Calibration Procedure : CP-AC-01 **Cert. No. :** ACL22043
Job No. : VC65AC0041 **Pages :** 2 of 8

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	34461A	MY60024273	1-15180725251-1	15-Sep-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KAI	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
- 3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL22043
Job No. : VC65AC0041
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

Continuation of Calibration Certificate

Cert. No. : ACL22043
Job No. : VC65AC0041
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.6

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	13.1
C - weight	19.4
Flat	24.7

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight
125	0.2	0.2	0.2
1000	0.0	0.0	0.0
8000	-0.3	-0.3	-0.2
Acceptance Limits			
± 1.5			
± 1.0			
±5.0			

Continuation of Calibration Certificate

Cert. No. : ACL22043
Job No. : VC65AC0041
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight
63	0.0	-0.1	0.0
125	0.0	0.0	0.0
250	0.0	0.0	0.0
500	0.0	0.0	-0.1
1000	0.0	0.0	0.0
2000	0.0	0.0	0.0
4000	0.0	0.0	0.0
8000	0.0	0.1	0.1
Acceptance Limits			
±2.0			
±1.5			
±1.5			
±1.5			
±1.0			
±2.0			
±3.0			
±5.0			

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

Continuation of Calibration Certificate

Cert. No. : ACL22043
Job No. : VC65AC0041
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.1	0.1	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.1	0.1	± 1.1
69.0	69.1	0.1	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.1	0.1	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.0	0.0	± 1.1
26.0	25.9	-0.1	± 1.1
25.0	25.0	0.0	± 1.1

Continuation of Calibration Certificate

Cert. No. : ACL22043
Job No. : VC65AC0041
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	93.9	-0.1	± 1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.0	0.0	± 1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	± 1.0
	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
SEL	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	± 1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.3	-0.1	± 3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	± 2.0
Negative half cycle	135.4	135.2	-0.2	± 2.0

Continuation of Calibration Certificate

Cert. No. : ACL22043
Job No. : VC65AC0041
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	89.6	0.0	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

— End of Calibration Certificate —



451-451/1 Sirinthorn Rd., Bangumru, Bangkok 10700 THAILAND.
Tel: 0-2435-8800 Fax: 0-2433-1679 e-mail: cal-center@sithiporn.com http://www.sithiporn.com

Cert. No. : ACL22036
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42 / Microphone UC-52 / Preamplifier NH-24
Serial No.: 00296512 / 179113 / 87521
ID No.: BKK_FS0969

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -

Ambient Temperature : (23.0 ± 3) °C

Pressure : (101.3 ± 3) kPa

Relative Humidity : (50.0 ± 20) %

Received Date : 05 JANUARY 2022

Calibration Date : 12-14 JANUARY 2022

Date of Issue : 17 JANUARY 2022

Calibrated by :

Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

REVIEW BY	<i>Nathakorn P</i>
APPROVED BY	<i>T. Petchurai</i>
NEXT CAL. DATE	12/1/23

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Continuation of Calibration Certificate

Cert. No. : ACL22036
Job No. : VC65AC0041
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	34461A	MY60024273	1-15180725251-1	15-Sep-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KAI	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
- 3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

Continuation of Calibration Certificate

Cert. No. : ACL22036
Job No. : VC65AC0041
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.6

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	13.1
C - weight	19.1
Flat	24.6

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			Acceptance Limits
	Flat	C-weight	A-weight	
125	0.1	0.2	0.2	± 1.5
1000	-0.1	-0.1	-0.1	± 1.0
8000	-0.3	-0.1	-0.1	±5.0

Continuation of Calibration Certificate

Cert. No. : ACL22036
Job No. : VC65AC0041
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	-0.1	-0.1	±2.0
125	0.0	0.0	0.0	±1.5
250	0.0	0.0	0.0	±1.5
500	0.0	0.0	0.0	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.1	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

Continuation of Calibration Certificate

Cert. No. : ACL22036
Job No. : VC65AC0041
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	28.9	-0.1	± 1.1
28.0	27.9	-0.1	± 1.1
27.0	26.9	-0.1	± 1.1
26.0	26.0	0.0	± 1.1
25.0	24.9	-0.1	± 1.1

Continuation of Calibration Certificate

Cert. No. : ACL22036
Job No. : VC65AC0041
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	± 1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.1	0.1	± 1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	± 1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.1	0.1	± 1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, Lcpeak (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.4	0.0	± 3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.2	-0.2	± 2.0
Negative half cycle	135.4	135.2	-0.2	± 2.0

Continuation of Calibration Certificate

Cert. No. : ACL22036
 Job No. : VC65AC0041
 Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle	0.0	±1.5
89.6	89.6		

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate



451-451/1 Sirinthorn Rd, Bangbunru, Bangkok 10700 THAILAND.
 Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiphorn.com http://www.sithiphorn.com

Cert. No. : ACL21144
 Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00572551 / 170383 / 72889
ID No.: BKK_FS0876

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
 104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
 KHWAEANG PHATTHANAKAN, KHET SUAN LUANG,
 BANGKOK, 10250 THAILAND.

Location :
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 01 NOVEMBER 2021
Calibration Date : 02-04 NOVEMBER 2021
Date of Issue : 05 NOVEMBER 2021

Calibrated by : Nathakorn Pisutpaisan

REVIEW BY	<i>Nathakorn P.</i>
APPROVED BY	<i>T. Petchurai</i>
NEXT CAL. DATE	2/11/22

Approved by :
T. Petchurai
 (Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Continuation of Calibration Certificate

Cert. No. : ACL21144
Job No. : VC65AC0011
Pages : 2 of 8Cert. No. : ACL21144
Job No. : VC65AC0011
Pages : 3 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	8846A	1997025	EEL.BP. 06/0264	05-Feb-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KAI	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
- 3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

Continuation of Calibration Certificate

Cert. No. : ACL21144
Job No. : VC65AC0011
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
15.7

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	12.6
C - weight	18.7
Flat	24.3

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight
125	0.4	0.4	0.4
1000	0.0	0.0	0.0
8000	-1.0	-1.0	-0.9
			±5.0

Continuation of Calibration Certificate

Cert. No. : ACL21144
Job No. : VC65AC0011
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight
63	-0.1	-0.1	-0.1
125	-0.1	0.0	0.0
250	0.0	0.0	-0.1
500	0.0	0.0	-0.1
1000	0.0	0.0	0.0
2000	0.0	0.0	0.0
4000	0.0	0.0	0.0
8000	0.0	0.1	0.1
			±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	±0.2
Flat	94.0	0.0	±0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	±0.1
Leq	94.0	0.0	±0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	±0.3

Continuation of Calibration Certificate

Cert. No. : ACL21144
Job No. : VC65AC0011
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	±1.1
136.0	136.0	0.0	±1.1
135.0	135.0	0.0	±1.1
134.0	134.0	0.0	±1.1
133.0	133.0	0.0	±1.1
132.0	132.0	0.0	±1.1
131.0	131.0	0.0	±1.1
129.0	129.0	0.0	±1.1
124.0	124.0	0.0	±1.1
119.0	119.0	0.0	±1.1
114.0	114.0	0.0	±1.1
109.0	109.0	0.0	±1.1
104.0	104.0	0.0	±1.1
99.0	99.0	0.0	±1.1
94.0	94.0	0.0	±1.1
89.0	89.1	0.1	±1.1
84.0	84.1	0.1	±1.1
79.0	79.0	0.0	±1.1
74.0	74.1	0.1	±1.1
69.0	69.1	0.1	±1.1
64.0	64.0	0.0	±1.1
59.0	59.1	0.1	±1.1
54.0	54.0	0.0	±1.1
49.0	49.0	0.0	±1.1
44.0	44.0	0.0	±1.1
39.0	39.0	0.0	±1.1
34.0	34.0	0.0	±1.1
30.0	30.0	0.0	±1.1
29.0	29.0	0.0	±1.1
28.0	28.0	0.0	±1.1
27.0	27.0	0.0	±1.1
26.0	26.0	0.0	±1.1
25.0	24.9	-0.1	±1.1

Continuation of Calibration Certificate

Cert. No. : ACL21144
Job No. : VC65AC0011
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 : -5.0
	2	8	117.0	117.0	0.0	1.0 : -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 : -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 : -5.0
	2	8	108.0	108.0	0.0	1.0 : -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, Lepeak (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.1	-0.3	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.2	-0.2	±2.0

Continuation of Calibration Certificate

Cert. No. : ACL21144
Job No. : VC65AC0011
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.6	89.6	0.0	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate



451-451/1 Sirinthorn Rd., Bangbunmu, Bangplud Bangkok 10700 THAILAND.
Tel.0-2435-8800 Fax.0-2433-1679 e-mail:cal-center@sithiphorn.com http://www.sithiphorn.com

Cert. No. : ACL22039
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00572452 / 171618 / 72790
ID No.: BKK_FS0922

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 05 JANUARY 2022
Calibration Date : 12-14 JANUARY 2022
Date of Issue : 17 JANUARY 2022

Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

REVIEW BY	<i>Nathakorn P.</i>
APPROVED BY	<i>[Signature]</i>
NEXT CAL. DATE	12 / 1 / 23

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Continuation of Calibration Certificate

Cert. No. : ACL22039
Job No. : VC65AC0041
Pages : 2 of 8Cert. No. : ACL22039
Job No. : VC65AC0041
Pages : 3 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL-BP_05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL-BP_03/0264	08-Feb-22
Digital Multimeter	34461A	MY60024273	1-15180725251-1	15-Sep-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KAI	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
- 3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

Continuation of Calibration Certificate

Cert. No. : ACL22039
Job No. : VC65AC0041
Pages : 4 of 8

Cert. No. : ACL22039
Job No. : VC65AC0041
Pages : 5 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.6

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	12.0
C - weight	18.8
Flat	24.4

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight
125	0.2	0.2	0.2
1000	-0.1	-0.1	-0.1
8000	-0.5	-0.5	-0.4
			Acceptance Limits
			± 1.5
			± 1.0
			± 5.0

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	0.0	0.0	±2.0
125	0.0	0.0	0.0	±1.5
250	0.0	0.0	-0.1	±1.5
500	0.0	0.0	0.0	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

Continuation of Calibration Certificate

Cert. No. : ACL22039
Job No. : VC65AC0041
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	53.9	-0.1	± 1.1
49.0	49.0	0.0	± 1.1
44.0	43.9	-0.1	± 1.1
39.0	38.9	-0.1	± 1.1
34.0	33.9	-0.1	± 1.1
30.0	29.9	-0.1	± 1.1
29.0	28.9	-0.1	± 1.1
28.0	27.9	-0.1	± 1.1
27.0	26.9	-0.1	± 1.1
26.0	25.9	-0.1	± 1.1
25.0	24.8	-0.2	± 1.1

Continuation of Calibration Certificate

Cert. No. : ACL22039
Job No. : VC65AC0041
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	± 1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.0	0.0	± 1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	± 1.0
	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
SEL	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	± 1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, Lcpeak (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	135.8	-0.6	± 3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.2	-0.2	± 2.0
Negative half cycle	135.4	135.2	-0.2	± 2.0

Continuation of Calibration Certificate

Cert. No. : ACL22039
Job No. : VC65AC0041
Pages : 8 of 8

11. Overload indication

Measured value (dB)	Deviated Value (dB)		Acceptance Limits (dB)
	Positive one-half cycle	Negative one-half cycle	
89.5	89.6	0.1	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)		SLM Display at final (dB)		Deviated Value (dB)	Acceptance Limits (dB)
	137.0		137.0			
A - weight					0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

451-451/1 Sirinthorn Rd, Bangbunru, Bangkok 10700 THAILAND.
Tel.0-2435-8800 Fax.0-2433-1679 e-mail:cal-center@sithiphorn.com http://www.sithiphorn.com

Cert. No. : ACL21123
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42 / Microphone UC-52 / Preamplifier NH-24
Serial No.: 00572609 / 157781 / 48096
ID No.: BKK_FS0924

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location :

Ambient Temperature : (23.0 ± 3) °C

Pressure : (101.3 ± 3) kPa

Relative Humidity : (50.0 ± 20) %

Received Date : 29 SEPTEMBER 2021

Calibration Date : 12-14 OCTOBER 2021

Date of Issue : 15 OCTOBER 2021

Calibrated by :

Nathakorn Pisupaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

REVIEW BY <i>Nathakorn P.</i>
APPROVED BY <i>T. Petchurai</i>
NEXT CAL. DATE 12/10/22

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Continuation of Calibration Certificate

Calibration Procedure : CP-AC-01
Cert. No. : ACL21123
Job No. : VC64AC0071
Pages : 2 of 8

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	8846A	1997025	EEL.BP. 06/0264	05-Feb-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KAI	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
- 3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL21123
Job No. : VC64AC0071
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

Continuation of Calibration Certificate

Cert. No. : ACL21123
Job No. : VC64AC0071
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
17.2

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	12.6
C - weight	18.7
Flat	24.6

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight
125	0.5	0.5	0.5
1000	0.0	0.0	0.0
8000	-1.0	-0.9	-0.9

Continuation of Calibration Certificate

Cert. No. : ACL21123
Job No. : VC64AC0071
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight
63	-0.1	-0.1	-0.1
125	0.0	0.0	-0.1
250	0.0	0.0	-0.1
500	-0.1	0.0	-0.1
1000	0.0	0.0	-0.1
2000	-0.1	0.0	0.0
4000	-0.1	0.0	0.0
8000	0.0	0.0	0.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.1	0.1	± 0.3

Continuation of Calibration Certificate

Cert. No. : ACL21123
Job No. : VC64AC0071
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.1	0.1	± 1.1
134.0	134.1	0.1	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.1	0.1	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.1	0.1	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.1	0.1	± 1.1
28.0	28.1	0.1	± 1.1
27.0	27.1	0.1	± 1.1
26.0	26.1	0.1	± 1.1
25.0	25.2	0.2	± 1.1

Continuation of Calibration Certificate

Cert. No. : ACL21123
Job No. : VC64AC0071
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	116.9	-0.1	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	107.9	-0.1	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.8	-0.2	1.5 ; -5.0
	2	8	108.0	107.9	-0.1	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, Lepeak (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.4	0.0	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

Continuation of Calibration Certificate

Cert. No. : ACL21123
Job No. : VC64AC0071
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.8	89.5	-0.3	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3


The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

T. Ratan

CERTIFICATE OF CALIBRATION

ISSUED BY Cirrus Research plc
DATE OF ISSUE 07/07/21
CERTIFICATE NUMBER 159633



Cirrus Research plc
Acoustic House
Bridlington Road
Hunmanby
North Yorkshire
YO14 0PH
United Kingdom

Page 1 of 1
Test engineer:
Nigel Smith
Electronically signed:

doseBadge Reader

Instrument

Manufacturer: Cirrus Research plc
Model Number: RC:110A
Serial Number: 77670
Notes:

Calibration Procedure

The tests were carried out in accordance with the requirements of IEC 60942:2003 where applicable.

Date of Calibration: 07 July 2021

Functionality Results

Function	Result
Keypad	Pass
Battery Power	Pass
Display	Pass
Communication	Pass
2 way IR link	Pass
Clock	Pass

REVIEW BY
APPROVED BY
NEXT CAL. DATE 7/7/22

Calibration Results

	Level (dB)	Frequency (Hz)	Distortion (% THD + Noise)
Initial	113.80	1000.0	0.26
Adjusted	114.00	1000.0	0.26
Uncertainty	± 0.11	± 0.14	± 0.10
Tolerances	± 0.60	± 2.00	± 4.00

Environmental Conditions

Pressure: 100.10 kPa
Temperature: 21.7 °C
Humidity: 60.6 %

Notes

This certificate provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory. The results within this certificate relate only to the items calibrated. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%.



63/14 15,67/35-36, Soi Petchkasem 7.7/1, Petchkasem Rd,
Wathapra, Bangkokkhai, Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranatee.com



63/14 15,67/35-36, Soi Petchkasem 7.7/1, Petchkasem Rd,
Wathapra, Bangkokkhai, Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranatee.com

CERTIFICATE OF CALIBRATION

Certificate No. : CL-037-64
Page 1 of 2

Equipment Name : Heat Stress Monitor with Sensor
Manufacturer : DeltaOHM
Model: HD32.2
Serial No.: 15006638
ID No.: BKK_FS0869

Customer
Name : ALS laboratory group (thailand) Co.,Ltd.
Address : 104 Phatthanakan 40, Phatthanakan
Rd., Khwaeng Suan Luang, Khet Suan Luang, Bangkok
10250 Thailand.

Received date : 26 May 2021
Calibration date : 07 June 2021
Issue date : 10 June 2021

Reference Used During Calibration

1. Standard Temperature Probe Model : STS-100 A500,
Serial No. : 687682-09, Due date : 25 Mar 2022
2. Digital Temperature Indicator Model : DTI-1000-A MK
II, Serial No.: 671407-00591 Due date : 20 May 2021

Calibration Condition

Temperature : (23±3) °C
Relative Humidity : (65±15) %

Calibration Procedure

The temperature calibration was done by in-house calibration method as WI-CL-001 according to comparison method with standard digital temperature indicator and standard temperature probe. The temperature scale use was based on ITS-90.

Traceability


The measurement results are traceable to the international system of units (SI) through National Institute of Metrology, Thailand (NIMT) Certificate number : IT-0036-21, Certificate number : ER-0071-20

Calibrated by

☐ Mr. Sorawit Thachalad
☒ Miss Orathai Wiwatwattaya

Approved Signatory:

Mr. Parinya Booncharoen
Technical Support
And Calibration Manager



63/14 15,67/35-36, Soi Petchkasem 7.7/1, Petchkasem Rd,
Wathapra, Bangkokkhai, Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranatee.com



Certificate No. : CL-037-64
Page 2 of 2

Result of Calibration:

☒ Without Adjustment ☐ With Adjustment

Calibration Range: 20 °C - 40 °C

Function:

Table 1: This equipment was connected with wet bulb probe Model: HP3201.2 S/N: 15015840.
Dimension: Diameter 14 mm. Length 170 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
30	20.039	20.2	0.2	0.14
30	25.037	25.2	0.2	0.14
30	30.026	30.2	0.2	0.14
30	35.015	35.2	0.2	0.14
30	40.004	40.2	0.2	0.14

Table 2: This equipment was connected with temperature probe Model: TP3207.2 S/N: 15015490.
Dimension: Diameter 14 mm. Length 150 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
70	20.052	20.5	0.4	0.17
70	24.875	25.1	0.2	0.14
70	29.824	29.9	0.1	0.14
70	34.767	34.7	-0.1	0.14
70	39.715	39.6	-0.1	0.14


Table 3: This equipment was connected with Globe thermometer probe Model: TP3276.2 S/N: 20019644.
Dimension: Diameter 8 mm. Length 170 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
110	20.039	20.1	0.1	0.14
110	25.037	25.1	0.1	0.14
110	30.026	30.1	0.1	0.14
110	35.015	35.2	0.2	0.17
110	40.004	40.2	0.2	0.14

UUC* - Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

★ End of Certificate ★



63/14-15,67/35-36, Soi Petchkasem 7,7/1, Petchkasem Rd,
Wathapra, Bangkokkhai,Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranalee.com



63/14-15,67/35-36, Soi Petchkasem 7,7/1, Petchkasem Rd,
Wathapra, Bangkokkhai,Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranalee.com

CERTIFICATE OF CALIBRATION

Certificate No. : CL-042-64
Page 1 of 2

Equipment Name : Heat Stress Monitor with Sensor
Manufacturer : DeltaOHM
Model: HD32.2
Serial No: 15006704
ID No: BKK_FS0670

Customer
Name : ALS laboratory group (thailand) Co.,Ltd.
Address : 104 Phatthanakan 40, Phatthanakan
Rd.,Khwaeng Suan Luang, Khat Suan Luang,Bangkok
10250 Thailand.

Received date : 04 June 2021
Calibration date : 11 June 2021
Issue date : 11 June 2021

Reference Used During Calibration

- 1.Standard Temperature Probe Model : STS-100 A500,
Serial No. : 667682-09, Due date : 25 Mar 2022
- 2.Digital Temperature Indicator Model : DTI-1000-A MK
II, Serial No.: 671407-00591 Due date : 20 May 2021

Calibration Procedure

The temperature calibration was done by In-House
calibration method as WI-CL-001 according to
comparison method with standard digital temperature
indicator and standard temperature probe. The
temperature scale use was based on ITS-90.

Traceability

The measurement results are traceable to the
international system of units (SI) through National
Institute of Metrology Thailand (NIMT) Certificate
number : IT-0036-21, Certificate number : ER-007-1-
20

Calibration Condition

Temperature : (23±3) °C
Relative Humidity : (55±15)%

REVIEW BY : *[Signature]*
APPROVED BY : *[Signature]*
NEXT CAL DATE : 11/6/2022

Calibrated by
☐ Mr. Sorawit Thachalad
☒ Miss Orathai Wiwatwittaya

Approved Signatory: *[Signature]*
Mr. Pannya Booncharoen
Technical Support
And Calibration Manager

Result of Calibration : ☒ Without Adjustment ☐ With Adjustment

Calibration Range: 20 °C ~ 40 °C

Function:

Table 1: This equipment was connected with wet bulb probe Model: HP3201.2 S/N: 15017681.
Dimension : Diameter 14 mm. Length 170 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
30	20.044	20.1	0.1	0.14
30	25.029	25.1	0.1	0.14
30	30.016	30.1	0.1	0.14
30	35.003	35.1	0.1	0.14
30	39.984	40.1	0.1	0.14

Table 2: This equipment was connected with temperature probe Model: TP3207.2 S/N: 15015488.
Dimension : Diameter 14 mm. Length 150 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
70	20.043	20.2	0.2	0.14
70	24.853	24.9	0.0	0.14
70	29.812	29.7	-0.1	0.17
70	34.773	34.6	-0.1	0.19
70	39.715	39.5	-0.2	0.14

Table 3: This equipment was connected with Globe thermometer probe Model: TP3276.2 S/N: 15021838.
Dimension : Diameter 8 mm Length 170 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
110	20.043	20.1	0.1	0.14
110	25.029	25.1	0.1	0.14
110	30.016	30.1	0.1	0.14
110	35.003	35.1	0.1	0.14
110	39.984	40.1	0.1	0.14

UUC* : Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing
a level of confidence of approximately 95%.

★ End of Certificate ★

CERTIFICATE OF CALIBRATION

Certificate No. : CL-094-64
Page 1 of 2

Equipment Name : Heat Stress Monitor with Sensor
Manufacturer : DeltaOHM
Model: HD32.2
Serial No: 15006300
ID No: BKK_FS0661

Customer
Name: ALS laboratory group (thailand) Co.,Ltd.
Address: 104 Phatthanakan 40, Phatthanakan
Rd., Khwaeng Suan Luang, Khet Suan Luang, Bangkok
10250 Thailand.

Received date: 8 DEC 2021
Calibration date: 13 DEC 2021
Issue date: 14 DEC 2021

Reference Used During Calibration

1. Standard Temperature Probe Model: STS-100 A500,
Serial No.: 667652-09, Due date: 25 Mar 2022
2. Digital Temperature Indicator Model: DTI-1000-A MK
II, Serial No.: 671407-00591 Due date: 04 June 2022

Calibration Procedure

The temperature calibration was done by In-House calibration method as WI-CL-001 according to comparison method with standard digital temperature indicator and standard temperature probe. The temperature scale use was based on ITS-90.

Calibration Condition

Temperature: (23±3)°C
Relative Humidity: (55±15)%

Traceability

The measurement results are traceable to the international system of units (SI) through National Institute of Metrology Thailand (NIMT) Certificate number: IT-0036-21, Certificate number: ER-0032-21

REVIEW BY *Harikarn P.*
APPROVED BY *466*
NEXT CAL. DATE *13/12/22*

Calibrated by
☒ Mr. Sorawit Thachalad
☐ Miss Orathai Wiwatwittaya

Approved Signatory

Kongth
Mr. Parinya Booncharoen
Technical Support
and Calibration Manager



Certificate No. : CL-094-64
Page 2 of 2

Result of Calibration: ☒ Without Adjustment ☐ With Adjustment
Calibration Range: 20 - 40 °C
Function:

Table 1: This equipment was connected with wet bulb probe Model: HP3201.2 S/N: 15003275.
Dimension: Diameter 14 mm. Length 170 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
30	20.061	20.1	0.0	0.099
30	25.051	25.1	0.0	0.099
30	30.036	30.1	0.1	0.099
30	35.033	35.1	0.1	0.099
30	40.025	40.1	0.1	0.099

Table 2: This equipment was connected with temperature probe Model: TP3207.2 S/N: 150031950.
Dimension: Diameter 14 mm. Length 150 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
70	20.059	20.1	0.0	0.099
70	24.877	24.9	0.0	0.099
70	29.822	29.8	0.0	0.099
70	34.778	34.6	-0.2	0.099
70	39.739	39.6	-0.1	0.099

Table 3: This equipment was connected with Globe thermometer probe Model: TP3276.2 S/N: 15008168.
Dimension: Diameter 8 mm. Length 170 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
110	20.061	20.1	0.0	0.099
110	25.053	25.1	0.0	0.099
110	30.035	30.1	0.1	0.099
110	35.033	35.1	0.1	0.099
110	40.025	40.1	0.1	0.099

UUC* : Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

* End of Certificate *



CERTIFICATE OF CALIBRATION

Certificate No. : CL-070-64
Page 1 of 2

Equipment Name : Heat Stress Monitor with Sensor
Manufacturer : DeltaOHM
Model: HD32.2
Serial No: 15006708
ID No: BKK_FS0671

Customer
Name : ALS laboratory group (thailand) Co.,Ltd.
Address : 104 Phatthanakan 40, Phatthanakan
Rd, Khwaeng Suan Luang, Khet Suan Luang Bangkok
10250 Thailand.

Received date : 8 SEP 2021
Calibration date : 30 SEP 2021
Issue date : 4 OCT 2021

Reference Used During Calibration

1. Standard Temperature Probe Model: STS-100 A500,
Serial No: 667682.09, Due date: 25 Mar 2022
2. Digital Temperature Indicator Model: DTI-1000-A MK
II, Serial No.: 671407-00591 Due date: 04 June 2022

Calibration Procedure

The temperature calibration was done by In-House
calibration method as WI-CL-001 according to
comparison method with standard digital temperature
indicator and standard temperature probe. The
temperature scale use was based on ITS-90.

Traceability

The measurement results are traceable to the
international system of units (SI) through National
Institute of Metrology Thailand (NIMT) Certificate
number: IT-0036-21, Certificate number: ER-0032-
21

Calibration Condition

Temperature : (23±3)°C
Relative Humidity : (55±15)%

REVIEW BY	Mr. Pinyan Booncharoen
APPROVED BY	Mr. Pinyan Booncharoen
NEXT CAL DATE	30/9/22

Calibrated by

☐ Mr. Sorawit Thachalad
☒ Miss Orathai Wiwatwittaya

Approved Signatory:

Mr. Pinyan Booncharoen
Technical Support
and Calibration Manager



Certificate No. : CL-070-64
Page 2 of 2

Result of Calibration : ☒ Without Adjustment ☐ With Adjustment
Calibration Range: 20 °C - 40 °C

Function:

Table 1: This equipment was connected with wet bulb probe Model: HP3201.2 S/N: 15015843.
Dimension: Diameter 14 mm, Length 170 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
30	20.051	20.1	0.0	0.099
30	25.047	25.1	0.1	0.099
30	30.035	30.1	0.1	0.099
30	35.026	35.1	0.1	0.099
30	40.014	40.1	0.1	0.099

Table 2: This equipment was connected with temperature probe Model: TP3207.2 S/N: 15015489.
Dimension: Diameter 14 mm, Length 150 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
70	20.045	20.2	0.2	0.099
70	24.874	25.0	0.1	0.099
70	29.821	29.8	0.0	0.099
70	34.779	34.7	-0.1	0.099
70	39.735	39.6	-0.1	0.099

Table 3: This equipment was connected with Globe thermometer probe Model: TP3276.2 S/N: 15015979.
Dimension: Diameter 8 mm, Length 170 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
110	20.050	19.9	-0.2	0.099
110	25.047	24.9	-0.1	0.099
110	30.035	29.9	-0.1	0.099
110	35.026	34.9	-0.1	0.099
110	40.014	39.8	-0.2	0.099

UUC* : Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing
a level of confidence of approximately 95%.

★ End of Certificate ★



CERTIFICATE OF CALIBRATION

Certificate No. : CL-032-64
Page 1 of 2

Equipment Name : Heat Stress Monitor with Sensor
Manufacturer : Delta OHM
Model: HD32.2
Serial No: 15006302
ID No: BKK_FS0663

Customer
Name : ALS laboratory group (thailand) Co.,Ltd.
Address : 104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand.

Received date : 14 MAY 2021
Calibration date : 19 MAY 2021
Issue date : 21 MAY 2021

REVIEW BY *Parinya P.*
APPROVED BY *[Signature]*
EXT CAL DATE 19/10/22

Reference Used During Calibration

- Standard Temperature Probe Model : STS-100 A500, Serial No. : 667682-09, Due date : 25 Mar 2022
- Digital Temperature Indicator Model : DTI-1000-A MK II, Serial No.: 671407-00591 Due date : 20 May 2021

Calibration Procedure

The temperature calibration was done by in-house calibration method as WI-CL-001 according to comparison method with standard digital temperature indicator and standard temperature probe. The temperature scale use was based on ITS-90.

Traceability

The measurement results are traceable to the international system of units (SI) through National Institute of Metrology Thailand (NIMT) Certificate number : TT-0036-21, Certificate number : ER-0071-20

Calibrated by

- ☐ Mr. Sorawit Thachalad
☒ Mr. Bongkoch Maithong

Approved Signatory:

[Signature]
Mr. Parinya Booncharoen
Technical Support
And Calibration Manager



THIS CERTIFICATE REPORT MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.

Certificate No. : CL-032-64
Page 2 of 2

Result of Calibration : ☒ Without Adjustment ☐ With Adjustment

Calibration Range: 20 °C – 40 °C

Function:

Table 1: This equipment was connected with wet bulb probe Model: HP3201.2 S/N: 15003284.
Dimension : Diameter 14 mm. Length 170 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
30	20.041	20.1	0.1	0.10
30	25.035	25.0	0.0	0.16
30	30.018	30.0	0.0	0.10
30	35.004	35.0	0.0	0.10
30	39.988	40.0	0.0	0.10

Table 2: This equipment was connected with temperature probe Model: TP3207.2 S/N: 15008018.
Dimension : Diameter 14 mm. Length 150 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
70	20.041	20.3	0.3	0.13
70	25.034	25.0	0.0	0.13
70	30.019	29.8	-0.2	0.13
70	35.004	34.7	-0.3	0.17
70	39.991	39.5	-0.5	0.13

Table 3: This equipment was connected with Globe thermometer probe Model: TP3276.2 S/N: 15008170.
Dimension : Diameter 8 mm. Length 170 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
110	20.078	20.3	0.2	0.28
110	25.034	25.3	0.3	0.10
110	30.018	30.3	0.3	0.10
110	35.005	35.3	0.3	0.10
110	39.989	40.3	0.3	0.10

UUC* : Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

*** End of Certificate ***



CERTIFICATE OF CALIBRATION

Certificate No.: CL-014-65
Page 1 of 2

Equipment Name: Heat Stress Monitor with Sensor
Manufacturer: DeltaOHM
Model: HD32.2
Serial No: 15006309
ID No: BKK_FS0667

Customer
Name: ALS laboratory group (thailand) Co., Ltd.
Address: 104 Phatthanakan 40, Phatthanakan
Rd., Khwaeng Suan Luang, Khet Suan Luang, Bangkok
10250 Thailand.

Received date: 10 JAN 2022
Calibration date: 14 FEB 2022
Issue date: 17 FEB 2022

Reference Used During Calibration

1. Standard Temperature Probe Model: STS-100 A500,
Serial No.: 667682-09, Due date: 25 Mar 2022
2. Digital Temperature Indicator Model: DTI-1000-A MK
II, Serial No.: 671407-00591 Due date: 04 June 2022

Calibration Procedure

The temperature calibration was done by In-House calibration method as WI-CL-001 according to comparison method with standard digital temperature indicator and standard temperature probe. The temperature scale use was based on ITS-90.

Calibration Condition

Temperature: (23±3) °C
Relative Humidity: (55±15)%

Traceability

The measurement results are traceable to the international system of units (SI) through National Institute of Metrology Thailand (NIMT) Certificate number: TT-0036-21, Certificate number: ER-0032.

21

REVIEW BY	<i>Wanpim P</i>
APPROVED BY	<i>W</i>
NEXT CAL. DATE	14/2/23

Calibrated by
☐ Mr. Sorawit Thachalad
☒ Miss Orathai Wiwatwittaya

Approved Signatory: *Wanpim P*
Mr. Panniya Booncharoen
Calibration Department Manager

J NAC
JIRANATE ASSOCIATES CO., LTD.

Result of Calibration: ☒ Without Adjustment ☐ With Adjustment
Calibration Range: 20 - 40 °C
Function:

Table 1: This equipment was connected with wet bulb probe Model: HP3201.2 S/N: 15017683.
Dimension: Diameter 14 mm, Length 170 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
30	20.056	20.2	0.1	0.099
30	25.046	25.2	0.2	0.099
30	30.036	30.2	0.2	0.099
30	35.032	35.2	0.2	0.099
30	40.027	40.2	0.2	0.099

Table 2: This equipment was connected with temperature probe Model: TP3207.2 S/N: 15015506.
Dimension: Diameter 14 mm, Length 150 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
70	20.050	20.1	0.1	0.099
70	24.970	24.9	-0.1	0.099
70	29.932	29.8	-0.1	0.099
70	34.876	34.6	-0.3	0.099
70	39.805	39.5	-0.3	0.099

Table 3: This equipment was connected with Globe thermometer probe Model: TP3276.2 S/N: 15009822.
Dimension: Diameter 8 mm, Length 170 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
110	20.055	20.0	-0.1	0.099
110	25.045	25.0	0.0	0.099
110	30.037	30.0	0.0	0.099
110	35.033	35.0	0.0	0.099
110	40.027	40.0	0.0	0.099

UUC*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

* End of Certificate *

J NAC
JIRANATE ASSOCIATES CO., LTD.



ROTA METER CALIBRATION RESULT JANUARY 2021

Rotameter ID.	Calibration Date	Regression Result	Coefficient (R ²)
BKK_FS0577	04 Jan 21	Y = 1.5816x - 11.072	0.9924
BKK_FS0579	04 Jan 21	Y = 1.0782x - 5.4528	1.0000
BKK_FS0583	04 Jan 21	Y = 0.9961x - 1.6228	0.9989
BKK_FS0584	04 Jan 21	Y = 1.0457x - 25.58	0.9978
BKK_FS0585	04 Jan 21	Y = 0.9194x + 3.669	0.9429
BKK_FS0586	04 Jan 21	Y = 1.0487x - 3.6038	0.9999
BKK_FS0587	04 Jan 21	Y = 0.9611x + 22.64	0.9951
BKK_FS0588	04 Jan 21	Y = 0.8525x + 28.083	0.9131
BKK_FS0589	04 Jan 21	Y = 0.9846x + 2.2929	0.9848
BKK_FS0590	04 Jan 21	Y = 0.9896x + 1.1571	0.9999
BKK_FS0591	04 Jan 21	Y = 1.0586x - 154.323	0.9983
BKK_FS0592	04 Jan 21	Y = 1.0243x - 130.39	0.9951
BKK_FS0593	04 Jan 21	Y = 1.0118x - 46.331	0.9991
BKK_FS0594	04 Jan 21	Y = 1.0068x - 16.51	0.9996
BKK_FS0595	04 Jan 21	Y = 1.0142x - 44.708	0.9990
BKK_FS0596	04 Jan 21	Y = 1.0026x - 12.106	0.9999
BKK_FS0597	04 Jan 21	Y = 1.0124x - 31.882	0.9999
BKK_FS1004	01 Jan 21	Y = 1.018x + 6.4121	0.9998
BKK_FS1005	01 Jan 21	Y = 0.9653x + 20.281	0.9977
BKK_FS1006	01 Jan 21	Y = 1.2167x - 6.4807	0.9992
BKK_FS1007	04 Jan 21	Y = 0.9928x - 2.3044	0.9997
BKK_FS1008	04 Jan 21	Y = 0.9867x - 1.2	0.9988
BKK_FS1009	04 Jan 21	Y = 1.1108x - 5.5888	0.9965
BKK_FS1010	04 Jan 21	Y = 1.0071x + 0.4769	0.9722
BKK_FS1011	04 Jan 21	Y = 1.0505x - 8.134	0.9998
BKK_FS1012	04 Jan 21	Y = 1.0119x - 7.4569	0.9999
BKK_FS1013	04 Jan 21	Y = 1.0003x - 31.261	0.9998
BKK_FS1014	04 Jan 21	Y = 0.9497x - 4.162	0.9942
BKK_FS1015	04 Jan 21	Y = 0.9737x - 3.7733	0.9996
BKK_FS1016	04 Jan 21	Y = 0.9883x + 3.7911	0.9996
BKK_FS1017	04 Jan 21	Y = 1.0689x - 5.2468	0.9998
BKK_FS1018	04 Jan 21	Y = 1.009x - 3.9231	0.9995
BKK_FS1019	04 Jan 21	Y = 1.0079x - 106.39	0.9992
BKK_FS1020	04 Jan 21	Y = 1.1045x - 11.412	0.9776
BKK_FS1021	04 Jan 21	Y = 0.8797x + 31.9	0.9617
BKK_FS1022	04 Jan 21	Y = 1.0179x - 55.217	0.9999
BKK_FS1023	04 Jan 21	Y = 1.0856x - 6.2755	0.9973
BKK_FS1024	04 Jan 21	Y = 0.8119x + 54.124	0.9771
BKK_FS1025	04 Jan 21	Y = 1.0353x - 100.17	0.9998



ROTA METER CALIBRATION RESULT JANUARY 2021

Rotameter ID.	Calibration Date	Regression Result	Coefficient (R ²)
BKK_FS1026	04 Jan 21	Y = 1.1049x - 4.335	0.9989
BKK_FS1027	04 Jan 21	Y = 0.9684x + 13.807	0.9992
BKK_FS1028	04 Jan 21	Y = 1.0279x - 60.396	0.9998
BKK_FS1029	04 Jan 21	Y = 1.0303x - 5.0748	0.9883
BKK_FS1030	04 Jan 21	Y = 1.0215x - 25.512	0.9897
BKK_FS1031	04 Jan 21	Y = 1.0211x - 51.359	0.9998
BKK_FS1039	01 Jan 21	Y = 0.9959x + 15.57	0.9998
BKK_FS1040	01 Jan 21	Y = 1.0061x - 1.3876	0.9999
BKK_FS1041	01 Jan 21	Y = 1.0987x - 4.5109	0.9992
BKK_FS1042	01 Jan 21	Y = 0.9554x + 25.905	0.9970
BKK_FS1043	01 Jan 21	Y = 0.9967x + 14.697	0.9997
BKK_FS1044	01 Jan 21	Y = 1.0401x - 1.0979	0.9992
RYG_FS0197	01 Jan 21	Y = 1.0031x + 6.9733	0.9999
RYG_FS0198	01 Jan 21	Y = 0.9837x + 68.26	0.9995
RYG_FS0199	01 Jan 21	Y = 0.9851x + 1.3543	0.9960

Review By :

(Mr. Wichan Choonharat)

Enviro Field Services Manager

Approved By :

(Mr. Sarayuth Jitranont)

Assistant General Manager



Bara Scientific Co., Ltd.
968 U Chu Liang Building Floor 7 Rama4 Road
Silom Bangkok Bangkok Thailand 10500
Tel : 02-6324300 Fax : 02-6375496-7
www.barascientific.com




Bara Scientific
CALIBRATION LABORATORY

Certificate of Calibration

Number of Page(s) 1 of 3

Certificate No. BSCC-UV-290/21
Equipment UVVis Spectrophotometer
Model UV-1800
Manufacturer Shimadzu
Serial No. A11454908533CD
ID No. BKK_EN0018
Date of receipt 15 October 2021
Date of calibration 15 October 2021
Date of issue 25 October 2021
Customer name ALS Laboratory Group (Thailand) Co., Ltd.
Address 104 Soi Phatthanakan 40, Phatthanakan Road, Phatthanakan, Suan Luang, Bangkok 10250



Temperature (25.0 - 26.4) °C (On site)
Humidity (49.5 - 53.4) %RH (On site)

Equipment condition Good Operation
Calibration Location Organic Prep

Calibration Procedure In-house method WI-UV-702-01 based on ASTM E275-01
Traceability 87839 and 87844
Wavelength Accuracy is traceable to certificate No. 87846 and 87877
Photometric Accuracy is traceable to certificate No. 87825
Stray Light is traceable to certificate No.
The above certificate are traceable to SI unit through Starna Scientific Ltd.
(UKAS accredited calibration laboratory NO. 0659)

Calibrated by Mr.Wandhana Janloey

Approved by

Mr.Kanchit Choothep
Technical Manager

The above results are valid exclusively for the calibrated item(s) as mention in this report / certificate.
Advertising the report / Certificate and publicity of the results are prohibited and also shall not be reproduced
except in full, without written approval of the Bara Scientific Co., Ltd.



Bara Scientific Co., Ltd.
968 U Chu Liang Building Floor 7 Rama4 Road
Silom Bangkok Bangkok Thailand 10500
Tel : 02-6324300 Fax : 02-6375496-7
www.barascientific.com



Certificate of Calibration

Number of Page(s) 2 of 3

Certificate No. BSCC-UV-290/21

Calibration Results:

1.Wavelength Accuracy

Certified Wavelength (nm)	UUC (nm)	Error (nm)	Uncertainty (±nm)
241.70	241.55	-0.15	0.18
334.02	333.80	-0.22	0.18
418.53	418.40	-0.13	0.18
572.99	572.85	-0.14	0.18
879.41	879.15	-0.26	0.18

2.Photometric Accuracy (UV)

Wavelength (nm)	Certified Absorbance (A)	UUC (A)	Error (A)	Uncertainty (±A)
235	0.0000	0.0000	0.0000	0.0075
	0.7174	0.7198	0.0024	0.0075
257	0.0000	-0.0001	-0.0001	0.0075
	0.8362	0.8377	0.0015	0.0075
313	0.0000	0.0000	0.0000	0.0075
	0.2778	0.2803	0.0025	0.0075
350	0.0000	-0.0001	-0.0001	0.0075
	0.6202	0.6221	0.0019	0.0075

*CNR = Customer not request

The above results are valid exclusively for the calibrated item(s) as mention in this report / certificate.
Advertising the report / Certificate and publicity of the results are prohibited and also shall not be reproduced
except in full, without written approval of the Bara Scientific Co., Ltd.



Bara Scientific Co., Ltd.
968 U Chu Liang Building Floor7 Rama4 Road
Siam Bangkok Bangkok Thailand 10500
Tel : 02-6324300 Fax : 02-6375496-7
www.barascientific.com



Bara Scientific
SOLUTION OF SUCCESS

Certificate of Calibration

3 of 3

Number of Page(s)

BSCC-UV-290/21

Certificate No.

Calibration Results:

3. Photometric Accuracy (Visible)

Wavelength (nm)	Certified Absorbance (A)	UUC (A)	Error (A)	Uncertainty (±A)
420.0	0.0000 0.9631 0.7390 1.0863	0.0000 0.5570 0.7334 1.0816	0.0000 -0.0061 -0.0056 -0.0047	0.0042 0.0042 0.0042 0.0042
440.0	0.0000 0.5524 0.7217 1.0606	0.0000 0.5489 0.7166 1.0570	0.0000 -0.0055 -0.0051 -0.0038	0.0042 0.0042 0.0042 0.0042
465.0	0.0000 0.5018 0.6657 0.9775	0.0000 0.4966 0.6610 0.9740	0.0000 -0.0052 -0.0047 -0.0035	0.0042 0.0042 0.0042 0.0042
546.1	0.0000 0.5147 0.6743 0.9909	0.0000 0.5113 0.6705 0.9890	0.0000 -0.0034 -0.0038 -0.0019	0.0042 0.0042 0.0042 0.0042
590.0	0.0000 0.5427 0.7037 1.0338	0.0000 0.5394 0.7001 1.0323	0.0000 -0.0033 -0.0036 -0.0015	0.0042 0.0042 0.0042 0.0042
635.0	0.0000 0.5268 0.6720 0.9864	0.0000 0.5235 0.6685 0.9847	0.0000 -0.0033 -0.0035 -0.0017	0.0042 0.0042 0.0042 0.0042

*CNIR - Customer not request

4. Stray Light*

Standard cut-off wavelength (nm)	Unit Under Calibration(UUC)	
	Wavelength (nm)	Absorbance (A)
200.91±0.11nm	200.31	2.0274

The Stray light transmission reference is less than 1.0%T and Stray light absorbance reference is greater than 2.00A

*Stray Light not NSC-ONSC Accredited.

The measurement uncertainty is base on a standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

End of Certificate

The above results are valid exclusively for the calibrated item(s) as mention in this report / certificate.
Advertising the report / Certificate and publicity of the results are prohibited and also shall not be reproduced except in full, without written approval of the Bara Scientific Co., Ltd.

© 2021 by Agilent Technologies

Agilent CrossLab Compliance Services

Certificate of System Qualification

GC-OQ + GCMS-OQ

REVIEW BY	Sorarat M.
APPROVED BY	Ch
NEXT CAL DATE	1 April 23

System ID:

GM-2

Organization Name:

ALS Laboratory Group (Thailand) Co., Ltd.

Organization Location:

104 Phatthanakan 40, Phatthanakan Rd., Kheiwang Suan Luang, Khet Suan Luang, Bangkok 10250

Date:

October 1, 2021 1:10:17 PM

EQP Name:

AgilentRecommended, AgilentRecommended

EQP Revision:

GC.02.51, GCMS.02.51

Overall Qualification Status:

Pass

System Inspection and Basic Safety and Operation

Name:

7890

Setpoint Status:

Pass

Overall System Inspection and Basic Safety and Operation Test Status

Pass

Inlet Pressure Accuracy

Name:

7890

Front

MMI

Setpoint Status:

Pass

Setpoint

Actual

Inlet Pressure:

25.0 psi

24.9 psi

Accuracy:

0.1 psi

Agilent Recommended:

<=

1.2

Overall Inlet Pressure Accuracy Test Status

Pass

GC Oven Temperature Accuracy

Name:

7890

Date: October 1, 2021 1:10:17 PM

System ID: GM-2

Setpoint Status:
Zone:

Temperature: °C

Accuracy: °C

Agilent Recommended: % setpoint in K % setpoint in K °C °C

Setpoint Status:
Zone:

Temperature: °C

Accuracy: °C

Agilent Recommended: % setpoint in K % setpoint in K °C °C

Overall GC Oven Temperature Accuracy Test Status

GC Oven Temperature Stability

Name:

Setpoint Status:

Temperature: °C

Stability: °C

Agilent Recommended: °C

Overall GC Oven Temperature Stability Test Status

Log Amp

Tested Combination1 Front MMI / External SQ

Name:

Setpoint Status:

Overall Log Amp Test Status

RPA

Tested Combination1

Name:

Setpoint Status:

Amu: m/z

Agilent Recommended: mV mV

Overall RPA Test Status

Tune EI

Tested Combination1

Name:

Setpoint Status:

Filament:

Setpoint Status:

Filament:

Overall Tune EI Test Status

Scouting Run

Tested Combination1

Name:

Source:

Setup Status:

Completed

Injection Volume on Column:

1.0 μ L

Overall Scouting Run Status

Completed

Signal to Noise EI

Tested Combination1 Front MMI / External SQ

Name: 5975C inert XL with TAD

Source: EI - Inert Filament: 1

Setup Status:

Pass

Signal to Noise:

819
320

Agilent Recommended:

>=

Source: EI - Inert Filament: 2

Setup Status:

Pass

Signal to Noise:

647
320

Agilent Recommended:

>=

Overall Signal to Noise EI Test Status

Pass

Injection Precision

Tested Combination1 Front MMI / External SQ

Name: 7693A

Source: EI - Inert

Setup Status:

Pass

Injection Volume on Column:

1.0 μ L

Area RSD:

4.75 %
5.00

Agilent Recommended:

<=

Retention Time RSD:

0.02 %
1.00

Overall Injection Precision Test Status

Pass

Mass Ratio Precision

Tested Combination1

Front MMI / External SQ

Injection Tower

Name:

7693A

Source:

EI - Inert

Setup Status:

Pass

Injection Volume on Column:

1.0 μ L

Area Mass 1

Abundance's

RSD:

4.75 %
5.00

Agilent Recommended:

Pass

Mass Ratio

0.81
<= 5.00
Pass

Overall Mass Ratio Precision Test Status

Pass

Instrument Details

Purpose

This section describes the as found system configuration.

Details

System

System ID GM-2
Manufacturer Agilent Technologies

Name 7890

Flow Data Input Manual Data

Temperature Data Input Manual Data or Other Data Logging

Tested Combination1

Injection Technique Injection Tower

Inlet Front

Detector External

LTM Included? No

Sampler 1

Manufacturer Agilent Technologies

Type Injection Tower

Name 7693A

Model Number G4513A

Serial Number CN10120123

Firmware Revision A.10.08

Usage Sample Injection

Location Front

Syringe Volume (µL) 10

Sampler 2

Manufacturer Agilent Technologies

Type Tray

Name 7693A

Model Number G4514A

Serial Number CN10060099

Firmware Revision A.10.16

Vial Heater Not installed

Mainframe 1

Manufacturer Agilent Technologies

Name 7690

Model Number G3440A

Serial Number CN10141049

Firmware Revision A.01.16

Oven Type Standard

Inlet 1

Manufacturer Agilent Technologies

Name 7890

Type MMI

Location Front

Carrier Gas Helium

Control Type Electronic Pressure Control (EPC)

Purged Inlet Yes

Detector 1

Manufacturer Agilent Technologies

Name Mass Spectrometer

Type Mass Spectrometer

Location External

Mass Spectrometer 1

Manufacturer	Agilent Technologies
Type	SQ
Name	5975C inert XL with TAD
Serial Number	US10153217
Firmware Revision	5.02.12
High Vacuum System	Turbo Pump
Scouting Run Standard	OFN Std

MS EI Source 1

Manufacturer	Agilent Technologies
Source Type	EI - Inert
Number of filaments	2

Electronic Signature

Purpose

This signature page was created and published because the ACE sign-off action was executed, which is valid for the entire document, including attachments. The ACE sign-off is an electronic signature that requires two distinct identification components: unique username and personal password. The Agilent representative who has delivered this service understands the meaning and legal status of an electronic signature. As a trained official operator, the Agilent representative has a unique password and login to access ACE and electronically sign this document. (Other e-signatures can be applied to this document using a Document Content Management or other suitable method defined in your data access and control procedures.)

Details

Full Name of Signer:	Supasak Nimsongtham
Logged On User Name:	supasak.nimsongtham@agilent.com
Signature Creation Date:	October 1, 2021
Reason for Signature:	Executed protocol and published this original version of document

Regulatory Disclaimer

This document provides a protocol to verify and record instrument configuration and evidence of proper operation. It has been prepared from our interpretation of applicable regulations as well as industry best practices. The document is designed to provide an important component of a complete compliance package. Validation depends upon many factors and use of this protocol alone does not assure compliance. Agilent Technologies makes no promises or representations as to its sufficiency for any specific regulatory program.

Warranty

Agilent Technologies makes no warranty of any kind to this material, including but not limited to, the implied warranties or merchantability and fitness for a particular purpose. Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

User Name: supasak.nimsongtham
Hostname: SCG115HKC

System ID: GM-2
Print Date: October 1, 2021 1:10:19 PM

ALS_GM2 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 1, 2021 12:42:37 PM	Audit	Session Created	Session	None
October 1, 2021 12:42:37 PM	Start	Configuration	Session	None
October 1, 2021 12:42:37 PM	Audit	Enrollment	Licensing	User is Field Engineer and does not require an unlock code
October 1, 2021 12:44:21 PM	Audit	Exp. loaded	Session	EOP details for primary technique (GC) - File path: [Protocol\Facts\GCs\Conf\lens\02.51\GCs\02.51.eop] EOP File Name: [GC\02.51.eop], EOP Name: [AgilentRecommended] EOP details for hypothetical technique (GCs) - File path: [Protocol\Facts\GCs\Conf\lens\02.51\GCs\02.51.eop] EOP File Name: [GCs\02.51.eop], EOP Name: [AgilentRecommended]
October 1, 2021 12:44:24 PM	End	Configuration	Session	None
October 1, 2021 12:44:28 PM	Start	Qualification	Session	QC
October 1, 2021 12:44:28 PM	Start	Execution	System Inspection and Basic Safety and Operation - 7890 - Qualitative Test - No setpoints associated	None

Page 1 / 7

User Name: supasak.nimsongtham
Hostname: SCG115HKC

System ID: GM-2
Print Date: October 1, 2021 1:10:19 PM

ALS_GM2 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 1, 2021 12:47:35 PM	End	Execution	System Inspection and Basic Safety and Operation - 7890 - Qualitative Test - No setpoints associated	Run Count : 1
October 1, 2021 12:47:37 PM	Start	Execution	Inlet Pressure Accuracy - Front MMIL - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	None
October 1, 2021 12:47:42 PM	End	Execution	Inlet Pressure Accuracy - Front MMIL - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	Run Count : 1
October 1, 2021 12:47:44 PM	Start	Execution	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	None
October 1, 2021 12:48:04 PM	Audit	Data	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Manual Data Entry
October 1, 2021 12:48:05 PM	End	Execution	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Run Count : 1
October 1, 2021 12:48:07 PM	Start	Execution	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	None
October 1, 2021 12:48:34 PM	Audit	Data	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Manual Data Entry
October 1, 2021 12:48:36 PM	End	Execution	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Run Count : 1

Page 2 / 7

User Name: supasak.niamsongtham
Hostname: SC0115HNC

System ID: GM-2
Print Date: October 1, 2021 1:16:19 PM

ALS_GM2 Transaction log:

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 1, 2021 12:46:38 PM	Start	Execution	GC Oven Temperature Stability	None - 78.00°C - Temperature: Oven - S: 100.0°C - L: <= 0.5°C
October 1, 2021 12:49:34 PM	Audit	Data	GC Oven Temperature Stability	Manual Data Entry - 78.00°C - Temperature: Oven - S: 100.0°C - L: <= 0.5°C
October 1, 2021 12:49:35 PM	End	Execution	GC Oven Temperature Stability	Run Count: 1 - 78.00°C - Temperature: Oven - S: 100.0°C - L: <= 0.5°C
October 1, 2021 12:48:37 PM	Start	Execution	Log Amp - 6975C Inert XL with TAD SQ - Source: EI - Inert	None
October 1, 2021 12:48:47 PM	End	Execution	Log Amp - 6975C Inert XL with TAD SQ - Source: EI - Inert	Run Count: 1
October 1, 2021 12:49:48 PM	Start	Execution	RPFA - 6975C Inert XL with TAD SQ - Source: EI - Inert	None
October 1, 2021 12:50:23 PM	End	Execution	RPFA - 6975C Inert XL with TAD SQ - Source: EI - Inert	Run Count: 1
October 1, 2021 12:50:25 PM	Start	Execution	Tune EI - 6975C Inert XL with TAD SQ - Source: EI - Inert Filament 1 (Qualitative - No setpoints associated)	None
October 1, 2021 12:50:49 PM	End	Execution	Tune EI - 6975C Inert XL with TAD SQ - Source: EI - Inert Filament 1 (Qualitative - No setpoints associated)	Run Count: 1
October 1, 2021 12:50:50 PM	Start	Execution	Tune EI - 6975C Inert XL with TAD SQ - Source: EI - Inert Filament 2 (Qualitative - No setpoints associated)	None
October 1, 2021 12:50:59 PM	End	Execution	Tune EI - 6975C Inert XL with TAD SQ - Source: EI - Inert Filament 2 (Qualitative - No setpoints associated)	Run Count: 1

Page 3/7

Page 3 / 7

User Name: supasak.niamsongtham
 Hostname: SC0115HNC

System ID: GM-2
 Print Date: October 1, 2021 1:16:19 PM

ALS_GM2 Transaction log:

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 1, 2021 12:51:01 PM	Start	Execution	Scouting Run - Injection Tower, Front MM, SQ - Source: EI - Inert- Part of GCMS System Preparation	None
October 1, 2021 12:51:18 PM	Audit	Data	Scouting Run - Injection Tower, Front MM, SQ - Source: EI - Inert- Part of GCMS System Preparation	Data File Path : E:\GM202021\SCOUTING RUN\01.D\DATA.MS
October 1, 2021 12:51:42 PM	Audit	Data	Scouting Run - Injection Tower, Front MM, SQ - Source: EI - Inert- Part of GCMS System Preparation	Data File Path : E:\GM202021\SCOUTING RUN\01.D\DATA.MS
October 1, 2021 12:52:42 PM	Audit	Data	Scouting Run - Injection Tower, Front MM, SQ - Source: EI - Inert- Part of GCMS System Preparation	Data File Path : E:\GM202021\SCOUTING RUN\01.D\DATA.MS
October 1, 2021 12:53:25 PM	End	Execution	Scouting Run - Injection Tower, Front MM, SQ - Source: EI - Inert- Part of GCMS System Preparation	Run Count : 1
October 1, 2021 12:53:27 PM	Start	Execution	Signal to Noise EI - Injection Tower, Front MM, SQ - Source: EI - Inert using Filament 1 - L: >= 320	None
October 1, 2021 12:53:40 PM	Audit	Data	Signal to Noise EI - Injection Tower, Front MM, SQ - Source: EI - Inert using Filament 1 - L: >= 320	Data File Path : E:\GM202021\SNF1_001.D\DATA.MS
October 1, 2021 12:53:58 PM	End	Execution	Signal to Noise EI - Injection Tower, Front MM, SQ - Source: EI - Inert using Filament 1 - L: >= 320	Run Count : 1

Page 4 / 7

Page 4 / 7

Print Date: October 1, 2021 1:10:19 PM

Time	Transaction	Activity	Type of Transaction	Optional information
11:00

State		Performed	Data		Injection Precision - Injection		Data file Path :	
October 1, 2021 12:50:37 PM	Audit				Tower: Front MM1, SQ: -	E:\GM20Q02021WP_MRP007	E:\GM20Q02021WP_MRP007	
					Source: EI - Inert-L (Ave): <= 1.00%	Source: EI - Inert-L (Ave): <= 1.00%	DI\DATA.MS	
					5.00% - L (Rel. Time): <= 1.00%			
October 1, 2021 12:50:37 PM	Audit				Tower: Front MM1, SQ: -	E:\GM20Q02021WP_MRP008	E:\GM20Q02021WP_MRP008	
					Source: EI - Inert-L (Ave): <= 1.00%	Source: EI - Inert-L (Ave): <= 1.00%	DI\DATA.MS	
					5.00% - L (Rel. Time): <= 1.00%			
October 1, 2021 12:50:52 PM	End				Tower: Front MM1, SQ: -		Run Count: 1:1	
					Source: EI - Inert-L (Ave): <= 1.00%	Source: EI - Inert-L (Ave): <= 1.00%		
					5.00% - L (Rel. Time): <= 1.00%			
October 1, 2021 12:50:55 PM	Start				Mass Ratio Precision - Injection	None		
					Tower: Front MM1, SQ: -			
					Source: EI - Inert-L (RSD): <= 5.00%	Source: EI - Inert-L (RSD): <= 5.00%		
October 1, 2021 12:55:06 PM	Audit				Mass Ratio Precision - Injection		Data file Path :	
					Tower: Front MM1, SQ: -	E:\GM20Q02021WP_MRP003	E:\GM20Q02021WP_MRP003	
					Source: EI - Inert-L (RSD): <= 5.00%	Source: EI - Inert-L (RSD): <= 5.00%	DI\DATA.MS	
October 1, 2021 12:55:06 PM	Audit				Mass Ratio Precision - Injection		Data file Path :	
					Tower: Front MM1, SQ: -	E:\GM20Q02021WP_MRP004	E:\GM20Q02021WP_MRP004	
					Source: EI - Inert-L (RSD): <= 5.00%	Source: EI - Inert-L (RSD): <= 5.00%	DI\DATA.MS	
October 1, 2021 12:55:06 PM	Audit				Mass Ratio Precision - Injection		Data file Path :	
					Tower: Front MM1, SQ: -	E:\GM20Q02021WP_MRP005	E:\GM20Q02021WP_MRP005	
					Source: EI - Inert-L (RSD): <= 5.00%	Source: EI - Inert-L (RSD): <= 5.00%	DI\DATA.MS	
October 1, 2021 12:55:06 PM	Audit				Mass Ratio Precision - Injection		Data file Path :	
					Tower: Front MM1, SQ: -	E:\GM20Q02021WP_MRP006	E:\GM20Q02021WP_MRP006	
					Source: EI - Inert-L (RSD): <= 5.00%	Source: EI - Inert-L (RSD): <= 5.00%	DI\DATA.MS	
October 1, 2021 12:55:06 PM	Audit				Mass Ratio Precision - Injection		Data file Path :	
					Tower: Front MM1, SQ: -	E:\GM20Q02021WP_MRP008	E:\GM20Q02021WP_MRP008	
					Source: EI - Inert-L (RSD): <= 5.00%	Source: EI - Inert-L (RSD): <= 5.00%	DI\DATA.MS	

Page 617

User Name: supasak.narsongtham
Host Name: SCG1115HKC

System ID: GM-2
Print Date: October 1, 2021 1:10:17 PM

ALS_QM2 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 1, 2021 12:55:06 PM	Audit	Data	Mass Ratio Precision - Injection Tower: Front MM, SQ - Source: EI - Inlet - L (RSD): <= 5.00%	Data file Path: E:\QM2\Q2021\UP_MRP007.D\DATA.MS
October 1, 2021 12:55:06 PM	Audit	Data	Mass Ratio Precision - Injection Tower: Front MM, SQ - Source: EI - Inlet - L (RSD): <= 5.00%	Data file Path: E:\QM2\Q2021\UP_MRP008.D\DATA.MS
October 1, 2021 12:55:10 PM	End	Execution	Mass Ratio Precision - Injection Tower: Front MM, SQ - Source: EI - Inlet - L (RSD): <= 5.00%	Run Count: 1
October 1, 2021 12:55:13 PM	End	Qualification	Session	QC
October 1, 2021 12:55:13 PM	Start	Reporting	Session	None
October 1, 2021 1:05:11 PM	Audit	Reporting	Session	Report Generated: Certificate

Page 7 / 7

REVIEW BY: Autcharawan S.
APPROVED BY: Somnart M.
NEXT CAL. DATE: 12/30/22



Certificate of Calibration

ICS-2100: Anion (ID#659)

This certificate is to verify that instrument below are calibrated

by Archemica Lab Co., Ltd.

ICS-2100 S/N: 15010977

AS-HV S/N: 5450A36659

For

ALS Laboratory Group (Thailand) Co., Ltd.



Operator Signature: _____ Date: Jan 12, 2022

(Mr. Thitipong Piromkripuk)

Applications Chemist

Certificate of System Qualification

GC-OQ

System ID:	GC-6
Organization Name:	ALS Laboratory Group (Thailand) Co., Ltd.
Organization Location:	104 Phattanakan Rd., Suan Luang, Bangkok 10250
Date:	October 21, 2021 10:05:40 AM
EQP Name:	AgilentRecommended
EQP Revision:	GC.02.50
Overall Qualification Status:	Pass

REVIEW BY

APPROVED BY

NEXT CAL. DATE

Suchada T.

Sarat M.

21 Apr 2023

System Inspection and Basic Safety and Operation

Name:	7890
Setpoint Status:	Pass

Overall System Inspection and Basic Safety and Operation Test Status

Pass

Inlet Pressure Decay

Name:	7890
Front	SSL

Setpoint Status:	Pass				
Pressure:	25.0	psi			
Pressure Change:	0.0	psi	/5 minutes		
Agilent Recommended:	>=	-2.0	and	<=	0.5

Overall Inlet Pressure Decay Test Status

Pass

Inlet Pressure Accuracy

Name:	7890
Front	SSL

Date:	October 21, 2021 10:05:40 AM
System ID:	GC-6

Setpoint Status:

Setpoint	Pass	
Inlet Pressure:	25.0	psi
Actual	24.9	psi
Accuracy:	0.1	psi
Agilent Recommended:	<=	1.2

Overall Inlet Pressure Accuracy Test Status

Pass

Inlet Pressure Decay

Name:	7890
Back	SSL

Setpoint Status:	Pass				
Pressure:	25.0	psi			
Pressure Change:	0.0	psi	/5 minutes		
Agilent Recommended:	>=	-2.0	and	<=	0.5

Overall Inlet Pressure Decay Test Status

Pass

Inlet Pressure Accuracy

Name:	7890
Back	SSL

Setpoint Status:	Pass	
Setpoint	Pass	
Inlet Pressure:	25.0	psi
Actual	24.9	psi
Accuracy:	0.1	psi
Agilent Recommended:	<=	1.2

Overall Inlet Pressure Accuracy Test Status

Pass

Detector Flow Accuracy

Date:	October 21, 2021 10:05:40 AM
System ID:	GC-6

Name: 7890 Front FID

Setpoint Status:

Flow Type: Fuel
Setpoint: 30.0 mL/min Measured Flow: 30.5 mL/min
Accuracy: 0.5 mL/min
Agilent Recommended: ≤ 10.0 % setpoint (3.0 mL/min)
Limit is percentage of setpoint or 0.5 mL/minute, whichever is largest.

Setpoint Status:

Flow Type: Oxidizer
Setpoint: 400.0 mL/min Measured Flow: 394.0 mL/min
Accuracy: 6.0 mL/min
Agilent Recommended: ≤ 10.0 % setpoint (40.0 mL/min)
Limit is percentage of setpoint or 0.5 mL/minute, whichever is largest.

Setpoint Status:

Flow Type: Makeup
Setpoint: 25.0 mL/min Measured Flow: 24.2 mL/min
Accuracy: 0.8 mL/min
Agilent Recommended: ≤ 10.0 % setpoint (2.5 mL/min)
Limit is percentage of setpoint or 0.5 mL/minute, whichever is largest.

Overall Detector Flow Accuracy Test Status

Pass

Detector Flow Accuracy

Name: 7890 Back FID

Setpoint Status:

Flow Type: Fuel
Setpoint: 30.0 mL/min Measured Flow: 29.1 mL/min
Accuracy: 0.9 mL/min
Agilent Recommended: ≤ 10.0 % setpoint (3.0 mL/min)
Limit is percentage of setpoint or 0.5 mL/minute, whichever is largest.

Setpoint Status:

Flow Type: Oxidizer
Setpoint: 400.0 mL/min Measured Flow: 397.3 mL/min
Accuracy: 2.7 mL/min
Agilent Recommended: ≤ 10.0 % setpoint (40.0 mL/min)
Limit is percentage of setpoint or 0.5 mL/minute, whichever is largest.

Setpoint Status:

Flow Type: Makeup
Setpoint: 25.0 mL/min Measured Flow: 24.4 mL/min
Accuracy: 0.6 mL/min
Agilent Recommended: ≤ 10.0 % setpoint (2.5 mL/min)
Limit is percentage of setpoint or 0.5 mL/minute, whichever is largest.

Overall Detector Flow Accuracy Test Status

Pass

GC Oven Temperature Accuracy

Name: 7890

Setpoint Status: **Pass**

Zone: Oven

Temperature: Setpoint/Actual 230.0 231.5 °C

Accuracy: 1.5 °C

Agilent Recommended: >= -1.0 % setpoint in K (-5.0 °C)
<= 1.0 % setpoint in K (5.0 °C)

Setpoint Status: **Pass**

Zone: Oven

Temperature: Setpoint/Actual 100.0 100.5 °C

Accuracy: 0.5 °C

Agilent Recommended: >= -1.0 % setpoint in K (-3.7 °C)
<= 1.0 % setpoint in K (3.7 °C)

Overall GC Oven Temperature Accuracy Test Status

Pass

GC Oven Temperature Stability

Name: 7890

Setpoint Status: **Pass**

Setpoint/Average

Temperature: 100.0 100.4667 °C

Stability: 0.1 °C

Agilent Recommended: <= 0.5 °C

Overall GC Oven Temperature Stability Test Status

Pass

Scouting Run

Tested Combination1 Front SSL / Front FID

Injection Tower

Name: 7693A

Setpoint Status: **Completed**

Injection Volume on Column: 1.0 uL

Overall Scouting Run Status
Completed

Noise and Drift

Tested Combination1 Front SSL / Front FID

Name: 7890

Setpoint Status: **Pass**

Base Signal: 12.7 pA

ASTM Noise

Agilent Recommended: <= 0.10 pA

Status: Pass

Drift pA/Hr

<= 0.10

<= 2.50

Overall Noise and Drift Test Status

Pass

Injection Precision

Tested Combination1 Front SSL / Front FID

Name: 7693A

Setpoint Status: **Pass**

Injection Volume on Column: 1.0 uL

Area RSD: 0.42 %

Agilent Recommended: <= 3.00 %

Retention Time RSD: 0.16 %

<= 1.00

Overall Injection Precision Test Status

Pass

Signal to Noise

Tested Combination1 Front SSL / Front FID
Injection Tower
Name: 7890

Setpoint Status: Pass
Signal to Noise: 1174861
Agilent Recommended: >= 300000

Overall Signal to Noise Test Status
Pass

Scouting Run
Tested Combination2 Back SSL / Back FID
Injection Tower
Name: 7693A

Setpoint Status: Completed
Injection Volume on Column: 1.0 uL

Overall Scouting Run Status
Completed

Noise and Drift
Tested Combination2 Back SSL / Back FID
Name: 7890

Setpoint Status: Pass
Base Signal: 10.4 pA

ASTM Noise pA
0.05
0.10
Agilent Recommended: <= 0.10
Status: Pass

Drift pA/Hr
0.00
2.50
Agilent Recommended: <= 2.50
Status: Pass

Overall Noise and Drift Test Status
Pass

Injection Precision

Tested Combination2 Back SSL / Back FID
Name: 7693A

Setpoint Status: Pass
Injection Volume on Column: 1.0 uL

Area RSD: 1.16 %
Agilent Recommended: <= 3.00

Retention Time RSD: 0.12 %
<= 1.00

Overall Injection Precision Test Status
Pass

Signal to Noise

Tested Combination2 Back SSL / Back FID
Injection Tower
Name: 7890

Setpoint Status: Pass
Signal to Noise: 805466
Agilent Recommended: >= 300000

Overall Signal to Noise Test Status
Pass

Instrument Details

Purpose

This section describes the as found system configuration.

Details

System	
System ID	GC-6
Manufacturer	Agilent Technologies
Name	7890
Flow Data Input	Manual Data
Temperature Data Input	Manual Data or Other Data Logging
Tested Combination1	
Injection Technique	Injection Tower
Sampler Identifier	Sampler 2
Inlet	Front
Detector	Front
LTM Included?	No
Tested Combination2	
Injection Technique	Injection Tower
Sampler Identifier	Sampler 3
Inlet	Back
Detector	Back
LTM Included?	No
Sampler 1	
Manufacturer	Agilent Technologies
Type	Tray
Name	7693A
Model Number	G4514A
Serial Number	CN15380030
Firmware Revision	A.11.01
Vial Heater	Not installed

Sampler 2

Manufacturer	Agilent Technologies
Type	Injection Tower
Name	7693A
Model Number	G4513A
Serial Number	CN10340103
Firmware Revision	A.10.09
Usage	Sample Injection
Location	Front
Syringe Volume (µL)	10

Sampler 3

Manufacturer	Agilent Technologies
Type	Injection Tower
Name	7693A
Model Number	G4513A
Serial Number	CN16280128
Firmware Revision	A.10.09
Usage	Sample Injection
Location	Back
Syringe Volume (µL)	10

Mainframe 1

Manufacturer	Agilent Technologies
Name	7890
Model Number	G3440A
Serial Number	CN11461066
Firmware Revision	Version 4.27
Component ID/Asset No.	GC-6
Oven Type	Standard

Inlet 1

Manufacturer	Agilent Technologies
Name	7890
Type	SSL
Location	Front
Carrier Gas	Helium
Control Type	Electronic Pressure Control (EPC)
Purged Inlet	Yes

Inlet 2

Manufacturer	Agilent Technologies
Name	7890
Type	SSL
Location	Back
Carrier Gas	Helium
Control Type	Electronic Pressure Control (EPC)
Purged Inlet	Yes

Detector 1

Manufacturer	Agilent Technologies
Name	7890
Type	FID
Adapter	Capillary
Control Type	Electronic Pressure Control (EPC)
Location	Front
Makeup Gas	Nitrogen

Detector 2

Manufacturer	Agilent Technologies
Name	7890
Type	FID
Adapter	Capillary
Control Type	Electronic Pressure Control (EPC)
Location	Back
Makeup Gas	Nitrogen

Electronic Signature

Purpose

This signature page was created and published because the ACE sign-off action was executed, which is valid for the entire document, including attachments. The ACE sign-off is an electronic signature that requires two distinct identification components: unique username and personal password. The Agilent representative who has delivered this service understands the meaning and legal status of an electronic signature. As a trained official operator, the Agilent representative has a unique password and login to access ACE and electronically sign this document. (Other e-signatures can be applied to this document using a Document Content Management or other suitable method defined in your data access and control procedures.)

Details

Full Name of Signer:	Suriya Thongkaew
Logged On User Name:	suriya.thongkaew@non.agilent.com
Signature Creation Date:	October 21, 2021
Reason for Signature:	Executed protocol and published this original version of document

Regulatory Disclaimer

This document provides a protocol to verify and record instrument configuration and evidence of proper operation. It has been prepared from our interpretation of applicable regulations as well as industry best practices. The document is designed to provide an important component of a complete compliance package. Validation depends upon many factors and use of this protocol alone does not assure compliance. Agilent Technologies makes no promises or representations as to its sufficiency for any specific regulatory program.

Warranty

Agilent Technologies makes no warranty of any kind to this material, including but not limited to, the implied warranties or merchantability and fitness for a particular purpose. Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

User Name: suriyathongkaew
Hostname: ASBKKW7015
System Id: GC-6
Print Date: October 21, 2021 10:05:46 AM

OQ GC ALS CN11461066 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 20, 2021 12:16:50 PM	Audit	SessionCreated	Session	None
October 20, 2021 12:18:50 PM	Start	Configuration	Session	None
October 20, 2021 12:19:50 PM	Audit	Entitlement	Licensing	User is Nonpaying and does not require an unlock code.
October 20, 2021 12:24:57 PM	Audit	EqlLoaded	Session	EQP details for primary technique [Gc] - File path: (ProtocolPacks\Gc\Configurations\02.51\Gc-02.51.eqp) EQP File Name: (Gc-02.51.eqp) EQP Name: (AgilentRecommended)
October 20, 2021 12:25:02 PM	End	Configuration	Session	None
October 20, 2021 12:25:09 PM	Start	Qualification	Session	OQ
October 20, 2021 12:25:09 PM	Start	Execution	System Inspection and Basic Safety and Operation - 7890 - Qualitative Test - No setpoints associated	None
October 20, 2021 12:30:25 PM	End	Execution	System Inspection and Basic Safety and Operation - 7890 - Qualitative Test - No setpoints associated	Run Count : 1
October 20, 2021 12:58:29 PM	Start	Execution	Inlet Pressure Decay - Front SSL - Pressure Controlled Inlet - S: 25.0 psi - L: >= 2.0 psi and <= 0.5 psi	None

Page 1 / 10

User Name: suriyathongkaew
Hostname: ASBKKW7015
System Id: GC-6
Print Date: October 21, 2021 10:05:46 AM

OQ GC ALS CN11461066 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 20, 2021 1:02:16 PM	End	Execution	Inlet Pressure Decay - Front SSL - Pressure Controlled Inlet - S: 25.0 psi - L: >= 2.0 psi and <= 0.5 psi	Run Count : 1
October 20, 2021 1:02:18 PM	Start	Execution	Inlet Pressure Accuracy - Front SSL - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	None
October 20, 2021 1:02:26 PM	End	Execution	Inlet Pressure Accuracy - Front SSL - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	Run Count : 1
October 20, 2021 1:02:29 PM	Start	Execution	Inlet Pressure Decay - Back SSL - Pressure Controlled Inlet - S: 25.0 psi - L: >= 2.0 psi and <= 0.5 psi	None
October 20, 2021 1:04:21 PM	End	Execution	Inlet Pressure Decay - Back SSL - Pressure Controlled Inlet - S: 25.0 psi - L: >= 2.0 psi and <= 0.5 psi	Run Count : 1
October 20, 2021 1:07:53 PM	Start	Execution	Inlet Pressure Accuracy - Back SSL - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	None
October 20, 2021 1:08:11 PM	End	Execution	Inlet Pressure Accuracy - Back SSL - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	Run Count : 1
October 20, 2021 1:08:16 PM	Start	Execution	Detector Flow Accuracy - Front FID - Type: Fuel - S: 30.0 mL/min - L: <= 10.0% setpoint	None
October 20, 2021 1:20:23 PM	Audit	Data	Detector Flow Accuracy - Front FID - Type: Fuel - S: 30.0 mL/min - L: <= 10.0% setpoint	Manual Data Entry
October 20, 2021 1:20:26 PM	End	Execution	Detector Flow Accuracy - Front FID - Type: Fuel - S: 30.0 mL/min - L: <= 10.0% setpoint	Run Count : 1

Page 2 / 10

User Name: suriyathongkiew
Hostname: ASBKW7015

System Id: GC-6
Print Date: October 21, 2021 10:05:46 AM

QQ GC ALS CN11461066 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 20, 2021 1:20:29 PM	Start	Execution	Detector Flow Accuracy - Front FID - Type: Oxidizer - S: 400.0 mL/min - L <= 10.0% setpoint	None
October 20, 2021 1:23:27 PM	Audit	Data	Detector Flow Accuracy - Front FID - Type: Oxidizer - S: 400.0 mL/min - L <= 10.0% setpoint	Manual Data Entry
October 20, 2021 1:23:29 PM	End	Execution	Detector Flow Accuracy - Front FID - Type: Oxidizer - S: 400.0 mL/min - L <= 10.0% setpoint	Run Count: 1
October 20, 2021 1:23:31 PM	Start	Execution	Detector Flow Accuracy - Front FID - Type: Makeup - S: 25.0 mL/min - L <= 10.0% setpoint	None
October 20, 2021 1:27:40 PM	Audit	Data	Detector Flow Accuracy - Front FID - Type: Makeup - S: 25.0 mL/min - L <= 10.0% setpoint	Manual Data Entry
October 20, 2021 1:27:42 PM	End	Execution	Detector Flow Accuracy - Front FID - Type: Makeup - S: 25.0 mL/min - L <= 10.0% setpoint	Run Count: 1
October 20, 2021 1:27:46 PM	Start	Execution	Detector Flow Accuracy - Back FID - Type: Fuel - S: 30.0 mL/min - L <= 10.0% setpoint	None
October 20, 2021 1:32:10 PM	Audit	Data	Detector Flow Accuracy - Back FID - Type: Fuel - S: 30.0 mL/min - L <= 10.0% setpoint	Manual Data Entry
October 20, 2021 1:32:12 PM	End	Execution	Detector Flow Accuracy - Back FID - Type: Fuel - S: 30.0 mL/min - L <= 10.0% setpoint	Run Count: 1
October 20, 2021 1:32:14 PM	Start	Execution	Detector Flow Accuracy - Back FID - Type: Oxidizer - S: 400.0 mL/min - L <= 10.0% setpoint	None
October 20, 2021 1:34:13 PM	Audit	Data	Detector Flow Accuracy - Back FID - Type: Oxidizer - S: 400.0 mL/min - L <= 10.0% setpoint	Manual Data Entry

Page 3 / 10

User Name: suriyathongkiew
Hostname: ASBKW7015

System Id: GC-6
Print Date: October 21, 2021 10:05:46 AM

QQ GC ALS CN11461066 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 20, 2021 1:34:16 PM	End	Execution	Detector Flow Accuracy - Back FID - Type: Oxidizer - S: 400.0 mL/min - L <= 10.0% setpoint	Run Count: 1
October 20, 2021 1:34:46 PM	Start	Execution	Detector Flow Accuracy - Back FID - Type: Makeup - S: 25.0 mL/min - L <= 10.0% setpoint	None
October 20, 2021 1:36:33 PM	Audit	Data	Detector Flow Accuracy - Back FID - Type: Makeup - S: 25.0 mL/min - L <= 10.0% setpoint	Manual Data Entry
October 20, 2021 1:36:36 PM	End	Execution	Detector Flow Accuracy - Back FID - Type: Makeup - S: 25.0 mL/min - L <= 10.0% setpoint	Run Count: 1
October 20, 2021 1:36:38 PM	Start	Execution	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	None
October 20, 2021 2:04:31 PM	Audit	Data	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Manual Data Entry
October 20, 2021 2:04:32 PM	End	Execution	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Run Count: 1
October 20, 2021 2:04:34 PM	Start	Execution	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	None
October 20, 2021 2:10:47 PM	Audit	Data	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Manual Data Entry

Page 4 / 10

User Name: suriyathongkiew
Hostname: ASBKKW7015

System ID: GC-6
Print Date: October 21, 2021 10:05:46 AM

OQ GC ALS CNI1461066 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 20, 2021 2:10:49 PM	End	Execution	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % slipoint In K	Run Count : 1
October 20, 2021 2:10:51 PM	Start	Execution	GC Oven Temperature Stability - 7890 - Temperature : Oven - S: 100.0°C - L: <= 0.5°C	None
October 20, 2021 2:31:39 PM	Audit	Data	GC Oven Temperature Stability - 7890 - Temperature : Oven - S: 100.0°C - L: <= 0.5°C	Manual Data Entry
October 20, 2021 2:31:41 PM	End	Execution	GC Oven Temperature Stability - 7890 - Temperature : Oven - S: 100.0°C - L: <= 0.5°C	Run Count : 1
October 21, 2021 2:31:44 PM	Start	Execution	GC Scouting Run - Injection Tower, Front SSL, Front FID - Part of System Preparation - No limits associated	None
October 20, 2021 2:43:06 PM	Audit	AccClosed	Session	None
October 21, 2021 8:18:59 AM	Audit	AccRestarted	Session	None
October 21, 2021 8:19:02 AM	Audit	SessionReloaded	Session	None
October 21, 2021 8:19:09 AM	Start	Qualification	Session	OQ
October 21, 2021 8:19:09 AM	Start	Execution	GC Scouting Run - Injection Tower, Front SSL, Front FID - Part of System Preparation - No limits associated	None
October 21, 2021 8:19:41 AM	Audit	AccClosed	Session	None

Page 5 / 10

User Name: suriyathongkiew
Hostname: ASBKKW7015

System ID: GC-6
Print Date: October 21, 2021 10:05:46 AM

OQ GC ALS CNI1461066 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 21, 2021 9:20:08 AM	Audit	AccRestarted	Session	None
October 21, 2021 9:20:09 AM	Audit	SessionReloaded	Session	None
October 21, 2021 9:20:13 AM	Start	Qualification	Session	OQ
October 21, 2021 9:20:13 AM	Start	Execution	GC Scouting Run - Injection Tower, Front SSL, Front FID - Part of System Preparation - No limits associated	None
October 21, 2021 9:29:45 AM	Audit	Data	GC Scouting Run - Injection Tower, Front SSL, Front FID - Part of System Preparation - No limits associated	Data files Path : C:\Chem32\1\DATA\AQGPV20_21\OQPV2021_F_2021-10-20 15:49:01\SCOUT_F001.D\F1 DIA.ch
October 21, 2021 9:30:05 AM	End	Execution	GC Scouting Run - Injection Tower, Front SSL, Front FID - Part of System Preparation - No limits associated	Run Count : 1
October 21, 2021 9:30:08 AM	Start	Execution	Noise and Drift - Front FID - Detector FID - L (Noise) <= 0.10 pA - L (Drift) <= 2.50 pA/hour	None
October 21, 2021 9:30:41 AM	Audit	Data	Noise and Drift - Front FID - Detector FID - L (Noise) <= 0.10 pA - L (Drift) <= 2.50 pA/hour	Data files Path : C:\Chem32\1\DATA\AQGPV20_21\OQPV2021_F_2021-10-20 15:49:01\SIGNSDRF_F001.D\F1 DIA.ch
October 21, 2021 9:31:10 AM	End	Execution	Noise and Drift - Front FID - Detector FID - L (Noise) <= 0.10 pA - L (Drift) <= 2.50 pA/hour	Run Count : 1

Page 6 / 10

User Name: suriyathongkiew
Hostname: ASBKKW7015

System ID: GC-6
Print Date: October 21, 2021 10:05:46 AM

OQ GC ALS CN11461066 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 21, 2021 9:31:42 AM	Start	Execution	Injection Precision - Injection Tower, Front SSL, Front FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	None
October 21, 2021 9:32:55 AM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Front FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Data files Path : C:\Chem321\DATA\AOQP\20 21\OQP\2021_F 2021-10-20 16-51-16\NUPREC_F002.D\N FIDIA.ch
October 21, 2021 9:32:55 AM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Front FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Data files Path : C:\Chem321\DATA\AOQP\20 21\OQP\2021_F 2021-10-20 16-51-16\NUPREC_F003.D\N FIDIA.ch
October 21, 2021 9:32:56 AM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Front FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Data files Path : C:\Chem321\DATA\AOQP\20 21\OQP\2021_F 2021-10-20 16-51-16\NUPREC_F004.D\N FIDIA.ch
October 21, 2021 9:32:56 AM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Front FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Data files Path : C:\Chem321\DATA\AOQP\20 21\OQP\2021_F 2021-10-20 16-51-16\NUPREC_F005.D\N FIDIA.ch
October 21, 2021 9:32:56 AM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Front FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Data files Path : C:\Chem321\DATA\AOQP\20 21\OQP\2021_F 2021-10-20 16-51-16\NUPREC_F006.D\N FIDIA.ch
October 21, 2021 9:32:56 AM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Front FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Data files Path : C:\Chem321\DATA\AOQP\20 21\OQP\2021_F 2021-10-20 16-51-16\NUPREC_F007.D\N FIDIA.ch

Page 7 / 10

User Name: suriyathongkiew
Hostname: ASBKKW7015

System ID: GC-6
Print Date: October 21, 2021 10:05:46 AM

OQ GC ALS CN11461066 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 21, 2021 9:33:07 AM	End	Execution	Injection Precision - Injection Tower, Front SSL, Front FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Run Count : 1
October 21, 2021 9:35:23 AM	Start	Execution	Signal to Noise - Injection Tower, Front SSL, Front FID - Detector FID - L >= 300000	None
October 21, 2021 9:34:01 AM	Audit	Data	Signal to Noise - Injection Tower, Front SSL, Front FID - Detector FID - L >= 300000	Data files Path : C:\Chem321\DATA\AOQP\20 21\OQP\2021_F 2021-10-20 16-51-16\SIGTONS_F001.D\N FIDIA.ch
October 21, 2021 9:34:15 AM	End	Execution	Signal to Noise - Injection Tower, Front SSL, Front FID - Detector FID - L >= 300000	Run Count : 1
October 21, 2021 9:34:19 AM	Start	Execution	GC Scouting Run - Injection Tower, Back SSL, Back FID - Part of System Preparation - No limits associated	None
October 21, 2021 9:35:04 AM	Audit	Data	GC Scouting Run - Injection Tower, Back SSL, Back FID - Part of System Preparation - No limits associated	Data files Path : C:\Chem321\DATA\AOQP\20 21\OQP\2021_B 2021-10-20 17-19-49\SCOUT_B001.D\FI DZB.ch
October 21, 2021 9:35:27 AM	End	Execution	GC Scouting Run - Injection Tower, Back SSL, Back FID - Part of System Preparation - No limits associated	Run Count : 1
October 21, 2021 9:35:32 AM	Start	Execution	Noise and Drift - Back FID - Detector FID - L (Noise) <= 0.10 pA - L (Drift) <= 2.50 pA/hour	None

Page 8 / 10

User Name: suriyathongkhaew
Hostname: ASBKW7015

System Id: GC-6
Print Date: October 21, 2021 10:05:46 AM

QQ GC ALS CN11461066 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 21, 2021 9:36:06 AM	Audit	Data	Noise and Drift - Back FID - Detector FID - L (Noise) <= 0.10 pA - L (Drift) <= 2.50 pA/hour	Data files Path : C:\Chem32\1\DATA\AQOPV20 21\OQP\2021_B 2021-10-20 17-13-46\SIGNSOFF_B001.D DFID2B.ch
October 21, 2021 9:36:16 AM	End	Execution	Noise and Drift - Back FID - Detector FID - L (Noise) <= 0.10 pA - L (Drift) <= 2.50 pA/hour	Run Count : 1
October 21, 2021 9:36:20 AM	Start	Execution	Injection Precision - Injection Tower, Back SSL, Back FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	None
October 21, 2021 9:36:57 AM	Audit	Data	Injection Precision - Injection Tower, Back SSL, Back FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Data files Path : C:\Chem32\1\DATA\AQOPV20 21\OQP\2021_B 2021-10-20 17-13-46\INUPREC_B002.D FID2B.ch
October 21, 2021 9:36:57 AM	Audit	Data	Injection Precision - Injection Tower, Back SSL, Back FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Data files Path : C:\Chem32\1\DATA\AQOPV20 21\OQP\2021_B 2021-10-20 17-13-46\INUPREC_B003.D FID2B.ch
October 21, 2021 9:36:57 AM	Audit	Data	Injection Precision - Injection Tower, Back SSL, Back FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Data files Path : C:\Chem32\1\DATA\AQOPV20 21\OQP\2021_B 2021-10-20 17-13-46\INUPREC_B004.D FID2B.ch
October 21, 2021 9:36:57 AM	Audit	Data	Injection Precision - Injection Tower, Back SSL, Back FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Data files Path : C:\Chem32\1\DATA\AQOPV20 21\OQP\2021_B 2021-10-20 17-13-46\INUPREC_B005.D FID2B.ch

Page 9 / 10

User Name: suriyathongkhaew
Hostname: ASBKW7015

System Id: GC-6
Print Date: October 21, 2021 10:05:46 AM

QQ GC ALS CN11461066 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 21, 2021 9:38:57 AM	Audit	Data	Injection Precision - Injection Tower, Back SSL, Back FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Data files Path : C:\Chem32\1\DATA\AQOPV20 21\OQP\2021_B 2021-10-20 17-13-45\INUPREC_B006.D FID2B.ch
October 21, 2021 9:38:57 AM	Audit	Data	Injection Precision - Injection Tower, Back SSL, Back FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Data files Path : C:\Chem32\1\DATA\AQOPV20 21\OQP\2021_B 2021-10-20 17-13-45\INUPREC_B007.D FID2B.ch
October 21, 2021 9:39:06 AM	End	Execution	Injection Precision - Injection Tower, Back SSL, Back FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Run Count : 1
October 21, 2021 9:39:11 AM	Start	Execution	Signal to Noise - Injection Tower, Back SSL, Back FID - Detector FID - L >= 300000	None
October 21, 2021 9:39:28 AM	Audit	Data	Signal to Noise - Injection Tower, Back SSL, Back FID - Detector FID - L >= 300000	Data files Path : C:\Chem32\1\DATA\AQOPV20 21\OQP\2021_B 2021-10-20 17-13-45\SIGTONS_9001.D FID2B.ch
October 21, 2021 9:39:39 AM	End	Execution	Signal to Noise - Injection Tower, Back SSL, Back FID - Detector FID - L >= 300000	Run Count : 1
October 21, 2021 9:39:43 AM	End	Qualification	Session	QQ
October 21, 2021 9:39:43 AM	Start	Reporting	Session	None
October 21, 2021 10:04:15 AM	Audit	Reporting	Session	Report Generated : Certificate

Page 10 / 10

ภาคผนวก จ-2

เอกสารสอบเทียบเครื่องมือ

บริษัท ยูไนเต็ด แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด

List of Instruments Certification for Air & Noise Quality Analysis

No.	Instrument/Equipment	Parameter	Manufacturer	Model/Serial No.	Calibrator	Certification No.	Date of Calibration	Due date of Calibration	Remark
Stack									
1	Pre-Test Console	Total Suspended Particulate Hydrogen Chloride Dioxin/Furans-TEQ Mercury Cadmium Lead Arsenic Beryllium Chromium	Apex Instruments, USA.	XC-572-V 0807047	Envi Equipment Service Co., Ltd.	E21-0813	19 Aug 21	18 Aug 22	-
2	Flue gas Analyzer	Sulphur Dioxide Oxide of Nitrogen as Nitrogen Dioxide Carbon Monoxide	Testo	Testo 350 60899456	Entech Industrial Sulation Co., Ltd.	G 640441	5 Aug 21	4 Aug 22	-

List of Opacity Training Certification for Opacity Mesurement

No.	Name	Training Couse	Train	Date	Remark
1	Mr.Apiwich Toungee	Opacity	Pollution Control Department	12-13 March 2015	-
2	Mr.Ronnapob Putragulpattana	Opacity	Pollution Control Department	22-23 March 2018	-

List of Instruments Certification for Water Quality Analysis

No.	Instrument/Equipment	Parameter	Manufacturer	Model/Serial No.	Calibrator	Certification No.	Date of Calibration	Due date of Calibration	Remark
Water									
1	pH Meter	pH	Horiba	LAQUA-PH210 HA0C0025	Technology Promotion Association (Thailand-Japan)	22CH639	11 May 22	10 May 23	-
2	DO Meter	DO	Horiba	LAQUA-DO210 HE9M0013	Technology Promotion Association (Thailand-Japan)	22TW73	18 Mar 22	17 Mar 23	-
3	Conductivity Meter	Conductivity	Horiba	LAQUA-EC210 HC9L0014	Technology Promotion Association (Thailand-Japan)	22CH580	27 Apr 24	26 Apr 23	-

Enviro Equipment Service Co., Ltd.

110/254 Moo 3, Tambon Bang Rak Phathana, Amphur Bang Bua Thong, Nonthaburi 11110

Tel. 098 362 9152, 089 478 7885

E-mail: sales@enviro-ees.com

Certificate No. : E21-0813

Page : 1 of 6

CERTIFICATE OF CALIBRATION

Customer : United Analyst and Engineering Consultant Co., Ltd.

Address : 81 Soi Udomsuk 41, Sukhumvit Road, Bangehak, Phrakhanong, Bangkok 10260

Description of Equipment : Console meter

Manufacturer : Apex Instrument

Model Number : XC-572-V

Serial Number : 0807047

ID./Control No. : -

Environment Conditions : Temperature (25 ± 2) °C

: Humidity (50 ± 15) % RH

Cal. Date : 19/08/2021

Issue Date : 19/08/2021

Calibration Method or Calibration Procedure Used

US EPA Method (United State Environmental Protection Agency)

This certificate is traceable to national standard, which realize the units of measurement according to the International System of Units (SI).

Result of Calibration

This certificate may not be reproduced other than in full except with prior Written approval of the Technical Manager, Enviro Equipment Service Company Limited.

These reported uncertainties of measurement are expanded by a coverage factor of k=2, providing a 95% confidence level

Calibrated by : Mr. Sanya Sangnil

Approved by :

(Mr. Mana Fuekhud)
Technical Manager

เอกสารไม่ควบคุม

Certificate No. : E21-0813

Page : 2 of 6

METHOD 5 CONSOLE CALIBRATION USING REFERENCE WET GAS METER W-NK-2.5-B-Z No.547425 5-POINT METRIC UNIT

Meter Console Information		Calibration Conditions		Factors/Conversions	
Console Model Number	XC-572-V	Date	19/8/2021	Std Temp	293 K
Console Serial Number	0807047	Calibration Reference No.	-	Std Press	760 mm Hg
DGM Model Number	SK25EX	Barometric Pressure	761.00	K _i	0.386
DGM Serial Number	00003580	Calibration Meter Gamma	0.999	Console Leak Check	
				PASS	

Calibration Data											
Metering Console				Calibration Meter							
Run Time	DGM Orifice	Volume	Volume	Volume	Volume	Volume	Volume	Volume	Volume	Volume	Volume
Elapsed (Q)	(P _m)	Initial	Final	Initial	Final	Initial	Final	Initial	Final	Initial	Final
(Q)	(P _m)	(V _m)	(V _{ref})	(V _{ref})	(V _{ref})	(V _{ref})	(V _{ref})	(V _{ref})	(V _{ref})	(V _{ref})	(V _{ref})
min	mm H ₂ O	m ³	m ³	m ³	m ³	m ³	m ³	m ³	m ³	m ³	m ³
12.17	13.0	3.3600	3.5000	23	23	23.01236	23.15126	26	26	26	26
12.30	13.0	3.5000	3.6400	23	23	23.15126	23.28948	26	26	26	26
8.72	26.0	3.6490	3.7890	24	24	23.29826	23.43644	26	26	26	26
8.88	26.0	3.7890	3.9290	24	24	23.43644	23.57434	26	26	26	26
14.83	40.0	3.9350	4.2150	25	25	23.58012	23.85818	26	26	26	26
14.67	40.0	4.2150	4.4950	25	25	23.85818	24.13456	26	26	26	26
10.37	70.0	4.5080	4.7880	25	25	24.14158	24.41250	26	26	26	26
10.13	70.0	4.7880	5.0680	26	26	24.41250	24.68290	26	26	26	26
8.90	90.0	5.0750	5.3550	27	27	24.68856	24.96012	26	26	26	26
8.88	90.0	5.3550	5.6350	27	27	24.96012	25.23050	26	26	26	26

เอกสารไม่ควบคุม

Certificate No. : E21-0813
Page : 3 of 6

METHOD 5 CONSOLE CALIBRATION
USING REFERENCE WET GAS METER W-NK-2.5-B-Z No.547425
5-POINT METRIC UNIT

Calibration Data									
Results									
Standardized Data				Dry Gas Meter					
Dry Gas Meter		Calibration Meter		Calibration Factor		Flowrate		Variation	
(V _{gas})	(Q _{gas})	(V _{wet})	(Q _{wet})	Value	Variation	Std & Corr	.0212 m ³ /min	Std & Corr	Variation
m ³	m ³ /min	m ³	m ³ /min	(Y)	(ΔY)	(Q _{std} (corr))	(ΔH _g)	(Q _{std} (corr))	(ΔH _g)
0.138	0.011	0.136	0.011	0.990	0.015	0.011	46.043	0.011	-2.056
0.138	0.011	0.135	0.011	0.985	0.010	0.011	47.522	0.011	-0.577
0.138	0.016	0.135	0.016	0.984	0.009	0.016	47.880	0.016	-0.219
0.138	0.016	0.135	0.015	0.982	0.007	0.015	49.931	0.015	1.832
0.276	0.019	0.273	0.018	0.988	0.014	0.018	52.821	0.018	4.722
0.276	0.019	0.271	0.018	0.982	0.008	0.018	52.270	0.018	4.171
0.277	0.027	0.266	0.026	0.960	-0.014	0.026	47.835	0.026	-0.264
0.277	0.027	0.265	0.026	0.958	-0.016	0.026	45.881	0.026	-2.217
0.277	0.031	0.266	0.030	0.961	-0.014	0.030	45.290	0.030	-2.809
0.277	0.031	0.265	0.030	0.956	-0.018	0.030	45.515	0.030	-2.583
				0.975	Y Average			48.099	DH@ Average

Note: For Calibration Factor Y, the ratio of the reading of the calibration meter to the dry gas meter, acceptable tolerance of individual values from the average is ±0.02.
For ΔH_g, orifice pressure differential that equates to 0.75 cfm (0.0212 m³/min) at standard temperature and pressure, acceptable tolerance of individual values from the average is ±0.2 inches (5.1mm) H₂O.

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม

Certificate No. : E21-0813
Page : 4 of 6



Calibration Date: 19-8-2021

Calibration Reference No: Ser21-0813

Meter Pressure vs Flowrate



Console Serial: 0807047

Console Model: XC-572-V

THERMOCOUPLES SYSTEM CALIBRATION

Sampling System Equipment Information	
Console Model Number	XC-572-V
Console Serial Number	0807047
DGM Model Number	SK25EX
DGM Serial Number	00003580
Meter Box Model Number	JENCO 765 KF
Meter Box Serial Number	JC 17073

Calibration Conditions	
Date	19/8/2021 03:00 PM
Calibration Reference No.	
Reference Thermometer	DIGICON
Serial Number	183169105

Results

Console Thermocouple Simulator

Channel and test point	Meter Box Channel Temperature Reading (°C)									
	-18.0	25.0	38.0	93.0	149.0	260.0	371.0	482.0	593.0	816.0
Stack	-17.0	24.0	37.0	93.0	149.0	259.0	371.0	482.0	594.0	817.0
Aux	-17.0	24.0	37.0	93.0	149.0					
Probe	-18.0	24.0	37.0	93.0	149.0					
Filter	-18.0	24.0	37.0	93.0	149.0					
Oven	-	-	-	-	-					
Exit	-18.0	24.0	37.0							

Tolerance Range

	Stack	Probe	Filter	± 1.50%	± 3.0 °C	± 3.0 °C	Absolute	Meter Exit	± 3.0 °C	± 2.0 °C
--	-------	-------	--------	---------	----------	----------	----------	------------	----------	----------



Certificate No: G 640441
Date of issue : 05-Aug-21

Instrument description : Flue gas Analyzer
Instrument model : Testo 350 New
Instrument serial no. : 60899456
ID no. or control no. : UAE.EFM.005/2560
Manufacturer : testo SE
Probe description : -
Probe model : -
Probe serial : -
Customer name : UNITED ANALYST AND ENGINEERING CONSULTANT CO.,LTD.
Customer address : 81 SOI UDOMSU41,SUKHUMVIT ROAD,BANGCHAK PRAKANONG
BANGKOK 10260
Total pages of certificate : 2 Pages
Receiving no. : L-211963
Receiving date. : 14-Jul-21
Parameter of calibration : Gas Calibration(Oxygen 2.501,10.00,21.00 %vol, Carbon Monoxide 80.23,309.9,1003 ppm,
Nitric Oxide 10.08,150.9,320.6 ppm, Sulphur Dioxide 50.04,100.9,601.1 ppm,
Nitrogen Dioxide 10.20,80.62,202.2 ppm)
Condition of UUC. : Used
Ambient condition : All of the Measurement were carried out the stabilized laboratory
Temperature : 23 ± 5 °C
Humidity : 55 ± 15 %RH
Calibration place : 17/121 Soi Ngamwongwan 47 Yaek 48, Toongsonghong, Lakso, Bangkok 10210
Calibration procedure no. : WI-CL-28-C

*The calibration certificate expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.
This certificate is applied only to item under test Environmental condition.
This Calibration Certificate may not be reproduced other than in full except with the permission of the issuing laboratory.
Calibration certificates without signature and seal not valid.
This calibration certificate documents are traceability to national standards, which realize measurement according to the International System of Units (SI).*

Date of calibration : 04-Aug-21

Kumthai

Mr. Kwanchai Khandoung
Calibration Technician

D. Wuthu

Mrs. Nongluck Wongsettee
Technical Manager



Certificate No.: G 640441

Standard References (Table 1)

Standard	Certificate No.	Vendor	Due date
Oxygen (O2) 2.501 % Vol	24531/19	Linde	16-Jul-23
Oxygen (O2) 10.00 % Vol	2453/19	Linde	18-Jul-23
Oxygen (O2) 21.00 % Vol	2426/19	Linde	16-Jul-23
Carbon monoxide (CO) 80.97 ppm	2842/21	Linde	24-Jun-23
Carbon monoxide (CO) 309.9 ppm	2803/21	Linde	22-Jun-23
Carbon monoxide (CO) 1003 ppm	2829/21	Linde	23-Apr-23
Nitric Oxide (NO) 10.08 ppm	3241/21	Linde	25-Jul-23
Nitric Oxide (NO) 150.9 ppm	2857/21	Linde	27-Jun-23
Nitric Oxide (NO) 320.6 ppm	2944/21	Linde	02-Jul-23
Sulphur Dioxide (SO2) 50.04 ppm	3205/21	Linde	25-Jul-23
Sulphur Dioxide (SO2) 100.9 ppm	4942/20	Linde	20-Nov-22
Sulphur Dioxide (SO2) 601.1 ppm	3204/21	Linde	20 Jul 23
Nitrogen Dioxide (NO2) 10.20 ppm	2929/19	Linde	27-Aug-21
Nitrogen Dioxide (NO2) 80.62 ppm	3240/21	Linde	25-Jul-23
Nitrogen Dioxide (NO2) 202.2 ppm	3239/21	Linde	20-Jul-23

Measured room conditions

Temperature : 23.2 °C Humidity : 53.8 %RH Pressure : 1016.3 mbar

Calibration conditions

Gas Temperature : 23 °C Flow rate : 1,100 ml/min Gas pressure : 1021.6 mbar

Calibration Results (without adjustment) (Table 2)

Parameter of Standard	Standard Values	Mean of UUC	Error	Uncertainty (±)
O2 (%Vol)	2.501	2.47	-0.031	0.20
O2 (%Vol)	10.00	9.86	-0.14	0.40
O2 (%Vol)	21.00	21.14	0.14	0.80
CO (ppm)	80.97	82	1.03	2.8
CO (ppm)	309.9	310	0.1	11
CO (ppm)	1003	999	-4	34
NO (ppm)	10.08	9	-1.08	3.0
NO (ppm)	150.9	151	0.1	5.0
NO (ppm)	320.6	322	1.4	10
SO2 (ppm)	50.04	49	-1.04	5.0
SO2 (ppm)	100.9	101	0.1	5.0
SO2 (ppm)	601.1	599	-2.1	14
NO2 (ppm)	10.20	9.9	-0.30	1.5
NO2 (ppm)	80.62	80.3	-0.32	5.0
NO2 (ppm)	202.2	198.9	-3.3	5.0

Remark : 1 cmol/mol = 1 %vol , 1 µmol/mol = 1 ppm.

End of Report



กรมควบคุมมลพิษ
POLLUTION CONTROL DEPARTMENT

ขอมอบประกาศนียบัตรนี้เพื่อแสดงว่า

นายรณภพ ภู่อระกุลพัฒนา

ได้ผ่านการฝึกอบรมและทดสอบผู้ตรวจวัดความทึบแสงของเขม่าควันด้วยสายตา
และการใช้แผนภูมิเขม่าควันของริงเกิลมานน์ ประจำปี ๒๕๖๑

ระหว่างวันที่ ๒๒ - ๒๓ มีนาคม ๒๕๖๑

จัดโดย สำนักจัดการคุณภาพอากาศและเสียง กรมควบคุมมลพิษ
ให้ไว้ ณ วันที่ ๒๓ มีนาคม ๒๕๖๑

(นางสุณี ปิยะพันธุ์พงศ์)

อธิบดีกรมควบคุมมลพิษ

ขอมอบประกาศนียบัตรนี้เพื่อแสดงว่า

นายอภิวิชญ์ ท่วงที

ได้ผ่านการฝึกอบรมและทดสอบผู้ตรวจวัดความทึบแสงของเขม่าควันด้วยสายตา
และการใช้แผนภูมิเขม่าควันของริงเกิลมานน์

ระหว่างวันที่ ๑๒ - ๑๓ มีนาคม ๒๕๕๘

จัดโดย สำนักจัดการคุณภาพอากาศและเสียง กรมควบคุมมลพิษ

ให้ไว้ ณ วันที่ ๑๓ มีนาคม ๒๕๕๘

(นายวิเชียร จุ่งรุ่งเรือง)

อธิบดีกรมควบคุมมลพิษ



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert.No.: 22CH639
Page.: 1 of 3

Certificate of Calibration

Equipment : pH Meter
Manufacturer : Horiba
Model : LAQUA-PH210
Serial No. : HAOC0025
ID No. : UAE EFM.117/2563 (ENV.pH.07/63)
Condition As-Received:
Used Item
Received Date : 10 May 2022
Calibration Date : 11 May 2022
Reference : 2205-0278WSC-1
Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
3 Soi Udomsuk 41, Sukhumvit Road, Bangchak,
Phrakhanong, Bangkok 10260
Ambient Temperature : (25 ± 2.5) °C
Relative Humidity : (50 ± 15) %
Calibration Procedure : In - house method :
- CP-CH5 by direct measurement with standard
voltage calibrator and direct measurement with
certified reference material (CRM)
- CP-CH8 by comparison with standard thermometer

Calibrated by : Uthen Kankawi


Approved by : Malee
Approved Signatory

(✓) Malee Bulkruea
() Sathip Meangmai
() Warakorn Lerngagrakul

Issue Date : 17 May 2022

The Uncertainties are for a confidence probability of approximately 95 %

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.



เอกสารไม่ควบคุม

A 0041145

เอกสารไม่ควบคุม



Cert.No.: 22CH639
Page.: 2 of 3

Condition of this calibration result

1. Reference Standard Instrument : -

Instrument	Serial No.	ID No.	Cert. No.	Due Date
1) Document Process Calibrator	46530031	130RC098	21E3245	07 Oct 2022
2) Ref. Standard Thermometer	2188080	130RC044	2111273	21 Nov 2022

This certification is traceable to the International System of Unit maintained at:-
- Traceable to National Institute of Metrology (Thailand), NIMT

2. Certified Reference Materials : The measurement results are traceable to SI through CPA chem Ltd.,
ANSI-ASQ National Accreditation Board, Accredited No. AR-1835

Buffer Solution

Buffer Solution	Manufacturer	Lot No.	Exp. date
pH 4.008	CPA chem	788995	01 Jan 2024
pH 6.983	CPA chem	766822	04 Sep 2022
pH 10.015	CPA chem	794124	14 Feb 2023

3. This certificate is valid only to the item calibrated on date and place of calibration.

Calibration Results

Function : mV Measurement

Performing standard curve by Fluke at pH (4,7)(7,10)

Unit Under Calibration	Nominal Value	Standard Voltage Input	Actual Reading		Uncertainty of Measurement (\pm mV)	Coverage factor k
			mV	pH		
pH Meter S/N.: HA0C0025	pH					
	4.00	177.48	177.5	4.00	0.058	2.00
	7.00	0.00	0.0	7.00	0.058	2.00
	10.00	-177.48	-177.5	10.00	0.058	2.00

เอกสารไม่ควบคุม
a 1108310

๓๓๓



Cert.No.: 22CH639
Page.: 3 of 3

Calibration Results

Function : pH Measurement

Performing three buffers standard curve by using buffer nominal pH (4,7)(7,10)

Unit Under Calibration	Standard pH Buffer Solution	Actual pH Reading	Actual mV Reading (mV)	Uncertainty of pH measurement (\pm)	Coverage factor k
pH Electrode S/N.: 990C0199	4.008	4.00	131.4	0.0085	2.05
	6.983	6.98	-43.5	0.015	2.15
	6.983	6.98	-43.5	0.014	2.13
	10.015	10.02	-215.1	0.0096	2.00

Function : Temperature Measurement

(*) Without adjustment

This equipment was connected with Temperature Probe;

- Model : 9652
- Serial No. : 990C0199
- Dimension of probe;
 - Length : 95 mm.
 - Diameter : 12 mm.
 - Immersion Depth : 100 mm.

Calibration Point (°C)	Standard Temperature (°C)	UUC* Reading (°C)	Error (°C)	Uncertainty of measurement (\pm °C)	Coverage factor k
25.0	24.995	25.1	0.105	0.13	2.00
30.0	29.999	30.0	0.001	0.13	2.00
35.0	35.004	35.0	-0.004	0.13	2.00

Remark : - UUC* = Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-o0o-

เอกสารไม่ควบคุม
a 1108309

๓๓๓



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3 : EQUIPMENT CALIBRATION AND TESTING SERVICES

534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250

TEL. 0-2717-3000 FAX. 0-2719-9484

Cert.No.: 22TW73
Page.: 1 of 2

Certificate of Testing

Equipment : DO Meter
Manufacturer : Horiba
Model : LAQUA-DO210
Serial No. : HE9M0013
ID No. : UAE.EFM.016/2563 (EFM.DO.05/63)
Received Date : 15 March 2022
Test Date : 18 March 2022
Reference : 2203-0596WSC-5
Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
3 Soi Udomsuk 41, Sukhumvit Road, Bangkok,
Phrakhanong, Bangkok 10260
Laboratory Condition : Temperature (25 ± 5) °C
Humidity (50 ± 20) %
Test Procedure : In - house method : GP-CH9
by Comparison Technique with Azide Modification Method
Tested by : Walailak Sirithean

Approved by :

(☒) Malee Butkruea
(☐) Saithip Meangmai
(☐) Warakorn Lernagatrakul

Approved Signatory

Issue Date : 22 March 2022



Cert.No.: 22TW73
Page.: 2 of 2

Result : Dissolved Oxygen Meter Adjustment With Air 100 %
Dissolved Oxygen Probe No.: 9K9G0097

Titration Method (Azide Modification Method) (mg/L)	DO Meter Reading (mg/L)	Standard Deviation (mg/L)
8.04	8.03	0.0084

This report was certified only for the instrument we tested.It is allowable to use for study the system efficiency.The environmental impact control and present to organization it may concerned Intend to use for advertising and referral purpose is prohibited.This report may not be reproduced other in full,without written approval of the laboratory

-o0o-

Malee .

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม

B 0284372

a 1100966



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
53/44 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-27 FAX. 0-2719-9484



TPA CALIBRATION
NSD-TSI-TS17025
CALIBRATION 0008


Cert. No.: 22LM51
Page.: 1 of 2

Certificate of Calibration

Equipment : DO Meter with Sensor
Manufacturer : Horiba
Model : LAQUA-DO210
Serial No. : HE9M0013
ID No. : UAE.EFM.016/2563(EFM.DO.05/63)
Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
3 Soi Udomsuk 41, Sukhumvit Road,
Bangchak, Phrakhanong,
Bangkok 10260
Location : TPA On Site Calibration Laboratory

Received Order : 15 March 2022
Calibrated Date : 21 March 2022
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %
AC Line Voltage : (220 ± 22) V

Calibrated by : Malee Bulkruea

Approved by : 
() Ponnithippa Tameyakul
(✓) Suwit Injai
Approved Signatory

Issue Date : 23 March 2022

The Uncertainties are for a confidence probability of approximately 95 %

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3: Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม

A 0039391



Equipment : DO Meter with Sensor
Condition As-Received : Used Item
Reference : 2203-0596WSC-6
Procedure Used :-

Calibration were conducted using in-house calibration procedure CP-OT01 according to comparison with
Industrial Platinum Resistance Thermometer (IPT) into Temperature Bath.
The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument **Model** **Serial No.** **Cert. No.** **Due Date**
1) Digital Thermometer 1523 2188080 2111273 22 Nov 2022

2. This certificate is valid only to the item calibrated on date and place of calibration.
3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function : Temperature measurement.

This instrument was connected with temperature sensor, S/N.: 9K9G0097

Calibration Point (°C)	Immersion Depth (mm)	Standard Temperature (°C)	UUC* Reading (°C)	Error (°C)	Uncertainty (± °C)	Coverage Factor k
25.0	80	24.998	25.0	0.002	0.16	2.00
30.0	80	29.999	30.0	0.001	0.16	2.00
35.0	80	34.997	34.9	-0.097	0.16	2.00

UUC* : Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor **k**, providing a level of confidence of approximately 95 %.

-o0o-

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม

A 1101167



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-27 FAX. 0-2719-9484



NSC-ITS17025
CALIBRATION 0001

Cert.No.: 22CH580


Page.: 1 of 3

Certificate of Calibration

Equipment : Conductivity Meter
Manufacturer : Horiba
Model : LAQUA-EC210
Serial No. : HC9L0014
ID No. : UAE.EFM.007/2563(EFM.SCT.01/63)
Condition As-Received: Used Item
Received Date : 26 April 2022
Calibration Date : 27 April 2022
Reference : 2204-0676WSC-1
Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
3 Soi Udomsuk 41, Sukhumvit Road,
Bangchak, Phrakhanong, Bangkok 10260

Ambient Temperature : $(25 \pm 2.5) ^\circ\text{C}$
Relative Humidity : $(50 \pm 15) \%$
Calibration Procedure : In-house method :
- CP-CH6 by direct measurement
with certified reference material (CRM)
- CP-CH8 by comparison with standard thermometer

Calibrated by : Warakorn Lengagrakul

Approved by : 
Approved Signatory

() Malee Buikruea
(✓) Sathip Meangmal
() Warakorn Lengagrakul

Issue Date : 6 May 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม

A 0040861



Cert.No.: 22CH580
Page.: 2 of 3

Condition of this result of calibration

1. Reference Standard Instrument :-

Instrument	Serial No.	ID No.	Certificate No.	Due date
1) Thermometer	1963878	130RC095	211977	17 Sep 2022
2) Ref. Std. Thermometer	4982054	110RC044	2111201	26 Oct 2022

This certification is traceable to the International System of Unit maintained at:-

- Traceable to National Institute of Metrology (Thailand), NIMT

2. Certified Reference Materials :-

- Conductivity calibration solution, CPA chem Ltd., The measurement results are traceable to SI through CPA chem Ltd., ANSI-ASQ National Accreditation Board, Accredited No. AR-1835

Conductivity Solution	Manufacturer	Lot No.	Exp. date
1413.0 $\mu\text{S/cm}$	CPA Chem	766815	04 Sep 2022
12.880 mS/cm	CPA Chem	761022	02 Aug 2022

- Control Conductivity calibration solution temperature by Water bath $(25 \pm 0.1) ^\circ\text{C}$

3. This certificate is valid only to the item calibrated on date and place of calibration.

Calibration results

Function : Conductivity Measurement

(*) After Adjustment at 1413.0 $\mu\text{S/cm}$

Conductivity Electrode Serial No.: 9B9F0064

Standard Conductivity Solution	Before Adjustment UUC* Reading	After Adjustment UUC* Reading	Uncertainty of Measurement (\pm)	Coverage factor k
1413.0 $\mu\text{S/cm}$	996.2 $\mu\text{S/cm}$	1413 $\mu\text{S/cm}$	9.2 $\mu\text{S/cm}$	2.00
12.880 mS/cm	9.01 mS/cm	12.78 mS/cm	0.086 mS/cm	2.00

Remark - UUC* = Unit Under Calibration

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม

a 1107546



Cert.No.: 22CH580
Page.: 3 of 3

Calibration Results

Function : Temperature Measurement

(*) Without adjustment

This equipment was connected with Temperature Probe;

- Model : 9383

- Serial No. 9B9F0064

Dimension of probe;

- Length : 90 mm.

- Diameter : 9 mm.

- Immersion Depth : 80 mm.

Calibration Point (°C)	Standard Temperature (°C)	UUC* Reading (°C)	Error (°C)	Uncertainty of Measurement (\pm °C)	Coverage factor k
25.0	25.002	25.0	-0.002	0.13	2.00
30.0	30.002	30.0	-0.002	0.13	2.00
35.0	35.003	35.0	-0.003	0.13	2.00

Remark : - UUC* = Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

-o0o-

รายการเครื่องมือหลักประจำห้องปฏิบัติการวิเคราะห์สำหรับตรวจวัดคุณภาพสิ่งแวดล้อม

No.	Instrument/Equipment	Parameter	Manufacturer	Model/Serial No.	Calibrator	Certification No.	Date of Calibration	Due date of Calibration*	Remark
Laboratory Instrument/Equipments.(คุณภาพอากาศ และน้ำ)									
1	Analytical Balance (Readability 0.1 mg)	ฝุ่นละอองรวม (Particulate)	Mettler-Toledo	AB204-S / 1128312528	Mettler-Toledo (Thailand) Ltd.	TH2058-097-040722- ACC-TH	7 Apr 22	6 Apr 23	-
2	Analytical Balance (Readability 0.1 mg)		Mettler-Toledo	AB204-S/FACT / B108115858	Mettler-Toledo (Thailand) Ltd.	TH2058-098-040722- ACC-TH	7 Apr 22	6 Apr 23	-
3	Ion Chromatrography (IC)	ไฮโดรเจนคลอไรด์ (HCl)	Dionex	DX-120 / 03010223	Archemica Lab Co.Ltd.	Qualification Report Anion (ID#042)	8 Dec 21	7 Dec 22	-
4	UV-VIS Spectrophotometer	สี (Color), สารชักพอก (Surfactant) ซีโอดี (COD), ไครเมียมไตรวาเลนต์ ไซยาไนด์ (Cyanide) ฟอร์มัลดีไฮด์ (Formaldehyde)	Agilent Technologies	Cary60 G6860A / MY15410009	DOE Services Co.,Ltd.	SP22-016	31 May 22	30 May 23	-
5	UV-VIS Spectrophotometer		Hitachi	U-1900 / 2021-064	DOE Services Co.,Ltd.	SP22-007	20 Jan 22	19 Jan 23	-
6	UV-VIS Spectrophotometer	สารประกอบฟีนอล (Phenol Compound), Phosphate, ออกไซด์ของไนโตรเจน ในรูปของไนโตรเจนไดออกไซด์ (NO _x as NO ₂), ไนเตรตในหน่วยไนโตรเจน (NO ₃) แอมโมเนียในหน่วยไนโตรเจน (NH ₃)	Hitachi	U-2900 / 21E22-009	DOE Services Co.,Ltd.	SP22-008	20 Jan 22	19 Jan 23	-
7	Atomic Absorption Spectrometer (AAS)	ปรอท (Hg), ทองแดง (Cu), แคดเมียม (Cd), สังกะสี (Zn), ตะกั่ว (Pb), ไครเมียมไตรวาเลนต์ (Cr ³⁺), สารหนู (As)	Agilent Technologies	System ID:G8432A AA240FS / MY13160001	Thailand Institute of Scientific and Technological Research (TISTR).	MTC.ACL. No. 486/65	7 Mar 22	6 Mar 23	-
8	Inductively Coupled Plasma (ICP)	เบริลเลียม (Be), ไครเมียมเฮกซะ- วาเลนต์ (Cr ⁶⁺), ไครเมียม (Cr)	Agilent Technologies	System ID:G8015A G8015AA / MY18030001	Agilent Technologies (Thailand) Co.,Ltd.	Preventive Maintenance Checklist	9 Dec 21	8 Dec 22	-

รายการเครื่องมือหลักประจำห้องปฏิบัติการวิเคราะห์คุณภาพสิ่งแวดล้อม

No.	Instrument/Equipment	Parameter	Manufacturer	Model/Serial No.	Calibrator	Certification No.	Date of Calibration	Due date of Calibration*	Remark
		นิกเกิล (Ni), แบเรียม (Ba), ซีลีเนียม (Se), แมงกานีส (Mn), เงิน (Ag), เหล็กทั้งหมด (Total Iron), Total Chromium							
9	pH Meter	ความเป็นกรดและด่าง (pH)	Mettler-Toledo	Seven Easy S20 / 1231155210	National Food Institute, Ministry of Industry, Thailand	2201793-001-01	1 Mar 22	28 Feb 23	
10	pH Meter	อุณหภูมิ (Temperature)	Mettler-Toledo	Seven Easy S20 / 1230525212	National Food Institute, Ministry of Industry, Thailand	2202093-001-01	16 Mar 22	15 Mar 23	
11	Analytical Balance (Readability 0.01 mg)	ของแข็งละลายน้ำทั้งหมด (TDS) ของแข็งแขวนลอยทั้งหมด (TSS)	Mettler-Toledo	XSR205DU / C210685394	Mettler-Toledo (Thailand) Ltd.	TH2058-043-050622-ACC-TH	9 May 22	8 May 23	-
12	Hot Air Oven		Memmert	UF55 / B216.1666	Technology Promotion Association (Thailand-Japan)	21TM1876	29 Oct 21	28 Oct 22	-
13	BOD Incubator	บีโอดี (BOD)	Arco	UC4-1320 / (UAE.WAO.015/2561)	Technology Promotion Association (Thailand-Japan)	22TM90	17 Feb 22	16 Feb 23	-
14	BOD Incubator		Arco	UR-1320 / (UAE.WAO.018/2551)	Technology Promotion Association (Thailand-Japan)	22TM305	7 Apr 22	6 Apr 23	-
15	COD Reactor (Heating Block)	ซีโอดี (COD)	Hanna	HI839800-02 / H018500I	Technology Promotion Association (Thailand-Japan)	HIT-2209-0184	1 Mar 22	1 Mar 23	-
16	Analytical Balance (Readability 0.1 mg)	น้ำมันและไขมัน (Fat Oil & Grease)	Mettler-Toledo	AB-204S/FACT / 1129361010	National Food Institute, Ministry of Industry, Thailand	2203120-001-01	1 Jun 22	31 May 23	-
17	Digestor Unit	ทีเคเอ็น (TKN)	FOSS TECATOR	2520auto / 91794469	National Food Institute, Ministry of Industry, Thailand	2202361-001-01	4 Apr 22	3 Apr 23	-

รายการเครื่องมือหลักประจำห้องปฏิบัติการวิเคราะห์สำหรับตรวจวิเคราะห์คุณภาพสิ่งแวดล้อม

No.	Instrument/Equipment	Parameter	Manufacturer	Model/Serial No.	Calibrator	Certification No.	Date of Calibration	Due date of Calibration*	Remark
18	Distillation Unit (Kjeldahl Method)	ทีเคเอ็น (TKN)	FOSS TECATOR	KT200 / 91790524	FOSS South East Asia	5874	30 Nov 21	29 Nov 22	-
19	Gas Chromatography (GC)	สารกำจัดศัตรูพืชและสัตว์ (Pesticide)	Agilent Technologies	System ID:CN11021007 7890 / CN11021007	Agilent Technologies (Thailand) Co.,Ltd.	Certificate of System Qualification GC-OQ	11 Feb 22	10 Feb 23	-
20	Conductivity Meter	ค่าความนำไฟฟ้า (EC)	SI Analytics	Lab955 / 16300356	SPC Calibration Center Co.,Ltd.	C24220084	22 Mar 22	21 Mar 23	
21	Incubator	แบคทีเรียกลุ่มโคลิฟอร์มทั้งหมด (Total Coliform Bacteria)	Binder	KB400 / 20200000015535	Technology Promotion Association (Thailand-Japan)	22TM347	27 May 22	26 May 23	-
22	Incubator	แบคทีเรียกลุ่มฟีคอลโคลิฟอร์ม (Fecal Coliform Bacteria)	Memmert	IPP 260 / V616.0066	Technology Promotion Association (Thailand-Japan)	22TM672	5 May 22	4 May 23	
23	Water Bath		Memmert	WNE 14 / L416.0606	Technology Promotion Association (Thailand-Japan)	22TM333	17 Feb 22	16 Feb 23	-
24	Water Bath		Memmert	WNE 14 / L416.0612	Technology Promotion Association (Thailand-Japan)	22TM334	17 Feb 22	16 Feb 23	-
25	Analytical Balance		Mettler-Toledo	M5603S / B0070110311	Mettler-Toledo (Thailand) Ltd.	TH2058-096-040722-ACC-TH	7 Apr 22	6 Apr 23	-
26	Auto Clave		ALP	CL-40L / 802664	Technology Promotion Association (Thailand-Japan)	22TM89	17 Feb 22	16 Feb 23	-


Due Date of Calibration* : Schedule the program once a year at least once a year.

Measurement Results

Repeatability




Eccentricity



Mettler-Toledo (Thailand) Ltd.
846/4 - 846/5 Lasalle Rd., Bangna Tai Sub-District
Bangna District, Bangkok 10260
+66 2723 0382
MT-TH.ServicesSupport@mt.com

Accuracy Calibration Certificate

Customer

Company: United Analyst and Engineering Consultant Co., Ltd.
Address: 3 Soi Udom Suk 41, Sukhumvit Rd., Bang Chak
City: Phra Khanong
Zip / Postal: 10260
State / Province: Bangkok
Order Number: 
Contact: Suwit Chotnok

Weighing Device

Manufacturer: Mettler Toledo
Model: AB204-S
Serial No.: 1128312828
Building: N/A
Floor: 2
Room: Balance Room 2 (206)
Instrument Type: Weighing Instrument
Asset Number: UAE-AIR.019/2550
Terminal Model: N/A
Terminal Serial No.: N/A
Terminal Asset No.: N/A

Procedure

Calibration Guideline: EURAMET cg-18 v. 4.0 (11/2015)
Mettler Toledo Work Instruction: CPW002/20

This calibration certificate contains measurements for As Found calibration. No As Left calibration was performed because the device was not modified after As Found calibration. Therefore, results for As Left correspond to As Found.

The sensitivity/span of the weighing instrument was adjusted before calibration with a built-in weight.

In accordance with EURAMET cg-18 (11/2015), the test loads were selected to reflect the specific use of the weighing device or to accommodate specific calibration conditions.

As Found	Temperature		Humidity	
	Start: 22.5 °C	End: 21.4 °C	Start: 56.1 %	End: 63.2 %

As Found Calibration Date: 07-Apr-2022

As Left Calibration Date: N/A

Issue Date: 08-Apr-2022

Calibrator: 

Sirawit Chanchan

Approved Signatory:



☒ Kassakorn Tassanachaisakul
☐ Santi Jitnyom
☐ Surachet Sukkale

Remarks

Equipment condition: Good
Next calibration according to customer's procedure
Calibration data not decide by calibration laboratory
Test weight by Filter pan : 1 g = 0.9999 g, 3 g = 3.0000 g, 5 g = 5.0000 g

End of Accredited Section

The information below and any attachments to this calibration certificate are not part of the accredited calibration.

Error of Indication

As Found

	Reference Value	Indication	Error of Indication	Expanded Uncertainty	k
1	0.0000 g	0.0000 g	0.0000 g	0.18 mg	2
2	0.1000 g	0.1000 g	0.0000 g	0.19 mg	2
3	1.0000 g	0.9999 g	-0.0001 g	0.19 mg	2
4	5.0000 g	5.0000 g	0.0000 g	0.19 mg	2
5	10.0000 g	9.9999 g	-0.0001 g	0.20 mg	2
6	20.0000 g	20.0000 g	0.0000 g	0.21 mg	2
7	50.0000 g	50.0000 g	0.0000 g	0.23 mg	2
8	70.0001 g	70.0000 g	-0.0001 g	0.28 mg	2
9	100.0000 g	100.0000 g	0.0000 g	0.29 mg	2
10	150.0000 g	150.0002 g	0.0002 g	0.40 mg	2
11	200.0001 g	200.0003 g	0.0002 g	0.46 mg	2



The uncertainty stated is the expanded uncertainty at calibration obtained by multiplying the standard combined uncertainty by the coverage factor k – which can be larger than 2 according to EURAMET cg-18. The value of the measurand lies within the assigned range of values with a probability of approximately 95%.

The user is responsible for maintaining environmental conditions and the settings of the weighing instrument when it was calibrated.

Test Equipment

All weights used for metrological testing are traceable to national or international standards. The weights were calibrated and certified by an accredited calibration laboratory.

Weight Set 1: OIML E2


Weight Set No.:	WS80	Date of Issue:	23-Feb-2022
Certificate Number:	C208581631	Calibration Due Date:	14-Aug-2023
Thermo Hygrometer			
Equipment No.:	IN161	Date of Issue:	14-Jun-2021
Certificate Number:	21H1220	Calibration Due Date:	01-Jun-2022

Mettler-Toledo (Thailand) Ltd.
846/4 - 846/5 Lasalle Rd., Bangna Tai Sub-District,
Bangna District, Bangkok 10260
+66 2723 0382
MT-TH.ServiceSupport@mt.com



Accuracy Calibration Certificate

Customer

Company:	United Analyst and Engineering Consultant Co., Ltd.		
Address:	3 Soi Udom Suk 41, Sukhumvit Rd., Bang Chak		
City:	Phra Khanong	Contact:	Suwit Chotnok
Zip / Postal:	10260		
State / Province:	Bangkok		
Order Number:	 * 0 3 3 3 2 4 2 3 9 0 6 *		

Weighing Device

Manufacturer:	Mettler-Toledo	Instrument Type:	Weighting Instrument
Model:	AB204-S/FACT	Asset Number:	UAE-AIR.016/2555
Serial No.:	B108115858	Terminal Model:	N/A
Building:	N/A	Terminal Serial No.:	N/A
Floor:	2	Terminal Asset No.:	N/A
Room:	Balance Room 2 (206)		

Range	Max. Capacity	Readability (d)
1	220 g	0.0001 g

Procedure

Calibration Guideline:	EURAMET cg-18 v. 4.0 (11/2015)
METTLER TOLEDO Work Instruction:	CP/W002/20

This calibration certificate contains measurements for As Found and As Left calibrations.

The sensitivity/span of the weighing instrument was adjusted before As Found and As Left calibrations with a built-in weight.
In accordance with EURAMET cg-18 (11/2015), the test loads were selected to reflect the specific use of the weighing device or to accommodate specific calibration conditions.

	Temperature		Humidity	
As Found	Start: 22.6 °C	End: 22.1 °C	Start: 56.0 %	End: 51.9 %
As Left	Start: 22.3 °C	End: 22.4 °C	Start: 46.2 %	End: 55.8 %

As Found Calibration Date:	07-Apr-2022	Calibrator:	 Sirawit Chanchan
As Left Calibration Date:	07-Apr-2022		
Issue Date:	08-Apr-2022	Approved Signatory:	 Kassakorn Tassanachaisakul <input checked="" type="checkbox"/> Santi Jitnyom <input type="checkbox"/> Surachet Sukkate

Measurement Uncertainty of the Weighing Instrument in Use

Stated is the expanded uncertainty with k=2 in use. The formula shall be used for the estimation of the uncertainty under consideration of the errors of indication. The value R represents the net load indication in the unit of measure of the device.

Temperature coefficient for the evaluation of the measurement uncertainty in use:	$3.0 \cdot 10^{-6} / K$
Temperature range on site for the evaluation of the measurement uncertainty in use:	3 K


Uncertainties of Uncertainty Equation

Range		As Found	As Left
d	Max		
1	0.0001 g	220 g	$U_1 = 0.19 \text{ mg} + 0.00817 \text{ mg/g} \cdot R$
			N/A

To optimize the stability of the linearization, besides of the zero load only increasing measurement points with a test load of 5% of the measurement range or larger are taken for the calculation of the linear equation.

Absolute and Relative Measurement Uncertainty in Use for Various Net Indications (Examples)

Net Indication	As Found	As Left
0.0220 g	0.19 mg	0.86%
0.2200 g	0.19 mg	0.087%
2.2000 g	0.21 mg	0.0095%
22.0000 g	0.37 mg	0.0017%
220.0000 g	2.0 mg	0.00090%




Error of Indication

As Found				
1	Reference Value	Indication	Error of Indication	Expanded Uncertainty
1	0.0000 g	0.0000 g	0.0000 g	0.15 mg
2	0.1000 g	0.1001 g	0.0001 g	0.16 mg
3	1.0000 g	0.9999 g	-0.0001 g	0.16 mg
4	5.0000 g	5.0000 g	0.0000 g	0.16 mg
5	10.0000 g	10.0001 g	0.0001 g	0.17 mg
6	20.0000 g	20.0001 g	0.0001 g	0.18 mg
7	50.0000 g	50.0003 g	0.0003 g	0.20 mg
8	70.0001 g	70.0005 g	0.0004 g	0.26 mg
9	100.0000 g	100.0005 g	0.0005 g	0.27 mg
10	150.0000 g	150.0007 g	0.0007 g	0.38 mg
11	200.0001 g	200.0008 g	0.0007 g	0.44 mg

As Left				
1	Reference Value	Indication	Error of Indication	Expanded Uncertainty
1	0.0000 g	0.0000 g	0.0000 g	0.15 mg
2	0.1000 g	0.1000 g	0.0000 g	0.16 mg
3	1.0000 g	0.9999 g	-0.0001 g	0.17 mg
4	5.0000 g	5.0000 g	0.0000 g	0.17 mg
5	10.0000 g	10.0000 g	0.0000 g	0.17 mg
6	20.0000 g	20.0000 g	0.0000 g	0.18 mg
7	50.0000 g	50.0000 g	0.0000 g	0.21 mg
8	70.0001 g	70.0001 g	0.0000 g	0.26 mg
9	100.0000 g	100.0000 g	0.0000 g	0.28 mg
10	150.0000 g	150.0001 g	0.0001 g	0.39 mg
11	200.0001 g	200.0001 g	0.0000 g	0.45 mg

The uncertainty stated is the expanded uncertainty at calibration obtained by multiplying the standard combined uncertainty by the coverage factor k – which can be larger than 2 according to EURAMET cg-18. The value of the measurand lies within the assigned range of values with a probability of approximately 95%..

The user is responsible for maintaining environmental conditions and the settings of the weighing instrument when it was calibrated.



For improved legibility of the graphics only increasing measurement points are shown and measurement points close to zero are not displayed.

Measurement Results

Repeatability



The "g" in the graph represents the readability of the range interval in which the test was performed.
The results of this graph are based upon the absolute values of the differences from the mean value.

Eccentricity



The "g" in the graph represents the readability of the range interval in which the test was performed.

Measurement Uncertainty of the Weighing Instrument in Use

Stated is the expanded uncertainty with k=2 in use. The formula shall be used for the estimation of the uncertainty under consideration of the errors of indication. The value R represents the net load indication in the unit of measure of the device.

Temperature coefficient for the evaluation of the measurement uncertainty in use:	2.5 · 10 ⁻⁶ / K
Temperature range on site for the evaluation of the measurement uncertainty in use:	3 K


Uncertainty of Uncertainty Equation

Range	As Found	As Left
d		
1	0.0001 g 220 g	U ₁ = 0.16 mg + 0.0111 mg/g · R U ₁ = 0.16 mg + 0.00592 mg/g · R

To optimize the stability of the linearization, besides of the zero load only increasing measurement points with a test load of 5% of the measurement range or larger are taken for the calculation of the linear equation.

Absolute and Relative Measurement Uncertainty in Use for Various Net Indications (Examples)

Net Indication	As Found		As Left	
0.0220 g	0.16 mg	0.73%	0.16 mg	0.73%
0.2200 g	0.16 mg	0.074%	0.16 mg	0.073%
2.2000 g	0.18 mg	0.0084%	0.17 mg	0.0079%
22.0000 g	0.40 mg	0.0018%	0.29 mg	0.0013%
220.0000 g	2.6 mg	0.0012%	1.5 mg	0.00068%



Test Equipment

All weights used for metrological testing are traceable to national or international standards. The weights were calibrated and certified by an accredited calibration laboratory.

Weight Set 1: OIML E2

Weight Set No.:	WS80	Date of Issue:	23-Feb-2022
Certificate Number:	C208581631	Calibration Due Date:	14-Aug-2023

Thermo Hygrometer

Equipment No.:	IN161	Date of Issue:	14-Jun-2021
Certificate Number:	21H1220	Calibration Due Date:	01-Jun-2022

Remarks

FACT adjustment functionality activated
Value of the built-in weight adjusted
Equipment condition: Good
Next calibration according to customer's procedure
Calibration data not decide by calibration laboratory
Test weight by Filler pan : 1 g = 1.0000 g, 3 g = 3.0000 g, 5 g = 5.0000 g

End of Accredited Section

The information below and any attachments to this calibration certificate are not part of the accredited calibration.

Qualification Report

PM Check list ,CM_OQ and PQ
DX-120 : Anion (ID#042)

For
UAE
(2nd Contract)



Certificate of Calibration DX-120 : Anion (ID#042)

This certificate is to verify that instrument below are calibrated
by Archemica Lab Co.,Ltd.

DX-120

S/N : 03010223

for



บริษัท อารเคมีคัล แล็บ จำกัด
ARCHEMICA LAB CO.,LTD

Operator Signature : K. Channarong Date : Dec 08, 2021

(Mr.Channarong Khiao-un)

Test Engineer

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม



Dionex Ion Chromatography
Preventive Maintenance Report

Customer Organization	Name/Department
UAE (2nd Contract)	Khun Suwan
Engineer Name	Date
Mr.Channarong Khiao-un	8-Dec-2021

Instrument Detail

Instrument Model	Application
DX-120 (ID#042)	Anion
Instrument components	Serial Number
DX-120	03010223

Consumable Detail

Columns	Guard Columns	Suppressors	Concentrators	Etc.
AS22	AG22	ASRS 300	-	-
Remark:				

Perform By
Archimica Lab Co.,Ltd



Channarong

Archimica Lab Co.,Ltd
บริษัท อัคริคา แล็บ จำกัด
ARCHIMICA LAB CO.,LTD

8-Dec-21

Date

Suwan

Customer

8-Dec-21

Date

Preventive Maintenance Check List



General Inspection Checklist

Item	Description	Result		Action Taken	N.A.
		Pass	Fail		
1	Power line 220 Vac	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
2	Pneumatic Line	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
3	Pressure outlet 80-100 psi	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
4	Barbed fitting and tee fitting	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
5	Crimped and blocked tubing	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
6	Rheodyne Valve for Leak	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check&Clean	<input type="checkbox"/>
7	Slider valve for leak	<input type="checkbox"/>	<input type="checkbox"/>	-	<input checked="" type="checkbox"/>
8	Inspect slider	<input type="checkbox"/>	<input type="checkbox"/>	-	<input checked="" type="checkbox"/>
9	Inspect port face	<input type="checkbox"/>	<input type="checkbox"/>	-	<input checked="" type="checkbox"/>
10	Inspect pressure bolt	<input type="checkbox"/>	<input type="checkbox"/>	-	<input checked="" type="checkbox"/>
11	Inspect fitting and ferrule	<input type="checkbox"/>	<input type="checkbox"/>	-	<input checked="" type="checkbox"/>
12	Suppressor for leak	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
13	Cell for leak	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
14	Electronic cable connected	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
15	Column selection valve for leak	<input type="checkbox"/>	<input type="checkbox"/>	-	<input checked="" type="checkbox"/>
16	Inspect all fitting and line	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
17	Eluent reservoir	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
18	Inspect cap o-ring	<input type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
19	Inspect air for leak	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
20	Piston seal has been replaced	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
21	Back up seal has been replaced	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
22	Pump Lubricate	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
23	Front panel test	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
24	Low limit alarm	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
25	Hi limit alarm	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
26	Conductivity electronic test 160 \pm 1 μ S	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
27	Check noise for suppressor (pk to pk <0.005 μ S)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
28	Check column	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
29	Check suppressor	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check&Clean	<input type="checkbox"/>
30	Check pump	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
31	Check cell	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
32	Check leak sensor	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
33	Flow rate	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
34	System pressure	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>
35	Detector background	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Check	<input type="checkbox"/>

Chromeleon Operational Qualification
(CM_OQ)



Chromeleon Operational Qualification, Part 1

Verification of Selected Results

Calibration Type: LOff
Integration Type: Area
Standard Method: External
Calibration Mode: Total
Auto Recalibrate: ON

Report Variable	Peak Name	Status
Offset (c0)	n.a.	ok
	n.a.	ok
	n.a.	ok
Slope (c1)	Methylparabene	ok
	Ethylparabene	ok
	Propylparabene	ok
Correlation Coeffi.	Methylparabene	ok
	Ethylparabene	ok
	Propylparabene	ok
Variance	Methylparabene	ok
	Ethylparabene	ok
	Propylparabene	ok
Std. Deviation	Methylparabene	ok
	Ethylparabene	ok
	Propylparabene	ok
Rel. Std. Dev.	Methylparabene	ok
	Ethylparabene	ok
	Propylparabene	ok
Variance Coeffi.	Methylparabene	ok
	Ethylparabene	ok
	Propylparabene	ok

เอกสารไม่ควบคุม



Chromeleon Operational Qualification

General Information

Computer Name (Server): LAB-IC
Computer Name (Client): LAB-IC
Version Number: 6.80 SR12 Build 3578 (207169)
Operator: Mr.Channarong Khiao-Un

General System Suitability Test: *Test passed*

Comparison Formats:

All Parameters: (Exceptions see below)	Significant Digits: (They must match exactly)	10
Time Related Frac. Coll. Parameters: [The parameters are marked with *.]	Max. Deviation:	0.02 s



Samon 8-Dec-21

Reviewer's Signature // Date

Operator's Signature // Date

เอกสารไม่ควบคุม



Chromeleon Operational Qualification, Part 1

Verification of Selected Results

Report Variable	Peak Name	Status
Theoretical Plates (EP)	Methylparabene	ok
	Ethylparabene	ok
	Propylparabene	ok
Theoretical Plates (USP)	Methylparabene	ok
	Ethylparabene	ok
	Propylparabene	ok
Theoretical Plates (JP)	Methylparabene	ok
	Ethylparabene	ok
	Propylparabene	ok

Test Result:

Passed

Simon 8-Dec-21
Reviewer's Signature // Date

Operator's Signature // Date
K. Attard / 8-Dec-21



Chromeleon Operational Qualification, Part 1

Verification of Selected Results

Report Variable	Peak Name	Status
Calibration Point X	Methylparabene	ok
	Ethylparabene	ok
	Propylparabene	ok
Calibration Point Y	Methylparabene	ok
	Ethylparabene	ok
	Propylparabene	ok
Amount [ng]	Methylparabene	ok
	Ethylparabene	ok
	Propylparabene	ok
Resolution (EP)	Methylparabene	ok
	Ethylparabene	ok
Resolution (USP)	Methylparabene	ok
	Ethylparabene	ok
Peak Asymmetry (EP/USP)	Methylparabene	ok
	Ethylparabene	ok
	Propylparabene	ok
Peak Asymmetry (AIA)	Methylparabene	ok
	Ethylparabene	ok
	Propylparabene	ok




Chromeleon Operational Qualification, Part 2

Most Frequently Used Parameters: Comparison with Expected Results

Variable Category	Report Variable	Peak Name	Status
Peak Results	Ret.Dev.(abs)	Methylparabene	ok
	Ret.Dev.(abs)	Ethylparabene	ok
	Ret.Dev.(abs)	Propylparabene	ok
	Ret.Dev.(rel)	Methylparabene	ok
	Ret.Dev.(rel)	Ethylparabene	ok
	Ret.Dev.(rel)	Propylparabene	ok
	Area	Methylparabene	ok
	Area	Ethylparabene	ok
	Area	Propylparabene	ok
	Rel.Area (Total)	Methylparabene	ok
	Rel.Area (Total)	Ethylparabene	ok
	Rel.Area (Total)	Propylparabene	ok
	Height	Methylparabene	ok
	Height	Ethylparabene	ok
	Height	Propylparabene	ok
	Rel.Height (Total)	Methylparabene	ok
	Rel.Height (Total)	Ethylparabene	ok
	Rel.Height (Total)	Propylparabene	ok
	Amount	Methylparabene	ok
Chromatogram	Amount	Ethylparabene	ok
	Concentration	Propylparabene	ok
	Concentration	Methylparabene	ok
	Concentration	Ethylparabene	ok
	Rel.Amount	Propylparabene	ok
	Rel.Amount	Methylparabene	ok
	Rel.Amount	Ethylparabene	ok
	Peak Width (0%)	Propylparabene	ok
	Peak Width (0%)	Methylparabene	ok
	Peak Width (0%)	Ethylparabene	ok
	Peak Width (5%)	Propylparabene	ok
	Peak Width (5%)	Methylparabene	ok
Peak Results	Peak Width (5%)	Ethylparabene	ok
	Peak Width (10%)	Propylparabene	ok
	Peak Width (10%)	Methylparabene	ok
	Peak Width (10%)	Ethylparabene	ok

เอกสารไม่ควบคุม



Chromeleon Operational Qualification, Part 2

Most Frequently Used Parameters: Comparison with Expected Results

Variable Category	Report Variable	Peak Name	Status
Calibration Type	LOff		ok
	Integration Type		ok
	Area		ok
	Standard Method:	External	ok
	Total		ok
	Calibration Mode:		ok
	Auto Recalibrate:	ON	ok
			ok
			ok
			ok
Chromatogram	No.		ok
	Name		ok
	Sample Type		ok
	Position		ok
	Status		ok
	Inj.Vol.		ok
	Dil.Fac.		ok
	Weight		ok
	Amount		ok
	Program		ok
Peak Results	Quantification Method		ok
	Channel		ok
	No. of Peaks		ok
	Start Time		ok
	Signal Min.		ok
	Signal Max.		ok
	Signal Dimension		ok
	Noise 2.1-2.3		ok
	No.	Methylparabene	ok
	No.	Ethylparabene	ok
Peak Results	No.	Propylparabene	ok
	Peak Name	Methylparabene	ok
	Peak Name	Ethylparabene	ok
	Peak Name	Propylparabene	ok
	Ret.Time	Methylparabene	ok
	Ret.Time	Ethylparabene	ok
Peak Results	Ret.Time	Propylparabene	ok
			ok

เอกสารไม่ควบคุม

Smp: Parabenes



Chromeleon Operational Qualification, Part 2

Most Frequently Used Parameters: Comparison with Expected Results

Variable Category	Report Variable	Peak Name	Status
Peak Results	Asymmetry(AIA)	Methylparabene	ok
	Asymmetry(AIA)	Ethylparabene	ok
	Asymmetry(AIA)	Propylparabene	ok
	Theoretical Plates(EP)	Methylparabene	ok
	Theoretical Plates(EP)	Ethylparabene	ok
	Theoretical Plates(EP)	Propylparabene	ok
	Theoretical Plates(USP)	Methylparabene	ok
	Theoretical Plates(USP)	Ethylparabene	ok
	Theoretical Plates(USP)	Propylparabene	ok
	Theoretical Plates(JP)	Methylparabene	ok
	Theoretical Plates(JP)	Ethylparabene	ok
	Theoretical Plates(JP)	Propylparabene	ok
Peak Calibration	Cal.Mode	Methylparabene	ok
	Cal.Mode	Ethylparabene	ok
	Cal.Mode	Propylparabene	ok
	Auto.Recal.	Methylparabene	ok
	Auto.Recal.	Ethylparabene	ok
	Auto.Recal.	Propylparabene	ok
	Cal.Type	Methylparabene	ok
	Cal.Type	Ethylparabene	ok
	Cal.Type	Propylparabene	ok
	Weights	Methylparabene	ok
	Weights	Ethylparabene	ok
	Weights	Propylparabene	ok
	Offset	Methylparabene	ok
	Offset	Ethylparabene	ok
	Offset	Propylparabene	ok
	Slope	Methylparabene	ok
	Slope	Ethylparabene	ok
	Slope	Propylparabene	ok
	RF-Value	Methylparabene	ok
	RF-Value	Ethylparabene	ok
	RF-Value	Propylparabene	ok
	No. of Points	Methylparabene	ok
	No. of Points	Ethylparabene	ok

เอกสารไม่ควบคุม

Smp: Parabenes



Chromeleon Operational Qualification, Part 2

Most Frequently Used Parameters: Comparison with Expected Results

Variable Category	Report Variable	Peak Name	Status
Peak Results	Peak Width (50%)	Methylparabene	ok
	Peak Width (50%)	Ethylparabene	ok
	Peak Width (50%)	Propylparabene	ok
	Left Width (0%)	Methylparabene	ok
	Left Width (0%)	Ethylparabene	ok
	Left Width (0%)	Propylparabene	ok
	Right Width (0%)	Methylparabene	ok
	Right Width (0%)	Ethylparabene	ok
	Right Width (0%)	Propylparabene	ok
	Peak Start	Methylparabene	ok
	Peak Start	Ethylparabene	ok
	Peak Start	Propylparabene	ok
	Peak Stop	Methylparabene	ok
	Peak Stop	Ethylparabene	ok
	Peak Stop	Propylparabene	ok
	Peak Start Value	Methylparabene	ok
	Peak Start Value	Ethylparabene	ok
	Peak Start Value	Propylparabene	ok
	Peak Stop Value	Methylparabene	ok
	Peak Stop Value	Ethylparabene	ok
	Peak Stop Value	Propylparabene	ok
	BL-Value Peak Start	Methylparabene	ok
	BL-Value Peak Start	Ethylparabene	ok
	BL-Value Peak Start	Propylparabene	ok
	BL-Value Peak Stop	Methylparabene	ok
	BL-Value Peak Stop	Ethylparabene	ok
	BL-Value Peak Stop	Propylparabene	ok
	Type	Methylparabene	ok
	Type	Ethylparabene	ok
	Type	Propylparabene	ok
	Resolution(EP)	Methylparabene	ok
	Resolution(EP)	Ethylparabene	ok
	Resolution(USP)	Methylparabene	ok
	Resolution(USP)	Ethylparabene	ok
	Asymmetry(EP)	Methylparabene	ok
	Asymmetry(EP)	Ethylparabene	ok
	Asymmetry(EP)	Propylparabene	ok

เอกสารไม่ควบคุม



Chromeleon Operational Qualification, Part 2

Most Frequently Used Parameters: Comparison with Expected Results

Variable Category	Report Variable	Peak Name	Status
Peak Calibration	Residual for Cal.Point X	Methylparabene	ok
	Residual for Cal.Point X	Ethylparabene	ok
	Residual for Cal.Point X	Propylparabene	ok
	Calibration Point Status	Methylparabene	ok
	Calibration Point Status	Propylparabene	ok
	Amount	Methylparabene	ok
	Amount	Ethylparabene	ok
	Amount	Propylparabene	ok
	Peak Tab. Cal.Type	Methylparabene	ok
	Peak Tab. Peak Type	Methylparabene	ok
Peak Table	Peak Tab. Left Limit	Methylparabene	ok
	Peak Tab. Right Limit	Methylparabene	ok
	Peak Tab. Group	Methylparabene	ok
	Peak Tab. Resp.Factor	Methylparabene	ok
	Peak Tab. Amount	Methylparabene	ok
	Peak Tab. Amnt.Dim	Methylparabene	ok

เอกสารไม่ควบคุม



Chromeleon Operational Qualification, Part 2

Most Frequently Used Parameters: Comparison with Expected Results

Variable Category	Report Variable	Peak Name	Status
Peak Calibration	No. of Points	Propylparabene	ok
	No. of Points(disabled)	Methylparabene	ok
	No. of Points(disabled)	Ethylparabene	ok
	No. of Points(disabled)	Propylparabene	ok
	Variance	Methylparabene	ok
	Variance	Ethylparabene	ok
	Variance	Propylparabene	ok
	Var.Coeff	Methylparabene	ok
	Var.Coeff	Ethylparabene	ok
	Var.Coeff	Propylparabene	ok
	Std.Dev.	Methylparabene	ok
	Std.Dev.	Ethylparabene	ok
	Std.Dev.	Propylparabene	ok
	Rel.Std.Dev.	Methylparabene	ok
	Rel.Std.Dev.	Ethylparabene	ok
	Rel.Std.Dev.	Propylparabene	ok
	Corr.Coeff.	Methylparabene	ok
	Corr.Coeff.	Ethylparabene	ok
	Corr.Coeff.	Propylparabene	ok
	Coeff.Det.	Methylparabene	ok
	Coeff.Det.	Ethylparabene	ok
	Coeff.Det.	Propylparabene	ok
	Adj. Coeff.Det.	Methylparabene	ok
	Adj. Coeff.Det.	Ethylparabene	ok
	Adj. Coeff.Det.	Propylparabene	ok
	X	Methylparabene	ok
	X	Ethylparabene	ok
	X	Propylparabene	ok
	Y	Methylparabene	ok
	Y	Ethylparabene	ok
	Y	Propylparabene	ok
	W	Methylparabene	ok
	W	Ethylparabene	ok
	W	Propylparabene	ok
	F(X)	Methylparabene	ok
	F(X)	Ethylparabene	ok
	F(X)	Propylparabene	ok

เอกสารไม่ควบคุม



Chromeleon Operational Qualification, Part 2

Most Frequently Used Parameters: Comparison with Expected Results

Variable Category	Report Variable	Peak Name	Status
Peak Purity	PPI	Methylparabene	ok
	PPI	Ethylparabene	ok
	PPI	Propylparabene	ok
	RSD PPI	Methylparabene	ok
	RSD PPI	Ethylparabene	ok
	RSD PPI	Propylparabene	ok
	Match	Methylparabene	ok
	Match	Ethylparabene	ok
	Match	Propylparabene	ok
	RSD Match	Methylparabene	Deviation
	RSD Match	Ethylparabene	Deviation
	RSD Match	Propylparabene	Deviation
	Rel.Max at	Methylparabene	ok
	Rel.Max at	Ethylparabene	ok

Test Result: **Failed**

Reviewer's Signature // Date
Suman 8-Dec-21

Operator's Signature // Date
Kammarone 8-Dec-21

ARCHCHEMICAL LAB
ARCH-CHEMICAL LAB CO., LTD.

เอกสารไม่ควบคุม



Chromeleon Operational Qualification, Part 3

Post-Acquisition Steps: Comparison with Expected Results

Channel Name	Report Variable	Peak Name	Status
Extract UV Channel: EXT250NM	Area	Methylparabene	ok
	Area	Ethylparabene	ok
	Area	Propylparabene	ok
	Height	Methylparabene	ok
	Height	Ethylparabene	ok
	Height	Propylparabene	ok
	Base Peak Width	Methylparabene	ok
	Base Peak Width	Ethylparabene	ok
	Base Peak Width	Propylparabene	ok
	Area	Methylparabene	ok
	Area	Ethylparabene	ok
	Area	Propylparabene	ok
	Height	Methylparabene	ok
	Height	Ethylparabene	ok

Smooth Data: UV_VIS_1_MA_005_001 UV_VIS_1_OL_051_001 EXT290NM_SG_005_010	Noise (1.9-2.4 min)	ok
	Noise (1.9-2.4 min)	ok
	Noise (1.9-2.4 min)	ok
	Noise (1.9-2.4 min)	ok

เอกสารไม่ควบคุม



Chromeleon Operational Qualification, Part 4

System Suitability Test: Comparison with Expected Results

Calibration Type: LOF
Integration Type: Area
Standard Method: External
Calibration Mode: Total
Auto Recalibrate: ON

Variable Category	Report Variable	Status
SST	Test No.	ok
	Test Name	ok
	Sample Condition	ok
	Sample Condition Result	ok
	Test Condition	ok
	Peak Condition	ok
	Aggregate Condition	ok
	Compare Operator	ok
	Compare Value	ok
	Result of Compare Value	ok
	Channel	ok
	Aggregated Samples	ok
	List of Aggr. Smp.	ok
	Result List for Aggr. Smp.	ok
	Result of Test Condition or Aggregate	ok
	N.A.	ok
	Test Result	ok
	Fail-Action	ok

Test Result: Passed

Simon 8-Dec-21

Reviewer's Signature // Date

K. Khamrarporn 8-Dec-21

Operator's Signature // Date



เอกสารไม่ควบคุม



Chromeleon Operational Qualification, Part 3

Post-Acquisition Steps: Comparison with Expected Results

Channel Name	Report Variable	Peak Name	Status
Arith. Comb. of Channels:			
ADD_UV_VIS_1_UV_VIS_1	Area	Methylparabene	ok
ADD_UV_VIS_1_UV_VIS_1	Area	Ethylparabene	ok
ADD_UV_VIS_1_UV_VIS_1	Area	Propylparabene	ok
MUL_UV_VIS_1_UV_VIS_1	Area	Methylparabene	ok
MUL_UV_VIS_1_UV_VIS_1	Area	Ethylparabene	ok
MUL_UV_VIS_1_UV_VIS_1	Area	Propylparabene	ok

Test Result:

Passed

Simon 8-Dec-21

Reviewer's Signature // Date

K. Khamrarporn 8-Dec-21

Operator's Signature // Date



เอกสารไม่ควบคุม



Chromeleon Operational Qualification, Part 5

Fraction Collection: Comparison with Expected Results

Calibration Type: LOFF
Integration Type: Area
Standard Method: External
Calibration Mode: Total
Auto Recalibrate: ON

Variable Category	Report Variable	Status
Fraction Report	Fract. No.	ok
	Fract. Starttime *)	ok
	Fract. Endtime *)	ok
	No. of Tubes	ok
	Position	ok
	Peak Name	ok
Tube Report	No. of Peaks	ok
	Position	ok
	Tube Starttime *)	ok
	Tube Endtime *)	ok
	Max. Tube Volume	ok
	Peak Name	ok
	No. of Peaks	ok
	Fract. No.	ok
	Fract. Starttime *)	ok
	Fract. Endtime *)	ok
	No. of Tubes	ok
	No. of Peaks	ok



Test Result: Passed

Sinan 8-Dec-21
Reviewer's Signature // Date

Kaninapang 8-Dec-21
Operator's Signature // Date

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม

Limits:

Test	Customized Limits	Dionex Recommended Limits
DX120 Conductivity Noise (nS)	2	2
DX120 Conductivity Drift (nS/hr)	20	20
Injector Precision (Area %RSD)	1.0	1.0
Injector Carry Over (Area %)	0.1	0.1
DX120 Detector Linearity (Corr.)	0.999	0.999
DX120 Detector Linearity (%RSD)	5	5

Additional Information:

Customer/Company:	UAE Consultant Co.,Ltd.	Date:	8-Dec-2021
Qualification Executor:	Mr.Channarong / Archemica	Period between Qualifications:	6 months
		Next Qualification:	Jun-2022



Customer Signature
Suman
Date: 8-Dec-21

Qualification Executor
Kannumporn

เอกสารไม่ควบคุม



Performance Qualification

Instruments:

Instrument Name	Model	Supplier	Serial Number	Moduleware Version
Pump	DX120	Dionex	03010223	3.03
Detector	DX120	Dionex	03010223	3.03
Autosampler	AS40 or man. inj.	Dionex	n.a.	0.00
Eluent Generator	n.a.	Dionex	n.a.	0.00
Chromeleon	6.80 SR12 Build 3578 (207169)	Dionex	33308	n.a.

Accessories:

Name	Description	Lot / Serial Number	Expire Date
Backpressure Tubing	0.13 mm (0.005") ID PEEK, 13 m (512")	n.a.	n.a.
Blank	Water	n.a.	n.a.
Sample 1	Nitrate, 5 ppm	210719	Jul-2022
Sample 2	Nitrate, 10 ppm	210719	Jul-2022
Sample 3	Nitrate, 25 ppm	210719	Jul-2022
Sample 4	Nitrate, 50 ppm	210719	Jul-2022
Sample 5	Nitrate, 100 ppm	210719	Jul-2022
Sample 6	Nitrate, 1000 ppm	210719	Jul-2022
Eluent	Water	n.a.	n.a.
Autosampler Reservoir A	Water	n.a.	n.a.



Customer Signature
Suman
Date: 8-Dec-21

Qualification Executor
Kannumporn

Customer signature indicates that all information in the following reports has been reviewed and accepted.

เอกสารไม่ควบคุม

Data for detector noise

Segment number	Noise, nS
1	2.696
2	2.231
3	1.111
4	2.130
5	2.284
6	0.960
7	2.415
8	2.031
9	2.138
10	2.622
11	0.719
12	1.270
13	1.671
14	1.444
15	1.699
16	2.892
17	1.850
18	1.950
19	1.379
20	2.205
Average, nS	1.885
Limit, nS	2
Result	PASS

Data for detector drift

20 Minute drift, nS	Drift, nS/hr	Limit, nS/hr	Result
0.427	1.280	20.000	PASS



Customer Signature

Kittanarong

Qualification Executor

8-Dec-21

Date

เอกสารไม่ควบคุม



Performance Qualification

Detector Noise and Drift

Instruments:

Instrument Name	Model	Supplier	Serial Number	Moduleware Version
Pump	DX120	Dionex	03010223	3.03
Detector	DX120	Dionex	03010223	3.03
Autosampler	AS40 or man. inj.	Dionex	n.a.	0.00
Eluent Generator	n.a.	Dionex	n.a.	0.00

Accessories

Name	Description
Backpressure Tubing	0.13 mm (0.005") ID PEEK, 13 m (512")
Eluent	Water

Additional Information

Customer/Company:	UAE Consultant Co., Ltd.	Date:	8-Dec-2021
Qualification Executor:	Mr.Channarong / Archechemica	Next Qualification:	Jun-2022

Test Results Summary

Test	Result
DX120 Conductivity Noise (nS)	PASS
DX120 Conductivity Drift (nS/hr)	PASS



Customer Signature

Kittanarong

Qualification Executor

8-Dec-21

Date

เอกสารไม่ควบคุม



Performance Qualification

Injector Precision

Instruments:

Instrument Name	Model	Supplier	Serial Number	Moduleware Version
Pump	DX120	Dionex	03010223	3.03
Detector	DX120	Dionex	03010223	3.03
Autosampler	AS40 or man. inj.	Dionex	n.a.	0.00
Eluent Generator	n.a.	Dionex	n.a.	0.00

Accessories

Name	Description
Backpressure Tubing	0.13 mm (0.005") ID PEEK, 13 m (512")
Sample 5	Nitrate, 100 ppm
Eluent	Water

Additional Information

Customer/Company:	UAE Consultant Co., Ltd.	Date:	8-Dec-2021
Qualification Executor:	Mr.Channarong / Archemica	Next Qualification:	Jun-2022

Test Results Summary


Test	Result
Injector Precision (Area %RSD)	PASS



Customer Signature: Suran Qualification Executor: Kannarong Date: 8-Dec-21

เอกสารไม่ควบคุม

Chromatogram of Detector Noise and Drift



Customer Signature: Suran Qualification Executor: Kannarong Date: 8-Dec-21

เอกสารไม่ควบคุม



Performance Qualification

Injector Carry Over

Instruments:

Instrument Name	Model	Supplier	Serial Number	Moduleware Version
Pump	DX120	Dionex	03010223	3.03
Detector	DX120	Dionex	03010223	3.03
Autosampler	AS40 or man. inj.	Dionex	n.a.	0.00
Eluent Generator	n.a.	Dionex	n.a.	0.00

Accessories

Name	Description
Backpressure Tubing	0.13 mm (0.005") ID PEEK, 13 m (512")
Sample 6	Nitrate, 1000 ppm
Blank	Water
Eluent	Water

Additional Information

Customer/Company:	UAE Consultant Co., Ltd.	Date:	8-Dec-2021
Qualification Executor:	Mr.Channarong / Archemica	Next Qualification:	Jun-2022

Test Results Summary

Test	Result
Injector Carry Over (Area %)	PASS

Customer Signature  Qualification Executor  Date 8-Dec-21

เอกสารไม่ควบคุม

Data for Injector Precision test

Name	Area uS*min Nitrate ECD_1
Inj Precision_1	1.665
Inj Precision_2	1.688
Inj Precision_3	1.691
Inj Precision_4	1.698
Inj Precision_5	1.699
Inj Precision_6	1.699
Inj Precision_7	1.689
Inj Precision_8	1.711
Inj Precision_9	1.711
Inj Precision_10	1.700
Average:	1.695
Std. Dev:	0.013
% RSD:	0.783 %
Limit:	1.0 %
Result:	PASS



Customer Signature  Qualification Executor  Date 8-Dec-21

เอกสารไม่ควบคุม



Performance Qualification

Detector Linearity

Instruments:

Instrument Name	Model	Supplier	Serial Number	Moduleware Version
Pump	DX120	Dionex	03010223	3.03
Detector	DX120	Dionex	03010223	3.03
Autosampler	AS40 or man. inj.	Dionex	n.a.	0.00
Eluent Generator	n.a.	Dionex	n.a.	0.00

Accessories

Name	Description
Backpressure Tubing	0.13 mm (0.005") ID PEEK, 13 m (512")
Sample 1	Nitrate, 5 ppm
Sample 2	Nitrate, 10 ppm
Sample 3	Nitrate, 25 ppm
Sample 4	Nitrate, 50 ppm
Sample 5	Nitrate, 100 ppm
Eluent	Water

Additional Information

Customer/Company:	UAE Consultant Co.,Ltd.	Date:	8-Dec-2021
Qualification Executor:	Mr.Channarong / Archemica	Next Qualification:	Jun-2022


Test Results Summary

Test	Result
DX120 Detector Linearity (Corr.)	PASS
DX120 Detector Linearity (%RSD)	PASS

Customer Signature:  Qualification Executor:  Date: 8-Dec-21

เอกสารไม่ควบคุม

Chromatogram for Carry Over test



Data for Carry Over test

Name	Ret. Time (detected) min Nitrate ECD_1	Area uS*min Nitrate ECD_1
High Level	0.18	30.991
Carryover	0.18	0.028
Water	0.18	0.022
Carry over:		0.070 %
Limit:		0.1 %
Result:		PASS



Customer Signature:  Qualification Executor:  Date: 8-Dec-21


เอกสารไม่ควบคุม

Data for Detector Linearity

Name	Amount ppm Nitrate ECD_1	Height uS Nitrate ECD_1
Detector linearity_1	5.000	3.962
Detector linearity_2	10.000	7.384
Detector linearity_3	25.000	17.462
Detector linearity_4	50.000	34.143
Detector linearity_5	100.000	67.127


CERTIFICATE

Linearity Plot



Calibration Type	Number of Points	Offset	Slope
Lin	5	0.000	0.675

Linearity:	Correlation Coefficient	% RSD
	1.000	2.260
Limit:	0.999	5
Result:	PASS	PASS



Customer Signature: Siam Date: 8-Dec-21
Qualification Executor: K. Khamwong

Certificate of Completion

This certifies that

Channarong Khiao-Un

Has successfully completed

eLearn: RPG IC-Specific Qualification Service Training



Channarong

8-Dec-21

Valid for 3 years from:

Nov/19/2021

Issued electronically and approved by:

TFS - Learning Management System, Training, Mentoring, and Certification Group
tmc.training@thermofisher.com

ฉบับนี้ใช้ได้เฉพาะ

The world leader in serving science

Important note: The certificate is only valid during employment with the Thermo Fisher Scientific including its subsidiaries and certified contractors.

Better Separations Through
Better Chemistry

Certificate of Analysis

Dionex Nitrate OQ/PQ IC Standards Kit (Set of 6)

Product Number 060254
Certificate of Analysis

Lot Number 210719

Expiration of Certification
July 2022

The Dionex Nitrate Standard was developed to aid the analysis of anions by Ion Chromatography (IC). The single-ion standard was prepared by the dissolution of high-purity salt in ≥ 18.2 megohm deionized water, which was tested by IC for ionic contaminants. The bottle label states the nominal concentration value of the ionic component for informational purposes only. The actual ion concentration value was determined by Ion Chromatography. The IC system was standardized using the National Institute of Standards & Technology (NIST), Standard Reference Material, SRM 3185 (Nitrate Standard Solution). Actual concentration values determined for the single-ion is listed below.



Dionex Nitrate Standard

Vial #	Concentration (mg/L)
1	5.07 \pm 0.03
2	9.95 \pm 0.07
3	24.49 \pm 0.10
4	49.16 \pm 0.13
5	99.0 \pm 1
6	993 \pm 4

Channarong

8-Dec-21

The concentration value is based a proven reliable method of analysis. The estimated uncertainties are two standard deviations of the concentration value. The concentration value is warranted to be stable for one year from the date of manufacture.

The preparation and analyses of the Dionex Nitrate Standard was performed with extreme care by Thermo Scientific Corporation Consumables Manufacturing Department in Sunnyvale California.

Document No. 078690-01 20-Dec-2011

thermoscientific.com/dionex

© 2011 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific Inc. or its subsidiaries. All other trademarks are the property of their respective owners. This document is for informational purposes only. It is not intended to be used as a substitute for a contract. Please consult your local sales representative for details.

XX21149-EN 0216S 031318-10

Thermo Fisher Scientific
1228 Tinn Way
P.O. Box 3003
Sunnyvale, CA 94088-3803
(408) 732-0700

thermo
scientific

เอกสารไม่ควบคุม

REPORT OF CALIBRATION

Page 2 of 5

Certificate No. : SP22-016

Environment Condition : Ambient Temperature 25 ± 5 °C

Relative humidity 55 ± 20 %RH

Calibration method : In-house method CP-01 Based on ASTM E275-08

Certified Reference Materials :

Material	Serial No.	Certificate No.	Due date
Absorbance Standard set	25760	95935	22 October 2023
Absorbance Standard set	25757	95929	22 October 2023
Wavelength Standard set	25806	95916	22 October 2023
Wavelength Standard set	25758	95915	22 October 2023

Traceability : This certification is traceable to the International System of Unit maintained at National -
Institute of Standards and Technology (NIST) through Starna Scientific Limited

Spectral Band Width of UUC : 1.5 nm.

Scan Speed of UUC : 90 nm/min

Scan Interval of UUC : 0.15 nm.

Resolution of UUC : Photometric 0.0001 Abs.

Wavelength 0.1 nm.

CERTIFICATE OF CALIBRATION

Page 1 of 5

Certificate No. : SP22-016

Customer : United Analyst and Engineering Consultant Co.,Ltd. (Head Office)

Address : 3 Soi Udomsuk 41, Sukhumvit Road, Bangchak, Phrakhanong,
Bangkok 10260

Location of calibration : Laboratory 315

Equipment : UV-Vis Spectrophotometer

Manufacturer : Agilent Technologies

Model : Cary 60

Serial No. : MY15410009

ID No. : N/A

Received Date : 23 May 2022

Calibration Date : 23 May 2022

Issue Date : 26 May 2022

Condition Instrument : Good

Calibrated by :

ปัทมพร

(Mr.Tanawut Rittidach)

Technical Manager

Approved by :

จุลจิรา

(Ms.Chonthicha Sangngern)

Quality Manager

The calibration result is applied only to the above calibrated item and was found accurate as shown on date and place of calibration only.

The measurement capability of the laboratory and its traceability to recognized national standards and to the unit of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the DQE Services Co., Ltd.

REPORT OF CALIBRATION

Certificate No. : SP22-016

Page 4 of 5

Photometric Accuracy :

Wavelength (nm.)	CRMs Values (Abs)	UUC Reading (Abs)	Correction (Abs)	Uncertainty (Abs)	Coverage factor k
235	0.0000 0.7478	0.0001 0.7421	-0.0001 0.0057	0.0050 0.0056	2.00 2.00
257	0.0000 0.8686	0.0000 0.8619	0.0000 0.0067	0.0050 0.0059	2.00 2.00
313	0.0000 0.2912	0.0000 0.2896	0.0000 0.0016	0.0050 0.0051	2.00 2.00
350	0.0000 0.6448	0.0000 0.6403	0.0000 0.0045	0.0050 0.0055	2.00 2.00

REPORT OF CALIBRATION

Certificate No. : SP22-016

Page 3 of 5

Calibration Results : Without adjustment

Photometric Accuracy :

Wavelength (nm.)	CRMs Values (Abs)	UUC Reading (Abs)	Correction (Abs)	Uncertainty (Abs)	Coverage factor k
420	0.0000 0.5787 1.0490 2.1900	0.0000 0.5755 1.0436 2.1847	0.0000 0.0032 0.0054 0.0053	0.0028 0.0031 0.0029 0.0075	2.00 2.00 2.00 2.00
440	0.0000 0.5607 1.0247 2.1229	0.0000 0.5588 1.0232 2.1211	0.0000 0.0019 0.0015 0.0018	0.0028 0.0034 0.0035 0.0082	2.00 2.00 2.00 2.00
465	0.0000 0.5236 0.9634 1.9763	0.0000 0.5197 0.9625 1.9752	0.0000 0.0039 0.0009 0.0011	0.0028 0.0029 0.0028 0.0070	2.00 2.00 2.00 2.00
546.1	0.0000 0.5191 1.0003 1.9987	-0.0001 0.5171 0.9984 1.9946	0.0001 0.0020 0.0019 0.0041	0.0028 0.0031 0.0033 0.0084	2.00 2.00 2.00 2.00
590	0.0000 0.5523 1.0809 2.0391	0.0000 0.5509 1.0799 2.0329	0.0000 0.0014 0.0010 0.0062	0.0028 0.0030 0.0029 0.0080	2.00 2.00 2.00 2.00
635	0.0000 0.5601 1.0512 1.9294	0.0000 0.5584 1.0498 1.9265	0.0000 0.0017 0.0014 0.0029	0.0028 0.0031 0.0029 0.0082	2.00 2.00 2.00 2.00

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม

CERTIFICATE OF CALIBRATION

Certificate No. : SP22-007

Page 1 of 5

Customer : United Analyst and Engineering Consultant Co., Ltd. (Head Office)

Address : 3 Soi Udomsuk 41, Sukhumvit Road, Bangchak, Phrakhanong,

Bangkok 10260

Location of calibration : Laboratory 315

Equipment : UV-Vis Spectrophotometer

Manufacturer : Hitachi

Model : U-1900

Serial No. : 2021-064

ID No. : UAE.WAS.006/2552

Received Date : 20 January 2022

Calibration Date : 20 January 2022

Issue Date : 24 January 2022

Condition Instrument : Good

Calibrated by : จตุพร Approved by : ชลธิชา
 (Mr. Tanawat Rittidach) (Ms. Chonbicha Sangngern)
 Technical Manager Quality Manager

The calibration result is applied only to the above calibrated item and was found accurate as shown on date and place of calibration only.

The measurement capability of the laboratory and its traceability to recognized national standards and to the unit of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the DQE Services Co., Ltd.

FM-708-02 R01 1/11/2021

เอกสารไม่ควบคุม

REPORT OF CALIBRATION

Certificate No. : SP22-016

Page 5 of 5

Wavelength Accuracy :

CRMs Values (nm.)	UUC Reading (nm.)	Correction (nm.)	Uncertainty (nm.)	Coverage factor k
241.72	242.0	-0.28	0.18	2.00
279.45	279.5	-0.05	0.18	2.00
287.81	287.5	0.31	0.18	2.00
334.06	333.5	0.56	0.18	2.00
360.93	360.5	0.43	0.18	2.00
418.59	418.0	0.59	0.18	2.00
445.94	445.4	0.54	0.18	2.00
453.66	453.2	0.46	0.18	2.00
460.02	459.7	0.32	0.18	2.00
536.59	536.2	0.39	0.18	2.00
637.98	638.3	-0.32	0.18	2.00
431.38	431.0	0.38	0.18	2.00
472.50	472.5	0.00	0.18	2.00
513.47	513.5	-0.03	0.18	2.00
528.88	528.5	0.38	0.18	2.00
573.17	573.0	0.17	0.18	2.00
585.35	585.0	0.35	0.20	2.00
684.40	684.7	-0.30	0.18	2.00
740.72	740.8	-0.08	0.20	2.00
748.55	748.5	0.05	0.18	2.00
807.03	807.3	-0.27	0.18	2.00
879.28	879.0	0.28	0.18	2.00

Remark : - UUC = Unit Under Calibration

- N/A - Not Available

- The result expanded uncertainty of measurement U is stated as the standard uncertainty of measurement multiplied by the coverage factor k.

which for a normal distribution corresponds to a coverage probability of approximately 95%

- * Indicates non TISI accredited

- End of Certificate -

เอกสารไม่ควบคุม

FM-708-02 R01 1/11/2021

REPORT OF CALIBRATION

Certificate No. : SP22-007

Page 3 of 5

Calibration Results : Without adjustment

Photometric Accuracy :

Wavelength (nm.)	CRMs Values (Abs)	UUC Reading (Abs)	Correction (Abs)	Uncertainty (Abs)	Coverage factor <i>k</i>
420	0.0000	0.000	0.0000	0.0028	2.00
	0.5787	0.577	0.0017	0.0031	2.00
	1.0490	1.050	-0.0010	0.0029	2.00
	2.1900	2.183	0.0070	0.0080	2.00
440	0.0000	0.000	0.0000	0.0028	2.00
	0.5607	0.560	0.0007	0.0034	2.00
	1.0247	1.023	0.0017	0.0035	2.00
	2.1229	2.118	0.0049	0.0079	2.00
465	0.0000	0.000	0.0000	0.0028	2.00
	0.5236	0.521	0.0026	0.0030	2.00
	0.9634	0.963	0.0004	0.0029	2.00
	1.9763	1.974	0.0023	0.0070	2.00
546.1	0.0000	0.000	0.0000	0.0028	2.00
	0.5191	0.518	0.0011	0.0031	2.00
	1.0003	1.000	0.0003	0.0033	2.00
	1.9987	1.996	0.0027	0.0084	2.00
590	0.0000	0.000	0.0000	0.0028	2.00
	0.5523	0.552	0.0003	0.0030	2.00
	1.0809	1.082	-0.0011	0.0030	2.00
	2.0391	2.033	0.0061	0.0079	2.00
635	0.0000	0.000	0.0000	0.0028	2.00
	0.5601	0.562	-0.0019	0.0031	2.00
	1.0512	1.052	-0.0008	0.0030	2.00
	1.9294	1.925	0.0044	0.0079	2.00

REPORT OF CALIBRATION

Certificate No. : SP22-007

Page 2 of 5

Environment Condition : Ambient Temperature 25 ± 5 °C

Relative humidity 55 ± 20 %RH

Calibration method : In-house method CP-01 Based on ASTM E275-08

Certified Reference Materials :

Material	Serial No.	Certificate No.	Due date
Absorbance Standard set	25760	95935	22 October 2023
Absorbance Standard set	25757	95929	22 October 2023
Wavelength Standard set	25806	95916	22 October 2023
Wavelength Standard set	25758	95915	22 October 2023

Traceability This certification is traceable to the International System of Unit maintained at National -

Institute of Standards and Technology (NIST) through Siama Scientific Limited

Spectral Band Width of UUC : 4.0 nm.

Scan Speed of UUC : 200 nm/min

Scan Interval of UUC : 0.1 nm.

Resolution of UUC : Photometric 0.001 Abs.

Wavelength 0.1 nm.

REPORT OF CALIBRATION

Certificate No. : SP22-007

Page 5 of 5

Wavelength Accuracy :

CRMs Values (nm.)	UUC Reading (nm.)	Correction (nm.)	Uncertainty (nm.)	Coverage factor <i>k</i>
241.54	240.8	0.74	0.18	2.00
279.40	278.5	0.90	0.18	2.00
288.70	288.0	0.70	0.18	2.00
334.22	333.5	0.72	0.18	2.00
361.26	360.5	0.76	0.18	2.00
418.48	418.0	0.48	0.18	2.00
446.70	446.0	0.70	0.18	2.00
453.20	453.0	0.20	0.18	2.00
460.06	459.5	0.56	0.18	2.00
536.90	536.0	0.90	0.18	2.00
637.94	637.2	0.74	0.18	2.00
440.74	440.0	0.74	0.18	2.00
472.22	471.6	0.62	0.18	2.00
513.70	513.0	0.70	0.18	2.00
528.72	528.0	0.72	0.18	2.00
574.60	573.8	0.80	0.18	2.00
585.48	584.6	0.88	0.20	2.00
684.63	684.0	0.63	0.18	2.00
740.27	739.8	0.47	0.20	2.00
748.28	747.8	0.48	0.18	2.00
807.16	806.4	0.76	0.18	2.00
879.70	878.8	0.90	0.18	2.00

Remark : - UUC = Unit Under Calibration

- N/A = Not Available

- The result expanded uncertainty of measurement U is stated as the standard uncertainty of measurement multiplied by the coverage factor *k*, which for a normal distribution corresponds to a coverage probability of approximately 95%

- * Indicates non TISI accredited

- End of Certificate -

REPORT OF CALIBRATION

Certificate No. : SP22-007

Page 4 of 5

Photometric Accuracy :

Wavelength (nm.)	CRMs Values (Abs)	UUC Reading (Abs)	Correction (Abs)	Uncertainty (Abs)	Coverage factor <i>k</i>
235	0.0000	0.000	0.0000	0.0050	2.00
	0.7478	0.746	0.0018	0.0057	2.00
257	0.0000	0.000	0.0000	0.0050	2.00
	0.8686	0.861	0.0076	0.0059	2.00
313	0.0000	0.000	0.0000	0.0050	2.00
	0.2912	0.291	0.0002	0.0051	2.00
350	0.0000	0.000	0.0000	0.0050	2.00
	0.6448	0.638	0.0068	0.0055	2.00

REPORT OF CALIBRATION

Certificate No. : SP22-008 Page 2 of 5

Environment Condition : Ambient Temperature 25 ± 5 °C

Relative humidity 55 ± 20 %RH

Calibration method : In-house method CP-01 Based on ASTM E275-08

Certified Reference Materials :

Material	Serial No.	Certificate No.	Due date
Absorbance Standard set	25760	95935	22 October 2023
Absorbance Standard set	25757	95929	22 October 2023
Wavelength Standard set	25806	95916	22 October 2023
Wavelength Standard set	25758	95915	22 October 2023

Traceability This certification is traceable to the International System of Unit maintained at National -

Institute of Standards and Technology (NIST) through Sarna Scientific Limited

Spectral Band Width of UUC : 1.5 nm.

Scan Speed of UUC : 200 nm/min

Scan Interval of UUC : 0.1 nm.

Resolution of UUC : Photometric 0.001 Abs.

Wavelength 0.1 nm.

CERTIFICATE OF CALIBRATION

Certificate No. : SP22-008 Page 1 of 5

Customer : United Analyst and Engineering Consultant Co.,Ltd. (Head Office)

Address : 3 Soi Udomsuk 41, Sukhumvit Road, Banglehak, Phrakhanong,

Bangkok 10260

Location of calibration : Laboratory 213

Equipment : UV-Vis Spectrophotometer

Manufacturer : Hitachi

Model : U-2900

Serial No. : 21E22-009

ID No. : UAE.WAT.051/2564

Received Date : 20 January 2022

Calibration Date : 20 January 2022

Issue Date : 24 January 2022

Condition Instrument : Good

Calibrated by : 

(Mr. Tanawut Ritidach)

Technical Manager

Approved by :

(Ms. Chonhicha Sangngern)

Quality Manager

The calibration result is applied only to the above calibrated item and was found accurate as shown on date and place of calibration only.

The measurement capability of the laboratory and its traceability to recognized national standards and to the unit of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the DQE Services Co., Ltd.



Request No. 25-65 / 0398

MTC. ACL.No. 486 / 65

CALIBRATION CERTIFICATE

NOMENCLATURE : 1. Atomic Absorption Spectrophotometer "Agilent Technologies"

Model AA240FS, Serial No. MY13160001

2. Working standard solution "Inorganic Ventures"

Multi Analyte Custom Grade Solution, Lot No. P2-MEB675610

SUBMITTED BY : United Analyst and Engineering Consultant Co., Ltd.

3 Soi Udomsuk41, Sukhumvit Road, Bangchak, Prakanong, Bangkok 10260

CALIBRATION PROCEDURE : 1. Performance Verification of Atomic Absorption Spectrophotometer
(WI-500-02-30)

2. Estimation Uncertainty of Measurement in Analytical Chemistry (QP-513)

REFERENCE MATERIAL : Traceable to NIST "Agilent Technologies", "Carlo Erba"

Cadmium Lot No. 0108047046, Chromium Lot No. 0106315418, Copper Lot No. 0107480530, Iron Lot No. 0104697566,

Lead Lot No. 0104659473, Manganese Lot No. T109228A, Nickel Lot No. 0104978044, Zinc Lot No. 0100792297

CALIBRATION RANGE: 0.02,0.10,0.30,0.50,0.70 mg/l at 228.8 nm.Cd, 0.10,0.20,0.30,0.50,0.70 mg/l at 357.9 nm.Cr,

0.05,0.10,0.30,0.50,0.70 mg/l at 324.7 nm.Cu, 0.10,0.30,0.50,0.70,1.00 mg/l at 248.3 nm.Fe, 0.20,0.50,0.70,1.00,1.50 mg/l

at 217.0 nm.Pb, 0.05,0.10,0.30,0.50,0.70 mg/l at 279.5 nm.Mn, 0.10,0.30,0.50,0.70,1.00 mg/l at 232.0 nm.Ni,

0.05,0.10,0.30,0.50,0.70 mg/l at 213.9 nm.Zn

AMBIENT CONDITIONS : Temperature 22 °C Relative humidity 60 %

The Atomic Absorption Spectrophotometer set has been calibrated against Reference Material traceable to National Institute of Standards and Technology (NIST) by The Analytical Chemistry Laboratory. The results are attached herewith.



Calibrated by
(Mr. Danai Srithongkum)

Approved by
(Mrs. Thippaya Junvee Fortune)

Director of Analytical Chemistry Laboratory

Ref. 2025265020400522001

Calibration Date : 3 February 2022

The results relate only to the items tested/calibrated or value assigned.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpa@tistr.or.th Website:www.tistr.or.th

Office/Laboratory

Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office

196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2257 8599
Fax. (66) 0 2257 8599
E-mail : sumalee@tistr.or.th

FM.BL.MTC.002 Rev.4

เอกสารไม่ควบคุม

DQE Services Co.,Ltd.

32 Soi Ladprao-Wanghin 55, Ladprao-Wanghin Rd., Ladprao, Bangkok 10230

Phone : +66 (0)2 538 2054, Email : dqeservicesinfo@gmail.com



REPORT OF CALIBRATION

Certificate No. : SP22-008

Page 5 of 5

Wavelength Accuracy :

CRMs Values (nm.)	UUC Reading (nm.)	Correction (nm.)	Uncertainty (nm.)	Coverage factor k
241.72	241.0	0.72	0.18	2.00
279.45	279.0	0.45	0.18	2.00
287.81	287.0	0.81	0.18	2.00
334.06	333.5	0.56	0.18	2.00
360.93	360.0	0.93	0.18	2.00
418.59	418.0	0.59	0.18	2.00
445.94	445.5	0.44	0.18	2.00
453.66	453.0	0.66	0.18	2.00
460.02	459.5	0.52	0.18	2.00
536.59	536.0	0.59	0.18	2.00
637.98	637.5	0.48	0.18	2.00
431.38	431.0	0.38	0.18	2.00
472.50	472.0	0.50	0.18	2.00
513.47	513.0	0.47	0.18	2.00
528.88	528.5	0.38	0.18	2.00
573.17	573.0	0.17	0.18	2.00
585.35	585.0	0.35	0.20	2.00
684.40	684.0	0.40	0.18	2.00
740.72	740.5	0.22	0.20	2.00
748.55	748.5	0.05	0.18	2.00
807.03	807.0	0.03	0.18	2.00
879.28	879.5	-0.22	0.18	2.00

Remark : - UUC = Unit Under Calibration

- N/A = Not Available

- The result expanded uncertainty of measurement U is stated as the standard uncertainty of measurement multiplied by the coverage factor k,

which for a normal distribution corresponds to a coverage probability of approximately 95%

- * Indicates non TISf accredited

- End of Certificate -

FM-708-02 R01 1/1/2021

เอกสารไม่ควบคุม



Request No. 25-65 / 0398

2 / 5

MTC. ACL. No. 486 / 65

2. Precision

Element	Conc. (mg/l)	Absorbance										Ave. Abs.	SD	%RSD
Cd	0.02	0.0074	0.0062	0.0065	0.0062	0.0070	0.0068	0.0070	0.0065	0.0065	0.0069	0.007	0.0004	5.76
	0.30	0.0952	0.0959	0.0951	0.0957	0.0952	0.0950	0.0952	0.0948	0.0956	0.0943	0.095	0.0005	0.49
	0.70	0.2213	0.2180	0.2203	0.2208	0.2234	0.2211	0.2196	0.2219	0.2201	0.2194	0.221	0.0015	0.67
Cr	0.10	0.0096	0.0098	0.0097	0.0102	0.0106	0.0097	0.0098	0.0099	0.0103	0.0093	0.010	0.0004	3.83
	0.30	0.0309	0.0302	0.0300	0.0316	0.0306	0.0299	0.0309	0.0297	0.0311	0.0296	0.030	0.0007	2.20
	0.70	0.0659	0.0667	0.0664	0.0648	0.0656	0.0662	0.0658	0.0638	0.0674	0.0669	0.066	0.0011	1.70
Cu	0.05	0.0080	0.0075	0.0078	0.0075	0.0077	0.0081	0.0080	0.0075	0.0074	0.0076	0.008	0.0003	3.26
	0.30	0.0417	0.0419	0.0412	0.0421	0.0424	0.0420	0.0423	0.0403	0.0418	0.0415	0.042	0.0006	1.47
	0.70	0.0969	0.0965	0.0972	0.0957	0.0961	0.0958	0.0961	0.0963	0.0959	0.0972	0.096	0.0006	0.58
Fe	0.10	0.0090	0.0105	0.0078	0.0099	0.0091	0.0093	0.0096	0.0094	0.0093	0.0084	0.009	0.0007	8.11
	0.50	0.0462	0.0470	0.0464	0.0464	0.0467	0.0462	0.0467	0.0460	0.0468	0.0466	0.047	0.0003	0.67
	1.00	0.0867	0.0886	0.0910	0.0892	0.0897	0.0873	0.0892	0.0885	0.0888	0.0874	0.089	0.0013	1.43
Pb	0.20	0.0091	0.0095	0.0088	0.0087	0.0082	0.0094	0.0090	0.0087	0.0082	0.0090	0.009	0.0004	4.94
	0.70	0.0322	0.0321	0.0324	0.0318	0.0335	0.0326	0.0327	0.0315	0.0336	0.0321	0.032	0.0007	2.09
	1.50	0.0653	0.0645	0.0663	0.0664	0.0652	0.0671	0.0662	0.0666	0.0657	0.0648	0.066	0.0008	1.28
Mn	0.05	0.0092	0.0092	0.0097	0.0087	0.0085	0.0079	0.0096	0.0085	0.0084	0.0099	0.009	0.0007	7.33
	0.30	0.0616	0.0630	0.0632	0.0633	0.0634	0.0628	0.0640	0.0633	0.0640	0.0629	0.063	0.0007	1.08
	0.70	0.1396	0.1366	0.1386	0.1377	0.1386	0.1386	0.1396	0.1380	0.1374	0.1383	0.138	0.0009	0.67
Ni	0.10	0.0102	0.0092	0.0097	0.0104	0.0091	0.0105	0.0096	0.0098	0.0102	0.010	0.0005	5.22	
	0.50	0.0488	0.0489	0.0489	0.0495	0.0484	0.0490	0.0481	0.0492	0.0495	0.0492	0.049	0.0004	0.91
	1.00	0.0976	0.0979	0.0975	0.0992	0.0977	0.0973	0.0986	0.0962	0.0985	0.0982	0.098	0.0008	0.85
Zn	0.05	0.0340	0.0349	0.0340	0.0352	0.0337	0.0351	0.0344	0.0346	0.0349	0.0343	0.035	0.0005	1.49
	0.30	0.1669	0.1653	0.1628	0.1642	0.1657	0.1637	0.1659	0.1652	0.1654	0.1657	0.165	0.0012	0.72
	0.70	0.3456	0.3467	0.3445	0.3430	0.3422	0.3444	0.3437	0.3438	0.3435	0.3438	0.344	0.0013	0.37

Continue 3 / 5

INDUSTRIAL METROLOGY AND TESTING SERVICE CENTRE

Continue 2 / 5

INDUSTRIAL METROLOGY AND TESTING SERVICE CENTRE

Request No. 25-65 / 0398

MTC. ACL. No. 486 / 65

CALIBRATION DATA

1. Noise Level in term of standard deviation

Element	Cd	Cr	Cu	Fe	Pb	Mn	Ni	Zn
Absorbance	-0.0004	0.0002	0.0007	0.0002	-0.0016	-0.0001	-0.0004	-0.0001
	0.0002	-0.0005	0.0010	0.0007	0.0000	-0.0003	0.0007	-0.0014
	-0.0002	0.0001	0.0008	0.0000	-0.0001	-0.0003	-0.0012	-0.0006
	0.0000	-0.0007	0.0007	0.0000	-0.0005	-0.0004	-0.0004	-0.0012
	0.0001	0.0004	0.0013	0.0014	-0.0001	-0.0001	0.0003	-0.0008
	0.0000	-0.0004	0.0003	-0.0012	-0.0005	-0.0007	-0.0004	-0.0008
	0.0000	-0.0009	0.0009	-0.0002	-0.0010	-0.0008	0.0007	-0.0003
	-0.0004	-0.0003	0.0015	0.0010	-0.0005	-0.0003	-0.0002	-0.0004
	0.0004	0.0008	0.0014	-0.0004	-0.0014	-0.0005	-0.0006	-0.0003
	-0.0006	-0.0013	0.0012	-0.0006	-0.0006	-0.0006	-0.0007	-0.0007
	0.0005	-0.0003	0.0014	-0.0004	-0.0008	-0.0003	-0.0006	-0.0011
	-0.0007	-0.0014	0.0004	-0.0001	-0.0001	0.0000	0.0000	-0.0003
	0.0008	0.0004	0.0005	-0.0006	-0.0008	0.0000	-0.0005	-0.0009
	0.0011	0.0002	0.0005	0.0017	-0.0016	-0.0008	0.0004	-0.0005
	0.0002	0.0010	0.0014	-0.0002	-0.0010	-0.0010	0.0002	-0.0001
Average Absorbance	0.0001	-0.0011	0.0011	-0.0003	-0.0011	-0.0003	-0.0008	-0.0012
	0.0000	-0.0015	0.0009	-0.0010	-0.0011	-0.0013	0.0000	-0.0004
	0.0015	-0.0012	0.0005	0.0002	-0.0017	-0.0001	0.0005	-0.0002
	0.0006	0.0014	0.0010	0.0002	-0.0003	0.0001	-0.0006	-0.0010
	0.0001	0.0003	0.0003	-0.0001	-0.0004	-0.0002	-0.0001	-0.0001
	0.000	0.000	0.001	0.000	-0.001	0.000	0.000	-0.001
	0.0005	0.0008	0.0004	0.0007	0.0005	0.0004	0.0005	0.0004
Standard Deviation	0.0005	0.0008	0.0004	0.0007	0.0005	0.0004	0.0005	0.0004
	0.0005	0.0008	0.0004	0.0007	0.0005	0.0004	0.0005	0.0004
	0.0005	0.0008	0.0004	0.0007	0.0005	0.0004	0.0005	0.0004
	0.0005	0.0008	0.0004	0.0007	0.0005	0.0004	0.0005	0.0004
	0.0005	0.0008	0.0004	0.0007	0.0005	0.0004	0.0005	0.0004
	0.0005	0.0008	0.0004	0.0007	0.0005	0.0004	0.0005	0.0004
	0.0005	0.0008	0.0004	0.0007	0.0005	0.0004	0.0005	0.0004
	0.0005	0.0008	0.0004	0.0007	0.0005	0.0004	0.0005	0.0004
	0.0005	0.0008	0.0004	0.0007	0.0005	0.0004	0.0005	0.0004
	0.0005	0.0008	0.0004	0.0007	0.0005	0.0004	0.0005	0.0004
	0.0005	0.0008	0.0004	0.0007	0.0005	0.0004	0.0005	0.0004
	0.0005	0.0008	0.0004	0.0007	0.0005	0.0004	0.0005	0.0004
	0.0005	0.0008	0.0004	0.0007	0.0005	0.0004	0.0005	0.0004
	0.0005	0.0008	0.0004	0.0007	0.0005	0.0004	0.0005	0.0004
	0.0005	0.0008	0.0004	0.0007	0.0005	0.0004	0.0005	0.0004

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governing TISTR.



MTC. ACL. No. 486 / 65

Element	Standard Value of RM (mg/L)	Reading (mg/L)	Error of Measurement (mg/L)	Error of Measurement (%)	Uncertainty (mg/L)
Fe	0.1003	0.106	0.006	5.68	± 0.008
	0.5015	0.522	0.021	4.09	± 0.017
	1.0030	0.993	-0.010	1.00	$+ 0.032$

Element	Standard Value of RM (mg/l)	Reading (mg/L)	Error of Measurement (mg/l)	Error of Measurement (%)	Uncertainty (mg/l)
Pb	0.1988	0.197	-0.002	0.91	± 0.014
	0.6958	0.722	0.026	3.77	± 0.022
	1.4910	1.463	-0.028	1.88	± 0.041

Element	Standard Value of RM (mg/l)	Reading (mg/l)	Error of Measurement (mg/l)	Error of Measurement (%)	Uncertainty (mg/l)
Mn	0.04955	0.054	0.004	8.98	± 0.004
	0.29730	0.317	0.0197	6.63	± 0.006
	0.69370	0.682	-0.0117	1.69	± 0.012

TIRE

is obtained from the governor of the

หอสารไมควบคั



MTC, ACL, No. 486 / 65

Element	Standard Value of RM (mg/l)	Reading (mg/l)	Error of Measurement (mg/l)	Error of Measurement (%)	Uncertainty (mg/l)
Cd	0.02004	0.019	-0.001	5.19	± 0.004
	0.30060	0.291	-0.010	3.19	± 0.006
	0.70140	0.678	-0.023	3.34	± 0.012

Element	Standard Value of RM (mg/l)	Reading (mg/L)	Error of Measurement (mg/l)	Error of Measurement (%)	Uncertainty (mg/l)
Cr	0.1002	0.101	0.001	0.80	± 0.007
	0.3006	0.298	-0.003	0.86	± 0.012
	0.7014	0.635	-0.066	9.47	± 0.023

Element	Standard Value of RM (mg/l)	Reading (mg/l)	Error of Measurement (mg/l)	Error of Measurement (%)	Uncertainty (mg/l)
Cu	0.0502	0.046	-0.004	8.37	± 0.004
	0.3012	0.295	-0.006	2.06	± 0.010
	0.7028	0.694	-0.009	1.25	± 0.021

NTRE

[illegible]

ออกสารเฝ้าควบคุม

Agilent 5110 and 5100 ICP-OES Preventive Maintenance Checklist



Agilent Preventive Maintenance provides factory recommended service for your analytical systems to assure reliable operation and the accuracy of your results. Delivered by highly-trained and certified service engineers using genuine Agilent parts and supplies, Agilent Preventive Maintenance provides everything you need to reduce unplanned downtime and keep your systems operating at their peak.

For more information about Agilent Technologies services please visit our web site using the following URL <http://www.agilent.com/en-us/services/analytical-instrument-services>

Customer Information

- Customers should provide all necessary operating supplies upon request of the engineer.
- For customers using HF applications, the instrument should be returned to its standard sample introduction system.
- A customer representative should be available to the engineer while performing the preventive maintenance procedures.
- Any parts, not included in the Parts Lists section of this document, are not part of the recommended Preventive Maintenance service, nor are they included in the price of this service.
- If a system requires the use of additional or special procedures and/or parts for the instrument service, then these must be ordered separately and charged as a repair, which may incur additional

Service Engineer's Responsibilities

- Only complete/printout pages that relate to the system being serviced.
- Complete empty fields with the relevant information.
- Complete the relevant checkboxes in the checklist using a "X" or tick mark "✓" in the checkbox.
- Complete Not Applicable check boxes to indicate services not delivered, as needed.
- Complete the PM service in the order of the tasks listed.
- Complete the Service Review section together with the customer.

Issued: 4 March 2021, Revision: A.01

Copyright © 2021
Page 1 of 8

Agilent Technologies

เอกสารไม่ควบคุม

Request No. 25-65 / 0398

5 / 5

MTC. ACL. No. 486 / 65

3.7 Reading on wavelength- Nickel (Ni) at 232.0 nm.

Element	Standard Value of RM (mg/l)	Reading (mg/l)	Error of Measurement (mg/l)	Error of Measurement (%)	Uncertainty (mg/l)
Ni	0.099	0.102	0.003	3.03	± 0.007
	0.495	0.489	-0.006	1.21	± 0.010
	0.990	0.975	-0.015	1.52	± 0.020

3.8 Reading on wavelength- Zinc (Zn) at 213.9 nm.

Element	Standard Value of RM (mg/l)	Reading (mg/l)	Error of Measurement (mg/l)	Error of Measurement (%)	Uncertainty (mg/l)
Zn	0.050	0.050	0.000	0.00	± 0.012
	0.300	0.307	0.007	2.33	± 0.011
	0.700	0.660	-0.040	5.71	± 0.015

Remark : The reported uncertainty is an expanded uncertainty calculated using a coverage factor of 2 (k = 2)
which gives a level of confidence of approximately 95%

Calibrated by.....
(Mr. Danai Srithongkum)

Approved by.....
(Mrs. Thippaya Junvee Fortune)

Director of Analytical Chemistry Laboratory
Calibration date : 3 February 2022

INDUSTRIAL METROLOGY AND TESTING SERVICE CENTRE

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office

35 Mu.3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpal@tistr.or.th Website: www.tistr.or.th

Office/Laboratory

Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office

196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2257 9392
Fax. (66) 0 2257 9392
E-mail : sumalee@tistr.or.th

FM.BL.MTC.002 Rev.4

เอกสารไม่ควบคุม

Agilent 5110 and 5100 ICP-OES
Preventive Maintenance ChecklistGeneral Preparation

- ☒ Discuss any specific questions or issues with the customer prior to starting.
- ☒ Review the instrument logbook.
- ☒ Perform general external inspection of system for cleanliness.
- ☒ Check for proper installation of safety-related parts, assemblies, sensors etc.
- ☒ Check for required firmware/software updates and verify with customers if they would like it installed.
- ☐ For HF application systems, if standard sample introduction system was not installed, ask the customer to install it. *N/A*
- ☒ Run Instrument Performance test and record results in Instrument Performance Test Results Table - Pre PM.

Inspect and clean the system

- ☒ Look for any obvious external damage or problems.
- ☒ Inspect water cooling hoses, gas lines and power cord for excessive wear or damage.
- ☒ Perform a general internal inspection of the system for excessive dust accumulation, clean if necessary.
- ☒ Inspect sample introduction components and record any required maintenance in the Service Engineer Comments and notify the customer as the required actions required.
- ☒ Record the instrument operating conditions in the ICP-OES Status Results Table.
- ☒ Replace the polychromator purge filter.
- ☒ Replace the radial pre-optics window
- ☒ Replace the axial pre-optics window for SVDV and VDV instruments.
- ☒ Check exhaust flow for the correct positive extraction at the exhaust duct to insure they meet minimum specifications.
- ☒ Replace air inlet dust filter.
- ☐ Replace high capacity air inlet dust filter element if installed. *N/A*
- ☒ Remove and clean instrument water inlet filter.

Agilent Water Recirculator

- ☐ Section NOT Applicable
- ☒ Drain cooling fluid and remove any particles from the chiller reservoir
- ☒ Remove, clean, and reinstall water inlet metal mesh filter if present.
- ☒ Re fill with Polyclear Plus cooling fluid.
- ☒ Clean the cooling system Air filter and the condenser.

Agilent 5110 and 5100 ICP-OES
Preventive Maintenance ChecklistSystem Information

Instrument system name and ID	ICP-OES 5110 VOY
Instrument system site and location	UAE Consultant
List system component product numbers	List the serial numbers of each component
1. 63015A	1. MW 19030001
2.	2.
3.	3.
4.	4.
5.	5.
6.	6.
7.	7.
8.	8.
9.	9.
10.	10.

ICP-OES Configuration table	Circle the type or write in the type if other
Nebulizer Type	SeaSpray OneNeb other
Spray Chamber	Cyclonic Single Pass Cyclonic Double Pass other
Torch	Radial Dual View other
Injector Diameter	2.4mm 1.8mm 1.4mm 0.8mm other
Injector Material	Quartz Ceramic other

**Agilent 5110 and 5100 ICP-OES
Preventive Maintenance Checklist**

- ☒ Water Flow
- ☒ Gas Flows
- ☒ RF Generator
- ☒ Camera Test
- ☒ Optics Test
- ☒ Nebulizer Test

Instrument Performance Test Results Table

Note: These measurements do not form part of any specification and are for reference only.

	Pre PM Sensitivity Check		Post PM Sensitivity Check	
	Radial	Axial *	Radial	Axial *
Zn 213.857 nm SRBR	4 041.3	7 856.1	4 141.9	3 910.7
Mn 257.610 nm SRBR	11415.2	30 894.7	11493.6	34460.9
Al 396.152 nm SBR	7.8	15.7	8.7	13.5
K 766.491 nm SBR	5.3	34.9	5.7	44.6

* Axial result is not applicable for G8016AA, G8012AA Radial View instruments.

Instrument Test Results Table

Note: The Instrument Test results are for systems using ICP Expert version 7.3 and above only.

Instrument Test	Result
Subsystem Communications Test	Pass
Air Flow	Pass
Water Flow	Pass
Gas Flows	Pass
RF Generator	Pass
Camera Test	Pass
Optics Test	Pass
Nebulizer test	Pass

**Agilent 5110 and 5100 ICP-OES
Preventive Maintenance Checklist****SPS 3 Auto Sampler****Section NOT Applicable**

- ☐ Power cycle the autosampler and verify successful initialization.
- ☐ Inspect X and Z axis belts for wear. Replace is necessary.
- ☐ Clean X and Z axis slide shafts.
- ☐ Using customer's racks and the Agilent software move the sample probe to the 4 outermost corners and rinse port, ensure that the probe is approximately centered in the vial.

SPS 4 Auto Sampler**Section NOT Applicable**

- ☐ Clean the spill tray, rack location mat, end frames and chassis with a damp soft cloth and diluted mild detergent.
- ☐ Clean the auto sampler cover panels, if cover kit is installed, with domestic window cleaner
- ☐ Check the X-axis and Z-axis drive belts for cracks, splits, damaged teeth, excessive fraying, color changes or degradation from fumes.
- ☐ Check the X-axis, Theta-axis and Z-axis FFC cables for cracks, incorrect positioning, damaged edges or damaged connectors.
- ☐ Pump Tubing Replacement. Replace peristaltic pump tubing. Replace all tubing that goes from the rinse station to the pump and from the pump to the waste/rinse bottles

AVS 4, 6, 7**Section NOT Applicable**

- ☐ Replace valve rotor seal
- ☐ Check fittings for signs of leaks
- ☐ Check tubing including autosampler tubing for kinks or excessive wear
- ☐ Check high flow pump for signs of leaks

Instrument Adjustment

- ☒ Check position of Zn peak, adjust if required.
- ☒ Check Argon Ratio, adjust to specified value if required.
- ☒ Perform Detector Calibration.
- ☒ Perform Instrument Calibration.
- ☒ Run Instrument Performance Test and record results in Instrument Performance Test Results Table - Post PM.
- ☒ For systems using ICP Expert version 7.3 and above run the following Instrument tests and record the result in the Instrument Test Results Table
 - ☒ Subsystem Communications Test
 - ☒ Air Flow

Agilent 5110 and 5100 ICP-OES
Preventive Maintenance Checklist

ICP-OES Parts List Table

Part description	Part Number	Product /Model # where used	Quantity Consumed
Axial Pre-Optic Window	G8010-68014	G8010A, G8011A, G8014A/G8015A	1
Radial Pre-Optic Window	G8010-68015	All	1
Polyclear Plus Cooling Fluid	G3292-80012	Agilent Water Recirculator	-
Purge Gas Filter	G8010-60136	All	1
Air inlet filter	G8000-68002	All	1
High Capacity Air Filter	G8010-60189	Optional	-
Rotor seal for 6-7 port valve for AVS6/7	G8494-60002	G8494A/G8495	-
Rotor seal for 4 port valve for AVS4	G8493-60002	G8493A	-
Rinse solution to rinse station 2.5mm id x 1m	G8410-80123	SPS 4	-
Barb connector 2.5mm-1.5mm ID	G8410-80124	SPS 4	-
PVC waste tubing 8mm od x 5mm id, 2m	G8410-80122	SPS 4	-
Additional Parts may be required from engineers stock:			
X axis drive belt	5410047500	SPS 3	-
Z axis drive belt	5410047400	SPS 3	-
Peristaltic pump tubing, PVC Solvaflex, 3 bridged,	3710049000	SPS 4	-

Restore system

For HF applications, ask the customer to reinstall their sample introduction system.

Leave system in an idle state: on and purging.

Guidance: If the PM service is performed prior to a qualification service, then use the qualification procedure as a guide for final instrument set up and checkout.

Service Review

- ☒ Affix the PM sticker to the system or instrument logbook based on the customer's request.
- ☒ Complete the Service Engineer Comments section below if there are additional comments.

Agilent 5110 and 5100 ICP-OES
Preventive Maintenance Checklist


ICP-OES Status Results Table

Note: These measurements do not form part of any specification and are for reference only.

Measurement	Standby Mode	Plasma On
Mains Voltage	225.317	VAC 144.510
Mains Current	0.214	A 3.237
Instrument Temperature	23.4	°C 23.5
RF Air Flow (sensor speed)	14.0	Hz 14.0
Plasma Exhaust Temperature	No measurement	°C 65.0
Water Flow Oscillator	No measurement	L/min 2.03
Water Flow Detector	0.00	L/min 1.37
Water Inlet Temperature	14.2	°C 13.8
Polychromator Temperature	35.0	°C 35.0
CCD Temperature	26.9	°C -39.7
Thermal Stabilizer	35.0	°C 35.0
Argon Supply Pressure	644.15	kPa 644.54
Purge Gas Supply Pressure*1	641.44	kPa 655.67
Option Gas Supply Pressure*1	-	kPa -
Nebulizer Flow	No measurement	L/min 0.40
Nebulizer Back Pressure	No measurement	kPa 242.05
Plasma Gas Flow	No measurement	L/min 15.00
Auxiliary Gas Flow	No measurement	L/min 1.20
RF Power	No measurement	W 1201.1
RF Supply Current	No measurement	A 8.233
RF Supply Voltage	No measurement	V 144.510

*1 If option installed

Report Summary	
Instrument Model	Agilent 5110/5110 VDV ICP-OES
Instrument ID	G8011A/G8015A
Instrument Serial Number	MY18030001
Software Version	7.3.1.9507
Firmware Version	3442
Tested By	Nukoon L.
Test Completed On	12/9/2021 9:14:59 AM
Result Summary	
Subsystem Communications Test	Skipped
Air Flow Test	Skipped
Water Flow Test	Skipped
Gas Flows Test	Skipped
RF Generator Test	Skipped
Camera Test	Skipped
Optics Test	Skipped
Advanced Valve System Test	Skipped
Resolution Test	Pass
Sensitivity Test	Pass
Precision Test	Pass

**Agilent Technologies**

Agilent 5110 and 5100 ICP-OES Preventive Maintenance Checklist

☒ Review the service and any test results with the customer.
☒ If the Instrument firmware was updated, record the details of the change in the Service Engineer's Comments box below or if necessary, in the customer's IQ records.

Service Engineer Comments (optional)
If there are any specific points you wish to note as part of performing the installation or other items of interest for the customer, please write in this box.

Other Important Customer Web Links
How to get information on your product:
☐ Literature Library - <http://www.agilent.com/en-us/products/icp-oes/icp-oes-systems/5110-icp-oes#literature>
☐ Need to know more? - <http://www.agilent.com/crosslab/university/>
☐ Need technical support, FAQs? - <http://www.agilent.com/en-us/support/landing/icp-oes>
☐ Need supplies? - www.agilent.com/chem/supplies

Service Completion
Service request number 6004337217 Date service completed 09/12/21
Agilent signature Nukoon L. Customer signature Aphorn Onkong
Document part number: G8014-90075

Issued: 4 March 2021, Revision: A.01 Copyright © 2021 Agilent Technologies
Page 8 of 8

Sensitivity Test					Pass	
Radial						
Element Wavelength	Specification	Method	Ratio	Standard	Blank	
As (188.980 nm)	≥ 46.0	SRBR	167.2	1131.3	42.4	
Se (196.026 nm)	≥ 41.0	SRBR	119.1	1177.1	84.2	
Zn (213.857 nm)	≥ 1421.0	SRBR	4082.3	49908.2	148.6	
Pb (220.353 nm)	≥ 48.0	SRBR	191.1	2882.8	172.6	
Mn (257.610 nm)	≥ 3518.0	SRBR	11415.2	265002.2	536.8	
Al (396.152 nm)	≥ 3.4	SBR	7.8	49838.0	5676.5	
Ba (493.408 nm)	≥ 34.0	SBR	116.1	1999041.4	17066.5	
K (766.491 nm)	≥ 1.8	SBR	5.3	101078.4	16104.6	
Axial						
Element Wavelength	Specification	Method	Ratio	Standard	Blank	
As (188.980 nm)	≥ 208.0	SRBR	252.9	3214.2	147.0	
Se (196.026 nm)	≥ 159.0	SRBR	216.2	3839.7	272.2	
Zn (206.200 nm)	≥ 234.0	SRBR	1203.3	14046.1	133.7	
Zn (213.857 nm)	≥ 1743.0	SRBR	7856.1	171323.1	472.9	
Cd (214.439 nm)	≥ 4227.0	SRBR	7054.9	129539.3	335.4	
Pb (220.353 nm)	≥ 320.0	SRBR	531.7	19218.2	566.2	
Mn (257.610 nm)	≥ 10625.0	SRBR	30884.7	1314844.0	1807.4	
Cr (267.716 nm)	≥ 1048.0	SRBR	4442.1	174420.3	1515.1	
Cu (324.754 nm)	≥ 19.0	SBR	50.7	374603.6	7249.0	
Al (396.152 nm)	≥ 6.0	SBR	15.7	279915.3	16790.4	
Ba (493.408 nm)	≥ 50.0	SBR	209.7	10699956.6	51728.3	
K (766.491 nm)	> 24.0	SBR	38.9	1983197.5	49746.6	

Resolution Test				Pass	
Element Wavelength	Specification	Width			
N (174.213 nm)	≤ 9.40	7.27			
As (188.980 nm)	≤ 8.20	6.23			
C (193.027 nm)	≤ 11.50	8.26			
Mo (202.032 nm)	≤ 8.20	6.42			
Cr (206.158 nm)	≤ 13.40	9.27			
Zn (213.857 nm)	≤ 8.70	6.77			
Pb (220.353 nm)	≤ 9.50	7.12			
Co (228.615 nm)	≤ 17.20	11.88			
Ba (230.424 nm)	≤ 9.40	7.36			
Mn (257.610 nm)	≤ 13.30	9.52			
Mn (260.568 nm)	≤ 20.30	14.30			
Cr (267.716 nm)	≤ 11.00	7.99			
Cu (324.754 nm)	≤ 25.00	19.06			
Cu (327.395 nm)	≤ 14.20	11.32			
Sr (338.071 nm)	≤ 33.50	24.39			
Ba (455.403 nm)	≤ 44.00	33.86			
Sr (460.733 nm)	≤ 36.00	17.38			
Ba (493.408 nm)	≤ 36.00	25.53			
Ba (614.171 nm)	≤ 42.00	24.99			
Ar (675.283 nm)	≤ 74.00	59.49			
K (766.491 nm)	≤ 80.00	65.27			

Sensitivity Test				Pass	
Radial					
Element Wavelength	Specification	Method	Ratio	Standard	Blank
As (188.980 nm)	≥ 46.0	SRBR	154.8	1242.3	58.5
Se (196.026 nm)	≥ 41.0	SRBR	117.4	1259.6	97.9
Zn (213.857 nm)	≥ 1421.0	SRBR	4192.8	52402.6	155.3
Pb (220.353 nm)	≥ 46.0	SRBR	196.4	2814.2	179.9
Mn (257.610 nm)	≥ 3518.0	SRBR	11993.6	281210.1	547.6
Al (396.152 nm)	≥ 3.4	SBR	8.7	55103.6	5662.9
Ba (493.408 nm)	≥ 34.0	SBR	125.4	2152916.9	17032.2
K (766.491 nm)	≥ 1.8	SBR	5.7	107906.7	16079.8
Axial					
Element Wavelength	Specification	Method	Ratio	Standard	Blank
As (188.980 nm)	≥ 208.0	SRBR	297.5	4054.8	170.4
Se (196.026 nm)	≥ 159.0	SRBR	260.2	4794.9	298.5
Zn (206.200 nm)	≥ 234.0	SRBR	1305.9	16162.3	150.3
Zn (213.857 nm)	≥ 1743.0	SRBR	8920.7	200915.6	504.7
Cd (214.436 nm)	≥ 4227.0	SRBR	7958.3	149327.5	350.4
Pb (220.353 nm)	≥ 320.0	SRBR	606.7	15244.5	584.0
Mn (257.610 nm)	≥ 10625.0	SRBR	34460.9	1493092.8	1872.5
Cr (267.716 nm)	≥ 1048.0	SRBR	5018.6	198000.6	1632.6
Cu (324.754 nm)	≥ 19.0	SBR	57.5	423663.7	7248.6
Al (396.152 nm)	≥ 6.0	SBR	18.5	320004.9	16441.4
Ba (493.408 nm)	≥ 60.0	SBR	233.3	11862915.4	50714.5
K (766.491 nm)	≥ 24.0	SBR	44.6	2218974.4	48657.9

Pass

Resolution Test

Element Wavelength	Specification	Width
N (174.213 nm)	≤ 9.40	7.22
As (188.980 nm)	≤ 8.20	6.15
C (193.027 nm)	≤ 11.50	8.22
Mo (202.032 nm)	≤ 8.20	6.33
C- (206.158 nm)	≤ 13.40	9.21
Zn (213.857 nm)	≤ 8.70	6.87
Pb (220.353 nm)	≤ 9.50	7.02
Co (228.615 nm)	≤ 17.20	11.81
Ba (230.424 nm)	≤ 9.40	7.46
Mn (257.610 nm)	≤ 13.30	9.49
Mn (260.568 nm)	≤ 20.30	14.19
Cr (267.716 nm)	≤ 11.00	7.90
Cu (324.754 nm)	≤ 25.00	18.92
Cu (327.395 nm)	≤ 14.20	11.32
Sr (338.071 nm)	≤ 33.50	24.29
Ba (455.403 nm)	≤ 44.00	33.68
Sr (460.733 nm)	≤ 36.00	17.64
Ba (493.408 nm)	≤ 36.00	25.56
Ba (614.171 nm)	≤ 47.00	24.75
Al (675.283 nm)	≤ 74.00	59.18
K (766.491 nm)	≤ 80.00	65.19



Calibration Certificate

Certificate No.: 2201793-001-01
Client name: UNITED ANALYST AND ENGINEERING CONSULTANT CO.,LTD.
Address: 3 Soi Udomsuk 41, Sukhumvit Road,
Bangchack, Prakhong, Bangkok 10260

Page 1 of 5

Equipment: pH Meter
Manufacturer: METTLER TOLEDO
Model: SevenEasy pH
Serial No.: 1231155210
ID No.: UAE.WAT.010/2553

Order No.: 2201793
Operation No.: 2201793-001
Date of Receipt: 21 February 2022
Date of Calibration: 1 March 2022

Calibrated by Mr.Pheraphat Tuanjit
Scientist
Approved by 
(Mr.Nuttapol Niyomchart) (fn)
Specialist, Division of Calibration Laboratory
Responsible for the Technical Management Team
Date of Issue: 1 March 2022

The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the National Food Institute.

Gas Flows Test

	Actual Flow	Back Pressure	Auxiliary Target Flow	Actual Flow	Pass
Nebulizer Target Flow	0.70	203.80	2.00	1.99	108.66

	Actual Flow	Back Pressure	Plasma Target Flow	Actual Flow	Back Pressure
Makeup Target Flow	2.00	113.89	18.00	17.93	24.24

RF Generator Test

	Pass
RF Power Supply Test	Passed
RF Power Supply (V)	141.475

	Pass
RF Oscillator Test	Passed
RF Oscillator Frequency (MHz)	25.874
Work Coil Current (A)	45.931
RF Power Supply Current (A)	2.000

Camera Test

	Integration Time (ms)	Standard Deviation	Status
Electronic Offset Test	1000	5.261	Passed
Dark Current Test	6000	0.734	Passed
Array Test	5	0.024	Passed
Linearity Test		0.118	Passed

F-CS-012 Revision: 00 Date: 14-12-61

Calibration Report

Certificate No.:

2201793-001-01

Equipment:

Digital Thermometer with RTD (pH Meter)

Resolution: 0.1 °C Model: SevenEasy pH

Serial No.: 1231155210 ID No.: UAE.WAT.010/2553

Manufacturer: METTLER TOLEDO

Date of Calibration:

1 March 2022

Page 5 of 5

Calibration point:

15.0, 25.0 and 35.0 °C

Calibration result:

- The probe was immersed in liquid bath or dry bath to a minimum depth of 100 mm.

- Description of probe, model: N/A SIN: N/A

Dimension of probe: Diameter 4 mm., Length 100 mm.,

Sheath material: Stainless Steel

UUC* Reading (°C)	Standard Temperature (°C)	Correction Value (°C)	Uncertainty ± (°C)
15.1	15.006	-0.1	0.099
25.1	25.004	-0.1	0.099
35.1	35.003	-0.1	0.099

P. Sanghavit
1 March 2022

Note

* UUC* : Unit Under Calibration

The report uncertainty of measurement was based on standard uncertainty multiplied by coverage factor (k=2, providing a level of confidence of approximately 95 %).

***** End *****

Calibration Report

Certificate No.:

2201793-001-01

Equipment:

Digital Thermometer with RTD (pH Meter)

Resolution: 0.1 °C Model: SevenEasy pH

Serial No.: 1231155210 ID No.: UAE.WAT.010/2553

Manufacturer: METTLER TOLEDO

Date of Calibration:

1 March 2022

Page 4 of 5

Location:

Chemical Calibration Laboratory, NATIONAL FOOD INSTITUTE

Environment Condition:

Ambient Temperature 24 °C ± 1 °C

Relative Humidity 53 % ± 2 %

Condition of this results of Calibration:

1. Calibration Method :
 - In house method: W-TE-025 by comparison with standard thermometer.
 - The Calibration is determined by comparing with a known temperature from a standard resistance thermometer.
 - The temperature scale in use at this laboratory is the International Temperature scale of 1990 (ITS-90).
2. Reference Standard Instrument:

Instrument	Model	Serial No.	Certificate No.	Due Date	Through
HANDHELD THERMOMETER	1523	2118154	PSLT 0851/84	03-Jun-22	TISTR
Platinum Resistance Thermometer (PRT)	5527A	877332			

Support Equipment : - Low Temperature Bath (ISOCAL-6), Model: Europa-6 Plus Basic, SN: 3415922

3. This certificate is traceable to International System of Units (SI Units).

4. This certificate was certified only for the instrument we calibrated.

5. This result of calibration was found accurate as shown on date and place of calibration only.

6. Condition of Calibrated item : Good

7. Result of Calibration : ☒ Without adjustment ☐ After adjustment

P. Sanghavit
1 March 2022

Calibration Report

Certificate No.: 2202093-001-01
Equipment: pH Meter
Resolution: 0.01 pH ; 1 mV
Manufacturer: METTLER TOLEDO
Model: SevenEasy pH
Serial No.: 1230525212
Type: Bench top
ID No.: UAE.WAS.003/2553
Date of Calibration: 16 March 2022
Location: Chemical Calibration Laboratory, National Food Institute.
Ambient Temperature: (23.0 ± 1.5) °C
Relative Humidity: (49.5 ± 5) %
Condition of Equipment: Good Condition
Condition of this Results of Calibration

Page 2 of 5

1. Calibration Method
In house method : W-CC-002 based on direct measurement by using standard voltage calibrator and certified reference material (CRM)

2. Reference Standards / Certified Reference Material

Instruments	Serial / ID No.	Manufacturer	Certificate No.	Due Date
2.1 DC Voltage Calibrator	2709007	Fuke	SCL-21F-0687	24 June 2022
2.2 Digital Thermometer	2709007	Fuke	CC-640599-01	30 October 2022
2.3 Thermo-Hygro Meter	ana.ku.BTH 005/58	PONPE	QR21-2787	15 November 2022

Certified Reference Material	Lot No.	Manufacturer	Ref N	Expiry Date
2.4 pH buffer 4.008 (Primary pH buffer Solution)	780012	CPAchem	PH216.L5	21 November 2023
2.5 pH buffer 6.865 (Primary pH buffer Solution)	780013	CPAchem	PH217.L5	21 November 2023
2.6 pH buffer 10.01 (Primary pH buffer Solution)	780015	CPAchem	PH220.L5	21 November 2022
2.7 pH buffer 7.00 (Standard pH buffer Solution)	776840	CPAchem	PH107.L5	8 November 2022

3. This certification is traceable to The International System of Unit (SI Unit)
3.1 Instruments No.2.1 through
3.2 Instruments No.2.2 through
3.3 Instruments No.2.3 through
3.4 Certified Reference Material No. 2.4 to 2.6 traceable to
NSC-TISI-TIS 17025 Laboratory Accreditation of Calibration No.0075
NSC-TISI-TIS 17025 Laboratory Accreditation of Calibration No.0061
NSC-TISI-TIS 17025 Laboratory Accreditation of Calibration No.0292

Primary measurement method: Harned cell using calibrated thermometer, barometer, and nanovoltmeter. The Standard Solution preparation and certified by CPAchem Ltd is accredited to ISO 17034 and ISO/IEC 17025

3.5 Certified Reference Material No. 2.7 traceable to
BIM RefN HI-7 LoIN 30.04.2020; BIM RefN HI-9 LoIN 28.05.2020; BIM RefN HI-8 LoIN 30.04.2020; BIM RefN HI-10 LoIN 28.05.2020. The Standard Solution preparation and certified by CPAchem Ltd is accredited to ISO 17034 and ISO/IEC 17025

4. This certificate was certified only for the instrument we calibrated.

5. This result of calibration was found accurate as shown on date and place of calibration only.



เอกสารไม่ควบคุม

FCS-012 Revision: 00 Date: 14-12-61

Calibration Certificate

Certificate No.: 2202093-001-01
Client name: UNITED ANALYST AND ENGINEERING CONSULTANT CO.,LTD.
Address: 3 Soi Udomsuk 41, Sukhumvit Road,
Bangchack, Prakanong, Bangkok 10260

Page 1 of 5

Equipment: pH Meter

Manufacturer: METTLER TOLEDO

Model: SevenEasy pH

Serial No.: 1230525212

ID No.: UAE.WAS.003/2553

Order No.: 2202093

Operation No.: 2202093-001

Date of Receipt: 11 March 2022

Date of Calibration: 16 March 2022

Calibrated by Mr.Manas Somsak
Specialist
Approved by 
(Mr.Pheraphat Tuanjit)
Manager, Division of Calibration Laboratory

Date of Issue: 21 March 2022
Responsible for the Technical Management Team

The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the National Food Institute.

FCS-009 Revision: 00 Date: 14-12-61

เอกสารไม่ควบคุม

Calibration Report

Certificate No.: 2202093-001-01
Equipment: Digital Thermometer with RTD (pH Meter)
Resolution: 0.1 °C
Model: SevenEasy pH
Serial No.: 1230525212
ID No.: UAE.WAS.003/2553
Manufacturer: METTLER TOLEDO
Date of Calibration: 16 March 2022

Page 4 of 5

Location: Chemical Calibration Laboratory, National Food Institute.

Environment Condition:

Ambient Temperature (23.0 ± 1.0) °C
Relative Humidity (50 ± 4) %

Condition of this results of Calibration:

1. Calibration Method :
- In house method: W-TE-025 by comparison with standard thermometer.
- The Calibration is determined by comparing with a known temperature from a standard resistance thermometer.
- The temperature scale in use at this laboratory is the International Temperature scale of 1990 (ITS-90).
2. Reference Standard Instrument :

Instrument	Model	Serial No.	Certificate No.	Due Date	Through
HANDHELD THERMOMETER	1523	2118154	PSL-T 0951164	24-Jun-22	TISTR
Platinum Resistance Thermometer (PRT)	5627A	877332			

Support Equipment : - Low Temperature Bath (ISOCAL-6), Model: Europa-6 Plus Basic, S/N: 341592/2

3. This certificate is traceable to International System of Units (SI Units).

4. This certificate was certified only for the instrument we calibrated.

5. This result of calibration was found accurate as shown on date and place of calibration only.

6. Condition of Calibrated Item : Good

7. Result of Calibration : ☒ Without adjustment ☐ After adjustment

Hz

Calibration Report

Certificate No.: 2202093-001-01
Equipment: pH Meter
Resolution: 0.01 pH ; 1 mV
Model: SevenEasy pH
Serial No.: 1230525212
ID No.: UAE.WAS.003/2553
Manufacturer: METTLER TOLEDO
Type: Bench top
Date of Calibration: 16 March 2022

Page 3 of 5

Calibration Results:

1. Calibration of pH Meter (Manual Temperature Compensation at 25 °C)

Nominal pH	DC Voltage Standard (mV)	Average Indicator Reading		Uncertainty (±mV)	Coverage Factor (k)
		mV	pH		
0	414.117	414	0.00	0.58	2.00
2	295.811	296	2.00	0.58	2.00
4	177.462	178	4.00	0.58	2.00
6	59.159	59	6.00	0.58	2.00
7	-0.001	0	7.00	0.58	2.00
8	-59.159	-59	8.00	0.58	2.00
10	-177.463	-177	10.00	0.58	2.00
12	-295.812	-296	12.00	0.58	2.00
14	-414.119	-414	14.00	0.58	2.00

2. Calibration of pH Meter with Electrode (Manual Temperature Compensation at 25 °C)

Equipment: pH Electrode
Manufacturer: METTLER TOLEDO
Serial No.: 9453843
Type: Combined Electrode
Model: InLab Solids
ID No.: N/A

Performance of Electrode system

(Three-Point Calibration at pH4, pH7 and pH10)

Certified Value @25 °C (pH)	Average Indicator Reading		Relative Slope (%)	Uncertainty (± pH)	Coverage Factor (k)
	pH	mV			
4.008	4.01	172	98.1	0.0071	2.00
6.866	6.87	6	-	0.0074	2.00
10.015	10.01	-175	97.4	0.0090	2.00
6.983	6.98	-3	-	0.0092	2.00



NSC-TIS-15 17025
CALIBRATION 0061

Accuracy Calibration Certificate

Customer

Company: United Analyst and Engineering Consultant Co., Ltd.
Address: 3 Soi Udom Suk 41, Sukhumvit Rd., Bang Chak
City: Phra Khanong
Contact: Suwit Chonok
Zip / Postal: 10260
State / Province: Bangkok
Order Number: * 0 3 7 0 1 2 2 6 6 *

Weighing Device

Manufacturer: Mettler Toledo
Model: XSR205DU
Serial No.: C210685394
Building: N/A
Floor: 2
Room: Balance Room
Instrument Type: Weighing Instrument
Asset Number: UAE.WAO.010/2565
Terminal Model: SRAT
Terminal Serial No.: C210685394
Terminal Asset No.: N/A

Range	Max. Capacity	Readability (g)
1	81 g	0.00001 g
2	220 g	0.0001 g

Procedure

Calibration Guideline: EURAMET cg-18 v. 4.0 (11/2015)
METTLER TOLEDO Work Instruction: CP/W002/20

This calibration certificate contains measurements for As Found calibration. No As Left calibration was performed because the device was not modified after As Found calibration. Therefore, results for As Left correspond to As Found.

The sensitivity/span of the weighing instrument was adjusted before calibration with a built-in weight.

In accordance with EURAMET cg-18 (11/2015), the test loads were selected to reflect the specific use of the weighing device or to accommodate specific calibration conditions.

	Temperature		Humidity	
As Found	Start: 22.4 °C	End: 22.4 °C	Start: 47.5 %	End: 46.2 %

As Found Calibration Date: 06-May-2022
As Left Calibration Date: N/A
Issue Date: 09-May-2022
Calibrator: Sirawit Chamdhan

Approved Signatory:

Technical Manager / Head of Calibration Center

Calibration Report

Certificate No.: 2202093-001-01
Equipment: Digital Thermometer with RTD (pH Meter)
Resolution: 0.1 °C
Model: SevenEasy pH
Serial No.: 1230525212
ID No.: UAE.WAS.003/2553
Manufacturer: METTLER TOLEDO
Date of Calibration: 16 March 2022

Page 5 of 5

Calibration point: 15.0, 25.0 and 35.0 °C

Calibration result:

- The probe was immersed in liquid bath or dry bath to a minimum depth of 120 mm.

- Description of probe, model: N/A SN: N/A

Dimension of probe: Diameter 3.5 mm, Length 135 mm,

Sheath material: Stainless Steel

UUC* Reading (°C)	Standard Temperature (°C)	Correction Value (°C)	Uncertainty ± (°C)
15.2	15.001	-0.2	0.099
25.2	25.002	-0.2	0.099
35.2	35.002	-0.2	0.099

Note: - UUC*: Unit Under Calibration

The report uncertainty of measurement was based on standard uncertainty multiplied by coverage factor $k = 2$, providing a level of confidence of approximately 95 %.

----- End -----

F-CS-012 Revision: 00 Date: 14-12-61

เอกสารไม่ควบคุม

Error of Indication

As Found				
Reference Value	Indication	Error of Indication	Expanded Uncertainty	k
1 0.00000 g	0.00000 g	0.00000 g	0.020 mg	2
2 0.05000 g	0.05001 g	0.00001 g	0.023 mg	2
3 0.10001 g	0.10001 g	0.00000 g	0.025 mg	2
4 1.00001 g	1.00001 g	0.00001 g	0.034 mg	2
5 5.00001 g	5.00001 g	0.00000 g	0.049 mg	2
6 20.00002 g	20.00002 g	0.00000 g	0.082 mg	2
7* 50.00000 g	50.00002 g	0.00002 g	0.12 mg	2
8 80.00004 g	80.00009 g	0.00005 g	0.25 mg	2
9 100.0000 g	100.0000 g	0.0000 g	0.20 mg	2
10* 150.0000 g	150.0000 g	0.0000 g	0.31 mg	2
11 200.0000 g	199.9999 g	-0.0001 g	0.35 mg	2

*The calculated uncertainty was replaced by the CMC (Calibration and Measurement Capabilities) value because the calculated uncertainty was smaller than the CMC value.



For improved legibility of the graphics only increasing measurement points are shown and measurement points close to zero are not displayed.

The uncertainty stated is the expanded uncertainty at calibration obtained by multiplying the standard combined uncertainty by the coverage factor k – which can be larger than 2 according to EURAMET cg-18. The value of the measurand lies within the assigned range of values with a probability of approximately 95%.

The user is responsible for maintaining environmental conditions and the settings of the weighing instrument when it was calibrated.

Test Equipment

All weights used for metrological testing are traceable to national or international standards. The weights were calibrated and certified by an accredited calibration laboratory.

Weight Set 1: OIML E2


Weight Set No.:	WS54	Date of Issue:	17-Nov-2020
Certificate Number:	170240	Calibration Due Date:	15-May-2022

Thermo Hygrometer

Equipment No.:	IN161	Date of Issue:	14-Jun-2021
Certificate Number:	21H1220	Calibration Due Date:	01-Jun-2022

Measurement Results

Repeatability



The "g" in the graph represents the readability of the range interval in which the test was performed.
The results of this graph are based upon the absolute values of the differences from the mean values.

Eccentricity

Test Load: 100 g

Position	As Found	As Left
1	100.0000 g	N/A
2	100.0000 g	N/A
3	100.0000 g	N/A
4	100.0000 g	N/A
5	100.0000 g	N/A

Maximum Deviation: 0.0000 g

The "g" in the graph represents the readability of the range interval in which the test was performed.

Measurement Uncertainty of the Weighing Instrument in Use

Stated is the expanded uncertainty with k=2 in use. The formula shall be used for the estimation of the uncertainty under consideration of the errors of indication. The value R represents the net load indication in the unit of measure of the device.

Temperature coefficient for the evaluation of the measurement uncertainty in use: 1.5 · 10⁻⁵ / K

Temperature range on site for the evaluation of the measurement uncertainty in use: 3 K


Linearization of Uncertainty Equation

Range		As Found		As Left	
d	Max				
1	0.00001 g	81 g	$U_1 = 0.021 \text{ mg} + 0.00450 \text{ mg/g} \cdot R$	N/A	N/A
2	0.0001 g	220 g	$U_2 = 0.06 \text{ mg} + 0.00448 \text{ mg/g} \cdot R$	N/A	N/A

To optimize the stability of the linearization, besides of the zero load only increasing measurement points with a test load of 5% of the measurement range or larger are taken for the calculation of the linear equation.

Absolute and Relative Measurement Uncertainty in Use for Various Net Indications (Examples)

Net Indication	As Found	As Left
0.00220 g	0.021 mg	N/A
0.02200 g	0.021 mg	N/A
0.22000 g	0.022 mg	N/A
2.20000 g	0.031 mg	N/A
220.0000 g	1.0 mg	N/A



The weighing range shown in the absolute uncertainty graph refers to the first interval/range of the device.

Remarks

- FACT adjustment functionality activated
- Equipment condition: Good
- Calibration after installation
- Next calibration according to customer's procedure
- Calibration data not decide by calibration laboratory

End of Accredited Section

The information below and any attachments to this calibration certificate are not part of the accredited calibration.



Equipment : Hot Air Oven
Condition As-Received : Used Item
Reference : 2204-00150C-1

Procedure Used :-

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD) and Thermocouple Type T.

The temperature scale used was based on ITS-90.

Condition of this result of calibration


1. Reference standard instrument:-
- | Instrument | Model | Serial No. | Cert. No. | Due Date |
|----------------------|--------|------------|-----------|-------------|
| 1) Data Acquisition | 34970A | MY41021843 | 22LM4 | 10 Jan 2023 |
2. This certificate is valid only to the item calibrated on date and place of calibration.
3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Close

Environment during calibration		
	Beginning	Finished
Temp. (°C)	28	28
REL.Humid. (%)	56	55
AC Supply (Volt)	221	224



Probe Installation Details :

Dimension of Chamber :

a = 5.0 cm D = 0.50 m
b = 5.0 cm W = 0.80 m
c = 5.0 cm H = 0.75 m
Capacity = 0.30 m³

Ref. Std. ID No.: @ Calibration Point (°C)		
Position :	(120 , 180)	(104)
1	21-04TC-01	18-04RTD-01
2	21-04TC-02	18-04RTD-02
3	21-04TC-03	18-04RTD-03
4	21-04TC-04	18-04RTD-04
5	21-04TC-05	18-04RTD-05
6	21-04TC-06	18-04RTD-06
7	21-04TC-07	18-04RTD-07
8	21-04TC-08	18-04RTD-08
9 (ref.)	21-04TC-09	18-04RTD-09

Man Pattanapongpaiboon

เอกสารไม่ควบคุม

a 1104316



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert. No.: 22TM304
Page.: 1 of 3

Certificate of Calibration

Equipment : Hot Air Oven
Manufacturer : Memmert
Model : UF 55
Serial No. : B212.0411
ID No. : UAE.WAO.005/2556
Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
3 Soi Udomsuk 41, Sukhumvit Road,
Bangchak, Phrakhanong,
Bangkok 10260
Location : Lab Floor 2
Received Order : 7 April 2022
Calibration Date : 7 April 2022
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %
Calibrated by : Man Pattanapongpaiboon
Approved by : Man Pattanapongpaiboon
() Ponthippa Tameyakul
() Malee Butkruea
() Suwit Imjai

Issue Date : 18 April 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written

Approval of the head of Corporate Services 3 - Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม

A 0040245



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-27 FAX. 0-2719-9484



NSC-TISI-TIS17025
CALIBRATION 0098

Cert. No.: 22TM90
Page.: 1 of 3

Certificate of Calibration

Equipment : BOD Incubator
Manufacturer : Arco
Model : UC4-1320
Serial No. : 13URC4S013201
ID No. : UAE.WAO.015/2561

Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
3 Soi Udomsuk 41, Sukhumvit Road,
Bangchak, Phrakhanong,
Bangkok 10260
Location : Lab Floor 2

Received Order : 17 February 2022
Calibration Date : 17 February 2022
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %

Calibrated by : Kunchit Promprat

Approved by :  Approved Signatory

() Ponthippa Tameyakul
() Malee Butkruea
() Suwit Imjai

Issue Date : 22 February 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม
A 0038099



Equipment : Hot Air Oven
Condition As-Received : Used Item
Reference : 2204-0015OC-1
Result of Calibration : (*) Without Adjustment
Function of UUC* : Temperature Source
Fresh air setting : Close

Calibration Point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Temperature stability (± °C)	Temperature uniformity (°C)	Overall Variation (°C)	Uncertainty (± °C)	Coverage Factor k
104.0	104.0	104.0	0.040	0.57	0.80	0.42	2
120.0	120.0	120.0	0.11	0.82	1.1	1.1	2
180.0	180.0	180.0	0.12	1.4	2.0	1.1	2
Measured Temperature (°C)							
Position							
1	2	3	4	5	6	7	8
104.0	104.403	104.220	104.517	104.474	103.859	104.292	104.357
120.0	120.183	119.878	120.238	120.355	119.455	120.046	120.173
180.0	180.502	179.929	180.655	180.797	179.012	180.043	180.305
							180.340

Average* : The average of 30 values in each position.

Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor

Temperature uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-o0o-

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม
A 1104315



Equipment : BOD Incubator
Condition As-Received :
Reference : 2202-0446OC-1
Result of Calibration :-
Function of UUC* : Temperature Source
Fresh air setting : Not Available

Cert. No.: 22TM90
Page.: 3 of 3

Calibration Point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Temperature stability (± °C)	Temperature uniformity (°C)	Overall Variation (°C)	Uncertainty (± °C)	Coverage Factor k
20.0	19.5	19.4	0.30	0.58	1.0	0.55	2
Measured Temperature (°C)							
Position							
1	2	3	4	5	6	7	8
20.154	20.013	20.356	19.939	19.834	19.761	19.817	19.824
							9 (ref.)
							19.922

Average* : The average of 30 values in each position.

Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor.
Temperature uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

-o-o-



Equipment : BOD Incubator
Condition As-Received :
Reference : 2202-0446OC-1
Procedure Used :-

Cert. No.: 22TM90
Page.: 2 of 3

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD).
The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34970A	MY44035217	21LM30	23 Dec 2022


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Not Available



Probe Installation Details :

a =	10	cm	D =	0.62	m
b =	10	cm	W =	1.2	m
c =	10	cm	H =	1.2	m
			Capacity =	0.89	m ³

Dimension of Chamber :

Environment during calibration		
	Beginning	Finished
Temp. (°C)	28	28
REL.Humid. (%)	68	75
AC Supply (Volt)	226	226

Position :	Ref. Std. ID No.:
1	18-10RTD-01
2	18-10RTD-02
3	18-10RTD-03
4	18-10RTD-04
5	18-10RTD-05
6	22-10RTD-10
7	18-10RTD-07
8	18-10RTD-08
9 (ref.)	18-10RTD-09

๗๓๓.

เอกสารไม่ควบคุม
a 1096042

๗๓๓.

เอกสารไม่ควบคุม
a 1096042



Equipment : BOD Incubator
Condition As-Received : Used Item
Reference : 2204-00150C-2

Procedure Used :-

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD).

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34970A	MY41021843	22LM4	10 Jan 2023


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Not Available



Probe Installation Details :

Probe	Installation Details	Dimension of Chamber
a =	10 cm	D = 0.62 m
b =	10 cm	W = 1.2 m
c =	10 cm	H = 1.2 m
		Capacity = 0.89 m ³

Position	Ref. Std. ID No.
1	18-04RTD-01
2	18-04RTD-02
3	18-04RTD-03
4	18-04RTD-04
5	18-04RTD-05
6	18-04RTD-06
7	18-04RTD-07
8	18-04RTD-08
9 (ref.)	18-04RTD-09

Environment during calibration		
	Beginning	Finished
Temp. (°C)	27	27
REL.Humid. (%)	56	59
AC Supply (Volt)	222	221

Cert. No.: 22TM305
Page.: 2 of 3



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
53/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert. No.: 22TM305
Page.: 1 of 3

Certificate of Calibration

Equipment :	BOD Incubator
Manufacturer :	ARCO
Model :	UR-1320
Serial No. :	-
ID No. :	UAE.WAO.018/2551
Submitted by :	United Analyst and Engineering Consultant Co.,Ltd. 3 Soi Udomsuk 41, Sukhumvit Road, Bangchak, Phrakhanong, Bangkok 10260
Location :	Lab Floor 2
Received Order :	7 April 2022
Calibration Date :	7 April 2022
Ambient Temperature :	(26 ± 10) °C
Relative Humidity :	(50 ± 30) %
Calibrated by :	Man Pattanapongpaiboon

Approved by : 
Approved Signatory

() Ponthipha Tameyakul
() Malee Butkruea
() Suwit Imjai

Issue Date : 18 April 2022

The Uncertainties are for a confidence probability of approximately 95 %

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.



เอกสารไม่ควบคุม

a 1104314


เอกสารไม่ควบคุม

A 0040246

Certificate No. : HIT-2209-0184
Page : 1 of 3

CERTIFICATE OF CALIBRATION

Equipment : COD Test Tube Heater
Meter Model : HI839800-02 **Serial No. :** H0185001
Manufacturer : Hanna Instruments
Made in : Romania
Condition As-Received : Used Product
Reference : RE220234
Customer name : United Analyst and Engineering Consultant Co., Ltd.
3 Soi Udomsuk 41, Sukhumvit Rd., Bangchak,
Phrakhanong, Bangkok 10260
Received date : 21 February 2022
Calibrate date : 1 March 2022
Issue date : 2 March 2022
Ambient Temperature : $(25 \pm 2) ^\circ\text{C}$
Relative Humidity : $(50 \pm 15) \% \text{ RH}$
Calibrated Location : Hanna Instruments (Thailand) Ltd.

Calibrated by : *Pichit* **Approved by :** *ANAN*
Mr. Pichit Pethong
Calibration Engineer
Mr. Anan Suwanchaisakul
Authorized Signatory


This certificate was certified only for the instrument we calibrated.

This result of calibration was found accurate on date and place of calibration only.

** This certificate may not be reproduced other than in full, except with the prior written **
approval of the head of Hanna Instrument (Thailand)

เอกสารไม่ควบคุม



Equipment : BOD Incubator
Condition As-Received : Used Item
Reference : 2204-00150C-2
Result of Calibration :- (*) Without Adjustment
Function of UUC* : Temperature Source
Fresh air setting : Not Available

Cert. No.: 22TM305
Page.: 3 of 3

Calibration Point ($^{\circ}\text{C}$)	UUC* Setting ($^{\circ}\text{C}$)	UUC* Reading ($^{\circ}\text{C}$)	Temperature stability ($\pm ^{\circ}\text{C}$)	Temperature uniformity ($^{\circ}\text{C}$)	Overall Variation ($^{\circ}\text{C}$)	Uncertainty ($\pm ^{\circ}\text{C}$)	Coverage Factor k
20.0	20.0	20.0	0.50	0.44	1.1	0.64	2
Measured Temperature ($^{\circ}\text{C}$)							
Position							
1	2	3	4	5	6	7	8
20.080	20.056	19.866	19.826	19.655	19.656	19.819	19.979
						9 (ref.)	19.899

Average* : The average of 30 values in each position.

Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor
Temperature uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor **k**, providing a level of confidence of approximately 95 %.

-000-

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม

๑ 1104313

Result of Calibration :

Calibration Point	Unit Under Calibration Setting	Unit Under Calibration Reading	Temperature Stability	Uncertainty of Measurement
150.0 (°C)	- (°C)	150.6 (°C)	1.3 (°C)	±0.39 (°C)

Calibration Point (°C)	Average Standard Reading (°C)				
	Position				
150.0	1	2	3	4	5
	150.2	150.4	150.4	150.3	150.2
	6	7	8	9	10
	150.4	150.9	151.1	151.1	150.6
	11	12	13	14	15
	150.4	151.0	151.5	151.3	150.5
	16	17	18	19	20
	150.3	150.8	151.2	151.2	150.5
	21	22	23	24	25
	150.2	150.3	150.5	150.4	150.3

The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%

** End of certificate **

Condition of this result of calibration

Reference Standard Instruments :

Instruments	Model	Serial No.	Certificate No.	Traceable
Thermometer With Sensor	HI935005	03250060101	21T167	Technology Promotion Association (Thailand-Japan)

Reference / Procedure :

This equipment was calibration by comparison to the reference standard (Standard platinum resistance thermometer) whose accuracy is traceable to the national standard. The calibration was performed by generating the specified working point of temperature then recorded the temperature reading values against the reference standard according to Hanna Calibration Laboratory work Instruction No. 141.

This temperature scale used was based on ITS-90

All data shown below were as-received values without adjustment.

SITE CALIBRATION

1	2	3	4	5
6	7	8	9	10
11	12	13	14	15
16	17	18	19	20
21	22	23	24	25

Calibration Report

Certificate No.: 2203120-001-01
Equipment: Electronic Balance
Model: AB204-S/FACT
Serial No.: 1129361010
Capacity: 220 g

Manufacturer: METTLER TOLEDO
Resolution: 0.0001 g
ID No.: UAE.WAS.002/2552

Date of Calibration: 1 June 2022

Environment Condition: Ambient Temperature: 19.9 ± 0.3 °C Relative Humidity: 45 ± 1.5 %

Place of Calibration: 108, Balance Room, UNITED ANALYST AND ENGINEERING CONSULTANT CO., LTD.

Condition of Equipment: Good Condition

Condition of This Results of Calibration:

1. Calibration Method: NFI Method WMA-001 In-House Method based on UKAS Lab 14 : 2019

2. Reference Standards:

Reference Standard	Model	Serial No.	Calibrated By	Certificate No.	Due Date
Standard Weight Class E2	1-500mg	B308068554	TCS	M22010205	6 January 2023
Standard Weight Class E2	1-500g	B308068128	TCS	M22010215	6 January 2023
Instrument	Model	Serial No.	Calibrated By	Certificate No.	Due Date
Thermo-Hygro Meter	PONPE 490	NFLBTH 010/18	Quality Reborn	QR22-0350	18 February 2023

3. This certification is traceable to SI UNIT

4. This certificate was certified only for the instrument we calibrated.

5. This result of calibration was found accurate as shown on date and place of calibration only.

Calibration Results:




1. Repeatability of Reading:

Nominal Value (g)	Standard Deviation of Reading (g)
100	0.000048
200	0.000052

2. Off-Center Error:

A mass of 50 g was placed and moved to various position on pan.

The balance reading obtained is given in the table.

		
1 (g) 49.9999	2 (g) 49.9998	3 (g) 49.9998
4 (g) 49.9999	5 (g) 49.9998	6 (g) 49.9998
(Maximum Difference) (g) 0.0001		

F-CS-012 Revision: 01 Date: 20-04-65

2008 ๒๕๕๑-๒๕๕๒ 36 หมู่ ๖ ตำบลบ้านใหม่ อำเภอเมือง จังหวัดนนทบุรี 10700

2008 Soi 36, Arun Amarin Road, Bang Yi Khan Subdistrict, Bang Phat District, Bangkok 10700, Thailand

Tel: +66(0) 2422 8588 Fax: +66(0) 2422 8545

เอกสารไม่ควบคุม



nfi.orth

Calibration Certificate

Certificate No.: 2203120-001-01
Client name: UNITED ANALYST AND ENGINEERING CONSULTANT CO., LTD.
Address: 3 Soi Udomsuk 41, Sukhumvit Road,
Bangchack, Prakhnong, Bangkok 10260

Equipment: Electronic Balance

Manufacturer: METTLER TOLEDO

Model: AB204-S/FACT

Serial No.: 1129361010

ID No.: UAE.WAS.002/2552

Order No.: 2203120

Operation No.: 2203120-001

Date of Receipt: 1 June 2022

Date of Calibration: 1 June 2022

Calibrated by Mr.Tavesak Seilee
Scientist

Approved by

(Mr.Pheraphat Tuanjit)

Manager, Division of Calibration Laboratory

Responsible for the Technical Management Team

Date of Issue: 7 June 2022

The uncertainties are for a confidence probability of approximately 95%

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the National Food Institute.

F-CS-009 Revision: 01 Date: 20-04-65

2008 ๒๕๕๑-๒๕๕๒ 36 หมู่ ๖ ตำบลบ้านใหม่ อำเภอเมือง จังหวัดนนทบุรี 10700

2008 Soi 36, Arun Amarin Road, Bang Yi Khan Subdistrict, Bang Phat District, Bangkok 10700, Thailand

Tel: +66(0) 2422 8588 Fax: +66(0) 2422 8545

เอกสารไม่ควบคุม



nfi.orth

Verification Certificate

Certificate No.: 2202361-001-01
Client name: UNITED ANALYST AND ENGINEERING CONSULTANT CO., LTD.
Address: 3 Sol Udomsuk 41, Sukhumvit Road,
Bangchack, Prakhong, Bangkok 10260

Page 1 of 4

Equipment: HEATING BLOCK DIGESTION

Manufacturer: FOSS

Model: 2520

Serial No.: 91794469

ID No.: UAE.WAS.011/2560

Order No.: 2202361

Operation No.: 2202361-001

Date of Receipt: 4 April 2022

Date of Calibration: 4-6 April 2022

Calibrated by Mr.Nuttapol Niyomchat
Specialist

Approved by

(Mr.Pheraphat Tuanjit)

Manager, Division of Calibration Laboratory

Responsible for the Technical Management Team

Date of Issue: 11 April 2022

The uncertainties are for a confidence probability of approximately 95 %.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the National Food Institute.

FCS-009 Revision: 00 Date: 14-12-61

เอกสารไม่ควบคุม

Calibration Report

Certificate No.: 2203120-001-01
Equipment: Electronic Balance
Manufacturer: METTLER TOLEDO
Model: AB204-S/FACT
Resolution: 0.0001 g
Serial No.: 1129361010
ID No.: UAE.WAS.002/2552
Capacity: 220 g

Page 3 of 3

Date of Calibration: 1 June 2022

Calibration Results: (Continued)

Calibration Range: 0 - 200 g

Calibration Adjustment: Internal Calibration

3. Departure from Nominal Value:

Nominal Value (g)	Standard Value (g)	Average Reading (g)	Correction (g)	Uncertainty (± g)	Coverage Factor k
Unload	0.00000	0.0000	0.0000	0.000088	2.00
0.01	0.01000	0.0100	0.0000	0.000088	2.00
0.05	0.05000	0.0499	0.0001	0.000088	2.00
0.1	0.10000	0.1000	0.0000	0.000088	2.00
0.2	0.20000	0.2000	0.0000	0.000088	2.00
0.5	0.50000	0.5000	0.0000	0.000088	2.00
1	1.00000	0.9999	0.0001	0.000088	2.00
2	2.00000	1.9999	0.0001	0.000089	2.00
5	5.00000	5.0000	0.0000	0.000089	2.00
10	9.99998	9.9999	0.0001	0.000092	2.00
20	19.99999	19.9999	0.0001	0.000094	2.00
50	49.99990	49.9999	0.0000	0.00012	2.00
70	69.99989	69.9998	0.0001	0.00014	2.00
100	100.00001	99.9999	0.0001	0.00017	2.00
150	149.99991	149.9997	0.0002	0.00022	2.00
200	200.00007	199.9998	0.0003	0.00030	2.00

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

FCS-012 Revision: 01 Date: 20-04-65

FCS-009 Revision: 00 Date: 14-12-61

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม
2008 โซ 36, อรุณ อมารินทร์ ไรด์, แขวงบ้านช่าง, เขตบางเขน กรุงเทพมหานคร 10700, Thailand
2008 โซ 36, อรุณ อมารินทร์ ไรด์, แขวงบ้านช่าง, เขตบางเขน กรุงเทพมหานคร 10700, Thailand
Tel : +66(0) 2422 8545 Fax : +66(0) 2422 8545



nfi.or.th

Verification Report

Certificate No.: 2202361-001-01
Equipment: HEATING BLOCK DIGESTION
 Model: 2520 Serial No.: 91794469
 Resolution: 1 °C ID No.: UAE.WAS.011/2560
 Manufacturer: FOSS
Date of Calibration: 4-6 April 2022
 380 °C

Page 3 of 4

Calibration point:
Calibration result:

Reporting of Temperature

Block No.	UUC* Setting (°C)	UUC* Reading (°C)	Stability (±°C)	Standard Thermometer (°C)	Uncertainty (±°C)
1	380	380	0.13	376.48	1.5
2	380	380	0.12	376.58	1.5
3	380	380	0.12	376.51	1.5
4	380	380	0.14	376.70	1.6
5	380	380	0.18	376.81	1.6
6	380	380	0.12	377.23	1.6
7	380	380	0.12	377.37	1.5
8	380	380	0.13	376.68	1.5
9	380	380	0.14	376.72	1.5
10	380	380	0.18	378.97	1.6
11	380	380	0.25	378.79	1.6
12	380	380	0.11	377.14	1.6
13	380	380	0.19	379.65	1.6
14	380	380	0.16	379.61	1.6
15	380	380	0.16	378.66	1.6
16	380	380	0.15	379.18	1.6
17	380	380	0.23	377.39	1.6
18	380	380	0.11	377.71	1.6
19	380	380	0.22	376.64	1.6
20	380	380	0.16	376.56	1.6

Note:

- UUC* = Unit Under Calibration
- Immersion depth of standard thermometer in tube level high of sand is equal heater plate of UUC.
- Stability = One-half of the greatest maximum difference of measured temperatures at one sensors, for at least half an hour after reaching steady state.

Verification Report

Certificate No.: 2202361-001-01
Equipment: HEATING BLOCK DIGESTION
 Model: 2520 Serial No.: 91794469
 Resolution: 1 °C ID No.: UAE.WAS.011/2560
 Manufacturer: FOSS
Date of Calibration: 4-6 April 2022

Page 2 of 4

Location: Laboratory Room, NATIONAL FOOD INSTITUTE
Environment Condition: Ambient Temperature (25 ± 3) °C
 Relative Humidity (55 ± 15) %
 Line Voltage (220 ± 10) Volt

Condition of this results of Calibration:

- This instrument was calibrated by insert standard thermocouples type R into its heating block digestion and compared to temperature obtained from reference standards thermometer at calibrated point.
 - The temperature scale used was based on ITS - 90 .
 - All data show below were final values and the initial data may be obtained upon request.
- Reference Standard Instrument :

Instrument	Model	Serial No.	Certificate No.	Due Date	Through
Digital Thermometer with Thermocouple	34970A/34901A Type R	MY4005576/MY41194453 TC#101-103 / CH#101-103	TC21/0041	24-Apr-2022	N.M. Technical Center Laboratory

- This certificate is traceable to international system of units (SI Units).
- This certificate was certified only for the instrument we calibrated.
- This result of calibration was found accurate as shown on date and place of calibration only.
- Condition of Calibrated item : Good

UUC* Description

Time of Record - Hour 30 Minute At 380 °C

7. Result of Calibration : ☒ Without adjustment ☐ After adjustment

Customer Service Report

Date: 30/11/21
Customer: UAE
Instrument: KT200

Hours
Start 8.00
Finish 9.00

Travel To Customer
8.00
19.00

Labour
9.00
19.00

Travel From Customer
19.00
15.00

Report No: 5874

Address: 91 ถนนวิภาวดีรังสิต แขวง ดอนเมือง กรุงเทพฯ 10210
Serial: 91790529

Application		Special		Standard	
Normal	<input checked="" type="checkbox"/>	Courtesy Visit	<input checked="" type="checkbox"/>	Installation	<input checked="" type="checkbox"/>
Distributor	<input checked="" type="checkbox"/>	PMA Onboarding	<input checked="" type="checkbox"/>	Quote	<input checked="" type="checkbox"/>
Internal	<input checked="" type="checkbox"/>	Warranty	<input checked="" type="checkbox"/>	Repair	<input checked="" type="checkbox"/>
Digital Service	<input checked="" type="checkbox"/>	Sales Support	<input checked="" type="checkbox"/>	Remote	<input checked="" type="checkbox"/>
				Training	<input checked="" type="checkbox"/>
				In House	<input checked="" type="checkbox"/>
				PM	<input checked="" type="checkbox"/>
				Other	<input checked="" type="checkbox"/>

PO/Quote Number: *Foss complete Pro*

Details of Work / Test		Condition / Status
- Check instrument		OK
- Safety kit for KT200		Pass
- Safety Valve		Pass
- Rubber Grommet		Pass
- Heating element		Pass
- New panel PCB		Pass
- Safety door		Pass
- Clean & lubricant		Pass
- Check volum		Pass
- Check set 30ml for 28ml		Pass

Part No:	Batch	Description	Qty
10009965	11235-983	Foss PM kit KT200	1
1575-0029	20.08.21	Safety Valve for Hentengail	2
1549-0126	06.11.21	Rubber Grommet for Hentengail	2
10003512	02.08.21	Heating Element	1
10002762	16.11.20	KT200 new panel PCB	1
60096273	16.08.21	KT200 new panel PCB	1
100066385	22.04.21	Safety door complete	1

I confirm this report is accurate and complete

Signed FOSS: *[Signature]*
Name: *[Name]*
Email: *[Email]*

Signed Customer: *[Signature]*
Name: *[Name]*
Email: *[Email]*

Would you be willing to participate in a brief survey in order to tell us how we performed?


เอกสารไม่ควบคุม

Verification Report

Certificate No.: 2202361-001-01
Equipment: HEATING BLOCK DIGESTION
Model: 2520 Serial No.: 91794469
Resolution: 1 °C ID No.: UAE.WAS.011/2560
Manufacturer: FOSS
Date of Calibration: 4-6 April 2022
Calibration point: 380 °C
Calibration result: Continued

Page 4 of 4

Figure 1. Location of Reference Standard and Block Diagram of Digestion Unit
TOP VIEW



Note:

- UUC* = Unit Under Calibration
- Immersion depth of standard thermometer in tube level high of sand is equal heater plate of UUC.
- Stability = One-half of the greatest maximum difference of measured temperatures at one sensors, for at least half an hour after reaching steady state.

The report uncertainty of measurement was based on standard uncertainty multiplied by coverage factor k= 2, providing a level of confidence of approximately 95 %.

----- End -----

F-CS-012 Revision: 00 Date: 14-12-61

เอกสารไม่ควบคุม

FossCare™

FOSS Preventive Maintenance Protocol

Customer : UAE

Instrument	Kjeltec™ 2100 = kjeltec 200	
Recommended PM interval (whichever occurs first between interval and no. of samples analysed)	12 months	No. of samples analysed (if applicable):
Preventive maintenance kit (P/N)	10009965	5/N 917905-24

Introduction

A maintenance protocol provides systematic and functional means of maintaining a specific instrument type. The recommended PM interval depends on the operational conditions and is based on our extensive experience and knowledge of manufacturing and maintaining analytical instruments.

Apart from sample throughput, the environmental conditions also need to be considered. A demanding environment, such as high ambient temperature, humidity, dirtiness etc can measurably shorten component lifetime and also the maintenance and component replacement intervals.

NOTE!

The content of this protocol is subject to change over time. In order to safeguard that you obtain the correct parts, please make sure to indicate serial no and date of installation when contacting your FOSS representative.

Maintenance Procedure

Exchange of Parts and Cleaning

Step	Action	Part	P/N	OK
1	Replace	Adapter for dig. tube 250 ml	1000 0056	<input type="checkbox"/>
2	Replace	Non return valve	1000 3538	<input type="checkbox"/>
3	Replace valves in alkali pump	Valve kit reagent/water pump	1575 0093	<input type="checkbox"/>
4	Replace stream tubing	Silicone tubing 8/12 mm	1582 0006	<input type="checkbox"/>
5	Replace alkali tubing	Tubing reinforced for alkali	1582 0011	<input type="checkbox"/>
6	Replace water tubing	Tubing PVC 8/11 mm	1582 0004	<input type="checkbox"/>
7	Cleaning	Steam generator		<input type="checkbox"/>
8	Cleaning	Splash head		<input type="checkbox"/>

Check and Adjustments

Step	Action	Module	Measured	Limits	OK
1	Check alkali volume, 10 ml/stroke	Alkali pump	98	At 50 ml -0/+3 ml	<input checked="" type="checkbox"/>
2	Check distillation volume		120ml	100 – 150 ml/4 min	<input checked="" type="checkbox"/>
3	Check front panel switches				<input checked="" type="checkbox"/>
4	Check cables and electrical connections				<input checked="" type="checkbox"/>
5	Check level pins in steam generator				<input checked="" type="checkbox"/>
6	Check safety door switch				<input checked="" type="checkbox"/>

Dedicated Analytical Solutions

FOSS Analytical A/S 69 Slangerupgade DK-3400 Hillerød Denmark	FOSS Analytical AB Box 70 SE-262 21 Höganas Sweden	Tel +45 7010 3370 Fax +45 7010 3371 E-mail support@foss.dk Web www.foss.dk	Tel +46 42 361500 Fax +46 42 340349 E-mail support@foss.dk Web www.foss.dk
---	--	---	---

Scouting Run

Tested Combination2	Front	SSL	/	Front	UECD
Injection Tower					
Name:	7683A				
Setpoint Status:	Completed				
Injection Volume on Column:	1.0 uL				
Overall Scouting Run Status	Completed				

Noise and Drift

Tested Combination2	Front	SSL	/	Front	UECD
Injection Tower					
Name:	7683A				
Setpoint Status:	Pass				
Base Signal:	282.3 Hz				
ASTM Noise	0.60 Hz				
Drift	13.19 Hz/Hr				
Agilent Recommended:	<= 3.00				
Status:	Pass				

Setpoint Status: Pass

Base Signal:	11.2 pA	ASTM Noise	0.06 pA	Drift	0.31 pA/Hr
Agilent Recommended:	<= 0.10				<= 2.50
Status:	Pass				

Overall Noise and Drift Test Status

Tested Combination1	Front	SSL	/	Back	FID
Injection Tower					
Name:	7683A				
Setpoint Status:	Pass				
Injection Volume on Column:	1.0 uL				
Area RSD:	0.38 %				
Agilent Recommended:	<= 3.00				
Retention Time RSD:	0.65 %				
Status:	Pass				

Signal to Noise

Tested Combination1	Front	SSL	/	Back	FID
Injection Tower					
Name:	76890				
Setpoint Status:	Pass				
Signal to Noise:	784901				
Agilent Recommended:	>= 300000				
Overall Signal to Noise Test Status	Pass				

Overall Injection Precision Test Status

Tested Combination1	Front	SSL	/	Back	FID
Injection Tower					
Name:	76890				
Setpoint Status:	Pass				
Signal to Noise:	784901				
Agilent Recommended:	>= 300000				
Overall Signal to Noise Test Status	Pass				

Instrument Details

Purpose

This section describes the as found system configuration.

Details

System	UAE_TOX-007
System ID	
Manufacturer	Agilent Technologies
Name	7890
Flow Data Input	Manual Data
Temperature Data Input	Manual Data or Other Data Logging
Tested Combination1	
Injection Technique	Injection Tower
Inlet	Front
Detector	Back
LTM Included?	No
Tested Combination2	
Injection Technique	Injection Tower
Inlet	Front
Detector	Front
LTM Included?	No
Sampler 1	
Manufacturer	Agilent Technologies
Type	Injection Tower
Name	7893A
Model Number	G2813A
Serial Number	CN82149436
Firmware Revision	A.11.02
Usage	Sample Injection
Location	Front
Syringe Volume (µL)	10

Setpoint Status:

Pass

Injection Volume on Column:

1.0 uL

Area RSD:

1.14 %

Retention Time RSD:

0.04 %

Agilent Recommended:

<= 3.00

<= 1.00

Overall Injection Precision Test Status

Pass

Signal to Noise

Tested Combination2

Front

SSL

/ Front

UECD

Name:

7890

Setpoint Status:

Pass

Signal to Noise:

2250

Agilent Recommended:

>= 1500

Overall Signal to Noise Test Status

Pass

Detector 2		Sampler 2	
Manufacturer	Agilent Technologies	Manufacturer	Agilent Technologies
Name	7890	Type	Tray
Type	UECD	Name	7883A
Serial Number	U116886	Model Number	G2614A
Adapter	Capillary	Serial Number	CN82246787
Control Type	Electronic Pressure Control (EPC)	Firmware Revision	A.02.01
Location	Front	Mainframe 1	
Makeup Gas	Nitrogen	Manufacturer	Agilent Technologies
		Name	7890
		Model Number	G3440A
		Serial Number	CN11021007
		Firmware Revision	A.01.11
		Oven Type	Standard
		Inlet 1	
		Manufacturer	Agilent Technologies
		Name	7890
		Type	SSL
		Location	Front
		Carrier Gas	Helium
		Control Type	Electronic Pressure Control (EPC)
		Purged Inlet	Yes
		Detector 1	
		Manufacturer	Agilent Technologies
		Name	7890
		Type	FID
		Adapter	Capillary
		Control Type	Electronic Pressure Control (EPC)
		Location	Back
		Makeup Gas	Nitrogen

User Name: saenguthai.itsarak

Hostname: LAPTOP-CQ58K0W

OQ_UAE_TOX.007 Transaction log :

System Id: CN11021007

Print Date: February 11, 2022 3:57:30 PM

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
February 11, 2022 2:07:42 PM	End	Execution	Inlet Pressure Accuracy - Front SSL - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	Run Count : 1
February 11, 2022 2:07:44 PM	Start	Execution	Detector Flow Accuracy - Back FID - Type : Fuel - S: 30.0 mL/min - L: <= 10.0% setpoint	None
February 11, 2022 2:08:02 PM	Audit	Data	Detector Flow Accuracy - Back FID - Type : Fuel - S: 33.0 mL/min - L: <= 10.0% setpoint	Manual Data Entry
February 11, 2022 2:08:10 PM	End	Execution	Detector Flow Accuracy - Back FID - Type : Fuel - S: 30.0 mL/min - L: <= 10.0% setpoint	Run Count : 1
February 11, 2022 2:08:12 PM	Start	Execution	Detector Flow Accuracy - Back FID - Type : Oxidizer - S: 400.0 mL/min - L: <= 10.0% setpoint	None
February 11, 2022 2:08:46 PM	Audit	Data	Detector Flow Accuracy - Back FID - Type : Oxidizer - S: 400.0 mL/min - L: <= 10.0% setpoint	Manual Data Entry
February 11, 2022 2:08:52 PM	End	Execution	Detector Flow Accuracy - Back FID - Type : Oxidizer - S: 400.0 mL/min - L: <= 10.0% setpoint	Run Count : 1
February 11, 2022 2:08:56 PM	Start	Execution	Detector Flow Accuracy - Back FID - Type : Makeup - S: 25.0 mL/min - L: <= 10.0% setpoint	None
February 11, 2022 2:09:16 PM	Audit	Data	Detector Flow Accuracy - Back FID - Type : Makeup - S: 25.0 mL/min - L: <= 10.0% setpoint	Manual Data Entry
February 11, 2022 2:09:23 PM	End	Execution	Detector Flow Accuracy - Back FID - Type : Makeup - S: 25.0 mL/min - L: <= 10.0% setpoint	Run Count : 1

เอกสารไม่ควบคุม

User Name: saenguthai.itsarak

Hostname: LAPTOP-CQ58K0W

System Id: CN11021007

Print Date: February 11, 2022 3:57:30 PM

OQ_UAE_TOX.007 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
February 11, 2022 2:08:20 PM	Audit	Session Released	Session	None
February 11, 2022 2:08:21 PM	Audit	Exp loaded	Session	EQP details for primary technique (2x) - file path: [PracticalPack\GasConfiguration\ion6\2.5\7\Sc 02.51 Leap]. EQP File Name: [Sc 02.51 leap]. EQP Name: [AplicRecommaned]
February 11, 2022 2:08:55 PM	End	Configuration	Session	None
February 11, 2022 2:08:59 PM	Start	Qualification	Session	OQ
February 11, 2022 2:08:59 PM	Start	Execution	System Inspection and Basic Safety and Operation - 7490 - Qualitative Test - No setpoints associated	None
February 11, 2022 2:07:19 PM	End	Execution	System Inspection and Basic Safety and Operation - 7490 - Qualitative Test - No setpoints associated	Run Count: 1
February 11, 2022 2:07:21 PM	Start	Execution	Inlet Pressure Decay - Front: SSL - Pressure Controlled Inlet - S: 25.0 psi - L: >= -3.0 psi and <= 0.5 psi	None
February 11, 2022 2:07:35 PM	End	Execution	Inlet Pressure Decay - Front: SSL - Pressure Controlled Inlet - S: 25.0 psi - L: >= -3.0 psi and <= 0.5 psi	Run Count: 1
February 11, 2022 2:07:37 PM	Start	Execution	Inlet Pressure Accuracy - Front: SSL - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	None

เอกสารไม่ควบคุม

User Name: sangnathai.larak Hostname: LAPTOP-Q3SKOMV		System ID: CN11021007 Print Date: February 11, 2022 3:57:30 PM	
OQ_UAE_TOX.007 Transaction log:			
Time	Transaction State	Activity Performed	Optional Information
February 11, 2022 2:10:43 PM	End	Execution	GC Oven Temperature Accuracy - 7800 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K Run Count: 1
February 11, 2022 2:10:48 PM	Start	Execution	GC Oven Temperature Stability - 7800 - Temperature : Oven - S: 100.0°C - L: <= 0.5°C
February 11, 2022 2:11:37 PM	Audit	Data	GC Oven Temperature Stability Manual Data Entry - 7800 - Temperature : Oven - S: 100.0°C - L: <= 0.5°C
February 11, 2022 2:11:38 PM	End	Execution	GC Oven Temperature Stability - 7800 - Temperature : Oven - S: 100.0°C - L: <= 0.5°C Run Count: 1
February 11, 2022 2:11:44 PM	Start	Execution	GC Scouting Run - Injection Tower, Front SSL Back FID: - Part of System Preparation - No limits associated
February 11, 2022 2:14:13 PM	Start	Execution	GC Scouting Run - Injection Tower, Front SSL Back FID: - Part of System Preparation - No limits associated
February 11, 2022 2:14:45 PM	Audit	Data	GC Scouting Run - Injection Tower, Front SSL Back FID: - E:\DC2002\BSC_FID_Bsc401. Part of System Preparation - No D\FID18.ch limits associated
February 11, 2022 2:15:16 PM	End	Execution	GC Scouting Run - Injection Tower, Front SSL Back FID: - Part of System Preparation - No limits associated
February 11, 2022 2:15:22 PM	Start	Execution	Noise and Drift - Back FID: - Detector FID - L (Noise): <= 0.10 pA - L (Drift): <= 2.50

User Name: sangnathai.larak
Hostname: LAPTOP-Q3SKOMV

System ID: CN11021007
Print Date: February 11, 2022 3:57:20 PM

OQ_UAE_TOX.007 Transaction log:

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
February 11, 2022 2:09:25 PM	Start	Execution	Detector Flow Accuracy - Front UECD: - Type: Makeup - S: 25.0 mL/min - L: <= 10.0% setpoint	None
February 11, 2022 2:09:50 PM	Audit	Data	Detector Flow Accuracy - Front UECD: - Type: Makeup - S: 25.0 mL/min - L: <= 10.0% setpoint	Manual Data Entry
February 11, 2022 2:09:54 PM	End	Execution	Detector Flow Accuracy - Front UECD: - Type: Makeup - S: 25.0 mL/min - L: <= 10.0% setpoint	Run Count: 1
February 11, 2022 2:09:56 PM	Start	Execution	GC Oven Temperature Accuracy - 7800 - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	None
February 11, 2022 2:10:19 PM	Audit	Data	GC Oven Temperature Accuracy - 7800 - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Manual Data Entry
February 11, 2022 2:10:20 PM	End	Execution	GC Oven Temperature Accuracy - 7800 - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Run Count: 1
February 11, 2022 2:10:22 PM	Start	Execution	GC Oven Temperature Accuracy - 7800 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	None
February 11, 2022 2:10:41 PM	Audit	Data	GC Oven Temperature Accuracy - 7800 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Manual Data Entry

User Name: xiangzhihai.zhak
Hostname: LAPTOP-CQ3SK0WV
System ID: CN11021007
Print Date: February 11, 2022 3:57:30 PM

QO_UAE_TOX-007 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
February 11, 2022 2:17:12 PM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Back FID - GC - L (Area): <= 3.00% - L (Ret. Time): <= 1.00%	Data files Path : E:\QO2022\QO2022-2022-02-10 13-02-1400Q_GC7880_FID_Pro0100.D\FID18.ch
February 11, 2022 2:17:12 PM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Back FID - GC - L (Area): <= 3.00% - L (Ret. Time): <= 1.00%	Data files Path : E:\QO2022\QO2022-2022-02-10 13-02-1400Q_GC7880_FID_Pro0100.D\FID18.ch
February 11, 2022 2:17:13 PM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Back FID - GC - L (Area): <= 3.00% - L (Ret. Time): <= 1.00%	Data files Path : E:\QO2022\QO2022-2022-02-10 13-02-1400Q_GC7880_FID_Pro0100.D\FID18.ch
February 11, 2022 2:17:35 PM	End	Execution	Injection Precision - Injection Tower, Front SSL, Back FID - GC - L (Area): <= 3.00% - L (Ret. Time): <= 1.00%	Run Count : 1
February 11, 2022 2:17:45 PM	Start	Execution	Signal to Noise - Injection Tower, Front SSL, Back FID - Detector FID - L >= 300000	None
February 11, 2022 2:18:06 PM	Start	Execution	Signal to Noise - Injection Tower, Front SSL, Back FID - Detector FID - L >= 300000	None
February 11, 2022 2:18:23 PM	Audit	Data	Signal to Noise - Injection Tower, Front SSL, Back FID - Detector FID - L >= 300000	Data files Path : E:\QO2022\QO2022-2022-02-10 13-02-1400Q_GC7880_FID_SND11.D\FID18.ch
February 11, 2022 2:18:02 PM	Start	Execution	Signal to Noise - Injection Tower, Front SSL, Back FID - Detector FID - L >= 300000	None

User Name: xiangzhihai.zhak
Hostname: LAPTOP-CQ3SK0WV
System ID: CN11021007
Print Date: February 11, 2022 3:57:30 PM

QO_UAE_TOX-007 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
February 11, 2022 2:18:57 PM	Audit	Data	Noise and Drift - Back FID - Detector FID - L (Noise): <= 0.10 pA - L (Drift): <= 2.50 pA/hour	Data files Path : E:\QO2022\NC_FID_835401.D\FID18.ch
February 11, 2022 2:19:06 PM	End	Execution	Noise and Drift - Back FID - Detector FID - L (Noise): <= 0.10 pA - L (Drift): <= 2.50 pA/hour	Run Count : 1
February 11, 2022 2:19:39 PM	Start	Execution	Injection Precision - Injection Tower, Front SSL, Back FID - GC - L (Area): <= 3.00% - L (Ret. Time): <= 1.00%	None
February 11, 2022 2:19:47 PM	Start	Execution	Injection Precision - Injection Tower, Front SSL, Back FID - GC - L (Area): <= 3.00% - L (Ret. Time): <= 1.00%	None
February 11, 2022 2:19:12 PM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Back FID - GC - L (Area): <= 3.00% - L (Ret. Time): <= 1.00%	Data files Path : E:\QO2022\QO2022-2022-02-10 13-02-1400Q_GC7880_FID_Pro0100.D\FID18.ch
February 11, 2022 2:19:12 PM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Back FID - GC - L (Area): <= 3.00% - L (Ret. Time): <= 1.00%	Data files Path : E:\QO2022\QO2022-2022-02-10 13-02-1400Q_GC7880_FID_Pro0100.D\FID18.ch
February 11, 2022 2:19:12 PM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Back FID - GC - L (Area): <= 3.00% - L (Ret. Time): <= 1.00%	Data files Path : E:\QO2022\QO2022-2022-02-10 13-02-1400Q_GC7880_FID_Pro0100.D\FID18.ch

User Name: saenguthitarak
Hostname: LAPTOP-G3SKOMV
Print Date: February 11, 2022 3:37:30 PM
System ID: CN11021007

CO_UAE_TOX.007 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
February 11, 2022 3:34:01 PM	Start	Execution	Injection Precision - Injection Tower, Front SSL, Front UECD: - GC - L (Area): <= 3.00% - L (Ret. Time): <= 1.00%	None
February 11, 2022 3:34:42 PM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Front UECD: - GC - L (Area): <= 3.00% - L (Ret. Time): <= 1.00%	Data files Path : E:\002022\00_GC7890_uE CD_Pre0103.D\CD01A.ch
February 11, 2022 3:34:42 PM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Front UECD: - GC - L (Area): <= 3.00% - L (Ret. Time): <= 1.00%	Data files Path : E:\002022\00_GC7890_uE CD_Pre0104.D\CD01A.ch
February 11, 2022 3:34:42 PM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Front UECD: - GC - L (Area): <= 3.00% - L (Ret. Time): <= 1.00%	Data files Path : E:\002022\00_GC7890_uE CD_Pre0105.D\CD01A.ch
February 11, 2022 3:34:42 PM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Front UECD: - GC - L (Area): <= 3.00% - L (Ret. Time): <= 1.00%	Data files Path : E:\002022\00_GC7890_uE CD_Pre0106.D\CD01A.ch
February 11, 2022 3:34:42 PM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Front UECD: - GC - L (Area): <= 3.00% - L (Ret. Time): <= 1.00%	Data files Path : E:\002022\00_GC7890_uE CD_Pre0107.D\CD01A.ch
February 11, 2022 3:34:42 PM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Front UECD: - GC - L (Area): <= 3.00% - L (Ret. Time): <= 1.00%	Data files Path : E:\002022\00_GC7890_uE CD_Pre0108.D\CD01A.ch
February 11, 2022 3:35:20 PM	End	Execution	Injection Precision - Injection Tower, Front SSL, Front UECD: - GC - L (Area): <= 3.00% - L (Ret. Time): <= 1.00%	Run Count : 1


เอกสารไม่ควบคุม

User Name: saenguthitarak
Hostname: LAPTOP-G3SKOMV
Print Date: February 11, 2022 3:57:30 PM
System ID: CN11021007

CO_UAE_TOX.007 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
February 11, 2022 3:30:11 PM	End	Execution	Signal to Noise - Injection Tower, Front SSL, Sack FID : Detector FID - L >= 300000	Run Count : 1
February 11, 2022 3:31:43 PM	Start	Execution	GC Sealing Run - Injection Tower, Front SSL, Front UECD: - Part of System Preparation - No limits associated	None
February 11, 2022 3:32:10 PM	Audit	Data	GC Sealing Run - Injection Tower, Front SSL, Front UECD: - Part of System Preparation - No limits associated	Data files Path
February 11, 2022 3:33:01 PM	Audit	Data	GC Sealing Run - Injection Tower, Front SSL, Front UECD: - Part of System Preparation - No limits associated	Data files Path
February 11, 2022 3:33:33 PM	End	Execution	GC Sealing Run - Injection Tower, Front SSL, Front UECD: - Part of System Preparation - No limits associated	Run Count : 1
February 11, 2022 3:33:46 PM	Start	Execution	Noise and Drift - Front UECD : Detector UECD - L (Noise): <= 3.00 Hz - L (Drift): <= 15.00 Hz/hour	None
February 11, 2022 3:33:47 PM	Audit	Data	Noise and Drift - Front UECD : Detector UECD - L (Noise): <= 3.00 Hz - L (Drift): <= 15.00 Hz/hour	Data files Path : E:\002022\00_GC7890_uE CD_Pre0109.D\CD01A.ch
February 11, 2022 3:33:58 PM	End	Execution	Noise and Drift - Front UECD : Detector UECD - L (Noise): <= 3.00 Hz - L (Drift): <= 15.00 Hz/hour	Run Count : 1

เอกสารไม่ควบคุม



Agilent CrossLab Compliance

Qualification Type:	GC-QC
System ID:	CN11021007
EQP Name:	Agilent 7890B
EQP Revision:	GC-02.51
EQP Purchase Date:	November 2020
Date:	February 11, 2022 4:01:27 PM
Report Type:	Report
Org Name:	United Analytical and Engineering Consultan Co., Ltd.
Org Location:	3, Soi Lannaak 41 Sukhumvit Rd., Bangkok, Bangkok Thailand 10250

Table of Contents

Section	Page
Noise and Drift - Front UECD	39
Injection Precision - Injection Tower, Front SSL, Front UECD	41
Signal to Noise - Injection Tower, Front SSL, Front UECD	42
Declaration of Change Control	51
Attachments	52
Agilent Technologies Materials: GC00045	89
Electronic Signature	90
Transaction Log	91

Section	Page
Cover	1
Table of Contents	2
Test Summary	4
Service Details	5
Instrument Details	6
Calculation Formulas	9
Protocol Details	11
Tests	12
System Inspection and Basic Safety and Operation - 7890	12
Inlet Pressure Accuracy - Front SSL	13
Inlet Pressure Accuracy - Front SSL	14
Detector Flow Accuracy - Blank FID	15
Detector Flow Accuracy - Front UECD	17
GC Oven Temperature Accuracy - 7890	18
GC Oven Temperature Stability - 7890	20
GC Standing Run - Injection Tower, Front SSL, Blank FID	21
Noise and Drift - Blank FID	24
Injection Precision - Injection Tower, Front SSL, Blank FID	26
Signal to Noise - Injection Tower, Front SSL, Blank FID	32
GC Standing Run - Injection Tower, Front SSL, Front UECD	36

Test Summary

Purpose
This section includes a status for each scheduled test and the overall qualification. For each test that is not (1) the status is automatically determined based on pre-defined limits, and (2) the total number of times the test was run is displayed. For detailed results and specifications for a test, refer to the test results in this EQR.

Details	Status	Result
Test		
System Inspection and Basic Safety and Operation - 7890	Pass	1
Inlet Pressure Accuracy - Front SSL	Pass	1
Inlet Pressure Accuracy - Front SSL	Pass	1
Detector Flow Accuracy - Blank FID	Pass	1
Detector Flow Accuracy - Front UECD	Pass	1
GC Oven Temperature Accuracy - 7890	Pass	1
GC Oven Temperature Stability - 7890	Pass	1
GC Standing Run - Injection Tower, Front SSL, Blank FID	Pass	1
Noise and Drift - Blank FID	Pass	1
Injection Precision - Injection Tower, Front SSL, Blank FID	Pass	1
Signal to Noise - Injection Tower, Front SSL, Blank FID	Pass	1
GC Standing Run - Injection Tower, Front SSL, Front UECD	Pass	1
Noise and Drift - Front UECD	Pass	1
Injection Precision - Injection Tower, Front SSL, Front UECD	Pass	1
Signal to Noise - Injection Tower, Front SSL, Front UECD	Pass	1
Overall Qualification Status	Pass	1

User Name: saengudhulai.burak
Host Name: LAPTOP-QJ3SK0WY
Print Date: February 11, 2022 3:57:30 PM
System ID: CN11021007

QC_UAE_TOX.007 Transaction Log

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
February 11, 2022 3:56:55 PM	Start	Factorial on	Signal to Noise - Injection Tower, Front SSL, Front UECD, Detector UECD - L >= 1500	Note
February 11, 2022 3:58:13 PM	Aud:	Data	Signal to Noise - Injection Tower, Front SSL, Front UECD, Detector UECD - L >= 1500	Date files Path:
February 11, 2022 3:58:35 PM	End	Execution	Signal to Noise - Injection Tower, Front SSL, Front UECD, Detector UECD - L >= 1500	Run Count: 1
February 11, 2022 3:57:32 PM	End	Qualification	Session	CQ
February 11, 2022 3:57:32 PM	Stop	Reporting	Session	None
February 11, 2022 3:53:40 PM	Auto	Reporting	Session	Report Generated: Certificate
February 11, 2022 3:56:02 PM	Auto	Reporting	Session	Report Generated: Report

Calculation Formulas

Purpose
This section includes calculation formulas for all available tests. Depending upon which tests are scheduled, all or some apply to your qualification.

$$\begin{aligned} \text{Accuracy} &= (X_{\text{test}} - N_{\text{test}}) \\ \text{Absolute Accuracy} &= (N_{\text{test}} - X_{\text{test}}) \\ \text{Average (mean value of } n \text{ observations)} &= \frac{1}{n} \sum_{i=1}^n X_i \end{aligned}$$

$\bar{X}_{(i)} =$ Mean value
 $\bar{X}_{(i+1)} =$ Success point
 $\bar{X}_i =$ Mean value
 $\bar{X}_{(i)} = \text{Value, } i^{\text{th}}$ observation
 $n =$ Total number of observations

$$\boxed{\% \text{ Curry Over}} = \frac{N_A}{N_o} = 100$$

$$\Delta N_t = c \sum_{i=1}^n N_i^2 - \left(\sum_{i=1}^n N_i \right)^2$$

$$\boxed{\text{def}} \text{ slope of the regression} = \frac{1}{\sum_{i=1}^n X_i^2} \left(n \sum_{i=1}^n X_i Y_i - \sum_{i=1}^n X_i \sum_{i=1}^n Y_i \right)$$

Detector Flow Accuracy

Purpose: Detector flow accuracy is determined by measuring the flow with a calibrated mass flowmeter and comparing them to the test septums and the values displayed by the GC (if applicable).

Configuration Details
Name: 7190 Front UECO

Septum: Inlet Pressure: 25.0 psi

Measurements and Results
Time: 10.41 Flow: 24.8 mL/min
Accuracy: 0.2 mL/min
Agent Recommended: <= 10.0 % septum
Limit is percentage of septum or 0.5 mL/min, whichever is larger.

Septum Status: Pass

Overall Detector Flow Accuracy Test Status: Pass

Overall GC Oven Temperature Accuracy Test Status

Pass

GC Oven Temperature Accuracy

Purpose: This test uses a calibrated digital thermometer to determine the accuracy of the GC oven. Accuracy is calculated as the absolute difference between the measured temperature and septum.

Configuration Details
Name: 7190 Zone Oven

Septum: Temperature: 100.0 °C

Measurements and Results
Probe: A single probe is used for this service.
Time: 10.50 Temperature: 100.2 °C
Accuracy: 0.2 °C
Agent Recommended: <= 1.0 °C
Limit is 1.0 °C and 0.5 % septum, whichever is larger.

Septum Status: Pass

Overall GC Oven Temperature Accuracy Test Status: Pass

GC Oven Temperature Stability

Purpose: This test uses a calibrated digital thermometer to determine the stability of the oven temperature. Stability is expressed as the delta between the highest and lowest measured temperatures.

Configuration Details
Name: 7190 Zone Oven

Septum: Temperature: 100.0 °C

Measurements and Results
Probe: A single probe is used for this service.
Time: 10.50 Temperature: 100.2 °C
Accuracy: 0.2 °C
Agent Recommended: <= 1.0 °C
Limit is 1.0 °C and 0.5 % septum, whichever is larger.

Septum Status: Pass

Overall GC Oven Temperature Stability Test Status: Pass

Inlet Pressure Decay

Purpose: This test demonstrates the pressure integrity of the GC inlet (with a valve controlled injection system, if applicable) and all flows controlled by the GC inlet pressure.

Configuration Details
Name: 7190 Front GSL

Septum: Pressure: 25.0 psi

Measurements and Results
Initial Pressure: 25.0 psi
Final Pressure: 24.8 psi
Accuracy: 0.2 psi
Agent Recommended: <= 0.5 psi
Limit is percentage of septum or 0.5 mL/min, whichever is larger.

Septum Status: Pass

Overall Inlet Pressure Decay Test Status: Pass

Detector Flow Accuracy

Purpose: Detector flow accuracy is determined by measuring the flow with a calibrated mass flowmeter and comparing them to the test septums and the values displayed by the GC (if applicable).

Configuration Details
Name: 7190 Back FID

Septum: Flow Type: Flow Flow: 30.1 mL/min

Measurements and Results
Time: 10.50 Flow: 30.1 mL/min
Accuracy: 0.1 mL/min
Agent Recommended: <= 10.0 % septum
Limit is percentage of septum or 0.5 mL/min, whichever is larger.

Septum Status: Pass

Overall Detector Flow Accuracy Test Status: Pass

Inlet Pressure Accuracy

Purpose: This test uses a digital calibrated manometer to demonstrate the ability of the system to provide accurate pressure to the head of the column. Accuracy is calculated as the absolute difference between the measured pressure and septum.

Configuration Details
Name: 7190 Front GSL

Septum: Inlet Pressure: 25.0 psi

Measurements and Results
Initial Pressure: 25.0 psi
Final Pressure: 24.8 psi
Accuracy: 0.2 psi
Agent Recommended: <= 0.5 psi
Limit is percentage of septum or 0.5 mL/min, whichever is larger.

Septum Status: Pass

Overall Inlet Pressure Accuracy Test Status: Pass

Overall Inlet Pressure Accuracy Test Status


Pass

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม

Analysis Date Path: F:\ACE
230Agilent_ACEBaseline\agilent\Task\CO2_LME_TOX.07\CO2Test\CO2_LME_TOX_01\Peak
subRun1

Sample Name Task: CO2_GC_Base_NCM
Acquisition method: NO_FID_Base01.D
Data file subRun1 for 315 SEC: 15-Feb-22, 12:31:54
Acquisition Date: ASTM
Noise Type: 0.00779
Noise Value: 3.0
Noise Offset Time: 30.0
Noise duration: 0.30035
Drift Value:



Overall Noise and Drift Test Status

Results: Average: 10003.41 pA
STD Deviation: 41.00000 pA
RSD: 0.38 %
Agilent Recommendation: Pass



Date Audit Log: LAPTGP-CDSB000V
Post name: E:\CO2Test\CO2Test_2022-02-10_15-05-14
Original Data Path: F:\ACE
Analyzed Data Path: 230Agilent_ACEBaseline\agilent\Task\CO2_LME_TOX.07\CO2Test\CO2_LME_TOX_01\Peak
subRun1


Injection Precision
Purpose: This test uses a five-point standard to determine injection precision. The mean, standard deviation, and % RSD of six measured injections are calculated.

Sample Name Task: CO2_GC_Base_NCM
Acquisition method: NO_FID_Base01.D
Data file subRun1 for 315 SEC: 15-Feb-22, 12:31:54
Acquisition Date: ASTM
Noise Type: 0.00779
Noise Value: 3.0
Noise Offset Time: 30.0
Noise duration: 0.30035
Drift Value:

Injection Precision
Purpose: This test uses a five-point standard to determine injection precision. The mean, standard deviation, and % RSD of six measured injections are calculated.

Sample Name Task: CO2_GC_Base_NCM
Acquisition method: NO_FID_Base01.D
Data file subRun1 for 315 SEC: 15-Feb-22, 12:31:54
Acquisition Date: ASTM
Noise Type: 0.00779
Noise Value: 3.0
Noise Offset Time: 30.0
Noise duration: 0.30035
Drift Value:

Injection Precision
Purpose: This test uses a five-point standard to determine injection precision. The mean, standard deviation, and % RSD of six measured injections are calculated.



Scouting Run
Purpose: This test is used to determine the chromatogram for presence of expected peaks, sufficient run time, and proper integration of peaks. Peaks are analyzed for this test.

Sample Name Task: CO2_GC_Base_NCM
Acquisition method: NO_FID_Base01.D
Data file subRun1 for 315 SEC: 15-Feb-22, 12:31:54
Acquisition Date: ASTM
Noise Type: 0.00779
Noise Value: 3.0
Noise Offset Time: 30.0
Noise duration: 0.30035
Drift Value:

Scouting Run
Purpose: This test is used to determine the chromatogram for presence of expected peaks, sufficient run time, and proper integration of peaks. Peaks are analyzed for this test.

Sample Name Task: CO2_GC_Base_NCM
Acquisition method: NO_FID_Base01.D
Data file subRun1 for 315 SEC: 15-Feb-22, 12:31:54
Acquisition Date: ASTM
Noise Type: 0.00779
Noise Value: 3.0
Noise Offset Time: 30.0
Noise duration: 0.30035
Drift Value:

Scouting Run
Purpose: This test is used to determine the chromatogram for presence of expected peaks, sufficient run time, and proper integration of peaks. Peaks are analyzed for this test.




Sample Name Task: CO2_GC_Base_NCM
Acquisition method: NO_FID_Base01.D
Data file subRun1 for 315 SEC: 15-Feb-22, 12:31:54
Acquisition Date: ASTM
Noise Type: 0.00779
Noise Value: 3.0
Noise Offset Time: 30.0
Noise duration: 0.30035
Drift Value:

Sample Name Task: CO2_GC_Base_NCM
Acquisition method: NO_FID_Base01.D
Data file subRun1 for 315 SEC: 15-Feb-22, 12:31:54
Acquisition Date: ASTM
Noise Type: 0.00779
Noise Value: 3.0
Noise Offset Time: 30.0
Noise duration: 0.30035
Drift Value:

Sample Name Task: CO2_GC_Base_NCM
Acquisition method: NO_FID_Base01.D
Data file subRun1 for 315 SEC: 15-Feb-22, 12:31:54
Acquisition Date: ASTM
Noise Type: 0.00779
Noise Value: 3.0
Noise Offset Time: 30.0
Noise duration: 0.30035
Drift Value:

Sample Name Task: CO2_GC_Base_NCM
Acquisition method: NO_FID_Base01.D
Data file subRun1 for 315 SEC: 15-Feb-22, 12:31:54
Acquisition Date: ASTM
Noise Type: 0.00779
Noise Value: 3.0
Noise Offset Time: 30.0
Noise duration: 0.30035
Drift Value:

Sample Name Task: CO2_GC_Base_NCM
Acquisition method: NO_FID_Base01.D
Data file subRun1 for 315 SEC: 15-Feb-22, 12:31:54
Acquisition Date: ASTM
Noise Type: 0.00779
Noise Value: 3.0
Noise Offset Time: 30.0
Noise duration: 0.30035
Drift Value:





Certificate of Calibration

Equipment: CONDUCTIVITY METER
Model: Lab955
Serial No. (or ID.): 16300356
Manufacturer: SI Analytics
Electrode Serial No. 16070067
Condition: In Condition

Certificate No.: C24220084
Issued Date: 22 March 2022
Job No.: KSPR2203267
Page: 1 of 2
Model : LF413T
Brand : SI Analytics

Customer: United Analyst and Engineering Consultant Company Limited
3 Soi Udomsuk 41 Sukhumvit Road,
Bangkok, Prakanong, Bangkok 10260 Thailand

Environment Condition: Temperature 23 °C ± 2 °C
Humidity 50 %RH ± 15 %RH

Calibration Place: Environment Laboratory, SPC RT Co., Ltd.
1194 Soi Wachirathamsothit 57, Sukhumvit 101/1 Rd.,
Bangchak, Prakanong, Bangkok 10260 Thailand

Calibration By: Mr. Wasan Nuchabee
Calibration Date: 22 March 2022
The Method used: In house method, SPCC-WI-49, base on ASTM D 1125-14 and D 5391-14
Traceability: This certificate is traceable to the SI Units maintained by CRM of NIST(SRM) through CPA chem Co., Ltd. (ISO/IEC 17034) Certificate No. 794135, 794136, 772624

Person in charge

(Mr. Wasan Nuchabee)

Authorized signatory

(Mr. Dumrong Boonsopon)

This certificate is issued the units of measurement according to the International System of Units (SI), it provides traceability of measurement to international or national standard or other recognized national standard laboratories.
The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).
These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of SPC RT Co., Ltd.

Agent CrossLab Compliance Services				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				
System ID: CH11021007				
System Name: SPC Calibration Center				

ใบตรวจสอบสภาพเครื่องวัดสิ่งแวดล้อม

เลขที่ใบงาน: KSPR2203267

ชนิดเครื่องมือ: CONDUCTIVITY METER			รุ่น: Lab955	หมายเลขเครื่อง: 16300356			
ตรวจสอบ (รับ)		รายการตรวจเช็ค			ตรวจสอบ (ส่ง)		หมายเหตุ
22 Mar 2022					22 Mar 2022		
ปกติ	ไม่ปกติ				ปกติ	ไม่ปกติ	
		<i>General</i>					
<input checked="" type="checkbox"/>	<input type="checkbox"/>	1. ความสมบูรณ์เครื่อง			<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	2. ความสะอาด (ช้อนใส่ตัวอย่าง, ภายใน-นอกเครื่อง)			<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	3. สวิตช์ ปิด – เปิด เครื่อง (On-Off Switch)			<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	4. ปุ่มกด (Keypad)			<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	5. หน้าจอ (Display, Screen Contrast)			<input checked="" type="checkbox"/>	<input type="checkbox"/>	
		<i>Spectrophotometer</i>					
<input type="checkbox"/>	<input type="checkbox"/>	6. แรงดันไฟฟ้า (Battery Backup) >= 2.5 VDC			<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	7. ตัวหมุนเลือกความยาวคลื่น (Wavelength Control)			<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	8. ความยาวคลื่น (Wavelength Check)			<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	9. แหล่งกำเนิดแสง (UV < 3,000 hour)			<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	10. แหล่งกำเนิดแสง (Visible < 5,000 hour)			<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	11. ช่องวัดหลายตัวอย่าง (Carousel Module)			<input type="checkbox"/>	<input type="checkbox"/>	
		<i>pH Meter and Conductivity Meter</i>					
<input checked="" type="checkbox"/>	<input type="checkbox"/>	12. อิเล็กโทรด (Electrode and Connection Cable)			<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	13. ระดับสารละลายใน Electrode (Level KCl)			<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	14. ฝาปิดกันปลาย Electrode (Dust Protection Hood)			<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	15. ขาจับอิเล็กโทรด (Stand)			<input type="checkbox"/>	<input type="checkbox"/>	
		<i>Turbidimeter</i>					
<input type="checkbox"/>	<input type="checkbox"/>	16. ค่าความขุ่นที่ต่ำสุด (No Sample)			<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	17. ระดับการส่องสว่างของแสง (>= 2.5 ไม่นเกิน 3.0)			<input type="checkbox"/>	<input type="checkbox"/>	
		<i>Automatic titrator</i>					
<input type="checkbox"/>	<input type="checkbox"/>	18. สภาพ Piston Burettes			<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	19. Function Rinsing and Dosing			<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	20. ระบบท่อสายยางและอุปกรณ์ประกอบ			<input type="checkbox"/>	<input type="checkbox"/>	

ข้อเสนอแนะ : Electrode วัดอุณหภูมิได้24.9 °C โดย Control Waterbath ที่ 25.0 ±0.1°C

Mr. Wasan Nuchabee
Service Engineer

Certificate No.: C24220084

Page: 2 of 2

Calibration Results:

Before Adjustment

Standard	Unit Under Calibration	Reading	Correction	Coverage Factor (k)	Uncertainty (±)
Conductivity Solution					
25.000 µS/cm	µS/cm	25.9	-0.900	2.00	0.22 µS/cm
1413.0 µS/cm	µS/cm	1444	-31.0	2.00	8.9 µS/cm
111.3 mS/cm	mS/cm	107.9	3.40	2.00	0.66 mS/cm

After Adjustment ; at 1413 µS/cm

Standard	Unit Under Calibration	Reading	Correction	Coverage Factor (k)	Uncertainty (±)
Conductivity Solution					
25.000 µS/cm	µS/cm	25.0	0.000	2.00	0.22 µS/cm
1413.0 µS/cm	µS/cm	1413	0.0	2.00	8.9 µS/cm
111.3 mS/cm	mS/cm	107.2	4.10	2.00	0.66 mS/cm

The End of Certificate



Equipment : Cooled Incubator
Condition As-Received : Used Item
Reference : 2205-0764OC-1
Procedure Used :-

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD).
The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34970A	MY44067817	21LM10	20 Jul 2022

2. This certificate is valid only to the item calibrated on date and place of calibration.


3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Not Available

Fan Setting : 100.0%



Probe Installation Details :

Probe	a	b	c	H	W	D
1	10 cm	10 cm	14 cm	0.48 m	0.65 m	1.3 m
2				0.40 m ³		

Dimension of Chamber :

Dimension	Value
D	0.48 m
W	0.65 m
H	1.3 m
Capacity	0.40 m ³

Environment during calibration		
	Beginning	Finished
Temp. (°C)	21	21
REL.Humid. (%)	67	65
AC Supply (Volt)	233	234

Position :	Ref. Std. ID No.:
1	15RTD2/11
2	15RTD2/12
3	15RTD2/13
4	15RTD2/14
5	15RTD2/15
6	15RTD2/16
7	15RTD2/17
8	15RTD2/18
9 (ref.)	15RTD2/19



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert. No.: 22TM347
Page: 1 of 3

Certificate of Calibration

Equipment :	Cooled Incubator
Manufacturer :	Binder
Model :	KB 400 E6
Serial No. :	20200000015535
ID No. :	UAE.MIC.018/2564
Submitted by :	United Analyst and Engineering Consultant Co.,Ltd. 3 Soi Udomsuk 41, Sukhumvit Road, Bangchak, Phrakhanong, Bangkok 10260
Location :	Microbiology Laboratory
Received Order :	27 May 2022
Calibration Date :	27 May 2022
Ambient Temperature :	(26 ± 10) °C
Relative Humidity :	(50 ± 30) %
Calibrated by :	Suwit Imjai
Approved by :	 Ponthippa Tameyakul Malee Bulkruea
Approved Signatory	

Issue Date : 2 June 2022
The Uncertainties are for a confidence probability of approximately 95 %

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Service

เอกสารควบคุม

เอกสารควบคุม



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PAITANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-27 FAX. 0-2719-9484



HAEC-TLTEST17025
CALIBRATION 0008

Cert. No.: 22TM672
Page.: 1 of 3

Certificate of Calibration

Equipment : Incubator
Manufacturer : Memmert
Model : IPP 260
Serial No. : V616.0066
ID No. : UAE.MIC.032/2559
Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
3 Soi Udomsuk 41, Sukhumvit Road,
Bangchak, Phrakhanong,
Bangkok 10260
Location : Microbiology Laboratory (302)
Received Order : 3 May 2022
Calibration Date : 5 May 2022
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %
Calibrated by : Preecha Hiahib
Approved by : 
Approved Signatory
() Pornthippa Tameyakul
(✓) Malee Butkruea
() Suwit Imjai

Issue Date :

11 May 2022

The Uncertainties are for a confidence probability of approximately 95 %

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing

เอกสารไม่ควบคุม



Equipment : Cooled Incubator
Condition As-Received : Used Item
Reference : 2205-0764OC-1
Result of Calibration :- (*) Without Adjustment
Function of UUC* : Temperature Source
Fresh air setting : Not Available

Calibration Point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Temperature stability (± °C)	Temperature uniformity (°C)	Overall Variation (°C)	Uncertainty (±°C)	Coverage Factor k
35.0	34.9	34.9	0.017	0.31	0.38	0.30	2

Measured Temperature (°C)									
Calibration Point (°C)	Position								
	1	2	3	4	5	6	7	8	9 (ref.)
35.0	34.808	35.139	34.922	35.062	35.109	35.161	35.132	35.129	35.092

Average* : The average of 30 values in each position.

Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor.

Temperature uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-o-o-

เอกสารควบคุม



Equipment : Incubator
Condition As-Received : Used Item
Reference : 2205-0003OC-3
Result of Calibration :- (*) Without Adjustment
Function of UUC* : Temperature Source
Fresh air setting : Not Available

Cert. No.: 22TM672
Page.: 3 of 3

Calibration Point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Temperature stability (± °C)	Temperature uniformity (°C)	Overall Variation (°C)	Uncertainty (±°C)	Coverage Factor k
25.0	25.0	25.0	0.021	0.18	0.33	0.30	2
36.0	36.0	36.0	0.077	0.96	1.8	0.33	2

Measured Temperature (°C)								
Position								
1	2	3	4	5	6	7	8	9 (ref.)
25.221	25.146	25.127	25.113	24.968	24.986	24.933	25.017	25.047
35.637	35.238	36.130	36.515	36.928	36.845	36.630	36.761	36.113

Average* : The average of 30 values in each position.

Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor.
Temperature uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.
Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation.
UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

-o0o-



Equipment : Incubator
Condition As-Received : Used Item
Reference : 2205-0003OC-3
Procedure Used :-

Cert. No.: 22TM672
Page.: 2 of 3

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD).
The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument Model Serial No. Cert. No. Due Date
1) Data Acquisition 34970A MY44067817 21LM10 20 Jul 2022

2. This certificate is valid only to the item calibrated on date and place of calibration.


3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Not Available

Environment during calibration		
	Beginning	Finished
Temp. (°C)	25	23
REL.Humid. (%)	62	57
AC Supply (Volt)	221	221



Probe Installation Details :

a = 5.0 cm
b = 5.0 cm
c = 5.0 cm

Dimension of Chamber :
D = 0.50 m
W = 0.60 m
H = 0.80 m
Capacity = 0.24 m³

Position :	Ref. Std. ID No.:
1	19-15RTD-01
2	19-15RTD-02
3	19-15RTD-03
4	19-15RTD-04
5	19-15RTD-05
6	21-15RTD-06
7	19-15RTD-07
8	19-15RTD-08
9 (ref.)	19-15RTD-09

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม



Equipment : Water Bath
Condition As-Received : Used Item
Reference : 2202-0444OC-3
Procedure Used :-

Calibration were conducted using in-house calibration procedure CP-OT04 according to direct measurement method with Data Acquisition which connected with Industrial Platinum Resistance Thermometer (IPRT).

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34970A	MY44067817	21LM10	20 Jul 2022


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

	Environmental		AC Voltage Supply (Volt)
	(°C)	(%R.H.)	
Beginning of Calibration	21	65	229
Finished of Calibration	22	58	230



Front

Position :	Ref. Std. ID No.:
1	70RC143
2	70RC144
3	70RC145
4	70RC146
5(ref.)	70RC147



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert. No.: 22TM333
Page.: 1 of 3

Certificate of Calibration

Equipment :	Water Bath
Manufacturer :	Memmert
Model :	WNE 14
Serial No. :	L416.0606
ID No. :	UAE.MIC.002/2560
Submitted by :	United Analyst and Engineering Consultant Co.,Ltd. 3 Soi Udomsuk 41, Sukhumvit Road, Bangchak, Phrakhanong, Bangkok 10260
Location :	Microbiology Laboratory
Received Order :	17 February 2022
Calibration Date :	17 February 2022
Ambient Temperature :	(26 ± 10) °C
Relative Humidity :	(50 ± 30) %
Calibrated by :	Suwit Imjai

Approved by : 
Approved Signatory

() Pornthippa Tameyakul
() Malee Butkruea

Issue Date : 22 February 2022
The Uncertainties are for a confidence probability of approximately 95%.

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3: Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert. No.: 22TM334
Page.: 1 of 3

Certificate of Calibration

Equipment : Water Bath
Manufacturer : Memmert
Model : WNE 14
Serial No. : L416.0612
ID No. : UAE.MIC.003/2560
Submitted by : United Analyst and Engineering Consultant Co., Ltd.
3 Soi Udomsuk 41, Sukhumvit Road,
Bangchak, Phrakhanong,
Bangkok 10260
Location : Microbiology Laboratory
Received Order : 17 February 2022
Calibration Date : 17 February 2022
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %
Calibrated by : Suwit Imjai

Approved by : 
Approved Signatory

() Ponthippa Tameyakul
(/) Malee Butkruea

Issue Date : 22 February 2022

The Uncertainties are for a confidence probability of approximately 95%.

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม

A 0038095



Equipment : Water Bath
Condition As-Received : Used Item
Reference : 2202-0444OC-3
Result of Calibration :- (*) Without Adjustment
Function of UUC* : Temperature Source

Cert. No.: 22TM333
Page.: 3 of 3

Calibration point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Average* Standard Reading (°C)				
			Position			Uncertainty (± °C)	Coverage Factor k
44.5	44.5	44.5	1	2	3		
			44.498	44.481	44.482	44.518	44.534

Calibration point (°C)	Uniformity (°C)	Stability (± °C)	Uncertainty (± °C)	Coverage Factor k
44.5	0.13	0.057	0.15	2

Average* : The average of 30 values in each position.

Uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Stability : One-half of the greatest maximum difference of measured temperature at any one probe.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor **k**, providing a level of confidence of approximately 95 %.

-o0o-

เอกสารไม่ควบคุม



Equipment : Water Bath
Condition As-Received : Used Item
Reference : 2202-0444OC-4
Result of Calibration :- (*) Without Adjustment
Function of UUC* : Temperature Source

Cert. No.: 22TM334
Page.: 3 of 3

Calibration point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Average* Standard Reading (°C)				
			1	2	3	4	5 (ref.)
44.5	44.5	44.5	44.572	44.514	44.507	44.530	44.565

Calibration point (°C)	Uniformity (°C)	Stability (± °C)	Uncertainty (± °C)	Coverage Factor <i>k</i>
44.5	0.10	0.042	0.15	2

Average* : The average of 30 values in each position.

Uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Stability : One-half of the greatest maximum difference of measured temperature at any one probe.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-o0o-



Equipment : Water Bath
Condition As-Received : Used Item
Reference : 2202-0444OC-4

Cert. No.: 22TM334
Page.: 2 of 3

Procedure Used :-

Calibration were conducted using in-house calibration procedure CP-OT04 according to direct measurement method with Data Acquisition which connected with Industrial Platinum Resistance Thermometer (IPT).

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument Model Serial No. Cert. No. Due Date
1) Data Acquisition 34970A MY44067817 21LM10 20 Jul 2022


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

	Environmental		AC Voltage Supply (Volt)
	(°C)	(%R.H.)	
Beginning of Calibration	21	65	229
Finished of Calibration	22	57	230



Front

Position :	Ref. Std. ID No.:
1	70RC143
2	70RC144
3	70RC145
4	70RC146
5(ref.)	70RC147



เอกสารไม่ควบคุม

๑ 1096054



เอกสารไม่ควบคุม

๑ 1096055


Measurement Results

Repeatability



The "d" in the graph represents the readability of the range/interval in which the test was performed.
The results of this graph are based upon the absolute values of the differences from the mean value.

Eccentricity



The "d" in the graph represents the readability of the range/interval in which the test was performed.

Mettler-Toledo (Thailand) Ltd.
846/4 - 846/5 Lasalle Rd., Bangna Tai Sub-District
Bangna District, Bangkok 10260
+66 2723 0382
MT-TH.ServicesSupport@mt.com

Accuracy Calibration Certificate

Customer

Company: United Analyst and Engineering Consultant Co., Ltd.
Address: 3 Soi Udom Suk 41, Sukhumvit Rd., Bang Chak
City: Phra Khanong
Zip / Postal: 10260
State / Province: Bangkok
Order Number: 4 0 3 3 7 4 0 1 4 9 4

Contact: Suwit Chotnok

Weighing Device

Manufacturer: Mettler Toledo
Model: MS603S/01
Serial No.: B007010311
Building: N/A
Floor: 2
Room: Balance Room (206)
Instrument Type: Weighing Instrument
Asset Number: UAE.MIC.008/2553
Terminal Model: N/A
Terminal Serial No.: N/A
Terminal Asset No.: N/A

Range	Max. Capacity	Readability (d)
1	620 g	0.001 g

Procedure

Calibration Guideline:

METTLER TOLEDO Work Instruction:

EURAMET cg-18 v. 4.0 (11/2015)
CP/W002/20
This calibration certificate contains measurements for As Found calibration. No As Left calibration was performed because the device was not modified after As Found calibration. Therefore, results for As Left correspond to As Found.

The sensitivity/span of the weighing instrument was adjusted before calibration with a built-in weight.

In accordance with EURAMET cg-18 (11/2015), the test loads were selected to reflect the specific use of the weighing device or to accommodate specific calibration conditions.

As Found	Start: 22.8 °C	End: 23.0 °C	Start: 49.9 %	End: 58.3 %
Temperature	Humidity			

As Found Calibration Date: 07-Apr-2022

As Left Calibration Date: N/A

Issue Date: 08-Apr-2022

Calibrator:

Sirawit Chanchan

Approved Signatory:

Kassakorn Tassanachalsakul

☒ Kassakorn Tassanachalsakul
☐ Santi Jitnyom
☐ Surachet Sukkate

Test Equipment

All weights used for metrological testing are traceable to national or international standards. The weights were calibrated and certified by an accredited calibration laboratory.

Weight Set 1: OIML F1

Weight Set No.:	WS55	Date of Issue:	09-Jul-2021
Certificate Number:	CCM-0137-21-C	Calibration Due Date:	07-Jul-2022

Weight Set 2: OIML E2

Weight Set No.:	WS60	Date of Issue:	23-Feb-2022
Certificate Number:	C208581631	Calibration Due Date:	14-Aug-2023

Thermo Hygrometer

Equipment No.:	IN161	Date of Issue:	14-Jun-2021
Certificate Number:	21H1220	Calibration Due Date:	01-Jun-2022

Remarks

- FACT adjustment functionality activated
- Equipment condition: Good
- Next calibration according to customer's procedure
- Calibration data not decide by calibration laboratory


End of Accredited Section

The information below and any attachments to this calibration certificate are not part of the accredited calibration.

Error of Indication

As Found

	Reference Value	Indication	Error of Indication	Expanded Uncertainty	k
1	0.000 g	0.000 g	0.000 g	1.0 mg	2
2	0.500 g	0.500 g	0.000 g	1.2 mg	2
3	1.000 g	1.000 g	0.000 g	1.2 mg	2
4	50.000 g	50.000 g	0.000 g	1.2 mg	2
5	100.000 g	100.000 g	0.000 g	1.3 mg	2
6	150.000 g	150.000 g	0.000 g	1.5 mg	2
7	200.000 g	200.001 g	0.001 g	1.6 mg	2
8	300.001 g	300.001 g	0.000 g	2.0 mg	2
9	400.001 g	400.001 g	0.000 g	2.5 mg	2
10	500.001 g	500.002 g	0.001 g	2.9 mg	2
11	600.001 g	600.001 g	0.000 g	3.4 mg	2



The uncertainty stated is the expanded uncertainty at calibration obtained by multiplying the standard combined uncertainty by the coverage factor $k = 2$ which can be larger than 2 according to EURAMET cg-18. The value of the measurand lies within the assigned range of values with a probability of approximately 95%.

The user is responsible for maintaining environmental conditions and the settings of the weighing instrument when it was calibrated.

Measurement Uncertainty of the Weighing Instrument in Use

Stated is the expanded uncertainty with $k=2$ in use. The formula shall be used for the estimation of the uncertainty under consideration of the errors of indication. The value R represents the net load indication in the unit of measure of the device.

Temperature coefficient for the evaluation of the measurement uncertainty in use: $3.0 \cdot 10^{-6} / K$

Temperature range on site for the evaluation of the measurement uncertainty in use: $3 K$


Linearization of Uncertainty Equation

Range	As Found		As Left
	d	Max	
1	0.001 g	620 g	$U_1 = 1.2 \text{ mg} + 0.0186 \text{ mg/g} \cdot R$
			N/A

To optimize the stability of the linearization, besides of the zero load only increasing measurement points with a test load of 5% of the measurement range or larger are taken for the calculation of the linear equation.

Absolute and Relative Measurement Uncertainty in Use for Various Net Indications (Examples)

Net Indication	As Found	As Left
0.062 g	1.2 mg	1.9%
0.620 g	1.2 mg	0.20%
6.200 g	1.3 mg	0.021%
62.000 g	2.4 mg	0.0038%
620.000 g	13 mg	0.0021%
		N/A
		N/A



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
53/44 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL: 0-2717-3000-27 FAX: 0-2719-9484



Cert. No.: 22TM89
Page.: 1 of 3

Certificate of Calibration

Equipment : Autoclave

Manufacturer : ALP

Model : CL-40L

Serial No. : 802664

ID No. : UAE.MIC.014/2550

Submitted by : United Analyst and Engineering Consultant Co., Ltd.
3 Soi Udomsuk 41, Sukhumvit Road,
Bangchak, Phrakhanong,
Bangkok 10260

Location : Air Analysis Unit

Received Order : 17 February 2022

Calibration Date : 17 February 2022

Ambient Temperature : $(26 \pm 10) ^\circ C$

Relative Humidity : $(50 \pm 30) \%$

Calibrated by : Kunchit Promprat

Approved by : 
Approved Signatory

() Ponthippa Tameyakul
() Malee Butkruea
() Suwit Imjai

Issue Date : 22 February 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม



Equipment : Autoclave
Condition As-Received : Used Item
Reference : 2202-0444OC-1
Result of Calibration :- (*) Without Adjustment

Cert. No.: 22TM89
Page.: 3 of 3

Operating parameter Set : Temperature = 122 °C
Sterilization period = 30 minute

UUC* Setting (°C)	UUC* Reading (°C)	Position	Average* Standard Reading (°C)	Stability (± °C)	Pressure Reading (MPa)	Uncertainty (± °C)	Coverage Factor <i>k</i>
122	122	1	122.373	0.32	0.12	1.2	2
		2	122.421				
		3	122.292				

Average* : The average of 30 values in each position.

Stability : One-half of the greatest maximum difference of measured temperature at any one probe.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-oOo-



Equipment : Autoclave
Condition As-Received : Used Item
Reference : 2202-0444OC-1
Procedure Used :-

Cert. No.: 22TM89
Page.: 2 of 3

Calibration were conducted using in-house calibration procedure CP-QT03 according to direct measurement method with Data Acquisition which connected with Thermocouple Type T
The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument :-


Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34970A	MY44035217	21LM30	23 Dec 2022
2. This certificate is valid only to the item calibrated on date and place of calibration.				
3. This certification is traceable to the International System of Unit.				
4. This result of calibration covers laboratory autoclaves for the sterilization of goods and material which could be infected with organisms categorized as Hazard Group 1, 2 and 3**				

(** = Categorization of pathogens according to hazard and categories of containment, second edition, 1990)
It does not cover autoclaves for use with material infect with organisms in Hazard Group 4, for which complete containment and sterilization of infected condensate is considered to be essential.

This result of calibration does not apply to sterilizers or disinfectors used for medical, dental, pharmaceutical or veterinary purposes which are directly concerned with patient care, or those used for fabrics subjected to sterilization which are required to be dry at the end of cycle.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source



	Environmental	
	(°C)	(%R.H.) (Volt)
Beginning of Calibration	27	68 226
Finished of Calibration	27	65 226

Position	Description	Ref. Std. ID No.:
1 =	Center of chamber	22-10TC-01
2 =	Temperature sensor	22-10TC-02
3 =	Exhaust port	22-10TC-03