

ภาคผนวก ง

เอกสารสอบเทียบเครื่องมือที่ใช้ในการวิเคราะห์



รายการเครื่องมือที่ใช้ในการวิเคราะห์ / ทดสอบ

Sample Name	Parameter	Equipment Name	ID No.	Calibrated Date	Next Cal	Freq. Calibrate (Months)
Ambient	Total Suspended Particulate	High Volume	RYG_FS0291	-	-	On site Calibration
Ambient	Total Suspended Particulate	High Volume	RYG_FS0396	-	-	On site Calibration
Ambient	Total Suspended Particulate	High Volume	RYG_FS0174	-	-	On site Calibration
Ambient	Total Suspended Particulate	High Volume	RYG_FS0393	-	-	On site Calibration
Ambient	Total Suspended Particulate	Digital Balance	RYG_EN0001	23-Mar-22	23-Mar-23	12
Ambient	Particulate Matter (PM-10)	High Volume	RYG_FS0398	-	-	On site Calibration
Ambient	Particulate Matter (PM-10)	High Volume	RYG_FS0400	-	-	On site Calibration
Ambient	Particulate Matter (PM-10)	High Volume	RYG_FS0187	-	-	On site Calibration
Ambient	Particulate Matter (PM-10)	High Volume	RYG_FS0185	-	-	On site Calibration
Ambient	Particulate Matter (PM-10)	Digital Balance	RYG_EN0001	23-Mar-22	23-Mar-23	12
Ambient	Nitrogen Dioxide	NO ₂ Analyzer	RYG_FS0535	4-Jan-22	4-Jul-22	6
Ambient	Nitrogen Dioxide	NO ₂ Analyzer	RYG_FS0533	4-Jan-22	4-Jul-22	6
Ambient	Nitrogen Dioxide	NO ₂ Analyzer	RYG_FS0453	4-Jan-22	4-Jul-22	6
Ambient	Nitrogen Dioxide	NO ₂ Analyzer	RYG_FS0264	4-Jan-22	4-Jul-22	6
Ambient	Sulfur Dioxide	SO ₂ Analyzer	RYG_FS0534	4-Jan-22	4-Jul-22	6
Ambient	Sulfur Dioxide	SO ₂ Analyzer	RYG_FS0532	4-Jan-22	4-Jul-22	6
Ambient	Sulfur Dioxide	SO ₂ Analyzer	RYG_FS0452	4-Jan-22	4-Jul-22	6
Ambient	Sulfur Dioxide	SO ₂ Analyzer	RYG_FS0263	4-Jan-22	4-Jul-22	6
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_FS0089	13-Jul-21	11-Jan-23	18
Noise	Leq 24 hrs	Sound Calibrator	RYG_FS0215	9-Aug-21	9-Aug-22	12
Noise	Leq 24 hrs	Sound Level Meter	RYG_FS0434	21-Jan-22	21-Jan-23	12
Noise	Leq 24 hrs	Sound Level Meter	RYG_FS0018	4-Oct-21	4-Oct-22	12
Noise	Leq 5 min	Sound Calibrator	RYG_FS0215	9-Aug-21	9-Aug-22	12
Noise	Leq 5 min	Sound Level Meter	RYG_FS0434	21-Jan-22	21-Jan-23	12
Noise	Leq 5 min	Sound Level Meter	RYG_FS0018	4-Oct-21	4-Oct-22	12
Noise	Noise Annoyance	Sound Calibrator	RYG_FS0215	9-Aug-21	9-Aug-22	12
Noise	Noise Annoyance	Sound Level Meter	RYG_FS0434	21-Jan-22	21-Jan-23	12
Noise	Noise Annoyance	Sound Level Meter	RYG_FS0018	4-Oct-21	4-Oct-22	12
Noise	Noise Annoyance	Sound Calibrator	RYG_FS0215	9-Aug-21	9-Aug-22	12
Noise	Noise Annoyance	Sound Level Meter	RYG_FS0431	21-Jan-22	21-Jan-23	12
Noise	Noise Annoyance	Sound Level Meter	RYG_FS0020	10-Jan-22	10-Jan-23	12
Rayong Lab	Cyanide	SPECTROPHOTOMETER	RYG_EN0037	1-Apr-21	1-Oct-22	18
Rayong Lab	Phenol	SPECTROPHOTOMETER	RYG_EN0037	1-Apr-21	1-Oct-22	18
Rayong Lab	pH at 25 °C	pH meter	RYG_EN0183	17-Mar-22	17-Mar-23	12
Rayong Lab	Nitrate	SPECTROPHOTOMETER	RYG_EN0037	1-Apr-21	1-Oct-22	18
Rayong Lab	Ammonia Nitrogen	SPECTROPHOTOMETER	RYG_EN0037	1-Apr-21	1-Oct-22	18
Rayong Lab	BOD (5 days at 20°C)	DO meter with Sensor	RYG_EN0140	2-Feb-21	3-Aug-22	18
Rayong Lab	BOD (5 days at 20°C)	Incubator	RYG_EN0154	22-Apr-22	21-Oct-23	12
Rayong Lab	Temperature	Digital Thermometer	RYG_FS0467	7-Jul-21	7-Jul-22	18
Rayong Lab	Dissolved Oxygen	Chamber (Cold Room)	RYG_EN0184	22-Feb-22	22-Feb-23	12
Water Lab	Hexavalent Chromium	Spectrophotometer	BKK_EN0018	15-Oct-21	15-Oct-22	12
Water Lab	Silver	ICP-MS	BKK_EL0043	30-Sep-21	29-Mar-23	18
Water Lab	Silver	Hot Block	BKK_EL0054	7-Apr-22	7-Oct-23	18
Water Lab	Silver	Chamber (Cold Room)	BKK_EN0167	18-May-21	16-Nov-22	18
Water Lab	Aluminium	ICP-MS	BKK_EL0043	30-Sep-21	29-Mar-23	18
Water Lab	Aluminium	Hot Block	BKK_EL0054	7-Apr-22	7-Oct-23	18
Water Lab	Aluminium	Chamber (Cold Room)	BKK_EN0167	18-May-21	16-Nov-22	18
Water Lab	Barium	ICP-MS	BKK_EL0043	30-Sep-21	29-Mar-23	18
Water Lab	Barium	Hot Block	BKK_EL0054	7-Apr-22	7-Oct-23	18
Water Lab	Barium	Chamber (Cold Room)	BKK_EN0167	18-May-21	16-Nov-22	18
Water Lab	Lead	ICP-MS	BKK_EL0043	30-Sep-21	29-Mar-23	18
Water Lab	Lead	Hot Block	BKK_EL0054	7-Apr-22	7-Oct-23	18
Water Lab	Lead	Chamber (Cold Room)	BKK_EN0167	18-May-21	16-Nov-22	18
Water Lab	Iron	ICP-MS	BKK_EL0043	30-Sep-21	29-Mar-23	18
Water Lab	Iron	Hot Block	BKK_EL0054	7-Apr-22	7-Oct-23	18
Water Lab	Iron	Chamber (Cold Room)	BKK_EN0167	18-May-21	16-Nov-22	18
Water Lab	Copper	ICP-MS	BKK_EL0043	30-Sep-21	29-Mar-23	18
Water Lab	Copper	Hot Block	BKK_EL0054	7-Apr-22	7-Oct-23	18
Water Lab	Copper	Chamber (Cold Room)	BKK_EN0167	18-May-21	16-Nov-22	18
Water Lab	Nickel	ICP-MS	BKK_EL0043	30-Sep-21	29-Mar-23	18
Water Lab	Nickel	Hot Block	BKK_EL0054	7-Apr-22	7-Oct-23	18
Water Lab	Nickel	Chamber (Cold Room)	BKK_EN0167	18-May-21	16-Nov-22	18
Water Lab	Arsenic	ICP-MS	BKK_EL0043	30-Sep-21	29-Mar-23	18
Water Lab	Arsenic	Hot Block	BKK_EL0054	7-Apr-22	7-Oct-23	18
Water Lab	Arsenic	Chamber (Cold Room)	BKK_EN0167	18-May-21	16-Nov-22	18



ALS Laboratory Group (Thailand) Co., Ltd.
104 Phatthanakan 40, Phatthanakan Rd.,
Khwaeng Phatthanakan, Khet Suan Luang,
Bangkok 10250 Thailand
T +66 2 760 3000 F +66 2 760 3197

รายการเครื่องมือที่ใช้ในการวิเคราะห์ / ทดสอบ

Sample Name	Parameter	Equipment Name	ID No.	Calibrated Date	Next Cal	Freq. Calibrate (Months)
Water Lab	Cadmium	ICP-MS	BKK_EL0043	30-Sep-21	29-Mar-23	18
Water Lab	Cadmium	Hot Block	BKK_EL0054	7-Apr-22	7-Oct-23	18
Water Lab	Cadmium	Chamber (Cold Room)	BKK_EN0167	18-May-21	16-Nov-22	18
Water Lab	Zinc	ICP-MS	BKK_EL0043	30-Sep-21	29-Mar-23	18
Water Lab	Zinc	Hot Block	BKK_EL0054	7-Apr-22	7-Oct-23	18
Water Lab	Zinc	Chamber (Cold Room)	BKK_EN0167	18-May-21	16-Nov-22	18
Water Lab	Trivalent Chromium	ICP-MS	BKK_EL0043	30-Sep-21	29-Mar-23	18
Water Lab	Trivalent Chromium	Hot Block	BKK_EL0054	7-Apr-22	7-Oct-23	18
Water Lab	Trivalent Chromium	Chamber (Cold Room)	BKK_EN0167	18-May-21	16-Nov-22	18
Water Lab	Mercury	CVAFS	BKK_EL0011	7-Jun-22	6-Jun-23	12
Water Lab	Total Coliform	Autoclave	BKK_ML0043	1-Dec-21	1-Jun-23	18
Water Lab	Total Coliform	Incubator	BKK_ML0010	5-Jan-21	6-Jul-22	18
Water Lab	Total Coliform	Hot Air Oven	BKK_ML0013	7-Jun-21	6-Dec-22	18
Water Lab	Fecal Coliform	Autoclave	BKK_ML0043	1-Dec-21	1-Jun-23	18
Water Lab	Fecal Coliform	Incubator	BKK_ML0010	5-Jan-21	6-Jul-22	18
Water Lab	Fecal Coliform	Hot Air Oven	BKK_ML0013	7-Jun-21	6-Dec-22	18
Water Lab	Fecal Coliform	Water Bath	BKK_ML0052	21-Feb-22	21-Feb-23	12
Water Lab	<i>Escherichia coli</i>	Autoclave	BKK_ML0043	1-Dec-21	1-Jun-23	18
Water Lab	<i>Escherichia coli</i>	Incubator	BKK_ML0010	5-Jan-21	6-Jul-22	18
Water Lab	<i>Escherichia coli</i>	Hot Air Oven	BKK_ML0013	7-Jun-21	6-Dec-22	18
Water Lab	<i>Escherichia coli</i>	Water Bath	BKK_ML0052	21-Feb-22	21-Feb-23	12
Rayong Lab	Sulfate	SPECTROPHOTOMETER	RYG_EN0037	1-Apr-21	1-Oct-22	18
Rayong Lab	Fluoride	pH ISE Meter	RYG_EN0152	23-Dec-21	23-Dec-22	12
Rayong Lab	Chloride	Burette	243007	21-Sep-18	21-Sep-23	60
Rayong Lab	Color	Chamber (Cold Room)	RYG_EN0184	22-Feb-22	22-Feb-23	12
Water Lab	Total Dissolved Solids 180°C	Electronic Top-Loading Balance	BKK_EN0002	25-Feb-22	25-Feb-23	12
Rayong Lab	Total Dissolved Solids 180°C	Chamber Oven	RYG_EN0010	5-May-21	3-Nov-22	18
Rayong Lab	Total Hardness	Chamber (Cold Room)	RYG_EN0184	22-Feb-22	22-Feb-23	12
Water Lab	Permanent Hardness	Burette	BKK_EN0171	30-Mar-21	28-Sep-22	18
Rayong Lab	Turbidity	Chamber (Cold Room)	RYG_EN0184	22-Feb-22	22-Feb-23	12
Water Lab	Total Alkalinity	Burette	BKK_EN0171	30-Mar-21	28-Sep-22	18
Water Lab	Manganese	ICP-MS	BKK_EL0043	30-Sep-21	29-Mar-23	18
Water Lab	Manganese	Hot Block	BKK_EL0054	7-Apr-22	7-Oct-23	18
Water Lab	Manganese	Chamber (Cold Room)	BKK_EN0167	18-May-21	16-Nov-22	18
Water Lab	Selenium	ICP-MS	BKK_EL0043	30-Sep-21	29-Mar-23	18
Water Lab	Selenium	Hot Block	BKK_EL0054	7-Apr-22	7-Oct-23	18
Water Lab	Selenium	Chamber (Cold Room)	BKK_EN0167	18-May-21	16-Nov-22	18



High Volume Air Sampler Calibration Worksheet

Project Site: Rojana Industrial Park Prachinburi Co., Ltd. (Prachinburi 1) Barometric Pressure (mm Hg): 755

Calibrate Location: โหล่งน้อย (A1) Temperature (°C): 32

Calibrate Date: 1-Jun-22 High Volume ID: RYG_FS0291


Calibration Sheet No.: C-010622-RYG_FS0291 High Volume Model: TE-5170D

Calibrator ID: RYG_FS0206 High Volume S/N: 5333

Calibrator Model: TE-5028A Calibrator Slope: 1.4867

Calibrator S/N: 1543 Calibrator Intercept: -0.0445

Test No.	Delta H ₂ O (inch)	Q _{std} (m ³ /min)	I: Chart (CFM)	Linear Regression
1	3.2	1.2299	44	Slope: 32.3732 Intercept: 4.2031 Correlation Coefficient: 0.9972
2	3.6	1.3018	46	
3	4.2	1.4026	50	
4	5.0	1.5263	54	
5	5.6	1.6127	56	



Calibrated by: Jan
(Mr. Jaradrawee Sriruksa)
Field Scientist(1)

Approved by: Mr. Noppong Juntarupan
(Mr. Noppong Juntarupan)
Enviro Field Coordinator Scientist (3)

FORM NO.: F 06-073 REVISION NO.: - ISSUE DATE: 14/03/16



High Volume Air Sampler Calibration Worksheet

Project Site: Rojana Industrial Park Prachinburi Co., Ltd. (Prachinburi 2) Barometric Pressure (mm Hg): 755

Calibrate Location: A2: บ้านนาบึง (A2) Temperature (°C): 32

Calibrate Date: 1-Jun-22 High Volume ID: RYG_FS0396


Calibration Sheet No.: C-010622-RYG_FS0396 High Volume Model: TE-5170D

Calibrator ID: RYG_FS0206 High Volume S/N: 5688

Calibrator Model: TE-5028A Calibrator Slope: 1.4867

Calibrator S/N: 1543 Calibrator Intercept: -0.0445

Test No.	Delta H ₂ O (inch)	Q _{std} (m ³ /min)	I: Chart (CFM)	Linear Regression
1	3.0	1.1923	44	Slope: 33.4939 Intercept: 3.9090 Correlation Coefficient: 0.9992
2	3.4	1.2664	46	
3	4.0	1.3698	50	
4	4.8	1.4963	54	
5	5.2	1.5556	56	



Calibrated by: Jan
(Mr. Jaradrawee Sriruksa)
Field Scientist(1)

Approved by: Mr. Noppong Juntarupan
(Mr. Noppong Juntarupan)
Enviro Field Coordinator Scientist (3)

FORM NO.: F 06-073 REVISION NO.: - ISSUE DATE: 14/03/16



High Volume Air Sampler Calibration Worksheet

Project Site: Rojana Industrial Park Prachinburi Co., Ltd. (Prachinburi 1) Barometric Pressure (mm Hg): 755

Calibrate Location: โหล่งน้อย (A2) Temperature (°C): 32

Calibrate Date: 1-Jun-22 High Volume ID: RYG_FS0174


Calibration Sheet No.: C-010622-RYG_FS0174 High Volume Model: TE-5170D

Calibrator ID: RYG_FS0206 High Volume S/N: 4800

Calibrator Model: TE-5028A Calibrator Slope: 1.4867

Calibrator S/N: 1543 Calibrator Intercept: -0.0445

Test No.	Delta H ₂ O (inch)	Q _{std} (m ³ /min)	I: Chart (CFM)	Linear Regression
1	3.6	1.3018	44	Slope: 35.9279 Intercept: -2.9023 Correlation Coefficient: 0.9992
2	4.0	1.3698	46	
3	4.6	1.4658	50	
4	5.4	1.5844	54	
5	5.8	1.6404	56	



Calibrated by: Jan
(Mr. Jaradrawee Sriruksa)
Field Scientist(1)

Approved by: Mr. Noppong Juntarupan
(Mr. Noppong Juntarupan)
Enviro Field Coordinator Scientist (3)

FORM NO.: F 06-073 REVISION NO.: - ISSUE DATE: 14/03/16



High Volume Air Sampler Calibration Worksheet

Project Site: Rojana Industrial Park Prachinburi Co., Ltd. (Prachinburi 2) Barometric Pressure (mm Hg): 755

Calibrate Location: A4: บ้านนาบึง (A4) Temperature (°C): 32

Calibrate Date: 1-Jun-22 High Volume ID: RYG_FS0393


Calibration Sheet No.: C-010622-RYG_FS0393 High Volume Model: TE-5170D

Calibrator ID: RYG_FS0206 High Volume S/N: 5682

Calibrator Model: TE-5028A Calibrator Slope: 1.4867

Calibrator S/N: 1543 Calibrator Intercept: -0.0445

Test No.	Delta H ₂ O (inch)	Q _{std} (m ³ /min)	I: Chart (CFM)	Linear Regression
1	2.2	1.0274	44	Slope: 23.2290 Intercept: 20.1163 Correlation Coefficient: 0.9976
2	2.6	1.1130	46	
3	3.6	1.3018	50	
4	4.4	1.4345	54	
5	5.2	1.5556	56	



Calibrated by: Jan
(Mr. Jaradrawee Sriruksa)
Field Scientist(1)

Approved by: Mr. Noppong Juntarupan
(Mr. Noppong Juntarupan)
Enviro Field Coordinator Scientist (3)

FORM NO.: F 06-073 REVISION NO.: - ISSUE DATE: 14/03/16



PENTA CALIBRATION CO., LTD.
66/124 The Connect 33 Village Kancharaphisek Road
Dokmai Prawe Bangkok 10250
Tel: +66 (0) 2069-9773
www.pentalab.com

RYG_EN0001

Certificate of Calibration

Represent to Certificate of Calibration PTC/07/22102

Certificate No.: PTC/07/22102 Page: 1 of 2
Equipment: Digital Balance Condition: Normal
Manufacturer: Sartorius Serial No.: 25409604
Model: LA130S-F ID No.: RYG_EN0001
Type of Balance: Single interval



Customer: ALS Laboratory Group (Thailand) Co., Ltd.
616/10 Moo 5 T. Maenamkoo, A. Pluakdaeng,
Rayong 21140, Thailand

REVIEW BY: *Thirakul*
APPROVED BY: *J. Noppong*
NEXT CAL. DATE: 23/03/23

Environment Condition: Temperature 23.9 °C ± 0.3 °C
Humidity 58.1 %RH ± 4.4 %RH
Air density 1.17 kg/m³

Calibration Place: ALS Laboratory Group (Thailand) Co., Ltd.
616/10 Moo 5 T. Maenamkoo, A. Pluakdaeng,
Rayong 21140, Thailand

The Method used: In house method PTC-WI-07, base on Euramet cg. 18

Traceability: This certificate is traceable to the SI Units through Thai Calibration Service Co., Ltd.
NSC-ONSC Accreditation No. Calibration 0189

Date Received: March 23, 2022

Calibration Date: March 23, 2022

Issued Date: March 25, 2022

Calibration By: Mr. Rungroey Metakul



Mr. Krongsak Kalasin
(Mr. Krongsak Kalasin)
Reviewed by

Mr. Keattisak Kiercio
(Mr. Keattisak Kiercio)
Laboratory Manager

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM). The effect that the results relate only to the items calibrated.

This calibration certificate shall not be reproduced except in full only, without written approval from Penta Calibration Co., Ltd.

PTC-FNC-01-02 2 Feb 2020



PENTA CALIBRATION CO., LTD.
66/124 The Connect 33 Village Kancharaphisek Road
Dokmai Prawe Bangkok 10250
Tel: +66 (0) 2069-9773
www.pentalab.com

Represent to Certificate of Calibration PTC/07/22102

Certificate No.: PTC/07/22102


Page: 2 of 2

Measurement Results:

Without Adjustment:

Function Calibration: Non Adjustment

Eccentric Error: Weight to be 1/3, 1/2 or of Maximum capacity



Eccentricity test 50 (g)

Position (g)				
1	2	3	4	5
0.0000	0.0000	-0.0001	0.0000	0.0001
Maximum deviation: 0.0001				

Repeatability Test: Weight to be 1/2 ≤ L₁ ≤ Maximum capacity

Determination of the standard deviation of weighing balance: Readability 0.0001 (g)

Nominal test value (g)	Standard Deviation
100	0.00009

Error of Indication: from nominal value, Readability 0.0001 (g)

Nominal Value (g)	Conventional Mass (g)	Indication (g)	Correction of Balance (g)	Uncertainty (g)	k
0	0.00000	0.0000	0.0000	0.00026	2.87
0.01	0.01000	0.0100	0.0000	0.00026	2.65
0.05	0.05000	0.0500	0.0000	0.00026	2.65
0.1	0.10000	0.1000	0.0000	0.00026	2.65
0.5	0.50000	0.4999	0.0001	0.00026	2.65
1	1.00000	0.9999	0.0001	0.00026	2.65
2	2.00000	1.9999	0.0001	0.00026	2.65
5	5.00001	5.0000	0.0000	0.00026	2.65
10	10.00000	10.0001	-0.0001	0.00026	2.65
20	20.00003	20.0001	-0.0001	0.00026	2.52
100	100.00004	100.0001	-0.0001	0.00027	2.18

Note: Weight of adjust (g)

The End of Certificate


PTC-FNC-01-02 2 Feb 2020



High Volume Air Sampler Calibration Worksheet

Project Site: Rojana Industrial Park Prachinburi Co., Ltd. (Prachinburi 1)
Calibrate Location: *3a Tanasak (A1)*
Calibrate Date: 1-Jun-22
Calibration Sheet No.: C-010622-RYG-FS0398
Calibrator ID: RYG-FS0206
Calibrator Model: TE-5028A
Calibrator S/N: 1543
Barometric Pressure (mm Hg): 755
Temperature (°C): 32
High Volume ID: RYG-FS0398
High Volume Model: TE-5009X
High Volume S/N: 5684
Calibrator Slope: 1.4867
Calibrator Intercept: -0.0445

Test No.	Delta H ₂ O (inch)	Q _{std} (m ³ /min)	I: Chart (CFM)	Linear Regression
1	2.0	0.9817	34	Slope: 31.8277 Intercept: 2.3368 Correlation Coefficient: 0.9976
2	2.4	1.0711	36	
3	3.0	1.1923	40	
4	3.6	1.3018	44	
5	4.0	1.3698	46	



Calibrated by: *J. Noppong*
(Mr. Jaradrawee Sritruxa)
Field Scientist(1)

Approved by: *J. Noppong*
(Mr. Noppong Juntarupan)
Enviro Field Coordinator Scientist (3)


FORM NO. F-06-074 REVISION NO.: ISSUE DATE: 14/03/16



High Volume Air Sampler Calibration Worksheet

Project Site: Rojana Industrial Park Prachinburi Co., Ltd. (Prachinburi 2)
Calibrate Location: *A2: 31WUE37*
Calibrate Date: 1-Jun-22
Calibration Sheet No.: C-010622-RYG-FS0400
Calibrator ID: RYG-FS0206
Calibrator Model: TE-5028A
Calibrator S/N: 1543
Barometric Pressure (mm Hg): 755
Temperature (°C): 32
High Volume ID: RYG-FS0400
High Volume Model: TE-5009X
High Volume S/N: 5691
Calibrator Slope: 1.4867
Calibrator Intercept: -0.0445

Test No.	Delta H ₂ O (inch)	Q _{std} (m ³ /min)	I: Chart (CFM)	Linear Regression
1	2.2	1.0274	34	Slope: 28.2904 Intercept: 4.8763 Correlation Coefficient: 0.9976
2	2.6	1.1130	36	
3	3.2	1.2299	40	
4	4.0	1.3698	44	
5	4.6	1.4658	46	



Calibrated by: *J. Noppong*
(Mr. Jaradrawee Sritruxa)
Field Scientist(1)

Approved by: *J. Noppong*
(Mr. Noppong Juntarupan)
Enviro Field Coordinator Scientist (3)


FORM NO. F-06-074 REVISION NO.: ISSUE DATE: 14/03/16



High Volume Air Sampler Calibration Worksheet

Project Site: Rojana Industrial Park Prachinburi Co., Ltd. (Prachinburi 1) Barometric Pressure (mm Hg): 755
 Calibrate Location: บ้านนาบ่อปลา (A2) Temperature (°C): 32
 Calibrate Date: 1-Jun-22 High Volume ID: RYG_F50187
 CalibrationSheet No.: C-010622-RYG_F50187 High Volume Model: TE-5009X
 Calibrator ID: RYG_F50206 High Volume S/N: 4795
 Calibrator Model: TE-5028A Calibrator Slope: 1.4867
 Calibrator S/N: 1543 Calibrator Intercept: -0.0445

Test No.	Delta H ₂ O (Inch)	Q _{std} (m ³ /min)	I: Chart (CFM)	Linear Regression
1	2.2	1.0274	34	Slope: 50.2214 Intercept: -17.7366 Correlation Coefficient: 0.9996
2	2.4	1.0711	36	
3	2.8	1.1534	40	
4	3.2	1.2299	44	
5	3.4	1.2664	46	



Calibrated by:
 (Mr. Jaradrawee Srituksa)
 Field Scientist(1)

Approved by:
 (Mr. Roppong Juntarupun)
 Enviro Field Coordinator Scientist (3)


FORM NO. F 06-074 REVISION NO. - ISSUE DATE: 14/03/16



High Volume Air Sampler Calibration Worksheet

Project Site: Rojana Industrial Park Prachinburi Co., Ltd. (Prachinburi 2) Barometric Pressure (mm Hg): 755
 Calibrate Location: A4: บ้านนาบ่อปลา Temperature (°C): 32
 Calibrate Date: 1-Jun-22 High Volume ID: RYG_F50185
 CalibrationSheet No.: C-010622-RYG_F50185 High Volume Model: TE-5009X
 Calibrator ID: RYG_F50206 High Volume S/N: 4793
 Calibrator Model: TE-5028A Calibrator Slope: 1.4867
 Calibrator S/N: 1543 Calibrator Intercept: -0.0445

Test No.	Delta H ₂ O (Inch)	Q _{std} (m ³ /min)	I: Chart (CFM)	Linear Regression
1	2.2	1.0274	34	Slope: 29.9152 Intercept: 3.0564 Correlation Coefficient: 0.9991
2	2.6	1.1130	36	
3	3.2	1.2299	40	
4	4.0	1.3698	44	
5	4.4	1.4345	46	



Calibrated by:
 (Mr. Jaradrawee Srituksa)
 Field Scientist(1)

Approved by:
 (Mr. Roppong Juntarupun)
 Enviro Field Coordinator Scientist (3)


FORM NO. F 06-074 REVISION NO. - ISSUE DATE: 14/03/16



MULTIPOINT CALIBRATION REPORT

Calibration Date: 4-Jan-22 Equipment Name: NOx Analyzer
 Manufacturer: Teledyne API Model: T200
 Serial No.: 7238 Equipment ID: RYG_F80535
 Calibrator Manufacturer: Teledyne API Model: 700
 Serial No.: 947
 Std. Gas Concentration (PPM): 51.33 Cylinder No.: LL36833
 Cylinder Pressure (psi): 1200 Certified By: Airgas Inc.
 Certified Date: 18-Mar-14 Expired Date: 18-Mar-22

Point	CALIBRATION RESULTS						
	Ideal	Actual NO	Error NO	%Error NO	Actual NOx	Error NOx	%Error NOx
ZERO	0.00	0.10	0.10	0.10	0.10	0.10	0.10
1	100.00	99.60	-0.40	-0.40	101.00	1.00	1.00
2	200.00	198.30	-1.70	-0.85	201.10	1.10	0.55
3	300.00	298.80	-1.20	-0.40	301.50	1.50	0.50
4	400.00	398.20	-1.80	-0.45	402.30	2.30	0.58
AVERAGE (%)				-0.40			0.55



Calibrated By:
 (Mr. Jirawut Sakarn)
 Field Environmental Scientist (3)

Approved By:
 (Mr. Sarayuth Jittrantorn)
 Assistant General Manager


ALS Laboratory Group
 FORM NO. F 06-056 REVISION NO. - ISSUE DATE: 02/04/12



MULTIPOINT CALIBRATION REPORT

Calibration Date: 4-Jan-22 Equipment Name: NOx Analyzer
 Manufacturer: Teledyne API Model: T200
 Serial No.: 7238 Equipment ID: RYG_F80533
 Calibrator Manufacturer: Teledyne API Model: 700
 Serial No.: 947
 Std. Gas Concentration (PPM): 51.33 Cylinder No.: LL36833
 Cylinder Pressure (psi): 1200 Certified By: Airgas Inc.
 Certified Date: 18-Mar-14 Expired Date: 18-Mar-22

Point	CALIBRATION RESULTS						
	Ideal	Actual NO	Error NO	%Error NO	Actual NOx	Error NOx	%Error NOx
ZERO	0.00	0.10	0.10	0.10	0.10	0.10	0.10
1	100.00	99.50	-0.50	-0.50	101.10	1.10	1.10
2	200.00	198.70	-1.30	-0.65	201.20	1.20	0.60
3	300.00	298.80	-1.20	-0.40	301.10	1.10	0.37
4	400.00	398.00	-2.00	-0.50	402.00	2.00	0.50
AVERAGE (%)				-0.39			0.53



Calibrated By:
 (Mr. Jirawut Sakarn)
 Field Environmental Scientist (3)

Approved By:
 (Mr. Sarayuth Jittrantorn)
 Assistant General Manager


ALS Laboratory Group
 FORM NO. F 06-056 REVISION NO. - ISSUE DATE: 02/04/12



MULTIPOINT CALIBRATION REPORT

Calibration Date	4-Jan-22	Equipment Name	NOx Analyzer
Manufacturer	HORIBA	Model	APNA-370
Serial No.	AWXG87CR	Equipment ID	RYG_FS0453
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	847		
Std. Gas Concentration (PPM)	51.33	Cylinder No.	LL36633
Cylinder Pressure (psi)	1200	Certified By	Airgas Inc.
Certified Date	18-Mar-14	Expired Date	18-Mar-22

Point	CALIBRATION RESULTS						
	Ideal	Actual NO	Error NO	%Error NO	Actual NOx	Error NOx	%Error NOx
ZERO	0.00	0.10	0.10	0.10	0.10	0.10	0.10
1	100.00	99.60	-0.40	-0.40	101.40	1.40	1.40
2	200.00	198.60	-1.40	-0.70	199.80	-0.20	-0.10
3	300.00	299.00	-1.00	-0.33	298.50	-1.50	-0.50
4	400.00	402.10	2.10	0.53	401.20	1.20	0.30
AVERAGE (%)				-0.16			0.24



Calibrated By

(Mr. Jirawut Sakam)
Field Environmental Scientist (3)

Approved By

(Mr. Sarayuth Jitranont)
Assistant General Manager


ALS Laboratory Group
FORM NO. F 06-056 REVISION NO. - ISSUE DATE: 02/04/12



MULTIPOINT CALIBRATION REPORT

Calibration Date	4-Jan-22	Equipment Name	NOx Analyzer
Manufacturer	HORIBA	Model	APNA-370
Serial No.	8G314J3K	Equipment ID	RYG_FS0284
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	847		
Std. Gas Concentration (PPM)	51.33	Cylinder No.	LL36633
Cylinder Pressure (psi)	1200	Certified By	Airgas Inc.
Certified Date	18-Mar-14	Expired Date	18-Mar-22

Point	CALIBRATION RESULTS						
	Ideal	Actual NO	Error NO	%Error NO	Actual NOx	Error NOx	%Error NOx
ZERO	0.00	0.05	0.05	0.05	0.10	0.10	0.10
1	100.00	99.20	-0.80	-0.80	100.10	0.10	0.10
2	200.00	198.40	-1.60	-0.80	199.10	-0.90	-0.45
3	300.00	298.60	-1.40	-0.47	301.50	1.50	0.50
4	400.00	398.10	-1.90	-0.47	398.00	-2.00	-0.50
AVERAGE (%)				-0.50			-0.05



Calibrated By

(Mr. Jirawut Sakam)
Field Environmental Scientist (3)

Approved By

(Mr. Sarayuth Jitranont)
Assistant General Manager


ALS Laboratory Group
FORM NO. F 06-056 REVISION NO. - ISSUE DATE: 02/04/12



MULTIPOINT CALIBRATION REPORT

Calibration Date	4-Jan-22	Equipment Name	SO2 Analyzer
Manufacturer	Teledyne API	Model	T100
Serial No.	8061	Equipment ID	RYG_FS0534
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	847		
Std. Gas Concentration (PPM)	50.87	Cylinder No.	LL36633
Cylinder Pressure (psi)	1200	Certified By	Airgas Inc.
Certified Date	18-Mar-14	Expired Date	18-Mar-22

Point	CALIBRATION RESULTS			
	Ideal	Actual	Error	%Error
ZERO	0.00	0.10	0.10	0.10
1	100.00	98.80	-1.20	-1.20
2	200.00	198.70	-1.30	-0.65
3	300.00	298.30	-1.70	-0.57
4	400.00	397.30	-2.70	-0.67
AVERAGE (%)				-0.60



Calibrated By

(Mr. Jirawut Sakam)
Field Environmental Scientist (3)

Approved By

(Mr. Sarayuth Jitranont)
Assistant General Manager


ALS Laboratory Group
FORM NO. F 06-056 REVISION NO. - ISSUE DATE: 02/04/12



MULTIPOINT CALIBRATION REPORT

Calibration Date	4-Jan-22	Equipment Name	SO2 Analyzer
Manufacturer	Teledyne API	Model	T100
Serial No.	8060	Equipment ID	RYG_FS0532
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	847		
Std. Gas Concentration (PPM)	50.87	Cylinder No.	LL36633
Cylinder Pressure (psi)	1200	Certified By	Airgas Inc.
Certified Date	18-Mar-14	Expired Date	18-Mar-22

Point	CALIBRATION RESULTS			
	Ideal	Actual	Error	%Error
ZERO	0.00	0.10	0.10	0.10
1	100.00	98.80	-1.20	-1.20
2	200.00	198.60	-1.40	-0.70
3	300.00	298.30	-1.70	-0.57
4	400.00	397.60	-2.40	-0.60
AVERAGE (%)				-0.59



Calibrated By

(Mr. Jirawut Sakam)
Field Environmental Scientist (3)

Approved By

(Mr. Sarayuth Jitranont)
Assistant General Manager


ALS Laboratory Group
FORM NO. F 06-056 REVISION NO. - ISSUE DATE: 02/04/12



MULTIPOINT CALIBRATION REPORT

Calibration Date	4-Jan-22	Equipment Name	SO2 Analyzer
Manufacturer	HORIBA	Model	APSA-370
Serial No.	90U0XJ31	Equipment ID	RYG_FS0452
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	50.87	Cylinder No.	LL36833
Cylinder Pressure (psi)	1200	Certified By	Algas Inc.
Certified Date	18-Mar-14	Expired Date	18-Mar-22

Point	CALIBRATION RESULTS			
	Ideal	Actual	Error	%Error
ZERO	0.00	0.10	0.10	0.10
1	100.00	98.60	-1.40	-1.40
2	200.00	198.00	-2.00	-1.00
3	300.00	298.10	-1.90	-0.63
4	400.00	398.20	-1.80	-0.45
AVERAGE (%)				-0.68



Calibrated By

(Mr.Jirawut Sakam)
Field Environmental Scientist (3)

Approved By

(Mr.Sarayuth Jitranont)
Assistant General Manager


ALS Laboratory Group
FORM NO. F 06-056 REVISION NO. - ISSUE DATE 02/04/12



MULTIPOINT CALIBRATION REPORT

Calibration Date	4-Jan-22	Equipment Name	SO2 Analyzer
Manufacturer	HORIBA	Model	APSA-370
Serial No.	YPRXJ20	Equipment ID	RYG_FS0263
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	50.87	Cylinder No.	LL36833
Cylinder Pressure (psi)	1200	Certified By	Algas Inc.
Certified Date	18-Mar-14	Expired Date	18-Mar-22

Point	CALIBRATION RESULTS			
	Ideal	Actual	Error	%Error
ZERO	0.00	0.10	0.10	0.10
1	100.00	99.80	-0.20	-0.20
2	200.00	199.40	-0.60	-0.30
3	300.00	298.20	-1.80	-0.60
4	400.00	398.00	-2.00	-0.50
AVERAGE (%)				-0.30



Calibrated By

(Mr.Jirawut Sakam)
Field Environmental Scientist (3)

Approved By

(Mr.Sarayuth Jitranont)
Assistant General Manager

ALS Laboratory Group
FORM NO. F 06-056 REVISION NO. - ISSUE DATE 02/04/12



63/14-15,67/35-36, Soi Petchkasem 7,7/1, Petchkasem Rd,
Wathapra, Bangkokyai,Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranatee.com

CERTIFICATE OF CALIBRATION

Certificate No. WS 04012001
Page 1 of 2 pages

Measurement Item	Cup anemometer with data logger		
Manufacturer	Data logger: Novaya Cup anemometer: Novaya		
Model/Type	Data logger: P00 WS 2515 Cup anemometer: WS 0201		
Serial Number	Data logger: A8861 Cup anemometer:		
ID No.	Data logger: P00 WS 2515 Cup anemometer:		
Customer	ALS Laboratory Group (Thailand) Co., Ltd. 104 Phramasarak 40, Phramasarak Rd, Klongkum, Bangkok 10250, Thailand		
Test Conditions	Wind tunnel cross test section area	900	m ²
	Anemometer frontal area	100	m ²
	Exposure of mounting pipe	1	m
	Roofage ratio of test object	0.111	1
Test Conditions	Air temperature	24.0	±0.8 °C
	Air pressure	1005.9	±0.4 hPa
	Relative air humidity	63.3	±0.8 %RH
Calibration Procedure	Calibrator was carried out base on: ISO 61420-12-1 (C1), 2008 Power Performance Measurements on Electricity Producing Wind Turbines MAGNET Anemometer Calibration Procedure - Version 2 2009		
Traceability	This calibration documents the traceability to national standard, which makes the unit of measurements according to the international system of units (SI) through National Institute of Metrology (Thailand (NMI))		
Measurement Date	1 Jul 13, 2021		
Issued Date	1 Jul 14, 2021		
Calibrated by	<input checked="" type="checkbox"/> Mr. Borak Thachad <input type="checkbox"/> Miss Chutai Wianthayee		
Approved Signature	 Mr. Panyai Boonlarnorn Technical Support and Calibration Manager		



THIS CERTIFICATE MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.



63/14-15,67/35-36, Soi Petchkasem 7,7/1, Petchkasem Rd,
Wathapra, Bangkokyai,Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranatee.com

Continuation of Certificate of Calibration Number

Certificate No. WS 04012001
Page 2 of 2 Pages

Result of calibration: ☒ With adjustment ☐ With adjustment
 Calibration in the range of 1 - 10 m/s at a calibration interval of 1 m/s.
 The results of calibration and associated measurement uncertainties are reported in the table below.

V _{ref} Reading m/s	V _{ref} Reading m/s	Error m/s	Uncertainty (%)
2.584	1.8	0.8	2.7
4.112	4.0	0.1	1.4
6.50	6.0	0.5	7.7
8.57	8.1	0.4	4.7
10.02	10.1	-0.1	0.70
11.98	12.3	-0.3	2.63
13.98	14.2	-0.2	1.49
16.02	16.5	-0.5	3.13
18.02	18.4	-0.4	2.22
19.99	19.3	0.6	3.05
11.02	11.1	-0.1	0.66
6.02	6.1	-0.1	1.66
7.02	7.1	-0.1	1.43
8.17	8.0	0.2	2.45
3.007	3.0	0.0	1.7
1.023	1.0	0.0	5.4

MUC* (Uncertainty)

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%.

Appendix 1: Instrumentation

NO	Sensor	Manufacturer	Model/Type	Calibration Date	Certificate Report Number	Range
1	Flow static	TSSTO INC.	0552143	July 16, 2020	MS 0035 20	5 - 30 m/s
2	Pressure Differential Pressure Meter	Zigzag	DPM2500	July 16, 2020	MS 0035 20	5 - 30 m/s
3	Air velocity measurement Pitot-static	TS INC.	8455 12	July 20, 2020	MS 0035 20	0 - 8 m/s
4	Temperature	Zigzag	DS17B01	March 30, 2021	SC 0271 04	32 - 70 °C
5	Relative humidity	Zigzag	DSH17H1	March 30, 2021	SC 0271 04	0 - 100 %RH
6	Atmospheric pressure	Zigzag	DSH17H1	March 30, 2021	SC 0271 04	500 - 1100 mPa
7	Wind turbine	QSGW	MP3300			0 - 50 m/s

End of certificate of calibration



CERTIFICATE OF CALIBRATION

Certificate No. WD-04072021
Page 1 of 2 pages

Measurement Item : Wind direction sensor with data logger.
Manufacturer : Data logger: Novolynx.
Wind direction sensor: Novolynx.
Model/Type : Data logger: 200 WS-25DL.
Wind direction sensor: WS-02F.
Serial Number : Data logger: A4987.
Wind direction sensor: .
ID No : Data logger: RVD_F90089.
Wind direction sensor: .
Customer : A.S. Laboratory Group (Thailand) Co., Ltd.
104 Phatthanakan 40, Phatthanakan Rd, Muang Suan Luang, Khet Suan Luang, Bangkok 10250
Thailand

Environmental Condition
The measurement was carried out in an ambient temperature of (23±3)°C and relative humidity of (60±10)%.

Measurement Method
The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and the laser is used for axis control. The measurement were taken at 45° intervals in clockwise and counterclockwise direction.

Note: The UUC was warmed up for 1 hour prior to the calibration being performed.

Traceability
The measurement results are traceable to the international system of units (SI) through Certificate No. EC545-D7-0045.
Certificate No. K086370044.

Measurement Date : 12/14/2021
Issued Date : 12/14/2021



Approved Signatory:

[Signature]
Ms. Pichya Booncharoen
Technical Support
and Calibration Manager

Performed by
☒ Mr. Somrat Thachad
☐ Miss Oranai Wathwattaya

THIS CERTIFICATE REPORT MAY NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT PERMISSION FOR REPRODUCTION HAS BEEN
OBTAINED BY WRITING FROM THE LABORATORY.

Continuation of Certificate of Calibration Number:

Certificate No. WD-04072021
Page 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment.
Calibration in the range of 0 - 360 ° at a calibration interval of 45°.
The results of calibration and associated measurement uncertainties are reported in table below:

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC* Reading (°)	Error (°)	Uncertainty ±(°)
1	Clockwise	0/360	0	0	0	3.0
2		45	45	42	-3	3.0
3		90	90	88	-2	3.0
4		135	135	133	-2	3.0
5		180	180	181	1	3.0
6		225	225	228	3	3.0
7		270	270	273	3	3.0
8	Counter Clockwise	315	315	318	3	3.0
9		0/360	0	0	0	3.0
10		45	45	42	-3	3.0
11		90	90	88	-2	3.0
12		135	135	133	-2	3.0
13		180	180	181	1	3.0
14		225	225	228	3	3.0
15		270	270	273	3	3.0
16		315	315	318	3	3.0

UUC*: Unit Under Calibration. The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor for providing a level of confidence of approximately 95%.

End of Certificate of Calibration



SITHIPORN ASSOCIATES CO.,LTD. CALIBRATION LABORATORY

451-451/1 Sirinthorn Rd, Bangbunru, Bangplud Bangkok 10700 THAILAND
Tel: 0-2435-8800 Fax: 0-2433-1679 e-mail: cal-center@sithiporn.com http://www.sithiporn.com



Cert. No. : ACC21009
Pages : 1 of 3

Calibration Certificate

Equipment : SOUND CALIBRATOR
Manufacturer : RION
Model : NC-74
Serial No. : 34178123
ID No. : RYG_FS0215

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location :
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 05 AUGUST 2021
Calibration Date : 09 AUGUST 2021
Date of Issue : 11 AUGUST 2021

Calibrated by : Nathakorn Pistupaisan

Approved by :

[Signature]
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced
other than in full, except with the prior written approval of the head of Calibration Laboratory.

SITHIPORN ASSOCIATES CO.,LTD. CALIBRATION LABORATORY

Continuation of Calibration Certificate

Cert. No. : ACC21009
Job No. : VC64AC0058
Pages : 2 of 3

Calibration Procedure : CP-AC-03

Calibration Method :

This equipment was calibrated by based on IEC-60942:2003 Standard.

The sound pressure level, frequency and total distortion of the sound calibrator was measured using the reference microphone.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EELBP_05/0264	10-Feb-22
Digital Multimeter	8846A	1997025	EELBP_06/0264	05-Feb-22
Digital Multimeter	33461A	MY53220116	EELBP_04/0264	10-Feb-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KA1	34560495	AA-3003-21	16-Feb-22
Audio Analyzer	AVR-3360A	V744B6069	EF-0010-21	10-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACC21009
Job No. : VC64AC0058
Pages : 3 of 3

Result of calibration :

1. Sound pressure level

Specified sound pressure level (dB)	Measured value (dB)	Deviated value (dB)	Uncertainty (dB)	Tolerance limit (dB)
94	94.06	0.06	0.14	0.40

2. Frequency

Specified Frequency (Hz)	Measured value (Hz)	Deviated value (%)	Uncertainty (%)	Tolerance limit (%)
1000	1001.5	0.1	0.1	1.0

3. Total distortion

Measured value (%)	Uncertainty (%)	Tolerance limit (%)
1.67	0.10	3.0

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate


QF-TS12-04-04-020664

451-451/1 Sirinthorn Rd, Bangbunmu, Bangplud Bangkok 10700 THAILAND
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:center@sithiphorn.com http://www.sithiphorn.comCert. No. : ACL22057
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00296517 / 179120 / 87527
ID No.: RYG_FS0434

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.Location :
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %Received Date : 14 JANUARY 2022
Calibration Date : 21-24 JANUARY 2022
Date of Issue : 25 JANUARY 2022

Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22057
Job No. : VC65AC0043
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.
For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	34461A	MY60024273	1-15180725251-1	15-Sep-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	29779900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KAI	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22057
Job No. : VC65AC0043
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long-term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22057
Job No. : VC65AC0043
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
15.4

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	11.6
C - weight	17.7
Flat	23.4

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.1	0.1	0.1	± 1.5
1000	0.0	0.0	0.0	± 1.0
8000	0.6	0.7	0.7	±5.0

QF-TS12-04-04-020664

T. Rth...

Continuation of Calibration Certificate

Cert. No. : ACL22057
Job No. : VC65AC0043
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	0.0	-0.1	0.0	±2.0
125	0.0	0.1	0.0	±1.5
250	0.0	0.0	0.0	±1.5
500	0.0	0.1	0.0	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.1	0.0	±2.0
4000	0.0	0.1	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SI M Display at initial (dB)	SI M Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.1	0.1	± 0.3

QF-TS12-04-04-020664

T. Rth...

Continuation of Calibration Certificate

Cert. No. : ACL22057
Job No. : VC65AC0043
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.1	0.1	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.1	0.1	± 1.1
26.0	26.1	0.1	± 1.1
25.0	25.1	0.1	± 1.1

QF-TS12-04-04-020664

T. Rth...

Continuation of Calibration Certificate

Cert. No. : ACL22057
Job No. : VC65AC0043
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	108.0	0.0	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.1	0.1	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
SEL	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.1	0.1	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.3	-0.1	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

QF-TS12-04-04-020664

T. Rth...

Continuation of Calibration Certificate

Cert. No. : ACL22057
Job No. : VC65AC0043
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.6	89.6	0.0	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation providing a level of confidence of approximately 95 %

End of Calibration Certificate

QF-TS12-04-04-020664

451-451/1 Sirinthon Rd., Bangbunru, Bangkok Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.comCert. No. : ACL21119
Pages : 1 of 8

Calibration Certificate


Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No. : 01122579 / 172172 / 74022
ID No. : RYG FS0018

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 21 SEPTEMBER 2021
Calibration Date : 04-06 OCTOBER 2021
Date of Issue : 11 OCTOBER 2021



Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL21119
Job No. : VC64AC0070
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY53202742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL-BP_03/0264	08-Feb-22
Digital Multimeter	33461A	MY53220076	EEL-BP_06/0264	05-Feb-22
Digital Multimeter	8846A	1997025	EEL-BP_06/0264	05-Feb-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KAI	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL21119
Job No. : VC64AC0070
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.3	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long-term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL21119
Job No. : VC64AC0070
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.8

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	11.6
C - weight	19.5
Flat	26.2

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.2	0.3	0.2	± 1.5
1000	0.0	-0.1	0.0	± 1.0
8000	-1.0	-0.9	-0.9	±5.0

QF-TS12-04-04-020664

T. Petch...

Continuation of Calibration Certificate

Cert. No. : ACL21119
Job No. : VC64AC0070
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	0.0	-0.1	±2.0
125	0.0	0.0	-0.1	±1.5
250	0.0	0.0	-0.1	±1.5
500	0.0	0.0	-0.1	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Lcq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

T. Petch...

Continuation of Calibration Certificate

Cert. No. : ACL21119
Job No. : VC64AC0070
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	132.9	-0.1	± 1.1
132.0	131.9	-0.1	± 1.1
131.0	130.9	-0.1	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.1	0.1	± 1.1
26.0	26.1	0.1	± 1.1
25.0	25.1	0.1	± 1.1

QF-TS12-04-04-020664

T. Petch...

Continuation of Calibration Certificate

Cert. No. : ACL21119
Job No. : VC64AC0070
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.1	-0.3	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

QF-TS12-04-04-020664

T. Petch...

Continuation of Calibration Certificate

Cert. No. : ACL21119
Job No. : VC64AC0070
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated	Acceptance
Positive	Negative	Value	Limits
one-half cycle	one-half cycle	(dB)	(dB)
89.5	89.5	0.0	±1.5

12. High level stability

Frequency	SLM Display	SLM Display	Deviated	Acceptance
Weighting	at initial	at final	Value	Limits
	(dB)	(dB)	(dB)	(dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$
or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

QF-TS12-04-04-020664

451-451/1 Srinthorn Rd, Bangbunru, Bangplud Bangkok 10700 THAILAND
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.comCert. No. : ACL22058
Pages : 1 of 8

Calibration Certificate


Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00296518 / 179118 / 87525
ID No.: RYG FS0431

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHUANG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location :
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 14 JANUARY 2022
Calibration Date : 21-24 JANUARY 2022
Date of Issue : 25 JANUARY 2022



Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced
other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22058
Job No. : VC65AC0043
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference
Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY 48017076	EE-0012-21	10-Feb-22
Waveform Generator	33511B	MY 52302742	EE-0011-21	10-Feb-22
Digital Multimeter	33461A	MY 53220104	EE-LBP-05-0264	10-Feb-22
Digital Multimeter	33461A	MY 53220076	EE-LBP-03-0264	08-Feb-22
Digital Multimeter	34461A	MY 60024273	1-15180725251-1	15-Sep-22
Programmable Attenuator	MA1-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KA1	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22058
Job No. : VC65AC0043
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long-term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22058
Job No. : VC65AC0043
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.6

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	11.6
C - weight	17.6
Flat	23.2

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.1	0.1	0.1	± 1.5
1000	0.0	0.0	0.0	± 1.0
8000	0.8	0.9	0.9	±5.0

QI-TS12-04-04-020664

P. 1

Continuation of Calibration Certificate

Cert. No. : ACL22058
Job No. : VC65AC0043
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	-0.1	0.0	±2.0
125	0.0	0.0	-0.1	±1.5
250	-0.1	0.0	-0.1	±1.5
500	0.0	0.0	-0.1	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.0	0.0	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Lcq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SIM Display at initial (dB)	SIM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QI-TS12-04-04-020664

P. 1

Continuation of Calibration Certificate

Cert. No. : ACL22058
Job No. : VC65AC0043
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.1	0.1	± 1.1
135.0	135.1	0.1	± 1.1
134.0	134.1	0.1	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.1	0.1	± 1.1
114.0	114.1	0.1	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.1	0.1	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.1	0.1	± 1.1
26.0	26.1	0.1	± 1.1
25.0	25.1	0.1	± 1.1

QI-TS12-04-04-020664

P. 1

Continuation of Calibration Certificate

Cert. No. : ACL22058
Job No. : VC65AC0043
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	116.9	-0.1	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
SEL	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, Lepeak (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	135.8	-0.6	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

QI-TS12-04-04-020664

P. 1

Continuation of Calibration Certificate

Cert. No. : ACL22058
Job No. : VC65AC0043
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.7	89.7	0.0	±1.5

12. High level stability

Frequency	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Weighting				
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

QF-TS12-04-04-020664

451-451/1 Sirdinthern Rd., Bangbunru, Bangplud Bangkok 10700 THAILAND
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.comCert. No. : ACL22031
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 01222716 / 143832 / 22763
ID No.: RYG_FS0020

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 05 JANUARY 2022
Calibration Date : 10-12 JANUARY 2022
Date of Issue : 13 JANUARY 2022

REVIEW BY	<i>Thakorn P.</i>
APPROVED BY	<i>T. Petchurai</i>
NEXT CAL DATE	10/1/23

Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22031
Job No. : VC65AC0040
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL_BP_05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL_BP_03/0264	08-Feb-22
Digital Multimeter	34461A	MY60024273	1-15180725251-1	15-Sep-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KA1	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22031
Job No. : VC65AC0040
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22031
Job No. : VC65AC0040
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.6

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	12.6
C - weight	19.2
Flat	24.6

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.7	0.7	0.7	± 1.5
1000	0.0	0.0	0.0	± 1.0
8000	-2.1	-2.0	-2.0	±5.0

QF-TS12-04-04-020664

T P.T.A.

Continuation of Calibration Certificate

Cert. No. : ACL22031
Job No. : VC65AC0040
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	0.0	0.0	-0.1	±2.0
125	0.0	0.0	0.0	±1.5
250	0.0	0.0	0.0	±1.5
500	0.0	0.1	0.0	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.1	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	±0.2
Flat	94.0	0.0	±0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	±0.1
Leq	94.0	0.0	±0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	±0.3

QF-TS12-04-04-020664

T P.T.A.

Continuation of Calibration Certificate

Cert. No. : ACL22031
Job No. : VC65AC0040
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	±1.1
136.0	136.0	0.0	±1.1
135.0	135.0	0.0	±1.1
134.0	134.0	0.0	±1.1
133.0	133.0	0.0	±1.1
132.0	132.0	0.0	±1.1
131.0	131.0	0.0	±1.1
129.0	129.0	0.0	±1.1
124.0	124.0	0.0	±1.1
119.0	119.0	0.0	±1.1
114.0	114.0	0.0	±1.1
109.0	109.0	0.0	±1.1
104.0	104.0	0.0	±1.1
99.0	99.0	0.0	±1.1
94.0	94.0	0.0	±1.1
89.0	89.0	0.0	±1.1
84.0	84.0	0.0	±1.1
79.0	79.0	0.0	±1.1
74.0	74.0	0.0	±1.1
69.0	69.0	0.0	±1.1
64.0	64.0	0.0	±1.1
59.0	59.0	0.0	±1.1
54.0	54.0	0.0	±1.1
49.0	49.0	0.0	±1.1
44.0	44.0	0.0	±1.1
39.0	38.9	-0.1	±1.1
34.0	34.0	0.0	±1.1
30.0	29.9	-0.1	±1.1
29.0	28.9	-0.1	±1.1
28.0	27.9	-0.1	±1.1
27.0	26.9	-0.1	±1.1
26.0	25.9	-0.1	±1.1
25.0	24.8	-0.2	±1.1

QF-TS12-04-04-020664

T P.T.A.

Continuation of Calibration Certificate

Cert. No. : ACL22031
Job No. : VC65AC0040
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.1	0.1	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
SEL	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.1	0.1	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.3	-0.1	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.1	0.1	-
Positive half cycle	135.4	135.2	-0.2	±2.0
Negative half cycle	135.4	135.2	-0.2	±2.0

QF-TS12-04-04-020664

T P.T.A.

Continuation of Calibration Certificate

Cert. No. : ACL22031
Job No. : VC65AC0040
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.6	89.6	0.0	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3


The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

QF-TS12-04-04-020664



Certificate of Calibration

Equipment: SPECTROPHOTOMETER
Model: DR6000
Serial No. (or ID.): 1627845 (RYG_EN0037)
Manufacturer: HACH
Condition: In ConditionCertificate No.: C06210159
Issued Date: 01 April 2021
Job No.: KSPR2104738
Page: 1 of 3Customer: ALS Laboratory Group (Thailand) Co., Ltd. (Rayong Branch)
616/10 Moo 5 T.Maenam Khu,
A.Pluakdaeng, Rayong 21140, Thailand.Environment Condition: Temperature 25.1 °C ± 0.4 °C
Humidity 48.8 %RH ± 3.7 %RHCalibration Place: ALS Laboratory Group (Thailand) Co., Ltd. (Rayong Branch) (Wet Chemistry Lab)
616/10 Moo 5 T.Maenam Khu,
A.Pluakdaeng, Rayong 21140, Thailand.

Calibration By: Mr. Chattaphon Foithong

Calibration Date: 01 April 2021

The Method used: In house method, SPCC-WI-24, base on ASTM E 275-08 and ASTM E 387-04

Traceability: This certificate is traceable to the CRM maintained by National Institute of Standards and Technology (NIST) through Starna Scientific Limited.

The standard for Wavelength Certificate No. 87146 and 87152

The standard for Photometric Certificate No. 87220 and 87139

The standard for Stray light Certificate No. 87163 and 87161

The standard for Spectral resolution Certificate No. 87173

(Mr. Chattaphon Foithong)
SPC RT Co., Ltd.
(Mr. Dumrong Boonsopon)
Authorized signatory

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.
The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor ($k=2$) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).
These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of SPC RT Co., Ltd.

SPC Calibration Center
1134 Soi Woonwongmaethit 51 Sukhumvit 101 Road Bangkok Phrakhanong Bangkok 10260 Thailand
Tel: 02-0003 1134 Fax: 02-0003 1134

Certificate No. : C06210159 Page 2 of 3

Calibration Results:
Without Adjustment

Wavelength Accuracy (nm), The spectral bandwidth of Std at 2 nm and UUC at 2 nm

Standard Wavelength	Unit Under Calibration	Correction	Uncertainty
418.61	418.4	0.21	0.13
536.66	536.7	-0.04	0.13
637.98	638.3	-0.32	0.14
748.48	748.7	-0.22	0.14
807.03	807.4	-0.37	0.14

Photometric Accuracy (Absorbance)

Wavelength	Standard absorbance	Unit Under Calibration	Correction	Uncertainty
420 nm	0.0000	0.000	0.0000	0.0045
	0.5890	0.590	-0.0010	0.0045
	0.7616	0.762	-0.0004	0.0045
	1.0263	1.027	-0.0007	0.0045
440 nm	0.0000	0.000	0.0000	0.0045
	0.5787	0.579	-0.0003	0.0045
	0.7442	0.744	0.0002	0.0045
	1.0039	1.004	-0.0001	0.0045
465 nm	0.0000	0.000	0.0000	0.0045
	0.5292	0.530	-0.0008	0.0045
	0.6865	0.687	-0.0005	0.0045
	0.9534	0.954	-0.0006	0.0045
546.1 nm	0.0000	0.000	0.0000	0.0045
	0.5468	0.546	0.0008	0.0045
	0.6957	0.695	0.0007	0.0045
	0.9991	0.998	0.0011	0.0045
590 nm	0.0000	0.000	0.0000	0.0045
	0.5851	0.584	0.0011	0.0045
	0.7238	0.723	0.0006	0.0045
	1.0957	1.094	0.0017	0.0045
635 nm	0.0000	0.000	0.0000	0.0045
	0.5692	0.568	0.0012	0.0045
	0.6914	0.691	0.0004	0.0045
	1.0881	1.087	0.0011	0.0045

SPC Calibration Center
1134 Soi Woonwongmaethit 51 Sukhumvit 101 Road Bangkok Phrakhanong Bangkok 10260 Thailand
Tel: 02-0003 1134 Fax: 02-0003 1134

Certificate No. : C06210159 Page 3 of 3

Calibration Results:
Without Adjustment

Photometric Accuracy (Absorbance)

Wavelength	Standard absorbance	Unit Under Calibration	Correction	Uncertainty
235 nm	0.0000	0.000	0.0000	0.0080
	0.7307	0.730	0.0007	0.0080
257 nm	0.0000	0.000	0.0000	0.0080
	0.8516	0.850	0.0016	0.0080
313 nm	0.0000	0.000	0.0000	0.0080
	0.2836	0.285	-0.0014	0.0080
350 nm	0.0000	0.000	0.0000	0.0080
	0.6319	0.629	0.0029	0.0080

Stray light *

Standard: cut-off	UUC: Wavelength (nm)	UUC: Transmission (%)	Absorbance (A)
260.57 ± 0.11 nm	260.6	1.5	1.824
392.03 ± 0.11 nm	392.0	1.5	1.824

The stray light transmission reference is less than 1.0 T(%) and absorbance is greater than 2.0 (A)

Spectral Resolution *

Nominal Concentration 0.02 % v/v	Peak	Trough	Ratio	SBW
Standard Wavelength (nm)	268.72	266.76	1.39	2.00
UUC: Wavelength (nm)	268.2	266.1		
Std Absorbance (A)	0.4616	0.2797		
Absorbance (A)	0.416	0.300		

* Calibration Marked * Not TISI Accredited * in this Certificate have been included for completeness.

The End of Certificate

SPC Calibration Center
1134 Soi Woonwongmaethit 51 Sukhumvit 101 Road Bangkok Phrakhanong Bangkok 10260 Thailand
Tel: 02-0003 1134 Fax: 02-0003 1134



Certificate of Calibration

Certificate No.: 22E986
Page: 1 of 2

Equipment: pH Meter
Manufacturer: Mettler Toledo
Model: SevenCompact S220
Serial No.: C104059460
ID No.: RYG_EN0183

This certificate may not be reproduced other than in full,
except with the prior written approval of the head of
Corporate Services 3: Equipment Calibration and Testing Services.

Condition As-Received: Used Item
Received Date: 16 March 2022
Calibration Date: 21 March 2022

Reference: 2203-0511DSC
Ambient Temperature: (23 ± 2) °C
Relative Humidity: (50 ± 10) %
Submitted by: ALS Laboratory Group (Thailand) Co., Ltd. Rayong Branch
616/10 Moo 5 T. Maenam Khu, A. Pluakdaeng, Rayong
21140, Thailand

Procedure used: Calibration was conducted using in-house calibration Procedure CP-E17 According to direct measurement method with Multi-Product Calibrator.

Condition of this result of calibration

1. Reference standards instruments:


Instrument	Model	Serial No.	Certificate No.	Due Date
1) Multi-Product Calibrator	5500A	6440007	21E1444	07 May 2022

2. This result of calibration was made on requested at the point specified by customer.

3. The certificate is valid only to the item calibrated on date and place of calibration.

4. This Certification is traceable to the International System of Unit maintained at:-

-National Institute of Metrology Thailand (NIMT)



Calibrated by: Pongsagorn Boonyaporn
Issue Date: 22 March 2022

Approved Signatory:

[✓] Phalinee Prabpaijal
[] Nuntawat Khamchai
[] Ponthipha Tameyakul

B 0284414



Cert. No.: 22E986
Page.: 2 of 2

Result of calibration: () Without adjustment () After adjustment

Function:	DC voltage measurement	Range:	2000 mV	
	Standard Value	UUC Reading	Error	Uncertainty
	(mV)	(mV)	(mV)	(± μV)
	-200.0000	-200.0	0.0	72
	-150.0000	-150.0	0.0	69
	-100.0000	-100.0	0.0	65
	-50.0000	-50.0	0.0	62
	0.0000	0.0	0.0	58
	50.0000	50.0	0.0	62
	100.0000	100.0	0.0	65
	150.0000	150.0	0.0	69
	200.0000	200.0	0.0	72

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95 %

*UUC= Unit Under Calibration.

-o0o-


a 1101070



Cert.No.: 21TW20
Page.: 1 of 2

Certificate of Testing

Equipment: DO Meter
Manufacturer: YSI
Model: 5100
Serial No.: 15L102139
ID No.: RYG_EN0140
Received Date: 29 January 2021
Test Date: 02 February 2021
Reference: 2101-0817DSC-1
Submitted by: ALS Laboratory Group (Thailand) Co., Ltd.
Rayong Branch
Eastern Seaboard Industrial Estate (Rayong)
64/77 Moo 4, Building No.B1, Highway 331,
Km91.5, T. Pluakdaeng, A. Pluakdaeng,
Rayong 21140 Thailand



Laboratory Condition: Temperature (25 ± 5) °C
Humidity (50 ± 20) %
Test Procedure: In - house method : CP-CH9
by Comparison Technique with Azide Modification Method

Calibrated by: Walalak Sirithean

Approved by:
Approved Signatory

(/) Malee Butkruea
() Sathip Meangmai
() Warakorn Lernagatrakul

Issue Date: 3 February 2021

B 0252485



Cert.No.: 21TW20
Page.: 2 of 2

Result: Dissolved Oxygen Meter Adjustment With Air 100 %

Dissolved Oxygen Probe No.: 16C100647

Titration Method (Azide Modification Method)	DO Meter Reading	Standard Deviation
(mg/L)	(mg/L)	(mg/L)
8.02	8.02	0.0055

This report was certified only for the instrument we tested. It is allowable to use for study the system efficiency. The environmental impact control and present to organization it may concerned. Intend to use for advertising and referral purpose is prohibited. This report may not be reproduced other in full, without written approval of the laboratory

-o0o-

a 1038971



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES & EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL: 0-2717-3000-27 FAX: 0-2719-9484



Cert. No.: 21TM271
Page.: 1 of 2

Certificate of Calibration

Equipment : DO Meter with Sensor
Manufacturer : YSI
Model : 5100
Serial No. : 15L102139
ID No. : RYG_EN0140
Submitted by : ALS Laboratory Group (Thailand) Co., Ltd. Rayong Branch
Eastern Seaboard Industrial Estate (Rayong)
64/77 Moo 4 Building No.B1, Highway 331 km. 91.5,
T. Pluakdaeng, A. Pluakdaeng, Rayong 21140 Thailand
Location : TPA On Site Calibration Laboratory
Received Order : 29 January 2021
Calibrated Date : 3 February 2021
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %
AC Line Voltage : (220 ± 22) V

Calibrated by : Malee Butkruea

Approved by :
Approved Signatory
() Pornthippa Tameyakul
(✓) Suwit Imjai

Issue Date : 4 February 2021

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services & Equipment Calibration and Testing Services

A 0024028



Equipment : DO Meter with Sensor
Condition As-Received : Used Item
Reference : 2101-0817DSC-2

Cert. No.: 21TM271
Page.: 2 of 2

Procedure Used :-

Calibration were conducted using in-house calibration procedure CP-OT01 according to comparison with Industrial Platinum Resistance Thermometer (IPRT) into Temperature Bath.

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
Digital Thermometer	1523	2188080	201389	20 Nov 2021

2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit maintained at:-
- National Institute of Metrology Thailand (NIMT)

Result of Calibration :- (*) Without Adjustment

Function : Temperature measurement.

This instrument was connected with temperature sensor, S/N : 16C100647

Calibration Point (°C)	Immersion Depth (mm)	Standard Temperature (°C)	UUC* Reading (°C)	Error (°C)	Uncertainty (± °C)	Coverage Factor k
20.00	60	20.008	19.96	-0.048	0.15	2.00

UUC* : Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-00-

a 1038626



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES & EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL: 0-2717-3000-27 FAX: 0-2719-9484



Cert. No.: 22TM317
Page.: 1 of 3

Certificate of Calibration

Equipment : Low Temp. Incubator
Manufacturer : Memmert
Model : IPP750
Serial No. : V818.0084
ID No. : RYG_EN0154
Submitted by : ALS Laboratory Group (Thailand) Co., Ltd.
(Rayong Branch)
616/10 Moo 5 T. Maenam Khu.
A. Pluakdaeng, Rayong 21140, Thailand
Location : BOD Room
Received Order : 22 April 2022
Calibration Date : 22 April 2022
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %

Calibrated by : Man Pattanapongpaiboon

Approved by :
Approved Signatory
() Pornthippa Tameyakul
(✓) Malee Butkruea
() Suwit Imjai

Issue Date : 3 May 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services & Equipment Calibration and Testing Services

A 0040735



Equipment : Low Temp. Incubator
Condition As-Received : Used Item
Reference : 2204-0146OC-1

Cert. No.: 22TM317
Page.: 2 of 3

Procedure Used :-

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement
The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
Data Acquisition	34970A	MY44031769	21LM12	02 Sep 2022


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Close



Probe Installation Details : Dimension of Chamber :

a = 10 cm	D = 0.60 m
b = 10 cm	W = 1.0 m
c = 10 cm	H = 1.2 m
	Capacity = 0.75 m ³

Environment during calibration		
	Beginning	Finished
Temp. (°C)	25	25
REL. Humid. (%)	54	58
AC Supply (Volt)	221	223

Position :	Ref. Std. ID No.:
1	9RTD-2/1
2	9RTD-2/2
3	9RTD-2/3
4	9RTD-2/4
5	9RTD-2/5
6	9RTD-2/6
7	9RTD-2/7
8	9RTD-2/8
9 (ref.)	9RTD-2/9

a 1106485



Equipment : Low Temp. Incubator
Condition As-Received : Used Item
Reference : 2204-0146OC-1
Result of Calibration : (*) Without Adjustment
Function of UUC* : Temperature Source
Fresh air setting : Close

Cert. No.: 22TM317
Page.: 3 of 3

Calibration Point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Temperature stability (± °C)	Temperature uniformity (°C)	Overall Variation (°C)	Uncertainty (±°C)	Coverage Factor
20.0	20.0	20.0	0.022	0.20	0.22	0.30	2

Calibration Point (°C)	Measured Temperature (°C)								
	Position								
20.0	20.209	20.174	20.199	20.110	20.075	20.062	20.027	20.069	20.030

Average* : The average of 30 values in each position.
Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor.
Temperature uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location, which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.
Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation.
UUC* : Unit Under Calibration
Note : The reported uncertainty of measurement was included stability and excluded uniformity.
The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

-o0o-

a 1106484



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG, BANGKOK 10250
TEL. 0-2717-3000-34 FAX. 0-2719-9484



Certificate of Calibration

Certificate No.: 21T1200
Page : 1 of 2

Equipment : Digital Thermometer With Sensor
Manufacturer : Testo
Model : 106
Serial No.: 31281494/504
ID No.: RYG_FS0467

This certificate may not be reproduced other than in full, except with the prior written approval of the head of Corporate Services 3: Equipment Calibration and Testing Services.

Condition As-Received: Used Item
Received Date: 02 July 2021
Calibration Date: 07 July 2021
to 08 July 2021

Reference: 2107-0068DSC
Ambient Temperature: (25 ± 3) °C
Relative Humidity: (50 ± 20) %

Submitted by: ALS Laboratory Group (Thailand) Co.,Ltd. Rayong Branch
616/10 Moo 5 T.Maenam Khu, A.Pluaekdaeng, Rayong
21140, Thailand

Procedure used: Calibration were conducted using in-house calibration procedure CP-T01 according to comparison with Platinum Resistance Thermometer (PRT) into liquid bath temperature controller.
The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standards instruments :

Instrument	Model	Serial No.	Certificate No.	Due Date
1) Digital Thermometer	1529-R	B19520	21680	26 Jun 2022
2) Platinum Resistance Thermometer	935-14-95	261589/1	21680	26 Jun 2022

2. The certificate is valid only to the item calibrated on date and place of calibration.
3. This Certification is traceable to the International System of Unit maintained at:-
National Institute of Metrology Thailand (NIMT)

REVIEW BY	Thirawat
APPROVED BY	Sut S
NEXT CAL DATE	7/19/22

Calibrated by : Yossapon Poljorn
Issue Date : 09 July 2021

Approved Signatory :

[] Phalinee Prabpaipal
[] Chatchawan Khunpluek
[x] Wanlop Larpkum

B 0265214



Cert. No.: 21T1200
Page.: 2 of 2

Result of Calibration:- Without Adjustment
Function: Temperature measurement
Dimension of probe : Diameter 3 mm., Length 55 mm. Sheath material : Stainless Steel

Immersion Depth (mm.)	Standard Temperature (°C)	UUC* Reading (°C)	Error (°C)	Uncertainty of Measurement (±°C)
50	25.0029	24.9	-0.1029	0.12
50	30.0018	29.9	-0.1018	0.12
50	40.0035	40.0	-0.0035	0.12

UUC* : Unit Under Calibration
The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.

-o0o-

a 1063351



Metrological Center

SCI ECO Services Company Limited

33/2 Moo 3, T. Banpa, A. Kaengkhioi, Saraburi 18110, Thailand.
Saraburi Tel : +66 3627 3096 Fax : +66 3627 3100
Bangkok Tel : +668 9205 6851 , +669 8247 2360
Website : www.scieco.co.th E-Mail : calibrate@scg.co.th

RYG_EN0184



Certificate No. T220384101 "Substitute for Calibration Certificate Number T220384" Page 1 of 4

Certificate of Calibration

Equipment : Chamber (Cold Room)

Manufacturer : MODULAR

Model : IREVCOHCOO

Serial No. : C00351459

Customer Code : RYG_EN0184

ID No. : T1939A5

Customer : ALS Laboratory Group (Thailand) Co.,Ltd. (Rayong Branch)

616/10 Moo 5 T.Maenam Khu,

A.Pluaekdaeng, Rayong 21140

Customer Location : Laboratory

Date of Receipt : 18 February 2022

Calibrated By : Boonchai Suriyawong (Site Calibration Manager)

Approved By : / Sujjar Naknakred (Site Calibration Manager)

Date of Issue : 18 MAR 2022

REVIEW BY	Thirawat
APPROVED BY	D
NEXT CAL DATE	18 Feb 2023

The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standard laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Metrological Center.

FM-1.14 117-01-02-64

Certificate No. T220384101

Page 2 of 4

Calibration Report

Equipment : Chamber (Cold Room)
Date of Calibration : 22 February 2022
Environment : Temperature : 23.2-24.3 °C
Line Voltage : 221.8-227.2 V
Relative Humidity : 55 - 65 %RH

Condition of this results of calibration :

1. This equipment was calibrated by insert 16 standard thermocouples type T into its chamber . the other one standard thermocouples type T use for ambient temperature measurement . The calibration was done in according to WI-T20 (based on ASTM E145-94 (Reapproved 2001) and AS2853-1986) .
All data show below were final values and the initial data from customer request . The temperature scale used was based on ITS - 90 .

2. Reference Standard Instrument :

Instrument	Model	Instrument No.	Certificate No.	Due Date
TC	TYPE T	TN141-TN150	T210743	21 April 2022
TC	TYPE T	TN151-TN160	T210743	21 April 2022
DATA LOGGER	34970A	T150	T210743	21 April 2022

3. This certificate is traceable to :

National Institute of Metrology (Thailand) through Metrological Center (NSC-TIS-17525 CALIBRATION 0244)

4. Condition of calibrated item : good

Equipment Description :

Time Constant : - Hour 40 Minute At 3 °C
Fresh Air Damper : ☐ Open ☐ Min ☐ Medium ☐ Max
☐ Close
☒ Not Available

5. Adjustment :

(X) without adjustment () after adjustment


Approved By: _____

FM-L15-117-15-05-63

Certificate No. T220384101

Page 3 of 4

Calibration Report



C = Centre, F = Centre of Face, A = Corner, E = Centre of Edge

1C = TN141	12F = TN152
2A = TN142	13A = TN153
3A = TN143	14A = TN154
4F = TN144	15F = TN155
5A = TN145	16E = TN156
6A = TN146	
7F = TN147	
8F = TN148	
9A = TN149	
10A = TN150	
11F = TN151	

Approved By: _____

FM-L15-117-15-05-63

Certificate No. T220384101

Page 4 of 4

Calibration Report

Measurement Results

Calibration Point	Average Standard Reading at each position (°C)									
	TN141	TN142	TN143	TN144	TN145	TN146	TN147	TN148	TN149	TN150
3.0	2.80	2.96	2.98	2.97	3.16	3.29	2.95	3.14	3.10	3.45
	TN151	TN152	TN153	TN154	TN155	TN156				
	3.04	3.19	3.03	3.34	3.21	3.11				

Chamber (Cold Room)			Temperature Distribution				
Setting (°C)	Reading (°C)		Average (°C)	Stability (± °C)	Uniformity (°C)	Uncertainty (± °C)	Coverage Factor k
	Min	Max					
3.0	2.7	4.1	3.5	3.11	1.30	1.30	2.00

* The Acquired uncertainty exclude "uniformity"

The calibration result apply only the above calibrated item.

The result of test was found accurate as shown on date and place of test only.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k which for a t-distribution, providing a level of confidence of approximately 95 %.

Approved By: _____

FM-L15-117-15-05-63

Certificate of Calibration

Number of Page(s) 1 of 3

Certificate No. BSCC-UV-290/21
Equipment UV/Vis Spectrophotometer
Model UV-1800
Manufacturer Shimadzu
Serial No. A11454908533CD
ID No. BKK_EN0018
Date of receipt 15 October 2021
Date of calibration 15 October 2021
Date of issue 25 October 2021

Customer name ALS Laboratory Group (Thailand) Co., Ltd

Address 104 Soi Phatthanakan 40, Phatthanakan Road, Phatthanakan, Suan Luang, Bangkok 10250

Temperature (25.0 - 26.4) °C (On site)
Humidity (49.5 - 53.4) %RH (On site)

Equipment condition Good Operation

Calibration Location Organic Prep

Calibration Procedure In-house method WI-UV-702-01 based on ASTM E275-01

Traceability Wavelength Accuracy is traceable to certificate No. 87839 and 87844
Photometric Accuracy is traceable to certificate No. 87846 and 87877
Stray Light is traceable to certificate No. 87825
The above certificate are traceable to SI unit through Starna Scientific Ltd
(UKAS accredited calibration laboratory NO. 0659)

Calibrated by Mr. Wanchana Janloey

Approved by

Mr. Kanchit Choothep
Technical Manager

The above results are valid exclusively for the calibrated item(s) as mention in this report / certificate.
Advertising the report / Certificate and publicity of the results are prohibited and also shall not be reproduced
except in full, without written approval of the Bara Scientific Co., Ltd.

FM-UV-708-02 Rev 01 (23/01/83)

Certificate of Calibration

Certificate No. BSCC-UV-290/21 Number of Page(s) 2 of 3

Calibration Results:

1.Wavelength Accuracy

Certified Wavelength (nm)	UUC (nm)	Error (nm)	Uncertainty (±nm)
241.70	241.55	-0.15	0.18
334.02	333.80	-0.22	0.18
418.53	418.40	-0.13	0.18
572.99	572.85	-0.14	0.18
879.41	879.15	-0.26	0.18

2.Photometric Accuracy (UV)

Wavelength (nm)	Certified Absorbance (A)	UUC (A)	Error (A)	Uncertainty (±A)
235	0.0000	0.0000	0.0000	0.0075
	0.7174	0.7198	0.0024	0.0075
257	0.0000	-0.0001	-0.0001	0.0075
	0.8362	0.8377	0.0015	0.0075
313	0.0000	0.0000	0.0000	0.0075
	0.2778	0.2803	0.0025	0.0075
350	0.0000	-0.0001	-0.0001	0.0075
	0.6202	0.6221	0.0019	0.0075

*CNR = Customer not request

The above results are valid exclusively for the calibrated item(s) as mention in this report / certificate. Advertising the report / Certificate and publicity of the results are prohibited and also shall not be reproduced except in full without written approval of the Bara Scientific Co., Ltd.

FM-UV-708-02 Rev 01 (23/01/63)

Certificate of Calibration

Certificate No. BSCC-UV-290/21 Number of Page(s) 3 of 3

Calibration Results:

3.Photometric Accuracy (Visible)

Wavelength (nm)	Certified Absorbance (A)	UUC (A)	Error (A)	Uncertainty (±A)
420.0	0.0000	0.0000	0.0000	0.0042
	0.5631	0.5570	-0.0061	0.0042
	0.7390	0.7334	-0.0056	0.0042
	1.0863	1.0816	-0.0047	0.0042
440.0	0.0000	0.0000	0.0000	0.0042
	0.5524	0.5469	-0.0055	0.0042
	0.7217	0.7166	-0.0051	0.0042
	1.0606	1.0570	-0.0036	0.0042
455.0	0.0000	0.0000	0.0000	0.0042
	0.5018	0.4966	-0.0052	0.0042
	0.6657	0.6610	-0.0047	0.0042
	0.9775	0.9740	-0.0035	0.0042
546.1	0.0000	0.0000	0.0000	0.0042
	0.5147	0.5113	-0.0034	0.0042
	0.6743	0.6705	-0.0038	0.0042
	0.9909	0.9890	-0.0019	0.0042
590.0	0.0000	0.0000	0.0000	0.0042
	0.5427	0.5394	-0.0033	0.0042
	0.7037	0.7001	-0.0036	0.0042
	1.0338	1.0323	-0.0015	0.0042
635.0	0.0000	0.0000	0.0000	0.0042
	0.5268	0.5235	-0.0033	0.0042
	0.6720	0.6685	-0.0035	0.0042
	0.9864	0.9847	-0.0017	0.0042

*CNR = Customer not request

4.Stray Light*

Standard cut-off wavelength (nm)	Unit Under Calibration(UUC) Wavelength (nm)	Transmission (%T)	Absorbance (A)
200 91±0.11nm	200.31	0.9399	2.0274

The Stray light transmission reference is less than 1.0%T and Stray light absorbance reference is greater than 2.00A


*Stray Light not NSC-ONSC Accredited.

The measurement uncertainty is base on a standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

End of Certificate

The above results are valid exclusively for the calibrated item(s) as mention in this report / certificate. Advertising the report / Certificate and publicity of the results are prohibited and also shall not be reproduced except in full without written approval of the Bara Scientific Co., Ltd.

FM-UV-708-02 Rev 01 (23/01/63)



Agilent CrossLab Compliance

Qualification Type: ICPMS-OQ
System ID: JP15471169
EQP Name: AgilentRecommended
EQP Revision: ICPMS 02.50
EQP Publish Date: March 2020
Date: September 30, 2021 4:07:16 PM
Report Type: Report
Org. Name: ALS Laboratory Group (Thailand) Co.,Ltd.
Org. Location: 104 Phatnakarn 40, Suan Luang, Bangkok 10250.

REVIEW BY: *Suphan H.*
APPROVED BY: *Suphan H.*
NEXT CAL DATE: 29 March 2023

Table of Contents

Section	Page
Cover	1
Table of Contents	2
Test Summary	3
Service Details	4
Instrument Details	5
Calculation Formulas	7
Protocol Details	8
Tests	9
Autosampler Check : SPS4	9
Integrated Sample Introduction System (ISIS) Check : ISIS3	10
Autotune : G8403A	11
Background (No Gas Mode) : G8403A	13
Background (Gas Modes) : G8403A	14
20-Minute Stability (No Gas Mode) : G8403A	15
Declaration of Change Control	16
Attachments	17
Electronic Signature	31
Transaction Logs	32

Test Summary

Purpose

This section includes a status for each scheduled test and the overall qualification. For each test that is run, (1) the status is automatically determined based on pre-defined limits, and (2) the total number of times the test was run is displayed. For detailed results and specifications for a test, refer to the test results in this EQR.

Details

Test	Status	Runs
Autosampler Check : SPS4	Pass	1
Integrated Sample Introduction System (ISIS) Check : ISIS3	Pass	1
Autotune : G8403A	Pass	1
Background (No Gas Mode) : G8403A	Pass	1
Background (Gas Modes) : G8403A	Pass	1
20-Minute Stability (No Gas Mode) : G8403A	Pass	1

Overall Qualification Status

Pass

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 3 / 34

Service Details

Purpose

This section includes local contact and delivery details for this service.

General Details

Service Order No./Request: 6004837154
EQP Name: Agilent/Recommended
EQP Revision: ICPMS.02.50
Report Type: Report

Organization Details

Name: ALS Laboratory Group (Thailand) Co., Ltd.
Location: 104 Phattanakarn 40, Suan Luang, Bangkok 10250.

Local Contact Details

Name: Chatchanai Komarakul
Job Title: Manager
Qualification Location: Laboratory

Operator Details

Name: Panthep Kurasathain
Job Title: Field Service Engineer.

Data Acquisition Details

Acquisition Software Name: MassHunter
Acquisition Software Revision: C.01.04

Customer Data System (CDS): IcpMs: MassHunter

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 4 / 34

Instrument Details

Purpose

This section describes the as found system configuration.

Details

ICP-MS 1

Manufacturer: Agilent Technologies
Name: 7900
Model Number: G8403A
Installed Options: #100H: Standard Package with Hydrogen option
Detector Type: SQ
Nebulizer: Mira Mist (G3161)
Spray Chamber: Quartz
Torch: Quartz
Sampling Cone: Ni
Skimmer Cone: Ni
Serial Number: JP15471169
Firmware Revision: C.01.04

ISIS 1

Manufacturer: Agilent Technologies
Name: ISIS3
Model Number: G8411A
Type: Peristaltic pump system
Serial Number: JP15510227

Autosampler 1

Manufacturer: Agilent Technologies
Name: SPS4
Model Number: G8410A
Serial Number: AU18430722

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 5 / 34

Chiller 1

Manufacturer: Agilent Technologies
Name: Chiller
Model Number: G3292A
Serial Number: 3U1610713

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 6 / 34

Calculation Formulas

Purpose

This section includes calculation formulas for all available tests. Depending upon which tests are scheduled, all or some apply to your qualification.

For a description of calculations for ICP-MS tests performed by the MassHunter software, refer to the MassHunter application and documentation.

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 7 / 34

Protocol Details

Purpose

This section lists the revisions for all test units used in this report. For complete test-specific and high-level change details, refer to the Revision History document.

Test Revision	Test
ICPMS.02.50	20-Minute Stability (No Gas Mode)
ICPMS.02.50	Autosampler Check
ICPMS.02.50	Autotune
ICPMS.02.50	Background (Gas Modes)
ICPMS.02.50	Background (No Gas Mode)
ICPMS.02.50	Integrated Sample Introduction System (ISIS) Check

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 8 / 34

Autosampler Check

Purpose

This test demonstrates that the autosampler module is correctly installed and connected. It does not test module performance.

Setpoint

Results	Criteria	Observed Result	Expected Result	Status
---------	----------	-----------------	-----------------	--------

After the self test, is probe in the home position?

Yes	Yes	Pass
-----	-----	------

As commanded, is the probe positioned at vial 2?

Yes	Yes	Pass
-----	-----	------

Setpoint Status: Pass

Runs: 1

Overall Autosampler Check Test Status

Pass

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 9 / 34

Integrated Sample Introduction System (ISIS) Check

Purpose

This test demonstrates that the ISIS module is correctly installed and connected. It does not test module performance.

Setpoint

Results	Criteria	Observed Result	Expected Result	Status
---------	----------	-----------------	-----------------	--------

As commanded, does the pump rotate?

Yes	Yes	Pass
-----	-----	------

As commanded, do the valves load and inject?

Yes	Yes	Pass
-----	-----	------

Setpoint Status: Pass

Runs: 1

Overall Integrated Sample Introduction System (ISIS) Check Test Status

Pass

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 10 / 34

Autotune

Purpose

This test uses traceable checkout standards to run a software-executed autotune in all modes. The tune report provides values for peak width, mass axis, sensitivity, oxide species, and doubly-charged species tests.

Setpoint

Results

Peakwidth Mass 7

0.719	AMU
>=	0.65
<=	0.80

Status:

Pass

Peakwidth Mass 89

0.750	AMU
>=	0.65
<=	0.80

Status:

Pass

Peakwidth Mass 205

0.713	AMU
>=	0.65
<=	0.80

Status:

Pass

Mass Axis 7

7.05	AMU
>=	6.9
<=	7.1

Status:

Pass

Mass Axis 89

88.95	AMU
>=	88.9
<=	89.1

Status:

Pass

Mass Axis 205

205.00	AMU
>=	204.9
<=	205.1

Status:

Pass

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 11 / 34

Mass 7 Sensitivity No Gas

94.28	Mcps/ppm
>=	25.5

Status:

Pass

Mass 89 Sensitivity No Gas

307.16	Mcps/ppm
>=	127.5

Status:

Pass

Mass 205 Sensitivity No Gas

203.77	Mcps/ppm
>=	76.5

Status:

Pass

Mass 59 Sensitivity He

28.38	Mcps/ppm
>=	23.8

Status:

Pass

Mass 89 Sensitivity H2

129.27	Mcps/ppm
>=	68

Status:

Pass

Oxide Ratio 156/140

1.047	%
<=	1.38

Status:

Pass

Doubly Charged Species Ratio 70/140

1.482	%
<=	2.3

Status:

Pass

Setpoint Status:

Pass

Runs: 1

Overall Autotune Test Status

Pass

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 12 / 34

Background (No Gas Mode)

Purpose

This test examines the background of the ICP-MS in no gas mode by monitoring ions during a blank run.

Setpoint

Conditions

7	AMU
89	AMU
205	AMU

Measurements and Results

Masses (AMU):

Measured Value:

Agilent Recommended:

Status:

7	89	205
3.200	3.300	9.900
<=	<=	<=
6.9	4.6	11.5
Pass	Pass	Pass

Setpoint Status:

Pass

Runs: 1

Overall Background (No Gas Mode) Test Status

Pass

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 13 / 34

Background (Gas Mode)

Purpose

This test examines the background of the ICP-MS in the various gas modes by monitoring ions during a blank run.

Setpoint

Gas Mode: Helium

Conditions

78	AMU
1.0	sec
20	

Measurements and Results

Mass (AMU):

Measured Value:

Agilent Recommended:

Status:

78
42.8500
<=
115
Pass

Setpoint Status:

Pass

Runs: 1

Overall Background (Gas Mode) Test Status

Pass

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 14 / 34

20-Minute Stability (No Gas Mode)

Purpose

This test monitors the abundance of ions present in the checkout standard over a 20-minute period to verify that the signal is stable. The %RSD of the abundance of given ions is calculated internally by the software and compared to the limit.

Setpoint

Conditions

Mode: Spectrum
Masses: 7, 9, 59, 89, 140, 205
Integration Time: 9.99 sec
Peak Pattern: .3 points/peak
Repetitions: 20
Sweeps/Replicates: 100

Measurements and Results

Masses (AMU):
Stability RSD:
Agilent Recommended:
Status:

Masses (AMU)	Stability RSD	Agilent Recommended	Status
7	0.96490	≤ 2.3	Pass
89	0.51495	≤ 2.3	Pass
205	0.73011	≤ 2.3	Pass

Setpoint Status:

Pass

Runs: 1

Overall 20-Minute Stability (No Gas Mode) Test Status

Pass

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 15 / 34

Declaration of Change Control

This document is under change control. Revision history is maintained and printed on each document. Access to the master documents is limited to process owners. Documents receive periodic review and cannot be assigned an evergreen status. The qualification performed according to this document refers only to the hardware/software configuration in place at the time of the qualification. Agilent Technologies recommends that instrument configuration change management procedures be in place in order to maintain the validation process. Any changes to the analytical or computer hardware or software must be clearly specified. A change management system provides a means for determining the degree of requalification required according to the extent of the changes made. All details of the changes must be thoroughly recorded and documented, together with details of completed tests and their results. Note: Hardware/software configuration management is the customer's responsibility.

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 16 / 34

Attachments

Training requirements note: The delivery engineer attaches an ACE technique-specific training certificate to the Equipment Qualification Report (EQR). Obtaining ACE technique-specific certification includes pre-requisite trainings for Data Integrity, General Compliance topics (GMP, GLP, ALCOA, etc.), instrument hardware and software components, and the ACE technique itself. The one certificate encompasses all pre-requisite trainings as documented in the Agilent Learning Management System called Success Factors.


Location	Category	Document Name	Page
EQR	General	Certificate of System Qualification	18
EQR	General	Operator's training certificate and qualifications	19
EQR	General	Certificate of Qualification for ACE	20
EQR	General	Certificate of Qualification for ACE	21
EQR	General	Tune reports	22
EQR	General	Test Report	25
EQR	General	Test Report	27
EQR	General	Test Report	29

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 17 / 34

General

Document Name: Certificate of System Qualification

 Agilent Technologies

Agilent Compliance Engine Self Qualification

Date: September 16, 2021 4:59:15 PM
Drive Serial #: ACA029C9 Platform Revision: ACE 3.11

Individual self-qualification reports for each specific technique installed are also available upon request. They provide additional details on the general report from the console summary and are structured by the actual algorithms challenged during the process. There is not a one-to-one relationship between algorithms and OQ program tests because some algorithms are used by several tests and across multiple similar hardware components of the qualified systems.

Technique Type	Tests Completed	Result
Atomic Absorption	7	Conforms
Capillary Electrophoresis	10	Conforms
Dissemination	6	Conforms
Emission Spectroscopy	3	Conforms
Gas Chromatography - GC/MS	17	Conforms
Gas Chromatography	29	Conforms
Gas Permeation Chromatography	9	Conforms
ICP-MS	6	Conforms
Infrared Spectroscopy	7	Conforms
Liquid Chromatography	17	Conforms
Liquid Chromatography - LC/MS	6	Conforms
Microfluidics	18	Conforms
Sample Preparation - Gas Chromatography	9	Conforms
Sample Preparation - Liquid Chromatography	6	Conforms
Supercritical Fluid Chromatography	15	Conforms
Software	6	Conforms
UV-Vis Spectrophotometer	13	Conforms

Overall Qualification Status

Conforms

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 18 / 34

© 2021 by Agilent Technologies

Agilent CrossLab Compliance Services

Document Name: Certificate of Qualification for ACE



Certificate of Completion

Learner Name:	Pandey, Kuranthain
Title Of Course:	AN-CE-SS-II-030-A: ACE 3.X User Update Training
Completion Date	July 7, 2020
Certified By Company	Learner or Auditor

All Service and Support training certificates have the following specific limitations:

A certificate for Service and Support training is only valid while employed by Agilent Technologies or while working as an Agilent-authorized service provider, through which the service employee has ongoing access to Agilent's Safety Alerts, Service Notes, internal technical updates, update training, current documentation, technical support, current parts, and parts updates. Completion of training alone, without being employed by Agilent Technologies, does not qualify an individual to safely install, service or maintain Agilent equipment.

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 20 / 34

General

Document Name: Tune reports




Tune Report

Operator Name	Suzukiwari Mak
AcqData Batch	C:\Mglen\5C\PMW\1\1\UserTime_7950.b
Acq. Date-Time	2021-09-20 14:44:08
Report Comment	GQ 30 Sep 2021
Instrument Name	CS403A JP15471169

[No Gas]

Sensitivity



Mass	Range	Count	RSD%	Background
7	1900	1428	2.430	3.220
16	5000	2015	2.820	3.330
200	1900	20277	3.310	8.630

Sampling Period (s) 0.311
Isolation Time (s) 0.1

Oxidized/Doubly Charged Ratio	
Oxidized	156 / 143 1.047 %
Doubly Charged	70 / 140 1.482 %

Mass	Peak Height	Axis	W-50%	W-10%
7	9474.89	7.05	0.82	0.716
89	30716.43	88.55	0.89	0.750
203	20586.12	205.00	0.82	0.713

Integration Time [sec]	0.1
Acquisition Time [sec]	22.74
Y Axis	Linear


Time Parameters			
Plasma Parameters			
RFs Mode	—	Helium/Gas	1.00 Liter
RF Power	1000 W	Option Gas	—
RF Matching	1.10 V	Nitrogen Pump	6.10 psi
Sample Depth	9.0 mm	21C Target	2 °C
Lens Parameters			
Extract 1	0.0 V	Omega Lens	0.1 V
Extract 2	-200.0 V	Cell Entrance	-30 V
Omega Bias	-90 V	Cell Exit	-50 V
Gas Parameters			
Use Gas	No	3rd Gas Flow	—
Use Draw	0.0 g/draw	4th Gas Flow	-0.0 V
			Energy Decay

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 22 / 34

Document Name:

Tune reports




Date: September 30, 2021 4:07:18 PM
 System ID: JP15471169

Page 23 / 34

Document Name:

Tune reports



Date: September 30, 2021 4:07:18 PM
 System ID: JP15471169

Page 24 / 34

General

Document Name:

Test Report

Batch Summary Report

Batch Folder: C:\Batch\321\85 He 31
 Analysis File: 85 He Batch bin
 Tune Step: #1 He

Ref	Acq Date/Time	Data File	Sample Name	Type	Level	Calibration
1	2021-09-30 14:24:07	85 He 31	85 He	Sample		1.70000

Page 1 / 2 2021-09-30 14:28:39

Date: September 30, 2021 4:07:18 PM
 System ID: JP15471169

Page 25 / 34

Document Name:

Test Report

Batch Summary Report

Analyte Table

Sample Name	Conc
85 He 31	42.8500

Page 2 / 2 2021-09-30 14:28:40

Date: September 30, 2021 4:07:18 PM
 System ID: JP15471169

Page 26 / 34

General

Document Name: Test Report

Batch Summary Report

Batch Folder: D:\Agilent Services\DO 30 Sep 2021\B5 H2 new\B5
Analysis File: B5 H2 new batch\bin
Tune Step: #1 H2

Run	Acq. Date-Time	Data File	Sample Name	Type	Level	Dilution
1	2021-09-30 15:08:58	B5 H2.d	B5 H2	Sample		1.0000

Page 1 / 2

2021-09-30 15:10:31

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 27 / 34

Document Name: Test Report

Batch Summary Report

Sample Name	TS
B5 H2	2.1300

Page 2 / 2

2021-09-30 15:10:31

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 28 / 34

General

Document Name: Test Report

Batch Summary Report

Batch Folder: D:\Agilent Services\DO 30 Sep 2021\20 Min\B5
Analysis File: 20 Min batch\bin
Tune Step: #1 No Gas

Run	Acq. Date-Time	Data File	Sample Name	Type	Level	Dilution
1	2021-09-30 15:17:44	20 Min.d	20 Min	Sample		1.0000

Page 1 / 2

2021-09-30 15:48:42

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 29 / 34

Document Name: Test Report

Batch Summary Report

Sample Name	7 - 130 Sec 1	8 - 130 Sec 1	10 - 130 Sec 1	12 - 130 Sec 1	140 - 130 Sec 1	170 - 130 Sec 1
20 Min	0.00402	7.80464	0.48812	0.11495	0.41014	0.73211

Page 2 / 2

2021-09-30 15:48:43

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 30 / 34

Electronic Signature

Purpose

This signature page was created and published because the ACE sign-off action was executed, which is valid for the entire document, including attachments. The ACE sign-off is an electronic signature that requires two distinct identification components: unique username and personal password. The Agilent representative who has delivered this service understands the meaning and legal status of an electronic signature. As a trained official operator, the Agilent representative has a unique password and login to access ACE and electronically sign this document. (Other e-signatures can be applied to this document using a Document Content Management or other suitable method defined in your data access and control procedures.)

Details

Full Name of Signer: Panthep Kurasathain
Logged On User Name: panthep_kurasathain@agilent.com
Signature Creation Date: September 30, 2021
Reason for Signature: Executed protocol and published this original version of document

Regulatory Disclaimer

This document provides a protocol to verify and record instrument configuration and evidence of proper operation. It has been prepared from our interpretation of applicable regulations as well as industry best practices. The document is designed to provide an important component of a complete compliance package. Validation depends upon many factors and use of this protocol alone does not assure compliance. Agilent Technologies makes no premises or representations as to its sufficiency for any specific regulatory program.

Warranty

Agilent Technologies makes no warranty of any kind to this material, including but not limited to, the implied warranties of merchantability and fitness for a particular purpose. Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Date: September 30, 2021 4:07:16 PM
System ID: JP15471169

Page 31 / 34

User Name: panthep_kurasathain
Host Name: ASBKWX315
System ID: JP15471169
Print Date: September 30, 2021 4:07:22 PM

ALS QGHW 7900 30Sep21 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
September 30, 2021 3:50:07 PM	Audit	SessionCreated	Session	None
September 30, 2021 3:50:07 PM	Start	Configuration	Session	None
September 30, 2021 3:50:07 PM	Audit	Entitlement	Licensing	User is FieldEngineer and does not require an unlock code
September 30, 2021 3:52:52 PM	Audit	Equipment	Session	EQP details for primary technique [eqpMn] - File path: [ProtocolPacks\eqpMn\Conf\unifocal02.200eqpMn.02.50.eqp], EQP File Name: [eqpMn.02.50.eqp], EQP Name: [AgilentRecommended]
September 30, 2021 3:52:54 PM	End	Configuration	Session	None
September 30, 2021 3:52:57 PM	Start	Qualification	Session	OQ
September 30, 2021 3:52:57 PM	Start	Execution	Autosampler Check : SP54: Autosampler Check	None
September 30, 2021 3:53:39 PM	End	Execution	Autosampler Check : SP54: Autosampler Check	Run Count : 1
September 30, 2021 3:53:04 PM	Start	Execution	Integrated Sample Introduction System (ISIS) Check : ISIS: Integrated Sample Introduction System (ISIS) Check	None
September 30, 2021 3:53:08 PM	End	Execution	Integrated Sample Introduction System (ISIS) Check : ISIS: Integrated Sample Introduction System (ISIS) Check	Run Count : 1

Page 1 / 3

Date: September 30, 2021 4:07:16 PM
System ID: JP15471169

Page 32 / 34

User Name: panthep_kurasathain
Host Name: ASBKWX315
System ID: JP15471169
Print Date: September 30, 2021 4:07:22 PM

ALS QGHW 7900 30Sep21 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
September 30, 2021 3:53:19 PM	Start	Execution	Audit: G8403A: Audit: 1	None
September 30, 2021 3:55:08 PM	End	Execution	Audit: G8403A: Audit: 1	Run Count : 1
September 30, 2021 3:55:12 PM	Start	Execution	Background (No Gas Mode): G8403A: No Gas Mode Background 1	None
September 30, 2021 3:55:40 PM	End	Execution	Background (No Gas Mode): G8403A: No Gas Mode Background 1	Run Count : 1
September 30, 2021 3:56:43 PM	Start	Execution	Background (Gas Modes): G8403A: Gas Mode Background :Helium	None
September 30, 2021 3:56:17 PM	End	Execution	Background (Gas Modes): G8403A: Gas Mode Background :Helium	Run Count : 1
September 30, 2021 3:56:19 PM	Start	Execution	Background (Gas Modes): G8403A: Gas Mode Background :Hydrogen	None
September 30, 2021 3:56:39 PM	End	Execution	Background (Gas Modes): G8403A: Gas Mode Background :Hydrogen	Run Count : 1
September 30, 2021 3:56:41 PM	Start	Execution	20 Minute Stability (No Gas Mode): G8403A: 20 Minute Stability (No Gas Mode) 1	None
September 30, 2021 3:57:22 PM	End	Execution	20 Minute Stability (No Gas Mode): G8403A: 20 Minute Stability (No Gas Mode) 1	Run Count : 1
September 30, 2021 3:57:24 PM	End	Qualification	Session	OQ
September 30, 2021 3:57:24 PM	Start	Reporting	Session	None

Page 2 / 3

Date: September 30, 2021 4:07:16 PM
System ID: JP15471169

Page 33 / 34

User Name: panthep_kurasathain
Host Name: ASBKWX315
System ID: JP15471169
Print Date: September 30, 2021 4:07:22 PM

ALS QGHW 7900 30Sep21 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
September 30, 2021 4:03:07 PM	Audit	Reporting	Session	Report Generated : Certificate
September 30, 2021 4:03:17 PM	Audit	Reporting	Session	Report Generated : Report
September 30, 2021 4:03:19 PM	Start	Qualification	Session	OQ
September 30, 2021 4:04:08 PM	End	Qualification	Session	OQ
September 30, 2021 4:04:08 PM	Start	Reporting	Session	None
September 30, 2021 4:04:28 PM	Audit	Reporting	Session	Report Generated : Certificate
September 30, 2021 4:04:28 PM	Audit	Reporting	Session	Report Generated : Report

Page 3 / 3

Date: September 30, 2021 4:07:16 PM
System ID: JP15471169

Page 34 / 34



Certificate No. T220730

Page 1 of 6

Certificate of Calibration

Equipment : HEATING BLOCK
Manufacturer : Environmental Express
Model : SC 196
Serial No. : 6974CECW3285
Customer Code : BKK_EL0054
ID No. : T5306A3
Customer : ALS Laboratory Group (Thailand) Co.,Ltd.
104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Phatthanakan,
Khet Suan Luang, Bangkok 10250
Customer Location : Acid Digestion Lab
Date of Receipt : 30 March 2022
Calibrated By : Watcharapon Sangtong (Technician)
Approved By : [Signature] / Sujjar Nakuakred (Site Calibration Manager)
Date of Issue : 12 APR 2022

REVIEW BY [Signature]
APPROVED BY [Signature]
NEXT CAL DATE 7/10/23

The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standard laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Metrological Center.

FM-L13 108 30-05-57



Certificate No. T220730

Page 2 of 6

Calibration Report

Equipment : HEATING BLOCK
Date of Calibration : 7 April 2022
Environment : Temperature : 21.8-23.1 °C
Line Voltage : 221.6-226.3 V
Relative Humidity : 55 - 65 %RH

Condition of this results of calibration :

1. This equipment was calibrated by insert nine standard thermocouples type T into its chamber , the other one standard thermocouples type T use for ambient temperature measurement . The calibration was done in according to WI-T20.

All data show below were final values and the initial data from customer request . The temperature scale used was based on ITS - 90 .

2. Reference Standard Instrument :

Instrument	Model	Instrument No.	Certificate No.	Due Date
TC	TYPE T	TN221-TN230	T210008	08 June 2022
TC	TYPE T	TN231-TN240	T210008	08 June 2022
DATA LOGGER	34970A	T149	T210008	08 June 2022

3. This certificate is traceable to :

National Institute of Metrology (Thailand) through Metrological Center (NSC-TISI-TIS 17025 CALIBRATION 0244)

4. Condition of calibrated item : good

Equipment Description :

Time Constant : 2 Hour 25 Minute At 95 °C
Fresh Air Damper ☐ Open ☐ Min ☐ Medium ☐ Max
☐ Close
☒ Not Available

5. Adjustment :

() without adjustment (X) after adjustment

Approved By. [Signature]


FM-L13 108 30-05-57



Certificate No. T220730

Page 3 of 6

Calibration Report



FRONT CONTROL

Approved By. [Signature]

FM-L13 108 30-05-57



Certificate No. T220730

Page 4 of 6

Calibration Report

Measurement Results

Calibration Point	Average Standard Reading at each position (°C)					
R1 Hole1-Hole6	TN221	TN222	TN223	TN224	TN225	TN226
CAL POINT	Max	93.60	93.82	94.05	94.20	94.36
	Min	93.07	93.26	93.51	93.66	93.82
	Average	93.33	93.54	93.78	93.93	94.09
R2 Hole7-Hole12	TN227	TN228	TN229	TN230	TN231	TN232
	Max	94.59	94.79	94.63	94.55	94.82
	Min	94.05	94.25	94.08	93.97	94.26
	Average	94.32	94.52	94.36	94.26	94.54
R3 Hole13-Hole18	TN233	TN234	TN235	TN236	TN237	TN238
	Max	95.03	94.54	94.78	94.84	95.06
	Min	94.46	93.98	94.20	94.28	94.49
	Average	94.74	94.26	94.49	94.56	94.78
R4 Hole19-Hole24	TN239	TN240	TN221	TN222	TN223	TN224
	Max	94.89	94.82	95.73	95.85	95.72
	Min	94.33	94.26	95.51	95.62	95.51
	Average	94.61	94.54	95.62	95.73	95.62
R5 Hole25-Hole30	TN225	TN226	TN227	TN228	TN229	TN230
	Max	96.28	96.39	96.37	96.54	96.19
	Min	96.01	96.10	96.02	96.20	95.89
	Average	96.15	96.24	96.20	96.37	96.04
R6 Hole31-Hole36	TN231	TN232	TN233	TN234	TN235	TN236
	Max	96.84	96.97	97.03	96.48	96.33
	Min	96.53	96.65	96.71	96.08	95.98
	Average	96.68	96.81	96.87	96.28	96.16
R7 Hole37-Hole42	TN237	TN238	TN239	TN240	TN221	TN222
	Max	96.46	96.15	96.19	96.06	96.95
	Min	96.13	95.84	95.85	95.72	96.64
	Average	96.30	95.99	96.02	95.89	96.80
R8 Hole43-Hole48	TN223	TN224	TN225	TN226	TN227	TN228
	Max	96.91	96.58	96.13	96.19	96.34
	Min	96.55	96.21	95.80	95.87	96.03
	Average	96.73	96.40	95.96	96.03	96.18

Approved By. [Signature]

FM-L13 108 30-05-57



Metrological Center

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhohi, Saraburi 18110

Telephone : +66 2 586 5792-4 Fax : +66 2 586 5109

Website : www.scieco.co.th E-Mail : calibrate@scg.co.th

Certificate No. T220730

Page 5 of 6

Measurement Results

Calibration Report

Calibration Point	Average Standard Reading at each position (°C)					
R1 Hole1-Hole6	TN221	TN222	TN223	TN224	TN225	TN226
CAL POINT	Max	104.47	104.65	104.79	105.31	105.47
105	Min	104.15	104.27	104.45	104.98	105.14
	Average	104.31	104.46	104.62	105.15	105.31
R2 Hole7-Hole12	TN227	TN228	TN229	TN230	TN231	TN232
	Max	105.55	105.73	105.65	105.84	105.97
	Min	105.28	105.43	105.35	105.52	105.68
	Average	105.42	105.58	105.50	105.68	105.82
R3 Hole13-Hole18	TN233	TN234	TN235	TN236	TN237	TN238
	Max	106.14	106.06	105.81	106.05	105.81
	Min	105.85	105.81	105.55	105.80	105.53
	Average	106.00	105.94	105.68	105.92	105.67
R4 Hole19-Hole24	TN239	TN240	TN221	TN222	TN223	TN224
	Max	105.86	105.60	104.44	104.51	104.28
	Min	105.61	105.37	104.27	104.35	104.12
	Average	105.74	105.48	104.35	104.43	104.20
R5 Hole25-Hole30	TN225	TN226	TN227	TN228	TN229	TN230
	Max	104.94	104.93	104.97	105.08	104.68
	Min	104.77	104.75	104.76	104.90	104.51
	Average	104.85	104.84	104.86	104.99	104.60
R6 Hole31-Hole36	TN231	TN232	TN233	TN234	TN235	TN236
	Max	105.44	105.45	105.61	104.95	104.84
	Min	105.27	105.27	105.44	104.76	104.66
	Average	105.36	105.36	105.53	104.86	104.73
R7 Hole37-Hole42	TN237	TN238	TN239	TN240	TN221	TN222
	Max	105.17	104.70	104.59	104.51	105.22
	Min	105.00	104.53	104.41	104.35	105.04
	Average	105.08	104.62	104.50	104.43	105.13
R8 Hole43-Hole48	TN223	TN224	TN225	TN226	TN227	TN228
	Max	105.61	105.45	105.10	104.77	104.87
	Min	105.44	105.28	104.92	104.60	104.70
	Average	105.53	105.37	105.01	104.69	104.93

Approved By.

FM-L13 108-30-05-57



Metrological Center

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhohi, Saraburi 18110

Telephone : +66 2 586 5792-4 Fax : +66 2 586 5109

Website : www.scieco.co.th E-Mail : calibrate@scg.co.th

Certificate No. T220730

Page 5 of 6

Calibration Report

Measurement Results:

HEATING BLOCK			Temperature Distribution	
Setting (°C)	Reading (°C)		Stability (±°C)	Uncertainty (±°C)
	Min, Max	Average		
100.0	100.0, 100.4	100.1	0.29	0.83
105.0	105.0, 105.4	105.1	0.20	0.79

* The quoted uncertainty exclude " uniformity "

The calibration result apply only the above calibrated item.

The result of test was found accurate as shown on date and place of test only.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k which for a t-distribution, providing a level of confidence of approximately 95 %.

Approved By.

FM-L13 108-30-05-57



Metrological Center

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhohi, Saraburi 18110, Thailand.

Saraburi Tel : +66 3627 3096 Fax : +66 3627 3100

Bangkok Tel : +668 9205 6851 , +669 8247 2360

Website : www.scieco.co.th E-Mail : calibrate@scg.co.th



Certificate No. T211009

Page 1 of 4

Certificate of Calibration

Equipment : Chamber (Cold Room)

Manufacturer : KOLDTECH

Model : KM 320

Serial No. : TBN-1012061/05

Customer Code : BKK_EN0167

ID No. : T2463A3

Customer : ALS Laboratory Group (Thailand) Co.,Ltd.

104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Phatthanakan,

Khet Suan Luang, Bangkok 10250

Customer Location : Laboratory

Date of Receipt : 6 May 2021

Calibrated By : Watcharapon Songthong (Technician)

Approved By : / Boonchai Suriyawong (Site Calibration Manager)

Date of Issue : 20 MAY 2021

The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standard laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Metrological Center.

FM-L14 11701-02-64



Metrological Center

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhohi, Saraburi 18110, Thailand.



Certificate No. T211009

Page 2 of 4

Calibration Report

Equipment : Chamber (Cold Room)
 Date of Calibration : 18 May 2021
 Environment : Temperature : 23.4-24.9 °C
 Line Voltage : 221.4-230.2 V
 Relative Humidity : 55 - 65 %RH

Condition of this results of calibration :

1. This equipment was calibrated by insert 16 standard thermocouples type T into its chamber , the other one standard thermocouples type T use for ambient temperature measurement . The calibration was done in according to WI-T20 (based on ASTM E145-94 (Reapproved 2001) and AS2853-1986).

All data show below were final values and the initial data from customer request . The temperature scale used was based on ITS - 90 .

2. Reference Standard Instrument :

Instrument	Model	Instrument No.	Certificate No.	Due Date
TC	TYPE T	TN161-TN170	T210009	8 January 2022
TC	TYPE T	TN171-TN180	T210009	8 January 2022
DATA LOGGER	34970A	T149	T210009	8 January 2022

3. This certificate is traceable to :

National Institute of Metrology (Thailand) through Metrological Center (NSC-TISI-TIS 17025 CALIBRATION 0244).

4. Condition of calibrated item : good

Equipment Description :

Time Constant ☒ 1 Hour ☐ Minute At ☐ 3 °C
 Fresh Air Damper ☐ Open ☐ Min ☐ Medium ☐ Max
☒ Close ☒ Not Available

5. Adjustment :

(X) without adjustment () after adjustment

Approved By.

FM-L15 11715-05-63

- lightness visual check inside the Mercur
- visual check if gold-traps are broken
- visual check if spectrometer is contaminated
- reactor cleaning
- check pump-hose, if necessary change it
- check drying-hose, output gas-liquid-separator
- test Bubble-Sensor
- check gas flows
- check volume flows, reagents
- recording stray light values
- measurement with 30 ng/l

Serial No.: 701 239

- lubricate the dosing-winding (Teflon-grease-spray)
- clean the dosing cylinder, if necessary exchange it
- lubricate the winding system of the height drive with some drops of oil
- check the toothed belt
- check the position of the mechanical stopper (height: 13mm)
- check the pump rate of mixing pump (<14s AS52, typ.7s/<20s AS52S, typ.10s)
- check the pump rate of washing cup
- check the electrical hose connections for good contact
- check the connectors of the magnetic valves
- check the dosing hose for buckling, if necessary exchange it

Device parameter		nominal value	actual value
visual check general tightness inside the Mercur		o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
visual check Goldtraps		o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
visual check spectrometer			
	cuvette	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
	lens	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check pump hoses		o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check hoses and hose connectors		o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check and clean reactor		o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check drying hose output Gas-liquid-separator		o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check bubble-sensor		o.k.: <input checked="" type="checkbox"/>	not o.k.: <input type="checkbox"/>
Check gasflow			
	Argon pressure valve 4	1.2 – 1.5 bar	1.5 bar
	Valve 1	10 Nl/h or 0.166 NL/min	0.167 NL/min
	Valve 2	50 Nl/h or 0.833 NL/min	0.83 NL/min
	Valve 3	5 Nl/h or 0.083 NL/min	0.083 NL/min
	Valve 4	10 Nl/h or 0.166 NL/min	0.166 NL/min
Check liquid flow			
	Acid	2.5ml/min ± 1 ml	2.5 ml/min
	Red.-agent	2.5ml/min ± 1 ml	2.5 ml/min
	Sample	10ml/min ± 2 ml	10 ml/min
Adventitious light - values		(V)	from file
		100	0
		200	0
		300	0
		350	0
		400	1
		450	2
		500	6
		550	14
		575	20
		600	28

Device parameter	nominal value	actual value
Analytical parameters		
Conditions: max conc.: 10 µg/L PMT-voltage: <u>404</u> V		
Blank-solution		Int. <u>0.0005</u>
without enrichment / FBR 30 ng/L	Int > 0.0015 RSD < 3 %	Int. <u>0.0026</u> RSD <u>1.12</u> %
Conditions: max conc.: 1.7 µg/L PMT-voltage: <u>595</u> V		
Blank-solution		Int. <u>0.0018</u>
with enrichment / FBR 30 ng/L	Int > 0.008 RSD < 3 %	Int. <u>0.0104</u> RSD <u>0.59</u> %
Fok - factor (Int ₂ / Int ₁)	> 3.5	<u>4</u>
Comments		

Bangkok, 7/06/2022.
Place, Date (DD/MM/YYYY)

06/06/2022



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CALIBRATION AND TESTING EQUIPMENT SERVICES

534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2217-3000-34 FAX 0-2219-9484

Cert. No.: 21TM2188
Page.: 1 of 3

Certificate of Calibration

Equipment :	Autoclave
Manufacturer :	AES Laboratory
Model :	Masterclave 528
Serial No. :	34677152
ID No. :	BKK ML0043

Submitted by : ALS Laboratory Group (Thailand) Co.,Ltd.
104 Phatthanakan 40, Phatthanakan Rd.,
Khwaeng Phatthanakan, Khet Suan Luang,
Bangkok 10250 Thailand

Location : Media Preparation Room

Received Order : 1 December 2021
Calibration Date : 1 December 2021
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %

Calibrated by : Khit Ruttanaprapachai

Approved by : Malu.
Approved Signatory

() Pornthippa Tameyakul
(✓) Malee Butkruea
() Suwit Imjai

Issue Date : 7 December 2021

The Uncertainties are for a confidence probability of approximately 95%.

This certificate may not be reproduced other than in full, except with the prior written approval of the head of Calibration and Testing Equipment Services.



Equipment : Autoclave
Condition As-Received : Used Item
Reference : 2112-0002OC-2
Procedure Used :-

Cert. No.: 21TM2188
Page.: 2 of 3

Calibration was conducted using in-house calibration procedure CP-OT03 according to direct measurement method with Data Acquisition which connected with Thermocouple Type T
The temperature scale used was based on ITS-90.


Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34970A	MY44060450	21LM4/1	06 Mar 2022

- This certificate is valid only to the item calibrated on date and place of calibration.
 - This certification is traceable to the International System of Unit.
 - This result of calibration covers laboratory autoclaves for the sterilization of goods and material which could be infected with organisms categorized as Hazard Group 1, 2 and 3**
 - (** = Categorization of pathogens according to hazard and categories of containment, second edition, 1990)
- It does not cover autoclaves for use with material infect with organisms in Hazard Group 4, for which complete containment and sterilization of infected condensate is considered to be essential.
This result of calibration does not apply to sterilizers or disinfectors used for medical, dental, pharmaceutical or veterinary purposes which are directly concerned with patient care, or those used for fabrics subjected to sterilization which are required to be dry at the end of cycle.

Result of Calibration :- (*) Without Adjustment
Function of UUC* :- Temperature Source



	Environmental		
	(°C)	(%R.H.)	(Volt)
Beginning of Calibration	24	51	220
Finished of Calibration	25	53	221

Position	Description	Ref. Std. ID No.:
1 =	Center of chamber	19-14TC-01
2 =	Temperature sensor	19-14TC-02
3 =	Exhaust port	19-14TC-03

Maha

a 1085616



Equipment : Autoclave
Condition As-Received : Used Item
Reference : 2112-0002OC-2
Result of Calibration :- (*) Without Adjustment

Cert. No.: 21TM2188
Page.: 3 of 3

Operating parameter Set : Temperature = 121.0 °C
Sterilization period = 15 minute

UUC* Setting (°C)	UUC* Reading (°C)	Position	Average* Standard Reading (°C)	Stability (± °C)	Pressure Reading (bar)	Uncertainty (± °C)	Coverage Factor k
121.0	120.7	1	120.792	0.078	1.1	0.75	2
		2	120.674				
		3	120.815				

Average* : The average of 30 values in each position.

Stability : One-half of the greatest maximum difference of measured temperature at any one probe.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-000-

Maha

a 1085615



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES & EQUIPMENT CALIBRATION AND TESTING SERVICES
534-1 PATTANAKARN ROAD SOI 18, SUANLUANG, BANGKOK 10250
TEL: 0-2713-3000-27 FAX: 0-2719-9484



Cert. No.: 21TM64
Page.: 1 of 3

Certificate of Calibration

Equipment : Incubator

Manufacturer : ShelLab

Model : 1915A

Serial No. : 0200599

ID No. : BKK_ML0010

Submitted by : ALS Laboratory Group (Thailand) Co., Ltd.
104 Phatthanakan 40, Phatthanakan Rd.,
Khwaeng Phatthanakan, Khet Suan Luang,
Bangkok 10250 Thailand

Location : Incubation & Microbiological Reading

Received Order : 5 January 2021

Calibration Date : 5 January 2021

Ambient Temperature : (26 ± 10) °C


Relative Humidity : (50 ± 30) %

Calibrated by : Krisda Malee

Approved by :
Approved Signatory

- () Pornthippa Tameyakul
(✓) Malee Butkruea
() Suwit Imjai

Issue Date : 13 January 2021



The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services - Equipment Calibration and Testing Services.

A 0023539



Equipment : Incubator
Condition As-Received : Used Item
Reference : 2101-0012OC-1
Procedure Used :-

Cert. No.: 21TM64
Page.: 2 of 3

Calibration was conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD).

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Serial No.	Cert. No.	Traceable	Due Date
1) Data Acquisition	MY44067817	20LMB	NIST, NIMT	29 Jul 2021

2. This certification is traceable to the SI unit.

3. This certificate is valid only to the item calibrated on date and place of calibration.


Remark : NIST : National Institute of Standards and Technology, The United State of America.

NIMT : National Institute of Metrology Thailand.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Close



Probe installation Details :

a = 10 cm
b = 10 cm
c = 10 cm

Dimension of Chamber :
D = 0.90 m
W = 0.75 m
H = 1.2 m
Capacity = 0.81 m³

	Environment during calibration	
	Beginning	Finished
Temp. (°C)	25	24
REL. Humid. (%)	53	51
AC Supply (Volt)	221	220

Position :	Ref. Std./ID No.:
1	15RTD2/11
2	15RTD2/12
3	15RTD2/13
4	15RTD2/14
5	15RTD2/15
6	15RTD2/16
7	15RTD2/17
8	15RTD2/18
9 (ref.)	15RTD2/19

Maha

a 1036401

a 1036400

Λ 0029135

a 1059245

a 1059244



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534-4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-27 FAX 0-2718-9484



Cert. No.: 22TM453
Page: 1 of 3

Certificate of Calibration

Equipment : Water Bath

Manufacturer : Memmert

Model : WB 45

Serial No. : I704.0285

ID No. : BKK_ML0052

Submitted by : ALS Laboratory Group (Thailand) Co., Ltd.
104 Phatthanakan 40, Phatthanakan Rd.,
Khwaeng Phatthanakan, Khet Suan Luang,
Bangkok 10250 Thailand
Incubation & Microbiological Reading

Location :

Received Order : 21 February 2022

Calibration Date : 21 February 2022

Ambient Temperature : (26 ± 10) °C

Relative Humidity : (50 ± 30) %

Calibrated by : Prawit Sodavitchit

Approved by :

() Pornthippa Tameyakul
() Malee Butkruea
() Suwit Imjai

Issue Date : 25 February 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3: Equipment Calibration and Testing Services

A 0038346



Equipment : Water Bath
Condition As-Received : Used Item
Reference : 2202-0615OC-2
Procedure Used :-

Cert. No.: 22TM453
Page: 2 of 3

Calibration were conducted using in-house calibration procedure CP-OT04 according to direct measurement method with Data Acquisition which connected with Industrial Platinum Resistance Thermometer (IPRT).

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34970A	MY44035217	21LM30	23 Dec 2022


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Result of Calibration :- () Without Adjustment

Function of UUC* : Temperature Source

	Environmental		AC Voltage Supply
	(°C)	(%R.H.)	(Volt)
Beginning of Calibration	25	51	220
Finished of Calibration	25	50	220



Front

Position :	Ref. Std. S/N.:
1	N37P300726
2	N37P300727
3	N37P300728
4	N37P300729
5(ref.)	N37P300730

Malee

a 1097103



Equipment : Water Bath
Condition As-Received : Used Item
Reference : 2202-0615OC-2
Result of Calibration :- () Without Adjustment
Function of UUC* : Temperature Source

Cert. No.: 22TM453
Page: 3 of 3

Calibration point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Average* Standard Reading (°C)				
			1	2	3	4	5 (ref.)
44.5	45.1	45.1	44.503	44.454	44.497	44.519	44.476

Calibration point (°C)	Uniformity (°C)	Stability (± °C)	Uncertainty (± °C)	Coverage Factor k
44.5	0.13	0.063	0.15	2

Average* : The average of 30 values in each position.

Uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Stability : One-half of the greatest maximum difference of measured temperature at any one probe.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-000-

Malee

a 1097102



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534-4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-27 FAX 0-2718-9484



Cert.No.: 21CH1733
Page: 1 of 3

Certificate of Calibration

Equipment : pH Meter

Manufacturer : Mettler Toledo

Model : SevenExcellence

Serial No. : B834291445

ID No. : RYG_EN0152

Condition As-Received : Used Item

Received Date : 22 December 2021

Calibration Date : 23 December 2021

Reference : 2112-0636DSC-2

Submitted by : ALS Laboratory Group (Thailand) Co., Ltd. Rayong Branch
616/10 Moo 5 T.Maenam Khu, A Pluakdaeng,
Rayong 21140, Thailand

Ambient Temperature : (25 ± 2.5) °C

Relative Humidity : (50 ± 15) %

Calibration Procedure :
In-house method :
- CP-CH5 by direct measurement with standard voltage calibrator and direct measurement with certified reference material (CRM)
- CP-CH8 by comparison with standard thermometer

Calibrated by : Warakorn Lerngagrakul

Approved by :

() Malee Butkruea
() Sathip Meangmai
() Warakorn Lerngagrakul

Issue Date : 24 December 2021

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3: Equipment Calibration and Testing Services

A 0036356



Cert.No.: 21CH1733
Page.: 2 of 3

Condition of this calibration result

1. Reference Standard Instrument

Instrument	Serial No.	ID No.	Cert. No.	Due Date
1) Document Process Calibrator	54030049	130RC116	21E2682	25 Aug 2022
2) Ref. Standard Thermometer	4982054	110RC044	2111201	26 Oct 2022

This certification is traceable to the International System of Unit maintained at:-
- Traceable to National Institute of Metrology (Thailand), NIMT

2. Certified Reference Materials

The measurement results are traceable to SI through CPA chem Ltd., ANSI-ASQ National Accreditation Board, Accredited No. AR-1835

Buffer Solution	Manufacturer	Lot No.	Exp. date
pH 4.008	CPA chem	761016	02 Aug 2023
pH 6.982	CPA chem	761017	02 Aug 2022
pH 10.015	CPA chem	761018	02 Aug 2022

3. This certificate is valid only to the item calibrated on date and place of calibration.

Calibration Results

Function : mV Measurement

Performing standard curve by Fluke at pH (4,7,10)

Unit Under Calibration	Nominal Value		Standard Voltage Input		Actual Reading		Uncertainty of Measurement (\pm mV)	Coverage factor k
	pH	mV	mV	pH	mV	pH		
pH Meter S/N.: B934291445	4.000	177.48	177.3	4.000	177.3	4.000	0.058	2.00
	7.000	0.00	-0.1	7.000	0.058	7.000	0.058	2.00
	10.000	-177.48	-177.5	10.000	0.058	10.000	0.058	2.00

a 1087319



Cert.No.: 21CH1733
Page.: 3 of 3

Calibration Results

Function : pH Measurement

Performing three buffers standard curve by using buffer nominal pH (4,7,10)

Unit Under Calibration	Standard pH Buffer Solution	Actual pH Reading	Actual mV Reading (mV)	Uncertainty of pH measurement (\pm)	Coverage factor k
pH Electrode S/N.: 1475518	4.008	4.011	180.6	0.0049	2.05
	6.982	6.984	5.3	0.0077	2.00
	10.015	10.014	-171.3	0.0065	2.00

Function : Temperature Measurement

(*) Without adjustment

This equipment was connected with Temperature Probe;

- Model : InLab Expert Pro-ISM

- Serial No. : 1475518

Dimension of probe:

- Length : 120 mm.

- Diameter : 12 mm.

- Immersion Depth : 100 mm.

Calibration Point (°C)	Standard Temperature (°C)	UUC* Reading (°C)	Error (°C)	Uncertainty of measurement (\pm °C)	Coverage factor k
25.0	25.002	24.9	-0.102	0.13	2.00

Remark : - UUC* = Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %

-o0o-

a 1087180



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
53/44 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG, BANGKOK 10250
TEL. 0-2717-3000-24 FAX 0-2719-9484



Certificate of Calibration

Certificate No.: 21E4151
Page: 1 of 2

Equipment : pH Meter
Manufacturer: Mettler Toledo
Model : SevenExcellence
Serial No.: B834291445
ID No.: RYG_END152

Condition As-Received: Used item
Received Date: 22 December 2021
Calibration Date: 28 December 2021

Reference: 2112-0636DSC
Ambient Temperature: (23 \pm 2) °C
Relative Humidity: (50 \pm 10) %
Submitted by: ALS Laboratory Group (Thailand) Co., Ltd. Rayong Branch
616/10 Moo 5 T.Maenam Khu, A.Puakdaeng, Rayong
21140, Thailand

Procedure used: Calibration was conducted using In-house calibration Procedure CP-E17 According to direct measurement method with Multi-Product Calibrator.

Condition of this result of calibration

1. Reference standards instruments :

Instrument	Model	Serial No.	Certificate No.	Due Date
1) Multi-Product Calibrator	5500A	6440007	21E1444	07 May 2022

2. This result of calibration was made on requested at the point specified by customer.

3. The certificate is valid only to the item calibrated on date and place of calibration.

4. This Certification is traceable to the International System of Unit maintained at:-
- National Institute of Metrology Thailand (NIMT)

Calibrated by: Wutthareeporn Wongchutikrane
Issue Date: 07 January 2022
Approved Signatory: [Signature]
[Signature]
[Signature]
[Signature]

B 0278122



Cert. No.: 21E4151
Page.: 2 of 2

Result of calibration :- (*) Without adjustment () After adjustment

Function: DC voltage measurement	Range: 2000 mV	Standard Value (mV)	UUC* Reading (mV)	Error (mV)	Uncertainty (\pm μ V)
		-100.0000	-100.0	0.0	65
		-50.0000	-50.0	0.0	62
		0.0000	0.0	0.0	58
		50.0000	50.0	0.0	62
		100.0000	100.0	0.0	65
		150.0000	150.0	0.0	69
		200.0000	199.9	-0.1	72

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95 %

*UUC= Unit Under Calibration.

-o0o-

a 1088610



Cert.No.: 18CG4595
Page: 1 of 2

Certificate of Calibration

Equipment : Burette
Capacity : 50 mL
Serial No. :
ID. No. : 243007
Manufacturer : Wileg
Made in : Germany
Submitted by : ALS Laboratory Group (Thailand) Co., Ltd.
Eastern Seaboard Industrial Estate (Rayong)
64/77 Moo 4, Building No.B1, Highway 331, km 91.5
T.Pluakdaeng, A.Pluakdaeng, Rayong 21140
Ambient Temperature : (22 ± 2.5) °C
Relative Humidity : (50 ± 10) %
Barometric Pressure : 757 mmHg
Calibration Procedure : ASTM E 542 - 01
Calibrated by : Natcha Chayingcheiw

Approved by :
() Pornthippa Tameyakul
(✓) Malee Bulkruea
() Ponpan Paipim
() Srisuda Khamtha

Issue Date : 27 September 2018

The Uncertainties are for a confidence probability of approximately 95 %.

This certificate may not be reproduced other than in full, except with the prior written approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

A 0087224



Equipment : Burette
Capacity : 50 mL
Serial No. :
ID. No. : 243007
Manufacturer : Wileg
Received Date : 10 September 2018
Condition As-Received : Used Item
Calibration Date : 21 September 2018
Reference : 1809-0411DPC

Cert.No.: 18CG4595
Page: 2 of 2

Condition of this result of calibration

- Reference Standard Instruments:

Instruments	Model	Serial No.	ID. No.	Certificate No.	Traceability	Due date
1) Balance	XP205DR	1126143764	140RC004	18MM1	NIMT	2 Jan 2019

This certification is traceable to SI Unit
- This certificate was certified only for the measuring instrument we calibrated.
- This result of calibration was found accurate as shown on date and place of calibration only.
- True value is converted to true volume at the standard temperature of 20 °C

Calibration result :

Nominal capacity (mL)	Reading (mL)	Uncertainty (± mL)	k Factor
50	49.9901	0.010	2.00

Remark mL = cm³

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-000-

a 0901034



PENTA CALIBRATION CO., LTD.
66/124 The Connect 33 Village Kanchanaphisek Road
Dokmai Prawat Bangkok 10250
Tel : +66 (0) 2069-9773
www.pentalcal.com

Certificate of Calibration

Represent to Certificate of Calibration : PTC/07/22071

Certificate No. : PTC/07/22071
Equipment : Digital Balance
Manufacturer : Sartorius
Model : MSE224-100-DU
Type of Balance : Single interval
Condition : Normal
Serial No. : 26207042
ID No. : BKK, EN0002

Customer : ALS Laboratory Group (Thailand) Co., Ltd.
104 Phatthanakarn 40 Phatthanakarn Rd.,
khwaeng Phatthanakarn, Khet Suan Luang, Bangkok 10250.

Environment Condition : Temperature 21.5 °C ± 0.7 °C
Humidity 61.8 %RH ± 4.7 %RH
Air density 1.19 kg/m³

Calibration Place : ALS Laboratory Group (Thailand) Co., Ltd.
104 Phatthanakarn 40 Phatthanakarn Rd.,
khwaeng Phatthanakarn, Khet Suan Luang, Bangkok 10250.

The Method used : In house method, PTC-WI-07, base on Euramet cg 18
Traceability : This certificate is traceable to the SI Units through Thai Calibration Service Co., Ltd.
Date Received : February 25, 2022
Calibration Date : February 25, 2022
Issued Date : March 01, 2022
Calibration By : Mr Rungroje Metakul

Reviewed by :
(Mr Kiangsak Kalasri)

Approved By :
(Mr Keattisak Kerdio)
Laboratory Manager

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognised national standard laboratories.
The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to the Expression of Uncertainty in Measurement (GUM). The effect that the results relate only to the items calibrated.
This calibration certificate shall not be reproduced except in full only, without written approval from penta calibration co., ltd.

PTC-FMC-01-02-14-03-2021



PENTA CALIBRATION CO., LTD.
66/124 The Connect 33 Village Kanchanaphisek Road
Dokmai Prawat Bangkok 10250
Tel : +66 (0) 2069-9773
www.pentalcal.com

Represent to Certificate of Calibration : PTC/07/22071


Certificate No. : PTC/07/22071
Page : 2 of 2

Measurement Results:

Without Adjustment :

Function Calibration : Non Adjustment

Eccentric Error: Weight to be 1/3, 1/2 or of Maximum capacity



Eccentricity test 100 (g)

Position (g)				
1	2	3	4	5
0.0000	-0.0002	0.0001	0.0001	-0.0001
Maximum deviation: 0.0002				

Repeatability Test : Weight to be 1/2 ≤ L₁ ≤ Maximum capacity

Determination of the standard deviation of weighing balance : Readability 0.0001 (g)

Nominal test value (g)	Standard Deviation
200	0.00005

Error of Indication : from nominal value, Readability 0.0001 (g)

Nominal Value (g)	Conventional Mass (g)	Indication (g)	Correction of Balance (g)	Uncertainty (g)	k
0	0.00000	0.0000	0.0000	0.00016	2.52
0.1	0.10000	0.1000	0.0000	0.00017	2.20
0.5	0.50000	0.5000	0.0000	0.00016	2.28
1	1.00001	1.0000	0.0000	0.00016	2.28
2	2.00001	2.0000	0.0000	0.00016	2.28
5	5.00001	5.0000	0.0000	0.00016	2.28
10	10.00002	10.0000	0.0000	0.00016	2.28
20	20.00002	20.0000	0.0000	0.00016	2.23
50	50.00001	50.0000	0.0000	0.00017	2.15
100	100.00002	99.9999	0.0001	0.00020	2.06
120	120.00004	120.0000	0.0000	0.00023	2.03
150	150.00003	150.0000	0.0000	0.00026	2.00
200	200.00003	199.9999	0.0001	0.00030	2.00

Note: Weight of adjust (g)

The End of Certificate

PTC-FMC-01-02-14-03-2021



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534-4 PATTANAKARN ROAD SRI 18 SUANLUANG SUANLUANG BANGKOK 10250
TEL: 0-2717-3000-27 FAX: 0-2719-9484



Cert. No.: 21TM827
Page: 1 of 3

Certificate of Calibration

Equipment : Hot Air Oven
Manufacturer : Memmert
Model : UFE 500
Serial No. : G511.1572
ID No. : RYG_EN0010
Submitted by : ALS Laboratory Group (Thailand) Co., Ltd. (Rayong Branch)
616/10 Moo 5 T. Maenam Khu.
A. Pluakdaeng,
Rayong 21140 Thailand
Location : Oven Room
Received Order : 5 May 2021
Calibration Date : 5 May 2021
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %
Calibrated by : Khit Rutanaprapachai
Approved by :
() Pornthippa Tameyakul
(✓) Malee Butkruea
() Suwit Imjai
Issue Date : 14 May 2021



The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3: Equipment Calibration and Testing Services

A 0028099



Equipment : Hot Air Oven
Condition As-Received : Used Item
Reference : 2105-0005OC-4

Cert. No.: 21TM827
Page: 2 of 3

Procedure Used :-

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD) and Thermocouple Type T.

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34972A	MY57013823	21LM3	26 Feb 2022

2. This certificate is valid only to the item calibrated on date and place of calibration.


3. This certification is traceable to the International System of Unit.

Result of Calibration :- (°) Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Close

Environment during calibration		
	Beginning	Finished
Temp. (°C)	28	29
REL Humid. (%)	59	56
AC Supply (Volt)	220	221



Probe Installation Details : Dimension of Chamber :
a = 5.0 cm D = 0.40 m
b = 5.0 cm W = 0.56 m
c = 5.0 cm H = 0.48 m
Capacity = 0.11 m³

Ref. Std. ID No.: @ Calibration Point		
Position :	(104) °C	(180) °C
1	21-17RTD-01	19-17TC-01
2	21-17RTD-02	19-17TC-02
3	17RTD-03	19-17TC-03
4	17RTD-04	19-17TC-04
5	17RTD-05	19-17TC-05
6	17RTD-06	19-17TC-06
7	17RTD-07	19-17TC-07
8	17RTD-08	19-17TC-08
9 (ref.)	17RTD-09	19-17TC-09

a 1054287



Equipment : Hot Air Oven
Condition As-Received : Used Item
Reference : 2105-0005OC-4
Result of Calibration :- (°) Without Adjustment
Function of UUC* : Temperature Source

Cert. No.: 21TM827
Page: 3 of 3

Calibration Point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Temperature stability (± °C)	Temperature uniformity (°C)	Overall Variation (°C)	Uncertainty (± °C)	Coverage Factor k
104.0	104.0	104.0	0.063	0.54	0.70	0.42	2
180.0	180.0	180.0	0.15	0.89	1.3	1.1	2

Calibration Point (°C)	Measured Temperature (°C)								
	1	2	3	4	5	6	7	8	9 (ref.)
104.0	104.243	103.732	103.760	103.742	103.863	103.743	104.311	103.689	103.815
180.0	180.101	180.481	179.401	179.692	179.980	179.943	180.127	179.915	179.709

Average* : The average of 30 values in each position.

Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor.

Temperature uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95%.

-000-

a 1054286




TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534-4 PATTANAKARN ROAD SRI 18 SUANLUANG SUANLUANG BANGKOK 10250
TEL: 0-2717-3000-27 FAX: 0-2719-9484



Cert. No.: 21CG1446
Page: 1 of 2

Certificate of Calibration

Equipment : Burette
Capacity : 50 mL
Serial No. :
ID. No. : BKK_EN0171
Manufacturer : Witeg
Made in : Germany
Submitted by : ALS Laboratory Group (Thailand) Co., Ltd.
104 Phatthanakan 40, Phatthanakan Rd.
Khwaeng Phatthanakan, Khet Suan Luang
Bangkok 10250 Thailand
Ambient Temperature : (20 ± 2.5) °C
Relative Humidity : (50 ± 10) %
Barometric Pressure : 755 mmHg
Calibration Procedure : ASTM E 542 - 01
Calibrated by : Sa-nguankam Wongsa



Approved by :
Approved Signatory

() Pornthippa Tameyakul
(✓) Malee Butkruea
() Ponpan Paipim
() Srisuda Khamtha

Issue Date : 31 March 2021

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3: Equipment Calibration and Testing Services

A 0026589



Equipment : Burette
Received Date : 24 March 2021
Condition As-Received : Used Item
Calibration Date : 30 March 2021
Reference : 2103-1006DSC-5

Cert.No.: 21CG1446
Page.: 2 of 2

Condition of this result of calibration

1. Reference Standard Instruments :

Instruments	Model	Serial No.	ID. No.	Certificate No.	Traceability	Due date
1) Balance	XP205	B134206712	140RC007	21MM181	NIMT	02 Mar 2022
2) Thermo-Hygrograph	TH 803	09153022	140EC004	20H1434	NIST,NIMT	19 June 2021
3) Thermometer		1594592	140EC010	20I1191	NIMT	08 Oct 2021

This certification is traceable to SI Unit

2. The certificate is valid only to the item calibrated on date and place of calibration.
3. True value is converted to true volume at the standard temperature of 20 °C

Calibration result :

Nominal capacity (mL)	Reading (mL)	Uncertainty (\pm mL)	k Factor
50	50.0041	0.011	2.00

Remark mL = cm³

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

-000-

Malu

a 1048960

ภาคผนวก จ

สำเนาหนังสือรับรองห้องปฏิบัติการวิเคราะห์เอกชน

ที่ อก ๐๓๑๐(๑)/ ๑๐๖๕



กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ เขตราชเทวี
กรุงเทพมหานคร ๑๐๕๐๐

๒๘ มกราคม ๒๕๖๕

เรื่อง ต่อยอดหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
เรียน กรรมการผู้จัดการ บริษัท เอแอลเอส แลบบอราทอรี กรุ๊ป (ประเทศไทย) จำกัด
อ้างถึง คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และชนิดสารเคมีของห้องปฏิบัติการวิเคราะห์เอกชน
ลงวันที่ ๓๐ กรกฎาคม ๒๕๖๓

สิ่งที่ส่งมาด้วย ๑. รายชื่อผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ จำนวน ๑ แผ่น
๒. รายชื่อเจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๕ แผ่น
๓. ขอบข่ายสารเคมีที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๓๑ แผ่น

ตามหนังสือที่อ้างถึง บริษัท เอแอลเอส แลบบอราทอรี กรุ๊ป (ประเทศไทย) จำกัด ขอต่ออายุ
หนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน เลขทะเบียน ๖-๒๐๔-๙-๕๐๔
ขอพัฒนาการ ๔๐ ถนนพัฒนาการ แขวงพัฒนาการ เขตสวนหลวง กรุงเทพมหานคร
ต่อกรมโรงงานอุตสาหกรรม นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว ให้บริษัท เอแอลเอส แลบบอราทอรี กรุ๊ป (ประเทศไทย)
จำกัด ต่อยอดหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน โดยมีองค์ประกอบดังนี้
ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ จำนวน ๖ ราย ตามสิ่งที่ส่งมาด้วย ๑
ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๑๖๒ ราย ตามสิ่งที่ส่งมาด้วย ๒
ค. ขอบข่ายสารเคมีที่ได้รับขึ้นทะเบียนไว้วิเคราะห์ในน้ำเสีย จำนวน ๕๔ รายการ น้ำได้ดิน
จำนวน ๑๒๖ รายการ อากาศเสีย ๑๖ รายการ สิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว จำนวน ๓๕ รายการ และดิน
จำนวน ๑๒๕ รายการ รวมทั้งสิ้นจำนวน ๓๖๑ รายการ ตามสิ่งที่ส่งมาด้วย ๓

หนังสือฉบับนี้จะหมดอายุในวันที่ ๒ กันยายน ๒๕๖๖ หากประสงค์จะต่ออายุหนังสือ
รับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ให้ยื่นคำขอต่ออายุพร้อมเอกสารประกอบคำขอ
ต่อกรมโรงงานอุตสาหกรรม ภายใน ๓๐ วัน ก่อนวันสิ้นอายุของหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์
เอกชน ซึ่งคำขอต่ออายุดังกล่าวขอรับได้ที่กรมโรงงานอุตสาหกรรม

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ

(นายศิระ จันทน์นิล)

นักวิทยาศาสตร์ชำนาญการพิเศษ รักษาการแทน
ผู้อำนวยการกองวิจัยและเฝ้าระวังมลพิษโรงงาน
ปฏิบัติการตามหน้าที่และอำนาจตามกฎหมาย

กองวิจัยและเฝ้าระวังมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๒๐๒ ๔๔๖๖ ๐ ๒๒๐๒ ๔๐๐๒

โทรสาร ๐ ๒๒๕๕ ๓๒๐๐ ๐ ๒๒๕๕ ๓๔๑๕

เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท เอแอลเอส แลบบอราทอรี กรุ๊ป (ประเทศไทย) จำกัด

เลขทะเบียน ๖-๒๐๔

ที่ อก ๐๓๑๐(๑)/

ลงวันที่ ๒๘ มกราคม ๒๕๖๕

ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ จำนวน ๖ ราย

- ๑) นางสาวยุพพร จันทร์ปลั่ง
- ๒) นางสาวชัชฌิยา โจนารกุล ณ นคร
- ๓) นายศรายุทธ จิตราภรณ์
- ๔) นางสาวกนกกร เอนก
- ๕) นายสุริยา สอนแก้ว
- ๖) นายวิชาญ ชุมพรี

- ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๐๐
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๐๑
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๐๒
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๐๓
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๐๔
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๐๕

(นายศิระ จันทน์นิล)

นักวิทยาศาสตร์ชำนาญการพิเศษ รักษาการแทน
ผู้อำนวยการกองวิจัยและเฝ้าระวังมลพิษโรงงาน
ปฏิบัติการตามหน้าที่และอำนาจตามกฎหมาย

เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท เอแอลเอส แลบบอราทอรี กรุ๊ป (ประเทศไทย) จำกัด

เลขทะเบียน ๖-๒๐๔

ที่ อก ๐๓๑๐(๑)/ ๑๐๖๕

ลงวันที่ ๒๘ มกราคม ๒๕๖๕

ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๑๖๒ ราย

- ๑) นางสาวจินดา จุลอรรถ
- ๒) นางสาววชิร น้อยเสียม
- ๓) นางสาวณัฏฐาญจน์ อิมหม
- ๔) นางสาววิมลพร สายสิง
- ๕) นางสาวนันทิยา สมบูรณ์
- ๖) นางสาวศรัณยา เจริญอักษร
- ๗) นางสาวสรวิทย์ มงคลจิรวุฒิ
- ๘) นางสาวศิริลักษณ์ พึ่งเพ
- ๙) นายณพพงศ์ จันทพินทุ์
- ๑๐) นายเกรียงไกร โกมลก
- ๑๑) นายอริสา จิราย
- ๑๒) นางสาวกนกพร แก้วมัน
- ๑๓) นางสาวสุวิมล ชัยเรืองวุฒิ
- ๑๔) นางสาวสุชาดา ธรรมการ
- ๑๕) นางสาวปณิภา ชัยเดชอนกุล
- ๑๖) นางสาวศศิธร หนูสวัสดิ์
- ๑๗) นางสาวเสาวลักษณ์ ภูนาอำพร
- ๑๘) นายอภิสิทธิ์ สิงหา
- ๑๙) นายศักดิ์สิทธิ์ ไพศาลพิสุทธิ
- ๒๐) ว่าที่ร้อยตรีหญิง พรหมนิภา ขันเจริญ
- ๒๑) นางจิตตา คำแก้ว
- ๒๒) นางสาวอรรณพ รักย
- ๒๓) นางสาวนพรัตน์ แยมกรานนท์
- ๒๔) นายจุลเดช วารินทร์
- ๒๕) นางสาวตาจุติรัตน์ รือคำ
- ๒๖) นายนคร สุขเจริญ
- ๒๗) นายบัญชา นามเขตต์
- ๒๘) นายพรมณ์ ศรีปิ่นพร
- ๒๙) นายอุทิศ จันทิมา
- ๓๐) ว่าที่ร้อยตรี เจริญเกียรติ อมรศรีเสริม
- ๓๑) นางสาววิภา สรวิภา
- ๓๒) นายอนุพงศ์ รัตนศรีประเสริฐ
- ๓๓) นางสาวจุฑาทิพย์ โอนสินเกียรติ
- ๓๔) นางสาวจางวรณ พินพิศกุลพิทยา

- ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๐๖
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๐๗
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๐๘
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๐๙
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๑๐
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๑๑
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๑๒
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๑๓
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๑๔
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๑๕
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๑๖
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๑๗
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๑๘
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๑๙
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๒๐
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๒๑
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๒๒
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๒๓
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๒๔
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๒๕
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๒๖
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๒๗
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๒๘
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๒๙
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๓๐
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๓๑
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๓๒

(นายศิระ จันทน์นิล)

๓๕) นางสาวปรางค์ทิพย์...

นักวิทยาศาสตร์ชำนาญการพิเศษ รักษาการแทน
ผู้อำนวยการกองวิจัยและเฝ้าระวังมลพิษโรงงาน
ปฏิบัติการตามหน้าที่และอำนาจตามกฎหมาย

- ๒ -

- ๓๕) นางสาวปรางค์ทิพย์ กิจไพศาลศักดิ์
- ๓๖) นางสาวเดือนใจ ทางกลาง
- ๓๗) นางสาวจิราพร ศิริเวช
- ๓๘) นายวรากร สุภักดิ์
- ๓๙) นายทนง วีระะศักดิ์
- ๔๐) นายณิธิ เจนจบ
- ๔๑) นายณิศร ขำเพชร
- ๔๒) นายอรรถพล นิยมวิทย์พันธ์
- ๔๓) นายภูวิช พรหมเสนา
- ๔๔) นายธนกร โคกาศพิพัฒน์
- ๔๕) นายชวฤทธิ์ วงษ์จันทร์
- ๔๖) นายอภิสิทธิ์ ศรีเสน
- ๔๗) นายเจตน์นรินทร์ คงศักดิ์ไทย
- ๔๘) นายจรูญ บุญชัย
- ๔๙) นายณณัติ เอนก
- ๕๐) นายอภิวัฒน์ ทุมพู
- ๕๑) นางสาวสุภาวรี งาม
- ๕๒) นางสาวพัชร ขวาลสมบุญ
- ๕๓) นางสาววิมล ปุณยเพ็ญ
- ๕๔) นางสาวกนกพร เข้มเพ็ชร
- ๕๕) นางสาวพัชรี หงษ์สมิต
- ๕๖) นางสาวกานิตา สุวงศ์ตระกูล
- ๕๗) นางสาวภาณุมาศ นามวัฒน์
- ๕๘) นางสาวอุไรรัตน์ ทิสรังแป้น
- ๕๙) นายธีรวัฒน์ ปังสุข
- ๖๐) นายอภิสิทธิ์ ยะโส
- ๖๑) นายประพนธ์ วรรณสุขชัย
- ๖๒) นายชยธร พวงทิพย์
- ๖๓) นางสาวกนกวรรณ จันทบาล
- ๖๔) นางสาวนภาพร หลีกบุญ
- ๖๕) นายสิทธิโชค ธงเงิน
- ๖๖) นางศศิวรรณ ใจบุญ
- ๖๗) นางสาวพรอนิศา ทุมคง
- ๖๘) นางสาวศรวิทย์ ยี่ดี
- ๖๙) นายบวรวิทย์ ศรีวิริยะ
- ๗๐) นายสุวิภา ทองอ่อน
- ๗๑) นายวิญญู บุญชนะ

- ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๓๓
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๓๔
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๓๕
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๓๖
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๓๗
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๓๘
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๓๙
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๔๐
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๔๑
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๔๒
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๔๓
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๔๔
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๔๕
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๔๖
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๔๗
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๔๘
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๔๙
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๕๐
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๕๑
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๕๒
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๕๓
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๕๔
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๕๕
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๕๖
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๕๗
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๕๘
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๕๙
ทะเบียนเลขที่ ๖-๒๐๔-๙-๕๐๔๖๐

(นายศิระ จันทน์นิล)

นักวิทยาศาสตร์ชำนาญการพิเศษ รักษาการแทน
ผู้อำนวยการกองวิจัยและเฝ้าระวังมลพิษโรงงาน
ปฏิบัติการตามหน้าที่และอำนาจตามกฎหมาย

๓๖) นายสมบุญ...

๓๖) นายสมบุรณ์ บุตรจันทร์
๓๗) นายวิรัตน์ ไชยธรรม
๓๘) นายอนุพันธ์ เพิ่มขุน
๓๙) นายจิรณัฐ ขาวละออ
๔๐) นายสมโภช วิสา
๔๑) นายอัสริ นามบุรี
๔๒) นายณัฐนันท์ ปานประเสริฐ
๔๓) นายอัครเดช จอัสว
๔๔) นายประเสริฐ สุระพันธ์
๔๕) นายบุญลือ จันทร์เนียม
๔๖) นายพิรพงษ์ ทองอุณปรีดา
๔๗) นายอนุพล ทองนุช
๔๘) นายอนุวัฒน์ ม่วงแพ
๔๙) นายเจตตราวุฒิ ปิตตะมะ
๕๐) นายคุณธนะ สายวรรณ
๕๑) นายพิชัย บุญยงค์
๕๒) นายภาณุพงศ์ โสมวงศ์
๕๓) นายสามารถ คุ้มปลี
๕๔) นายสัณฐิ์ โกศรีนาม
๕๕) นายณัฐวุฒิ ศรีประเสริฐ
๕๖) นายชวรัตน์ นาคพนม
๕๗) นายพงษ์กร จันทิพย์
๕๘) ว่าที่ร้อยตรี ภาณุพงศ์ แสนศรี
๕๙) นายสิทธิโชค ทาสีดา
๖๐) นายอนันต์ อินสุตา
๖๑) นางสาวกรรณิชา ขาดิวันชัย
๖๒) นางสาวพิมพ์ตะวัน มีนาถ
๖๓) นางสาวเพชรรัตน์ สิงห์สมบุญ
๖๔) นางสาวศุภณีย์ พรหมจันทร์
๖๕) นายกริณี พริ้ววษา
๖๖) นายจักริน นันท์วิชา
๖๗) นายจักรชัย สุขเปี้ย
๖๘) นายณรรณห์ ต๊ะทองคำ
๖๙) นายคุณพล สมนอก
๗๐) นายกฤษณ์ อุบลศรี
๗๑) นายอนันต์ นามะกุล
๗๒) นายอติพนธ์ บัวแดง

ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๓๔
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๓๕
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๓๖
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๓๗
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๓๘
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๓๙
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๔๐
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๔๑
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๔๒
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๔๓
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๔๔
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๔๕
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๔๖
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๔๗
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๔๘
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๔๙
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๕๐
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๕๑
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๕๒
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๕๓
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๕๔
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๕๕
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๕๖
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๕๗
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๕๘
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๕๙
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๖๐

(นายศิระ จันทร์เิก)

อธิบดีกรมการศึกษานานาชาติ
ผู้อำนวยการกองวิจัยและพัฒนาระบบ
ปฏิบัติการทางเทคโนโลยีสารสนเทศ

๑๐๔) นายณนพชัย...

๑๐๕) นายณนพชัย อุบลรัตน์
๑๐๖) นายณัฐพล คุณสุทธิ
๑๐๗) นายณัฐวัฒน์ สรร
๑๐๘) นายปิยะนัฐ พลมะศรี
๑๐๙) นายพณศักร โสมเขียว
๑๑๐) นายพิพัฒน์ กำคำ
๑๑๑) นายภาณุพงศ์ มาดิษฐ์
๑๑๒) นายมงคล ผลาพิชัย
๑๑๓) นายณัฐภัทร พูลศิริ
๑๑๔) นายสิริรัตน์ ทองอิน
๑๑๕) นายอนันต์ พันสมัย
๑๑๖) นายอติศักดิ์ ฝั่ม
๑๑๗) นายอนันต์ พงษ์
๑๑๘) นายณัฐวัฒน์ เกื้อทอง
๑๑๙) นายวราวุธ ติง
๑๒๐) นายแสงตะวัน นະะสิ
๑๒๑) นายสุทธพงศ์ รัตน
๑๒๒) นายชณัฐวิทย์ ไชยชนะ
๑๒๓) นายวิศรุต ศรีธรรมมา
๑๒๔) นายณนทกร เมื่อยทอง
๑๒๕) นายเกษิ์ สุทธ
๑๒๖) นางสาวณัฐกรณีย์ รักทะเล
๑๒๗) นางสาวประภากรณีย์ บุตรพร
๑๒๘) นางสาวณัฐวิทย์ นามพร
๑๒๙) นางสาวพัชรีนทร์ แสนสร้อย
๑๓๐) นายไพโรจน์ เปี่ยมพิมาย
๑๓๑) นางสาวศุภณีย์ ทองมาก
๑๓๒) นางสาวกิตติกา จิตสว่าง
๑๓๓) นางสาวชไมพร เล็กจิ๋ว
๑๓๔) นางสาวกัญญาพร คำแก้ว
๑๓๕) นางสาวสุกัญญา ภาณุ
๑๓๖) นางสาวกัญญาภา คุ้ม
๑๓๗) นางสาวไทรินทร์ ศรีวิ
๑๓๘) นางสาวกัทธนาพร หุสปัญญา
๑๓๙) นางสาวกิตติกา ปานทอง
๑๔๐) นางสาวอริสา ทองนวล
๑๔๑) นางสาวอรยา คำค้อม

ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๖๑
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๖๒
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๖๓
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๖๔
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๖๕
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๖๖
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๖๗
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๖๘
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๖๙
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๗๐
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๗๑
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๗๒
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๗๓
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๗๔
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๗๕
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๗๖
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๗๗
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๗๘
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๗๙
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๘๐
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๘๑
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๘๒
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๘๓
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๘๔
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๘๕
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๘๖
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๘๗
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๘๘
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๘๙
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๙๐

(นายศิระ จันทร์เิก)

อธิบดีกรมการศึกษานานาชาติ
ผู้อำนวยการกองวิจัยและพัฒนาระบบ
ปฏิบัติการทางเทคโนโลยีสารสนเทศ

๑๖๖) นางสาวสุภากรณ...

๑๖๗) นางสาวสุภากรณีย์ สุนทรสน
๑๖๘) นางสาวสุภากรณีย์ นนทบุรี
๑๖๙) นางสาวสุภากรณีย์ นนทบุรี
๑๗๐) นางสาวสุภากรณีย์ นนทบุรี
๑๗๑) นางสาวสุภากรณีย์ นนทบุรี
๑๗๒) นางสาวสุภากรณีย์ นนทบุรี
๑๗๓) นางสาวสุภากรณีย์ นนทบุรี
๑๗๔) นางสาวสุภากรณีย์ นนทบุรี
๑๗๕) นางสาวสุภากรณีย์ นนทบุรี
๑๗๖) นางสาวสุภากรณีย์ นนทบุรี
๑๗๗) นางสาวสุภากรณีย์ นนทบุรี
๑๗๘) นางสาวสุภากรณีย์ นนทบุรี
๑๗๙) นางสาวสุภากรณีย์ นนทบุรี
๑๘๐) นางสาวสุภากรณีย์ นนทบุรี
๑๘๑) นางสาวสุภากรณีย์ นนทบุรี
๑๘๒) นางสาวสุภากรณีย์ นนทบุรี
๑๘๓) นางสาวสุภากรณีย์ นนทบุรี
๑๘๔) นางสาวสุภากรณีย์ นนทบุรี
๑๘๕) นางสาวสุภากรณีย์ นนทบุรี
๑๘๖) นางสาวสุภากรณีย์ นนทบุรี
๑๘๗) นางสาวสุภากรณีย์ นนทบุรี
๑๘๘) นางสาวสุภากรณีย์ นนทบุรี
๑๘๙) นางสาวสุภากรณีย์ นนทบุรี
๑๙๐) นางสาวสุภากรณีย์ นนทบุรี

ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๙๑
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๙๒
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๙๓
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๙๔
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๙๕
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๙๖
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๙๗
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๙๘
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๕๙๙
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๐๐
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๐๑
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๐๒
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๐๓
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๐๔
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๐๕
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๐๖
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๐๗
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๐๘
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๐๙
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๑๐
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๑๑
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๑๒
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๑๓
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๑๔
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๑๕
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๑๖
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๑๗
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๑๘
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๑๙
ทะเบียนเลขที่ ๖-๒๐๔-๖-๕๖๒๐

(นายศิระ จันทร์เิก)

อธิบดีกรมการศึกษานานาชาติ
ผู้อำนวยการกองวิจัยและพัฒนาระบบ
ปฏิบัติการทางเทคโนโลยีสารสนเทศ

เอกสารแนบท้ายหนังสือรับรองการขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกสาร
บริษัท เอแอลเอส แล็บอราทอรี กรุ๊ป (ประเทศไทย) จำกัด เลขทะเบียน ๖-๒๐๔
ที่ ๑๐๔๑๑(๑) ๑๐๖๕ ลงวันที่ ๒๕ มกราคม ๒๕๖๕

ขอขยายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๒๖ รายการ

แนบท้าย จำนวน 59 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Aldicarb	High-Performance Liquid Chromatographic Method ⁽⁴⁾
2	Aldicarb Sulfone	High-Performance Liquid Chromatographic Method ⁽⁴⁾
3	Aldicarb Sulfide	High-Performance Liquid Chromatographic Method ⁽⁴⁾
4	Aldrin	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
5	Arsenic	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
6	Barium	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
7	α-BHC	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
8	β-BHC	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
9	γ-BHC	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
10	γ-BHC	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
11	Biochemical Oxygen Demand	1) 5-Day BOD Test, Azide Modification Method ⁽⁴⁾ 2) 5-Day BOD Test, Membrane Electrode Method ⁽⁴⁾
12	Carbaryl	High-Performance Liquid Chromatographic Method ⁽⁴⁾
13	Carbofuran	High-Performance Liquid Chromatographic Method ⁽⁴⁾
14	Cadmium	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
15	Chemical Oxygen Demand	1) Closed Reflux, Colorimetric Method ⁽⁴⁾ 2) Closed Reflux, Titrimetric Method ⁽⁴⁾
16	Chlordane	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
17	Chromium	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
18	Color	ADMI Weighted-Ordinate Spectrophotometric Method

(นายวิชาญธน นักรศกฤติ)

ผู้อำนวยการศูนย์มาตรฐานการวิเคราะห์เอกสาร
และระบบข้อมูลสารสนเทศ

19 Copper...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
19	Copper	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
20	Cyanide	Distillation, Colorimetric Method ⁽⁴⁾
21	2,4'-DDD	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
22	4,4'-DDD	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
23	2,4'-DDE	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
24	4,4'-DDE	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
25	2,4'-DDT	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
26	4,4'-DDT	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
27	Dieldrin	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
28	Endosulfan Sulfate	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
29	Endosulfan I	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
30	Endosulfan II	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
31	Endrin	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
32	Endrin Aldehyde	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
33	Formaldehyde	Distillation, Colorimetric Method ⁽⁴⁾
34	Free Chlorine	1) DPD Ferrous Titrimetric Method ⁽⁴⁾ 2) Iodometric Method ⁽⁴⁾
35	Heptachlor	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
36	Heptachlor epoxide	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
37	Hexavalent Chromium	Filtration, Colorimetric Method ⁽⁴⁾
38	3-Hydroxycarbofuran	High-Performance Liquid Chromatographic Method ⁽⁴⁾
39	Lead	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
40	Manganese	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
41	Mercury	1) Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/Mass spectrometric Method ⁽⁴⁾
42	Methiocarb	High-Performance Liquid Chromatographic Method ⁽⁴⁾
43	Methoxychlor	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾

(นางสาวกัญจน์ นัครสกุลวิไล)
ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
กรมควบคุมมลพิษ

44 Methomyl...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
44	Methomyl	High-Performance Liquid Chromatographic Method ⁽⁴⁾
45	Nickel	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
46	Oil & Grease	1) Liquid-Liquid, Partition-Gravimetric Method ⁽⁴⁾ 2) Soxhlet Extraction Method ⁽⁴⁾
47	Oxamyl	High-Performance Liquid Chromatographic Method ⁽⁴⁾
48	Propoxur	High-Performance Liquid Chromatographic Method ⁽⁴⁾
49	pH	Electrometric Method ⁽⁴⁾
50	Phenols	1) Distillation, Chloroform Extraction Method ⁽⁴⁾ 2) Distillation, Direct Photometric Method ⁽⁴⁾
51	Selenium	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
52	Sulfide	Iodometric Method ⁽⁴⁾
53	Temperature	Laboratory and Field Methods ⁽⁴⁾
54	Total Dissolved Solids	Dried at 180 °C ⁽⁴⁾
55	Total Kjeldahl Nitrogen	Semi-Micro Kjeldahl Method ⁽⁴⁾
56	Total Suspended Solids	Dried at 103-105 °C ⁽⁴⁾
57	Toxaphene	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
58	Trivalent Chromium	1) Digestion, Inductively Coupled Plasma Method; Colorimetric Method; Calculation ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method; Colorimetric Method; Calculation ⁽⁴⁾
59	Zinc	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ⁽⁴⁾

ไม่ได้ขึ้น จำนวน 126 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Acenaphthene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
2	Acetone	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾

(นางสาวกัญจน์ นัครสกุลวิไล)
ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
กรมควบคุมมลพิษ

3 Aldrin...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
3	Aldrin	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
4	Anthracene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
5	Antimony	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
6	Arsenic	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
7	Atrazine	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
8	Barium	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
9	Benz(a)anthracene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
10	Benzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
11	Benzo(b)fluoranthene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
12	Benzo(k)fluoranthene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
13	Benzoic Acid	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
14	Benzo(a)pyrene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
15	Benzo(g,h,i)perylene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
16	Beryllium	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
17	Bis(2-chloroethyl)ether	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾

(นางสาวกัญจน์ นัครสกุลวิไล)
ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
กรมควบคุมมลพิษ

18 Bis(2-ethylhexyl)phthalate...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
18	Bis(2-ethylhexyl)phthalate	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
19	Bromodichloromethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
20	Bromoform	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
21	Butanol	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
22	Butyl Benzyl Phthalate	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
23	Cadmium	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
24	Carbazole	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
25	Carbon Disulfide	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
26	Carbon tetrachloride	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
27	Chlordane	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
28	p-Chloroaniline	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
29	Chlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
30	Chlorodibromomethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
31	Chloroform	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
32	2-Chlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
33	Chromium	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾

(นางสาวกัญจน์ นัครสกุลวิไล)
ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
กรมควบคุมมลพิษ

34 Chromium (III)...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
34	Chromium (III)	1) Digestion, Inductively Coupled Plasma Method; Colorimetric Method; Calculation ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method; Colorimetric Method; Calculation ⁽⁴⁾
35	Chromium (VI)	Colorimetric Method ⁽⁴⁾
36	Chrysene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
37	Cyanide	Distillation, Colorimetric Method ⁽⁴⁾
38	2,4-D	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
39	DDD	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
40	DDE	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
41	DDT	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
42	Dibenz(a,h)anthracene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
43	Di-n-Butyl Phthalate	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
44	1,2-Dichlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
45	1,3-Dichlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
46	1,4-Dichlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
47	3,3-Dichlorobenzidine	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
48	1,1-Dichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
49	1,2-Dichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
50	1,1-Dichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾


 (นางจิราภรณ์ นัตถกุลวิไล)
 ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
 กรมควบคุมมลพิษ


51 cis-1,2-Dichloroethylene...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
51	cis-1,2-Dichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
52	trans-1,2-Dichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
53	2,4-Dichlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
54	1,2-Dichloropropane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
55	1,3-Dichloropropane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
56	1,3-Dichloropropene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
57	Dieldrin	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
58	Diethyl Phthalate	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
59	2,4-Dimethylphenol	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
60	2,4-Dinitrophenol	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
61	2,4-Dinitrotoluene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
62	2,6-Dinitrotoluene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
63	Di-n-Octyl Phthalate	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
64	Endosulfan	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
65	Endrin	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
66	Ethylbenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
67	Fluoranthene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾


 (นางจิราภรณ์ นัตถกุลวิไล)
 ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
 กรมควบคุมมลพิษ


68 Fluorene...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
68	Fluorene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
69	Heptachlor	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
70	Heptachlor epoxide	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
71	Hexachlorobenzene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
72	Hexachloro-1,3-butadiene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
73	n-Hexane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
74	α-HCH	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
75	β-HCH	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
76	γ-HCH	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
77	Hexachlorocyclopentadiene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
78	Hexachloroethane	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
79	Indeno(1,2,3-cd)pyrene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
80	Isophorone	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
81	Lead	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
82	Manganese	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
83	Mercury	1) Cold Vapor Atomic Absorption Spectrometric Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾


 (นางจิราภรณ์ นัตถกุลวิไล)
 ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
 กรมควบคุมมลพิษ

84 Methanol...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
84	Methanol	1) Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾ 2) Equilibrium Headspace, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
85	Methoxychlor	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
86	Methyl Bromide	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
87	Methylene Chloride	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
88	2-Methylphenol	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
89	2-Methylnaphthalene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
90	Methyl tert-Butyl Ether	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
91	Naphthalene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
92	Nickel	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
93	Nitrobenzene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
94	N-Nitrosodiphenylamine	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
95	N-Nitrosodi-n-Propylamine	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
96	Polychlorinated Biphenyls - PCB 1016 - PCB 1221 - PCB 1232 - PCB 1242 - PCB 1248 - PCB 1254 - PCB 1260	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾


 (นางจิราภรณ์ นัตถกุลวิไล)
 ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
 กรมควบคุมมลพิษ

97 Pentachlorophenol...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
97	Pentachlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
98	pH	Electrometric Method ⁽⁴⁾
99	Phenanthrene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
100	Phenol	1) Distillation, Direct Photometric Method ⁽⁴⁾ 2) Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
101	Pyrene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
102	Selenium	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
103	Silver	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
104	Styrene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
105	1,1,2,2-Tetrachloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
106	Tetrachloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
107	Toluene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
108	Toxaphene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
109	TPH (C ₅ -C ₆)	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,24)
110	TPH (C ₇ -C ₁₀)	Solvent Extraction, Gas Chromatographic Method ^(9,21)
111	TPH (C ₁₁ -C ₃₀)	Solvent Extraction, Gas Chromatographic Method ^(9,21)
112	1,2,4-Trichlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
113	1,1,1-Trichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾

วิมล
(นางวิภากร วัชรกุลกิจ)
ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์มลพิษ
และประเมินภัยพิบัติ

114 1,1,2-Trichloroethane...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
114	1,1,2-Trichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
115	Trichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
116	2,4,5-Trichlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
117	2,4,6-Trichlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
118	1,3,5-Trimethylbenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
119	Vanadium	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾
120	Vinyl Acetate	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
121	Vinyl Chloride	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
122	m-Xylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
123	o-Xylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
124	p-Xylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
125	Xylene (Total)	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
126	Zinc	1) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ⁽⁴⁾

จากผลเสีย (เบื้องต้น) จำนวน 16 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Antimony	Isokinetic, Digestion, Inductively Coupled Plasma Method ⁽⁵⁾
2	Arsenic	Isokinetic, Digestion, Inductively Coupled Plasma Method ⁽⁵⁾

วิมล
(นางวิภากร วัชรกุลกิจ)
ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์มลพิษ
และประเมินภัยพิบัติ

3 Carbon Monoxide...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
3	Carbon Monoxide	1) Sampling Bag Non-Dispersive Infrared Method ⁽⁵⁾ 2) Non-Dispersive Infrared Method ⁽⁵⁾ 3) Instrumental Analyzer Method ⁽⁵⁾
4	Chlorine	1) Absorption Sampling, Ion Chromatographic Method ⁽⁵⁾ 2) Isokinetic Sampling, Ion Chromatographic Method ⁽⁵⁾
5	Copper	Isokinetic, Digestion, Inductively Coupled Plasma Method ⁽⁵⁾
6	Dioxins	Isokinetic Sampling, Analysis by ISO/IEC 17025 Accredited Laboratory or Analysis by Department of Industrial Works Registered Laboratory (Dioxins/Furans Analysis Approved) ⁽⁵⁾
7	Hydrogen Chloride	1) Absorption Sampling, Ion Chromatographic Method ⁽⁵⁾ 2) Isokinetic Sampling, Ion Chromatographic Method ⁽⁵⁾
8	Hydrogen Sulfide	Absorption Sampling, Iodometric Method ⁽⁵⁾
9	Lead	Isokinetic, Digestion, Inductively Coupled Plasma Method ⁽⁵⁾
10	Mercury	1) Isokinetic Sampling, Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ⁽⁵⁾ 2) Isokinetic, Digestion, Inductively Coupled Plasma Method ⁽⁵⁾
11	Opacity	Ringelmann's Method ⁽⁴⁾
12	Oxides of Nitrogen	1) Absorption Sampling, Phenoldisulfonic Acid Method ⁽⁵⁾ 2) Chemiluminescence Method ⁽⁵⁾ 3) Instrumental Analyzer Method ⁽⁵⁾
13	Sulfur Dioxide	1) Absorption Sampling, Barium-Thorin Titrimetric Method ⁽⁵⁾ 2) UV Fluorescence Method ⁽⁵⁾ 3) Instrumental Analyzer Method ⁽⁵⁾
14	Sulfuric Acid	Isokinetic Sampling, Barium-Thorin Titrimetric Method ⁽⁵⁾
15	Total Suspended Particulate	Isokinetic Sampling, Gravimetric Method ⁽⁵⁾
16	Xylene	Absorption Sampling, Gas Chromatographic Method ⁽⁵⁾

วิมล
(นางวิภากร วัชรกุลกิจ)
ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์มลพิษ
และประเมินภัยพิบัติ

สิ่งปลูก...

สิ่งปลูกหรือวัตถุที่ไม่ใช่แล้ว จำนวน 35 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Aldrin	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^(1,5,25) 2) Soxhlet Extraction, Gas Chromatographic Method ^(16,22) 3) Automated Soxhlet Extraction, Gas Chromatographic Method ^(22,31)
2	Antimony	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,6,13) 2) Waste Extraction, Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(1,6,14) 3) Digestion, Inductively Coupled Plasma Method ^(7,13) 4) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ^(7,14)
3	Arsenic	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,6,15) 2) Waste Extraction, Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(1,6,16) 3) Digestion, Inductively Coupled Plasma Method ^(7,13) 4) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ^(7,14)
4	Barium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,6,15) 2) Waste Extraction, Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(1,6,16) 3) Digestion, Inductively Coupled Plasma Method ^(7,13) 4) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ^(7,14)
5	Beryllium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,6,15) 2) Waste Extraction, Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(1,6,16) 3) Digestion, Inductively Coupled Plasma Method ^(7,13) 4) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ^(7,14)

วิมล
(นางวิภากร วัชรกุลกิจ)
ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์มลพิษ
และประเมินภัยพิบัติ

6 Cadmium...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
6	Cadmium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1.6.15) 2) Waste Extraction, Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(1.6.16) 3) Digestion, Inductively Coupled Plasma Method ^(7.15) 4) Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(7.16)
7	Chlordane	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^(1.9.25) 2) Soxhlet Extraction, Gas Chromatographic Method ^(10.22) 3) Automated Soxhlet Extraction, Gas Chromatographic Method ^(22.31)
8	Chromium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1.6.15) 2) Waste Extraction, Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(1.6.16) 3) Digestion, Inductively Coupled Plasma Method ^(7.15) 4) Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(7.16)
9	Chromium (III)	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method; Waste Extraction, Colorimetric Method; Calculation Method ^(1.6.15,17) 2) Waste Extraction, Digestion, Inductively Coupled Plasma/Mass Spectrometric Method; Waste Extraction, Colorimetric Method; Calculation Method ^(1.6.16,17) 3) Digestion, Inductively Coupled Plasma Method; Alkaline Digestion, Colorimetric Method; Calculation Method ^(7.8,15,17) 4) Digestion, Inductively Coupled Plasma/Mass Spectrometric Method; Alkaline Digestion, Colorimetric Method; Calculation Method ^(7.8,16,17)
10	Chromium (VI)	1) Waste Extraction, Colorimetric Method ^(1.6.17) 2) Alkaline Digestion, Colorimetric Method ^(8.17)

(นางธิกาญจน์ นัครสกุลวิไล)
ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์สิ่งแวดล้อม
กรมส่งเสริมการค้าระหว่างประเทศ

11 Cobalt...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
11	Cobalt	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1.6.15) 2) Waste Extraction, Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(1.6.16) 3) Digestion, Inductively Coupled Plasma Method ^(7.15) 4) Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(7.16)
12	Copper	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1.6.15) 2) Waste Extraction, Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(1.6.16) 3) Digestion, Inductively Coupled Plasma Method ^(7.15) 4) Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(7.16)
13	2,4-D	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^(1.9.25) 2) Soxhlet Extraction, Gas Chromatographic Method ^(10.22) 3) Automated Soxhlet Extraction, Gas Chromatographic Method ^(22.31)
14	DDD	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^(1.9.25) 2) Soxhlet Extraction, Gas Chromatographic Method ^(10.22) 3) Automated Soxhlet Extraction, Gas Chromatographic Method ^(22.31)
15	DDE	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^(1.9.25) 2) Soxhlet Extraction, Gas Chromatographic Method ^(10.22) 3) Automated Soxhlet Extraction, Gas Chromatographic Method ^(22.31)
16	DDT	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^(1.9.25)

(นางธิกาญจน์ นัครสกุลวิไล)
ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์สิ่งแวดล้อม
กรมส่งเสริมการค้าระหว่างประเทศ

2) Soxhlet...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
17	Dieldrin	2) Soxhlet Extraction, Gas Chromatographic Method ^(10.22) 3) Automated Soxhlet Extraction, Gas Chromatographic Method ^(22.31) 1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^(1.9.25) 2) Soxhlet Extraction, Gas Chromatographic Method ^(10.22) 3) Automated Soxhlet Extraction, Gas Chromatographic Method ^(22.31)
18	Endrin	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^(1.9.25) 2) Soxhlet Extraction, Gas Chromatographic Method ^(10.22) 3) Automated Soxhlet Extraction, Gas Chromatographic Method ^(22.31)
19	Heptachlor	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^(1.9.25) 2) Soxhlet Extraction, Gas Chromatographic Method ^(10.22) 3) Automated Soxhlet Extraction, Gas Chromatographic Method ^(22.31)
20	Lead	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1.6.15) 2) Waste Extraction, Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(1.6.16) 3) Digestion, Inductively Coupled Plasma Method ^(7.15) 4) Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(7.16)
21	Lindane	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^(1.9.25) 2) Soxhlet Extraction, Gas Chromatographic Method ^(10.22) 3) Automated Soxhlet Extraction, Gas Chromatographic Method ^(22.31)
22	Mercury	1) Waste Extraction, Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^(1.6.18)

(นางธิกาญจน์ นัครสกุลวิไล)
ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์สิ่งแวดล้อม
กรมส่งเสริมการค้าระหว่างประเทศ

2) Waste Extraction...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
23	Methoxychlor	2) Waste Extraction, Thermal Decomposition Amalgamation and Atomic Absorption Spectrometric Method ^(1.6.19) 3) Waste Extraction, Digestion, Cold-Vapor Atomic Fluorescence Spectrometric Method ^(1.6.20) 4) Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ⁽¹⁸⁾ 5) Thermal Decomposition Amalgamation and Atomic Absorption Spectrometric Method ⁽¹⁹⁾ 6) Digestion, Cold-Vapor Atomic Fluorescence Spectrometric Method ⁽²⁰⁾
24	Mirex	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^(1.9.25) 2) Soxhlet Extraction, Gas Chromatographic Method ^(10.22) 3) Automated Soxhlet Extraction, Gas Chromatographic Method ^(22.31)
25	Molybdenum	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1.6.15) 2) Waste Extraction, Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(1.6.16) 3) Digestion, Inductively Coupled Plasma Method ^(7.15) 4) Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(7.16)
26	Nickel	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1.6.15) 2) Waste Extraction, Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(1.6.16) 3) Digestion, Inductively Coupled Plasma Method ^(7.15) 4) Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(7.16)

(นางธิกาญจน์ นัครสกุลวิไล)
ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์สิ่งแวดล้อม
กรมส่งเสริมการค้าระหว่างประเทศ

27 Polychlorinated...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
27	Polychlorinated biphenyls (PCBs) - Aroclor 1016 - Aroclor 1221 - Aroclor 1232 - Aroclor 1242 - Aroclor 1248 - Aroclor 1254 - Aroclor 1260 - 2-Chlorobiphenyl - 2,3-Dichlorobiphenyl - 2,2',5-Trichlorobiphenyl - 2,4',5-Trichlorobiphenyl - 2,2',3,5'-Tetrachlorobiphenyl - 2,2',5,5'-Tetrachlorobiphenyl - 2,3',4,4'-Tetrachlorobiphenyl - 2,2',3,4,5-Pentachlorobiphenyl - 2,2',4,5,5-Pentachlorobiphenyl - 2,3,3',4',6-Pentachlorobiphenyl - 2,2',3,4,4',5-Hexachlorobiphenyl - 2,2',3,4,5,5-Hexachlorobiphenyl - 2,2',3,5,5,6-Hexachlorobiphenyl - 2,2',4,4',5,5'-Hexachlorobiphenyl - 2,2',3,3',4,4',5-Heptachlorobiphenyl - 2,2',3,4,4',5,5'-Heptachlorobiphenyl - 2,2',3,4,4',5,6-Heptachlorobiphenyl - 2,2',3,4',5,5',6-Heptachlorobiphenyl - 2,2',3,3',4,4',5,5',6-Nonachlorobiphenyl	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^(1,5,23) 2) Soxhlet Extraction, Gas Chromatographic Method ^(10,23) 3) Automated Soxhlet Extraction, Gas Chromatographic Method ^(22,31)

(นางธิษฐาน นิตยกุลวิไล)
ผู้อำนวยการศูนย์มาตรฐานวิชาการและห้องปฏิบัติการ

28 Pentachlorophenol...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
28	Pentachlorophenol	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^(1,9,23) 2) Soxhlet Extraction, Gas Chromatographic Method ^(10,22) 3) Automated Soxhlet Extraction, Gas Chromatographic Method ^(22,31)
29	pH	Electrometric Method ^(29,30)
30	Selenium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,6,15) 2) Waste Extraction, Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(1,6,16) 3) Digestion, Inductively Coupled Plasma Method ^(7,15) 4) Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(7,16)
31	Silver	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,6,15) 2) Waste Extraction, Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(1,6,16)
32	Thallium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,6,15) 2) Waste Extraction, Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(1,6,16) 3) Digestion, Inductively Coupled Plasma Method ^(7,15) 4) Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(7,16)
33	Toxaphene	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^(1,9,23) 2) Soxhlet Extraction, Gas Chromatographic Method ^(10,22) 3) Automated Soxhlet Extraction, Gas Chromatographic Method ^(22,31)
34	Vanadium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,6,15) 2) Waste Extraction, Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(1,6,16) 3) Digestion, Inductively Coupled Plasma Method ^(7,15)

(นางธิษฐาน นิตยกุลวิไล)
ผู้อำนวยการศูนย์มาตรฐานวิชาการและห้องปฏิบัติการ

4) Digestion...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
35	Zinc	4) Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(7,16) 1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,6,15) 2) Waste Extraction, Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(1,6,16) 3) Digestion, Inductively Coupled Plasma Method ^(7,15) 4) Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(7,16)

คืน จำนวน 125 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Acenaphthene	Automated Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^(25,31)
2	Acetone	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^(14,24)
3	Aldrin	1) Soxhlet Extraction, Gas Chromatographic Method ^(10,22) 2) Automated Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^(25,31)
4	Anthracene	Automated Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^(25,31)
5	Antimony	1) Digestion, Inductively Coupled Plasma Method ^(7,15) 2) Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(7,16)
6	Arsenic	1) Digestion, Inductively Coupled Plasma Method ^(7,15) 2) Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(7,16)
7	Atrazine	1) Soxhlet Extraction, Gas Chromatographic Method ^(10,22) 2) Automated Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^(25,31)
8	Barium	1) Digestion, Inductively Coupled Plasma Method ^(7,15) 2) Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(7,16)

(นางธิษฐาน นิตยกุลวิไล)
ผู้อำนวยการศูนย์มาตรฐานวิชาการและห้องปฏิบัติการ

9 Benz(a)anthracene...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
9	Benz(a)anthracene	Automated Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^(25,31)
10	Benzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^(14,24)
11	Benzo(b)fluoranthene	Automated Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^(25,31)
12	Benzo(k)fluoranthene	Automated Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^(25,31)
13	Benzoic acid	Automated Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^(25,31)
14	Benzo(a)pyrene	Automated Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^(25,31)
15	Benzo(g,h,i)perylene	Automated Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^(25,31)
16	Beryllium	1) Digestion, Inductively Coupled Plasma Method ^(7,15) 2) Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(7,16)
17	Bis(2-chloroethyl)ether	Automated Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^(25,31)
18	Bis(2-ethylhexyl)phthalate	Automated Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^(25,31)
19	Bromodichloromethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^(14,24)
20	Bromoform	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^(14,24)
21	Butanol	Equilibrium Headspace, Gas Chromatographic/Mass Spectrometric Method ^(12,24)
22	Butyl Benzyl Phthalate	Automated Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^(25,31)
23	Cadmium	1) Digestion, Inductively Coupled Plasma Method ^(7,15) 2) Digestion, Inductively Coupled Plasma/Mass Spectrometric Method ^(7,16)
24	Carbazole	Automated Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^(25,31)
25	Carbon Disulfide	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^(14,24)

(นางธิษฐาน นิตยกุลวิไล)
ผู้อำนวยการศูนย์มาตรฐานวิชาการและห้องปฏิบัติการ

26 Carbon tetrachloride...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
26	Carbon tetrachloride	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
27	Chlordane	1) Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(10,22) 2) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
28	p-Chloroaniline	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
29	Chlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
30	Chlorodibromomethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
31	Chloroform	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
32	2-Chlorophenol	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
33	Chromium	1) Digestion, Inductively Coupled Plasma Method ^(7,15) 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ^(7,14)
34	Chromium (III)	1) Digestion, Inductively Coupled Plasma Method; Alkaline Digestion, Colorimetric Method; Calculation Method ^(7,8,13,17) 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method; Alkaline Digestion, Colorimetric Method; Calculation Method ^(7,8,16,17)
35	Chromium (VI)	Alkaline Digestion, Colorimetric Method ^(8,17)
36	Chrysene	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
37	Cyanide	Extraction, Distillation, Colorimetric Method ^(26,27,28)
38	2,4-D	1) Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(10,22) 2) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
39	DDD	1) Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(10,22) 2) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)

สำเนา
(นางวิภาดา ชัยกุลกิจ)

40 DDE...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
40	DDE	1) Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(10,22) 2) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
41	DDT	1) Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(10,22) 2) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
42	Dibenz(a,h)anthracene	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
43	Di-n-Butyl Phthalate	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
44	1,2-Dichlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
45	1,3-Dichlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
46	1,4-Dichlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
47	3,3-Dichlorobenzidine	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
48	1,1-Dichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
49	1,2-Dichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
50	1,1-Dichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
51	cis-1,2-Dichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
52	trans-1,2-Dichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
53	2,4-Dichlorophenol	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
54	1,2-Dichloropropane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
55	1,3-Dichloropropane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
56	1,3-Dichloropropene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)

สำเนา
(นางวิภาดา ชัยกุลกิจ)

57 Dieldrin...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
57	Dieldrin	1) Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(10,22) 2) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
58	Diethyl Phthalate	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
59	2,4-Dimethylphenol	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
60	2,4-Dinitrophenol	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
61	2,4-Dinitrotoluene	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
62	2,6-Dinitrotoluene	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
63	Di-n-Octyl Phthalate	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
64	Endosulfan	1) Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(10,22) 2) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
65	Endrin	1) Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(10,22) 2) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
66	Ethylbenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
67	Fluoranthene	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
68	Fluorene	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
69	Heptachlor	1) Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(10,22) 2) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
70	Heptachlor Epoxide	1) Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(10,22) 2) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)

สำเนา
(นางวิภาดา ชัยกุลกิจ)

71 Hexachlorobenzene...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
71	Hexachlorobenzene	1) Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(10,22) 2) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
72	Hexachloro-1,3-butadiene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
73	n-Hexane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
74	α-HCH	1) Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(10,22) 2) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
75	β-HCH	1) Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(10,22) 2) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
76	γ-HCH	1) Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(10,22) 2) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
77	Hexachlorocyclopentadiene	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
78	Hexachloroethane	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
79	Indeno(1,2,3-cd)pyrene	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
80	Isophorone	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
81	Lead	1) Digestion, Inductively Coupled Plasma Method ^(7,15) 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ^(7,14)
82	Manganese	1) Digestion, Inductively Coupled Plasma Method ^(7,15) 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ^(7,14)
83	Mercury	1) Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ⁽¹⁸⁾

สำเนา
(นางวิภาดา ชัยกุลกิจ)

2) Thermal...

ลำดับที่	สารเคมี	วิธีวิเคราะห์
84	Methanol	2) Thermal Decomposition, Amalgamation, and Atomic Absorption Spectrophotometry ⁽¹⁹⁾ 3) Digestion, Cold-Vapor Atomic Fluorescence Spectrometric Method ⁽²⁰⁾
85	Methoxychlor	Equilibrium Headspace, Gas Chromatographic/ Mass Spectrometric Method ^(12,24) 1) Soxhlet Extraction, Gas Chromatographic/ Method ^(10,22) 2) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
86	Methyl Bromide	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
87	Methylene Chloride	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
88	2-methylphenol	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
89	2-Methylnaphthalene	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
90	Methyl tert-Butyl Ether	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
91	Naphthalene	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
92	Nickel	1) Digestion, Inductively Coupled Plasma Method ^(7,15) 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ^(7,16)
93	Nitrobenzene	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
94	N-Nitrosodiphenylamine	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
95	N-Nitrosodi-n-propylamine	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
96	Polychlorinated biphenyls (PCBs) - Aroclor 1016 - Aroclor 1221 - Aroclor 1232	1) Soxhlet Extraction, Gas Chromatographic Method ^(10,23) 2) Automated Soxhlet Extraction, Gas Chromatographic Method ^(23,32)

วิบูลย์
(นางวิภาดา จิตร์สกุลกิจ)
ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์ทดสอบผลิตภัณฑ์

- Aroclor 1242...

ลำดับที่	สารเคมี	วิธีวิเคราะห์
	- Aroclor 1242 - Aroclor 1248 - Aroclor 1254 - Aroclor 1260 - 2-Chlorobiphenyl - 2,2',3,5'-Tetrachlorobiphenyl - 2,2',5,5'-Tetrachlorobiphenyl - 2,3',4,4'-Tetrachlorobiphenyl - 2,2',3,4,5'-Pentachlorobiphenyl - 2,2',4,5,5'-Pentachlorobiphenyl - 2,3',3',4,6'-Pentachlorobiphenyl - 2,2',3,4,4',5'-Hexachlorobiphenyl - 2,2',3,4,5,5'-Hexachlorobiphenyl - 2,2',3,5,5',6'-Hexachlorobiphenyl - 2,2',4,4',5,5'-Hexachlorobiphenyl - 2,2',3,3',4,4',5'-Heptachlorobiphenyl - 2,2',3,4,4',5,5'-Heptachlorobiphenyl - 2,2',3,4,4',5,6'-Heptachlorobiphenyl - 2,2',3,4,4',5,6'-Heptachlorobiphenyl - 2,2',3,3',4,4',5,6'-Nonachlorobiphenyl - Pentachlorophenol	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
97	Phenanthrene	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
98	Phenol	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
99	Pyrene	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
100		

วิบูลย์
(นางวิภาดา จิตร์สกุลกิจ)
ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์ทดสอบผลิตภัณฑ์
และทะเบียนผลิตภัณฑ์

101 Selenium...

ลำดับที่	สารเคมี	วิธีวิเคราะห์
101	Selenium	1) Digestion, Inductively Coupled Plasma Method ^(7,15) 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ^(7,16)
102	Silver	1) Digestion, Inductively Coupled Plasma Method ^(7,15) 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ^(7,16)
103	Styrene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
104	1,1,2,2-Tetrachloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
105	Tetrachloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
106	Toluene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
107	Toxaphene	1) Soxhlet Extraction, Gas Chromatographic Method ^(10,22) 2) Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
108	TPH (C ₅ -C ₆)	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
109	TPH (C ₅ -C ₁₀)	1) Solvent Extraction, Gas Chromatographic Method ^(11,21) 2) Automated Soxhlet Extraction, Gas Chromatographic Method ^(23,31)
110	TPH (C ₁₀ -C ₃₃)	1) Solvent Extraction, Gas Chromatographic Method ^(11,21) 2) Automated Soxhlet Extraction, Gas Chromatographic Method ^(23,31)
111	1,2,4-Trichlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
112	1,1,1-Trichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
113	1,1,2-Trichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
114	Trichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
115	2,4,5-Trichlorophenol	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)

วิบูลย์
(นางวิภาดา จิตร์สกุลกิจ)
ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์ทดสอบผลิตภัณฑ์

116 2,4,6-Trichlorophenol...

ลำดับที่	สารเคมี	วิธีวิเคราะห์
116	2,4,6-Trichlorophenol	Automated Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(25,31)
117	1,3,5-Trimethylbenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
118	Vanadium	1) Digestion, Inductively Coupled Plasma Method ^(7,15) 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ^(7,16)
119	Vinyl Acetate	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
120	Vinyl Chloride	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
121	m-Xylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
122	o-Xylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
123	p-Xylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
124	Xylene (Total)	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(14,24)
125	Zinc	1) Digestion, Inductively Coupled Plasma Method ^(7,15) 2) Digestion, Inductively Coupled Plasma/ Mass Spectrometric Method ^(7,16)

เอกสารอ้างอิง

- กระทรวงอุตสาหกรรม. ประกาศกระทรวงอุตสาหกรรม, พ.ศ. 2548. เรื่อง การกำหนดสิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว. ราชกิจจานุเบกษา. 25 มกราคม 2549. เล่มที่ 123 ตอนพิเศษ 113.
- กระทรวงอุตสาหกรรม. ประกาศกระทรวงอุตสาหกรรม, พ.ศ. 2549. เรื่อง กำหนดค่าปริมาณเขม่าควันที่เจือปนในอากาศที่ระบายออกจากรถยนต์ของรถยนต์ที่จดทะเบียนเป็นเชื้อเพลิง. ราชกิจจานุเบกษา. 4 ธันวาคม 2549. เล่มที่ 123 ตอนพิเศษ 125.
- สมาคมวิศวกรรมสิ่งแวดล้อมแห่งประเทศไทย. คู่มือวิเคราะห์น้ำเสีย. พิมพ์ครั้งที่ 4. กรุงเทพฯ: เรือนแก้วการพิมพ์, 2547.
- APHA, AWWA, WEF. Standard Methods for the Examination of Water and Wastewater. 23rd ed. Washington, DC: APHA, 2017.
- United States Environmental Protection Agency. Standards of Performance for New Stationary Sources. 40 CFR 60. Appendix A, 2019.
- United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. SW-846, 1997.

วิบูลย์
(นางวิภาดา จิตร์สกุลกิจ)
ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์ทดสอบผลิตภัณฑ์

7. United States...

7. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Acid Digestion of Sludges and Sediments and Soils. SW-846 Method 3050B, 1996.
8. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Alkaline Digestion for Hexavalent Chromium. SW-846 Method 3060A, 1996.
9. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Separatory Funnel Liquid-Liquid Extraction. SW-846 Method 3510C, 1996.
10. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Soxhlet Extraction. SW-846 Method 3540C, 1996.
11. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Microscale Solvent Extraction (MSE). SW-846 Method 3570, 2002.
12. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Volatile Organic Compounds (VOCs) in Various Sample Matrices Using Equilibrium Headspace Analysis. SW-846 Method 5021A, 2014.
13. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Purge-and-Trap for Aqueous Samples. SW-846 Method 5030B, 1996.
14. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Closed-System Purge-and-Trap and Extraction for Volatile Organics in Soil and Waste Samples. SW-846 Method 5035, 1996.
15. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Inductively Coupled Plasma- Atomic Emission Spectrometry. SW-846 Method 6010B, 1996.
16. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Inductively Coupled Plasma-Mass Spectrometry. SW-846 Method 6020A, 2007.
17. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Chromium, Hexavalent (Colorimetric). SW-846 Method 7196A, 1992.
18. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Mercury in Solid or Semisolid Waste (Manual Cold-Vapor Technique). SW-846 Method 7471B, 2007.
19. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Mercury in Solids and Solutions by Thermal Decomposition, Amalgamation, and Atomic Absorption Spectrophotometry. SW-846 Method 7473, 2007.

(นางจิราภรณ์ อัครกุลศิริ)

ผู้อำนวยการศูนย์บริการวิชาการและพัฒนาระบบข้อมูล
มหาวิทยาลัยเทคโนโลยีสุรนารี

20. United States...

20. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Mercury in Sediment and Tissue Sample by Atomic Fluorescence Spectrometry. SW-846 Method 7474, 2007.
21. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Nonhalogenated Organics Using GC/FID. SW-846 Method 8015B, 1996.
22. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Organochlorine Pesticides by Gas Chromatography. SW-846 Method 8081B, 2007.
23. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Polychlorinated Biphenyls (PCBs) by Gas Chromatography. SW-846 Method 8082, 1996.
24. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Volatile Organic Compounds by Gas Chromatography/Mass Spectrometry (GC/MS). SW-846 Method 8260D, 2018.
25. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Semivolatile Organic Compounds by Gas Chromatography/Mass Spectrometry (GC/MS). SW-846 Method 8270E, 2018.
26. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Total and Amenable Cyanide: Distillation SW-846 Method 9010B, 1996.
27. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Cyanide Extraction Procedure for Solids and Oil. SW-846 Method 9013A, 1996.
28. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Cyanide in Waters and Extracts Using Titrimetric and Manual Spectrophotometric Procedures. SW-846 Method 9014, 2014.
29. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. pH Electrometric Measurement. SW-846 Method 9040C, 2004.
30. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Soil and Waste pH. SW-846 Method 9045D, 2004.
31. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Automated Soxhlet Extraction. SW-846 Method 3541, 1994.

(นางจิราภรณ์ อัครกุลศิริ)

ผู้อำนวยการศูนย์บริการวิชาการและพัฒนาระบบข้อมูล
มหาวิทยาลัยเทคโนโลยีสุรนารี

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบและประเมินผลปฏิบัติการ การวิจัยและพัฒนาระบบข้อมูล การบริการวิชาการและพัฒนาระบบข้อมูล โทร. ๐ ๒๒๒๒ ๔๐๐๖, ๔๐๐๖



ที่ บก ๐๓๓๐(๓)/ ๒๕๖๐

กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท
เขตราชเทวี กรุงเทพฯ ๑๐๕๐๐

๒๘ มิถุนายน ๒๕๖๕

เรื่อง ขันทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

เรียน กรรมการผู้จัดการ บริษัท เอแอลเอส แลบบอราทอรี กรุ๊ป (ประเทศไทย) จำกัด

อ้างถึง คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และขอขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
ลงวันที่ ๒๕ เมษายน ๒๕๖๔

สิ่งที่ส่งมาด้วย เอกสารแนบท้ายหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
บริษัท เอแอลเอส แลบบอราทอรี กรุ๊ป (ประเทศไทย) จำกัด จำนวน ๒ แผ่น

ตามหนังสือที่อ้างถึง บริษัท เอแอลเอส แลบบอราทอรี กรุ๊ป (ประเทศไทย) จำกัด ขอขึ้นทะเบียน
ห้องปฏิบัติการวิเคราะห์เอกชน พร้อมรายชื่อผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ เจ้าหน้าที่ประจำ
ห้องปฏิบัติการวิเคราะห์ และรายการสารเคมีที่จะทำการวิเคราะห์ ต่อกรมโรงงานอุตสาหกรรม นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว ให้บริษัท เอแอลเอส แลบบอราทอรี กรุ๊ป
(ประเทศไทย) จำกัด ขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน มีเลขทะเบียน ๖-๒๒๒๓-๖-๒๕๖๕
๖๒๖/๑๐ หมู่ที่ ๕ ตำบลแม่ไม้ อำเภอปลวกแดง จังหวัดระยอง โดยมีองค์ประกอบดังนี้

ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์

- | | |
|--------------------------|-----------------------------|
| ๑) นายเดช ช้างชน | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๒) นางวิลาวัลย์ บริรักษ์ | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๓) นายสุพจน์ สลามเต๊ะ | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |

ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์

- | | |
|-----------------------------|-----------------------------|
| ๑) นางสาวณัฏฐ์ บรรจงกิจ | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๒) นางพจนา สิตา | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๓) นางสาวนิตยา กุลสุริวงค์ | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๔) นายพิทยา ทองแดง | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๕) นางชลธิชา สูงสงข | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๖) ว่าที่ ร.ต.ณชัย ม่วงมา | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๗) นายวรวิทย์ พันพา | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๘) นายศักดิ์วิวัฒน์ จรัสกาย | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๙) นายสุรศักดิ์ สาธิน | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๑๐) นางสาวพรพรรณ ภาณุตานนท์ | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๑๑) นายสถาพร ถาวแก้ว | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๑๒) นายสุพจน์ดำรง โชคบัณฑิต | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |

๑๓) นายวิมล...


- | | |
|--------------------------|-----------------------------|
| ๑๓) นายวิมล... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๑๔) นางสาวนิตยา บรรจงกิจ | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๑๕) นางสาวนิตยา บรรจงกิจ | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๑๖) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๑๗) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๑๘) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๑๙) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๒๐) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๒๑) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๒๒) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๒๓) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๒๔) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๒๕) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๒๖) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๒๗) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๒๘) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๒๙) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๓๐) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๓๑) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๓๒) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๓๓) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๓๔) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๓๕) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๓๖) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๓๗) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๓๘) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๓๙) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๔๐) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๔๑) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๔๒) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๔๓) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |
| ๔๔) นายธนกร... | ทะเบียนเลขที่ ๖-๒๒๒๓-๖-๒๕๖๕ |

ค. ขอบข่ายสารเคมีที่ได้รับขึ้นทะเบียนให้วิเคราะห์ที่น้ำเสีย จำนวน ๑๔ รายการ
อากาศเสีย (ปล่องระบาย) จำนวน ๘ รายการ และน้ำใต้ดิน จำนวน ๓ รายการ รวมทั้งสิ้นจำนวน ๒๕ รายการ
ตามสิ่งที่ส่งมาด้วย

หนังสือฉบับนี้มีอายุ ๓ ปี นับจากวันที่กรมโรงงานอุตสาหกรรมออกหนังสือ หากประสงค์จะต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ให้ยื่นคำขอต่ออายุพร้อมเอกสารประกอบคำขอต่อกรมโรงงานอุตสาหกรรมภายใน ๓๐ วัน ก่อนวันสิ้นอายุของหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ซึ่งคำขอต่ออายุดังกล่าวขอรับได้ที่กรมโรงงานอุตสาหกรรม

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ


(นางจินดา เศษะศรีจันทร์)
ผู้อำนวยการกองประเมินผลสิ่งแวดล้อม
ผู้ตรวจราชการแผ่นดินกระทรวงอุตสาหกรรม

๒๘ มิ.ย. ๒๕๖๕

กองวิจัยและเตือนภัยมลพิษโรงงาน
ศูนย์วิจัยและเตือนภัยมลพิษโรงงานภาคตะวันออก
โทร. ๐ ๒๘๐๕ ๗๐๖๑-๓
ไปรษณีย์อิเล็กทรอนิกส์ envw@dlw.mail.go.th


เอกสารแนบท้ายหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
บริษัท เอนแอลเอส แลบบอราทอรี กรุ๊ป (ประเทศไทย) จำกัด เลขทะเบียน ว-๗๒๒๓
ที่ อก ๐๓๑๐(๓)/ ๒๕๗๐ ลงวันที่ ๒๘ มิถุนายน ๒๕๖๕

ขอขยายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๒๔ รายการ
น้ำเสีย จำนวน 14 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Biochemical Oxygen Demand	1) 5-Day BOD Test, Membrane Electrode Method ⁽²⁾ 2) 5-Day BOD Test, Azide Modification Method ⁽²⁾
2	Chemical Oxygen Demand	1) Open Reflux, Titrimetric Method ⁽²⁾ 2) Closed Reflux, Colorimetric Method ⁽²⁾ 3) Closed Reflux, Titrimetric Method ⁽²⁾
3	Color	ADMI Weighted - Ordinate Spectrophotometric Method ⁽²⁾
4	Cyanide	Distillation, Colorimetric Method ⁽²⁾
5	Formaldehyde	Distillation, Colorimetric Method ⁽¹⁾
6	Free Chlorine	DPD-Ferrous Titrimetric Method ⁽²⁾
7	Oil and Grease	Liquid-Liquid Partition-Gravimetric Method ⁽²⁾
8	pH	Electrometric Method ⁽²⁾
9	Phenols	1) Distillation, Chloroform Extraction Method ⁽²⁾ 2) Distillation, Direct Photometric Method ⁽²⁾
10	Sulfide	ZnS Precipitation, Iodometric Method ⁽²⁾
11	Temperature	Laboratory and Field Method ⁽²⁾
12	Total Dissolved Solids	Dried at 180 °C ⁽²⁾
13	Total Kjeldahl Nitrogen	Semi-Micro Kjeldahl Method ⁽²⁾
14	Total Suspended Solids	Dried at 103-105 °C ⁽²⁾

อากาศเสีย (ปล่อยระบาย) จำนวน 7 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Carbon Monoxide	1) Sampling Bag, Non-Dispersive Infrared Method ⁽⁵⁾ 2) Instrumental Analyzer Method ⁽⁶⁾
2	Hydrogen Sulfide	Absorption Sampling, Iodometric Method ⁽⁵⁾
3	Opacity	Ringelmann's Method ^(3,4)
4	Oxide of Nitrogen	1) Absorption Sampling, Phenoldisulfonic Acid Method ⁽⁶⁾ 2) Instrumental Analyzer Method ⁽⁶⁾
5	Sulfur Dioxide	1) Absorption Sampling, Barium-Thorin Titrimetric Method ⁽⁵⁾ 2) Instrumental Analyzer Method ⁽⁶⁾


(นางสาวจินดา เศษะศรีจันทร์)
ผู้อำนวยการ
ศูนย์วิจัยและเตือนภัยมลพิษโรงงานภาคตะวันออก Sulfuric Acid...


ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
6	Sulfuric Acid	Isokinetic Sampling, Barium - Thorin Titrimetric Method ⁽⁶⁾
7	Total Suspended Particulate	Isokinetic Sampling, Gravimetric Method ⁽⁷⁾

น้ำใต้ดิน จำนวน 3 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Cyanide	Distillation, Colorimetric Method ⁽²⁾
2	pH	Electrometric Method ⁽²⁾
3	Phenols	Distillation, Direct Photometric Method ⁽²⁾

เอกสารอ้างอิง

1. ธงชัย พรหมสวัสดิ์ และวิบูลย์ลักษณ์ วิสุทธิศักดิ์, บรรณาธิการ. (2547) คู่มือวิเคราะห์น้ำเสีย. พิมพ์ครั้งที่ 4. กรุงเทพฯ: สมาคมวิศวกรรมสิ่งแวดล้อมแห่งประเทศไทย.
2. APHA, AWWA, WEF. Standard Methods for the Examination of Water and Wastewater. 23rd ed. Washington, DC : APHA, 2017
3. กระทรวงอุตสาหกรรม. ประกาศกระทรวงอุตสาหกรรม, พ.ศ. 2549. เรื่อง กำหนดค่าปริมาณเขม่าควันที่เจือปนในอากาศที่ระบายจากปล่องของหม้อน้ำโรงสีข้าวที่ใช้แก๊สเป็นเชื้อเพลิง.ราชกิจจานุเบกษา. 4 ธันวาคม 2549. เล่มที่ 123 ตอนพิเศษ 125ง.
4. กระทรวงอุตสาหกรรม. ประกาศกระทรวงอุตสาหกรรม, พ.ศ. 2549. เรื่อง กำหนดค่าปริมาณเขม่าควันที่เจือปนในอากาศที่ระบายจากปล่องของหม้อน้ำของโรงงาน.ราชกิจจานุเบกษา. 4 ธันวาคม 2549. เล่มที่ 123 ตอนพิเศษ 125ง.
5. United States Environmental Protection. Standards of Performance for New Stationary Sources. 40 CFR 60. Appendix A, 2017.
6. United States Environmental Protection. Standards of Performance for New Stationary Sources. 40 CFR 60. Appendix A, 2019.
7. United States Environmental Protection. Standards of Performance for New Stationary Sources. 40 CFR 60. Appendix A, 2020.
8. United States Environmental Protection Agency. Determination of Carbon Monoxide Emissions from Stationary Sources; Instrumental Analyzer Procedure. 40 CFR 60. Appendix A Method 10, 2017.
9. United States Environmental Protection Agency. Determination of Oxide of Nitrogen Emissions from Stationary Sources; Instrumental Analyzer Procedure. 40 CFR 60. Appendix A Method 7E, 2019.
10. United States Environmental Protection Agency. Determination of Sulfur Dioxide Emissions from Stationary Sources; Instrumental Analyzer Procedure. 40 CFR 60. Appendix A Method 6C, 2017.


(นางสาวจินดา เศษะศรีจันทร์)
ผู้อำนวยการ

ศูนย์วิจัยและเตือนภัยมลพิษโรงงานภาคตะวันออก

ศูนย์วิจัยและเตือนภัยมลพิษโรงงานภาคตะวันออก กองวิจัยและเตือนภัยมลพิษโรงงาน กรมโรงงานอุตสาหกรรม โทร ๐ ๒๘๐๕ ๗๐๖๑-๓



จัดทำโดย บริษัท เอแอลเอส แลบบอราทอรี กรุ๊ป (ประเทศไทย) จำกัด

104 ซอยพัฒนาการ 40 ถนนพัฒนาการ แขวงพัฒนาการ เขตสวนหลวง กรุงเทพฯ 10250

โทรศัพท์ 0-2760-3000 โทรสาร 0-2760-3197

www.alsglobal.com