

ภาคผนวก 14

เอกสารขึ้นทะเบียนห้องปฏิบัติการ
และเอกสารสอบเทียบเครื่องมือ



๐๔ พฤศจิกายน ๒๕๖๔

เรื่อง ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

เรียน กรรมการผู้จัดการ บริษัท โอกลา เทสต์ติ้ง แอนด์ คอนซัลติ้ง เซอร์วิส จำกัด

อ้างถึง คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และชนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน
ลงวันที่ ๑๒ กรกฎาคม ๒๕๖๔

สิ่งที่ส่งมาด้วย เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
บริษัท โอกลา เทสต์ติ้ง แอนด์ คอนซัลติ้ง เซอร์วิส จำกัด จำนวน ๑ แผ่น

ตามหนังสือที่อ้างถึง บริษัท โอกลา เทสต์ติ้ง แอนด์ คอนซัลติ้ง เซอร์วิส จำกัด ขอต่ออายุหนังสือ
รับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน เลขทะเบียน ว-๒๑๙ สถานที่ตั้งเลขที่ ๖๓/๑๓ ซอยเพชรเกษม ๗
แขวงวัดท่าพระ เขตบางกอกใหญ่ กรุงเทพมหานคร ต่อกรมโรงงานอุตสาหกรรม นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว ให้บริษัท โอกลา เทสต์ติ้ง แอนด์ คอนซัลติ้ง เซอร์วิส จำกัด
ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน โดยมีองค์ประกอบดังนี้

ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์

- | | |
|-----------------------------|----------------------------|
| ๑) นายธวัชชัย จงวุฒิชัย | ทะเบียนเลขที่ ว-๒๑๙-ค-๕๑๒๔ |
| ๒) นางสาวปนัดดา พันธุ์กะจับ | ทะเบียนเลขที่ ว-๒๑๙-ค-๖๖๙๙ |
| ๓) นางสาวจามจุรี คำปุย | ทะเบียนเลขที่ ว-๒๑๙-ค-๙๖๖๓ |

ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์

- | | |
|--|----------------------------|
| ๑) นางสาวธัญชนก ขำขุน | ทะเบียนเลขที่ ว-๒๑๙-จ-๙๔๑๖ |
| ๒) ว่าที่ร้อยตรีหญิงสาวตรี เวียงจันทร์ | ทะเบียนเลขที่ ว-๒๑๙-จ-๙๔๑๗ |
| ๓) นางสาวภาณุชนารถ เชี่ยวชาญ | ทะเบียนเลขที่ ว-๒๑๙-จ-๙๔๑๘ |
| ๔) นางสาววันวิสา หวังแวกลาง | ทะเบียนเลขที่ ว-๒๑๙-จ-๙๔๑๙ |
| ๕) นางสาวธิดารัตน์ กลัดตลาด | ทะเบียนเลขที่ ว-๒๑๙-จ-๙๔๒๐ |
| ๖) นางสาวรัตตชา ศรีปราสาท | ทะเบียนเลขที่ ว-๒๑๙-จ-๙๔๒๑ |
| ๗) นางสาวแพรวพรรณ กองกะแซง | ทะเบียนเลขที่ ว-๒๑๙-จ-๙๔๒๒ |
| ๘) นางสาวจุลภา สมบุญ | ทะเบียนเลขที่ ว-๒๑๙-จ-๙๔๒๓ |
| ๙) นางสาวนิจินา มะติยาภักดิ์ | ทะเบียนเลขที่ ว-๒๑๙-จ-๙๔๒๔ |
| ๑๐) นางสาวเบญจพร อินแก้ว | ทะเบียนเลขที่ ว-๒๑๙-จ-๙๖๖๔ |
| ๑๑) นายธนทัต เวชกิจ | ทะเบียนเลขที่ ว-๒๑๙-จ-๙๖๖๕ |
| ๑๒) นายปริญญา กล้าน้อย | ทะเบียนเลขที่ ว-๒๑๙-จ-๙๖๖๖ |

ค. ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนให้วิเคราะห์ในน้ำเสีย จำนวน ๙ รายการ และ
อากาศเสีย จำนวน ๕ รายการ รวมทั้งสิ้น ๑๔ รายการ ตามสิ่งที่ส่งมาด้วย

หนังสือฉบับนี้จะหมดอายุในวันที่ ๑๕ สิงหาคม ๒๕๖๗ หากประสงค์จะต่ออายุหนังสือ
รับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ให้ยื่นคำขอต่ออายุพร้อมเอกสารประกอบคำขอต่อ
กรมโรงงานอุตสาหกรรมภายใน ๓๐ วัน ก่อนวันสิ้นอายุของหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
ซึ่งคำขอต่ออายุดังกล่าวขอรับได้ที่กรมโรงงานอุตสาหกรรม

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ

(นางจินดา เตชะศรีนทร์)

ผู้อำนวยการกองวิจัยและเตือนภัยมลพิษโรงงาน
ปฏิบัติราชการแทนอธิบดีกรมโรงงานอุตสาหกรรม

กองวิจัยและเตือนภัยมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๐๔-๖

โทรสาร ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๙๙

ไปรษณีย์อิเล็กทรอนิกส์ saraban@diw.gmail.go.th

เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
บริษัท โอกลา เทสติ้ง แอนด์ คอนซัลติ้ง เซอร์วิส จำกัด เลขทะเบียน ว-๒๑๙
ที่ อก ๐๓๑๐(๑)/ ๑๐๔๓๑ ลงวันที่ ๐๔ พฤศจิกายน ๒๕๖๔

ขอขยายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๑๔ รายการ

น้ำเสีย จำนวน 9 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Biochemical Oxygen Demand	1) 5-Day BOD Test, Azide Modification Method ^[3] 2) 5-Day BOD Test, Membrane Electrode Method ^[3]
2	Free Chlorine	Iodometric Method ^[3]
3	Oil & Grease	Liquid-Liquid, Partition-Gravimetric Method ^[3]
4	pH	Electrometric Method ^[3]
5	Sulfide	Iodometric Method ^[3]
6	Temperature	Laboratory and Field Methods ^[3]
7	Total Dissolved Solids	Dried at 180 °C ^[3]
8	Total Kjeldahl Nitrogen	Macro-Kjeldahl Method ^[3]
9	Total Suspended Solids	Dried at 103-105 °C ^[3]

อากาศเสีย (ปล่องระบาย) จำนวน 5 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Carbon Monoxide	Instrumental Analyzer ^[4]
2	Opacity	Ringelmann's Method ^[1,2]
3	Oxides of Nitrogen	Instrumental Analyzer ^[4]
4	Sulfur Dioxide	Instrumental Analyzer ^[4]
5	Total Suspended Particulate	Isokinetic Sampling, Gravimetric Method ^[4]

เอกสารอ้างอิง

1. กระทรวงอุตสาหกรรม. ประกาศกระทรวงอุตสาหกรรม, พ.ศ. 2549. เรื่อง กำหนดค่าปริมาณ
เขม่าควันที่เจือปนในอากาศที่ระบายออกจากปล่องของหม้อน้ำโรงสีข้าวที่ใช้กลบเป็นเชื้อเพลิง.

ราชกิจจานุเบกษา. 4 ธันวาคม 2549. เล่มที่ 123 ตอนพิเศษ 125 ง.

2. กระทรวงอุตสาหกรรม. ประกาศกระทรวงอุตสาหกรรม, พ.ศ. 2549. เรื่อง กำหนดค่าปริมาณ
เขม่าควันที่เจือปนในอากาศที่ระบายออกจากปล่องของหม้อน้ำโรงงาน. ราชกิจจานุเบกษา. 4 ธันวาคม 2549.
เล่มที่ 123 ตอนพิเศษ 125 ง.

3. APHA, AWWA, WEF. **Standard Methods for the Examination of Water and Wastewater**. 23rd ed. Washington, DC: APHA, 2017.

4. United States Environmental Protection Agency. **Standards of Performance for New Stationary Sources**. 40 CFR 60. Appendix A, 2018.

(นางริกาญจน์ จัตรสกุลวิไล)

ผู้อำนวยการกลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
และทะเบียนห้องปฏิบัติการ



TSP Calibration Report

Location: OKLA-Testing Lab

Date: 6/5/2022

Sampler: EM-TSP-02

Serial No:

Barometric pressure, mm Hg (Pa): 760

Temperature, Deg C (Ta): 33

Transfer Standard Type: Tisch TE 5025A

Serial No: 1758

Last Calibration Date: 2-May-19

Operator: Mr.Parinya

Qstd Slope: 2.03370

Qstd Intercept: -0.03624


Plate No	H2O (in)	Qstd		I (Chart)	IC (corrected)
		(m3/min)	(cfm)		
1	9.60	1.540	54.375	50.00	49.34
2	7.60	1.374	48.521	44.00	43.42
3	6.40	1.264	44.631	40.00	39.47
4	4.00	1.007	35.552	31.00	30.59
5	2.00	0.722	25.514	22.00	21.71

Linear Regression

Slope: 0.9589

Intecept: -3.1000

Corr. Coeff: 0.9996



CALIBRATION BY :

DATE :

6/5/2022

APPROVED BY :

DATE :

6/5/22

ต้องการข้อมูลทางด้านเทคนิคเพิ่มเติม : นายปริญญา กล้าน้อย 02-8681246 ต่อ 21

67/35-36, 3rd Petkasem 7/1 Rd., Thapra, Bangkokyai, Bangkok 10600 Thailand

Tel: (66) 0-28681246 Fax: (66) 0-2868-0860

**OKLA TESTING & CONSULTING SERVICE CO., LTD. (Head office)****บริษัท โอคลา เทสติ้ง แอนด์ คอนซัลติ้ง เซอร์วิส จำกัด (สำนักงานใหญ่)**67/35-36, 3rd Floor., Phetkasem 7/1 Rd., Wat Tha Pra, Bangkokyai, Bangkok, THAILAND 10600

Tel: (66) 02 868 1246

67/35-36 ชั้น 3 เพชรเกษม ซอย 7/1 แขวงวัดท่าพระ เขตบางกอกใหญ่ กรุงเทพฯ 10600

FAX: (66) 02 868 0860

E-MAIL: sales@okla-testing.com

Website: www.okla-testing.com

J-NAC Group

Tax ID: 0105553003058



PM10 Calibration Report

Location: OKLA-Testing Lab **Date:** 6/5/2022

Sampler: EM-PM10-02 **Serial No:**


Barometric pressure, mm Hg (Pa): 760 **Temperature, Deg C (Ta):** 33

Transfer Standard Type: Tisch TE 5025A **Serial No:** 1758

Last Calibration Date: 2-May-19 **Operator:** Mr.Parinya

Qstd Slope: 2.0337 **Qstd Intercept:** -0.03624

Plate No	H2O (in)	Qstd		I (Chart)	IC (corrected)
		(m3/min)	(cfm)		
1	10.00	1.571	55.469	56.00	55.26
2	8.00	1.409	49.748	51.00	50.33
3	6.40	1.264	44.631	45.00	44.41
4	4.60	1.077	38.033	38.00	37.50
5	2.40	0.788	27.827	28.00	27.63

Linear Regression**Slope:** 1.0146**Intecept:** -0.7400**Corr. Coeff:** 0.9994

CALIBRATION BY :		DATE :	6/5/2022
APPROVED BY :		DATE :	6/5/22
ต้องการข้อมูลทางด้านเทคนิคเพิ่มเติม : นายปริญญา กล้าน้อย 02-8681246 ต่อ 21			
67/35-36,3rd Petkasem 7/1 Rd.,Thapra, Bangkokyai, Bangkok 10600 Thailand			
Tel: (66) 0-28681246 Fax:(66) 0-2868-0860			



OKLA TESTING & CONSULTING SERVICE CO., LTD. (Head office)

บริษัท โอคลา เทสติ้ง แอนด์ คอนซัลติ้ง เซอร์วิส จำกัด (สำนักงานใหญ่)

67/35-36, 3rd Floor., Phetkasem 7/1 Rd., Wat Tha Pra, Bangkokyai, Bangkok, THAILAND 10600

Tel: (66) 02 868 1246

67/35-36 ชั้น 3 เพชรเกษม ซอย 7/1 แขวงวัดท่าพระ เขตบางกอกใหญ่ กรุงเทพฯ 10600

FAX: (66) 02 868 0860

E-MAIL: sales@okla-testing.com

Website: www.okla-testing.com

J-NAC Group

Tax ID: 010553003058

Sound level Calibration report

Instrument Calibrated

Description : Sound Level Meter
Manufacturer : BSWA
Model : BSWA 309
Serial No. : 540029
Microphone No. : 392149
Peramplifier No. : 540228

Ambient Enviroment

Temperature : (24 \pm 3) $^{\circ}$ C
Relative Humidity : (55 \pm 15)%
Ambient Pressure : (1008 \pm 5)hPa

Standards Used

- Sound Level Calibrator Delta ohm HD 2020 S/N 17000992

Measurement Result

Method	Standard Reference (dB)	Reading (dB)	Error (dB)	After Adjustment \pm (dB)
Sound level Calibrator HD-2020	94.0	93.8	+0.2	94.0

CALIBRATION BY :		6/5/2022.
APPROVED BY :		6/5/22
67/35-36, 3 rd Petkasem 7/1 Rd., T		0600 Thailand Tel: (66) 0-28681246 Fax: (66) 0-2868-0860



OKLA TESTING & CONSULTING SERVICE CO., LTD. (Head office)

บริษัท โอคลา เทสติ้ง แอนด์ คอนซัลติง เซอร์วิส จำกัด (สำนักงานใหญ่)

67/35-36, 3rd Floor., Phetkasem 7/1 Rd., Wat Tha Pra, Bangkokkya, Bangkok, THAILAND 10600

Tel: (66) 02 868 1246

67/35-36 ชั้น 3 เพชรเกษม ซอย 7/1 แขวงวัดท่าพระ เขตบางกอกใหญ่ กรุงเทพฯ 10600

FAX: (66) 02 868 0860

E-MAIL: sales@okla-testing.com

Website: www.okla-testing.com

J-NAC Group


Tax ID: 010553003058

CALIBRATION REPORT

EQUIPMENT NAME : SO ₂ Analyzer			
MANUFACTURER : Horiba	MODEL : APSA-370	SERIAL NO : SFBMA08V	
STANDARD GAS CONCENTRATION (PPM) 45.25 PPM		CYLINDER NO : EB0124972	
CYLINDER PRESSURE (PSI) : 1,000 PSI		CERTIFIED DATE : 24/09/2019	
CERTIFIED BY : Airgas		EXPIRED DATE : 24/09/2022	

CALIBRATION RESULTS

POINT NO	CALIBRATION RESULTS			
	IDEAL	ACTUAL	ERROR	%ERROR
ZERO	0.00	0.0001	0.00	-
1	100.00	100.0000	0.0	0.00
2	200.00	200.7000	0.7	0.35
3	300.00	301.0000	1.0	0.33
4	400.00	399.0000	-1.0	-0.25
AVERAGE (%)				0.11



CALIBRATED BY :

DATE : 6/5/22

CHECKED BY :

DATE : 6/5/22

ต้องการข้อมูลทางด้านเทคนิคเพิ่มเติม : นายปริญญา กล้าน้อย 02-8681246 ต่อ 22

67/35-36, 3rd Petkasem 7/1 Rd., Thapra, Bangkokkya, Bangkok 10600 Thailand Tel: (66) 0-28681246 Fax: (66) 0-2868-0860



OKLA TESTING & CONSULTING SERVICE CO., LTD. (Head office)

บริษัท โอคลา เทสติ้ง แอนด์ คอนซัลติง เซอร์วิส จำกัด (สำนักงานใหญ่)

67/35-36, 3rd Floor., Phetkasem 7/1 Rd., Wat Tha Pra, Bangkokyal, Bangkok, THAILAND 10600

67/35-36 ชั้น 3 เพชรเกษม ซอย 7/1 แขวงวัดท่าพระ เขตบางกอกใหญ่ กรุงเทพฯ 10600

E-MAIL: sales@okla-testing.com

Website: www.okla-testing.com

J-NAC Group

Tel: (66) 02 868 1246

FAX: (66) 02 868 0860


Tax ID: 0105533003058

CALIBRATION REPORT

EQUIPMENT NAME : NOx Analyzer			
MANUFACTURER : HORIBA	MODEL : APNA-370	SERIAL NO : 36WA70V3	
STANDARD GAS CONCENTRATION (PPM) : 44.78 PPM		CYLINDER NO : EB0124972	
CYLINDER PRESSURE (psig) : 1000 PSI		CERTIFIED DATE : 24/09/2019	
CERTIFIED BY : Airgas		EXPIRED DATE : 24/09/2022	

CALIBRATION RESULTS

POINT NO	CALIBRATION RESULTS						
	IDEAL	ACTUAL NO	ERROR NO	%ERROR NO	ACTUAL NO _x	ERROR NO _x	%ERROR NO _x
ZERO	0.00	0.00	0.00	-	0.20	0.20	-
1	100.00	100.10	0.10	0.10	100.10	0.10	0.10
2	200.00	200.20	0.20	0.10	200.10	0.10	0.05
3	300.00	300.10	0.10	0.03	300.10	0.10	0.03
4	400.00	400.10	0.10	0.03	400.20	0.20	0.05
				0.06			0.06



CALIBRATED BY :

DATE : 6/5/2022

CHECKED BY :

DATE : 6/5/22

ต้องการข้อมูลทางด้านเทคนิคเพิ่มเติม : นายปริญญ์ กล้าน้อย 02-8681246 ต่อ 22

67/35-36, 3rd Petkasem 7/1 Rd., Thapra, Bangkokyal, Bangkok 10600 Thailand Tel: (66) 0-28681246 Fax: (66) 0-2868-0860

CERTIFICATE OF CALIBRATION

FOR

NOMENCLATURE : VIBRATION METER
MANUFACTURER : VIBROCK
MODEL / TYPE : V9000
SERIAL NO. : 2342
CLID. NO. : 252200819
JOB CONTROL NO. : 220330033647

CUSTOMER : OKLA TESTING & CONSULTING SERVICE CO., LTD.
67/35-36, 3RD FLOOR, PHETKASEM 7/1 RD., WATTHAPRA,
BANGKOKYAI, BANGKOK 10600 THAILAND

DATE OF RECEIVED : 30 March 2022

DATE OF ISSUED : 04 April 2022

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Suwit Phuanbusabong
Calibration Engineer



Approved By : Mongkol Yotsoontorn
Authorized Signatory
04 April 2022



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q22033647

F3-011-04/01-12

page 1 of 3



@clccalibration

REPORT OF CALIBRATION FOR

NOMENCLATURE : **VIBRATION METER**
MANUFACTURER : **VIBROCK**
MODEL / TYPE : **V9000**
SERIAL NO. : **2342**
DATE OF CALIBRATION : **31 March 2022**

ENVIRONMENT CONDITIONS :

Temperature : $(23 \pm 2) ^\circ\text{C}$

Relative Humidity : $(55 \pm 15) \% \text{RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **CLC-CPEE-08** based on **ISO 16063-21** as calibration guideline.
The calibration was performed by using Digital Multimeter, High Resolution Programmable Timer/Counter and Portable Vibration Calibrator which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

1. Digital Multimeter, Hewlett Packard Model 34401A S/N. 3146A75935.
2. High Resolution Programmable Timer/Counter, Philips Model PM6680B S/N. SM607101.
3. Portable Vibration Calibrator, The Modal Shop Model 9110D S/N. 11424.

TRACEABILITY :

1. The measurements are traceable to International System of Units (SI), through National Institute of Metrology (Thailand) Certificate No. EE-0112-21, Due Date 26 October 2022.
2. The measurements are traceable to International System of Units (SI), through Aeronautical Radio of Thailand Ltd. Certificate No. 07-0001/22, Due Date 22 February 2023.
3. The measurements are traceable to International System of Units (SI), through The Modal Shop, Inc. Certificate No. 2649.01, Due Date 10 November 2022.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2,00$ which for a normal distribution corresponds to a coverage probability of approximately 95 %.
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2021)"

Certificate No. Q22033647

F3-011-04/01-12

page 2 of 3



@clccalibration



CLC
Accredited
ISO/IEC 17025

CALIBRATION LABORATORY Co., LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230

Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment () adjustment

CALIBRATION DATA

VELOCITY RESULT

Test point		Mode	STD Reading	DUC Reading	Correction	Uncertainty
(mm/s)	(frequency)		(mm/s)	(mm/s)	(mm/s)	± (% of rdg.)
10	160 Hz	peak	10.00	10.10	-0.10	1.8
20	160 Hz		20.00	19.92	+0.08	1.8
40	160 Hz		40.00	39.10	+0.90	1.0
60	160 Hz		60.00	58.80	+1.20	1.0
80	160 Hz		80.00	78.71	+1.29	1.0
100	160 Hz		100.00	98.65	+1.35	1.0

Note. The Scope of Accredited ANAB Certificate No. ACDM-2814 Version 008 Page 1 of 54

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q22033647

F3-011-04/01-12

page 3 of 3



@clccalibration

**OKLA TESTING & CONSULTING SERVICE CO., LTD.****บริษัท โอกลา เทสติ้ง แอนด์ คอนซัลติง เซอร์วิส จำกัด**

63/13 PHETKASEM 7 Rd., THAPRA, BANGKOKYAI, BANGKOK 10600, THAILAND

63/13 เพชรเกษม ซอย 7 แขวงวัดท่าพระ เขตบางกอกใหญ่ กรุงเทพฯ 10600

E-MAIL: sales@okla-testing.com Website: www.okla-testing.com

Tel: (66) 02 868 1246

FAX: (66) 02 868 0860

J-NAC Group



TSP Calibration Report

Location: OKLA-Testing Lab

Date: 6/6/2022

Sampler: EM-TSP-02

Serial No:

Barometric pressure, mm Hg (Pa): 760

Temperature, Deg C (Ta): 33

Transfer Standard Type: Tisch TE 5025A

Serial No: 1758

Last Calibration Date: 2-May-19

Operator: Mr.Parinya

Qstd Slope: 2.00373

Qstd Intercept: -0.00705


Plate No	H2O (in)	Qstd		I (Chart)	IC (corrected)
		(m3/min)	(cfm)		
1	9.60	1.533	54.138	50.00	49.34
2	7.60	1.365	48.197	44.00	43.42
3	6.40	1.253	44.249	41.00	40.46
4	4.10	1.004	35.466	32.00	31.58
5	2.10	0.721	25.453	23.00	22.70

Linear Regression

Slope: 0.9302

Intecept: -1.1000

Corr. Coeff: 0.9996



CALIBRATION BY :

DATE :

6/6/2022

APPROVED BY :

DATE :

6/6/22

ต้องการข้อมูลทางด้านเทคนิคเพิ่มเติม : นายปริญญา กล้าน้อย 02-8681246 ต่อ 21

67/35-36, 3rd Petkasem 7/1 Rd., Thapra, Bangkokyai, Bangkok 10600 Thailand

Tel: (66) 0-28681246 Fax: (66) 0-2868-0860



OKLA TESTING & CONSULTING SERVICE CO., LTD. (Head office)

บริษัท โอคลา เทสติ้ง แอนด์ คอนซัลติง เซอร์วิส จำกัด (สำนักงานใหญ่)

67/35-36, 3rd Floor., Phetkasem 7/1 Rd., Wat Tha Pra, Bangkokyai, Bangkok, THAILAND 10600

Tel: (66) 02 868 1246

67/35-36 ชั้น 3 เพชรเกษม ซอย 7/1 แขวงวัดท่าพระ เขตบางกอกใหญ่ กรุงเทพฯ 10600

FAX: (66) 02 868 0860

E-MAIL: sales@okla-testing.com

Website: www.okla-testing.com

J-NAC Group

Tax ID: 0105553003058



PM10 Calibration Report

Location: OKLA-Testing Lab Date: 6/6/2022
Sampler: EM-PM10-02 Serial No: _____
Barometric pressure, mm Hg (Pa): 760 Temperature, Deg C (Ta): 33
Transfer Standard Type: Tisch TE 5025A Serial No: 1758
Last Calibration Date: 2-May-19 Operator: Mr.Parinya
Qstd Slope: 2.0337 Qstd Intercept: -0.03624


Plate No	H2O (in)	Qstd		I (Chart)	IC (corrected)
		(m3/min)	(cfm)		
1	10.00	1.571	55.469	56.00	55.26
2	8.00	1.409	49.748	50.00	49.34
3	6.40	1.264	44.631	44.00	43.42
4	4.70	1.088	38.430	39.00	38.49
5	2.50	0.803	28.375	28.00	27.63

Linear Regression

Slope: 1.0095

Intecept: -0.9100

Corr. Coeff: 0.9990



CALIBRATION BY :		DATE :	6/6/2022
APPROVED BY :		DATE :	6/6/22
ต้องการข้อมูลทางด้านเทคนิคเพิ่มเติม : นายปริญญา กล้าน้อย 02-8681246 ต่อ 21			
67/35-36,3rd Petkasem 7/1 Rd.,Thapra, Bangkokyai, Bangkok 10600 Thailand			
Tel: (66) 0-28681246 Fax:(66) 0-2868-0860			



OKLA TESTING & CONSULTING SERVICE CO., LTD. (Head office)

บริษัท โอคลา เทสติ้ง แอนด์ คอนซัลติ้ง เซอร์วิส จำกัด (สำนักงานใหญ่)

67/35-36, 3rd Floor., Phetkasem 7/1 Rd., Wat Tha Pra, Bangkokyai, Bangkok, THAILAND 10600

Tel: (66) 02 868 1246

67/35-36 ชั้น 3 เพชรเกษม ซอย 7/1 แขวงวัดท่าพระ เขตบางกอกใหญ่ กรุงเทพฯ 10600

FAX: (66) 02 868 0860

E-MAIL: sales@okla-testing.com

Website: www.okla-testing.com

J-NAC Group

Tax ID: 0105553003058

Sound level Calibration report

Instrument Calibrated

Description : Sound Level Meter
Manufacturer : BSWA
Model : BSWA 309
Serial No. : 540029
Microphone No. : 392149
Peramplifier No. : 540228

Ambient Enviroment

Temperature : (24 \pm 3) $^{\circ}$ C
Relative Humidity : (55 \pm 15)%
Ambient Pressure : (1008 \pm 5)hPa

Standards Used

- Sound Level Calibrator Delta ohm HD 2020 S/N 17000992

Measurement Result

Method	Standard Reference (dB)	Reading (dB)	Error (dB)	After Adjustment \pm (dB)
Sound level Calibrator HD-2020	94.0	93.8	+0.2	94.0

CALIBRATION BY :		6/6/2022.
APPROVED BY :		6/6/22
67/35-36, 3 rd Petkasem 7/1 Rd., Thapra, Bangkokyai, Bangkok 10600 Thailand Tel: (66) 0-28681246 Fax: (66) 0-2868-0860		



OKLA TESTING & CONSULTING SERVICE CO., LTD. (Head office)

บริษัท โอคลา เทสติ้ง แอนด์ คอนซัลติง เซอร์วิส จำกัด (สำนักงานใหญ่)

67/35-36, 3rd Floor., Phetkasem 7/1 Rd., Wat Tha Pra, Bangkokkyl, Bangkok, THAILAND 10600

Tel: (66) 02 868 1246

67/35-36 ชั้น 3 เพชรเกษม ซอย 7/1 แขวงวัดท่าพระ เขตบางกอกใหญ่ กรุงเทพฯ 10600

FAX: (66) 02 868 0860

E-MAIL: sales@okla-testing.com

Website: www.okla-testing.com

J-NAG Group


Tax ID: 010553003058

CALIBRATION REPORT

EQUIPMENT NAME :	SO ₂ Analyzer		
MANUFACTURER :	Horiba	MODEL :	APSA-370
SERIAL NO :		SFBMA08V	
STANDARD GAS CONCENTRATION (PPM)	45.25 PPM	CYLINDER NO :	EB0124972
CYLINDER PRESSURE (PSI) :	1,000 PSI	CERTIFIED DATE :	24/09/2019
CERTIFIED BY :	Airgas	EXPIRED DATE :	24/09/2022

CALIBRATION RESULTS

POINT NO	CALIBRATION RESULTS			
	IDEAL	ACTUAL	ERROR	%ERROR
ZERO	0.00	0.0001	0.00	-
1	100.00	100.0000	0.0	0.00
2	200.00	200.7000	0.7	0.35
3	300.00	301.0000	1.0	0.33
4	400.00	399.0000	-1.0	-0.25
AVERAGE (%)				0.11



CALIBRATED BY :

DATE : 6/6/22

CHECKED BY :

DATE : 6/6/22

ต้องการข้อมูลทางด้านเทคนิคเพิ่มเติม : นายปริญญา กล้าน้อย 02-8681246 ต่อ 22

67/35-36, 3rd Petkasem 7/1 Rd., Thapra, Bangkokkyl, Bangkok 10600 Thailand Tel: (66) 0-28681246 Fax: (66) 0-2868-0860



OKLA TESTING & CONSULTING SERVICE CO., LTD. (Head office)

บริษัท โอคลา เทสติ้ง แอนด์ คอนซัลติง เซอร์วิส จำกัด (สำนักงานใหญ่)

67/35-36, 3rd Floor., Phetkasem 7/1 Rd., Wat Tha Pra, Bangkokyal, Bangkok, THAILAND 10600

Tel: (66) 02 868 1246

67/35-36 ชั้น 3 เพชรเกษม ซอย 7/1 แขวงวัดท่าพระ เขตบางกอกใหญ่ กรุงเทพฯ 10600

FAX: (66) 02 868 0860

E-MAIL: sales@okla-testing.com

Website: www.okla-testing.com

J-NAC Group


Tax ID: 010553003058

CALIBRATION REPORT

EQUIPMENT NAME : NOx Analyzer			
MANUFACTURER : HORIBA	MODEL : APNA-370	SERIAL NO : 36WA70V3	
STANDARD GAS CONCENTRATION (PPM) : 44.78 PPM		CYLINDER NO : EB0124972	
CYLINDER PRESSURE (psig) : 1000 PSI		CERTIFIED DATE : 24/09/2019	
CERTIFIED BY : Airgas		EXPIRED DATE : 24/09/2022	

CALIBRATION RESULTS

POINT NO	CALIBRATION RESULTS						
	IDEAL	ACTUAL NO	ERROR NO	%ERROR NO	ACTUAL NO _x	ERROR NO _x	%ERROR NO _x
ZERO	0.00	0.00	0.00	-	0.20	0.20	-
1	100.00	100.10	0.10	0.10	100.10	0.10	0.10
2	200.00	200.20	0.20	0.10	200.10	0.10	0.05
3	300.00	300.10	0.10	0.03	300.10	0.10	0.03
4	400.00	400.10	0.10	0.03	400.20	0.20	0.05
				0.06			0.06



CALIBRATED BY :

DATE : 6/6/22

CHECKED BY :

DATE : 6/6/22

ต้องการข้อมูลทางด้านเทคนิคเพิ่มเติม : นายปริญญา กล้าน้อย 02-8681246 ต่อ 22

67/35-36, 3rd Petkasem 7/1 Rd., Thapra, Bangkokyal, Bangkok 10600 Thailand Tel: (66) 0-28681246 Fax: (66) 0-2868-0860

CERTIFICATE OF CALIBRATION

FOR

NOMENCLATURE : VIBRATION METER
MANUFACTURER : VIBROCK
MODEL / TYPE : V9000
SERIAL NO. : 2342
CLID. NO. : 252200819
JOB CONTROL NO. : 220330033647

CUSTOMER : OKLA TESTING & CONSULTING SERVICE CO., LTD.
67/35-36, 3RD FLOOR, PHETKASEM 7/1 RD., WATTHAPRA,
BANGKOKYAI, BANGKOK 10600 THAILAND

DATE OF RECEIVED : 30 March 2022

DATE OF ISSUED : 04 April 2022

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Suwit Phuanbusabong
Calibration Engineer



Approved By : Mongkol Yotsoontorn
Authorized Signatory
04 April 2022



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q22033647

F3-011-04/01-12

page 1 of 3



@clccalibration

REPORT OF CALIBRATION FOR

NOMENCLATURE : **VIBRATION METER**
MANUFACTURER : **VIBROCK**
MODEL / TYPE : **V9000**
SERIAL NO. : **2342**
DATE OF CALIBRATION : **31 March 2022**

ENVIRONMENT CONDITIONS :

Temperature : $(23 \pm 2) ^\circ\text{C}$

Relative Humidity : $(55 \pm 15) \% \text{RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **CLC-CPEE-08** based on **ISO 16063-21** as calibration guideline.
The calibration was performed by using Digital Multimeter, High Resolution Programmable Timer/Counter and Portable Vibration Calibrator which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

1. Digital Multimeter, Hewlett Packard Model 34401A S/N. 3146A75935.
2. High Resolution Programmable Timer/Counter, Philips Model PM6680B S/N. SM607101.
3. Portable Vibration Calibrator, The Modal Shop Model 9110D S/N. 11424.

TRACEABILITY :

1. The measurements are traceable to International System of Units (SI), through National Institute of Metrology (Thailand) Certificate No. EE-0112-21, Due Date 26 October 2022.
2. The measurements are traceable to International System of Units (SI), through Aeronautical Radio of Thailand Ltd. Certificate No. 07-0001/22, Due Date 22 February 2023.
3. The measurements are traceable to International System of Units (SI), through The Modal Shop, Inc. Certificate No. 2649.01, Due Date 10 November 2022.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2,00$ which for a normal distribution corresponds to a coverage probability of approximately 95 %.
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2021)"

Certificate No. Q22033647

F3-011-04/01-12

page 2 of 3



@clccalibration



CLC
Accredited
ISO/IEC 17025

CALIBRATION LABORATORY Co., LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230

Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment () adjustment

CALIBRATION DATA

VELOCITY RESULT

Test point		Mode	STD Reading	DUC Reading	Correction	Uncertainty
(mm/s)	(frequency)		(mm/s)	(mm/s)	(mm/s)	± (% of rdg.)
10	160 Hz	peak	10.00	10.10	-0.10	1.8
20	160 Hz		20.00	19.92	+0.08	1.8
40	160 Hz		40.00	39.10	+0.90	1.0
60	160 Hz		60.00	58.80	+1.20	1.0
80	160 Hz		80.00	78.71	+1.29	1.0
100	160 Hz		100.00	98.65	+1.35	1.0

Note. The Scope of Accredited ANAB Certificate No. ACDM-2814 Version 008 Page 1 of 54

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q22033647

F3-011-04/01-12

page 3 of 3



@clccalibration



OKLA TESTING & CONSULTING SERVICE CO., LTD. (Head office)

บริษัท โอคลา เทสติ้ง แอนด์ คอนซัลติง เซอร์วิส จำกัด (สำนักงานใหญ่)

67/35-36, 3rd Floor., Phetkasem 7/1 Rd., Wat Tha Pra, Bangkokyai, Bangkok, THAILAND 10600

Tel: (66) 02 868 1246

67/35-36 ชั้น 3 เพชรเกษม ซอย 7/1 แขวงวัดท่าพระ เขตบางกอกใหญ่ กรุงเทพฯ 10600

FAX: (66) 02 868 0860

E-MAIL: sales@okla-testing.com

Website: www.okla-testing.com

J-NAC Group

Tax ID: 0105553003058



TSP Calibration Report

Location: OKLA-Testing Lab

Date: 6/6/2022

Sampler: EM-TSP-03

Serial No:

Barometric pressure, mm Hg (Pa): 760

Temperature, Deg C (Ta): 33

Transfer Standard Type: Tisch TE 5025A

Serial No: 1758

Last Calibration Date: 2-May-19

Operator: Mr.Parinya

Qstd Slope: 2.00373

Qstd Intercept: -0.00705


Plate No	H2O (in)	Qstd		I (Chart)	IC (corrected)
		(m3/min)	(cfm)		
1	9.30	1.509	53.289	50.00	49.34
2	7.10	1.319	46.593	44.00	43.42
3	6.10	1.223	43.205	41.00	40.46
4	4.00	0.992	35.034	32.00	31.58
5	2.10	0.721	25.453	22.00	21.71

Linear Regression

Slope: 1.0035

Intecept: -3.5600

Corr. Coeff: 0.9990



CALIBRATION BY:

DATE:

6/6/2022

APPROVED BY:

DATE:

6/6/22

ต้องการข้อมูลทางด้านเทคนิคเพิ่มเติม : นายปริญญา กล้าน้อย 02-8681246 ต่อ 21

67/35-36, 3rd Petkasem 7/1 Rd., Thapra, Bangkokyai, Bangkok 10600 Thailand Tel: (66) 0-28681246



OKLA TESTING & CONSULTING SERVICE CO., LTD. (Head office)

บริษัท โอคลา เทสติ้ง แอนด์ คอนซัลติง เซอร์วิส จำกัด (สำนักงานใหญ่)

67/35-36, 3rd Floor., Phetkasem 7/1 Rd., Wat Tha Pra, Bangkokyai, Bangkok, THAILAND 10600

Tel: (66) 02 868 1246

67/35-36 ชั้น 3 เพชรเกษม ซอย 7/1 แขวงวัดท่าพระ เขตบางกอกใหญ่ กรุงเทพฯ 10600

FAX: (66) 02 868 0860

E-MAIL: sales@okla-testing.com

Website: www.okla-testing.com

J-NAC Group

Tax ID: 010553003058



PM10 Calibration Report

Location: OKLA-Testing Lab

Date: 6/6/2022

Sampler: EM-PM10-03

Serial No:

Barometric pressure, mm Hg (Pa): 760

Temperature, Deg C (Ta): 33

Transfer Standard Type: Tisch TE 5025A

Serial No: 1758

Last Calibration Date: 2-May-19

Operator: Mr.Parinya

Qstd Slope: 2.03370

Qstd Intercept: -0.03624


Plate No	H2O (in)	Qstd		I (Chart)	IC (corrected)
		(m3/min)	(cfm)		
1	10.00	1.571	55.469	56.00	55.26
2	8.20	1.426	50.351	50.00	49.34
3	6.20	1.244	43.949	45.00	44.41
4	4.50	1.066	37.631	38.00	37.50
5	2.20	0.756	26.697	28.00	27.63

Linear Regression

Slope: 0.9523

Intecept: 2.0500

Corr. Coeff: 0.9989



CALIBRATION BY :		DATE :	6/6/2022.
APPROVED BY :		DATE :	6/6/22
ต้องการข้อมูลทางด้านเทคนิคเพิ่มเติม : นายปริญญา กล้าน้อย 02-8681246 ต่อ 21			
67/35-36,3rd Petkasem 7/1 Rd., Thapra, Bangkokyai, Bangkok 10600 Thailand Tel: (66) 0-28681246			



OKLA TESTING & CONSULTING SERVICE CO., LTD. (Head office)

บริษัท โอคลา เทสติ้ง แอนด์ คอนซัลติง เซอร์วิส จำกัด (สำนักงานใหญ่)

67/35-36, 3rd Floor., Phetkasem 7/1 Rd., Wat Tha Pra, Bangkokyai, Bangkok, THAILAND 10600

Tel: (66) 02 868 1246

67/35-36 ชั้น 3 เพชรเกษม ซอย 7/1 แขวงวัดท่าพระ เขตบางกอกใหญ่ กรุงเทพฯ 10600

FAX: (66) 02 868 0860

E-MAIL: sales@okla-testing.com

Website: www.okla-testing.com

J-NAC Group

Tax ID: 0105553003058

Sound level Calibration report

Instrument Calibrated

Description : Sound Level Meter
Manufacturer : BSWA
Model : BSWA 309
Serial No. : 540008
Microphone No. : 391925
Peramplifier No. : 540069

Ambient Enviroment

Temperature : (24 \pm 3) $^{\circ}$ C
Relative Humidity : (55 \pm 15)%
Ambient Pressure : (1008 \pm 5)hPa

Standards Used

- Sound Level Calibrator Delta ohm HD 2020 S/N 17000992

Measurement Result

Method	Standard Reference (dB)	Reading (dB)	Error (dB)	After Adjustment \pm (dB)
Sound level Calibrator HD-2020	94.0	93.6	+0.4	94.0

CALIBRATION BY :

APPROVED BY :

6/6/22

6/6/22

67/35-36, 3rd Petkasem 7/1 Rd., Wattapra, Bangkokyai, Bangkok 10600 Thailand Tel: (66) 0-28681246 Fax:(66) 0-2868-0860



OKLA TESTING & CONSULTING SERVICE CO., LTD. (Head office)

บริษัท โอคลา เทสติ้ง แอนด์ คอนซัลติง เซอร์วิส จำกัด (สำนักงานใหญ่)

67/35-36, 3rd Floor., Phetkasem 7/1 Rd., Wat Tha Pra, Bangkokyai, Bangkok, THAILAND 10600

Tel: (66) 02 868 1246

67/35-36 ชั้น 3 เพชรเกษม ซอย 7/1 แขวงวัดท่าพระ เขตบางกอกใหญ่ กรุงเทพฯ 10600

FAX: (66) 02 868 0860

E-MAIL: sales@okla-testing.com

Website: www.okla-testing.com

J-NAG Group


Tax ID: 0105553003058

CALIBRATION REPORT

EQUIPMENT NAME :	SO ₂ Analyzer		
MANUFACTURER :	Horiba	MODEL :	APSA-370
SERIAL NO :		SFBMA08V	
STANDARD GAS CONCENTRATION (PPM)	45.25 PPM	CYLINDER NO :	EB0124972
CYLINDER PRESSURE (PSI) :	1,000 PSI	CERTIFIED DATE :	24/09/2019
CERTIFIED BY :	Airgas	EXPIRED DATE :	24/09/2022

CALIBRATION RESULTS

POINT NO	CALIBRATION RESULTS			
	IDEAL	ACTUAL	ERROR	%ERROR
ZERO	0.00	0.0001	0.00	-
1	100.00	100.0000	0.0	0.00
2	200.00	200.7000	0.7	0.35
3	300.00	301.0000	1.0	0.33
4	400.00	399.0000	-1.0	-0.25
AVERAGE (%)				0.11



CALIBRATED BY :

DATE : 6/6/22

CHECKED BY :

DATE : 6/6/22

ต้องการข้อมูลทางด้านเทคนิคเพิ่มเติม : นายปริญญา กล้าน้อย 02-8681246 ต่อ 22

67/35-36, 3rd Petkasem 7/1 Rd., Thapra, Bangkokyai, Bangkok 10600 Thailand Tel: (66) 0-28681246 Fax: (66) 0-2868-0860



OKLA TESTING & CONSULTING SERVICE CO., LTD. (Head office)

บริษัท โอคลา เทสติ้ง แอนด์ คอนซัลติง เซอร์วิส จำกัด (สำนักงานใหญ่)

67/35-36, 3rd Floor., Phetkasem 7/1 Rd., Wat Tha Pra, Bangkok, THAILAND 10600

67/35-36 ชั้น 3 เพชรเกษม ซอย 7/1 แขวงวัดท่าพระ เขตบางกอกใหญ่ กรุงเทพฯ 10600

E-MAIL: sales@okla-testing.com

Website: www.okla-testing.com

J-NAC Group

Tel: (66) 02 060 1246

FAX: (66) 02 060 0860


Tax ID: 010553003058

CALIBRATION REPORT

EQUIPMENT NAME : NOx Analyzer			
MANUFACTURER : HORIBA	MODEL : APNA-370	SERIAL NO : 36WA70V3	
STANDARD GAS CONCENTRATION (PPM) : 44.78 PPM		CYLINDER NO : EB0124972	
CYLINDER PRESSURE (psig) : 1000 PSI		CERTIFIED DATE : 24/09/2019	
CERTIFIED BY : Airgas		EXPIRED DATE : 24/09/2022	

CALIBRATION RESULTS

POINT NO	CALIBRATION RESULTS						
	IDEAL	ACTUAL NO	ERROR NO	%ERROR NO	ACTUAL NO _x	ERROR NO _x	%ERROR NO _x
ZERO	0.00	0.00	0.00	-	0.20	0.20	-
1	100.00	100.10	0.10	0.10	100.10	0.10	0.10
2	200.00	200.20	0.20	0.10	200.10	0.10	0.05
3	300.00	300.10	0.10	0.03	300.10	0.10	0.03
4	400.00	400.10	0.10	0.03	400.20	0.20	0.05
				0.06			0.06



CALIBRATED BY :

DATE : 6/6/22

CHECKED BY :

DATE : 6/6/22

ต้องการข้อมูลทางด้านเทคนิคเพิ่มเติม : นายปริญญา กล้าน้อย 02-8681246 ต่อ 22

67/35-36, 3rd Petkasem 7/1 Rd., Thapra, Bangkok, Thailand 10600 Thailand Tel: (66) 0-28681246 Fax: (66) 0-2868-0860



GIIC Calibration Laboratory

700/20-21 Phaholyothin Rd., Samsennai, Phayathai,
Bangkok 10400 Thailand

Tel : +66 (02) 615 4999

Fax : +66 (02) 615 4644

E-mail : cal@giic.co.th



NSC-TISI-TIS 17025
CALIBRATION 6256

CERTIFICATE No.CAL00639-22..... PAGE1..... OF3.....

Certificate of Calibration

Equipment : DIGITAL THERMO-HYGROMETER

Manufacturer : DIGICON

Model / Type : TH-03

Serial No. : 115092766

ID No. : -

Customer : OKLA TESTING & CONSULTING SERVICE CO., LTD.
67/35-36, 3rd Floor, Phetkasem 7/1, Watthapra,
Bangkokyai, Bangkok 10600 Thailand.

C.S.R. No. : H0000639-22

Received Date : 04 May 2022

Calibration Date : 05 May 2022

Calibrated By : TONTRAKARN SRIKACHA

Approved By : WIWAT CHAMNANDEE

Issue Date : 05 May 2022

The uncertainties are for a level of confidence of approximately 95%.

This certificate may not be reproduced except in full unless permission for the reproduction has been obtained in writing from the laboratory.

CERTIFICATE No. CAL00639-22 PAGE 2 OF 3

CALIBRATION REPORT

Condition of this calibration result:

1. Environment : Temperature : $(25 \pm 3) ^\circ\text{C}$
 Relative Humidity : $(50 \pm 15) \% \text{RH}$

2. Reference / procedure Used :

- This equipment was calibrated by comparison to precision humidity measuring instrument into humidity chamber for humidity measurement and a platinum resistance thermometer into temperature chamber for temperature measurement according to GILC Calibration Laboratory
- Calibration Procedure No. GILCLAB-CP-H01, GILCLAB-CP-H03.

3. Reference Standard Instrument :

Instrument	Model	Serial No	Certificate No	Due Dated
Platinum Resistance Thermometer	PCR-1	RB-31604	21I703	6 Jul 22
Data Logger	HC2-S	60936993	21T9467	11 Oct 22
Dual Measurement Multimeter	GDM 8261A	GEP925925	CAL00436-22	19 Mar 23

4. This Certification is traceable to the SI unit through :

- Technology Promotion Association (Thailand-Japan) Calibration Services and Environmental Analysis Department.
- Quality Calibration
- GILC Calibration Laboratory

5. Uncertainty :

- The reported uncertainty of measurement was estimated and based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95 %.

6. Disclaimer :

- The laboratory accepted that was we has done in our calibration method. It with no guarantee that it works as you believe that it should and user accept the risks that occur. We accept no liability for any damage or financial losses.

CERTIFICATE No. CAL00639-22 PAGE 3 OF 3

CALIBRATION REPORT

The temperature scale used was based on ITS-90.

All data shown below were as-received values without adjustment.

Calibration result :

Function : Temperature Measurement.

Standard Temperature (°C)	¹ U.U.C. Reading (°C)	Error (°C)	Uncertainty of Measurement (± °C)
9.986	10	0.014	0.88
24.989	25	0.011	0.88
40.028	40	-0.028	0.88

Function : Humidity Measurement. : (25.01 °C)

Standard Humidity (% rh)	¹ U.U.C. Reading (% rh)	Error (% rh)	Uncertainty of Measurement (± % rh)
24.99	23	-1.99	1.8
49.94	44	-5.94	1.8
85.94	81	-4.94	2.9

¹U.U.C. = Unit Under Calibration

This result of calibration was found accurate as show on data and place of calibration only.

- END -

CERTIFICATE OF CALIBRATION

FOR

NOMENCLATURE : DO METER
MANUFACTURER : HANNA INSTRUMENTS
MODEL / TYPE : HI5421/HI76483
SERIAL NO. : 04240005101/KC1A11T8H
CLID. NO. : 272101220
JOB CONTROL NO. : 220426042326

CUSTOMER : OKLA TESTING & CONSULTING SERVICE CO., LTD.
67/35-36, 3RD FLOOR, PHETKASEM 7/1 RD., WATTHAPRA,
BANGKOKYAI, BANGKOK 10600 THAILAND

DATE OF RECEIVED : 26 April 2022

DATE OF ISSUED : 28 April 2022

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Prapaporn Khanchalee
Calibration Engineer



Approved By : Mongkol Yotsoontorn
Authorized Signatory

28 April 2022



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q22042326

F3-011-04/01-12

page 1 of 3



@clccalibration

REPORT OF CALIBRATION

FOR

NOMENCLATURE	:	DO METER
MANUFACTURER	:	HANNA INSTRUMENTS
MODEL / TYPE	:	HI5421/HI76483
SERIAL NO.	:	04240005101/KC1A11T8H
DATE OF CALIBRATION	:	27 April 2022

ENVIRONMENT CONDITIONS :

Temperature : $(25 \pm 2.5) ^\circ\text{C}$

Relative Humidity : $(50 \pm 15) \% \text{RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **CLC-CPCH-06**. The calibration was performed by direct measurement with Certified Reference Material (CRM).

REFERENCE STANDARD USED :

Dissolved Oxygen, Sigma-Alorich Product ID QC3077-500ML .

TRACEABILITY :

The measurements are traceable to International System of Units (SI) , through Merck Co., Ltd.

Lot LRAD0713.01 , Due Date September 2023.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2,00$ which for a normal distribution corresponds to a coverage probability of approximately 95 %.

It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2021)"

Certificate No. Q22042326

F3-011-04/01-12

page 2 of 3



CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment () adjustment

The table in the following gives the calibration results and associated measurement uncertainties of Do Meter.

CALIBRATION DATA

Nominal Value (mg/L)	DUC Reading (mg/L)	Correction (mg/L)	Uncertainty (mg/L)
5.91	5.88	+0.03	± 0.22

Note. The Scope of Accredited ANAB Certificate No. ACDM-2814 Version 008 Page 4 of 54

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q22042326

F3-011-04/01-12

page 3 of 3



@clccalibration

CERTIFICATE OF CALIBRATION

FOR

NOMENCLATURE : pH METER
MANUFACTURER : HANNA INSTRUMENTS
MODEL / TYPE : HI5521/HI1131
SERIAL NO. : 04160019101/061334CN
CLID. NO. : 272101219
JOB CONTROL NO. : 220426042324

CUSTOMER : OKLA TESTING & CONSULTING SERVICE CO., LTD.
67/35-36, 3RD FLOOR, PHETKASEM 7/1 RD., WATTHAPRA,
BANGKOKYAI, BANGKOK 10600 THAILAND

DATE OF RECEIVED : 26 April 2022

DATE OF ISSUED : 29 April 2022

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Prapaporn Khanchalee
Calibration Engineer



Approved By : Mongkol Yotsoontorn
Authorized Signatory
29 April 2022



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q22042324

F3-011-04/01-12

page 1 of 3



@clccalibration

REPORT OF CALIBRATION

FOR

NOMENCLATURE : pH METER
MANUFACTURER : HANNA INSTRUMENTS
MODEL / TYPE : HI5521/HI1131
SERIAL NO. : 04160019101/061334CN
DATE OF CALIBRATION : 27 April 2022

ENVIRONMENT CONDITIONS :

Temperature : $(25 \pm 2.5) ^\circ\text{C}$

Relative Humidity : $(50 \pm 15) \% \text{ RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **WI-305-128**. The calibration was performed by direct measurement with Certified Reference Material (CRM).

REFERENCE STANDARD USED :

1. pH Standard Solution, NIMT TRM CODE TRM-S-2003, TRM CODE TRM-S-2007.
2. pH Standard Solution, Control Company Catalog Number 06-664-260,11754256, Lot Number CC728484.

TRACEABILITY :

1. The measurements are traceable to International System of Units (SI) , through National Institute of Metrology (Thailand).
Lot Number. 160221 , 180121. Due Date 14 June 2022.
2. The measurements are traceable to International System of Units (SI) , through Control Company.
Certificate No. 4281-12405788 , Due Date 30 June 2023.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.

It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2021)"

Certificate No. **Q22042324**

F3-011-04/01-12

page 2 of 3



@clccalibration



CLC
Accredited
ISO/IEC 17025

CALIBRATION LABORATORY Co., LTD.

2/10-11,14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230

Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



NSC-TISI-TIS 17025
CALIBRATION 0059
CLC

CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment () adjustment

The table in the following gives the calibration results and associated measurement uncertainties of pH meter.

CALIBRATION DATA

pH METER RESULT @ 25 °C

Standard pH Buffer Solution (pH)	pH Meter Reading (pH)	pH Meter Reading (mV)	Correction (pH)	Uncertainty of pH Measurement (\pm pH)	k Factor
4.000	4.01	121.0	-0.010	0.023	2,87
6.996	7.01	-47.4	-0.014	0.015	2,06
10.007	10.04	-203.7	-0.033	0.100	2,25

Note. The Scope of Accredited TISI Certificate No. 19C087/0655 Issue 1 Page 79 of 111

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q22042324

F3-011-04/01-12

page 3 of 3



@clccalibration

CERTIFICATE OF CALIBRATION FOR

NOMENCLATURE : DIGITAL THERMOMETER
MANUFACTURER : HANNA INSTRUMENTS
MODEL / TYPE : HI5521/HI7662-W
SERIAL NO. : 04160019101/0615024N
CLID. NO. : 232202088
JOB CONTROL NO. : 220426042327

CUSTOMER : OKLA TESTING & CONSULTING SERVICE CO., LTD.
67/35-36, 3RD FLOOR, PHETKASEM 7/1 RD., WATTHAPRA,
BANGKOKYAI, BANGKOK 10600 THAILAND

DATE OF RECEIVED : 26 April 2022

DATE OF ISSUED : 29 April 2022

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Tanawan Seenam-Ngoen
Calibration Engineer

Approved By : Mongkol Yotsoontorn
Authorized Signatory
29 April 2022



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to
the International System of Units (SI)

Certificate No. Q22042327

F3-011-04/01-12

page 1 of 3



@clccalibration

REPORT OF CALIBRATION FOR

NOMENCLATURE : **DIGITAL THERMOMETER**
MANUFACTURER : **HANNA INSTRUMENTS**
MODEL / TYPE : **HI5521/HI7662-W**
SERIAL NO. : **04160019101/0615024N**
DATE OF CALIBRATION : **28 April 2022**

ENVIRONMENT CONDITIONS :

Temperature : $(23 \pm 2) ^\circ\text{C}$

Relative Humidity : $(55 \pm 10) \% \text{ RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **WI-305-187** based on **ASTM E 644-04** as calibration guidelines.

The calibration was performed by using Calibration Bath, Precision Thermometer and IPRT which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

1. Calibration Bath, Kambic Model OB-22/2 ULT S/N. 17115653.
2. Precision Thermometer, ASL Model F200-A-8 S/N. 014433/03.
3. IPRT, ASL Model T100-250-1D S/N. L0193A-1-1.

TRACEABILITY :

1. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22007520, Due Date 22 January 2023.
2. The measurements are traceable to International System of Units (SI), through Thailand Institute of Scientific and Technological Research (TISTR). Certificate No. PSL-T 0717/64, Due Date 14 June 2022.
3. The measurements are traceable to International System of Units (SI), through National Institute of Metrology (Thailand). Certificate No. TT-0121-21, Due Date 24 November 2022.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2,00$ which for a normal distribution corresponds to a coverage probability of approximately 95 %. It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2021)"

Certificate No. **Q22042327**

F3-011-04/01-12

page 2 of 3



@clccalibration



CLC
Accredited
ISO/IEC 17025

CALIBRATION LABORATORY Co., LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230

Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



NSC-TISI-TIS 17025
CALIBRATION 0059
CLC

CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment () adjustment

The DUC Reading were recorded and the means value were reported of five times measurement in the table below.

CALIBRATION DATA

CORRECTION OF TEMPERATURE [THERMISTOR]

Immersion depth (mm)	Actual Temperature (°C)	DUC Reading (°C)	Correction (°C)	Uncertainty \pm (°C)
105	24.00	24.1	- 0.10	0.07
	25.00	25.1	- 0.10	
	27.00	27.1	- 0.10	

Note. Probe \varnothing 3.5 mm

Materials : Metal Sheath.

The Scope of Accredited TISI Certificate No. 19C087/0655 Issue 1 Page 28 of 111

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q22042327

F3-011-04/01-12

page 3 of 3



@clccalibration



CLC
Accredited
ISO/IEC 17025

CALIBRATION LABORATORY Co., LTD.

2/10-11,14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230

Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



NSC-TISI-TIS 17025
CALIBRATION 0059
CLC

CERTIFICATE OF CALIBRATION

FOR

NOMENCLATURE : CONDUCTIVITY METER
MANUFACTURER : HANNA INSTRUMENTS
MODEL / TYPE : HI5521/HI76312
SERIAL NO. : 04160019101/0614117M
CLID. NO. : 272201302
JOB CONTROL NO. : 220426042325

CUSTOMER : OKLA TESTING & CONSULTING SERVICE CO., LTD.
67/35-36, 3RD FLOOR, PHETKASEM 7/1 RD., WATTHAPRA,
BANGKOKYAI, BANGKOK 10600 THAILAND

DATE OF RECEIVED : 26 April 2022

DATE OF ISSUED : 29 April 2022

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Prapaporn Khanchalee
Calibration Engineer



Approved By : Mongkol Yotsoontorn
Authorized Signatory
29 April 2022



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q22042325

F3-011-04/01-12

page 1 of 3



@clccalibration

REPORT OF CALIBRATION

FOR

NOMENCLATURE : CONDUCTIVITY METER
MANUFACTURER : HANNA INSTRUMENTS
MODEL / TYPE : HI5521/HI76312
SERIAL NO. : 04160019101/0614117M
DATE OF CALIBRATION : 28 April 2022

ENVIRONMENT CONDITIONS :

Temperature : $(25 \pm 2.5) ^\circ\text{C}$

Relative Humidity : $(50 \pm 15) \% \text{ RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **WI-305-130**. The calibration was performed by direct measurement with Certified Reference Material (CRM).

REFERENCE STANDARD USED :

Potassium Chloride Solution (nominal 0.147 mS/cm)

Potassium Chloride Solution (nominal 1.41 mS/cm)

Potassium Chloride Solution (nominal 12.8 mS/cm)

TRACEABILITY :

The measurements are traceable to International System of Units (SI) , through Merck Co., Ltd.

Certificate No. HC90696057 , HC02139203 , HC04515254. Due Date 31 August 2022 , 30 June 2023 , 30 November 2023.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2,00$ which for a normal distribution corresponds to a coverage probability of approximately 95 % .
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2021)"

Certificate No. **Q22042325**

F3-011-04/01-12

page 2 of 3



@clccalibration



CLC
Accredited
ISO/IEC 17025

CALIBRATION LABORATORY Co., LTD.

2/10-11,14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230

Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



NSC-TISI-TIS 17025
CALIBRATION 0059
CLC

CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment () adjustment

The table in the following gives the calibration results and associated measurement uncertainties of Conductivity Meter.

CALIBRATION DATA

Conductivity Solution Test @ 25°C

Standard Conductivity Solution	DUC Reading	Uncertainty of Measurement
146.00 μ S/cm	146.0 μ S/cm [Cell Constant 1.1165]	$\pm 2.10 \mu$ S/cm
1.412 mS/cm	1.412 mS/cm [Cell Constant 1.1200]	± 0.021 mS/cm
12.85 mS/cm	12.85 mS/cm [Cell Constant 1.1550]	± 0.19 mS/cm

Note. The Scope of Accredited TISI Certificate No. 19C087/0655 Issue 1 Page 79 of 111

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q22042325

F3-011-04/01-12

page 3 of 3



@clccalibration

CERTIFICATE OF CALIBRATION

FOR

NOMENCLATURE : DIGITAL THERMOMETER
MANUFACTURER : HANNA INSTRUMENTS
MODEL / TYPE : HI5521/HI7662-W
SERIAL NO. : 04160019101/0615024N
CLID. NO. : 232202088
JOB CONTROL NO. : 220426042327

CUSTOMER : OKLA TESTING & CONSULTING SERVICE CO., LTD.
67/35-36, 3RD FLOOR, PHETKASEM 7/1 RD., WATTHAPRA,
BANGKOKYAI, BANGKOK 10600 THAILAND

DATE OF RECEIVED : 26 April 2022

DATE OF ISSUED : 29 April 2022

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Tanawan Seenam-Ngoen

Calibration Engineer

Approved By :

Mongkol Yotsornom

Authorized Signatory

29 April 2022



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q22042327

F3-011-04/01-12

page 1 of 3



@clccalibration

REPORT OF CALIBRATION FOR

NOMENCLATURE : DIGITAL THERMOMETER
MANUFACTURER : HANNA INSTRUMENTS
MODEL / TYPE : HI5521/HI7662-W
SERIAL NO. : 04160019101/0615024N
DATE OF CALIBRATION : 28 April 2022

ENVIRONMENT CONDITIONS :

Temperature : $(23 \pm 2) ^\circ\text{C}$

Relative Humidity : $(55 \pm 10) \% \text{ RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **WI-305-187** based on **ASTM E 644-04** as calibration guidelines.

The calibration was performed by using Calibration Bath, Precision Thermometer and IPRT which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

1. Calibration Bath, Kambic Model OB-22/2 ULT S/N. 17115653.
2. Precision Thermometer, ASL Model F200-A-8 S/N. 014433/03.
3. IPRT, ASL Model T100-250-1D S/N. L0193A-1-1.

TRACEABILITY :

1. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22007520, Due Date 22 January 2023.

2. The measurements are traceable to International System of Units (SI), through Thailand Institute of Scientific and Technological Research (TISTR). Certificate No. PSL-T 0717/64, Due Date 14 June 2022.

3. The measurements are traceable to International System of Units (SI), through National Institute of Metrology (Thailand). Certificate No. TT-0121-21, Due Date 24 November 2022.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2,00$ which for a normal distribution corresponds to a coverage probability of approximately 95 %. It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2021)"

Certificate No. Q22042327

F3-011-04/01-12

page 2 of 3



CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment () adjustment

The DUC Reading were recorded and the means value were reported of five times measurement in the table below.

CALIBRATION DATA

CORRECTION OF TEMPERATURE [THERMISTOR]

Immersion depth (mm)	Actual Temperature (°C)	DUC Reading (°C)	Correction (°C)	Uncertainty \pm (°C)
105	24.00	24.1	- 0.10	0.07
	25.00	25.1	- 0.10	
	27.00	27.1	- 0.10	

Note. Probe \varnothing 3.5 mm

Materials : Metal Sheath.

The Scope of Accredited TISI Certificate No. 19C087/0655 Issue 1 Page 28 of 111

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q22042327

F3-011-04/01-12

page 3 of 3



CERTIFICATE OF CALIBRATION

FOR

NOMENCLATURE : MEASURING PIPETTE
MANUFACTURER : GLASSCO
MODEL / TYPE : 1 ml
SERIAL NO. : N/A[EM-MER01001/19]
CLID. NO. : 272201297
JOB CONTROL NO. : 220426042207

CUSTOMER : OKLA TESTING & CONSULTING SERVICE CO., LTD.
67/35-36, 3RD FLOOR, PHETKASEM 7/1 RD., WATTHAPRA,
BANGKOKYAI, BANGKOK 10600 THAILAND

DATE OF RECEIVED : 26 April 2022

DATE OF ISSUED : 03 May 2022

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Prapaporn Khanchalee
Calibration Engineer



Approved By : Mongkol Yotsoontorn
Authorized Signatory

03 May 2022



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the
International System of Units (SI)

Certificate No. Q22042207

F3-011-04/01-12

page 1 of 3



@clccalibration

REPORT OF CALIBRATION FOR

NOMENCLATURE : MEASURING PIPETTE
MANUFACTURER : GLASSCO
MODEL / TYPE : 1 ml
SERIAL NO. : N/A[EM-MER01001/19]
DATE OF CALIBRATION : 28 April 2022

ENVIRONMENT CONDITIONS :

Temperature : $(20 \pm 2.5) ^\circ\text{C}$

Relative Humidity : $(50 \pm 10) \% \text{ RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **WI-305-89** based on **ASTM E542-01** as calibration guidelines.
The calibration was performed by using Electronic Balance, Thermo-hygrograph, Barometer and Thermometer which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

1. Barometer, Barigo S/N.001.
2. Electronic Balance, Sartorius Model CPA224S S/N.23908487.
3. Thermo-hygrograph, Isuzu Model 3-3126 S/N.30760420.
4. Thermometer, Brannan S/N. 1.

TRACEABILITY :

1. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22011065, Due Date 03 February 2023.
2. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22020944, Due Date 03 March 2023.
3. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22007505, Due Date 26 January 2023.
4. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22017270, Due Date 21 February 2023.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2021)"

Certificate No. Q22042207

F3-011-04/01-12

page 2 of 3



@clccalibration

CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment () adjustment

The calibration was performed by applied volume to the Device Under Calibration (DUC) . The actual volume readings from STD were reported in average of seven times measurements.

CALIBRATION DATA

CORRECTION OF VOLUME

DUC Test point (ml)	Actual volume (ml)	Correction (ml)	Uncertainty \pm (ml)	Coverage factor k
*0.1	0.1006	+0.0006	0.0024	2,00
*0.5	0.4964	-0.0036	0.0024	2,00
1	1.0045	+0.0045	0.0024	2,00

Type of glassware : ☐ to Contain ☒ to Deliver

Note. * means Calibrations marked " Not TISI Accredited " in this Certificate have been included for completeness.

The Scope of Accredited TISI Certificate No. 19C087/0655 Issue 1 Page 79 of 111

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q22042207

F3-011-04/01-12

page 3 of 3



@clccalibration



CLC
Accredited
ISO/IEC 17025

CALIBRATION LABORATORY Co., LTD.

2/10-11,14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230

Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



NSC-TISI-TIS 17025
CALIBRATION 0059
CLC

CERTIFICATE OF CALIBRATION FOR

NOMENCLATURE : MEASURING PIPETTE
MANUFACTURER : GLASSCO
MODEL / TYPE : 5 ml
SERIAL NO. : N/A[EM-MER01001/18]
CLID. NO. : 272201296
JOB CONTROL NO. : 220426042206

CUSTOMER : OKLA TESTING & CONSULTING SERVICE CO., LTD.
67/35-36, 3RD FLOOR, PHETKASEM 7/1 RD., WATTHAPRA,
BANGKOKYAI, BANGKOK 10600 THAILAND

DATE OF RECEIVED : 26 April 2022

DATE OF ISSUED : 03 May 2022

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Prapaporn Khanchalee
Calibration Engineer



Approved By : Mongkol Yotsoontorn
Authorized Signatory

03 May 2022



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the
International System of Units (SI)

Certificate No. Q22042206

F3-011-04/01-12

page 1 of 3



@clccalibration

REPORT OF CALIBRATION FOR

NOMENCLATURE : **MEASURING PIPETTE**
MANUFACTURER : **GLASSCO**
MODEL / TYPE : **5 ml**
SERIAL NO. : **N/A[EM-MER01001/18]**
DATE OF CALIBRATION : **28 April 2022**

ENVIRONMENT CONDITIONS :

Temperature : $(20 \pm 2.5) ^\circ\text{C}$

Relative Humidity : $(50 \pm 10) \% \text{ RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **WI-305-89** based on **ASTM E542-01** as calibration guidelines.
The calibration was performed by using Electronic Balance, Thermo-hygrograph, Barometer and Thermometer which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

1. Barometer, Barigo S/N.001.
2. Electronic Balance, Sartorius Model CPA224S S/N.23908487.
3. Thermo-hygrograph, Isuzu Model 3-3126 S/N.30760420.
4. Thermometer, Brannan S/N. 1.

TRACEABILITY :

1. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22011065, Due Date 03 February 2023.
2. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22020944, Due Date 03 March 2023.
3. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22007505, Due Date 26 January 2023.
4. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22017270, Due Date 21 February 2023.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2021)"

Certificate No. Q22042206

F3-011-04/01-12

page 2 of 3



@clccalibration

CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment () adjustment

The calibration was performed by applied volume to the Device Under Calibration (DUC) . The actual volume readings from STD were reported in average of seven times measurements.

CALIBRATION DATA

CORRECTION OF VOLUME

DUC Test point (ml)	Actual volume (ml)	Correction (ml)	Uncertainty \pm (ml)	Coverage factor k
*0.5	0.4969	-0.0031	0.0024	2,00
2.5	2.4842	-0.0158	0.0029	2,00
5	4.9808	-0.0192	0.0029	2,00

Type of glassware : ☐ to Contain ☒ to Deliver

Note. * means Calibrations marked " Not TISI Accredited " in this Certificate have been included for completeness.

The Scope of Accredited TISI Certificate No. 19C087/0655 Issue 1 Page 79 of 111

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q22042206

F3-011-04/01-12

page 3 of 3



@clccalibration

CERTIFICATE OF CALIBRATION FOR

NOMENCLATURE : MEASURING PIPETTE
MANUFACTURER : GLASSCO
MODEL / TYPE : 10 ml
SERIAL NO. : N/A[EM-MER01001/17]
CLID. NO. : 272000237
JOB CONTROL NO. : 220426042205

CUSTOMER : OKLA TESTING & CONSULTING SERVICE CO., LTD.
67/35-36, 3RD FLOOR, PHETKASEM 7/1 RD., WATTHAPRA,
BANGKOKYAI, BANGKOK 10600 THAILAND

DATE OF RECEIVED : 26 April 2022

DATE OF ISSUED : 03 May 2022

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Prapaporn Khanchalee
Calibration Engineer



Approved By : Mongkol Yotsoontorn
Authorized Signatory

03 May 2022



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the
International System of Units (SI)

Certificate No. Q22042205

F3-011-04/01-12

page 1 of 3



@clccalibration

REPORT OF CALIBRATION FOR

NOMENCLATURE : MEASURING PIPETTE
MANUFACTURER : GLASSCO
MODEL / TYPE : 10 ml
SERIAL NO. : N/A[EM-MER01001/17]
DATE OF CALIBRATION : 28 April 2022

ENVIRONMENT CONDITIONS :

Temperature : $(20 \pm 2.5) ^\circ\text{C}$

Relative Humidity : $(50 \pm 10) \% \text{ RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **WI-305-89** based on **ASTM E542-01** as calibration guidelines.
The calibration was performed by using Electronic Balance, Thermo-hygrograph, Barometer and Thermometer which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

1. Barometer, Barigo S/N.001.
2. Electronic Balance, Sartorius Model CPA224S S/N.23908487.
3. Thermo-hygrograph, Isuzu Model 3-3126 S/N.30760420.
4. Thermometer, Brannan S/N. 1.

TRACEABILITY :

1. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22011065, Due Date 03 February 2023.
2. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22020944, Due Date 03 March 2023.
3. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22007505, Due Date 26 January 2023.
4. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22017270, Due Date 21 February 2023.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2021)"

Certificate No. Q22042205

F3-011-04/01-12



CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment () adjustment

The calibration was performed by applied volume to the Device Under Calibration (DUC) . The actual volume readings from STD were reported in average of seven times measurements.

CALIBRATION DATA

CORRECTION OF VOLUME

DUC Test point (ml)	Actual volume (ml)	Correction (ml)	Uncertainty \pm (ml)	Coverage factor k
1	0.9936	-0.0064	0.0024	2,00
5	4.9781	-0.0219	0.0029	2,00
10	9.9693	-0.0307	0.0039	2,00

Type of glassware : ☐ to Contain ☒ to Deliver

Note. The Scope of Accredited TISI Certificate No. 19C087/0655 Issue 1 Page 79 of 111

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q22042205

F3-011-04/01-12

page 3 of 3





CLC
Accredited
ISO/IEC 17025

CALIBRATION LABORATORY Co., LTD.

2/10-11,14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230

Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



NSC-TISI-TIS 17025
CALIBRATION 0059
CLC

CERTIFICATE OF CALIBRATION

FOR

NOMENCLATURE : VOLUMETRIC PIPETTE
MANUFACTURER : GLASSCO
MODEL / TYPE : 5 ml
SERIAL NO. : N/A[EM-Mbro1051/17]
CLID. NO. : 272101210
JOB CONTROL NO. : 220426042204

CUSTOMER : OKLA TESTING & CONSULTING SERVICE CO., LTD.
67/35-36, 3RD FLOOR, PHETKASEM 7/1 RD., WATTHAPRA,
BANGKOKYAI, BANGKOK 10600 THAILAND

DATE OF RECEIVED : 26 April 2022

DATE OF ISSUED : 03 May 2022

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Prapaporn Khanchalee
Calibration Engineer



Approved By : Mongkol Yotsoontorn
Authorized Signatory
03 May 2022



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the
International System of Units (SI)

Certificate No. Q22042204

F3-011-04/01-12

page 1 of 3



@clccalibration

REPORT OF CALIBRATION FOR

NOMENCLATURE : **VOLUMETRIC PIPETTE**
MANUFACTURER : **GLASSCO**
MODEL / TYPE : **5 ml**
SERIAL NO. : **N/A[EM-Mbro1051/17]**
DATE OF CALIBRATION : **27 April 2022**

ENVIRONMENT CONDITIONS :

Temperature : $(20 \pm 2.5) ^\circ\text{C}$

Relative Humidity : $(50 \pm 10) \% \text{ RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **WI-305-89** based on **ASTM E542-01** as calibration guidelines.
The calibration was performed by using Electronic Balance, Thermo-hygrograph, Barometer and Thermometer which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

1. Barometer, Barigo S/N.001.
2. Electronic Balance, Sartorius Model CPA224S S/N.23908487.
3. Thermo-hygrograph, Isuzu Model 3-3126 S/N.30760420.
4. Thermometer, Brannan S/N. 1.

TRACEABILITY :

1. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22011065, Due Date 03 February 2023.
2. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22020944, Due Date 03 March 2023.
3. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22007505, Due Date 26 January 2023.
4. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22017270, Due Date 21 February 2023.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.

It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2021)"

Certificate No. Q22042204

F3-011-04/01-12





CLC
Accredited
ISO/IEC 17025

CALIBRATION LABORATORY Co., LTD.

2/10-11,14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230

Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



NSC-TISI-TIS 17025
CALIBRATION 0059
CLC

CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment () adjustment

The calibration was performed by applied volume to the Device Under Calibration (DUC) . The actual volume readings from STD were reported in average of seven times measurements.

CALIBRATION DATA

CORRECTION OF VOLUME

DUC Test point (ml)	Actual volume (ml)	Correction (ml)	Uncertainty \pm (ml)	Coverage factor k
5	4.9817	-0.0183	0.0030	2,00

Type of glassware : ☐ to Contain ☒ to Deliver

Note. The Scope of Accredited TISI Certificate No. 19C087/0655 Issue 1 Page 78 of 111

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q22042204

F3-011-04/01-12

page 3 of 3



@clccalibration

CERTIFICATE OF CALIBRATION

FOR

NOMENCLATURE : VOLUMETRIC PIPETTE
MANUFACTURER : GLASSCO
MODEL / TYPE : 10 ml
SERIAL NO. : N/A[EM-Mbro1001/17]
CLID. NO. : 272101209
JOB CONTROL NO. : 220426042203

CUSTOMER : OKLA TESTING & CONSULTING SERVICE CO., LTD.
67/35-36, 3RD FLOOR, PHETKASEM 7/1 RD., WATTHAPRA,
BANGKOKYAI, BANGKOK 10600 THAILAND

DATE OF RECEIVED : 26 April 2022

DATE OF ISSUED : 03 May 2022

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By :

Prapaporn Khanchalee
Calibration Engineer



Approved By :

Mongkol Yotsoontorn
Authorized Signatory
03 May 2022



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q22042203

F3-011-04/01-12

page 1 of 3



@clccalibration

REPORT OF CALIBRATION FOR

NOMENCLATURE : VOLUMETRIC PIPETTE
MANUFACTURER : GLASSCO
MODEL / TYPE : 10 ml
SERIAL NO. : N/A[EM-Mbro1001/17]
DATE OF CALIBRATION : 27 April 2022

ENVIRONMENT CONDITIONS :

Temperature : $(20 \pm 2.5) ^\circ\text{C}$

Relative Humidity : $(50 \pm 10) \% \text{ RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **WI-305-89** based on **ASTM E542-01** as calibration guidelines.
The calibration was performed by using Electronic Balance, Thermo-hygrograph, Barometer and Thermometer which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

1. Barometer, Barigo S/N.001.
2. Electronic Balance, Sartorius Model CPA224S S/N.23908487.
3. Thermo-hygrograph, Isuzu Model 3-3126 S/N.30760420.
4. Thermometer, Brannan S/N. 1.

TRACEABILITY :

1. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22011065, Due Date 03 February 2023.
2. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22020944, Due Date 03 March 2023.
3. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22007505, Due Date 26 January 2023.
4. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22017270, Due Date 21 February 2023.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2021)"

Certificate No. Q22042203

F3-011-04/01-12

page 2 of 3



@clccalibration

CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment () adjustment

The calibration was performed by applied volume to the Device Under Calibration (DUC) . The actual volume readings from STD were reported in average of seven times measurements.

CALIBRATION DATA

CORRECTION OF VOLUME

DUC Test point (ml)	Actual volume (ml)	Correction (ml)	Uncertainty \pm (ml)	Coverage factor k
10	10.0007	+0.0007	0.0043	2,00

Type of glassware : ☐ to Contain ☒ to Deliver

Note. The Scope of Accredited TISI Certificate No. 19C087/0655 Issue 1 Page 78 of 111

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q22042203

F3-011-04/01-12

page 3 of 3



@clccalibration

CERTIFICATE OF CALIBRATION

FOR

NOMENCLATURE : VOLUMETRIC PIPETTE
MANUFACTURER : GLASSCO
MODEL / TYPE : 20 ml
SERIAL NO. : N/A[EM-VPP20201/17]
CLID. NO. : 272101208
JOB CONTROL NO. : 220426042202

CUSTOMER : OKLA TESTING & CONSULTING SERVICE CO., LTD.
67/35-36, 3RD FLOOR, PHETKASEM 7/1 RD., WATTHAPRA,
BANGKOKYAI, BANGKOK 10600 THAILAND

DATE OF RECEIVED : 26 April 2022

DATE OF ISSUED : 03 May 2022

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By :

Prapaporn Khanchalee
Calibration Engineer



Approved By :

Mongkol Yotsoontorn
Authorized Signatory
03 May 2022



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q22042202

F3-011-04/01-12

page 1 of 3



@clccalibration

REPORT OF CALIBRATION FOR

NOMENCLATURE : VOLUMETRIC PIPETTE
MANUFACTURER : GLASSCO
MODEL / TYPE : 20 ml
SERIAL NO. : N/A[EM-VPP20201/17]
DATE OF CALIBRATION : 27 April 2022

ENVIRONMENT CONDITIONS :

Temperature : $(20 \pm 2.5) ^\circ\text{C}$

Relative Humidity : $(50 \pm 10) \% \text{ RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **WI-305-89** based on **ASTM E542-01** as calibration guidelines.
The calibration was performed by using Electronic Balance, Thermo-hygrograph, Barometer and Thermometer which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

1. Barometer, Barigo S/N.001.
2. Electronic Balance, Sartorius Model CPA224S S/N.23908487.
3. Thermo-hygrograph, Isuzu Model 3-3126 S/N.30760420.
4. Thermometer, Brannan S/N. 1.

TRACEABILITY :

1. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22011065, Due Date 03 February 2023.
2. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22020944, Due Date 03 March 2023.
3. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22007505, Due Date 26 January 2023.
4. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22017270, Due Date 21 February 2023.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2021)"

Certificate No. Q22042202

F3-011-04/01-12

page 2 of 3



@clccalibration

CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment () adjustment

The calibration was performed by applied volume to the Device Under Calibration (DUC) . The actual volume readings from STD were reported in average of seven times measurements.

CALIBRATION DATA

CORRECTION OF VOLUME

DUC Test point (ml)	Actual volume (ml)	Correction (ml)	Uncertainty \pm (ml)	Coverage factor k
20	20.0094	+0.0094	0.0072	2,00

Type of glassware : ☐ to Contain ☒ to Deliver

Note. The Scope of Accredited TISI Certificate No. 19C087/0655 Issue 1 Page 78 of 111

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q22042202

F3-011-04/01-12

page 3 of 3



CERTIFICATE OF CALIBRATION

FOR

NOMENCLATURE : VOLUMETRIC PIPETTE
MANUFACTURER : HBG
MODEL / TYPE : 25 ml
SERIAL NO. : N/A[EM-VPP02501/17]
CLID. NO. : 272000238
JOB CONTROL NO. : 220426042201

CUSTOMER : OKLA TESTING & CONSULTING SERVICE CO., LTD.
67/35-36, 3RD FLOOR, PHETKASEM 7/1 RD., WATTHAPRA,
BANGKOKYAI, BANGKOK 10600 THAILAND

DATE OF RECEIVED : 26 April 2022

DATE OF ISSUED : 03 May 2022

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Prapaporn Khanchalee
Calibration Engineer



Approved By : Mongkol Yotsoontorn
Authorized Signatory
03 May 2022



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the
International System of Units (SI)

Certificate No. Q22042201

F3-011-04/01-12

page 1 of 3



@clccalibration

REPORT OF CALIBRATION FOR

NOMENCLATURE : **VOLUMETRIC PIPETTE**
MANUFACTURER : **HBG**
MODEL / TYPE : **25 ml**
SERIAL NO. : **N/A[EM-VPP02501/17]**
DATE OF CALIBRATION : **27 April 2022**

ENVIRONMENT CONDITIONS :

Temperature : $(20 \pm 2.5) ^\circ\text{C}$

Relative Humidity : $(50 \pm 10) \% \text{ RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **WI-305-89** based on **ASTM E542-01** as calibration guidelines.
The calibration was performed by using Electronic Balance, Thermo-hygrograph, Barometer and Thermometer which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

1. Barometer, Barigo S/N.001.
2. Electronic Balance, Sartorius Model CPA224S S/N.23908487.
3. Thermo-hygrograph, Isuzu Model 3-3126 S/N.30760420.
4. Thermometer, Brannan S/N. 1.

TRACEABILITY :

1. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22011065, Due Date 03 February 2023.
2. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22020944, Due Date 03 March 2023.
3. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22007505, Due Date 26 January 2023.
4. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22017270, Due Date 21 February 2023.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.

It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2021)"

Certificate No. Q22042201

F3-011-04/01-12

page 2 of 3



@clccalibration

CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment () adjustment

The calibration was performed by applied volume to the Device Under Calibration (DUC) . The actual volume readings from STD were reported in average of seven times measurements.

CALIBRATION DATA

CORRECTION OF VOLUME

DUC Test point (ml)	Actual volume (ml)	Correction (ml)	Uncertainty \pm (ml)	Coverage factor k
25	24.9629	-0.0371	0.0076	2,00

Type of glassware : ☐ to Contain ☒ to Deliver

Note. The Scope of Accredited TISI Certificate No. 19C087/0655 Issue 1 Page 78 of 111

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q22042201

F3-011-04/01-12

page 3 of 3



@clccalibration



CLC
Accredited
ISO/IEC 17025

CALIBRATION LABORATORY Co., LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230

Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



NSC-TISI-TIS 17025
CALIBRATION 0059
CLC

CERTIFICATE OF CALIBRATION

FOR

NOMENCLATURE : VOLUMETRIC FLASK
MANUFACTURER : SCI
MODEL / TYPE : 100 ml
SERIAL NO. : N/A[EM-VPP02501/17]
CLID. NO. : 272101212
JOB CONTROL NO. : 220426042200

CUSTOMER : OKLA TESTING & CONSULTING SERVICE CO., LTD.
67/35-36, 3RD FLOOR, PHETKASEM 7/1 RD., WATTHAPRA,
BANGKOKYAI, BANGKOK 10600 THAILAND

DATE OF RECEIVED : 26 April 2022

DATE OF ISSUED : 04 May 2022

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By :

Prapaporn Khanchalee

Calibration Engineer



Approved By :

Mongkol Yotsoontorn

Authorized Signatory

04 May 2022



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the
International System of Units (SI)

Certificate No. Q22042200

F3-011-04/01-12

page 1 of 3



@clccalibration



CLC
Accredited
ISO/IEC 17025

CALIBRATION LABORATORY Co., LTD.

2/10-11,14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



REPORT OF CALIBRATION

FOR

NOMENCLATURE : VOLUMETRIC FLASK
MANUFACTURER : SCI
MODEL / TYPE : 100 ml
SERIAL NO. : N/A[EM-VPP02501/17]
DATE OF CALIBRATION : 30 April 2022

ENVIRONMENT CONDITIONS :

Temperature : $(20 \pm 2.5) ^\circ\text{C}$

Relative Humidity : $(50 \pm 10) \% \text{RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **WI-305-88** based on **ASTM E542-01** as calibration guidelines.
The calibration was performed by using Electronic Balance, Thermo-hygrograph, Barometer and Thermometer which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

1. Barometer, Barigo S/N.001.
2. Electronic Balance, Sartorius Model CPA224S S/N.23908487.
3. Thermo-hygrograph, Isuzu Model 3-3126 S/N.30760420.
4. Thermometer, Brannan S/N. 1.

TRACEABILITY :

1. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22011065, Due Date 03 February 2023.
2. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22020944, Due Date 03 March 2023.
3. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22007505, Due Date 26 January 2023.
4. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22017270, Due Date 21 February 2023.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2021)"

Certificate No. Q22042200

F3-011-04/01-12

page 2 of 3



@clccalibration



CLC
Accredited
ISO/IEC 17025

CALIBRATION LABORATORY Co., LTD.

2/10-11,14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230

Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



NSC-TISI-TIS 17025
CALIBRATION 0059
CLC

CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment () adjustment

The calibration was performed by applied volume to the Device Under Calibration (DUC) . The actual volume readings from STD were reported in average of seven times measurements.

CALIBRATION DATA

CORRECTION OF VOLUME

DUC Test point (ml)	Actual volume (ml)	Correction (ml)	Uncertainty \pm (ml)	Coverage factor k
100	99.9637	-0.0363	0.0190	2,00

Type of glassware : ☒ to Contain ☐ to Deliver

Note. The Scope of Accredited TISI Certificate No. 19C087/0655 Issue 1 Page 78 of 111

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q22042200

F3-011-04/01-12

page 3 of 3



@clccalibration

CERTIFICATE OF CALIBRATION

FOR

NOMENCLATURE	:	VOLUMETRIC FLASK
MANUFACTURER	:	BOROSIL
MODEL / TYPE	:	500 ml
SERIAL NO.	:	N/A[EM-VPP02501/18]
CLID. NO.	:	272201295
JOB CONTROL NO.	:	220426042199

CUSTOMER : OKLA TESTING & CONSULTING SERVICE CO., LTD.
67/35-36, 3RD FLOOR, PHETKASEM 7/1 RD., WATTHAPRA,
BANGKOKYAI, BANGKOK 10600 THAILAND

DATE OF RECEIVED : 26 April 2022

DATE OF ISSUED : 04 May 2022

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Prapaporn Khanchalee
Calibration Engineer

Approved By : Mongkol Yotsoontorn
Authorized Signatory
04 May 2022



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q22042199

F3-01 1-04/01-12

page 1 of 3



@clccalibration

REPORT OF CALIBRATION FOR

NOMENCLATURE : VOLUMETRIC FLASK
MANUFACTURER : BOROSIL
MODEL / TYPE : 500 ml
SERIAL NO. : N/A[EM-VPP02501/18]
DATE OF CALIBRATION : 30 April 2022

ENVIRONMENT CONDITIONS :

Temperature : $(20 \pm 2.5) ^\circ\text{C}$

Relative Humidity : $(50 \pm 10) \% \text{RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **WI-305-88** based on **ASTM E542-01** as calibration guidelines.
The calibration was performed by using Electronic Balance, Thermo-hygrograph, Barometer and Thermometer which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

1. Barometer, Barigo S/N.001.
2. Electronic Balance, Sartorius Model Secura6102-1s S/N.0042104938.
3. Thermo-hygrograph, Isuzu Model 3-3126 S/N.30760420.
4. Thermometer, Brannan S/N. 1.

TRACEABILITY :

1. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22011065, Due Date 03 February 2023.
2. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q21110986, Due Date 08 December 2022.
3. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22007505, Due Date 26 January 2023.
4. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q22017270, Due Date 21 February 2023.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2021)"

Certificate No. Q22042199

F3-011-04/01-12

page 2 of 3



@clccalibration

CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment () adjustment

The calibration was performed by applied volume to the Device Under Calibration (DUC) . The actual volume readings from STD were reported in average of seven times measurements.

CALIBRATION DATA

CORRECTION OF VOLUME

DUC Test point (ml)	Actual volume (ml)	Correction (ml)	Uncertainty \pm (ml)	Coverage factor k
500	499.96	-0.04	0.08	2,00

Type of glassware : ☒ to Contain ☐ to Deliver

Note. The Scope of Accredited TISI Certificate No. 19C087/0655 Issue 1 Page 78 of 111

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q22042199

F3-011-04/01-12

page 3 of 3

