

ภาคผนวก ช

เอกสารสอบเทียบเครื่องมือที่ใช้ในการวิเคราะห์

CERTIFICATE OF CALIBRATION

Certificate No.: C0-2008004/21 Page 1 of total 4 pages


Customer
WATER ANALYSIS CENTER CO., LTD.
30/5 Soi Vipavadee 60, Vipavadee Rangsit Road,
Kwaeng Taladbangkhen, Khet Laksi, Bangkok 10210


Equipment
pH Meter
Manufacturer
METTLER TOLEDO
Model
SevenCompact
Serial No.
B327527211
ID No.
WWL 0068
Description
Range : 0 - 14 pH, Resolution : 0.01 pH

Environmental Conditions
Ambient Temperature: (20 ± 2) °C
Relative Humidity: (50 ± 10) %
Atmospheric Pressure:

Calibration Location
Jayhawks Laboratory (CL&GL)
Received Date
20 August 2021
Calibration Date
20 August 2021

Date of Issue
23 August 2021

Checked by

Act as Technical Manager
() (Krisyos K.) () (Sakda Y.)
() (Patiphan K.) () (Onnapa P.)
() (Pongsak H.) () (Nitiphong K.)
() (Kanung C.) () (Nonthachai K.)
() (Pramong P.) () (Noppoi P.)

Approved by

Representative of Managing Director
(Dr. Ekachai Puttibwong)

This calibration certificate shall not be reproduced other than in full except with the prior written approval of the Thai Heart Calibration Co., Ltd.

FE-169

REV.02 02/24/21

Certificate No.: C0-2008004/21

Page 2 of total 4 pages

Reference Method:

- The calibration method used was CP-178 based on an in-house method.
- This certificate can be traceable to the national standards, which is realized the shown measurement units according to the International System of Units (SI Units).

Reference Standard:

Type	pH Value	Lot No.	Due Date	Traceability
pH Standard Solution	4.01	081020	Feb. 1, 2022	NIMT
	7.01	020221	Dec. 25, 2021	
	10.00	091020	Jan. 19, 2022	

Type	Model	Serial No.	Certificate No.	Due Date	Traceability
Documenting Process Calibrator	753	3101007	10-0804001/21	Apr. 7, 2022	THC
Digital Thermometer with Sensor	1523 / 5622	1709138 / 4605984-005	10-1006001/21	Jun. 10, 2022	

Remark: This certificate is traceable to the International System of Unit (SI Unit) through:

- NIMT, National Institute of Metrology (Thailand).
- THC, Thai Heart Calibration Co., Ltd.

Measurement Results:

1. Function Simulated pH Meter

Standard Applied (mV)	Nominal Value (pH)	UUC Reading		Uncertainty (± mV)
		pH	mV	
177.48	4.00	4.01	177.4	0.060
0.00	7.00	7.00	0.0	0.060
-177.48	10.00	10.01	-177.4	0.060

UUC : Unit Under Calibration

Note : Adjust Curve to simulate pH (4.7,10)

Calibrated by Kittipong

REV.02 02/24/21

FE-169

Certificate No.: C0-2008004/21

Page 3 of total 4 pages

Measurement Results (Cont.):

2. Calibration of pH Electrode

pH Standard Solution (pH)	Measured Value		Uncertainty (± pH)
	(pH)	(mV)	
4.01	4.00	187.0	0.013
7.01	7.00	11.1	0.013
10.00	10.02	-161.6	0.013

Note : Adjust Curve to Buffer Solution pH (4.7,10)

Temperature stability of micro bath : $25 \pm 0.2^{\circ}\text{C}$

The above reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor $k = 2.00$, providing a level of confidence approximately 95%.

Calibrated by Kitipong
REV.02 02/24/21

FE-169

Certificate No.: C0-2008004/21

Page 4 of total 4 pages

Reference Method:

- The calibration method used was CP-096 based on an in-house method.
- The temperature scale used was an ITS-90.
- This certificate can be traceable to the national standards, which is realized the shown measurement units according to the International System of Units (SI Units).

Reference Standard Instruments:

Type	Model	Serial No.	Cert. No.	Due Date	Traceability
Thermometer Readout	1529-R	B7C853	20E3985	Nov. 9, 2021	TPA
Platinum Resistance Thermometer	5626	4853	C0A300046	Oct. 28, 2023	FLUKE
Liquid Bath	XORTS-40A	XO111019	10-0306002/21	Jun. 3, 2023	THC

Remark: This certificate is traceable to the International System of Unit (SI Unit) through:

- TPA, Technology Promotion Association (Thailand-Japan).
- FLUKE, Fluke Corporation, U.S.A.
- THC, Thai Heart Calibration Co., Ltd.

Measurement Results:

(X) Without Adjustment

Dimension of probe : Diameter 4 mm. Sensor Type : RTD (PT100)

Immersion Depth (mm.)	Standard Reading ($^{\circ}\text{C}$)	UUC Reading ($^{\circ}\text{C}$)	Correction ($^{\circ}\text{C}$)	Uncertainty ($\pm ^{\circ}\text{C}$)
120	22.00	22.0	0.00	0.058
120	25.00	25.0	0.00	0.058
120	28.00	28.0	0.00	0.058

UUC : Unit Under Calibration

The above reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor $k = 2.00$, providing a level of confidence approximately 95%.

- End of Certificate -

Calibrated by Pongsak
REV.02 02/24/21

FE-169



THAI HEART CALIBRATION CO., LTD.
112/1 Moo 5, Phase 5, Muang Samut Prakan 10250
Tel: 02-380-2182, 02-357-0815, 02-357-8964 Fax: 02-357-8837



ANAB
ACCREDITED
CALIBRATION LABORATORY
LC-2685

CERTIFICATE OF CALIBRATION

Certificate No.: CO-2107005/21 Page 1 of total 2 pages

Customer
WATER ANALYSIS CENTER CO., LTD.
30/5 Soi Viphavadee 60, Viphavadee Rangsit Road,
Kwaeng Taladbangkhen, Khet Laksi, Bangkok 10210

Equipment	Conductivity Meter
Manufacturer	EUTECH
Serial No.	2657889
Description	
	Model CON 2700
	ID No. WWL 0136

Environmental Conditions Ambient Temperature: $(20 \pm 2) ^\circ\text{C}$
Relative Humidity: $(50 \pm 10) \%$
Atmospheric Pressure: -

Calibration Location Jayhawks Laboratory (CL&GL)
Received Date 21 July 2021
Calibration Date 21 July 2021

Date of Issue 22 July 2021

Checked by

Approved by

Act as Technical Manager

Representative of Managing Director

()	(Krisyos K.)	()	(Sakda Y.)
()	(Patiphan K.)	()	(Onnapa P.)
()	(Pongsak H.)	()	(Nitiphong K.)
()	(Kanung C.)	()	(Nonthachai K.)
()	(Pramong P.)	()	(Noppol P.)

(Dr. Ekachai Puttitwong)

This calibration certificate shall not be reproduced other than in full except with the prior written approval of the Thai Heart Calibration Co., Ltd.

FE-169

REV.02.02/24/21

Certificate No : AD2012-017-0001

Environment : Ambient Temperature : $(25 \pm 2) ^\circ\text{C}$
Relative Humidity : $(50 \pm 15)\% \text{RH}$

STD Reading (mg/l)	UUC Reading Before (mg/l)	UUC Reading After (mg/l)	Error (mg/l)	Uncertainty (\pm mg/l)
9.046	9.07	-	0.024	0.013

STD = Standard

UUC = Unit Under Calibration

Description of UUC : Range 0.00 to 60.00 mg/l
Resolution 0.01 mg/l

Measurement Standards Used & Traceability :

The International System of Units (SI) through

MIT Certificate No. L2002-756.L2002-757 for Data Logger (Lutron Temperature & Humid & Baro) Serial No. B014887, Due 28-Feb-21

MIT Certificate No. L2001-629 for HI Accuracy Thermometer Serial No. 130508834, Due 07-Jan-21

End of Certificate

Page 2 of 2

Instrument : DO Meter
Model : DO-31P
Serial No. : 780065

Calibrate Procedure

- ☐ This instrument was calibrated by comparison with standard solution (PH/ORP)
☐ This instrument was calibrated by comparison with scattering plate value (Turbidity)
☐ This instrument was calibrated by comparison with conductivity (Conductivity)
☒ This instrument was calibrated by comparison with Sodium sulfite anhydrous (DO)
Condition of this result of calibration

1). Reference Standard Solution

Standard	Lot No	Batch	Cert. No.	Due Date
Sodium Sulfite Power	1.06657.0500	K52300357	-	31 Mar 2022

- 2). Traceability This certification is traceable to
☒ Merck KGaA 64271 Darmstadt
☐ DKK Corporation

Result Of Calibration

Standard Solution		Before Adjust		After Adjust	
(mg/l) at 26.0°C		Indicator	Error	Indicator	Error
Zero	0.00	0.10	+ 0.10	0.00	-
Span	7.99	8.21	+ 0.22	7.99	-

DO Electrode No. OE270AA(5) S/N 111F0029

Calibrated By P.
(Ms. Phanee Yooyen)
Technician

929,929/1 Soi Pattanakarn 30, Pattanakarn Rd., Suanluang, Suanluang, Bangkok 10260
Head Office : Tel. 02-319-9994 ext.1 Fax.02-318-4961 E-mail : also@automation.co.th
Rajong Branch : 1/15 Huayong Rd., A. Muang, Rayong 21150 Tel. 033-892-152 Fax. 033-892-345
Lamphun Branch : 122/5 M.4, T.Ban Klang, A.Muang, Lamphun 51000 Tel/Fax. 053-591-376
website : www.automation.co.th

Cert. No. WAC-065
Page 1 of 2

CERTIFICATE OF CALIBRATION

Instrument : DO Meter
Model : DO-31P
Serial No. : 780065
Manufacturer : TOA-DKK
Measuring Range : 0.00 ~ 20.00 mg/l

Customer : Water Analysis Center Co.,Ltd.
1/94 Moo.5 T.Kanham, A.U.-Thai
Ayuthaya 13210 Thailand

Date Of Received : 03 / 12 / 2021
Date Of Calibration : 03 / 12 / 2021

Ambient Condition : Temperature 24 °C
Humidity 47 % RH

Calibrated By : P.
(Ms. Phanee Yooyen)
Technician

Approved By : P.
(Mr. Nipon Phungsomsak)
Technical Manager

Date Of Issue : 03 / 12 / 2021

This Certificate may not be reproduced other than in full, except with the prior written approval of the head of the industrial instruments calibration center.

Certificate No.: MC 2107214

Page 2 of 3

The Reference Standard :

Description	Certificate No.	Serial No.	Due date
Data Acquisition/Switch Unit	MC 2009600	MY44095818	8 August 2021
With Thermocouple Type " T " ID. No.6/1 to 6/9			

This certificate is traceable to the international system of units maintained at:

- Master Calibration Co., Ltd. And Quality Reborn Co., Ltd.

1. Calibration Procedure:

This instrument was calibration according to TLAS G-20 by comparison with calibrated thermocouple type T under no load condition. The Thermocouples were placed on nine points and located one thermocouple in each of the eight corners of the chamber and was away from the each wall of 5 cm to 10 cm. And placed the ninth thermocouple within 2.5 cm of the geometric center of the chamber.

Temperature Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.

Temperature Stability - one-half of the greatest maximum difference of measured temperatures at any one sensor.

Overall Variation - The Difference of the maximum and minimum measured temperatures throughout observation.




Figure 1 : Sensor Installation Location

Checked by: *Thangorn*

[MCF-Q-077 ; Rev.6 ; Date : 22/04/2021]

Master Calibration Co., Ltd.

547 Soi Ratchadaniw, Kwang Samnong, Khet Huaykwang, Bangkok 10310
Tel. : (02) 274 2978-9, (02) 274 2987-8 Fax : (02) 274 2518, (02) 274 2989
Website : www.mastercalibration.com E-mail : calibrate@mastercalibration.com



**TEMPERATURE
CONTROLLER ENCLOSURES**

Certificate No.: MC 2107214

Page 1 of 3

Customer : Water Analysis Center Co., Ltd.
1/94 Moo 5, T.Kantham, A.U.-Thai, Ayutthaya 13210.

Reference Job No. : 21-1565 Received Date : 13 July 2021
Description : Refrigerator
Manufacturer : SANDENTINTERCOOL Model : SEC-1500SBD
Serial No. : SEC1500201A-0708-00304 ID. No. : WW1.0038
Marking : Additionally for the purpose of identification by this laboratory a label marked with this certificate number (MC 2107214) has been attached to the case.
Method : In-House calibration procedure MWI-T-033 this method is reference to TLAS G-20 "Temperature Controlled Enclosures"

Location of Calibration : Water Analysis Center Co., Ltd. : Laboratory.

Environmental Conditions : Ambient Temperature : (26.3) °C

Relative Humidity : (56.4 to 59.3) %

Date of Calibration : 13 July 2021 Date of Issue : 14 July 2021

Checked by: *Thangorn* Approved by: *Aittipong*
Thangorn Jinchacharoen Aittipong Kanjanasit
(Calibration Supervisor) (Technical Manager)

The uncertainties are for a confidence probability of approximately 95%

This certificate is issued in accordance with the conditions of accreditation granted by the National Standardization Council of Thailand-Office of the National Standardization Council that has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of Master Calibration Co., Ltd.

[MCF-Q-077 ; Rev.6 ; Date : 22/04/2021]

Certificate No.: MC 2107214

Page 3 of 3

2. Result of calibration :

Temperature Measurement Accuracy Test

Indicating Temperature (°C)	Measured Temperature (°C) at Spread Locations									Uncertainty (±°C)
	#1	#2	#3	#4	#5	#6	#7	#8	Ref. #9	
2.6	4.0	4.0	4.1	4.0	3.9	3.8	3.7	3.8	3.4	1.2

Chamber Characterization Result

Controller Temperature (°C)	Indicating Temperature (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
2.0	2.6	2.7	1.4	5.8

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95 %.

This report will certify of the calibrated equipment only.

End of Certificate

Checked by : Thangorn

[MCF-Q-077 ; Rev.6 ; Date : 22/04/2021]



TEMPERATURE CONTROLLER ENCLOSURES

Report No. : MC 2103787

Page 1 of 3
QR Code

Customer : Water Analysis Center Co., Ltd.
1/94 Moo 5, T.Kantham, A.U.-Thai, Ayutthaya 13210.

Reference Job No. : 21-0710 Received Date : 25 March 2021
Description : Oven
Manufacturer : Memmert Model : UF260
Serial No. : B620.0814 ID. No. : N/A
Marking : Additionally for the purpose of identification by this laboratory a label marked with this report number (MC 2103787) has been attached to the case.

Method : In-House calibration procedure MWI-T-033 this method is reference to

TLAS G-20 "Temperature Controlled Enclosures"

Location of Calibration : Water Analysis Center Co., Ltd. ; Laboratory.

Environmental Conditions : Ambient Temperature : (31.8 to 35.3) °C

Relative Humidity : (44.7 to 55.9) %

Date of Calibration : 25 March 2021 Date of Issue : 26 March 2021

Checked by : Thangorn Approved by : Aitipong
Thangorn Limchaicharoen Aitipong Kailanavasi
(Calibration Supervisor) (Technical Manager)

The uncertainties are for a confidence probability of approximately 95%

This certificate is issued in accordance with the condition of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full expect with the prior written approval of the issuing laboratory.

[MCF-Q-077 ; Rev.5 ; Date : 15/07/2014]

Continuation of Report No. : MC 2103787

Page 2 of 3

The Reference Standard :

Description	Report No.	Serial No.	Due date
Data Acquisition/Switch Unit	MC 2016027	MY41010916	10 January 2022
With Thermocouple Type "T" ID. No.17/1 to 17/9			

This certificate is traceable to the international system of units maintained at:

- Master Calibration Co., Ltd.


1. Calibration Procedure:

This Instrument was calibration according to TLAS G-20 by comparison with calibrated thermocouple type T under no load condition. The Thermocouples were placed on nine points and located one thermocouple in each of the eight corners of the chamber and was away from the each wall of 5 cm to 10 cm. And placed the ninth thermocouple within 2.3 cm of the geometric center of the chamber.

Temperature Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.

Temperature Stability - one-half of the greatest maximum difference of measured temperatures at any one sensor.

Overall Variation - The Difference of the maximum and minimum measured temperatures throughout observation.



Overall Ambient Temperature around the Chamber variation : 1.1 °C
Overall Line Voltage variation : 1.2 V
Chamber Size (W*H*D) : 65 cm x 80 cm x 50 cm

Checked by : Thanigorn

[MCF-Q-077 ; Rev.5 ; Date : 15/07/2014]

Continuation of Report No. : MC 2103787

Page 3 of 3

2. Result of calibration :

Temperature Measurement Accuracy Test

Indicating Temperature (°C)	Measured Temperature (°C) at Spread Locations									Uncertainty (±°C)
	#1	#2	#3	#4	#5	#6	#7	#8	Ref. #9	
104.0	103.7	103.7	103.7	103.9	104.2	104.3	104.3	104.3	104.0	0.67
180.0	179.1	179.1	179.0	179.2	180.4	180.5	180.6	180.6	180.2	0.99

Chamber Characterization Result

Controller Temperature (°C)	Indicating Temperature (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
104.0	104.0	0.27	0.44	1.0
180.0	180.0	0.29	1.31	1.9

3. Uncertainty of Measurement

The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95 %.

This report will certify of the calibrated equipment only.

End of Calibration Report

Checked by : Thanigorn

[MCF-Q-077 ; Rev.5 ; Date : 15/07/2014]

Continuation of Report No. : MC 2103787

The Reference Standard :

Description	Report No.	Serial No.	Due date
Data Acquisition/Switch Unit With Thermocouple Type "T" ID. No.17/1 to 17/9	MC 2016027	MY41010916	10 January 2022

Report No. : MC 2103787

Customer : Water Analysis Center Co., Ltd.
1/94 Moo 5, T.Kanham, A.U.-Thai, Ayuthaya 13210.

Reference Job No. : 21-0710 Received Date : 25 March 2021

Description : Oven

Manufacturer : Memmert Model : UF260

Serial No. : B620.0814 ID. No. : N/A

Marking : Additionally for the purpose of identification by this laboratory a label marked with this report number (MC 2103787) has been attached to the case.

Method : In-House calibration procedure MWI-T-033 this method is reference to

TLAS G-20 "Temperature Controlled Enclosures"

Location of Calibration : Water Analysis Center Co., Ltd. ; Laboratory.

Environmental Conditions : Ambient Temperature : (31.8 to 35.3) °C

Relative Humidity : (44.7 to 55.9) %

Date of Calibration : 25 March 2021 Date of Issue : 26 March 2021

Checked by : Thangorn Approved by : Aitipong
Thangorn Limchaichroen Aitipong Ka Janasawat
(Calibration Supervisor) (Technical Manager)

The uncertainties are for a confidence probability of approximately 95%

This certificate is issued in accordance with the condition of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full expect with the prior written approval of the issuing laboratory.

This certificate is traceable to the international system of units maintained at:

- Master Calibration Co., Ltd.

1. Calibration Procedure:

This Instrument was calibration according to TLAS G-20 by comparison with calibrated thermocouple type T under no load condition. The Thermocouples were placed on nine points and located one thermocouple in each of the eight corners of the chamber and was away from the each wall of 5 cm to 10 cm. And placed the ninth thermocouple within 2.5 cm of the geometric center of the chamber.

Temperature Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.

Temperature Stability - one-half of the greatest maximum difference of measured temperatures at any one sensor.

Overall Variation - The Difference of the maximum and minimum measured temperatures throughout observation.




Figure 1 : Sensor Installation Location

Checked by : Thangorn

Continuation of Report No. : MC-2103787

Page 3 of 3

2. Result of calibration :

Temperature Measurement Accuracy Test

Indicating Temperature (°C)	Measured Temperature (°C) at Spread Locations									Uncertainty (±°C)
	#1	#2	#3	#4	#5	#6	#7	#8	Ref. #9	
104.0	103.7	103.7	103.7	103.9	104.2	104.3	104.3	104.3	104.0	0.67
180.0	179.1	179.1	179.0	179.2	180.4	180.5	180.6	180.6	180.2	0.99

Chamber Characterization Result

Controller Temperature (°C)	Indicating Temperature (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
104.0	104.0	0.27	0.44	1.0
180.0	180.0	0.29	1.31	1.9

3. Uncertainty of Measurement

The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95 %.

This report will certify of the calibrated equipment only.

End of Calibration Report

Checked by : *Thongorn*

[MCF-Q-071 ; Rev.5 ; Date : 15/07/2014]



Certificate of Calibration

Equipment: Balance
Model: BL210S
Serial No. (or ID.): 15808131 (WWL 0022)
Manufacturer: Sartorius
Condition: In condition

Certificate No.: C01211841
Issued Date: 24 June 2021
Job No.: KSPR2107969
Page: 1 of 2

Customer: Water Analysis Center Co., Ltd.
1/94 Moo 5, Rojana Industrial Park, Rojana Road,
Tambol Kanham, Amphur U-Thai, Ayutthaya 13210 Thailand

Environment Condition: Temperature 27 °C ± 0.3 °C
Humidity 40 %RH ± 1.7 %RH

Calibration Place: Water Analysis Center Co., Ltd. (ฟองเครื่องชั่ง)
1/94 Moo 5, Rojana Industrial Park, Rojana Road,
Tambol Kanham, Amphur U-Thai, Ayutthaya 13210 Thailand

Calibration By: Mr. Phakapol Donnin
Calibration Date: 10 June 2021
The Method used: In house method, SPCC-WI-47, base on UKAS Lab 14
Traceability: This certificate is traceable to the SI Units maintained by National Institute of Metrology (NIMT), Thailand through SPC RT Co., Ltd. Certificate No. C02210017

SERT
บริษัท เซอร์ติฟายด์ อาร์ท จำกัด
SPC RT Co., Ltd.

Mr. Phakapol Donnin
(Mr. Phakapol Donnin)

Rungrod
(Mr. Rungrod Jenkitrakulchai)

Person in charge


Authorized signatory

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to International or national standard or other recognized national standard laboratories.
The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor ($k=2$) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).
These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of SPC RT Co., Ltd.

Calibration Results:

Without Adjustment

Eccentric Error: Weight to be 1/4 or 1/3 of Maximum capacity, taken from the center of the pan as a zero reference.



Nominal Test Value		50			(g)
Reference Points (g)					
A	B	C	D	E	
-	0.0000	-0.0001	-0.0001	0.0000	

Repeatability: Determination of the standard deviation of weighing balance., Readability

Nominal test value (g)	Standard Deviation
20	0.00004
200	0.00005

0.0001 (g)

Departure of indication from nominal value, Readability 0.0001 (g)

Nominal Value	Conventional Mass	Displayed Value	Correction of Balance	Uncertainty	k
(g)	(g)	(g)	(g)	(g)	
1	1.00001	1.0000	0.0000	0.00010	2.03
2	2.00002	2.0000	0.0000	0.00010	2.03
5	4.99989	5.0000	0.0000	0.00010	2.03
10	10.00000	10.0000	0.0000	0.00011	2.02
20	19.99999	20.0000	0.0000	0.00011	2.02
50	49.99997	50.0000	0.0000	0.00012	2.01
70	69.99996	70.0000	0.0000	0.00015	2.00
100	100.00000	100.0000	0.0000	0.00017	2.00
120	119.99999	120.0001	-0.0001	0.00021	2.00
150	149.99997	150.0000	0.0000	0.00023	2.00
200	199.99990	200.0003	-0.0004	0.00029	2.00

The End of Certificate



Bara Scientific
Education of Success

Bara Scientific Co., Ltd.
 368 U Chu Liang Building Floor7 Rama4 Road
 Silom Bangkok Bangkok Thailand 10500
 Tel: 02-8324300 Fax: 02-6376496-7
www.barascientific.com



N3C-T187-T19 17026

Certificate of Calibration

Number of Page(s) 1 of 3

Certificate No.	BSCC-LUV-135/21
Equipment	UV/VIS Spectrophotometer
Model	UV-1800
Manufacturer	SHIMADZU
Serial No.	A11B93405598CD
ID No.	WWVL0082
Date of receipt	30 April 2021
Date of calibration	30 April 2021
Date of issue	10 May 2021

Customer name Water Analysis Center Co., Ltd.

Address
1/94 Moo 5, T.Kantham, A.Uthai, Ayutthaya 13210

Temperature (29.9-32.1) °C (On site)
Humidity (48.7-52.6) %RH (On site)

Equipment condition Good Operation

Calibration Location Laboratory Room Water Analysis Center

Calibration Procedure In-house method WI-UV-702-01 based on ASTM E275-01

Traceability

Wavelength Accuracy is traceable to certificate No. 79670 and 79671

Photometric Accuracy is traceable to certificate No. 79672 and 79673

Stray Light is traceable to certificate No. 79669

The above certificate are traceable to SI unit through Sigma Scientific Ltd. (UKAS accredited calibration laboratory NO. 0656)

Calibrated by
Mr. Waruth Janphung

Approved by _____

Mr. Kanchit Choothep
Technical Manager

The above results are valid exclusively for the calibrated item(s) as mention in this report / certificate.
Advertising the report / certificate and publicity of the results are prohibited and also shall not be reproduced except in full, without written approval of the Bara Scientific Co., Ltd.



Bara Scientific Co., Ltd.
968 U Chu Liang Building Floor 7 Rama4 Road
Siam Bangkok Bangkok Thailand 10500
Tel : 02-6324300 Fax : 02-6375496-7
www.barascientific.com



Certificate of Calibration

Certificate No.

BSSC-UV-135/21

Number of Page(s)

2 of 3

Calibration Results:

1. Wavelength Accuracy

Certified Wavelength (nm)	UUC (nm)	Error (nm)	Uncertainty (±nm)
360.89	360.82	-0.07	0.18
418.53	418.71	0.18	0.18
446.38	446.38	0.00	0.18
453.67	453.58	-0.09	0.18
459.99	459.80	-0.19	0.18
638.00	638.10	0.10	0.18
431.22	431.38	0.16	0.18
513.39	513.57	0.18	0.18
528.82	528.82	-0.00	0.18
572.99	572.58	-0.41	0.18
585.25	585.37	0.12	0.18
684.50	684.65	0.15	0.18
741.02	741.14	0.12	0.18
879.41	879.29	-0.12	0.18

2. Photometric Accuracy (UV)

Wavelength (nm)	Certified Absorbance (A)	UUC (A)	Error (A)	Uncertainty (±A)
235	CNR	CNR	CNR	CNR
257	CNR	CNR	CNR	CNR
313	CNR	CNR	CNR	CNR
350	0.0000	0.0000	0.0000	0.0075
	0.6358	0.6310	-0.0048	0.0075

*CNR = Customer not request

The above results are valid exclusively for the calibrated item(s) as mention in this report / certificate. Advertising the report / Certificate and publicity of the results are prohibited and also shall not be reproduced except in full, without written approval of the Bara Scientific Co., Ltd.



Bara Scientific Co., Ltd.
968 U Chu Liang Building Floor 7 Rama4 Road
Siam Bangkok Bangkok Thailand 10500
Tel : 02-6324300 Fax : 02-6375496-7
www.barascientific.com



Certificate of Calibration

Certificate No.

BSSC-UV-135/21

Number of Page(s)

3 of 3

Calibration Results:

3. Photometric Accuracy (Visible)

Wavelength (nm)	Certified Absorbance (A)	UUC (A)	Error (A)	Uncertainty (±A)
420.0	0.0000	0.0000	0.0000	0.0042
	0.5473	0.5511	0.0038	0.0042
	0.7625	0.7655	0.0030	0.0042
	1.0484	1.0525	0.0041	0.0042
440.0	0.0000	0.0000	0.0000	0.0042
	0.5388	0.5401	0.0013	0.0042
	0.7446	0.7464	0.0018	0.0042
	1.0235	1.0259	0.0024	0.0042
465.0	0.0000	0.0000	0.0000	0.0042
	0.4873	0.4808	0.0065	0.0042
	0.6868	0.6882	0.0014	0.0042
	0.9433	0.9453	0.0020	0.0042
546.1	0.0000	0.0000	0.0000	0.0042
	0.5009	0.5027	0.0018	0.0042
	0.6952	0.6950	-0.0002	0.0042
	0.9558	0.9558	-0.0012	0.0042
590.0	0.0000	0.0000	0.0000	0.0042
	0.5292	0.5303	0.0011	0.0042
	0.7228	0.7218	-0.0010	0.0042
	0.9893	0.9874	-0.0019	0.0042
635.0	0.0000	0.0000	0.0000	0.0042
	0.5140	0.5151	0.0011	0.0042
	0.6902	0.6889	-0.0013	0.0042
	0.9539	0.9519	-0.0020	0.0042

*CNR = Customer not request

4. Stray Light*

Standard cut-off wavelength (nm)	Wavelength (nm)	Transmission (%T)	Absorbance (A)
200.87±0.11nm	200.40	0.9524	2.0215

The Stray light transmission reference is less than 1.0%T and Stray light absorbance reference is greater than 2.00A

*Stray Light not NSC-ONSC Accredited.

The measurement uncertainty is base on a standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

End of Certificate

The above results are valid exclusively for the calibrated item(s) as mention in this report / certificate. Advertising the report / Certificate and publicity of the results are prohibited and also shall not be reproduced except in full, without written approval of the Bara Scientific Co., Ltd.



บริษัท ไทยยูนิค จำกัด THAI UNIQUE CO., LTD.

80-82 ถนนประชาธิปไตย แขวงบางขุนพรหม เขตพระนคร กรุงเทพฯ 10200
80-82 Prachathipatai Rd., Bangkokumprorn, Pranakorn, Bangkok 10200

Tel. 0-2629-0191-6, 0-2280-1787, Fax. 0-2280-1788, E-mail : thawan@thaiunique.com, Website : www.thaiunique.com

PREVENTATIVE MAINTENANCE (PM) CHECK LIST

FOR ATOMIC ABSORPTION SPECTROMETER

Model & Serial Number: 240FS AA X 1741830004

Customer: Water Analysis Center Co., Ltd.

Date: 7 Apr 2021

Safety

- ☒ Flame, Inspect/replace o-ring nebulizer, spray chamber and burner
- ☒ Flame, Clean nebulizer, spray chamber and burner
- ☒ Flame, Check liquid trap interlock, burner interlock, pressure relief bung interlock and shield interlock

Furnace, Clean work head, electrode and shroud N/A

Furnace, Clean PSD and PSD tray N/A

Furnace, Check water pressure N/A

Check drain tube

Check exhaust system

Check gas pressure sensor interlock

Check and all gas hoses for SpectrAA

Clean computer control

Optics

- ☒ Inspect/Replace that external optics surfaces
- ☒ Check Wavelength Accuracy the copper line at 323.0-326.0 nm = 324.5 nm
- ☒ Check that PMT % Gain the copper at 324.8 nm, 4 mA, 0.5 nm slit width, Gain = 22.7 (should be $\leq 64\%$ or $\leq 380V$)
- ☒ Flame, Check D2 lamp is work



บริษัท ไทยยูนิค จำกัด THAI UNIQUE CO., LTD.

80-82 ถนนประชาธิปไตย แขวงบางขุนพรหม เขตพระนคร กรุงเทพฯ 10200
80-82 Prachathipatai Rd., Bangkokumprorn, Pranakorn, Bangkok 10200

Tel. 0-2629-0191-6, 0-2280-1787, Fax. 0-2280-1788, E-mail : thawan@thaiunique.com, Website : www.thaiunique.com

Electronics

- ☒ Check power supply voltage
- ☒ Check cables and connectors
- ☒ Check/Clean all boards in the instrument
- ☐ Furnace, Check camera and align** N/A

**Option for Graphite Zeeman only

Mechanisms

- ☒ Flame, Check the burner adjuster
- ☐ Furnace, Check PSD accessories N/A

Analytical performance

- ☒ Clear the sample compartment
- ☒ Flame, Check uptake rate form 7.2-10.6 mL per minute = 10.4 mL/min
- ☒ Test Photometric noise, STDV = 0.0002 Abs (should be ≤ 0.00050 Abs)
- ☒ Flame, Test high solids nebulizer setting use
- ☐ -Air/acet Cu 5 ppm = 0.76 Abs, and Precision (%RSD) = 0.5 % (should be > 0.55 Abs and $< 0.5\%$ RSD)

or

- ☐ -N2O/Acet Cu 5 ppm = _____ Abs, and Precision (%RSD) = _____ % (should be > 0.3 Abs and $< 0.5\%$ RSD)
- ☐ Furnace, Characteristic mass and sensitivity Cu 25 ppb = _____ Abs, and Precision (%RSD) = _____ % (should be ≥ 0.15 Abs and $\leq 4.0\%$ RSD)

SIGN :

Engineer

Don Hor
(Sungja Nakhon...)

Customer :

Kot
(...)

SVD Results Report

4

VARIAN

Report ID: 4

Diagnostic Start Time: 4/7/2021 11:21:56 AM

Diagnostic End Time: 4/7/2021 12:04:17 PM

Customer: Water Analysis Center Co., Ltd.

Service Engineer: Suriya Nacharoen

Contact Details: Kanitsara

Instrument Configuration

Configuration:

Serial Number: MY18230004
 Instrument Model: Varian AA140/240/280
 Turret Type: Automatic
 Number Of Lamps: 4
 Flame Instrument: True
 Mono Type: Automatic
 Gasbox Type: 'Y' Gas Box
 Furnace Instrument: True
 Auto Burner Adjuster: False
 Zeeman Present: False
 Mains Frequency: 50
 Internal Zeeman: False
 Firmware Version: 2.12
 Internal UltraAA: False
 Optics Type: Double Beam
 Photomultiplier Type: Normal(900nm)
 D2 BG Correction Fitted: True
 PWB Version: 181
 Boot Block Version: 2.02

EEPROM Data:

Instrument Run Hours: 16347.950
 D2 Run Hours: 4626.033
 Zero Wavelength Offset: 25.877
 D2 Serial Number: not set
 Mono Correction: -0.605
 D2 Install Date: 1/1/1970
 Flame Hours: 5481.417
 D2 Original Intensity: 1.000
 D2 Last Intensity: 400.000

Frequency:

Averaging Period: 30.0
 Datapoint Count: 20
 Upper Limit: 51.00
 Highest Measured Frequency: 50.00
 Average Frequency: 50.00
 Lower Limit: 48.00
 Lowest Measured Frequency: 50.00
 Result: **Passed**

Power Supply:


Averaging Period: 30.0
 Datapoint Count: 20
 Lower Limit (V) 10.80
 Upper Limit (V) 13.20
 Actual (V) 12.10
 Result: **Passed**
 -13.20
 -10.80
 -11.90
 Result: **Passed**
 4.50
 5.50
 5.00
 Result: **Passed**
 279.00
 341.00
 318.00
 Result: **Passed**
 310.00V Rail

Beam Balance:


Lamp Type: Copper
Lamp Socket Used: 3

Peak Selected: 324.80
Lamp Alignment: **Performed**

X' Lamp Screw



Y' Lamp Screw



Grating Squareness:

Lamp Element(s): Copper
Lamp Turret Position: 3
Lamp Current(mA): 4.00
Slit Width(nm): 0.5
1st Order Wavelength(nm): 324.80
Lamp Alignment: **Performed**

	Lower Limit (nm)	Actual (nm)	Upper Limit (nm)	Result:
Zero Order	-0.10	0.00	0.10	Passed
First Order	324.45	324.62	325.15	Passed
Second Order	649.23	649.54	649.97	Passed

Wavelength Repeatability:

Lamp Used: Copper
Peak Used(nm): 324.750
Connected to Socket: 3
Lamp Current(mA): 4
Slit Width(nm): 0.2
Slit Height: Normal

Lamp Alignment: **Performed**

Lower Limit(nm) 324.551
Upper Limit(nm) 324.671
(Approach from Zero Order)
Sample 1: 324.611
Sample 2: 324.611
Sample 3: 324.611
Sample 4: 324.611
Sample 5: 324.615
Sample 6: 324.611
Sample 7: 324.615
Sample 8: 324.611
Sample 9: 324.615
Sample 10: 324.611

Mean: 324.613
Standard Deviation: 0.002

Result: **Passed**

Mechanical

Wavelength Drive:

Passed

Slit Drive:

Passed

Turret Drive:

Passed

Auto Burner Adjuster Drive:

Untested

Miscellaneous

Signal Processing Linearity:

Calculate Mode: New Calc Mode

	Lower Limit	Actual	Upper Limit	Result:
S0	114	268	297	Passed
S1	156	166	191	Passed
S2	271	300	332	Passed
S3	474	516	579	Passed
S4	825	933	1008	Passed
S5	1435	1555	1754	Passed
S6	2498	2802	3053	Passed
S7	4347	4795	5313	Passed

Interlocks:

Burner Fitted:	Working
N2O Burner Fitted:	Working
Flame Shield Closed:	Working
Gas Control Fitted:	Untested
Pressure Release Bung Fitted:	Working
Liquid Trap Fitted:	Working

Flame Detect:	Working
GCU Active:	Working
Oxidant Pressure:	Working
Oxidant Changeover:	Working
Ignition:	Working

Auto Lamp Recognition:

Lamp 1: 42 - Potassium (K)
Lamp 2: 53 - Sodium (Na)
Lamp 3: 14 - Copper (Cu)
Lamp 4: 37 - Nickel (Ni)

Lamp 5: Not Supported
Lamp 6: Not Supported
Lamp 7: Not Supported
Lamp 8: Not Supported

Result: Passed

GTA Temperature Monitoring:

Not Performed

Notes:

C2104SU29 PM 1/2

Signatures:

Water Analysis Center Co., Ltd Date

Suriya Nacharoen

Date



บริษัท ไทยยูนิค จำกัด THAI UNIQUE CO., LTD.

80-82 ถนนประชาวิทย์ แขวงบางขุนพรหม เขตพระนคร กรุงเทพฯ 10200
80-82 Prachathipat Rd., Bangkokphrom, Pranakorn, Bangkok 10200

Tel. 0-2629-0191-6, 0-2280-1787, Fax. 0-2280-1788, E-mail : thawat@thaiunique.com, Website : www.thaiunique.com

PREVENTATIVE MAINTENANCE (PM) CHECK LIST

FOR ATOMIC ABSORPTION SPECTROMETER

Model & Serial Number: AA 240FS & AA0911 M073

Customer: Water Analysis Center Co., Ltd.

Date: 04 Feb 2021

Safety

- ☒ Flame, Inspect/replace o-ring nebulizer, spray chamber and burner
- ☒ Flame, Clean nebulizer, spray chamber and burner
- ☒ Flame, Check liquid trap interlock, burner interlock, pressure relief bung interlock and shield interlock
- ☐ Furnace, Clean work head, electrode and shroud N/A
- ☐ Furnace, Clean PSD and PSD tray N/A
- ☐ Furnace, Check water pressure N/A
- ☒ Check drain tube
- ☒ Check exhaust system
- ☒ Check gas pressure sensor interlock
- ☒ Check and all gas hoses for SpectraAA
- ☒ Clean computer control

Optics

- ☒ Inspect/Replace that external optics surfaces
- ☒ Check Wavelength Accuracy the copper line at 323.0-326.0 nm = 324.8 nm
- ☒ Check that PMT % Gain the copper at 324.8 nm, 4 mA, 0.5 nm slit width, Gain = 56.7 (should be $\leq 64\%$ or $\leq 380V$)
- ☒ Flame, Check D2 lamp is work



บริษัท ไทยยูนิค จำกัด THAI UNIQUE CO., LTD.

80-82 ถนนประชาวิทย์ แขวงบางขุนพรหม เขตพระนคร กรุงเทพฯ 10200
80-82 Prachathipat Rd., Bangkokphrom, Pranakorn, Bangkok 10200

Tel. 0-2629-0191-6, 0-2280-1787, Fax. 0-2280-1788, E-mail : thawat@thaiunique.com, Website : www.thaiunique.com

Electronics

- ☒ Check power supply voltage
- ☒ Check cables and connectors
- ☒ Check/Clean all boards in the instrument
- ☐ Furnace, Check camera and align** N/A

**Option for Graphite Zeeman only

Mechanisms

- ☒ Flame, Check the burner adjuster
- ☐ Furnace, Check PSD accessories N/A


Analytical performance

- ☒ Clear the sample compartment
- ☒ Flame, Check uptake rate form 7.2-10.6 mL per minute = 9.8 mL/min
- ☒ Test Photometric noise, STDV = 0.0001 Abs (should be ≤ 0.00050 Abs)
- ☒ Flame, Test high solids nebulizer setting use
 - Air/acet Cu 5 ppm = 0.75 Abs, and Precision (%RSD) = 0.5 % (should be > 0.55 Abs and $< 0.5\%$ RSD)
 - or
 - N2O/Acet Cu 5 ppm = Abs, and Precision (%RSD) = % (should be > 0.3 Abs and $< 0.5\%$ RSD)
- ☐ Furnace, Characteristic mass and sensitivity Cu 25 ppb = Abs, and Precision (%RSD) = % (should be ≥ 0.15 Abs and $\leq 4.0\%$ RSD)

SIGN :

Engineer (Sriyapa Mahachareon) (นายสุวิภา มหัชชาเรณ)

Customer : (นายสุวิภา มหัชชาเรณ)


VARIAN

SVD Results Report

Report ID:3 Diagnostic Start Time:05-Dec-20 10:38:28 AM Diagnostic End Time:04-Feb-21 11:58:44 AM

Customer: Water Analysis Center Co., Ltd. Service Engineer: SuriyaNachaporn

Address: Prana khron si ayuthaya Contact Details: K_Kanitsara

Instrument Configuration

Configuration:

Serial Number: AA0911M073

Instrument Model: Varian AA140/240/280

Flame Instrument: True

Furnace Instrument: True

Zeeman Present: False

Internal Zeeman: False

Internal UltraAA: False

Optics Type: Double Beam

D2 BG Correction Fitted: True

Boot Block Version: 2.02

Turret Type: Automatic

Number Of Lamps: 4

Mono Type: Automatic

Gasbox Type: 'Y' Gas Box

Auto Burner Adjuster: False

Mains Frequency: 50

Firmware Version: 2.12

Photomultiplier Type: Normal(900nm)

PWB Version: 181

EEPROM Data:

Instrument Run Hours: 20613.650

Zero Wavelength Offset: -18.731

Mono Correction: -0.370

Flame Hours: 4025.000

D2 Run Hours: 2906.000

D2 Serial Number: not set 1

D2 Install Date: 01-Jan-70

D2 Original Intensity: 1.000

D2 Last Intensity: 661.000

Frequency:

Averaging Period: 30.0

Datapoint Count: 20

Upper Limit: 51.00

Lower Limit: 49.00

Average Frequency: 50.00

Highest Measured Frequency: 50.00

Lowest Measured Frequency: 50.00

Result: **Passed**

Power Supply:

Averaging Period: 30.0

Datapoint Count: 20

	Lower Limit (V)	Actual (V)	Upper Limit (V)	Result:
12.00V Rail	10.80	12.20	13.20	Passed
-12.00V Rail	-13.20	-12.00	-10.80	Passed
5.00V Rail	4.50	5.10	5.50	Passed
310.00V Rail	279.00	318.00	341.00	Passed

Optics

Beam Balance:

Lamp Type: Copper

Lamp Socket Used: 3

Peak Selected: 324.80

Lamp Alignment:

Performed

X' Lamp Screw

Sample Peak

0.748

80% Sample Peak Mark

0.748

Upper Limit

1.145

Ref Pt 1:

1.070

Lower Limit

0.995

Ref Pt 2:

1.073

X Lamp Screw Result:

Passed

Y' Lamp Screw

Sample Peak

0.778

80% Sample Peak Mark

0.778

Upper Limit

1.139

Ref Pt 1:

1.085

Lower Limit

0.990

Ref Pt 2:

1.085

Y Lamp Screw Result:

Passed

Lamp Element(s): Copper

Lamp Turret Position: 3

Lamp Current(mA): 4.00

Slit Width(nm): 0.5

1st Order Wavelength(nm): 324.80

Lamp Alignment:

Performed

	Lower Limit (nm)	Actual (nm)	Upper Limit (nm)	Result:
Zero Order	-0.10	0.00	0.10	<div>Passed</div>
First Order	324.45	324.73	325.15	<div>Passed</div>
Second Order	649.23	649.56	649.97	<div>Passed</div>

Wavelength Repeatability:

Lamp Used: Copper

Peak Used(nm): 324.750

Connected to Socket: 3

Lamp Current(mA): 4

Slit Width(nm): 0.2

Slit Height: Normal

Lamp Alignment:

Performed

Lower Limit(nm)

324.751

Upper Limit(nm)

324.871

(Approach from Zero Order)

Sample 1: 324.811

Sample 3: 324.811

Sample 5: 324.815

Sample 7: 324.815

Sample 9: 324.815

(Approach from end)

Sample 2: 324.807

Sample 4: 324.811

Sample 6: 324.811

Sample 8: 324.811

Sample 10: 324.815

Mean: 324.812

Standard Deviation: 0.003

Result:

Passed

Mechanical

Wavelength Drive:

Passed

Slit Drive:

Passed

Turret Drive:

Passed

Auto Burner Adjuster Drive:

Untested

Miscellaneous

Signal Processing Linearity:

Calculate Mode: New Calc Mode

	Lower Limit	Actual	Upper Limit	Result:
S0	114	248	297	Passed
S1	158	165	181	Passed
S2	271	294	332	Passed
S3	474	506	579	Passed
S4	825	908	1008	Passed
S5	1435	1516	1754	Passed
S6	2498	2723	3053	Passed
S7	4347	4681	5313	Passed

Interlocks:

Burner Fitted:

Working

N2O Burner Fitted:

Working

Flame Shield Closed:

Working

Gas Control Fitted:

Untested

Pressure Release Bung Fitted:

Working

Liquid Trap Fitted:

Working

Flame Detect:

Working

GCU Active:

Working

Oxidant Pressure:

Working

Oxidant Changeover:

Working

Ignition:

Working

Auto Lamp Recognition:

Lamp 1: 50 - Selenium (Se)

Lamp 2: 3 - Arsenic (As)

Lamp 3: 14 - Copper (Cu)

Lamp 4: 42 - Potassium (K)

Lamp 5: Not Supported

Lamp 6: Not Supported

Lamp 7: Not Supported

Lamp 8: Not Supported

Result:

Passed

GTA Temperature Monitoring:

Not Performed

Notes:

C2102SU11 PM1/2

Signatures:

Suriya Nacharoen

04 Feb 21

Date

Water Analysis Center Co., Ltd.

Date

BSC Certification Test Report

Page 1 of 6

Certificate No. : M0979/21
Customer Name : LABORATORY WATER ANALYSIS CENTER COMPANY LIMITED
Customer Address : 1/94 Moo 5 T.Kanbarn, A.U.-Thai,
Phra Nakhon Si Ayutthaya 13210

Equipment : Biological Safety Cabinet **Class II Type A2**

Manufacturer : Microtech

Model : V6-T

Serial No : 0972

ID No. : WWL0084

Were in accordance with ☒ EN 12469 ☐ NSF 49 ☐ Manufacturer's specification

Test Date : 23/09/2021

Due Date : 23/09/2022 **or after HEPA filters are replaced or until is moved**

Test by : Mr. Puwadol Keawkia

Approved by :

(Mr. Kridsada Thinhuaet)
Authorized Signatory

Issued Date : 24/09/2021

This calibration certificate documents the traceability to national standards, which realize the unit of measurement according to the International System of Units (SI).

This certificate may not be reproduced other than in full except with the prior written approval of the Megafil Company Limited.

Page 2 of 6

Certificate No. : M0979/21

Procedure Used :
: European Standard EN12469 : 2000 has the status of British Standard,
Biotechnology Performance criteria for microbiological safety cabinets.
: NSF International Standard / American National Standard NSF / ANSI 49-2008
Biosafety Cabinet : Design, Construction, Performance and Field Certification.
: Australian Standard : AS 1807.23-2000 Determination of intensity of radiation
from germicidal ultraviolet lamps.
: Manufacturer's specification.

1. Downflow velocity test.

Measurement Information

No. of Rows	No. of Readings	Grid Spacing Front-Back	Grid Spacing Side-Side	Probe height Above sash
2	8	1/4, 3/4	1/8, 3/8	100 mm.

Measurement Data.

0.34	0.37	0.36	0.35
0.32	0.33	0.32	0.34

Average velocity 0.34 m/s (67 FPM.) **Velocity range** 0.25-0.50 m/s (49-98 FPM.)

Uniformity(EN: +/-20%avg. 0.27 - 0.41 m/s (54 - 80 FPM.)

Supply filter dimension 24 x 72 (inch x inch) **Supply filter area** 10.69 SQ.FT

Downflow volume (Q) 716 CFM.

Result Summary ☒ Pass ☐ Fail

Equipment used : Thermo Anemometer **Model** 425 **S/N :** 03004786 **Calibration date :** 23/02/2021

Certificate No. : M0979/21

2. Inflow velocity test.

Select method : ☐ DIM ☒ Exhaust velocity. ☐ MFG's Specifications

0.57	0.59	0.56	0.61	0.63
0.58	0.6	0.56	0.59	0.58
0.61	0.57	0.56	0.54	0.56
0.59	0.56	0.62	0.59	0.62
0.59	0.57	0.63	0.59	0.58

Average Inflow velocity 0.50 m/s (98 FPM.) Velocity range 20.40 m/s (>79 FPM.)

Inflow dimension 8 x 72 (inch x inch) Inflow area 4.00 SQ.FT

Inflow volume(Q) 392 CFM

Result Summary ☒ Pass ☐ Fail

Adjustments Required ☐ Fan Speed ☐ Damper

Equipment used : Thermo Anemometer Model 425 S/N : 03004786 Calibration date 23/02/2021

3. HEPA filter leak test.

Measurement Data

HEPA Filter	PAO Upstream Conc.(calculated)	Specification	Measured leak penetration
Supply HEPA Filter	18 µg/L	<0.003%	<0.003%
Exhaust HEPA Filter	18 µg/L	<0.003%	<0.003%

Certificate No. : M0979/21

Leak location

Supply HEPA Filter

Back



Exhaust HEPA Filter

Back



Result Summary ☒ Pass ☐ Fail

Equipment used : Aerosol Photometer Model TDA-2H S/N : 21683 Calibration date 24/02/2021

Equipment used : Smoke Generator Model TDA-6C S/N : 21623

4. Airflow smoke patterns test

Measurement Information

- Downflow Pattern test : Smoke shall be passed from one end of the cabinet to the other, along the centerline of the work surface, at a height of 4 inch (10 cm) above the top of the access opening.
- View screen retention test : Smoke shall be passed from one end of the cabinet to the other, 1.0 in (2.5 cm) behind the view screen, at a height 6.0 inch (15 cm) above the top of the access opening.
- Work opening edge retention test : Smoke shall be passed along the entire perimeter of the work opening. Particular attention should be paid to corners and vertical edges.
- Sash/window seal test : Smoke shall be passed up the inside of the window 2 in (5 cm) from the sides and along the top of the work area.

Certificate No. : M0979/21

Result Summary

Downflow Pattern test	<input checked="" type="checkbox"/> Accept	<input type="checkbox"/> Non-Conforming
View screen retention test	<input checked="" type="checkbox"/> Accept	<input type="checkbox"/> Non-Conforming
Work opening edge retention test	<input checked="" type="checkbox"/> Accept	<input type="checkbox"/> Non-Conforming
Sash/window seal test	<input checked="" type="checkbox"/> Accept	<input type="checkbox"/> Non-Conforming

5. Site Installation

Sash Alarm.	<input type="checkbox"/> Pass	<input type="checkbox"/> Fail	<input checked="" type="checkbox"/> N/A
Interlock System.	<input type="checkbox"/> Pass	<input type="checkbox"/> Fail	<input checked="" type="checkbox"/> N/A
Exhaust System Performance	<input type="checkbox"/> Pass	<input type="checkbox"/> Fail	<input checked="" type="checkbox"/> N/A

Remark / Recommendation

ระบบ Site installation ไม่มีการตรวจสอบ เนื่องจากตู้ไม่ฟังก์ชัน

6. Illumination Test (Lighting) : Option

Lighting should be adequate for safe working within the cabinet. Illumination measured at the work surface.

Lux		
632	1000	997 630
947	1456	1449 921

Remark :

Certificate No. : M0979/21

7. Ultraviolet Lamp Test (UV) : Option

Ultraviolet radiation where UV Lamp are fitted, the intensity of radiation at a wavelength of 254 nm. Shall be not less than 400 mW/m² when measures at work floor surface.

mW/m ²		
740	1580	1570 750
480	1040	1020 480

Remark :

-000-



63/14-15, 67/35-36, Soi Petchkasem 7/1, Petchkasem Rd,
Wattapra, Bangkokyai, Bangkok 10600 Thailand.

Tel.: (66) 02-8680812#13 Fax.: (66) 02-8680860 www.jiranatee.com

Continuation of Certificate of Calibration Number

Certificate No: WS-01022022

Page 2 of 2 Pages

Result of calibration: ☒ Without adjustment ☐ With adjustment

Result of calibration: ☒ Without adjustment ☐

Result of calibration: ☒ Without adjustment ☐ With adjustment

Result of calibration: ☒ Without adjustment ☐ With adjustment

Result of calibration: ☒ Without adjustment ☐ With adjustment

Result of calibration: ☒ Without adjustment ☐ With adjustment

Result of calibration: ☒ Without adjustment ☐ With adjustment

Result of calibration: ☒ Without adjustment ☐ With adjustment

Result of calibration: ☒ Without adjustment ☐ With adjustment

Result of calibration: ☒ Without adjustment ☐ With adjustment

Result of calibration: ☒ Without adjustment ☐ With adjustment

Result of calibration: ☒ Without adjustment ☐ With adjustment

Result of calibration: ☒ Without adjustment ☐ With adjustment

Result of calibration: ☒ Without adjustment ☐ With adjustment

Result of calibration: ☒ Without adjustment ☐ With adjustment

Result of calibration: ☒ Without adjustment ☐ With adjustment

Result of calibration: ☒ Without adjustment ☐ With adjustment

Result of calibration: ☒ Without adjustment ☐ With adjustment

Appendix 1: Instrumentations

Appendix 1: Instrumentations

Appendix 1: Instrumentations

Appendix 1: Instrumentations

1991c, 2022.

1991c, 2022.

Approved Signature: _____

Approved Signature: _____

Approved Signature: _____

Approved Signature: _____

End of certificate of calibration



THIS CERTIFICATE MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.



63/14-15,67/35-36, Soi Petchkasem 7/1, Petchkasem Rd,
Wattapra, Bangkokkai, Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranatee.com

CERTIFICATE OF CALIBRATION

Certificate No.: WD-01022022
Page 1 of 2 pages

Measurement Item : Wind direction sensor with data logger
Manufacturer : Data logger: Novatek
: Wind direction sensor: Novatek
Model/Type : Date logger: 200-WS-25LR
: Wind direction sensor: WS-02P
Serial Number : Data logger: A5040
Wind direction sensor: R6-040
ID No : Date logger: -
: Wind direction sensor:
Customer : Water Analysis Center Co., Ltd.
: 94/1 Moo 5, Thanham, A-Uthai, Ayutthaya 13210

Environmental Condition:
The measurement was carried out in an ambient temperature of (23±2) °C, and relative humidity of (40±10) %.

Measurement Method:
The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and line laser is used for axis control. The measurement were taken at 45° intervals in clockwise and counterclockwise directions.

Note: The UUC was warmed up for 1 hour prior to the calibration being performed

Traceability:
The measurement results are traceable to the international system of units (SI) through Certificate No: Q2108001-4, Certificate No: RW344/0025.

Measurement Date : FEB 18, 2022.
Issued Date : FEB 21, 2022.

Performed by
☒ Mr. Somsak Thaisiad
☐ Mrs. Oranai Wattanasak

Approved Signatory:
Mr. Panyee Boonphong
Calibration Department Manager

JIRANTEE ASSOCIATES CO., LTD.

THIS CERTIFICATE REPORT CAN NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.



63/14-15,67/35-36, Soi Petchkasem 7/1, Petchkasem Rd,
Wattapra, Bangkokkai, Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranatee.com

Continuation of Certificate of Calibration Number

Certificate No.: WD-01022022
Pages 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment
Calibration in the range of 0 - 360° at a calibration interval of 45°.
The results of calibration and associated measurement uncertainties are reported in table below.

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC* Reading (°)	Error (°)	Uncertainty ±(°)
1		0/360	360	359	-1	3.0
2		45	45	41	-4	3.0
3		90	90	67	-3	3.0
4	Clockwise	135	135	135	0	3.0
5		180	180	182	2	3.0
6		225	225	227	2	3.0
7		270	270	273	3	3.0
8		315	315	319	4	3.0
9		0/360	360	359	-1	3.0
10		45	45	41	-4	3.0
11		90	90	87	-3	3.0
12	Counter Clockwise	135	135	135	0	3.0
13		180	180	182	2	3.0
14		225	225	227	2	3.0
15		270	270	273	3	3.0
16		315	315	319	4	3.0

UUC*: Unit Under Calibration The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

End of Certificate of Calibration




Certificate No.: C01213617

Page: 2 of 2

Calibration Results:

Without Adjustment

Essential Error: Weight to be 1/4 or 1/3 of Maximum capacity, taken from the center of the pan as a zero reference.



Nominal Test Value		Reference Points (g)			5	(g)
A	B	C	D	E		
-	0.00000	0.00000	0.00000	0.00000	0.00000	0.00000

Repeatability: Determination of the standard deviation of weighing balance., Readability 0.00001 (g)

Nominal test value (g)	Standard Deviation
1	0.800007
10	0.000009

Departure of indication from nominal value., Readability 0.00001 (g)

Nominal Value (g)	Conventional Mass (g)	Displayed Value (g)	Correction of Balance (g)	Uncertainty (g)	k
0.001	0.001001	0.00100	0.00000	0.000014	2.10
0.005	0.005002	0.00500	0.00000	0.000014	2.10
0.01	0.010002	0.01000	0.00000	0.000014	2.08
0.05	0.049999	0.05000	0.00000	0.000015	2.07
0.1	0.100003	0.10000	0.00000	0.000015	2.06
0.5	0.499994	0.49998	0.00001	0.000018	2.03
1	0.999988	0.99999	0.00000	0.000020	2.02
2	1.999986	1.99999	0.00000	0.000023	2.01
5	4.999983	4.99998	0.00000	0.000028	2.00
10	10.000005	9.99998	0.00002	0.000035	2.00

The End of Certificate

High Volume Air Sampler Calibration Worksheet

Page 1 of 1

Project Site : งานอุตสาหกรรมโรงงานอุตสาหกรรม

Location : อ. บางพลี

Date of measurement : 8/4/2022

Worksheet No. : C-270220-WWL0093

High Volume ID : WWL0093

High Volume Model : TE-5170 (TSP)

High Volume S/N : 2729

Ambient Condition

Temperature (°C) : 27

Barometric Pressure (mmHg) : 757

Calibration Office

Calibrator ID : WWL0103

Calibrator Model : TE-5028A


Calibrator S/N : 3271

Calibrate Date : 11/02/22

Quality Standard Slope : 1.60965

Quality Standard Intercept : -0.04335

Test No.	delta H ₂ O (inchi)	Q _{act} (m³/min)	I (Chart)	IC (Corrected)	Linear Regression
1	4.90	1.397	50.0	49.80	Slope : 33.59
2	4.40	1.325	47.0	46.81	Intercept : 2.6512
3	3.20	1.134	41.0	40.83	Correlation Coefficient : 0.9995
4	2.40	0.985	36.0	35.85	
5	2.00	0.902	33.0	32.87	

Calibrated by : 
(Mr. RATTAPOL BAIKAD)
ChemistApproved by : (Mr. KUNGSASIKORN KOSUM)
Technical Management



บริษัท ศูนย์วิเคราะห์น้ำ จำกัด

WATER ANALYSIS CENTER COMPANY LIMITED

104 หมู่ 5 ต.สวนพริก อ.อุบล จ.พระนครศรีอยุธยา 13210

104 Moo 5, T. Kuanbun, A.U.-Thai, Ayutthaya 13210, Thailand

Tel: 0-35226-383, 0-35800-593 Fax: 0-35800-594

High Volume Air Sampler Calibration Worksheet

Project Site:

Location:

Date of measurement:

Worksheet No.:

High Volume ID:

High Volume S/N:

Ambient Condition

Temperature (°C):

Barometric Pressure (mmHg):

สถานที่ทดสอบโรงประปา

วัดสวนพริก

8/4/2022

C-270210-WVL0094

WVL0094

TE-5170 (TSP)

2736

11/02/22

1.60965

Quality Standard Slope:

Quality Standard Intercept:

Calibration Office

Calibrator ID:

Calibrator Model:

Calibrator S/N:

Calibrate Date:

Quality Standard Slope:

Quality Standard Intercept:

WWL0103

TE-5028A


3271

11/02/22

1.60965

-0.04335

Test No.	delta H ₂ O (inch)	Q _{act} (m ³ /min)	I (Chart)	IC (Corrected)	Linear Regression
1	5.80	1.517	54.0	53.77	Slope: 38.53
2	4.50	1.339	47.0	46.80	Intercept: -4.7881
3	3.70	1.217	42.0	41.82	Correlation Coefficient: 0.9998
4	2.75	1.053	36.0	35.85	
5	2.10	0.923	31.0	30.87	



Calibrated by:

Mr. Rungtassakorn Kosum

Chemist

Approved by:

Mr. Rungtassakorn Kosum

Technical Management

POLAB 64-1/25

แก้ข้อบกพร่อง วันที่แก้ไข: 1 มิ.ย. 2562 หน้า: 1 ของ 1



บริษัท ศูนย์วิเคราะห์น้ำ จำกัด

WATER ANALYSIS CENTER COMPANY LIMITED

104 หมู่ 5 ต.สวนพริก อ.อุบล จ.พระนครศรีอยุธยา 13210

104 Moo 5, T. Kuanbun, A.U.-Thai, Ayutthaya 13210, Thailand

Tel: 0-35226-383, 0-35800-593 Fax: 0-35800-594

High Volume Air Sampler Calibration Worksheet

Project Site:

Location:

Date of measurement:

Worksheet No.:

High Volume ID:

High Volume S/N:

Ambient Condition

Temperature (°C):

Barometric Pressure (mmHg):

สถานที่ทดสอบโรงประปา

วัดสวนพริก

8/4/2022

C-270210-WVL0095

WVL0095

TE-5170 (TSP)

2727

11/02/22

1.60965

Quality Standard Slope:

Quality Standard Intercept:

Calibration Office

Calibrator ID:

Calibrator Model:

Calibrator S/N:

Calibrate Date:

Quality Standard Slope:

Quality Standard Intercept:

WWL0103

TE-5028A


3271

11/02/22

1.60965

-0.04335

Test No.	delta H ₂ O (inch)	Q _{act} (m ³ /min)	I (Chart)	IC (Corrected)	Linear Regression
1	5.60	1.480	51.0	50.424	Slope: 36.37
2	4.70	1.359	46.0	45.480	Intercept: 1.5785
3	3.55	1.184	40.0	39.548	Correlation Coefficient: 0.9997
4	2.90	1.073	36.0	35.593	
5	2.10	0.917	30.0	29.661	



Calibrated by:

Mr. Rungtassakorn Kosum

Chemist

Approved by:

Mr. Rungtassakorn Kosum

Technical Management

POLAB 64-1/25

แก้ข้อบกพร่อง วันที่แก้ไข: 1 มิ.ย. 2562 หน้า: 1 ของ 1



บริษัท ศูนย์วิเคราะห์น้ำ จำกัด

WATER ANALYSIS CENTER COMPANY LIMITED

194 หมู่ 5 ต.สามพัน อ.อุทัย จ.พระนครศรีอยุธยา 13210
1/94 Moo 5, T.Kanham, A.U-Thai, Ayutthaya 13210, Thailand
Tel: 0-35226-383, 0-35800-593 Fax: 0-35800-594

High Volume Air Sampler Calibration Worksheet

Project Site :

Location :

Date of measurement :

Worksheet No. :

High Volume ID :

High Volume Model :

High Volume S/N :

Ambient Condition

Temperature (°C) :

Barometric Pressure (mmHg) :

สถานที่ทดสอบ : โรงงาน

วันที่ทดสอบ : 8/4/2022

หมายเลข : C-270220-WWL0096

High Volume ID : WWL0096

High Volume Model : TE-5170 (TSP)

High Volume S/N : 2730

Ambient Condition

Temperature (°C) : 27

Barometric Pressure (mmHg) : 757

Calibration Office

Calibrator ID :

Calibrator Model :

Calibrator S/N :

Calibrate Date :

Quality Standard Slope :

Quality Standard Intercept :

WWL0103

TE-5028A


3271

11/02/22

1.60965

-0.04335

Test No.	delta H ₂ O (inch)	Q _{std} (m ³ /min)	I (Chart)	IC (Corrected)	Linear Regression
1	5.90	1.530	54.0	53.77	Slope : 37.21
2	4.80	1.382	48.0	47.80	Intercept : -3.3516
3	3.85	1.241	43.0	42.82	Correlation Coefficient : 0.9998
4	2.90	1.080	37.0	36.84	
5	2.20	0.945	32.0	31.87	



Calibrated by :

Signature

(MR.BATTAPOL BAEZAI)

Chemist

Approved by :

Signature

(MR.RUNGSAKORN KOSUM)

Technical Management

FOLAB 64-125

แก้ไขครั้งสุดท้าย วันที่บันทึกใช้ : 1 เม.ย. 2562 หน้า : 1 ของ 1



บริษัท ศูนย์วิเคราะห์น้ำ จำกัด

WATER ANALYSIS CENTER COMPANY LIMITED

194 หมู่ 5 ต.สามพัน อ.อุทัย จ.พระนครศรีอยุธยา 13210
1/94 Moo 5, T.Kanham, A.U-Thai, Ayutthaya 13210, Thailand
Tel: 0-35226-383, 0-35800-593 Fax: 0-35800-594

High Volume Air Sampler Calibration Worksheet

Project Site :

Location :

Date of measurement :

Worksheet No. :

High Volume ID :

High Volume Model :

High Volume S/N :

Ambient Condition

Temperature (°C) :

Barometric Pressure (mmHg) :

สถานที่ทดสอบ : โรงงาน

วันที่ทดสอบ : 8/4/2022

หมายเลข : C-270220-WWL0097

High Volume ID : WWL0097

High Volume Model : TE-5170 (TSP)

High Volume S/N : 2726

Ambient Condition

Temperature (°C) : 27

Barometric Pressure (mmHg) : 757

Calibration Office

Calibrator ID :

Calibrator Model :

Calibrator S/N :

Calibrate Date :

Quality Standard Slope :

Quality Standard Intercept :

WWL0103

TE-5028A


3271

11/02/22

1.60965

-0.04335

Test No.	delta H ₂ O (inch)	Q _{std} (m ³ /min)	I (Chart)	IC (Corrected)	Linear Regression
1	5.30	1.451	52.0	51.78	Slope : 35.74
2	4.30	1.310	47.0	46.80	Intercept : -0.1557
3	3.30	1.151	41.0	40.83	Correlation Coefficient : 0.9997
4	2.70	1.043	37.0	36.84	
5	1.70	0.834	30.0	29.87	



Calibrated by :

Signature

(MR.BATTAPOL BAEZAI)

Chemist

Approved by :

Signature

(MR.RUNGSAKORN KOSUM)

Technical Management

FOLAB 64-125

แก้ไขครั้งสุดท้าย วันที่บันทึกใช้ : 1 เม.ย. 2562 หน้า : 1 ของ 1

TSP HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sampler Location	Date	Person
Project Site	27 Jun 2022	Mr.Chaimongkhon Siriphaeng
Transfer Standard Type	Office	Q _{sp} Slope (m)
Calibrator Model	TE-5025A	Q _{sp} Intercept (b)
Calibrator Serial Number	3092	


Calibration Information

Sampler Number		Motor Serial Number	Recorder Serial Number	576			
Test No.	Pressure Drop Across Orifice (ΔH ₂ O) (in H ₂ O)	ΔH ₂ O Potentiometer Reading	(A) $\frac{(\Delta H_2O)(P_{sp})^{1/2}}{a}$ $(\frac{1}{1000})(\frac{in H_2O}{lb/in^2})^{1/2}$	(X) $Q_{sp} = (\frac{1}{1000})(\frac{in H_2O}{lb/in^2})^{1/2}$	(I) Sample Flow Rate Indication	(Y) IC = $(\frac{P_{sp}}{P_{atm}})(\frac{T_{atm}}{T_{sp}})^{1/2}$	Temperature Barometric Pressure (°K = °C+273) (mmHg)
1	4.30	2.02646	0.99603	35.70	36.0	35.70	303.0
2	7.0	2.40936	1.18279	45.0	45.0	42.64	303.0
3	7.50	2.70741	1.30288	55.0	55.0	49.59	303.0
4	5.4	10.70	3.24998	1.50251	15.0	54.54	760.0
5	6.2	12.20	3.50624	1.65518	60.0	59.50	760.0
				Average		59.50	303.0

Linear Regression: y = mx + b

Slope (m)	
Intercept (b)	
R-Square (R ²)	0.995166
Correlation Coefficient (r)	0.997566

Andersen Instruments, Inc.



Calibrated By

Mr.Chaimongkhon Siriphaeng

TSP HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sampler Location	Date	Person
Project Site	27 Jun 2022	Mr.Chaimongkhon Siriphaeng
Transfer Standard Type	Office	Q _{sp} Slope (m)
Calibrator Model	TE-5025A	Q _{sp} Intercept (b)
Calibrator Serial Number	3092	


Calibration Information

Sampler Number	1, 2, 3, 4, 5	Motor Serial Number	1293, 4791	Recorder Serial Number	582
Test No.	Pressure Drop Across Orifice (ΔH ₂ O) (in-H ₂ O)	Pressure Drop Across Orifice (ΔH ₂ O) (in-H ₂ O)	Pressure Drop Across Orifice (ΔH ₂ O) (in-H ₂ O)	Pressure Drop Across Orifice (ΔH ₂ O) (in-H ₂ O)	Pressure Drop Across Orifice (ΔH ₂ O) (in-H ₂ O)
1	4.30	2.02646	0.99603	35.70	42.64
2	7.0	2.40936	1.18279	45.0	49.59
3	7.50	2.70741	1.30288	55.0	54.54
4	5.4	10.70	3.24998	1.50251	59.50
5	6.2	12.20	3.50624	1.65518	Average

Linear Regression: y = mx + b

Slope (m)	
Intercept (b)	
R-Square (R ²)	0.995664
Correlation Coefficient (r)	0.997664

Andersen Instruments, Inc.



Calibrated By

Mr.Chaimongkhon Siriphaeng

TSP HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sample Location หรือ Site	วันที่เก็บตัวอย่าง หรือวันที่	Date person	Mz. Name (คน) / ตัวอย่าง
Transfer Standard Type	Office	Q ₂ Steps (n)	2.1.137/2
Calibrator Model	TE-5025A	Q ₂ Interval (b)	-0.03080
Calibrator Serial Number	209C		


Calibration Information			
Sampler Number	TSP No. 11-	Motor Serial Number	Recorder Serial Number
		11173	

Test No.	Pressure Drop Across Orifice ($\Delta h_{\text{H}_2\text{O}}$) (mH ₂ O)	Pressure Drop Across Fouling Negl. ^a ($\Delta h_{\text{H}_2\text{O}}$)	(A) $(\Delta h_{\text{H}_2\text{O}}/\rho)P_{\text{H}_2\text{O}}/P_{\text{H}_2\text{O}}^{\text{ref}})^{1/3}$	(X) $Q_{\text{H}_2} = (1/\text{m})(A/\rho)$ (g/min)	(1) Sample Flow Rate Indication (°/min)	(Y) IC = $[(P/P_{\text{H}_2\text{O}})(T/T_{\text{H}_2\text{O}})]^{1/2}$ (°/min)	Temperature [°K = °C+273] (°C)	Barometric Pressure (mmHg)
1	2.1	4.80	2.05466	0.99003	1.11	55.70	29.16	740.0
2	3.1	5.0	2.41920	1.18279	1.11	42.64	29.16	740.0
3	6.0	6.50	2.90132	3.30288	1.11	-10.59	29.16	740.0
4	10.20	10.20	3.18277	1.53142	1.54	54.64	29.16	740.0
5	1.1	1.1	3.46980	1.65829	1.10	95.40	303.0	740.0
					Average			

Linear Regression : $y = mx + b$

Slope (m)	2.10423
Intercept (b)	
R-Square (R^2)	
Correlation Coefficient (r)	0.985861

Andersen Instruments, Inc.



Submitted By

Mr. Chittaranjan Sainthana

Calibrated By

Mr. Chairman, thank you.

SMILE
Laboratory Co., Ltd.

Laboratory Co., Ltd.

... and the ...

TSP HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Transfer Location	Project Site	UTM X	UTM Y	Date
		1000000	1000000	22 April 2022

Calibration		
Transfer Standard Type	Calibrator	Q ₅₁ Slope (m)
Calibrator Model	TE-5025A	Q ₅₁ Intercept (b)
Calibrator Serial Number	31907	-0.00390


Calibration Information		
Sampler Number	Tracer Serial Number	Recorder Serial Number
		4112

Test No.	Pressure Drop Across Positive (mbar)	ΔH_Q (mJ/m ²)	$[\Delta H_Q/P_{\text{eff}}]_{\text{eff}}^{\text{eff}} \cdot T_{\text{eff}}^{\text{eff}}$	$Q_{\text{eff}} = (2\pi m l / (A \cdot b)) \cdot \sigma$	(X)	(I)	(Y)	Temperature (°C)	Barometric Pressure (mmHg)
1	1.1	3.80	1.95231	0.03704	0.03704	0.03704	33.70	40.3.0	742.0
2	4.3	6.40	7.50698	1.27076	1.27076	0.2764	42.64	40.3.0	742.0
3	5.7	8.60	2.06828	1.40094	1.40094	49.59	49.59	40.3.0	742.0
4	5.2	10.30	3.18277	1.53142	1.53142	54.64	54.64	40.3.0	742.0
5	6.1	12.60	3.50024	1.69103	1.69103	59.50	59.50	40.3.0	742.0
						Average		303.0	691.0

Linear Regression: $y = mx + b$

Slope (m)	
Intercept (b)	
R-Square (R^2)	
Correlation Coefficient (r)	0.995088

Andersen Instruments, Inc.





บริษัท ศูนย์วิเคราะห์น้ำ จำกัด
WATER ANALYSIS CENTER COMPANY LIMITED
1/94 หมู่ 5 ต.สามพัน อ.อุทัย จ.พระนครศรีอยุธยา 13210
1/94 Moo 5, T.Kaibum, A.U-Thai, Ayutthaya 13210, Thailand
Tel: 0-35226-383, 0-35800-593 Fax: 0-35800-594

High Volume Air Sampler Calibration Worksheet

Project Site: สถานอุตสาหกรรมโฆะระอุยกะ Page 1 of 1

Location: วัดสามพัน

Date of measurement: 8/4/2022

Worksheet No.: C-280220-WWL0099 Calibration Orifice

High Volume ID: WWL0099 Calibrator ID: WWL0103

High Volume Model: TE-6070 (PM-10) Calibrator Model: TE-5028A


High Volume S/N: 2732 Calibrator S/N: 3271

Ambient Condition: 11/02/22

Temperature (°C): 27 Quality Standard Slope: 1.60965

Barometric Pressure (mmHg): 757 Quality Standard Intercept: -0.04335

Test No.	delta H ₂ O (inch)	Q _{ad} (m ³ /min)	I (Chart)	IC (Corrected)	Linear Regression
1	4.00	0.808	44.0	27.67	Slope: 41.83
2	3.40	0.747	40.0	25.15	Intercept: -6.1579
3	3.00	0.704	37.0	23.27	Correlation Coefficient: 0.9997
4	2.40	0.632	32.0	20.12	SSP: 65.3825
5	1.90	0.565	28.0	17.61	



Calibrated by: ฟ.ทวีป

(Mr.BATTAPOL BAIKAD)

Chemist

Approved by: ฟ.ทวีป

(Mr.RUNGSAKORN KOSUM)

Technical Management

FO.LAB 6.4-1/25

บันทึกข้อมูล วันที่บันทึกใช้ : 1 ม.ค. 2562 หน้า : 1 ของ 1



บริษัท ศูนย์วิเคราะห์น้ำ จำกัด
WATER ANALYSIS CENTER COMPANY LIMITED
1/94 หมู่ 5 ต.สามพัน อ.อุทัย จ.พระนครศรีอยุธยา 13210
1/94 Moo 5, T.Kaibum, A.U-Thai, Ayutthaya 13210, Thailand
Tel: 0-35226-383, 0-35800-593 Fax: 0-35800-594

High Volume Air Sampler Calibration Worksheet

Project Site: สถานอุตสาหกรรมโฆะระอุยกะ Page 1 of 1

Location: บ้านสามพัน

Date of measurement: 8/4/2022

Worksheet No.: C-280220-WWL0100 Calibration Orifice

High Volume ID: WWL0100 Calibrator ID: WWL0103

High Volume Model: TE-6070 (PM-10) Calibrator Model: TE-5028A


High Volume S/N: 2735 Calibrator S/N: 3271

Ambient Condition: 11/02/22

Temperature (°C): 27 Quality Standard Slope: 1.60965

Barometric Pressure (mmHg): 757 Quality Standard Intercept: -0.04335

Test No.	delta H ₂ O (inch)	Q _{ad} (m ³ /min)	I (Chart)	IC (Corrected)	Linear Regression
1	4.20	0.828	46.0	28.93	Slope: 36.37
2	3.50	0.758	42.0	26.41	Intercept: -1.1884
3	2.90	0.692	38.0	23.90	Correlation Coefficient: 0.9999
4	2.30	0.619	34.0	21.38	SSP: 63.47
5	1.80	0.551	30.0	18.87	



Calibrated by: ฟ.ทวีป

(Mr.BATTAPOL BAIKAD)

Chemist

Approved by: ฟ.ทวีป

(Mr.RUNGSAKORN KOSUM)

Technical Management

FO.LAB 6.4-1/25

บันทึกข้อมูล วันที่บันทึกใช้ : 1 ม.ค. 2562 หน้า : 1 ของ 1



High Volume Air Sampler Calibration Worksheet

Page 1 of 1

Project Site :

Location :

Date of measurement :

Worksheet No. :

High Volume ID :

High Volume Model :

High Volume S/N :

Ambient Condition :

Temperature (°C) :

Barometric Pressure (mmHg) :

สถานที่ตรวจวัด :
สำนักงานเทศบาลเมืองอุบลราชธานีวันที่ตรวจวัด :
8/4/2022วันที่ตรวจวัด :
8/4/2022หมายเลขใบตรวจวัด :
C-280220-WWL0101รหัสเครื่องวัด :
WWL0101รุ่นเครื่องวัด :
TE-6070 (PM-10)หมายเลขเครื่องวัด :
2733วันที่ตรวจวัด :
11/02/22อุณหภูมิ :
27ความดันบรรยากาศ :
757

Calibration Office :

Calibrator ID :

Calibrator Model :

Calibrator S/N :

Calibrate Date :

Quality Standard Slope :

Quality Standard Intercept :

WWL0103

TE-5028A


3271

11/02/22

1.60965

-0.04335

Test No.	delta H ₂ O (inch)	Q _{del} (m ³ /min)	I (Chart)	IC (Corrected)	Linear Regression
1	4.50	0.856	44.0	27.67	Slope : 31.91
2	3.60	0.768	40.0	25.15	Intercept : 0.4804
3	2.90	0.692	36.0	22.64	Correlation Coefficient : 0.9995
4	2.30	0.619	32.0	20.12	SSP : 58.11
5	1.70	0.536	28.0	17.61	



Calibrated by :

Approved by :

(MR.ATTAPOL BAIKAL)

Chemist

Approved by :

(MR. RUNGSAKORN KOSUM)

Technical Management

POLAB 64-125

วันที่ใช้เครื่องมือ : 1 ม.ค. 2562 หน้า : 1 ของ 1

Project Site :

Location :

Date of measurement :

Worksheet No. :

High Volume ID :

High Volume Model :

High Volume S/N :

Ambient Condition :

Temperature (°C) :

Barometric Pressure (mmHg) :

สถานที่ตรวจวัด :
สำนักงานเทศบาลเมืองอุบลราชธานีวันที่ตรวจวัด :
8/4/2022วันที่ตรวจวัด :
8/4/2022หมายเลขใบตรวจวัด :
C-280220-WWL0102รหัสเครื่องวัด :
WWL0102รุ่นเครื่องวัด :
TE-6070 (PM-10)หมายเลขเครื่องวัด :
2731วันที่ตรวจวัด :
11/02/22อุณหภูมิ :
27ความดันบรรยากาศ :
757

Calibration Office :

Calibrator ID :

Calibrator Model :

Calibrator S/N :

Calibrate Date :

Quality Standard Slope :

Quality Standard Intercept :

WWL0103

TE-5028A


3271

11/02/22

1.60965

-0.04335

Test No.	delta H ₂ O (inch)	Q _{del} (m ³ /min)	I (Chart)	IC (Corrected)	Linear Regression
1	4.10	0.828	44.0	28.03	Slope : 35.71
2	3.40	0.757	40.0	25.48	Intercept : -1.5704
3	2.80	0.689	36.0	22.94	Correlation Coefficient : 0.9999
4	2.20	0.614	32.0	20.39	SSP : 60.87
5	1.70	0.543	28.0	17.84	



Calibrated by :

Approved by :

(MR.ATTAPOL BAIKAL)

Chemist

Approved by :

(MR. RUNGSAKORN KOSUM)

Technical Management

POLAB 64-125

วันที่ใช้เครื่องมือ : 1 ม.ค. 2562 หน้า : 1 ของ 1

PM10 HIGH VOLUME AIR SAMPLER CALIBRATION REPORT


Site Information		Date	
Sampler Location Project Site	$\gamma_{transmission}$ (%/m ² /h)		
Calibration Office			
Transfer Standard Type	Office	Q_{d0} Slope (m)	2.1672
Calibrator Model	TE-5025A	Q_{d0} Intercept (b)	-0.03890
Calibrator Serial Number	3092		
		Mr.Chetmongkhon Siriphaeng	
		27 April 2022	

Calibration Information						
Sampler Number	141105	Motor Serial Number		Recorder Serial Number	19191	
Test No.	Pressure Drop Across Orifice (in-H ₂ O)	ΔH_2O	(in-H ₂ O)	ΔH_2O	(in-H ₂ O)	
	Positive leg-in	ΔH_2O	(in-H ₂ O)	ΔH_2O	(in-H ₂ O)	
1	2.6	1.5	3.10	1.74036	$[\Delta H_2O/P_{atm}T_{atm}T_{atm}]^{1/2}$	Barometric Pressure (mmHg)
2	5.0	2.9	5.09	2.40096	$[\Delta H_2O/P_{atm}T_{atm}T_{atm}]^{1/2}$	Temperature (°K = °C+273)
3	8.30	5.0	8.30	2.86771	$[\Delta H_2O/P_{atm}T_{atm}T_{atm}]^{1/2}$	
4	5.5	11.20	5.5	3.36601	$[\Delta H_2O/P_{atm}T_{atm}T_{atm}]^{1/2}$	
5	6.6	13.00	6.6	3.56594	$[\Delta H_2O/P_{atm}T_{atm}T_{atm}]^{1/2}$	
				1.71260	Average	

Linear Regression: $y = mx + b$

Slope (m)	
Intercept (b)	2.1×10^{-3} s
R-Square (R^2)	
Correlation Coefficient (r)	0.997006

Andersen Instruments, Inc.



Calibrated By

Mr. Chaitanankhon Sariphimong


Calibrated by

Mr. Chatmonkhan Sarichuan

SMILE
Laboratory Co., Ltd.

44
Mr. Chalmers van Sijpe

Andersen Instruments, Inc.



Calibrated by

Mr. Chatmonkhan Sarichuan

44
Mr. Chalmers van Sijpe

44
Mr. Chalmers van Strydom

บริษัท : บริษัท

วันที่ : 22 April 2022

Project Site : Mr.Chaimongkhon Saiphaeng


PM10 HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information	
Sampler Location	22 April 2022
Project Site	Mr.Chaimongkhon Saiphaeng
Calibration Office	
Transfer Standard Type	Office
Calibrator Model	TE-5025A
Calibrator Serial Number	3092

Sampler Number	Recorder Serial Number
11110	1116

Test No.	Pressure Drop Across Orifice (ΔH ₂ O) (mmH ₂ O)	ΔH ₂ O	(A)	(X)	(Y)	Temperature (°K = °C+273)	Brometric Pressure (mmHg)
1	1.8	3.70	1.90134	0.92220	31.63	305.0	760.0
2	3.0	6.10	2.44131	1.17896	41.32	305.0	760.0
3	4.0	8.20	2.85051	1.56397	51.00	305.0	760.0
4	5.1	11.20	3.30801	1.90995	57.33	305.0	760.0
5	6.0	12.10	3.63836	1.63291	61.28	305.0	760.0
				Average		305.0	760.0

Linear Regression : y = mx + b	
Slope (m)	
Intercept (b)	
R-Square (R ²)	0.995187
Correlation Coefficient (r)	0.997587



SMILE Laboratory Co., Ltd.

Calibrated By : Mr.Chaimongkhon Saiphaeng

บริษัท : บริษัท

วันที่ : 22 April 2022

Project Site : Mr.Chaimongkhon Saiphaeng


PM10 HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information	
Sampler Location	22 April 2022
Project Site	Mr.Chaimongkhon Saiphaeng
Calibration Office	
Transfer Standard Type	Office
Calibrator Model	TE-5025A
Calibrator Serial Number	3092

Sampler Number	Recorder Serial Number
11110	1116

Test No.	Pressure Drop Across Orifice (ΔH ₂ O) (mmH ₂ O)	ΔH ₂ O	(A)	(X)	(Y)	Temperature (°K = °C+273)	Brometric Pressure (mmHg)
1	1.8	3.70	1.90134	0.92220	31.63	305.0	760.0
2	3.0	6.10	2.44131	1.17896	41.32	305.0	760.0
3	4.0	8.20	2.85051	1.56397	51.00	305.0	760.0
4	5.1	11.20	3.30801	1.90995	57.33	305.0	760.0
5	6.0	12.10	3.63836	1.63291	61.28	305.0	760.0
				Average		305.0	760.0

Linear Regression : y = mx + b	
Slope (m)	
Intercept (b)	
R-Square (R ²)	0.995187
Correlation Coefficient (r)	0.997587



SMILE Laboratory Co., Ltd.

Calibrated By : Mr.Chaimongkhon Saiphaeng



ENVIRO SERVICE

42 Raminthra 14 yeak 9, Tha Raeng, Bangkok, Bangkok 10230

Tel : 02-9435814-5 Fax : 02-9438201 Tex id : 0105555170865

บริษัท เอ็นวิโรเซอร์วิส จำกัด
ENVIRO SERVICE CO., LTD.

PM10 HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information	
Sampler Location	22 April 2022
Project Site	Mr.Chaimongkhon Suriphaeng

Calibration Office	
Transfer Standard Type	Office
Calibrator Model	TE6025A
Calibrator Serial Number	3872

Calibration Information	
Sampler Number	14110
Recorder Serial Number	


Test No.	Pressure Drop Across Orifice (H ₂ O) (inH ₂ O)	(A)	(X)	(Y)	Temperature Pressure (mmHg)
1	1.8	1.7	3.50	1.84924	31.63
2	2.9	2.7	5.60	2.35912	41.52
3	4.1	4.1	8.30	2.89772	51.40
4	5.3	11.40	3.37492	1.60495	51.53
5	6.4	12.30	3.56020	1.70607	61.28
Average					305.0

Linear Regression: $y = mx + b$

Slope (m)	
Intercept (b)	
R-Square (R ²)	0.993811
Correlation Coefficient (r)	

Andersen Instruments, Inc.

$Q_{std} = 30.049161x + 0.2416567$
 $R^2 = 0.991639$



SMILE
Laboratory Co., Ltd.

Calibrated by

Mr.Chaimongkhon Suriphaeng

TEMPERATURE : 26.5 DEG.C

PRESSURE : 752 mmHg

TESTED BY :

Mr. Pasagon Samol



บริษัท เอ็นวิโรเซอร์วิส จำกัด
ENVIRO SERVICE CO., LTD.

REPORT QA. GAS-CALIBRATOR

CALIBRATE DATE: 18 Dec 21

GAS CALIBRATOR

MANUFACTURER : Environics MODEL : 6100 SN: 7462

FLOW CALIBRATOR : DryCal[®] DC-Lite MODEL : DCL-H SN: 107934

MODEL : DCLT 5K SN: 2105

MANUFACTURER : Bios International Corporation

REPORT QA. GAS-CALIBRATOR

AIR	SETTING	1000	2000	3000	4000	5000	6000	7000	8000	9000	10000	
FLOW	REF	997.00	2002.00	3005.00	3994.00	5000.00	5995.00	6990.00	7994.00	9004.00	9871.00	AVG
(CCM)	%ERROR	-0.300	0.100	0.167	-0.150	0.000	-0.083	-0.143	-0.075	0.044	-1.290	-0.17
GAS	SETTING	10	20	30	40	50	60	70	80	90	100	
FLOW	REF	10.18	20.19	30.27	40.16	50.34	60.33	70.54	80.69	90.28	100.7	AVG
(CCM)	%ERROR	1.800	0.950	0.900	0.400	0.680	0.550	0.771	0.862	0.311	0.700	0.79

Standard Reference

Reference Photometer Zero Air Brand : API Analyzer Model 701 S/N 349

Calibration Test Results

Expected Ozone (PPM)	REF Photometer Reading		% Error	Status
	before adjust	after adjust		
0.000	0.131	0.000	0.000	pass
0.100	0.164	0.102	2.000	pass
0.200	0.211	0.202	1.000	pass
0.300	0.242	0.298	-0.667	pass
0.400	0.377	0.397	-0.750	pass



บริษัท เอ็นไวร์ เซอร์วิส จำกัด
ENVIR SERVICE CO., LTD.

ENVIR SERVICE

42 Ramindra 14 yeak 9, Tha Reeng, Bangkok, Bangkok 10230

Tel : 02-9435814-5 Fax : 02-9438201 Tax id : 0105555170865

Standard Reference

Reference Photometer

Brand :

Zero Air

API Analyzer

Model 701 S/N 349

Calibration Test Results

Expected Ozone (PPM)	REF Photometer Reading		% Error	Status
	before adjust	after adjust		
0.000	0.020	0.000	0.000	pass
0.100	0.088	0.100	0.000	pass
0.200	0.176	0.199	-0.500	pass
0.300	0.286	0.298	-0.667	pass
0.400	0.388	0.396	-1.000	pass

TEMPERATURE : 26.3 DEGC

PRESSURE : 752 mmHg



TESTED BY :

Mr. Passagon Samol

บริษัท เอ็นไวร์ เซอร์วิส จำกัด
ENVIR SERVICE CO., LTD.

บริษัท ศูนย์วิเคราะห์น้ำ จำกัด

WATER ANALYSIS CENTER COMPANY LIMITED

194 หมู่ 5 ม.บางนาเหนือ อ.บางนา จ.กรุงเทพมหานคร 13210

194 Moo 5, T. Khamnang, A. U-Tad, Bangkok 13210, Thailand

Tel: 0-9226-583, 0-5800-593 Fax: 0-5800-594

Nitrogen Dioxide Analyzer Calibration Worksheet

Project Site :

Location :

Date of measurement :

Worksheet No. :

Ambient NOx Analyzer ID :

Manufacturer :

Ambient NOx Analyzer Model :

Ambient NOx Analyzer S/N :

Multi Gas Calibrator

Calibrator ID :

Calibrator Model :

Calibrator S/N :

Calibrate Date :

Cylinder Std. Gas

Std. Gas Concentration (PPM) :

Cylinder Pressure (psi)

Certified Date :


Expired Date :

Serial No. :

CALIBRATION RESULTS

Point	Ideal		Actual NO		%Error NO		Error NO _x		%Error NO _x	
	0.00	200.00	200.10	400.10	0.20	0.10	0.20	0.20	0.20	0.10
ZERO	0.00	0.20	0.20	-	0.20	0.20	0.20	0.20	0.20	-
SPAN 200 ppb	200.00	200.10	200.10	0.10	0.05	0.05	0.05	0.05	0.05	0.05
SPAN 400 ppb	400.00	400.10	400.10	0.10	0.03	0.03	0.03	0.03	0.03	0.03
AVERAGE (%)		0.04		0.04		0.04		0.04		0.07

Calibration Curve



Calibrated by

Sutut

(Mr. SUTIWAT JATHEERAPARKUL)

Chemist

Approved by

R

(Mr. RUNGSASIKORN KOSUM)

Technical Management



Nitrogen Dioxide Analyzer Calibration Worksheet

Project Site : สถานอุตสาหกรรมโรงงาน อุตสาหกรรม

Location : บ้านนา

Date of measurement : 02 April 2022

Worksheet No. : C-220422-WWL 0115

Ambient NOx Analyzer ID : WWL 0115

Manufacturer : HORIBA

Ambient NOx Analyzer Model : APNA-370

Ambient NOx Analyzer S/N : 705KA911

Multi Gas Calibrator

Calibrator ID : WWL0128

Calibrator Model : Series 6100

Calibrator S/N : S/N 7462

Calibrate Date : 20 June 2020

Cylinder Std. Gas

Std. Gas Concentration (PPM) : 50.90

Cylinder Pressure (psi) : 2000


Certified Date : 07 December 2017

Expired Date : 07 December 2021

Serial No. : CC241587

Point	CALIBRATION RESULTS				
	Ideal	Actual NO	Error NO	%Error NO	%Error NO _x
ZERO	0.00	0.20	0.20	-	-
SPAN 200 ppb	200.00	200.10	0.10	0.05	0.10
SPAN 400 ppb	400.00	400.10	0.10	0.03	0.05
AVERAGE (%)					0.04

Calibration Curve



Calibrated by

(Mr. SUTIWAT JATHARAPAKUL)

Chemist

Approved by

(Mr. RUNGSASIKORN KOSUM)

Technical Management



Nitrogen Dioxide Analyzer Calibration Worksheet

Project Site : สถานอุตสาหกรรมโรงงาน อุตสาหกรรม

Location : บ้านนา

Date of measurement : 22 April 2022

Worksheet No. : C-220422-WWL 0116

Ambient NOx Analyzer ID : WWL 0116

Manufacturer : HORIBA

Ambient NOx Analyzer Model : APNA-370

Ambient NOx Analyzer S/N : 9BRKGTUK

Multi Gas Calibrator

Calibrator ID : WWL0128

Calibrator Model : Series 6100

Calibrator S/N : S/N 7462

Calibrate Date : 28 February 2019

Cylinder Std. Gas

Std. Gas Concentration (PPM) : 50.90

Cylinder Pressure (psi) : 2000


Certified Date : 07 December 2017

Expired Date : 07 December 2021

Serial No. : CC241587

Point	CALIBRATION RESULTS				
	Ideal	Actual NO	Error NO	%Error NO	%Error NO _x
ZERO	0.00	0.20	0.20	-	-
SPAN 200 ppb	200.00	200.20	0.20	0.10	0.30
SPAN 400 ppb	400.00	400.30	0.30	0.08	0.20
AVERAGE (%)					0.09

Calibration Curve



Calibrated by

(Mr. SUTIWAT JATHARAPAKUL)

Chemist

Approved by

(Mr. RUNGSASIKORN KOSUM)

Technical Management




Nitrogen Dioxide Analyzer Calibration Worksheet

Project Site : ส่วนอุตสาหกรรมโรงโม่หิน
Location : บ้านใหม่ลำพูน
Date of measurement : 22 April 2022
Worksheet No. : C-220422-WWL 0117
Ambient NOx Analyzer ID : WWL 0117
Manufacturer : HORIBA
Ambient NOx Analyzer Model : AFNA-370
Ambient NOx Analyzer S/N : VKLYC3K0

Multi Gas Calibrator
Calibrator ID : WWL0128
Calibrator Model : Series 6100
Calibrator S/N : SN 7462
Calibrate Date : 28 February 2019
Cylinder Std. Gas
Std. Gas Concentration (PPM) : 50.90
Cylinder Pressure (psi) : 2000
Certified Date : 07 December 2017
Expired Date : 07 December 2021
Serial No. : CC241587

Point	CALIBRATION RESULTS				
	Ideal	Actual NO	Error NO	Actual NO _x	%Error NO _x
ZERO	0.00	0.20	0.20	0.20	-
SPAN 200 ppb	200.00	200.10	0.10	200.30	0.15
SPAN 400 ppb	400.00	400.20	0.20	400.20	0.05
AVERAGE (%)					0.10

Calibrated by Suttit

(Mr. SUTIWAT JAITHEERAPAPKUL)

Chemist

Approved by

(Mr. RUNGSASIKORN KOSUM)


Technical Management

Nitrogen Dioxide Analyzer Calibration Worksheet

Project Site : ส่วนอุตสาหกรรมโรงโม่หิน
Location : บ้านใหม่ลำพูน
Date of measurement : 22 April 2022
Worksheet No. : C-121121-WWL 0113
Ambient NOx Analyzer ID : WWL 0113
Manufacturer : HORIBA
Ambient NOx Analyzer Model : AFNA-370
Ambient NOx Analyzer S/N : WDMY8HT8

Multi Gas Calibrator
Calibrator ID : WWL0128
Calibrator Model : Series 6100
Calibrator S/N : SN 7462
Calibrate Date : 28 February 2019
Cylinder Std. Gas
Std. Gas Concentration (PPM) : 50.90
Cylinder Pressure (psi) : 2000
Certified Date : 07 December 2017
Expired Date : 07 December 2021
Serial No. : CC241587

Point	CALIBRATION RESULTS				
	Ideal	Actual NO	Error NO	Actual NO _x	%Error NO _x
ZERO	0.00	0.20	0.20	0.20	-
SPAN 200 ppb	200.00	200.20	0.20	200.30	0.15
SPAN 400 ppb	400.00	400.10	0.10	400.20	0.05
AVERAGE (%)					0.06

Calibrated by Suttit

(Mr. SUTIWAT JAITHEERAPAPKUL)

Chemist

Approved by

(Mr. RUNGSASIKORN KOSUM)

Technical Management



บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 ถนนสุขุมวิท 14 แขวง 9 เขตวัฒนา กรุงเทพมหานคร 10230 โทรศัพท์: 02-943814-5 โทรสาร: 02-9438201
เว็บไซต์: www.envia-service.com
ENVIA SERVICE CO., LTD. 42 Ramithira 14 road 9, The Rang, Bangkok, Bangkok 10230 Tel: 02-943814-5 Fax: 02-9438201

Analyzer Performance Test

Calibrated Date: 22 April 2022 Location: กรุงเทพมหานคร (AS)

Instruments Information

Analyzer Type: NDIR/NOx Analyzer
Model: 42C

Manufacturer: Thermo Environmental
S/N: 42C-8512-546

Calibration System

Calibrator Unit
Diluter Model: Thermo 5008
S/N: 705
ZERO AIR Generator API Model 701
S/N: 1924
Standard Gas
NO Conc 66.47 PPM
CO Conc 85.11 PPM
O₂ Conc 20.9% PPM
Cylinder number: E8012027
Expiry Date: 28 Oct 2027


Environment: Temperature: 25.5 °C Humidity: 51 %RH

Calibration Check (Before adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.1	0.0	0.1	393.4	400.0	-1.7
NOx	0.1	0.0	0.1	395.7	400.0	-1.1

Calibration Check (After adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.0	0.0	0.0	400.0	400.0	0.0
NOx	0.0	0.0	0.0	400.0	400.0	0.0



บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 ถนนสุขุมวิท 14 แขวง 9 เขตวัฒนา กรุงเทพมหานคร 10230 โทรศัพท์: 02-943814-5 โทรสาร: 02-9438201
เว็บไซต์: www.envia-service.com
ENVIA SERVICE CO., LTD. 42 Ramithira 14 road 9, The Rang, Bangkok, Bangkok 10230 Tel: 02-943814-5 Fax: 02-9438201

Analyzer Performance Test

Calibrated Date: 22 April 2022 Location: กรุงเทพมหานคร (AS)

Instruments Information

Analyzer Type: NDIR/NOx Analyzer
Model: 42C

Manufacturer: Thermo Environmental
S/N: 42C-7282-386

Calibration System

Calibrator Unit
Diluter Model: Thermo 5008
S/N: 705
ZERO AIR Generator API Model 701
S/N: 1924
Standard Gas
NO Conc 66.47 PPM
CO Conc 85.11 PPM
O₂ Conc 20.9% PPM
Cylinder number: E8012027
Expiry Date: 28 Oct 2027


Environment: Temperature: 25.5 °C Humidity: 51 %RH

Calibration Check (Before adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.1	0.0	0.1	393.3	400.0	-1.7
NOx	0.1	0.0	0.1	398.1	400.0	-1.0

Calibration Check (After adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.0	0.0	0.0	400.0	400.0	0.0
NOx	0.0	0.0	0.0	400.0	400.0	0.0





บริษัท เอ็นวีเอ เซอร์วิส จำกัด

42 ถนนสีลม 14 แขวง 9 เขตวชิรบุรี กรุงเทพมหานคร 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201
บริษัท เอ็นวีเอ เซอร์วิส จำกัด
ENVIA SERVICE CO., LTD. 42 Raminthra 14, 9th, The Rang, Bangkok, Bangkok 10230 Tel: 02-9435814-5 Fax: 02-9438201

Analyzer Performance Test

Calibrated Date: 22 April 2022 Location : หักเหล็ก (A9)

Instruments Information

Analyzer Type: NONINOX Analyzer
Model: 42C

Calibration System

Calibrator Unit
Oxizer Model Teats Model 6008
ZERO AIR Generator API Model 701
SN: 1624
Standard Gas
NO Conc 55.47 PPM
O2 Conc 55.11 PPM
CO Conc 4.535 PPM
Cylinder number EB0120027
Expire Date: 28 Oct. 2027


Environment: Temperature, 25.5 °C Humidity, 51 %RH

Calibration Check (Before adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.1	0.0	0.1	383.1	400.0	-1.7
NOx	0.1	0.0	0.1	386.8	400.0	-1.1

Calibration Check (After adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.0	0.0	0.0	400.0	400.0	0.0
NOx	0.0	0.0	0.0	400.0	400.0	0.0



Signature

Calibrate By : Mr. Pasagorn Samol



บริษัท เอ็นวีเอ เซอร์วิส จำกัด

42 ถนนสีลม 14 แขวง 9 เขตวชิรบุรี กรุงเทพมหานคร 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201
บริษัท เอ็นวีเอ เซอร์วิส จำกัด
ENVIA SERVICE CO., LTD. 42 Raminthra 14, 9th, The Rang, Bangkok, Bangkok 10230 Tel: 02-9435814-5 Fax: 02-9438201

Analyzer Performance Test

Calibrated Date: 22 April 2022 Location : หักเหล็ก (A9)

Instruments Information

Analyzer Type: NONINOX Analyzer
Model: 42C

Calibration System

Calibrator Unit
Oxizer Model Teats Model 6008
ZERO AIR Generator API Model 701
SN: 1624
Standard Gas
NO Conc 55.47 PPM
O2 Conc 55.11 PPM
CO Conc 4.535 PPM
Cylinder number EB0120027
Expire Date: 28 Oct. 2027


Environment: Temperature, 25.5 °C Humidity, 51 %RH

Calibration Check (Before adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.1	0.0	0.1	383.6	400.0	-1.8
NOx	0.1	0.0	0.1	389.4	400.0	-0.9

Calibration Check (After adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.0	0.0	0.0	400.0	400.0	0.0
NOx	0.0	0.0	0.0	400.0	400.0	0.0



Signature

Calibrate By : Mr. Pasagorn Samol

บริษัท ศูนย์วิเคราะห์น้ำ จำกัด



WATER ANALYSIS CENTER COMPANY LIMITED

194 หมู่ 5 ต.บางนาหนาด อ.เมือง จ.นครราชสีมา 32110
194 Moo 5, T.Bangna Nhand, A.U-Thai, Ayutthaya 12110, Thailand
Tel: 0-35226-583, 0-35800-593 Fax: 0-35800-594



บริษัท เอ็นวีเออร์ เซอร์วิส จำกัด

42 ถนนมิตรภาพ 114 หมู่ 9 ต.บางนาหนาด อ.เมือง จ.นครราชสีมา 32110 โทร: 02-9435814-5 โทรสาร: 02-94358201
บริษัท เอ็นวีเออร์ เซอร์วิส จำกัด
ENVIA SERVICE CO., LTD. 42 Raminthra 14, Moo 9, T.Bangna Nhand, A.U-Thai, Ayutthaya, Bangkok 12110 Tel: 02-9435814-5 Fax: 02-94358201

Analyzer Performance Test

Calibrated Date: 22 April 2022 Location: บ้านนาหนาด (A10)

Instruments Information
Analyzer Type: NONOXINOX Analyzer Model: APNA-380 Manufacturer: Horiba Environmental S/N: 857870112

Calibration System
Cylinder No.: 220422WWL 0113
Ambient SOx Analyzer ID: WWL 0113
Manufacturer: HORIBA
Ambient SOx Analyzer Model: AFSA-370
Ambient SOx Analyzer S/N: WDMY8HT8

Standard Gas
NO Conc 55.47 PPM
SO2 Conc 65.11 PPM
CO Conc 85.11 PPM
Cylinder No.: E80129027
Expire Date: 28 Oct 2027


Environment: Temperature 25.5 °C Humidity 51 %RH

Calibration Check (Before adjust)

Gas	Zero				Span			
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Drift (%)	Reading Value (ppb)	Expected Value (ppb)	Drift (%)	Drift (%)
NO	0.1	0.0	0.1	0.1	396.6	400.0	-0.9	-0.9
NOx	0.1	0.0	0.1	0.1	400.0	400.0	0.0	0.0

Calibration Check (After adjust)

Gas	Zero				Span			
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Drift (%)	Reading Value (ppb)	Expected Value (ppb)	Drift (%)	Drift (%)
NO	0.0	0.0	0.0	0.0	400.0	400.0	0.0	0.0
NOx	0.0	0.0	0.0	0.0	400.0	400.0	0.0	0.0



Calibrated By: Mr. Pongsorn Samrit

Sulfur Dioxide Analyzer Calibration Worksheet


Project Site: สวนอุตสาหกรรมโรจนะ อุตสาหกรรม
Location: จ.โคราช
Date of measurement: 22 April 2022
Worksheet No.: C-220422WWL 0113
Ambient SOx Analyzer ID: WWL 0113
Manufacturer: HORIBA
Ambient SOx Analyzer Model: AFSA-370
Ambient SOx Analyzer S/N: WDMY8HT8

Multi Gas Calibrator
Calibrator ID: WWL0128
Calibrator Model: Series 6100
Calibrator S/N: S/N 7462
Calibrate Date: 20 June 2020
Cylinder Std. Gas
Std. Gas Concentration (PPM): 49.58
Cylinder Pressure (psi): 2000
Certified Date: 07 December 2017
Expired Date: 07 December 2021
Serial No.: CC241587

CALIBRATION RESULTS

Point	Ideal	Actual SOx	Error Sox	%Error Sox
ZERO	0.00	0.00	0.00	-
SPAN 200 ppb	200.00	200.10	0.10	0.05
SPAN 400 ppb	400.00	400.10	0.10	0.03
AVERAGE (%)				
				0.04

Calibration Curve



Calibrated by: Sutut
(Mr. SUTWAT JAITHEERAPAPKUL)
Chemist

Approved by: [Signature]
(Mr. RUNGSASIKORN KOSUM)
Technical Management



Sulfur Dioxide Analyzer Calibration Worksheet

Project Site : สวนอุตสาหกรรมโรจนะ อุทัย

Location : วัดคันหัน

Date of measurement : 22 April 2022

Worksheet No. : C-220422-WWL 0110

Ambient SO_x Analyzer ID : WWL 0110

Manufacturer : HORIBA

Ambient SO_x Analyzer Model : AFS-A-370

Ambient SO_x Analyzer S/N : Y8SW7T00

Mult Gas Calibrator

Calibrator ID : WWL0128

Calibrator Model : Series 6100

Calibrator S/N : SN 7462

Calibrate Date : 23 February 2019

Cylinder Std. Gas

Std. Gas Concentration (PPM) : 49.68


Cylinder Pressure (psi) : 2000

Certified Date : 07 December 2017

Expired Date : 07 December 2021

Serial No. : CC241587

Point	CALIBRATION RESULTS			
	Ideal	Actual SO _x	Error Sox	%Error Sox
ZERO	0.00	0.00	0.00	-
SPAN 200 ppb	200.00	200.20	0.20	0.10
SPAN 400 ppb	400.00	400.20	0.20	0.05
AVERAGE (%)				0.07



Calibrated by Sutit

(Mr. SUTIWAT JAITHEERAPAPKUL)

Client

Approved by

(Mr. RUNGSAKORN KOSUM)

Technical Management



Sulfur Dioxide Analyzer Calibration Worksheet

Project Site : สวนอุตสาหกรรมโรจนะ อุทัย

Location : วัดคันหัน

Date of measurement : 22 April 2022

Worksheet No. : C-220422-WWL 0111

Ambient SO_x Analyzer ID : WWL 0111

Manufacturer : HORIBA

Ambient SO_x Analyzer Model : AFS-A-370

Ambient SO_x Analyzer S/N : PGRKTBDX

Mult Gas Calibrator

Calibrator ID : WWL0128

Calibrator Model : Series 6100

Calibrator S/N : SN 7462

Calibrate Date : 23 February 2019

Cylinder Std. Gas

Std. Gas Concentration (PPM) : 49.68


Cylinder Pressure (psi) : 2000

Certified Date : 07 December 2017

Expired Date : 07 December 2021

Serial No. : CC241587

Point	CALIBRATION RESULTS			
	Ideal	Actual SO _x	Error Sox	%Error Sox
ZERO	0.00	0.00	0.00	-
SPAN 200 ppb	200.00	200.20	0.20	0.10
SPAN 400 ppb	400.00	400.20	0.20	0.05
AVERAGE (%)				0.07



Calibrated by Sutit

(Mr. SUTIWAT JAITHEERAPAPKUL)

Client

Approved by

(Mr. RUNGSAKORN KOSUM)

Technical Management




Sulfur Dioxide Analyzer Calibration Worksheet

Project Site : สวนอุตสาหกรรมโรจนะ อุตสาหกรรม
Location : ต.บ้านนา อ.เมือง จ.พระนครศรีอยุธยา 13210
Date of measurement : 22 April 2022
Worksheet No. : C-220422-WWL-0112
Ambient SO_x Analyzer ID : WWL 0112
Manufacturer : HORIBA
Ambient SO_x Analyzer Model : APSA-370
Ambient SO_x Analyzer S/N : 8R181BBF

Multi Gas Calibrator
Calibrator ID : WWL0128
Calibrator Model : Series 6100
Calibrator S/N : SN 7462
Calibrate Date : 28 February 2019

Cylinder Std. Gas
Std. Gas Concentration (PPM) : 49.68
Cylinder Pressure (psi) : 2000
Certified Date : 07 December 2017
Expired Date : 07 December 2021
Serial No. : CC241587

Point	CALIBRATION RESULTS			
	Ideal	Actual SO _x	Error Sox	%Error Sox
ZERO	0.00	0.00	0.00	-
SPAN 200 ppb	200.00	200.10	0.10	0.05
SPAN 400 ppb	400.00	400.10	0.10	0.03
AVERAGE (%)				0.04



Calibrated by

S. Subit

(Mr. SUTWAT JATHEERAPAKKUL)

Chemist

Approved by

S. Subit

(Mr. RUNGSAKORN KOSUM)

Technical Management




Sulfur Dioxide Analyzer Calibration Worksheet

Project Site : สวนอุตสาหกรรมโรจนะ อุตสาหกรรม
Location : ต.บ้านนา อ.เมือง จ.พระนครศรีอยุธยา 13210
Date of measurement : 22 April 2022
Worksheet No. : C-220422-WWL-0118
Ambient SO_x Analyzer ID : WWL 0118
Manufacturer : HORIBA
Ambient SO_x Analyzer Model : APSA-370
Ambient SO_x Analyzer S/N : W2VNDX08

Multi Gas Calibrator
Calibrator ID : WWL0128
Calibrator Model : Series 6100
Calibrator S/N : SN 7462
Calibrate Date : 20 June 2020

Cylinder Std. Gas
Std. Gas Concentration (PPM) : 49.68
Cylinder Pressure (psi) : 2000
Certified Date : 07 December 2017
Expired Date : 07 December 2021
Serial No. : CC241587

Point	CALIBRATION RESULTS			
	Ideal	Actual SO _x	Error Sox	%Error Sox
ZERO	0.00	0.10	0.10	-
SPAN 200 ppb	200.00	200.10	0.10	0.05
SPAN 400 ppb	400.00	400.10	0.10	0.03
AVERAGE (%)				0.04



Calibrated by

S. Subit

(Mr. SUTWAT JATHEERAPAKKUL)

Chemist

Approved by

S. Subit

(Mr. RUNGSAKORN KOSUM)

Technical Management



บริษัท เซ็นไวร์ เซอร์วิส จำกัด

42 ซอยอินทรา 14 ถนน 9 แขวงจันทบุรี แขวงจันทบุรี กรุงเทพมหานคร 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201
ENVIRO SERVICE CO., LTD. 42 Ramindra 14, Joke 9, The Rang, Bangkok, Bangkok 10230 Tel : 02-9435814-5 Fax : 02-9438201

Analyzer Performance Test

Calibrated Date: 22 April 2022

Location : บ้านสวนใหญ่ (A6)

Instruments Information

Analyzer Type: SO2 Analyzer	Manufacturer API
Model: 100A	S/N: 319


Calibration System

Calibrator Unit	Standard Gas
Dilutor Model Dashi Model 5008	NO Conc 55.47 PPM
S/N: 705	SO2 Conc 55.11 PPM
ZERO AIR Generator API MODEL 701	CO Conc 4.835 PPM
S/N: 1924	Cylinder number EB0129027
Expiry Date: 28 Oct. 2027	

Environment: Temperature 28.6 °C Humidity 61 %RH

Calibration Report

Status	Reference	Reading	Drift	Reference	Reading	Drift%
Before	0.0	0.1	0.1	400.0	393.3	-1.7
After	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :

Mr. PASAGORN SAMOL



บริษัท เซ็นไวร์ เซอร์วิส จำกัด

42 ซอยอินทรา 14 ถนน 9 แขวงจันทบุรี แขวงจันทบุรี กรุงเทพมหานคร 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201
ENVIRO SERVICE CO., LTD. 42 Ramindra 14, Joke 9, The Rang, Bangkok, Bangkok 10230 Tel : 02-9435814-5 Fax : 02-9438201

Analyzer Performance Test

Calibrated Date: 22 April 2022

Location : บ้านสวนใหญ่ (A7)

Instruments Information

Analyzer Type: SO2 Analyzer	Manufacturer API
Model: 100A	S/N: 378


Calibration System

Calibrator Unit	Standard Gas
Dilutor Model Dashi Model 5008	NO Conc 55.47 PPM
S/N: 705	SO2 Conc 55.11 PPM
ZERO AIR Generator API MODEL 701	CO Conc 4.835 PPM
S/N: 1924	Cylinder number EB0129027
Expiry Date: 28 Oct. 2027	

Environment: Temperature 25.5 °C Humidity 61 %RH

Calibration Report

Status	Reference	Reading	Drift	Reference	Reading	Drift%
Before	0.0	0.1	0.1	400.0	393.1	-1.7
After	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :

Mr. PASAGORN SAMOL



บริษัท เอ็นวี เซอร์วิส จำกัด

43 ถนนมิตรภาพ 14 หมู่ 9 เขตรองรับบริการ กรุงเทพมหานคร 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201
บริษัท เอ็นวี เซอร์วิส จำกัด
ENVIA SERVICE CO., LTD. 42 Ramintra 14, group 9, The Reng, Bangkok, Bangkok 10230 Tel : 02-9435814-5 Fax : 02-9438201

Analyzer Performance Test

Calibrated Date: 22 April 2022

Location : ริมถนน (AS)

Instruments Information

Analyzer Type: SQ2 Analyzer	Manufacturer API
Model: 100A	S/N: 105


Calibration System

Calibrator Unit	Standard Gas
Dilutor Model Dashi Model 5008 S/N: 705 ZERO AIR Generator API MODEL 701 S/N: 1824	NO Conc 65.47 PPM SO2 Conc 65.11 PPM CO Conc 4.535 PPM Cylinder number EBO128027 Expire Date: 28 Oct. 2027

Environment: Temperature_26.6_°C Humidity_61_%RH

Calibration Report

Status	Reference (ppm)	Reading (ppm)	Drift (ppm)	Reference (ppm)	Reading (ppm)	Drift%
Before	0.0	0.1	0.1	400.0	383.5	-4.6
After	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :

Mr. PASAGORN SAMOL

Mr. PASAGORN SAMOL



บริษัท เอ็นวี เซอร์วิส จำกัด

43 ถนนมิตรภาพ 14 หมู่ 9 เขตรองรับบริการ กรุงเทพมหานคร 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201
บริษัท เอ็นวี เซอร์วิส จำกัด
ENVIA SERVICE CO., LTD. 42 Ramintra 14, group 9, The Reng, Bangkok, Bangkok 10230 Tel : 02-9435814-5 Fax : 02-9438201

Analyzer Performance Test

Calibrated Date: 22 April 2022

Location : ริมถนน (AS)

Instruments Information

Analyzer Type: SQ2 Analyzer	Manufacturer API
Model: 100A	S/N: 1192


Calibration System

Calibrator Unit	Standard Gas
Dilutor Model Dashi Model 5008 S/N: 705 ZERO AIR Generator API MODEL 701 S/N: 1824	NO Conc 65.47 PPM SO2 Conc 65.11 PPM CO Conc 4.535 PPM Cylinder number EBO128027 Expire Date: 28 Oct. 2027

Environment: Temperature_25.5_°C Humidity_61_%RH

Calibration Report

Status	Reference (ppm)	Reading (ppm)	Drift (ppm)	Reference (ppm)	Reading (ppm)	Drift%
Before	0.0	0.1	0.1	400.0	383.7	-4.6
After	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :

Mr. PASAGORN SAMOL

Mr. PASAGORN SAMOL

บริษัท เอ็นวีเอ เซอร์วิส จำกัด

Analyzer Performance Test

Calibrated Date: 22 April 2022 Location: บริษัท เอ็นวีเอ (A10)

Instruments Information

Analyzer Type: SC2 Analyzer Model: 100A	Manufacturer API Environmental S/N: 1818
--	---


Calibration System

Calibrator Unit	Standard Gas
Dilutor Model Dashi Model 5008 S/N: 705	NO Conc 55.47 PPM
ZERO AIR Generator API MODEL 701 S/N: 1624	SO2 Conc 55.11 PPM
	CO Conc 4.635 PPM
	Cylinder number EBO126227
	Expiry Date: 28 Oct. 2027

Environment: Temperature 25.5 °C Humidity: 61 %RH

Calibration Report

Status	Zero		Span		Drift%
	Reference (PPM)	Reading (PPM)	Drift (PPM)	Reading (PPM)	
Before	0.0	0.1	0.1	384.6	-1.3
After	0.0	0.0	0.0	400.0	0.0



Calibrated By: Mr. PASAGORN SAMOL

Request No. 21-64/0570

MTC No. EEL BP 86/0364

CALIBRATION CERTIFICATE

Submitted by : WATER ANALYSIS CENTER CO.,LTD

Address 1/94 MOO 5, T.KANHAM, A.U-THAI, AYUTTHAYA 13120.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.

Soi 10, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

Instrument Calibrated :

Description : Sound Calibrator	Ambient Environment
Manufacturer : BSWATECH	Temperature : (23 ± 3) °C
Model : CA111	Relative Humidity : (50 ± 15) %
Serial No. : 520272	Ambient Pressure : (101.325 ± 1.500) kPa

Standards used : 1. Digital Function Synthesizer NF Electronic DF-193A S/N 123037.

2. Measuring Amplifier Brüel&Kjær 2636 S/N 1537484.

3. Programmable Attenuator Tensgawa TPA-303A S/N OF 2214.

4. Digital Multimeter Agilent 34401A S/N MY44005560.

5. Pressure Transmitter Vaisala PTB202AD S/N T0650001.

6. Audio Analyzer Keithley 2015-P S/N 4106495.

7. Condenser Microphone Brüel&Kjær 4180 S/N 2889571.

Calibration Procedure: CP-102-04 based on IEC 60942-2003. The sound pressure level of instrument was measured by standard microphone using an insert voltage technique.

This instrument has been calibrated against standards maintained at Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

Date of Receipt 21 May 2021

Date of Calibration : 31 May 2021

The results relate only to the items tested/calibrated or value assigned. Adverting the Report/Certificate into public domain is prohibited unless written permission is obtained from the provider of TISTR.

Head Office 25 Mu 5 Tambon Ubontho Th. Amphoe Chong Luang, Chanyawat Sathumthani 12120, Thailand
Tel : 061 0 2577 9000 Fax : 061 0 2577 9009 E-mail : tistr@tistr.or.th
Office/Laboratory Soi 10, Bangpoo Industrial Estate, Sukhumvit Road, Amphoe Muang, Chanyawat Samutprakan 10280, Thailand
Tel : 061 0 2579 1121-50 ext. 115, 116 Fax : 061 0 2579 8992 E-mail : surtisee@tistr.or.th
Office 196 Phahonyothin Road, Chonburi, Bangkok 10000, Thailand
Tel : 061 0 2579 1121-50 ext. 525, 527 Fax : 061 0 2579 8992 E-mail : surtisee@tistr.or.th

FMJL/MTC002 Rev.4



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-64/0570 MTC No. EEL_BP 86/0564

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage

factor $k=2$, providing a level of confidence of approximately 95%.

Nominal Output of Unit Under Test = 94 dB re 20μPa at 1000 Hz

Acoustic Output in dB re 20μPa, Corrected to Reference Conditions : 101.325 kPa, 23.0 °C and 50 %RH

1. Sound Pressure Level

Standard Microphone Type	Sound Pressure Level			Tolerance limit
	Measured Sound Pressure Level (dB)	Deviated value (dB)	Uncertainty (dB)	
1/2 inch Brüel&Kjaer4180	93.80	-0.20	± 0.10	±0.40 dB

2. Frequency

Standard Microphone Type	Frequency			Tolerance limit
	Measured Frequency (Hz)	Deviated value (Hz)	Uncertainty (Hz)	
1/2 inch Brüel&Kjaer4180	1000.0	1.0	± 1.5	±1.0%

3. Total distortion

Standard Microphone Type	Total distortion			Tolerance limit
	Measured Total distortion (%)	Uncertainty (%)		
1/2 inch Brüel&Kjaer4180	1.99	± 0.30		±3.0%

Note : 1. No adjustment.

2. The calibrator pressure correction was not included.

3. The microphone volume correction was not included.

Date of Calibration : 31 May 2021

The results relate only to the items tested calibrated or value assigned.

Assembling the Report: Certificates and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office : 25 Mu 3 Tambon Phloeng Phai, Amphoe Phloeng Phai, Changwat Pathumthani 12120, Thailand
Tel: 060 0 2577 9000
Fax: 060 0 2577 9009
E-mail: tistr@tistr.or.th
Office/Laboratory : 50/10, Bangkoo Industrial Estate, Sukhumvit Road, Amphoe Bangkoo, Changwat Samutprakan 10260, Thailand
Tel: 060 0 2579 1121-30 ext. 5219, 5215, 5217
Fax: 060 0 2579 8592
E-mail: surasak@tistr.or.th

PM.BJ.MTC.002 Rev.4



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-64/0570 MTC No. EEL_BP 86/0564

Nominal Output of Unit Under Test = 114 dB re 20μPa at 1000 Hz

Acoustic Output in dB re 20μPa, Corrected to Reference Conditions : 101.325 kPa, 23.0 °C and 50 %RH

1. Sound Pressure Level

Standard Microphone Type	Sound Pressure Level			Tolerance limit
	Measured Sound Pressure Level (dB)	Deviated value (dB)	Uncertainty (dB)	
1/2 inch Brüel&Kjaer4180	113.88	-0.12	± 0.10	±0.40 dB

2. Frequency

Standard Microphone Type	Frequency			Tolerance limit
	Measured Frequency (Hz)	Deviated value (Hz)	Uncertainty (Hz)	
1/2 inch Brüel&Kjaer4180	1000.9	0.9	± 1.5	±1.0%

3. Total Distortion

Standard Microphone Type	Total Distortion			Tolerance limit
	Measured Total Distortion (%)	Uncertainty (%)		
1/2 inch Brüel&Kjaer4180	0.75	± 0.50		±3.0%

Note : 1. No adjustment.

2. The calibrator pressure correction was not included.

3. The microphone volume correction was not included.

Calibrated by :

Approved by :

Mr. Werasakchai Deschinyai



Acting Director

Electrical and Electronic Standards Laboratory

Industrial Metrology and Testing Service Centre

Date of Calibration : 31 May 2021

Date of Issue : 4 Jun 2021

Ref : 201204052102145001

End of Certificate

3 3

The results relate only to the items tested calibrated or value assigned.

Assembling the Report: Certificates and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office : 25 Mu 3 Tambon Phloeng Phai, Amphoe Phloeng Phai, Changwat Pathumthani 12120, Thailand
Tel: 060 0 2577 9000
Fax: 060 0 2577 9009
E-mail: tistr@tistr.or.th
Office/Laboratory : 50/10, Bangkoo Industrial Estate, Sukhumvit Road, Amphoe Bangkoo, Changwat Samutprakan 10260, Thailand
Tel: 060 0 2579 1121-30 ext. 5219, 5215, 5217
Fax: 060 0 2579 8592
E-mail: surasak@tistr.or.th

PM.BJ.MTC.002 Rev.4

W	FO.LAB 6.4-1/28	มาถึงครั้งที่ : 0	วันที่บังคับใช้ : 1 ม.ค. 2562	หน้า : 1 ของ 1
---	-----------------	-------------------	-------------------------------	----------------

แบบบันทึกการทวนสอบเครื่อง Sound Level Meter

เครื่อง CA111 Sound Calibrator S/N 520272 รหัสเครื่องมือ SR004 เกณฑ์การยอมรับ $93.80 \pm 0.3, 113.88 \pm 0.3$
วันที่สอบเทียบ 31/05/64
วันที่สอบเทียบครั้งสุดท้าย 30/05/65
เครื่อง Digital Thermohygro Meter S/N 105091609 รหัสเครื่องมือ WWL 0055
วันที่สอบเทียบ 02/12/64
วันที่สอบเทียบครั้งสุดท้าย 01/12/65
เครื่อง Sound Level Meter S/N 200051 รหัสเครื่องมือ WWL 0206
วันที่สอบเทียบ 17-28/09/63
วันที่สอบเทียบครั้งสุดท้าย 16/09/65

การทวนสอบก่อนออกให้งาน

อุณหภูมิ (°C) 24 เกณฑ์การยอมรับ 23.0±2.0
ความชื้นสัมพัทธ์ (%) 46 เกณฑ์การยอมรับ 50.0±15.0
วันที่ทวนสอบ 24/04/65

การทวนสอบหลังออกจากห้อง

อุณหภูมิ (°C) 24 เกณฑ์การยอมรับ 23.0±2.0
ความชื้นสัมพัทธ์ (%) 46 เกณฑ์การยอมรับ 50.0±15.0
วันที่ทวนสอบ 30/04/65

Item	ระดับเสียงที่วัดได้ (dB) (ค่าเฉลี่ยที่ 94.0dB)	ระดับเสียงที่วัดได้ (dB) (ค่าเฉลี่ยที่ 94.0dB)	ระดับเสียงที่วัดได้ (dB) (ค่าเฉลี่ยที่ 94.0dB)	ระดับเสียงที่วัดได้ (dB) (ค่าเฉลี่ยที่ 114.0dB)
1	93.8	113.9	93.8	113.8
2	93.8	113.9	93.8	113.8
3	93.8	113.9	93.8	113.8
4	93.8	113.9	93.8	113.8
5	93.8	113.9	93.8	113.8
6	93.8	113.9	93.8	113.8
7	93.8	113.9	93.8	113.8
8	93.8	113.9	93.8	113.8
9	93.8	113.9	93.8	113.8
10	93.8	113.9	93.8	113.8
X	93.80	113.90	93.80	113.80
SD	0.00	0.00	0.00	0.00
%RSD (≤ 10)	0.00	0.00	0.00	0.00
ผลการ ทวนสอบ	ผ่าน	ผ่าน	ผ่าน	ผ่าน

ผู้บันทึก
ผู้ตรวจสอบ
.....

ผู้บันทึก
ผู้ตรวจสอบ
.....



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-63/0837

MTC No. EEL. BP. 44/0963

CALIBRATION CERTIFICATE

Submitted by : Water Analysis Center Co., Ltd.
Address : 1/94 Moo 5, T. Kaoham, A. U-Thai, Ayutthaya 13210.
Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Center.
Soi 1, Bangpoo Industrial Estate, Sukhumvit Rd., A. Muang, Samprakan 10280.
Instrument Calibrated : Ambient Environment
Description : Integrating Sound Level Meter
Manufacturer : ACO
Model : 6226
Serial No. : 200051 Code: WWL0206
Microphone : Type 7052 No. 75988
Pre-amplifier : -
Standards used :
1. Band Pass Filter Stanford Research Systems SR 650 S/N 28712;
2. Condenser Microphone Brüel&Kjær 4180 S/N 2889871;
3. Decade Attenuator Ando AL-205 S/N 00464602;
4. Function/Arbitrary Waveform Generator Agilent 33220A S/N MY44042668;
5. Digital Function Synthesizer NF Electronic Instruments DF-193A S/N 122037;
6. Digital Multimeter Fluke 8520A S/N 4985007;
7. Pistonphone Rion NC-72 S/N 00402446;
8. Measuring Amplifier Brüel&Kjær 2636 S/N 1537484;
Temperature : (23 ± 3) °C
Relative Humidity : (50 ± 15) %
Ambient Pressure : (101.325 ± 1.5) kPa

Date of Receipt : 15 Sep. 2020

Date of Calibration : 17-28 Sep. 2020

1 / 8

The results relate only to the items tested or calibrated

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

PM.BLMTC.002 Rev.3

Head Office : 35/44, 3 Tambon Khlong Ha, Amphoe Khlong Luang, Chongwat Pichitthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : tumpai@tistr.or.th Website: www.tistr.or.th
Office : 196 Phahonyothin Road, Chatuchak Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 3217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-63/0837

MTC No. EEL. BP. 44/0963

9. Power Amplifier Britel&Kjier 2706 S/N 1517650;
10. Speaker Tannoy Limited, Great Britain British Patent No. 215300;
11. Digital Multimeter Agilent 34401A S/N MY44005560; and
12. Programmable Attenuator Tamagawa TPA-303A S/N 2212.

Calibration Procedure :

This instrument was calibrated by using calibration procedures no CP-102-02 and CP-102-03, which were based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). These calibration procedures were related to the electrical and acoustic signal tests. The electrical signal test was carried out with the direct measurement method. The acoustic signal test was performed in an anechoic room with the comparison measurement method.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%.

Date of Calibration : 17-28 Sep. 2020

2 / 8

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL/MTC.002 Rev.3

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax (66) 0 2577 9009
E-mail : rumpal@tistr.or.th Website:www.tistr.or.th

Office/Laboratory
Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax (66) 0 2579 1121-30 ext. 5219, 5225, 5217
E-mail : sumalee@tistr.or.th



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-63/0837

MTC No. EEL. BP. 44/0963

1. Absolute Sensitivity

Reference Acoustic Signal (dB)	Unit Under Test			Tolerance Limit Class 2 (\pm dB)
	Measured Value Before adjust	Measured Value After adjust	Deviation (dB)	Uncertainty (\pm dB)
113.91	113.4	113.9	0.0	0.30
				1.4

Note: The external calibration adjustment was firstly performed. The internal calibration adjustment was then completed at the display of 114.9 dB.

2. Self-generated noise

2.1 Normal test

Measured value (dB)	Uncertainty (\pm dB)
24.8	0.10

2.2 The microphone of the sound level meter was replaced by electrical signal input device

Frequency	Measured Value (dB)	Uncertainty (\pm dB)
Weighting		
A-Weighting	15.6	0.10
C-Weighting	26.5	0.10
Flat	31.8	0.10

Date of Calibration : 17-28 Sep. 2020

3 / 8

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL/MTC.002 Rev.3

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax (66) 0 2577 9009
E-mail : rumpal@tistr.or.th Website:www.tistr.or.th

Office/Laboratory
Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax (66) 0 2579 1121-30 ext. 5219, 5225, 5217
E-mail : sumalee@tistr.or.th

3. Acoustical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)	Flat (dB)		
125	-0.2	0.0	-0.1	0.40	2.0
1 000	0.6	0.6	0.7	0.40	1.4
4 000	0.1	0.2	0.3	0.40	3.6

4. Electrical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)	Flat (dB)		
63	0.2	0.1	-0.1	0.20	2.5
125	0.1	0.1	0.1	0.20	2.0
250	0.2	0.1	0.1	0.20	1.9
500	0.1	0.1	0.0	0.20	1.9
1 000	0.0	0.0	0.0	0.20	1.4
2 000	-0.2	0.0	-0.1	0.20	2.6
4 000	-0.3	-0.2	-0.1	0.20	3.6
8 000	-0.3	-0.1	-0.1	0.20	5.6

Date of Calibration : 17-28 Sep. 2020

4 / 8

The results relate only to the items tested or calibrated.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mai 3 Tambon Klong Ha, Amphoe Klong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpal@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FAJBLMTC-002 Rev.3

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
A-weighting	94.0	0.0	0.20	0.4
C-weighting	94.0	0.0	0.20	0.4
Flat	94.1	0.1	0.20	0.4

5.2 Time weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
Fast	94.0	0.0	0.20	0.3
Slow	94.0	0.0	0.20	0.3
Leq	91.9	-2.1	0.20	0.3

6. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
122	122.0	0.0	0.30	1.4
121	121.0	0.0	0.30	1.4
120	120.0	0.0	0.30	1.4
119	119.0	0.0	0.30	1.4
114	113.9	-0.1	0.30	1.4
109	108.9	-0.1	0.30	1.4

Date of Calibration : 17-28 Sep. 2020

5 / 8

The results relate only to the items tested or calibrated.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mai 3 Tambon Klong Ha, Amphoe Klong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpal@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FAJBLMTC-002 Rev.3

6. Level Linearity on the reference level range (continue)

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
104	103.9	-0.1	0.30	1.4
99	99.0	0.0	0.30	1.4
94	94.0	0.0	0.30	1.4
89	89.0	0.0	0.30	1.4
84	84.0	0.0	0.30	1.4
79	79.0	0.0	0.30	1.4
74	74.1	0.1	0.30	1.4
69	69.0	0.0	0.30	1.4
64	63.9	-0.1	0.30	1.4
59	59.0	0.0	0.30	1.4
54	53.9	-0.1	0.30	1.4
49	49.0	0.0	0.30	1.4
44	44.0	0.0	0.30	1.4
39	38.9	-0.1	0.30	1.4
34	34.0	0.0	0.30	1.4
33	33.0	0.0	0.30	1.4
32	32.1	0.1	0.30	1.4
31	31.2	0.2	0.30	1.4
30	30.2	0.2	0.30	1.4

Date of Calibration : 17-28 Sep. 2020

6 / 8

The results relate only to the items tested or calibrated.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Pa, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpal@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

FM.BL.MTC.002 Rev.3

7. Level Linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
40-130	125	125.0	0.0	0.30	1.4
30-120	115	115.0	0.0	0.30	1.4
20-110	105	105.0	0.0	0.30	1.4
20-100	95	95.0	0.0	0.30	1.4
20-90	85	85.0	0.0	0.30	1.4
20-80	75	74.9	-0.1	0.30	1.4

8. Tone burst response

Time Weighting	Toneburst Duration, Tb (ms)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (dB)
Fast	200	115.9	-0.1	0.20	±1.3
	2	98.3	-0.7	0.20	+1.3; -2.8
	0.25	89.8	-0.2	0.20	+1.8; -5.3
Slow	200	109.3	-0.3	0.20	±1.3
	2	89.9	-0.1	0.20	+1.3; -5.3
	200	110.1	0.1	0.20	±1.3
SEL	2	90.1	0.1	0.20	+1.3; -2.8
	0.25	81.1	0.1	0.20	+1.8; -5.3

Date of Calibration : 17-28 Sep. 2020

7 / 8

The results relate only to the items tested or calibrated.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Pa, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpal@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FM.BL.MTC.002 Rev.3



NSC-TIS-TIS 17025
CALIBRATION 0037

73-TESTR

THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-63/0837

MTC No. HEL. BP. 44/0963

9. Peak C sound level

Number of cycles in test signal	Anticipated value (dB)	Measured value (dB)	Deviated value (dB)	Uncertainty (±dB)	Tolerance limits Class 2 (±dB)
Complete cycle	125.4	125.8	0.4	0.20	2.4
Positive half cycle	124.4	124.2	-0.2	0.20	1.4
Negative half cycle	124.4	124.2	-0.2	0.20	1.4

10. Overload Indication

Measured value (dB)	Deviated value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
Positive one-half cycle	131.1	0.30	1.8

Calibrated by:

Korak Leachon

(Mr. Komkiet Laotassiri)

Approved by:



Electrical and Electronic Laboratory
Industrial Metrology and Testing Service Center

Date of Calibration : 17-28 Sep. 2020

Date of Issue : 29 Sep. 2020

Ref : 2011263091503604001

End of Certificate

8 / 8

The results relate only to the items tested or calibrated.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : tump@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C, Bangoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2579 1121-30 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office
196 Phayathai Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

ผู้บันทึก
ผู้ตรวจสอบ

ผู้บันทึก
ผู้ตรวจสอบ

W	FO.LAB 6.4-1/28	แก้ไขครั้งที่ : 0	วันที่บังคับใช้ : 1 ม.ค. 2562	หน้า : 1 ของ 1
---	-----------------	-------------------	-------------------------------	----------------

แบบบันทึกการตรวจสอบเครื่อง Sound Level Meter

เครื่อง CA111 Sound Calibrator S/N 520272 รหัสเครื่องมือ SR004 เกณฑ์การยอมรับ 93.80 ± 0.3, 113.88 ± 0.3
วันที่สอบเทียบ 31/05/64 วันที่สอบเทียบเครื่องต่อไป 30/05/65
เครื่อง Digital Thermohygro Meter S/N 105091609 รหัสเครื่องมือ WWL 0055
วันที่สอบเทียบ 02/12/64 วันที่สอบเทียบเครื่องต่อไป 01/12/65
เครื่อง Sound Level Meter S/N 200052 รหัสเครื่องมือ WWL 0207
วันที่สอบเทียบ 13-16/12/64 วันที่สอบเทียบเครื่องต่อไป 12/12/66
การตรวจสอบห้องจากนอกห้องงาน
อุณหภูมิ (°C) 24 เกณฑ์การยอมรับ 23.0 ± 2.0
ความชื้นสัมพัทธ์ (%) 46 เกณฑ์การยอมรับ 50.0 ± 15.0
วันที่ทำสอบ 24/04/65

Item	ระดับเสียงที่วัดได้ (dB) (ความดังที่ 94.0dB)	ระดับเสียงที่วัดได้ (dB) (ความดังที่ 114.0dB)	ระดับเสียงที่วัดได้ (dB) (ความดังที่ 94.0dB)	ระดับเสียงที่วัดได้ (dB) (ความดังที่ 114.0dB)
1	93.8	113.9	93.8	113.9
2	93.8	113.9	93.8	113.9
3	93.8	113.9	93.8	113.9
4	93.8	113.9	93.8	113.9
5	93.8	113.9	93.8	113.9
6	93.8	113.9	93.8	113.9
7	93.8	113.9	93.8	113.9
8	93.8	113.9	93.8	113.9
9	93.8	113.9	93.8	113.9
10	93.8	113.9	93.8	113.9
X	93.80	113.90	93.80	113.90
SD	0.00	0.00	0.00	0.00
%RSD (≤ 10)	0.00	0.00	0.00	0.00
ผลการ ตรวจสอบ	ผ่าน	ผ่าน	ผ่าน	ผ่าน



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-63/0837

MTC No. EEL. BP. 45/0963

CALIBRATION CERTIFICATE

Submitted by : Water Analysis Center Co., Ltd.
Address : 1/94 Moo 5, T. Kanham, A. U-Thai, Ayuthaya 13210.
Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Center.
Soi 1, Bangpoo Industrial Estate, Sukhumvit Rd., A. Muang, Samutprakan 10280.

Instrument Calibrated :

Description : Integrating Sound Level Meter
Manufacturer : ACO
Model : 6226
Serial No. : 200052 Code: WWL0207
Microphone : Type 7052 No. 75989
Preamplifier : -
Standards used :

Ambient Environment
Temperature : $(23 \pm 3) ^\circ\text{C}$
Relative Humidity : $(50 \pm 15) \%$
Ambient Pressure : $(101.325 \pm 1.5) \text{ kPa}$

1. Band Pass Filter Stanford Research Systems SR 650 S/N 28712;
2. Condenser Microphone Brüel&Kjær 4180 S/N 2889871;
3. Decade Attenuator Ardo AL-205 S/N 00464602;
4. Function/Arbitrary Waveform Generator Agilent 33220A S/N MY44042668;
5. Digital Function Synthesizer NF Electronic Instruments DF-193A S/N 122037;
6. Digital Multimeter Fluke 8520A S/N 4985007;
7. Pistonphone Rion NC-72 S/N 00402446;
8. Measuring Amplifier Brüel&Kjær 2636 S/N 1537484;

Date of Receipt : 15 Sep. 2020

Date of Calibration : 17-28 Sep. 2020

1 / 8

The results relate only to the items tested or calibrated.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu. 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : numpai@tistr.or.th Website: www.tistr.or.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FM.BL.MTC.002 Rev.3



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-63/0837

MTC No. EEL. BP. 45/0963

9. Power Amplifier Brüel&Kjær 2706 S/N 1517650;
10. Speaker Tamoy Limited, Great Britain British Patent No. 215300;
11. Digital Multimeter Agilent 34401A S/N MY44005560; and
12. Programmable Attenuator Tamagawa TPA-303A S/N 2212.

Calibration Procedure :

This instrument was calibrated by using calibration procedures no CP-102-02 and CP-102-03, which were based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). These calibration procedures were related to the electrical and acoustic signal tests. The electrical signal test was carried out with the direct measurement method. The acoustic signal test was performed in an anechoic room with the comparison measurement method.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%.

Date of Calibration : 17-28 Sep. 2020

2 / 8

The results relate only to the items tested or calibrated.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu. 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : numpai@tistr.or.th Website: www.tistr.or.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FM.BL.MTC.002 Rev.3

1. Absolute Sensitivity

Reference Acoustic Signal (dB)	Unit Under Test			Tolerance Limit Class 2 (±dB)
	Measured Value	Deviation	Uncertainty	
113.90	Before adjust 113.4	After adjust 113.9	(±dB) 0.0	0.30
				1.4

Note: The external calibration adjustment was firstly performed. The internal calibration adjustment was then completed at the display of 114.8 dB.

2. Self-generated noise

2.1 Normal test

Measured value (dB)	Uncertainty (±dB)
27.5	0.10

2.2 The microphone of the sound level meter was replaced by electrical signal input device

Frequency	Measured Value (dB)	Uncertainty (±dB)
Weighting		
A-Weighting	15.9	0.10
C-Weighting	25.6	0.10
Flat	29.3	0.10

Date of Calibration : 17-28 Sep. 2020

3 / 8

The results relate only to the items tested or calibrated.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FM.BLMTC.002 Rev.3

3. Acoustical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)	Flat (dB)	
125	0.5	0.3	0.3	0.40
1 000	-0.2	-0.1	-0.1	0.40
4 000	-0.4	0.6	-0.2	0.40
				3.6

4. Electrical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)	Flat (dB)	
63	0.1	0.1	0.0	0.20
125	0.1	0.1	0.1	0.20
250	0.1	0.0	0.0	0.20
500	0.1	0.0	0.0	0.20
1 000	0.0	0.0	0.0	0.20
2 000	-0.2	0.0	0.0	0.20
4 000	-0.3	-0.2	-0.1	0.20
8 000	-0.2	-0.1	-0.2	0.20
				5.6

Date of Calibration : 17-28 Sep. 2020

4 / 8

The results relate only to the items tested or calibrated.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FM.BLMTC.002 Rev.3

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
A-weighting	94.0	0.0	0.20	0.4
C-weighting	94.0	0.0	0.20	0.4
Flat	94.1	0.1	0.20	0.4

5.2 Time weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
Fast	94.0	0.0	0.20	0.3
Slow	94.0	0.0	0.20	0.3
Leq	92.1	-1.9	0.20	0.3

6. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
122	121.9	-0.1	0.30	1.4
121	120.9	-0.1	0.30	1.4
120	120.0	0.0	0.30	1.4
119	119.0	0.0	0.30	1.4
114	113.9	-0.1	0.30	1.4
109	108.9	-0.1	0.30	1.4

6. Level linearity on the reference level range (continue)

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
104	103.9	-0.1	0.30	1.4
99	99.0	0.0	0.30	1.4
94	94.0	0.0	0.30	1.4
89	89.0	0.0	0.30	1.4
84	84.0	0.0	0.30	1.4
79	79.0	0.0	0.30	1.4
74	74.1	0.1	0.30	1.4
69	69.0	0.0	0.30	1.4
64	63.9	-0.1	0.30	1.4
59	59.0	0.0	0.30	1.4
54	53.9	-0.1	0.30	1.4
49	49.0	0.0	0.30	1.4
44	44.0	0.0	0.30	1.4
39	38.9	-0.1	0.30	1.4
34	34.0	0.0	0.30	1.4
33	33.1	0.1	0.30	1.4
32	32.1	0.1	0.30	1.4
31	31.2	0.2	0.30	1.4
30	30.3	0.3	0.30	1.4

7. Level linearity including the level range control

Range	Anticipated		Measured		Deviated		Uncertainty		Tolerance Limit
	Value (dB)		Value (dB)		Value (dB)		(\pm dB)		
40-130	125		125.0		0.0		0.30		1.4
30-120	115		115.0		0.0		0.30		1.4
20-110	105		105.0		0.0		0.30		1.4
20-100	95		95.0		0.0		0.30		1.4
20-90	85		85.0		0.0		0.30		1.4
20-80	75		74.9		-0.1		0.30		1.4

8. Tone burst response

Time Weighting	Toneburst Duration, T _b (ms)		Measured Value (dB)		Deviated Value (dB)		Uncertainty (dB)		Tolerance Limit Class 2 (dB)
			Value (dB)		Value (dB)		(\pm dB)		
Fast	200		115.7		-0.3		0.20		± 1.3
	2		98.7		-0.3		0.20		+1.3; -2.8
	0.25		89.0		-1.0		0.20		+1.8; -5.3
Slow	200		109.5		-0.1		0.20		± 1.3
	2		89.9		-0.1		0.20		+1.3; -5.3
	200		110.1		0.1		0.20		± 1.3
SEL	2		90.2		0.2		0.20		+1.3; -2.8
	0.25		81.1		0.1		0.20		+1.8; -5.3

Date of Calibration : 17-28 Sep. 2020

7 / 8

The results relate only to the items tested or calibrated.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.3

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpal@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sri 1C, Bangpoo Industrial Estate, Sukhumvit Road, Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

9. Peak C sound level

Number of cycles in test signal	Anticipated value (dB)		Measured value (dB)		Deviated value (dB)		Uncertainty (dB)		Tolerance Limits Class 2 (\pm dB)
	Value (dB)		Value (dB)		Value (dB)		(\pm dB)		
Complete cycle	125.4		125.7		0.3		0.20		2.4
Positive half cycle	124.4		124.2		-0.2		0.20		1.4
Negative half cycle	124.4		124.2		-0.2		0.20		1.4

10. Overload indication

Measured value (dB)		Deviated value (dB)		Uncertainty (dB)		Tolerance Limits Class 2 (\pm dB)
Positive one-half cycle	Negative one-half cycle	Value (dB)		(\pm dB)		
133.0	133.0	0.0		0.30		1.8

Calibrated by :

.....
(Mr. Komkrit Laohasiri)

Approved by :

.....
(Mr. Nopadol Laohasiri)

Electrical and Electronic Standards Laboratory
Industrial Metrology and Testing Service Center

Date of Calibration : 17-28 Sep. 2020

Date of Issue : 29 Sep. 2020

Ref : 2011263091503604002

End of Certificate

8 / 8

The results relate only to the items tested or calibrated.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.3

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpal@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sri 1C, Bangpoo Industrial Estate, Sukhumvit Road, Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

W	FO-LAB 6.4-1/28	แก้ไขครั้งที่ : 0	วันที่ส่งกลับให้ : 1 ม.ค. 2562	หน้า : 1 ของ 1
---	-----------------	-------------------	--------------------------------	----------------

แบบบันทึกการทวนสอบเครื่อง Sound Level Meter

เครื่อง CA111 Sound Calibrator S/N 520272 รหัสเครื่องมือ SR004 เกณฑ์การยอมรับ 93.80 ± 0.3, 113.88 ± 0.3
วันที่สอบเทียบ 31/05/64 วันที่สอบเทียบเสร็จต่อไป 30/05/65

เครื่อง Digital Thermohygro Meter S/N 105091609 รหัสเครื่องมือ WWL 0055
วันที่สอบเทียบ 02/12/64 วันที่สอบเทียบเสร็จต่อไป 01/12/65

เครื่อง Sound Level Meter S/N 200053 รหัสเครื่องมือ WWL 0207
วันที่สอบเทียบ 17-28/09/63 วันที่สอบเทียบเสร็จต่อไป 16/09/65

การทวนสอบก่อนออกงาน

อุณหภูมิ (°C) 24 เกณฑ์การยอมรับ 23.0 ± 2.0
ความชื้นสัมพัทธ์ (%) 46 เกณฑ์การยอมรับ 50.0 ± 15.0
วันที่ทวนสอบ 24/04/65

การทวนสอบหลังจากออกงาน

อุณหภูมิ (°C) 24 เกณฑ์การยอมรับ 23.0 ± 2.0
ความชื้นสัมพัทธ์ (%) 46 เกณฑ์การยอมรับ 50.0 ± 15.0
วันที่ทวนสอบ 30/04/65

Item	ระดับเสียงที่วัดได้ (dB) (ความดังที่ 94.0dB)	ระดับเสียงที่วัดได้ (dB) (ความดังที่ 114.0dB)
1	93.9	113.9
2	93.9	113.9
3	93.9	113.9
4	93.9	113.9
5	93.9	113.9
6	93.9	113.9
7	93.9	113.9
8	93.9	113.9
9	93.9	113.9
10	93.9	113.9
X	93.90	113.90
SD	0.00	0.00
%RSD (≤ 10)	0.00	0.00
ผลการ ทวนสอบ	ผ่าน	ผ่าน

ผู้บันทึก
ผู้ตรวจสอบ
.....

ผู้บันทึก
ผู้ตรวจสอบ
.....



77-TISTR

THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-43/0837

MTC No. EEL. BR. 46/0963

CALIBRATION CERTIFICATE

Submitted by : Water Analysis Center Co., Ltd.
Address : 1/94 Moo 5, T. Kanham, A. U-Thai, Ayutthaya 13210.
Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Center.
Sol 1, Bangpoo Industrial Estate, Sukhumvit Rd., A. Muang, Samutprakan 10280.
Instrument Calibrated : Ambient Environment
Description : Integrating Sound Level Meter Temperature : (23 ± 3) °C
Manufacturer : ACO Relative Humidity : (50 ± 15) %
Model : 6226 Ambient Pressure : (101.325 ± 1.5) kPa
Serial No. : 200053 Code: WWL0208
Microphone : Type 7052 No. 75990
Preamplifier : -

- Standards used :
1. Band Pass Filter Stamford Research Systems SR 650 S/N 28712;
 2. Condenser Microphone Brüel&Kjær 4180 S/N 2889871;
 3. Decade Attenuator Ardo AL-205 S/N 00464602;
 4. Function/Arbitrary Waveform Generator Agilent 33220A S/N MY44042668;
 5. Digital Function Synthesizer NF Electronic Instruments DF-193A S/N 122037;
 6. Digital Multimeter Fluke 8520A S/N 4985007;
 7. Pistonphone Rion NC-72 S/N 00402446;
 8. Measuring Amplifier Brüel&Kjær 2636 S/N 1537484;

Date of Receipt : 15 Sep. 2020

Date of Calibration : 17-28 Sep. 2020

1 / 8

The results relate only to the items tested or calibrated.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

PA.BL.MTC.002 Rev.3

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Chengwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpal@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Chongwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtg@tistr.or.th

Office
196 Phahonyothin Road, Chatuchak Bangkok 10900,
Thailand
Tel. (66) 0 2579 121-30 ext. 5219, 5225, 3217
Fax. (66) 0 2579 8592
E-mail : sunate@tistr.or.th



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-63/0837

MTC No. EEL. BP. 46/0963

9. Power Amplifier Brief&Kjer 2706 S/N 1517650;
10. Speaker Tamoy Limited, Great Britain British Patent No. 215300;
11. Digital Multimeter Agilent 34401A S/N MY44005560; and
12. Programmable Attenuator Tamagawa TPA-303A S/N 2212.

Calibration Procedure :

This instrument was calibrated by using calibration procedures no CP-102-02 and CP-102-03, which were based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). These calibration procedures were related to the electrical and acoustic signal tests. The electrical signal test was carried out with the direct measurement method. The acoustic signal test was performed in an anechoic room with the comparison measurement method.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%.

Date of Calibration : 17-28 Sep. 2020

2 / 8

The results relate only to the items tested or calibrated.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.3

Head Office
35 Mu. 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : numpai@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-63/0837

MTC No. EEL. BP. 46/0963

1. Absolute Sensitivity

Reference Acoustic Signal (dB)	Unit Under Test			Tolerance Limit Class 2 (\pm dB)
	Measured Value (dB)	Deviation (dB)		
		Before adjust	After adjust	
113.90	133.3	113.9	0.0	0.30
				1.4

Note: The external calibration adjustment was firstly performed. The internal calibration adjustment was then completed at the display of 114.3 dB.

2. Self-generated noise

2.1 Normal test

Measured value (dB)	Uncertainty (\pm dB)
21.1	0.10

2.2 The microphone of the sound level meter was replaced by electrical signal input device

Frequency Weighting	Measured Value (dB)	Uncertainty (\pm dB)
A-Weighting	22.2	0.10
C-Weighting	32.5	0.10
Flat	37.2	0.10

3 / 8

Date of Calibration : 17-28 Sep. 2020

The results relate only to the items tested or calibrated.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.3

Head Office
35 Mu. 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : numpai@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th



NSC-TISTR 17025
CALIBRATION 0037

THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-63/0837

MTC No. EEL. BP. 46/0963

Request No. 21-63/0837

MTC No. EEL. BP. 46/0963

3. Acoustical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)	Flat (dB)		
125	0.2	0.1	-0.1	0.40	2.0
1 000	0.0	-0.1	0.1	0.40	1.4
4 000	0.2	0.4	0.4	0.40	3.6

4. Electrical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)	Flat (dB)		
63	-0.1	-0.4	-0.4	0.20	2.5
125	0.1	0.0	0.0	0.20	2.0
250	0.2	0.1	0.0	0.20	1.9
500	0.2	0.0	0.0	0.20	1.9
1 000	0.0	0.0	0.0	0.20	1.4
2 000	-0.2	0.0	0.1	0.20	2.6
4 000	-0.3	-0.2	0.0	0.20	3.6
8 000	-0.3	-0.2	-0.2	0.20	5.6

Date of Calibration : 17-28 Sep. 2020

4 / 8

The results relate only to the items tested or calibrated.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtg@tistr.or.th

Office
196 Phahonyothin Road, Chatuchak Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sunalee@tistr.or.th

PM.BLMTC.002 Rev.3



NSC-TISTR 17025
CALIBRATION 0037

THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-63/0837

MTC No. EEL. BP. 46/0963

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
A-weighting	94.0	0.0	0.20	0.4
C-weighting	94.0	0.0	0.20	0.4
Flat	94.1	0.1	0.20	0.4

5.2 Time weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
Fast	94.0	0.0	0.20	0.3
Slow	94.0	0.0	0.20	0.3
Leq	92.1	-1.9	0.20	0.3

6. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
122	121.9	-0.1	0.30	1.4
121	120.9	-0.1	0.30	1.4
120	120.0	0.0	0.30	1.4
119	118.9	-0.1	0.30	1.4
114	113.9	-0.1	0.30	1.4
109	108.8	-0.2	0.30	1.4

Date of Calibration : 17-28 Sep. 2020

5 / 8

The results relate only to the items tested or calibrated.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtg@tistr.or.th

Office
196 Phahonyothin Road, Chatuchak Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sunalee@tistr.or.th

PM.BLMTC.002 Rev.3



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-63/0837

MTC No. EEL, BP. 46/0963

THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-63/0837

MTC No. EEL, BP. 46/0963



6. Level linearity on the reference level range (continue)

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
104	103.9	-0.1	0.30	1.4
99	99.0	0.0	0.30	1.4
94	94.0	0.0	0.30	1.4
89	88.9	-0.1	0.30	1.4
84	83.9	-0.1	0.30	1.4
79	79.1	0.1	0.30	1.4
74	74.1	0.1	0.30	1.4
69	69.1	0.1	0.30	1.4
64	63.9	-0.1	0.30	1.4
59	59.0	0.0	0.30	1.4
54	53.9	-0.1	0.30	1.4
49	49.0	0.0	0.30	1.4
44	44.0	0.0	0.30	1.4
39	39.0	0.0	0.30	1.4
34	34.3	0.3	0.30	1.4
33	33.4	0.4	0.30	1.4
32	32.5	0.5	0.30	1.4
31	31.7	0.7	0.30	1.4
30	30.9	0.9	0.30	1.4

7. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
40-130	125	125.0	0.0	0.30	1.4
30-120	115	115.0	0.0	0.30	1.4
20-110	105	105.0	0.0	0.30	1.4
20-100	95	95.0	0.0	0.30	1.4
20-90	85	85.0	0.0	0.30	1.4
20-80	75	74.8	-0.2	0.30	1.4

8. Tone burst response

Time Weighting	Toneburst Duration, Tb (ms)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (dB)
Fast	200	115.5	-0.5	0.20	±1.3
	2	98.2	-0.8	0.20	+1.3; -2.8
	0.25	89.6	-0.4	0.20	+1.8; -5.3
Slow	200	109.4	-0.2	0.20	±1.3
	2	89.9	-0.1	0.20	+1.3; -5.3
	200	110.1	0.1	0.20	±1.3
SEL	2	90.1	0.1	0.20	+1.3; -2.8
	0.25	81.2	0.2	0.20	+1.8; -5.3

Date of Calibration : 17-28 Sep. 2020

6 / 8

7 / 8

The results relate only to the items tested or calibrated.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Klong Ha, Amphoe Klong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : tump@tistr.or.th Website: www.tistr.or.th

Office
196 Phahonyothin Road, Chatuchak Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FMBL MTC.002 Rev.3

Head Office
35 Mu 3 Tambon Klong Ha, Amphoe Klong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : tump@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang Chongwat Sanpradon 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office
196 Phahonyothin Road, Chatuchak Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FMBL MTC.002 Rev.3



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-63/0837 MTC No. EEL. BP. 46/0963

9. Peak C sound level

Number of cycles in test signal	Anticipated value (dB)	Measured value (dB)	Deviated value (dB)	Uncertainty (±dB)	Tolerance limits Class 2 (±dB)
Complete cycle	125.4	125.7	0.3	0.20	2.4
Positive half cycle	124.4	123.9	-0.5	0.20	1.4
Negative half cycle	124.4	124.0	-0.4	0.20	1.4

10. Overload indication

Measured value (dB)	Deviated value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
Positive one-half cycle 130.8	0.0	0.30	1.8

Approved by :

Komkrit Laohasiri
(Mr. Komkrit Laohasiri)

Electrical and Electronic Standards Laboratory
Industrial Metrology and Testing Service Center

Date of Calibration : 17-28 Sep. 2020
Date of Issue : 29 Sep. 2020
Ref : 2011263091503604003

End of Certificate 8 / 8

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu. 3 Tambon Withee Pa, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2377 9000
Fax. (66) 0 2377 9009
E-mail : rumpal@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C, Bangsoo Industrial Estate, Sukhumvit Road,
Amphoe Muang Chongwat Sanulakarn 10280, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumatee@tistr.or.th

PM.BLMTC.002 Rev.3

W	FO.LAB 6.4-1/28	แก้ไขครั้งที่ : 0	วันที่แก้ไข : 1 ม.ค. 2562	หน้า : 1 ของ 1
---	-----------------	-------------------	---------------------------	----------------

แบบบันทึกการตรวจสอบเครื่อง Sound Level Meter

เครื่อง CAL11 Sound Calibrator SN 520272 รหัสเครื่อง SN 93.80 ± 0.3, 113.88 ± 0.3
วันที่สอบเทียบ 31/05/64 วันที่สอบเทียบจริงไป 30/05/65
รหัสเครื่อง WVL 0055
วันที่สอบเทียบจริงไป 01/12/65
รหัสเครื่อง WVL 0160
วันที่สอบเทียบจริงไป 12/12/66
การตรวจสอบแหล่งออกเสียง
อุณหภูมิ (°C) 24 เกณฑ์การยอมรับ 23.0 ± 2.0
ความชื้นสัมพัทธ์ (%) 46 เกณฑ์การยอมรับ 50.0 ± 15.0
วันที่ทวนสอบ 30/04/65

Item	ระดับเสียงที่วัดได้ (dB) (ความดังที่ 94.0dB)	ระดับเสียงที่วัดได้ (dB) (ความดังที่ 114.0dB)	Item	ระดับเสียงที่วัดได้ (dB) (ความดังที่ 94.0dB)	ระดับเสียงที่วัดได้ (dB) (ความดังที่ 114.0dB)
1	93.8	113.9	1	93.8	113.9
2	93.8	113.9	2	93.8	113.9
3	93.8	113.9	3	93.8	113.9
4	93.8	113.9	4	93.8	113.9
5	93.8	113.9	5	93.8	113.9
6	93.8	113.9	6	93.8	113.9
7	93.8	113.9	7	93.8	113.9
8	93.8	113.9	8	93.8	113.9
9	93.8	113.9	9	93.8	113.9
10	93.8	113.9	10	93.8	113.9
X	93.80	113.90	X	93.80	113.90
SD	0.00	0.00	SD	0.00	0.00
%RSD (≤ 10)	0.00	0.00	%RSD (≤ 10)	0.00	0.00
ผลการ ทวนสอบ	ผ่าน	ผ่าน	ผลการ ทวนสอบ	ผ่าน	ผ่าน

ผู้บันทึก *สมชาย* ผู้ตรวจสอบ *สมชาย*
ผู้บันทึก *สมชาย* ผู้ตรวจสอบ *สมชาย*



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-55/0137

MTC No. EEL. BP. 105/1164

CALIBRATION CERTIFICATE

Submitted by : WATER ANALYSIS CENTER CO., LTD.

Address : 1/94 MOO 5, T.KANHAM, A.U-THAI, AYUTHAYA 13210.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.

Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., A.Muang, Samutprakan 10280.

Instrument Calibrated :

Description : Sound Level Meter

Manufacturer : Rion

Model : NL-42

Serial No. : 00395803 (WWL 0160)

Microphone : Type UC-52 No.180449

Preamplifier : Type NH-24 No.87814

Standards used :

1. Band Pass Filter Stamford Research Systems SR 650 S/N 28712.
2. Condenser Microphone Brüel&Kjær 4180 S/N 2889871.
3. Decade Attenuator Audio AL-205 S/N 00464602.
4. Function/Arbitrary Waveform Generator Agilent 33220A S/N MY4402668.
5. Digital Function Synthesizer NF Electronic Instruments DF-193A S/N 122037.
6. Digital Multimeter Fluke 8520A S/N 4985007.
7. Pistonphone Rion NC-72 S/N 00402446.
8. Measuring Amplifier Brüel&Kjær 2636 S/N 1557484.

Date of Receipt : 26 Nov. 2021

Date of Calibration : 13-16 Dec. 2021

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu.3 Tambon Khlong Ha, Amphoe Khlong Luang
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rump@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : rmtc@tistr.or.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FM.BL.MTC.002 Rev.4



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-55/0137

MTC No. EEL. BP. 105/1164

9. Power Amplifier Brüel&Kjær 2706 S/N 1517650.
10. Speaker Tannoy Limited, Great Britain British Patent No. 215300.
11. Digital Multimeter Agilent 34401A S/N MY44005560.
12. Programmable Attenuator Tamagawa TPA-303A S/N 2212.

Calibration Procedure :

This instrument has been calibrated by using calibration procedures no CP-102-02 and CP-102-03, which were based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). These calibration procedures were related to the electrical and acoustic signal tests. The electrical signal test was carried out with the direct measurement method. The acoustic signal test was performed in an anechoic room with the comparison measurement method.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%.

Date of Calibration : 13-16 Dec. 2021

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu.3 Tambon Khlong Ha, Amphoe Khlong Luang
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rump@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : rmtc@tistr.or.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FM.BL.MTC.002 Rev.4

1. Absolute Sensitivity

Reference Acoustic Signal (dB)	Unit Under Test			Tolerance Limit Class 2 (±dB)
	Measured Value (dB)	Deviation (dB)	Uncertainty (±dB)	
113.91	Before adjust 114.1	After adjust 113.9	0.0	0.30
				1.4

Note: The external calibration adjustment was firstly performed. The internal calibration adjustment was then completed at the display of 113.9 dB.

2. Self-generated noise

2.1 Normal test

Measured value (dB)	Uncertainty (±dB)
16.5	0.10

2.2 The microphone of the sound level meter was replaced by electrical signal input device

Frequency	Measured Value (dB)	Uncertainty (±dB)
Weighting	12.6	0.10
A-Weighting	17.8	0.10
C-Weighting	23.2	0.10

Date of Calibration : 13-16 Dec.2021

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office : 35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpal@tistr.or.th Website: www.tistr.or.th

Office/Laboratory : Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road, Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

PM.BLMTC.002 Rev.4

4/8

3. Acoustical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)	Flat (dB)	
125	-0.2	-0.1	-0.1	0.40
1 000	-0.1	-0.1	-0.1	0.40
4 000	-0.8	-0.7	-0.7	0.40
				2.0
				1.4
				3.6

4. Electrical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)	Flat (dB)	
63	0.0	-0.1	-0.1	0.20
125	-0.1	0.0	-0.1	0.20
250	0.0	0.0	0.0	0.20
500	0.0	0.0	0.0	0.20
1 000	0.0	0.0	0.0	0.20
2 000	-0.1	0.0	-0.1	0.20
4 000	0.0	0.0	0.0	0.20
8 000	0.1	0.1	0.0	0.20
				1.9
				1.4
				2.6
				3.6
				5.6

Date of Calibration : 13-16 Dec.2021

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office : 35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpal@tistr.or.th Website: www.tistr.or.th

Office/Laboratory : Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road, Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

PM.BLMTC.002 Rev.4

4/8

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
A-weighting	94.0	0.0	0.20	0.4
C-weighting	94.0	0.0	0.20	0.4
Flat	94.0	0.0	0.20	0.4

5.2 Time weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
Fast	94.0	0.0	0.20	0.3
Slow	94.0	0.0	0.20	0.3
Leq	94.0	0.0	0.20	0.3

6. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
137	137.0	0.0	0.30	1.4
136	136.1	0.1	0.30	1.4
135	135.0	0.0	0.30	1.4
134	134.1	0.1	0.30	1.4
133	133.1	0.1	0.30	1.4
132	132.0	0.0	0.30	1.4
131	131.0	0.0	0.30	1.4

Date of Calibration : 13-16 Dec.2021

5/8

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax (66) 0 2577 9009
E-mail : numpai@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FM.B/LMTC.002 Rev.4

6. Level linearity on the reference level range (cont.)

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
130	130.0	0.0	0.30	1.4
129	129.0	0.0	0.30	1.4
124	124.0	0.0	0.30	1.4
119	119.0	0.0	0.30	1.4
114	114.0	0.0	0.30	1.4
109	109.0	0.0	0.30	1.4
104	104.0	0.0	0.30	1.4
99	99.0	0.0	0.30	1.4
94	94.0	0.0	0.30	1.4
89	89.0	0.0	0.30	1.4
84	84.1	0.1	0.30	1.4
79	79.0	0.0	0.30	1.4
74	74.0	0.0	0.30	1.4
69	69.0	0.0	0.30	1.4
64	64.0	0.0	0.30	1.4
59	59.0	0.0	0.30	1.4
54	54.0	0.0	0.30	1.4
49	48.9	-0.1	0.30	1.4
44	44.0	0.0	0.30	1.4
39	39.0	0.0	0.30	1.4
34	34.0	0.0	0.30	1.4
29	28.9	-0.1	0.30	1.4
28	28.0	0.0	0.30	1.4

Date of Calibration : 13-16 Dec.2021

6/8

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax (66) 0 2577 9009
E-mail : numpai@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FM.B/LMTC.002 Rev.4

6. Level linearity on the reference level range (cont.)

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
27	27.0	0.0	0.30	1.4
26	25.9	-0.1	0.30	1.4
25	25.0	0.0	0.30	1.4

7. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)		Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
		Value	Value			
20-130	125	125.0	0.0	0.0	0.30	1.4

8. Tone burst response

Time Weighting	Toneburst Duration, T _b (ms)	Measured Value (dB)		Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (dB)
		Value	Value			
Fast	200	126.0	0.0	0.0	0.20	±1.3
	2	109.0	0.0	0.0	0.20	+1.3; -2.8
	0.25	99.9	-0.1	0.0	0.20	+1.8; -5.3
Slow	200	119.5	-0.1	0.0	0.20	±1.3
	2	99.9	-0.1	0.0	0.20	+1.3; -5.3
	0.25	120.0	0.0	0.0	0.20	±1.3
SEL	2	100.0	0.0	0.0	0.20	+1.3; -2.8
	0.25	90.9	-0.1	0.0	0.20	+1.8; -5.3

Date of Calibration : 13-16 Dec.2021

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office 35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Changwat Pathumthani 12120, Thailand Tel. (66) 0 2577 9000 Fax. (66) 0 2577 9009 E-mail : numpai@tistr.or.th Website: www.tistr.or.th

Office 196 Phahonyothin Road, Chatuchak, Bangkok 10900, Thailand Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217 Fax. (66) 0 2579 8592 E-mail : surmalee@tistr.or.th

FM.BI-MTC.002 Rev.4

7/8

9. Peak C sound level

Number of cycles in test signal	Anticipated value (dB)	Measured value (dB)	Deviated value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
Complete cycle	125.4	125.4	0.0	0.20	2.4
Positive half cycle	124.4	124.1	-0.3	0.20	1.4
Negative half cycle	124.4	124.1	-0.3	0.20	1.4

10. Overload indication

Measured value (dB)	Deviated value (dB)		Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
	Positive	Negative		
136.6	136.6	0.0	0.30	1.8

Calibrated by :

Panya Phasingert
(Mr. Panya Phasingert)

Approved by :

Tawikiat Jamsanran
(Mr. Tawikiat Jamsanran)

Electrical and Electronic Standards Laboratory
Industrial Metrology and Testing Service Centre

Date of Calibration : 13-16 Dec.2021

Date of Issue : 17 Dec. 2021

Ref : 2011264112604939002

End of Certificate

8/8

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office 35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Changwat Pathumthani 12120, Thailand Tel. (66) 0 2577 9000 Fax. (66) 0 2577 9009 E-mail : numpai@tistr.or.th Website: www.tistr.or.th

Office 196 Phahonyothin Road, Chatuchak, Bangkok 10900, Thailand Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217 Fax. (66) 0 2579 8592 E-mail : surmalee@tistr.or.th

FM.BI-MTC.002 Rev.4



NSC-TISTR 17025
CALIBRATION 0037

TISTR

THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-65/0137

MTC No. EEL BP. 104/1164

CALIBRATION CERTIFICATE

Submitted by : WATER ANALYSIS CENTER CO., LTD.

Address : 1/94 MOO 5, T.KANHAM, A.U-THAI, AYUTTHAYA 13210.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.

Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., A.Muang, Samutprakan 10280.

Instrument Calibrated :

Description : Sound Level Meter

Manufacturer : Rion

Model : NL-42

Serial No. : 00396923 (WWL 0161)

Microphone : Type UC-52 No.180583

Preamplifier : Type NH-24 No.87936

Standards used :

1. Band Pass Filter Stanford Research Systems SR 650 S/N 28712.
2. Condenser Microphone Brüel&Kjær 4180 S/N 2889871.
3. Decade Attenuator Ando AL-205 S/N 00464602.
4. Function/Arbitrary Waveform Generator Agilent 33220A S/N MY44042668.
5. Digital Function Synthesizer NF Electronic Instruments DF-193A S/N 122037.
6. Digital Multimeter Fluke 8520A S/N 4985007.
7. Pistonphone Rion NC-72 S/N 00402446.
8. Measuring Amplifier Brüel&Kjær 2636 S/N 1537484.

Ambient Environment

Temperature : (23 ± 3) °C

Relative Humidity : (50 ± 15) %

Ambient Pressure : (101.325 ± 1.5) kPa

Date of Receipt : 26 Nov. 2021

Date of Calibration : 13-16 Dec.2021

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rump@tistr.or.th Website: www.tistr.or.th

Office/Laboratory

Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office

196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

ผู้รับใช้ :

ผู้ตรวจสอบ :

ผู้รับใช้ :

ผู้ตรวจสอบ :

W	FO-LAB 6.4-1 /28	แก้ไขครั้งที่ : 0	วันที่บังคับใช้ : 1 ม.ค. 2562	หน้า : 1 ของ 1
---	------------------	-------------------	-------------------------------	----------------

แบบบันทึกการทวนสอบเครื่อง Sound Level Meter

เครื่อง CA111 Sound Calibrator SN 520272 รหัสเครื่องมือ SR004 เกณฑ์การยอมรับ 93.80 ± 0.3, 113.98 ± 0.3
วันที่สอบเทียบ 31/05/64 วันที่สอบเทียบจริงไป 30/05/65
เครื่อง Digital Thermohygrometer S/N 105091609 รหัสเครื่องมือ WWL 0055
วันที่สอบเทียบ 02/12/64 วันที่สอบเทียบจริงไป 01/12/65
เครื่อง Sound Level Meter S/N 00396923 รหัสเครื่องมือ WWL 0161
วันที่สอบเทียบ 13-16/12/64 วันที่สอบเทียบจริงไป 12/12/66

การทวนสอบก่อนออกให้งาน

อุณหภูมิ (°C) 24 เกณฑ์การยอมรับ 23.0 ± 2.0

ความชื้นสัมพัทธ์ (%) 46 เกณฑ์การยอมรับ 50.0 ± 15.0

วันที่ทวนสอบ 24/04/65

การทวนสอบหลังทดสอบให้งาน

อุณหภูมิ (°C) 24 เกณฑ์การยอมรับ 23.0 ± 2.0

ความชื้นสัมพัทธ์ (%) 46 เกณฑ์การยอมรับ 50.0 ± 15.0

วันที่ทวนสอบ 30/04/65

Item	ระดับเสียงที่วัดได้ (dB) (ความดังที่ 94.0dB)	ระดับเสียงที่วัดได้ (dB) (ความดังที่ 114.0dB)
1	93.8	113.9
2	93.8	113.9
3	93.8	113.9
4	93.8	113.9
5	93.8	113.9
6	93.8	113.9
7	93.8	113.9
8	93.8	113.9
9	93.8	113.9
10	93.8	113.9
X	93.80	113.90
SD	0.00	0.00
%RSD (≤ 10)	0.00	0.00
ผลการ ทวนสอบ	ผ่าน	ผ่าน



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-65/0137

MTC No. EEL. BP. 104/1164

9. Power Amplifier Brüel&Kjær 2706 S/N 1517650.
10. Speaker Tannoy Limited, Great Britain British Patent No. 215300.
11. Digital Multimeter Agilent 34401A S/N MY44005560.
12. Programmable Attenuator Tamagawa TPA-303A S/N 2212.

Calibration Procedure .

This instrument was calibrated by using calibration procedures no CP-102-02 and CP-102-03, which were based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). These calibration procedures were related to the electrical and acoustic signal tests. The electrical signal test was carried out with the direct measurement method. The acoustic signal test was performed in an anechoic room with the comparison measurement method.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%.

Date of Calibration : 13-16 Dec.2021

2/8
Forming

The results relate only to the items tested/calibrated or value assigned.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.EEL.MTC.002 Rev.4

Head Office
35 Mu. 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 8992
E-mail : numpai@tistr.or.th Websites:www.tistr.or.th

Office
196 Phahonyothin Road, Chatuchak Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8992
E-mail : sumalee@tistr.or.th



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-65/0137

MTC No. EEL. BP. 104/1164

1. Absolute Sensitivity

Reference Acoustic Signal (dB)	Unit Under Test			Tolerance Limit Class 2 (±dB)
	Measured Value (dB)	Deviation (dB)	Uncertainty (±dB)	
113.91	Before adjust 114.2	After adjust 113.9	0.0	0.30
				1.4

Note: The external calibration adjustment was firstly performed. The internal calibration adjustment was then completed at the display of 124.9 dB.

2. Self-generated noise

2.1 Normal test

Measured value (dB)	Uncertainty (±dB)
16.4	0.10

2.2 The microphone of the sound level meter was replaced by electrical signal input device

Frequency	Measured Value (dB)	Uncertainty (±dB)
Weighting	12.5	0.10
A-Weighting	17.7	0.10
C-Weighting	23.4	0.10

Date of Calibration : 13-16 Dec.2021

3/8
Forming

The results relate only to the items tested/calibrated or value assigned.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.EEL.MTC.002 Rev.4

Head Office
35 Mu. 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 8992
E-mail : numpai@tistr.or.th Websites:www.tistr.or.th

Office
196 Phahonyothin Road, Chatuchak Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8992
E-mail : sumalee@tistr.or.th

3. Acoustical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)	Flat (dB)		
125	-0.1	0.0	0.0	0.40	2.0
1 000	-0.3	-0.3	-0.3	0.40	1.4
4 000	-0.6	-0.6	-0.6	0.40	3.6

4. Electrical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)	Flat (dB)		
63	0.0	0.0	0.0	0.20	2.5
125	0.0	0.0	0.0	0.20	2.0
250	0.0	0.0	0.0	0.20	1.9
500	0.0	0.0	0.0	0.20	1.9
1 000	0.0	0.0	0.0	0.20	1.4
2 000	0.0	0.1	0.0	0.20	2.6
4 000	0.0	0.1	0.0	0.20	3.6
8 000	0.1	0.2	0.0	0.20	5.6

Date of Calibration : 13-16 Dec.2021

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Head Office
35 Mu. 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpat@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office
196 Phichayothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8992
E-mail : sumalee@tistr.or.th

4/8

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
A-weighting	94.0	0.0	0.20	0.4
C-weighting	94.0	0.0	0.20	0.4
Flat	94.0	0.0	0.20	0.4

5.2 Time weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
Fast	94.0	0.0	0.20	0.3
Slow	94.0	0.0	0.20	0.3
Leq	94.0	0.0	0.20	0.3

6. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
137	137.0	0.0	0.30	1.4
136	136.0	0.0	0.30	1.4
135	135.0	0.0	0.30	1.4
134	134.0	0.0	0.30	1.4
133	133.0	0.0	0.30	1.4
132	132.0	0.0	0.30	1.4
131	131.0	0.0	0.30	1.4

Date of Calibration : 13-16 Dec.2021

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Head Office
35 Mu. 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpat@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office
196 Phichayothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8992
E-mail : sumalee@tistr.or.th

5/8

6. Level linearity on the reference level range (cont.)

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
130	130.0	0.0	0.30	1.4
129	129.0	0.0	0.30	1.4
124	124.0	0.0	0.30	1.4
119	119.0	0.0	0.30	1.4
114	114.0	0.0	0.30	1.4
109	109.0	0.0	0.30	1.4
104	104.0	0.0	0.30	1.4
99	99.0	0.0	0.30	1.4
94	94.0	0.0	0.30	1.4
89	89.0	0.0	0.30	1.4
84	84.1	0.1	0.30	1.4
79	79.0	0.0	0.30	1.4
74	74.0	0.0	0.30	1.4
69	69.0	0.0	0.30	1.4
64	64.0	0.0	0.30	1.4
59	59.0	0.0	0.30	1.4
54	54.0	0.0	0.30	1.4
49	49.0	0.0	0.30	1.4
44	44.0	0.0	0.30	1.4
39	39.0	0.0	0.30	1.4
34	34.0	0.0	0.30	1.4
29	28.9	-0.1	0.30	1.4
28	27.9	-0.1	0.30	1.4

6/8

Date of Calibration : 13-16 Dec.2021

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road, Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office
196 Phayathai Road, Chatchak, Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

6. Level linearity on the reference level range (cont.)

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
27	26.9	-0.1	0.30	1.4
26	25.9	-0.1	0.30	1.4
25	24.9	-0.1	0.30	1.4

7. Level linearity including the level range control

Range	Measured Value (dB)		Deviated Value (dB)		Uncertainty (±dB)		Tolerance Limits Class 2 (±dB)
	Anticipated Value (dB)	Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Value (dB)	
20-130	125	125.0	0.0	0.0	0.30	0.0	1.4

8. Tone burst response

Time Weighting	Toneburst Duration, Tb (ms)	Measured Value (dB)		Deviated Value (dB)		Uncertainty (±dB)		Tolerance Limits Class 2 (dB)
		Anticipated Value (dB)	Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Value (dB)	
Fast	200	126.0	0.0	0.0	0.0	0.20	±1.3	±1.3
	2	108.9	-0.1	0.20	-0.1	0.20	+1.3; -2.8	+1.3; -2.8
	0.25	99.9	-0.1	0.20	-0.1	0.20	+1.8; -5.3	+1.8; -5.3
Slow	200	119.5	-0.1	0.20	-0.1	0.20	±1.3	±1.3
	2	99.9	-0.1	0.20	-0.1	0.20	+1.3; -5.3	+1.3; -5.3
	0.25	120.0	0.0	0.20	0.0	0.20	±1.3	±1.3
SEL	200	100.0	0.0	0.20	0.0	0.20	+1.3; -2.8	+1.3; -2.8
	2	90.9	-0.1	0.20	-0.1	0.20	+1.8; -5.3	+1.8; -5.3
	0.25	90.9	-0.1	0.20	-0.1	0.20	+1.8; -5.3	+1.8; -5.3

7/8

Date of Calibration : 13-16 Dec.2021

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sol 1C, Bangpoo Industrial Estate, Sukhumvit Road, Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office
196 Phayathai Road, Chatchak, Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th



NSC-TISTR 17025
CALIBRATION 0037

THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

MTC No. EEL BP. 104/1164

Request No. 21-65/0137

9. Peak C sound level

Number of cycles in test signal	Anticipated value (dB)	Measured value (dB)	Deviated value (dB)	Uncertainty (±dB)	Tolerance limits Class 2 (±dB)
Complete cycle	125.4	125.4	0.0	0.20	2.4
Positive half cycle	124.4	124.1	-0.3	0.20	1.4
Negative half cycle	124.4	124.1	-0.3	0.20	1.4

10. Overhead indication

Measured value (dB)		Deviated value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
Positive one-half cycle	Negative one-half cycle	0.0	0.30	1.8
136.7	136.7			

Approved by :

Calibrated by : *Panya Phasingst*
(Mr. Panya Phasingst)
Tanya Phasingst
(Mr. Tawakiat Jansamran)

Electrical and Electronic Standards Laboratory
Industrial Metrology and Testing Service Centre

Date of Calibration : 13-16 Dec.2021
Date of Issue : 17 Dec. 2021
Ref : 201126412604939001

End of Certificate

8 / 8

The results relate only to the items tested/calibrated or value assigned.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.LMTC.002 Rev.4

Head Office
35 Mu. 3 Tambon Khlong Ha, Amphoe Khlong Luang, Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9009
Fax (66) 0 2577 9009
E-mail : tump@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sri Jit, Bangpoo Industrial Estate, Sukhumvit Road, Amphoe Muang, Changwat Samutprakarn 10280, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

Office
196 Phahonyothin Road, Chatuchak Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

Certificate of System Qualification

GC-OQ + GCMS-OQ

System ID: CN10830014
Organization Name: S.P.S. Consulting Service Co., Ltd.
Organization Location: 7 Soi Paholyothin 24 Bangkok 10900
Date: September 8, 2021 11:48:04 AM
EQP Name: AgilentRecommended, AgilentRecommended
EQP Revision: GC.02.51, GCMS.02.51
Overall Qualification Status: Pass

System Inspection and Basic Safety and Operation

Name: 6890
Setup Status: Pass

Overall System Inspection and Basic Safety and Operation Test Status

Pass

Inlet Pressure Decay

Name: 6890 Front SSL
Setup Status: Pass
Pressure: 25.0 psi
Pressure Change: -0.1 psi / 5 minutes
Agilent Recommended: >= -2.0 and <= 0.5

Overall Inlet Pressure Decay Test Status

Pass

Inlet Pressure Accuracy

Name: 6890 Front SSL

Date: September 8, 2021 11:48:04 AM
System ID: CN10830014

Setpoint Status: Pass
Flow Type: Oxidizer
Setpoint: 400.0 mL/min Measured Flow: 401.6 mL/min
Accuracy: 1.6 mL/min
Agilent Recommended: <= 10.0 % setpoint (40.0 mL/min)
Limit is percentage of setpoint or 0.5 mL/minute, whichever is largest.

Setpoint Status: Pass
Flow Type: Makeup
Setpoint: 25.0 mL/min Measured Flow: 25.7 mL/min
Accuracy: 0.7 mL/min
Agilent Recommended: <= 10.0 % setpoint (2.5 mL/min)
Limit is percentage of setpoint or 0.5 mL/minute, whichever is largest.

Overall Detector Flow Accuracy Test Status

Pass

GC Oven Temperature Accuracy

Name: 6890
Setpoint Status: Pass
Zone: Oven
Setpoint/Actual
Temperature: 230.0 230.5 °C
Accuracy: 0.5 °C
Agilent Recommended: >= -1.0 % setpoint in K (-5.0 °C)
<= 1.0 % setpoint in K (5.0 °C)

Setpoint Status: Pass
Inlet Pressure: 25.0 psi Actual 24.9 psi
Accuracy: 0.1 psi
Agilent Recommended: <= 1.2 psi

Overall Inlet Pressure Accuracy Test Status

Pass

Inlet Pressure Accuracy

Name: 6890 Back SSL
Setpoint Status: Pass
Setpoint Actual
Inlet Pressure: 25.0 psi 25.0 psi
Accuracy: 0.0 psi
Agilent Recommended: <= 1.2 psi

Overall Inlet Pressure Accuracy Test Status

Pass

Detector Flow Accuracy

Name: 6890 Front FID
Setpoint Status: Pass
Flow Type: Fuel
Setpoint: 30.0 mL/min Measured Flow: 30.8 mL/min
Accuracy: 0.8 mL/min
Agilent Recommended: <= 10.0 % setpoint (3.0 mL/min)
Limit is percentage of setpoint or 0.5 mL/minute, whichever is largest.

Name: 6890

Setpoint Status: Pass

Base Signal: 20.2 pA

ASTM Noise
counts

788.00

Drift
counts/Hr

321.58

195.20

19200.00

Agilent Recommended: <=

Status: Pass

Overall Noise and Drift Test Status

Pass

Injection Precision

Tested Combination1

Name: 7683B

Setpoint Status: Pass

Injection Volume on Column:

1.0

uL

Area RSD: 0.41 %

Agilent Recommended: <= 3.00

Retention Time RSD: 0.13 %

Agilent Recommended: <= 1.00

Overall Injection Precision Test Status

Pass

Signal to Noise

Tested Combination1

Name: 6890

Setpoint Status: Pass

Signal to Noise: 1019853

Agilent Recommended: >= 300000

Date: September 8, 2021 11:48:04 AM

System ID: CN10630014

Setpoint Status: Pass

Zone: Oven

Temperature: 100.0 100.1 °C

Accuracy: 0.1 °C

Agilent Recommended: >= -3.7 °C

Agilent Recommended: <= 3.7 °C

Overall GC Oven Temperature Accuracy Test Status

Pass

GC Oven Temperature Stability

Name: 6890

Setpoint Status: Pass

Temperature: 100.0 100.1 °C

Stability: 0.0 °C

Agilent Recommended: <= 0.5

Overall GC Oven Temperature Stability Test Status

Pass

Scouting Run

Tested Combination1

Name: 7683B

Setpoint Status: Completed

Injection Volume on Column: 1.0 uL

Overall Scouting Run Status: Completed

Noise and Drift

Tested Combination1

Name: 6890

Setpoint Status: Pass

Signal to Noise: 1019853

Agilent Recommended: >= 300000

Date: September 8, 2021 11:48:04 AM

System ID: CN10630014

Signal to Noise EI

Tested Combination2 Back SSL / External SQ

Name: 5975C

Source: EI - Inert Filament 1

Setpoint Status: Pass

Signal to Noise: 809

Agilent Recommended: ≥ 160

Source: EI - Inert Filament 2

Setpoint Status: Pass

Signal to Noise: 288

Agilent Recommended: ≥ 160

This test's 0 comment(s) and 2 deviation(s) are available in the Attachments section.

Overall Signal to Noise EI Test Status

Pass

Date: September 8, 2021 11:48:04 AM

System ID: CN10630014

Overall Signal to Noise Test Status

Pass

Log Amp

Tested Combination2 Back SSL / External SQ

Name: 5975C

Setpoint Status: Pass

Overall Log Amp Test Status

Pass

RFPA

Tested Combination2 Back SSL / External SQ

Name: 5975C

Setpoint Status: Pass

Amu: 1050 m/z Drift After Five Minutes: 10 mV and ≤ 100 RFPA Voltage: 498 mV ≤ 1100

Agilent Recommended: ≥ -100

Overall RFPA Test Status

Pass

Tune EI

Tested Combination2 Back SSL / External SQ

Name: 5975C

Setpoint Status: Pass

Filament 1

Setpoint Status: Pass

Filament 2

Overall Tune EI Test Status

Pass

Date: September 8, 2021 11:48:04 AM

System ID: CN10630014

Sampler 2

Manufacturer	Agilent Technologies
Type	Manual Injection
Usage	Sample Injection
Syringe Volume (µL)	10
Mainframe 1	
Manufacturer	Agilent Technologies
Name	6890
Model Number	G1530N
Serial Number	CN10630014
Firmware Revision	N.02.01
Oven Type	Standard
Inlet 1	
Manufacturer	Agilent Technologies
Name	6890
Type	SSL
Location	Front
Carrier Gas	Helium
Control Type	Electronic Pressure Control (EPC)
Purged Inlet	Yes
Inlet 2	
Manufacturer	Agilent Technologies
Name	6890
Type	SSL
Location	Back
Carrier Gas	Helium
Control Type	Electronic Pressure Control (EPC)
Purged Inlet	Yes

Instrument Details

Purpose
This section describes the as found system configuration.

Details

System	System ID	CN10630014
Manufacturer	Agilent Technologies	
Name	6890	
Flow Data Input	Manual Data	
Temperature Data Input	Manual Data or Other Data Logging	
Tested Combination1		
Injection Technique	Injection Tower	
Inlet	Front	
Detector	Front	
LTM Included?	No	
Tested Combination2		
Injection Technique	Manual Injection	
Inlet	Back	
Detector	External	
LTM Included?	No	
Sampler 1		
Manufacturer	Agilent Technologies	
Type	Injection Tower	
Name	7683B	
Model Number	G2813A	
Serial Number	CN64138101	
Usage	Sample Injection	
Location	Front	
Syringe Volume (µL)	10	

Electronic Signature

Purpose

This signature page was created and published because the ACE sign-off action was executed, which is valid for the entire document, including attachments. The ACE sign-off is an electronic signature that requires two distinct identification components: unique username and personal password. The Agilent representative who has delivered this service understands the meaning and legal status of an electronic signature. As a trained official operator, the Agilent representative has a unique password and login to access ACE and electronically sign this document. (Other e-signatures can be applied to this document using a Document Content Management or other suitable method defined in your data access and control procedures.)

Details

Full Name of Signer: Adirek Rattanawijit
Logged On User Name: adirek.rattanawijit@non.agilent.com
Signature Creation Date: September 8, 2021
Reason for Signature: Executed protocol and published this original version of document

Regulatory Disclaimer

This document provides a protocol to verify and record instrument configuration and evidence of proper operation. It has been prepared from our interpretation of applicable regulations as well as industry best practices. The document is designed to provide an important component of a complete compliance package. Validation depends upon many factors and use of this protocol alone does not assure compliance. Agilent Technologies makes no promises or representations as to its sufficiency for any specific regulatory program.

Warranty

Agilent Technologies makes no warranty of any kind to this material, including but not limited to, the implied warranties or merchantability and fitness for a particular purpose. Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Date: September 8, 2021 11:48:04 AM
System ID: CN10630014

Detector 1

Manufacturer: Agilent Technologies
Name: Mass Spectrometer
Type: Mass Spectrometer
Location: External

Detector 2

Manufacturer: Agilent Technologies
Name: 8890
Type: FID
Adapter: Capillary
Control Type: Electronic Pressure Control (EPC)
Location: Front
Makeup Gas: Nitrogen

Mass Spectrometer 1

Manufacturer: Agilent Technologies
Type: SQ
Name: 5875C
Serial Number: US61633454
Firmware Revision: 5.02.04
High Vacuum System: Turbo Pump
Scouting Run Standard: OFN Std

MS EI Source 1

Manufacturer: Agilent Technologies
Source Type: EI - Inert
Number of filaments: 2

Date: September 8, 2021 11:48:04 AM
System ID: CN10630014

User Name: admin@crosslab.com

Host Name: DESKTOP-9M7V99

SP8_OOQCMS Transaction log:

System ID: CN10530014

Print Date: September 8, 2021 11:48:04 AM

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
September 7, 2021 3:12:48 PM	End	Execution	System Inspection and Basic Safety and Operation - 6850: - Qualitative Test - No endpoints associated	Run Count: 1
September 7, 2021 3:13:47 PM	Start	Execution	GC Solving Run - Injection Tower, Front SSL, Front FID: - Part of System Preparation - No limits associated	None
September 7, 2021 3:40:18 PM	Start	Execution	GC Solving Run - Injection Tower, Front SSL, Front FID: - Part of System Preparation - No limits associated	None
September 7, 2021 3:40:29 PM	Start	Execution	Initial Pressure Decay - Front SSL - Pressure Controlled Inlet - S: 25.0 psi - L: >= 2.0 psi and <= 0.5 psi	None
September 7, 2021 3:40:39 PM	End	Execution	Initial Pressure Decay - Front SSL - Pressure Controlled Inlet - S: 25.0 psi - L: >= 2.0 psi and <= 0.5 psi	Run Count: 1
September 7, 2021 3:40:41 PM	Start	Execution	Initial Pressure Accuracy - Front SSL - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	None
September 7, 2021 3:40:45 PM	End	Execution	Initial Pressure Accuracy - Front SSL - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	Run Count: 1
September 7, 2021 3:40:49 PM	Start	Execution	Initial Pressure Accuracy - Back SSL - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	None
September 7, 2021 3:40:53 PM	End	Execution	Initial Pressure Accuracy - Back SSL - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	Run Count: 1

Page 2 / 12

Page 2 / 12

Date: September 8, 2021 11:48:04 AM
System ID: CN10530014

Page 13 / 23

User Name: admin@crosslab.com
Host Name: DESKTOP-9M7V99

System ID: CN10530014
Print Date: September 8, 2021 11:48:04 AM

SP8_OOQCMS Transaction log:

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
September 7, 2021 2:28:20 PM	Start	Session Created	Session	None
September 7, 2021 2:28:20 PM	Start	Configuration	Session	None
September 7, 2021 2:28:20 PM	Audit	Entitlement	Licensing	User is Field Engineer and does not require an unlock code
September 7, 2021 3:12:19 PM	Audit	Exploded	Session	EQP details for primary technique (GC) - File path: [Protocol]PackagingConfiguration072.51(05.02.51.eap). EQP File Name: [GC.02.51.eap], EQP Name: [AgilentRecommended] EQP details for hybridized technique (GC/MS) - File path: [Protocol]PackagingConfiguration072.51(05.02.51.eap). EQP File Name: [GC/MS.02.51.eap], EQP Name: [AgilentRecommended]
September 7, 2021 3:12:25 PM	End	Configuration	Session	None
September 7, 2021 3:12:30 PM	Start	Qualification	Session	OQ
September 7, 2021 3:12:30 PM	Start	Erection	System Inspection and Basic Safety and Operation - 6850: - Qualitative Test - No endpoints associated	None

Page 1 / 12

Page 1 / 12

Date: September 8, 2021 11:48:04 AM
System ID: CN10530014

Page 12 / 23

User Name: admin@netmanwille
Host Name: DESKTOP-68M7V96
System ID: CN10630014
Print Date: September 8, 2021 11:45:08 AM

SPS_OGCMS Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
September 7, 2021 3:42:09 PM	Audit	Data	GC Oven Temperature Accuracy - 6890 - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Manual Data Entry
September 7, 2021 3:42:11 PM	End	Execution	GC Oven Temperature Accuracy - 6890 - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Run Count : 1
September 7, 2021 3:42:13 PM	Start	Execution	GC Oven Temperature Accuracy - 6890 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	None
September 7, 2021 3:42:34 PM	Audit	Data	GC Oven Temperature Accuracy - 6890 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Manual Data Entry
September 7, 2021 3:42:36 PM	End	Execution	GC Oven Temperature Accuracy - 6890 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Run Count : 1
September 7, 2021 3:42:38 PM	Start	Execution	GC Oven Temperature Stability - 6890 - Temperature : Oven - S: 100.0°C - L: <= 0.5°C	None
September 7, 2021 3:43:24 PM	Audit	Data	GC Oven Temperature Stability - 6890 - Temperature : Oven - S: 100.0°C - L: <= 0.5°C	Manual Data Entry
September 7, 2021 3:43:27 PM	End	Execution	GC Oven Temperature Stability - 6890 - Temperature : Oven - S: 100.0°C - L: <= 0.5°C	Run Count : 1
September 7, 2021 3:43:33 PM	Start	Execution	GC Scouting Run - Injection Tower, Front SSI, Front FID - Part of System Preparation - No limits associated	None

User Name: admin@netmanwille
Host Name: DESKTOP-68M7V96
System ID: CN10630014
Print Date: September 8, 2021 11:45:08 AM

SPS_OGCMS Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
September 7, 2021 3:40:55 PM	Start	Execution	Detector Flow Accuracy - Front FID - Type: Fuel - S: 30.0 mL/min - L: <= 10.0% setpoint	None
September 7, 2021 3:41:11 PM	Audit	Data	Detector Flow Accuracy - Front FID - Type: Fuel - S: 30.0 mL/min - L: <= 10.0% setpoint	Manual Data Entry
September 7, 2021 3:41:12 PM	End	Execution	Detector Flow Accuracy - Front FID - Type: Fuel - S: 30.0 mL/min - L: <= 10.0% setpoint	Run Count : 1
September 7, 2021 3:41:14 PM	Start	Execution	Detector Flow Accuracy - Front FID - Type: Oxidizer - S: 400.0 mL/min - L: <= 10.0% setpoint	None
September 7, 2021 3:41:34 PM	Audit	Data	Detector Flow Accuracy - Front FID - Type: Oxidizer - S: 400.0 mL/min - L: <= 10.0% setpoint	Manual Data Entry
September 7, 2021 3:41:36 PM	End	Execution	Detector Flow Accuracy - Front FID - Type: Oxidizer - S: 400.0 mL/min - L: <= 10.0% setpoint	Run Count : 1
September 7, 2021 3:41:38 PM	Start	Execution	Detector Flow Accuracy - Front FID - Type: Makeup - S: 25.0 mL/min - L: <= 10.0% setpoint	None
September 7, 2021 3:41:50 PM	Audit	Data	Detector Flow Accuracy - Front FID - Type: Makeup - S: 25.0 mL/min - L: <= 10.0% setpoint	Manual Data Entry
September 7, 2021 3:41:52 PM	End	Execution	Detector Flow Accuracy - Front FID - Type: Makeup - S: 25.0 mL/min - L: <= 10.0% setpoint	Run Count : 1
September 7, 2021 3:41:54 PM	Start	Execution	GC Oven Temperature Accuracy - 6890 - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	None

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
September 7, 2021 3:49:08 PM	Start	Execution	Injection Prediction - Injection Tower, Front SSL, Front FID: - GC - L (Ave): <= 3.00% - L (Ret. Time): <= 1.00%	None
September 7, 2021 3:49:22 AM	Audit	Data	Injection Prediction - Injection Tower, Front SSL, Front FID: - GC - L (Ave): <= 3.00% - L (Ret. Time): <= 1.00%	Data File Path: K:\SP9\QCPV2021\UNUPREC 02.D\FID1A.CH
September 7, 2021 3:49:22 AM	Audit	Data	Injection Prediction - Injection Tower, Front SSL, Front FID: - GC - L (Ave): <= 3.00% - L (Ret. Time): <= 1.00%	Data File Path: K:\SP9\QCPV2021\UNUPREC 03.D\FID1A.CH
September 7, 2021 3:49:22 AM	Audit	Data	Injection Prediction - Injection Tower, Front SSL, Front FID: - GC - L (Ave): <= 3.00% - L (Ret. Time): <= 1.00%	Data File Path: K:\SP9\QCPV2021\UNUPREC 04.D\FID1A.CH
September 7, 2021 3:49:22 AM	Audit	Data	Injection Prediction - Injection Tower, Front SSL, Front FID: - GC - L (Ave): <= 3.00% - L (Ret. Time): <= 1.00%	Data File Path: K:\SP9\QCPV2021\UNUPREC 05.D\FID1A.CH
September 7, 2021 3:49:22 AM	Audit	Data	Injection Prediction - Injection Tower, Front SSL, Front FID: - GC - L (Ave): <= 3.00% - L (Ret. Time): <= 1.00%	Data File Path: K:\SP9\QCPV2021\UNUPREC 06.D\FID1A.CH
September 7, 2021 3:49:22 AM	Audit	Data	Injection Prediction - Injection Tower, Front SSL, Front FID: - GC - L (Ave): <= 3.00% - L (Ret. Time): <= 1.00%	Data File Path: K:\SP9\QCPV2021\UNUPREC 07.D\FID1A.CH
September 7, 2021 3:49:27 End	End	Execution	Injection Prediction - Injection Tower, Front SSL, Front FID: - GC - L (Ave): <= 3.00% - L (Ret. Time): <= 1.00%	Run Count: 1

Page 6 / 12

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
September 7, 2021 3:48:05 AM PM	Auto	Data	GC Scouting Run - Injection Tower, Front SSL, Front FID: - Part of System Preparation - No Inlets associated	Data File Path: K:\P9\GC\P\2021\SCOUT1 No NG01.D\WD1A.CH
September 7, 2021 3:48:38 PM	End	Execution	GC Scouting Run - Injection Tower, Front SSL, Front FID: - Part of System Preparation - No Inlets associated	Run Count: 1
September 7, 2021 3:48:41 PM	Start	Execution	Noise and Drift - Front FID: - Detector FID - L (Noise): <= 0.10 pA - L (Drift): <= 2.50 pA/hour	None
September 7, 2021 3:47:00 PM	Auto	Data	Noise and Drift - Front FID: - Detector FID - L (Noise): <= 0.10 pA - L (Drift): <= 2.50 pA/hour	Data File Path: K:\P9\GC\P\2021\9\GNBD RF01.D\WD1A.CH
September 7, 2021 3:47:51 PM	End	Qualification	Session	OO
September 7, 2021 3:47:51 PM	Start	Configuration	Session	None
September 7, 2021 3:48:45 PM	End	Configuration	Session	None
September 7, 2021 3:48:45 PM	Start	Qualification	Session	OO
September 7, 2021 3:48:46 PM	Start	Execution	Noise and Drift - Front FID: - Detector FID - L (Noise): <= 0.10 pA - L (Drift): <= 2.50 pA/hour	None
September 7, 2021 3:49:02 PM	End	Execution	Noise and Drift - Front FID: - Detector FID - L (Noise): <= 0.10 pA - L (Drift): <= 2.50 pA/hour	Run Count: 1

User Name: admin@crosslab.com
 Hostname: DESKTOP-ASH7V9S

System ID: CN10530014
 Print Date: September 8, 2021 11:48:04 AM

SPS_OQGCMS Transaction log:

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
September 7, 2021 4:17:12 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Back SSL, SQ: - Source: EI - Inert using Filament 1 - L: >= 180	Data file Path: K:\SPS\OQPV2021\SN\F1_002.D\DATA.MS
September 7, 2021 4:17:21 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Back SSL, SQ: - Source: EI - Inert using Filament 1 - L: >= 180	Data file Path: K:\SPS\OQPV2021\SN\F1_001.D\DATA.MS
September 7, 2021 4:17:38 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Back SSL, SQ: - Source: EI - Inert using Filament 1 - L: >= 180	Data file Path: K:\SPS\OQPV2021\SN\F1_001.D\DATA.MS
September 7, 2021 4:18:10 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Back SSL, SQ: - Source: EI - Inert using Filament 1 - L: >= 180	Data file Path: K:\SPS\OQPV2021\SN\F1_001.D\DATA.MS
September 7, 2021 4:20:24 PM	End	Execution	Run Count: 1	
September 7, 2021 4:20:42 PM	Audit	Test Unprocessed	Deviation filed for Run Count: 1	
September 7, 2021 4:20:42 PM	Start	Execution	None	
September 7, 2021 4:20:53 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Back SSL, SQ: - Source: EI - Inert using Filament 1 - L: >= 180	Data file Path: K:\SPS\OQPV2021\SN\F1_001.D\DATA.MS

Page 10 / 12

Page 10 / 12

User Name: admin@crosslab.com
Host Name: DESKTOP-ASH7V9S

System ID: CN10530014
Print Date: September 8, 2021 11:48:04 AM

SPS_OQGCMS Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
September 7, 2021 4:20:58 PM	End	Execution	Signal to Noise EI - Liquid Injection, Back SSL, SQ: - Source: EI - Inert using Filament 2 - L: >= 180	Run Count: 1
September 7, 2021 4:20:58 PM	Start	Execution	Signal to Noise EI - Liquid Injection, Back SSL, SQ: - Source: EI - Inert using Filament 1 - L: >= 180	None
September 7, 2021 4:20:58 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Back SSL, SQ: - Source: EI - Inert using Filament 1 - L: >= 180	Data file Path: K:\SPS\OQPV2021\SN\F1_001.D\DATA.MS
September 7, 2021 4:13:01 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Back SSL, SQ: - Source: EI - Inert using Filament 1 - L: >= 180	Data file Path: K:\SPS\OQPV2021\SN\F1_001.D\DATA.MS
September 7, 2021 4:13:28 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Back SSL, SQ: - Source: EI - Inert using Filament 1 - L: >= 180	Data file Path: K:\SPS\OQPV2021\SN\F1_001.D\DATA.MS
September 7, 2021 4:15:40 PM	Audit	AssClosed	Session	None
September 7, 2021 4:18:47 PM	Audit	AssReopened	Session	None
September 7, 2021 4:18:48 PM	Audit	SessionReloading	Session	None
September 7, 2021 4:18:53 PM	Start	Qualification	Session	OQ
September 7, 2021 4:18:53 PM	Start	Execution	Signal to Noise EI - Liquid Injection, Back SSL, SQ: - Source: EI - Inert using Filament 1 - L: >= 180	None

Page 9 / 12

Page 9 / 12

User Name: admin@ratnaswaji

Hostname: DESKTOP-4647V66

System ID: CN10630014

Print Date: September 8, 2021 11:48:04 AM

SPS_OGOCMS Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
September 7, 2021 4:27:56 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Back SSL, SC: - Source: EI - Inert using Filament 1 - L: >= 160	Data file Path : K:\SPSSN_F1_001.D\DATA.MS
September 7, 2021 4:30:00 PM	End	Execution	Signal to Noise EI - Liquid Injection, Back SSL, SC: - Source: EI - Inert using Filament 1 - L: >= 160	Run Count: 3
September 7, 2021 4:30:10 PM	Audit	AcqClosed	Session	None
September 8, 2021 11:31:05 AM	Audit	AcqRestarted	Session	None
September 8, 2021 11:31:07 AM	Audit	SessionReloaded	Session	None
September 8, 2021 11:31:15 AM	Start	Qualification	Session	OQ
September 8, 2021 11:31:32 AM	End	Qualification	Session	OQ
September 8, 2021 11:31:32 AM	Start	Reporting	Session	None
September 8, 2021 11:48:01 AM	Audit	Reporting	Session	Report Generated : Certificate

Page 12 / 12

Page 12 / 12

Date: September 8, 2021 11:48:04 AM
System ID: CN10630014

Page 23 / 23

User Name: admin@ratnaswaji

Hostname: DESKTOP-4647V66

SPS_OGOCMS Transaction log :

System ID: CN10630014

Print Date: September 8, 2021 11:46:08 AM


Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
September 7, 2021 4:21:20 PM	End	Execution	Signal to Noise EI - Liquid Injection, Back SSL, SC - Source: EI - Inert using Filament 1 - L: >= 160	Run Count : 2
September 7, 2021 4:21:33 PM	Audit	TestUnlocked	Signal to Noise EI - Liquid Injection, Back SSL, SC - Source: EI - Inert using Filament 1 - L: >= 160	Deviation filed for Run Count : 2
September 7, 2021 4:21:38 PM	Start	Execution	Signal to Noise EI - Liquid Injection, Back SSL, SC - Source: EI - Inert using Filament 1 - L: >= 160	None
September 7, 2021 4:21:45 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Back SSL, SC - Source: EI - Inert using Filament 1 - L: >= 160	Data file Path : K:\SPSSN_F1_001.D\DATA.MS
September 7, 2021 4:23:35 PM	Audit	AcqClosed	Session	None
September 7, 2021 4:23:37 PM	Audit	AcqRestarted	Session	None
September 7, 2021 4:23:58 PM	Audit	SessionReloaded	Session	None
September 7, 2021 4:24:03 PM	Start	Qualification	Session	CQ
September 7, 2021 4:24:03 PM	Start	Execution	Signal to Noise EI - Liquid Injection, Back SSL, SC - Source: EI - Inert using Filament 1 - L: >= 160	None
September 7, 2021 4:26:26 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Back SSL, SC - Source: EI - Inert using Filament 1 - L: >= 160	Data file Path : K:\SPSSN_F1_001.D\DATA.MS

Page 11 / 12

Page 11 / 12

Date: September 8, 2021 11:46:04 AM
System ID: CN10630014

Page 22 / 23



บริษัท ศูนย์วิเคราะห์น้ำ จำกัด

1/94 หมู่ที่ 5 ต.คานหาม อ.อุทัย จ.พระนครศรีอยุธยา 13210

โทรศัพท์ 035-800593, 081-9917119 โทรสาร 035-800594

Email : wac@wacthai.com Website : www.wacthai.com