
เอกสารสอบเทียบเครื่องมือที่ใช้ในการวิเคราะห์

CERTIFICATE OF CALIBRATION

Certificate No.: C0-2008004/21 Page 1 of total 4 pages

Customer
WATER ANALYSIS CENTER CO., LTD.
30/5 Soi Vipavadee 60, Vipavadee Rangsit Road,
Kwaeng Taladbangkhen, Khet Laksi, Bangkok 10210

Equipment
pH Meter
Manufacturer
METTLER TOLEDO
Model
B327527211
ID No.
SevenCompact
WWL 0068
Description
Range : 0 - 14 pH, Resolution : 0.01 pH

Environmental Conditions
Ambient Temperature: (20 ± 2) °C
Relative Humidity: (50 ± 10) %
Atmospheric Pressure: -

Calibration Location
Jayhawks Laboratory (CL&GL)

Received Date
20 August 2021

Calibration Date
20 August 2021

Date of Issue
23 August 2021



() () () () () () () () () ()
() () () () () () () () () ()
() () () () () () () () () ()
() () () () () () () () () ()

This calibration certificate shall not be reproduced other than in full except with the prior written approval of the Thai Heart Calibration Co., Ltd.

Certificate No.: C0-2008004/21 Page 2 of total 4 pages

Reference Method:

- The calibration method used was CP-178 based on an in-house method.
- This certificate can be traceable to the national standards, which is realized the shown measurement units according to the International System of Units (SI Units).

Reference Standard:

Type	pH Value	Lot No.	Due Date	Traceability
pH Standard Solution	4.01	081020	Feb. 1, 2022	NIMT
	7.01	020221	Dec. 25, 2021	
	10.00	091020	Jan. 19, 2022	

Type	Model	Serial No.	Certificate No.	Due Date	Traceability
Documenting Process Calibrator	753	3101007	10-0804001/21	Apr. 7, 2022	THC
Digital Thermometer with Sensor	1523 / 5622	1709138 / 4605984-005	10-1006001/21	Jun. 10, 2022	

Remark: This certificate is traceable to the International System of Unit (SI Unit) through:

- NIMT, National Institute of Metrology (Thailand).
- THC, Thai Heart Calibration Co., Ltd.

Measurement Results:

1. Function Simulated pH Meter

Standard Applied (mV)	Nominal Value (pH)	UUC Reading		Uncertainty (± mV)
		pH	mV	
177.48	4.00	4.01	177.4	0.060
0.00	7.00	7.00	0.0	0.060
-177.48	10.00	10.01	-177.4	0.060

UUC : Unit Under Calibration

Note : Adjust Curve to simulate pH (4,7,10)

Certificate No.: C0-2008004/21

Page 3 of total 4 pages

Measurement Results (Cont.):

2. Calibration of pH Electrode

pH Standard Solution (pH)	Measured Value		Uncertainty (± pH)
	(pH)	(mV)	
4.01	4.00	187.0	0.013
7.01	7.00	11.1	0.013
10.00	10.02	-161.6	0.013

Note : Adjust Curve to Buffer Solution pH (4.7,10)

Temperature stability of micro bath : $25 \pm 0.2^{\circ}\text{C}$

The above reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor $k = 2.00$, providing a level of confidence approximately 95%.

Certificate No.: C0-2008004/21

Page 4 of total 4 pages

Reference Method:

- The calibration method used was CP-096 based on an in-house method.

- The temperature scale used was an ITS-90.

- This certificate can be traceable to the national standards, which is realized the shown measurement units according to the International System of Units (SI Units).

Reference Standard Instruments:

Type	Model	Serial No.	Cert. No.	Due Date	Traceability
Thermometer Readout	1529-R	B7C853	20E3985	Nov. 9, 2021	TPA
Platinum Resistance Thermometer	5626	4853	C0A30046	Oct. 28, 2023	FLUKE
Liquid Bath	XORTS-40A	XO111019	IO-0306002/21	Jun. 3, 2023	THC

Remark: This certificate is traceable to the International System of Unit (SI Unit) through:

- TPA, Technology Promotion Association (Thailand-Japan).
- FLUKE, Fluke Corporation, U.S.A.
- THC, Thai Heart Calibration Co., Ltd.

Measurement Results:

(X) Without Adjustment

Dimension of probe : Diameter 4 mm. Sensor Type : RTD (PT100)

Immersion Depth (mm.)	Standard Reading ($^{\circ}\text{C}$)	UUC Reading ($^{\circ}\text{C}$)	Correction ($^{\circ}\text{C}$)	Uncertainty ($\pm ^{\circ}\text{C}$)
120	22.00	22.0	0.00	0.058
120	25.00	25.0	0.00	0.058
120	28.00	28.0	0.00	0.058

UUC : Unit Under Calibration

The above reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor $k = 2.00$, providing a level of confidence approximately 95%.

- End of Certificate -



THAI HEART CALIBRATION CO., LTD.

112/1 Moo 5, Phrak Sai, Muang, Samut Prakan 10280
Tel. 0-2394-2162, 0-2757-8435, 0-2757-8496 Fax. 0-2757-8507



CERTIFICATE OF CALIBRATION

Certificate No.: C0-2107005/21 Page 1 of total 2 pages

Customer
WATER ANALYSIS CENTER CO., LTD.
30/5 Soi Viphavadee 60, Viphavadee Rangsit Road,
Kwaeng Taladbangkhen, Khet Laksi, Bangkok 10210

Equipment Conductivity Meter
Manufacturer EUTECH **Model** CON 2700
Serial No. 2657889 **ID No.** WWL 0136
Description -

Environmental Conditions Ambient Temperature: $(20 \pm 2) ^\circ\text{C}$
Relative Humidity: $(50 \pm 10) \%$
Atmospheric Pressure: -

Calibration Location Jayhawks Laboratory (CL&GL)
Received Date 21 July 2021
Calibration Date 21 July 2021

Date of Issue 22 July 2021

() (Patiphan K.) () (Onnapa P.)
() (Pongsak H.) () (Niriphong K.)
() (Kanung C.) () (Nonthachai K.)
() (Pramong P.) () (Noppol P.)

This calibration certificate shall not be reproduced other than in full except with the prior written approval of the Thai Heart Calibration Co., Ltd.

FE-169

REV.02 02/24/21

Certificate No : AD2012-017-0001

Environment : Ambient Temperature : $(25 \pm 2) ^\circ\text{C}$

Relative Humidity : $(50 \pm 15)\% \text{RH}$

STD Reading (mg/l)	UUC Reading Before (mg/l)	UUC Reading After (mg/l)	Error (mg/l)	Uncertainty (\pm mg/l)
9.046	9.07	-	0.024	0.013

STD = Standard

UUC = Unit Under Calibration

Description of UUC : Range 0.00 to 60.00 mg/l
Resolution 0.01 mg/l

Measurement Standards Used & Traceability :

The International System of Units (SI) through

MIT Certificate No. L2002-756.L2002-757 for Data Logger (Lutron Temperature & Humid & Baro) Serial No. B014887, Due 28-Feb-21

MIT Certificate No. L2001-629 for Hi Accuracy Thermometer Serial No. 130508834, Due 07-Jan-21

End of Certificate

Page 2 of 2

929/929/ Sol Pattanakarn 30, Pattanakarn Rd., Suanluang, Bangkok 10250
Head Office : Tel. 02-319-9994 ext.1 Fax.02-319-4961 E-mail : ats@automation.co.th
Rayong Branch : 1/15 Huaypong Rd., A. Muang, Rayong 21150 Tel. 038-692-152 Fax. 038-692-345
Lamphun Branch : 122/5 M.4, T.Ban Klang, A.Muang, Lamphun 51000 Tel/Fax. 053-581-876
website : www.automation.co.th

SV 212001/2021

Cert. No. WAC-065
Page 1 of 2

CERTIFICATE OF CALIBRATION

Instrument : DO Meter
Model : DO-31P
Serial No. : 780065
Manufacturer : TOA-DKK
Measuring Range : 0.00 ~ 20.00 mg/l
Customer : Water Analysis Center Co.,Ltd.
1/94 Moo.5 T.Kanham, A.U.-Thai
Ayutthaya 13210 Thailand

Date Of Received : 03 / 12 / 2021
Date Of Calibration : 03 / 12 / 2021

Ambient Condition : Temperature 24 °C
Humidity 47 % RH

Calibrated By :

Approved By :

Date Of Issue : 03 / 12 / 2021

This Certificate may not be reproduced other than in full, except with the prior written approval of the head of the industrial instruments calibration center.

Instrument : DO Meter
Model : DO-31P
Serial No. : 780065

Cert. No. WAC-065
Page 2 of 2

Calibrate Procedure

- ☐ This instrument was calibrated by comparison with standard solution (PH/ORP)
- ☐ This instrument was calibrated by comparison with scattering plate value (Turbidity)
- ☐ This instrument was calibrated by comparison with conductivity (Conductivity)
- ☒ This instrument was calibrated by comparison with Sodium sulfite anhydrous (DO)

Condition of this result of calibration

1). Reference Standard Solution

Standard Lot No Batch Cert. No. Due Date

Sodium Sulfite Power 1.06657.0500 K52300357 - 31 Mar 2022

2). Traceability This certification is traceable to

- ☒ Merck KGaA 64271 Darmstadt
- ☐ DKK Corporation

Result Of Calibration

Standard Solution (mg/l) at 26.0°C	Before Adjust		After Adjust	
	Indicator	Error	Indicator	Error
Zero	0.00	+ 0.10	0.00	-
Span	7.99	+ 0.22	7.99	-

DO Electrode No. OE270AA(5) S/N 111F0029

Calibrated By

Certificate No.: MC 2107214

Page 2 of 3

The Reference Standard :

Description	Certificate No.	Serial No.	Due date
Data Acquisition/Switch Unit	MC 2009600	MY44095818	8 August 2021
With Thermocouple Type "T" ID. No.6/1 to 6/9			

This certificate is traceable to the international system of units maintained at:

- Master Calibration Co., Ltd. And Quality Reborn Co., Ltd.

1. Calibration Procedure:

This Instrument was calibration according to TLAS G-20 by comparison with calibrated thermocouple type T under no load condition. The Thermocouples were placed on nine points and located one thermocouple in each of the eight corners of the chamber and was away from the each wall of 5 cm to 10 cm. And placed the ninth thermocouple within 2.5 cm of the geometric center of the chamber.

Temperature Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.

Temperature Stability - one-half of the greatest maximum difference of measured temperatures at any one sensor.

Overall Variation - The Difference of the maximum and minimum measured temperatures throughout observation.




Figure 1: Sensor Installation Location

Checked by: *Thangum*

[MCF-Q-077 ; Rev.6 ; Date : 22/04/2021]

Certificate of Calibration

**TEMPERATURE
CONTROLLER ENCLOSURES**



Certificate No.: MC 2107214

Page 1 of 3



Customer	: Water Analysis Center Co., Ltd.	Received Date	: 13 July 2021
Reference Job No.	: 21-1565		
Description	: Refrigerator		
Manufacturer	: SANDEN/TERCOOL	Model	: SEC-1500SBD
Serial No.	: SEC1500201A-0708-00304	ID. No.	: WWL0038
Marking	: Additionally for the purpose of identification by this laboratory a label marked with this certificate number (MC 2107214) has been attached to the case.		
Method	: In-House calibration procedure MWI-T-033 this method is reference to TLAS G-20 "Temperature Controlled Enclosures".		
Location of Calibration	: Water Analysis Center Co., Ltd. ; Laboratory.		
Environmental Conditions	: Ambient Temperature : (26.3) °C		
	Relative Humidity : (56.4 to 59.3) %		
Date of Calibration	: 13 July 2021	Date of Issue	: 14 July 2021

Checked by :

Approved by :

The uncertainties are for a confidence probability of approximately 95%

This certificate is issued in accordance with the conditions of accreditation granted by the National Standardization Council of Thailand-Office of the National Standardization Council that has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of Master Calibration Co., Ltd.

[MCF-Q-077 ; Rev.6 ; Date : 22/04/2021]

Certificate No.: MC 2107214

Page 3 of 3

2. Result of calibration :

Temperature Measurement Accuracy Test

Indicating Temperature (°C)	Measured Temperature (°C) at Spread Locations								Uncertainty (±°C)
	#1	#2	#3	#4	#5	#6	#7	#8	Ref. #9
2.6	4.0	4.0	4.1	4.0	3.9	3.8	3.7	3.8	3.4
									1.2

Chamber Characterization Result

Controller Temperature (°C)	Indicating Temperature (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
2.0	2.6	2.7	1.4	5.8

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95 %.

This report will certify of the calibrated equipment only.

End of Certificate

Checked by: *Thungorn*

[MCF-Q-077 ; Rev.6 ; Date : 22/04/2021]

TEMPERATURE CONTROLLER ENCLOSURES

Report No. : MC 2103787

Page 1 of 3



Customer : Water Analysis Center Co., Ltd.
1/94 Moo 5, T.Kantham, A.U-Thai, Ayutthaya 13210.

Reference Job No. : 21-0710 Received Date : 25 March 2021

Description : Oven

Manufacturer : Memmert Model : UF260

Serial No. : B620.0814 ID. No. : N/A

Marking : Additionally for the purpose of identification by this laboratory a label marked with this report number (MC 2103787) has been attached to the case.

Method : In-House calibration procedure MWI-T-033 this method is reference to TLAS G-20 "Temperature Controlled Enclosures".

Location of Calibration : Water Analysis Center Co., Ltd. ; Laboratory.

Environmental Conditions : Ambient Temperature : (31.8 to 35.3) °C
Relative Humidity : (44.7 to 55.9) %

Date of Calibration : 25 March 2021 Date of Issue : 26 March 2021

Checked by:

Approved by:

The uncertainties are for a confidence probability of approximately 95%

This certificate is issued in accordance with the condition of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full expect with the prior written approval of the issuing laboratory.

[MCF-Q-077 ; Rev.5 ; Date : 15/07/2014]

Continuation of Report No. : MC 2103787

Page 2 of 3

The Reference Standard :

Description	Report No.	Serial No.	Due date
Data Acquisition/Switch Unit With Thermocouple Type "T" ID. No.17/1 to 17/9	MC 2016027	MY41010916	10 January 2022

This certificate is traceable to the international system of units maintained at:

- Master Calibration Co., Ltd.

1. Calibration Procedure:

This Instrument was calibration according to TLAS G-20 by comparison with calibrated thermocouple type T under no load condition. The Thermocouples were placed on nine points and located one thermocouple in each of the eight corners of the chamber and was away from the each wall of 5 cm to 10 cm. And placed the ninth thermocouple within 2.5 cm of the geometric center of the chamber.

Temperature Uniformity - the maximum difference of measured temperatures at any sensors and the

measured temperature at the reference location which are observed at the same time or at as close an

observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.

Temperature Stability - one-half of the greatest maximum difference of measured temperatures at any one sensor.

Overall Variation - The Difference of the maximum and minimum measured temperatures throughout observation.




Figure 1: Sensor Installation Location

Checked by : *Thanasorn*

[MCF-Q-077 ; Rev.5 ; Date : 15/07/2014]

Continuation of Report No. : MC 2103787

Page 3 of 3

2. Result of calibration :

Temperature Measurement Accuracy Test

Indicating Temperature (°C)	Measured Temperature (°C) at Spread Locations									Uncertainty (±°C)
	#1	#2	#3	#4	#5	#6	#7	#8	Ref. #9	
104.0	103.7	103.7	103.7	103.9	104.2	104.3	104.3	104.3	104.0	0.67
180.0	179.1	179.1	179.0	179.2	180.4	180.5	180.6	180.6	180.2	0.99

Chamber Characterization Result

Controller Temperature (°C)	Indicating Temperature (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
104.0	104.0	0.27	0.44	1.0
180.0	180.0	0.29	1.31	1.9

3. Uncertainty of Measurement

The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95 %.

This report will certify of the calibrated equipment only.

End of Calibration Report

Checked by : *Thanasorn*

[MCF-Q-077 ; Rev.5 ; Date : 15/07/2014]

Continuation of Report No. : MC 2103787

The Reference Standard :

Description	Report No.	Serial No.	Due date
Data Acquisition/Switch Unit	MC 2016027	MY41010916	10 January 2022
With Thermocouple Type " T " ID. No.171 to 17/9			

This certificate is traceable to the international system of units maintained at:

- Master Calibration Co., Ltd.

1. Calibration Procedure:

This Instrument was calibration according to TLAS G-20 by comparison with calibrated thermocouple type T under no load condition. The Thermocouples were placed on nine points and located one thermocouple in each of the eight corners of the chamber and was away from the each wall of 5 cm to 10 cm. And placed the ninth thermocouple within 2.5 cm of the geometric center of the chamber.

Temperature Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.

Temperature Stability - one-half of the greatest maximum difference of measured temperatures at any one sensor.

Overall Variation - The Difference of the maximum and minimum measured temperatures throughout observation.





Figure 1 : Sensor Installation Location

Checked by : *Thangon*

Calibration Report

**TEMPERATURE
CONTROLLER ENCLOSURES**



Report No. : MC 2103787

Page 1 of 3

Customer : Water Analysis Center Co., Ltd.
1/94 Moo 5, T.Kantham, A.U.-Thai, Ayutthaya 13210.

Reference Job No. : 21-0710 Received Date : 25 March 2021
Description : Oven
Manufacturer : Memmert Model : UF260
Serial No. : B620.0814 ID. No. : N/A
Marking : Additionally for the purpose of identification by this laboratory a label marked with this report number (MC 2103787) has been attached to the case.

Method : In-House calibration procedure MWI-T-033 this method is reference to

TLAS G-20 "Temperature Controlled Enclosures".

Location of Calibration : Water Analysis Center Co., Ltd. ; Laboratory.

Environmental Conditions : Ambient Temperature : (31.8 to 35.3) °C

Relative Humidity : (44.7 to 55.9) %

Date of Calibration : 25 March 2021 Date of Issue : 26 March 2021

Checked by :

Approved by :

The uncertainties are for a confidence probability of approximately 95%

This certificate is issued in accordance with the condition of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full expect with the prior written approval of the issuing laboratory.

Continuation of Report No. : MC 2103787

Page 3 of 3

2. Result of calibration :

Temperature Measurement Accuracy Test

Indicating Temperature (°C)	Measured Temperature (°C) at Spread Locations									Uncertainty (±°C)
	#1	#2	#3	#4	#5	#6	#7	#8	Ref. #9	
104.0	103.7	103.7	103.7	103.9	104.2	104.3	104.3	104.3	104.0	0.67
180.0	179.1	179.1	179.0	179.2	180.4	180.5	180.6	180.6	180.2	0.99

Chamber Characterization Result

Controller Temperature (°C)	Indicating Temperature (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
104.0	104.0	0.27	0.44	1.0
180.0	180.0	0.29	1.31	1.9

3. Uncertainty of Measurement


The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95 %

This report will certify of the calibrated equipment only.

End of Calibration Report

Checked by : *Thanagorn*

[MCF-Q-077 ; Rev.5 ; Date : 15/07/2014]



Certificate of Calibration

Equipment: Balance
Model: BL210S
Serial No. (or ID.): 15808131 (WWL 0022)
Manufacturer: Sartorius
Condition: In condition

Certificate No.: C01211841
Issued Date: 24 June 2021
Job No.: KSPR2107969
Page: 1 of 2

Customer: Water Analysis Center Co., Ltd.
1/94 Moo 5, Rojana Industrial Park, Rojana Road,
Tambol Kanham, Amphur U-Thai, Ayutthaya 13210 Thailand

Environment Condition: Temperature 27 °C ± 0.3 °C
Humidity 40 %RH ± 1.7 %RH

Calibration Place: Water Analysis Center Co., Ltd. (ห้องเครื่องชั่ง)
1/94 Moo 5, Rojana Industrial Park, Rojana Road,
Tambol Kanham, Amphur U-Thai, Ayutthaya 13210 Thailand

Calibration By: Mr. Phakapol Dommin

Calibration Date: 10 June 2021

The Method used: In house method, SPC-WI-47, base on UKAS Lab 14

Traceability: This certificate is traceable to the SI Units maintained by National Institute of Metrology (NIMT), Thailand through SPC RT Co., Ltd. Certificate No. C02210017

SERT
บริษัท เอสอาร์ที แอร์ที จำกัด
SPC RT Co., Ltd.

This certificate is issued the units of measurement according to the International System of Units (SI), the international or national standard or other recognized national standard laboratories.
The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor ($k=2$) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).
These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of SPC RT Co., Ltd.

Certificate No.: C01211841

Page: 2 of 2

Calibration Results:

Without Adjustment

Eccentric Error: Weight to be 1/4 or 1/3 of Maximum capacity, taken from the center of the pan as a zero reference.

Nominal Test Value (g)	Reference Points (g)				
	A	B	C	D	E
-	0.0000	-0.0001	-0.0001	-0.0001	0.0000

Repeatability: Determination of the standard deviation of weighing balance., Readability 0.0001 (g)

Nominal test value (g)	Standard Deviation
20	0.00004
200	0.00005

Departure of indication from nominal value., Readability 0.0001 (g)

Nominal Value (g)	Conventional Mass (g)	Displayed Value (g)	Correction of Balance (g)	Uncertainty (g)	k
1	1.00001	1.0000	0.0000	0.00010	2.03
2	2.00002	2.0000	0.0000	0.00010	2.03
5	4.99999	5.0000	0.0000	0.00010	2.03
10	10.00000	10.0000	0.0000	0.00011	2.02
20	19.99999	20.0000	0.0000	0.00011	2.02
50	49.99997	50.0000	0.0000	0.00012	2.01
70	69.99996	70.0000	0.0000	0.00015	2.00
100	100.00000	100.0000	0.0000	0.00017	2.00
120	119.99999	120.0001	-0.0001	0.00021	2.00
150	149.99997	150.0000	0.0000	0.00023	2.00
200	199.99990	200.0003	-0.0004	0.00029	2.00

The End of Certificate