

IC-THERMO

Serial No. : 20053176




This certificate is to verify that instrument below are calibrated

by Archemica Lab Co., Ltd.

Integriön	S/N: 20053176
AS-DV	S/N: 2008880131

For

Esternthai Consulting 1992 CO., Ltd.



Operator Signature: B. E. D. A. Date: Oct 19, 2021

(Mr. Chaivong Wongphatcharaporn)

Applications Chemist

22
COPY

Primary Flow Calibrator

Serial No. : 110619

SOUND LEVEL CALIBRATOR

MODEL : NC-75

SERIAL No. : 34802645

Request No. 21-65/0018

MTC No. EEL. BP. 24/1064

CALIBRATION CERTIFICATE

Submitted by : Eastern Thai Consulting 1992 Co.,Ltd.
Address : 683 Moo 11 Sukaphibal 8 Rd., Nongkham, Sriracha, Chonburi 20230.
Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.
: Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

Instrument Calibrated :

Description : Sound Calibrator
Manufacturer : Rion
Model : NC-75
Serial No. : 34802645
Ambient Environment
Temperature : (23 ± 3) °C
Relative Humidity : (50 ± 15) %
Ambient Pressure : (101.325 ± 1.500) kPa

Standards used :

1. Digital Function Synthesizer NF Electronic DF-193A S/N 122037.
2. Measuring Amplifier Bruel&Kjaer 2636 S/N 1537484.
3. Programmable Attenuator Tamagawa TPA-303A S/N OF 2214.
4. Digital Multimeter Agilent 34401A S/N MY44005560.
5. Pressure Transmitter Vaisala PTB202AD S/N T0650001.
6. Audio Analyzer Keithley 2015-P S/N 4106495.
7. Condenser Microphone B&K 4180 S/N 2889871.

Calibration Procedure: CP-102-04 based on IEC 60942-2003; The sound pressure level generated by sound calibrator under test shall be measured by standard microphone using an insert voltage technique.

This instrument has been calibrated against standards maintained at Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

Date of Receipt : 11 Oct. 2021

Date of Calibration : 21 Oct. 2021

The results relate only to the items tested/calibrated or value assigned.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Office : Office
Office/Laboratory : Office
196 Phahonyothin Road, Chituchak, Bangkok 10900.
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road, Thailand
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2577 9000
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : mtc@tistr.or.th
E-mail : sumalee@tistr.or.th

Request No. 21-65/0018

MTC No. EEL. BP. 24/1064

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%.

Nominal Output of Unit Under Test = 94 dB re 20μPa at 1000 Hz

Acoustic Output in dB re 20μPa, Corrected to Reference Conditions: 101.325 kPa, 23.0 °C and 50 %RH.

1. Sound Pressure Level

Standard Microphone Type	Measured Sound Pressure Level (dB)	Deviated value (dB)	Uncertainty (dB)	Tolerance limit
1/2 inch B&K 4180	93.97	-0.03	± 0.10	±0.40 dB

2. Frequency

Standard Microphone Type	Measured Frequency (Hz)	Deviated value (Hz)	Uncertainty (Hz)	Tolerance limit
1/2 inch B&K 4180	1000.0	0.0	± 1.5	±1.0%

3. Total Distortion

Standard Microphone Type	Measured Total Distortion (%)	Uncertainty (%)	Tolerance limit
1/2 inch B&K 4180	0.50	± 0.50	±3.0%

Note : 1. No adjustment.

2. The calibrator pressure correction was not included.

3. The microphone volume correction was not included.

Calibrated by :

(Mr. Weerachai Deechaiyae)

Approved by :

(Mr. Prawate Khaiyapa)
Acting Director

Electrical and Electronic Standards Laboratory
Industrial Metrology and Testing Service Centre

Date of Calibration : 21 Oct. 2021

Date of Issue : 26 Oct. 2021

End of Certificate

Ref.: 2011264101104187003

The results relate only to the items tested/calibrated or value assigned.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Office : Office
Office/Laboratory : Office
196 Phahonyothin Road, Chituchak, Bangkok 10900.
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road, Thailand
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2577 9000
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : mtc@tistr.or.th
E-mail : sumalee@tistr.or.th

SOUND LEVEL METER

MODEL : NL-21

SERIAL No. : 01209912

Request No. 21-65/0018

MTC No. EEL. BP. 22/1064

CALIBRATION CERTIFICATE

Submitted by : Eastern Thai Consulting 1992 Co., Ltd.

Address : 683 Moo 11 Sukaphibal 8 Rd., Nongkham, Sriracha, Chonburi 20230.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.

Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., A.Muang, Samutprakan 10280.

Instrument Calibrated :

Description : Sound Level Meter

Manufacturer : Rion

Model : NL-21

Serial No. : 01209912 (No.18)

Microphone : Type UC-52 No.157154

Preamplifier : Type NH-19 No.54250

Standards used :

1. Band Pass Filter Stanford Research Systems SR 650 S/N 28712.
2. Condenser Microphone Brüel&Kjær 4180 S/N 2889871.
3. Decade Attenuator Ando AL-205 S/N 00464602.
4. Function/Arbitrary Waveform Generator Agilent 33220A S/N MY44042668.
5. Digital Function Synthesizer NF Electronic Instruments DF-193A S/N 122037.
6. Digital Multimeter Fluke 8520A S/N 4985007.
7. Pistonphone Rion NC-72 S/N 00402446.
8. Measuring Amplifier Brüel&Kjær 2636 S/N 1537484.

Date of Receipt : 11 Oct. 2021

Date of Calibration : 20-27 Oct. 2021

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Lead Office

5 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang

angwat Pathumthani 12120, Thailand

el. (66) 0 2577 9000

ax. (66) 0 2577 9009

E-mail : rumpai@tistr.or.th Website:www.tistr.or.th

Office/Laboratory

Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,

Amphoe Muang, Changwat Samutprakan 10280, Thailand

Tel. (66) 0 2323 1672-80 ext. 115, 116

Fax. (66) 0 2323 9165

E-mail : mtc@tistr.or.th

Office

196 Phahonyothin Road, Chatuchak, Bangkok 10900,

Thailand

Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217

Fax. (66) 0 2579 8592

E-mail : sumalee@tistr.or.th

FM.BLMTC.002 Rev.4

Request No. 21-65/0018

MTC No. EEL. BP. 22/1064

9. Power Amplifier Brüel&Kjær 2706 S/N 1517650.

10. Speaker Tannoy Limited, Great Britain British Patent No. 215300.

11. Digital Multimeter Agilent 34401A S/N MY44005560.

12. Programmable Attenuator Tamagawa TPA-303A S/N 2212.

Calibration Procedure :

This instrument was calibrated by using calibration procedures no CP-102-02 and CP-102-03, which were based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). These calibration procedures were related to the electrical and acoustic signal tests. The electrical signal test was carried out with the direct measurement method. The acoustic signal test was performed in an anechoic room with the comparison measurement method.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%.

Date of Calibration : 20-27 Oct. 2021

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Lead Office

5 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang

angwat Pathumthani 12120, Thailand

el. (66) 0 2577 9000

ax. (66) 0 2577 9009

E-mail : rumpai@tistr.or.th Website:www.tistr.or.th

Office/Laboratory

Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,

Amphoe Muang, Changwat Samutprakan 10280, Thailand

Tel. (66) 0 2323 1672-80 ext. 115, 116

Fax. (66) 0 2323 9165

E-mail : mtc@tistr.or.th

Office

196 Phahonyothin Road, Chatuchak, Bangkok 10900,

Thailand

Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217

Fax. (66) 0 2579 8592

E-mail : sumalee@tistr.or.th

FM.BLMTC.002 Rev.4

COPY 2 / 8

1. Absolute Sensitivity

Reference Acoustic Signal (dB)	Unit Under Test				Tolerance Limit Class 2 (±dB)
	Measured Value (dB)		Deviation (dB)	Uncertainty (±dB)	
	Before adjust	After adjust			
113.94	114.4	113.9	0.0	0.30	1.4

Note: The external calibration adjustment was firstly performed. The internal calibration adjustment was then completed at the display of 117.8 dB.

2. Self-generated noise

2.1 Normal test

Measured value (dB)	Uncertainty (±dB)
16.8	0.10

2.2 The microphone of the sound level meter was replaced by electrical signal input device

Frequency Weighting	Measured Value (dB)	Uncertainty (±dB)
A-Weighting	15.9	0.10
C-Weighting	22.7	0.10
Flat	27.5	0.10

Date of Calibration : 20-27 Oct. 2021

The results relate only to the items tested/calibrated or value assigned.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Lead Office
5 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Bangkok Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax (66) 0 2579 8592
E-mail : sunalee@tistr.or.th

3. Acoustical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)	Flat (dB)		
125	0.1	0.2	0.1	0.40	2.0
1 000	-0.4	-0.3	-0.3	0.40	1.4
4 000	0.8	0.7	0.7	0.40	3.6

4. Electrical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)	Flat (dB)		
63	0.0	-0.1	-0.1	0.20	2.5
125	-0.1	0.0	0.0	0.20	2.0
250	0.0	-0.1	-0.1	0.20	1.9
500	0.0	0.0	0.0	0.20	1.9
1 000	0.0	0.0	0.0	0.20	1.4
2 000	0.1	0.1	0.1	0.20	2.6
4 000	0.0	0.1	0.1	0.20	3.6
8 000	0.2	0.2	0.1	0.20	5.6

Date of Calibration : 20-27 Oct. 2021

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Lead Office
5 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Bangkok Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax (66) 0 2579 8592
E-mail : sunalee@tistr.or.th

COPY 4 / 8

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
A-weighting	94.0	0.0	0.20	0.4
C-weighting	94.0	0.0	0.20	0.4
Flat	94.0	0.0	0.20	0.4

5.2 Time weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
Fast	94.0	0.0	0.20	0.3
Slow	94.0	0.0	0.20	0.3
Leq	94.0	0.0	0.20	0.3

6. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
125	125.0	0.0	0.30	1.4
124	124.0	0.0	0.30	1.4
123	123.0	0.0	0.30	1.4
122	122.0	0.0	0.30	1.4
121	121.0	0.0	0.30	1.4
120	120.0	0.0	0.30	1.4
119	119.0	0.0	0.30	1.4
114	114.0	0.0	0.30	1.4
109	109.0	0.0	0.30	1.4

Date of Calibration : 20-27 Oct. 2021

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Lead Office : 5 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpal@tistr.or.th Website: www.tistr.or.th

Office/Laboratory : Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road, Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2323 1672-80 ext. 115, 116
E-mail : mtc@tistr.or.th

Office : 196 Phahonyothin Road, Chatuchak Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FM.BL.MTC.002 Rev.4

6. Level linearity on the reference level range (continue)

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
104	104.0	0.0	0.30	1.4
99	99.0	0.0	0.30	1.4
94	94.0	0.0	0.30	1.4
89	89.0	0.0	0.30	1.4
84	84.1	0.1	0.30	1.4
79	79.0	0.0	0.30	1.4
74	74.1	0.1	0.30	1.4
69	69.1	0.1	0.30	1.4
64	64.0	0.0	0.30	1.4
59	59.1	0.1	0.30	1.4
54	54.0	0.0	0.30	1.4
49	49.0	0.0	0.30	1.4
44	44.1	0.1	0.30	1.4
39	38.9	-0.1	0.30	1.4
34	34.0	0.0	0.30	1.4
33	33.0	0.0	0.30	1.4
32	31.9	-0.1	0.30	1.4
31	31.0	0.0	0.30	1.4
30	30.0	0.0	0.30	1.4
29	28.9	-0.1	0.30	1.4
28	27.9	-0.1	0.30	1.4

Date of Calibration : 20-27 Oct. 2021

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Lead Office : 5 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpal@tistr.or.th Website: www.tistr.or.th

Office/Laboratory : Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road, Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office : 196 Phahonyothin Road, Chatuchak Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FM.BL.MTC.002 Rev.4

Request No. 21-65/0018

MTC No. EEL. BP. 22/1064

7. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
40-130	125	125.0	0.0	0.30	1.4
30-120	115	115.0	0.0	0.30	1.4
20-110	105	105.0	0.0	0.30	1.4
20-100	95	95.0	0.0	0.30	1.4
20-90	85	85.0	0.0	0.30	1.4
20-80	75	75.1	0.1	0.30	1.4

8. Tone burst response

Time Weighting	Toneburst Duration, Tb (ms)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (dB)
Fast	200	116.1	0.1	0.20	±1.3
	2	99.0	0.0	0.20	+1.3; -2.8
	0.25	89.9	-0.1	0.20	+1.8; -5.3
Slow	200	109.6	0.0	0.20	±1.3
	2	90.0	0.0	0.20	+1.3; -5.3
	200	110.0	0.0	0.20	±1.3
SEL	2	90.0	0.0	0.20	+1.3; -2.8
	0.25	80.9	-0.1	0.20	+1.8; -5.3

Date of Calibration : 20-27 Oct. 2021

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office : 5 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office : 196 Phahonyothin Road, Chatuchak, Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FM.BLMTC.002 Rev.4

Request No. 21-65/0018

MTC No. EEL. BP. 22/1064

9. Peak C sound level

Number of cycles in test signal	Anticipated value (dB)	Measured value (dB)	Deviated value (dB)	Uncertainty (±dB)	Tolerance limits Class 2 (±dB)
Complete cycle	125.4	125.0	-0.4	0.20	2.4
Positive half cycle	124.4	124.1	-0.3	0.20	1.4
Negative half cycle	124.4	124.1	-0.3	0.20	1.4

10. Overload indication

Measured value (dB)	Deviated value (dB)		Tolerance Limits Class 2 (±dB)	
	Positive one-half cycle	Negative one-half cycle	Uncertainty (±dB)	
135.6	135.6	0.0	0.30	1.8

Calibrated by :

Phanngsi
(Mr. Panya Phasingst)

Approved by :

T. M. 9
(Mr. Tawikiat Iamsanran)



Electrical and Electronic Standards Laboratory

Date of Calibration : 20-27 Oct. 2021

Date of Issue : 28 Oct. 2021

Industrial Metrology and Testing Service Centre

Ref : 2011264101104187001

End of Certificate

8/8

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office : 5 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office : 196 Phahonyothin Road, Chatuchak, Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FM.BLMTC.002 Rev.4

SOUND LEVEL METER

MODEL : NL-42

SERIAL No. : 01147299

Request No. 21-65/0018

MTC No. EEL. BP. 23/1064

CALIBRATION CERTIFICATE

Submitted by : Eastern Thai Consulting 1992 Co., Ltd.
Address : 683 Moo 11 Sukaphibal 8 Rd., Nongkham, Sriracha, Chonburi 20230.
Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., A.Muang, Samutprakan 10280.

Instrument Calibrated :
Description : Sound Level Meter
Manufacturer : Rion
Model : NL-42
Serial No. : 01147299 (No.26)
Microphone : Type UC-52 No.186065
Preamplifier : Type NH-24 No.46657

Standards used :

1. Band Pass Filter Stanford Research Systems SR 650 S/N 28712.
2. Condenser Microphone Brüel&Kjær 4180 S/N 2889871.
3. Decade Attenuator Ando AL-205 S/N 00464602.
4. Function/Arbitrary Waveform Generator Agilent 33220A S/N MY44042668.
5. Digital Function Synthesizer NF Electronic Instruments DF-193A S/N 122037.
6. Digital Multimeter Fluke 8520A S/N 4985007.
7. Pistonphone Rion NC-72 S/N 00402446.
8. Measuring Amplifier Brüel&Kjær 2636 S/N 1537484.

Date of Receipt : 11 Oct. 2021

Date of Calibration : 20-27 Oct. 2021

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Lead Office : 5 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office/Laboratory : Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road, Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office : 196 Phahonyothin Road, Chatuchak, Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FMBL.MTC.002 Rev.4

Request No. 21-65/0018

MTC No. EEL. BP. 23/1064

9. Power Amplifier Brüel&Kjær 2706 S/N 1517650.
10. Speaker Tannoy Limited, Great Britain British Patent No. 215300.
11. Digital Multimeter Agilent 34401A S/N MY44005560.
12. Programmable Attenuator Tamagawa TPA-303A S/N 2212.

Calibration Procedure :

This instrument was calibrated by using calibration procedures no CP-102-02 and CP-102-03, which were based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). These calibration procedures were related to the electrical and acoustic signal tests. The electrical signal test was carried out with the direct measurement method. The acoustic signal test was performed in an anechoic room with the comparison measurement method.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%.

Date of Calibration : 20-27 Oct. 2021

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Lead Office : 5 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office/Laboratory : Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road, Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office : 196 Phahonyothin Road, Chatuchak, Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FMBL.MTC.002 R4

1. Absolute Sensitivity

Reference Acoustic Signal (dB)	Unit Under Test			Tolerance Limit Class 2 (±dB)
	Measured Value (dB)	Deviation (dB)	Uncertainty (±dB)	
113.92	Before adjust 114.1	After adjust 113.9	0.0	0.30
				1.4

Note: The external calibration adjustment was firstly performed. The internal calibration adjustment was then completed at the display of 114.4 dB.

2. Self-generated noise

2.1 Normal test

Measured value (dB)	Uncertainty (±dB)
16.3	0.10

2.2 The microphone of the sound level meter was replaced by electrical signal input device

Frequency Weighting	Measured Value (dB)	Uncertainty (±dB)
A-Weighting	12.9	0.10
C-Weighting	18.0	0.10
Flat	24.1	0.10

Date of Calibration : 20-27 Oct. 2021

3 / 8

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

FMBL/MTC.002 Rev.4

3. Acoustical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)	Flat (dB)		
125	0.0	0.1	0.0	0.40	2.0
1 000	-0.1	-0.1	0.0	0.40	1.4
4 000	0.5	0.4	0.3	0.40	3.6

4. Electrical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)	Flat (dB)		
63	-0.1	-0.1	-0.1	0.20	2.5
125	-0.1	0.1	0.0	0.20	2.0
250	-0.1	0.0	0.0	0.20	1.9
500	0.0	0.0	0.0	0.20	1.9
1 000	0.0	0.0	0.0	0.20	1.4
2 000	0.1	0.1	0.0	0.20	2.6
4 000	0.1	0.0	0.0	0.20	3.6
8 000	0.1	0.1	0.0	0.20	5.6

Date of Calibration : 20-27 Oct. 2021

4 / 8

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

FMBL/MTC.002 Rev

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
A-weighting	94.0	0.0	0.20	0.4
C-weighting	94.0	0.0	0.20	0.4
Flat	94.0	0.0	0.20	0.4

5.2 Time weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
Fast	94.0	0.0	0.20	0.3
Slow	94.0	0.0	0.20	0.3
Leq	94.0	0.0	0.20	0.3

6. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
137	137.0	0.0	0.30	1.4
136	136.1	0.1	0.30	1.4
135	135.0	0.0	0.30	1.4
134	134.0	0.0	0.30	1.4
133	133.0	0.0	0.30	1.4
132	132.0	0.0	0.30	1.4
131	131.0	0.0	0.30	1.4

Date of Calibration : 20-27 Oct. 2021

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

6. Level linearity on the reference level range (cont.)

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
130	130.0	0.0	0.30	1.4
129	129.0	0.0	0.30	1.4
124	124.0	0.0	0.30	1.4
119	119.0	0.0	0.30	1.4
114	114.0	0.0	0.30	1.4
109	109.0	0.0	0.30	1.4
104	104.0	0.0	0.30	1.4
99	99.0	0.0	0.30	1.4
94	94.0	0.0	0.30	1.4
89	89.0	0.0	0.30	1.4
84	84.1	0.1	0.30	1.4
79	79.0	0.0	0.30	1.4
74	74.0	0.0	0.30	1.4
69	69.0	0.0	0.30	1.4
64	64.0	0.0	0.30	1.4
59	58.9	-0.1	0.30	1.4
54	54.0	0.0	0.30	1.4
49	49.0	0.0	0.30	1.4
44	44.0	0.0	0.30	1.4
39	39.0	0.0	0.30	1.4
34	33.9	-0.1	0.30	1.4
29	28.9	-0.1	0.30	1.4
28	28.0	0.0	0.30	1.4

Date of Calibration : 20-27 Oct. 2021

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

6. Level linearity on the reference level range (cont.)

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
27	27.0	0.0	0.30	1.4
26	26.0	0.0	0.30	1.4
25	25.0	0.0	0.30	1.4

7. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
20-130	125	125.0	0.0	0.30	1.4

8. Tone burst response

Time Weighting	Toneburst Duration, Tb (ms)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (dB)
Fast	200	126.0	0.0	0.20	±1.3
	2	108.9	-0.1	0.20	+1.3; -2.8
	0.25	99.9	-0.1	0.20	+1.8; -5.3
Slow	200	119.5	-0.1	0.20	±1.3
	2	99.9	-0.1	0.20	+1.3; -5.3
	0.25	120.0	0.0	0.20	±1.3
SEL	2	100.0	0.0	0.20	+1.3; -2.8
	0.25	90.8	-0.2	0.20	+1.8; -5.3

Date of Calibration : 20-27 Oct. 2021

7 / 8

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sri 1C, Bangpoo Industrial Estate, Sukhumvit Road, Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FM.BL.MTC.002 Rev.4

9. Peak C sound level

Number of cycles in test signal	Anticipated value (dB)	Measured value (dB)	Deviated value (dB)	Uncertainty (±dB)	Tolerance limits Class 2 (±dB)
Complete cycle	125.4	125.4	0.0	0.20	2.4
Positive half cycle	124.4	124.1	-0.3	0.20	1.4
Negative half cycle	124.4	124.1	-0.3	0.20	1.4

10. Overload indication

Measured value (dB)	Deviated value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
Positive one-half cycle	136.5	0.0	0.30
Negative one-half cycle	136.5	0.0	0.30

Calibrated by:

Panya Phasing Sri

(Mr. Panya Phasing Sri)

.....

(Mr. Tawikiat Iamsamman)

Date of Calibration : 20-27 Oct. 2021

Date of Issue : 28 Oct. 2021

Approved by:



Director

Electrical and Electronic Standards Laboratory

Industrial Metrology and Testing Service Centre

Ref : 2011264101104187002

End of Certificate

8 / 8

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Sri 1C, Bangpoo Industrial Estate, Sukhumvit Road, Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FM.BL.MTC.002 Rev.4

SOUND LEVEL METER

MODEL : NL-21

SERIAL No. : 00209072



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-65/0156

MTC No. EEL. BP. 29/1264

CALIBRATION CERTIFICATE

Submitted by : Eastern Thai Consulting 1992 Co., Ltd.

Address : 683 Moo 11 Sukaphibai 8 Rd., Nongkham, Sriracha, Chonburi 20230.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.

Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., A.Muang, Samutprakarn 10280.

Instrument Calibrated :

Description : Sound Level Meter
Manufacturer : Rion
Model : NL-21
Serial No. : 00209072 (No.16)
Microphone : Type UC-52 No.128251
Preamplifier : Type NH-21 No.32087

Ambient Environment

Temperature : $(23 \pm 3) ^\circ\text{C}$
Relative Humidity : $(50 \pm 15) \%$
Ambient Pressure : $(101.325 \pm 1.5) \text{ kPa}$

Standards used :

1. Band Pass Filter Wavetek 752A S/N 90010494.
2. Condenser Microphone Brüel&Kjær 4180 S/N 2633526.
3. Decade Attenuator Ando AL-205 S/N 00464602.
4. Function/Arbitrary Waveform Generator Agilent 33220A S/N MY44042668.
5. Digital Function Synthesizer NF Electronic Instruments DF-193A S/N 122037.
6. Digital Multimeter Fluke 8520A S/N 4985007.
7. Pistonphone Rion NC-72 S/N 00402446.
8. Measuring Amplifier Brüel&Kjær 2636 S/N 1537484.

Date of Receipt : 3 Dec. 2021

Date of Calibration : 7 Jan. 2022

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

lead Office
5 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Bangkok 10900
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : simalee@tistr.or.th

Office/Laboratory
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakarn 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : miracle@tistr.or.th

Office
196 Phahonyothin Road, Chaluchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : simalee@tistr.or.th

FM.BL.MTC.002 Rev.4



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-65/0156

MTC No. EEL. BP. 29/1264

9. Power Amplifier Brüel&Kjær 2706 S/N 1517650.

10. Speaker Tannoy Limited, Great Britain British Patent No. 215300.

11. Digital Multimeter Agilent 34401A S/N MY44005560.

12. Programmable Attenuator Tamagawa TPA-303A S/N 2212.

Calibration Procedure :

This instrument was calibrated by using calibration procedures no CP-102-02 and CP-102-03, which were based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). These calibration procedures were related to the electrical and acoustic signal tests. The electrical signal test was carried out with the direct measurement method. The acoustic signal test was performed in an anechoic room with the comparison measurement method.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%.

Date of Calibration : 7 Jan. 2022

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

lead Office
5 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Bangkok 10900
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : simalee@tistr.or.th

Office/Laboratory
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakarn 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : miracle@tistr.or.th

Office
196 Phahonyothin Road, Chaluchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : simalee@tistr.or.th

FM.BL.MTC.002 Rev.4

1. Absolute Sensitivity

Reference Acoustic Signal (dB)	Unit Under Test			Tolerance Limit Class 2 (±dB)
	Measured Value (dB)	Deviation (dB)	Uncertainty (±dB)	
113.98	Before adjust 113.9	After adjust 114.0	0.0	0.30
			0.0	1.4

Note: The external calibration adjustment was firstly performed. The internal calibration adjustment was then completed at the display of 116.5 dB.

2. Self-generated noise

2.1 Normal test

Measured value (dB)	Uncertainty (±dB)
19.5	0.10

2.2 The microphone of the sound level meter was replaced by electrical signal input device

Frequency Weighting	Measured Value (dB)	Uncertainty (±dB)
A-Weighting	15.5	0.80
C-Weighting	21.5	1.30
Flat	23.8	0.10

Date of Calibration : 7 Jan. 2022

COPY 3/8

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Lead Office
5 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Angsawattana District, Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : tistr@tistr.go.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592

FM.BLMTC.002 Rev.4

3. Acoustical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)	Flat (dB)		
125	-0.1	0.0	0.0	0.40	2.0
1 000	-0.1	-0.1	-0.1	0.40	1.4
4 000	-1.0	-0.9	-0.9	0.40	3.6

4. Electrical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)	Flat (dB)		
63	0.0	-0.1	-0.2	0.20	2.5
125	-0.1	-0.1	-0.2	0.20	2.0
250	0.0	-0.1	0.0	0.20	1.9
500	0.0	0.0	0.0	0.20	1.9
1 000	0.0	0.0	0.0	0.20	1.4
2 000	0.1	0.0	0.1	0.20	2.6
4 000	0.1	0.0	0.1	0.20	3.6
8 000	0.3	0.2	0.0	0.20	5.6

Date of Calibration : 7 Jan. 2022

COPY 4/8

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Lead Office
5 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Angsawattana District, Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : tistr@tistr.go.th

Office/Laboratory
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road, Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592

FM.BLMTC.002 Rev.

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (+dB)	Tolerance Limits Class 2 (+dB)
A-weighting	94.0	0.0	0.20	0.4
C-weighting	94.0	0.0	0.20	0.4
Flat	94.0	0.0	0.20	0.4

5.2 Time weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (+dB)	Tolerance Limits Class 2 (+dB)
Fast	94.0	0.0	0.20	0.3
Slow	94.0	0.0	0.20	0.3
Leq	94.0	0.0	0.20	0.3

6. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (+dB)	Tolerance Limits Class 2 (+dB)
125	125.0	0.0	0.30	1.4
124	124.0	0.0	0.30	1.4
123	123.0	0.0	0.30	1.4
122	122.0	0.0	0.30	1.4
121	121.0	0.0	0.30	1.4
120	120.0	0.0	0.30	1.4
119	119.0	0.0	0.30	1.4
114	114.0	0.0	0.30	1.4
109	109.0	0.0	0.30	1.4

Date of Calibration : 7 Jan. 2022

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009

Office
196 Phahonyothin Road, Chatchak, Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592

6. Level linearity on the reference level range (con.)

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (+dB)	Tolerance Limits Class 2 (+dB)
104	104.0	0.0	0.30	1.4
99	99.0	0.0	0.30	1.4
94	94.0	0.0	0.30	1.4
89	89.0	0.0	0.30	1.4
84	84.1	0.1	0.30	1.4
79	79.1	0.1	0.30	1.4
74	74.0	0.0	0.30	1.4
69	69.0	0.0	0.30	1.4
64	64.0	0.0	0.30	1.4
59	59.0	0.0	0.30	1.4
54	54.0	0.0	0.30	1.4
49	49.0	0.0	0.30	1.4
44	44.0	0.0	0.30	1.4
39	38.9	-0.1	0.30	1.4
34	33.9	-0.1	0.30	1.4
33	33.0	0.0	0.30	1.4
32	31.9	-0.1	0.30	1.4
31	31.0	0.0	0.30	1.4
30	29.9	-0.1	0.30	1.4
29	28.9	-0.1	0.30	1.4
28	27.8	-0.2	0.30	1.4

Date of Calibration : 7 Jan. 2022

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009

Office
196 Phahonyothin Road, Chatchak, Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592

7. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
40-130	125	125.0	0.0	0.30	1.4
30-120	115	115.0	0.0	0.30	1.4
20-110	105	104.9	-0.1	0.30	1.4
20-100	95	94.9	-0.1	0.30	1.4
20-90	85	85.0	0.0	0.30	1.4
20-80	75	75.0	0.0	0.30	1.4

8. Tone burst response

Time Weighting	Toneburst Duration, Tb (ms)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (dB)
Fast	200	116.0	0.0	0.20	±1.3
	2	98.9	-0.1	0.20	+1.3; -2.8
	0.25	89.8	-0.2	0.20	+1.8; -5.3
Slow	200	109.5	-0.1	0.20	±1.3
	2	89.9	-0.1	0.20	+1.3; -5.3
	0.25	80.8	-0.2	0.20	+1.8; -5.3

Date of Calibration : 7 Jan. 2022

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Lead Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

9. Peak C-sound level

Number of cycles in test signal	Anticipated value (dB)	Measured value (dB)	Deviated value (dB)	Uncertainty (±dB)	Tolerance limits Class 2 (±dB)
Complete cycle	125.4	125.0	-0.4	0.20	2.4
Positive half cycle	124.4	124.1	-0.3	0.20	1.4
Negative half cycle	124.4	124.1	-0.3	0.20	1.4

10. Overload indication

Measured value (dB)		Deviated value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
Positive one-half cycle	Negative one-half cycle			
135.7	135.6	0.1	0.30	1.8

Calibrated by :

Panya Phasingri
(Mr. Panya Phasingri)

Approved by :

Wittawat Supanich
(Mr. Wittawat Supanich)

Electrical and Electronic Standards Laboratory
Industrial Metrology and Testing Service Centre

Date of Calibration : 7 Jan. 2022

Date of Issue : 7 Jan. 2022

Ref : 2011264120305037002

End of Certificate

8 / 8

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Lead Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

SOUND LEVEL METER

MODEL : NL-21

SERIAL No. : 00443359



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-64/0485

MTC No. EEL. BP. 33/0464

CALIBRATION CERTIFICATE

Submitted by : Eastern Thai Consulting 1992 Co., Ltd.

Address : 683 Moo 11 Sukaphibal 8 Rd., Nongkham, Sriracha, Chonburi, 20230.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., A.Muang, Samutprakan 10280.

Instrument Calibrated :

Description : Sound Level Meter
Manufacturer : Rion
Model : NL-21
Serial No. : 00443359 (No.14)
Microphone : Type UC-52 No.153408
Preamplifier : Type NH-19 No.76773

Ambient Environment

Temperature : $(23 \pm 3) ^\circ\text{C}$
Relative Humidity : $(50 \pm 15) \%$
Ambient Pressure : $(101.325 \pm 1.5) \text{ kPa}$

Standards used :

1. Band Pass Filter Stanford Research Systems SR 650 S/N 28712.
2. Condenser Microphone Brüel&Kjær 4180 S/N 2889871.
3. Decade Attenuator Ando AL-205 S/N 00464602.
4. Function/Arbitrary Waveform Generator Agilent 33220A S/N MY44042668.
5. Digital Function Synthesizer NF Electronic Instruments DF-193A S/N 122037.
6. Digital Multimeter Fluke 8520A S/N 4985007.
7. Pistonphone Rion NC-72 S/N 00402446.
8. Measuring Amplifier Brüel&Kjær 2636 S/N 1537484.

Date of Receipt : 19 Apr. 2021

Date of Calibration : 5-13 May 2021

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website:www.tistr.or.th

Office/Laboratory
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FMBL-MTC.002 Rev.4



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-64/0485

MTC No. EEL. BP. 33/0464

9. Power Amplifier Brüel&Kjær 2706 S/N 1517650.
10. Speaker Tammy Limited, Great Britain British Patent No. 215300.
11. Digital Multimeter Agilent 34401A S/N MY44005560.
12. Programmable Attenuator Tamagawa TPA-303A S/N 2212.

Calibration Procedure :

This instrument was calibrated by using calibration procedures no CP-102-02 and CP-102-03, which were based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). These calibration procedures were related to the electrical and acoustic signal tests. The electrical signal test was carried out with the direct measurement method. The acoustic signal test was performed in an anechoic room with the comparison measurement method.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%.

Date of Calibration : 5-13 May 2021

2 / 8

COPY

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website:www.tistr.or.th

Office/Laboratory
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FMBL-MTC.002 Rev.4

1. Absolute Sensitivity

Reference Acoustic Signal (dB)	Unit Under Test				Tolerance Limit Class 2 (±dB)
	Measured Value (dB)		Deviation (dB)	Uncertainty (±dB)	
	Before adjust	After adjust			
	113.93	114.0			
					1.4

Note: The external calibration adjustment was firstly performed. The internal calibration adjustment was then completed at the display of 115.3 dB.

2. Self-generated noise

2.1 Normal test

Measured value (dB)	Uncertainty (±dB)
20.1	0.10

2.2 The microphone of the sound level meter was replaced by electrical signal input device

Frequency Weighting	Measured Value (dB)	Uncertainty (±dB)
A-Weighting	23.5	0.10
C-Weighting	30.1	0.10
Flat	35.5	0.10

Date of Calibration : 5-13 May 2021

3 / 8

COPY

The results relate only to the items tested/calibrated or value assigned.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpa@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

FMBL.MTC.002 Rev.4

3. Acoustical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)	Flat (dB)		
125	0.1	0.1	0.1	0.40	2.0
1 000	0.1	0.2	0.2	0.40	1.4
4 000	0.6	0.6	0.6	0.40	3.6

4. Electrical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)	F at (dB)		
63	-0.1	-0.3	-0.3	0.20	2.5
125	-0.2	0.0	-0.1	0.20	2.0
250	-0.2	-0.1	-0.1	0.20	1.9
500	-0.1	0.0	0.0	0.20	1.9
1 000	0.0	0.0	0.0	0.20	1.4
2 000	0.1	0.1	0.1	0.20	2.6
4 000	0.0	0.0	0.1	0.20	3.6
8 000	0.1	0.2	0.0	0.20	5.6

Date of Calibration : 5-13 May 2021

4 / 8

COPY

The results relate only to the items tested/calibrated or value assigned.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpa@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

FMBL.MTC.002 Rev.4



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-64/0485

MTC No. EEL. BP. 33/0464

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
A-weighting	94.0	0.0	0.20	0.4
C-weighting	94.0	0.0	0.20	0.4
Flat	94.0	0.0	0.20	0.4

5.2 Time weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
Fast	94.0	0.0	0.20	0.3
Slow	94.0	0.0	0.20	0.3
Leq	94.0	0.0	0.20	0.3

6. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
125	125.0	0.0	0.30	1.4
124	124.0	0.0	0.30	1.4
123	123.0	0.0	0.30	1.4
122	122.0	0.0	0.30	1.4
121	121.0	0.0	0.30	1.4
120	120.0	0.0	0.30	1.4
119	119.0	0.0	0.30	1.4
114	114.0	0.0	0.30	1.4
109	109.0	0.0	0.30	1.4

Date of Calibration : 5-13 May 2021

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FMBL-MTC.002 Rev.4



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-64/0485

MTC No. EEL. BP. 33/0464

6. Level linearity on the reference level range (con.)

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
104	104.0	0.0	0.30	1.4
99	99.0	0.0	0.30	1.4
94	94.0	0.0	0.30	1.4
89	89.0	0.0	0.30	1.4
84	84.1	0.1	0.30	1.4
79	79.1	0.1	0.30	1.4
74	74.0	0.0	0.30	1.4
69	69.0	0.0	0.30	1.4
64	64.0	0.0	0.30	1.4
59	59.0	0.0	0.30	1.4
54	54.0	0.0	0.30	1.4
49	49.0	0.0	0.30	1.4
44	44.1	0.1	0.30	1.4
39	39.0	0.0	0.30	1.4
34	34.2	0.2	0.30	1.4
33	33.4	0.4	0.30	1.4
32	32.4	0.4	0.30	1.4
31	31.5	0.5	0.30	1.4
30	30.6	0.6	0.30	1.4
29	29.9	0.9	0.30	1.4
28	29.1	1.1	0.30	1.4

Date of Calibration : 5-13 May 2021

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang, Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website: www.tistr.or.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FMBL-MTC.002 Rev.4

COPY

6/8

7. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
40-130	125	124.9	-0.1	0.30	1.4
30-120	115	115.0	0.0	0.30	1.4
20-110	105	105.0	0.0	0.30	1.4
20-100	95	95.0	0.0	0.30	1.4
20-90	85	85.1	0.1	0.30	1.4
20-80	75	75.1	0.1	0.30	1.4

8. Tone burst response

Time Weighting	Toneburst Duration, Tb (ms)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (dB)
Fast	200	116.0	0.0	0.20	±1.3
	2	99.0	0.0	0.20	+1.3; -2.8
	0.25	89.9	-0.1	0.20	+1.8; -5.3
Slow	200	109.6	0.0	0.20	±1.3
	2	90.0	0.0	0.20	+1.3; -5.3
	200	110.0	0.0	0.20	±1.3
SEL	2	90.0	0.0	0.20	+1.3; -2.8
	0.25	80.9	-0.1	0.20	+1.8; -5.3

Date of Calibration : 5-13 May 2021

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Klong Ha, Amphoe Klong Luang, Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpak@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road, Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

FM.BLMTC.002 Rev.4

9. Peak C sound level

Number of cycles in test signal	Anticipated value (dB)	Measured value (dB)	Deviated value (dB)	Uncertainty (±dB)	Tolerance limits Class 2 (±dB)
Complete cycle	125.4	125.0	-0.4	0.20	2.4
Positive half cycle	124.4	124.1	-0.3	0.20	1.4
Negative half cycle	124.4	124.1	-0.3	0.20	1.4

10. Overload indication

Measured value (dB)		Deviated value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
Positive one-half cycle	Negative one-half cycle	0.2	0.30	1.8
135.7	135.5			

Calibrated by :

Komkrit Laohasiri
(Mr. Komkrit Laohasiri)

Approved by :

(Mr. Prawit Klasya)
Acting Director
Electrical and Electronic Standards Laboratory
Industrial Metrology and Testing Service Centre

Date of Calibration : 5-13 May 2021

Date of Issue : 14 May 2021

Ref : 2011264041901629002

End of Certificate

8 / 8

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office
35 Mu 3 Tambon Klong Ha, Amphoe Klong Luang, Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpak@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road, Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

FM.BLMTC.002 Rev.4

SOUND LEVEL METER

MODEL : 6236

SERIAL No. : NSMAC623690376



NSC-TISTR 1025
CALIBRATION 0037

THAI AND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-65/0274

MTC No. EEL. BP. 109/0165

CALIBRATION CERTIFICATE

Submitted by : Envilab Co., Ltd.
Address : 540, 540/1 Soi Bangkhae 7, Bangkhae, Bangkok 10160, Thailand.
Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., A.Muang, Samutprakan 10280.

Instrument Calibrated :
Description : Integrating Sound Level Meter
Manufacturer : ACO
Model : 6236
Serial No. : NSMAC623690376
Microphone : Type 7052 No.78593
Preamplifier : -
Standards used :

- Ambient Environment**
Temperature : $(23 \pm 3) ^\circ\text{C}$
Relative Humidity : $(50 \pm 15) \%$
Ambient Pressure : $(101.325 \pm 1.5) \text{ kPa}$
1. Band Pass Filter Stanford Research Systems SR 650 S/N 28712.
 2. Condenser Microphone Brüel&Kjær 4180 S/N 2889871.
 3. Decade Attenuator Ando AL-205 S/N 00464602.
 4. Function/Arbitrary Waveform Generator Agilent 33220A S/N MY44042668.
 5. Digital Function Synthesizer NF Electronic Instruments DF-193A S/N 122037.
 6. Digital Multimeter Fluke 8520A S/N 4985007.
 7. Pistonphone Rion NC-72 S/N 00402446.
 8. Measuring Amplifier Brüel&Kjær 2636 S/N 1537484.

Date of Receipt : 26 Jan. 2021

Date of Calibration : 8-9 Feb. 2022

The results relate only to the items tested/calibrated or value assigned
Advising the Report/Certificate and publicity at the results except in full are prohibited unless written permission is obtained from the governor of TISTR

2/8

Head Office
35 Mu 3 Tambon Khlong Luang, Amphoe Khlong Luang
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rump@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10950,
Thailand
Tel. (66) 0 2579 1121 30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FM.BL.MTC.002 Rev.4



NSC-TISTR 1025
CALIBRATION 0037

THAI AND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-65/0274

MTC No. EEL. BP. 109/0165

9. Power Amplifier Brüel&Kjær 2706 S/N 1517650.
10. Speaker Tannoy Limited, Great Britain British Patent No. 215300.
11. Digital Multimeter Agilent 34401A S/N MY44005560.
12. Programmable Attenuator Tamagawa TPA-303A S/N 2212.

Calibration Procedure :

This instrument was calibrated by using calibration procedures no CP-102-02 and CP-102-03, which were based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). These calibration procedures were related to the electrical and acoustic signal tests. The electrical signal test was carried out with the direct measurement method. The acoustic signal test was performed in an anechoic room with the comparison measurement method.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%.

Date of Calibration : 8-9 Feb. 2022

2/8

The results relate only to the items tested/calibrated or value assigned
Advising the Report/Certificate and publicity at the results except in full are prohibited unless written permission is obtained from the governor of TISTR

Head Office
35 Mu 3 Tambon Khlong Luang, Amphoe Khlong Luang
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rump@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10950,
Thailand
Tel. (66) 0 2579 1121 30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

FM.BL.MTC.002 Rev.4



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-65/0274 MTC No. EEL.BP. 109/0165

1. Absolute Sensitivity

Reference Acoustic Signal (dB)	Unit Under Test			Tolerance Limit Class 2 (±dB)
	Measured Value (dB)	Deviation (dB)	Uncertainty (±dB)	
113.97	Before adjust 114.5	After adjust 114.0	0.0	0.30
				1.4

Note: The external calibration adjustment was firstly performed. The internal calibration adjustment was then completed at the display of 113.7 dB.

2. Self-generated noise

2.1 Normal test

Measured value (dB)	Uncertainty (±dB)
19.9	0.10

2.2 The microphone of the sound level meter was replaced by electrical signal input device

Frequency	Measured Value (dB)	Uncertainty (±dB)
Weighting		
A-Weighting	18.6	0.10
C-Weighting	22.0	0.10
Flat	24.6	0.10

Date of Calibration : 8-9 Feb. 2022

3/8

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FMBL MTC.002 Rev.4

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax (66) 0 2577 9009
E-mail : tuncpa@tistr.or.th

Office/Laboratory
Sri 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakarn 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax (66) 0 2323 9165
E-mail : mtc@tistr.or.th



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-65/0274 MTC No. EEL.BP. 109/0165

3. Acoustical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Tolerance Limits
	A-weighting (dB)	C-weighting (dB)	Flat (dB)	
125	0.4	0.1	0.1	Class 2 (±dB) 2.0
1 000	-0.5	-0.5	-0.5	1.4
4 000	-0.9	-0.9	-0.9	3.6

4. Electrical signal test of frequency weightings

Frequency (Hz)	Deviation from response curve			Tolerance Limits
	A-weighting (dB)	C-weighting (dB)	Flat (dB)	
63	0.2	0.1	0.0	Class 2 (±dB) 2.5
125	0.1	0.1	0.0	2.0
250	0.1	0.0	0.0	1.9
500	0.1	0.0	0.0	1.9
1 000	0.0	0.0	0.0	1.4
2 000	-0.1	0.0	0.0	2.6
4 000	-0.5	-0.4	-0.1	3.6
8 000	-0.6	-0.6	-0.2	5.6

Date of Calibration : 8-9 Feb. 2022

4/8

The results relate only to the items tested/calibrated or value assigned. Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FMBL MTC.002 Rev.4

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax (66) 0 2577 9009
E-mail : tuncpa@tistr.or.th

Office/Laboratory
Sri 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakarn 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax (66) 0 2323 9165
E-mail : mtc@tistr.or.th

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (+dB)	Tolerance Limits Class 2 (±dB)
A-weighting	94.0	0.0	0.20	0.4
C-weighting	94.0	0.0	0.20	0.4
Flat	94.1	0.1	0.20	0.4

5.2 Time weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Uncertainty (+dB)	Tolerance Limits Class 2 (±dB)
Fast	94.0	0.0	0.20	0.3
Slow	94.0	0.0	0.20	0.3
Leq	94.0	0.0	0.20	0.3

6. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (+dB)	Tolerance Limits Class 2 (±dB)
122	122.1	0.1	0.30	1.4
121	121.1	0.1	0.30	1.4
120	120.1	0.1	0.30	1.4
119	119.0	0.0	0.30	1.4
114	114.0	0.0	0.30	1.4
109	108.9	-0.1	0.30	1.4

6. Level linearity on the reference level range (cont.)

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (+dB)	Tolerance Limits Class 2 (±dB)
104	103.9	-0.1	0.30	1.4
99	99.0	0.0	0.30	1.4
94	94.0	0.0	0.30	1.4
89	89.0	0.0	0.30	1.4
84	83.9	-0.1	0.30	1.4
79	79.0	0.0	0.30	1.4
74	74.1	0.1	0.30	1.4
69	69.1	0.1	0.30	1.4
64	64.0	0.0	0.30	1.4
59	59.0	0.0	0.30	1.4
54	54.0	0.0	0.30	1.4
49	49.0	0.0	0.30	1.4
44	44.0	0.0	0.30	1.4
39	39.0	0.0	0.30	1.4
34	34.2	0.2	0.30	1.4
33	33.4	0.4	0.30	1.4
32	32.4	0.4	0.30	1.4
31	31.6	0.6	0.30	1.4
30	30.6	0.6	0.30	1.4



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-65/0274 MTC No. EEL BP. 109/0165

7. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
40-130	125	125.0	0.0	0.30	1.4
30-120	115	115.0	0.0	0.30	1.4
20-110	105	105.0	0.0	0.30	1.4
20-100	95	95.0	0.0	0.30	1.4
20-90	85	85.0	0.0	0.30	1.4
20-80	75	74.9	-0.1	0.30	1.4

8. Tone burst response

Time Weighting	Toneburst Duration, T _b (ms)	Measured Value (dB)	Deviated Value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (dB)
Fast	200	115.7	-0.3	0.20	±1.3
	2	98.7	-0.3	0.20	+1.3; -2.8
	0.25	89.5	-0.5	0.20	+1.8; -5.3
Slow	200	109.5	-0.1	0.20	±1.3
	2	89.9	-0.1	0.20	+1.3; -5.3
	200	110.1	0.1	0.20	±1.3
SEL	2	90.1	0.1	0.20	+1.3; -2.8
	0.25	81.3	0.3	0.20	+1.8; -5.3

Date of Calibration : 8-9 Feb. 2022

7/8

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.4

Head Office
35 Mu 3 Tambon Kiluan Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax (66) 0 2577 9009
E-mail : rumpak@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Soi 1C, Bangpoo Industrial Estate, Suburmit Road,
Amphoe Muang Chiangwat Samutprakarn 10280, Thailand
Tel. (66) 0 2323 1672 80 ext. 115, 116
Fax (66) 0 2323 9165
E-mail : mtc@tistr.or.th



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-65/0274 MTC No. EEL BP. 109/0165

9. Peak C sound level

Number of cycles in test signal	Anticipated value (dB)	Measured value (dB)	Deviated value (dB)	Uncertainty (±dB)	Tolerance limits Class 2 (±dB)
Complete cycle	125.4	125.7	0.3	0.20	2.4
Positive half cycle	124.4	124.2	-0.2	0.20	1.4
Negative half cycle	124.4	124.2	-0.2	0.20	1.4

10. Overload indication

Measured value (dB)		Deviated value (dB)	Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
Positive one-half cycle	133.2	0.0	0.30	1.8
Negative one-half cycle	133.2			

Calibrated by:

Pannasit Phasingst
(Mr. Pannasit Phasingst)

Approved by:

Prasit
(Mr. Prasit Prasit)
TISTR

Electrical and Electronic Standards Laboratory
Industrial Metrology and Testing Service Centre

Ref: 2011265013600351007

Date of Calibration : 8-9 Feb. 2022

Date of Issue : 10 Feb. 2022

The results relate only to the items tested/calibrated or value assigned.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.4

Head Office
35 Mu 3 Tambon Kiluan Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax (66) 0 2577 9009
E-mail : rumpak@tistr.or.th Website: www.tistr.or.th

Office/Laboratory
Soi 1C, Bangpoo Industrial Estate, Suburmit Road,
Amphoe Muang Chiangwat Samutprakarn 10280, Thailand
Tel. (66) 0 2323 1672 80 ext. 115, 116
Fax (66) 0 2323 9165
E-mail : mtc@tistr.or.th



7890 GC Preventive Maintenance Checklist – Standard

Agilent Preventive Maintenance provides factory recommended service for your analytical systems to assure reliable operation and the accuracy of your results. Delivered by highly-trained and certified service engineers using genuine Agilent parts and supplies, Agilent Preventive Maintenance provides everything you need to reduce unplanned downtime and keep your systems operating at their peak. For more information about Agilent Technologies GC Support please visit our web site using the following URL:

<http://www.agilent.com/en-us/products/gas-chromatography/gc-systems/7890b-gc#support>

Customer Information

- Customers should provide all necessary operating supplies upon request of the engineer.
- A customer representative should be available to the engineer while performing the preventive maintenance procedures.
- Any parts, not included in the Parts Lists section of this document, are not part of the recommended Preventive Maintenance service, nor are they included in the price of this service.
- If a system requires the use of additional or special procedures and/or parts for the instrument service, then these must be ordered separately and charged as a repair, which may incur additional costs.

Service Engineer's Responsibilities

- Only complete sections that relate to the system or module being serviced.
- Complete empty fields with the relevant information.
- Complete the relevant checkboxes in the checklist using a "X" or tick mark "✓" in the checkbox.
- Complete Not Applicable check boxes to indicate services not delivered, as needed.
- Complete the PM Service in the order of the tasks listed.
- Complete the Service Review section together with the customer.

Additional Instruction Notes

- Check for any active service notes for this unit. If there are any applicable "Safety" or "Modification Recommended" Service notes, plan to implement the changes on this unit before doing any qualification service.
- Do not implement firmware updates, unless you get approval from the customer and are sure that they are compatible with the instrument control software.



Certificate of Calibration

ICS-1000 : Cation & Anion (ID#057)

This certificate is to verify that instrument below are calibrated
by Archemica Lab Co.,Ltd.

ICS-1000 S/N : 04090295

RFC-30 S/N : 04100666

for

SECOT Co.,Ltd.



Operator Signature : _____ Date : Dec 16, 2021

(Mr.Channarong Khiao-un)

Test Engineer

System Information

Guidance

- ✓ Check this box if an instrument configuration report is attached instead of completing the table.

Instrument system name and ID	6C 7340 B
Instrument system site and location	1 N 57.2
List system component product numbers	List the serial numbers of each component
1. 63440 B	1. CN 1564347
2. 6 4513 A	2. CN 11350133
3. 6 4514 A	3. CN 13080006
4.	4.
5.	5.
6.	6.
7.	7.
8.	8.
9.	9.
10.	10.

Preparation

- ✓ Discuss any specific issues with the customer prior to starting.
- ✓ Review the instrument logbook.
- ✓ Save instrument control settings before starting the procedure.
- ✓ Perform general inspection of system for cleanliness
- ✓ Check for proper installation of safety-related parts, assemblies, sensors etc.
- ✓ Check for required firmware updates and verify with customers if they would like it installed.
- ✓ Before starting the following procedures, record the Detector Signal Output(s) in the results table. If the GC is turned OFF or in a service mode, comparing the detector outputs before and after the service is not possible.



Clean and inspect GC

- ✓ Unplug power cord from the power source.
- ✓ Open GC covers and vacuum/remove any dust/debris. Pay particular attention to cooling fans.
- ✓ Inspect internal connectors for proper contact and placement.
- ✓ Reconnect Power to the GC. Power the GC on and verify the power on self-test passed.
- ✓ Verify oven motor spins freely and turns on with the oven door closed; off when the door is opened.
- ✓ Verify operation of all other fans - the inlet and EPC cooling fans.
- ✓ Verify oven intake/outlet flap assembly is operating smoothly while heating and cooling the oven

Inlet and detector consumable replacement

- ✓ For the inlets installed, perform inlet maintenance as defined in the 7890 manual - "Maintaining Your GC" - for the inlet(s) installed.
- ✓ Replace the split vent trap cartridge filter on units with these inlets: Split/Splitless Capillary (SSL), Multi-Mode Inlet (MMI), Programmed Temperature Vaporizer (PTV), Volatiles Interface (VI).
- ✓ If the inlet system is used in Split Mode with viscous samples, inspect and clean the split vent tube on the inlet and flush or replace the tubing between the inlet and the split vent trap.
- ✓ If the GC includes a Flame Ionization Detector (FID), replace the jet. If the ignitor shows any buildup of sample or corrosion, replace the ignitor. Examine the FID collector and castle assemblies for contamination - clean as necessary.

Zero Sensors and Leak test

- ✓ Zero all pressure sensors per the procedure in the 7890 "Advanced User Guide".
- ✓ Perform inlet pressure decay test(s) as defined in the 7890 "Troubleshooting Manual". If the PM is done in preparation for an Operational Qualification, then the pressure decay test defined within that protocol can be used for the PM.
- ✓ Record if test passed or failed in the results table.

ALS Maintenance

- ☐ Section NOT applicable
- ✓ Check all cabling and configuration settings between GC tray, and injectors.
- ✓ Vacuum or removed any dust, especially around fans.
- ✓ Check operation of all fans.
- ✓ Check syringe for smooth plunger operation.
- ✓ Check for smooth operation of the needle support - clean if necessary
- ✓ Check for correct operation of syringe volume settings.



Restore Instrument

- ☒ Restore the normal operating conditions or customer method using the Keyboard or Data System.
- ☒ Purge the system with carrier flow for 15 minutes
- ☒ Bake out the system, then restore the normal operating conditions
- ☒ After equilibration, check and record the post PM detector signal output values.
- ☒ Results should be similar or lower than the detector outputs recorded prior to PM.
- ☒ Perform a chemical checkout. If this is a routine PM, inject the customer's sample using the ALS if applicable. This will act as a final checkout of both the ALS and the GC.

Guidance

If the PM Service is performed prior to a qualification service, then use the qualification procedure as a guide for final instrument set up and checkout.



Service Review

- ☒ Attach available reports/printouts of all tests to this documentation.
- ☒ Record the PM service activity in the customer's instrument records/logbook
- ☒ Update/reset instrument maintenance counters as appropriate
- ☒ Affix the PM sticker to the system or instrument logbook based on the customer's request.
- ☒ Complete the Service Review Comments section below if there are additional comments
- ☒ Review the service and any test results with the customer.
- ☐ If the Instrument firmware was updated, record the details of the change in the Service Engineer's Comments box below or if necessary, in the customer's IQ records.
- ☐ Please ask the customer if they would like to have Smart Alerts installed on their computer.

7890 GC Test Results Table

Detector Signal Outputs	Before PM service	After PM service
Front detector output	N/A	290.4
Back detector output	N/A	16.8
AUX detector output	N/A	N/A
Pressure decay test	Expected result	Actual result or N/A
Front inlet pressure decay test	Pass	Pass
Back inlet pressure decay test	Pass	Pass



7890 GC Parts List Table

The following kits are recommended for capillary and purged packed inlets. If this is a general PM and the customer has a preferred set of consumables, you may use the customer's consumables.

Part Description	Part Number	Model# where used	Quantity Consumed
SSL Capillary Inlet PM kit, Splitless	5188-6497	7890A/B	1
SSL Capillary Inlet PM kit, split	5188-6496	7890A/B	1
SSL Capillary Ultra Inert Inlet Gold Seal with Washer	5190-6144	7890A/B	-
SSL Capillary Ultra Inert Inlet Splitless Liner - Single taper with Glass Wool	5190-2293	7890A/B	-
SSL Capillary Ultra Inert Inlet Low Pressure Drop Split Liner - with Glass Wool	5190-2295	7890A/B	-
PP Inlet PM kit	5188-6498	7890A/B	-
Split vent trap PM kit, single cartridge (for MMI, PTV & VI)	5188-6495	7890A/B	-
MMI Cleaning Kit	G3510-60820	7890A/B	-
PTV Septumless Head Rebuild Kit	5182-9747	7890A/B	-
PTV Septumless Head Teflon Guide	5182-9748	7890A/B	-
Ignitor (glow plug) assembly with O-ring	19231-60680	7890A/B	1
FID Collector Rebuild/Cleaning Kit	G1531-67000	7890A/B	-
FID Collector Replacement Kit	G1531-67001	7890A/B	-
Standard .011-inch FID Jet for capillary FID base	G1531-80560*	7890A/B	-
High Temperature .018-inch FID Jet for capillary FID base	G1531-80620*	7890A/B	-
Standard .018-inch FID Jet for packed column with packed FID base	18710-20119*	7890A/B	-
Standard .011-inch FID Jet for capillary column with packed/adaptable FID base	19244-80560*	7890A/B	1
High Temperature .018-inch FID Jet for capillary column with packed/adaptable FID base	19244-80620*	7890A/B	-

* The jets (G1531-80560, G1531-80620, 18710-20119, 19244-80560 and 19244-80620) are recommended for 7890A/B PM. Please refer to the service note "COLUMNS/SUPPLIES-197A" for detailed information.



Service Engineer Comments (optional)

If there are any specific points you wish to note as part of performing the service or other items of interest for the customer, please write in this box.

N/A

Other Important Customer Web Links

- ☐ 7890 GC manual "Maintaining Your GC" - http://www.agilent.com/cs/library/usermanuals/public/33430-90052%207890B_Maintaining%20Guide.pdf
- ☐ Need to know more? - <http://www.agilent.com/crosslab/university/>
- ☐ Need supplies? - www.agilent.com/chem/supplies

Service Completion

Service request number 60454741

Date service completed 43 June 2021

Agilent signature

George T.

Customer signature

Marin Camacho

Document part number: G3430-90004