

ภาคผนวก ง

เอกสารรับรองการสอบเทียบเครื่องมือ



right solutions.
right partner.

รายการเครื่องมือที่ใช้ในการตรวจ / ทดสอบ

Sample Name	Parameter	Equipment Name	ID No.	Calibrated Date	Next Cal	Freq. Calibrate (Months)
Ambient	Total Suspended Particulate	High Volume	RYG_F50179	-	-	On site Calibration
Ambient	Total Suspended Particulate	High Volume	RYG_F50292	-	-	On site Calibration
Ambient	Total Suspended Particulate	High Volume	RYG_F50175	-	-	On site Calibration
Ambient	Total Suspended Particulate	High Volume	RYG_F50173	-	-	On site Calibration
Ambient	Total Suspended Particulate	High Volume	RYG_F50177	-	-	On site Calibration
Ambient	Total Suspended Particulate	Digital Balance	RYG_EN0001	23-Mar-22	23-Mar-23	12
Ambient	Particulate Matter (PM-10)	High Volume	RYG_F50398	-	-	On site Calibration
Ambient	Particulate Matter (PM-10)	High Volume	RYG_F50189	-	-	On site Calibration
Ambient	Particulate Matter (PM-10)	High Volume	RYG_F50184	-	-	On site Calibration
Ambient	Particulate Matter (PM-10)	High Volume	RYG_F50183	-	-	On site Calibration
Ambient	Particulate Matter (PM-10)	High Volume	RYG_F50187	-	-	On site Calibration
Ambient	Particulate Matter (PM-10)	High Volume	RYG_F50400	-	-	On site Calibration
Ambient	Particulate Matter (PM-10)	High Volume	RYG_F50185	-	-	On site Calibration
Ambient	Particulate Matter (PM-10)	Digital Balance	RYG_EN0001	23-Mar-22	23-Mar-23	12
Ambient	Arsenic	ICP-OES	RYG_F50174	13-Sep-21	12-Mar-23	18
Ambient	Arsenic	ICP-OES	RYG_F50174	13-Sep-21	12-Mar-23	18
Ambient	Lead	ICP-OES	RYG_F50174	13-Sep-21	12-Mar-23	18
Ambient	Lead	ICP-OES	RYG_F50174	13-Sep-21	12-Mar-23	18
Ambient	Mercury	CVAFS	RYG_F50174	13-Sep-21	12-Mar-23	18
Ambient	Mercury	CVAFS	RYG_F50174	13-Sep-21	12-Mar-23	18
Ambient	Nitrogen Dioxide	NO ₂ Analyzer	RYG_F50461	1-Jul-22	1-Jan-23	6
Ambient	Nitrogen Dioxide	NO ₂ Analyzer	RYG_F51064	1-Jul-22	1-Jan-23	6
Ambient	Nitrogen Dioxide	NO ₂ Analyzer	RYG_F50272	1-Jul-22	1-Jan-23	6
Ambient	Nitrogen Dioxide	NO ₂ Analyzer	RYG_F50252	1-Jul-22	1-Jan-23	6
Ambient	Nitrogen Dioxide	NO ₂ Analyzer	RYG_F50463	1-Jul-22	1-Jan-23	6
Ambient	Nitrogen Dioxide	NO ₂ Analyzer	RYG_F50469	1-Jul-22	1-Jan-23	6
Ambient	Nitrogen Dioxide	NO ₂ Analyzer	RYG_F50463	1-Jul-22	1-Jan-23	6
Ambient	Sulfur Dioxide	SO ₂ Analyzer	RYG_F50460	1-Jul-22	1-Jan-23	6
Ambient	Sulfur Dioxide	SO ₂ Analyzer	RYG_F50266	1-Jul-22	1-Jan-23	6
Ambient	Sulfur Dioxide	SO ₂ Analyzer	RYG_F50271	1-Jul-22	1-Jan-23	6
Ambient	Sulfur Dioxide	SO ₂ Analyzer	RYG_F50251	1-Jul-22	1-Jan-23	6
Ambient	Sulfur Dioxide	SO ₂ Analyzer	RYG_F50462	1-Jul-22	1-Jan-23	6
Ambient	Sulfur Dioxide	SO ₂ Analyzer	RYG_F50458	1-Jul-22	1-Jan-23	6
Ambient	Sulfur Dioxide	SO ₂ Analyzer	RYG_F50462	1-Jul-22	1-Jan-23	6
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_F50412	29-Jul-21	27-Jan-23	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_F50089	13-Jul-21	11-Jan-23	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_F50143	23-Feb-21	24-Aug-22	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_F50413	29-Jul-21	27-Jan-23	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_F50141	7-Jun-21	6-Sep-22	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_F50414	29-Jul-21	27-Jan-23	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_F50530	14-Jul-21	12-Jan-23	18
Stack	Arsenic	Console Control Unit	RYG_F50527	12-Jul-22	12-Jan-23	6
Stack	Arsenic	Console Control Unit	RYG_F50518	12-Jul-22	12-Jan-23	6
Stack	Arsenic	Flue gas Analyzer	RYG_F50465	19-Jan-22	19-Jan-23	12
Stack	Arsenic	Flue gas Analyzer	RYG_F50566	9-Dec-21	9-Dec-22	12
Stack	Arsenic	ICP-OES	RYG_EI0037	13-Sep-21	12-Mar-23	18
Stack	Cadmium	Console Control Unit	RYG_F50527	12-Jul-22	12-Jan-23	6
Stack	Cadmium	Console Control Unit	RYG_F50518	12-Jul-22	12-Jan-23	6
Stack	Cadmium	Flue gas Analyzer	RYG_F50465	19-Jan-22	19-Jan-23	12
Stack	Cadmium	Flue gas Analyzer	RYG_F50566	9-Dec-21	9-Dec-22	12
Stack	Cadmium	ICP-OES	RYG_EI0037	13-Sep-21	12-Mar-23	18
Stack	Lead	Console Control Unit	RYG_F50527	12-Jul-22	12-Jan-23	6
Stack	Lead	Console Control Unit	RYG_F50518	12-Jul-22	12-Jan-23	6
Stack	Lead	Flue gas Analyzer	RYG_F50465	19-Jan-22	19-Jan-23	12
Stack	Lead	Flue gas Analyzer	RYG_F50566	9-Dec-21	9-Dec-22	12
Stack	Lead	ICP-OES	RYG_EI0037	13-Sep-21	12-Mar-23	18

1

alsglobal.com



right solutions.
right partner.

รายการเครื่องมือที่ใช้ในการตรวจ / ทดสอบ

Sample Name	Parameter	Equipment Name	ID No.	Calibrated Date	Next Cal	Freq. Calibrate (Months)
Noise	Noise Dose, TWA	Noise Badge Reader	RYG_F50440	7-Sep-21	7-Sep-22	12
Noise	Noise Dose, TWA	Noise Badge Reader	RYG_F50210	7-Sep-21	7-Sep-22	12
Heat	Heat Stress	Heat Stress Monitor	RYG_F50231	3-Aug-22	3-Aug-23	12
Heat	Heat Stress	Heat Stress Monitor	RYG_F50230	3-Aug-22	3-Aug-23	12
Heat	Heat Stress	Heat Stress Monitor	RYG_F50231	3-Aug-22	3-Aug-23	12
Heat	Heat Stress	Heat Stress Monitor	RYG_F50356	16-Feb-22	16-Feb-23	12
Humance	Humance	Lux Meter	RYG_F50471	08-Apr-22	08-Apr-23	12
Rayong Lab	pH at 25 °C	pH Meter	RYG_EN0183	17-Mar-22	17-Mar-23	12
Rayong Lab	Sulfate	Spectrophotometer	RYG_EI0037	27-Sep-22	27-Mar-24	18
Rayong Lab	Sulfate	Chamber (Gold Room)	RYG_FI0184	22-Feb-22	22-Feb-23	12
Rayong Lab	Turbidity	Chamber (Gold Room)	RYG_FI0184	22-Feb-22	22-Feb-23	12
Rayong Lab	Dissolved Oxygen	Chamber (Gold Room)	RYG_EI0182	22-Feb-22	22-Feb-23	12
Rayong Lab	BOD	DO meter with sensor	RYG_EN0032	14-Feb-22	14-Aug-23	18
Rayong Lab	BOD	Incubator	RYG_EN0154	22-Apr-22	21-Oct-23	18
Rayong Lab	Total Suspended Solids	Electronic Balance	RYG_EN0002	23-Mar-22	23-Mar-23	12
Rayong Lab	Total Suspended Solids	Hot Air Oven	RYG_EN0010	20-Oct-22	20-Apr-24	18
Rayong Lab	Total Dissolved Solids 180°C	Electronic Balance	RYG_EN0002	23-Mar-22	23-Mar-23	12
Rayong Lab	Total Dissolved Solids 180°C	Hot Air Oven	RYG_EN0010	20-Oct-22	20-Apr-24	18
Rayong Lab	Conductivity	Conductivity meter	RYG_EN0029	23-Feb-22	24-Aug-23	18
Rayong Lab	Residual Chlorine	ORP Meter	RYG_EN0183	17-Mar-22	17-Mar-23	12
Rayong Lab	Salinity	Conductivity meter	RYG_EN0029	23-Feb-22	24-Aug-23	18
Rayong Lab	Oil & Grease	Electronic Balance	RYG_EN0002	23-Mar-22	23-Mar-23	12
Rayong Lab	Oil & Grease	Hot Air Oven	RYG_EN0010	20-Oct-22	20-Apr-24	18
Rayong Lab	Oil & Grease	Filter Bath	RYG_EN0061	20-Oct-22	20-Apr-24	18
Rayong Lab	Temperature	Digital Thermometer with Probe	RYG_EN0065	4-Jun-22	5-Jul-23	18
Water Lab	THM C ₂ Cl ₄	Gas Chromatography (MSD)	RYG_EN0059	21-Jun-22	21-Dec-23	18
Water Lab	THM C ₂ Cl ₄	Gas Chromatography (FID)	RYG_EN0103	20-Oct-21	20-Apr-23	18
Water Lab	THM C ₂ Cl ₄	Gas Chromatography (FID)	RYG_EN0103	20-Oct-21	20-Apr-23	18
Water Lab	THM C ₂ Cl ₄	Gas Chromatography (FID)	RYG_EN0103	20-Oct-21	20-Apr-23	18
Water Lab	Iron	ICP-AES	RYG_EI0043	30-Sep-21	29-Mar-23	18
Water Lab	Iron	Hot Block	RYG_EI0054	7-Apr-22	7-Oct-23	18
Water Lab	Iron	Chamber (Gold Room)	RYG_EI0167	30-Jun-22	30-Dec-23	18
Water Lab	Lead	ICP-AES	RYG_EI0043	30-Sep-21	29-Mar-23	18
Water Lab	Lead	Hot Block	RYG_EI0054	7-Apr-22	7-Oct-23	18
Water Lab	Lead	Chamber (Gold Room)	RYG_EI0167	30-Jun-22	30-Dec-23	18
Water Lab	Arsenic	ICP-AES	RYG_EI0043	30-Sep-21	29-Mar-23	18
Water Lab	Arsenic	Hot Block	RYG_EI0054	7-Apr-22	7-Oct-23	18
Water Lab	Arsenic	Chamber (Gold Room)	RYG_EI0167	30-Jun-22	30-Dec-23	18
Water Lab	Cadmium	ICP-AES	RYG_EI0043	30-Sep-21	29-Mar-23	18
Water Lab	Cadmium	Hot Block	RYG_EI0054	7-Apr-22	7-Oct-23	18
Water Lab	Cadmium	Chamber (Gold Room)	RYG_EI0167	30-Jun-22	30-Dec-23	18
Water Lab	Selenium	ICP-AES	RYG_EI0043	30-Sep-21	29-Mar-23	18
Water Lab	Selenium	Hot Block	RYG_EI0054	7-Apr-22	7-Oct-23	18
Water Lab	Selenium	Chamber (Gold Room)	RYG_EI0167	30-Jun-22	30-Dec-23	18
Water Lab	Mercury	CVAFS	RYG_EI0011	7-Jun-22	6-Jun-23	12
Coal	Lead	ICP-OES	RYG_EI0037	13-Sep-21	12-Mar-23	18
Coal	Arsenic	ICP-OES	RYG_EI0037	13-Sep-21	12-Mar-23	18
Coal	Cadmium	ICP-OES	RYG_EI0037	13-Sep-21	12-Mar-23	18
Coal	Mercury	CVAFS	RYG_EI0011	7-Jun-22	6-Jun-23	12

3

alsglobal.com



right solutions.
right partner.

รายการเครื่องมือที่ใช้ในการตรวจ / ทดสอบ


Sample Name	Parameter	Equipment Name	ID No.	Calibrated Date	Next Cal	Freq. Calibrate (Months)
Stack	Mercury	Console Control Unit	RYG_F50527	12-Jul-22	12-Jan-23	6
Stack	Mercury	Console Control Unit	RYG_F50518	12-Jul-22	12-Jan-23	6
Stack	Mercury	Flue gas Analyzer	RYG_F50465	19-Jan-22	19-Jan-23	12
Stack	Mercury	Flue gas Analyzer	RYG_F50566	9-Dec-21	9-Dec-22	12
Stack	Mercury	ICP-OES	RYG_EI0011	7-Jun-22	6-Jun-23	12
Stack	Total Suspended Particulate	Console Control Unit	RYG_F50527	12-Jul-22	12-Jan-23	6
Stack	Total Suspended Particulate	Console Control Unit	RYG_F50507	3-Jul-22	3-Jan-23	6
Stack	Total Suspended Particulate	Console Control Unit	RYG_F51093	3-Jul-22	3-Jan-23	6
Stack	Total Suspended Particulate	Flue gas Analyzer	RYG_F50465	19-Jan-22	19-Jan-23	12
Stack	Total Suspended Particulate	Flue gas Analyzer	RYG_F51156	10-Dec-21	10-Dec-22	12
Stack	Total Suspended Particulate	Digital Balance	RYG_EN0002	25-Feb-22	25-Feb-23	12
Stack	PM10	Console Control Unit	RYG_F50527	12-Jul-22	12-Jan-23	6
Stack	PM10	Flue gas Analyzer	RYG_F50465	19-Jan-22	19-Jan-23	12
Stack	PM10	Digital Balance	RYG_EN0004	23-Mar-22	23-Mar-23	12
Stack	Volatile Organic Compounds	Console Control Unit	RYG_F50527	12-Jul-22	12-Jan-23	6
Stack	Volatile Organic Compounds	Flue gas Analyzer	RYG_F50465	19-Jan-22	19-Jan-23	12
Stack	Volatile Organic Compounds	Field Rotameter	RYG_F51039	1-Jul-22	1-Oct-22	3
Stack	Volatile Organic Compounds	Field Rotameter	RYG_EI0119	1-Oct-21	1-Apr-23	18
Stack (CEM)	Oxides of Nitrogen	Analyzer, System calibration, S	-	-	-	-
Stack (CEM)	Sulfur Dioxide	Analyzer, System calibration, S	-	-	-	-
Workplace	Total Dust	Field Rotameter	RYG_F50198	1-Jul-22	1-Oct-22	3
Workplace	Total Dust	Field Rotameter	RYG_F50198	1-Oct-22	1-Jan-23	3
Workplace	Total Dust	Digital Balance	RYG_EN0004	23-Mar-22	23-Mar-23	12
Noise	Leq 24 hrs	Sound Calibrator	RYG_F50313	26-Sep-22	26-Sep-23	12
Noise	Leq 24 hrs	Sound Level Meter	RYG_F50305	21-Jan-22	21-Jan-23	12
Noise	Leq 24 hrs	Sound Level Meter	RYG_F50387	13-Sep-21	13-Sep-22	12
Noise	Leq 24 hrs	Sound Calibrator	RYG_F50316	31-Aug-22	31-Aug-23	12
Noise	Leq 24 hrs	Sound Level Meter	RYG_F50027	10-Jan-22	10-Jan-23	12
Noise	Leq 24 hrs	Sound Level Meter	RYG_F50031	20-Jun-22	20-Jun-23	12
Noise	Leq 8 hrs	Sound Calibrator	RYG_F50495	10-Jan-22	10-Jan-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50302	21-Jan-22	21-Jan-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50029	24-May-22	24-May-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50030	21-Jan-22	21-Jan-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50303	11-Jul-22	11-Jul-23	12
Noise	Leq 8 hrs	Sound Calibrator	RYG_F50495	10-Jan-22	10-Jan-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50302	4-Oct-21	4-Oct-22	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50028	21-Jan-22	21-Jan-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50301	13-Sep-21	13-Sep-22	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50302	11-Jul-22	11-Jul-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50304	11-Jul-22	11-Jul-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50305	11-Jul-22	11-Jul-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50387	13-Sep-21	13-Sep-22	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50388	13-Sep-21	13-Sep-22	12
Noise	Leq 8 hrs	Sound Calibrator	RYG_F50316	31-Aug-22	31-Aug-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50303	11-Jul-22	11-Jul-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50304	11-Jul-22	11-Jul-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50305	11-Jul-22	11-Jul-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50387	13-Sep-21	13-Sep-22	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50388	13-Sep-21	13-Sep-22	12
Noise	Leq 8 hrs	Sound Calibrator	RYG_F50316	31-Aug-22	31-Aug-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50302	19-Oct-22	19-Oct-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50302	11-Jul-22	11-Jul-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50304	11-Jul-22	11-Jul-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50491	10-Jan-22	10-Jan-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50492	10-Jan-22	10-Jan-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50493	10-Jan-22	10-Jan-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50494	10-Jan-22	10-Jan-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50495	10-Jan-22	10-Jan-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50496	10-Jan-22	10-Jan-23	12
Noise	Leq 8 hrs	Sound Level Meter	RYG_F50497	10-Jan-22	10-Jan-23	12



High Volume Air Sampler Calibration Worksheet

Project Site : GHECO-ONE COMPANY LIMITED
Calibrate Location : กรุงเทพมหานคร
Calibrate Date : 15-Aug-22
CalibrationSheet No.: C-150822-RYG_FS0292
Calibrator ID: RYG_FS0205
Calibrator Model: TE-5028A
Calibrator S/N: 1166
Barometric Pressure (mm Hg) : 755
Temperature (°C) : 30
High Volume ID : RYG_FS0292
High Volume Model : TE-5170D
High Volume S/N : 5497
Calibrator Slope : 1.53016
Calibrator Intercept : -0.0468

Test No.	Delta H ₂ O (inch)	Q _{std} (m ³ /min)	I : Chart (CFM)	Linear Regression
1	2.4	1.0475	40	Slope : 35.5471 Intercept : 2.5606 Correlation Coefficient : 0.9989
2	3.0	1.1657	44	
3	4.0	1.3387	50	
4	4.8	1.4621	54	
5	5.8	1.6025	60	



Calibrated by :
(Mr. Apichart Wilars)
Field Scientist (1)

Approved by :
(Mr. Noppong Juntarupan)
Enviro Field Coordinator Scientist (3)


FORM NO.: F 06-073 REVISION NO.: ISSUE DATE: 14/03/16



High Volume Air Sampler Calibration Worksheet

Project Site : GHECO-ONE COMPANY LIMITED
Calibrate Location : กรุงเทพมหานคร
Calibrate Date : 15-Aug-22
CalibrationSheet No.: C-150822-RYG_FS0176
Calibrator ID: RYG_FS0205
Calibrator Model: TE-5028A
Calibrator S/N: 1166
Barometric Pressure (mm Hg) : 755
Temperature (°C) : 30
High Volume ID : RYG_FS0176
High Volume Model : TE-5170D
High Volume S/N : 4802
Calibrator Slope : 1.53016
Calibrator Intercept : -0.0468

Test No.	Delta H ₂ O (inch)	Q _{std} (m ³ /min)	I : Chart (CFM)	Linear Regression
1	2.8	1.1277	40	Slope : 38.6663 Intercept : -4.0017 Correlation Coefficient : 0.9988
2	3.5	1.2553	44	
3	4.4	1.4018	50	
4	5.0	1.4912	54	
5	6.2	1.6553	60	



Calibrated by :
(Mr. Apichart Wilars)
Field Scientist (1)

Approved by :
(Mr. Noppong Juntarupan)
Enviro Field Coordinator Scientist (3)


FORM NO.: F 06-073 REVISION NO.: ISSUE DATE: 14/03/16



High Volume Air Sampler Calibration Worksheet

Project Site : GHECO-ONE COMPANY LIMITED
Calibrate Location : กรุงเทพมหานคร
Calibrate Date : 15-Aug-22
CalibrationSheet No.: C-150822-RYG_FS0175
Calibrator ID: RYG_FS0205
Calibrator Model: TE-5028A
Calibrator S/N: 1166
Barometric Pressure (mm Hg) : 755
Temperature (°C) : 30
High Volume ID : RYG_FS0175
High Volume Model : TE-5170D
High Volume S/N : 4801
Calibrator Slope : 1.53016
Calibrator Intercept : -0.0468

Test No.	Delta H ₂ O (inch)	Q _{std} (m ³ /min)	I : Chart (CFM)	Linear Regression
1	2.6	1.0684	40	Slope : 37.8390 Intercept : -1.2802 Correlation Coefficient : 0.9957
2	3.2	1.2023	44	
3	4.0	1.3387	50	
4	5.0	1.4912	54	
5	5.8	1.6025	60	



Calibrated by :
(Mr. Apichart Wilars)
Field Scientist (1)

Approved by :
(Mr. Noppong Juntarupan)
Enviro Field Coordinator Scientist (3)


FORM NO.: F 06-073 REVISION NO.: ISSUE DATE: 14/03/16



High Volume Air Sampler Calibration Worksheet

Project Site : GHECO-ONE COMPANY LIMITED
Calibrate Location : กรุงเทพมหานคร
Calibrate Date : 15-Aug-22
CalibrationSheet No.: C-150822-RYG_FS0173
Calibrator ID: RYG_FS0205
Calibrator Model: TE-5028A
Calibrator S/N: 1166
Barometric Pressure (mm Hg) : 755
Temperature (°C) : 30
High Volume ID : RYG_FS0173
High Volume Model : TE-5170D
High Volume S/N : 4799
Calibrator Slope : 1.53016
Calibrator Intercept : -0.0468

Test No.	Delta H ₂ O (inch)	Q _{std} (m ³ /min)	I : Chart (CFM)	Linear Regression
1	2.5	1.0682	40	Slope : 33.0117 Intercept : 4.6155 Correlation Coefficient : 0.9997
2	3.2	1.2023	44	
3	4.2	1.3706	50	
4	5.0	1.4912	54	
5	6.4	1.6810	60	



Calibrated by :
(Mr. Apichart Wilars)
Field Scientist (1)

Approved by :
(Mr. Noppong Juntarupan)
Enviro Field Coordinator Scientist (3)


FORM NO.: F 06-073 REVISION NO.: ISSUE DATE: 14/03/16



High Volume Air Sampler Calibration Worksheet

Project Site :	GHECO-DNE COMPANY LIMITED	Barometric Pressure (mm Hg) :	755
Calibrate Location :	บริษัท เจริญโภคภัณฑ์อาหาร จำกัด (มหาชน)	Temperature (°C) :	30
Calibrate Date :	15-Aug-22	High Volume ID :	RYG_FS0177
Calibration Sheet No. :	C-150822-RYG_FS0177	High Volume Model :	TE-5170D
Calibrator ID :	RYG_FS0205	High Volume S/N :	4803
Calibrator Model :	TE-5028A	Calibrator Slope :	1.53016
Calibrator S/N :	1166	Calibrator Intercept :	-0.0468

Test No.	Delta H ₂ O (Inch)	Q _{vol} (m ³ /min)	I : Chart (CFM)	Linear Regression
1	2.6	1.0884	40	Slope : 33.2228 Intercept : 3.7965 Correlation Coefficient : 0.9992
2	3.2	1.2023	44	
3	4.4	1.4018	50	
4	5.2	1.5198	54	
5	6.4	1.6810	60	



Calibrated by
(Mr. Apichart Wilars)
Field Scientist (1)

Approved by
(Mr. Noppong Juntanapan)
Enviro Field Coordinator Scientist (3)

FORM NO.: F 06-073 REVISION NO.: ISSUE DATE: 14/03/16



PENTA
CALIBRATION

PENTA CALIBRATION CO., LTD.
66/124 The Connect 33 Village Kanchanaphisek Road
Dokmai Prawat Bangkok 10250
Tel: +66 (0) 2669-9773
www.pentalab.co.th

Certificate of Calibration

Represent to Certificate of Calibration ,PTC/07/22102

Certificate No.:	PTC/07/22102	Page:	1 of 2
Equipment:	Digital Balance	Condition:	Normal
Manufacturer:	Sartorius	Serial No.:	25409664
Model:	LA130S-F	ID No.:	RYG_EN0001
Type of Balance:	Single interval		

Customer: ALS Laboratory Group (Thailand) Co.,Ltd.
616/10 Moo 5 T.Maenamkoo, A.Pluakdaeng,
Rayong 21140, Thailand

REVIEW BY
APPROVED BY
NEXT CAL. DATE 03/03/23

Environment Condition: Temperature 23.9 °C ± 0.3 °C
Humidity 58.1 %RH ± 4.4 %RH
Air density 1.17 kg/m³

Calibration Place: ALS Laboratory Group (Thailand) Co.,Ltd.
616/10 Moo 5 T.Maenamkoo, A.Pluakdaeng,
Rayong 21140, Thailand


The Method used: In house method, PTC-WI-07, base on Euramet cg 18
Traceability: This certificate is traceable to the SI Units through Thai Calibration Service Co.,Ltd.
. NSC-ONSC Accreditation No.: Calibration 0189

Date Received: March 23, 2022

Calibration Date: March 23, 2022

Issued Date: March 25, 2022

Calibration By: Mr. Rungraje Metakul



Approved By
(Mr. Keattisak Kerdto)
Laboratory Manager

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.
The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM). The effect that the results relate only to the items calibrated.
This calibration certificate shall not be reproduced except in full only, without written approval from penta calibration co., ltd

PTC-FMC-07-02 21 Feb 2019



PENTA
CALIBRATION

PENTA CALIBRATION CO., LTD.
66/124 The Connect 33 Village Kanchanaphisek Road
Dokmai Prawat Bangkok 10250
Tel: +66 (0) 2669-9773
www.pentalab.co.th

Represent to Certificate of Calibration ,PTC/07/22102

Certificate No.: PTC/07/22102

Page: 2 of 2

Measurement Results:

Without Adjustment :

Function Calibration: Non Adjustment

Eccentric Error: Weight to be 1/3, 1/2 or of Maximum capacity

Eccentricity test		50 (g)				
		Position (g)				
		1	2	3	4	5
0.0000	0.0000	-0.0001	0.0000	0.0001		
Maximum deviation:		0.0001				

Repeatability Test : Weight to be 1/2 ≤ L₁ ≤ Maximum capacity

Determination of the standard deviation of weighing balance, Readability 0.0001 (g)

Nominal test value (g)	Standard Deviation
100	0.00009

Error of indication : from nominal value., Readability 0.0001 (g)

Nominal Value (g)	Conventional Mass (g)	Indication (g)	Correction of Balance (g)	Uncertainty (g)	k
0	0.00000	0.0000	0.0000	0.00026	2.07
0.01	0.01000	0.0100	0.0000	0.00026	2.65
0.05	0.05000	0.0500	0.0000	0.00026	2.65
0.1	0.10000	0.1000	0.0000	0.00026	2.65
0.5	0.50000	0.4999	0.0001	0.00026	2.65
1	1.00000	0.9999	0.0001	0.00026	2.65
2	2.00000	1.9999	0.0001	0.00026	2.65
5	5.00001	5.0000	0.0000	0.00026	2.65
10	10.00000	10.0001	-0.0001	0.00026	2.65
20	20.00003	20.0001	-0.0001	0.00026	2.52
100	100.00004	100.0001	-0.0001	0.00027	2.18

Note: Weight of adjust (g)

The End of Certificate


PTC-FMC-07-02 21 Feb 2019



High Volume Air Sampler Calibration Worksheet

Project Site :	GHECO-DNE COMPANY LIMITED	Barometric Pressure (mm Hg) :	755
Calibrate Location :	บริษัท เจริญโภคภัณฑ์อาหาร จำกัด (มหาชน)	Temperature (°C) :	30
Calibrate Date :	15-Aug-22	High Volume ID :	RYG_FS0398
Calibration Sheet No. :	C-150822-RYG_FS0398	High Volume Model :	TE-5009X
Calibrator ID :	RYG_FS0205	High Volume S/N :	5684
Calibrator Model :	TE-5028A	Calibrator Slope :	1.53016
Calibrator S/N :	1166	Calibrator Intercept :	-0.0468

Test No.	Delta H ₂ O (Inch)	Q _{vol} (m ³ /min)	I : Chart (CFM)	Linear Regression
1	1.8	0.9135	32	Slope : 28.7720 Intercept : 5.6799 Correlation Coefficient : 0.9993
2	2.4	1.0475	36	
3	3.2	1.2023	40	
4	4.0	1.3387	44	
5	4.8	1.4621	49	



Calibrated by
(Mr. Apichart Wilars)
Field Scientist (1)

Approved by
(Mr. Noppong Juntanapan)
Enviro Field Coordinator Scientist (3)


FORM NO.: F 06-074 REVISION NO.: ISSUE DATE: 14/03/16



High Volume Air Sampler Calibration Worksheet

Project Site: GHECO-ONE COMPANY LIMITED Barometric Pressure (mm Hg): 755
 Calibrate Location: ศูนย์บริการลูกค้า Temperature (°C): 30
 Calibrate Date: 15-Aug-22 High Volume ID: RYG_FS0189
 CalibrationSheet No.: C-150822-RYG_FS0189 High Volume Model: TE-5009X
 Calibrator ID: RYG_FS0205 High Volume S/N: 4797
 Calibrator Model: TE-5028A Calibrator Slope: 1.53016
 Calibrator S/N: 1166 Calibrator Intercept: -0.0468

Test No.	Delta H ₂ O (inch)	Q _{air} (m ³ /min)	I: Chart (CFM)	Linear Regression
1	1.6	0.8639	32	Slope: 27.7765 Intercept: 7.9305 Correlation Coefficient: 0.9992
2	2.2	1.0049	36	
3	3.0	1.1657	40	
4	3.8	1.3060	44	
5	4.6	1.4323	48	



Calibrated by: Mr. Apichart Wilars
 Field Scientist (1)

Approved by: Mr. Noppong Juntarupan
 Enviro Field Coordinator Scientist (3)


FORM NO.: F 06-074 REVISION NO.: - ISSUE DATE: 14/03/16



High Volume Air Sampler Calibration Worksheet

Project Site: GHECO-ONE COMPANY LIMITED Barometric Pressure (mm Hg): 755
 Calibrate Location: ศูนย์บริการลูกค้า Temperature (°C): 30
 Calibrate Date: 15-Aug-22 High Volume ID: RYG_FS0294
 CalibrationSheet No.: C-150822-RYG_FS0294 High Volume Model: TE-5009X
 Calibrator ID: RYG_FS0205 High Volume S/N: 5501
 Calibrator Model: TE-5028A Calibrator Slope: 1.53016
 Calibrator S/N: 1166 Calibrator Intercept: -0.0468

Test No.	Delta H ₂ O (inch)	Q _{air} (m ³ /min)	I: Chart (CFM)	Linear Regression
1	1.8	0.9135	32	Slope: 24.8765 Intercept: 9.4830 Correlation Coefficient: 0.9983
2	2.5	1.0682	36	
3	3.2	1.2023	40	
4	4.4	1.4018	44	
5	5.4	1.5479	48	



Calibrated by: Mr. Apichart Wilars
 Field Scientist (1)

Approved by: Mr. Noppong Juntarupan
 Enviro Field Coordinator Scientist (3)


FORM NO.: F 06-074 REVISION NO.: - ISSUE DATE: 14/03/16



High Volume Air Sampler Calibration Worksheet

Project Site: GHECO-ONE COMPANY LIMITED Barometric Pressure (mm Hg): 755
 Calibrate Location: ศูนย์บริการลูกค้า Temperature (°C): 30
 Calibrate Date: 15-Aug-22 High Volume ID: RYG_FS0183
 CalibrationSheet No.: C-150822-RYG_FS0183 High Volume Model: TE-5009X
 Calibrator ID: RYG_FS0205 High Volume S/N: 4791
 Calibrator Model: TE-5028A Calibrator Slope: 1.53016
 Calibrator S/N: 1166 Calibrator Intercept: -0.0468

Test No.	Delta H ₂ O (inch)	Q _{air} (m ³ /min)	I: Chart (CFM)	Linear Regression
1	1.8	0.9135	32	Slope: 26.0416 Intercept: 8.0946 Correlation Coefficient: 0.9968
2	2.6	1.0084	36	
3	3.2	1.2023	40	
4	4.4	1.4018	44	
5	5.2	1.5198	48	



Calibrated by: Mr. Apichart Wilars
 Field Scientist (1)

Approved by: Mr. Noppong Juntarupan
 Enviro Field Coordinator Scientist (3)


FORM NO.: F 06-074 REVISION NO.: - ISSUE DATE: 14/03/16



High Volume Air Sampler Calibration Worksheet

Project Site: GHECO-ONE COMPANY LIMITED Barometric Pressure (mm Hg): 755
 Calibrate Location: ศูนย์บริการลูกค้า Temperature (°C): 30
 Calibrate Date: 15-Aug-22 High Volume ID: RYG_FS0187
 CalibrationSheet No.: C-150822-RYG_FS0187 High Volume Model: TE-5009X
 Calibrator ID: RYG_FS0205 High Volume S/N: 4795
 Calibrator Model: TE-5028A Calibrator Slope: 1.53016
 Calibrator S/N: 1166 Calibrator Intercept: -0.0468

Test No.	Delta H ₂ O (inch)	Q _{air} (m ³ /min)	I: Chart (CFM)	Linear Regression
1	1.6	0.8639	32	Slope: 28.3165 Intercept: 7.3928 Correlation Coefficient: 0.9982
2	2.2	1.0049	36	
3	3.0	1.1657	40	
4	3.8	1.3060	44	
5	4.5	1.4171	48	



Calibrated by: Mr. Apichart Wilars
 Field Scientist (1)

Approved by: Mr. Noppong Juntarupan
 Enviro Field Coordinator Scientist (3)


FORM NO.: F 06-074 REVISION NO.: - ISSUE DATE: 14/03/16



High Volume Air Sampler Calibration Worksheet

Project Site : GHECO-ONE COMPANY LIMITED
 Calibrate Location : กรุงเทพมหานคร
 Calibrate Date : 15-Aug-22
 CalibrationSheet No.: C-150822-RYG_FS0400
 Calibrator ID: RYG_FS0205
 Calibrator Model: TE-5028A
 Calibrator S/N: 1166
 Barometric Pressure (mm Hg) : 755
 Temperature (°C) : 30
 High Volume ID : RYG_FS0400
 High Volume Model: TE-5009X
 High Volume S/N: 5691
 Calibrator Slope : 1.53016
 Calibrator Intercept: -0.0468

Test No.	Delta H ₂ O (inch)	Q _{std} (m ³ /min)	I: Chart (CFM)	Linear Regression
1	1.8	0.9135	32	Slope: 25.2583 Intercept: 8.7328 Correlation Coefficient: 0.9996
2	2.6	1.0884	36	
3	3.4	1.2379	40	
4	4.4	1.4018	44	
5	5.4	1.5479	48	



Calibrated by: [Signature]
 (Mr. Apichart Wilars)
 Field Scientist (1)

Approved by: [Signature]
 (Mr. Noppong Juntarupan)
 Enviro Field Coordinator Scientist (3)


FORM NO.: F-06-074 REVISION NO.: ISSUE DATE: 14/03/16



High Volume Air Sampler Calibration Worksheet

Project Site : GHECO-ONE COMPANY LIMITED
 Calibrate Location : กรุงเทพมหานคร
 Calibrate Date : 15-Aug-22
 CalibrationSheet No.: C-150822-RYG_FS0186
 Calibrator ID: RYG_FS0205
 Calibrator Model: TE-5028A
 Calibrator S/N: 1166
 Barometric Pressure (mm Hg) : 755
 Temperature (°C) : 30
 High Volume ID : RYG_FS0186
 High Volume Model: TE-5009X
 High Volume S/N: 4794
 Calibrator Slope : 1.53016
 Calibrator Intercept: -0.0468

Test No.	Delta H ₂ O (inch)	Q _{std} (m ³ /min)	I: Chart (CFM)	Linear Regression
1	1.9	0.9372	32	Slope: 33.5905 Intercept: 0.6438 Correlation Coefficient: 0.9987
2	2.4	1.0475	36	
3	3.0	1.1657	40	
4	3.8	1.3060	44	
5	4.4	1.4018	48	



Calibrated by: [Signature]
 (Mr. Apichart Wilars)
 Field Scientist (1)

Approved by: [Signature]
 (Mr. Noppong Juntarupan)
 Enviro Field Coordinator Scientist (3)

FORM NO.: F-06-074 REVISION NO.: ISSUE DATE: 14/03/16


BKK_EL0037



High Volume Air Sampler Calibration Worksheet

Project Site : GHECO-ONE COMPANY LIMITED
 Calibrate Location : กรุงเทพมหานคร
 Calibrate Date : 15-Aug-22
 CalibrationSheet No.: C-150822-RYG_FS0174
 Calibrator ID: RYG_FS0205
 Calibrator Model: TE-5028A
 Calibrator S/N: 1166
 Barometric Pressure (mm Hg) : 755
 Temperature (°C) : 30
 High Volume ID : RYG_FS0174
 High Volume Model: TE-5170D
 High Volume S/N: 4800
 Calibrator Slope : 1.53016
 Calibrator Intercept: -0.0468

Test No.	Delta H ₂ O (inch)	Q _{std} (m ³ /min)	I: Chart (CFM)	Linear Regression
1	2.4	1.0475	40	Slope: 36.4386 Intercept: 1.1135 Correlation Coefficient: 0.9965
2	3.2	1.2023	44	
3	4.0	1.3387	50	
4	4.8	1.4621	54	
5	5.8	1.6025	60	



Calibrated by: [Signature]
 (Mr. Apichart Wilars)
 Field Scientist (1)

Approved by: [Signature]
 (Mr. Noppong Juntarupan)
 Enviro Field Coordinator Scientist (3)

FORM NO.: F-06-073 REVISION NO.: ISSUE DATE: 14/03/16

© 2020 by Agilent Technologies

Agilent CrossLab Compliance Services

Certificate of System Qualification

ES-OQ

System ID: MY16010005
 Organization Name: ALS Laboratory Group (Thailand) Co., Ltd.
 Organization Location: 104 Phatthanakan 40 Phatthanakan Rd., Bangkok 10250
 Date: September 13, 2021 5:49:11 PM
 EQP Name: Agilent/Recommended
 EQP Revision: ES.02.50
 Overall Qualification Status: Pass

Preparation

Pass

Instrument Tests

Pass

Autosampler Operation

Pass

REVIEW BY: Thitima B.
 APPROVED BY: [Signature]
 NEXT CAL. DATE: 12 Mar 23

Date: September 13, 2021 5:49:11 PM
 System ID: MY16010005

REVIEW BY Sudarat N.
APPROVED BY Sudarat N.
NEXT CAL. DATE 01/06/2023

Serial-No.: 1700124 Customer-No.: 004-002
Date: 9/06/2022 Carried out by: Mr. Stephen Poku

Maintenance with following Operational Qualification (OQ)
(requires a separate OQ protocol)

Maintenance Protocol

Atomic Fluorescence Spectrometer
mercur / mercur plus

[illegible]

Maintenance works basic unit

- | | |
|--|-------------------------------------|
| maintenances workbooks | <input checked="" type="checkbox"/> |
| lightness visual check inside the Mercur | <input checked="" type="checkbox"/> |
| visual check if gold-traps are broken | <input checked="" type="checkbox"/> |
| visual check if spectrometer is contaminated | <input checked="" type="checkbox"/> |
| reactor cleaning | <input checked="" type="checkbox"/> |
| check pump-hose, if necessary change it | <input checked="" type="checkbox"/> |
| check drying-hose, output gas-liquid-separator | <input checked="" type="checkbox"/> |
| test Bubble-Sensor | <input checked="" type="checkbox"/> |
| check gas flows | <input checked="" type="checkbox"/> |
| check volume flows, reagents | <input checked="" type="checkbox"/> |
| recording stray light values | <input checked="" type="checkbox"/> |
| measurement with 30 ng/l | <input checked="" type="checkbox"/> |

Maintenance works Autosampler

Serial No.: 701 229

- | | |
|--|-------------------------------------|
| lubricate the dosing-winding (Teflon-grease-spray) | <input checked="" type="checkbox"/> |
| clean the dosing cylinder, if necessary exchange it | <input checked="" type="checkbox"/> |
| lubricate the winding system of the height drive with some drops of oil | <input checked="" type="checkbox"/> |
| check the toothed belt | <input checked="" type="checkbox"/> |
| check the position of the mechanical stopper (height: 13mm) | <input checked="" type="checkbox"/> |
| check the pump rate of mixing pump (<14s AS52, typ.7s/<20s AS52S, typ.10s) | <input checked="" type="checkbox"/> |
| check the pump rate of washing cup | <input checked="" type="checkbox"/> |
| check the electrical hose connections for good contact | <input checked="" type="checkbox"/> |
| check the connectors of the magnetic valves | <input checked="" type="checkbox"/> |
| check the dosing hose for buckling, if necessary exchange it | <input checked="" type="checkbox"/> |

Device parameter		nominal value	actual value
visual check general tightness inside the Mercur		o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
visual check Goldtraps		o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
visual check spectrometer			
	cuvette	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
	lens	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check pump hoses		o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check hoses and hose connectors		o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check and clean reactor		o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check drying hose output Gas-liquid-seperator		o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check bubble-sensor		o.k.: <input checked="" type="checkbox"/>	not o.k.: <input type="checkbox"/>
Check gasflow			
	Argon pressure valve 4	1.2 – 1.5 bar	1.5 bar
	Valve 1	10 Nl/h or 0.166 NL/min	0.167 NL/min
	Valve 2	50 Nl/h or 0.833 NL/min	0.33 NL/min
	Valve 3	5 Nl/h or 0.083 NL/min	0.083 NL/min
	Valve 4	10 Nl/h or 0.166 NL/min	0.166 NL/min
Check liquid flow			
	Acid	2.5ml/min ± 1 ml	1.5 ml/min
	Red.-agent	2.5ml/min ± 1 ml	1.5 ml/min
	Sample	10ml/min ± 2 ml	10 ml/min
Adventitious light - values		(V) from file	
	100	0	0
	200	0	0
	300	0	0
	350	0	0
	400	1	1
	450	2	2
	500	6	6
	550	13	14
	575	19	20
	600	27	28




MULTIPOINT CALIBRATION REPORT

Device parameter	nominal value	actual value
Analytical parameters		
Conditions.: max conc.: 10µg/L PMT-voltage: 404 V		
Blank-solution without enrichment / FBR 30 ng/L	Int > 0.0015 RSD < 3 %	Int... 0.0005 Int... 0.0026 RSD... 1.12 %
Conditions.: max conc.: 1.7µg/L PMT-voltage: 396 V		
Blank-solution with enrichment / FBR 30 ng/L	Int > 0.008 RSD < 3 %	Int... 0.0018 Int... 0.0104 RSD... 0.89 %
Fok.- factor (Int ₂ / Int ₁)	> 3,5	4
Comments		

Calibration Date	1-Jul-22	Equipment Name	NOx Analyzer
Manufacturer	HORIBA	Model	APNA-370
Serial No.	T95HWM41	Equipment ID	RYG_FS0461
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	55.88	Cylinder No.	GN0027222
Cylinder Pressure (psi)	1800	Certified By	Algas Inc.
Certified Date	9-Feb-22	Expired Date	9-Feb-30

Point	CALIBRATION RESULTS						
	Ideal	Actual NO	Error NO	%Error NO	Actual NOx	Error NOx	%Error NOx
ZERO	0.00	0.10	0.10	0.10	0.10	0.10	0.10
1	100.00	98.70	-1.30	-1.30	100.10	0.10	0.10
2	200.00	201.00	1.00	0.50	201.40	1.40	0.70
3	300.00	298.30	-1.70	-0.57	302.10	2.10	0.70
4	400.00	398.40	-1.60	-0.40	403.50	3.50	0.88
AVERAGE (%)				-0.33			0.50



Mr. Srikar Palkon
Signature Technician

Bangkok, 8/06/2022
Place, Date (DD/MM/YYYY)

Signature Customer
06/06/2022
Place, Date (DD/MM/YYYY)

Calibrated By
(Mr. Jirawat Sakam)
Field Environmental Scientist (3)


Approved By
(Mr. Sarayuth Jittrantont)
Assistant General Manager



MULTIPOINT CALIBRATION REPORT

Calibration Date	1-Jul-22	Equipment Name	NOx Analyzer
Manufacturer	HORIBA	Model	APNA-370
Serial No.	148EH0E0	Equipment ID	BKK_FS1064
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	55.88	Cylinder No.	GN0027222
Cylinder Pressure (psi)	1800	Certified By	Algas Inc.
Certified Date	9-Feb-22	Expired Date	9-Feb-30

Point	CALIBRATION RESULTS						
	Ideal	Actual NO	Error NO	%Error NO	Actual NOx	Error NOx	%Error NOx
ZERO	0.00	0.10	0.10	0.10	0.10	0.10	0.10
1	100.00	99.00	-1.00	-1.00	100.70	0.70	0.70
2	200.00	199.40	-0.60	-0.30	201.50	1.50	0.75
3	300.00	298.60	-1.40	-0.47	302.30	2.30	0.77
4	400.00	401.40	1.40	0.35	398.00	-2.00	-0.50
AVERAGE (%)				-0.26			0.36



Calibrated By
(Mr. Jirawat Sakam)
Field Environmental Scientist (3)


Approved By
(Mr. Sarayuth Jittrantont)
Assistant General Manager



MULTIPOINT CALIBRATION REPORT

Calibration Date	1-Jul-22	Equipment Name	NOx Analyzer
Manufacturer	HORIBA	Model	APNA-370
Serial No.	7AV89544	Equipment ID	RYG_FS0272
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	55.88	Cylinder No.	GN0027222
Cylinder Pressure (psi)	1800	Certified By	Algas Inc.
Certified Date	9-Feb-22	Expired Date	9-Feb-30

Point	CALIBRATION RESULTS						
	Ideal	Actual NO	Error NO	%Error NO	Actual NOx	Error NOx	%Error NOx
ZERO	0.00	0.05	0.05	0.05	0.10	0.10	0.10
1	100.00	99.10	-0.90	-0.90	100.10	0.10	0.10
2	200.00	198.60	-1.40	-0.70	199.00	-1.00	-0.50
3	300.00	298.70	-1.30	-0.43	300.50	0.50	0.17
4	400.00	398.00	-2.00	-0.50	398.70	-1.30	-0.33
AVERAGE (%)				-0.50			-0.09



Calibrated By
(Mr. Jirawat Sakam)
Field Environmental Scientist (3)


Approved By
(Mr. Sarayuth Jittrantont)
Assistant General Manager



MULTIPOINT CALIBRATION REPORT

Calibration Date	1-Jul-22	Equipment Name	NOx Analyzer
Manufacturer	Teledyne API	Model	T200
Serial No.	2198	Equipment ID	RYG_FS0252
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	55.88	Cylinder No.	GN0027222
Cylinder Pressure (psi)	1800	Certified By	Airgas Inc.
Certified Date	9-Feb-22	Expired Date	9-Feb-30

Point	CALIBRATION RESULTS						
	Ideal	Actual NO	Error NO	%Error NO	Actual NOx	Error NOx	%Error NOx
ZERO	0.00	0.10	0.10	0.10	0.10	0.10	0.10
1	100.00	98.80	-1.20	-1.20	101.00	1.00	1.00
2	200.00	198.00	-2.00	-1.00	201.30	1.30	0.65
3	300.00	298.10	-1.90	-0.63	302.30	2.30	0.77
4	400.00	398.20	-1.80	-0.45	398.80	-1.20	-0.30
AVERAGE (%)				-0.64			0.44



Calibrated By

(Mr.Jirawut Sakam)
Field Environmental Scientist (3)

Approved By

(Mr.Sarayuth Jittrantont)
Assistant General Manager

ALS Laboratory Group


FORM NO.: F 06-056 REVISION NO.: - ISSUE DATE: 02/04/12



MULTIPOINT CALIBRATION REPORT

Calibration Date	1-Jul-22	Equipment Name	NOx Analyzer
Manufacturer	HORIBA	Model	APNA-370
Serial No.	AWXG87CR	Equipment ID	RYG_FS0453
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	55.88	Cylinder No.	GN0027222
Cylinder Pressure (psi)	1800	Certified By	Airgas Inc.
Certified Date	9-Feb-22	Expired Date	9-Feb-30

Point	CALIBRATION RESULTS						
	Ideal	Actual NO	Error NO	%Error NO	Actual NOx	Error NOx	%Error NOx
ZERO	0.00	0.10	0.10	0.10	0.10	0.10	0.10
1	100.00	99.60	-0.40	-0.40	101.40	1.40	1.40
2	200.00	198.60	-1.40	-0.70	199.60	-0.20	-0.10
3	300.00	299.00	-1.00	-0.33	298.50	-1.50	-0.50
4	400.00	402.10	2.10	0.53	401.20	1.20	0.30
AVERAGE (%)				-0.18			0.24



Calibrated By

(Mr.Jirawut Sakam)
Field Environmental Scientist (3)

Approved By

(Mr.Sarayuth Jittrantont)
Assistant General Manager

ALS Laboratory Group


FORM NO.: F 06-056 REVISION NO.: - ISSUE DATE: 02/04/12



MULTIPOINT CALIBRATION REPORT

Calibration Date	1-Jul-22	Equipment Name	NOx Analyzer
Manufacturer	HORIBA	Model	APNA-370
Serial No.	NV0ER3YH	Equipment ID	RYG_FS0459
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	55.88	Cylinder No.	GN0027222
Cylinder Pressure (psi)	1800	Certified By	Airgas Inc.
Certified Date	9-Feb-22	Expired Date	9-Feb-30

Point	CALIBRATION RESULTS						
	Ideal	Actual NO	Error NO	%Error NO	Actual NOx	Error NOx	%Error NOx
ZERO	0.00	0.05	0.05	0.05	0.10	0.10	0.10
1	100.00	99.50	-0.50	-0.50	101.80	1.80	1.80
2	200.00	198.70	-1.30	-0.65	199.70	-0.30	-0.15
3	300.00	301.10	1.10	0.37	301.50	1.50	0.50
4	400.00	401.30	1.30	0.33	399.90	-1.10	-0.28
AVERAGE (%)				-0.08			0.39



Calibrated By

(Mr.Jirawut Sakam)
Field Environmental Scientist (3)

Approved By

(Mr.Sarayuth Jittrantont)
Assistant General Manager

ALS Laboratory Group


FORM NO.: F 06-056 REVISION NO.: - ISSUE DATE: 02/04/12



MULTIPOINT CALIBRATION REPORT

Calibration Date	1-Jul-22	Equipment Name	NOx Analyzer
Manufacturer	HORIBA	Model	APNA-370
Serial No.	R06K0177	Equipment ID	RYG_FS0463
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	55.88	Cylinder No.	GN0027222
Cylinder Pressure (psi)	1800	Certified By	Airgas Inc.
Certified Date	9-Feb-22	Expired Date	9-Feb-30

Point	CALIBRATION RESULTS						
	Ideal	Actual NO	Error NO	%Error NO	Actual NOx	Error NOx	%Error NOx
ZERO	0.00	0.10	0.10	0.10	0.10	0.10	0.10
1	100.00	98.80	-1.20	-1.20	101.10	1.10	1.10
2	200.00	201.80	1.80	0.90	201.50	1.50	0.75
3	300.00	299.40	-0.60	-0.20	302.60	2.60	0.87
4	400.00	398.10	-1.90	-0.47	401.90	1.90	0.47
AVERAGE (%)				-0.18			0.66



Calibrated By

(Mr.Jirawut Sakam)
Field Environmental Scientist (3)

Approved By

(Mr.Sarayuth Jittrantont)
Assistant General Manager

ALS Laboratory Group


FORM NO.: F 06-056 REVISION NO.: - ISSUE DATE: 02/04/12



MULTIPOINT CALIBRATION REPORT

Calibration Date	1-Jul-22	Equipment Name	SO2 Analyzer
Manufacturer	HORIBA	Model	APSA-370
Serial No.	VABF9LSH	Equipment ID	RYG_FS0480
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	56.3	Cylinder No.	GN0027222
Cylinder Pressure (psi)	1800	Certified By	Airgas Inc.
Certified Date	9-Feb-22	Expired Date	9-Feb-30

Point	CALIBRATION RESULTS			
	Ideal	Actual	Error	%Error
ZERO	0.00	0.10	0.10	0.10
1	100.00	98.70	-1.30	-1.30
2	200.00	197.80	-2.20	-1.10
3	300.00	295.50	-3.50	-1.17
4	400.00	398.30	-1.70	-0.42
AVERAGE (%)				-0.78



Calibrated By

(Mr. Jirawut Sakam)
Field Environmental Scientist (3)

Approved By

(Mr. Sarayuth Jitranont)
Assistant General Manager

ALS Laboratory Group


FORM NO.: F 05-056 REVISION NO.: ISSUE DATE: 02/04/12



MULTIPOINT CALIBRATION REPORT

Calibration Date	1-Jul-22	Equipment Name	SO2 Analyzer
Manufacturer	HORIBA	Model	APSA-370
Serial No.	NM3M2D5M	Equipment ID	RYG_FS0288
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	56.3	Cylinder No.	GN0027222
Cylinder Pressure (psi)	1800	Certified By	Airgas Inc.
Certified Date	9-Feb-22	Expired Date	9-Feb-30

Point	CALIBRATION RESULTS			
	Ideal	Actual	Error	%Error
ZERO	0.00	0.10	0.10	0.10
1	100.00	98.90	-1.10	-1.10
2	200.00	202.40	2.40	1.20
3	300.00	302.30	2.30	0.77
4	400.00	398.00	-2.00	-0.50
AVERAGE (%)				0.09



Calibrated By

(Mr. Jirawut Sakam)
Field Environmental Scientist (3)

Approved By

(Mr. Sarayuth Jitranont)
Assistant General Manager

ALS Laboratory Group


FORM NO.: F 05-056 REVISION NO.: ISSUE DATE: 02/04/12



MULTIPOINT CALIBRATION REPORT

Calibration Date	1-Jul-22	Equipment Name	SO2 Analyzer
Manufacturer	HORIBA	Model	APSA-370
Serial No.	1092NYKM	Equipment ID	RYG_FS0271
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	56.3	Cylinder No.	GN0027222
Cylinder Pressure (psi)	1800	Certified By	Airgas Inc.
Certified Date	9-Feb-22	Expired Date	9-Feb-30

Point	CALIBRATION RESULTS			
	Ideal	Actual	Error	%Error
ZERO	0.00	0.10	0.10	0.10
1	100.00	99.50	-0.50	-0.50
2	200.00	198.20	-1.80	-0.90
3	300.00	297.60	-2.40	-0.80
4	400.00	398.00	-2.00	-0.50
AVERAGE (%)				-0.52



Calibrated By

(Mr. Jirawut Sakam)
Field Environmental Scientist (3)

Approved By

(Mr. Sarayuth Jitranont)
Assistant General Manager

ALS Laboratory Group


FORM NO.: F 05-056 REVISION NO.: ISSUE DATE: 02/04/12



MULTIPOINT CALIBRATION REPORT

Calibration Date	1-Jul-22	Equipment Name	SO2 Analyzer
Manufacturer	Teledyne API	Model	T100
Serial No.	1773	Equipment ID	RYG_FS0251
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	56.3	Cylinder No.	GN0027222
Cylinder Pressure (psi)	1800	Certified By	Airgas Inc.
Certified Date	9-Feb-22	Expired Date	9-Feb-30

Point	CALIBRATION RESULTS			
	Ideal	Actual	Error	%Error
ZERO	0.00	0.10	0.10	0.10
1	100.00	99.60	-0.40	-0.40
2	200.00	201.80	1.80	0.90
3	300.00	297.20	-2.80	-0.93
4	400.00	395.00	-5.00	-1.00
AVERAGE (%)				-0.27



Calibrated By

(Mr. Jirawut Sakam)
Field Environmental Scientist (3)

Approved By

(Mr. Sarayuth Jitranont)
Assistant General Manager

ALS Laboratory Group


FORM NO.: F 05-056 REVISION NO.: ISSUE DATE: 02/04/12



MULTIPOINT CALIBRATION REPORT

Calibration Date	1-Jul-22	Equipment Name	SO2 Analyzer
Manufacturer	HORIBA	Model	APSA-370
Serial No.	90U0J31	Equipment ID	RYG_FS0452
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	56.3	Cylinder No.	GN0027222
Cylinder Pressure (psi)	1800	Certified By	Airgas Inc.
Certified Date	9-Feb-22	Expired Date	9-Feb-30

Point	CALIBRATION RESULTS			
	Ideal	Actual	Error	%Error
ZERO	0.00	0.10	0.10	0.10
1	100.00	98.60	-1.40	-1.40
2	200.00	198.00	-2.00	-1.00
3	300.00	298.10	-1.90	-0.63
4	400.00	398.20	-1.80	-0.45
AVERAGE (%)				-0.88



Calibrated By

(Mr. Jirawut Sakam)
Field Environmental Scientist (3)

Approved By

(Mr. Sarayuth Jitranont)
Assistant General Manager

ALS Laboratory Group


FORM NO.: F 06-056 REVISION NO.: - ISSUE DATE: 02/04/12



MULTIPOINT CALIBRATION REPORT

Calibration Date	1-Jul-22	Equipment Name	SO2 Analyzer
Manufacturer	HORIBA	Model	APSA-370
Serial No.	PAUY077A	Equipment ID	RYG_FS0458
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	56.3	Cylinder No.	GN0027222
Cylinder Pressure (psi)	1800	Certified By	Airgas Inc.
Certified Date	9-Feb-22	Expired Date	9-Feb-30

Point	CALIBRATION RESULTS			
	Ideal	Actual	Error	%Error
ZERO	0.00	0.10	0.10	0.10
1	100.00	98.90	-1.10	-1.10
2	200.00	201.10	1.10	0.55
3	300.00	302.30	2.30	0.77
4	400.00	398.60	-1.40	-0.35
AVERAGE (%)				-0.01



Calibrated By

(Mr. Jirawut Sakam)
Field Environmental Scientist (3)

Approved By

(Mr. Sarayuth Jitranont)
Assistant General Manager

ALS Laboratory Group


FORM NO.: F 06-056 REVISION NO.: - ISSUE DATE: 02/04/12



MULTIPOINT CALIBRATION REPORT

Calibration Date	1-Jul-22	Equipment Name	SO2 Analyzer
Manufacturer	HORIBA	Model	APSA-370
Serial No.	XL29Y85B	Equipment ID	RYG_FS0462
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	56.3	Cylinder No.	GN0027222
Cylinder Pressure (psi)	1800	Certified By	Airgas Inc.
Certified Date	9-Feb-22	Expired Date	9-Feb-30

Point	CALIBRATION RESULTS			
	Ideal	Actual	Error	%Error
ZERO	0.00	0.10	0.10	0.10
1	100.00	99.10	-0.90	-0.90
2	200.00	198.10	-1.90	-0.95
3	300.00	297.90	-2.10	-0.70
4	400.00	403.20	3.20	0.80
AVERAGE (%)				-0.33



Calibrated By

(Mr. Jirawut Sakam)
Field Environmental Scientist (3)

Approved By

(Mr. Sarayuth Jitranont)
Assistant General Manager

ALS Laboratory Group

FORM NO.: F 06-056 REVISION NO.: - ISSUE DATE: 02/04/12



63/14-15,67/35-36, Soi Petchkasem7,7/1, Petchkasem Rd,
Walthapra, Bangkokkhai, Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranatee.com

CERTIFICATE OF CALIBRATION

Certificate No: WS 12072221
Page 1 of 2 pages

Measurement Item	Cup anemometer with data logger	
Manufacturer	Data logger: Navisys Cup anemometer: Navisys	
Model/Type	Data logger: 230 WS 26LB Cup anemometer: WS 28A	
Serial Number	Data logger: A5374 Cup anemometer:	
Lot No.	Data logger: RYG FS0412 Cup anemometer:	
Customer	ALS laboratory group (Thailand) co., ltd 104 Phahonrachon Rd Phahonrachon Rd Khwaeng Bhum Buang Khet Bhum Buang Bangkok 10250 Thailand	
Test Conditions	<ul style="list-style-type: none"> Wind tunnel cross test section area Anemometer height area Diameter of mounting pole Blockage ratio of test object 	900 cm ² 100 cm ² 111 mm 0.111
Test Conditions	<ul style="list-style-type: none"> Air temperature Air pressure Relative air humidity 	29.9 ±0.5 °C 1027.7 ±0.4 hPa 67.7 ±3.5 %RH
Calibration Procedure	Calibration was carried out based on: ISO 61422-12: 1911, 2005 Power Performance Measurements of Battery-Powered Wind Turbines. NACNET Anemometer Calibration Procedure - version 2 2020.	
Traceability	This instruction documents the traceable to national standard which require the UKA of measurements according to the international system of units (SI) through National Institute of Metrology, Thailand (NIMT)	
Measurement Date	Jul 29, 2021	
Issued Date	Jul 29, 2021	
Calibrated By	<input checked="" type="checkbox"/> Mr. Sritrak Thachasat <input type="checkbox"/> Miss Chaita Waisakulaya	
Approved Signature	 Mr. Panyu Booncharoen Technical Support and Calibration Manager	



THIS CERTIFICATE MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.

Continuation of Certificate of Calibration Number

Certificate No: WD-12072021
Page 2 of 2 Pages

Result of calibration: ☒ Without adjustment ☐ With adjustment
Calibration in the range of 1 - 16 m/s at a calibration interval of 1 m/s.

The results of calibration and associated measurement uncertainties are reported in the table below.

V _{act} Reading m/s	V _{act} Reading m/s	Error (m/s)	Uncertainty (%)
2.075	2.0	-0.1	2.5
4.150	4.1	0.1	1.3
6.225	6.0	0.0	0.6
8.300	8.1	0.1	0.6
10.375	10.1	0.1	0.7
12.450	12.2	0.2	0.6
14.525	14.3	0.3	0.4
16.600	16.4	0.4	0.6
18.675	18.5	0.3	0.4
20.750	20.5	0.2	0.6
22.825	22.6	0.1	0.6
24.900	24.7	0.2	0.6
26.975	26.7	0.3	0.6
29.050	28.8	0.2	0.6
31.125	30.9	0.3	0.4

UUC* Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

Appendix 1: Instrumentations

NO	Sensor	Manufacturer	Model/Type	Calibration Date	Certificate Report Number	Range
1	Pressure	TECOT INC	84309148	July 16, 2020	MA/0035/20	5 - 30 mPa
2	Precision Differential Pressure Meter	Zepac	DPM5500	July 16, 2020	MA/0035/20	5 - 30 Pa
3	Air velocity transducer (hot wire)	TSI INC	8445-12	July 20, 2020	MA/0035/20	0 - 5 m/s
4	Temperature	Zepac	CPA 100	March 30, 2021	CL/027/24	30 - 70°C
5	Relative humidity	Zepac	CPA 100	March 30, 2021	RH/030/2021	0 - 100 %RH
6	Atmospheric pressure	Zepac	CPA 100	March 30, 2021	BP/010/2021	500 - 1100 mPa
7	Wind tunnel	CGOON	MP3300			0 - 20 m/s

End of certificate of calibration



CERTIFICATE OF CALIBRATION

Certificate No: WD-12072021
Page 1 of 2 pages

Measurement Item : Wind direction sensor with data logger.

Manufacturer : Data logger: Novayn
Wind direction sensor: Novayn

Model/Type : Data logger: 200-WD-25LB
Wind direction sensor: WS-02R

Serial Number : Data logger: A5374
Wind direction sensor: -

ID No : Data logger: R10_F80412
Wind direction sensor: -

Customer : AIS laboratory group (Thailand) Co., Ltd.
104 Phatthanawan 40, Phatthanawan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand

Environmental Condition:
The measurement was carried out in an ambient temperature of (23±3)°C and relative humidity of (40-10)%

Measurement Method:
The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and the laser is used for axis control. The measurement were taken at 45° intervals in clockwise and counterclockwise directions.

Note: The UUC was warmed up for 1 hour prior to the calibration being performed.

Traceability:
The measurement results are traceable to the international system of units (SI) through Certificate No: OC565-07-0045, Certificate No: HW863/G044.

Measurement Date : Jul 29, 2021.
Issued Date : Jul 29, 2021.

Performed by
☒ Mr. Sornrat Thachaisat
☐ Miss Graha Wathavajay



Approved Signatory:
Mr. Panya Booncharoen
Technical Support
and Calibration Manager

THIS CERTIFICATE REPORT MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.

Continuation of Certificate of Calibration Number

Certificate No: WD-12072021
Page 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment.

Calibration in the range of 0 - 360 ° at a calibration interval of 45°.

The results of calibration and associated measurement uncertainties are reported in table below.

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC* Reading (°)	Error (°)	Uncertainty ±(°)
1	Clockwise	0/360	360	359	-1	2.0
2		45	45	42	-3	3.0
3		90	90	87	-3	3.0
4		135	135	133	-2	2.0
5		180	180	178	-2	3.0
6		225	225	226	1	3.0
7		270	270	273	3	3.0
8		315	315	318	3	3.0
9	Counter Clockwise	0/360	360	359	-1	2.0
10		45	45	42	-3	3.0
11		90	90	87	-3	3.0
12		135	135	133	-2	2.0
13		180	180	178	-2	3.0
14		225	225	226	1	3.0
15		270	270	273	3	3.0
16		315	315	318	3	3.0

UUC* Unit Under Calibration The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

End of Certificate of Calibration



CERTIFICATE OF CALIBRATION

Certificate No: WS 04072021
Page 1 of 2 pages

Measurement Item : Cup anemometer with data logger.

Manufacturer : Data logger: Novayn
Cup anemometer: Novayn

Model/Type : Data logger: 200-WS 20FL
Cup anemometer: WS 02P

Serial Number : Data logger: A4967
Cup anemometer: -

ID No : Data logger: R10_F80389
Cup anemometer: -

Customer : AIS laboratory group (Thailand) Co., Ltd.
104 Phatthanawan 40, Phatthanawan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand

Test Conditions : Wind tunnel cross section area : 220 cm²
Anemometer frontal area : 120 cm²
Diameter of mounting pipe : 75 mm
Blockage ratio of test object : 0.11 (%)

Test Conditions : Air temperature : 24.0 ±0.8 °C
Air pressure : 100.9 ±0.4 hPa
Relative air humidity : 60.0 ±3.6 %RH

Calibration Procedure : Calibration was carried out base on:
ISO 91403-12-1 (D1): 2005 Power Performance Measurements of Continuously Rotating Wind Turbines
MCA/001 Anemometer Calibration Procedure - Version 2, 2009.

Traceability : This calibration documents the traceable to national standard, which relates the unit of measurements according to the international system of units (SI) through National Institute of Metrology (NIM) (NMI).

Measurement Date : Jul 13, 2021.
Issued Date : Jul 14, 2021.

Calibrated by
☒ Mr. Sornrat Thachaisat
☐ Miss Graha Wathavajay



Approved Signatory:
Mr. Panya Booncharoen
Technical Support
and Calibration Manager

THIS CERTIFICATE MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.

Continuation of Certificate of Calibration Number

Certificate No: WB-04072021
Page 2 of 2 pages

Result of calibration: ☒ Without adjustment ☐ With adjustment
Calibration in the range of 1 ~ 16 m/s at a calibration interval of 1 m/s.

The results of calibration and associated measurement uncertainties are reported in the table below.

V _{ref} Reading m/s	V _{cal} Reading m/s	Error (m/s)	Uncertainty (%)
2.084	3.8	-0.3	2.7
4.112	4.0	0.1	1.4
6.00	6.0	0.0	1.2
8.02	8.1	0.1	0.70
10.02	10.1	0.1	0.63
11.98	12.0	0.3	0.57
13.98	14.2	0.2	0.49
16.02	16.5	0.5	0.63
16.03	16.4	0.4	0.63
12.92	13.3	0.3	0.63
11.02	11.1	0.1	0.65
9.02	9.1	0.1	0.63
7.02	7.1	0.1	0.77
5.177	5.0	-0.2	0.97
3.027	3.0	0.0	1.7
1.253	0.5	-0.5	5.4

UUC: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

Appendix 1: Instrumentation

NO	Sensor	Manufacturer	Model/Type	Calibration Date	Certificate Report Number	Range
1	Pin probe	TCSTO INC	DS322145	July 16, 2020	MA-0035-20	5 ~ 30 m/s
2	Precision Differential Pressure Meter	Zedao	DM2500	July 16, 2020	MA-0035-20	5 ~ 30 m/s
3	Air velocity transducer (hot wire)	TSI INC	8445-12	July 26, 2020	MA-0035-20	0 ~ 5 m/s
4	Temperature	Zedao	DS322145	March 30, 2021	CL-027-24	32 ~ 70 °C
5	Relative humidity	Zedao	DS322145	March 30, 2021	PH-03032021	0 ~ 100 %RH
6	Atmospheric pressure	Zedao	DS322145	March 30, 2021	BP-01032021	800 ~ 1100 hPa
7	Wind turbine	6550V	MP3300			0 ~ 55 m/s

End of certificate of calibration



CERTIFICATE OF CALIBRATION

Certificate No: WD-04072021
Page 1 of 2 pages

Measurement Item : Wind direction sensor with data logger.

Manufacturer : Data logger Novolyx.
Wind direction sensor: Novolyx.

Model/Type : Data logger: 200-WB-25DL.
Wind direction sensor: WS-02F.

Serial Number : Data logger: A4987.
Wind direction sensor: -

ID No : Data logger: RYG-F50089.
Wind direction sensor: -

Customer : AIS laboratory group (Thailand) Co.Ltd.
104 Phatthanasak 40, Phatthanasak Rd,Khwasang Suan Luang, Khut Suan Luang, Bangkok 10250 Thailand.

Environmental Condition:
The measurement was carried out in an ambient temperature of (23±3)°C and relative humidity of (60±10)%.

Measurement Method:
The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and the laser is used for axis control. The measurement were taken at 45° intervals in clockwise and counter-clockwise directions.

Note: The UUC was warmed up for 1 hour prior to the calibration being performed.

Traceability:
The measurement results are traceable to the international system of units (SI) through Certificate No: 03503-07-0045.
Certificate No: KW503/0044

Measurement Date : Jul 14, 2021.
Issued Date : Jul 14, 2021.



Performed by
☒ Mr. Sornchai Thacholad
☐ Miss Orana Wivattakanya

Approved Signatory:
Mr. Panyas Booncharoen,
Technical Support
and Calibration Manager

THIS CERTIFICATE REPORT MAY NOT BE REPRODUCED, EXCEPT IN FULL, UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.

Page 1 of 2 pages

Continuation of Certificate of Calibration Number

Certificate No: WD-04072021
Page 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment
Calibration in the range of 0 ~ 360 ° at a calibration interval of 45°.

The results of calibration and associated measurement uncertainties are reported in table below.

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC ^a Reading (°)	Error (°)	Uncertainty ±(°)
1	Clockwise	0/360	0	0	0	3.0
2		45	45	42	-3	3.0
3		90	90	88	-2	3.0
4		135	135	133	-2	3.0
5		180	180	181	1	3.0
6		225	225	228	3	3.0
7		270	270	273	3	3.0
8		315	315	318	3	3.0
9	Counter Clockwise	0/360	0	0	0	3.0
10		45	45	42	-3	3.0
11		90	90	88	-2	3.0
12		135	135	133	-2	3.0
13		180	180	181	1	3.0
14		225	225	228	3	3.0
15		270	270	273	3	3.0
16		315	315	318	3	3.0

UUC: Unit Under Calibration The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

End of Certificate of Calibration



CERTIFICATE OF CALIBRATION

Certificate No: WD-04072021
Page 1 of 2 pages

Measurement Item : Wind direction sensor with data logger.

Manufacturer : Data logger Novolyx.
Wind direction sensor: Novolyx.

Model/Type : Data logger: 200-WB-25DL.
Wind direction sensor: WS-02F.

Serial Number : Data logger: A4987.
Wind direction sensor: -

ID No : Data logger: RYG-F50089.
Wind direction sensor: -

Customer : AIS laboratory group (Thailand) Co.Ltd.
104 Phatthanasak 40, Phatthanasak Rd,Khwasang Suan Luang, Khut Suan Luang, Bangkok 10250 Thailand.

Measurement Method:
The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and the laser is used for axis control. The measurement were taken at 45° intervals in clockwise and counter-clockwise directions.

Note: The UUC was warmed up for 1 hour prior to the calibration being performed.

Traceability:
The measurement results are traceable to the international system of units (SI) through Certificate No: 03503-07-0045.
Certificate No: KW503/0044

Measurement Date : July 14, 2021.
Issued Date : July 14, 2021.



Performed by
☒ Mr. Sornchai Thacholad
☐ Miss Orana Wivattakanya

Approved Signatory:
Mr. Panyas Booncharoen,
Technical Support
and Calibration Manager

THIS CERTIFICATE REPORT MAY NOT BE REPRODUCED, EXCEPT IN FULL, UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.

Continuation of Certificate of Calibration Number

Certificate No: WS 01032021
Page: 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment

Calibration in the range: 0 - 180° at a calibration interval of 45°

The maximum tolerances and absolute measurement uncertainties are indicated in table below

No.2	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UNC Reading (°)	Error (°)	Uncertainty (°)
1	Clockwise	0/225	23.0	23.0	0	1.0
2		45	45	44	1	1.0
3		90	90	90	0	1.0
4		135	136	135	1	1.0
5		180	180	179	1	1.0
6	Counter clockwise	225	224	225	1	1.0
7		180	180	180	0	1.0
8		135	134	135	1	1.0
9		90	90	90	0	1.0
10		45	45	45	0	1.0

UNC: Uncertainty Coefficient. The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor of 2 (k=2), which corresponds to a confidence of approximately 95%.

End of Certificate of Calibration



CERTIFICATE OF CALIBRATION

Certificate No: WS 01032021
Page: 1 of 2 pages

Measurement Item: Cup anemometer with data logger

Manufacturer: Data logger Nivayak
Cup anemometer Nivayak

Model/Type: Data logger 200 WS25LB
Cup anemometer WS25LB

Serial Number: Data logger 45375
Cup anemometer

ID No: Data logger P46/130413
Cup anemometer

Customer: A.S. Technology group (Thailand) Co., Ltd.
104 Prachasarak 40, Petchkasem Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand

Test Conditions: Wind tunnel cross test section area: 900 cm²
Anemometer frontal area: 100 cm²
Thickness of measuring edge: 6mm
Protrusion height of test object: 3111 mm

Test Conditions: Air temperature: 24.5 ±0.5 °C
Air pressure: 1027.4 ±0.4 hPa
Relative air humidity: 52.4 ±0.3 RH

Calibration Procedure: Calibration was carried out based on:
ISO 61400-12-1 (Ed.3: 2015) Power Performance Measurements of Electricity Producing Wind Turbines
MCA/NET Anemometer Calibration Procedure - Version 2: 2020

Traceability: The calibration is based on the reference to national standard, which is part of the list of measurements according to the international system of units (SI) through National Institute of Metrology (Thailand NMI).

Measurement Date: Jul 29, 2021
Issued Date: Jul 29, 2021

REVIEW BY: N. Bantik
APPROVED BY: [Signature]
NEXT CAL. DATE: 24/6/22

Calibrated by:
Jirana Tee, Jirana Tee
Walthapra, Bangkok, Thailand



Approved Signatory: [Signature]
Mr. Panya Boonchuen
Technical Support
and Calibration Manager

THIS CERTIFICATE MAY NOT BE REPRODUCED EXCEPT IN FULL, UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.

Continuation of Certificate of Calibration Number

Certificate No: WS 13072021
Page: 1 of 2 pages

Result of calibration: ☒ Without adjustment ☐ With adjustment

Calibration in the range: 0 - 180° at a calibration interval of 45°

The maximum tolerances and absolute measurement uncertainties are indicated in table below

No.2	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UNC Reading (°)	Error (°)	Uncertainty (°)
1	Clockwise	0/225	23.0	23.0	0	1.0
2		45	45	44	1	1.0
3		90	90	90	0	1.0
4		135	136	135	1	1.0
5		180	180	179	1	1.0
6	Counter clockwise	225	224	225	1	1.0
7		180	180	180	0	1.0
8		135	134	135	1	1.0
9		90	90	90	0	1.0
10		45	45	45	0	1.0

UNC: Uncertainty Coefficient

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor of 2 (k=2), which corresponds to a confidence of approximately 95%.

Appendix: Turning direction

ID	Device	Manufacturer	Model/Type	Calibration Date	Setting/Report Number	Range
1	Wind speed sensor (Wind)	HOBO	HOBO-2301	2021/07/29	2021/07/29	0-10m/s
2	Wind speed sensor (Wind)	HOBO	HOBO-2301	2021/07/29	2021/07/29	0-10m/s
3	Wind speed sensor (Wind)	HOBO	HOBO-2301	2021/07/29	2021/07/29	0-10m/s
4	Wind speed sensor (Wind)	HOBO	HOBO-2301	2021/07/29	2021/07/29	0-10m/s
5	Wind speed sensor (Wind)	HOBO	HOBO-2301	2021/07/29	2021/07/29	0-10m/s

End of certificate of calibration



CERTIFICATE OF CALIBRATION

Certificate No: WS 13072021
Page: 1 of 2 pages

Measurement Item: Cup anemometer with data logger

Manufacturer: Data logger Nivayak
Cup anemometer Nivayak

Model/Type: Data logger 200 WS25LB
Cup anemometer WS25LB

Serial Number: Data logger 45375
Cup anemometer

ID No: Data logger P46/130413
Cup anemometer

Customer: A.S. Technology group (Thailand) Co., Ltd.
104 Prachasarak 40, Petchkasem Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand

Test Conditions: Wind tunnel cross test section area: 900 cm²
Anemometer frontal area: 100 cm²
Thickness of measuring edge: 6mm
Protrusion height of test object: 3111 mm

Test Conditions: Air temperature: 24.5 ±0.5 °C
Air pressure: 1027.4 ±0.4 hPa
Relative air humidity: 52.4 ±0.3 RH

Calibration Procedure: Calibration was carried out based on:
ISO 61400-12-1 (Ed.3: 2015) Power Performance Measurements of Electricity Producing Wind Turbines
MCA/NET Anemometer Calibration Procedure - Version 2: 2020

Traceability: The calibration is based on the reference to national standard, which is part of the list of measurements according to the international system of units (SI) through National Institute of Metrology (Thailand NMI).

Measurement Date: Jul 29, 2021
Issued Date: Jul 29, 2021

REVIEW BY: [Signature]
APPROVED BY: [Signature]
NEXT CAL. DATE: 24/6/22

Calibrated by:
☒ Mr. Panya Tee, Jirana Tee
☐ Mrs. Panya Tee, Jirana Tee



Approved Signatory: [Signature]
Mr. Panya Boonchuen
Technical Support
and Calibration Manager

THIS CERTIFICATE MAY NOT BE REPRODUCED EXCEPT IN FULL, UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.

Continuation of Certificate of Calibration Number

Certificate No: WS-13072021
Page 2 of 2 Pages

Result of calibration: ☒ Without adjustment ☐ With adjustment

Calibration in the range of 1 - 16 m/s at a calibration interval of 1 m/s

The results of calibration and associated measurement uncertainties are reported in the table below

V _{avg} Reading m/s	V _{unc} Reading m/s	Error (m/s)	Uncertainty (%)
2.067	2.0	-0.1	2.4
4.188	4.1	0.0	1.2
6.305	6.1	0.1	0.97
7.99	8.0	0.0	0.84
10.00	10.1	0.1	0.69
12.03	12.2	0.2	0.72
13.69	14.3	0.3	0.47
15.98	16.4	0.4	0.35
16.03	16.3	0.3	0.38
12.99	13.1	0.1	0.69
11.01	11.1	0.1	0.67
9.01	9.0	0.0	0.67
6.93	7.1	0.1	0.81
5.177	5.1	-0.1	0.97
2.972	3.1	0.1	1.6
1.044	0.9	-0.1	2.3

UUC* End of Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

Appendix 1: Instrumentations

NO	Sensor	Manufacturer	Model/Type	Calibration Date	Certificate Report Number	Range
1	Pressure	KOYO INC	CS352143	July 16, 2020	MX 003520	5 - 20 m/s
2	Precision Differential Pressure Meter	Zogas	DMV2000	July 16, 2020	MX 003520	5 - 30 m/s
3	Air velocity (transducer not used)	16 INC.	8455-12	July 20, 2020	MX 003520	0 - 5 m/s
4	Temperature	Zogas	DSH-10	March 30, 2021	DL-027566	-30 - 120°C
5	Relative humidity	Zogas	DSH-10	March 30, 2021	DL-027566	0 - 100 %RH
6	Atmospheric pressure	Zogas	DSH-10	March 30, 2021	DL-027566	500 - 1100 mPa
7	Wind turbine	ELCON	MP3302			0 - 60 m/s

End of certificate of calibration



CERTIFICATE OF CALIBRATION

Certificate No: WD-13072021
Page 1 of 2 pages

Measurement Item : Wind direction sensor with data logger.

Manufacturer : Data logger: Novaynx
: Wind direction sensor: Novaynx

Model/Type : Data logger: 200-WS-251B
: Wind direction sensor: WS-02P

Serial Number : Data logger: A55375
: Wind direction sensor: -

ID No : Data logger: R10_F80413
: Wind direction sensor: -

Customer : ALS laboratory group (Thailand) Co.Ltd
: 104 Prathanan 40, Prathanan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand

Environmental Condition:
The measurement was carried out in an ambient temperature of (23±3)°C, and relative humidity of (40±10)%

Measurement Method:
The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and line laser is used for axis control. The measurement were taken at 45° intervals in clockwise and counterclockwise directions.

Note: The UUC was warmed up for 1 hour prior to the calibration being performed

Traceability:
The measurement results are traceable to the international system of units (SI) through Certificate No: CG563-07/0045, Certificate No: KW563/0044

Measurement Date : Jul. 29, 2021
Issued Date : Jul. 29, 2021

Performed by
☒ Mr. Sorawit Thachaid
☐ Ms. Oranrat Wistatayaya



Approved Signatory: _____
Mr. Parinya Booncharoen
Technical Support
and Calibration Manager

THIS CERTIFICATE REPORT MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY

Continuation of Certificate of Calibration Number

Certificate No: WD-13072021
Page 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment

Calibration in the range of 0 - 360 ° at a calibration interval of 45°

The results of calibration and associated measurement uncertainties are reported in table below

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC* Reading (°)	Error (°)	Uncertainty ±(°)
1		0/360	360	359	-1	3.0
2		45	45	42	-3	3.0
3		90	90	87	-3	3.0
4		135	135	134	-1	3.0
5		180	180	181	1	3.0
6		225	225	228	3	3.0
7		270	270	273	3	3.0
8		315	315	318	3	3.0
9		0/360	360	359	-1	3.0
10		45	45	42	-3	3.0
11		90	90	87	-3	3.0
12		135	135	134	-1	3.0
13	Counter Clockwise	180	180	181	1	3.0
14		225	225	228	3	3.0
15		270	270	273	3	3.0
16		315	315	318	3	3.0

UUC* Unit Under Calibration The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

End of Certificate of Calibration



CERTIFICATE OF CALIBRATION

Certificate No: WS 01022021
Page 1 of 2 pages

Measurement Item : Cup anemometer with data logger

Manufacturer : Data logger: Novaynx
: Cup anemometer: Novaynx

Model/Type : Data logger: WS-25CL
: Cup anemometer: WS-02P

Serial Number : Data logger: A4461
: Cup anemometer: -

ID No : Data logger: RMX_F80141
: Cup anemometer: -

Customer : ALS laboratory group (Thailand) Co.Ltd
: 104 Prathanan 40, Prathanan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand

Test Conditions : Wind tunnel cross test section area : 900 cm²
: Anemometer frontal area : 100 cm²
: Diameter of mounting pipe : 1 mm
: Blockage ratio of test object : 0.111 [-]

Test Conditions : Air temperature : 23.7 ±0.8 °C
: Air pressure : 1010.3 ±0.4 hPa
: Relative air humidity : 53.7 ±3.5 %RH

Calibration Procedure : Calibration was carried out base on
ISO 61400-12-1 ED1: 2015- Power Performance Measurements of Electricity Producing Wind Turbines
MCA/NET Anemometer Calibration Procedure - Version 2: 2009.

Traceability : The calibration documents the traceable to national standard which realize the unit of measurements according to the international system of units (SI) through National Institute of Metrology (NIMT)

Measurement Date : Jun 07, 2021
Issued Date : Jun 07, 2021

Calibrated by
☒ Mr. Sorawit Thachaid
☐ Ms. Oranrat Wistatayaya



Approved Signatory: _____
Mr. Parinya Booncharoen
Technical Support
and Calibration Manager

THIS CERTIFICATE MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY

Continuation of Certificate of Calibration Number

Certificate No: WB-01062021
Page 2 of 2 Pages

Result of calibration: ☒ Without adjustment ☐ With adjustment
Calibration in the range of 1 - 16 m/s at a calibration interval of 1 m/s

The results of calibration and associated measurement uncertainties are reported in the table below.

V _{ref} Reading m/s	V _{ref} Reading m/s	Error (m/s)	Uncertainty (%)
2.055	2.0	0.1	2.9
4.124	4.0	-0.1	1.2
5.59	6.0	0.2	10.1
8.03	8.0	0.2	0.74
9.59	10.1	0.1	0.60
11.66	12.2	0.2	0.67
14.02	14.4	0.4	0.45
15.05	16.2	0.6	0.35
15.01	15.3	0.3	2.8
12.99	15.3	0.3	0.41
10.99	11.2	0.2	0.53
9.01	9.3	0.3	1.2
7.05	7.0	0.0	0.77
5.121	5.0	-0.1	0.66
3.048	3.0	0.0	1.6
1.058	1.2	0.1	5.3

UUC*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

Appendix II: Instrumentations

NO	Sensor	Manufacturer	Model/Type	Calibration Date	Certificate Report Number	Range
1	Wind static	TEDEL INC.	00302146	July 16, 2020	MA-0035-20	5 - 30 m/s
2	Pressure Differential Pressure Meter	Zepal	DFM2500	July 16, 2020	MA-0035-20	5 - 30 m/s
3	Air velocity transducer (hot wire)	TS INC.	8455 12	July 20, 2020	MA-0035MA-20	0 - 5 m/s
4	Temperature	Zepal	DDH-TMP	March 30, 2021	C-02764	30 - 70°C
5	Relative humidity	Zepal	DDH-TMP	March 30, 2021	PH-C0302021	0 - 100 %RH
6	Atmospheric pressure	Zepal	DDH-TMP	March 30, 2021	BP-01032021	500 - 1100 mPa
7	Wind tunnel	ESCOM	MP3300			0 - 20 m/s

End of certificate of calibration



CERTIFICATE OF CALIBRATION

Certificate No: WD-01062021
Page 1 of 2 pages

Measurement Item : Wind direction sensor with data logger.

Manufacturer : Data logger: Novolyx
: Wind direction sensor: Novolyx

Model/Type : Data logger: WS-250L
: Wind direction sensor: WS-02P

Serial Number : Data logger: A4481.
: Wind direction sensor: -

ID No : Data logger: BKH_RS0141.
: Cup anemometer: -

Customer : AIS Laboratory group (Thailand) Co.Ltd.
104 Phathanakarn 40, Phathanakarn Rd,Khwaeng Suan Luang, Khet Suan Luang,Bangkok 10250 Thailand

Environmental Condition:
The measurement was carried out in an ambient temperature of (23±3)°C and relative humidity of (40±10)%,

Measurement Method.

The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and the laser is used for data control. The measurement were taken at 45° intervals in clockwise and counterclockwise directions.

Note: The UUC was warmed up for 1 hour prior to the calibration being performed

Traceability:

The measurement results are traceable to the international system of units (SI) through Certificate No: C0563-07-0045, Certificate No: RW563/0044

Measurement Date : Jun 07, 2021.
Issued Date : Jun 07, 2021.

Performed by
☒ Mr. Sorakrit Thongmalat
☐ Miss Orathai Wikanawitaya



Approved Signatory:
Mr. Panya Booncharoen
Technical Support
and Calibration Manager

THIS CERTIFICATE REPORT MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.

Continuation of Certificate of Calibration Number

Certificate No: WD-01062021
Page 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment
Calibration in the range of 0 - 360 ° at a calibration interval of 45°.

The results of calibration and associated measurement uncertainties are reported in table below.

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC* Reading (°)	Error (°)	Uncertainty ±(°)
1	Clockwise	0/360	0	0	0	3.0
2		45	45	42	3	3.0
3		90	90	90	0	3.0
4		135	135	136	1	3.0
5		180	180	182	2	3.0
6		225	225	227	2	3.0
7		270	270	273	3	3.0
8		315	315	314	1	3.0
9	Counter Clockwise	0/360	0	0	0	3.0
10		45	45	42	3	3.0
11		90	90	90	0	3.0
12		135	135	136	1	3.0
13		180	180	182	2	3.0
14		225	225	227	2	3.0
15		270	270	273	3	3.0
16		315	315	314	1	3.0

UUC*: Unit Under Calibration The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

End of Certificate of Calibration



CERTIFICATE OF CALIBRATION

Certificate No: WB-14072021
Page 1 of 2 pages

Measurement Item : Cup anemometer with data logger

Manufacturer : Data logger: Novolyx
: Cup anemometer: Novolyx

Model/Type : Data logger: 200-WS-25L6
: Cup anemometer: WS-02P

Serial Number : Data logger: A4376
: Cup anemometer: -

ID No : Data logger: BKH_RS0141
: Cup anemometer: -

Customer : AIS Laboratory group (Thailand) Co.Ltd.
104 Phathanakarn 40, Phathanakarn Rd,Khwaeng Suan Luang, Khet Suan Luang,Bangkok 10250 Thailand

Test Conditions : Wind tunnel cross test section area : 400 cm²
: Anemometer probe area : 100 cm²
: Diameter of mounting pipe : mm
: Blockage ratio of test object : 0.11

Test Conditions : Air temperature : 26.0 ±0.8 °C
: Air pressure : 1035.5 ±0.4 hPa
: Relative air humidity : 51.4 ±5.6 %RH

Calibration Procedure : Calibration was carried out base on:
ISO 31450-12-1 Part 1: 2001 Power Performance Measurements of Factory-Producing Wind Turbines
IEC61400-12-1 Part 1: 2001 Power Performance Measurements of Factory-Producing Wind Turbines
IEC61400-12-1 Part 1: 2001 Power Performance Measurements of Factory-Producing Wind Turbines

Traceability : The calibration documents are traceable to national standard, which realize the unit of measurement according to the international system of units (SI) through National Institute of Metrology (NIM) of Thailand.

Measurement Date : Jun 29, 2021.
Issued Date : Jun 29, 2021.

Calibrated by
☒ Mr. Sorakrit Thongmalat
☐ Miss Orathai Wikanawitaya



Approved Signatory:
Mr. Panya Booncharoen
Technical Support
and Calibration Manager

THIS CERTIFICATE MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.

Continuation of Certificate of Calibration Number

Certificate No: WS-14072021
Page 2 of 2 Pages

Result of calibration: ☒ Without adjustment ☐ With adjustment
Calibration in the range of 1 ~ 16 m/s at a calibration interval of 1 m/s.
The results of calibration and associated measurement uncertainties are reported in the table below.

V _{ref} Reading m/s	V _{ref} Reading m/s	Error (m/s)	Uncertainty (m/s)
2057	1.8	-0.3	3.1
4136	4.0	-0.1	1.3
602	6.0	0.0	2.1
769	8.0	0.0	0.74
1000	10.1	0.1	0.67
1159	12.3	0.0	0.72
1358	14.2	0.2	0.48
1598	16.2	0.2	0.77
1499	15.2	0.2	0.49
1300	13.1	0.1	0.52
1101	11.0	0.0	0.94
901	9.0	0.0	0.81
699	7.0	0.0	2.0
5189	5.1	-0.1	0.96
2987	3.0	0.0	2.0
1054	0.8	-0.2	0.5

UUC: Unit Under Calibration
The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%.

Appendix 1: Instrumentation

NO	Sensor	Manufacturer	Model/Type	Calibration Date	Certificate Report Number	Range
1	Hot plate	TESO NO.	06352145	July 16, 2020	MR-0035-20	5 ~ 30 m/s
2	Pressure Differential Pressure Meter	Zepac	DP42250	July 16, 2020	MR-0035-20	5 ~ 30 m/s
3	Air velocity transducer (hot wire)	TSI INC.	8445-12	July 20, 2020	MR-003644-20	0 ~ 5 m/s
4	Temperature	Zepac	05A-TM	March 30, 2021	PL-027-24	-32 ~ 70 °C
5	Relative humidity	Zepac	05A-TM	March 30, 2021	PL-027-24	0 ~ 100 %RH
6	Airspeed pressure	Zepac	05A-TM	March 30, 2021	PL-01C32021	500 ~ 1100 m/s
7	Wind tunnel	COLOM	MP3302	-	-	0 ~ 25 m/s

End of certificate of calibration



CERTIFICATE OF CALIBRATION

Certificate No: WD-14072021
Page 1 of 2 pages

Measurement Item : Wind direction sensor with data logger.
Manufacturer : Data logger: Novayaex.
Wind direction sensor: Novayaex.
Model/Type : Data logger: 200-WS-26LB.
Wind direction sensor: WS-02P.
Serial Number : Data logger: A5376.
Wind direction sensor: -.
ID No : Data logger: RYG_F50414.
Wind direction sensor: -.
Customer : ALS laboratory group (Thailand) Co., Ltd.
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luing, Khet Suan Luing, Bangkok 10250 Thailand.

Environmental Condition:
The measurement was carried out in an ambient temperature of (23±3)°C and relative humidity of (40±10)%.

Measurement Method:
The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and data logger is used for axis control. The measurement were taken at 45° intervals in clockwise and counterclockwise directions.

Note: This UUC was warmed up for 1 hour prior to the calibration being performed.

Traceability:
The measurement results are traceable to the international system of units (SI) through Certificate No: 00543-07-0045.
Certificate No: RWS53/0044.

Measurement Date : Jul. 29, 2021.
Issued Date : Jul. 29, 2021.

Performed by
☒ Mr. Soravit Thairattal
☐ Miss Oranai Wisetkittaya



Approved Signatory:
Mr. Panya Booncharoen
Technical Support
and Calibration Manager

THIS CERTIFICATE REPORT MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.

Continuation of Certificate of Calibration Number

Certificate No: WD-14072021
Page 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment
Calibration in the range of 0 ~ 360 ° at a calibration interval of 45°.
The results of calibration and associated measurement uncertainties are reported in table below.

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC* Reading (°)	Error (°)	Uncertainty ±(°)
1	Clockwise	0/360	360	359	-1	3.0
2		45	46	43	-2	3.0
3		90	90	87	-3	3.0
4		135	135	132	-3	3.0
5		180	180	179	-1	3.0
6		225	225	226	1	3.0
7		270	270	273	3	3.0
8		315	315	318	3	3.0
9	Counter Clockwise	0/360	360	369	+1	3.0
10		45	46	43	-2	3.0
11		90	90	87	-3	3.0
12		135	135	132	-3	3.0
13		180	180	179	-1	3.0
14		225	225	226	1	3.0
15		270	270	273	3	3.0
16		315	315	318	3	3.0

UUC: Unit Under Calibration The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%.


End of Certificate of Calibration



CERTIFICATE OF CALIBRATION

Certificate No: WS-09072021
Page 1 of 2 pages

Measurement Item : Cup anemometer with data logger
Manufacturer : Data logger: Novayaex.
Cup anemometer: Novayaex.
Model/Type : Data logger: 110-WS-250L-0.
Cup anemometer: WS-02P.
Serial Number : Data logger: A5559.
Cup anemometer: WS-014.
ID No : Data logger: -.
Cup anemometer: -.
Customer : ALS laboratory group (Thailand) Co., Ltd.
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luing, Khet Suan Luing, Bangkok 10250 Thailand.
Test Conditions : Wind tunnel cross test section area : 900 m²
Anemometer frame area : 100 m²
Diameter of mounting pole : 1 mm
Blockage ratio of test object : 0.111 %
Test Conditions : Air temperature : 23.4 ±0.8 °C
Air pressure : 1006.2 ±0.4 hPa
Relative air humidity : 59.3 ±3.5 %RH
Calibration Procedure : Calibration was carried out base on
ISO 91423-12-1 (2013) Power Performance Measurements of Electrod, Producing Wind
Turbines.
NAC/NCT Anemometer Calibration Procedure - Version 2: 2019
Traceability : This calibration documents the traceable to national standards, which relate the unit of
measurements according to the international system of units (SI) through National Institute of
Metrology, Thailand (NIMT).
Measurement Date : Jul. 14, 2021
Issued Date : Jul. 15, 2021



Calibrated by
☒ Mr. Soravit Thairattal
☐ Miss Oranai Wisetkittaya



Approved Signatory:
Mr. Panya Booncharoen
Technical Support
and Calibration Manager

THIS CERTIFICATE MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.

Continuation of Certificate of Calibration Number

Certificate No: WS-08072021
Page 2 of 2 Pages

Result of calibration: ☒ Without adjustment ☐ With adjustment
Calibration in the range of 1 - 16 m/s at a calibration interval of 1 m/s.
The results of calibration and associated measurement uncertainties are reported in the table below.

V _{ref} Reading m/s	V _{ref} Reading m/s	Error m/s	Uncertainty (m)
2.069	1.9	0.2	2.5
4.122	4.0	0.1	1.2
6.02	6.0	0.0	0.55
7.97	8.0	0.0	0.84
9.98	10.0	0.0	0.59
12.02	12.1	0.1	0.47
13.99	14.2	0.2	0.46
15.98	16.2	0.2	0.55
14.99	15.2	0.2	0.39
13.00	13.1	0.1	0.45
11.02	11.1	0.1	0.53
8.99	9.0	0.0	0.70
6.98	7.0	0.0	0.95
5.112	5.0	0.1	1.2
2.975	3.0	0.0	1.5
1.023	0.9	0.1	5.3

UUC: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

Appendix 1: Instrumentations

NO	Sensor	Manufacturer	Model/Type	Calibration Date	Certificate Report Number	Range
1	Plot data	TEBTO INC	06359145	July 16, 2020	MA-0035-20	5 - 30 m/s
2	Precision Differential Pressure Meter	Zepac	DPV2500	July 16, 2020	MA-0035-20	5 - 30 m/s
3	Air velocity transducer (hot wire)	TD INC	8455-12	July 20, 2020	MA-0034A-20	0 - 5 m/s
4	Temperature	Zepac	DS9TH7	March 30, 2021	CL-007-24	-30 - 70°C
5	Relative humidity	Zepac	DS9TH7	March 30, 2021	RH-03030201	0 - 100 %RH
6	Atmospheric pressure	Zepac	DS9TH7	March 30, 2021	BP-01030201	500 - 1100 hPa
7	Wind Lure	0050W	MN3600			0 - 50 m/s

End of certificate of calibration



CERTIFICATE OF CALIBRATION

Certificate No: WS-08072021
Page 1 of 2 pages

Measurement Item : Wind direction sensor with data logger

Manufacturer : Data logger: Novalynx
: Wind direction sensor: Novalynx

Model/Type : Data logger: 110-WS-250U-0
: Wind direction sensor: WS-05P

Serial Number : Data logger: A5660
: Wind direction sensor: WS0-D14

ID No : Data logger :
: Wind direction sensor :

Customer : ALS laboratory group (Thailand) Co.,Ltd.
104 Phatthanakan 40, Phatthanakan Rd,Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand

Environmental Condition:
The measurement was carried out in an ambient temperature of 23±3°C and relative humidity of 40±10%.

Measurement Method:
The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and the laser is used for axis control. The measurement were taken at 45° intervals in clockwise and counterclockwise directions.

Note: The UUC was warmed up for 1 hour prior to the calibration being performed.

Traceability:
The measurement results are traceable to the international system of units (SI) through Certificate No: CC563-07-0045, Certificate No: KWS63/G044.

Measurement Date : July 14, 2021.
Issued Date : July 15, 2021.



Performed by
☒ Mr. Sravit Thachad
☐ Miss Orathai Waiwattaya

Approved Signatory: _____
Mr. Panya Booncharoen
Technical Support
and Calibration Manager

THIS CERTIFICATE REPORT MAY NOT BE REPRODUCED EXCEPT IN FULL, UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.

Continuation of Certificate of Calibration Number

Certificate No: WS-08072021
Pages 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment.
Calibration in the range of 0 - 360 ° at a calibration interval of 45°.
The results of calibration and associated measurement uncertainties are reported in the table below.

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC* Reading (°)	Error (°)	Uncertainty ±(°)
1	Clockwise	0/360	360	359	-1	3.0
2		45	45	42	-3	3.0
3		90	90	87	-3	3.0
4		135	135	134	-1	3.0
5		180	180	182	2	3.0
6		225	225	228	3	3.0
7		270	270	273	3	3.0
8	Counter Clockwise	315	315	318	3	3.0
9		0/360	360	359	-1	3.0
10		45	45	42	-3	3.0
11		90	90	87	-3	3.0
12		135	135	134	-1	3.0
13		180	180	182	2	3.0
14		225	225	228	3	3.0
15		270	270	273	3	3.0
16		315	315	318	3	3.0

UUC: Unit Under Calibration The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

End of Certificate of Calibration



CERTIFICATE OF CALIBRATION

Certificate No.: CL-050 64
Page 1 of 2

Equipment Name : Data Logger with Temperature
Sensor
Manufacturer : Novalynx
Model : 110-WS-25
Serial No. : A5660
ID No. : .

Customer
Name : ALS laboratory group (Thailand) Co.,Ltd.
Address : 104 Phatthanakan 40, Phatthanakan Rd.,Khwaeng Suan Luang, Khet Suan Luang,Bangkok 10250 Thailand.

Received date : 12 JUL 2021
Calibration date : 13 JUL 2021
Issue date : 13 JUL 2021

Reference Used During Calibration
1.Standard Temperature Probe Model : STS-100 A500,
Serial No. : 667682 09, Due date : 25 Mar 2022
2.Digital Temperature Indicator Model : DTI-1000-A MK II, Serial No: 671407-00591 Due date : 04 June 2022

Calibration Condition
Temperature : (23±3)°C
Relative Humidity : (55±15)%

Calibration Procedure
The temperature calibration was done by In House calibration method as WI-CL-001 according to comparison method with standard digital temperature indicator and standard temperature probe. The temperature scale use was based on ITS-90.

Traceability
The measurement results are traceable to the international system of units (SI) through National Institute of Metrology Thailand (NIMT) Certificate number: IT-0036-21, Certificate number : ER-0032-21

Calibrated by
☐ Mr. Sravit Thachad
☒ Miss Orathai Waiwattaya



Approved Signatory: _____
Mr. Panya Booncharoen
Technical Support
and Calibration Manager

THIS CERTIFICATE MAY NOT BE REPRODUCED EXCEPT IN FULL, UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.



63/14-15,67/35-36, Soi Petchkasem 7,7/1, Petchkasem Rd,
Walthapra, Bangkokyai,Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranotee.com



Certificate No.: CL-050-64
Page 2 of 2



63/14-15,67/35-36, Soi Petchkasem 7,7/1, Petchkasem Rd,
Walthapra, Bangkokyai,Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranotee.com

CALIBRATION REPORT

Calibration No.: RH-02072021
Page 1 of 1 Pages

Result of Calibration: ☒ Without Adjustment ☐ With Adjustment
Calibration Range: 20°C - 40°C

Function:
This equipment was connected with temperature sensor Model: HMP60 S/N: T0210901
Dimension: Diameter 12mm. Length 80 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
60	20.050	19.7	-0.3	0.060
60	24.975	24.5	-0.4	0.13
60	29.864	29.5	-0.4	0.080
60	34.829	34.3	-0.5	0.090
60	39.831	39.4	-0.5	0.95

UUC*: Unit Under Calibration
The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

☆ End of Certificate ☆

Measurement Item: Relative humidity with data logger.
Manufacturer: Data logger: Navaynuk
Relative humidity sensor: Navaynuk.
Model/Type: Data logger: H10WS-26Du-D.
Relative humidity sensor: HMP60.
Serial Number: Data logger: A5660.
Relative humidity sensor: T0210901.
ID No: Data logger: -
Relative humidity sensor: -
Customer: ALS laboratory group (Thailand) co., Ltd.
104 Phatthanakan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand.

Environmental Condition:
The measurement was carried out in an ambient temperature of (25.3)°C, and relative humidity of (50.1)5%.

Measurement Method:
The Relative humidity with data logger, Unit Under Calibration (UUC) was calibrated by comparison method with the equilibrium of standard salt solution: CH₃COOK; Potassium Acetate, Mg(NO₃)₂ Magnesium Nitrate, KCl, Potassium Chloride to determine the errors.

Measurement Date: Jul 14, 2021
Issued Date: Jul 14, 2021

Measurement Results:

The results of calibration are reported in table below.

Standard salt solution.	Standard (RH%)	UUC Reading	Error
CH ₃ COOK Potassium Acetate	22.51	22.2	-0.3
Mg(NO ₃) ₂ Magnesium Nitrate	52.89	52.3	-0.6
KCl Potassium Chloride	84.34	83.8	-0.5

Performed by
☒ Mr. Sorak Thachad
☐ Miss Orath Waiwattaya



Approved Signature: *Orath Waiwattaya*
Mr. Panyia Booncharoen
Technical Support
and Calibration Manager

THIS CALIBRATION REPORT MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THIS LABORATORY.

CONSOLE CONTROL UNIT CALIBRATION TEST REPORT

Barometric Pressure (mm.Hg): 755
Relative Humidity (%): 70.0
Temperature (°C): 30.0

Reference Dry Gas Meter Data

Serial No.: 1507009
Model No.: SK25EXSR-OC6
Correction Factor (Y): 1.0050
Next Calibration Date: 7 Oct 22

ΔH (mm.H ₂ O)	Θ Minutes	Reference Dry Gas Meter Calibration					Console Control: Drygas Meter					Dry Gas Meter Correction Factor (Y)	Office Calibration Factor ΔSig
		Vr (liters)			Tr (°C)	Ti (°C)			To (°C)	Avg.Tr (°C)			
		Final	Initial	Total		Final	Initial	Total					
15	11.28	150.00	0.00	150.00	30.0	284586.6	284222.0	146.60	30.0	30.0	30.0	1.0278	39.1530
25	8.83	150.00	0.00	150.00	31.0	264526.8	264390.0	146.80	31.0	31.0	31.0	1.0254	40.1496
50	6.32	150.00	0.00	150.00	31.0	264495.2	264339.0	145.20	31.0	31.0	31.0	1.0271	41.1361
100	4.45	150.00	0.00	150.00	31.0	264495.0	264339.0	146.00	31.0	31.0	31.0	1.0236	40.7897
150	3.07	150.00	0.00	150.00	31.0	265059.0	264912.0	147.00	31.0	31.0	31.0	1.0118	41.0143
												Avg.	40.5743

Y Rule of reading of reference to dry gas meter: tolerance for individual values ± 0.02 from average.

ΔHg: Office pressure differential that equates to 21.24 in of air @ 25°C and 760 mm of mercury. mmH₂O: tolerance for individual values ± 5.08 from average.

Procedure: 40 CFR 60 APP A METH. SEC 5.3.3.7

Calibrated by: *Saksit Phaisangphit*

(Mr. Saksit Phaisangphit)

Field Scientist(4)

Approved by: *Nathaporn Jangwarewong*

(Mr. Nathaporn Jangwarewong)

Field Specialist(1)

FORM NO. F-0024 REVISION NO. 1 ISSUE DATE 03/19



DIGITAL TEMPERATURE CALIBRATION DATA SHEET

Calibration Date : 12 Jul 22		Ambient Temperature (°C) : 30		
Calibration sheet No. : C-120722-BKK_FS0528		Relative Humidity (%) : 70		
Digital Temperature ID BKK_FS0508		Reference Temperature ID : BKK_FS0609		
Console Serial No. : 1503017		Serial No. : 7688004		
Console Model : XC-572-V		Model : FLUKE 714		
		Next Calibrate : 26 Jul 23		
Location	Reference Temperature °C	Digital Temperature °C	Error °C	Remark
Stack	0	2	2	
	25	24	-1	
	50	51	1	
	100	103	3	
	150	151	1	
	200	202	2	
	250	251	1	
	300	301	1	
	500	503	3	
	1000	1001	1	
Probe	1200	1202	2	
	100	101	1	
	125	126	1	
Oven	150	153	3	
	100	101	1	
	125	126	1	
Filler	150	151	1	
	100	102	2	
	125	125	0	
Exit	150	152	2	
	0	0	0	
	10	10	0	
Meter	20	20	0	
	0	0	0	
	25	25	0	
AUX	50	50	0	
	0	0	0	
	25	25	0	
	50	50	0	

Calibrated by: *Saksit Phaisangphit*

(Mr. Saksit Phaisangphit)

Field Scientist (4)

Approved by: *Nathaporn Jangwarewong*

(Mr. Nathaporn Jangwarewong)

Field Specialist(1)

Form 281-049 012/05/02



Pitot Tube Calibration Data

Pitot Tube Identification Number : BKK_FS0531 Calibration Date : 12 Jul 22
Lab test duct Number : 258-1-13-01 Standard Pitot ID : BKK_FS0441
Calibration Sheet No. : C-120722-BKK_FS0531 Cp Standard : 0.99

Type S Pitot Tube Coefficient Data					
	Type s pitot tube Leg A,B	Standard pitot tube (ΔP, mm.H ₂ O)	Type s pitot tube (ΔP, mm.H ₂ O)	Cp (s) Leg A	Cp (s) Leg B
Test 1	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
Test 2	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
Test 3	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
Cp				0.842	0.842

$$Cp(S) = Cp = \sqrt{\frac{\Delta P(s)}{\Delta P(s)}}$$

$$[Cp(A) - Cp(B)] \text{ must BE } \leq 0.01$$

$$\text{Average deviation(A or B)} = \frac{\sum [Cp(s) - Cp(A \text{ or } B)]}{3} \text{ must BE } \leq 0.01$$

Calibrated by Saksit Phaisanphisit Approved by Nattapon Jengwareewong
(Mr Saksit Phaisanphisit) (Mr Nattapon Jengwareewong)
Field Scientist (4) Field Specialist(1)

Form ZB1-046 (04/03/02)



Pitot Tube Calibration Data

Pitot Tube Identification Number : BKK_FS0532 Calibration Date : 12 Jul 22
Lab test duct Number : 258-1-13-01 Standard Pitot ID : BKK_FS0441
Calibration Sheet No. : C-120722-BKK_FS0532 Cp Standard : 0.99

Type S Pitot Tube Coefficient Data					
	Type s pitot tube Leg A,B	Standard pitot tube (ΔP, mm.H ₂ O)	Type s pitot tube (ΔP, mm.H ₂ O)	Cp (s) Leg A	Cp (s) Leg B
Test 1	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
Test 2	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
Test 3	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
Cp				0.842	0.842

$$Cp(S) = Cp = \sqrt{\frac{\Delta P(s)}{\Delta P(s)}}$$

$$[Cp(A) - Cp(B)] \text{ must BE } \leq 0.01$$

$$\text{Average deviation(A or B)} = \frac{\sum [Cp(s) - Cp(A \text{ or } B)]}{3} \text{ must BE } \leq 0.01$$

Calibrated by Saksit Phaisanphisit Approved by Nattapon Jengwareewong
(Mr Saksit Phaisanphisit) (Mr Nattapon Jengwareewong)
Field Scientist (4) Field Specialist(1)

Form ZB1-046 (04/03/02)



PROBE NOZZLE DIAMETER CALIBRATION DATA SHEET

Calibration Date 12 Jul 22 Nozzle Set ID. : BKK_FS0533
Calibration Sheet No. : C-120722-BKK_FS0533 Vernier Caliper ID. : BKK_FS0626


Nozzle ID #	Nozzle Diameter (mm.)			Hi - Lo ΔD	(D ₁ + D ₂ + D ₃) / 3 D _{avg}
	D ₁	D ₂	D ₃		
1	0.318	0.318	0.318	0.000	0.318
2	0.475	0.475	0.475	0.000	0.475
3	0.635	0.635	0.635	0.000	0.635
4	0.792	0.792	0.792	0.000	0.792
5	0.952	0.952	0.952	0.000	0.952
6	1.110	1.110	1.110	0.000	1.110
7	1.270	1.270	1.270	0.000	1.270

Where :

D₁, D₂, D₃ = Three different nozzle diameters at 60 degrees to each other, each measured the nearest 0.025 mm.

ΔD = Maximum distance between any two diameters, must be ≤ 0.100 mm.

D_{avg} = (D₁ + D₂ + D₃) / 3



Calibrated by Saksit Phaisanphisit Approved by Nattapon Jengwareewong
(Mr Saksit Phaisanphisit) (Mr Nattapon Jengwareewong)
Field Scientist (4) Field Specialist(1)

Form No. QN ZB1-028 (13/01/01)

CONSOLE CONTROL UNIT CALIBRATION TEST REPORT

Barometric Pressure (mm.Hg) : 755
Relative Humidity (%) : 70.0
Temperature (°C) : 30.0

Reference Dry Gas Meter Data

Serial No. : A2003240
Model No. : DGM-SK05RM-036
Correction Factor (Yr) : 1.0160
Next Calibration Date : 27 May 23

Calibration of Date : 12 Jul 22
Next Cal. Date : 12 Jun 23

Calibration No. : C-120722-BKK_FS0518
Dry Gas Meter No. : BKK_FS0518
Console Serial No. : 1504025
Console Model No. : XC-572-V

Console Control Unit Dry Gas Meter Calibration				Dry Gas Meter Calibration			
AH (mm.H ₂ O)	θ (Minutes)	Vr (liters)		Tt (°C)		Orifice Calibration Factor (Y)	ΔVig
		Final	Initial	Final	Initial		
15	12.41	150.00	0.00	150.00	31.0	1.0012	47.6984
25	9.38	150.00	0.00	150.00	31.0	1.0002	45.6070
50	6.61	150.00	0.00	150.00	31.0	0.9978	45.2959
100	4.65	150.00	0.00	150.00	30.0	0.9900	44.9700
120	4.17	150.00	0.00	150.00	30.0	0.9781	43.4066
Avg				Avg		0.9915	44.4374

Y = Ratio of reading of reference to dry gas meter tolerance for individual values ± 0.02 from average.
ΔVig = Orifice pressure differential that equates to 21.24 in. of air @ 25°C and 760 mm of mercury, mmH₂O tolerance for individual values ± 5.08 from average.

Procedure: 49 GPR (UAPP AMETH) SEC 5.3 & 7

Calibrated by: Saksit Phaisanphisit Approved by: Nattapon Jengwareewong
(Mr Saksit Phaisanphisit) (Mr Nattapon Jengwareewong)
Field Scientist (4) Field Specialist(1)


Form No. QN ZB1-028 (13/01/01)




DIGITAL TEMPERATURE CALIBRATION DATA SHEET

Calibration Date :	12-Jul-22	Ambient Temperature (°C) :	30
Calibration sheet No. :	C-120722-BKK_FS0519	Relative Humidity (%) :	70
Digital Temperature ID	BKK_FS0519	Reference Temperature ID :	BKK_FS1144
Console Serial No. :	1504025	Serial No. :	20109000013
Console Model :	XC-572-V	Model :	Digicon-CC-VT-MS
		Next Calibrate :	31 Jan 23

Location	Reference Temperature °C	Digital Temperature °C	Error °C	Remark
Stack	0	0	0	
	25	25	0	
	50	50	0	
	100	101	1	
	150	153	3	
	200	202	2	
	250	252	2	
	300	302	2	
	500	503	3	
	1000	1004	4	
Probe	1200	1205	5	
	100	101	1	
	125	127	2	
	150	153	3	
	100	101	1	
Oven	125	127	2	
	150	153	3	
	100	101	1	
Filter	125	127	2	
	150	153	3	
	100	101	1	
Exit	125	127	2	
	150	153	3	
	100	101	1	
Meter	0	0	0	
	25	24	-1	
	50	50	0	
AUX	0	0	0	
	25	25	0	
	50	50	0	

Calibrated by : 
(Mr. Tinnakorn Kulchart)
Field Scientist (1)

Approved by : 
(Mr. Natthapol Jengwareewong)
Field Specialist (1)




Pitot Tube Calibration Data


Pitot Tube Identification Number : BKK_FS0522 Calibration Date : 12 Jul 22
Lab test duct Number : 258-1-13-01 Standard Pitot ID : BKK_FS0441
Calibration Sheet No. : C-120722-BKK_FS0522 Cp Standard : 0.99

Type S Pitot Tube Coefficient Data					
	Type s pitot tube Leg A,B	Standard pitot tube (ΔP, mm.H ₂ O)	Type s pitot tube (ΔP, mm.H ₂ O)	Cp (s) Leg A	Cp (s) Leg B
Test 1	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
Test 2	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
Test 3	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
			\bar{C}_p	0.842	0.842

$$Cp(S) = Cp = \sqrt{\frac{\Delta P(sld)}{\Delta P(s)}}$$

$$\left[\bar{C}_{p(A)} - \bar{C}_{p(B)} \right] \text{ must BE } \leq 0.01$$
$$\text{Average deviation(A or B)} = \frac{\sum_{i=1}^3 [Cp(s) - Cp(A \text{ or } B)]}{3} \text{ must BE } \leq 0.01$$

Calibrated by : 
(Mr. Tinnakorn Kulchart)
Field Scientist (1)

Approved by : 
(Mr. Natthapol Jengwareewong)
Field Specialist (1)

Form 281-016 (04/03/02)




Pitot Tube Calibration Data


Pitot Tube Identification Number : BKK_FS0523 Calibration Date : 12 Jul 22
Lab test duct Number : 258-1-13-01 Standard Pitot ID : BKK_FS0441
Calibration Sheet No. : C-120722-BKK_FS0523 Cp Standard : 0.99

Type S Pitot Tube Coefficient Data					
	Type s pitot tube Leg A,B	Standard pitot tube (ΔP, mm.H ₂ O)	Type s pitot tube (ΔP, mm.H ₂ O)	Cp (s) Leg A	Cp (s) Leg B
Test 1	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
Test 2	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
Test 3	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
			\bar{C}_p	0.842	0.842

$$Cp(S) = Cp = \sqrt{\frac{\Delta P(sld)}{\Delta P(s)}}$$

$$\left[\bar{C}_{p(A)} - \bar{C}_{p(B)} \right] \text{ must BE } \leq 0.01$$
$$\text{Average deviation(A or B)} = \frac{\sum_{i=1}^3 [Cp(s) - Cp(A \text{ or } B)]}{3} \text{ must BE } \leq 0.01$$

Calibrated by : 
(Mr. Tinnakorn Kulchart)
Field Scientist (1)

Approved by : 
(Mr. Natthapol Jengwareewong)
Field Specialist (1)

Form 281-016 (04/03/02)





PROBE NOZZLE DIAMETER CALIBRATION DATA SHEET


Calibration Date 12 Jul 22 Nozzle Set ID : BKK_FS0524
Calibration Sheet No. : C-120722-BKK_FS0524 Vernier Caliper ID : BKK_FS0626

Nozzle ID #	Nozzle Diameter (cm.)			Ht - Lo ΔD	(D ₁ + D ₂ + D ₃) : 3 D _{avg}
	D ₁	D ₂	D ₃		
1	0.318	0.318	0.318	0.000	0.318
2	0.475	0.475	0.475	0.000	0.475
3	0.635	0.635	0.635	0.000	0.635
4	0.792	0.792	0.792	0.000	0.792
5	0.952	0.952	0.952	0.000	0.952
6	1.110	1.110	1.110	0.000	1.110
7	1.270	1.270	1.270	0.000	1.270

Where :
D₁, D₂, D₃ > There different nozzle diameters at 60 degrees to each other, each measured the nearest 0.025 mm.
ΔD = Maximum distance between any two diameters, must be ≤ 0.100 mm.
D_{avg} = (D₁ + D₂ + D₃) : 3



Calibrated by : 
(Mr. Tinnakorn Kulchart)
Field Scientist (1)

Approved by : 
(Mr. Natthapol Jengwareewong)
Field Specialist (1)

Form No. Q8 281-025 (12/01/03)

Certificate No: G 650023
Date of issue : 21-Jan-22

Instrument description : Fuel Gas Analyzer
Instrument model : Testo 340
Instrument serial no. : 62150585
ID no. or control no. : RYG_F50465
Manufacturer : testo SE
Probe description : -
Probe model : -
Probe serial : -
Customer name : ALS LABORATORY GROUP (THAILAND) CO.,LTD.
Customer address : 104 Phatthanakan 40, Phatthanakan Road, Khwaeng Phatthanakan, Khet Suan Luang, Bangkok, 10250 Thailand
Total pages of certificate : 3 Pages
Receiving no. : L-220082
Receiving date : 14-Jan-22
Parameter of calibration : Gas Calibration(Oxygen 2.501,10.00,21.00 %Vol, Carbon Monoxide 60.97,309.9,1003 ppm, Nitric Oxide 10.08,150.9,320.6 ppm, Sulphur Dioxide 50.04,100.9,601.1 ppm)

REVIEW BY: *Mankorn P.*
APPROVED BY: *[Signature]*
NEXT CAL DATE: 19/1/23

Condition of UUC : Used
Ambient condition : All of the Measurement were carried out the stabilized laboratory
Temperature : 23 ± 5 °C
Humidity : 55 ± 15 %RH
Calibration place : 17/121 Soi Ngamwongwan 47 Yeak 48, Toongsonghong, Laksi, Bangkok 10210
Calibration procedure no. : WI-CL-28-C

The calibration certificate expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.
This certificate is applied only to item under test Environmental condition.
This Calibration Certificate may not be reproduced other than in full except with the permission of the issuing laboratory.
Calibration certificates without signature and seal not valid.
This calibration certificate documents are traceability to national standards, which realize measurement according to the International System of Units (SI).
Date of calibration : 19-Jan-22

[Signature]
Mr. Sedtawut Nueathong
Calibration Technician

[Signature]
Mrs. Nongluck Wongsettee
Technical Manager

FM-CL-09-C Rev.8

Page 1 of 3

Issued Date 26/02/16

ENTECH INDUSTRIAL SOLUTION CO.,LTD.

17/121 Soi Ngamwongwan 47 Yeak 48, Toongsonghong, Laksi, Bangkok 10210 THAILAND Tel: 0-2779-8888 Fax: 0-2779-8899 info@entech.co.th
Tax ID : 0105536035591 www.entech.co.th

Certificate No.: G 650023

Calibration Results After Adjustment (Table 3)

Parameter of Standard	Standard Values	Mean of UUC	Error	Uncertainty (±)
O2 (%Vol)	2.501	2.47	-0.031	0.20
O2 (%Vol)	10.00	9.89	-0.11	0.40
O2 (%Vol)	21.00	21.11	0.12	0.80
CO (ppm)	60.97	83	2.03	2.8
CO (ppm)	309.9	314	4.1	11
CO (ppm)	1003	1016	13	34
NO (ppm)	10.08	9	-1.08	3.0
NO (ppm)	150.90	150	-0.9	5.0
NO (ppm)	320.6	318	-2.6	10
SO2 (ppm)	50.04	48	-2.04	5.0
SO2 (ppm)	100.9	101	0.1	5.0
SO2 (ppm)	601.1	604	2.9	14

Remark : 1 cmol/mol = 1 %vol., 1 µmol/mol = 1 ppm.

End of Report

FM-CL-09-C Rev.8

Page 3 of 3

Issued Date 26/02/16

ENTECH INDUSTRIAL SOLUTION CO.,LTD.

17/121 Soi Ngamwongwan 47 Yeak 48, Toongsonghong, Laksi, Bangkok 10210 THAILAND Tel: 0-2779-8888 Fax: 0-2779-8899 info@entech.co.th
Tax ID : 0105536035591 www.entech.co.th

Certificate No.: G 650023

Standard References (Table 1)

Standard	Certificate No.	Vendor	Due date
Oxygen (O2) 2.501 % Vol	2431/19	Unde	16-Jul-23
Oxygen (O2) 10.00 % Vol	2453/19	Unde	18-Jul-23
Oxygen (O2) 21.00 % Vol	2426/19	Unde	16-Jul-23
Carbon monoxide (CO) 60.97 ppm	2842/21	Unde	24-Jun-23
Carbon monoxide (CO) 309.9 ppm	2603/21	Unde	22-Jun-23
Carbon monoxide (CO) 1003 ppm	2029/21	Unde	23-Apr-23
Nitric Oxide (NO) 10.08 ppm	3241/21	Unde	25-Jul-23
Nitric Oxide (NO) 150.9 ppm	2857/21	Unde	27-Jun-23
Nitric Oxide (NO) 320.6 ppm	2944/21	Unde	2 Jul 23
Sulphur Dioxide (SO2) 50.04 ppm	3205/21	Unde	25-Jul-23
Sulphur Dioxide (SO2) 100.9 ppm	4942/20	Unde	20-Nov-22
Sulphur Dioxide (SO2) 601.1 ppm	3204/21	Unde	20-Jul-23

Measured room conditions

Temperature : 22.5 °C Humidity : 56.2 %RH Pressure : 1018.3 mbar

Calibration conditions

Gas Temperature : 23 °C Flow rate : 600 ml/min Gas pressure : 1021.9 mbar

Calibration Results Before Adjustment (Table 2)

Parameter of Standard	Standard Values	Mean of UUC	Error	Uncertainty (±)
O2 (%Vol)	2.501	2.47	-0.031	0.20
O2 (%Vol)	10.00	9.89	-0.11	0.40
O2 (%Vol)	21.00	21.12	0.12	0.80
CO (ppm)	60.97	83	2.03	2.8
CO (ppm)	309.9	314	4.1	11
CO (ppm)	1003	1016	13	34
NO (ppm)	10.08	9	-2.08	3.0
NO (ppm)	150.9	162	11.1	5.0
NO (ppm)	320.6	345	24.4	10
SO2 (ppm)	50.04	38	-12.04	5.0
SO2 (ppm)	100.9	92	-8.9	5.0
SO2 (ppm)	601.1	593	-8.1	14

FM-CL-09-C Rev.8

Page 2 of 3

Issued Date 26/02/16

ENTECH INDUSTRIAL SOLUTION CO.,LTD.

17/121 Soi Ngamwongwan 47 Yeak 48, Toongsonghong, Laksi, Bangkok 10210 THAILAND Tel: 0-2779-8888 Fax: 0-2779-8899 info@entech.co.th
Tax ID : 0105536035591 www.entech.co.th

Calibration certificate Kalibrier-Zertifikat

4319339

Object	Controlunit 1350	Measuring Box testo 350	Hereby we confirm that the performing calibration laboratory is working with a management system according to ISO 9001:2015 and ISO/IEC 17025:2018 Accreditation certificates can be found under www.tiinfo.de . The measuring installations used for calibration are regularly calibrated and traceable to the national standards of the German Federal Physical Technical Institute (PTB) or other national standards. Should no national standards exist, the measuring procedure corresponds with the technical regulations and norms valid at the time of the measurement. The documents established for this procedure are available for viewing. At the necessary measured data can be found on this calibration certificate.
Gegenstand			
Manufacturer	TESTO SE & Co. KGaA	TESTO SE & Co. KGaA	
Hersteller			
Type description	0632 3511	0632 3510	
Typ			
Serial no.	03580182	62985049	
Serien Nr.			
Inventory no.	---	---	
Inventar Nr.			
Test equipment no.	---	---	
Prüfmittel Nr.			
Equipment no.	14672444	14674793	
Equipment Nr.			
Location	---	---	
Standort			
Customer	ALS Laboratory Group (Thailand) Co., Ltd		
Auftraggeber			
Customer ID no.	TH-21140 A. Pluekdaeng, Rayong, Thailand		
Kunden Nr.			
Order no.	1031994		
Auftrags Nr.			
Order no.	10842828 / 0520 0055		
Auftrags Nr.			
Date of calibration		09.12.2021	
Datum der Kalibrierung			
Date of the recommended re-calibration		09.12.2022	
Datum der empfohlenen Rekalibrierung			

Conformity statement Konformitätsaussage

- ☒ Measured value(s) within the allowable deviation¹. Messwert(e) innerhalb der zulässigen Abweichung¹.
☐ Measured value(s) outside of the allowable deviation¹. Messwert(e) außerhalb der zulässigen Abweichung¹.

¹ The expanded measurement uncertainty was calculated according to EA-4-02 M:2013 with a coverage probability of approx. 95% and contains the uncertainty of the reference, the method and the uncertainty of the unit under test. The statement of conformity is based on the decision rule "Vertrauensniveau 50" (confidence level 50).
² Die erweiterte Messunsicherheit wurde nach EA-4-02 M:2013 mit einer Überdeckungswahrscheinlichkeit von etwa 95% berechnet und enthält die Unsicherheit der Referenz, des Verfahrens sowie die Unsicherheit des Prüfings. Die Konformitätsaussage erfolgt nach der Entscheidungsregel "Vertrauensniveau 50".

This calibration certificate may not be reproduced other than in full except with the permission of the issuing laboratory. Calibration certificates without signature and seal are not valid.
Dieses Kalibrierzertifikat kann nur vollständig reproduziert werden. Auszüge oder Änderungen bedürfen der Genehmigung des ausstellenden Kalibrierlaboratoriums. Kalibrierzertifikate ohne Unterschrift und Siegel haben keine Gültigkeit.

Seal Stamp



Supervisor Fachverantwortlicher
[Signature]
Martin Förderer

Technician Geprüfter
[Signature]
Johannes Wängler

Testo Industrial Services GmbH

Gewerbepforte 3
17198 Ritzschdorf
Tel: +49 7951 93001-5000
Fax: +49 7951 93001-6010
www.testo.de
info@testo.de
Page 1/3

Calibration certificate Kalibrier-Zertifikat

4319339

Measuring equipment Messeinrichtungen

Index	Reference	Traceability	Next cal.	Certificate-no.	Eq.-no.
a	Test gas medium 1 Prüfgas Medium 1	SCS-SC50028 2021-03	2024-03	4017585	12898976
b	Test gas medium 3 Prüfgas Medium 3	SCS-SC50028 2021-03	2022-03	4017588	12898982
c	Test gas medium 5 Prüfgas Medium 5	SCS-SC50028 2021-03	2022-03	4017591	12898984
d	Test gas medium 8 Prüfgas Medium 8	SCS-SC50028 2021-07	2022-10	4220471	12898987
e	Test gas medium 7 Prüfgas Medium 7	SCS-SC50028 2021-03	2022-03	4017596	12898986
f	Test gas medium 11 Prüfgas Medium 11	ISO-ISO 6141 2021-04	2022-04	4017632	14087964
g	Digistat 4420 Digistat 4420	15070-01-01 2021-01	2022-07	616673	12866901
h	Pneumator Pneumator	15070-01-01 2021-07	2022-07	D53219	12866547

Reference certificates are available at www.primasonline.com Referenzzertifikate sind auf www.primasonline.com einsehbar

Ambient conditions Umgebungsbedingungen

Temperature Temperatur (20...26) °C Humidity Feuchte (20...60) % RH % rF

Measuring procedure Messverfahren

The calibration was carried out by comparison measurement with calibrated test gases, a calibrator of temperature and pressure.
Die Kalibrierung erfolgte durch Vergleichsmessung mit kalibrierten Prüfgasen, einem Temperatur- und Druckkalibrator.

Measuring results Messergebnisse

Channel Kanal ---

Unit	Reference value	Indicated measured value	Deviation	Allowed deviation ²⁾	Measurement uncertainty (k=2)	Confirmation
Einheit	Bezugswert	Angezeigter Messwert	Abweichung	Zulässige Abweichung ²⁾	Messunsicherheit (k=2)	Bestätigung
CO						
ppm	100,1 ^a	104	3,9	± 11	3,3	pass
ppm	401,0 ^b	404	3,0	± 21	8,5	pass
ppm	700,0 ^c	729	29,0	± 36	14,4	pass
NO						
ppm	150,3 ^a	151	0,7	± 9	4,0	pass
ppm	300 ^a	301	1	± 16	6,9	pass
NO2						
ppm	99,9 ^a	102,3	2,4	± 5,1	3,20	pass
SO2						
ppm	100,1 ^f	97	-3,1	± 6	3,5	pass
O2						
Vol.-%	0,0 ^a	0,04	0,0	± 0,21	0,027	pass
Vol.-%	2,510 ^a	2,54	0,030	± 0,21	0,055	pass
Vol.-%	5,000 ^b	5,06	0,060	± 0,21	0,102	pass
Temperatur						
°C	100,0 ^a	99,1	-0,9	± 1,1	0,24	pass
°C	200,0 ^a	199,9	-0,1	± 1,1	0,24	pass
Druck						
hPa	50,0 ^b	50,0	0,0	± 0,9	0,52	pass
hPa	100,0 ^b	100,0	0,0	± 1,6	0,52	pass

²⁾ In accordance with the manufacturer gemäß Hersteller

Testo Industrial Services GmbH

Gewerbestraße 3 Tel: +49 7651 9201-5000 Fax: +49 7651 9201-5010 www.testo.de info@testo.de Page 2/3

Calibration certificate Kalibrier-Zertifikat

4319339

Special remarks Besondere Bemerkungen



Stopwatch Calibration Test Report

Calibration Date : 3 Jul 22 Next Cal. Date : 3 Jan 23
Barometric Pressure (mmHg) : 756 Temperature (°C) : 31.0
Relative Humidity (%) : 62.0

Reference Stopwatch Data

Stopwatch ID No. : E18061
Model : F808
Serial No. : -
Calibration Date : 8 Sep 20
Certificate No. : E-2009218

Console Control Meter Data

Dry Gas Meter No. : BKK_FS0507
Model : XC-572-V
Serial No. : 1503017

Run No.	Time Actual (m:ss.ms)	Time Reading (m:ss)	Diff. (ms)	Diff. (min)
1	5:00:08	5:00	8	0.00013
2	5:00:11	5:00	11	0.00018
3	5:00:10	5:00	10	0.00017
4	5:00:10	5:00	10	0.00017
5	5:00:10	5:00	10	0.00017
6	5:00:10	5:00	10	0.00017
7	5:00:08	5:00	8	0.00013
8	5:00:09	5:00	9	0.00016
9	5:00:11	5:00	11	0.00018
10	5:00:12	5:00	12	0.00020
Average			0.00017	
SD			0.00002	

Calibrate by :

Mr. Prasant Surakhan

Field Scientist (3)

Approved by :

Mr. Samart Roo-ngan

Specialist (1)

CONSOLE CONTROL UNIT CALIBRATION TEST REPORT

Barometric Pressure (mm.Hg) : 756
Relative Humidity (%) : 62.0
Temperature (°C) : 31.0
Reference Dry Gas Meter Data
Reference Dry Gas Meter ID : BKK_FS0507
Serial No. : 1503017
Correction Factor (Y) : 1.0060
Next Calibration Date : 7 Oct 22

Reference Dry Gas Meter Calibration		Dry Gas Meter Correction Factor		Office Calibration Factor	
ΔH (mm H ₂ O)	ΔH (mm H ₂ O)	ΔH (mm H ₂ O)	ΔH (mm H ₂ O)	ΔH (mm H ₂ O)	ΔH (mm H ₂ O)
15	15.00	15.00	15.00	15.00	15.00
25	25.00	25.00	25.00	25.00	25.00
50	50.00	50.00	50.00	50.00	50.00
80	80.00	80.00	80.00	80.00	80.00
120	120.00	120.00	120.00	120.00	120.00

Y Ratio of reading of reference to dry gas meter : tolerance for individual values : ± 0.02 from average

ΔH₀ Difference pressure differential that equates to 0.21 in of air @ 25 °C and 760 mm of mercury, mmHgO : tolerance for individual values : ± 5.08 from average

Procedure: 40 CFR 60 APP A, MET-1, vers 5.3.4.7

Calibrated by :

Mr. Prasant Surakhan
Field Scientist(3)

Approved by :

Mr. Samart Roo-ngan
Specialist(1)



DIGITAL TEMPERATURE CALIBRATION DATA SHEET

Calibration Date :	3 Jul 22	Ambient Temperature (°C) :	31
Calibration sheet No. :	C-030722-BKK_FS0508	Relative Humidity (%) :	62
Digital Temperature ID	BKK_FS0508	Reference Temperature ID :	BKK_FS0600
Serial No. :	1503017	Serial No. :	7688004
Model :	XC-572-V	Model :	FLUKE 714
		Next Calibrate :	26 Jul 23

Location	Reference Temperature °C	Digital Temperature °C	Error °C	Remark
Slack	0	1	1	
	25	26	1	
	50	51	1	
	100	101	1	
	150	151	1	
	200	201	1	
	250	252	2	
	300	302	2	
	500	503	3	
	1000	1003	3	
Probe	1200	1203	3	
	100	101	1	
	125	126	1	
	150	151	1	
Oven	100	100	0	
	125	125	0	
	150	151	1	
Filter	100	100	0	
	125	125	0	
Exit	150	151	1	
	0	1	1	
	10	11	1	
Meter	20	21	1	
	0	0	0	
	25	25	0	
AUX	50	50	0	
	0	0	0	
	25	25	0	
	50	50	0	

Calibrated by : Prasert S.
(Mr. Prasert Surakhon)
Field Scientist (3)

Approved by : Mr. Samart Roo-ngan
(Mr. Samart Roo-ngan)
Specialist (1)

FORM NO. F-06-027 REVISION NO. - ISSUE DATE 2/6/02



Pitot Tube Calibration Data

Pitot Tube Identification Number : BKK_FS0511
Calibration Date : 3 Jul 22
Lab test duct Number : 258-1-13-01
Standard Pitot ID : BKK_FS0441
Calibration Sheet No. : C-030722-BKK_FS0511
Cp Standard : 0.99

Type S Pitot Tube Coefficient Data					
	Type s pitot tube Leg A,B	Standard pitot tube (ΔP, mm.H ₂ O)	Type s pitot tube (ΔP, mm.H ₂ O)	Cp (s) Leg A	Cp (s) Leg B
Test 1	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
Test 2	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
Test 3	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
\bar{C}_p				0.842	0.842

$$Cp(S) = Cp_{std} \sqrt{\frac{\Delta P(Std)}{\Delta P(s)}}$$

$$| \bar{C}_{p(A)} - \bar{C}_{p(B)} | \text{ must } BE \leq 0.01$$

$$\text{Average deviation(A or B)} = \frac{\sum [Cp(s) - Cp(A \text{ or } B)]}{3} \text{ must } BE \leq 0.01$$

Calibrated by : Prasert S.
(Mr. Prasert Surakhon)
Field Scientist (3)

Approved by : Mr. Samart Roo-ngan
(Mr. Samart Roo-ngan)
Specialist (1)

FORM NO. F-06-016 REVISION NO. - ISSUE DATE 2/6/02



Pitot Tube Calibration Data

Pitot Tube Identification Number : BKK_FS0512
Calibration Date : 3 Jul 22
Lab test duct Number : 258-1-13-01
Standard Pitot ID : BKK_FS0441
Calibration Sheet No. : C-030722-BKK_FS0512
Cp Standard : 0.99

Type S Pitot Tube Coefficient Data					
	Type s pitot tube Leg A,B	Standard pitot tube (ΔP, mm.H ₂ O)	Type s pitot tube (ΔP, mm.H ₂ O)	Cp (s) Leg A	Cp (s) Leg B
Test 1	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
Test 2	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
Test 3	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
\bar{C}_p				0.842	0.842

$$Cp(S) = Cp_{std} \sqrt{\frac{\Delta P(Std)}{\Delta P(s)}}$$

$$| \bar{C}_{p(A)} - \bar{C}_{p(B)} | \text{ must } BE \leq 0.01$$

$$\text{Average deviation(A or B)} = \frac{\sum [Cp(s) - Cp(A \text{ or } B)]}{3} \text{ must } BE \leq 0.01$$

Calibrated by : Prasert S.
(Mr. Prasert Surakhon)
Field Scientist (3)

Approved by : Mr. Samart Roo-ngan
(Mr. Samart Roo-ngan)
Specialist (1)

FORM NO. F-06-027 REVISION NO. - ISSUE DATE 2/6/02



PROBE NOZZLE DIAMETER CALIBRATION DATA SHEET

Calibration Date : 3 Jul 22
Nozzle Set ID : BKK_FS0513
Calibration Sheet No. : C-030722-BKK_FS0513
Vernier Caliper ID : BKK_FS0628


Nozzle ID #	Nozzle Diameter (cm.)			Hi - Lo ΔD	(D ₁ + D ₂ + D ₃) / 3 D _{avg}
	D ₁	D ₂	D ₃		
1	0.315	0.315	0.315	0.000	0.315
2	0.475	0.475	0.475	0.000	0.475
3	0.635	0.635	0.635	0.000	0.635
4	0.790	0.790	0.790	0.000	0.790
5	0.950	0.950	0.950	0.000	0.950
6	1.110	1.110	1.110	0.000	1.110
7	1.270	1.270	1.270	0.000	1.270

Where :

D₁, D₂, D₃ = Three different nozzle diameters at 60 degrees to each other, each measured the nearest 0.025 mm.

ΔD = Maximum distance between any two diameters, must be ≤ 0.100 mm.

D_{avg} = (D₁ + D₂ + D₃) / 3



Calibrated by : Prasert S.
(Mr. Prasert Surakhon)
Field Scientist (3)

Approved by : Mr. Samart Roo-ngan
(Mr. Samart Roo-ngan)
Specialist (1)

FORM NO. F-06-028 REVISION NO. - ISSUE DATE 2/6/02



CONSOLE CONTROL UNIT CALIBRATION TEST REPORT

Calibration of Date : 3 Jul 22
 Next Cal. Date : 3 Jan 23
 Barometric Pressure (mm.Hg) : 756
 Relative Humidity (%) : 54.0
 Temperature : 31.0
 Console Control Meter Data.
 Calibration No. : C-030722-BKK_FS1093
 Dry Gas Meter No. : BKK_FS1093
 Serial No. : 1706090
 Model No. : XC-572-V
 Reference Dry Gas Meter ID : BKK_FS1122
 Serial No. : A2003240
 Correction Factor (Yr) : 1.0160
 Next Calibration Date : 27 May 23

ΔH (mm.H ₂ O)	Θ Minutes	Reference Dry Gas Meter Calibration						Console Control - Drygas Meter						Dry Gas Meter Correction Factor (Yr)	Office Calibration Factor (Yr)
		Vr (Liters)			Tr (°C)			Vm (Liters)			Ti (°C)				
		Final	Initial	Total	Final	Initial	Total	Final	Initial	Total	Final	Initial	Total		
15	12.21	150.00	0.00	150.00	23.0	31.0	31.0	148478.2	148325.0	153.20	30.0	30.0	30.0	0.9866	46.5475
25	9.25	150.00	0.00	150.00	31.0	31.0	31.0	148462.2	148493.0	153.20	31.0	31.0	31.0	0.9824	44.0016
50	6.56	150.00	0.00	150.00	32.0	32.0	32.0	148415.2	148662.0	153.20	32.0	32.0	32.0	0.9800	44.2714
80	5.10	150.00	0.00	150.00	32.0	32.0	32.0	149258.6	149105.0	153.60	33.0	33.0	33.0	0.9878	42.8035
120	4.18	150.00	0.00	150.00	32.0	32.0	32.0	149421.0	149271.0	153.00	36.0	36.0	36.0	0.9950	42.8503
														Avg	43.8949

Y Ratio of reading of reference to dry gas meter; tolerance for individual values ± 0.02 from average.
 Avg. Office pressure differential that equates to 21.24 in. of air @ 25°C and 760 mm of mercury, mmH₂O; tolerance for individual values ± 5.00 from average.

Procedure: 40 CFR 60 APP A, METH. 3EC 5.3 & 7

Calibrated by:

(Mr. Prasert Surakhan)
Field Scientist(3)

Approved by:

(Mr. Samart Roo-ngan)
Specialist(1)



DIGITAL TEMPERATURE CALIBRATION DATA SHEET

Calibration Date :	3 Jul 22	Ambient Temperature (°C)	31
Calibration sheet No. :	C-030722-BKK_FS1093	Relative Humidity (%) :	62
Digital Temperature ID :	BKK_FS1093	Reference Temperature ID :	BKK_FS0809
Console Serial No. :	1706090	Serial No. :	7688004
Console Model :	XC-572-V	Model :	FLUKE714
		Last Calibrate :	26 Jul 23

Location	Reference Temperature °C	Digital Temperature °C	Error °C	Remark
Stack	0	0	0	
	25	25	0	
	50	50	0	
	100	100	0	
	150	150	0	
	200	200	0	
	250	250	0	
	300	300	0	
	500	501	1	
	1000	1001	1	
Probe	1200	1201	1	
	100	100	0	
	125	126	1	
	150	151	1	
Oven	100	101	1	
	125	125	0	
	150	150	0	
Filter	100	100	0	
	125	125	0	
	150	151	1	
Exit	0	0	0	
	10	10	0	
	20	20	0	
Moler	0	0	0	
	25	25	0	
	50	50	0	
AUX	0	0	0	
	25	25	0	
	50	50	0	

Calibrated by :

(Mr. Prasert Surakhan)
Field Scientist (3)

Approved by :

(Mr. Samart Roo-ngan)
Specialist (1)

REVISION: FORM 001 - REVISION: 01 - DATE: 2-1-17



Stopwatch Calibration Test Report

Calibration Date : 3 Jul 22
 Next Cal. Date : 3 Jan 23
 Barometric Pressure (mm.Hg) : 756
 Temperature (°C) : 31.0
 Relative Humidity (%) : 62.0

Reference Stopwatch Data

Stopwatch ID No. : E18051
 Model : F808
 Serial No. :
 Calibration Date : 8 Sep 20
 Certificate No. : E-2009018

Console Control Meter Data

Dry Gas Meter No. : BKK_FS1093
 Model : XC-572-V
 Serial No. : 1706090

Run No.	Time Actual (m:ss.ms)	Time Reading (m:ss)	Diff. (ms)	Diff. (min)
1	5:00:11	5:00	11	0.00018
2	5:00:10	5:00	10	0.00017
3	5:00:11	5:00	11	0.00018
4	5:00:10	5:00	10	0.00017
5	5:00:12	5:00	12	0.00020
6	5:00:12	5:00	12	0.00020
7	5:00:10	5:00	10	0.00017
8	5:00:10	5:00	10	0.00017
9	5:00:08	5:00	8	0.00013
10	5:00:09	5:00	9	0.00015
Average				0.00017
SD				0.00002

Calibrate by :

Mr. Prasert Surakhan

Field Scientist (3)

Approved by :

Mr. Samart Roo-ngan

Specialist (1)



Pitot Tube Calibration Data

Pitot Tube Identification Number : BKK_FS1104
 Calibration Date : 3 Jul 22
 Lab test duct Number : 258-1-13-01
 Standard Pitot ID : BKK_FS0441
 Calibration Sheet No. : C-030722-BKK_FS1104
 Cp Standard : 0.99

Type S Pitot Tube Coefficient Data					
	Type s pitot tube Leg A,B	Standard pitot tube (ΔP, mm.H ₂ O)	Type s pitot tube (ΔP, mm.H ₂ O)	Cp (s) Leg A	Cp (s) Leg B
Test 1	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
Test 2	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
Test 3	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
			Cp	0.842	0.842

$$Cp(S) = Cp_{std} \sqrt{\frac{\Delta P_{std}}{\Delta P (s)}}$$

$$[Cp_{(A)} - Cp_{(B)}] \text{ must BE } \leq 0.01$$

$$\text{Average deviation(A or B)} = \frac{\sum [Cp (s) - Cp(A \text{ or } B)]}{3} \text{ must BE } \leq 0.01$$

Calibrated by :

Mr. Prasert Surakhan

Field Scientist (3)

Approved by :

Mr. Samart Roo-ngan

Specialist (1)

REVISION: FORM 001 - REVISION: 01 - DATE: 2-1-17



Pitot Tube Calibration Data

Pitot Tube Identification Number : BKK_FS1105 Calibration Date : 3 Jul 22
 Lab test duct Number : 258-1-13-01 Standard Pitot ID : BKK_FS0441
 Calibration Sheet No. : C-030722-BKK_FS1105 Cp Standard : 0.99

Type S Pitot Tube Coefficient Data					
	Type s pitot tube Leg A,B	Standard pitot tube (ΔP , mm.H ₂ O)	Type s pitot tube (ΔP , mm.H ₂ O)	Cp (s) Leg A	Cp (s) Leg B
Test 1	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
Test 2	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
Test 3	A	12.00	16.60	0.842	-
	B	12.00	16.60	-	0.842
\bar{C}_p				0.842	0.842

$$C_p(S) = C_p \cdot \sqrt{\frac{\Delta P(s)}{\Delta P(s)}}$$

$$[\bar{C}_p(A) - \bar{C}_p(B)] \text{ must BE } \leq 0.01$$

$$\text{Average deviation(A or B)} = \frac{\sum [C_p(s) - C_p(A \text{ or } B)]}{3} \text{ must BE } \leq 0.01$$

Calibrated by: Mr. Prasen Surakhan
Field Scientist (3)
 Approved by: Mr. Samart Roo-ngan
Specialist (1)

INTERNAL FILE NAME: REFERENCE NO.: ISSUE DATE: 01/07/2022



PROBE NOZZLE DIAMETER CALIBRATION DATA SHEET

Calibration Date : 3 Jul 22 Nozzle Set ID : BKK_FS1106
 Calibration Sheet No. : C-030722-BKK_FS1105 Vermer Caliper ID : BKK_FS0626


Nozzle ID #	Nozzle Diameter (mm.)			Hi - Lo	(D ₁ + D ₂ + D ₃) / 3
	D ₁	D ₂	D ₃	ΔD	D _{avg}
1	0.300	0.300	0.300	0.000	0.300
2	0.450	0.450	0.450	0.000	0.450
3	0.630	0.630	0.630	0.000	0.630
4	0.780	0.780	0.780	0.000	0.780
5	0.950	0.950	0.950	0.000	0.950
6	1.090	1.090	1.090	0.000	1.090
7	1.250	1.250	1.250	0.000	1.250

Where:

D₁, D₂, D₃ = Three different nozzle diameters at 60 degrees to each other, each measured the nearest 0.025 mm.


ΔD = Maximum distance between any two diameters, must be ≤ 0.100 mm.

D_{avg} = (D₁ + D₂ + D₃) / 3



Calibrated by: Mr. Prasen Surakhan
Field Scientist (3)
 Approved by: Mr. Samart Roo-ngan
Specialist (1)

INTERNAL FILE NAME: REFERENCE NO.: ISSUE DATE: 01/07/2022



Calibration certificate Kalibrier-Zertifikat

4321031

Object: Controlunit t350 Measuring Box testo 350
 Manufacturer: TESTO SE & Co. KGaA
 Type description: 0632 3511 0632 3510
 Serial no.: 03580090 62985022
 Inventory no.: ---
 Test equipment no.: ---
 Equipment no.: 14672696 14675708
 Location: ---
 Customer: ALS Laboratory Group (Thailand) Co., Ltd
 Auftraggeber: 104 Phatthanasakan 40, Phatthanasakan Rd.,
 TH-10250 Bangkok, Thailand
 Customer ID no.: 1031994
 Order no.: 10842906 / 0520 0055

Hereby we confirm that the performing calibration laboratory is working with a management system according to ISO 9001:2015 and ISO/IEC 17025:2018. Accreditation certificates can be found under www.testo.de. The measuring installations used for calibration are regularly calibrated and traceable to the national standards of the German Federal Physical Technical Institute (PTB) or other national standards. Should no national standards exist, the measuring procedure corresponds with the technical regulations and norms valid at the time of the measurement. The documents established for this procedure are available for viewing. At the necessary measured data can be found on this calibration certificate.

Hiermit bestätigen wir, dass das durchführende Kalibrierlabor ein Managementsystem nach ISO 9001:2015, sowie ISO/IEC 17025:2018 eingeführt hat. Die Urkunden finden Sie auf www.testo.de. Die für die Kalibrierung verwendeten Messmittel sind regelmäßig kalibriert und sind rückführbar auf die nationalen Normale der Physikalisch-Technischen Bundesanstalt (PTB). Deutschland oder auf andere nationale Normale. Wo keine nationalen Normale existieren, entspricht das Messverfahren den derzeit gültigen technischen Regeln und Normen. Die für diesen Vorgang angefertigte Dokumentation kann eingesehen werden. Alle erforderlichen Messdaten sind in diesem Kalibrier-Zertifikat aufgeführt.

Date of calibration: 10.12.2021
 Date of the recommended re-calibration: 10.12.2022

Conformity statement Konformitätsaussage

☒ Measured value(s) within the allowable deviation¹. Messwert(e) innerhalb der zulässigen Abweichung¹.
☐ Measured value(s) outside of the allowable deviation¹. Messwert(e) außerhalb der zulässigen Abweichung¹.

¹ The expanded measurement uncertainty was calculated according to EA-4/02:2013 with a coverage probability of approx. 95% and contains the uncertainty of the reference, the method and the uncertainty of the used test. The statement of conformity is based on the decision rule "Vierstufenregel 50" (confidence level: 50%).

² Die erwartete Messunsicherheit wurde nach EA-4/02:2013 mit einer Überdeckungswahrscheinlichkeit von etwa 95% berechnet und enthält die Unsicherheit der Referenz, des Verfahrens sowie die Unsicherheit des Prüfmittels. Die Konformitätsaussage erfolgt nach der Einstufungsregel "Vierstufenregel 50".
 This calibration certificate may not be reproduced other than in full except with the permission of the issuing laboratory. Calibration certificates without signature and seal are not valid.
 Dieser Kalibrierchein darf nur vollständig weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung des ausstellenden Kalibrierlaboratoriums. Kalibrierzertifikate ohne Unterschrift und Stempel haben keine Gültigkeit.

Seal Stempel:
 Supervisor Fachverantwortlicher: Martin Förderer
 Technician Bearbeiter: Samuel Garcia Zloti

Calibration certificate Kalibrier-Zertifikat

4321031

Measuring equipment Messeinrichtungen

Index	Reference	Traceability	Next cal.	Certificate-no.	Eq.-no.
Referenz	Rückführung	Relativ	Zertif.-Nr.	Eq.-Nr.	
a	Test gas medium 1 ProGas Medium 1	SCS-SCS0026 2021-03	2024-03	4017585	12899976
b	Test gas medium 3 ProGas Medium 3	SCS-SCS0026 2021-03	2022-03	4017588	12899982
c	Test gas medium 5 ProGas Medium 5	SCS-SCS0026 2021-03	2022-03	4017591	12899984
d	Test gas medium 6 ProGas Medium 6	SCS-SCS 2021-07	2022-10	4220471	12899987
e	Test gas medium 7 ProGas Medium 7	SCS-SCS0026 2021-03	2022-03	4017596	12899988
f	Test gas medium 11 ProGas Medium 11	ISO-ISO 8141 2021-04	2022-04	4017602	14281964
g	Digistart 4420 Digistart 4420	19070-01-01 2021-07	2022-07	E166873	12066601
h	Pneumat PneuMaster	16070-01-01 2021-07	2022-07	D53219	12066547

Reference certificates are available at www.primasonline.com Referenzzertifikate sind auf www.primasonline.com abrufbar

Ambient conditions Umgebungsbedingungen

Temperature Temperatur (20...26) °C Humidity Feuchte (20...60) % RH % rF

Measuring procedure Messverfahren

The calibration was carried out by comparison measurement with calibrated test gases, a calibrator of temperature and pressure. Die Kalibrierung erfolgte durch Vergleichsmessung mit kalibrierten Prüfgasen, einem Temperatur- und Druckkalibrator.

Measuring results Messergebnisse

Channel Kanal: ---

Unit	Reference value	Indicated measured value probe	Deviation	Allowed deviation ²	Measurement uncertainty (k=2)	Confirmation
Einheit	Bezugswert	Angezeigter Messwert Kalibrierobjekt	Abweichung	Zulässige Abweichung ²	Messunsicherheit (k=2)	Überprüfung
CO	ppm	100.1 ^a	103	± 11	3.3	pass
	ppm	401.0 ^b	402	± 21	8.5	pass
	ppm	700.0 ^c	726	± 36	14.4	pass
NO	ppm	150.3 ^a	152	± 9	4.0	pass
	ppm	301 ^d	304	± 16	6.9	pass
NO2	ppm	99.9 ^e	104.1	± 5.1	3.20	pass
SO2	ppm	100.1 ^f	98	± 6	3.5	pass
O2	Vol.-%	0.0 ^a	0.05	± 0.21	0.027	pass
	Vol.-%	2.510 ^b	2.55	± 0.21	0.055	pass
	Vol.-%	5.000 ^c	5.06	± 0.21	0.102	pass
Temperatur	°C	100.0 ^g	98.9	± 1.1	0.24	pass
	°C	200.0 ^h	199.9	± 1.1	0.24	pass
Druck	hPa	50.0 ^a	49.9	± 0.9	0.52	pass
	hPa	100.0 ^b	99.9	± 1.6	0.52	pass

² in accordance with the manufacturer (gemäß Hersteller)

Calibration certificate Kalibrier-Zertifikat

4321031

Special remarks Besondere Bemerkungen



PENTA
CALIBRATION

PENTA CALIBRATION CO., LTD.
66/124 The Connect 33 Village Kanchanaphisek Road
Dokmai Prawet Bangkok 10250
Tel: +66 (0) 2069-9773
www.pentalcal.com

Certificate of Calibration



Represent to Certificate of Calibration ,PTC/07/22071

Certificate No.: PTC/07/22071 Page: 1 of 2
Equipment: Digital Balance Condition: Normal
Manufacturer: Sartorius Serial No: 26207042
Model: MSE224-100-DU ID No: BKK_EN0002
Type of Balance: Single interval


Customer: ALS Laboratory Group (Thailand) Co.,Ltd.
104 Phatthanakarn 40 Phatthanakarn Rd.,
khwaeng Phatthanakarn, Khet Suan Luang, Bangkok 10250.

Environment Condition: Temperature 21.5 °C ± 0.7 °C
Humidity 61.8 %RH ± 4.7 %RH
Air density 1.19 kg/m³

Calibration Place: ALS Laboratory Group (Thailand) Co.,Ltd.
104 Phatthanakarn 40 Phatthanakarn Rd.,
khwaeng Phatthanakarn, Khet Suan Luang, Bangkok 10250.

REVIEW BY *Sarant M.*
APPROVED BY *KL AL*
NEXT CAL. DATE *25/02/23*

The Method used: In house method, PTC-WI-07, base on Euramet cg. 18
Traceability: This certificate is traceable to the SI Units through Thai Calibration Service Co.,Ltd.
, NSC-ONSC Accreditation No.: Calibration 0189
Date Received: February 25, 2022
Calibration Date: February 25, 2022
Issued Date: March 01, 2022
Calibration By: Mr. Rungroje Metakul



Mr. Kiangsak Kalasri
(Mr.Kiangsak Kalasri)
Reviewed by

Mr. Keattisak Kerdto
(Mr. Keattisak Kerdto)
Laboratory Manager

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognised national standard laboratories.

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM). The effect that the results relate only to the items calibrated.

This calibration certificate shall not be reproduced except in full only, without written approval from penta calibration co., ltd

PTC-FMC-07-02 2 Feb 2020

Testo Industrial Services GmbH

Overseasstraße 3 Tel: +49 7631 20601-8000 www.kontakt.de Page 3/3
79108 Kirschsteden Fax: +49 7631 20601-6010 info@kirschsteden.de



PENTA CALIBRATION CO., LTD.
66/124 The Connect 33 Village Kanchanaphisek Road
Dokmai Prawet Bangkok 10250
Tel: +66 (0) 2069-9773
www.pentalcal.com

Represent to Certificate of Calibration ,PTC/07/22071


Certificate No.: PTC/07/22071 Page 2 of 2

Measurement Results:

Without Adjustment :

Function Calibration: Non Adjustment

Eccentric Error: Weight to be 1/3 1/2 or of Maximum capacity



Eccentricity test 100 (g)				
Position (g)				
1	2	3	4	5
0.0000	-0.0002	-0.0001	0.0001	-0.0001
Maximum deviation:				0.0002

Repeatability Test : Weight to be 1/2 ≤ L, ≤ Maximum capacity

Determination of the standard deviation of weighing balance.. Readability 0.0001 (g)

Nominal test value (g)	Standard Deviation
200	0.00005

Error of indication : from nominal value.. Readability 0.0001 (g)

Nominal Value (g)	Conventional Mass (g)	Indication (g)	Correction of Balance (g)	Uncertainty (g)	k
0	0.00000	0.0000	0.0000	0.00016	2.52
0.1	0.10000	0.1000	0.0000	0.00017	2.20
0.5	0.50000	0.5000	0.0000	0.00016	2.28
1	1.00001	1.0000	0.0000	0.00016	2.28
2	2.00001	2.0000	0.0000	0.00016	2.28
5	5.00001	5.0000	0.0000	0.00016	2.28
10	10.00002	10.0000	0.0000	0.00016	2.28
20	20.00002	20.0000	0.0000	0.00016	2.23
50	50.00001	50.0000	0.0000	0.00017	2.15
100	100.00002	99.9999	0.0001	0.00020	2.06
120	120.00004	120.0000	0.0000	0.00023	2.03
150	150.00003	150.0000	0.0000	0.00026	2.00
200	200.00003	199.9999	0.0001	0.00030	2.00

Note: Weight of adjust (g)

The End of Certificate

PTC-FMC-07-02 2 Feb 2020



PENTA
CALIBRATION

PENTA CALIBRATION CO., LTD.
66/124 The Connect 33 Village Kanchanaphisek Road
Dokmai Prawet Bangkok 10250
Tel: +66 (0) 2069-9773
www.pentalcal.com

Certificate of Calibration



Represent to Certificate of Calibration ,PTC/07/22104

Certificate No.: PTC/07/22104 Page: 1 of 3
Equipment: Digital Balance Condition: Normal
Manufacturer: Sartorius Serial No: 33108993
Model: MSE125P-100-DU ID No: RYG_EN0004
Type of Balance: Single interval

Customer: ALS Laboratory Group (Thailand) Co.,Ltd.
616/10 Moo 5 T.Maenamkoo, A.Pluakdaeng,
Rayong 21140, Thailand

Environment Condition: Temperature 23.9 °C ± 0.3 °C
Humidity 58.1 %RH ± 4.4 %RH
Air density 1.17 kg/m³

Calibration Place: ALS Laboratory Group (Thailand) Co.,Ltd.
616/10 Moo 5 T.Maenamkoo, A.Pluakdaeng,
Rayong 21140, Thailand

REVIEW BY *Thanyakul*
APPROVED BY *P. S.*
NEXT CAL. DATE *28/06/23*

The Method used: In house method, PTC-WI-07, base on Euramet cg. 18
Traceability: This certificate is traceable to the SI Units through Thai Calibration Service Co.,Ltd.
, NSC-ONSC Accreditation No.: Calibration 0189

Date Received: March 23, 2022
Calibration Date: March 23, 2022
Issued Date: March 25, 2022
Calibration By: Mr. Rungroje Metakul



Mr. Kiangsak Kalasri
(Mr.Kiangsak Kalasri)
Reviewed by

Mr. Keattisak Kerdto
(Mr. Keattisak Kerdto)
Laboratory Manager

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognised national standard laboratories.

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM). The effect that the results relate only to the items calibrated.

This calibration certificate shall not be reproduced except in full only, without written approval from penta calibration co., ltd

PTC-FMC-07-02 2 Feb 2020

Represent to Certificate of Calibration ,PTC/07/22104

Certificate No.: PTC/07/22104


Page: 2 of 3

Measurement Results:

Without Adjustment :

Function Calibration: Non Adjustment

Eccentric Error: Weight to be 1/3, 1/2 or of Maximum capacity



Eccentricity test 50 (g)				
Position (g)				
1	2	3	4	5
0.00000	-0.00004	-0.00001	0.00000	0.00001
Maximum deviation: 0.00004				

Repeatability Test : Weight to be $1/2 \leq L_1 \leq$ Maximum capacity

Determination of the standard deviation of weighing balance., Readability 0.00001 (g)

Nominal test value (g)	Standard Deviation
50	0.000007

Error of indication : from nominal value., Readability 0.00001 (g)

Nominal Value (g)	Conventional Mass (g)	Indication (g)	Correction of Balance (g)	Uncertainty (g)	k
0	0.000000	0.00000	0.00000	0.000020	2.65
0.01	0.010001	0.01000	0.00000	0.000022	2.17
0.05	0.050002	0.04999	0.00001	0.000022	2.17
0.1	0.099999	0.09999	0.00001	0.000022	2.17
0.5	0.500001	0.50001	-0.00001	0.000022	2.17
1	1.000004	0.99999	0.00001	0.000022	2.14
2	1.999999	1.99999	0.00001	0.000022	2.14
5	5.000015	4.99999	0.00002	0.000023	2.14
10	10.000004	10.00000	0.00000	0.000024	2.10
20	20.000029	20.00000	0.00003	0.000032	2.00
50	50.000043	49.99999	0.00005	0.000069	2.00

Note: Weight of adjust - (g)

Represent to Certificate of Calibration ,PTC/07/22104

Certificate No.: PTC/07/22104


Page: 3 of 3

Measurement Results:

Without Adjustment :

Function Calibration: Non Adjustment

Eccentric Error: Weight to be 1/3, 1/2 or of Maximum capacity



Eccentricity test 50 (g)				
Position (g)				
1	2	3	4	5
0.0000	0.0000	0.0000	0.0000	0.0000
Maximum deviation: 0.0000				

Repeatability Test : Weight to be $1/2 \leq L_1 \leq$ Maximum capacity

Determination of the standard deviation of weighing balance., Readability 0.0001 (g)

Nominal test value (g)	Standard Deviation
100	0.00000

Error of indication : from nominal value., Readability 0.0001 (g)

Nominal Value (g)	Conventional Mass (g)	Indication (g)	Correction of Balance (g)	Uncertainty (g)	k
65	65.00006	65.0000	0.0001	0.00013	2.00
70	70.00007	70.0000	0.0001	0.00013	2.00
75	75.00009	75.0000	0.0001	0.00014	2.00
80	80.00008	80.0000	0.0001	0.00014	2.00
85	85.00009	85.0000	0.0001	0.00015	2.00
90	90.00010	90.0000	0.0001	0.00015	2.00
95	95.00012	95.0000	0.0001	0.00016	2.00
100	100.00004	100.0000	0.0000	0.00014	2.00
110	110.00004	110.0000	0.0000	0.00015	2.00
120	120.00007	120.0000	0.0001	0.00016	2.00

Note: Weight of adjust - (g)

The End of Certificate

PTC-FNC-02-02-24-01-2020

PTC-FNC-02-02-24-01-2020



ROTA METER CALIBRATION RESULT JULY 2022

Rotameter ID.	Calibration Date	Regression Result	Coefficient (R ²)
BKK_FS0577	01 Jul 22	Y = 1.0202x + 0.1976	1.0000
BKK_FS0579	01 Jul 22	Y = 1.0078x + 0.4769	0.9998
BKK_FS0583	01 Jul 22	Y = 1.016x + 0.3922	1.0000
BKK_FS0584	01 Jul 22	Y = 1.0036x + 2.2262	0.9997
BKK_FS0585	01 Jul 22	Y = 1.0189x - 5.6476	0.9997
BKK_FS0586	01 Jul 22	Y = 1.0095x - 1.1524	0.9995
BKK_FS0587	01 Jul 22	Y = 1.013x - 3.6619	0.9996
BKK_FS0588	01 Jul 22	Y = 1.0154x + 4.8357	0.9990
BKK_FS0589	01 Jul 22	Y = 0.9918x + 4.8069	0.9999
BKK_FS0590	01 Jul 22	Y = 1.0038x - 0.4857	0.9996
BKK_FS0591	01 Jul 22	Y = 0.9705x - 52.174	0.9986
BKK_FS0592	01 Jul 22	Y = 0.9646x - 37.642	0.9985
BKK_FS0593	01 Jul 22	Y = 0.9767x - 58.445	0.9988
BKK_FS0594	01 Jul 22	Y = 0.9902x - 62.87	0.9999
BKK_FS0595	01 Jul 22	Y = 1.0249x - 98.162	0.9999
BKK_FS0596	01 Jul 22	Y = 0.9843x - 26.806	0.9991
BKK_FS0597	01 Jul 22	Y = 0.9802x - 61.653	0.9978
BKK_FS1004	01 Jul 22	Y = 0.9696x + 17.69	0.9990
BKK_FS1005	01 Jul 22	Y = 1.0092x + 2.4571	0.9990
BKK_FS1006	01 Jul 22	Y = 1.168x - 5.566	0.9997
BKK_FS1007	01 Jul 22	Y = 0.9917x + 1.6592	1.0000
BKK_FS1008	01 Jul 22	Y = 1.0132x + 0.7207	1.0000
BKK_FS1009	01 Jul 22	Y = 1.0132x + 1.1633	0.9990
BKK_FS1010	01 Jul 22	Y = 1.0033x + 0.5758	0.9999
BKK_FS1011	01 Jul 22	Y = 1.0234x + 0.1759	0.9996
BKK_FS1012	01 Jul 22	Y = 1.0106x - 2.0048	0.9997
BKK_FS1013	01 Jul 22	Y = 0.9677x - 35.851	0.9997
BKK_FS1014	01 Jul 22	Y = 1.0021x + 0.3148	0.9998
BKK_FS1015	01 Jul 22	Y = 0.9994x + 1.786	1.0000
BKK_FS1016	01 Jul 22	Y = 1.0105x - 80.256	0.9998
BKK_FS1017	01 Jul 22	Y = 0.9995x + 0.649	1.0000
BKK_FS1018	01 Jul 22	Y = 1.0011x + 1.1786	1.0000
BKK_FS1019	01 Jul 22	Y = 1.0023x - 68.424	0.9996
BKK_FS1020	01 Jul 22	Y = 1.0547x - 0.666	0.9998
BKK_FS1021	01 Jul 22	Y = 1.018x - 3.3286	0.9998
BKK_FS1022	01 Jul 22	Y = 0.9932x - 57.035	0.9986
BKK_FS1023	01 Jul 22	Y = 1.0094x + 0.0717	0.9999
BKK_FS1024	01 Jul 22	Y = 1.0042x + 0.4086	0.9997
BKK_FS1025	01 Jul 22	Y = 1.0132x - 88.507	0.9996



ROTA METER CALIBRATION RESULT JULY 2022

Rotameter ID.	Calibration Date	Regression Result	Coefficient (R ²)
BKK_FS1026	01 Jul 22	Y = 1.0018x + 1.0776	0.9997
BKK_FS1027	01 Jul 22	Y = 1.0053x + 0.231	0.9995
BKK_FS1028	01 Jul 22	Y = 0.9792x - 60.312	0.9982
BKK_FS1029	01 Jul 22	Y = 0.9935x + 0.8234	1.0000
BKK_FS1030	01 Jul 22	Y = 1.0039x + 0.515	0.9999
BKK_FS1031	01 Jul 22	Y = 1.009x - 79.295	0.9998
BKK_FS1039	01 Jul 22	Y = 0.9879x + 7.3524	0.9996
BKK_FS1040	01 Jul 22	Y = 0.9704x + 88.336	0.9987
BKK_FS1041	01 Jul 22	Y = 1.0645x - 1.7878	0.9999
BKK_FS1042	01 Jul 22	Y = 0.9983x + 3.6262	0.9998
BKK_FS1043	01 Jul 22	Y = 1.0069x - 6.9619	1.0000
BKK_FS1044	01 Jul 22	Y = 1.0355x - 0.6214	0.9997
BKK_FS1161	01 Jul 22	Y = 1.0126x + 0.7738	0.9999
BKK_FS1162	01 Jul 22	Y = 0.9994x + 2.6357	0.9995
BKK_FS1163	01 Jul 22	Y = 0.977x - 55.03	0.9987
BKK_FS1164	01 Jul 22	Y = 0.9914x + 0.8427	0.9997
BKK_FS1165	01 Jul 22	Y = 0.9893x + 6.5919	0.9998
BKK_FS1166	01 Jul 22	Y = 1.0031x - 77.881	0.9996
BKK_FS1200	01 Jul 22	Y = 1.0313x - 0.4602	0.9995
BKK_FS1201	01 Jul 22	Y = 1.0045x + 0.15	0.9996
BKK_FS1202	01 Jul 22	Y = 0.9702x - 44.156	0.9994
RYG_FS0197	01 Jul 22	Y = 1.0039x - 0.179	0.9999
RYG_FS0198	01 Jul 22	Y = 0.9971x + 16.648	0.9999
RYG_FS0199	01 Jul 22	Y = 1.0832x - 2.6367	1.0000

Review By : Wichan Choonharat
(Mr. Wichan Choonharat)
Enviro Field Services Manager

Approved By : Mr. Sarayuth Jitranont
(Mr. Sarayuth Jitranont)
Assistant General Manager

Certificate of System Qualification

GC-QQ + GCMS-QQ

System ID: GM-2
Organization Name: ALS Laboratory Group (Thailand) Co., Ltd.
Organization Location: 104 Phatthanakan 40, Phattanakarn Rd., Kheiwang Sun Luang, Khel Sun Luang, Bangkok 10250
Date: October 1, 2021 1:10:17 PM
EOP Name: AgilentRecommended, AgilentRecommended
EOP Revision: GC.02.51, GCMS.02.51
Overall Qualification Status: Pass

REVIEW BY: Santana M.
APPROVED BY: Al
CAL DATE: 1 April 23

System Inspection and Basic Safety and Operation

Name: 7890
Setpoint Status: Pass

Overall System Inspection and Basic Safety and Operation Test Status
Pass

Inlet Pressure Accuracy

Name: 7890
Front MMI
Setpoint Status: Pass
Setpoint Actual
Inlet Pressure: 25.0 psi 24.9 psi
Accuracy: 0.1 psi
Agilent Recommended: <= 1.2

Overall Inlet Pressure Accuracy Test Status
Pass

GC Oven Temperature Accuracy

Name: 7890

Date: October 1, 2021 1:10:17 PM
System ID: GM-2

Setpoint Status: Pass
Zone: Oven
Setpoint/Actual
Temperature: 230.0 230.5 °C
Accuracy: 0.5 °C
Agilent Recommended: >= -1.0 % setpoint in K (-5.0 °C)
<= 1.0 % setpoint in K (5.0 °C)

Setpoint Status: Pass
Zone: Oven
Setpoint/Actual
Temperature: 100.0 101.5 °C
Accuracy: 1.5 °C
Agilent Recommended: >= -1.0 % setpoint in K (-3.7 °C)
<= 1.0 % setpoint in K (3.7 °C)

Overall GC Oven Temperature Accuracy Test Status

Pass

GC Oven Temperature Stability

Name: 7890
Setpoint Status: Pass
Setpoint/Average
Temperature: 100.0 101.5 °C
Stability: 0.0 °C
Agilent Recommended: <= 0.5

Overall GC Oven Temperature Stability Test Status

Pass

Log Amp

Tested Combination1 Front MMI / External SQ
Name: 5975C Inert XL with TAD
Setpoint Status: Pass

Date: October 1, 2021 1:10:17 PM
System ID: GM-2

Overall Log Amp Test Status

Pass

RPPA

Tested Combination1 Front MMI / External SQ
Name: 5975C Inert XL with TAD
Setpoint Status: Pass
Amu: 1050 m/z
Drift After Five Minutes: 6 mV
RPPA Voltage: 461 mV
Agilent Recommended: >= -100 and <= 100 <= 1100

Overall RPPA Test Status

Pass

Tune EI

Tested Combination1 Front MMI / External SQ
Name: 5975C Inert XL with TAD
Setpoint Status: Pass
Filament: 1
Setpoint Status: Pass
Filament: 2

Overall Tune EI Test Status

Pass

Scouting Run

Tested Combination1 Front MMI / External SQ
Injection Tower
Name: 7693A
Source: EI - Inert

Date: October 1, 2021 1:10:17 PM
System ID: GM-2

Setpoint Status: Completed
Injection Volume on Column: 1.0 µL

Overall Scouting Run Status

Completed

Signal to Noise EI

Tested Combination1 Front MMI / External SQ
Name: 5975C Inert XL with TAD
Source: EI - Inert Filament: 1
Setpoint Status: Pass
Signal to Noise: 619
Agilent Recommended: >= 320
Source: EI - Inert Filament: 2
Setpoint Status: Pass
Signal to Noise: 647
Agilent Recommended: >= 320

Overall Signal to Noise EI Test Status

Pass

Injection Precision

Tested Combination1 Front MMI / External SQ
Name: 7693A
Source: EI - Inert
Setpoint Status: Pass
Injection Volume on Column: 1.0 µL
Area RSD: 4.75 % Retention Time RSD: 0.02 %
Agilent Recommended: <= 5.00 <= 1.00

Overall Injection Precision Test Status

Pass

Date: October 1, 2021 1:10:17 PM
System ID: GM-2

Mass Ratio Precision

Tested Combination1

Front MMI / External SQ

Injection Tower

Name: 7693A

Source: EI - Inert

Setpoint Status: Pass

Injection Volume on Column: 1.0 µL

Area Mass 1

Abundance's

4.75 %

<= 5.00

Pass

Mass Ratio

0.81 %

<= 5.00

Pass

RSD:

Agilent Recommended:

Overall Mass Ratio Precision Test Status

Pass

Date: October 1, 2021 1:10:17 PM
System ID: GM-2

Page 5 / 15

Instrument Details

Purpose

This section describes the as found system configuration.

Details

System

System ID GM-2
Manufacturer Agilent Technologies
Name 7690
Flow Data Input Manual Data
Temperature Data Input Manual Data or Other Data Logging

Tested Combination1

Injection Technique Injection Tower
Inlet Front
Detector External
LTM Included? No

Sampler 1

Manufacturer Agilent Technologies
Type Injection Tower
Name 7693A
Model Number G4513A
Serial Number CN10120123
Firmware Revision A.10.08
Usage Sample Injection
Location Front
Syringe Volume (µL) 10Date: October 1, 2021 1:10:17 PM
System ID: GM-2

Page 6 / 16

Sampler 2

Manufacturer Agilent Technologies
Type Tray
Name 7693A
Model Number G4514A
Serial Number CN10060009
Firmware Revision A.10.16
Vial Heater Not installed

Mainframe 1

Manufacturer Agilent Technologies
Name 7890
Model Number G3440A
Serial Number CN10141049
Firmware Revision A.01.16
Oven Type Standard

Inlet 1

Manufacturer Agilent Technologies
Name 7890
Type MMI
Location Front
Carrier Gas Helium
Control Type Electronic Pressure Control (EPC)
Purged Inlet Yes

Detector 1

Manufacturer Agilent Technologies
Name Mass Spectrometer
Type Mass Spectrometer
Location ExternalDate: October 1, 2021 1:10:17 PM
System ID: GM-2

Page 7 / 19

Mass Spectrometer 1

Manufacturer Agilent Technologies
Type SQ
Name 5975C Inert XL with TAD
Serial Number US10153217
Firmware Revision 5.02.12
High Vacuum System Turbo Pump
Scouting Run Standard OFN Std

MS EI Source 1

Manufacturer Agilent Technologies
Source Type EI - Inert
Number of filaments 2Date: October 1, 2021 1:10:17 PM
System ID: GM-2

Page 8 / 16

Electronic Signature

Purpose

This signature page was created and published because the ACE sign-off action was executed, which is valid for the entire document, including attachments. The ACE sign-off is an electronic signature that requires two distinct identification components: unique username and personal password. The Agilent representative who has delivered this service understands the meaning and legal status of an electronic signature. As a trained official operator, the Agilent representative has a unique password and login to access ACE and electronically sign this document. (Other e-signatures can be applied to this document using a Document Content Management or other suitable method defined in your data access and control procedures.)

Details

Full Name of Signer:	Supasak Nimsongtham
Logged On User Name:	supasak.nimsongtham@agilent.com
Signature Creation Date:	October 1, 2021
Reason for Signature:	Executed protocol and published this original version of document

Regulatory Disclaimer

This document provides a protocol to verify and record instrument configuration and evidence of proper operation. It has been prepared from our interpretation of applicable regulations as well as industry best practices. The document is designed to provide an important component of a complete compliance package. Validation depends upon many factors and use of this protocol alone does not assure compliance. Agilent Technologies makes no promises or representations as to its sufficiency for any specific regulatory program.

Warranty

Agilent Technologies makes no warranty of any kind to this material, including but not limited to, the implied warranties or merchantability and fitness for a particular purpose. Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Date: October 1, 2021 1:10:17 PM
System ID: GM-2

Page 9 / 16

User Name: supakorn.nimnangtham

Host Name: SCG115HKC

System Id: 610

Print Date: October 1, 2021 1:10:19 PM

ALIS_GMS Transaction log:

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 1, 2021 12:42:37 PM	Audit	Session Created	Session	None
October 1, 2021 12:42:37 PM	Start	Configuration	Session	None
October 1, 2021 12:42:37 PM	Audit	Entitlement	Licensing	User is FieldEngineer and does not require an unlock code
October 1, 2021 12:44:21 PM	Audit	EspLoaded	Session	EQP details for primary technique [Gc] - File path: P:\Process\ProData\ToC\Conf\gms02.S1VG\02.S1eq.eqp EQP File Name: [Gc.02.S1eq.eqp] EQP Name: [AgentRecommended] EQP details for hypothesisid technique [Gc] - File path: P:\Process\ProData\ToC\Conf\gms02.S1VG\02.S1eq.eqp EQP File Name: [Gc.S1eq.02.S1eq.eqp] EQP Name: [AgentRecommended]
October 1, 2021 12:44:24 PM	End	Configuration	Session	None
October 1, 2021 12:44:28 PM	Start	Qualification	Session	QQ
October 1, 2021 12:44:28 PM	Start	Execution	System Inspection and Basic Safety and Operation - 7800 - Qualitative Test - No serious concerns	None

Page 1/7

Date: October 1, 2021 1:10:17 PM
System ID: GM-2

Page 10 / 16

User Name: supasak.hosinsomgitham

Host Name: SGC6115HNC

System ID: GMS-1

Print Date: October 1, 2021 1:10:19 PM

ALS_GM2 Transaction log :

Time	Transaction State	Activity Performed	Optional Information
October 1, 2021 12:47:35 PM	End	Execution	System Inspection and Static Safety and Operation - 7890 - Qualitative Test - No warnings associated Run Count : 1
October 1, 2021 12:47:37 PM	Start	Execution	Inlet Pressure Accuracy - Front MM: - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi None
October 1, 2021 12:47:42 PM	End	Execution	Inlet Pressure Accuracy - Front MM: - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi Run Count : 1
October 1, 2021 12:47:44 PM	Start	Execution	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K None
October 1, 2021 12:48:04 PM	Audit	Data	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K Manual Data Entry
October 1, 2021 12:48:05 PM	End	Execution	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K Run Count : 1
October 1, 2021 12:48:07 PM	Start	Execution	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K None
October 1, 2021 12:48:34 PM	Audit	Data	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K Manual Data Entry
October 1, 2021 12:48:36 PM	End	Execution	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K Run Count : 1

Page 212

Date: October 1, 2021 1:10:17 PM
System ID: GM-2

Page 11 / 16

User Name: kushalsamkanchim

System ID: 99

Hostname: SC01115HSC

Print Date: October 1, 2021 10:10 PM

ALM_0M2 Transaction log:

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 1, 2021 12:48:38 PM	Start	Execution	GC Oven Temperature Stability -7890 - Temperature - Oven - S: 100.0°C - L: <= 0.5°C	None
October 1, 2021 12:48:34 PM	Abort	Data	GC Oven Temperature Stability -7890 - Temperature - Oven - S: 100.0°C - L: <= 0.5°C	Manual Data Entry
October 1, 2021 12:49:36 PM	End	Execution	GC Oven Temperature Stability -7890 - Temperature - Oven - S: 100.0°C - L: <= 0.5°C	Run Count: 1
October 1, 2021 12:49:37 PM	Start	Execution	Log Amp - 5975C Inert XL with TAD SQ - Source: EI - Inert	None
October 1, 2021 12:49:47 PM	End	Execution	Log Amp - 5975C Inert XL with TAD SQ - Source: EI - Inert	Run Count: 1
October 1, 2021 12:49:48 PM	Start	Execution	RFP4 - 5975C Inert XL with TAD SQ - Source: EI - Inert	None
October 1, 2021 12:50:23 PM	End	Execution	RFP4 - 5975C Inert XL with TAD SQ - Source: EI - Inert	Run Count: 1
October 1, 2021 12:50:25 PM	Start	Execution	Tune EI - 5975C Inert XL with TAD SQ - Source: EI - Inert Flameant 1 (Qualitative - No separata associated)	None
October 1, 2021 12:50:49 PM	End	Execution	Tune EI - 5975C Inert XL with TAD SQ - Source: EI - Inert Flameant 1 (Qualitative - No separata associated)	Run Count: 1
October 1, 2021 12:50:50 PM	Start	Execution	Tune EI - 5975C Inert XL with TAD SQ - Source: EI - Inert Flameant 2 (Qualitative - No separata associated)	None
October 1, 2021 12:50:53 PM	End	Execution	Tune EI - 5975C Inert XL with TAD SQ - Source: EI - Inert Flameant 2 (Qualitative - No separata associated)	Run Count: 1

Page 3 / 7

Date: October 1, 2021 1:10:17 PM
System ID: GM-2

Page 12 / 16

User Name: supasak.kunisingtham

Hostname: SCG115HNC

System ID: GM-2

Print Date: October 1, 2021 1:10:19 PM

ALS_GW2 Transaction log:

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 1, 2021 12:51:01 PM	Start	Execution	Scouting Run - Injection Tower, Front MM, SQ - Source - EI - Inert-Part of GCMS System Preparation	None
October 1, 2021 12:51:18 PM	Audit	Data	Scouting Run - Injection Tower, Front MM, SQ - Source - EI - Inert-Part of GCMS System Preparation	Data File Path: E:\GM2002021\SCOUTING\RUN001.D\DATA.MS
October 1, 2021 12:51:42 PM	Audit	Data	Scouting Run - Injection Tower, Front MM, SQ - Source - EI - Inert-Part of GCMS System Preparation	Data File Path: E:\GM2002021\SCOUTING\RUN001.D\DATA.MS
October 1, 2021 12:52:42 PM	Audit	Data	Scouting Run - Injection Tower, Front MM, SQ - Source - EI - Inert-Part of GCMS System Preparation	Data File Path: E:\GM2002021\SCOUTING\RUN001.D\DATA.MS
October 1, 2021 12:53:25 PM	End	Execution	Scouting Run - Injection Tower, Front MM, SQ - Source - EI - Inert-Part of GCMS System Preparation	Run Count: 1
October 1, 2021 12:53:27 PM	Start	Execution	Signal to Noise EI - Injection Tower, Front MM, SQ - Source: EI - Inert using Filament 1 - L >= 320	None
October 1, 2021 12:53:40 PM	Audit	Data	Signal to Noise EI - Injection Tower, Front MM, SQ - Source: EI - Inert using Filament 1 - L >= 320	Data File Path: E:\GM2002021\SNF1_001.D\DATA.MS
October 1, 2021 12:53:56 PM	End	Execution	Signal to Noise EI - Injection Tower, Front MM, SQ - Source: EI - Inert using Filament 1 - L >= 320	Run Count: 1

Page 4 / 7

Page 4 / 7

User Name: supasak.kunisingtham

Hostname: SCG115HNC

System ID: GM-2

Print Date: October 1, 2021 1:10:19 PM

ALS_GW2 Transaction log:

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 1, 2021 12:53:59 PM	Start	Execution	Signal to Noise EI - Injection Tower, Front MM, SQ - Source: EI - Inert using Filament 2 - L >= 320	None
October 1, 2021 12:54:04 PM	Audit	Data	Signal to Noise EI - Injection Tower, Front MM, SQ - Source: EI - Inert using Filament 2 - L >= 320	Data File Path: E:\GM2002021\SNF2_001.D\DATA.MS
October 1, 2021 12:54:32 PM	End	Execution	Signal to Noise EI - Injection Tower, Front MM, SQ - Source: EI - Inert using Filament 2 - L >= 320	Run Count: 1
October 1, 2021 12:54:35 PM	Start	Execution	Injection Precision - Injection Tower, Front MM, SQ - Source: - EI - Inert L (Area): <= 5.00% - L (Ret. Time): <= 1.00%	None
October 1, 2021 12:54:37 PM	Audit	Data	Injection Precision - Injection Tower, Front MM, SQ - Source: - EI - Inert L (Area): <= 5.00% - L (Ret. Time): <= 1.00%	Data File Path: E:\GM2002021\IP_MRP003.D\DATA.MS
October 1, 2021 12:54:37 PM	Audit	Data	Injection Precision - Injection Tower, Front MM, SQ - Source: - EI - Inert L (Area): <= 5.00% - L (Ret. Time): <= 1.00%	Data File Path: E:\GM2002021\IP_MRP004.D\DATA.MS
October 1, 2021 12:54:37 PM	Audit	Data	Injection Precision - Injection Tower, Front MM, SQ - Source: - EI - Inert L (Area): <= 5.00% - L (Ret. Time): <= 1.00%	Data File Path: E:\GM2002021\IP_MRP005.D\DATA.MS
October 1, 2021 12:54:37 PM	Audit	Data	Injection Precision - Injection Tower, Front MM, SQ - Source: - EI - Inert L (Area): <= 5.00% - L (Ret. Time): <= 1.00%	Data File Path: E:\GM2002021\IP_MRP006.D\DATA.MS

Page 5 / 7

Page 5 / 7

Date: October 1, 2021 1:10:17 PM
System ID: GM-2

Page 13 / 10

Date: October 1, 2021 1:10:17 PM
System ID: GM-2

Page 14 / 16

User Name: supasak.kunisingtham

Hostname: SCG115HNC

System ID: GM-2

Print Date: October 1, 2021 1:10:19 PM

ALS_GW2 Transaction log:

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 1, 2021 12:54:37 PM	Audit	Data	Injection Precision - Injection Tower, Front MM, SQ - Source: - EI - Inert L (Area): <= 5.00% - L (Ret. Time): <= 1.00%	Data File Path: E:\GM2002021\IP_MRP007.D\DATA.MS
October 1, 2021 12:54:37 PM	Audit	Data	Injection Precision - Injection Tower, Front MM, SQ - Source: - EI - Inert L (Area): <= 5.00% - L (Ret. Time): <= 1.00%	Data File Path: E:\GM2002021\IP_MRP008.D\DATA.MS
October 1, 2021 12:54:52 PM	End	Execution	Injection Precision - Injection Tower, Front MM, SQ - Source: - EI - Inert L (Area): <= 5.00% - L (Ret. Time): <= 1.00%	Run Count: 1
October 1, 2021 12:54:55 PM	Start	Execution	Mass Ratio Precision - Injection Tower, Front MM, SQ - Source: EI - Inert - L (RSD): <= 5.00%	None
October 1, 2021 12:55:06 PM	Audit	Data	Mass Ratio Precision - Injection Tower, Front MM, SQ - Source: EI - Inert - L (RSD): <= 5.00%	Data File Path: E:\GM2002021\IP_MRP003.D\DATA.MS
October 1, 2021 12:55:08 PM	Audit	Data	Mass Ratio Precision - Injection Tower, Front MM, SQ - Source: EI - Inert - L (RSD): <= 5.00%	Data File Path: E:\GM2002021\IP_MRP004.D\DATA.MS
October 1, 2021 12:55:06 PM	Audit	Data	Mass Ratio Precision - Injection Tower, Front MM, SQ - Source: EI - Inert - L (RSD): <= 5.00%	Data File Path: E:\GM2002021\IP_MRP005.D\DATA.MS
October 1, 2021 12:55:06 PM	Audit	Data	Mass Ratio Precision - Injection Tower, Front MM, SQ - Source: EI - Inert - L (RSD): <= 5.00%	Data File Path: E:\GM2002021\IP_MRP006.D\DATA.MS

Page 6 / 7

Page 6 / 7

User Name: supasak.nimsongitam Hostname: SCG115HNC		System Id: GM-2 Print Date: October 1, 2021 1:10:19 PM		
ALS_GW2 Transaction log:				
Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 1, 2021 12:55:06 PM	Audit	Data	Mass Ratio Precision - Injection Tower, Front MM, SQ - Source: EI - Inert - L (RSD): <= 5.00%	Data file Path: E:\GM2002021\IP_MRP007.D\DATA.MS
October 1, 2021 12:55:08 PM	Audit	Data	Mass Ratio Precision - Injection Tower, Front MM, SQ - Source: EI - Inert - L (RSD): <= 5.00%	Data file Path: E:\GM2002021\IP_MRP008.D\DATA.MS
October 1, 2021 12:55:10 PM	End	Execution	Mass Ratio Precision - Injection Tower, Front MM, SQ - Source: EI - Inert - L (RSD): <= 5.00%	Run Count: 1
October 1, 2021 12:55:13 PM	End	Qualification	Session	OD
October 1, 2021 12:55:13 PM	Start	Reporting	Session	None
October 1, 2021 1:09:11 PM	Audit	Reporting	Session	Report Generated: Certificate

Page 7 / 7

Page 7 / 7

Date: October 1, 2021 1:10:17 PM
System ID: GM-2

Page 15 / 10

Date: October 1, 2021 1:10:17 PM
System ID: GM-2

Page 16 / 16



EMISSION TEST RESULT

Client	GHECO-ONE Co., Ltd.	Run #	3
Date	19 Aug 22	Location	1st Furnace PC Boiler
Start Time	11:42	Test Operator	Arurat M.
SO ₂ Analyzer Model	TELEDYNE API T100H	Finish Time	12:02
NO _x /O ₂ Analyzer Model	TELEDYNE API T200H	Serial No.	324
CO/CO ₂ Analyzer Model	TELEDYNE API T300M	Serial No.	482
		Serial No.	377

Time (min)	O ₂ (%)	CO ₂ (%)	NO _x (ppm)	SO ₂ (ppm)	CO (ppm)	Remark
11:42	5.34	14.88	47.19	10.65	13.01	
11:43	5.35	14.88	47.10	10.62	10.05	
11:44	5.33	14.82	47.45	10.67	6.56	
11:45	5.31	14.92	48.14	10.69	7.30	
11:46	5.32	14.90	48.64	10.65	12.49	
11:47	5.32	14.98	49.74	10.74	13.52	
11:48	5.35	14.99	49.43	11.38	11.88	
11:49	5.29	14.93	48.94	11.05	10.45	
11:50	5.24	14.85	48.54	10.79	9.84	
11:51	5.29	14.87	47.71	10.71	12.00	
11:52	5.35	14.91	46.44	10.83	11.25	
11:53	5.33	14.83	45.28	10.80	12.31	
11:54	5.32	14.93	45.42	10.72	13.64	
11:55	5.29	14.79	47.01	10.88	9.98	
11:56	5.31	14.85	48.89	10.89	8.30	
11:57	5.33	15.00	49.93	11.13	13.62	
11:58	5.35	14.77	50.22	11.25	11.37	
11:59	5.34	14.73	50.04	11.44	7.02	
12:00	5.32	14.94	49.41	11.38	7.49	
12:01	5.32	14.95	49.28	11.48	6.68	
12:02	5.34	15.08	49.80	11.49	6.06	
Average	5.32	14.89	48.32	10.98	10.23	

Arurat M

(Mr. Arurat Moungphai)

Environmental Field Scientist (2)

FORM NO. F-08-02 REVISION NO. 2 ISSUE DATE 3/05/19

ALS Laboratory Group

PM CEMs Data (@7%O₂)

Client Name	GHECO-ONE CO., LTD.	Date	15-Nov-22
Plant Name	GHECO-ONE	Location	PC Boiler

Run No.1			Run No.2			Run No.3			Run No.4			Run No.5		
Time	Dust	Opacity	Time	Dust	Opacity	Time	Dust	Opacity	Time	Dust	Opacity	Time	Dust	Opacity
mg/m ³	%		mg/m ³	%		mg/m ³	%		mg/m ³	%		mg/m ³	%	
11:25	3.10	1.75	12:30	3.10	1.80	13:30	3.20	2.13	14:30	3.10	1.90	15:30	3.20	2.07
11:26	3.10	1.70	12:31	3.10	1.79	13:31	3.20	2.08	14:31	3.10	1.86	15:31	3.20	2.09
11:27	3.10	1.82	12:32	3.10	1.83	13:32	3.10	2.02	14:32	3.10	1.94	15:32	3.10	2.03
11:28	3.10	1.75	12:33	3.20	1.92	13:33	3.10	1.91	14:33	3.20	2.15	15:33	3.20	2.21
11:29	3.10	1.69	12:34	3.20	1.98	13:34	3.10	1.87	14:34	3.20	2.54	15:34	3.20	2.20
11:30	3.10	1.56	12:35	3.20	1.93	13:35	3.10	1.82	14:35	3.20	2.49	15:35	3.20	2.15
11:31	3.00	1.56	12:36	3.20	1.98	13:36	3.10	2.04	14:36	3.20	2.40	15:36	3.10	2.10
11:32	3.00	1.56	12:37	3.10	1.93	13:37	3.20	2.06	14:37	3.20	2.33	15:37	3.10	2.10
11:33	3.10	1.60	12:38	3.10	1.96	13:38	3.10	1.96	14:38	3.20	2.28	15:38	3.10	2.13
11:34	3.10	1.71	12:39	3.10	1.95	13:39	3.20	1.94	14:39	3.20	2.45	15:39	3.10	2.22
11:35	3.20	1.73	12:40	3.10	1.82	13:40	3.10	1.98	14:40	3.20	2.57	15:40	3.20	2.15
11:36	3.20	1.73	12:41	3.10	1.73	13:41	3.10	1.94	14:41	3.20	2.48	15:41	3.20	2.17
11:37	3.20	1.87	12:42	3.10	1.82	13:42	3.10	1.99	14:42	3.10	2.33	15:42	3.20	2.17
11:38	3.20	1.87	12:43	3.10	1.90	13:43	3.10	1.98	14:43	3.10	2.25	15:43	3.20	2.26
11:39	3.00	1.93	12:44	3.10	1.81	13:44	3.10	1.92	14:44	3.10	2.21	15:44	3.20	2.25
11:40	3.10	1.86	12:45	3.10	2.11	13:45	3.10	1.77	14:45	3.10	2.19	15:45	3.20	2.35
11:41	3.20	1.81	12:46	3.10	2.04	13:46	3.10	1.79	14:46	3.20	2.25	15:46	3.20	2.48
11:42	3.20	1.86	12:47	3.10	1.84	13:47	3.10	1.70	14:47	3.20	2.44	15:47	3.20	2.47
11:43	3.10	1.86	12:48	3.10	1.80	13:48	3.10	1.70	14:48	3.20	2.65	15:48	3.20	2.53
11:44	3.10	2.03	12:49	3.10	1.88	13:49	3.10	1.79	14:49	3.20	2.58	15:49	3.20	2.59
11:45	3.10	2.18	12:50	3.10	1.83	13:50	3.10	1.95	14:50	3.20	2.40	15:50	3.20	2.54
11:46	3.10	2.02	12:51	3.20	1.80	13:51	3.10	1.99	14:51	3.20	2.35	15:51	3.20	2.37
11:47	3.10	2.03	12:52	3.10	1.79	13:52	3.10	1.98	14:52	3.10	2.32	15:52	3.20	2.41
11:48	3.20	2.07	12:53	3.10	1.84	13:53	3.10	2.09	14:53	3.20	2.34	15:53	3.20	2.40
11:49	3.20	1.99	12:54	3.10	1.85	13:54	3.20	2.36	14:54	3.20	2.45	15:54	3.20	2.38
11:50	3.20	2.08	12:55	3.10	1.98	13:55	3.20	2.20	14:55	3.20	2.33	15:55	3.20	2.36
11:51	3.20	2.34	12:56	3.10	2.11	13:56	3.20	2.11	14:56	3.10	2.16	15:56	3.20	2.30
11:52	3.10	2.11	12:57	3.10	1.93	13:57	3.20	2.17	14:57	3.10	2.07	15:57	3.20	2.35
11:53	3.10	1.89	12:58	3.10	2.00	13:58	3.10	2.14	14:58	3.10	2.12	15:58	3.20	2.26
11:54	3.00	1.73	12:59	3.10	1.95	13:59	3.10	2.11	14:59	3.10	2.14	15:59	3.20	2.19
11:55	3.10	1.81	13:00	3.10	1.92	14:00	3.10	2.14	15:00	3.20	2.16	16:00	3.20	2.36
11:56	3.10	1.87	13:01	3.10	1.95	14:01	3.10	2.02	15:01	3.10	2.15	16:01	3.20	2.36
11:57	3.10	1.92	13:02	3.20	2.19	14:02	3.20	2.05	15:02	3.10	2.13	16:02	3.20	2.27
11:58	3.10	1.84	13:03	3.20	2.24	14:03	3.20	2.15	15:03	3.10	2.13	16:03	3.20	2.29
11:59	3.10	1.86	13:04	3.20	2.08	14:04	3.20	2.30	15:04	3.10	2.14	16:04	3.20	2.37
12:00	3.00	1.68	13:05	3.10	2.04	14:05	3.20	2.27	15:05	3.20	2.43	16:05	3.20	2.31
12:01	3.10	1.57	13:06	3.10	1.96	14:06	3.20	2.31	15:06	3.20	2.38	16:06	3.20	2.26
12:02	3.10	2.02	13:07	3.10	1.97	14:07	3.20	2.29	15:07	3.10	2.18	16:07	3.20	2.32
12:03	3.10	2.04	13:08	3.20	1.95	14:08	3.10	2.37	15:08	3.10	2.16	16:08	3.20	2.22
12:04	3.10	2.01	13:09	3.20	2.21	14:09	3.10	2.31	15:09	3.20	2.14	16:09	3.10	2.23
12:05	3.10	1.95	13:10	3.20	2.04	14:10	3.10	2.22	15:10	3.20	2.11	16:10	3.20	2.25
12:06	3.10	1.84	13:11	3.10	1.96	14:11	3.10	2.19	15:11	3.20	2.12	16:11	3.20	2.18
12:07	3.10	1.79	13:12	3.10	1.86	14:12	3.10	2.23	15:12	3.10	2.08	16:12	3.10	2.09
12:08	3.10	1.80	13:13	3.10	1.87	14:13	3.20	2.26	15:13	3.10	1.97	16:13	3.20	2.11
12:09	3.10	1.93	13:14	3.10	1.87	14:14	3.20	2.27	15:14	3.10	2.05	16:14	3.20	2.31
12:10	3.10	2.02	13:15	3.10	1.85	14:15	3.10	2.29	15:15	3.10	1.95	16:15	3.10	2.29
12:11	3.10	1.83	13:16	3.10	1.86	14:16	3.10	2.40	15:16	3.10	1.97	16:16	3.10	2.29
12:12	3.10	1.70	13:17	3.10	1.85	14:17	3.20	2.35	15:17	3.10	1.94	16:17	3.10	2.16
12:13	3.10	1.72	13:18	3.10	1.82	14:18	3.20	2.23	15:18	3.20	1.98	16:18	3.10	2.17
Avg.	3.11	1.85	Avg.	3.12	1.93	Avg.	3.14	2.08	Avg.	3.15	2.23	Avg.	3.18	2.27

CERTIFICATE OF ANALYSIS
Grade of Product: EPA Protocol

Part Number:	E04N199E15A0664	Reference Number:	160-401907846-1
Cylinder Number:	EB0136209	Cylinder Volume:	144.4 CF
Laboratory:	124 - Plumsteadville - PA	Cylinder Pressure:	2015 PSIG
PGVP Number:	A12020	Valve Outlet:	650
Gas Code:	CO,NO,NOX,SO2,BALN	Certification Date:	Oct 06, 2020

Expiration Date: Oct 06, 2028

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 820-R-12-031, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a mole/mole basis unless otherwise noted.

Do Not Use This Cylinder below 100 psig i.e. 0.7 megapascals

Component	Requested Concentration	Actual Concentration	Protocol Method	Total Relative Uncertainty	Assay Dates
NOX	55.00 PPM	54.64 PPM	G1	+/- 1.3% NIST Traceable	09/29/2020, 10/06/2020
CARBON MONOXIDE	55.00 PPM	54.42 PPM	G1	+/- 0.8% NIST Traceable	09/29/2020
NITRIC OXIDE	55.00 PPM	54.64 PPM	G1	+/- 1.3% NIST Traceable	09/29/2020, 10/06/2020
SULFUR DIOXIDE	55.00 PPM	54.34 PPM	G1	+/- 1.0% NIST Traceable	09/29/2020, 10/06/2020
NITROGEN	Balance				


Type	Lot ID	Cylinder No	Concentration	Uncertainty	Expiration Date
NTRM	11010130	KAL004536	97.31 PPM CARBON MONOXIDE/NITROGEN	+/- 0.4%	Oct 04, 2022
PRM	12365	D656025	9.91 PPM AIR/NITROGEN DIOXIDE	2.0%	Feb 20, 2020
NTRM	17060222	EB0078169	100.3 PPM NITRIC OXIDE/NITROGEN	+/- 1.0%	Jul 23, 2023
GMIS	124200889	CC0323707	4.026 PPM NITROGEN DIOXIDE/NITROGEN	2.1%	Aug 15, 2021
NTRM	11010416	KAL004802	59.6 PPM SULFUR DIOXIDE/NITROGEN	+/- 0.8%	Jul 23, 2023
NTRM	16010203	KAL003087	97.69 PPM SULFUR DIOXIDE/NITROGEN	+/- 0.8%	Dec 23, 2021

The SRM, PRM or RGM noted above is only in reference to the GMIS used in the assay and not part of the analysis.

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
Nicolet iS50 FTIR AUP2010245 CO	FTIR	Sep 21, 2020
Nicolet iS50 FTIR AUP2010245 NO	FTIR	Sep 14, 2020
Nicolet iS50 FTIR AUP2010245 NO2	FTIR	Sep 22, 2020
Nicolet iS50 FTIR AUP2010245 SO2	FTIR	Sep 16, 2020

Triad Data Available Upon Request

NOTES: Gross Weight: 27.8 Kg, Net Weight: 4.6 Kg



Approved for Release

Page 1 of 160-401907846-1

CERTIFICATE OF ANALYSIS
Grade of Product: EPA Protocol

Part Number:	E04N199E3HA0002	Reference Number:	160-402138465-1
Cylinder Number:	ND11218	Cylinder Volume:	247.2 Cubic Feet
Laboratory:	124 - Plumsteadville - PA	Cylinder Pressure:	2215 PSIG
PGVP Number:		Valve Outlet:	660
Gas Code:	CO,NO,NOX,SO2,BALN	Certification Date:	Jul 15, 2021

CERTIFICATE OF ANALYSIS

Grade of Product: EPA Protocol

Part Number: E02N184E15A07B7 Reference Number: 160-401948145-1
Cylinder Number: CC740041 Cylinder Volume: 145.8 CF
Laboratory: 124 - Plumsteadville - PA Cylinder Pressure: 2015 PSIG
PGVP Number: A12020 Valve Outlet: 590
Gas Code: O2,BALN Certification Date: Nov 11, 2020

Expiration Date: Nov 11, 2028

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 600/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analyte uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a mole/mole basis unless otherwise noted.

Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

ANALYTICAL RESULTS				
Component	Requested Concentration	Actual Concentration	Protocol Method	Total Relative Uncertainty
OXYGEN	16.00 %	16.17 %	G1	+/- 0.3% NIST Traceable
NITROGEN	Balance			

CALIBRATION STANDARDS				
Type	Lot ID	Cylinder No	Concentration	Uncertainty
NTRM	16060503	CC108542	23.204 % OXYGEN/NITROGEN	+/- 0.2% Dec 24, 2021

ANALYTICAL EQUIPMENT		
Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
SIEMENS OXYMAT 6 - N1-W5-951 - O2	PARAMAGNETIC	Oct 26, 2020

Triad Data Available Upon Request

NOTES:
Gross Weight: 27.8 Kg
Net Weight: 4.7 Kg



Approved for Release

Page 1 of 160-401948145-1

CERTIFICATE OF ANALYSIS

Grade of Product: EPA Protocol

Part Number: E02N192E3HA0000 Reference Number: 160-401948144-1
Cylinder Number: GN0025083 Cylinder Volume: 248.4 CF
Laboratory: 124 - Plumsteadville - PA Cylinder Pressure: 2214 PSIG
PGVP Number: A12020 Valve Outlet: 590
Gas Code: O2,BALN Certification Date: Nov 11, 2020

Expiration Date: Nov 11, 2028

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 600/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analyte uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a mole/mole basis unless otherwise noted.

Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

ANALYTICAL RESULTS				
Component	Requested Concentration	Actual Concentration	Protocol Method	Total Relative Uncertainty
OXYGEN	8.000 %	8.019 %	G1	+/- 0.3% NIST Traceable
NITROGEN	Balance			

CALIBRATION STANDARDS				
Type	Lot ID	Cylinder No	Concentration	Uncertainty
NTRM	10010602	1D38055	9.987 % OXYGEN/NITROGEN	+/- 0.3% Apr 19, 2022

ANALYTICAL EQUIPMENT		
Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
SIEMENS OXYMAT 6 - N1-W5-951 - O2	PARAMAGNETIC	Oct 26, 2020

Triad Data Available Upon Request

NOTES:
Gross Weight: 48.1 Kg
Net Weight: 8.2 Kg



Approved for Release

Page 1 of 160-401948



ROTA METER CALIBRATION RESULT OCTOBER 2022

Rotameter ID.	Calibration Date	Regression Result	Coefficient (R ²)
BKK_FS0577	01 Oct 22	Y = 1.0202x + 0.1976	1.0000
BKK_FS0579	01 Oct 22	Y = 1.0078x + 0.4789	0.9998
BKK_FS0583	01 Oct 22	Y = 1.016x + 0.3922	1.0000
BKK_FS0584	01 Oct 22	Y = 1.0036x + 2.2262	0.9997
BKK_FS0585	01 Oct 22	Y = 1.0189x - 5.6476	0.9997
BKK_FS0586	01 Oct 22	Y = 1.0095x - 1.1524	0.9995
BKK_FS0587	01 Oct 22	Y = 1.013x - 3.6619	0.9996
BKK_FS0588	01 Oct 22	Y = 1.0154x + 4.8357	0.9999
BKK_FS0589	01 Oct 22	Y = 0.9918x + 4.8069	0.9999
BKK_FS0590	01 Oct 22	Y = 1.0038x - 0.4857	0.9996
BKK_FS0591	01 Oct 22	Y = 0.9705x - 52.174	0.9986
BKK_FS0592	01 Oct 22	Y = 0.9646x - 37.642	0.9985
BKK_FS0593	01 Oct 22	Y = 0.9767x - 58.445	0.9988
BKK_FS0594	01 Oct 22	Y = 0.9902x - 62.87	0.9999
BKK_FS0595	01 Oct 22	Y = 1.0249x - 98.162	0.9999
BKK_FS0596	01 Oct 22	Y = 0.9843x - 26.806	0.9991
BKK_FS0597	01 Oct 22	Y = 0.9802x - 61.653	0.9978
BKK_FS1004	01 Oct 22	Y = 0.9762x + 11.724	0.9998
BKK_FS1005	01 Oct 22	Y = 1.0081x + 1.5143	1.0000
BKK_FS1006	01 Oct 22	Y = 1.098x - 2.9327	0.9999
BKK_FS1007	01 Oct 22	Y = 0.9917x + 1.6592	1.0000
BKK_FS1008	01 Oct 22	Y = 1.0132x + 0.7207	1.0000
BKK_FS1009	01 Oct 22	Y = 1.0132x + 1.1633	0.9960
BKK_FS1010	01 Oct 22	Y = 1.0033x + 0.5758	0.9999
BKK_FS1011	01 Oct 22	Y = 1.0234x + 0.1759	0.9996
BKK_FS1012	01 Oct 22	Y = 1.0106x - 2.0048	0.9997
BKK_FS1013	01 Oct 22	Y = 0.9677x - 35.851	0.9997
BKK_FS1014	01 Oct 22	Y = 1.0021x + 0.3148	0.9998
BKK_FS1015	01 Oct 22	Y = 0.9994x + 1.786	1.0000
BKK_FS1016	01 Oct 22	Y = 1.0105x - 80.256	0.9998
BKK_FS1017	01 Oct 22	Y = 0.9995x + 0.649	1.0000
BKK_FS1018	01 Oct 22	Y = 1.0011x + 1.1786	1.0000
BKK_FS1019	01 Oct 22	Y = 1.0023x - 68.424	0.9996
BKK_FS1020	01 Oct 22	Y = 1.0547x - 0.666	0.9998
BKK_FS1021	01 Oct 22	Y = 1.018x - 3.3286	0.9998
BKK_FS1022	01 Oct 22	Y = 0.9932x - 57.035	0.9986
BKK_FS1023	01 Oct 22	Y = 1.0094x + 0.0717	0.9999
BKK_FS1024	01 Oct 22	Y = 1.0042x + 0.4086	0.9997
BKK_FS1025	01 Oct 22	Y = 1.0132x - 88.507	0.9996



ROTA METER CALIBRATION RESULT OCTOBER 2022

Rotameter ID.	Calibration Date	Regression Result	Coefficient (R ²)
BKK_FS1026	01 Oct 22	Y = 1.0018x + 1.0776	0.9997
BKK_FS1027	01 Oct 22	Y = 1.0053x + 0.231	0.9995
BKK_FS1028	01 Oct 22	Y = 0.9792x - 60.312	0.9982
BKK_FS1029	01 Oct 22	Y = 0.9935x + 0.8234	1.0000
BKK_FS1030	01 Oct 22	Y = 1.0039x + 0.515	0.9999
BKK_FS1031	01 Oct 22	Y = 1.009x - 79.295	0.9998
BKK_FS1039	01 Oct 22	Y = 0.9967x + 4.5048	0.9999
BKK_FS1040	01 Oct 22	Y = 0.9936x + 32.694	0.9998
BKK_FS1041	01 Oct 22	Y = 1.067x - 1.999	1.0000
BKK_FS1042	01 Oct 22	Y = 1.0019x + 2.1571	1.0000
BKK_FS1043	01 Oct 22	Y = 1.1569x - 96.479	0.8412
BKK_FS1044	01 Oct 22	Y = 1.0318x - 0.9374	0.9999
BKK_FS1161	01 Oct 22	Y = 1.0126x + 0.7738	0.9999
BKK_FS1162	01 Oct 22	Y = 0.9994x + 2.6357	0.9995
BKK_FS1163	01 Oct 22	Y = 0.977x - 55.03	0.9987
BKK_FS1164	01 Oct 22	Y = 0.9914x + 0.8427	0.9997
BKK_FS1165	01 Oct 22	Y = 0.9893x + 6.5919	0.9998
BKK_FS1166	01 Oct 22	Y = 1.0031x - 77.881	0.9996
BKK_FS1200	01 Oct 22	Y = 1.0313x - 0.4602	0.9995
BKK_FS1201	01 Oct 22	Y = 1.0045x + 0.15	0.9996
BKK_FS1202	01 Oct 22	Y = 0.9702x - 44.156	0.9994
RYG_FS0197	01 Oct 22	Y = 1.0039x - 0.179	0.9999
RYG_FS0198	01 Oct 22	Y = 0.9964x + 21.757	1.0000
RYG_FS0199	01 Oct 22	Y = 1.0577x - 1.7486	1.0000

Review By: Wichan Choonharat
(Mr. Wichan Choonharat)
Enviro Field Services Manager

Approved By: Mr. Sarayuth Jitranont
(Mr. Sarayuth Jitranont)
Assistant General Manager

SITHIPORN ASSOCIATES CO.,LTD. CALIBRATION LABORATORY

451-451/1 Sirinthorn Rd, Bangbunmu, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.com



Cert. No. : ACC22013
Pages : 1 of 3

Calibration Certificate


Equipment : SOUND CALIBRATOR
Manufacturer : RION
Model : NC-74
Serial No.: 34178121
ID No.: RYG_FS0213

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 22 APRIL 2022
Calibration Date : 26 APRIL 2022
Date of Issue : 29 APRIL 2022



Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchur
(Thanakul Petchur)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

SITHIPORN ASSOCIATES CO.,LTD. CALIBRATION LABORATORY

Continuation of Calibration Certificate

Cert. No. : ACC22013
Job No. : VC65AC0054
Pages : 2 of 3

Calibration Procedure : CP-AC-03

Calibration Method :

This equipment was calibrated by based on IEC-60942-2003 Standard.

The sound pressure level, frequency and total distortion of the sound calibrator was measured using the reference microphone.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33511B	MY52302742	EF-0008-22	04-Feb-23
Digital Multimeter	33461A	MY53220104	EEL.BP. 04/0265	09-Feb-23
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0265	09-Feb-23
Digital Multimeter	33461A	MY60024273	EEL.BP. 05/0265	09-Feb-23
Programmable Attenuator	MAT-1070	62100114	EF-0009-22	07-Feb-23
Condenser Microphone	4180	2977900	AA-1013-22	24-Feb-23
Measuring Amplifier	NA-42KAI	34560495	AA-3005-22	22-Feb-23
Audio Analyzer	AVR-3360A	V744B6069	EF-0010-22	07-Feb-23

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

SITHIPORN ASSOCIATES CO.,LTD. CALIBRATION LABORATORY

Continuation of Calibration Certificate

Cert. No. : ACC22013
Job No. : VC65AC0054
Pages : 3 of 3

SITHIPORN ASSOCIATES CO.,LTD. CALIBRATION LABORATORY

451-451/1 Sirinthorn Rd, Bangbunmu, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.com



Cert. No. : ACL22059
Pages : 1 of 8

Result of calibration :

1. Sound pressure level

Specified sound pressure level (dB)	Measured value (dB)	Deviated value (dB)	Uncertainty (dB)	Tolerance limit (dB)
94	94.11	0.11	0.14	0.40

2. Frequency

Specified Frequency (Hz)	Measured value (Hz)	Deviated value (%)	Uncertainty (%)	Tolerance limit (%)
1000	1003.1	0.3	0.1	1.0

3. Total distortion

Measured value (%)	Uncertainty (%)	Tolerance limit (%)
2.02	0.10	3.0

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate


Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00734220 / 145272 / 34370
ID No.: RYG_FS0026

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 14 JANUARY 2022
Calibration Date : 21-24 JANUARY 2022
Date of Issue : 25 JANUARY 2022



Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchur
(Thanakul Petchur)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22059
Job No. : VC65AC0043
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2015) Standard for sound level meter (SLM).

The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	34461A	MY60024273	1-15180725251-1	15-Sep-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KA1	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

QT-TS12-04-04-020664

T. P. K.

Continuation of Calibration Certificate

Cert. No. : ACL22059
Job No. : VC65AC0043
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.6

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	11.6
C - weight	17.8
Flat	23.7

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.3	0.3	0.3	± 1.5
1000	0.0	0.0	0.0	± 1.0
8000	-1.6	-1.5	-1.5	±5.0

QT-TS12-04-04-020664

T. P. K.

Continuation of Calibration Certificate

Cert. No. : ACL22059
Job No. : VC65AC0043
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QT-TS12-04-04-020664

T. P. K.

Continuation of Calibration Certificate

Cert. No. : ACL22059
Job No. : VC65AC0043
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	0.0	0.0	0.0	±2.0
125	0.0	0.0	0.0	±1.5
250	0.0	0.0	0.0	±1.5
500	0.0	0.1	0.0	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.1	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QT-TS12-04-04-020664

T. P. K.

Continuation of Calibration Certificate

Cert. No. : ACL22059
Job No. : VC65AC0043
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	±1.1
136.0	136.0	0.0	±1.1
135.0	135.0	0.0	±1.1
134.0	134.0	0.0	±1.1
133.0	133.0	0.0	±1.1
132.0	132.0	0.0	±1.1
131.0	131.0	0.0	±1.1
129.0	129.0	0.0	±1.1
124.0	124.0	0.0	±1.1
119.0	119.0	0.0	±1.1
114.0	114.0	0.0	±1.1
109.0	109.0	0.0	±1.1
104.0	104.0	0.0	±1.1
99.0	99.0	0.0	±1.1
94.0	94.0	0.0	±1.1
89.0	89.0	0.0	±1.1
84.0	84.0	0.0	±1.1
79.0	79.0	0.0	±1.1
74.0	74.0	0.0	±1.1
69.0	69.0	0.0	±1.1
64.0	64.0	0.0	±1.1
59.0	59.0	0.0	±1.1
54.0	54.0	0.0	±1.1
49.0	49.0	0.0	±1.1
44.0	44.0	0.0	±1.1
39.0	38.9	-0.1	±1.1
34.0	33.9	-0.1	±1.1
30.0	30.0	0.0	±1.1
29.0	29.0	0.0	±1.1
28.0	28.0	0.0	±1.1
27.0	27.0	0.0	±1.1
26.0	26.1	0.1	±1.1
25.0	25.0	0.0	±1.1

QF-1S12-04-04-020664

T. Petchurai

Continuation of Calibration Certificate

Cert. No. : ACL22059
Job No. : VC65AC0043
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.6	89.7	0.1	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A-weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation providing a level of confidence of approximately 95 %

End of Calibration Certificate

QF-1S12-04-04-020664

T. Petchurai

Continuation of Calibration Certificate

Cert. No. : ACL22059
Job No. : VC65AC0043
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.1	0.1	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, Lepeak (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.1	-0.3	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.2	-0.2	±2.0
Negative half cycle	135.4	135.2	-0.2	±2.0

QF-1S12-04-04-020664

T. Petchurai

451-451/1 Sirinthorn Rd., Bangbunru, Bangplud Bangkok 10700 THAILAND
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:center@sithiporn.com http://www.sithiporn.comCert. No. : ACL21099
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 01073608 / 172153 / 85748
ID No.: RYG_FS0387

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location :
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 01 SEPTEMBER 2021
Calibration Date : 13-15 SEPTEMBER 2021
Date of Issue : 16 SEPTEMBER 2021

Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-1S12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL21099
Job No. : VC64AC0066
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	8846A	1997025	EEL.BP. 06/0264	05-Feb-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KA1	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

QF-TS12-04-04-020664

T. R. S. L.

Continuation of Calibration Certificate

Cert. No. : ACL21099
Job No. : VC64AC0066
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.1	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

T. R. S. L.

Continuation of Calibration Certificate

Cert. No. : ACL21099
Job No. : VC64AC0066
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
15.1

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	11.6
C - weight	17.7
Flat	23.3

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Flat	C-weight	A-weight	Acceptance Limits
125	0.7	0.7	0.7	± 1.5
1000	0.0	0.0	0.0	± 1.0
8000	-2.9	-2.8	-2.8	±5.0

QF-TS12-04-04-020664

T. R. S. L.

Continuation of Calibration Certificate

Cert. No. : ACL21099
Job No. : VC64AC0066
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	-0.1	-0.1	±2.0
125	0.0	0.0	-0.1	±1.5
250	0.0	-0.1	-0.1	±1.5
500	0.0	0.0	-0.1	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.0	0.0	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.1	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.1	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.1	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.1	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

T. R. S. L.

Continuation of Calibration Certificate

Cert. No. : ACL21099
Job No. : VC64AC0066
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	±1.1
136.0	136.0	0.0	±1.1
135.0	135.0	0.0	±1.1
134.0	134.0	0.0	±1.1
133.0	132.9	-0.1	±1.1
132.0	131.9	-0.1	±1.1
131.0	130.9	-0.1	±1.1
129.0	128.9	-0.1	±1.1
124.0	123.9	-0.1	±1.1
119.0	119.0	0.0	±1.1
114.0	114.0	0.0	±1.1
109.0	109.0	0.0	±1.1
104.0	104.0	0.0	±1.1
99.0	99.0	0.0	±1.1
94.0	94.0	0.0	±1.1
89.0	89.0	0.0	±1.1
84.0	84.0	0.0	±1.1
79.0	79.0	0.0	±1.1
74.0	74.0	0.0	±1.1
69.0	69.0	0.0	±1.1
64.0	64.0	0.0	±1.1
59.0	59.0	0.0	±1.1
54.0	54.0	0.0	±1.1
49.0	49.0	0.0	±1.1
44.0	44.0	0.0	±1.1
39.0	39.0	0.0	±1.1
34.0	34.0	0.0	±1.1
30.0	29.9	-0.1	±1.1
29.0	28.9	-0.1	±1.1
28.0	27.9	-0.1	±1.1
27.0	26.9	-0.1	±1.1
26.0	25.9	-0.1	±1.1
25.0	24.9	-0.1	±1.1

Continuation of Calibration Certificate

Cert. No. : ACL21099
Job No. : VC64AC0066
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 : -5.0
	2	8	117.0	116.9	-0.1	1.0 : -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 : -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 : -5.0
	2	8	108.0	108.0	0.0	1.0 : -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.3	-0.1	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL21099
Job No. : VC64AC0066
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.5	89.5	0.0	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A-weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

451-451/1 Srinthom Rd, Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.com



Cert. No. : ACC22024
Pages : 1 of 3

Calibration Certificate

Equipment : SOUND CALIBRATOR
Manufacturer : RION
Model : NC-74
Serial No. : 34178124
ID No. : RYG_FS0216

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location :
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %
Received Date : 22 AUGUST 2022
Calibration Date : 31 AUGUST 2022
Date of Issue : 02 SEPTEMBER 2022

Calibrated by : Nithakorn Pisutpaisan

Approved by : *T. Petchurai*
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACC22024
Job No. : VC65AC0077
Pages : 2 of 3

Calibration Procedure : CP-AC-03

Calibration Method :

This equipment was calibrated by based on IEC-60942-2003 Standard.

The sound pressure level, frequency and total distortion of the sound calibrator was measured using the reference microphone.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33511B	MY52302742	EF-0008-22	04-Feb-23
Digital Multimeter	33461A	MY53220104	EEL.BP. 04/0265	09-Feb-23
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0265	09-Feb-23
Digital Multimeter	33461A	MY60024273	EEL.BP. 05/0265	09-Feb-23
Programmable Attenuator	MAT-1070	62100114	EF-0009-22	07-Feb-23
Condenser Microphone	4180	2977900	AA-1013-22	24-Feb-23
Measuring Amplifier	NA-42KAI	34560495	AA-3005-22	22-Feb-23
Audio Analyzer	AVR-3360A	V744B6069	EF-0010-22	07-Feb-23

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

QF-TS12-04-04-020664

SITHIPORN ASSOCIATES CO.,LTD.
CALIBRATION LABORATORY451-451/1 Sirinthorn Rd.,Bangbunru, Bangkok 10700 THAILAND
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:center@sithiporn.com http://www.sithiporn.comREC-TIS-TIS 17025
CALIBRATION 0394Cert. No. : ACL22025
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00734221 / 145286 / 34371
ID No.: RYG. FS0027

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 05 JANUARY 2022
Calibration Date : 10-12 JANUARY 2022
Date of Issue : 13 JANUARY 2022

Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACC22024
Job No. : VC65AC0077
Pages : 3 of 3

Result of calibration :

1. Sound pressure level

Specified sound pressure level (dB)	Measured value (dB)	Deviated value (dB)	Uncertainty (dB)	Tolerance limit (dB)
94	94.21	0.21	0.14	0.40

2. Frequency

Specified Frequency (Hz)	Measured value (Hz)	Deviated value (%)	Uncertainty (%)	Tolerance limit (%)
1000	1001.3	0.1	0.1	1.0

3. Total distortion

Measured value (%)	Uncertainty (%)	Tolerance limit (%)
1.95	0.10	3.0

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22025
Job No. : VC65AC0040
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).

The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	34461A	MY60024273	1-15180723251-1	15-Sep-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KAI	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22025
Job No. : VC65AC0040
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

Continuation of Calibration Certificate

Cert. No. : ACL22025
Job No. : VC65AC0040
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
16.2

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	11.6
C - weight	18.0
Flat	23.9

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			Acceptance Limits
	Flat	C-weight	A-weight	
125	0.5	0.5	0.6	± 1.5
1000	0.0	0.0	0.1	± 1.0
8000	-2.3	-2.3	-2.3	±5.0

QF-TS12-04-04-020664

T P.L.

QF-TS12-04-04-020664

T P.L.

Continuation of Calibration Certificate

Cert. No. : ACL22025
Job No. : VC65AC0040
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	0.0	0.0	0.0	±2.0
125	0.0	0.0	0.0	±1.5
250	0.0	0.0	0.0	±1.5
500	0.0	0.1	0.0	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.1	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.1	0.1	± 0.3

QF-TS12-04-04-020664

T P.L.

Continuation of Calibration Certificate

Cert. No. : ACL22025
Job No. : VC65AC0040
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	38.9	-0.1	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.0	0.0	± 1.1
26.0	26.0	0.0	± 1.1
25.0	25.0	0.0	± 1.1

QF-TS12-04-04-020664

T P.L.

Continuation of Calibration Certificate.

Cert. No. : ACL22025
Job No. : VC65AC0040
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	108.0	0.0	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.1	0.1	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.1	0.1	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, Lepeak (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.3	-0.1	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

QF-TS12-04-04-020664

SITHIPORN ASSOCIATES CO.,LTD.
CALIBRATION LABORATORY451-451/1 Sirinthorn Rd, Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.comCert. No. : ACL22154
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00734218 / 146937 / 34368
ID No.: RYG_FS0031

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location :
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 17 JUNE 2022
Calibration Date : 20-22 JUNE 2022
Date of Issue : 27 JUNE 2022

Calibrated by : Nahaikom Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22025
Job No. : VC65AC0040
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle	0.1	±1.5
89.5	89.6		

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22154
Job No. : VC65AC0068
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0007-22	04-Feb-23
Waveform Generator	33511B	MY52302742	EF-0008-22	04-Feb-23
Digital Multimeter	33461A	MY53220104	EEL.BP. 04/0265	09-Feb-23
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0265	09-Feb-23
Digital Multimeter	34461A	MY60024273	EEL.BP. 05/0265	09-Feb-23
Programmable Attenuator	MAT-1070	62100114	EF-0009-22	07-Feb-23
Condenser Microphone	4180	2977900	AA-1013-22	24-Feb-23
Measuring Amplifier	NA-42KA1	34560495	AA-3005-22	22-Feb-23

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
- 3.2 Thailand Institute of Scientific and Technological Research (TISTR).

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22154
Job No. : VC65AC0068
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

T. Petch.

Continuation of Calibration Certificate

Cert. No. : ACL22154
Job No. : VC65AC0068
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	-0.1	0.0	±2.0
125	0.0	0.0	0.0	±1.5
250	0.0	0.0	-0.1	±1.5
500	0.0	0.0	-0.1	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

T. Petch.

Continuation of Calibration Certificate

Cert. No. : ACL22154
Job No. : VC65AC0068
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.95)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
20.1

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	11.6
C - weight	17.4
Flat	23.1

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.5	0.5	0.5	± 1.5
1000	0.1	0.1	0.1	± 1.0
8000	-1.5	-1.5	-1.4	±5.0

QF-TS12-04-04-020664

T. Petch.

Continuation of Calibration Certificate

Cert. No. : ACL22154
Job No. : VC65AC0068
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	132.9	-0.1	± 1.1
132.0	131.9	-0.1	± 1.1
131.0	130.9	-0.1	± 1.1
129.0	128.9	-0.1	± 1.1
124.0	123.9	-0.1	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	29.9	-0.1	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.0	0.0	± 1.1
26.0	25.9	-0.1	± 1.1
25.0	24.9	-0.1	± 1.1

QF-TS12-04-04-020664

T. Petch.

Continuation of Calibration Certificate

Cert. No. : ACL22154
Job No. : VC65AC0068
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
SEL	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	135.7	-0.7	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

Continuation of Calibration Certificate

Cert. No. : ACL22154
Job No. : VC65AC0068
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle	(dB)	(dB)
89.6	89.6	0.0	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

QF-TS12-04-04-020664

T. Petchuraj

SITHIPORN ASSOCIATES CO.,LTD.
CALIBRATION LABORATORY

451-451/1 Sirinthorn Rd., Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.com



Cert. No. : ACC22001
Pages : 1 of 3

Calibration Certificate

Equipment : SOUND CALIBRATOR
Manufacturer : RION
Model : NC-75
Serial No.: 35002736
ID No.: - 6 5 0 7 3 6

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location :
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 05 JANUARY 2022
Calibration Date : 10 JANUARY 2022
Date of Issue : 13 JANUARY 2022

Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchuraj
(Thanakul Petchuraj)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

QF-TS12-04-04-020664

T. Petchuraj

Continuation of Calibration Certificate

Cert. No. : ACC22001
Job No. : VC65AC0040
Pages : 2 of 3

Calibration Procedure : CP-AC-03

Calibration Method :

This equipment was calibrated by based on IEC-60942-2003 Standard.

The sound pressure level, frequency and total distortion of the sound calibrator was measured using the reference microphone.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	33461A	MY60024273	1-15180725251-1	15-Sep-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KA1	34560495	AA-3003-21	16-Feb-22
Audio Analyzer	AVR-3360A	V744B6069	EF-0010-21	10-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
- 3.2 Thailand Institute of Scientific and Technological Research (TISTR).

QF-TS12-04-04-020664

T. Petchuraj

Continuation of Calibration Certificate

Cert. No. : ACC22001
Job No. : VC65AC0040
Pages : 3 of 3

451-451/1 Sirinthern Rd, Bangbunru, Bangplud Bangkok 10700 THAILAND
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.com



Cert. No. : ACL22062
Pages : 1 of 8

Result of calibration :

1. Sound pressure level

Specified sound pressure level (dB)	Measured value (dB)	Deviated value (dB)	Uncertainty (dB)	Tolerance limit (dB)
94	93.99	-0.01	0.14	0.40

2. Frequency

Specified Frequency (Hz)	Measured value (Hz)	Deviated value (%)	Uncertainty (%)	Tolerance limit (%)
1000	1000.0	0.0	0.1	1.0

3. Total distortion

Measured value (%)	Uncertainty (%)	Tolerance limit (%)
0.28	0.10	3.0

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22062
Job No. : VC65AC0043
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).

The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05-0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03-0264	08-Feb-22
Digital Multimeter	34461A	MY60024273	1-15180725251-1	15-Sep-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KA1	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).


Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42 / Microphone UC-52 / Preamplifier NH-24
Serial No. : 01222723 / 143841 / 22770
ID No. : RYG FS0022

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHUWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location :
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %
Received Date : 14 JANUARY 2022
Calibration Date : 21-24 JANUARY 2022
Date of Issue : 25 JANUARY 2022



Calibrated by : Nathakorn Pisutpaisan

Approved by : T. Petchuraj
(Thanakul Petchuraj)

This certificate is issued in accordance with the requirements of ISO IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22062
Job No. : VC65AC0043
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long-term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22062
Job No. : VC65AC0043
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.8

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	12.0
C - weight	18.0
Flat	24.1

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.4	0.4	0.4	± 1.5
1000	-0.1	-0.1	-0.1	± 1.0
8000	-0.4	-0.3	-0.3	±5.0

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22062
Job No. : VC65AC0043
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	0.0	-0.1	±2.0
125	0.0	0.0	0.0	±1.5
250	0.0	0.0	0.0	±1.5
500	0.0	0.0	0.0	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22062
Job No. : VC65AC0043
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.1	0.1	± 1.1
26.0	26.1	0.1	± 1.1
25.0	25.1	0.1	± 1.1

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22062
Job No. : VC65AC0043
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, Lpeak (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	135.8	-0.6	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.2	-0.2	±2.0
Negative half cycle	135.4	135.2	-0.2	±2.0

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22062
Job No. : VC65AC0043
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.6	89.6	0.0	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

451-451/1 Sirinthorn Rd, Bangbunru, Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.comCert. No. : ACL22115
Pages : 1 of 8

Calibration Certificate


Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00734223 / 157777 / 22653
ID No.: RYG_FS0029

Condition As Found : GOOD

Customer : A.S. LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 17 MAY 2022
Calibration Date : 24-27 MAY 2022
Date of Issue : 30 MAY 2022



Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22115
Job No. : VC65AC0060
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0007-22	04-Feb-23
Waveform Generator	33511B	MY52302742	EF-0008-22	04-Feb-23
Digital Multimeter	33461A	MY53220104	EEL_BP_04/0265	09-Feb-23
Digital Multimeter	33461A	MY53220076	EEL_BP_03/0265	09-Feb-23
Digital Multimeter	34461A	MY60024273	EEL_BP_05/0265	09-Feb-23
Programmable Attenuator	MAT-1070	62100114	EF-0009-22	07-Feb-23
Condenser Microphone	4180	2977900	AA-1013-22	24-Feb-23
Measuring Amplifier	NA-42KAI	34560495	AA-3005-22	22-Feb-23

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL22115
Job No. : VC65AC0060
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22115
Job No. : VC65AC0060
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.95)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
18.0

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	13.8
C - weight	20.3
Flat	25.8

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.7	0.7	0.7	± 1.5
1000	0.1	0.1	0.1	± 1.0
8000	-1.5	-1.5	-1.5	±5.0

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22115
Job No. : VC65AC0060
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	-0.2	-0.1	±2.0
125	-0.1	0.0	0.0	±1.5
250	0.0	0.0	-0.1	±1.5
500	0.0	0.0	-0.1	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.0	0.0	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Lcq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22115
Job No. : VC65AC0060
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	132.9	-0.1	± 1.1
132.0	131.9	-0.1	± 1.1
131.0	130.9	-0.1	± 1.1
129.0	128.9	-0.1	± 1.1
124.0	123.9	-0.1	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	29.9	-0.1	± 1.1
29.0	28.9	-0.1	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.0	0.0	± 1.1
26.0	25.9	-0.1	± 1.1
25.0	24.9	-0.1	± 1.1

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22115
Job No. : VC65AC0060
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	116.9	-0.1	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.8	-0.2	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	135.8	-0.6	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22115
Job No. : VC65AC0060
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.6	89.6	0.0	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$
or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

451-451/1 Srinthorn Rd, Bangbunru, Bangplud Bangkok 10700 THAILAND
Tel: 0-2435-8800 Fax: 0-2433-1679 e-mail: cal-center@sithiporn.com http://www.sithiporn.com



Cert. No. : ACL22060
Pages : 1 of 8


Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No. : 00734225 / 169439 / 72460
ID No. : RYG FS0030

Condition As Found : GOOD

Customer : A/S LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %
Received Date : 14 JANUARY 2022
Calibration Date : 21-24 JANUARY 2022
Date of Issue : 25 JANUARY 2022



Calibrated by : Nathakorn Pisutpaisan

Approved by : *T. Petchur*
(Thanakul Petchur)

This certificate is issued in accordance with the requirements of ISO IEC 17025 standard, may not be reproduced
other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22060
Job No. : VC65AC0043
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference
Standard Instruments.
For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	1-F-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	1-F-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL_BP_03.0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL_BP_03.0264	08-Feb-22
Digital Multimeter	34461A	MY60024273	1-15180725251-1	15-Sep-22
Programmable Attenuator	MAT-1070	62109114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KA1	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL22060
Job No. : VC65AC0043
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N A
2. Self-generated noise	✓	-	0.2	N A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.4	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long-term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22060
Job No. : VC65AC0043
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.2

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	10.8
C - weight	17.2
Flat	23.0

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.4	0.5	0.5	± 1.5
1000	-0.1	-0.1	-0.1	± 1.0
8000	-1.6	-1.5	-1.5	±5.0

QF-TS12-04-04-020664

T. R. P. A.

Continuation of Calibration Certificate

Cert. No. : ACL22060
Job No. : VC65AC0043
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	0.0	-0.1	0.0	±2.0
125	0.0	0.0	0.0	±1.5
250	0.0	0.0	-0.1	±1.5
500	0.0	0.0	-0.1	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

T. R. P. A.

Continuation of Calibration Certificate

Cert. No. : ACL22060
Job No. : VC65AC0043
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	132.9	-0.1	± 1.1
132.0	131.9	-0.1	± 1.1
131.0	130.9	-0.1	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.0	0.0	± 1.1
26.0	26.1	0.1	± 1.1
25.0	25.1	0.1	± 1.1

QF-TS12-04-04-020664

T. R. P. A.

Continuation of Calibration Certificate

Cert. No. : ACL22060
Job No. : VC65AC0043
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
SEL	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	135.8	-0.6	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

QF-TS12-04-04-020664

T. R. P. A.

Continuation of Calibration Certificate

Cert. No. : ACL22060
Job No. : VC65AC0043
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.6	89.5	-0.1	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22159
Job No. : VC65AC0069
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM). The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0007-22	04-Feb-23
Waveform Generator	33511B	MY52302742	EF-0608-22	04-Feb-23
Digital Multimeter	33461A	MY53220104	EEL.BP. 04/0265	09-Feb-23
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0265	09-Feb-23
Digital Multimeter	34461A	MY60024273	EEL.BP. 05/0265	09-Feb-23
Programmable Attenuator	MAT-1070	62100114	EF-0009-22	07-Feb-23
Condenser Microphone	4180	2977900	AA-1013-22	24-Feb-23
Measuring Amplifier	NA-42KA1	34560495	AA-3005-22	22-Feb-23

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

451-451/1 Sirinthon Rd, Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:center@sithiporn.com http://www.sithiporn.comCert. No. : ACL22159
Pages : 1 of 8


Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00472130 / 157774 / 72464
ID No.: RYG_FS0303

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %
Received Date : 06 JULY 2022
Calibration Date : 11-18 JULY 2022
Date of Issue : 19 JULY 2022



Calibrated by : Nathakorn Pisutpaisan

Approved by :

(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22159
Job No. : VC65AC0069
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22159
Job No. : VC65AC0069
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.95)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
23.4

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	15.4
C - weight	21.0
Flat	26.9

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.0	0.0	0.0	± 1.5
1000	-0.1	-0.1	-0.1	± 1.0
8000	-0.3	-0.2	-0.2	±5.0

QF-TS12-04-04-020664

P.T.A.

Continuation of Calibration Certificate

Cert. No. : ACL22159
Job No. : VC65AC0069
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	-0.2	-0.1	±2.0
125	-0.1	0.0	-0.1	±1.5
250	0.0	0.0	-0.1	±1.5
500	0.0	0.0	-0.1	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.0	0.0	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	93.9	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	93.9	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	93.9	94.0	0.1	± 0.3

QF-TS12-04-04-020664

P.T.A.

Continuation of Calibration Certificate

Cert. No. : ACL22159
Job No. : VC65AC0069
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.1	0.1	± 1.1
136.0	136.1	0.1	± 1.1
135.0	135.1	0.1	± 1.1
134.0	134.1	0.1	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.1	0.1	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.1	0.1	± 1.1
114.0	114.1	0.1	± 1.1
109.0	109.1	0.1	± 1.1
104.0	104.1	0.1	± 1.1
99.0	99.1	0.1	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	26.9	-0.1	± 1.1
26.0	26.0	0.0	± 1.1
25.0	25.0	0.0	± 1.1

QF-TS12-04-04-020664

P.T.A.

Continuation of Calibration Certificate

Cert. No. : ACL22159
Job No. : VC65AC0069
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	116.9	-0.1	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.8	-0.2	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	135.6	-0.8	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

QF-TS12-04-04-020664

P.T.A.

Continuation of Calibration Certificate

Cert. No. : ACL22159
Job No. : VC65AC0069
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.6	89.6	0.0	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

451-451/1 Sirinithorn Rd, Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:center@sithiporn.com http://www.sithiporn.com



Cert. No. : ACL21117
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No. : 00233183 / 144835 / 23230
ID No. : RYG_FS0024

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 21 SEPTEMBER 2021
Calibration Date : 04-06 OCTOBER 2021
Date of Issue : 11 OCTOBER 2021

REVIEW BY	<i>Nathakorn P.</i>
APPROVED BY	<i>Th. Petchur</i>
NEXT CAL DATE	4/10/22

Calibrated by : Nathakorn Pisutpaisan

Approved by :

Th. Petchur
(Thanakul Petchur)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL21117
Job No. : VC64AC0070
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	8846A	1997025	EEL.BP. 06/0264	05-Feb-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KAI	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
- 3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL21117
Job No. : VC64AC0070
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.4	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL21117
Job No. : VC64AC0070
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
22.9

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting (Hz)	Measured value (dB)
A - weight	13.8
C - weight	19.7
Flat	25.4

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	-0.1	-0.1	-0.1	± 1.5
1000	0.0	-0.1	0.0	± 1.0
8000	0.3	0.4	0.4	±5.0

QF-TS12-04-04-020664

T. P. P. P.

Continuation of Calibration Certificate

Cert. No. : ACL21117
Job No. : VC64AC0070
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	0.0	0.0	0.0	±2.0
125	0.0	0.0	0.0	±1.5
250	0.0	0.0	0.0	±1.5
500	0.0	0.1	0.0	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.1	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

T. P. P. P.

Continuation of Calibration Certificate

Cert. No. : ACL21117
Job No. : VC64AC0070
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.1	0.1	± 1.1
28.0	28.1	0.1	± 1.1
27.0	27.1	0.1	± 1.1
26.0	26.1	0.1	± 1.1
25.0	25.2	0.2	± 1.1

QF-TS12-04-04-020664

T. P. P. P.

Continuation of Calibration Certificate

Cert. No. : ACL21117
Job No. : VC64AC0070
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.1	0.1	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
SEL	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.1	0.1	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, Lepenk (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.2	-0.2	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.1	0.1	-
Positive half cycle	135.4	135.2	-0.2	±2.0
Negative half cycle	135.4	135.2	-0.2	±2.0

QF-TS12-04-04-020664

T. P. P. P.

Continuation of Calibration Certificate

Cert. No. : ACL21117
Job No. : VC64AC0070
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.6	89.5	-0.1	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$
or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

451-451/1 Sirinthorn Rd., Bangumru, Bangplud Bangkok 10700 THAILAND.
Tel.0-2435-8800 Fax.0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.comCert. No. : ACL22054
Pages : 1 of 8


Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No. : 00233184 / 144837 / 23232
ID No. : RYG_FS0025

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %
Received Date : 14 JANUARY 2022
Calibration Date : 21-24 JANUARY 2022
Date of Issue : 25 JANUARY 2022



Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced
other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22054
Job No. : VC65AC0043
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference
Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	34461A	MY60024273	I-15180725251-1	15-Sep-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KAI	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL22054
Job No. : VC65AC0043
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.4	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long-term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22054
Job No. : VC65AC0043
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.2

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	10.8
C - weight	17.0
Flat	22.8

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			Acceptance Limits
	Flat	C-weight	A-weight	
125	0.2	0.2	0.2	± 1.5
1000	0.0	0.0	0.0	± 1.0
8000	-0.9	-0.8	-0.8	±5.0

QF-TS12-04-04-020664

T. P. A.

Continuation of Calibration Certificate

Cert. No. : ACL22054
Job No. : VC65AC0043
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	0.0	-0.1	0.0	±2.0
125	0.0	0.0	0.0	±1.5
250	0.0	0.0	0.0	±1.5
500	0.0	0.0	0.0	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
1eq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-1S12-04-04-020664

T. P. A.

Continuation of Calibration Certificate

Cert. No. : ACL22054
Job No. : VC65AC0043
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	131.9	-0.1	± 1.1
131.0	130.9	-0.1	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.1	0.1	± 1.1
26.0	26.1	0.1	± 1.1
25.0	25.1	0.1	± 1.1

QF-TS12-04-04-020664

T. P. A.

Continuation of Calibration Certificate

Cert. No. : ACL22054
Job No. : VC65AC0043
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.4	0.0	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.2	-0.2	±2.0

QF-TS12-04-04-020664

T. P. A.

Continuation of Calibration Certificate

Cert. No. : ACL22054
Job No. : VC65AC0043
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.6	89.5	-0.1	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

451-451/1 Sirinthorn Rd., Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiphorn.com http://www.sithiphorn.com



Cert. No. : ACL21098
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00472126 / 180410 / 88180
ID No.: RYG_FS0301

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 01 SEPTEMBER 2021
Calibration Date : 13-15 SEPTEMBER 2021
Date of Issue : 16 SEPTEMBER 2021

REVIEW BY	<i>[Signature]</i>
APPROVED BY	<i>[Signature]</i>
NEXT CAL. DATE	15/11/22

Calibrated by : Nathakorn Pisutpaisan

Approved by :

[Signature]
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL21098
Job No. : VC64AC0066
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.
For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	8846A	1997025	EEL.BP. 06/0264	05-Feb-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KA1	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL21098
Job No. : VC64AC0066
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL21098
Job No. : VC64AC0066
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
26.1

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	12.0
C - weight	18.4
Flat	23.9

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	-1.6	-1.6	-1.6	± 1.5
1000	0.0	0.0	0.0	± 1.0
8000	0.4	0.5	0.5	±5.0

Continuation of Calibration Certificate

Cert. No. : ACL21098
Job No. : VC64AC0066
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	-0.2	-0.1	±2.0
125	0.0	0.0	0.0	±1.5
250	0.0	0.0	-0.1	±1.5
500	0.0	0.0	-0.1	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL21098
Job No. : VC64AC0066
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.1	0.1	± 1.1
84.0	84.1	0.1	± 1.1
79.0	79.1	0.1	± 1.1
74.0	74.1	0.1	± 1.1
69.0	69.1	0.1	± 1.1
64.0	64.1	0.1	± 1.1
59.0	59.1	0.1	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.1	0.1	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.1	0.1	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.1	0.1	± 1.1
29.0	29.1	0.1	± 1.1
28.0	28.1	0.1	± 1.1
27.0	27.2	0.2	± 1.1
26.0	26.2	0.2	± 1.1
25.0	25.2	0.2	± 1.1

Continuation of Calibration Certificate

Cert. No. : ACL21098
Job No. : VC64AC0066
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, Lepeak (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	135.8	-0.6	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL21098
Job No. : VC64AC0066
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.5	89.5	0.0	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$
or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

451-451/1 Sirinthorn Rd, Bangbunru, Bangkok 10700 THAILAND.
Tel: 0-2435-8800 Fax: 0-2433-1679 e-mail: cal-center@sithiporn.com http://www.sithiporn.comCert. No. : ACL22158
Pages : 1 of 8


Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RJON
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00472127 / 169440 / 72461
ID No.: RYG_FS0302

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %
Received Date : 06 JULY 2022
Calibration Date : 11-18 JULY 2022
Date of Issue : 19 JULY 2022



Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced
other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22158
Job No. : VC65AC0069
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference
Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0007-22	04-Feb-23
Waveform Generator	33511B	MY52302742	EF-0008-22	04-Feb-23
Digital Multimeter	33461A	MY53220104	EEL.BP. 04/0265	09-Feb-23
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0265	09-Feb-23
Digital Multimeter	33461A	MY60024273	EEL.BP. 05/0265	09-Feb-23
Programmable Attenuator	MAT-1070	62100114	EF-0009-22	07-Feb-23
Condenser Microphone	4180	2977900	AA-1013-22	24-Feb-23
Measuring Amplifier	NA-42KA1	34560495	AA-3005-22	22-Feb-23

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL22158
Job No. : VC65AC0069
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22158
Job No. : VC65AC0069
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.95)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
17.3

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	12.0
C - weight	18.1
Flat	23.9

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.0	0.0	0.1	± 1.5
1000	-0.1	-0.1	-0.1	± 1.0
8000	0.4	0.4	0.5	±5.0

QF-TS12-04-04-020664

T. B. A.

Continuation of Calibration Certificate

Cert. No. : ACL22158
Job No. : VC65AC0069
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	53.9	-0.1	± 1.1
49.0	48.9	-0.1	± 1.1
44.0	43.9	-0.1	± 1.1
39.0	38.9	-0.1	± 1.1
34.0	33.9	-0.1	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.0	0.0	± 1.1
26.0	26.0	0.0	± 1.1
25.0	25.0	0.0	± 1.1

QF-TS12-04-04-020664

T. B. A.

Continuation of Calibration Certificate

Cert. No. : ACL22158
Job No. : VC65AC0069
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	0.0	-0.1	-0.1	±2.0
125	0.0	0.1	0.0	±1.5
250	0.0	0.0	0.0	±1.5
500	0.0	0.0	0.0	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

T. B. A.

Continuation of Calibration Certificate

Cert. No. : ACL22158
Job No. : VC65AC0069
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tnne burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.0	-0.4	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.2	-0.2	±2.0
Negative half cycle	135.4	135.2	-0.2	±2.0

QF-TS12-04-04-020664

T. B. A.

Continuation of Calibration Certificate

Cert. No. : ACL22158
Job No. : VC65AC0069
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.6	89.5	-0.1	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$
or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

451-451/1 Sirinthorn Rd., Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiphorn.com http://www.sithiphorn.comCert. No. : ACL22160
Pages : 1 of 8

Calibration Certificate


Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00472132 / 169445 / 72466
ID No.: RYG_FS0304

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTANAKAN 40, PHATTANAKAN ROAD,
KHWAENG PHATTANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 06 JULY 2022
Calibration Date : 11-18 JULY 2022
Date of Issue : 19 JULY 2022



Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced
other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22160
Job No. : VC65AC0069
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference
Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0007-22	04-Feb-23
Waveform Generator	33511B	MY52302742	EF-0008-22	04-Feb-23
Digital Multimeter	33461A	MY53220104	EEL_BP_04/0265	09-Feb-23
Digital Multimeter	33461A	MY53220076	EEL_BP_03/0265	09-Feb-23
Digital Multimeter	34461A	MY60024273	EEL_BP_05/0265	09-Feb-23
Programmable Attenuator	MAT-1070	62100114	EF-0009-22	07-Feb-23
Condenser Microphone	4180	2977900	AA-1013-22	24-Feb-23
Measuring Amplifier	NA-42KA1	34560495	AA-3005-22	22-Feb-23

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL22160
Job No. : VC65AC0069
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long-term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22160
Job No. : VC65AC0069
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.95)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.2

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting (dB)	Measured value (dB)
A - weight	9.9
C - weight	16.3
Flat	22.1

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			Acceptance Limits
	Flat	C-weight	A-weight	
125	0.2	0.2	0.2	± 1.5
1000	-0.1	-0.1	-0.1	± 1.0
8000	-1.1	-1.1	-1.1	±5.0

QF-TS12-04-04-020664

T P.T.A.

Continuation of Calibration Certificate

Cert. No. : ACL22160
Job No. : VC65AC0069
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	0.0	0.0	0.0	±2.0
125	0.1	0.1	0.1	±1.5
250	0.0	0.0	0.0	±1.5
500	0.0	0.1	0.0	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.1	0.0	±2.0
4000	0.0	0.1	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Lcq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.1	0.1	± 0.3

QF-TS12-04-04-020664

T P.T.A.

Continuation of Calibration Certificate

Cert. No. : ACL22160
Job No. : VC65AC0069
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.1	0.1	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.0	0.0	± 1.1
26.0	26.0	0.0	± 1.1
25.0	25.0	0.0	± 1.1

QF-TS12-04-04-020664

T P.T.A.

Continuation of Calibration Certificate

Cert. No. : ACL22160
Job No. : VC65AC0069
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	108.0	0.0	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.1	0.1	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.1	0.1	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	135.3	-1.1	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

QF-TS12-04-04-020664

T P.T.A.

Continuation of Calibration Certificate

Cert. No. : ACL22160
Job No. : VC65AC0069
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.5	89.6	0.1	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$
or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

451-451/1 Sirinthorn Rd, Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2433-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.comCert. No. : ACL22161
Pages : 1 of 8

Calibration Certificate


Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00572561 / 170398 / 72899
ID No.: RYG_FS0300

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 06 JULY 2022
Calibration Date : 11-18 JULY 2022
Date of Issue : 19 JULY 2022



Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchuraj
(Thanakul Petchuraj)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced
other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

P.E.A.

Continuation of Calibration Certificate

Cert. No. : ACL22161
Job No. : VC65AC0069
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference
Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0007-22	04-Feb-23
Waveform Generator	33511B	MY52302742	EF-0008-22	04-Feb-23
Digital Multimeter	33461A	MY53220104	EEL.BP. 04/0265	09-Feb-23
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0265	09-Feb-23
Digital Multimeter	34461A	MY60024273	EEL.BP. 05/0265	09-Feb-23
Programmable Attenuator	MAT-1070	62100114	EF-0009-22	07-Feb-23
Condenser Microphone	4180	2977900	AA-1013-22	24-Feb-23
Measuring Amplifier	NA-42KAI	34560495	AA-3005-22	22-Feb-23

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL22161
Job No. : VC65AC0069
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

P.E.A.

QF-TS12-04-04-020664

P.E.A.

Continuation of Calibration Certificate

Cert. No. : ACL22161
Job No. : VC65AC0069
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.95)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
16.2

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	12.0
C - weight	18.3
Flat	24.2

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.4	0.4	0.4	± 1.5
1000	0.0	0.0	0.0	± 1.0
8000	0.4	0.5	0.5	±5.0

QF-TS12-04-04-020664

P.T.A

Continuation of Calibration Certificate

Cert. No. : ACL22161
Job No. : VC65AC0069
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	0.0	0.0	±2.0
125	0.0	0.0	0.0	±1.5
250	0.0	0.0	0.0	±1.5
500	0.0	0.0	0.0	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

P.T.A

Continuation of Calibration Certificate

Cert. No. : ACL22161
Job No. : VC65AC0069
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	53.9	-0.1	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	29.9	-0.1	± 1.1
29.0	28.9	-0.1	± 1.1
28.0	27.9	-0.1	± 1.1
27.0	26.9	-0.1	± 1.1
26.0	25.9	-0.1	± 1.1
25.0	24.9	-0.1	± 1.1

QF-TS12-04-04-020664

P.T.A

Continuation of Calibration Certificate

Cert. No. : ACL22161
Job No. : VC65AC0069
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.1	0.1	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.0	-0.4	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.2	-0.2	±2.0
Negative half cycle	135.4	135.2	-0.2	±2.0

QF-TS12-04-04-020664

P.T.A

Continuation of Calibration Certificate

Cert. No. : ACL22161
Job No. : VC65AC0069
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.7	89.5	-0.2	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$
or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

451-451/1 Sirinshorn Rd., Bangbunru, Bangplad Bangkok 10700 THAILAND.
Tel.0-2435-8800 Fax.0-2433-1679 e-mail:cal-center@sithiphorn.com http://www.sithiphorn.comCert. No. : ACL22058
Pages : 1 of 8


Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42 / Microphone UC-52 / Preamplifier NH-24
Serial No.: 00296518 / 179118 / 87525
ID No.: RYG_FS0431

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %
Received Date : 14 JANUARY 2022
Calibration Date : 21-24 JANUARY 2022
Date of Issue : 25 JANUARY 2022



Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced
other than in full, except with the prior written approval of the head of Calibration Laboratory.

QI-TS12-04-04-020664

QI-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22058
Job No. : VC65AC0043
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference
Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05-0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03-0264	08-Feb-22
Digital Multimeter	34461A	MY60024273	1-15180725251-1	15-Sep-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KA1	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

QI-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22058
Job No. : VC65AC0043
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For ≥ 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long-term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QI-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22058
Job No. : VC65AC0043
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.6

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	11.6
C - weight	17.6
Flat	23.2

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.1	0.1	0.1	± 1.5
1000	0.0	0.0	0.0	± 1.0
8000	0.8	0.9	0.9	±5.0

QI-TS12-04-04-020664

P.L.A

Continuation of Calibration Certificate

Cert. No. : ACL22058
Job No. : VC65AC0043
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	-0.1	0.0	±2.0
125	0.0	0.0	-0.1	±1.5
250	-0.1	0.0	-0.1	±1.5
500	0.0	0.0	-0.1	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.0	0.0	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QI-TS12-04-04-020664

P.L.A

Continuation of Calibration Certificate

Cert. No. : ACL22058
Job No. : VC65AC0043
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.1	0.1	± 1.1
135.0	135.1	0.1	± 1.1
134.0	134.1	0.1	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.1	0.1	± 1.1
114.0	114.1	0.1	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.1	0.1	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.1	0.1	± 1.1
26.0	26.1	0.1	± 1.1
25.0	25.1	0.1	± 1.1

QI-TS12-04-04-020664

P.L.A

Continuation of Calibration Certificate

Cert. No. : ACL22058
Job No. : VC65AC0043
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	116.9	-0.1	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, Lepeak (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	132.0	133.0	0.0	-
One	136.4	135.8	-0.6	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

QI-TS12-04-04-020664

P.L.A

Continuation of Calibration Certificate

Cert. No. : ACL22058
Job No. : VC65AC0043
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.7	89.7	0.0	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$
or any value following calculation providing a level of confidence of approximately 95 %

End of Calibration Certificate

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL21100
Job No. : VC64AC0066
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	8846A	1997025	EEL.BP. 06/0264	05-Feb-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KA1	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

451-451/1 Sirinthorn Rd, Bangumru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.comCert. No. : ACL21100
Pages : 1 of 8


Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No. : 01173609 / 172170 / 74021
ID No. : RYG_FS0388

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTANAKAN 40, PHATTANAKAN ROAD,
KHAENG PHATTANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %
Received Date : 01 SEPTEMBER 2021
Calibration Date : 13-15 SEPTEMBER 2021
Date of Issue : 16 SEPTEMBER 2021



Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced
other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL21100
Job No. : VC64AC0066
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL21100
Job No. : VC64AC0066
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
15.1

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	12.0
C - weight	18.1
Flat	23.9

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.4	0.4	0.4	± 1.5
1000	0.0	0.0	0.0	± 1.0
8000	-0.7	-0.6	-0.6	±5.0

Continuation of Calibration Certificate

Cert. No. : ACL21100
Job No. : VC64AC0066
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	0.0	-0.1	-0.1	±2.0
125	0.0	0.0	0.0	±1.5
250	0.0	0.0	-0.1	±1.5
500	0.0	0.0	-0.1	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL21100
Job No. : VC64AC0066
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.1	0.1	± 1.1
27.0	27.1	0.1	± 1.1
26.0	26.1	0.1	± 1.1
25.0	25.1	0.1	± 1.1

Continuation of Calibration Certificate

Cert. No. : ACL21100
Job No. : VC64AC0066
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.3	-0.1	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.2	-0.2	±2.0

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL21100
Job No. : VC64AC0066
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated	Acceptance
Positive	Negative	Value	Limits
one-half cycle	one-half cycle	(dB)	(dB)
89.6	89.6	0.0	±1.5

12. High level stability

Frequency	SLM Display	SLM Display	Deviated	Acceptance
Weighting	at initial	at final	Value	Limits
(dB)	(dB)	(dB)	(dB)	(dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$
or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

451-451/1 Sirinthorn Rd, Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.comCert. No. : ACL22182
Pages : 1 of 8


Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42 / Microphone UC-52 / Preamplifier NH-24
Serial No. : 00873109 / 171842 / 73485
ID No. : RYG_FS0384

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %
Received Date : 22 AUGUST 2022
Calibration Date : 26-31 AUGUST 2022
Date of Issue : 02 SEPTEMBER 2022



Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchara
(Thanakul Petchara)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced
other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22182
Job No. : VC65AC0077
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference
Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0007-22	04-Feb-23
Waveform Generator	33511B	MY52302742	EF-0008-22	04-Feb-23
Digital Multimeter	33461A	MY53220104	EEL.BP. 04/0265	09-Feb-23
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0265	09-Feb-23
Digital Multimeter	34461A	MY60024273	EEL.BP. 05/0265	09-Feb-23
Programmable Attenuator	MAT-1070	62100114	EF-0009-22	07-Feb-23
Condenser Microphone	4180	2977900	AA-1013-22	24-Feb-23
Measuring Amplifier	NA-42KAI	34560495	AA-3005-22	22-Feb-23

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
- 3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL22182
Job No. : VC65AC0077
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.4	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22182
Job No. : VC65AC0077
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.95)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
16.5

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	11.2
C - weight	17.6
Flat	23.3

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.4	0.5	0.5	± 1.5
1000	0.0	0.0	0.0	± 1.0
8000	-2.4	-2.4	-2.4	±5.0

Continuation of Calibration Certificate

Cert. No. : ACL22182
Job No. : VC65AC0077
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	-0.1	-0.1	±2.0
125	0.0	0.0	-0.1	±1.5
250	-0.1	-0.1	-0.1	±1.5
500	-0.1	0.0	-0.1	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.0	0.0	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22182
Job No. : VC65AC0077
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.1	0.1	± 1.1
136.0	136.1	0.1	± 1.1
135.0	135.1	0.1	± 1.1
134.0	134.1	0.1	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.1	0.1	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.1	0.1	± 1.1
114.0	114.1	0.1	± 1.1
109.0	109.1	0.1	± 1.1
104.0	104.1	0.1	± 1.1
99.0	99.1	0.1	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	29.9	-0.1	± 1.1
29.0	28.9	-0.1	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.0	0.0	± 1.1
26.0	26.0	0.0	± 1.1
25.0	24.9	-0.1	± 1.1

Continuation of Calibration Certificate

Cert. No. : ACL22182
Job No. : VC65AC0077
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	116.9	-0.1	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
SEL	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.3	-0.1	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22182
Job No. : VC65AC0077
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated	Acceptance
Positive	Negative	Value	Limits
one-half cycle	one-half cycle	(dB)	(dB)
89.6	89.5	-0.1	±1.5

12. High level stability

Frequency	SLM Display	SLM Display	Deviated	Acceptance
Weighting	at initial	at final	Value	Limits
	(dB)	(dB)	(dB)	(dB)
A-weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$
or any value following calculation, providing a level of confidence of approximately 95 %


End of Calibration Certificate

451-451/1 Sirinthorn Rd., Sangbunru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:center@sithiporn.com http://www.sithiporn.comCert. No. : ACL22193
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00597167 / 157778 / 34375
ID No.: RYG_FS0437

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.Location :
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %
Received Date : 06 SEPTEMBER 2022
Calibration Date : 07-09 SEPTEMBER 2022
Date of Issue : 14 SEPTEMBER 2022

Calibrated by : Nuthakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced
other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22193
Job No. : VC65AC0081
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference
Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0007-22	04-Feb-23
Waveform Generator	33511B	MY52302742	EF-0008-22	04-Feb-23
Digital Multimeter	33461A	MY53220104	EEL.BP. 04/0265	09-Feb-23
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0265	09-Feb-23
Digital Multimeter	34461A	MY60024273	EEL.BP. 05/0265	09-Feb-23
Programmable Attenuator	MAT-1070	62100114	EF-0009-22	07-Feb-23
Condenser Microphone	4180	2977900	AA-1013-22	24-Feb-23
Measuring Amplifier	NA-42KAJ	34560495	AA-3005-22	22-Feb-23

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL22193
Job No. : VC65AC0081
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long-term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22193
Job No. : VC65AC0081
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.95)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
18.3

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	15.1
C - weight	21.4
Flat	27.0

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.3	0.4	0.3	± 1.5
1000	-0.1	-0.1	-0.1	± 1.0
8000	-0.6	-0.6	-0.6	±5.0

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22193
Job No. : VC65AC0081
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	-0.1	-0.1	±2.0
125	0.0	0.0	0.0	±1.5
250	0.0	0.0	-0.1	±1.5
500	0.0	0.0	0.0	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22193
Job No. : VC65AC0081
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.1	0.1	± 1.1
84.0	84.1	0.1	± 1.1
79.0	79.1	0.1	± 1.1
74.0	74.1	0.1	± 1.1
69.0	69.1	0.1	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.1	0.1	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.1	0.1	± 1.1
26.0	26.0	0.0	± 1.1
25.0	24.9	-0.1	± 1.1

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22193
Job No. : VC65AC0081
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.1	0.1	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.3	-0.1	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.2	-0.2	±2.0
Negative half cycle	135.4	135.2	-0.2	±2.0

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22193
Job No. : VC65AC0081
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.6	89.6	0.0	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22231
Job No. : VC65AC0088
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM). The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0907-22	04-Feb-23
Waveform Generator	33511B	MY52302742	EF-0008-22	04-Feb-23
Digital Multimeter	33461A	MY53220104	EEL.BP. 04/0265	09-Feb-23
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0265	09-Feb-23
Digital Multimeter	33461A	MY60024273	EEL.BP. 05/0265	09-Feb-23
Programmable Attenuator	MAT-1070	62100114	EF-0009-22	07-Feb-23
Condenser Microphone	4180	2977900	AA-1013-22	24-Feb-23
Measuring Amplifier	NA-42KAJ	34560495	AA-3005-22	22-Feb-23

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
- 3.2 Thailand Institute of Scientific and Technological Research (TISTR).

QF-TS12-04-04-020664

451-451/1 Sirinthorn Rd, Bangbunru, Banglue Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:center@sithiporn.com http://www.sithiporn.com



Cert. No. : ACL22231
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RJON
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00472126 / 176915 / 88180
ID No.: RYG_FS0301

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %
Received Date : 03 OCTOBER 2022
Calibration Date : 18-19 OCTOBER 2022
Date of Issue : 20 OCTOBER 2022

REVIEW BY	<i>Nathakorn P.</i>
APPROVED BY	<i>Thanakul P.</i>
NEXT CAL. DATE	11/10/23

Calibrated by : Nathakorn Pisutpaian

Approved by : *T. Petchurni*
(Thanakul Petchurni)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22231
Job No. : VC65AC0088
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22231
Job No. : VC65AC0088
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.95)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.6

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	9.9
C - weight	16.7
Flat	22.5

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	-7.6	-7.6	-7.3	± 1.5
1000	0.5	0.5	0.5	± 1.0
8000	-5.1	-5.1	-5.1	±5.0

QF-TS12-04-04-020664

T. R. R.

Continuation of Calibration Certificate

Cert. No. : ACL22231
Job No. : VC65AC0088
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	-0.1	-0.1	±2.0
125	0.0	0.0	-0.1	±1.5
250	0.0	-0.1	-0.1	±1.5
500	0.0	0.0	-0.1	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

T. R. R.

Continuation of Calibration Certificate

Cert. No. : ACL22231
Job No. : VC65AC0088
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	131.9	-0.1	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.0	0.0	± 1.1
26.0	25.9	-0.1	± 1.1
25.0	24.9	-0.1	± 1.1

QF-TS12-04-04-020664

T. R. R.

Continuation of Calibration Certificate

Cert. No. : ACL22231
Job No. : VC65AC0088
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.4	0.0	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

QF-TS12-04-04-020664

T. R. R.

Continuation of Calibration Certificate

Cert. No. : ACL22231
Job No. : VC65AC0088
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.6	89.6	0.0	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$
or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

451-451/1 Sirinthorn Rd, Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.com



Cert. No. : ACL22024
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00709746 / 187332 / 01297
ID No.: RYG_FS0491

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWANG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %
Received Date : 05 JANUARY 2022
Calibration Date : 10-12 JANUARY 2022
Date of Issue : 13 JANUARY 2022

NEW BY : *[Signature]*
APPROVED BY : *[Signature]*
EXT CAL DATE : 10/1/23

Calibrated by : Nathakorn Pisutpaisan

Approved by : *[Signature]*
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced
other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22024
Job No. : VC65AC0040
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference
Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	34461A	MY60024273	1-15180725251-1	15-Sep-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KAJ	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL22024
Job No. : VC65AC0040
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22024
Job No. : VC65AC0040
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
15.1

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	13.4
C - weight	19.8
Flat	25.2

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.1	0.1	0.1	± 1.5
1000	0.0	0.0	0.0	± 1.0
8000	1.1	1.2	1.2	±5.0

Continuation of Calibration Certificate

Cert. No. : ACL22024
Job No. : VC65AC0040
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	0.0	-0.1	±2.0
125	0.0	0.0	0.0	±1.5
250	0.0	0.0	0.0	±1.5
500	0.0	0.1	0.0	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.1	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22024
Job No. : VC65AC0040
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.1	0.1	± 1.1
26.0	26.1	0.1	± 1.1
25.0	25.1	0.1	± 1.1

Continuation of Calibration Certificate

Cert. No. : ACL22024
Job No. : VC65AC0040
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	108.0	0.0	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.1	0.1	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.1	0.1	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	135.5	-0.9	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.1	0.1	-
Positive half cycle	135.4	135.2	-0.2	±2.0
Negative half cycle	135.4	135.2	-0.2	±2.0

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22024
Job No. : VC65AC0040
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.6	89.6	0.0	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	136.9	0.1	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$
or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

451-451/1 Sirinthorn Rd, Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.comNSC-TSM-TS 17025
CALIBRATION 0394Cert. No. : ACL22026
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00900071 / 188464 / 01733
ID No.: RYG_FS0492

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %
Received Date : 05 JANUARY 2022
Calibration Date : 10-12 JANUARY 2022
Date of Issue : 13 JANUARY 2022

REVIEW BY	Manfon P.
APPROVED BY	
NEXT CAL DATE	10/1/23

Calibrated by : Nathakorn Pisutpaisan

Approved by :
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced
other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22026
Job No. : VC65AC0040
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference
Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53230104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	34461A	MY60024273	1-15180725251-1	15-Sep-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KAI	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL22026
Job No. : VC65AC0040
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fall	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22026
Job No. : VC65AC0040
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.8

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	11.6
C - weight	17.8
Flat	23.6

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.2	0.2	0.2	± 1.5
1000	-0.1	-0.1	-0.1	± 1.0
8000	0.2	0.3	0.3	±5.0

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22026
Job No. : VC65AC0040
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	0.0	0.0	-0.1	±2.0
125	0.0	0.0	0.0	±1.5
250	0.0	0.0	0.0	±1.5
500	0.0	0.1	0.0	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.1	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22026
Job No. : VC65AC0040
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	38.9	-0.1	± 1.1
34.0	33.9	-0.1	± 1.1
30.0	29.9	-0.1	± 1.1
29.0	28.9	-0.1	± 1.1
28.0	27.9	-0.1	± 1.1
27.0	26.8	-0.2	± 1.1
26.0	25.9	-0.1	± 1.1
25.0	24.9	-0.1	± 1.1

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22026
Job No. : VC65AC0040
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.1	0.1	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.2	-0.2	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.1	0.1	-
Positive half cycle	135.4	135.2	-0.2	±2.0
Negative half cycle	135.4	135.2	-0.2	±2.0

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22026
Job No. : VC65AC0040
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.6	89.5	-0.1	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$
or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22027
Job No. : VC65AC0040
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	34461A	MY60024273	1-15180725251-1	15-Sep-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KA1	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

QF-TS12-04-04-020664

451-451/1 Sirinthorn Rd, Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.comCert. No. : ACL22027
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No. : 00900072 / 188465 / 01734
ID No. : RYG_FS0493

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 05 JANUARY 2022
Calibration Date : 10-12 JANUARY 2022
Date of Issue : 13 JANUARY 2022

Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced
other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22027
Job No. : VC65AC0040
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C' sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22027
Job No. : VC65AC0040
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.8

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	9.9
C - weight	16.9
Flat	22.6

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.4	0.4	0.4	± 1.5
1000	0.0	0.0	0.0	± 1.0
8000	-1.2	-1.1	-1.1	±5.0

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22027
Job No. : VC65AC0040
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	-0.1	-0.1	±2.0
125	0.0	0.0	-0.1	±1.5
250	0.0	0.0	-0.1	±1.5
500	0.0	0.0	-0.1	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.0	0.0	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22027
Job No. : VC65AC0040
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.1	0.1	± 1.1
134.0	134.1	0.1	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.1	0.1	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.1	0.1	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.0	0.0	± 1.1
26.0	26.1	0.1	± 1.1
25.0	25.1	0.1	± 1.1

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22027
Job No. : VC65AC0040
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	116.9	-0.1	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.8	-0.2	1.5 ; -5.0
	2	8	108.0	107.9	-0.1	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, Lepeak (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.0	-0.4	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22027
Job No. : VC65AC0040
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.6	89.6	0.0	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$
or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

451-451/1 Sirinthorn Rd, Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.comCert. No. : ACL22028
Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RJON
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00900073 / 188466 / 01735
ID No.: RYG_FS0494

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %
Received Date : 05 JANUARY 2022
Calibration Date : 10-12 JANUARY 2022
Date of Issue : 13 JANUARY 2022

REVIEW BY	<i>Thakorn P.</i>
APPROVED BY	<i>Thakorn P.</i>
NEXT CAL. DATE	10/1/23

Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchum
(Thanakul Petchumai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced
other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22028
Job No. : VC65AC0040
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference
Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	34461A	MY60024273	1-15180725251-1	15-Sep-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KA1	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL22028
Job No. : VC65AC0040
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fall	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22028
Job No. : VC65AC0040
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.8

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	11.6
C - weight	17.8
Flat	23.5

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.1	0.2	0.2	± 1.5
1000	0.0	0.0	0.0	± 1.0
8000	0.0	0.1	0.1	±5.0

Continuation of Calibration Certificate

Cert. No. : ACL22028
Job No. : VC65AC0040
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	-0.1	-0.1	±2.0
125	0.0	0.0	-0.1	±1.5
250	-0.1	-0.1	-0.1	±1.5
500	0.0	0.0	-0.1	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.0	0.0	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22028
Job No. : VC65AC0040
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.1	0.1	± 1.1
135.0	135.1	0.1	± 1.1
134.0	134.1	0.1	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.1	0.1	± 1.1
114.0	114.1	0.1	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.1	0.1	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	30.0	0.0	± 1.1
29.0	29.0	0.0	± 1.1
28.0	28.1	0.1	± 1.1
27.0	27.1	0.1	± 1.1
26.0	26.1	0.1	± 1.1
25.0	25.1	0.1	± 1.1

Continuation of Calibration Certificate

Cert. No. : ACL22028
Job No. : VC65AC0040
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	116.9	-0.1	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.8	-0.2	1.5 ; -5.0
	2	8	108.0	107.9	-0.1	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.4	0.0	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22028
Job No. : VC65AC0040
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.5	89.7	0.2	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$
or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

451-451/1 Sirinthorn Rd., Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.comCert. No. : ACL22029
Pages : 1 of 8

Calibration Certificate


Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00900074 / 188467 / 01736
ID No.: RYG_FS0495

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 05 JANUARY 2022
Calibration Date : 10-12 JANUARY 2022
Date of Issue : 13 JANUARY 2022



Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchur
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced
other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22029
Job No. : VC65AC0040
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference
Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0012-21	10-Feb-22
Waveform Generator	33511B	MY52302742	EF-0011-21	10-Feb-22
Digital Multimeter	33461A	MY53220104	EEL.BP. 05/0264	10-Feb-22
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0264	08-Feb-22
Digital Multimeter	34461A	MY60024273	1-15180725251-1	15-Sep-22
Programmable Attenuator	MAT-1070	62100114	1500-07774E	08-Mar-22
Condenser Microphone	4180	2977900	AA-1008-21	05-Feb-22
Measuring Amplifier	NA-42KA1	34560495	AA-3003-21	16-Feb-22

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL22029
Job No. : VC65AC0040
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22029
Job No. : VC65AC0040
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.96)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
14.2

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	10.8
C - weight	17.4
Flat	23.3

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.2	0.2	0.2	± 1.5
1000	0.0	0.0	0.0	± 1.0
8000	-0.8	-0.7	-0.7	±5.0

Continuation of Calibration Certificate

Cert. No. : ACL22029
Job No. : VC65AC0040
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	-0.2	0.0	±2.0
125	-0.1	0.0	0.0	±1.5
250	0.0	0.0	-0.1	±1.5
500	0.0	0.0	-0.1	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.1	0.1	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Lcq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

T. Bha.

QF-TS12-04-04-020664

T. Bha.

Continuation of Calibration Certificate

Cert. No. : ACL22029
Job No. : VC65AC0040
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	± 1.1
136.0	136.0	0.0	± 1.1
135.0	135.0	0.0	± 1.1
134.0	134.0	0.0	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.0	0.0	± 1.1
124.0	124.0	0.0	± 1.1
119.0	119.0	0.0	± 1.1
114.0	114.0	0.0	± 1.1
109.0	109.0	0.0	± 1.1
104.0	104.0	0.0	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.1	0.1	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.1	0.1	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	29.9	-0.1	± 1.1
29.0	28.9	-0.1	± 1.1
28.0	28.0	0.0	± 1.1
27.0	27.0	0.0	± 1.1
26.0	25.9	-0.1	± 1.1
25.0	24.9	-0.1	± 1.1

Continuation of Calibration Certificate

Cert. No. : ACL22029
Job No. : VC65AC0040
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.1	0.1	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	135.8	-0.6	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

QF-TS12-04-04-020664

T. Bha.

QF-TS12-04-04-020664

T. Bha.

Continuation of Calibration Certificate

Cert. No. : ACL22029
Job No. : VC65AC0040
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.6	89.6	0.0	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$
or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

451-451/1 Sirinthorn Rd., Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:center@sithiporn.com http://www.sithiporn.comCert. No. : ACL22181
Pages : 1 of 8


Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-42/ Microphone UC-52 / Preamplifier NH-24
Serial No.: 00873057 / 171591 / 73333
ID No.: RYG_FS0381

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAENG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %
Received Date : 22 AUGUST 2022
Calibration Date : 26-31 AUGUST 2022
Date of Issue : 02 SEPTEMBER 2022



Calibrated by : Nathakom Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced
other than in full, except with the prior written approval of the head of Calibration Laboratory.

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22181
Job No. : VC65AC0077
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on JEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference
Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0007-22	04-Feb-23
Waveform Generator	33511B	MY52302742	EF-0008-22	04-Feb-23
Digital Multimeter	33461A	MY53220104	EEL.BP. 04/0265	09-Feb-23
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0265	09-Feb-23
Digital Multimeter	34461A	MY60024273	EEL.BP. 05/0265	09-Feb-23
Programmable Attenuator	MAT-1070	62100114	EF-0009-22	07-Feb-23
Condenser Microphone	4180	2977900	AA-1013-22	24-Feb-23
Measuring Amplifier	NA-42KAI	34560495	AA-3005-22	22-Feb-23

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
- 3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL22181
Job No. : VC65AC0077
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.4	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	-	-	-	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

QF-TS12-04-04-020664

QF-TS12-04-04-020664

Continuation of Calibration Certificate

Cert. No. : ACL22181
Job No. : VC65AC0077
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.95)	93.9	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
15.4

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	12.0
C - weight	18.3
Flat	24.0

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0.2	0.2	0.2	± 1.5
1000	0.0	0.0	0.0	± 1.0
8000	0.2	0.2	0.2	±5.0

QF-TS12-04-04-020664

T. Petch.

Continuation of Calibration Certificate

Cert. No. : ACL22181
Job No. : VC65AC0077
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.1	0.1	± 1.1
136.0	136.1	0.1	± 1.1
135.0	135.1	0.1	± 1.1
134.0	134.1	0.1	± 1.1
133.0	133.0	0.0	± 1.1
132.0	132.0	0.0	± 1.1
131.0	131.0	0.0	± 1.1
129.0	129.1	0.1	± 1.1
124.0	124.1	0.1	± 1.1
119.0	119.1	0.1	± 1.1
114.0	114.1	0.1	± 1.1
109.0	109.1	0.1	± 1.1
104.0	104.1	0.1	± 1.1
99.0	99.0	0.0	± 1.1
94.0	94.0	0.0	± 1.1
89.0	89.0	0.0	± 1.1
84.0	84.0	0.0	± 1.1
79.0	79.0	0.0	± 1.1
74.0	74.0	0.0	± 1.1
69.0	69.0	0.0	± 1.1
64.0	64.0	0.0	± 1.1
59.0	59.0	0.0	± 1.1
54.0	54.0	0.0	± 1.1
49.0	49.0	0.0	± 1.1
44.0	44.0	0.0	± 1.1
39.0	39.0	0.0	± 1.1
34.0	34.0	0.0	± 1.1
30.0	29.9	-0.1	± 1.1
29.0	28.9	-0.1	± 1.1
28.0	28.0	0.0	± 1.1
27.0	26.9	-0.1	± 1.1
26.0	25.9	-0.1	± 1.1
25.0	25.0	0.0	± 1.1

QF-TS12-04-04-020664

T. Petch.

Continuation of Calibration Certificate

Cert. No. : ACL22181
Job No. : VC65AC0077
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	-0.1	-0.1	-0.1	±2.0
125	-0.1	0.0	-0.1	±1.5
250	0.0	0.0	-0.1	±1.5
500	0.0	0.0	-0.1	±1.5
1000	0.0	0.0	0.0	±1.0
2000	0.0	0.0	0.0	±2.0
4000	0.0	0.0	0.0	±3.0
8000	0.0	0.0	0.0	±5.0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	0.0	-
C - weight	94.0	0.0	± 0.2
Flat	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	0.0	-
Slow	94.0	0.0	± 0.1
Leq	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.3

QF-TS12-04-04-020664

T. Petch.

Continuation of Calibration Certificate

Cert. No. : ACL22181
Job No. : VC65AC0077
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±1.1

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, Lepeak (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
One	136.4	136.1	-0.3	±3.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	-
Positive half cycle	135.4	135.1	-0.3	±2.0
Negative half cycle	135.4	135.1	-0.3	±2.0

QF-TS12-04-04-020664

T. Petch.

Continuation of Calibration Certificate

Cert. No. : ACL22181
Job No. : VC65AC0077
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.5	89.6	0.1	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.3

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate

QF-TS12-04-04-020664

7. Rtn.

CERTIFICATE OF CALIBRATION

ISSUED BY: Cirrus Research plc
DATE OF ISSUE: 07/09/21
CERTIFICATE NUMBER: 162333

REVIEW BY: *Markem P.*
APPROVED BY: *M.A. Ch.*
NEXT CAL. DATE: 7/9/22

Page 1 of 1

Test engineer:
Nigel Smith
Electronically signed:

Nigel Smith

doseBadge Reader

Instrument

Manufacturer: Cirrus Research plc
Model Number: RC:110A
Serial Number: 73729
Notes:

Calibration Procedure

The tests were carried out in accordance with the requirements of IEC 60942:2003 where applicable.

Date of Calibration: 07 September 2021

Functionality Results

Function	Result
Keypad	Pass
Battery Power	Pass
Display	Pass
Communication	Pass
Charge Bk Up	Pass
Clock	Pass

Calibration Results

	Level (dB)	Frequency (Hz)	Distortion (% THD + Noise)
Initial	113.90	998.2	0.46
Adjusted	114.00	998.2	0.46
Uncertainty	± 0.11	± 0.14	± 0.10
Tolerances	± 0.60	± 2.00	± 4.00

Environmental Conditions

Pressure: 101.40 kPa
Temperature: 22.3 °C
Humidity: 59.4 %

Notes

This certificate provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory. The results within this certificate relate only to the items calibrated. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%.

CERTIFICATE OF CALIBRATION

ISSUED BY: Cirrus Research plc
DATE OF ISSUE: 07/09/21
CERTIFICATE NUMBER: 162335

REVIEW BY: *Markem P.*
APPROVED BY: *M.A. Ch.*
NEXT CAL. DATE: 7/9/22

Page 1 of 1

Test engineer:
Nigel Smith
Electronically signed:

Nigel Smith

doseBadge Reader

Instrument

Manufacturer: Cirrus Research plc
Model Number: RC:110A
Serial Number: 89107
Notes:

Calibration Procedure

The tests were carried out in accordance with the requirements of IEC 60942:2003 where applicable.

Date of Calibration: 07 September 2021

Functionality Results

Function	Result
Keypad	Pass
Battery Power	Pass
Display	Pass
Communication	Pass
Charge Bk Up	Pass
Clock	Pass

Calibration Results

	Level (dB)	Frequency (Hz)	Distortion (% THD + Noise)
Initial	113.90	1000.2	0.20
Adjusted	114.00	1000.2	0.20
Uncertainty	± 0.11	± 0.14	± 0.10
Tolerances	± 0.60	± 2.00	± 4.00

Environmental Conditions

Pressure: 101.40 kPa
Temperature: 22.4 °C
Humidity: 60.2 %

Notes

This certificate provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory. The results within this certificate relate only to the items calibrated. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%.



63/14-15,67/35-36, Soi Petchkasem 7,7/1, Petchkasem Rd,
Walthapra, Bangkokyoi, Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranatee.com



CERTIFICATE OF CALIBRATION

Certificate No.: CL-134-65
Page 1 of 2

Equipment Name: Heat Stress Monitor
Manufacturer: DeltaOHM
Model: HD32.2
Serial No: 15020735
ID No: RYG_FS0231

Customer
Name: ALS laboratory group (thailand) Co., Ltd.
Address: 104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand.

Received date: 27 Jul 2022
Calibration date: 3 Aug 2022
Issue date: 8 Aug 2022

Reference Used During Calibration
1. Standard Temperature Probe Model: STS-100 A500, Serial No: 667682-09, Due date: 23 Mar 2023
2. Digital Temperature Indicator Model: DTI-1000A MK II, Serial No: 671407-00591 Due date: 04 June 2022

Calibration Condition
Temperature: (23±3) °C
Relative Humidity: (55±15)%

Calibration Procedure
The temperature calibration was done by In-House calibration method as WI-CL-001 according to comparison method with standard digital temperature indicator and standard temperature probe. The temperature scale use was based on ITS-90.

Traceability
The measurement results are traceable to the international system of units (SI) through National Institute of Metrology Thailand (NIMT) Certificate number: TT-0034-22, Certificate number: ER-0032-21

Calibrated by
☒ Mr. Sorawit Thachalad
☒ Miss Jitraporn Lertsomphol

Approved Signatory:
Mr. Perinya Booncharoen
Calibration Department Manager





63/14-15,67/35-36, Soi Petchkasem 7/71, Petchkasem Rd.
Walthapra, Bangkokhyai, Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranatee.com



Certificate No.: CL-134-65
Page 2 of 2

Result of Calibration: ☒ Without Adjustment ☐ With Adjustment

Calibration Range: 20 - 40 °C

Function: Table 1: This equipment was connected with wet bulb probe Model: HP3201.2 S/N: 15035050.
Dimension: Diameter 14 mm, Length 170 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
30	20.043	20.1	0.1	0.099
30	25.037	25.1	0.0	0.14
30	30.027	30.0	0.0	0.14
30	35.021	35.0	0.0	0.099
30	40.012	40.0	0.0	0.099

Table 2: This equipment was connected with temperature probe Model: TP3207.2 S/N: 15033221.
Dimension: Diameter 14 mm, Length 150 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
70	20.043	20.1	0.1	0.099
70	25.037	24.9	-0.1	0.099
70	30.028	29.7	-0.3	0.099
70	35.021	34.5	-0.5	0.099
70	40.011	39.5	-0.5	0.099

Table 3: This equipment was connected with Globe thermometer probe Model: TP3276.2 S/N: 17023218.
Dimension: Diameter 8 mm, Length 170 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
110	20.043	20.1	0.1	0.099
110	25.037	25.1	0.1	0.099
110	30.028	30.1	0.1	0.099
110	35.021	35.1	0.1	0.099
110	40.012	40.1	0.1	0.099

UUC*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

☆ End of Certificate ☆



63/14-15,67/35-36, Soi Petchkasem 7/71, Petchkasem Rd.
Walthapra, Bangkokhyai, Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranatee.com



CERTIFICATE OF CALIBRATION

Certificate No.: CL-135-65
Page 1 of 2

Equipment Name: Heat Stress Monitor
Manufacturer: DeltaOHM
Model: H032.2
Serial No: 15020736
ID No: RYG_FS0232

Customer
Name: ALS laboratory group (thailand) Co., Ltd.
Address: 104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand.

Received date: 27 Jul 2022
Calibration date: 3 Aug 2022
Issue date: 8 Aug 2022

Reference Used During Calibration
1. Standard Temperature Probe Model: STS-100 A500, Serial No: 667682-09, Due date: 23 Mar 2023
2. Digital Temperature Indicator Model: DTI-1000-A MK II, Serial No: 671407-00591 Due date: 04 June 2022


Calibration Condition
Temperature: (23±3) °C
Relative Humidity: (55±15)%

Calibration Procedure
The temperature calibration was done by in-house calibration method as WI-CL-001 according to comparison method with standard digital temperature indicator and standard temperature probe. The temperature scale use was based on ITS-90.

Traceability
The measurement results are traceable to the international system of units (SI) through National Institute of Metrology Thailand (NIMT) Certificate number: TT-0034-22, Certificate number: ER-0032-21

REVIEW BY: *Makorn P.*
APPROVED BY: *Mr. Pannya Booncharoen*
NEXT CAL DATE: 3/8/23

Calibrated by
☒ Mr. Soravit Thachalad
☒ Miss Jitraporn Lertsomphol



Approved Signatory: *Mr. Pannya Booncharoen*
Calibration Department Manager

THIS CERTIFICATE REPORT MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.



63/14-15,67/35-36, Soi Petchkasem 7/71, Petchkasem Rd.
Walthapra, Bangkokhyai, Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranatee.com



Certificate No.: CL-135-65
Page 2 of 2

Result of Calibration: ☒ Without Adjustment ☐ With Adjustment

Calibration Range: 20 - 40 °C

Function: Table 1: This equipment was connected with wet bulb probe Model: HP3201.2 S/N: 15027737.
Dimension: Diameter 14 mm, Length 170 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
30	20.043	20.1	0.1	0.099
30	25.037	25.2	0.1	0.16
30	30.028	30.1	0.1	0.099
30	35.021	35.1	0.1	0.099
30	40.012	40.1	0.1	0.099

Table 2: This equipment was connected with temperature probe Model: TP3207.2 S/N: 15015503.
Dimension: Diameter 14 mm, Length 150 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
70	20.044	20.1	0.1	0.099
70	25.037	24.9	-0.1	0.099
70	30.027	29.8	-0.2	0.099
70	35.021	34.6	-0.4	0.099
70	40.012	39.5	-0.5	0.099

Table 3: This equipment was connected with Globe thermometer probe Model: TP3276.2 S/N: 15031164.
Dimension: Diameter 8 mm, Length 170 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
110	20.043	20.1	0.1	0.099
110	25.037	25.1	0.1	0.099
110	30.027	30.1	0.1	0.099
110	35.021	35.1	0.1	0.099
110	40.013	40.1	0.1	0.099

UUC*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

☆ End of Certificate ☆



63/14-15,67/35-36, Soi Petchkasem 7/71, Petchkasem Rd.
Walthapra, Bangkokhyai, Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranatee.com



CERTIFICATE OF CALIBRATION

Certificate No.: CL-019-65
Page 1 of 2

Equipment Name: Heat Stress Monitor with Sensor
Manufacturer: DeltaOHM
Model: H032.2
Serial No: 18018311
ID No: RYG_FS0356

Customer
Name: ALS laboratory group (thailand) Co., Ltd.
Address: 104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand.

Received date: 10 JAN 2022
Calibration date: 16 FEB 2022
Issue date: 17 FEB 2022

Reference Used During Calibration
1. Standard Temperature Probe Model: STS-100 A500, Serial No: 667682-09, Due date: 23 Mar 2022
2. Digital Temperature Indicator Model: DTI-1000-A MK II, Serial No: 671407-00591 Due date: 04 June 2022

Calibration Condition
Temperature: (23±3) °C
Relative Humidity: (55±15)%

Calibration Procedure
The temperature calibration was done by in-house calibration method as WI-CL-001 according to comparison method with standard digital temperature indicator and standard temperature probe. The temperature scale use was based on ITS-90.

Traceability
The measurement results are traceable to the international system of units (SI) through National Institute of Metrology Thailand (NIMT) Certificate number: TT-0036-21, Certificate number: ER-0032-21

REVIEW BY: *Makorn P.*
APPROVED BY: *Mr. Pannya Booncharoen*
NEXT CAL DATE: 16 FEB 2023

Calibrated by
☒ Mr. Soravit Thachalad
☒ Miss Orathai Wiatwattaya



Approved Signatory: *Mr. Pannya Booncharoen*
Calibration Department Manager

THIS CERTIFICATE REPORT MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.



63/14-15,67/35-36, Soi Petchkasem 7/71, Petchkasem Rd,
Wathapra, Bangkokhyai, Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranate.com



Certificate No.: CL-019-65
Page 2 of 2

Result of Calibration: ☒ Without Adjustment ☐ With Adjustment

Calibration Range: 20 ~ 40 °C

Function:

Table 1: This equipment was connected with wet bulb probe Model: HP3201.2 S/N: 18021466.
Dimension: Diameter 14 mm. Length 170 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
30	20.054	20.0	-0.1	0.099
30	25.043	25.0	0.0	0.099
30	30.036	30.0	0.0	0.099
30	35.026	35.0	0.0	0.099
30	40.027	40.0	0.0	0.099

Table 2: This equipment was connected with temperature probe Model: TP3207.2 S/N: 18021258.
Dimension: Diameter 14 mm. Length 150 mm.


Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
70	20.051	20.1	0.0	0.099
70	24.990	25.1	0.1	0.099
70	29.917	29.9	0.0	0.099
70	34.873	34.7	0.2	0.099
70	39.864	39.6	0.3	0.099

Table 3: This equipment was connected with Globe thermometer probe Model: TP3276.2 S/N: 18020493.
Dimension: Diameter 8 mm. Length 170 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
110	20.054	20.1	0.0	0.099
110	25.044	25.1	0.1	0.099
110	30.036	30.1	0.1	0.099
110	35.029	35.1	0.1	0.099
110	40.029	40.1	0.1	0.099

UUC* : Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor $k=2$ providing a level of confidence of approximately 95%.



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
5344 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG, BANGKOK 10250
TEL. 0-2717-3000-24 FAX. 0-2719-9484



Certificate of Calibration

Certificate No.: 22PH188
Page: 1 of 2

Equipment: Lux Meter
Manufacturer: Tenmars
Model: TM-201L
Serial No.: 190702490
ID No.: RYG_FS0471
Condition As-Received: Used Item
Received Date: 07 April 2022
Calibration Date: 08 April 2022

This certificate may not be reproduced other than in full, except with the prior written approval of the head of Corporate Services 3: Equipment Calibration and Testing Services.

Reference: 2204-0185WSC Submitted by: ALS Laboratory Group (Thailand) Co., Ltd.
Ambient Temperature: (23 ± 2) °C
Relative Humidity: (50 ± 15) %

104 Phatthanakan 40, Phatthanakan Rd.,
Khwaeng Phatthanakan, Khet Suan Luang,
Bangkok 10250 Thailand

Procedure used: Calibration were conducted using In-house calibration procedure CP-PH01 by measuring against luminous-intensity standard lamp (source-based method) According to the inverse square law measurement method.

Condition of this result of calibration

1. Reference standards instruments:

Instrument	Model	Serial No.	Certificate No.	Due Date
1) Photometry & Encoder	LMguide 9.6 m	120RC003	61-140006-1	30 Apr 2022
2) Luminous intensity standard lamp	OL FEL-U	F-1543	TP-1020-21	02 May 2022


2. This result of calibration was made on requested at the point specified by customer.

3. Test Equipment : Programmable Voltage/Current Source (Model : OL83A, S/N : 16221394).

4. Test Equipment : Illuminance Meter (Model : 51002, S/N : 080129).

5. The certificate is valid only to the item calibrated on date and place of calibration.

6. This Certification is traceable to the International System of Unit maintained at:-
National Institute of Metrology Thailand (NIMT)



Calibrated by: Nivat Nitis
Issue Date: 20 April 2022

Approved Signatory:
[*] Phalinee Prabpaipal
[] Nuntawat Khamchai

B 0285815



Cert. No.: 22PH188
Page: 2 of 2

Result of calibration: (*) Without adjustment () After adjustment

Function: Illuminance Measurement Range: 200 lx

Standard Value	UUC* Reading	Error	Uncertainty
(lx)	(lx)	(lx)	(± lx)
0	0.0	0.0	0.060
20	20.0	0.0	0.26
50	50.0	0.0	0.65
100	100.0	0.0	1.3
150	150.0	0.0	2.0
190	190.0	0.0	2.5

Function: Illuminance Measurement Range: 2000 lx

Standard Value	UUC* Reading	Error	Uncertainty
(lx)	(lx)	(lx)	(± lx)
200	200	0	2.6
500	500	0	6.5
1000	1000	0	13
1500	1500	0	20
1900	1900	0	25

Function: Illuminance Measurement Range: 20000 lx

Standard Value	UUC* Reading	Error	Uncertainty
(lx)	(lx)	(lx)	(± lx)
2000	1990	-10	26
3000	2990	-10	39
4000	3990	-10	52
5000	4990	-20	65

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%

UUC* = Unit Under Calibration.

-00-



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
5344 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG, BANGKOK 10250
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert.No.: 22CH405
Page: 1 of 3

Certificate of Calibration

Equipment: pH Meter
Manufacturer: Mettler Toledo
Model: Seven Compact S220
Serial No.: C104059460
ID No.: RYG_EN0183
Condition As-Received: Used Item
Received Date: 16 March 2022
Calibration Date: 17 March 2022
Reference: 2203-0611DSC-4
Submitted by: ALS Laboratory Group (Thailand) Co., Ltd.
Rayong Branch
616/10 Moo 5 T.Moenam Khu,
A.Phuakdaeng, Rayong 21140, Thailand

Ambient Temperature: (25 ± 2.5) °C
Relative Humidity: (50 ± 15) %
Calibration Procedure: In-house method:
- CP-CH5 by direct measurement with standard voltage calibrator and direct measurement with certified reference material (CRM)
- CP-CH8 by comparison with standard thermometer

Calibrated by: Warakorn Lerngratukul

Approved by:
Approved Signatory

(/) Mailee Bulkruea
() Saitip Meangmai
() Warakorn Lerngratukul

Issue Date: 22 March 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written approval of the head of Corporate Services 3: Equipment Calibration and Testing Services

1104092

A 0037307



Cert.No.: 22CH405
Page.: 2 of 3

Condition of this calibration result

1. Reference Standard Instrument :-

Instrument	Serial No.	ID No.	Cert. No.	Due Date
1) Document Process Calibrator	54030049	130RC116	21E2682	25 Aug 2022
2) Ref. Standard Thermometer	4982054	110RC044	21H1201	26 Oct 2022

This certification is traceable to the International System of Unit maintained at:-
- Traceable to National Institute of Metrology (Thailand), NIMT

2. Certified Reference Materials : The measurement results are traceable to SI through CPA chem Ltd.,
ANSI-ASQ National Accreditation Board, Accredited No. AR-1835

Buffer Solution	Manufacturer	Lot No.	Exp. date
pH 4.008	CPA chem	788995	01 Jan 2024
pH 6.982	CPA chem	761017	02 Aug 2022
pH 10.015	CPA chem	766824	04 Sep 2022

3. This certificate is valid only to the item calibrated on date and place of calibration.

Calibration Results

Function : mV Measurement

Performing standard curve by Fluke at pH (4,7,10)

Unit Under Calibration	Nominal Value	Standard Voltage Input	Actual Reading	Uncertainty of Measurement	Coverage factor
	pH	mV	mV	pH	k
pH Meter S/N.: C104059460	4.000	177.48	177.4	0.058	2.00
	7.000	0.00	-0.1	0.058	2.00
	10.000	-177.48	-177.5	0.058	2.00



Cert.No.: 22CH405
Page.: 3 of 3

Calibration Results

Function : pH Measurement

Performing three buffers standard curve by using buffer nominal pH (4,7,10)

Unit Under Calibration	Standard pH Buffer Solution	Actual pH Reading	Actual mV Reading (mV)	Uncertainty of pH measurement (±)	Coverage factor k
pH Electrode S/N.: 1453404	4.008	4.010	177.7	0.0046	2.00
	6.982	6.988	3.6	0.0084	2.00
	10.015	10.010	-172.9	0.0073	2.05

Function : Temperature Measurement

(*) Without adjustment

This equipment was connected with Temperature Probe;

- Model :	InLab Expert Pro-ISM
- Serial No. :	1453404
Dimension of probe:	
- Length :	120 mm.
- Diameter :	12 mm.
- Immersion Depth :	100 mm.

Calibration Point (°C)	Standard Temperature (°C)	UUC* Reading (°C)	Error (°C)	Uncertainty of measurement (± °C)	Coverage factor k
25.0	25.002	24.9	-0.102	0.13	2.00

Remark : - UUC* = Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-000-

a 1100955

a 1100954



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
53/44 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG, BANGKOK 10250
TEL. 0-2717-3000-24 FAX. 0-2719-9484



Certificate of Calibration

Certificate No.: 22E986
Page : 1 of 2

Equipment : pH Meter
Manufacturer: Mettler Toledo
Model : SevenCompact S220
Serial No.: C104059460
ID No.: RYG_EN0183

Condition As-Received: Used Item

Received Date: 16 March 2022
Calibration Date: 21 March 2022

Reference: 2203-08110SG
Ambient Temperature: (23 ± 2) °C
Relative Humidity: (50 ± 10) %
Submitted by: ALS Laboratory Group (Thailand) Co.,Ltd. Rayong Branch
616/10 Moo 5 T.Maenam Khu, A.Plusikdaeng, Rayong
21140, Thailand

Procedure used: Calibration were conducted using in-house calibration Procedure CP-E17 According to direct measurement method with Multi-Product Calibrator.

Condition of this result of calibration

1.Reference standards instruments :


Instrument	Model	Serial No.	Certificate No.	Due Date
1) Multi-Product Calibrator	5500A	6440007	21E1444	07 May 2022

2.This result of calibration was made on requested at the point specified by customer.

3.The certificate is valid only to the item calibrated on date and place of calibration.

4.This Certification is traceable to the International System of Unit maintained at:-

-National Institute of Metrology Thailand (NIMT)



Calibrated by: Pongsagorn Boonyapom
Issue Date: 22 March 2022

Approved Signatory :
[x] Phalinee Prabpaipai
[] Nuntawat Khamchai
[] Pornthippa Tarneyakul

B 0284414



Cert.No.: 22E986
Page.: 2 of 2

Result of calibration:- (*) Without adjustment () After adjustment

Function:	DC voltage measurement	Range:	2000 mV	
	Standard Value	UUC* Reading	Error	Uncertainty
	(mV)	(mV)	(mV)	(± μV)
	-200.0000	-200.0	0.0	72
	-150.0000	-150.0	0.0	69
	-100.0000	-100.0	0.0	65
	-50.0000	-50.0	0.0	62
	0.0000	0.0	0.0	58
	50.0000	50.0	0.0	62
	100.0000	100.0	0.0	65
	150.0000	150.0	0.0	69
	200.0000	200.0	0.0	72

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95 %

*UUC= Unit Under Calibration.

-000-

a 1101070




Certificate of Calibration

Certificate No.: C06210159 Page 2 of 3

Equipment: SPECTROPHOTOMETER
 Model: OR6000
 Serial No. (or ID.): 1627845 (RYG_EN0037)
 Manufacturer: HACH
 Condition: In Condition

Certificate No.: C06210159
 Issued Date: 01 April 2021
 Job No.: KSPR2104738
 Page: 1 of 3

Customer: ALS Laboratory Group (Thailand) Co., Ltd. (Rayong Branch)
 616/10 Moo 5 T.Maenam Khu,
 A.Pluakdaeng, Rayong 21140, Thailand.




Environment Condition: Temperature 25.1 °C ± 0.4 °C
 Humidity 48.8 %RH ± 3.7 %RH


Calibration Place: ALS Laboratory Group (Thailand) Co., Ltd. (Rayong Branch) (Wet Chemistry Lab)
 616/10 Moo 5 T.Maenam Khu,
 A.Pluakdaeng, Rayong 21140, Thailand.

Calibration By: Mr. Chattuphon Foithong
 Calibration Date: 01 April 2021
 The Method used: In house method, SPCC-WI-24, base on ASTM E 275-08 and ASTM E 387-04
 Traceability: This certificate is traceable to the CRM maintained by National Institute of Standards and Technology (NIST) through Starna Scientific Limited.

The standard for Wavelength Certificate No. 87146 and 87152
 The standard for Photometric Certificate No. 87220 and 87139
 The standard for Stray light Certificate No. 87163 and 87161
 The standard for Spectral resolution Certificate No. 87173


 (Mr. Chattuphon Foithong)
 Person in charge


 บริษัท เอสพีซี แอร์ที จำกัด
 SPC RT Co., Ltd.


 (Mr. Dumrong Boonsopon)
 Authorized signatory

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.
 The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).
 These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of SPC RT Co., Ltd.

บริษัท เอสพีซี แอร์ที จำกัด
 ๔ RT CO., LTD.
 รหัส 00003 1134 ซอยวิภาวดีรังสิต 37 แขวงจตุจักร เขตจตุจักร กรุงเทพมหานคร 12060
 รหัส 00003 1134 ซอยวิภาวดีรังสิต 37 แขวงจตุจักร เขตจตุจักร กรุงเทพมหานคร 12060 Thailand

บริษัท เอสพีซี แอร์ที จำกัด
 ๔ RT CO., LTD.
 รหัส 00003 1134 ซอยวิภาวดีรังสิต 37 แขวงจตุจักร เขตจตุจักร กรุงเทพมหานคร 12060
 รหัส 00003 1134 ซอยวิภาวดีรังสิต 37 แขวงจตุจักร เขตจตุจักร กรุงเทพมหานคร 12060 Thailand

Certificate No.: C06210159 Page 3 of 3

Calibration Results:
Without Adjustment

Photometric Accuracy (Absorbance)				
Wavelength	Standard absorbance	Unit Under Calibration	Correction	Uncertainty
235 nm	0.0000	0.000	0.0000	0.0080
	0.7307	0.730	0.0007	0.0080
257 nm	0.0000	0.000	0.0000	0.0080
	0.8516	0.850	0.0016	0.0080
313 nm	0.0000	0.000	0.0000	0.0080
	0.2836	0.285	-0.0014	0.0080
350 nm	0.0000	0.000	0.0000	0.0080
	0.6319	0.629	0.0029	0.0080
Stray light *				
Standard: cut-off	UUC: Wavelength (nm)	UUC: Transmission (%T)	Absorbance (A)	
260.57 +/- 0.11 nm	260.6	1.5	1.824	
392.03 +/- 0.11 nm	392.0	1.5	1.824	
The stray light transmission reference is less than 1.0 T(%) and absorbance is greater than 2.0 (A)				
Spectral Resolution *				
Nominal Concentration 0.02 % v/v	Peak	Trough	Ratio	SBW
Standard Wavelength (nm)	268.72	266.76	1.39	2.00
UUC: Wavelength (nm)	269.2	266.1		
Std Absorbance (A)	0.4616	0.2797		
Absorbance (A)	0.415	0.300		

* Calibration Marked * Not TISI Accredited * in this Certificate have been included for completeness.

The End of Certificate

Calibration Results:
Without Adjustment

Wavelength Accuracy (nm), The spectral bandwidth of Std at 2 nm and UUC at 2 nm				
Standard Wavelength	Unit Under Calibration	Correction	Uncertainty	
418.61	418.4	0.21	0.13	
536.66	536.7	-0.04	0.13	
637.98	638.3	-0.32	0.14	
748.48	748.7	-0.22	0.14	
807.03	807.4	-0.37	0.14	
Photometric Accuracy (Absorbance)				
Wavelength	Standard absorbance	Unit Under Calibration	Correction	Uncertainty
420 nm	0.0000	0.000	0.0000	0.0045
	0.5890	0.590	-0.0010	0.0045
	0.7616	0.762	-0.0004	0.0045
	1.0263	1.027	-0.0007	0.0045
440 nm	0.0000	0.000	0.0000	0.0045
	0.5787	0.579	-0.0003	0.0045
	0.7442	0.744	0.0002	0.0045
	1.0039	1.004	-0.0001	0.0045
465 nm	0.0000	0.000	0.0000	0.0045
	0.5292	0.530	-0.0008	0.0045
	0.6865	0.687	-0.0005	0.0045
	0.9534	0.954	-0.0006	0.0045
546.1 nm	0.0000	0.000	0.0000	0.0045
	0.5468	0.546	0.0008	0.0045
	0.6957	0.695	0.0007	0.0045
	0.9991	0.998	0.0011	0.0045
590 nm	0.0000	0.000	0.0000	0.0045
	0.5851	0.584	0.0011	0.0045
	0.7238	0.723	0.0008	0.0045
	1.0957	1.094	0.0017	0.0045
635 nm	0.0000	0.000	0.0000	0.0045
	0.5692	0.568	0.0012	0.0045
	0.6914	0.691	0.0004	0.0045
	1.0881	1.087	0.0011	0.0045

ใบตรวจสอบสภาพเครื่องวัดสิ่งแวดล้อม

เลขที่ใบงาน: KSPR2104738

ชนิดเครื่องมือ: SPECTROPHOTOMETER รุ่น: DR6000 หมายเลขเครื่อง: 1627845

ตรวจสอบ (วัน)		รายการตรวจเช็ค	ตรวจสอบ (สิ่ง)		หมายเหตุ
01 Apr 2021			01 Apr 2021		
ปกติ	ไม่ปกติ		ปกติ	ไม่ปกติ	
		General			
<input checked="" type="checkbox"/>	<input type="checkbox"/>	1. ความสมบูรณ์เครื่อง	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	2. ความสะอาด (ช่องใส่ตัวอย่าง, ภายใน-นอกเครื่อง)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	3. สวิทช์ ปิด – เปิด เครื่อง (On-Off Switch)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	4. ปุ่มกด (Keypad)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	5. หน้าจอ (Display, Screen Contrast)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
		Spectrophotometer			
<input type="checkbox"/>	<input type="checkbox"/>	6. แรงดันไฟฟ้า (Battery Backup) >= 2.5 VDC	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	7. หัวหมุนเลือกความยาวคลื่น (Wavelength Control)	<input type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	8. ความยาวคลื่น (Wavelength Check)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	656.1=656.1 nm
<input checked="" type="checkbox"/>	<input type="checkbox"/>	9. แหล่งกำเนิดแสง (UV < 3,000 hour)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	10. แหล่งกำเนิดแสง (Visible < 5,000 hour)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	11. ช่องวัดหลายตัวอย่าง (Carousel Module)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
		pH Meter and Conductivity Meter			
<input type="checkbox"/>	<input type="checkbox"/>	12. อิเล็กโทรด (Electrode and Connection Cable)	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	13. ระดับสารละลายใน Electrode (Level KCl)	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	14. ฝาปิดกันปลาย Electrode (Dust Protection Hood)	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	15. ขาจับอิเล็กโทรด (Stand)	<input type="checkbox"/>	<input type="checkbox"/>	
		Turbidimeter			
<input type="checkbox"/>	<input type="checkbox"/>	16. ค่าความขุ่นต่ำสุด (No Sample)	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	17. ระดับการส่องสว่างของแสง (>= 2.5 ไม่น้อย 3.0)	<input type="checkbox"/>	<input type="checkbox"/>	
		Automatic titrator			
<input type="checkbox"/>	<input type="checkbox"/>	18. สภาพ Piston Burettes	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	19. Function Rinsing and Dosing	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	20. ระบบท่อสายยางและอุปกรณ์ประกอบ	<input type="checkbox"/>	<input type="checkbox"/>	

เห็นด้วยและนำเข้า

Mr. Chattuphon Foithong
Service Engineer

บริษัท เอสพีซี แอร์ที จำกัด
 ๔ RT CO., LTD.
 รหัส 00003 1134 ซอยวิภาวดีรังสิต 37 แขวงจตุจักร เขตจตุจักร กรุงเทพมหานคร 12060
 รหัส 00003 1134 ซอยวิภาวดีรังสิต 37 แขวงจตุจักร เขตจตุจักร กรุงเทพมหานคร 12060 Thailand

บริษัท เอสพีซี แอร์ที จำกัด
 ๔ RT CO., LTD.
 รหัส 00003 1134 ซอยวิภาวดีรังสิต 37 แขวงจตุจักร เขตจตุจักร กรุงเทพมหานคร 12060
 รหัส 00003 1134 ซอยวิภาวดีรังสิต 37 แขวงจตุจักร เขตจตุจักร กรุงเทพมหานคร 12060 Thailand



Certificate of Calibration

Equipment: SPECTROPHOTOMETER
Model: DR6000
Serial No. (or ID.): 1627845 (RYG_EN0037)
Manufacturer: HACH
Condition: In Condition

Certificate No.: C06220464
Issued Date: 27 September 2022
Job No.: KSPR2212224
Page: 1 of 3

Customer: ALS Laboratory Group (Thailand) Co., Ltd. (Rayong Branch)
616/10 Moo 5 T.Maenam Khu,
A.Pluakdaeng, Rayong 21140, Thailand.

Environment Condition: Temperature 23.1 °C ±
Humidity 65.4 %RH ±

Calibration Place: ALS Laboratory Group (Thailand) Co., Ltd. (Rayong Branch) (Wet Chemistry)
616/10 Moo 5 T.Maenam Khu,
A.Pluakdaeng, Rayong 21140, Thailand.

Calibration By: Mr. Chattaphon Foithong
Calibration Date: 27 September 2022
The Method used: In house method, CAL-WI-24, base on ASTM E 275-06 and ASTM E 387-04
Traceability: This certificate is traceable to the CRM maintained by National Institute of Standards and Technology (NIST) through Stama Scientific Limited.

The standard for Wavelength Certificate No. 91418 and 91435
The standard for Photometric Certificate No. 91441 and 101038
The standard for Stray light Certificate No. 101041 and 101040
The standard for Spectral resolution Certificate No. 101037

(Mr. Chattaphon Foithong)
Person in charge

(Mr. Thaisongkiet Pongngam)
Authorized signatory

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.
The measurement uncertainty stated in the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).
These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of DKSH Technology Limited.

DKSH Technology Limited
2523 หมู่ 5 ตำบลบ้านนา อำเภอเมือง ราชบุรี 76100
2523 Sukhumvit Road, Bangkok, Prachinburi, Bangkok 10200
Phone: +66 2528 7020 Email: info.calibration@dksh.com Website: www.dksh.com/calibration-thailand

Delivering Growth - In Asia and Beyond.

CALFM-C06-13: 20 Jul 2022

Calibration Results: Without Adjustment

Wavelength Accuracy (nm), The spectral bandwidth of Std at 2 nm and UUC at 2 nm				
Standard Wavelength	Unit Under Calibration	Correction	Uncertainty	
418.61	418.4	0.21	0.14	
536.66	536.7	-0.04	0.14	
637.99	638.3	-0.32	0.14	
748.48	748.6	-0.32	0.14	
807.03	807.4	-0.37	0.13	

Photometric Accuracy (Absorbance)				
Wavelength	Standard absorbance	Unit Under Calibration	Correction	Uncertainty
420 nm	0.0000	0.000	0.0000	0.0045
	0.5805	0.563	-0.0025	0.0045
	0.7334	0.737	-0.0036	0.0045
	1.0534	1.067	-0.0036	0.0045
440 nm	0.0000	0.000	0.0000	0.0045
	0.5503	0.553	-0.0027	0.0045
	0.7179	0.720	-0.0021	0.0045
	1.0312	1.034	-0.0028	0.0045
465 nm	0.0000	0.000	0.0000	0.0045
	0.5024	0.506	-0.0036	0.0045
	0.6693	0.672	-0.0027	0.0045
	0.9504	0.964	-0.0035	0.0045
546.1 nm	0.0000	0.000	0.0000	0.0045
	0.5168	0.519	-0.0022	0.0045
	0.6903	0.691	-0.0007	0.0045
	0.9904	0.992	-0.0016	0.0045
590 nm	0.0000	0.000	0.0000	0.0045
	0.5525	0.554	-0.0015	0.0045
	0.7175	0.718	-0.0005	0.0045
	1.0301	1.031	-0.0009	0.0045
635 nm	0.0000	0.000	0.0000	0.0045
	0.5367	0.538	-0.0013	0.0045
	0.6947	0.695	-0.0003	0.0045
	0.9823	0.983	-0.0007	0.0045

บริษัท ดีเคเอส เอช จำกัด
DKSH Technology Limited
2523 หมู่ 5 ตำบลบ้านนา อำเภอเมือง ราชบุรี 76100
2523 Sukhumvit Road, Bangkok, Prachinburi, Bangkok 10200
Phone: +66 2528 7020 Email: info.calibration@dksh.com Website: www.dksh.com/calibration-thailand

Delivering Growth - In Asia and Beyond.

CALFM-C06-13: 20 Jul 2022



ใบตรวจสอบสภาพเครื่องวัดสิ่งแวดล้อม

Certificate No.: C06220464 Page 3 of 3

เลขที่ใบงาน: KSPR2212224

Calibration Results: Without Adjustment

Photometric Accuracy (Absorbance)				
Wavelength	Standard absorbance	Unit Under Calibration	Correction	Uncertainty
235 nm	0.0000	0.000	0.0000	0.0000
	0.7423	0.744	-0.0017	0.0033
257 nm	0.0000	0.000	0.0000	0.0080
	0.8609	0.861	-0.0001	0.0084
313 nm	0.0000	0.000	0.0000	0.0060
	0.2655	0.262	-0.0025	0.0080
350 nm	0.0000	0.000	0.0000	0.0050
	0.6361	0.636	0.0001	0.0080
Stray light *				
Standard: cut-off	UUC: Wavelength (nm)	UUC: Transmission (%)	Absorbance (A)	
260.67 +/- 0.11 nm	260.7	2.1	1.670	
391.94 +/- 0.11 nm	391.9	1.7	1.770	
Spectral Resolution *				
Nominal Concentration 0.02 % v/v	Peak	Trough	Ratio	SSW
Standard Wavelength (nm)	268.60	266.63	1.39	2.00
UUC: Wavelength (nm)	268.2	266.1		
Std Absorbance (A)	0.4810	0.3176		
Absorbance (A)	0.373	0.268		

* Calibration Marked * Not TISI Accredited * In this Certificate have been included for completeness.

The End of Certificate

ชนิดเครื่องมือ: SPECTROPHOTOMETER		รุ่น: DR6000	หมายเลขเครื่อง: 1627845		
ตรวจสอบ (วัน)		รายการตรวจเช็ค	ตรวจสอบ (ส่ง)		หมายเหตุ
27 Sep 2022			27 Sep 2022		
ปกติ	ไม่ปกติ		ปกติ	ไม่ปกติ	
		General			
<input checked="" type="checkbox"/>	<input type="checkbox"/>	1. ความสมบูรณ์เครื่อง	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	2. ความสะอาด (ช่องใส่ตัวอย่าง, ภายใน-นอกเครื่อง)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	3. สวิทช์ เปิด – ปิด เครื่อง (On-Off Switch)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	4. ปุ่มกด (Keypad)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	5. หน้าจอ (Display, Screen Contrast)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
		Spectrophotometer			
<input type="checkbox"/>	<input type="checkbox"/>	6. แบตเตอรี่สำรอง (Battery Backup) >= 2.5 VDC	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	7. ตัวหมุนเลือกความยาวคลื่น (Wavelength Control)	<input type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	8. ความยาวคลื่น (Wavelength Check)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	655.1 ถึง 656.1 nm
<input checked="" type="checkbox"/>	<input type="checkbox"/>	9. แหล่งกำเนิดแสง (UV < 3,000 hour)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	10. แหล่งกำเนิดแสง (Visible < 5,000 hour)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	11. ช่องวัดหลายตัวอย่าง (Carousel Module)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
		pH Meter and Conductivity Meter			
<input type="checkbox"/>	<input type="checkbox"/>	12. อิเล็กโทรด (Electrode and Connection Cable)	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	13. ระดับสารละลายใน Electrode (Level KCl)	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	14. ฝาปิดกันฝุ่น Electrode (Dust Protection Hood)	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	15. ขาตั้งอิเล็กโทรด (Stand)	<input type="checkbox"/>	<input type="checkbox"/>	
		Turbidimeter			
<input type="checkbox"/>	<input type="checkbox"/>	16. ค่าความขุ่นที่ต่ำสุด (No Sample)	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	17. ระดับการส่องสว่างของแสง (>= 2.5 ไมล์ 3.0)	<input type="checkbox"/>	<input type="checkbox"/>	
		Automatic titrator			
<input type="checkbox"/>	<input type="checkbox"/>	18. สภาท Piston Burettes	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	19. Function Rinsing and Dosing	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	20. ระบบท่อสายยางและอุปกรณ์ประกอบ	<input type="checkbox"/>	<input type="checkbox"/>	

เซ็นเซอร์และค่า:

Mr. Chattaphon Foithong
Service Engineer

บริษัท ดีเคเอส เอช จำกัด
DKSH Technology Limited
2523 หมู่ 5 ตำบลบ้านนา อำเภอเมือง ราชบุรี 76100
2523 Sukhumvit Road, Bangkok, Prachinburi, Bangkok 10200
Phone: +66 2528 7020 Email: info.calibration@dksh.com Website: www.dksh.com/calibration-thailand

Delivering Growth - In Asia and Beyond.

CALFM-C06-13: 20 Jul 2022

บริษัท ดีเคเอส เอช จำกัด
DKSH Technology Limited
2523 หมู่ 5 ตำบลบ้านนา อำเภอเมือง ราชบุรี 76100
2523 Sukhumvit Road, Bangkok, Prachinburi, Bangkok 10200
Phone: +66 2528 7020 Email: info.calibration@dksh.com Website: www.dksh.com/calibration-thailand

Delivering Growth - In Asia and Beyond.

CALFM-C06-13: 20 Jul 2022



Metrological Center

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhoei, Saraburi 18110, Thailand.

Saraburi Tel : +66 3627 3096 Fax : +66 3627 3100

Bangkok Tel : +668 9205 6851 , +669 8247 2360

Website : www.scieco.co.th

E-Mail : calibrate@scg.co.th

RYG_EN0184



Metrological Center

SCI ECO Services Company Limited


33/2 Moo 3, T.Banpa, A.Kaengkhoei, Saraburi 18110, Thailand.



Certificate No. T220384101 "Substitute for Calibration Certificate Number T220384" Page 1 of 4

Certificate of Calibration

Equipment : Chamber (Cold Room)
 Manufacturer : MODULAR
 Model : IREVCOHCOO
 Serial No. : C00351459
 Customer Code : RYG_EN0184
 ID No. : T1939A5
 Customer : ALS Laboratory Group (Thailand) Co.,Ltd. (Rayong Branch)
 616/10 Moo 5 T.Maenam Khu,
 A.Pluakdaeng, Rayong 21140
 Customer Location : Laboratory
 Date of Receipt : 18 February 2022
 Calibrated By : Boonchai Suriyawong (Site Calibration Manager)
 Approved By : / Sujjar Nakhakred (Site Calibration Manager)
 Date of Issue : 18 MAR 2022



The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standard laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Metrological Center.

FM-L14 117-01-02-64

FM-L15 117-15-05-63



Metrological Center

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhoei, Saraburi 18110, Thailand.



Metrological Center

SCI ECO Services Company Limited


33/2 Moo 3, T.Banpa, A.Kaengkhoei, Saraburi 18110, Thailand.



Certificate No. T220384101

Page 3 of 4

Calibration Report



C = Centre, F = Centre of Face, A = Corner, E = Centre of Edge

1C = TN141	12F = TN152
2A = TN142	13A = TN153
3A = TN143	14A = TN154
4F = TN144	15F = TN155
5A = TN145	16E = TN156
6A = TN146	
7F = TN147	
8F = TN148	
9A = TN149	
10A = TN150	
11F = TN151	

Approved By:

FM-L14 117-01-02-64

Certificate No. T220384101

Page 2 of 4

Calibration Report

Equipment : Chamber (Cold Room)
 Date of Calibration : 22 February 2022
 Environment : Temperature : 23.2-24.3 °C
 Line Voltage : 221.8-227.2 V
 Relative Humidity : 55 - 65 %RH

Condition of this results of calibration :

1. This equipment was calibrated by insert 16 standard thermocouples type T into its chamber, the other one standard thermocouples type T use for ambient temperature measurement. The calibration was done in according to WI-T20 (based on ASTM E145-94 (Reapproved 2001) and AS2853-1986).
 All data show below were final values and the initial data from customer request. The temperature scale used was based on ITS - 90 .

2. Reference Standard Instrument :

Instrument	Model	Instrument No.	Certificate No.	Due Date
TC	TYPE T	TN141-TN150	T210743	21 April 2022
TC	TYPE T	TN151-TN160	T210743	21 April 2022
DATA LOGGER	34970A	TI50	T210743	21 April 2022

3. This certificate is traceable to :

National Institute of Metrology (Thailand) through Metrological Center (NSC-TISI-TIS 17025 CALIBRATION 0244)

4. Condition of calibrated item : good

Equipment Description :

Time Constant : - Hour 40 Minute At 3 °C
 Fresh Air Damper : ☐ Open ☐ Min ☐ Medium ☐ Max
☐ Close
☒ Not Available

5. Adjustment :

(X) without adjustment () after adjustment

Approved By:

Certificate No. T220384101

Page 4 of 4

Calibration Report

Measurement Results

Calibration Point	Average Standard Reading at each position (°C)									
	TN141	TN142	TN143	TN144	TN145	TN146	TN147	TN148	TN149	TN150
3.0	2.80	2.96	2.98	2.97	3.16	3.29	2.95	3.14	3.10	3.45
	TN151	TN152	TN153	TN154	TN155	TN156				
	3.04	3.19	3.03	3.34	3.21	3.11				

Chamber (Cold Room)			Temperature Distribution				
Setting (°C)	Reading (°C)		Average (°C)	Stability (± °C)	Uniformity (°C)	Uncertainty (± °C)	Coverage Factor k
	Min , Max	Average					
3.0	2.7 - 4.1	3.5	3.11	1.30	1.30	2.00	2.05

* The Accuted uncertainty exclude "uniformity"

The calibration result apply only the above calibrated item.

The result of test was found accurate as shown on date and place of test only.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k which for a t-distribution, providing a level of confidence of approximately 95 % .

Approved By:


FM-L15 117-15-05-63



Cert.No.: 22TW34
Page.: 1 of 2

Certificate of Testing

Equipment : DO Meter
Manufacturer : YSI
Model : 5000-115V
Serial No. : 15E102796
ID No. : RYG_EN0032
Received Date : 11 February 2022
Test Date : 14 February 2022
Reference : 2202-0404DSC-4
Submitted by : ALS Laboratory Group (Thailand) Co.,Ltd.
(Rayong Branch)
616/10 Moo 5 T.Maenam Khu, A.Pluakdaeng,
Rayong 21140, Thailand
Laboratory Condition : Temperature (25 ± 5) °C
Humidity (50 ± 20) %
Test Procedure : In - house method : CP-CH9
by Comparison Technique with Azide Modification Method
Tested by : Walalak Sirithean
Approved by : Saithip
Approved Signatory
() Malee Bulkruea
(✓) Saithip Meangmai
() Warakorn Lengagtrakul
Issue Date : 18 February 2022



B 0281285



Cert.No.: 22TW34
Page.: 2 of 2

Result : Dissolved Oxygen Meter Adjustment With Air 100 %
Dissolved Oxygen Probe No.: 15E100464

Titration Method (Azide Modification Method) (mg/L)	DO Meter Reading (mg/L)	Standard Deviation (mg/L)
8.02	8.02	0.0084

This report was certified only for the instrument we tested. It is allowable to use for study the system efficiency. The environmental impact control and present to organization it may concerned intend to use for advertising and referral purpose is prohibited. This report may not be reproduced other in full, without written approval of the laboratory

-00-

Saithip

a 1094744



Cert. No.: 22LM12
Page.: 1 of 2

Certificate of Calibration

Equipment : DO Meter with Sensor
Manufacturer : YSI
Model : 5000-115V
Serial No. : 15E102796
ID No. : RYG_EN0032
Submitted by : ALS Laboratory Group (Thailand) Co.,Ltd. (Rayong Branch)
616/10 Moo 5 T.Maenam Khu, A.Pluakdaeng,
Rayong 21140, Thailand
Location : TPA On Site Calibration Laboratory
Received Order : 11 February 2022
Calibrated Date : 21 February 2022
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %
AC Line Voltage : (220 ± 22) V
Calibrated by : Kunchit Promprat
Approved by : Mu
Approved Signatory
() Pornthippa Tameyakul
(✓) Malee Bulkruea
() Suwit Injai
Issue Date : 21 February 2022



Equipment : DO Meter with Sensor
Condition As-Received : Used Item
Reference : 2202-0404DSC-5
Procedure Used : Calibration were conducted using in-house calibration procedure CP-OT01 according to comparison with Industrial Platinum Resistance Thermometer (IPRT) into Temperature Bath.
The temperature scale used was based on ITS-90.

Cert. No.: 22LM12
Page.: 2 of 2

Condition of this result of calibration

1. Reference standard instrument:-
- | Instrument | Model | Serial No. | Cert. No. | Due Date |
|------------------------|-------|------------|-----------|-------------|
| 1) Digital Thermometer | 1523 | 2188080 | 2111273 | 22 Nov 2022 |
2. This certificate is valid only to the item calibrated on date and place of calibration.
3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment
Function : Temperature measurement.

This instrument was connected with temperature sensor, S/N.: 15E100464

Calibration Point (°C)	Immersion Depth (mm)	Standard Temperature (°C)	UUC* Reading (°C)	Error (°C)	Uncertainty (± °C)	Coverage Factor k
20.00	45	20.001	19.88	-0.121	0.15	2.00

UUC* : Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-00-

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services

A 0030008

a 1095714



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-27 FAX. 0-2719-9464



Certificate of Calibration

Cert. No.: 22TM317
Page: 1 of 3

Equipment : Low Temp. Incubator
Manufacturer : Memmert
Model : IPP750
Serial No. : V818.0084
ID No. : RYG_EN0154
Submitted by : ALS Laboratory Group (Thailand) Co., Ltd.
(Rayong Branch)
616/10 Moo 5 T. Maenam Khu,
A. Pluakdaeng, Rayong 21140, Thailand
Location : BOD Room
Received Order : 22 April 2022
Calibration Date : 22 April 2022
Ambient Temperature : $(26 \pm 10) ^\circ\text{C}$
Relative Humidity : $(50 \pm 30) \%$
Calibrated by : Man Pattanapongpaiboon

Approved by :
Approved Signatory

() Ponthippa Tameyakul
() Maleo Bulkruea
() Suwit Imjai

Issue Date : 3 May 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced either in full or in part without the prior written approval of the head of Corporate Services 3: Equipment Calibration and Testing Services

A 0040735



Equipment : Low Temp. Incubator
Condition As-Received : Used Item
Reference : 2204-0146OC-1
Procedure Used :-

Cert. No.: 22TM317
Page: 2 of 3

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement
The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34970A	MY44031769	21LM12	02 Sep 2022

2. This certificate is valid only to the item calibrated on date and place of calibration.


3. This certification is traceable to the International System of Unit.

Result of Calibration :- () Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Close

Environment during calibration		
	Beginning	Finished
Temp. (°C)	25	25
REL.Humid. (%)	54	58
AC Supply (Volt)	221	223



Probe Installation Details :

a = 10 cm
b = 10 cm
c = 10 cm

Dimension of Chamber :

D = 0.60 m
W = 1.0 m
H = 1.2 m
Capacity = 0.75 m³

Position :	Ref. Std. ID No.:
1	9RTD-2/1
2	9RTD-2/2
3	9RTD-2/3
4	9RTD-2/4
5	9RTD-2/5
6	9RTD-2/6
7	9RTD-2/7
8	9RTD-2/8
9 (ref.)	9RTD-2/9

a 1106485

RYG_EN0002



PENTA
CALIBRATION

PENTA CALIBRATION CO., LTD.
66/124 The Connect 33 Village Kanchanaphisek Road
Dokmai Praweit Bangkok 10250
Tel: +66 (0) 2059-9773
www.pentacal.co.th

Certificate of Calibration

Represent to Certificate of Calibration : PTC/07/22103

Certificate No.:	PTC/07/22103	Page:	1 of 2
Equipment:	Digital Balance	Condition:	Normal
Manufacturer:	Sartorius	Serial No.	26207038
Model:	MSE2245-100-DU	ID No.	RYG_EN0002
Type of Balance:	Single interval		

Customer: ALS Laboratory Group (Thailand) Co., Ltd.
616/10 Moo 5 T. Maenamkoo, A. Pluakdaeng,
Rayong 21140, Thailand

Environment Condition: Temperature 23.9 °C ± 0.3 °C
Humidity 58.1 %RH ± 4.4 %RH
Air density 1.17 kg/m³

Calibration Place: ALS Laboratory Group (Thailand) Co., Ltd.
616/10 Moo 5 T. Maenamkoo, A. Pluakdaeng,
Rayong 21140, Thailand

The Method used: In house method. PTC-WI-07. base on Euramet cg 18

Traceability: This certificate is traceable to the SI Units through Thai Calibration Service Co., Ltd
, NSG-ONS Accreditation No. : Calibration 0189

Date Received: March 23, 2022

Calibration Date: March 23, 2022

Issued Date: March 25, 2022

Calibration By: Mr. Rungroj Metakul



REVIEW BY
APPROVED BY
NEXT CAL. DATE 23/03/25

Equipment : Low Temp. Incubator
Condition As-Received : Used Item
Reference : 2204-0146OC-1
Result of Calibration :- () Without Adjustment
Function of UUC* : Temperature Source
Fresh air setting : Close

Cert. No.: 22TM317
Page: 3 of 3

Calibration Point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Temperature stability (± °C)	Temperature uniformity (°C)	Overall Variation (°C)	Uncertainty (± °C)	Coverage Factor k
20.0	20.0	20.0	0.022	0.20	0.22	0.30	2

Calibration Point (°C)	Measured Temperature (°C)								
	Position								
20.0	1	2	3	4	5	6	7	8	9 (ref.)
	20.209	20.174	20.199	20.110	20.075	20.062	20.027	20.069	20.030

Average* : The average of 30 values in each position.

Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor.

Temperature uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-o0o-

(Mr. Kiangsak Kalasi)
Reviewed by

Approved By :
(Mr. Keattisak Kerdtio)
Laboratory Manager

This certificate is issued in triplicate of measurement according to the International System of Units (SI). It provides traceability of measurement to international standard or other recognized national standard laboratories.

The measurement uncertainty is stated as the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to the Expression of Uncertainty in Measurement (GUM). This effect that the results relate only to the items calibrated.

This certificate and its result may not be reproduced except in full only, without written approval from penta calibration co., Ltd



PENTA CALIBRATION CO., LTD.
66/124 The Connect 33 Village Kanchanaphisek Road
Dokmai Praweit Bangkok 10250
Tel: +66 (0) 2059-9773
www.pentacal.com

Represent to Certificate of Calibration ,PTC/07/22103

Certificate No.: PTC/07/22103


Page: 2 of 2

Measurement Results:

Without Adjustment :

Function Calibration: Non Adjustment

Eccentric Error: Weight to be 1/3, 1/2 or of Maximum capacity



Eccentricity test 100 (g)				
Position (g)				
1	2	3	4	5
0.0000	0.0000	-0.0002	0.0002	0.0002
Maximum deviation:			0.0002	

Repeatability Test : Weight to be $1/2 \leq L \leq$ Maximum capacity

Determination of the standard deviation of weighing balance., Readability 0.0001 (g)

Nominal test value (g)	Standard Deviation
200	0.00003

Error of indication : from nominal value., Readability 0.0001 (g)

Nominal Value (g)	Conventional Mass (g)	Indication (g)	Correction of Balance (g)	Uncertainty (g)	k
0	0.00000	0.0000	0.0000	0.000086	2.16
0.01	0.01000	0.0100	0.0000	0.00010	2.06
0.1	0.10000	0.1000	0.0000	0.00010	2.06
1	1.00000	1.0000	0.0000	0.00010	2.06
2	2.00000	1.9999	0.0001	0.00010	2.06
5	5.00001	5.0000	0.0000	0.00010	2.06
10	10.00000	10.0000	0.0000	0.00010	2.06
20	20.00003	19.9999	0.0001	0.00011	2.05
50	50.00004	49.9999	0.0001	0.00012	2.00
100	100.00004	100.0001	-0.0001	0.00017	2.00
200	200.00011	200.0000	0.0001	0.00027	2.00

Note: Weight of adjust (g)

The End of Certificate



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES & EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL: 0-2717-3000-27 FAX: 0-2719-9484



Cert. No.: 21TM827
Page: 1 of 3

Certificate of Calibration

Equipment : Hot Air Oven

Manufacturer : Memmert

Model : UFE 500

Serial No. : G511.1572

ID No. : RYG_EN0010

Submitted by : ALS Laboratory Group (Thailand) Co.,Ltd. (Rayong Branch)
616/10 Moo 5 T. Maenam Khu,
A. Pluakdaeng,
Rayong 21140 Thailand

Location : Oven Room

Received Order : 5 May 2021

Calibration Date : 5 May 2021

Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %

Calibrated by : Khit Rutianaprapachai

Approved by :
Approved Signatory

() Ponthipha Tameyakul
(/) Malee Butkruea
() Suwit Imjai

Issue Date : 14 May 2021

The Uncertainties are for a confidence probability of approximately 95%

This certificate must not be reproduced or used in full except with the prior written
Approval of the head of Corporate Services & Equipment Calibration and Testing Services.

A 0028099



Equipment : Hot Air Oven

Cert. No.: 21TM827

Condition As-Received : Used Item

Page: 2 of 3

Reference : 2105-0005OC-4

Procedure Used :-

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD) and Thermocouple Type T.

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34972A	MY57013823	21LM3	26 Feb 2022


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Close



Probe Installation Details : Dimension of Chamber :
a = 5.0 cm D = 0.40 m
b = 5.0 cm W = 0.56 m
c = 5.0 cm H = 0.48 m
Capacity = 0.11 m³

Environment during calibration		
	Beginning	Finished
Temp. (°C)	28	29
REL.Humid. (%)	59	56
AC Supply (Volt)	220	221

Ref. Std. ID No.: @ Calibration Point		
Position :	(104) °C	(180) °C
1	21-17RTD-01	19-17TC-01
2	21-17RTD-02	19-17TC-02
3	17RTD-03	19-17TC-03
4	17RTD-04	19-17TC-04
5	17RTD-05	19-17TC-05
6	17RTD-06	19-17TC-06
7	17RTD-07	19-17TC-07
8	17RTD-08	19-17TC-08
9 (ref.)	17RTD-09	19-17TC-09



Equipment : Hot Air Oven

Cert. No.: 21TM827

Condition As-Received : Used Item

Page: 3 of 3

Reference : 2105-0005OC-4

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Calibration Point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Temperature stability (± °C)	Temperature uniformity (°C)	Overall Variation (°C)	Uncertainty (± °C)	Coverage Factor k
104.0	104.0	104.0	0.063	0.54	0.70	0.42	2
180.0	180.0	180.0	0.15	0.89	1.3	1.1	2

Calibration Point (°C)	Measured Temperature (°C)								
	Position								
	1	2	3	4	5	6	7	8	9 (ref.)
180.0	180.101	180.481	179.401	179.692	179.980	179.943	180.127	179.915	179.709

Average* : The average of 30 values in each position.

Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor.

Temperature uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as steady an observation time as possible to determine the temperature pattern or homogeneity within the chamber under close-station conditions.

Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-00-



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES & EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3009-27 FAX. 0-2719-9184



Cert. No.: 22TM1517
Page: 1 of 3

Certificate of Calibration

Equipment : Hot Air Oven
Manufacturer : Memmert
Model : UFE 500
Serial No. : G511.1572
ID No. : RYG_EN0010
Submitted by : ALS Laboratory Group (Thailand) Co., Ltd. (Rayong Branch)
616/10 Moo 5 T. Maenam Khu, A. Pluakdaeng, Rayong 21140 Thailand
Location : Oven Room
Received Order : 20 October 2022
Calibration Date : 20 October 2022
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %
Calibrated by : Man Pattanapongpailoon

Approved by :
Approved Signatory

() Pornthippa Tameyakul
(x) Malee Bulkruea
() Suwit Imjai

Issue Date : 2 November 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services & Equipment Calibration and Testing Services.

A 0046908



Equipment : Hot Air Oven
Condition As-Received : Used Item
Reference : 2210-0376DC-2
Procedure Used :-

Cert. No.: 22TM1517
Page: 2 of 3

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTO) and Thermocouple Type T.

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34972A	MY49023932	22LM97	29 Jul 2023


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source
Fresh air setting : Close

Environment during calibration		
	Beginning	Finished
Temp. (°C)	25	25
REL.Humid. (%)	54	59
AC Supply (Volt)	223	225



Probe Installation Details : Dimension of Chamber :
a = 5.0 cm D = 0.40 m
b = 5.0 cm W = 0.56 m
c = 5.0 cm H = 0.48 m
Capacity = 0.11 m³

Ref. Std. ID No.: @ Calibration Point		
Position :	(180) °C	(104) °C
1	21-16TC-01	20-16RTD-01
2	21-16TC-02	20-16RTD-02
3	21-16TC-03	20-16RTD-03
4	21-16TC-04	20-16RTD-04
5	21-16TC-05	22-16RTD-05
6	21-16TC-06	20-16RTD-06
7	21-16TC-07	20-16RTD-07
8	21-16TC-08	22-16RTD-08
9 (ref.)	21-16TC-09	22-16RTD-09

a 1132466

RYG_EN0029



Equipment : Hot Air Oven
Condition As-Received : Used Item
Reference : 2210-0376DC-2
Result of Calibration :- (*) Without Adjustment
Function of UUC* : Temperature Source
Fresh air setting : Close

Cert. No.: 22TM1517
Page: 3 of 3

Calibration Point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Temperature stability (± °C)	Temperature uniformity (°C)	Overall Variation (°C)	Uncertainty (± °C)	Coverage Factor k
104.0	104.0	104.0	0.076	0.52	0.60	0.42	2
180.0	180.0	180.0	0.13	0.88	1.2	1.1	2

Calibration Point (°C)	Measured Temperature (°C)								
	Position								
	1	2	3	4	5	6	7	8	9 (ref.)
104.0	103.768	103.734	103.723	103.800	104.215	104.131	104.132	103.740	103.747
180.0	179.723	179.359	179.439	179.489	180.361	180.114	180.131	180.243	179.605

Average* : The average of 30 values in each position.

Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor

Temperature uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-000-

a 1132465



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES & EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3009-27 FAX. 0-2719-9184



Cert.No.: 22CH253
Page: 1 of 2

Certificate of Calibration

Equipment : Conductivity Meter
Manufacturer : Mettler Toledo
Model : S230
Serial No. : B241407147
ID No. : RYG_EN0029
Condition As-Received : Used Item
Received Date : 22 February 2022
Calibration Date : 23 February 2022
Reference : 2202-0732DSC-1
Submitted by : ALS Laboratory Group (Thailand) Co., Ltd. (Rayong Branch)
616/10 Moo 5 T. Maenam Khu, A. Pluakdaeng, Rayong 21140, Thailand
Ambient Temperature : (25 ± 2.5) °C
Relative Humidity : (50 ± 15) %
Calibration Procedure : In -house method :
- CP-CH6 : based on direct measurement by using certified reference material (CRM)
Calibrated by : Walalak Sinthean
Approved by :
Approved Signatory
() Malee Bulkruea
() Saithip Meangmai
() Warakorn Lemgagrakul
Issue Date : 25 February 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services & Equipment Calibration and Testing Services.

A 0038145



Cert.No.: 22CH283

Page: 2 of 2

Condition of this result of calibration**1. Reference Standard Instrument :-**

Instrument	Serial No.	ID No.	Certificate No.	Due date
1) Thermometer	9549224	130RC003	211451	15 Apr 2022

This certification is traceable to the International System of Unit maintained at:-

- Traceable to National Institute of Metrology (Thailand), NIMT

2. Certified Reference Materials :-

- Conductivity calibration solution, CPA chem Ltd., The measurement results are traceable to SI through CPA chem Ltd., ANSI-ASQ National Accreditation Board, Accredited No. AR-1635

Conductivity Solution	Manufacturer	Lot No.	Exp. date
84.000 $\mu\text{S/cm}$	CPA Chem	754034	26 June 2022
1413.0 $\mu\text{S/cm}$	CPA Chem	766815	04 Sep 2022
12.880 mS/cm	CPA Chem	761022	02 Aug 2022

- Control Conductivity calibration solution temperature by Water bath (25 ± 0.1) $^{\circ}\text{C}$

3. This certificate is valid only to the item calibrated on date and place of calibration.

Calibration results**Function : Conductivity Measurement**(*) After Adjustment at 1413.0 $\mu\text{S/cm}$

Conductivity Electrode Serial No.: 5821441030

Standard Conductivity Solution	Before Adjustment UUC* Reading	After Adjustment UUC* Reading	Uncertainty of Measurement (\pm)	Coverage factor k
84.000 $\mu\text{S/cm}$	82.4 $\mu\text{S/cm}$	84.4 $\mu\text{S/cm}$	0.62 $\mu\text{S/cm}$	2.00
1413.0 $\mu\text{S/cm}$	1375 $\mu\text{S/cm}$	1413 $\mu\text{S/cm}$	9.2 $\mu\text{S/cm}$	2.00
12.880 mS/cm	12.54 mS/cm	12.81 mS/cm	0.086 mS/cm	2.00

Remark - UUC* = Unit Under Calibration- Cell constant = 0.555236 cm^{-1} The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

-o0o-

Mali

a 1090534



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES & EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert. No.: 21TM829

Page: 1 of 3

Certificate of Calibration

Equipment : Hot Air Oven

Manufacturer : Mommert

Model : UM 400

Serial No. : b495.0899

ID No. : RYG_EN0006

Submitted by : ALS Laboratory Group (Thailand) Co.,Ltd. (Rayong Branch)
616/10 Moo 5 T. Maenam Khu,
A. Pluakdaeng,
Rayong 21140 Thailand

Location : Oven Room

Received Order : 5 May 2021

Calibration Date : 5 - 6 May 2021

Ambient Temperature : (26 ± 10) $^{\circ}\text{C}$ Relative Humidity : (50 ± 30) %

Calibrated by : Khit Ruitanaprapachai

Approved by : Mali
Approved Signatory

() Pornthippa Tameyakul
() Malee Bulkruea
() Suwit Imjai

Issue Date : 14 May 2021

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the Head of Corporate Services & Equipment Calibration and Testing Services

A 0028096



Equipment : Hot Air Oven

Cert. No.: 21TM829

Condition As-Received : Used Item

Page: 2 of 3

Reference : 2105-0005OC-1

Procedure Used :-

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD).

The temperature scale used was based on ITS-90.

Condition of this result of calibration**1. Reference standard instrument:-**

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34872A	MY57013823	211M3	26 Feb 2022


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Close



Probe Installation Details : Dimension of Chamber :
a = 5.0 cm D = 0.33 m
b = 5.0 cm W = 0.40 m
c = 5.0 cm H = 0.40 m
Capacity = 0.053 m^3

Environment during calibration		
	Beginning	Finished
Temp. ($^{\circ}\text{C}$)	29	30
REL.Humid. (%)	56	58
AC Supply (Volt)	221	222

Position :	Ref. Std. ID No.:
1	21-17RTD-01
2	21-17RTD-02
3	17RTD-03
4	17RTD-04
5	17RTD-05
6	17RTD-06
7	17RTD-07
8	17RTD-08
9 (ref.)	17RTD-09

Mali

a 1054310



Equipment : Hot Air Oven

Cert. No.: 21TM829

Condition As-Received : Used Item

Page: 3 of 3

Reference : 2105-0005OC-1

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Calibration Point ($^{\circ}\text{C}$)	UUC* Setting ($^{\circ}\text{C}$)	UUC* Reading ($^{\circ}\text{C}$)	Temperature stability (\pm $^{\circ}\text{C}$)	Temperature uniformity ($^{\circ}\text{C}$)	Overall Variation ($^{\circ}\text{C}$)	Uncertainty (\pm $^{\circ}\text{C}$)	Coverage Factor k
70.0	70.0	70.0	0.21	1.8	2.0	0.55	2

Calibration Point ($^{\circ}\text{C}$)	Measured Temperature ($^{\circ}\text{C}$)								
	Position								
70.0	1	2	3	4	5	6	7	8	9 (ref.)
	70.404	70.277	70.607	70.307	68.789	69.257	68.846	69.331	70.495

Average* : The average of 30 values in each position.**Temperature stability :** One-half of the greatest maximum difference of measured temperature at any one sensor.**Temperature uniformity :** The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.**Overall Variation :** The Difference of the maximum and minimum measured temperatures throughout observation.**UUC* :** Unit Under Calibration**Note :** The reported uncertainty of measurement was included stability and excluded uniformity.The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

-o0o-

Mali

a 1054309



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
5344 PATTANAKARN ROAD 501 B, SUANLUANG, SUANLUANG BANGKOK 10250
TEL: 0-2717-3008-27 FAX: 0-2719-9184



Cert. No.: 22TM1492
Page: 1 of 3

Certificate of Calibration

Equipment : Hot Air Oven
Manufacturer : Memmert
Model : UM 400
Serial No. : 6495.0899
ID No. : RYG_EN0006
Submitted by : ALS Laboratory Group (Thailand) Co.,Ltd. (Rayong Branch)
616/10 Moo 5, T. Maenam Khu,
A. Pluakdaeng,
Rayong 21140, Thailand
Location : Oven Room
Received Order : 20 October 2022
Calibration Date : 20 October 2022
Ambient Temperature : $(26 \pm 10) ^\circ\text{C}$
Relative Humidity : $(50 \pm 30) \%$
Calibrated by : Preecha Hlathib
Approved by :
() Pornhippa Tameyakul
() Malee Butkruea
() Suwit Imjai
Issue Date : 2 November 2022

REVIEW BY
APPROVED BY
NEXT CAL DATE 30/04/24

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3: Equipment Calibration and Testing Services

A 0046905



Equipment : Hot Air Oven
Condition As-Received : Used Item
Reference : 2210-0376OC-1
Procedure Used :-

Cert. No.: 22TM1492
Page: 2 of 3

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD).

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34970A	MY44035217	21LM30	23 Dec 2022


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Result of Calibration :- () Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Close



Probe Installation Details :

a = 5.0 cm
b = 5.0 cm
c = 5.0 cm

Dimension of Chamber :

D = 0.33 m
W = 0.40 m
H = 0.40 m
Capacity = 0.053 m³

Environment during calibration		
	Beginning	Finished
Temp. (°C)	28	29
REL.Humid. (%)	43	47
AC Supply (Volt)	220	221

Position :	Ref. Std. ID No.:
1	18-10RTD-01
2	18-10RTD-02
3	18-10RTD-03
4	18-10RTD-04
5	18-10RTD-05
6	18-10RTD-06
7	18-10RTD-07
8	18-10RTD-08
9 (ref.)	18-10RTD-09

a 1132473

RYG_EN0061



Equipment : Hot Air Oven
Condition As-Received : Used Item
Reference : 2210-0376OC-1
Result of Calibration :- () Without Adjustment
Function of UUC* : Temperature Source
Fresh air setting : Close

Cert. No.: 22TM1492
Page: 3 of 3

Calibration Point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Temperature stability (± °C)	Temperature uniformity (°C)	Overall Variation (°C)	Uncertainty (± °C)	Coverage Factor k
70.0	70.0	70.0	0.079	0.47	0.77	0.42	2

Calibration Point (°C)	Measured Temperature (°C)								
	1	2	3	4	5	6	7	8	9 (ref.)
70.0	70.262	69.995	70.079	70.177	70.654	70.039	70.688	70.149	70.328

Average* : The average of 30 values in each position.

Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-000-

a 1132472



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
5344 PATTANAKARN ROAD 501 B, SUANLUANG, SUANLUANG BANGKOK 10250
TEL: 0-2717-3008-27 FAX: 0-2719-9184



Cert. No.: 21TM673
Page: 1 of 3

Certificate of Calibration

Equipment : Water Bath
Manufacturer : Memmert
Model : WNB22
Serial No. : L513.0648
ID No. : RYG_EN0061
Submitted by : ALS Laboratory Group (Thailand) Co.,Ltd. (Rayong Branch)
616/10 Moo 5 T. Maenam Khu,
A. Pluakdaeng,
Rayong 21140 Thailand
Location : Wat Chemistry Lab
Received Order : 5 May 2021
Calibration Date : 5 May 2021
Ambient Temperature : $(26 \pm 10) ^\circ\text{C}$
Relative Humidity : $(50 \pm 30) \%$
Calibrated by : Tawatchai Pama
Approved by :
() Pornhippa Tameyakul
() Malee Butkruea
() Suwit Imjai
Issue Date : 14 May 2021

REVIEW BY
APPROVED BY
NEXT CAL DATE 3/1/23

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3: Equipment Calibration and Testing Services

A 0028098



Equipment : Water Bath
Condition As-Received : Used Item
Reference : 2105-0005OC-3
Procedure Used :-

Cert. No.: 21TM673
Page: 2 of 3

Calibration were conducted using in-house calibration procedure CP-OT04 according to direct measurement method with Data Acquisition which connected with Industrial Platinum Resistance Thermometer (IPRT).

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34970A	MY44060450	Z1LM4	06 Mar 2022


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

	Environmental		AC Voltage Supply
	(°C)	(%R.H.)	(Volt)
Beginning of Calibration	22	68	230
Finished of Calibration	20	64	231



Front

Position :	Ref. Std. S/N.:
1	4803988-001
2	4803988-002
3	4803988-003
4	4803988-004
5(ref.)	4803988-005



Equipment : Water Bath
Condition As-Received : Used Item
Reference : 2105-0005OC-3
Result of Calibration :- (*) Without Adjustment
Function of UUC* : Temperature Source

Cert. No.: 21TM673
Page: 3 of 3

Calibration point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Average* Standard Reading (°C)				
			Position				
			1	2	3	4	5 (ref.)
85.0	85.0	85.0	84.891	84.893	84.890	84.892	84.917

Calibration point (°C)	Uniformity (°C)	Stability (± °C)	Uncertainty (± °C)	Coverage Factor k
85.0	0.089	0.052	0.22	2

Average* : The average of 30 values in each position.

Uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Stability : One-half of the greatest maximum difference of measured temperature at any one probe.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-00-

Male

a 1054289

RYG_EN0061



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES & EQUIPMENT CALIBRATION AND TESTING SERVICES
554/4 PATTANAKARN ROAD SOI JK, SUAN LUANG, SUAN LUANG BANGKOK 10250
TEL. 0 2715-3808-27 FAX 0 2710-9484



Cert. No.: 22TM1491
Page : 1 of 3

Certificate of Calibration

Equipment : Water Bath

Manufacturer : Memmert

Model : WNB22

Serial No. : L513.0648

ID No. : RYG_EN0061

Submitted by : ALS Laboratory Group (Thailand) Co., Ltd. (Rayong Branch)
616/10 Moo 5, T. Maenam Khu,
A. Pluakdaeng,
Rayong 21140, Thailand

Location : Wet Chemistry Lab

Received Order : 20 October 2022

Calibration Date : 20 October 2022

Ambient Temperature : (26 ± 10) °C

Relative Humidity : (50 ± 30) %

Calibrated by : Preecha Hiahilb

Approved by :
Approved Signatory

() Pornthippa Tameyakul
(/) Malee Bulkruea
() Suwit Imjai

Issue Date : 2 November 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced either in full or in part without the prior written approval of the head of Corporate Services & Equipment Calibration and Testing Services



Equipment : Water Bath
Condition As-Received : Used Item
Reference : 2210-0376OC-4
Procedure Used :-

Cert. No.: 22TM1491
Page : 2 of 3

Calibration were conducted using in-house calibration procedure CP-OT04 according to direct measurement method with Data Acquisition which connected with Industrial Platinum Resistance Thermometer (IPRT).

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34970A	MY44035217	21LM30	23 Dec 2022


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

	Environmental		AC Voltage Supply
	(°C)	(%R.H.)	(Volt)
Beginning of Calibration	24	53	222
Finished of Calibration	24	50	221



Front

Position :	Ref. Std. S/N.:
1	N37P300728
2	N37P300727
3	N37P300728
4	N37P300729
5(ref.)	N37P300730

Male

a 1054288

Male

A 0046906

a 1132471



Equipment : Water Bath
 Condition As-Received : Used Item
 Reference : 2210-0376OC-4
 Result of Calibration : (*) Without Adjustment
 Function of UUC* : Temperature Source

Cert. No.: 22TM1491
 Page : 3 of 3

Calibration point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Average* Standard Reading (°C)				
			Position				
			1	2	3	4	5 (ref.)
85.0	85.0	85.0	84.527	84.563	84.628	84.516	84.580

Calibration point (°C)	Uniformity (°C)	Stability (± °C)	Uncertainty (± °C)	Coverage Factor k
85.0	0.12	0.081	0.18	2

Average* : The average of 30 values in each position.

Uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Stability : One-half of the greatest maximum difference of measured temperature at any one probe.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

-o0o-

Mdu.

a 1132470



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
 CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
 53/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANI LIANG, BANGKOK 10250
 TEL. 0-2717-3000-24 FAX. 0-2719-9484



Certificate of Calibration

Certificate No.: 22T13
 Page : 1 of 2

Equipment : Digital Thermometer With Sensor
 Manufacturer : Fluke
 Model : 50S
 Serial No.: 76950190
 ID No.: RYG_EN0065

This certificate may not be reproduced other than in full, except with the prior written approval of the head of Corporate Services 3: Equipment Calibration and Testing Services.

Condition As-Received: Used Item
 Received Date: 22 December 2021
 Calibration Date: 04 January 2022
 Reference: 2112-0636DSC
 Ambient Temperature: (25 ± 3) °C
 Relative Humidity: (50 ± 20) %

Submitted by: ALS Laboratory Group (Thailand) Co., Ltd. Rayong Branch
 616/10 Moo 5 T.Maenam Khu, A Pluakdaeng, Rayong
 21140, Thailand

Procedure used: Calibration were conducted using in-house calibration procedure CP-T01 according to comparison with Industrial Platinum Resistance Thermometer (IPRT) into liquid bath temperature controller. The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standards instruments :

Instrument	Model	Serial No.	Certificate No.	Due Date
1) Digital Thermometer	1529	A7A6C9	2111126	14 Oct 2022
2) Industrial Platinum Resistance Thermometer	5027-12	571975	2111126	14 Oct 2022
3) Digital Thermometer	1529	A66176	2111248	16 Nov 2022
4) Industrial Platinum Resistance Thermometer	5027	739435	2111248	16 Nov 2022

2. The certificate is valid only to the item calibrated on date and place of calibration.

3. This Certification is traceable to the International System of Unit maintained at:-

-National Institute of Metrology Thailand (NIMT)

REVIEW BY	Tranitol
APPROVED BY	D. K.
NEXT CAL. DATE	06/07/23

Calibrated by : Thatchanan Chankong
 Issue Date : 12 January 2022

Approved Signatory :

☐ Phalinee Pratsolpal
☒ Chatchawan Khunpluek
☐ Wanlop Larplum

B 0278021



Cert. No.: 22T13
 Page.: 2 of 2

Result of Calibration:- Before Adjustment
 Function: Temperature measurement
 This equipment was connected with Thermocouple Type K ID No. 76950190/T

Immersion Depth (mm.)	Standard Temperature (°C)	UUC* Reading (°C)	Error (°C)	Uncertainty of Measurement (± °C)
150	3.0037	3.7	0.6963	0.36
150	60.0061	59.4	-0.6061	0.38
150	104.0031	103.0	-1.0031	0.49
150	180.0031	177.5	-2.5031	0.72

Result of Calibration:- After Adjustment

Immersion Depth (mm.)	Standard Temperature (°C)	UUC* Reading (°C)	Error (°C)	Uncertainty of Measurement (± °C)
150	3.0026	4.6	1.5974	0.36
150	20.0035	21.2	1.1965	0.36
150	60.0019	60.3	0.2981	0.38
150	70.0021	70.2	0.1979	0.40
150	85.0050	85.2	0.1950	0.44
150	104.0028	104.0	-0.0028	0.49
150	150.0049	149.4	-0.6049	0.62
150	180.0015	178.5	-1.5015	0.72

UUC* : Unit Under Calibration

The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.

-o0o-

© 2022 by Agilent Technologies

Certificate of System Qualification

GC-OQ + GCMS-OQ

System ID: GM-7
 Organization Name: ALS Laboratory Group (Thailand) Co., Ltd.
 Organization Location: 104 Patthanakarn 40, Patthanakarn rd., Khwang Suan Luang, Khet Suan Luang, Bangkok 10250

Date: June 21, 2022 2:04:12 PM
 EOP Name: AgilentRecommended, AgilentRecommended
 EOP Revision: GC.02.50, GCMS.02.50
 Overall Qualification Status: Pass

System Inspection and Basic Safety and Operation

Name: 7690
 Setpoint Status: Pass

Overall System Inspection and Basic Safety and Operation Test Status
 Pass

Inlet Pressure Accuracy

Name: 7690
 Front: SSL
 Setpoint Status: Pass

Setpoint: 25.0 psi
 Actual: 25.0 psi
 Accuracy: 0.0 psi
 Agilent Recommended: 1.2

Overall Inlet Pressure Accuracy Test Status

Pass

GC Oven Temperature Accuracy

Name: 7690

Date: June 21, 2022 2:04:12 PM
 System ID: GM-7

Setpoint Status: **Pass**

Zone: **Oven**

Setpoint/Actual

Temperature: **230.0** **230.0** °C

Accuracy: **0.0** °C

Agilent Recommended: **>=** **-1.0** % setpoint in K (**-5.0** °C)

<= **1.0** % setpoint in K (**5.0** °C)

Setpoint Status: **Pass**

Zone: **Oven**

Setpoint/Actual

Temperature: **100.0** **100.4** °C

Accuracy: **0.4** °C

Agilent Recommended: **>=** **-1.0** % setpoint in K (**-3.7** °C)

<= **1.0** % setpoint in K (**3.7** °C)

Overall GC Oven Temperature Accuracy Test Status

Pass

GC Oven Temperature Stability

Name: **7890**

Setpoint Status: **Pass**

Setpoint/Average

Temperature: **100.0** **100.0333** °C

Stability: **0.1** °C

Agilent Recommended: **<=** **0.5**

Overall GC Oven Temperature Stability Test Status

Pass

Log Amp

Tested Combination1 **Front** **SSL** **/ External** **SQ**

Name: **5977A**

Setpoint Status: **Pass**

Date: June 21, 2022 2:04:12 PM
System ID: GM-7

Page 2 / 15

Overall Log Amp Test Status

Pass

RFPA

Tested Combination1 **Front** **SSL** **/ External** **SQ**

Name: **5977A**

Setpoint Status: **Pass**

Amu: **1050** m/z

Drift After Five Minutes: **22** mV

RFPA Voltage: **569** mV

Agilent Recommended: **>=** **-100** and **<=** **100** **<=** **1100**

Overall RFPA Test Status

Pass

Tune EI

Tested Combination1 **Front** **SSL** **/ External** **SQ**

Name: **5977A**

Setpoint Status: **Pass**

Filament: **1**

Setpoint Status: **Pass**

Filament: **2**

Overall Tune EI Test Status

Pass

Signal to Noise EI

Tested Combination1 **Front** **SSL** **/ External** **SQ**

Name: **5977A**

Date: June 21, 2022 2:04:12 PM
System ID: GM-7

Page 3 / 15

Source: **EI - Extractor** Filament: **1**

Setpoint Status: **Pass**

Signal to Noise: **51283**

Agilent Recommended: **>=** **1200**

Source: **EI - Extractor** Filament: **2**

Setpoint Status: **Pass**

Signal to Noise: **7088**

Agilent Recommended: **>=** **1200**

This test's 0 comment(s) and 1 deviation(s) are available in the Attachments section.

Overall Signal to Noise EI Test Status

Pass

Date: June 21, 2022 2:04:12 PM
System ID: GM-7

Page 4 / 15

Instrument Details

Purpose

This section describes the as found system configuration.

Details

System

System ID **GM-7**

Manufacturer **Agilent Technologies**

Name **7890**

Tested Combination1

Injection Technique **Manual Injection**

Inlet **Front**

Detector **External**

LTM Included? **No**

Sampler 1

Manufacturer **Agilent Technologies**

Type **Manual Injection**

Usage **Sample Injection**

Syringe Volume (µL) **10**

Mainframe 1

Manufacturer **Agilent Technologies**

Name **7890**

Model Number **G3442B**

Serial Number **CN14133181**

Firmware Revision **B.02.03**

Oven Type **Standard**

Date: June 21, 2022 2:04:12 PM
System ID: GM-7

Page 5 / 15

Inlet 1

Manufacturer	Agilent Technologies
Name	7880
Type	SSL
Location	Front
Carrier Gas	Helium
Control Type	Electronic Pressure Control (EPC)
Purged Inlet	Yes

Detector 1

Manufacturer	Agilent Technologies
Name	Mass Spectrometer
Type	Mass Spectrometer
Location	External

Mass Spectrometer 1

Manufacturer	Agilent Technologies
Type	SQ
Name	5977A
Serial Number	U51415M209
Firmware Revision	5977 6.00.21
High Vacuum System	Turbo Pump
Scouting Run Standard	OFN Std

MS EI Source 1

Manufacturer	Agilent Technologies
Source Type	EI - Extractor
Number of filaments	2

Date: June 21, 2022 2:04:12 PM
System ID: GM-7

Page 6 / 15

Electronic Signature

Purpose

This signature page was created and published because the ACE sign-off action was executed, which is valid for the entire document, including attachments. The ACE sign-off is an electronic signature that requires two distinct identification components: unique username and personal password. The Agilent representative who has delivered this service understands the meaning and legal status of an electronic signature. As a trained official operator, the Agilent representative has a unique password and login to access ACE and electronically sign this document. (Other e-signatures can be applied to this document using a Document Content Management or other suitable method defined in your data access and control procedures.)

Details

Full Name of Signer:	Supasak Nimsongtham
Logged On User Name:	supasak.nimsongtham@agilent.com
Signature Creation Date:	June 21, 2022
Reason for Signature:	Executed protocol and published this original version of document

Regulatory Disclaimer

This document provides a protocol to verify and record instrument configuration and evidence of proper operation. It has been prepared from our interpretation of applicable regulations as well as industry best practices. The document is designed to provide an important component of a complete compliance package. Validation depends upon many factors and use of this protocol alone does not assure compliance. Agilent Technologies makes no promises or representations as to its sufficiency for any specific regulatory program.

Warranty

Agilent Technologies makes no warranty of any kind to this material, including but not limited to, the implied warranties or merchantability and fitness for a particular purpose. Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Date: June 21, 2022 2:04:12 PM
System ID: GM-7

Page 7 / 15

User Name: supasak.nimsongtham Hostname: SC01115HNC			System Id: GM-7 Print Date: June 21, 2022 2:04:17 PM	
ALS-GM7-2022 Transaction log:				
Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 21, 2022 10:25:05 AM	Audit	Session Created	Session	None
June 21, 2022 10:25:05 AM	Start	Configuration	Session	None
June 21, 2022 10:25:05 AM	Audit	Entitlement	Licensing	User is Fiddling/peer and does not require an unlock code
June 21, 2022 10:25:26 AM	Audit	Exp/Loaded	Session	EQP details for primary technique [GC] - File path: [ProtocolPackage\Config\Inst02.50\GC.02.50.ecp] EQP File Name: [GC.02.50.ecp] EQP Name: [Agilent Recommended] EQP details for hyphenated technique [GC/MS] - File path: [ProtocolPackage\Config\Inst02.50\GC/MS.02.50.ecp] EQP File Name: [GC/MS.02.50.ecp] EQP Name: [Agilent Recommended]
June 21, 2022 10:25:38 AM	End	Configuration	Session	None
June 21, 2022 10:25:43 AM	Start	Qualification	Session	QQ
June 21, 2022 10:25:43 AM	Start	Execution	System Inspection and Basic Safety and Operation - 7880 - Qualitative Test - No septacts associated	None
June 21, 2022 10:25:54 AM	End	Execution	System Inspection and Basic Safety and Operation - 7880 - Qualitative Test - No septacts associated	Run Count: 1

Page 1/8

Page 1 / 8

Date: June 21, 2022 2:04:12 PM
System ID: GM-7

Page 8 / 15

User Name: supasak.nimsongtham

Hostname: SC01115HNC

System ID: GM-7

Print Date: June 21, 2022 2:04:17 PM

ALS-GM7-2022 Transaction log:

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 21, 2022 10:25:50 AM	Start	Execution	Inlet Pressure Accuracy - Front SSL - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.3 psi	None
June 21, 2022 10:26:10 AM	End	Execution	Inlet Pressure Accuracy - Front SSL - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	Run Count: 1
June 21, 2022 10:26:12 AM	Start	Execution	GC Oven Temperature Accuracy - 7880 - Temperature : Oven - S: 250.0°C - L: >= -1.0 AND <= 1.0 % septact in K	None
June 21, 2022 10:24:09 AM	Audit	Data	GC Oven Temperature Accuracy - 7880 - Temperature : Oven - S: 250.0°C - L: >= -1.0 AND <= 1.0 % septact in K	Manual Data Entry
June 21, 2022 10:24:10 AM	End	Execution	GC Oven Temperature Accuracy - 7880 - Temperature : Oven - S: 250.0°C - L: >= -1.0 AND <= 1.0 % septact in K	Run Count: 1
June 21, 2022 10:24:11 AM	Start	Execution	GC Oven Temperature Accuracy - 7880 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % septact in K	None
June 21, 2022 10:28:42 AM	Audit	Data	GC Oven Temperature Accuracy - 7880 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % septact in K	Manual Data Entry
June 21, 2022 10:28:44 AM	End	Execution	GC Oven Temperature Accuracy - 7880 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % septact in K	Run Count: 1
June 21, 2022 10:28:45 AM	Start	Execution	GC Oven Temperature Stability - 7880 - Temperature : Oven - S: 100.0°C - L: <= 0.5°C	None

Page 2 / 8

Page 2 / 8

Date: June 21, 2022 2:04:12 PM
System ID: GM-7

Page 9 / 15

User Name: supasak.nimsongtham
Hostname: SCG115HKC
System ID: GM-7
Print Date: June 21, 2022 2:04:17 PM

ALS-GM7-2022 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 21, 2022 11:01:00 AM Audit	AcceClosed	Session	Session	None
June 21, 2022 11:01:47 AM Audit	AcceRestarted	Session	Session	None
June 21, 2022 11:01:48 AM Audit	SessionRelocated	Session	Session	None
June 21, 2022 11:01:51 AM Start	Qualification	Session	Session	OQ
June 21, 2022 11:01:51 AM Start	Execution	GC Oven Temperature Stability	None	- 7850 : - Temperature : Oven - S: 100.0°C - L <= 0.5°C
June 21, 2022 11:03:14 AM Audit	Data	DataManager	DataManager	DataManager was in a data verification state but the user chose to start over.
June 21, 2022 11:04:19 AM Audit	Data	GC Oven Temperature Stability	Manual Data Entry	- 7850 : - Temperature : Oven - S: 100.0°C - L <= 0.5°C
June 21, 2022 11:04:22 AM End	Execution	GC Oven Temperature Stability	Run Count : 1	- 7850 : - Temperature : Oven - S: 100.0°C - L <= 0.5°C
June 21, 2022 11:04:24 AM Start	Execution	Log Amp - 5977A SQ: - Source: None	None	EI - Extractor
June 21, 2022 11:04:34 AM End	Execution	Log Amp - 5977A SQ: - Source: Run Count: 1	None	EI - Extractor
June 21, 2022 11:04:37 AM Start	Execution	RPPA - 5977A SQ: - Source: EI None	None	- Extractor
June 21, 2022 11:07:49 AM End	Execution	RPPA - 5977A SQ: - Source: EI Run Count: 1	None	- Extractor
June 21, 2022 11:07:52 AM Start	Execution	Tune EI - 5977A SQ: - Source: - None	None	EI - Extractor Filament 1 (Qualitative - No septums associated)

Page 3 / 8

Date: June 21, 2022 2:04:12 PM
System ID: GM-7

Page 10 / 15

User Name: supasak.nimsongtham
Hostname: SCG115HKC
System ID: GM-7
Print Date: June 21, 2022 2:04:17 PM

ALS-GM7-2022 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 21, 2022 11:08:35 AM End	Execution	Tune EI - 5977A SQ: - Source: - Run Count: 1	None	EI - Extractor Filament 1 (Qualitative - No septums associated)
June 21, 2022 11:14:59 AM Start	Execution	Tune EI - 5977A SQ: - Source: - None	None	EI - Extractor Filament 2 (Qualitative - No septums associated)
June 21, 2022 11:15:48 AM End	Execution	Tune EI - 5977A SQ: - Source: - Run Count: 1	None	EI - Extractor Filament 2 (Qualitative - No septums associated)
June 21, 2022 11:16:49 AM Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None	
June 21, 2022 11:17:05 AM Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	None	
June 21, 2022 11:17:10 AM Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None	
June 21, 2022 11:26:09 AM Audit	AcceClosed	Session	Session	None
June 21, 2022 12:30:20 PM Audit	AcceRestarted	Session	Session	None
June 21, 2022 12:36:22 PM Audit	SessionRelocated	Session	Session	None
June 21, 2022 12:36:26 PM Start	Qualification	Session	Session	OQ
June 21, 2022 12:36:26 PM Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None	

Page 4 / 8

Date: June 21, 2022 2:04:12 PM
System ID: GM-7

Page 11 / 15

User Name: supasak.nimsongtham
Hostname: SCG115HKC
System ID: GM-7
Print Date: June 21, 2022 2:04:17 PM

ALS-GM7-2022 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 21, 2022 12:37:07 PM Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	None	
June 21, 2022 12:37:08 PM Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None	
June 21, 2022 12:38:54 PM Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data File Path: H:\ALSGM7_2022\SNF1_001.D	
June 21, 2022 12:39:24 PM Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data File Path: H:\ALSGM7_2022\SNF1_001.D	
June 21, 2022 12:40:09 PM Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data File Path: H:\ALSGM7_2022\SNF1_001.D	
June 21, 2022 12:42:04 PM Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data File Path: H:\ALSGM7_2022\SNF1_001.D	
June 21, 2022 12:42:17 PM Audit	AcceClosed	Session	Session	None
June 21, 2022 12:33:31 PM Audit	AcceRestarted	Session	Session	None
June 21, 2022 12:33:33 PM Audit	SessionRelocated	Session	Session	None
June 21, 2022 12:35:37 PM Start	Qualification	Session	Session	OQ
June 21, 2022 12:35:37 PM Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None	

Page 5 / 8

Date: June 21, 2022 2:04:12 PM
System ID: GM-7

Page 12 / 15

User Name: supasak.nimsongtham
Hostname: SCG115HKC
System ID: GM-7
Print Date: June 21, 2022 2:04:17 PM

ALS-GM7-2022 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 21, 2022 12:34:44 PM Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data File Path: E:\ALSGM7_2022\SNF1_001.D	
June 21, 2022 12:35:26 PM End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Run Count: 1	
June 21, 2022 12:37:11 PM Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	None	
June 21, 2022 12:38:15 PM Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data File Path: E:\ALSGM7_2022\SNF2_001.D	
June 21, 2022 12:38:30 PM Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data File Path: E:\ALSGM7_2022\SNF2_001.D	
June 21, 2022 12:38:45 PM Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data File Path: E:\ALSGM7_2022\SNF2_001.D	
June 21, 2022 12:39:00 PM Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data File Path: E:\ALSGM7_2022\SNF2_001.D	
June 21, 2022 12:39:14 PM Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data File Path: E:\ALSGM7_2022\SNF2_001.D	

Page 6 / 8

Date: June 21, 2022 2:04:12 PM
System ID: GM-7

Page 13 / 15

User Name: supasak.nimsongtham

Hostname: SCG1115HKC

System ID: GM-7

Print Date: June 21, 2022 2:04:17 PM

ALS-GM7-2022 Transaction Log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 21, 2022 12:39:45 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data File Path: E:\ALSGM7_2022\SNF2_001.D
June 21, 2022 12:40:16 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data File Path: E:\ALSGM7_2022\SNF2_001.D
June 21, 2022 12:40:40 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data File Path: E:\ALSGM7_2022\SNF2_001.D
June 21, 2022 12:41:00 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data File Path: E:\ALSGM7_2022\SNF2_001.D
June 21, 2022 12:41:29 PM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Run Count: 1
June 21, 2022 12:42:30 PM	Audit	TestUnlocked	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Deviation Ref for Run Count: 1
June 21, 2022 12:42:30 PM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	None
June 21, 2022 12:42:35 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data File Path: E:\ALSGM7_2022\SNF2_001.D

Page 7 / 8

Page 7 / 8

Date: June 21, 2022 2:04:12 PM
System ID: GM-7

Page 14 / 15

User Name: supasak.nimsongtham
 Hostname: SCG1115HKC

System ID: GM-7
 Print Date: June 21, 2022 2:04:17 PM

ALS-GM7-2022 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 21, 2022 12:42:45 PM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Run Count: 2
June 21, 2022 12:42:50 PM	End	Qualification	Session	OQ
June 21, 2022 12:42:50 PM	Start	Reporting	Session	None
June 21, 2022 12:45:17 PM	Audit	AcqGated	Session	None
June 21, 2022 1:07:47 PM	Audit	AcqRestarted	Session	None
June 21, 2022 1:07:50 PM	Audit	SessionRelapsed	Session	None
June 21, 2022 1:07:55 PM	Start	Qualification	Session	OQ
June 21, 2022 2:02:42 PM	Audit	Reporting	Session	Report Generated: Certificate

Page 6 / 8


Page 8 / 8

Date: June 21, 2022 2:04:12 PM
System ID: GM-7

Page 15 / 15

Certificate of System Qualification

GC-QQ

System ID: GC-5
Organization Name: ALS Laboratory Group (Thailand) Co., Ltd.
Organization Location: 104 Phattanakan 40, Phattanakan Rd., Suan Luang, Bangkok 10250Date: October 20, 2021 10:15:57 AM
EQP Name: AgilentRecommended
EQP Revision: GC-02.50
Overall Qualification Status: Pass

System Inspection and Basic Safety and Operation

Name: 7890
Setpoint Status: PassOverall System Inspection and Basic Safety and Operation Test Status
Pass

Inlet Pressure Decay

Name: 7890
Front SSL
Setpoint Status: Pass
Pressure: 25.0 psi
Pressure Change: 0.3 psi / 5 minutes
Agilent Recommended: >= -2.0 and <= 0.5Overall Inlet Pressure Decay Test Status
Pass

Inlet Pressure Accuracy

Name: 7890
Front SSLDate: October 20, 2021 10:15:57 AM
System ID: GC-5

Page 1 / 15

Setpoint Status: Pass
Inlet Pressure: 25.0 psi Actual: 24.9 psi
Accuracy: 0.1 psi
Agilent Recommended: <= 1.2

Overall Inlet Pressure Accuracy Test Status

Pass

Detector Flow Accuracy

Name: 7890
Front FID
Setpoint Status: Pass
Flow Type: Fuel
Setpoint: 30.0 mL/min Measured Flow: 30.0 mL/min
Accuracy: 0.0 mL/min
Agilent Recommended: <= 10.0 % setpoint (3.0 mL/min)
Limit is percentage of setpoint or 0.5 mL/minute, whichever is largest.Setpoint Status: Pass
Flow Type: Oxidizer
Setpoint: 400.0 mL/min Measured Flow: 350.3 mL/min
Accuracy: 9.7 mL/min
Agilent Recommended: <= 10.0 % setpoint (40.0 mL/min)
Limit is percentage of setpoint or 0.5 mL/minute, whichever is largest.Setpoint Status: Pass
Flow Type: Makeup
Setpoint: 25.0 mL/min Measured Flow: 24.5 mL/min
Accuracy: 0.5 mL/min
Agilent Recommended: <= 10.0 % setpoint (2.5 mL/min)
Limit is percentage of setpoint or 0.5 mL/minute, whichever is largest.Date: October 20, 2021 10:15:57 AM
System ID: GC-5

Page 2 / 15

Overall Detector Flow Accuracy Test Status
Pass

GC Oven Temperature Accuracy

Name: 7890

Setpoint Status: Pass

Zone: Oven

Setpoint/Actual

Temperature: 230.0 231.2 °C

Accuracy: 1.2 °C

Agilent Recommended: >= -1.0 % setpoint in K (-5.0 °C)
<= 1.0 % setpoint in K (5.0 °C)

Setpoint Status: Pass

Zone: Oven

Setpoint/Actual

Temperature: 100.0 100.4 °C

Accuracy: 0.4 °C

Agilent Recommended: >= -1.0 % setpoint in K (-3.7 °C)
<= -1.0 % setpoint in K (3.7 °C)

Overall GC Oven Temperature Accuracy Test Status
Pass

GC Oven Temperature Stability

Name: 7890

Setpoint Status: Pass

Setpoint/Average

Temperature: 100.0 100.4 °C

Stability: 0.0 °C

Agilent Recommended: <= 0.5

Overall GC Oven Temperature Stability Test Status
Pass

Date: October 20, 2021 10:15:57 AM
System ID: GC-5

Scouting Run

Tested Combination1 Front SSL / Front FID

Injection Tower

Name: 7683B

Setpoint Status: Completed

Injection Volume on Column: 1.0 uL

Overall Scouting Run Status
Completed

Noise and Drift

Tested Combination1 Front SSL / Front FID

Name: 7890

Setpoint Status: Pass

Base Signal: 20.2 pA

ASTM Noise

Drift

Agilent Recommended: <= 0.10 <= 2.50

Status: Pass Pass

Overall Noise and Drift Test Status
Pass

Injection Precision

Tested Combination1 Front SSL / Front FID

Name: 7683B

Date: October 20, 2021 10:15:57 AM
System ID: GC-5

Setpoint Status: Pass

Injection Volume on Column: 1.0 uL

Area RSD: 0.52 %

Retention Time RSD: 0.22 %

Agilent Recommended: <= 3.00 <= 1.00

Overall Injection Precision Test Status

Pass

Signal to Noise

Tested Combination1 Front SSL / Front FID

Injection Tower

Name: 7890

Setpoint Status: Pass

Signal to Noise: 1258310

Agilent Recommended: >= 300000

Overall Signal to Noise Test Status

Pass

Date: October 20, 2021 10:15:57 AM
System ID: GC-5

Instrument Details

Purpose

This section describes the as found system configuration.

Details

System

System ID: GC-5

Manufacturer: Agilent Technologies

Name: 7890

Flow Data Input: Manual Data

Temperature Data Input: Manual Data or Other Data Logging

Tested Combination1

Injection Technique: Injection Tower

Inlet: Front

Detector: Front

LTM Included?: No

Sampler 1

Manufacturer: Agilent Technologies

Type: Injection Tower

Name: 7683B

Model Number: G2913A

Serial Number: CN00259643

Firmware Revision: A.11.03

Usage: Sample Injection

Location: Front

Syringe Volume (uL): 10

Sampler 2

Manufacturer: Agilent Technologies

Type: Tray

Name: 7683A

Model Number: G2614A

Serial Number: CN81347892

Firmware Revision: A.02.01

Date: October 20, 2021 10:15:57 AM
System ID: GC-5

User Name: suriya.thongkaew		System Id: GC-5		
Hostname: ASDKKW7029		Print Date: October 20, 2021 10:16:00 AM		
DQ GC ALS US10813027 Transaction log :				
Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 19, 2021 11:52:55 AM	Start	Execution	Detector Flow Accuracy - Front FID - Type: Makeup - S 25.0 mL/min - L <= 10.0% setpoint	None
October 19, 2021 12:02:38 PM	Audit	Data	Detector Flow Accuracy - Front FID - Type: Makeup - S 25.0 mL/min - L <= 10.0% setpoint	Manual Data Entry
October 19, 2021 12:03:39 PM	End	Execution	Detector Flow Accuracy - Front FID - Type: Makeup - S 25.0 mL/min - L <= 10.0% setpoint	Run Count: 1
October 19, 2021 12:03:42 PM	Start	Execution	GC Oven Temperature Accuracy - 7850 - Temperature Oven - S 230.0°C - L >= -1.0 AND <= 1.0 % setpoint n K	None
October 19, 2021 12:23:20 PM	Audit	Data	GC Oven Temperature Accuracy - 7850 - Temperature Oven - S 230.0°C - L >= -1.0 AND <= 1.0 % setpoint n K	Manual Data Entry
October 19, 2021 12:23:24 PM	End	Execution	GC Oven Temperature Accuracy - 7850 - Temperature Oven - S 230.0°C - L >= -1.0 AND <= 1.0 % setpoint n K	Run Count: 1
October 19, 2021 12:23:28 PM	Start	Execution	GC Oven Temperature Accuracy - 7850 - Temperature Oven - S 100.0°C - L >= -1.0 AND <= 1.0 % setpoint n K	None
October 19, 2021 12:33:48 PM	Audit	Data	GC Oven Temperature Accuracy - 7850 - Temperature Oven - S 100.0°C - L >= -1.0 AND <= 1.0 % setpoint n K	Manual Data Entry
October 19, 2021 12:33:50 PM	End	Execution	GC Oven Temperature Accuracy - 7850 - Temperature Oven - S 100.0°C - L >= -1.0 AND <= 1.0 % setpoint n K	Run Count: 1

Page 3 / 7

Page 3 / 7

User Name: suriya.thongkaew

Hostname: ASDKKW7029

System Id: GC-5

Print Date: October 20, 2021 10:16:00 AM

DQ GC ALS US10813027 Transaction log:

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 19, 2021 12:33:53 PM	Start	Execution	GC Oven Temperature Stability - 7850 - Temperature Oven - S 100.0°C - L <= 0.5°C	None
October 19, 2021 12:54:48 PM	Audit	Data	GC Oven Temperature Stability - 7850 - Temperature Oven - S 100.0°C - L <= 0.5°C	Manual Data Entry
October 19, 2021 12:54:49 PM	End	Execution	GC Oven Temperature Stability - 7850 - Temperature Oven - S 100.0°C - L <= 0.5°C	Run Count: 1
October 19, 2021 12:54:52 PM	Start	Execution	GC Screening Run - Injection Tower, Front SSL, Front FID - Part of System Preparation - No limits associated	None
October 19, 2021 4:48:40 PM	Audit	AccClosed	Session	None
October 20, 2021 9:34:06 AM	Audit	AccRestarted	Session	None
October 20, 2021 9:34:08 AM	Audit	SessionReloaded	Session	None
October 20, 2021 9:34:12 AM	Start	Qualification	Session	OQ
October 20, 2021 9:34:12 AM	Start	Execution	GC Screening Run - Injection Tower, Front SSL, Front FID - Part of System Preparation - No limits associated	None
October 20, 2021 9:35:51 AM	Audit	Data	GC Screening Run - Injection Tower, Front SSL, Front FID - Part of System Preparation - No limits associated	Data files Path: C:\Chem32\1\DATA\QGPV20 21\QGPV2021 2021-10-19 17-23-14\SC04\IN\PREC004.D\FID1A.ch

Page 4 / 7

Page 4 / 7

Date: October 20, 2021 10:15:57 AM
System ID: GC-5

Page 11 / 15

Date: October 20, 2021 10:15:57 AM
System ID: GC-5

Page 12 / 15

User Name: suriya.thongkaew			System Id: GC-5	
Hostname: ASDKKW7029			Print Date: October 20, 2021 10:16:00 AM	
DQ GC ALS US10813027 Transaction log:				
Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 20, 2021 9:36:24 AM	End	Execution	GC Screening Run - Injection Tower, Front SSL, Front FID - Part of System Preparation - No limits associated	Run Count: 1
October 20, 2021 9:36:27 AM	Start	Execution	Noise and Drift - Front FID - Detector FID - L (Noise) <= 0.10 pA - L (Drift) <= 2.50 pA/hour	None
October 20, 2021 9:37:19 AM	Audit	Data	Noise and Drift - Front FID - Detector FID - L (Noise) <= 0.10 pA - L (Drift) <= 2.50 pA/hour	Data files Path: C:\Chem32\1\DATA\QGPV20 21\QGPV2021 2021-10-19 17-23-14\IN\PREC002.D\FID1A.ch
October 20, 2021 9:37:30 AM	End	Execution	Noise and Drift - Front FID - Detector FID - L (Noise) <= 0.10 pA - L (Drift) <= 2.50 pA/hour	Run Count: 1
October 20, 2021 9:37:32 AM	Start	Execution	Injection Precision - Injection Tower, Front SSL, Front FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	None
October 20, 2021 9:37:53 AM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Front FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Data files Path: C:\Chem32\1\DATA\QGPV20 21\QGPV2021 2021-10-19 17-23-14\IN\PREC002.D\FID1A.ch
October 20, 2021 9:37:53 AM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Front FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Data files Path: C:\Chem32\1\DATA\QGPV20 21\QGPV2021 2021-10-19 17-23-14\IN\PREC003.D\FID1A.ch

Page 5 / 7

Page 5 / 7

User Name: suriya.thongkaew			System Id: GC-5	
Hostname: ASDKHW7029			Print Date: October 20, 2021 10:16:00 AM	
DQ GC ALS US10813027 Transaction log:				
Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
October 20, 2021 9:37:53 AM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Front FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Data files Path: C:\Chem32\1\DATA\QGPV20 21\QGPV2021 2021-10-19 17-23-14\IN\PREC004.D\FID1A.ch
October 20, 2021 9:37:53 AM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Front FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Data files Path: C:\Chem32\1\DATA\QGPV20 21\QGPV2021 2021-10-19 17-23-14\IN\PREC005.D\FID1A.ch
October 20, 2021 9:37:53 AM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Front FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Data files Path: C:\Chem32\1\DATA\QGPV20 21\QGPV2021 2021-10-19 17-23-14\IN\PREC006.D\FID1A.ch
October 20, 2021 9:37:53 AM	Audit	Data	Injection Precision - Injection Tower, Front SSL, Front FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Data files Path: C:\Chem32\1\DATA\QGPV20 21\QGPV2021 2021-10-19 17-23-14\IN\PREC007.D\FID1A.ch
October 20, 2021 9:38:21 AM	End	Execution	Injection Precision - Injection Tower, Front SSL, Front FID - GC - L (Area) <= 3.00% - L (Ret. Time) <= 1.00%	Run Count: 1
October 20, 2021 9:38:28 AM	Start	Execution	Signal to Noise - Injection Tower, Front SSL, Front FID - Detector FID - L >= 300000	None
October 20, 2021 9:38:42 AM	Audit	Data	Signal to Noise - Injection Tower, Front SSL, Front FID - Detector FID - L >= 300000	Data files Path: C:\Chem32\1\DATA\QGPV20 21\QGPV2021 2021-10-19 17-23-14\IS\QTONS001.D\FID1A.ch
October 20, 2021 9:38:56 AM	End	Execution	Signal to Noise - Injection Tower, Front SSL, Front FID - Detector FID - L >= 300000	Run Count: 1

Page 6 of 7

Page 6 / 7

Date: October 20, 2021 10:15:57 AM
System ID: GC-5

Page 13 / 15

Date: October 20, 2021 10:15:57 AM
System ID: GC-5

Page 14 / 15

User Name: suriya.thongkiew		System ID: GC-5		
Hostname: ASBKKW7029		Print Date: October 28, 2021 10:16 00 AM		
OQ GC ALS US10813027 Transaction log :				
Time	Transaction Date	Activity Performed	Type of Transaction	Optional Information
October 20, 2021 9:38:54 AM	End	Qualification	Session	OQ
October 20, 2021 9:38:54 AM	Start	Reporting	Session	None
October 20, 2021 10:15:14 AM	Audit	Reporting	Session	Report Generated Certificate

Page 7 / 7

Page 7 / 7

Date: October 20, 2021 10:15:57 AM
System ID: GC-5

Page 15 / 15



Agilent CrossLab Compliance

Qualification Type: ICPMS-OQ

System ID: JP15471169

EQP Name: AgilentRecommended

EQP Revision: ICPMS.02.50

EQP Publish Date: March 2020

Date: September 30, 2021 4:07:18 PM

Report Type: Report

Org. Name: ALS Laboratory Group (Thailand) Co., Ltd.

Org. Location: 104 Phattanakarn 40, Suan Luang, Bangkok 10250.

REVIEW BY Sophan M.

APPROVED BY Somchai N.

NEXT CAL. DATE 29 March 2023

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 1 / 34

Table of Contents

Section	Page
Cover	1
Table of Contents	2
Test Summary	3
Service Details	4
Instrument Details	5
Calculation Formulas	7
Protocol Details	8
Tests	9
Autosampler Check : SPS4	9
Integrated Sample Introduction System (ISIS) Check : ISIS3	10
Autotune : G8403A	11
Background (No Gas Mode) : G8403A	13
Background (Gas Modes) : G8403A	14
20-Minute Stability (No Gas Mode) : G8403A	15
Declaration of Change Control	16
Attachments	17
Electronic Signature	31
Transaction Logs	32

Test Summary

Purpose

This section includes a status for each scheduled test and the overall qualification. For each test that is run, (1) the status is automatically determined based on pre-defined limits, and (2) the total number of times the test was run is displayed. For detailed results and specifications for a test, refer to the test results in this EQR.

Details	Status	Runs
Test		
Autosampler Check : SPS4	Pass	1
Integrated Sample Introduction System (ISIS) Check : ISIS3	Pass	1
Autotune : G8403A	Pass	1
Background (No Gas Mode) : G8403A	Pass	1
Background (Gas Modes) : G8403A	Pass	1
20-Minute Stability (No Gas Mode) : G8403A	Pass	1
Overall Qualification Status	Pass	

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 2 / 34

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 3 / 34

Service Details

Purpose

This section includes local contact and delivery details for this service.

General Details

Service Order No./Request: 0004837154
EQP Name: Agilent/Recommended
EQP Revision: ICPMS.02.50
Report Type: Report

Organization Details

Name: ALS Laboratory Group (Thailand) Co., Ltd.
Location: 104 Phattanakarn 40, Suan Luang, Bangkok 10250.

Local Contact Details

Name: Chatchanal Komarakul.
Job Title: Manager
Qualification Location: Laboratory

Operator Details

Name: Panthep Kurasathain
Job Title: Field Service Engineer.

Data Acquisition Details

Acquisition Software Name: MassHunter
Acquisition Software Revision: C.01.04

Customer Data System (CDS): IcpMs: MassHunter

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Instrument Details

Purpose

This section describes the as found system configuration.

Details

ICP-MS 1

Manufacturer: Agilent Technologies
Name: 7900
Model Number: G8403A
Installed Options: #100H: Standard Package with Hydrogen option
Detector Type: SQ
Nebulizer: Mira Mist (G3161)
Spray Chamber: Quartz
 Torch: Quartz
Sampling Cone: Ni
Skimmer Cone: Ni
Serial Number: JP15471169
Firmware Revision: C.01.04

ISIS 1

Manufacturer: Agilent Technologies
Name: ISIS3
Model Number: G8411A
Type: Peristaltic pump system
Serial Number: JP15510227

Autosampler 1

Manufacturer: Agilent Technologies
Name: SPS4
Model Number: G8410A
Serial Number: AU15430722

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Chiller 1

Manufacturer: Agilent Technologies
Name: Chiller
Model Number: G3292A
Serial Number: 3U1610713

Calculation Formulas

Purpose

This section includes calculation formulas for all available tests. Depending upon which tests are scheduled, all or some apply to your qualification.

For a description of calculations for ICP-MS tests performed by the MassHunter software, refer to the MassHunter application and documentation.

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Protocol Details

Purpose

This section lists the revisions for all test units used in this report. For complete test-specific and high-level change details, refer to the Revision History document.

Test Revision	Test
ICPMS.02.50	20-Minute Stability (No Gas Mode)
ICPMS.02.50	Autosampler Check
ICPMS.02.50	Autotune
ICPMS.02.50	Background (Gas Modes)
ICPMS.02.50	Background (No Gas Mode)
ICPMS.02.50	Integrated Sample Introduction System (ISIS) Check

Autosampler Check

Purpose

This test demonstrates that the autosampler module is correctly installed and connected. It does not test module performance.

Setpoint			
Results			
Criteria	Observed Result	Expected Result	Status
After the self test, is probe in the home position?	Yes	Yes	Pass
As commanded, is the probe positioned at vial 2?	Yes	Yes	Pass
Setpoint Status:	Pass		Runs: 1
Overall Autosampler Check Test Status			
Pass			

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 8 / 34

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 9 / 34

Integrated Sample Introduction System (ISIS) Check

Purpose

This test demonstrates that the ISIS module is correctly installed and connected. It does not test module performance.

Setpoint			
Results			
Criteria	Observed Result	Expected Result	Status
As commanded, does the pump rotate?	Yes	Yes	Pass
As commanded, do the valves load and inject?	Yes	Yes	Pass
Setpoint Status:	Pass		Runs: 1
Overall Integrated Sample Introduction System (ISIS) Check Test Status			
Pass			

Autotune

Purpose

This test uses traceable checkout standards to run a software-executed autotune in all modes. The tune report provides values for peak width, mass axis, sensitivity, oxide species, and doubly-charged species tests.

Setpoint			
Results			
Peakwidth Mass 7	0.719	AMU	
Agilent Recommended:	>= 0.65		
	<= 0.80		
Status:	Pass		
Peakwidth Mass 89	0.750	AMU	
Agilent Recommended:	>= 0.65		
	<= 0.80		
Status:	Pass		
Peakwidth Mass 205	0.713	AMU	
Agilent Recommended:	>= 0.65		
	<= 0.80		
Status:	Pass		
Mass Axis 7	7.05	AMU	
Agilent Recommended:	>= 6.9		
	<= 7.1		
Status:	Pass		
Mass Axis 89	88.85	AMU	
Agilent Recommended:	>= 88.9		
	<= 89.1		
Status:	Pass		
Mass Axis 205	205.00	AMU	
Agilent Recommended:	>= 204.9		
	<= 205.1		
Status:	Pass		

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 10 / 34

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 11 / 34

Mass 7 Sensitivity No Gas

Agilent Recommended:

Status:

94.28	Mcps/ppm
>= 25.5	
Pass	

Mass 89 Sensitivity No Gas

Agilent Recommended:

Status:

307.15	Mcps/ppm
>= 127.5	
Pass	

Mass 205 Sensitivity No Gas

Agilent Recommended:

Status:

203.77	Mcps/ppm
>= 76.5	
Pass	

Mass 59 Sensitivity He

Agilent Recommended:

Status:

28.38	Mcps/ppm
>= 23.8	
Pass	

Mass 89 Sensitivity H2

Agilent Recommended:

Status:

129.27	Mcps/ppm
>= .68	
Pass	

Oxide Ratio 156/140

Agilent Recommended:

Status:

1.047	%
<= 1.38	
Pass	

Doubly Charged Species Ratio 70/140

Agilent Recommended:

Status:

1.482	%
<= 2.3	
Pass	

Setpoint Status:

Pass

Runs: 1

Overall Autotune Test Status

Pass

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 12 / 34

Background (No Gas Mode)

Purpose

This test examines the background of the ICP-MS in no gas mode by monitoring ions during a blank run.

Setpoint

Conditions

Masses:	7	AMU
	89	AMU
	205	AMU

Measurements and Results

Masses (AMU):

Measured Value:

Agilent Recommended:

Status:

7	89	205
3.200	3.300	9.900
cps		
<= 6.9	<= 4.6	<= 11.5
Pass	Pass	Pass

Setpoint Status:

Pass

Runs: 1

Overall Background (No Gas Mode) Test Status

Pass

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 13 / 34

Background (Gas Mode)

Purpose

This test examines the background of the ICP-MS in the various gas modes by monitoring ions during a blank run.

Setpoint Gas Mode: Helium

Conditions

Mass:	78	AMU
Integration Time:	1.0	sec
Cycles:	20	

Measurements and Results

Mass (AMU):

Measured Value:

Agilent Recommended:

Status:

78	
42.8500	cps
<= 115	
Pass	

Setpoint Status:

Pass

Runs: 1

Setpoint Gas Mode: Hydrogen

Conditions

Mass:	78	AMU
Integration Time:	1.0	sec
Cycles:	20	

Measurements and Results

Mass (AMU):

Measured Value:

Agilent Recommended:

Status:

78	
2.1500	cps
<= 4.6	
Pass	

Setpoint Status:

Pass

Runs: 1

Overall Background (Gas Mode) Test Status

Pass

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 14 / 34

20-Minute Stability (No Gas Mode)

Purpose

This test monitors the abundance of ions present in the checkout standard over a 20-minute period to verify that the signal is stable. The %RSD of the abundance of given ions is calculated internally by the software and compared to the limit.

Setpoint

Conditions

Mode:	Spectrum	
Masses:	7, 9, 59, 89, 140, 205	
Integration Time:	9.99	sec
Peak Pattern:	3	points/peak
Repetitions:	20	
Sweeps/Replicates:	100	

Measurements and Results

Masses (AMU):

Stability RSD:

Agilent Recommended:

Status:

7	89	205
0.98400	0.51495	0.73011
%		
<= 2.3	<= 2.3	<= 2.3
Pass	Pass	Pass

Setpoint Status:

Pass

Runs: 1

Overall 20-Minute Stability (No Gas Mode) Test Status

Pass

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 15 / 34

Declaration of Change Control


This document is under change control. Revision history is maintained and printed on each document. Access to the master documents is limited to process owners. Documents receive periodic review and cannot be assigned an evergreen status. The qualification performed according to this document refers only to the hardware/software configuration in place at the time of the qualification. Agilent Technologies recommends that instrument configuration change management procedures be in place in order to maintain the validation process. Any changes to the analytical or computer hardware or software must be clearly specified. A change management system provides a means for determining the degree of requalification required according to the extent of the changes made. All details of the changes must be thoroughly recorded and documented, together with details of completed tests and their results. Note: Hardware/software configuration management is the customer's responsibility.

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 16 / 34

General

Document Name: Certificate of System Qualification

		
Agilent Compliance Engine Self Qualification		
Date: September 14, 2021 4:50:10 PM		
Drive Serial #: ACA25C9	Platform Revision: ACE 3.11	
<small>Individual self-qualification reports for each specific technique installed are also available upon request. They provide additional details on the general report from the concise summary and are structured by the actual algorithms challenged during the process. There is not a one-to-one relationship between algorithms and OQ program tests because some algorithms are used by several tests and across multiple similar hardware components of the qualified systems.</small>		
Technique Type	Tests Completed	Result
Atomic Absorption	7	Conforms
Capillary Electrophoresis	10	Conforms
Discolorimetry	6	Conforms
Emission Spectroscopy	3	Conforms
Gas Chromatography - GC/MS	17	Conforms
Gas Chromatography	29	Conforms
HPLC/MS Chromatography	9	Conforms
ICP-MS	6	Conforms
Infrared Spectroscopy	7	Conforms
Liquid Chromatography	17	Conforms
Liquid Chromatography - LC/MS	6	Conforms
Microfluidics	18	Conforms
Sample Preparation - Gas Chromatography	9	Conforms
Sample Preparation - Liquid Chromatography	6	Conforms
Supercritical Fluid Chromatography	15	Conforms
Software	6	Conforms
UV-Vis Spectrophotometer	13	Conforms
Overall Qualification Status		
Conforms		

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 16 / 34

Attachments

Training requirements note: The delivery engineer attaches an ACE technique-specific training certificate to the Equipment Qualification Report (EQR). Obtaining ACE technique-specific certification includes pre-requisite trainings for Data Integrity, General Compliance topics (GMP, GLP, ALCOA, etc.), instrument hardware and software components, and the ACE technique itself. The one certificate encompasses all pre-requisite trainings as documented in the Agilent Learning Management System called Success Factors.

Location	Category	Document Name	Page
EQR	General	Certificate of System Qualification	18
EQR	General	Operator's training certificate and qualifications	19
EQR	General	Certificate of Qualification for ACE	20
EQR	General	Certificate of Qualification for ACE	21
EQR	General	Tune reports	22
EQR	General	Test Report	25
EQR	General	Test Report	27
EQR	General	Test Report	29

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 17 / 34

General

Document Name: Operator's training certificate and qualifications

	
Certificate of Completion	
Learner Name:	Pamela Kurstin
Title Of Course:	AN-CE-ICPMS-2-018-A: Agilent 7900 ICPMS FSE update training
Completion Date:	June 7, 2014
Certified By Company:	Learning at Agilent
<small>All Service and Support training certificates have the following specific limitations: A certificate for Service and Support training is only valid while employed by Agilent Technologies or while working as an Agilent-authorized service provider, through which the service employee has ongoing access to Agilent's Safety Alerts, Service Portal, internal technical updates, update training, current documentation, technical support, internal parts, and parts updates. Completion of training alone, without being employed by Agilent Technologies, does not qualify an individual to safely install, service or maintain Agilent products.</small>	

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 19 / 34

General

Document Name: Certificate of Qualification for ACE



Certificate of Completion

Learner Name: Panthep Kurusetham

Title Of Course: AN-CE-SS-II-030-A: ACE 3.X User Update Training

Completion Date: July 7, 2020

Certified By Company: Learning at Agilent

All Service and Support training certificates have the following specific limitations.

A certificate for Service and Support training is only valid while employed by Agilent Technologies or while working as an Agilent-authorized service provider, through which the service employee has ongoing access to Agilent's Safety Alerts, Service Notes, internal technical updates, update training, course development, technical support, course parts, and parts updates. Completion of training alone, without being employed by Agilent Technologies, does not qualify an individual to safely install, service or maintain Agilent products.

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 20 / 34

General

Document Name: Certificate of Qualification for ACE



Certificate of Completion

Learner Name: Panthep Kurusetham

Title Of Course: AN-CE-ICPMAS-2-035-B: CrossLab Compliance Hardware Specific Delivery for Agilent ICP-MS Systems

Completion Date: October 31, 2020

Certified By Company: Learning at Agilent

All Service and Support training certificates have the following specific limitations.

A certificate for Service and Support training is only valid while employed by Agilent Technologies or while working as an Agilent-authorized service provider, through which the service employee has ongoing access to Agilent's Safety Alerts, Service Notes, internal technical updates, update training, course development, technical support, course parts, and parts updates. Completion of training alone, without being employed by Agilent Technologies, does not qualify an individual to safely install, service or maintain Agilent products.

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 21 / 34

General

Document Name: Tune reports

Tune Report

Operator Name: Siddhant Khat

AcqDate Batch: C:\Agilent\ICPMS\HW\Tune\Tune_79023


Acq. Date-Time: 2021-09-30 14:44:58

Report Comment: OD 30 Sec 7021

Instrument Name: GS453A-JP15471169

He Gas

Sensitivity



Mass	Range	Count	100%	Background
7	1000	1428	0.62	0.228
88	10000	10716	1.425	1.366
202	10000	20377	5.315	5.155

Sampling Period [sec]: 0.311
Integration Time [sec]: 0.1

Oxidizer/Double Charged Ratio

Grade: 156 / 140: 1.047 %
Doubly Charged: 70 / 140: 1.482 %

Resolution/Auto



Mass	Peak Height	Abs	10-50%	90-10%
7	1471.48	1.26	0.82	0.718
88	14714.82	14.55	0.77	0.796
202	22514.12	125.00	0.12	0.713

Integration Time [sec]: 0.1
Acquisition Time [sec]: 22.34
Y Axis: Linear

Tune Parameters

Plasma Mode	—	He Gas Flow	1.00 L/min	Makeup Gas	0.10 L/min
RF Power	150 W	Optim Gas	—	Auxiliary Gas	0.90 L/min
RF Matching	1.18 V	Nebulizer P. Mode	0.10 rpm	Plasma Gas	15.0 L/min
Sample Depth	9.0 mm	S/C Temp	2°C		

Lens Parameters		Omega Lens	9.1 V	Deflect	13.0 V
Extract 1	8.0 V	Cell Entrance	-30 V	Plate Bias	-35 V
Extract 2	-205.0 V	Cell Exit	-50 V		
Omega Bias	-80 V				

Cell Parameters		Set Gas Flow	—	Energy Discrimination	5.0 V
Use Gas	No	OCF Bias	-8.0 V		
He Flow	0.0 mL/min				

1 of 3

2021-09-30 2:44 PM

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 22 / 34

Document Name: Tune reports

Tune Report


H2 Flow	0.0 mL/min	OCF RF	190 V
OP Parameters			
Vari Gas	154	Aux Gas	0.900
Mass Offset	125	Aux Offset	0.01
OP Bias			-3.0 V

Hardware Settings

Torch H	-0.3 mm	Torch V	0.1 mm
EM			
Discharge	4.0 mV	Analog HV	2247 V
Plate HV			1318 V

ICP

Sensitivity



Mass	Range	Count	100%	Background
59	1000	1453	0.402	0.402
88	10000	10517	0.826	0.793
202	10000	10815	2.445	0.713

Sampling Period [sec]: 0.31
Integration Time [sec]: 0.1

Oxidizer/Double Charged Ratio

Grade: 156 / 140: 0.824 %
Doubly Charged: 70 / 140: 1.025 %

Tune Parameters

Plasma Mode	—	Nebulizer Gas	1.00 L/min	Makeup Gas	0.10 L/min
RF Power	150 W	Optim Gas	—	Auxiliary Gas	0.90 L/min
RF Matching	1.18 V	Nebulizer Pump	0.10 rpm	Plasma Gas	15.0 L/min
Sample Depth	9.0 mm	S/C Temp	2°C		

Lens Parameters		Omega Lens	9.0 V	Deflect	13.0 V
Extract 1	0.0 V	Cell Entrance	-30 V	Plate Bias	-35 V
Extract 2	-210.0 V	Cell Exit	-60 V		
Omega Bias	-105 V				

Cell Parameters		Set Gas Flow	—	Energy Discrimination	5.0 V
Use Gas	Yes	OCF Bias	-22.5 V		
He Flow	0.0 mL/min	OCF RF	200 V		
H2 Flow	0.0 mL/min				

OP Parameters		Aux Gas	0.900	OP Bias	-18.5 V
Vari Gas	124	Aux Offset	0.01		
Mass Offset	125				

Hardware Settings

Torch H	-0.3 mm	Torch V	0.1 mm
---------	---------	---------	--------

2 of 3


2021-09-30 2:44 PM

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 23 / 34

Document Name:

iTune reports




Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 24 / 34

General

Document Name:

Test Report




Date: September 30, 2021 4:07:18 PM
System ID: JP15471163

Page 25 / 34

Document Name:

(Test Report




Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 28 / 34

General

Document Name:

Test Report



Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 27 / 34

Document Name: Test Report

Batch Summary Report

Analyte Table	
Sample Name	TS
1.85152	2.1300

Page 2 / 2

2021-09-30 15:10:31

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 28 / 34

General

Document Name: Test Report

Batch Summary Report

Batch Folder: D:\Agilent Service\2021-09-30 Sep 2021\20 Main
Analyte File: 20 Main\Batch18
Tune Step: #1 XG G36

Run	Acq. Date/Time	Data File	Sample Name	Time	Level	Division
1	2021-09-30 15:17:44	2014-e	20 Main	Sample		10000

Page 1 / 2

2021-09-30 15:46:42

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 29 / 34

Document Name: Test Report

Batch Summary Report

Analyte Table	
Sample Name	TS
1.85152	2.1300

Page 2 / 2

2021-09-30 15:44:43

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 30 / 34

Electronic Signature

Purpose

This signature page was created and published because the ACE sign-off action was executed, which is valid for the entire document, including attachments. The ACE sign-off is an electronic signature that requires two distinct identification components: unique username and personal password. The Agilent representative who has delivered this service understands the meaning and legal status of an electronic signature. As a trained official operator, the Agilent representative has a unique password and login to access ACE and electronically sign this document. (Other e-signatures can be applied to this document using a Document Content Management or other suitable method defined in your data access and control procedures.)

Details

Full Name of Signer: Panthep Kurasathain
Logged On User Name: panthep_kurasathain@agilent.com
Signature Creation Date: September 30, 2021
Reason for Signature: Executed protocol and published this original version of document

Regulatory Disclaimer

This document provides a protocol to verify and record instrument configuration and evidence of proper operation. It has been prepared from our interpretation of applicable regulations as well as industry best practices. The document is designed to provide an important component of a complete compliance package. Validation depends upon many factors and use of this protocol alone does not assure compliance. Agilent Technologies makes no promises or representations as to its sufficiency for any specific regulatory program.

Warranty

Agilent Technologies makes no warranty of any kind to this material, including but not limited to, the implied warranties or merchantability and fitness for a particular purpose. Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 31 / 34

User Name: panthep_kurassathin
Host Name: ASDKWK315
System ID: JP15471169
Print Date: September 30, 2021 4:07:22 PM

ALS OQHW 7909 20Sep21 Transaction log:

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
September 30, 2021 3:50:07 PM	Audit	Session Created	Session	None
September 30, 2021 3:50:07 PM	Start	Configuration	Session	None
September 30, 2021 3:50:07 PM	Audit	Enrollment	Licensing	User is Field Engineer and does not require an unlock code
September 30, 2021 3:52:52 PM	Audit	Epoch loaded	Session	EQP details for primary to (none) (2018) - File path: (Protocol Path) c:\msd\config\protocols\02.50\csm\02.50.a (eq). EQP File Name: (top\02.50.a), EQP Name: (Agilent Recommended)
September 30, 2021 3:52:54 PM	End	Configuration	Session	None
September 30, 2021 3:52:57 PM	Start	Qualification	Session	OQ
September 30, 2021 3:52:57 PM	Start	Execution	Autosampler Check: SP54	None
September 30, 2021 3:53:03 PM	End	Execution	Autosampler Check: SP54	Run Count: 1
September 30, 2021 3:53:04 PM	Start	Execution	Integrated Sample Introduction System (ISIS) Check: (ISIS)	None
September 30, 2021 3:53:06 PM	End	Execution	Integrated Sample Introduction System (ISIS) Check: (ISIS)	Run Count: 1

Page 1 / 3

User Name: panthep_kurassathin
Host Name: ASDKWK315
System ID: JP15471169
Print Date: September 30, 2021 4:07:22 PM

ALS OQHW 7909 20Sep21 Transaction log:

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
September 30, 2021 3:53:10 PM	Start	Execution	Autoburn: G8403A: Autoburn 1	None
September 30, 2021 3:55:08 PM	End	Execution	Autoburn: G8403A: Autoburn 1	Run Count: 1
September 30, 2021 3:55:12 PM	Start	Execution	Background (No Gas Mode): G8403A: No Gas Mode Background 1	None
September 30, 2021 3:55:40 PM	End	Execution	Background (No Gas Mode): G8403A: No Gas Mode Background 1	Run Count: 1
September 30, 2021 3:56:43 PM	Start	Execution	Background (Gas Mode): G8403A: Gas Mode Background Helium	None
September 30, 2021 3:56:17 PM	End	Execution	Background (Gas Mode): G8403A: Gas Mode Background Helium	Run Count: 1
September 30, 2021 3:56:16 PM	Start	Execution	Background (Gas Mode): G8403A: Gas Mode Background Hydrogen	None
September 30, 2021 3:56:36 PM	End	Execution	Background (Gas Mode): G8403A: Gas Mode Background Hydrogen	Run Count: 1
September 30, 2021 3:56:41 PM	Start	Execution	20-Minute Stability (No Gas Mode): G8403A: 20-Minute Stability (No Gas Mode) 1	None
September 30, 2021 3:57:22 PM	End	Execution	20-Minute Stability (No Gas Mode): G8403A: 20-Minute Stability (No Gas Mode) 1	Run Count: 1
September 30, 2021 3:57:24 PM	End	Qualification	Session	OQ
September 30, 2021 3:57:24 PM	Start	Reporting	Session	None

Page 2 / 3

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 32 / 34

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 33 / 34

User Name: panthep_kurassathin
Host Name: ASDKWK315
System ID: JP15471169
Print Date: September 30, 2021 4:07:22 PM

ALS OQHW 7909 20Sep21 Transaction log:

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
September 30, 2021 4:03:07 PM	Audit	Reporting	Session	Report Generated: Certificate
September 30, 2021 4:03:17 PM	Audit	Reporting	Session	Report Generated: Report
September 30, 2021 4:03:59 PM	Start	Qualification	Session	OQ
September 30, 2021 4:04:08 PM	End	Qualification	Session	OQ
September 30, 2021 4:04:08 PM	Start	Reporting	Session	None
September 30, 2021 4:04:28 PM	Audit	Reporting	Session	Report Generated: Certificate
September 30, 2021 4:04:36 PM	Audit	Reporting	Session	Report Generated: Report

Page 3 / 3



Metrological Center

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhoei, Saraburi 18110

Telephone : +66 2 586 5792-4 Fax : +66 2 586 5109

Website : www.scieco.co.th E-Mail : calibrate@scg.co.th

Certificate No. T220730

Page 1 of 6

Certificate of Calibration

Equipment : HEATING BLOCK

Manufacturer : Environmental Express

Model : SC 196

Serial No. : 6974CECW3285

Customer Code : BKK_EL0054

ID No. : T5306A3


Customer : ALS Laboratory Group (Thailand) Co.,Ltd.

104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Phatthanakan,
Khet Suan Luang, Bangkok 10250



Customer Location : Acid Digestion Lab

Date of Receipt : 30 March 2022

Calibrated By : Watcharapon Sangtong (Technician)

Approved By :  / Sujjar Naknakred (Site Calibration Manager)

Date of Issue : 12 APR 2022

REVIEW BY	
APPROVED BY	
NEXT CAL. DATE	7/10/23

The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standard laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Metrological Center.

Date: September 30, 2021 4:07:18 PM
System ID: JP15471169

Page 34 / 34



Certificate No. T220730

Page 2 of 6

Calibration Report

Equipment : HEATING BLOCK
Date of Calibration : 7 April 2022
Environment : Temperature : 21.8-23.1 °C
Line Voltage : 221.6-226.3 V
Relative Humidity : 55-65 %RH

Condition of this results of calibration :

1. This equipment was calibrated by insert nine standard thermocouples type T into its chamber , the other one standard thermocouples type T use for ambient temperature measurement . The calibration was done in according to WI-T20.

All data show below were final values and the initial data from customer request . The temperature scale used was based on ITS - 90 .

2. Reference Standard Instrument :

Instrument	Model	Instrument No.	Certificate No.	Due Date
TC	TYPE T	TN221-TN230	T210008	08 June 2022
TC	TYPE T	TN231-TN240	T210008	08 June 2022
DATA LOGGER	34970A	T149	T210008	08 June 2022

3. This certificate is traceable to :

National Institute of Metrology (Thailand) through Metrological Center (NSC-TISI-TIS 17025 CALIBRATION 0244.)

4. Condition of calibrated item : good

Equipment Description :

Time Constant 2 Hour 25 Minute At 95 °C
Fresh Air Damper ☐ Open ☐ Min ☐ Medium ☐ Max
☐ Close
☒ Not Available

5. Adjustment :


() without adjustment (X) after adjustment

Approved By. _____

Certificate No. T220730

Page 3 of 6

Calibration Report



FRONT CONTROL

Approved By. _____

FM-L13 108/30-05-57

FM-L13 108/30-05-57



Certificate No. T220730

Page 4 of 6

Calibration Report

Measurement Results		Average Standard Reading at each position (°C)						
Calibration Point		TN221	TN222	TN223	TN224	TN225	TN226	
R1 Hole1-Hole6	CAL POINT	Max	93.60	93.82	94.05	94.20	94.36	94.26
		Min	93.07	93.26	93.51	93.66	93.82	93.71
	Average	93.33	93.54	93.78	93.93	94.09	94.26	94.09
R2 Hole7-Hole12	CAL POINT	Max	94.59	94.79	94.63	94.55	94.82	95.00
		Min	94.05	94.25	94.08	93.97	94.26	94.44
	Average	94.32	94.52	94.36	94.26	94.54	94.72	94.72
R3 Hole13-Hole18	CAL POINT	Max	95.03	94.54	94.78	94.84	95.06	94.73
		Min	94.46	93.98	94.20	94.28	94.49	94.18
	Average	94.74	94.26	94.49	94.56	94.78	94.84	94.45
R4 Hole19-Hole24	CAL POINT	Max	94.89	94.82	95.73	95.85	95.73	96.10
		Min	94.33	94.26	95.51	95.62	95.51	95.85
	Average	94.61	94.54	95.62	95.73	95.62	95.97	95.97
R5 Hole25-Hole30	CAL POINT	Max	96.28	96.39	96.37	96.54	96.19	96.04
		Min	95.01	96.10	95.02	96.20	95.89	95.71
	Average	96.15	96.24	96.20	96.37	96.04	95.88	95.88
R6 Hole31-Hole36	CAL POINT	Max	96.84	96.97	97.03	96.48	96.33	95.76
		Min	96.53	96.65	96.71	96.08	95.98	95.43
	Average	96.68	96.81	96.87	96.28	96.16	95.69	95.69
R7 Hole37-Hole42	CAL POINT	Max	96.46	96.15	96.19	96.06	96.95	97.09
		Min	96.13	95.84	95.85	95.72	96.04	96.78
	Average	96.30	95.99	96.02	95.89	96.80	96.93	96.93
R8 Hole43-Hole48	CAL POINT	Max	96.91	96.58	96.13	96.19	96.34	96.19
		Min	96.55	96.21	95.80	95.87	96.03	95.88
	Average	96.73	96.40	95.96	96.03	96.18	96.03	96.03

Approved By. _____

FM-L13 108/30-05-57

Certificate No. T220730

Page 5 of 6

Calibration Report

Measurement Results		Average Standard Reading at each position (°C)						
Calibration Point		TN221	TN222	TN223	TN224	TN225	TN226	
R1 Hole1-Hole6	CAL POINT	Max	104.47	104.65	104.79	105.31	105.47	105.46
		Min	104.15	104.27	104.45	104.98	105.14	105.20
	Average	104.31	104.46	104.62	105.15	105.31	105.33	105.33
R2 Hole7-Hole12	CAL POINT	Max	105.55	105.73	105.65	105.84	105.97	106.07
		Min	105.28	105.43	105.35	105.52	105.68	105.83
	Average	105.42	105.58	105.50	105.68	105.82	105.95	105.95
R3 Hole13-Hole18	CAL POINT	Max	106.14	106.06	105.81	106.05	105.81	105.87
		Min	105.85	105.81	105.55	105.80	105.53	105.64
	Average	106.00	105.94	105.68	105.92	105.67	105.75	105.75
R4 Hole19-Hole24	CAL POINT	Max	105.86	105.60	104.44	104.51	104.28	104.78
		Min	105.61	105.37	104.27	104.35	104.12	104.61
	Average	105.74	105.48	104.35	104.43	104.20	104.69	104.69
R5 Hole25-Hole30	CAL POINT	Max	104.94	104.93	104.97	105.08	104.68	104.69
		Min	104.77	104.75	104.76	104.90	104.51	104.49
	Average	104.85	104.84	104.86	104.99	104.60	104.59	104.59
R6 Hole31-Hole36	CAL POINT	Max	105.44	105.45	105.61	104.95	104.84	104.42
		Min	105.27	105.27	105.44	104.76	104.66	104.25
	Average	105.36	105.36	105.53	104.86	104.75	104.75	104.33
R7 Hole37-Hole42	CAL POINT	Max	105.17	104.70	104.59	104.51	105.22	105.23
		Min	105.00	104.53	104.41	104.35	105.04	105.37
	Average	105.08	104.62	104.50	104.43	105.13	105.45	105.45
R8 Hole43-Hole48	CAL POINT	Max	105.61	105.45	105.10	104.77	104.87	105.02
		Min	105.44	105.28	104.92	104.60	104.70	104.83
	Average	105.53	105.37	105.01	104.69	104.79	104.93	104.93

Approved By. _____

FM-L13 108/30-05-57

Certificate No. T220730

Page 5 of 6

Calibration Report

Measurement Results:

HEATING BLOCK			Temperature Distribution	
Setting (°C)	Reading (°C)		Stability (±°C)	Uncertainty (±°C)
	Min, Max	Average		
100.0	100.0, 100.4	100.1	0.29	0.83
105.0	105.0, 105.4	105.1	0.20	0.79

* The quoted uncertainty exclude " uniformity "

The calibration result apply only the above calibrated item.

The result of test was found accurate as shown on date and place of test only.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k which for a t-distribution, providing a level of confidence of approximately 95 % .

Approved By: _____

Certificate No. T221644

Page 1 of 4

Certificate of Calibration

Equipment : Chamber (Cold Room)

Manufacturer : KOLDTECH

Model : KM 320

Serial No. : TBN-1012061/05

Customer Code : BKK_EN0167

ID No. : T2463A3

Customer : ALS Laboratory Group (Thailand) Co.,Ltd.

104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Phatthanakan,
Khet Suan Luang, Bangkok 10250

Customer Location : Environmental Laboratory

Date of Receipt : 27 June 2022

Calibrated By : Sujjar Nakhakred (Site Calibration Manager)

Approved By : _____ / Boonchai Suriyawong (Site Calibration Manager)

Date of Issue : 04 JUL 2022

The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standard laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Metrological Center.

FM-L13 108/30-05-57

FM-L14 117/01-02-64

Certificate No. T221644

Page 2 of 4

Calibration Report

Equipment : Chamber (Cold Room)
Date of Calibration : 30 June - 1 July 2022
Environment : Temperature : 18.9-23.7 °C
Line Voltage : 222.9-226.5 V
Relative Humidity : 55 - 65 %RH

Condition of this results of calibration :

- This equipment was calibrated by insert nine standard thermocouples type T into its chamber , the other one standard thermocouples type T use for ambient temperature measurement . The calibration was done in according to WI-T20 (based on ASTM E145-94 (Reapproved 2001) and AS2853-1986). All data show below were final values and the initial data from customer request . The temperature scale used was based on ITS - 90 .
- Reference Standard Instrument :


Instrument	Model	Instrument No.	Certificate No.	Due Date
TC	TYPE T	TN161-TN170	T210009	30 July 2022
TC	TYPE T	TN171-TN180	T210009	30 July 2022
DATA LOGGER	34970A	T149	T210009	30 July 2022
- This certificate is traceable to :
National Institute of Metrology (Thailand) through Metrological Center (NSC-TISI-TIS 17025 CALIBRATION 0244.)
- Condition of calibrated item : good
Equipment Description :
Time Constant : 3 Hour - Minute At 3 °C
Fresh Air Damper : ☐ Open ☐ Min ☐ Medium ☐ Max
☐ Close
☒ Not Available
- Adjustment :
() without adjustment (X) after adjustment

Approved By: _____

Certificate No. T221644

Page 3 of 4

Calibration Report



1C =	TN161
2A =	TN162
3A =	TN163
4F =	TN164
5A =	TN165
6A =	TN166
7F =	TN167
8F =	TN168
9A =	TN169
10A =	TN170

11F =	TN171
12F =	TN172
13A =	TN173
14A =	TN174
15F =	TN175
16E =	TN176

Approved By: _____

FM-L15 117/15-05-63

FM-L15 117/15-05-63