

ภาคผนวก

ภาคผนวก ก	มาตรฐานน้ำทิ้งจากอาคาร
ภาคผนวก ข	เอกสารการป้องกันและระงับอัคคีภัย
ภาคผนวก ค	อุปกรณ์เก็บตัวอย่างและเครื่องมือตรวจวัด
ภาคผนวก ง	ใบรายงานผลการตรวจวิเคราะห์
ภาคผนวก จ	เอกสารสอบเทียบเครื่องมือ
ภาคผนวก ฉ	หนังสืออนุญาตขึ้นทะเบียน ห้องปฏิบัติการวิเคราะห์เอกชน

ภาคผนวก ก
มาตรฐานน้ำทิ้งจากอาคาร



ประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม

เรื่อง กำหนดมาตรฐานควบคุมการระบายน้ำทิ้ง
จากอาคารบางประเภทและบางขนาด

โดยที่ได้มีการปฏิรูประบบราชการ โดยให้มีการจัดตั้งกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อมขึ้นมา และให้อิโณการกิจของกระทรวงวิทยาศาสตร์ เทคโนโลยีและสิ่งแวดล้อม ในส่วนที่เกี่ยวข้องกับพระราชบัญญัติส่งเสริมและรักษาคุณภาพสิ่งแวดล้อมแห่งชาติ พ.ศ. ๒๕๓๕ ไปเป็นของกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม ประกอบกันเป็นการสมควรให้คณะกรรมการควบคุมลพิษเป็นผู้พิจารณาเห็นชอบกับวิธีการตรวจหาค่ามาตรฐานการระบายน้ำทิ้ง นอกเหนือจากวิธีการที่กำหนดไว้ เทนกรรมควบคุมลพิษ จึงสมควรแก้ไขปรับปรุงประกาศกระทรวงวิทยาศาสตร์ เทคโนโลยีและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุมการระบายน้ำทิ้งจากอาคารบางประเภทและบางขนาด

อาศัยอำนาจตามความในมาตรา ๕๕ แห่งพระราชบัญญัติส่งเสริมและรักษาคุณภาพสิ่งแวดล้อมแห่งชาติ พ.ศ. ๒๕๓๕ แก้ไขโดยมาตรา ๑๑๔ แห่งพระราชกฤษฎีกาแก้ไขบทบัญญัติให้สอดคล้องกับการโอนอำนาจหน้าที่ของส่วนราชการ ให้เป็นไปตามพระราชบัญญัติปรับปรุงกระทรวง ทบวง กรม พ.ศ. ๒๕๔๕ พ.ศ. ๒๕๔๕ อันเป็นพระราชบัญญัติที่มีบทบัญญัติบางประการเกี่ยวกับการจำกัดสิทธิและเสรีภาพของบุคคล ซึ่งมาตรา ๒๕ ประกอบกับมาตรา ๓๕ มาตรา ๔๔ มาตรา ๕๐ และมาตรา ๕๑ ของรัฐธรรมนูญแห่งราชอาณาจักรไทยบัญญัติให้กระทำได้ โดยอาศัยอำนาจตามบทบัญญัติแห่งกฎหมาย รัฐมนตรีว่าการกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม โดยคณะนัของคณะกรรมการควบคุมลพิษ และโดยความเห็นชอบของคณะกรรมการสิ่งแวดล้อมแห่งชาติ จึงออกประกาศไว้ ดังต่อไปนี้

ข้อ ๑ ให้ยกเลิกประกาศกระทรวงวิทยาศาสตร์ เทคโนโลยีและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุมการระบายน้ำทิ้งจากอาคารบางประเภทและบางขนาด ลงวันที่ ๑๐ มกราคม พ.ศ. ๒๕๓๗

ข้อ ๒ ในประกาศนี้

“อาคาร” หมายความว่า อาคารที่ก่อสร้างขึ้น ไม่ว่าจะมีลักษณะเป็นอาคารหลังเดียว หรือเป็นกลุ่มของอาคารซึ่งตั้งอยู่ภายในพื้นที่ซึ่งเป็นบริเวณเดียวกัน และไม่ว่าจะมีระยะขนาน้ำท่อเดียว หรือมีหลายท่อ ที่เชื่อมติดต่อกันระหว่างอาคารหรือไม่ก็ตาม ซึ่งได้แก่

(๑) อาคารชุด ตามกฎหมายว่าด้วยอาคารชุด

(๒) โรงแรม ตามกฎหมายว่าด้วยโรงแรม

(๔) อาคารโรงเรียนเอกชน โรงเรียนของทางราชการ สถาบันอุดมศึกษาของเอกชน หรือสถาบันอุดมศึกษาของทางราชการที่มีพื้นที่ใช้สอยรวมกันทุกชั้นของอาคารหรือกลุ่มของอาคารตั้งแต่ ๒๕,๐๐๐ ตารางเมตรขึ้นไป

(๕) อาคารที่ทำการของทางราชการ รัฐวิสาหกิจ องค์การระหว่างประเทศ หรือของเอกชนที่มีพื้นที่ใช้สอยรวมกันทุกชั้นของอาคารหรือกลุ่มของอาคารตั้งแต่ ๕๕,๐๐๐ ตารางเมตรขึ้นไป

(๖) อาคารของศูนย์การค้าหรือห้างสรรพสินค้าที่มีพื้นที่ใช้สอยรวมกันทุกชั้นของอาคารหรือกลุ่มของอาคารตั้งแต่ ๒๕,๐๐๐ ตารางเมตรขึ้นไป

(๗) ตลาดที่มีพื้นที่ใช้สอยรวมกันทุกชั้นของอาคารหรือกลุ่มของอาคารตั้งแต่ ๒,๕๐๐ ตารางเมตรขึ้นไป

(๘) กัดาคารหรือร้านอาหารที่มีพื้นที่ให้บริการรวมกันทุกชั้นของอาคารหรือกลุ่มของอาคารตั้งแต่ ๒,๕๐๐ ตารางเมตรขึ้นไป

ข้อ ๕ อาคารประเภท ข. หมายความว่า อาคารดังต่อไปนี้

(๑) อาคารชุดที่มีจำนวนห้องสำหรับใช้เป็นที่อยู่อาศัยรวมกันทุกชั้นของอาคาร หรือกลุ่มของอาคารตั้งแต่ ๑๐๐ ห้องนอน แต่ไม่ถึง ๕๐๐ ห้องนอน

(๒) โรงแรมที่มีจำนวนห้องสำหรับใช้เป็นห้องพักรวมกันทุกชั้นของอาคาร หรือกลุ่มของอาคารตั้งแต่ ๖๐ ห้อง แต่ไม่ถึง ๒๐๐ ห้อง

(๓) หอพักที่มีจำนวนห้องสำหรับใช้เป็นที่อยู่อาศัยรวมกันทุกชั้นของอาคาร หรือกลุ่มของอาคารตั้งแต่ ๒๕๐ ห้องขึ้นไป

(๔) สถานบริการที่มีพื้นที่ใช้สอยรวมกันทุกชั้นของอาคาร หรือกลุ่มของอาคารตั้งแต่ ๕,๐๐๐ ตารางเมตรขึ้นไป

(๕) โรงพยาบาลของทางราชการ รัฐวิสาหกิจ หรือสถานพยาบาล ตามกฎหมายว่าด้วยสถานพยาบาล ที่มีเตียงสำหรับผู้ป่วยไว้ค้างคืนรวมกันทุกชั้นของอาคารหรือกลุ่มของอาคารตั้งแต่ ๑๐ เตียง แต่ไม่ถึง ๓๐ เตียง

(๖) อาคารโรงเรียนเอกชน โรงเรียนของทางราชการ สถาบันอุดมศึกษาของเอกชน หรือสถาบันอุดมศึกษาของทางราชการที่มีพื้นที่ใช้สอยรวมกันทุกชั้นของอาคารหรือกลุ่มของอาคารตั้งแต่ ๕,๐๐๐ ตารางเมตร แต่ไม่ถึง ๒๕,๐๐๐ ตารางเมตร

(๓) หอพัก ตามกฎหมายว่าด้วยหอพัก

(๔) สถานบริการประเภทสถานอาบน้ำ นวดหรืออบตัว ซึ่งมีผู้ให้บริการแก่ลูกค้า ตามกฎหมายว่าด้วยสถานบริการ

(๕) โรงพยาบาลของทางราชการหรือสถานพยาบาล ตามกฎหมายว่าด้วยสถานพยาบาล

(๖) อาคารโรงเรียนเอกชน ตามกฎหมายว่าด้วยโรงเรียนเอกชน โรงเรียนของทางราชการ อาคารสถาบันอุดมศึกษาของเอกชน ตามกฎหมายว่าด้วยสถาบันอุดมศึกษาของเอกชนและสถาบันอุดมศึกษาของทางราชการ

(๗) อาคารที่ทำการของทางราชการ รัฐวิสาหกิจ หรือองค์การระหว่างประเทศและของเอกชน

(๘) อาคารของศูนย์การค้าหรือห้างสรรพสินค้า

(๙) ตลาด ตามกฎหมายว่าด้วยการสาธารณสุข แต่ไม่รวมถึง ท่าเทียบเรือประมง สะพานปลา หรือกิจการแพปลา

(๑๐) กัดาคารหรือร้านอาหาร

“น้ำทิ้ง” หมายความว่า น้ำเสียที่ผ่านระบบบำบัดน้ำเสียแล้วจนเป็นไปตามมาตรฐานควบคุมการระบายน้ำทิ้งตามที่กำหนดไว้ในประกาศนี้

ข้อ ๓ ให้แบ่งประเภทของอาคารตามข้อ ๒ ออกเป็น ๕ ประเภท คือ

(๑) อาคารประเภท ก.

(๒) อาคารประเภท ข.

(๓) อาคารประเภท ค.

(๔) อาคารประเภท ง.

(๕) อาคารประเภท จ.

ข้อ ๔ อาคารประเภท ก. หมายความว่า อาคารดังต่อไปนี้

(๑) อาคารชุดที่มีจำนวนห้องสำหรับใช้เป็นที่อยู่อาศัยรวมกันทุกชั้นของอาคาร หรือกลุ่มของอาคารตั้งแต่ ๕๐๐ ห้องนอนขึ้นไป

(๒) โรงแรมที่มีจำนวนห้องสำหรับใช้เป็นห้องพักรวมกันทุกชั้นของอาคาร หรือกลุ่มของอาคารตั้งแต่ ๒๐๐ ห้องขึ้นไป

(๓) โรงพยาบาลของทางราชการ รัฐวิสาหกิจหรือสถานพยาบาล ตามกฎหมายว่าด้วยสถานพยาบาล ที่มีเตียงสำหรับผู้ป่วยไว้ค้างคืนรวมกันทุกชั้นของอาคารหรือกลุ่มของอาคารตั้งแต่ ๓๐ เตียงขึ้นไป

(๗) อาคารที่ทำการของทางราชการ รัฐวิสาหกิจ องค์การระหว่างประเทศ หรือของเอกชนที่มีพื้นที่ใช้สอยรวมกันทุกชั้นของอาคารหรือกลุ่มของอาคารตั้งแต่ ๑๐,๐๐๐ ตารางเมตร แต่ไม่ถึง ๕๕,๐๐๐ ตารางเมตร

(๘) อาคารของศูนย์การค้าหรือห้างสรรพสินค้าที่มีพื้นที่ใช้สอยรวมกันทุกชั้นของอาคารหรือกลุ่มของอาคารตั้งแต่ ๕,๐๐๐ ตารางเมตร แต่ไม่ถึง ๒๕,๐๐๐ ตารางเมตร

(๙) ตลาดที่มีพื้นที่ใช้สอยรวมกันทุกชั้นของอาคารหรือกลุ่มของอาคารตั้งแต่ ๑,๕๐๐ ตารางเมตร แต่ไม่ถึง ๒,๕๐๐ ตารางเมตร

(๑๐) กัดาคารหรือร้านอาหารที่มีพื้นที่ให้บริการรวมกันทุกชั้นของอาคารหรือกลุ่มของอาคารตั้งแต่ ๕๐๐ ตารางเมตร แต่ไม่ถึง ๒,๕๐๐ ตารางเมตร

ข้อ ๖ อาคารประเภท ค. หมายความว่า อาคารดังต่อไปนี้

(๑) อาคารชุดที่มีจำนวนห้องสำหรับใช้เป็นที่อยู่อาศัยรวมกันทุกชั้นของอาคาร หรือกลุ่มของอาคารไม่ถึง ๑๐๐ ห้องนอน

(๒) โรงแรมที่มีจำนวนห้องสำหรับใช้เป็นห้องพักรวมกันทุกชั้นของอาคาร หรือกลุ่มของอาคารไม่ถึง ๖๐ ห้อง

(๓) หอพักที่มีจำนวนห้องสำหรับใช้เป็นที่อยู่อาศัยรวมกันทุกชั้นของอาคาร หรือกลุ่มของอาคารตั้งแต่ ๕๐ ห้อง แต่ไม่ถึง ๒๕๐ ห้อง

(๔) สถานบริการที่มีพื้นที่ใช้สอยรวมกันทุกชั้นของอาคาร หรือกลุ่มของอาคารตั้งแต่ ๑,๐๐๐ ตารางเมตร แต่ไม่ถึง ๕,๐๐๐ ตารางเมตร

(๕) อาคารที่ทำการของทางราชการ รัฐวิสาหกิจ องค์การระหว่างประเทศ หรือของเอกชนที่มีพื้นที่ใช้สอยรวมกันทุกชั้นของอาคารหรือกลุ่มของอาคารตั้งแต่ ๕,๐๐๐ ตารางเมตร แต่ไม่ถึง ๑๐,๐๐๐ ตารางเมตร

(๖) ตลาดที่มีพื้นที่ใช้สอยรวมกันทุกชั้นของอาคารหรือกลุ่มของอาคารตั้งแต่ ๑,๐๐๐ ตารางเมตร แต่ไม่ถึง ๑,๕๐๐ ตารางเมตร

(๗) กัดาคารหรือร้านอาหารที่มีพื้นที่ให้บริการรวมกันทุกชั้นของอาคาร หรือกลุ่มของอาคารตั้งแต่ ๒๕๐ ตารางเมตร แต่ไม่ถึง ๕๐๐ ตารางเมตร

ข้อ ๗ อาคารประเภท ง. หมายความว่า อาคารดังต่อไปนี้

(๑) หอยพิคที่มีจำนวนห้องสำหรับใช้เป็นที่อยู่อาศัยร่วมกันทุกชั้นของอาคาร หรือกลุ่มของอาคาร ตั้งแต่ ๑๐ ห้อง แต่ไม่ถึง ๕๐ ห้อง

(๒) ตลาดที่มีพื้นที่ใช้สอยรวมกันทุกชั้นของอาคาร หรือกลุ่มของอาคารตั้งแต่ ๕๐๐ ตารางเมตร แต่ไม่ถึง ๑,๐๐๐ ตารางเมตร

(๓) กัดดาการหรือร้านอาหารที่มีพื้นที่ให้บริการรวมกันทุกชั้นของอาคาร หรือกลุ่มของอาคาร ตั้งแต่ ๑๐๐ ตารางเมตร แต่ไม่ถึง ๒๕๐ ตารางเมตร

ข้อ ๘ อาคารประเภท จ. หมายความว่า กัดดาการหรือร้านอาหารที่มีพื้นที่ให้บริการรวมกันทุกชั้นไม่ถึง ๑๐๐ ตารางเมตร

ข้อ ๙ มาตรฐานควบคุมการระบายน้ำทิ้งจากอาคาร ประเภท ก. ต้องมีค่าดังต่อไปนี้

(๑) ความเป็นกรดและด่าง (PH) ต้องมีค่าระหว่าง ๕-๘

(๒) บีโอดี (BOD) ต้องมีค่าไม่เกิน ๒๐ มิลลิกรัมต่อลิตร

(๓) สารแขวนลอย (Suspended Solids) ต้องมีค่าไม่เกิน ๓๐ มิลลิกรัมต่อลิตร

(๔) ซัลไฟด์ (Sulfide) ต้องมีค่าไม่เกิน ๑.๐ มิลลิกรัมต่อลิตร

(๕) สารที่ละลายได้ทั้งหมด (Total Dissolved Solids) ต้องมีค่าเพิ่มขึ้นจากปริมาณสารละลายในน้ำใช้ตามปกติไม่เกิน ๕๐๐ มิลลิกรัมต่อลิตร

(๖) ตะกอนหนัก (Settleable Solids) ต้องมีค่าไม่เกิน ๐.๕ มิลลิตรต่อลิตร

(๗) น้ำมันและไขมัน (Fat Oil and Grease) ต้องมีค่าไม่เกิน ๒๐ มิลลิกรัมต่อลิตร

(๘) ทีเคเอ็น (TKN) ต้องมีค่าไม่เกิน ๓๕ มิลลิกรัมต่อลิตร

ข้อ ๑๐ มาตรฐานควบคุมการระบายน้ำทิ้งจากอาคาร ประเภท ข. ต้องเป็นไปตามข้อ ๘ เว้นแต่

(๑) บีโอดี ต้องมีค่าไม่เกิน ๓๐ มิลลิกรัมต่อลิตร

(๒) สารแขวนลอย ต้องมีค่าไม่เกิน ๔๐ มิลลิกรัมต่อลิตร

ข้อ ๑๑ มาตรฐานควบคุมการระบายน้ำทิ้งจากอาคาร ประเภท ก. ต้องเป็นไปตามข้อ ๘ เว้นแต่

(๑) บีโอดี ต้องมีค่าไม่เกิน ๔๐ มิลลิกรัมต่อลิตร

(๒) สารแขวนลอย ต้องมีค่าไม่เกิน ๕๐ มิลลิกรัมต่อลิตร

(๑) การตรวจสอบค่าน้ำมันและไขมันให้กระทำโดยใช้วิธีการสกัดด้วยตัวทำละลาย แอลกอฮอล์ น้ำหนักของน้ำมันและไขมัน

(๘) การตรวจสอบค่าทีเคเอ็นให้กระทำโดยใช้วิธีการเจลดาค์ด (Kjeldahl)

ข้อ ๑๕ การคิดคำนวณพื้นที่ใช้สอย จำนวนอาคารและจำนวนห้องของอาคาร หรือกลุ่มของอาคาร ให้เป็นไปตามวิธีการที่คณะกรรมการควบคุมมลพิษกำหนด โดยประกาศในราชกิจจานุเบกษา

ข้อ ๑๖ วิธีการเก็บตัวอย่างน้ำ ความถี่ และระยะเวลาในการเก็บตัวอย่างน้ำ ให้เป็นไปตามที่คณะกรรมการควบคุมมลพิษกำหนด โดยประกาศในราชกิจจานุเบกษา

ข้อ ๑๗ ประกาศนี้ให้ใช้บังคับตั้งแต่วันถัดจากวันประกาศในราชกิจจานุเบกษาเป็นต้นไป

ประกาศ ณ วันที่ ๑๗ พฤศจิกายน พ.ศ. ๒๕๔๔

ยงยุทธ ติยะไพรัช

รัฐมนตรีว่าการกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม

(๓) ซัลไฟด์ ต้องมีค่าไม่เกิน ๓.๐ มิลลิกรัมต่อลิตร

(๔) ค่าทีเคเอ็น ต้องมีค่าไม่เกิน ๔๐ มิลลิกรัมต่อลิตร

ข้อ ๑๒ มาตรฐานควบคุมการระบายน้ำทิ้งจากอาคาร ประเภท ง. ต้องเป็นไปตามข้อ ๘ เว้นแต่

(๑) บีโอดี ต้องมีค่าไม่เกิน ๕๐ มิลลิกรัมต่อลิตร

(๒) สารแขวนลอย ต้องมีค่าไม่เกิน ๕๐ มิลลิกรัมต่อลิตร

(๓) ซัลไฟด์ ต้องมีค่าไม่เกิน ๔.๐ มิลลิกรัมต่อลิตร

(๔) ค่าทีเคเอ็น ต้องมีค่าไม่เกิน ๔๐ มิลลิกรัมต่อลิตร

ข้อ ๑๓ มาตรฐานควบคุมการระบายน้ำทิ้งจากอาคาร ประเภท จ. ต้องมีค่าดังต่อไปนี้

(๑) ความเป็นกรดและด่างต้องมีค่าระหว่าง ๕-๘

(๒) บีโอดี ต้องมีค่าไม่เกิน ๒๐๐ มิลลิกรัมต่อลิตร

(๓) สารแขวนลอย ต้องมีค่าไม่เกิน ๖๐ มิลลิกรัมต่อลิตร

(๔) น้ำมันและไขมัน ต้องมีค่าไม่เกิน ๑๐๐ มิลลิกรัมต่อลิตร

ข้อ ๑๔ การตรวจสอบมาตรฐานการระบายน้ำทิ้งจากอาคาร ให้ใช้วิธีการดังต่อไปนี้

(๑) การตรวจสอบค่าความเป็นกรดและด่างให้กระทำโดยใช้เครื่องวัดความเป็นกรดและด่างของน้ำ (PH Meter)

(๒) การตรวจสอบค่าบีโอดีให้กระทำโดยใช้วิธีการอะไซด์โมดิฟิเคชัน (Azide Modification) ที่อุณหภูมิ ๒๐ องศาเซลเซียส เป็นเวลา ๕ วัน ติดต่อกันหรือวิธีการอื่นที่คณะกรรมการควบคุมมลพิษให้ความเห็นชอบ

(๓) การตรวจสอบค่าสารแขวนลอยให้กระทำโดยใช้วิธีการกรองผ่านกระดาษกรองใยแก้ว (Glass Fibre Filter Disc)

(๔) การตรวจสอบค่าซัลไฟด์ให้กระทำโดยใช้วิธีการไทเตรท (Titrate)

(๕) การตรวจสอบค่าสารที่ละลายได้ทั้งหมดให้กระทำโดยใช้วิธีการระเหยแห้งระหว่างอุณหภูมิ ๑๐๓ องศาเซลเซียส ถึงอุณหภูมิ ๑๐๕ องศาเซลเซียส ในเวลา ๑ ชั่วโมง

(๖) การตรวจสอบค่าตะกอนหนักให้กระทำโดยใช้วิธีการกรวยอิมมอฟฟ์ (Imhoff cone) ขนาดบรรจุ ๑,๐๐๐ ลูกบาศก์เซนติเมตร ในเวลา ๑ ชั่วโมง

ภาคผนวก ข
เอกสารการป้องกันและระงับอัคคีภัย



© 2000 Blackwell Science Ltd
Journal of Internal Medicine 247: 105–112

Downloaded from <http://ajphaphapublications.org/>

[illegible]

1. The first step is to identify the problem or question that needs to be answered. This involves understanding the context and the specific requirements of the task.

Scene Title: ERT Training

Date: 9-10 December 2023

Time: 8.30.18.30 hrs

Department: Fire City Safety

Trainer: Fire Brigade Dept. - Rangsit District

Version: Revised: None

No.	Emp. No.	Unit	Name - Signature	Position	Department	Signature 9 Dec 2023	Signature 10 Dec 2023
1	807009						
2	807010						
3	807011						
4	807012						
5	807013						
6	807014						
7	807015						
8	807016						
9	807017						
10	807018						
11	807019						
12	807020						
13	807021						
14	807022						
15	807023						
16	807024						
17	807025						
18	807026						
19	807027						
20	807028						
21	807029						
22	807030						
23	807031						
24	807032						
25	807033						
26	807034						
27	807035						
28	807036						
29	807037						
30	807038						
31	807039						
32	807040						
33	807041						
34	807042						
35	807043						
36	807044						
37	807045						
38	807046						
39	807047						
40	807048						
41	807049						
42	807050						
43	807051						
44	807052						
45	807053						
46	807054						
47	807055						
48	807056						
49	807057						
50	807058						
51	807059						
52	807060						
53	807061						
54	807062						
55	807063						
56	807064						
57	807065						
58	807066						
59	807067						
60	807068						
61	807069						
62	807070						
63	807071						
64	807072						
65	807073						
66	807074						
67	807075						
68	807076						
69	807077						
70	807078						
71	807079						
72	807080						
73	807081						
74	807082						
75	807083						
76	807084						
77	807085						
78	807086						
79	807087						
80	807088						
81	807089						
82	807090						
83	807091						
84	807092						
85	807093						
86	807094						
87	807095						
88	807096						
89	807097						
90	807098						
91	807099						
92	807100						
93	807101						
94	807102						
95	807103						
96	807104						
97	807105						
98	807106						
99	807107						
100	807108						
101	807109						
102	807110						
103	807111						
104	807112						
105	807113						
106	807114						
107	807115						
108	807116						
109	807117						
110	807118						
111	807119						
112	807120						
113	807121						
114	807122						
115	807123						
116	807124						
117	807125						
118	807126						
119	807127						
120	807128						
121	807129						
122	807130						
123	807131						
124	807132						
125	807133						
126	807134						
127	807135						
128	807136						
129	807137						
130	807138						
131	807139						
132	807140						
133	807141						
134	807142						
135	807143						
136	807144						
137	807145						
138	807146						
139	807147						
140	807148						
141	807149						
142	807150						
143	807151						
144	807152						
145	807153						
146	807154						
147	807155						
148	807156						
149	807157						
150	807158						
151	807159						
152	807160						
153	807161						
154	807162						
155	807163						
156	807164						
157	807165						
158	807166						
159	807167						
160	807168						
161	807169						
162	807170						
163	807171						
164	807172						
165	807173						
166	807174						
167	807175						
168	807176						
169	807177						
170	807178						
171	807179						
172	807180						
173	807181						
174	807182						
175	807183						
176	807184						
177	807185						
178	807186						
179	807187						
180	807188						
181	807189						
182	807190						
183	807191						
184	807192						
185	807193						
186	807194						
187	807195						
188	807196						
189	807197						
190	807198						
191	807199						
192	807200						
193	807201						
194	807202						
195	807203						
196	807204						
197	807205						
198	807206						
199	807207						
200	807208						
201	807209						
202	807210						
203	807211						
204	807212						
205	807213						
206	807214						
207	807215						
208	807216						
209	807217						
210	807218						
211	807219						
212	807220						
213	807221						
214	807222						
215	807223						
216	807224						
217	807225						
218	807226						
219	807227						
220	807228						
221	807229						
222	807230						
223	807231						
224	807232						
225	807233						
226	807234						
227	807235						
228	807236						
229	807237						
230	807238						
231	807239						
232	807240						
233	807241						
234	807242						
235	807243						
236	807244						
237	807245						
238	807246						
239	807247						
240	807248						
241	807249						
242	807250						
243	807251						
244	807252						
245	807253						
246	807254						
247	807255						
248	807256						
249	807257						
250	807258						
251	807259						
252	807260						

807260

807260

807260

807260

807260

807260</

ภาคผนวก ค
อุปกรณ์เก็บตัวอย่างและเครื่องมือตรวจวัด





(ก) เครื่องวัดค่าความเป็นกรด-ด่าง (pH Meter)



(ข) อุปกรณ์เก็บตัวอย่างแบบเข็อก



(ค) ถุงมือยาง



(ง) ภาชนะบรรจุตัวอย่าง

อุปกรณ์เก็บตัวอย่าง และเครื่องมือตรวจวัด

ภาคผนวก ง
ใบรายงานผลการตรวจวิเคราะห์



ANALYSIS REPORT

COMMISSION NAME	TAIPEI 2025
ADDRESS	134, Jinnong Road, Section 4, Neihu District, Taipei City 114, Taiwan
CONTACT INFORMATION	Tel: 886-2-2652-1111 Fax: 886-2-2652-1112 Email: liae@liae.com.tw Website: www.liae.com.tw
ANALYST NAME	LIANG, JIN-CHENG
ANALYST DATE	2024/05/15
ANALYST TIME	10:00 AM
ANALYST BY	LIANG, JIN-CHENG
ANALYST BY	LIANG, JIN-CHENG

ANALYST	TEST	ANALYST NAME	ANALYST DATE	ANALYST TIME	ANALYST BY
LIANG, JIN-CHENG	TAIPEI 2025	LIANG, JIN-CHENG	2024/05/15	10:00 AM	LIANG, JIN-CHENG
LIANG, JIN-CHENG	TAIPEI 2025	LIANG, JIN-CHENG	2024/05/15	10:00 AM	LIANG, JIN-CHENG

134, Jinnong Road, Section 4, Neihu District, Taipei City 114, Taiwan

Page 3

Linkbelt Analytic and Engineering Consultant Co., Ltd.

2024/05/15

1. BY THE TEST PARTIAL IN THIS ANALYSIS REPORT WITHOUT OFFICIAL APPROVAL.

2. ANALYST NAME AND DATE TO SUBMITTED ANALYST NAME.

LIANG, JIN-CHENG

ANALYSIS REPORT

COMMISSION NAME	TAIPEI 2025
ADDRESS	134, Jinnong Road, Section 4, Neihu District, Taipei City 114, Taiwan
CONTACT INFORMATION	Tel: 886-2-2652-1111 Fax: 886-2-2652-1112 Email: liae@liae.com.tw Website: www.liae.com.tw
ANALYST NAME	LIANG, JIN-CHENG
ANALYST DATE	2024/05/15
ANALYST TIME	10:00 AM
ANALYST BY	LIANG, JIN-CHENG
ANALYST BY	LIANG, JIN-CHENG

ANALYST	TEST	ANALYST NAME	ANALYST DATE	ANALYST TIME	ANALYST BY
LIANG, JIN-CHENG	TAIPEI 2025	LIANG, JIN-CHENG	2024/05/15	10:00 AM	LIANG, JIN-CHENG
LIANG, JIN-CHENG	TAIPEI 2025	LIANG, JIN-CHENG	2024/05/15	10:00 AM	LIANG, JIN-CHENG

134, Jinnong Road, Section 4, Neihu District, Taipei City 114, Taiwan

Page 3

Linkbelt Analytic and Engineering Consultant Co., Ltd.

2024/05/15

1. BY THE TEST PARTIAL IN THIS ANALYSIS REPORT WITHOUT OFFICIAL APPROVAL.

2. ANALYST NAME AND DATE TO SUBMITTED ANALYST NAME.

LIANG, JIN-CHENG

1. BY THE TEST PARTIAL IN THIS ANALYSIS REPORT WITHOUT OFFICIAL APPROVAL.

2. ANALYST NAME AND DATE TO SUBMITTED ANALYST NAME.

LIANG, JIN-CHENG

ภาคผนวก จ
เอกสารสอบเทียบเครื่องมือ



ภาคผนวก จ-1
เอกสารสอบเทียบเครื่องมือวิเคราะห์



รายการใบรับรองสอบเทียบ ทวนสอบ เครื่องมือหลักประจำห้องปฏิบัติการวิเคราะห์ สำหรับคุณภาพน้ำ

No.	Instrument/Equipment	Parameter	Manufacturer	Model/Serial No.	Calibrator	Certification No.	Date of Calibration	Due date of Calibration*	Remark
เครื่องมือหลักประจำห้องปฏิบัติการวิเคราะห์สำหรับคุณภาพน้ำ									
1	pH Meter	อุณหภูมิ (Temperature) ความเป็นกรดและด่าง (pH)	Mettler-Toledo	Seven Easy S20 / 1231155210	National Food Institute, Ministry of Industry, Thailand	2201793-001-01	1 Mar 22	1 Mar 23	-
2	pH Meter		Mettler-Toledo	Seven Easy S20 / 1230525212	National Food Institute, Ministry of Industry, Thailand	2202093-001-01	16 Mar 22	16 Mar 23	-
3	UV-VIS Spectrophotometer	ซัลไฟด์ (Sulphide), ซีโอดี (COD)	Agilent Technologies	Cary60 G6860A / MY15410009	DQE Services Co.,Ltd.	SP22-016	31 May 22	30 May 23	
4	UV-VIS Spectrophotometer		Hitachi	U-1900 / 2021-064	DQE Services Co.,Ltd.	SP22-007	20 Jan 22	19 Jan 23	-
5	UV-VIS Spectrophotometer		Hitachi	U-2900 / 21E22-009	DQE Services Co.,Ltd.	SP22-008	20 Jan 22	20 Jan 23	-
6	Analytical Balance (Repeatability 0.01 mg)	สารแขวนลอย (Total Suspended Solids) สารที่ละลายได้ทั้งหมด (TDS)	Mettler-Toledo	XSR205DU / C009071872	Technology Promotion Association (Thailand-Japan)	22MM210	26 Apr 22	25 Apr 23	-
7	Hot Air Oven		Memmert	UF55 / B212.0411	Technology Promotion Association (Thailand-Japan)	22TM304	7 Apr 22	6 Apr 23	-
8	Analytical Balance (Repeatability 0.1 mg)	น้ำมันและไขมัน (Fat,Oil & Grease)	Mettler-Toledo	AB-204S/FACT / 1129361010	National Food Institute, Ministry of Industry, Thailand	2203120-001-01	1 Jun 22	31 May 23	-
9	BOD Incubator	บีโอดี (BOD)	Arco	UC4-1320 / (UAE.WAO.015/2561)	Technology Promotion Association (Thailand-Japan)	22TM90	17 Feb 22	16 Feb 23	-
10	BOD Incubator		Arco	UR-1320 / (UAE.WAO.006/2553)	Technology Promotion Association (Thailand-Japan)	22TM306	7 Apr 22	6 Apr 23	-
11	BOD Incubator		Arco	UR-1320 / (UAE.WAO.018/2551)	Technology Promotion Association (Thailand-Japan)	22TM305	7 Apr 22	7 Apr 23	-
12	COD Reactor (Heating Block)	ซีโอดี (COD)	Hanna	HI839800-02 / 4500052101	Hanna Instruments (Thailand) Ltd.	HIT-2219-0480	9 May 22	8 May 23	-

รายการใบรับรองสอบเทียบ ทวนสอบ เครื่องมือหลักประจำห้องปฏิบัติการวิเคราะห์ สำหรับคุณภาพน้ำ

No.	Instrument/Equipment	Parameter	Manufacturer	Model/Serial No.	Calibrator	Certification No.	Date of Calibration	Due date of Calibration*	Remark
เครื่องมือหลักประจำห้องปฏิบัติการวิเคราะห์สำหรับคุณภาพน้ำ									
13	COD Reactor (Heating Block)	ซีโอดี (COD)	Hanna	HI839800-02 / 4500052101	Hanna Instruments (Thailand) Ltd.	HIT-2219-0480	9 May 22	9 May 23	-
14	Digestor Unit	ทีเคเอ็น (Total Kjeldahl Nitrogen)	FOSS TECATOR	2520auto / 91794469	National Food Institute, Ministry of Industry, Thailand	2202361-001-01	4 Apr 22	4 Apr 23	-
15	Digestor Unit		Velp	DKL20 / 213517	National Food Institute, Ministry of Industry, Thailand	2203368-001-01	23 Jun 22	22 Jun 23	-
16	Distillation Unit (Kjeldahl Method)		FOSS TECATOR	KT200 / 91790524	FOSS South East Asia	5874	30 Nov 21	29 Nov 22	-
17	Distillation Unit (Kjeldahl Method)		FOSS TECATOR	KT8100 / 91889052	FOSS South East Asia	6623	25 Jul 22	24 Jul 23	-
18	Incubator (Cooled Incubator)	แบคทีเรียกลุ่มฟิโคไลโคลิฟอร์ม (Fecal Coliform Bacteria)	Memmert	IPP 260 / V616.0066	Technology Promotion Association (Thailand-Japan)	22TM672	5 May 22	4 May 23	-
19	Incubator (Cooled Incubator)		Memmert	IPP 260 / V618.0033	Technology Promotion Association (Thailand-Japan)	22TM503	3 May 22	2 May 23	-
20	Incubator		Memmert	IF 75 / D317.0305	Technology Promotion Association (Thailand-Japan)	22TM670	3 May 22	3 May 23	-
21	Incubator		Memmert	IN 75 / D317.0307	Technology Promotion Association (Thailand-Japan)	22TM671	3 May 22	3 May 23	-
22	Water Bath		Memmert	WNE 14 / L416.0606	Technology Promotion Association (Thailand-Japan)	22TM333	17 Feb 22	17 Feb 23	-
23	Water Bath		Memmert	WNE 14 / L416.0612	Technology Promotion Association (Thailand-Japan)	22TM334	17 Feb 22	17 Feb 23	-

รายการใบรับรองสอบเทียบ ทวนสอบ เครื่องมือหลักประจำห้องปฏิบัติการวิเคราะห์ สำหรับคุณภาพน้ำ

No.	Instrument/Equipment	Parameter	Manufacturer	Model/Serial No.	Calibrator	Certification No.	Date of Calibration	Due date of Calibration*	Remark
เครื่องมือหลักประจำห้องปฏิบัติการวิเคราะห์สำหรับคุณภาพน้ำ									
24	Water Bath	แบคทีเรียกลุ่มฟีคอลโคลิฟอร์ม (Fecal Coliform Bacteria)	Memmert	WNE 14 / L414.1407	Technology Promotion Association (Thailand-Japan)	22TM565	7 Apr 22	7 Apr 23	-
25	Water Bath		Memmert	WNE 14 / L414.1410	Technology Promotion Association (Thailand-Japan)	22TM564	7 Apr 22	7 Apr 23	-
26	Water Bath		Memmert	WNE 14 / L416.0614	Technology Promotion Association (Thailand-Japan)	22TM332	17 Feb 22	17 Feb 23	-
27	Analytical Balance		Mettler-Toledo	MS603S / B0070110311	Mettler-Toledo (Thailand) Ltd.	TH2058-096-040722-ACC-TH	7 Apr 22	6 Apr 23	-
28	Autoclave		ALP	CL-40L / 807298	Technology Promotion Association (Thailand-Japan)	22TM1121	11 Jul 22	10 Jul 23	-
29	Autoclave		ALP	CL-40L / 808763	Technology Promotion Association (Thailand-Japan)	22TM681	27 May 22	26 May 23	-
30	Refrigerator	เก็บรักษาตัวอย่างสำหรับวิเคราะห์ทางจุลชีววิทยา	Sanyo	SBC-337KD(GYN) / 71100607	Technology Promotion Association (Thailand-Japan)	22TM562	7 Apr 22	7 Apr 23	-

Due Date of Calibration* : Based on the annual calibration plan. At least 1 time per year.

เอกสารใบความคืบหน้า

Calibration Report

Certificate No.: 0000000000
Equipment: pH Meter
Model: HI 9142
Serial No.: 1000000000
Date of Calibration: 10/01/2020

Calibration point: 7.00, 4.00, 10.00
Reference point: 7.00, 4.00, 10.00

Measurand	Standard Value	Measured Value	Uncertainty	Expanded Uncertainty
pH	7.00	7.00	±0.01	±0.02
pH	4.00	4.00	±0.01	±0.02
pH	10.00	10.00	±0.01	±0.02

Signature
Date: 10/01/2020

By: [Signature]
Title: [Title]
This report is valid for the period of 12 months from the date of calibration.
Date of Calibration: 10/01/2020

เอกสารไม่ควบคุม

Calibration Certificate

Certificate No.: 0000000000
Client name: [Client Name]
Address: [Address]
Regulatory Authority: [Regulatory Authority]

Equipment: pH Meter
Manufacturer: HI 9142
Model: HI 9142
Serial No.: 1000000000
Order No.: 0000000000
Operation No.: 0000000000
Date of Receipt: 10/01/2020
Date of Calibration: 10/01/2020

Calibrated by: [Signature]
Approved by: [Signature]
Date of Issue: 10/01/2020

This certificate is valid for the period of 12 months from the date of calibration.
This certificate is valid for the period of 12 months from the date of calibration.
Date of Calibration: 10/01/2020

เอกสารไม่ควบคุม

Calibration Report

Certificate No.: 0000000000
Equipment: pH Meter
Model: HI 9142
Serial No.: 1000000000
Date of Calibration: 10/01/2020

Calibration point: 7.00, 4.00, 10.00
Reference point: 7.00, 4.00, 10.00

Measurand	Standard Value	Measured Value	Uncertainty	Expanded Uncertainty
pH	7.00	7.00	±0.01	±0.02
pH	4.00	4.00	±0.01	±0.02
pH	10.00	10.00	±0.01	±0.02

Measurand	Standard Value	Measured Value	Uncertainty	Expanded Uncertainty
pH	7.00	7.00	±0.01	±0.02
pH	4.00	4.00	±0.01	±0.02
pH	10.00	10.00	±0.01	±0.02

This report is valid for the period of 12 months from the date of calibration.
This report is valid for the period of 12 months from the date of calibration.
Date of Calibration: 10/01/2020

เอกสารไม่ควบคุม

Calibration Report

Certificate No.: 0000000000
Equipment: pH Meter
Model: HI 9142
Serial No.: 1000000000
Date of Calibration: 10/01/2020

Calibration point: 7.00, 4.00, 10.00
Reference point: 7.00, 4.00, 10.00

Measurand	Standard Value	Measured Value	Uncertainty	Expanded Uncertainty
pH	7.00	7.00	±0.01	±0.02
pH	4.00	4.00	±0.01	±0.02
pH	10.00	10.00	±0.01	±0.02

Measurand	Standard Value	Measured Value	Uncertainty	Expanded Uncertainty
pH	7.00	7.00	±0.01	±0.02
pH	4.00	4.00	±0.01	±0.02
pH	10.00	10.00	±0.01	±0.02

This report is valid for the period of 12 months from the date of calibration.
This report is valid for the period of 12 months from the date of calibration.
Date of Calibration: 10/01/2020

เอกสารไม่ควบคุม

Calibration Report

Calibration No.: 00221115 Page 1 of 1
 Equipment: Digital thermometer with 1000 g weight
 Model: 100-10 Brand: HANNA
 Location: Laboratory
 Date of Calibration: 22 May 2022 Approved:

Location: General Laboratory, Forensic Science Department
 Environmental Condition: Ambient Temperature: 25.0 ± 0.5 °C
 Humidity: 65 ± 5 % RH

Condition of Reference at Calibration:
 - Calibration certificate is valid and traceable to SI units.
 - The equipment is used for the purpose of calibration.
 - The equipment is not damaged or altered.
 - The equipment is not used for any other purpose.
 - The equipment is not used for any other purpose.
 - The equipment is not used for any other purpose.

Material	Serial No.	Calibration No.	Expiry Date	Range
Standard Reference Material (SRM)	1000	00221115	22 May 2022	1000 g

Standard Reference Material (SRM) is used for the purpose of calibration.

- The equipment is used for the purpose of calibration.
- The equipment is not damaged or altered.
- The equipment is not used for any other purpose.
- The equipment is not used for any other purpose.
- The equipment is not used for any other purpose.

Signature: [Signature] Date: 22 May 2022
 Approved: [Signature] Date: 22 May 2022

เอกสารไม่ควบคุม

Calibration Report

Calibration No.: 00221115 Page 1 of 1
 Equipment: Digital thermometer with 1000 g weight
 Model: 100-10 Brand: HANNA
 Location: Laboratory
 Date of Calibration: 22 May 2022 Approved:

Location: General Laboratory, Forensic Science Department
 Environmental Condition: Ambient Temperature: 25.0 ± 0.5 °C
 Humidity: 65 ± 5 % RH

Material	Serial No.	Calibration No.	Expiry Date	Range
Standard Reference Material (SRM)	1000	00221115	22 May 2022	1000 g

Standard Reference Material (SRM) is used for the purpose of calibration.

Signature: [Signature] Date: 22 May 2022
 Approved: [Signature] Date: 22 May 2022

เอกสารไม่ควบคุม

NFI Forensic Institute of Thailand
 กรมวิทยาศาสตร์การแพทย์
 กระทรวงสาธารณสุข

CERTIFICATE OF CALIBRATION

Calibration No.: 00221115 Page 1 of 1
 Equipment: Digital thermometer with 1000 g weight
 Model: 100-10 Brand: HANNA
 Location: Laboratory
 Date of Calibration: 22 May 2022 Approved:

Location: General Laboratory, Forensic Science Department
 Environmental Condition: Ambient Temperature: 25.0 ± 0.5 °C
 Humidity: 65 ± 5 % RH

Condition of Reference at Calibration:
 - Calibration certificate is valid and traceable to SI units.
 - The equipment is used for the purpose of calibration.
 - The equipment is not damaged or altered.
 - The equipment is not used for any other purpose.
 - The equipment is not used for any other purpose.
 - The equipment is not used for any other purpose.

Signature: [Signature] Date: 22 May 2022
 Approved: [Signature] Date: 22 May 2022

เอกสารไม่ควบคุม

NFI Forensic Institute of Thailand
 กรมวิทยาศาสตร์การแพทย์
 กระทรวงสาธารณสุข

REPORT OF CALIBRATION

Calibration No.: 00221115 Page 1 of 1
 Equipment: Digital thermometer with 1000 g weight
 Model: 100-10 Brand: HANNA
 Location: Laboratory
 Date of Calibration: 22 May 2022 Approved:

Location: General Laboratory, Forensic Science Department
 Environmental Condition: Ambient Temperature: 25.0 ± 0.5 °C
 Humidity: 65 ± 5 % RH

Condition of Reference at Calibration:
 - Calibration certificate is valid and traceable to SI units.
 - The equipment is used for the purpose of calibration.
 - The equipment is not damaged or altered.
 - The equipment is not used for any other purpose.
 - The equipment is not used for any other purpose.
 - The equipment is not used for any other purpose.

Signature: [Signature] Date: 22 May 2022
 Approved: [Signature] Date: 22 May 2022

เอกสารไม่ควบคุม

REPORT OF CALIBRATION

1024

Wachstumsbedingung: relative Temperatur 14 ± 1 °C

² <http://www.fishbase.org>, 1998-2000. Updated 14/01/2004.

Cellulose acetate, by James Watson & Co., Bristol, no. 13178-2279-09.

Controlled Release Materials

Material	Acq. No.	Catalogue No.	Year-batch
Weakness Feedback on	2014	9900	22 October 2015
Weakness Feedback on	2015	9900	22 October 2016
Prerequisite Weakness on	2006	9911	22 October 2016
Strongly Weakness on	2014	9902	22 October 2015

Disclaimer: The author does not intend to be considered a provider of clinical advice or a physician.

Journal of Neurobiology and Biotechnology (JNNB) through Science Society of Japan

Downloaded from <http://www.sagepub.com> at 11:23 11 May 2015

Copyright © 2005 John Wiley & Sons, Ltd. <http://www.interscience.wiley.com>

Source: Bureau of Economic Analysis (1999), p. 10.

Downloaded from <http://jci.org/> on September 11, 2012

Page 111 of 122

เอกสารไม่ควบคุม

REPORT OF CALIBRATION

Page 10 of 10

Expanding Horizons: World Admissions

Discussion

Sub-Category	Category 1 Values	Category 2 Values	Category 3 Values	Category 4 Values	Category 5 Values
Sub A	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
Sub B	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
Sub C	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
Sub D	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
Sub E	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
Sub F	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
Sub G	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000
	0.0000	0.0000	0.0000	0.0000	0.0000

เอกสารไม่ควบคุม

REPORT OF CALIBRATION

Page 4479

Massachusetts, University of

Runway	CRJ Cases	FLC Rating	Extension	Accession	Coverage Status
000	0.000	0.000	0.000	0.000	0
010	0.000	0.000	0.000	0.000	0.00
020	0.000	0.000	0.000	0.000	0.00
030	0.000	0.000	0.000	0.000	0.00
040	0.000	0.000	0.000	0.000	0.00
050	0.000	0.000	0.000	0.000	0.00
060	0.000	0.000	0.000	0.000	0.00
070	0.000	0.000	0.000	0.000	0.00
080	0.000	0.000	0.000	0.000	0.00
090	0.000	0.000	0.000	0.000	0.00
100	0.000	0.000	0.000	0.000	0.00
110	0.000	0.000	0.000	0.000	0.00
120	0.000	0.000	0.000	0.000	0.00
130	0.000	0.000	0.000	0.000	0.00
140	0.000	0.000	0.000	0.000	0.00
150	0.000	0.000	0.000	0.000	0.00
160	0.000	0.000	0.000	0.000	0.00
170	0.000	0.000	0.000	0.000	0.00
180	0.000	0.000	0.000	0.000	0.00
190	0.000	0.000	0.000	0.000	0.00
200	0.000	0.000	0.000	0.000	0.00
210	0.000	0.000	0.000	0.000	0.00
220	0.000	0.000	0.000	0.000	0.00
230	0.000	0.000	0.000	0.000	0.00
240	0.000	0.000	0.000	0.000	0.00
250	0.000	0.000	0.000	0.000	0.00
260	0.000	0.000	0.000	0.000	0.00
270	0.000	0.000	0.000	0.000	0.00
280	0.000	0.000	0.000	0.000	0.00
290	0.000	0.000	0.000	0.000	0.00
300	0.000	0.000	0.000	0.000	0.00
310	0.000	0.000	0.000	0.000	0.00
320	0.000	0.000	0.000	0.000	0.00
330	0.000	0.000	0.000	0.000	0.00
340	0.000	0.000	0.000	0.000	0.00
350	0.000	0.000	0.000	0.000	0.00
360	0.000	0.000	0.000	0.000	0.00
370	0.000	0.000	0.000	0.000	0.00
380	0.000	0.000	0.000	0.000	0.00
390	0.000	0.000	0.000	0.000	0.00
400	0.000	0.000	0.000	0.000	0.00
410	0.000	0.000	0.000	0.000	0.00
420	0.000	0.000	0.000	0.000	0.00
430	0.000	0.000	0.000	0.000	0.00
440	0.000	0.000	0.000	0.000	0.00
450	0.000	0.000	0.000	0.000	0.00
460	0.000	0.000	0.000	0.000	0.00
470	0.000	0.000	0.000	0.000	0.00
480	0.000	0.000	0.000	0.000	0.00
490	0.000	0.000	0.000	0.000	0.00
500	0.000	0.000	0.000	0.000	0.00
510	0.000	0.000	0.000	0.000	0.00
520	0.000	0.000	0.000	0.000	0.00
530	0.000	0.000	0.000	0.000	0.00
540	0.000	0.000	0.000	0.000	0.00
550	0.000	0.000	0.000	0.000	0.00
560	0.000	0.000	0.000	0.000	0.00
570	0.000	0.000	0.000	0.000	0.00
580	0.000	0.000	0.000	0.000	0.00
590	0.000	0.000	0.000	0.000	0.00
600	0.000	0.000	0.000	0.000	0.00
610	0.000	0.000	0.000	0.000	0.00
620	0.000	0.000	0.000	0.000	0.00
630	0.000	0.000	0.000	0.000	0.00

เอกสารไม่ควบคุม

REPORT OF CALTHICATIONS

Page 10 of 10

Baylis, Andrew J.

2006 Values	2007 Realized	Consensus	Uncertainty	Percentage Error
(Jan.)	(Jan.)	(Jan.)	(Jan.)	%
211.51	222.2	210	2.2	5.08
226.66	229.5	228	1.5	2.58
262.74	269.0	258	11.2	2.35
236.22	251.7	272	25.5	2.08
351.30	362.7	378	16.3	3.35
405.48	419.9	418	1.5	3.64
466.76	481.3	478	3.5	3.09
602.38	611.9	628	16.5	2.96
688.92	697.9	678	19.9	3.64
766.66	786.9	808	22.3	2.20
1077.96	1077.2	878	200	2.96
666.76	666.0	678	11.2	2.21
1135.02	1111.8	818	300	3.01
1011.92	1111.9	818	300	0.99
1122.77	1226.9	1212	114	2.96
1766.05	1711.9	1388	323	2.58
1666.46	1661.9	1388	273	2.58
1666.02	1661.9	1388	273	0.95
766.07	766.9	678	89	0.95
766.20	767.2	678	90	0.96
877.14	886.9	678	209	2.0
2077.76	2111.9	808	1303	2.08

เอกสารไม่ควบคุม

ISO 9001:2015
Surat
 1141-10000-Project 19, Jalan Pongki No. 1, Jalan Lenggong, Bangkok 10110
 Phone : +66-21-24 214, Email : apponnong@gmail.com

CERTIFICATE OF CALIBRATION

Certificate No. : 31021-008 Page 1 of 1

Customer : United Technic and Engineering Co., Ltd. (Thailand Branch)

Address : 150 Mittrapha 41, Sukhumvit Road, Bangkok, Thailand
 Bangkok 10110

Location of Calibration : Laboratory 213

Equipment : 175 N/A Spring scales

Manufacturer : Hoshino

Model : 11204

Serial No. : 31021-008

SN No. : 1502100101204

Calibrated Date : 20 January 2022

Calibration Date : 20 January 2022

Expiry Date : 24 January 2022

Calibration Interval : 12 mo.

Calibrated by :  Approved by : 

Calibration Engineer : Customer Representative

Thailand branch is approved by the Ministry of Commerce for calibration of the equipment.

Our company is able to provide a calibration certificate for the equipment which is not in the scope of the calibration certificate. The certificate is not valid for the equipment which is not in the scope of the calibration certificate.

Form No. 001-00000

เอกสารไม่ควบคุม

ISO 9001:2015
Surat
 1141-10000-Project 19, Jalan Pongki No. 1, Jalan Lenggong, Bangkok 10110
 Phone : +66-21-24 214, Email : apponnong@gmail.com

REPORT OF CALIBRATION

Certificate No. : 31021-008 Page 2 of 1

Customer Calibration : Analog Temperature 20 x 1, 2

Model/Serial No. : 15 x 101100

Calibration method : 15 x 101100 (17 x 101100) (17 x 101100) (17 x 101100)

Calibration Reference Materials :

Material	Serial No.	Calibration No.	Expiry Date
Standard Pressure 10	21100	31021	20 January 2022
Standard Pressure 10	21101	31021	20 January 2022
Standard Pressure 10	21102	31021	20 January 2022
Standard Pressure 10	21103	31021	20 January 2022

Availability : The certificate is valid for the international system of units (SI) and is not valid for the international system of units (SI) and is not valid for the international system of units (SI).

Approved Date : 2022-01-20

Scale Span of 10110 : 10110

Scale Interval of 10110 : 10110

Resolution of 10110 : 10110

Repeatability : 10110

Form No. 001-00000

เอกสารไม่ควบคุม

ISO 9001:2015
Surat
 1141-10000-Project 19, Jalan Pongki No. 1, Jalan Lenggong, Bangkok 10110
 Phone : +66-21-24 214, Email : apponnong@gmail.com

REPORT OF CALIBRATION

Certificate No. 31021-008 Page 3 of 1

Customer Calibration : 175 N/A Spring scales

Manufacturer Accuracy :

Force (N)	175 N/A Value	175 N/A Reading	Correction	Expanded	175 N/A Error
10	1.0000	1.0000	0.0000	0.0000	0.00
20	2.0000	2.0000	0.0000	0.0000	0.00
30	3.0000	3.0000	0.0000	0.0000	0.00
40	4.0000	4.0000	0.0000	0.0000	0.00
50	5.0000	5.0000	0.0000	0.0000	0.00
60	6.0000	6.0000	0.0000	0.0000	0.00
70	7.0000	7.0000	0.0000	0.0000	0.00
80	8.0000	8.0000	0.0000	0.0000	0.00
90	9.0000	9.0000	0.0000	0.0000	0.00
100	10.0000	10.0000	0.0000	0.0000	0.00
110	11.0000	11.0000	0.0000	0.0000	0.00
120	12.0000	12.0000	0.0000	0.0000	0.00
130	13.0000	13.0000	0.0000	0.0000	0.00
140	14.0000	14.0000	0.0000	0.0000	0.00
150	15.0000	15.0000	0.0000	0.0000	0.00
160	16.0000	16.0000	0.0000	0.0000	0.00
170	17.0000	17.0000	0.0000	0.0000	0.00
180	18.0000	18.0000	0.0000	0.0000	0.00
190	19.0000	19.0000	0.0000	0.0000	0.00
200	20.0000	20.0000	0.0000	0.0000	0.00

Form No. 001-00000

เอกสารไม่ควบคุม

ISO 9001:2015
Surat
 1141-10000-Project 19, Jalan Pongki No. 1, Jalan Lenggong, Bangkok 10110
 Phone : +66-21-24 214, Email : apponnong@gmail.com

REPORT OF CALIBRATION

Certificate No. 31021-008 Page 4 of 1

Manufacturer Accuracy :

Force (N)	175 N/A Value	175 N/A Reading	Correction	Expanded	175 N/A Error
10	1.0000	1.0000	0.0000	0.0000	0.00
20	2.0000	2.0000	0.0000	0.0000	0.00
30	3.0000	3.0000	0.0000	0.0000	0.00
40	4.0000	4.0000	0.0000	0.0000	0.00
50	5.0000	5.0000	0.0000	0.0000	0.00
60	6.0000	6.0000	0.0000	0.0000	0.00
70	7.0000	7.0000	0.0000	0.0000	0.00
80	8.0000	8.0000	0.0000	0.0000	0.00
90	9.0000	9.0000	0.0000	0.0000	0.00
100	10.0000	10.0000	0.0000	0.0000	0.00
110	11.0000	11.0000	0.0000	0.0000	0.00
120	12.0000	12.0000	0.0000	0.0000	0.00
130	13.0000	13.0000	0.0000	0.0000	0.00
140	14.0000	14.0000	0.0000	0.0000	0.00
150	15.0000	15.0000	0.0000	0.0000	0.00
160	16.0000	16.0000	0.0000	0.0000	0.00
170	17.0000	17.0000	0.0000	0.0000	0.00
180	18.0000	18.0000	0.0000	0.0000	0.00
190	19.0000	19.0000	0.0000	0.0000	0.00
200	20.0000	20.0000	0.0000	0.0000	0.00

Form No. 001-00000

เอกสารไม่ควบคุม



Verification Report

Certificate No.: 220226-001-01
 Equipment: METACHASS (2022)014
 Model: 2022 Serial No.: 470460
 Revision: 1 to 2 (2 No.: 042460111)2022
 Manufacturer: 2022

Date of Calibration: 14 Jun 2022 Page 1 of 1

Location: Laboratory Room, NFI (2022) 0001 (2022)014
 Environmental Condition:
 Ambient Temperature: 28 ± 0.5 °C
 Relative Humidity: 55 ± 5 %
 Air Pressure: 1013 ± 0.5 hPa

Condition of Measurement of Calibration

- The instrument was calibrated by using standard temperature and pressure (sea level) after the operation and control is completed. (checked) (see manual and check measurement is calibrated good)
- The temperature was controlled at 20 ± 0.5 °C.
- All gas flow rates were checked and the value was in the range of the specification.

Reference Standard Information

Standard	Model	Serial No.	Certificate No.	Exp. Date	Through
Triple Temperature	202201014	470460	220226-001-01	2022-06-14	2022-06-14
Gas Flowmeter	202201014	470460	220226-001-01	2022-06-14	2022-06-14

- The certificate is attached to the instrument under the seal of the laboratory.
- The certificate was verified and the instrument was approved.
- The result of calibration was checked and the value was in the range of the specification.
- Condition of Calibration form: Good
- Calibration result:
 One of result: 1.00 ± 0.01 (Mean: 1.00 ± 0.01 %)
- Good of calibration: ☒ Initial adjustment: ☐ After adjustment

Ministry of Industry, Thailand
 National Food Institute, Bangkok 10110

เอกสารไม่ควบคุม



Verification Report

Certificate No.: 220226-001-01
 Equipment: METACHASS (2022)014
 Model: 2022 Serial No.: 470460
 Revision: 1 to 2 (2 No.: 042460111)2022
 Manufacturer: 2022

Date of Calibration: 14 Jun 2022 Page 1 of 1

Block No.	100% Heating (°C)	100% Heating (°C)	Stability (°C)	Standard Temperature (°C)	Uncertainty (°C)
1	100	100	0.01	100.00	0.2
2	100	100	0.01	100.00	0.2
3	100	100	0.01	100.00	0.2
4	100	100	0.01	100.00	0.2
5	100	100	0.01	100.00	0.2
6	100	100	0.01	100.00	0.2
7	100	100	0.01	100.00	0.2
8	100	100	0.01	100.00	0.2
9	100	100	0.01	100.00	0.2
10	100	100	0.01	100.00	0.2
11	100	100	0.01	100.00	0.2
12	100	100	0.01	100.00	0.2
13	100	100	0.01	100.00	0.2
14	100	100	0.01	100.00	0.2
15	100	100	0.01	100.00	0.2
16	100	100	0.01	100.00	0.2
17	100	100	0.01	100.00	0.2
18	100	100	0.01	100.00	0.2
19	100	100	0.01	100.00	0.2
20	100	100	0.01	100.00	0.2

100% = 100% (see manual)
 Standard Temperature: 100.00 ± 0.01 (see manual)
 Stability: 0.01 (see manual)
 Uncertainty: 0.2 (see manual)

Ministry of Industry, Thailand
 National Food Institute, Bangkok 10110

เอกสารไม่ควบคุม



Verification Report

Certificate No.: 220226-001-01
 Equipment: METACHASS (2022)014
 Model: 2022 Serial No.: 470460
 Revision: 1 to 2 (2 No.: 042460111)2022
 Manufacturer: 2022

Date of Calibration: 14 Jun 2022 Page 1 of 1

Location: Laboratory Room, NFI (2022) 0001 (2022)014
 Environmental Condition:
 Ambient Temperature: 28 ± 0.5 °C
 Relative Humidity: 55 ± 5 %
 Air Pressure: 1013 ± 0.5 hPa

Figure 1: Location of reference flowmeter and flowmeter of equipment used
 TOP VIEW



Figure 1: Location of reference flowmeter and flowmeter of equipment used

- The certificate is attached to the instrument under the seal of the laboratory.
- The certificate was verified and the instrument was approved.
- The result of calibration was checked and the value was in the range of the specification.
- Condition of Calibration form: Good
- Calibration result:
 One of result: 1.00 ± 0.01 (Mean: 1.00 ± 0.01 %)
- Good of calibration: ☒ Initial adjustment: ☐ After adjustment

The result of calibration was checked and the value was in the range of the specification.

Ministry of Industry, Thailand
 National Food Institute, Bangkok 10110

เอกสารไม่ควบคุม



Verification Certificate

Certificate No.: 220226-001-01
 Client name: METACHASS (2022)014
 Address: 10110, Bangkok 10110, Thailand

Equipment: METACHASS (2022)014
 Manufacturer: METACHASS (2022)014
 Model: 2022
 Serial No.: 470460
 ID No.: 042460111
 Order No.: 220226-001-01
 Operation No.: 220226-001-01
 Date of Receipt: 14 Jun 2022
 Date of Calibration: 14 Jun 2022

Calibrated by: NFI (2022) 0001 (2022)014
 Approved by: NFI (2022) 0001 (2022)014
 Date of Issue: 14 Jun 2022

The certificate is a reference certificate of equipment used.
 The certificate is issued by the National Food Institute, Bangkok 10110, Thailand. The certificate is valid for the equipment used for the purpose of the certificate. The certificate is valid for the equipment used for the purpose of the certificate. The certificate is valid for the equipment used for the purpose of the certificate.

Ministry of Industry, Thailand
 National Food Institute, Bangkok 10110

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม

4 Control of Received Equipment

4.1 Verify that the correct instrument type and accessory kit items are received and in proper condition.

The program has stopped and the maintenance period of the server. The monitor will verify that all devices connected to stopped for the pending time. For each time-out, notify that the maintenance stream has not. If no more "0" in the maintenance of the maintenance stream.

[illegible]

5 Installation

3.1 The equipment must be installed in a suitable location with power, water and drains available.

^aPart of the knowledge contained on some of the elements is given in the appendix (p. 10), while "C" is the full version of the full knowledge is given.

Learning Requirements	Assessment Evidence	Pass (Y/N)
Identifying species for cross breeding	Micrographs demonstrating	Y
all, suitable crosses for cross project	100-200x 1000x	Y
Explain	100x	Y
Total Age of cross project	21/00/00-00/00/00	Y
State	For cross breeding, one must understand the genetic material involved	Y
Explain, interpret, discuss	How, 100x	Y
Explain, interpret, discuss	How, 100x, 1000x	Y
Explain, interpret, discuss	1000x-100x	Y

5.3 The instrument must be assembled correctly

5440) One of the two commonest is *U. rubra* (L.) in the high ranges of the Himalayas.

Attachment Fitting Connections	Attachment Options	Part (Y/N)
<p>1. Disinfect water in plastic generator 2. Disinfect water in chlorine vessel 3. Tap water solution in 4. Alcohol 5. Water 6. Add water 7. External 3 liter module 8. Gas control 9. Cooling water in tap water 10. Water valve and stop tap vessel 11. Drain 12. Cooling water tap tap water 13. Air in cooling water</p>	<p>Flow rate indicator for a valve</p>	<p>5</p>

5.3 The instrument should be assembled and covered as

Consider the staffing and to the person apply. For the first part of question and second part the expected maximum is obtained if τ_1 is set to τ^* is the right solution of the right-hand side equation.

[illegible]

⁴ Hence, please observe that the entire calculation may be adjusted if other party levels of detection are desired. I utilized the two standard values shown in the table.

Energy	Wavelength (nm)
10.2 eV	121.6 nm
12.09 eV	102.6 nm
13.6 eV	91.2 nm
15.48 eV	80.1 nm
17.42 eV	70.9 nm
19.44 eV	63.5 nm

$$s_{\text{residual}} = \frac{\sum_{i=1}^n (y_i - \hat{y}_i)^2}{n-2} = \frac{1.0000}{17} = 0.0588$$

© 1999 Blackwell Publishers Ltd. *Journal of Internal Medicine* 245: 399–406

10. *Journal of Management Studies*, 1996, 33, 1, 1-15.

1998, 1999, 2000, 2001, 2002, 2003, 2004, 2005, 2006, 2007, 2008, 2009, 2010, 2011, 2012, 2013, 2014, 2015, 2016, 2017, 2018, 2019, 2020, 2021, 2022, 2023, 2024, 2025, 2026, 2027, 2028, 2029, 2030, 2031, 2032, 2033, 2034, 2035, 2036, 2037, 2038, 2039, 2040, 2041, 2042, 2043, 2044, 2045, 2046, 2047, 2048, 2049, 2050, 2051, 2052, 2053, 2054, 2055, 2056, 2057, 2058, 2059, 2060, 2061, 2062, 2063, 2064, 2065, 2066, 2067, 2068, 2069, 2070, 2071, 2072, 2073, 2074, 2075, 2076, 2077, 2078, 2079, 2080, 2081, 2082, 2083, 2084, 2085, 2086, 2087, 2088, 2089, 2090, 2091, 2092, 2093, 2094, 2095, 2096, 2097, 2098, 2099, 2100, 2101, 2102, 2103, 2104, 2105, 2106, 2107, 2108, 2109, 2110, 2111, 2112, 2113, 2114, 2115, 2116, 2117, 2118, 2119, 2120, 2121, 2122, 2123, 2124, 2125, 2126, 2127, 2128, 2129, 2130, 2131, 2132, 2133, 2134, 2135, 2136, 2137, 2138, 2139, 2140, 2141, 2142, 2143, 2144, 2145, 2146, 2147, 2148, 2149, 2150, 2151, 2152, 2153, 2154, 2155, 2156, 2157, 2158, 2159, 2160, 2161, 2162, 2163, 2164, 2165, 2166, 2167, 2168, 2169, 2170, 2171, 2172, 2173, 2174, 2175, 2176, 2177, 2178, 2179, 2180, 2181, 2182, 2183, 2184, 2185, 2186, 2187, 2188, 2189, 2190, 2191, 2192, 2193, 2194, 2195, 2196, 2197, 2198, 2199, 2200, 2201, 2202, 2203, 2204, 2205, 2206, 2207, 2208, 2209, 2210, 2211, 2212, 2213, 2214, 2215, 2216, 2217, 2218, 2219, 2220, 2221, 2222, 2223, 2224, 2225, 2226, 2227, 2228, 2229, 2230, 2231, 2232, 2233, 2234, 2235, 2236, 2237, 2238, 2239, 2240, 2241, 2242, 2243, 2244, 2245, 2246, 2247, 2248, 2249, 2250, 2251, 2252, 2253, 2254, 2255, 2256, 2257, 2258, 2259, 2260, 2261, 2262, 2263, 2264, 2265, 2266, 2267, 2268, 2269, 2270, 2271, 2272, 2273, 2274, 2275, 2276, 2277, 2278, 2279, 2280, 2281, 2282, 2283, 2284, 2285, 2286, 2287, 2288, 2289, 2290, 2291, 2292, 2293, 2294, 2295, 2296, 2297, 2298, 2299, 2300, 2301, 2302, 2303, 2304, 2305, 2306, 2307, 2308, 2309, 2310, 2311, 2312, 2313, 2314, 2315, 2316, 2317, 2318, 2319, 2320, 2321, 2322, 2323, 2324, 2325, 2326, 2327, 2328, 2329, 2330, 2331, 2332, 2333, 2334, 2335, 2336, 2337, 2338, 2339, 2340, 2341, 2342, 2343, 2344, 2345, 2346, 2347, 2348, 2349, 2350, 2351, 2352, 2353, 2354, 2355, 2356, 2357, 2358, 2359, 2360, 2361, 2362, 2363, 2364, 2365, 2366, 2367, 2368, 2369, 2370, 2371, 2372, 2373, 2374, 2375, 2376, 2377, 2378, 2379, 2380, 2381, 2382, 2383, 2384, 2385, 2386, 2387, 2388, 2389, 2390, 2391, 2392, 2393, 2394, 2395, 2396, 2397, 2398, 2399, 2400, 2401, 2402, 2403, 2404, 2405, 2406, 2407, 2408, 2409, 2410, 2411, 2412, 2413, 2414, 2415, 2416, 2417, 2418, 2419, 2420, 2421, 2422, 2423, 2424, 2425, 2426, 2427, 2428, 2429, 2430, 2431, 2432, 2433, 2434, 2435, 2436, 2437, 2438, 2439, 2440, 2441, 2442, 2443, 2444, 2445, 2446, 2447, 2448, 2449, 2450, 2451, 2452, 2453, 2454, 2455, 2456, 2457, 2458, 2459, 2460, 2461, 2462, 2463, 2464, 2465, 2466, 2467, 2468, 2469, 2470, 2471, 2472, 2473, 2474, 2475, 2476, 2477, 2478, 2479, 2480, 2481, 2482, 2483, 2484, 2485, 2486, 2487, 2488, 2489, 2490, 2491, 2492, 2493, 2494, 2495, 2496, 2497, 2498, 2499, 2500, 2501, 2502, 2503, 2504, 2505, 2506, 2507, 2508, 2509, 2510, 2511, 2512, 2513, 2514, 2515, 2516, 2517, 2518, 2519, 2520, 2521, 2522, 2523, 2524, 2525, 2526, 2527, 2528, 2529, 2530, 2531, 2532, 2533, 2534, 2535, 2536, 2537, 2538, 2539, 2540, 2541, 2542, 2543, 2544, 2545, 2546, 2547, 2548, 2549, 2550, 2551, 2552, 2553, 2554, 2555, 2556, 2557, 2558, 2559, 2560, 2561, 2562, 2563, 2564, 2565, 2566, 2567, 2568, 2569, 2570, 2571, 2572, 2573, 2574, 2575, 2576, 2577, 2578, 2579, 2580, 2581, 2582, 2583, 2584, 2585, 2586, 2587, 2588, 2589, 2590, 2591, 2592, 2593, 2594, 2595, 2596, 2597, 2598, 2599, 2600, 2601, 2602, 2603, 2604, 2605, 2606, 2607, 2608, 2609, 2610, 2611, 2612, 2613, 2614, 2615, 2616, 2617, 2618, 2619, 2620, 2621, 2622, 2623, 2624, 2625, 2626, 2627, 2628, 2629, 2630, 2631, 2632, 2633, 2634, 2635, 2636, 2637, 2638, 2639, 2640, 2641, 2642, 2643, 2644, 2645, 2646, 2647, 2648, 2649, 2650, 2651, 2652, 2653, 2654, 2655, 2656, 2657, 2658, 2659, 2660, 2661, 2662, 2663, 2664, 2665, 2666, 2667, 2668, 2669, 2670, 2671, 2672, 2673, 2674, 2675, 2676, 2677, 2678, 2679, 26

© 2006 Blackwell Publishing Ltd
Journal of Internal Medicine 260: 479–487

เอกสารไม่ควบคุม

3 Summary of Deviations/Comments

The above data were used to estimate the number of people who were exposed to the gas.

Declaration	Signature	Comments

6. OQ Documentation

†New laboratory specimens of bone, stone, egg and hair (see below). ‡Large(40%) common, brown-colored egg. x 10 magnification.

If patient's social situation requires further reporting or counseling of needs, consult with Social Work as needed.

Received 10/1/01

[illegible]

Fluoride level (ppm) _____

Name: _____

How complete?

Journal of Business Ethics (2015) 129:111–125

เอกสารไม่ควบคุม

7 Appendix A

3.1 Preparation of Reagents

2000年 第1期

† *For a more extensive discussion of the evidence on the effects of the 1990s on women, see the accompanying article by Susan M. Schaefer.*

from 100 g (100%) per liter of solution. Commercially available concentrations are 10%. The use of concentrations above 40% is not advised as crystals formed impeding the function of the device. It can also only be concentrated to 40% (200 g/l) before use.

3.1.3. Titrant and determination of concentration

To be able to achieve a consistent interpretation of physics results, one must be able to say that the H₂O molecule is *not* a transomer, or what it is supposed to be. Quine's argument is particularly useful in relation to problems of particle physics as described below in this section. In order for H₂O conservation to obtain in the context of particle physics:

- Standard substance:**
Weigh 4.95 g, 10 g of acetylacetone-acetone (anhydrous) (AcAc/CCl₄) into a volumetric flask to make a 100 mL solution. Dry it for 1 h at 25°C or 2 h at 200°C. After drying in a desiccator, transfer the acetylacetone solution to a beaker with a glass rod. Remove it to a desiccator.
- Indicators/reagents:**
Dimethyl 0.1 g acetylacetone (100 mL anhydrous), Dimethyl 0.1 g bromoacetic acid in 100 mL anhydrous.
- Procedure:**
Weigh 4.95 g, 10 g of the standard substance (acetylacetone) into a volumetric flask; weigh the weight (W₁). Weigh the sodium acetate in a beaker flask and add 10 mL of 10% (w/v) of dimethylacetone and 5 drops of water to the sodium acetate. Transfer to flask, then the solvent is all used. Add the indicator solution to the flask. The solution will turn pink. Cool slowly to room temperature under constant stirring. Continue the titration until the next pink color change occurs. Remove all the solvent.

Copyright © 2004 by John Wiley & Sons, Inc.

Notes: Temperature changes with latitude; the volume and the concentration of ice sheets varies. The modeling assumptions of the three world agreements that of the agreement during stabilization. If temperature variations are significant, sufficient resources must be obtained for the goal of a sustainable world. (LUTIG 1992:210)

Journal of Management Inquiry 22(1)

3-2 Cancellation

Answer: $\frac{13,891}{100} = 138.91\%$

Open-Cellular used in relation to the rights of the individual

*Note: The colour change of the effluent produced (HMC, HMC-1) may be difficult to see. Therefore a pH meter or a colour indicator (e.g. 5 g/L NaOH indicator 8.1 g/L Bromocresol green is 100 ml solution) with scale is used as a reference.

7.3 Receiver Solution

Dilute acid (4 % with formaldehyde) gives neutral red inclusive solution.

In order to derive accurate results, the involved vehicles subjected to the impact (215 kJ) and positive force is subjected along moving a fixed distance. The 0.4 m long moving system is provided by connecting 800 g of balls and is shown in Fig. 1(a). (a) Test set-up (b) ball and force and distance measured were in a distance of about 8.1 cm. The acceleration of mass transportation and position of the measured points (100 g) is 1/300 of acceleration and 1/3 of the position (100 g) is 1/300 of acceleration. Thus, it is 1/300 of the measured value and also, the velocity

Note: The addition of a144 is a reversible positive blood test; 75% blood serum, 16 (up from 0.05–0.20 at blood, body and oral excretion when using this).

Adjustment of the model will be made by the following procedure:

4. Transfer 25 mL of the reaction solution to a graduated 100-mL volumetric flask and dilute to the mark with distilled water. If the solution is too dark to read, dilute with 10 mL of 0.1 M sodium hydroxide solution and dilute to the mark with distilled water. Calculate the amount of sodium hydroxide solution necessary to adjust the reaction acid solution to the 10:1 ratio with the formula: $\text{mL NaOH} = (\text{mL reaction} - 40) \times 0.1$.
5. Add the calculated amount of 0.1 M NaOH solution to the reaction mixture (Step 4).
6. To check percent reduction using 2,2'-bipyridine, first transfer 10 mL of the reaction mixture to a 250-mL volumetric flask and dilute to the mark with distilled water. If the value of the blank is higher than 0.2, add 0.2 mL of 0.1 M sodium hydroxide solution to the flask and dilute to the mark with distilled water. For comparison, add 10 mL of distilled water to the flask and dilute to the mark with distilled water. Repeat with a reading of 0.1, 0.25 and 0.5 mL of the reaction mixture. If a good blank is not achieved, with further small additions of 0.1 M NaOH you can reduce the blank until a satisfactory value is achieved.

© 2004 Blackwell Publishing Ltd *Journal of Internal Medicine* 255: 109–116

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม



Equipment : Resistor
Condition As Received : Good Item
Reference : 2259-000004-2
Procedure Used :

Cert. No. : 22T6802
Page : 1 of 3

Calibration was conducted using calibration procedure SP-0772 according to other measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD).

The temperature used used was based on ITS-90

Summary of the results of calibration

1. Reference standard information

Instrument	Model	Serial No.	Exp. No.	Exp. Date
1. Data Acquisition	54470A	014622000	215,365	25.10.2022

2. The certificate is valid only to the item calibrated in date and place of calibration

3. The certificate is valid only to the International System of Unit

Result of Calibration : $\pm 1^\circ\text{C}$ without adjustment

Function of 44470 : Temperature Source

Probe set setting : Not Available

	Beginning	Final
Temp. $\pm 1^\circ\text{C}$	20	20
Rel. Humid. $\pm 1\%$	42	57
AC Supply (Vrms)	220	220

Position	Ref. No.
1	15-10010-01
2	15-10010-02
3	15-10010-03
4	15-10010-04
5	15-10010-05
6	15-10010-06
7	15-10010-07
8	15-10010-08
9	15-10010-09



Hypercube Data:

L	0.1	mm
W	0.1	mm
H	0.1	mm

Distance of Center:

X	0.05	mm
Y	0.05	mm
Z	0.05	mm
Distance r	0.08	mm

เอกสารไม่ควบคุม



Equipment : Resistor
Condition As Received : Good Item
Reference : 2259-000004-2
Procedure Used :

Cert. No. : 22T6802
Page : 1 of 3

Calibration was conducted using calibration procedure SP-0772 according to other measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD).

The temperature used used was based on ITS-90

Summary of the results of calibration

1. Reference standard information

Instrument	Model	Serial No.	Exp. No.	Exp. Date
1. Data Acquisition	54470A	014622000	215,365	25.10.2022

2. The certificate is valid only to the item calibrated in date and place of calibration

3. The certificate is valid only to the International System of Unit

Result of Calibration : $\pm 1^\circ\text{C}$ without adjustment

Function of 44470 : Temperature Source

Probe set setting : Not Available

	Beginning	Final
Temp. $\pm 1^\circ\text{C}$	20	20
Rel. Humid. $\pm 1\%$	42	57
AC Supply (Vrms)	220	220

Calibration Temp. ($^\circ\text{C}$)	Measured Temperature ($^\circ\text{C}$)								
	1	2	3	4	5	6	7	8	9
20.0	20.02	20.04	20.07	20.10	20.08	20.05	20.03	20.01	20.00
20.1	20.03	20.06	20.10	20.14	20.09	20.07	20.04	20.02	20.01

average) The average of 20 above is each position

Temperature stability : One half of the greatest maximum difference of measured temperatures at any one-point

Temperature uniformity : The standard difference of measured temperatures at any one-point and the measured temperatures at the reference location which was placed at the same time or at an equal interval time as

needed to determine the temperature uniformity within the reference corner stability value conditions

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

Measurement : The difference of the maximum and minimum measured temperatures throughout calibration

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม



Equipment: Instrument
 Condition As Received: Used Item
 Reference: 2000-000000-4
 Result of Calibration: (1) Within Adjustment
 Function of UUT: Temperature Source
 From as setting: 0.0 Celsius

Cert. No. 0274802
 Page: 2 of 2

Calibration Point (°C)	UUT Setting (°C)	UUT Reading (°C)	Temperature Adjusts (± °C)	Temperature Uncertainty (°C)	Uncert. Adjustment (°C)	Uncertainty (± °C)	Change Factor (%)
0.0	0.0	0.00	0.00	0.00	0.00	0.00	0
40.0	40.0	40.0	0.00	0.00	0.00	0.00	0

Range: The average of 30 reads is used for the
Temperature stability: One half of the greatest maximum difference of measured temperatures of any one sensor.
Temperature uniformity: The maximum difference of measured temperatures of any sensors and the measured
 temperature of the reference location, which are measured at the same time or at 40 second intervals for an
 average to determine the temperature pattern of homogeneity within the chamber under steady state conditions.
Repeat Variability: The difference of the measured and reference measured temperatures throughout observation.
UUT: 104 (Date Calibration)

Note: The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage
 factor k, providing a level of confidence of approximately 95%.

0.00

เอกสารไม่ควบคุม



Equipment: Instrument
 Condition As Received: Used Item
 Reference: 2000-000000-4
 Result of Calibration: (1) Within Adjustment
 Function of UUT: Temperature Source
 From as setting: 0.0 Celsius



Certificate of Calibration

Cert. No. 0274802
 Page: 1 of 2

Equipment: Instrument
 Reference: Instrument
 Make: P-75
 Serial No.: 1211 0000
 Mfr No.: 1048 0000 0000
 Submitted by: 1048 0000 0000
 Location: 1048 0000 0000
 Release Date: 1048 0000
 Calibration Date: 1048 0000
 Ambient Temperature: 1048 0000
 Relative Humidity: 1048 0000
 Calibrated by: 1048 0000

Approved by: 1048 0000
 1) 1048 0000
 2) 1048 0000
 3) 1048 0000

Issue Date: 1048 0000

The following are the conditions provided for the equipment:
 1) 1048 0000
 2) 1048 0000
 3) 1048 0000

เอกสารไม่ควบคุม



Equipment: Instrument
 Condition As Received: Used Item
 Reference: 2000-000000-4
 Result of Calibration: (1) Within Adjustment
 Function of UUT: Temperature Source
 From as setting: 0.0 Celsius

Cert. No. 0274802
 Page: 2 of 2

Range: The average of 30 reads is used for the
Temperature stability: One half of the greatest maximum difference of measured temperatures of any one sensor.
Temperature uniformity: The maximum difference of measured temperatures of any sensors and the measured
 temperature of the reference location, which are measured at the same time or at 40 second intervals for an
 average to determine the temperature pattern of homogeneity within the chamber under steady state conditions.
Repeat Variability: The difference of the measured and reference measured temperatures throughout observation.
UUT: 104 (Date Calibration)

Note: The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage
 factor k, providing a level of confidence of approximately 95%.

Calibration of the result of calibration
 1) Reference standard 4 instruments
 2) Data Acquisition
 3) Data Acquisition
 4) Data Acquisition
 5) Data Acquisition
 6) Data Acquisition
 7) Data Acquisition
 8) Data Acquisition
 9) Data Acquisition
 10) Data Acquisition
 11) Data Acquisition
 12) Data Acquisition
 13) Data Acquisition
 14) Data Acquisition
 15) Data Acquisition
 16) Data Acquisition
 17) Data Acquisition
 18) Data Acquisition
 19) Data Acquisition
 20) Data Acquisition
 21) Data Acquisition
 22) Data Acquisition
 23) Data Acquisition
 24) Data Acquisition
 25) Data Acquisition
 26) Data Acquisition
 27) Data Acquisition
 28) Data Acquisition
 29) Data Acquisition
 30) Data Acquisition

Calibration Point (°C)	UUT Setting (°C)	UUT Reading (°C)	Temperature Adjusts (± °C)	Temperature Uncertainty (°C)	Uncert. Adjustment (°C)	Uncertainty (± °C)	Change Factor (%)
0.0	0.0	0.00	0.00	0.00	0.00	0.00	0
40.0	40.0	40.0	0.00	0.00	0.00	0.00	0

Range: The average of 30 reads is used for the
Temperature stability: One half of the greatest maximum difference of measured temperatures of any one sensor.
Temperature uniformity: The maximum difference of measured temperatures of any sensors and the measured
 temperature of the reference location, which are measured at the same time or at 40 second intervals for an
 average to determine the temperature pattern of homogeneity within the chamber under steady state conditions.
Repeat Variability: The difference of the measured and reference measured temperatures throughout observation.
UUT: 104 (Date Calibration)

Note: The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage
 factor k, providing a level of confidence of approximately 95%.

0.00

เอกสารไม่ควบคุม



Equipment: Instrument
 Condition As Received: Used Item
 Reference: 2000-000000-4
 Result of Calibration: (1) Within Adjustment
 Function of UUT: Temperature Source
 From as setting: 0.0 Celsius

Cert. No. 0274802
 Page: 1 of 2

Calibration Point (°C)	UUT Setting (°C)	UUT Reading (°C)	Temperature Adjusts (± °C)	Temperature Uncertainty (°C)	Uncert. Adjustment (°C)	Uncertainty (± °C)	Change Factor (%)
0.0	0.0	0.00	0.00	0.00	0.00	0.00	0
40.0	40.0	40.0	0.00	0.00	0.00	0.00	0

Range: The average of 30 reads is used for the
Temperature stability: One half of the greatest maximum difference of measured temperatures of any one sensor.
Temperature uniformity: The maximum difference of measured temperatures of any sensors and the measured
 temperature of the reference location, which are measured at the same time or at 40 second intervals for an
 average to determine the temperature pattern of homogeneity within the chamber under steady state conditions.
Repeat Variability: The difference of the measured and reference measured temperatures throughout observation.
UUT: 104 (Date Calibration)

Note: The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage
 factor k, providing a level of confidence of approximately 95%.

0.00

เอกสารไม่ควบคุม



Cert. No.: 2019001
 Page: 1 of 3

Certificate of Calibration

Equipment: Resistance
Manufacturer: Kometec
Model: K9-70
Serial No.: 0217-2007
UIC No.: 00000000000000000000
Submitted by: United States and Engineering Consultant Co., Ltd.
 3-50-1 Sukhumvit 41, Sukhumvit Road,
 Bangkok, Phrasang, Bangkok 10250
Location: Metrology Laboratory (DCC)
Received Date: 23 May 2020
Calibration Date: 23 May 2020
Ambient Temperature: (20 ± 1) °C
Relative Humidity: (30 ± 5) %
Calibrated by: Praveen Nigam

Approved by:
 Praveen Nigam
 Head of Metrology Laboratory (DCC)

Issue Date: 23 May 2020

The Information on this certificate is provided for information only.

เอกสารไม่ควบคุม



Equipment: Resistance
Manufacturer: Kometec
Model: K9-70
Serial No.: 0217-2007
UIC No.: 00000000000000000000

Cert. No.: 2019001
 Page: 1 of 3

Calibration was conducted using calibration procedure CP-0002 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD).

The temperature measurement was based on ITS-90.

Condition of Measurement:

1. Resistance standard (Resistance)
Instrument: Model: 34475A Serial No.: 999420101
Condition: 1. Data Acquisition 2. 1.1 Without Adjustment

Result of Calibration: 1.1 Without Adjustment

Function of UIC: Temperature Sensor

Function of setting: 1.1 Without Adjustment



Value Indication (Value):

0.0 0.0 0.0

0.0 0.0 0.0

0.0 0.0 0.0

Dimension of Dimension:

0.0 0.0 0.0

0.0 0.0 0.0

0.0 0.0 0.0

Dimension	Beginning	Final
Value (°C)	20	20
RTD (mV)	0.0	0.0
AD (mV)	0.0	0.0

Position	Ref. No.
1	00000000000000000000
2	00000000000000000000
3	00000000000000000000
4	00000000000000000000
5	00000000000000000000
6	00000000000000000000
7	00000000000000000000
8	00000000000000000000
9	00000000000000000000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

0.000 0.000 0.000

Certificate of Calibration

Cert. No.: 2019001
 Page: 1 of 3

Equipment: Resistance
Manufacturer: Kometec
Model: K9-70
Serial No.: 0217-2007
UIC No.: 00000000000000000000
Submitted by: United States and Engineering Consultant Co., Ltd.
 3-50-1 Sukhumvit 41, Sukhumvit Road,
 Bangkok, Phrasang, Bangkok 10250
Location: Metrology Laboratory
Received Date: 23 February 2020
Calibration Date: 23 February 2020
Ambient Temperature: (20 ± 1) °C
Relative Humidity: (30 ± 5) %
Calibrated by: Praveen Nigam

Approved by:
 Praveen Nigam
 Head of Metrology Laboratory (DCC)

Issue Date: 23 February 2020

The Information on this certificate is provided for information only.

The Information on this certificate is provided for information only.

The Information on this certificate is provided for information only.

The Information on this certificate is provided for information only.

The Information on this certificate is provided for information only.

The Information on this certificate is provided for information only.

The Information on this certificate is provided for information only.

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม



Doc. No.: 2270009
 Page: 1 of 3

Certificate of Calibration

Equipment: Water Bath
Manufacturer: Memmert
Model: WM 14
Serial No.: 4314 1430
SI No.: UMS-000-00000000
Submitted by: United Asian and Engineering Consultant Co., Ltd.
 3-80/1 Mahachulalongkornrajavidyalaya Road
 Bangkok, Thailand
 Bangkok 10200
Location: Ministry Laboratory
Received Date: 7 April 2022
Calibration Date: 7 April 2022
Reference Temperature: (20 ± 1) °C
Relative Humidity: (30 ± 20) %
Calibrated by: Preeti S. Srinivasan

Approved by:
 Preeti S. Srinivasan
 Metrologist
 Head of Lab

Issue Date: 10 April 2022

The measurement uncertainty for a measurement point is approximately 0.01 °C.

The measurement uncertainty for a measurement point is approximately 0.01 °C.

The measurement uncertainty for a measurement point is approximately 0.01 °C.

เอกสารไม่ควบคุม



Equipment: Water Bath
Manufacturer: Memmert
Model: WM 14
Serial No.: 4314 1430
SI No.: UMS-000-00000000
Submitted by: United Asian and Engineering Consultant Co., Ltd.
 3-80/1 Mahachulalongkornrajavidyalaya Road
 Bangkok, Thailand
 Bangkok 10200

Doc. No.: 2270009
 Page: 2 of 3

Calibration was performed using a double calibration procedure (DPC) according to draft measurement method with Data Acquisition system connected with industrial Platform Scale (Mettler) (DPC).

The measurement uncertainty for a measurement point is approximately 0.01 °C.

Condition of the result of calibration:

1. Reference standard: Mettler

Equipment: Water Bath
Manufacturer: Memmert
Model: WM 14
Serial No.: 4314 1430
SI No.: UMS-000-00000000
Submitted by: United Asian and Engineering Consultant Co., Ltd.
 3-80/1 Mahachulalongkornrajavidyalaya Road
 Bangkok, Thailand
 Bangkok 10200


2. The certificate is valid only for the term calibration date and date of calibration.

3. The certificate is traceable to the International System of Unit.

Result of Calibration: (1) Water Adjustment

Function of WBC: Temperature Control

	Environmental	AC Voltage Supply
	1 °C ± 1	1 °C ± 1
Measuring of Calibration	20	20
Function of Calibration	20	20



Position	Ref. No.
1	1000144
2	1000144
3	1000144
4	1000144
5	1000144

เอกสารไม่ควบคุม



Equipment: Water Bath
Manufacturer: Memmert
Model: WM 14
Serial No.: 4314 1430
SI No.: UMS-000-00000000
Submitted by: United Asian and Engineering Consultant Co., Ltd.
 3-80/1 Mahachulalongkornrajavidyalaya Road
 Bangkok, Thailand
 Bangkok 10200

Doc. No.: 2270009
 Page: 3 of 3

Calibration point (°C)	Measuring (°C)	Heating (°C)	Average Standard Reading (°C)				
			1	2	3	4	5 (avg.)
44.2	44.2	44.2	44.200	44.200	44.200	44.200	44.200

Calibration point (°C)	Measuring (°C)	Heating (°C)	Uncertainty (°C)	Coverage Factor
44.2	44.2	44.2	0.01	2

Measuring: The average of 30 values is used for the average.

Uncertainty: The maximum difference of measured temperature at any point and the measured temperature at the reference location, which was measured at the same time as at any other in the calibration data, is possible to determine the temperature difference at any point and the measured temperature at the reference location.

Measuring: The average of the measured temperature difference of measured temperature at any point and the measured temperature at the reference location is used for the average.

Measuring: The average of 30 values is used for the average.

Measuring: The average of 30 values is used for the average.

Measuring: The average of 30 values is used for the average.

Measuring: The average of 30 values is used for the average.

Measuring: The average of 30 values is used for the average.

Measuring: The average of 30 values is used for the average.

เอกสารไม่ควบคุม



Certificate of Calibration

Doc. No.: 2270009
 Page: 1 of 3

Equipment: Water Bath
Manufacturer: Memmert
Model: WM 14
Serial No.: 4314 1430
SI No.: UMS-000-00000000
Submitted by: United Asian and Engineering Consultant Co., Ltd.
 3-80/1 Mahachulalongkornrajavidyalaya Road
 Bangkok, Thailand
 Bangkok 10200
Location: Ministry Laboratory
Received Date: 7 April 2022
Calibration Date: 7 April 2022
Reference Temperature: (20 ± 1) °C
Relative Humidity: (30 ± 20) %
Calibrated by: Preeti S. Srinivasan

Approved by:
 Preeti S. Srinivasan
 Metrologist
 Head of Lab

Issue Date: 10 April 2022

The measurement uncertainty for a measurement point is approximately 0.01 °C.

The measurement uncertainty for a measurement point is approximately 0.01 °C.

The measurement uncertainty for a measurement point is approximately 0.01 °C.

เอกสารไม่ควบคุม



Equipment: Balance
Certificate No./Version: 11001 100
Reference: ISO 9000:2015
Procedure Used:

Gen. No.: 2019001
Page: 1 of 1

Calibration was conducted using reference calibration procedure (P-010) according to good measurement method with Data Acquisition which connected with Thermocouple Type T.

The accuracy of each point was based on ITS-90.

Condition of the result of calibration

1. Reference related calibration

Equipment: Balance
Certificate No.: 2019001
Date: 20 May 2020

- The certificate is valid only to the item submitted to calibration.
- The certificate is according to the International System of Unit.
- The result of calibration is only valid for the calibration of goods and standard which used for reference with reference calibrated as Master Group 1, 2 and 3.
- The result of calibration is only valid for the calibration of goods and standard which used for reference with reference calibrated as Master Group 1, 2 and 3.
- The result of calibration is only valid for the calibration of goods and standard which used for reference with reference calibrated as Master Group 1, 2 and 3.

Result of Calibration: 1. Without Adjustment
Precision of UMC: Temperature Range



Performance			
Beginning of Calibration	1 °C	10 °C	20 °C
End of Calibration	20	20	20
Precision	0.01	0.01	0.01

เอกสารไม่ควบคุม



Equipment: Balance
Certificate No./Version: 11001 100
Reference: ISO 9000:2015
Procedure Used:

Gen. No.: 2019001
Page: 1 of 1

Calibration was conducted using reference calibration procedure (P-010) according to good measurement method with Data Acquisition which connected with Thermocouple Type T.

The accuracy of each point was based on ITS-90.

Condition of the result of calibration

1. Reference related calibration

Equipment: Balance
Certificate No.: 2019001
Date: 20 May 2020

- The certificate is valid only to the item submitted to calibration.
- The certificate is according to the International System of Unit.
- The result of calibration is only valid for the calibration of goods and standard which used for reference with reference calibrated as Master Group 1, 2 and 3.
- The result of calibration is only valid for the calibration of goods and standard which used for reference with reference calibrated as Master Group 1, 2 and 3.
- The result of calibration is only valid for the calibration of goods and standard which used for reference with reference calibrated as Master Group 1, 2 and 3.

Result of Calibration: 1. Without Adjustment
Precision of UMC: Temperature Range

Performance			
Beginning of Calibration	1 °C	10 °C	20 °C
End of Calibration	20	20	20
Precision	0.01	0.01	0.01

เอกสารไม่ควบคุม



ISO 9000:2015 CERTIFICATION (P-010) and ISO 9000:2015 CERTIFICATION (P-010) and ISO 9000:2015 CERTIFICATION (P-010)



Gen. No.: 2019001
Page: 1 of 1

Certificate of Calibration

Equipment: Balance
Certificate No.: 2019001
Date: 20 May 2020
Submitted by: 1. Without Adjustment
Precision of UMC: Temperature Range

Approved by: 1. Without Adjustment
Precision of UMC: Temperature Range

Issue Date: 2 June 2020

The Certificate is valid for a maximum of 12 months after the date of issue.

เอกสารไม่ควบคุม



Equipment: Balance
Certificate No./Version: 11001 100
Reference: ISO 9000:2015
Procedure Used:

Gen. No.: 2019001
Page: 1 of 1

Calibration was conducted using reference calibration procedure (P-010) according to good measurement method with Data Acquisition which connected with Thermocouple Type T.

The accuracy of each point was based on ITS-90.

Condition of the result of calibration

1. Reference related calibration

Equipment: Balance
Certificate No.: 2019001
Date: 20 May 2020

- The certificate is valid only to the item submitted to calibration.
- The certificate is according to the International System of Unit.
- The result of calibration is only valid for the calibration of goods and standard which used for reference with reference calibrated as Master Group 1, 2 and 3.
- The result of calibration is only valid for the calibration of goods and standard which used for reference with reference calibrated as Master Group 1, 2 and 3.
- The result of calibration is only valid for the calibration of goods and standard which used for reference with reference calibrated as Master Group 1, 2 and 3.

Result of Calibration: 1. Without Adjustment
Precision of UMC: Temperature Range

Performance			
Beginning of Calibration	1 °C	10 °C	20 °C
End of Calibration	20	20	20
Precision	0.01	0.01	0.01

เอกสารไม่ควบคุม



Equipment: Refrigerator
Condition As Received: Used Item
Reference: 2204-0700002
Model of Calibration: (1) Without Adjustment

Cert. No.: 2219491
Page: 1 of 2

Operating parameter Set: Temperature = 175.0 °C

Measurement point(s): 15 points

UVC Setting (°C)	IRRT Reading (°C)	Position	Average Standard Reading (°C)	Stability (± °C)	Percent Reading (± %)	Uncertainty (± °C)	Coverage Factor
175.0	175.0	1	175.000				
		2	175.000				
		3	175.000				
		4	175.000				

Operating parameter Set: Temperature = 181 °C

Measurement point(s): 15 points

UVC Setting (°C)	IRRT Reading (°C)	Position	Average Standard Reading (°C)	Stability (± °C)	Percent Reading (± %)	Uncertainty (± °C)	Coverage Factor
181.0	181.0	1	181.000				
		2	181.000				
		3	181.000				
		4	181.000				

Average: The average of 32 values in each position.

Stability: One half of the greatest maximum difference of measured temperatures at any one point.

UVC: Ultraviolet Calibration.

Note: The reported uncertainty of measurement includes random stability and standard conformity.

The reported uncertainty of measurement was based on a coverage probability included by a coverage factor k, providing a level of confidence of approximately 95 %.

—

เอกสารไม่ควบคุม



NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY
1001 DuPont Circle, NIST Building 210, Gaithersburg, MD 20899-0101
www.nist.gov



Cert. No.: 2219492
Page: 1 of 2

Certificate of Calibration

Customer: Rutgers
Model Name: Temo
Model: 220-2219492
Serial No.: 1110007
ID No.: 220-2219492
Submitted by: James Aragon and Engineering Services 4 Inc., LLC
Location: 1000 University Ave., Suite 100, Raleigh, North Carolina 27601
Received Date: 1 April 2022
Calibration Date: 1 April 2022
Ambient Temperature: 20 ± 0.1 °C
Relative Humidity: 30 ± 0.1 %
Calibrated by: Frank Gonzalez

Approved by: [Signature]
Approved Signature: [Signature]

Issue Date: 16 April 2022
For Uncertainty within a confidence probability of approximately 95 %

Measurement uncertainty based on a coverage probability of approximately 95 %

เอกสารไม่ควบคุม



Equipment: Refrigerator
Condition As Received: Used Item
Reference: 2204-0700002
Model of Calibration: (1) Without Adjustment

Cert. No.: 2219493
Page: 1 of 2

Calibration was conducted using calibration standards 22-0700002 according to NIST measurement method with data acquisition system connected with Reference Temperature Sensor (RTS).

The temperature was used and based on 110.00.

CONDITION (Model/Configuration)

Reference: 2204-0700002

1. Data Acquisition: Model: 2204-0700002, Serial No.: 2219493, Date: 21 April 2022, Rev: 1.000

2. Measurement is relative to the reference standard or data set used in calibration.

3. The certificate is issued to the International System of Units.

Model of Calibration: (1) Without Adjustment

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

เอกสารไม่ควบคุม



Equipment: Refrigerator
Condition As Received: Used Item
Reference: 2204-0700002
Model of Calibration: (1) Without Adjustment

Cert. No.: 2219493
Page: 1 of 2

Calibration was conducted using calibration standards 22-0700002 according to NIST measurement method with data acquisition system connected with Reference Temperature Sensor (RTS).

The temperature was used and based on 110.00.

CONDITION (Model/Configuration)

Reference: 2204-0700002

1. Data Acquisition: Model: 2204-0700002, Serial No.: 2219493, Date: 21 April 2022, Rev: 1.000

2. Measurement is relative to the reference standard or data set used in calibration.

3. The certificate is issued to the International System of Units.

Model of Calibration: (1) Without Adjustment

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

Function of UVC: Temperature Sensor

เอกสารไม่ควบคุม

ภาคผนวก จ-2
เอกสารสอบเทียบเครื่องมือตรวจวัด



List of Instruments Certification for Water Quality Analysis

No.	Instrument/Equipment	Parameter	Manufacturer	Model/Serial No.	Calibrator	Certification No.	Date of Calibration	Due date of Calibration	Remark
Water									
1	pH Meter	pH	Horiba	pH100A JC03354	Technology Promotion Association (Thailand-Japan)	22CH60	13 Jan 22	13 Jan 23	-
2	DO Meter	DO	YSI	Pro 20i 18K104053	Technology Promotion Association (Thailand-Japan)	22TW168	24 Jul 22	25 Jul 23	-



Cert.No.: 22CH60
Page: 1 of 3

Certificate of Calibration

Equipment : pH Meter
Manufacturer : EcoSense
Model : pH100A
Serial No. : JC03354
ID No. : UAE.EFM.063/2562(ENV pH 03/62)
Condition As-Received: Used Item
Received Date : 12 January 2022
Calibration Date : 13 January 2022
Reference : 2201-0350WSC-1
Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
3 Soi Udomsuk 41, Sukhumvit Road, Bangkok,
Phrakhanong, Bangkok 10260
Ambient Temperature : (25 ± 2.5) °C
Relative Humidity : (50 ± 15) %
Calibration Procedure : In-house method
CP-CH5 by direct measurement with standard
voltage calibrator and direct measurement with
certified reference material (CRM)
CP-CH8 by comparison with standard thermometer

Calibrated by : Warakorn Lerngagitrakul

Approved by :

Approved Signatory

(/) Malee Butkruea
() Saitip Meangmai
() Warakorn Lerngagitrakul

Issue Date : 17 January 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written

Approval of the head of Corporate Services 3: Equipment Calibration and Testing Services

เอกสารไม่ควบคุม



Condition of this calibration result

1 Reference Standard Instrument : -
Instrument Serial No. ID No. Cert. No. Due Date
1) Document Process Calibrator 54030049 130RC116 21E2682 25 Aug 2022
2) Ref. Standard Thermometer 4982054 110RC044 21H1201 26 Oct 2022
This certification is traceable to the International System of Unit maintained at:-
Traceable to National Institute of Metrology (Thailand), NIMT

2. Certified Reference Materials : The measurement results are traceable to SI through CPA chem Ltd.
ANSI-ASQ National Accreditation Board, Accredited No. AR-1835

Buffer Solution	Manufacturer	Lot No.	Ex. date
pH 4.008	CPA chem	766820	23 Sep 2023
pH 6.982	CPA chem	761017	02 Aug 2022
pH 10.015	CPA chem	766824	04 Sep 2022

3. This certificate is valid only to the item calibrated on date and place of calibration.

Calibration Results

Function : mV Measurement

Performing standard curve by Fluke at pH (4.7)(7.10)

Unit Under Calibration	Nominal Value	Standard Voltage Input	Actual Reading		Uncertainty of Measurement (± mV)	Coverage factor k
			mV	pH		
pH Meter	4.00	177.48	177	4.01	0.58	2.00
S/N: JC03354	7.00	0.00	0	7.00	0.58	2.00
	7.00	0.00	0	7.00	0.58	2.00
	10.00	-177.48	-178	-10.01	0.58	2.00

เอกสารไม่ควบคุม



Cert.No.: 22CH60
Page: 3 of 3

Calibration Results

Function : pH Measurement

Performing three buffers standard curve by using buffer nominal pH (4.7)(7.10)

Unit Under Calibration	Standard pH Buffer Solution	Actual pH Reading	Actual mV Reading (mV)	Uncertainty of pH measurement (±)	Coverage factor k
pH Electrode S/N: 200729SIA605377	4.008	4.01	144	0.0079	2.00
	6.982	6.98	-28	0.011	2.00
	6.982	6.98	-27	0.0099	2.00
	10.015	10.01	-200	0.0096	2.00

Function : Temperature Measurement

(*) Without adjustment

This equipment was connected with Temperature Probe:

Model :
Serial No. 200729SIA605377

Dimension of probe:

- Length : 112 mm
- Diameter : 12 mm
- Immersion Depth : 100 mm

Calibration Point (°C)	Standard Temperature (°C)	UUC* Reading (°C)	Error (°C)	Uncertainty of measurement (± °C)	Coverage factor k
25.0	25.003	25.1	0.097	0.13	2.00
30.0	30.002	30.1	0.098	0.13	2.00
35.0	35.004	35.0	-0.004	0.13	2.00

Remark : - UUC* = Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-000-

เอกสารไม่ควบคุม



Cert.No.: 21TW158
Page.: 1 of 2

Certificate of Testing

Equipment : DO Meter
Manufacturer : YSI
Model : Pro 20i
Serial No. : 18H110457
ID No. : UAE.EFM.202/2561(ENV DO.06/61)
Received Date : 27 July 2021
Test Date : 30 July 2021
Reference : 2107-0695WSC-3
Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
3 Soi Udomsuk 41, Sukhumvit Road, Bangkok,
Phrakhanong, Bangkok 10260
Laboratory Condition : Temperature (25 ± 5) °C
Humidity (50 ± 20) %
Test Procedure : In-house method : CP-CH9
by Comparison Technique with Azide Modification Method

Tested by : Walailak Sirtinean

Approved by :

Approved Signatory

() Malee Butkruea
(/) Saitip Meangmai
() Warakorn Lerngagitrakul

Issue Date : 4 August 2021

เอกสารไม่ควบคุม



Cert.No.: 21TW158
Page.: 2 of 2

Result : Dissolved Oxygen Meter Adjustment With Air 100 %

Dissolved Oxygen Probe No.: 18H110457

Titration Method (Azide Modification Method) (mg/L)	DO Meter Reading (mg/L)	Standard Deviation (mg/L)
8.04	8.03	0.0055

This report was certified only for the instrument we tested. It is allowable to use for study the system efficiency. The environmental impact control and present to organization it may concerned. Intend to use for advertising and referral purpose is prohibited. This report may not be reproduced other in full without written approval of the laboratory.

-o0o-

เอกสารไม่ควบคุม



TECHNOLOGY PROMOTION ASSOCIATION (THAI AND JAPAN)
CORPORATE SERVICES & EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL: 0-2719-3000-27 FAX: 0-2719-9484



Cert. No.: 21TM1366
Page : 1 of 2

Certificate of Calibration

Equipment : DO Meter With Sensor

Manufacturer : YSI

Model : Pro 20i

Serial No. : 18H110457

ID No. : UAE.EFM.202/2561 (ENV.DO.06/61)

Submitted by : United Analyst and Engineering Consultant Co.,Ltd
3 Soi Udomsuk 41, Sukhumvit Road,
Bangchak, Phrakhanong,
Bangkok 10260

Location : TPA On Site Calibration Laboratory

Received Order : 2 August 2021


Calibrated Date : 2 August 2021

Ambient Temperature : (26 ± 10) °C

Relative Humidity : (50 ± 30) %

AC Line Voltage : (220 ± 22) V

Calibrated by : Preecha Hlanib

Approved by : 
Approved Signatory

() Ponthippa Tarneyakul

() Malee Bulkruea

(✓) Suwit Imjai

Issue Date : 10 August 2021

The Uncertainties are for a confidence probability of approximately 95 %

This certificate may not be reproduced other than in full, except with the prior written

Approval of the head of Corporate Services 3, Equipment Calibration and Testing Services

เอกสารไม่ควบคุม



Equipment : DO Meter With Sensor
Condition As-Received : Used Item
Reference : 2107-0695WSC-4

Cert. No.: 21TM1366
Page.: 2 of 2

Procedure Used :-

Calibration were conducted using in-house calibration procedure CP-OT01 according to comparison with Industrial Platinum Resistance Thermometer (IPR T) into Temperature Bath.

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Digital Thermometer	1502A	A52847	2011246	14 Oct 2021

2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit

Result of Calibration :- (*) Without Adjustment

Function : Temperature measurement.

This instrument was connected with temperature sensor, S/N.: 18H100722

Calibration Point (°C)	Immersion Depth (mm)	Standard Temperature (°C)	UUC* Reading (°C)	Error (°C)	Uncertainty (± °C)	Coverage Factor k
25.0	100	25.005	25.0	-0.005	0.16	2.00
30.0	100	30.009	30.0	-0.009	0.16	2.00
35.0	100	35.003	35.0	-0.003	0.16	2.00

UUC* : Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor **k**, providing a level of confidence of approximately 95 %.

-o0o-

เอกสารไม่ควบคุม

ภาคผนวก จ
หนังสืออนุญาตขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน



[illegible][illegible][illegible][illegible]

2017

1000

[Signature]

[illegible]

11/11/11 11:11 AM

ATC	Product	Special
1	Acid	1 Acid-liquid interaction, for Chromatography Method?
2	Alkaline	2 Alkaline-liquid interaction, for Chromatography Method?
3	Neutral	3 Neutral-liquid interaction, for Chromatography Method?
4	Acidic	4 Acidic-liquid interaction, for Chromatography Method?
5	Basic	5 Basic-liquid interaction, for Chromatography Method?
6	Amphoteric	6 Amphoteric-liquid interaction, for Chromatography Method?
7	Non-polar	7 Non-polar-liquid interaction, for Chromatography Method?
8	Polar	8 Polar-liquid interaction, for Chromatography Method?
9	Amphiphilic	9 Amphiphilic-liquid interaction, for Chromatography Method?
10	Hydrophobic	10 Hydrophobic-liquid interaction, for Chromatography Method?
11	Hydrophilic	11 Hydrophilic-liquid interaction, for Chromatography Method?
12	Amphiphilic	12 Amphiphilic-liquid interaction, for Chromatography Method?
13	Amphiphilic	13 Amphiphilic-liquid interaction, for Chromatography Method?
14	Amphiphilic	14 Amphiphilic-liquid interaction, for Chromatography Method?
15	Amphiphilic	15 Amphiphilic-liquid interaction, for Chromatography Method?
16	Amphiphilic	16 Amphiphilic-liquid interaction, for Chromatography Method?
17	Amphiphilic	17 Amphiphilic-liquid interaction, for Chromatography Method?
18	Amphiphilic	18 Amphiphilic-liquid interaction, for Chromatography Method?
19	Amphiphilic	19 Amphiphilic-liquid interaction, for Chromatography Method?
20	Amphiphilic	20 Amphiphilic-liquid interaction, for Chromatography Method?
21	Amphiphilic	21 Amphiphilic-liquid interaction, for Chromatography Method?
22	Amphiphilic	22 Amphiphilic-liquid interaction, for Chromatography Method?
23	Amphiphilic	23 Amphiphilic-liquid interaction, for Chromatography Method?
24	Amphiphilic	24 Amphiphilic-liquid interaction, for Chromatography Method?
25	Amphiphilic	25 Amphiphilic-liquid interaction, for Chromatography Method?
26	Amphiphilic	26 Amphiphilic-liquid interaction, for Chromatography Method?
27	Amphiphilic	27 Amphiphilic-liquid interaction, for Chromatography Method?
28	Amphiphilic	28 Amphiphilic-liquid interaction, for Chromatography Method?
29	Amphiphilic	29 Amphiphilic-liquid interaction, for Chromatography Method?
30	Amphiphilic	30 Amphiphilic-liquid interaction, for Chromatography Method?
31	Amphiphilic	31 Amphiphilic-liquid interaction, for Chromatography Method?
32	Amphiphilic	32 Amphiphilic-liquid interaction, for Chromatography Method?
33	Amphiphilic	33 Amphiphilic-liquid interaction, for Chromatography Method?
34	Amphiphilic	34 Amphiphilic-liquid interaction, for Chromatography Method?
35	Amphiphilic	35 Amphiphilic-liquid interaction, for Chromatography Method?
36	Amphiphilic	36 Amphiphilic-liquid interaction, for Chromatography Method?
37	Amphiphilic	37 Amphiphilic-liquid interaction, for Chromatography Method?
38	Amphiphilic	38 Amphiphilic-liquid interaction, for Chromatography Method?
39	Amphiphilic	39 Amphiphilic-liquid interaction, for Chromatography Method?
40	Amphiphilic	40 Amphiphilic-liquid interaction, for Chromatography Method?
41	Amphiphilic	41 Amphiphilic-liquid interaction, for Chromatography Method?
42	Amphiphilic	42 Amphiphilic-liquid interaction, for Chromatography Method?
43	Amphiphilic	43 Amphiphilic-liquid interaction, for Chromatography Method?
44	Amphiphilic	44 Amphiphilic-liquid interaction, for Chromatography Method?
45	Amphiphilic	45 Amphiphilic-liquid interaction, for Chromatography Method?
46	Amphiphilic	46 Amphiphilic-liquid interaction, for Chromatography Method?
47	Amphiphilic	47 Amphiphilic-liquid interaction, for Chromatography Method?
48	Amphiphilic	48 Amphiphilic-liquid interaction, for Chromatography Method?
49	Amphiphilic	49 Amphiphilic-liquid interaction, for Chromatography Method?
50	Amphiphilic	50 Amphiphilic-liquid interaction, for Chromatography Method?

14-00000

Source: U.S. Census Bureau.

[illegible]

Author's Note: I thank the following people for their comments on earlier drafts of this article: David G. Myers, David A. Reardon, and two anonymous reviewers for *Journal of Management Inquiry*.

[illegible]

Acknowledgements

[illegible]

ST. BARNABAS' CATHEDRAL

#	name	file
1	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
2	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
3	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
4	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
5	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
6	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
7	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
8	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
9	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
10	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
11	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
12	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
13	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
14	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
15	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
16	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
17	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
18	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
19	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe
20	EmptyListGenerator	C:\Program Files\Foxit Software\Foxit Reader\Foxit Reader.exe

doi:10.1017/S0022292412001509

[illegible]

© 2004 Blackwell Publishing Ltd *Journal of Internal Medicine* 255: 105–112

[illegible]

#	problem	theorem
10	Time	1) Mark's Theorem, Davenport's Periodicity Lemma Algebra Method ^[1,10]
11	Ordering	2) Davenport, Intensity Coupled Theory Method ^[10] 3) Mark's Theorem, Davenport's Intensity Coupled Algebra Method ^[1,10]
12	Nonperiodic	2) Davenport, Intensity Coupled Theory Method ^[10] 3) Mark's Theorem, Davenport's Periodicity Lemma and Intensities, See Chronomorphisms Method ^[10] 2) Davenport's Intensity Coupled Theory Method ^[10]
13	Nonperiodic sums	1) Mark's Theorem, Poisson's Top, UK, Chronomorphisms Method ^[10] 2) Poisson's Top, See Chronomorphisms Method Chronomorphisms Method ^[10]
14	periodicity	1) Mark's Theorem, Davenport's Intensity Coupled Algebra Method ^[1,10]
15	Dir	2) Davenport, Intensity Coupled Theory Method ^[10] 2) Mark's Theorem, Davenport's Periodicity Lemma and Intensities, See Chronomorphisms Method ^[10] 2) Mark's Theorem, Davenport's Intensity Coupled Algebra Method ^[1,10] 2) Davenport's Periodicity Lemma, Davenport's Method ^[10] 4) Davenport, Intensity Coupled Theory Method ^[10]

Abstract

Index	Code	Text
1	1000000000	1000000000 (10^9) (10^9)
2	1000000000	1000000000 (10^9) (10^9)

[illegible]

#A	Profile	Related
1	Perceptual-motor	1. Hübner, Hübner, Die Chinesenpsychologie Stuttgart 1977
10	Intuition	2. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
11	EGG-stressful situation	3. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
18	EGG-stressful situation	4. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
2	Emotional-motor	5. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
3	Perception	6. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
4	Perception	7. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
5	Perception	8. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
6	Perception	9. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
7	Perception	10. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
8	Perception	11. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
9	Perception	12. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
12	Perception	13. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
13	Perception	14. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
14	Perception	15. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
15	Perception	16. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
16	Perception	17. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
17	Perception	18. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
19	Perception	19. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
20	Perception	20. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
21	Perception	21. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
22	Perception	22. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
23	Perception	23. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
24	Perception	24. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
25	Perception	25. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
26	Perception	26. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
27	Perception	27. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
28	Perception	28. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
29	Perception	29. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978
30	Perception	30. Hübner, Hübner, Die Chinesenpsychologie, Paderborn Schöningh Verlag 1978

[illegible]

© 2000 Blackwell Science Ltd *Journal of Internal Medicine* 247: 399–406

[illegible]

— 經濟學與政治學 —

[illegible]

2.2.2. Theoretical Basis

#Ref	variable	reference
[10]	1,2,3-Dichloropropane	Fraser and Topp, Gas Chromatography/HPLC Separation Methods TM
[11]	1,2,4-Trichlorobenzene	Fraser and Topp, Gas Chromatography/HPLC Separation Methods TM
[12]	1,2-Dichloroethane	Fraser and Topp, Gas Chromatography/HPLC Separation Methods TM
[13]	1,2,4-Trichlorobenzene	Chromatography Methods, Gas Chromatography/HPLC Separation Methods TM
[14]	1,2,4-Trichlorobenzene	Chromatography Methods, Gas Chromatography/HPLC Separation Methods TM
[15]	1,2,3,4-Tetrachlorobenzene	Fraser and Topp, Gas Chromatography/HPLC Separation Methods TM
[16]	Hexachloro-	Organic Chemistry, General Physical Methods TM
[17]	Chloroacetaldehyde	Fraser and Topp, Gas Chromatography/HPLC Separation Methods TM
[18]	Trichloroethylene	Fraser and Topp, Gas Chromatography/HPLC Separation Methods TM
[19]	1,2-Dichloro-	Fraser and Topp, Gas Chromatography/HPLC Separation Methods TM
[20]	1,2-Dichloro-	Fraser and Topp, Gas Chromatography/HPLC Separation Methods TM
[21]	1,2-Dichloro-	Fraser and Topp, Gas Chromatography/HPLC Separation Methods TM
[22]	1,2-Dichloro-	Fraser and Topp, Gas Chromatography/HPLC Separation Methods TM
[23]	1,2-Dichloro-	Fraser and Topp, Gas Chromatography/HPLC Separation Methods TM
[24]	1,2-Dichloro-	Fraser and Topp, Gas Chromatography/HPLC Separation Methods TM
[25]	1,2-Dichloro-	Fraser and Topp, Gas Chromatography/HPLC Separation Methods TM

continued...

1. *Journal of the American Medical Association*, 2000; 283: 2601-2606. (The authors conclude that the use of a computerized decision support system in the management of patients with heart failure is associated with a significant improvement in patient outcomes.)

1. Introduction

