



**CLC**  
Accredited  
ISO/IEC 17025

# CALIBRATION LABORATORY Co., LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



**CONDITION OF CALIBRATION ITEM : GOOD**

**MEASUREMENT RESULTS : ( ) without adjustment ( X ) adjustment**

## CALIBRATION DATA

### 1. Error of indications [ Before Adjustment ]

Nominal Test Value ( g )	Conventional mass ( g )	Display Value ( g )	Error of Balance ( g )	Uncertainty $\pm$ ( mg )	Coverage factor k
10.0000	10.0000	9.9999	-0.0001	-	-
20.0000	20.0000	19.9994	-0.0006	-	-
50.0000	50.0000	49.9985	-0.0015	-	-
100.0000	100.0000	99.9978	-0.0022	-	-
200.0000	199.9997	199.9954	-0.0043	-	-

### 2. Error of indications [ After Adjustment ]

Nominal Test Value ( g )	Conventional mass ( g )	Display Value ( g )	Error of Balance ( g )	Uncertainty $\pm$ ( mg )	Coverage factor k
Unload	0.0000	0.0000	0.0000	0.04	2,28
0.2000	0.2000	0.1999	-0.0001	0.07	2,00
0.5000	0.5000	0.5000	0.0000	0.07	2,00
1.0000	1.0000	1.0000	0.0000	0.07	2,00
2.0000	2.0000	2.0000	0.0000	0.07	2,00
5.0000	5.0000	5.0001	+0.0001	0.07	2,00
10.0000	10.0000	10.0000	0.0000	0.07	2,00
20.0000	20.0000	20.0000	0.0000	0.08	2,00
50.0000	50.0000	50.0000	0.0000	0.08	2,00
100.0000	100.0000	99.9999	-0.0001	0.10	2,00
200.0000	199.9997	199.9998	+0.0001	0.16	2,00

Certificate No. Q21065289

F3-011-04/01-12

page 3 of 4

