

ภาคผนวกที่ 4

สำเนาเอกสารขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
บริษัท เอ็นไวรอนเมนต์ รีเสิร์ช แอนด์ เทคโนโลยี

ที่ อก ๐๓๑๐(๑)/ ๗ ๓ ๒๕



กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท
เขตราชเทวี กรุงเทพฯ ๑๐๕๐๐

๒๙ กรกฎาคม ๒๕๖๕

เรื่อง ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

เรียน กรรมการผู้จัดการ บริษัท เอ็นไวรอนเมนต์ รีเสิร์ช แอนด์ เทคโนโลยี จำกัด

อ้างถึง คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และชนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน
ลงวันที่ ๓๐ มีนาคม ๒๕๖๔

- สิ่งที่ส่งมาด้วย ๑. รายชื่อผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ จำนวน ๑ แผ่น
๒. รายชื่อเจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๒ แผ่น
๓. ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๑๒ แผ่น

ตามหนังสือที่อ้างถึง บริษัท เอ็นไวรอนเมนต์ รีเสิร์ช แอนด์ เทคโนโลยี จำกัด ขอต่ออายุ
หนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน เลขทะเบียน ว-๐๙๙ สถานที่ตั้งเลขที่ ๒๕/๑๑๔ หมู่ที่ ๖
ซอยชินเขต ๑ ถนนงามวงศ์วาน แขวงทุ่งสองห้อง เขตหลักสี่ กรุงเทพมหานคร ต่อกรมโรงงานอุตสาหกรรม นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว ให้บริษัท เอ็นไวรอนเมนต์ รีเสิร์ช แอนด์ เทคโนโลยี จำกัด
ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน โดยมีองค์ประกอบดังนี้

- ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ จำนวน ๑๖ ราย ตามสิ่งที่ส่งมาด้วย ๑
ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๔๙ ราย ตามสิ่งที่ส่งมาด้วย ๒
ค. ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนให้วิเคราะห์ในน้ำเสีย จำนวน ๒๗ รายการ น้ำใต้ดิน
จำนวน ๕๘ รายการ อากาศเสีย จำนวน ๒๖ รายการ สิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว จำนวน ๒๐ รายการ และ
ดิน จำนวน ๕๖ รายการ รวมทั้งสิ้นจำนวน ๑๘๗ รายการ ตามสิ่งที่ส่งมาด้วย ๓

หนังสือฉบับนี้จะหมดอายุในวันที่ ๑๘ พฤษภาคม ๒๕๖๗ หากประสงค์จะต่ออายุหนังสือ
รับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ให้ยื่นคำขอต่ออายุพร้อมเอกสารประกอบคำขอต่อ
กรมโรงงานอุตสาหกรรมภายใน ๓๐ วัน ก่อนวันสิ้นอายุของหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
ซึ่งคำขอต่ออายุดังกล่าวขอรับได้ที่กรมโรงงานอุตสาหกรรม

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ

(นางจินดา เตชะศรีรินทร์)

ผู้อำนวยการกองวิจัยและเตือนภัยมลพิษโรงงาน
ปฏิบัติราชการแทนอธิบดีกรมโรงงานอุตสาหกรรม

กองวิจัยและเตือนภัยมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๒๐๒ ๔๐๐๒ ๐ ๒๒๐๒ ๔๑๔๖

โทรสาร ๐ ๒๓๕๔ ๓๔๑๕

เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท เอ็นไวรอนเมนต์ รีเสิร์ช แอนด์ เทคโนโลยี จำกัด

เลขทะเบียน ว-๐๙๙

ที่ อก ๐๓๑๐(๑)/พ.๒๕

ลงวันที่ ๒๙ กรกฎาคม ๒๕๖๕

ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ จำนวน ๑๖ ราย

๑) นางสาวปณิชา พรหมชัย	ทะเบียนเลขที่ ว-๐๙๙-ค-๒๔๑๔
๒) นางณัฐรดา เลี้ยงรักษา	ทะเบียนเลขที่ ว-๐๙๙-ค-๓๐๐๒
๓) นายมงคล บุรภักดี	ทะเบียนเลขที่ ว-๐๙๙-ค-๕๕๐๐
๔) นางสาวธนิดา บุญรุ่งเรือง	ทะเบียนเลขที่ ว-๐๙๙-ค-๗๐๒๓
๕) นางสาวมิตา แดงไทย	ทะเบียนเลขที่ ว-๐๙๙-ค-๗๖๖๔
๖) นางสาวไรวินทร์ โพธิ์สิทธิ์	ทะเบียนเลขที่ ว-๐๙๙-ค-๗๖๖๕
๗) นางสาวณัฐนิชา เสริมมตวงศ์	ทะเบียนเลขที่ ว-๐๙๙-ค-๗๖๖๖
๘) นายณพลสิทธิ์ ทวีพรประดิษฐ์	ทะเบียนเลขที่ ว-๐๙๙-ค-๗๖๖๗
๙) นางสาวธิดารัตน์ ปุ๊กกะ	ทะเบียนเลขที่ ว-๐๙๙-ค-๘๘๐๑
๑๐) นายอภิชาติ พูลพล	ทะเบียนเลขที่ ว-๐๙๙-ค-๘๘๐๒
๑๑) นายนิทัศน์ ศิริชาติ	ทะเบียนเลขที่ ว-๐๙๙-ค-๘๘๐๓
๑๒) นายสุทธิชาญ สังข์ทอง	ทะเบียนเลขที่ ว-๐๙๙-ค-๘๘๐๔
๑๓) นางสาวยุวดี ณ ระนอง	ทะเบียนเลขที่ ว-๐๙๙-ค-๘๘๐๕
๑๔) นางสาววาสนา ชันเงิน	ทะเบียนเลขที่ ว-๐๙๙-ค-๘๘๐๖
๑๕) นางสาวสุภาวรรณ สุวรรณภา	ทะเบียนเลขที่ ว-๐๙๙-ค-๘๘๐๗
๑๖) นางสาวนภาพร จรัส หมื่นวงศ์	ทะเบียนเลขที่ ว-๐๙๙-ค-๘๘๐๘

เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
บริษัท เอ็นไวรอนเม้นท์ รีเสิร์ช แอนด์ เทคโนโลยี จำกัด เลขทะเบียน ว-๐๙๙
ที่ อก ๐๓๑๐(๑)/ ลงวันที่

ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๔๙ ราย

๑) นางสาวเปรมวดี ปุริโธสง	ทะเบียนเลขที่	ว-๐๙๙-จ-๕๕๐๒
๒) นางสาวจิตตวรรณ ลิ้มสมบุญ	ทะเบียนเลขที่	ว-๐๙๙-จ-๗๐๒๖
๓) นางสาวณัชร คนแรง	ทะเบียนเลขที่	ว-๐๙๙-จ-๗๐๒๙
๔) นางสาวสุตารัตน์ เขจรักษ์	ทะเบียนเลขที่	ว-๐๙๙-จ-๗๐๓๗
๕) นางสาวลิตา โพธิ์เจริญ	ทะเบียนเลขที่	ว-๐๙๙-จ-๗๐๔๒
๖) นางสาวรัชนิวรรณ ภูประเสริฐ	ทะเบียนเลขที่	ว-๐๙๙-จ-๗๐๔๔
๗) นายภาณุพล โพธิ์แดง	ทะเบียนเลขที่	ว-๐๙๙-จ-๗๐๔๕
๘) นายวันชนะ สีหามาตร	ทะเบียนเลขที่	ว-๐๙๙-จ-๗๐๕๐
๙) นายโสพล ป้อยแก้ว	ทะเบียนเลขที่	ว-๐๙๙-จ-๗๐๕๔
๑๐) นายอภิวัฒน์ ชำนาญเวช	ทะเบียนเลขที่	ว-๐๙๙-จ-๗๐๕๖
๑๑) นางสาวอชิรญาณ์ฐ์ อ่อนน้อม	ทะเบียนเลขที่	ว-๐๙๙-จ-๗๖๗๑
๑๒) นายวัชรางกูร กองแสง	ทะเบียนเลขที่	ว-๐๙๙-จ-๗๖๗๒
๑๓) นางสาวสุธาทิพย์ อิ่มน้อย	ทะเบียนเลขที่	ว-๐๙๙-จ-๗๖๗๓
๑๔) นายชยณัฐ บุญก้านตง	ทะเบียนเลขที่	ว-๐๙๙-จ-๗๖๗๕
๑๕) นางสาวพิชดา เขียววรภัย	ทะเบียนเลขที่	ว-๐๙๙-จ-๗๖๗๖
๑๖) นางสาวสายใจ ลาดบัวขาว	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๑๐
๑๗) นางสาวรัตนภรณ์ วงศ์ประโคน	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๑๒
๑๘) นางสาวจารุวรรณ แป้นจำนงค์	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๑๓
๑๙) นางสาวชมพูนุท กสิชีวิน	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๑๕
๒๐) นางสาวรวีวรรณ สุขารมย์	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๑๖
๒๑) นางสาวนัฐภรณ์ กันสุข	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๑๗
๒๒) นางสาวอรอนงค์ นวนนุ้ม	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๑๘
๒๓) นางสาวสรวรรณ พุดพินมาต	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๑๙
๒๔) นางสาวกัญญาลักษณ์ กระทาง	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๒๑
๒๕) นางสาวปิยธิดา ประแดงโค	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๒๓
๒๖) นางสาวปวีตรา นาเหล็ก	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๒๔
๒๗) นางสาวชนิดา นิลผาย	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๒๕
๒๘) นางสาวพิยะดา จารุไชย	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๒๖
๒๙) นางสาวทักษพร ไกรสิงห์	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๒๗
๓๐) นางสาวฉวีวรรณ บุญจันทิก	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๒๘
๓๑) นางสาวเบญจวรรณ คำหงษา	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๒๙
๓๒) นางสาวพัชชา แก้วย้อย	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๓๐
๓๓) นางสาวณัฐชา สัมฤทธิ์ดี	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๓๑
๓๔) นางสาวอังคณา อุ่นตา	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๓๓
๓๕) นางสาวบุศดี มุภาษา	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๓๔

๓๖) นายรอมซี...

๓๖) นายรอมชี กาเต๊ะ	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๓๕
๓๗) นายสุริยะ ชูทอง	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๓๖
๓๘) นายศักรินทร์ นิภานันท์	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๓๗
๓๙) นายอภิเดช ยาสมดี	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๓๘
๔๐) นายฉันทวิทย์ เหลวกุล	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๓๙
๔๑) นายศิวาวุธ ธรรมนิทา	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๔๐
๔๒) นายณัฐพล สุทธิมล	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๔๑
๔๓) นายอาทิตย์ นุชบุษบา	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๔๒
๔๔) นายอนุวัฒน์ เรืองอ่อน	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๔๔
๔๕) นายฉัตรชัย โยวะผุย	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๔๕
๔๖) นายกลยุทธิ์ อินทร์คำ	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๔๗
๔๗) นางสาวนันทชา เนื่อนวล	ทะเบียนเลขที่	ว-๐๙๙-จ-๘๘๔๘
๔๘) นางสาวพิไลวรรณ แปงทา	ทะเบียนเลขที่	ว-๐๙๙-จ-๙๕๒๑
๔๙) นางสาวจารุวรรณ กระจำพันธุ์	ทะเบียนเลขที่	ว-๐๙๙-จ-๙๕๒๒

เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
บริษัท เอ็นไวรอนเม้นท์ รีเสิร์ช แอนด์ เทคโนโลยี จำกัด เลขทะเบียน ว-๐๙๙
ที่ อก ๐๓๑๐(๑)/ ลงวันที่

ขอขยายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๑๘๗ รายการ

น้ำเสีย จำนวน 27 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Arsenic	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[3]
2	Barium	2) Digestion, Inductively Coupled Plasma Method ^[3]
3	Biochemical Oxygen Demand	Digestion, Inductively Coupled Plasma Method ^[3] 1) 5-Day BOD Test, Azide Modification Method ^[3] 2) 5-Day BOD Test, Membrane Electrode Method ^[3]
4	Cadmium	Digestion, Inductively Coupled Plasma Method ^[3]
5	Chemical Oxygen Demand	Closed Reflux, Titrimetric Method ^[3]
6	Chromium	Digestion, Inductively Coupled Plasma Method ^[3]
7	Color	ADMI Weighted-Ordinate Spectrophotometric Method ^[3]
8	Copper	Digestion, Inductively Coupled Plasma Method ^[3]
9	Cyanide	Distillation, Colorimetric method ^[3]
10	Formaldehyde	Distillation, Colorimetric Method ^[2]
11	Free Chlorine	1) Iodometric Method ^[3] 2) DPD Colorimetric Method ^[3]
12	Hexavalent Chromium	Colorimetric Method ^[3]
13	Lead	1) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
14	Manganese	Digestion, Inductively Coupled Plasma Method ^[3]
15	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[3]
16	Nickel	Digestion, Inductively Coupled Plasma Method ^[3]
17	Oil & Grease	Liquid-Liquid, Partition-Gravimetric Method ^[3]
18	pH	Electrometric Method ^[3]
19	Phenols	Distillation, Direct Photometric Method ^[3]
20	Selenium	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
21	Sulfide	Iodometric method ^[3]



(นางจิราญญา จิตรสกุลวิไล)

ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
เลขทะเบียนห้องปฏิบัติการ

22 Temperature...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
22	Temperature	Laboratory and Field Methods ^[3]
23	Total Dissolved Solids	Dried at 180 °C ^[3]
24	Total Kjeldahl Nitrogen	1) Macro Kjeldahl Method ^[3] 2) Semi-Micro Kjeldahl Method ^[3]
25	Total Suspended Solids	Dried at 103-105 °C ^[3]
26	Trivalent Chromium	Digestion, Inductively Coupled Plasma Method; Colorimetric Method; Calculation ^[3]
27	Zinc	Digestion, Inductively Coupled Plasma Method ^[3]

น้ำใต้ดิน จำนวน 58 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Acetone	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ^[3]
2	Antimony	Digestion, Inductively Coupled Plasma Method ^[3]
3	Arsenic	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
4	Barium	Digestion, Inductively Coupled Plasma Method ^[3]
5	Benzene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
6	Beryllium	Digestion, Inductively Coupled Plasma Method ^[3]
7	Bromodichloromethane	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
8	Bromoform	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
9	Cadmium	Digestion, Inductively Coupled Plasma Method ^[3]
10	Carbon Disulfide	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
11	Carbon Tetrachloride	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
12	Chlorobenzene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
13	Chlorodibromomethane	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]



(นางวิชาญจน์ จิตรสกุลวิไล)

ผู้อำนวยการกลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
และทะเบียนห้องปฏิบัติการ

14 Chloroform...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
14	Chloroform	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
15	Chromium	Digestion, Inductively Coupled Plasma Method ^[3]
16	Chromium (III)	Digestion, Inductively Coupled Plasma Method; Colorimetric Method; Calculation ^[3]
17	Chromium (VI)	Colorimetric Method ^[3]
18	Cyanide	Colorimetric Method ^[3]
19	1,2-Dichlorobenzene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
20	1,3-Dichlorobenzene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
21	1,4-Dichlorobenzene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
22	1,1-Dichloroethane	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
23	1,2-Dichloroethane	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
24	1,1-Dichloroethylene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
25	cis-1,2-Dichloroethylene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
26	trans-1,2-Dichloroethylene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
27	1,2-Dichloropropane	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
28	1,3-Dichloropropane	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
29	1,3-Dichloropropene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
30	Ethylbenzene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
31	Hexachloro-1,3-butadiene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]



(นางริกาญจน์ จิตรสกุลไค)

ผู้อำนวยการกลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบเคมี
กระทรวงมหาดไทย

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
32	Lead	1) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
33	Manganese	Digestion, Inductively Coupled Plasma Method ^[3]
34	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[3]
35	Methyl Bromide	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
36	Methylene Chloride	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
37	Methyl Tert-Butyl Ether	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
38	Naphthalene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
39	Nickel	Digestion, Inductively Coupled Plasma Method ^[3]
40	pH	Electrometric method ^[3]
41	Selenium	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
42	Silver	Digestion, Inductively Coupled Plasma Method ^[3]
43	Styrene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
44	1,1,2,2-Tetrachloroethane	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
45	Tetrachloroethylene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
46	Toluene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
47	1,2,4-Trichlorobenzene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
48	1,1,1-Trichloroethane	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
49	1,1,2-Trichloroethane	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]

วิภา

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
50	Trichloroethylene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
51	1,3,5-Trimethylbenzene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
52	Vanadium	Digestion, Inductively Coupled Plasma Method ^[3]
53	Vinyl Chloride	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
54	m-Xylene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
55	o-Xylene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
56	p-Xylene	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
57	Xylene (Total)	Purge and Trap Gas Chromatographic/ Mass spectrometric Method ^[3]
58	Zinc	Digestion, Inductively Coupled Plasma Method ^[3]

อากาศเสีย (ปล่อยระบาย) จำนวน 26 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Antimony	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[4]
2	Arsenic	1) Isokinetic Sampling, Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[4] 2) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[4]
3	Beryllium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[4]
4	Cadmium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[4]
5	Carbon Monoxide	Instrumental Analyzer Method ^[4]
6	Chlorine	1) Absorption Sampling, Ion Chromatographic Method ^[4] 2) Isokinetic Sampling, Ion Chromatographic Method ^[4]

พิมพ์

(นางวิภาญจน์ ฉัตรสกุลวิไล)

ผู้อำนวยการกลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
สภามหาวิทยาลัยเทคโนโลยีพระจอมเกล้าธนบุรี

7 Chromium...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
7	Chromium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[4]
8	Cobalt	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[4]
9	Copper	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[4]
10	Dioxin/Furans	Isokinetic Sampling ^[4]
11	Hydrogen Chloride	1) Absorption Sampling, Ion Chromatographic Method ^[4] 2) Isokinetic Sampling, Ion Chromatographic Method ^[4]
12	Hydrogen Fluoride	1) Absorption Sampling, Ion Chromatographic Method ^[4] 2) Isokinetic Sampling, Ion Chromatographic Method ^[4]
13	Hydrogen Sulfide	Absorption Sampling, Iodometric Method ^[4]
14	Lead	1) Isokinetic Sampling, Digestion, Direct Air-Acetylene Flame Method ^[4] 2) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[4]
15	Manganese	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[4]
16	Mercury	Isokinetic Sampling, Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[4]
17	Nickel	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[4]
18	Opacity	Ringelmann's Method ^[1]
19	Oxide of Nitrogen	1) Absorption Sampling, Phenoldisulfonic acid Method ^[4] 2) Instrumental Analyzer Method ^[4]
20	Selenium	1) Isokinetic Sampling, Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[4] 2) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[4]

วิมล

(นางวิภาญจน์ อัครสกุลวิไล)

ผู้อำนวยการกลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
สถานะเขียนห้องปฏิบัติการ

21 Sulfur...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
21	Sulfur Dioxide	1) Absorption Sampling, Barium-Thorin Titrimetric Method ^[4] 2) Isokinetic Sampling, Barium-Thorin Titrimetric Method ^[4] 3) Instrumental Analyzer Method ^[4]
22	Sulfuric Acid	Isokinetic Sampling, Barium-Thorin Titrimetric Method ^[4]
23	Tin	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[4]
24	Total Suspended Particulate	Isokinetic Sampling, Gravimetric Method ^[4]
25	Vanadium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[4]
26	Xylene	Adsorption Sampling, Gas Chromatographic Method ^[4]

สิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว จำนวน 20 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Antimony	Digestion, Inductively Coupled Plasma Method ^[5,8]
2	Arsenic	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[5,9] 2) Digestion, Inductively Coupled Plasma Method ^[5,8]
3	Barium	Digestion, Inductively Coupled Plasma Method ^[5,8]
4	Beryllium	Digestion, Inductively Coupled Plasma Method ^[5,8]
5	Cadmium	Digestion, Inductively Coupled Plasma Method ^[5,8]
6	Chromium	Digestion, Inductively Coupled Plasma Method ^[5,8]
7	Chromium (III)	Digestion, Inductively Coupled Plasma Method; Alkaline Digestion, Colorimetric Method; Calculation Method ^[5,6,8,10]
8	Chromium (VI)	Alkaline Digestion, Colorimetric Method ^[6,10]
9	Cobalt	Digestion, Inductively Coupled Plasma Method ^[5,8]
10	Copper	Digestion, Inductively Coupled Plasma Method ^[5,8]
11	Lead	Digestion, Inductively Coupled Plasma Method ^[5,8]
12	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[11]
13	Molybdenum	Digestion, Inductively Coupled Plasma Method ^[5,8]
14	Nickel	Digestion, Inductively Coupled Plasma Method ^[5,8]

วิมล

(นางวิมล ธีรกุลวิไล)

ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
กระทรวงมหาดไทย

15 pH...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
15	pH	Electrometric Method ^[14]
16	Selenium	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[5,12] 2) Digestion, Inductively Coupled Plasma Method ^[5,8]
17	Silver	Digestion, Inductively Coupled Plasma Method ^[5,8]
18	Thallium	Digestion, Inductively Coupled Plasma Method ^[5,8]
19	Vanadium	Digestion, Inductively Coupled Plasma Method ^[5,8]
20	Zinc	Digestion, Inductively Coupled Plasma Method ^[5,8]

ดิน จำนวน 56 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Acetone	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[7,13]
2	Antimony	Digestion, Inductively Coupled Plasma Method ^[5,8]
3	Arsenic	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[5,9] 2) Digestion, Inductively Coupled Plasma Method ^[5,8]
4	Barium	Digestion, Inductively Coupled Plasma Method ^[5,8]
5	Benzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[7,13]
6	Beryllium	Digestion, Inductively Coupled Plasma Method ^[5,8]
7	Bromodichloromethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[7,13]
8	Bromoform	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[7,13]
9	Cadmium	Digestion, Inductively Coupled Plasma Method ^[5,8]
10	Carbon Disulfide	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[7,13]
11	Carbon Tetrachloride	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[7,13]
12	Chlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[7,13]
13	Chlorodibromomethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[7,13]

31/10/2561

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
14	Chloroform	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
15	Chromium	Digestion, Inductively Coupled Plasma Method ^[5,8]
16	Chromium (III)	Digestion, Inductively Coupled Plasma Method; Colorimetric Method; Calculation Method ^[5,7,9,11]
17	Chromium (VI)	Alkaline Digestion, Colorimetric Method ^[7,11]
18	1,2-Dichlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
19	1,3-Dichlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
20	1,4-Dichlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
21	1,1-Dichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
22	1,2-Dichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
23	1,1-Dichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
24	cis-1,2-Dichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
25	trans-1,2-Dichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
26	1,2-Dichloropropane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
27	1,3-Dichloropropane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
28	1,3-Dichloropropene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
29	Ethylbenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
30	Hexachloro-1,3-butadiene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
31	Lead	Digestion, Inductively Coupled Plasma Method ^[5,8]
32	Manganese	Digestion, Inductively Coupled Plasma Method ^[5,8]
33	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[11]

วิมล


ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
34	Methyl Bromide	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
35	Methylene Chloride	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
36	Methyl Tert-Butyl Ether	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
37	Naphthalene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
38	Nickel	Digestion, Inductively Coupled Plasma Method ^[5,8]
39	Selenium	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[5,12] 2) Digestion, Inductively Coupled Plasma Method ^[5,8]
40	Silver	Digestion, Inductively Coupled Plasma Method ^[5,8]
41	Styrene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
42	1,1,2,2-Tetrachloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
43	Tetrachloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
44	Toluene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
45	1,2,4-Trichlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
46	1,1,1-Trichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
47	1,1,2-Trichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
48	Trichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
49	1,3,5-Trimethylbenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
50	Vanadium	Digestion, Inductively Coupled Plasma Method ^[5,8]
51	Vinyl Chloride	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]

วิมล

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
52	m-Xylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
53	o-Xylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
54	p-Xylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
55	Xylene (Total)	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[7,13]
56	Zinc	Digestion, Inductively Coupled Plasma Method ^[5,8]

เอกสารอ้างอิง

- กระทรวงอุตสาหกรรม. ประกาศกระทรวงอุตสาหกรรม, พ.ศ. 2549. เรื่อง กำหนดค่าปริมาณเคมีภัณฑ์ที่เจือปนในอากาศที่ระบายออกจากปล่องของหม้อน้ำโรงสีข้าวที่ใช้กลบเป็นเชื้อเพลิง.
- ราชกิจจานุเบกษา. 4 ธันวาคม 2549. เล่มที่ 123 ตอนพิเศษ 125ง.
- สมาคมวิศวกรรมสิ่งแวดล้อมแห่งประเทศไทย. คู่มือวิเคราะห์น้ำเสีย. พิมพ์ครั้งที่ 4. กรุงเทพฯ: เรือนแก้วการพิมพ์, 2547.
- APHA, AWWA, WEF. **Standard Methods for the Examination of Water and Wastewater**. 23rd ed. Washington, DC: APHA, 2017.
- United States Environmental Protection Agency. **Standards of Performance for New Stationary Sources**. 40 CFR 60. Appendix A, 2019.
- United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Acid Digestion of Sediments, Sludges, and Soils. SW-846 Method 3050B**, 1996.
- United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Alkaline Digestion for Hexavalent Chromium. SW-846 Method 3060A**, 1996.
- United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Closed-System Purge-and-Trap and Extraction for Volatile Organics in Soil and Waste Samples. SW-846 Method 5035A**, 2002.
- United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Inductively Coupled Plasma-optical Emission Spectrometry. SW-846 Method 6010D**, 2018
- United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Antimony and Arsenic (Atomic Absorption, Borohydride ReductionX. SW-846 Method 7062**, 1992.


 (นางวิภาญจน์ นัตถกุลวิไล)
 ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
 และทะเบียนห้องปฏิบัติการ

10. United...

10. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Chromium, Hexavalent (Colorimetric), SW-846 Method 7196A**, 1992.
11. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Mercury in Solid or Semisolid Waste (Manual Cold-Vapor Technique, SW-846 Method 7471B**, 2007.
12. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Selenium (Atomic Absorption, Borohydride Reduction), SW-846 Method 7742**, 1994.
13. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Volatile Organic Compounds by Gas Chromatography/ Mass Spectrometry (GC/MS). SW-846 Method 8260D**, 2018.
14. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Solid and Waste pH. SW-846 Method 9045D**, 2004.



(นางวิภาญจน์ อัครสกุลวิไล)

ผู้อำนวยการกลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
และทะเบียนห้องปฏิบัติการ



ที่ อก ๐๓๑๐(๑)/ ๒๐๓ ๙

กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท
เขตราชเทวี กรุงเทพฯ ๑๐๔๐๐

๑๐ กุมภาพันธ์ ๒๕๖๕

เรื่อง เปลี่ยนแปลงสารมลพิษที่วิเคราะห์

เรียน กรรมการผู้จัดการ บริษัท เอ็นไวรอนเมนต์ รีเสิร์ช แอนด์ เทคโนโลยี จำกัด

อ้างถึง คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และชนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน
ลงวันที่ ๒๔ ธันวาคม ๒๕๖๔

สิ่งที่ส่งมาด้วย เอกสารแนบท้ายหนังสือเปลี่ยนแปลงสารมลพิษที่วิเคราะห์

บริษัท เอ็นไวรอนเมนต์ รีเสิร์ช แอนด์ เทคโนโลยี จำกัด จำนวน ๑ แผ่น

ตามหนังสือที่อ้างถึง บริษัท เอ็นไวรอนเมนต์ รีเสิร์ช แอนด์ เทคโนโลยี จำกัด ห้องปฏิบัติการ
วิเคราะห์เอกชน เลขทะเบียน ว-๐๙๙ สถานที่ตั้งเลขที่ ๒๕/๑๑๔ หมู่ที่ ๖ ซอยชินเขต ๑ ถนนงามวงศ์วาน
แขวงทุ่งสองห้อง เขตหลักสี่ กรุงเทพมหานคร ขอเปลี่ยนแปลงสารมลพิษที่วิเคราะห์ ความละเอียดแจ้งแล้ว นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้วให้ บริษัท เอ็นไวรอนเมนต์ รีเสิร์ช แอนด์ เทคโนโลยี
จำกัด เพิ่มขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนให้วิเคราะห์ในดิน ตามสิ่งที่ส่งมาด้วย

อนึ่ง หนังสือฉบับนี้จะหมดอายุพร้อมหนังสือต่ออายุรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
ที่ อก ๐๓๑๐(๑)/๗๓๒๕ ลงวันที่ ๒๙ กรกฎาคม ๒๕๖๔ คือในวันที่ ๑๘ พฤษภาคม ๒๕๖๗ ทั้งนี้ สามารถยื่น
คำขอผ่านระบบอิเล็กทรอนิกส์ได้ที่หน้าเว็บไซต์กรมโรงงานอุตสาหกรรม ตาม QR Code ท้ายหนังสือฉบับนี้

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ

(นางจินดา เดชะศรีนทร์)

ผู้อำนวยการกองวิจัยและเตือนภัยมลพิษโรงงาน
ปฏิบัติราชการแทนอธิบดีกรมโรงงานอุตสาหกรรม



ยื่นคำขอผ่านระบบอิเล็กทรอนิกส์

กองวิจัยและเตือนภัยมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๐๓-๕ โทรสาร ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๙๙

ไปรษณีย์อิเล็กทรอนิกส์ saraban@diw.mail.go.th

เอกสารแนบท้ายหนังสือเปลี่ยนแปลงสารมลพิษที่วิเคราะห์

บริษัท เอ็นไวรอนเมนต์ รีเสิร์ช แอนด์ เทคโนโลยี จำกัด

เลขทะเบียน ว-๐๙๙

ที่ อก ๐๓๑๐(๑)/ ๒๐๓๙

ลงวันที่ ๑๐ กุมภาพันธ์ ๒๕๖๕

ขอขยายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๓ รายการ

ดิน จำนวน 3 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	TPH (C ₅ – C ₈)	Purge and Trap, Gas Chromatographic Method ^[2,3]
2	TPH (C _{>8} – C ₁₆)	Ultrasonic Extraction, Gas Chromatographic Method ^[1,3]
3	TPH (C _{>16} – C ₃₅)	Ultrasonic Extraction, Gas Chromatographic Method ^[1,3]

เอกสารอ้างอิง

1. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Ultrasonic Extraction. SW-846 Method 3550C**, 2007.
2. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Closed System Purge and Trap and Extraction for Volatile Organics in Soil and Waste Sample. SW-846 Method 5035A**, 2002.
3. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Nonhalogenated Organics Using GC/FID. SW-846 Method 8015D**, 2003



ที่ อก ๐๓๑๐(๑)/ ๑๑๕๖๗

กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท
เขตราชเทวี กรุงเทพฯ ๑๐๔๐๐

๒๔ สิงหาคม ๒๕๖๕

เรื่อง เปลี่ยนแปลงบุคลากรของห้องปฏิบัติการวิเคราะห์

เรียน กรรมการผู้จัดการ บริษัท เอ็นไวรอนเมนต์ รีเสิร์ช แอนด์ เทคโนโลยี จำกัด

อ้างถึง ๑. คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และชนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน
ลงวันที่ ๑๑ สิงหาคม ๒๕๖๕

๒. หนังสือบริษัท เอ็นไวรอนเมนต์ รีเสิร์ช แอนด์ เทคโนโลยี จำกัด ลงวันที่ ๑๑ สิงหาคม ๒๕๖๕

ตามหนังสือที่อ้างถึง ๑ และ ๒ บริษัท เอ็นไวรอนเมนต์ รีเสิร์ช แอนด์ เทคโนโลยี จำกัด
ห้องปฏิบัติการวิเคราะห์เอกชน เลขทะเบียน ว-๐๙๙ สถานที่ตั้งเลขที่ ๒๕/๑๑๔ หมู่ที่ ๖ ซอยชินเขต ๑
ถนนงามวงศ์วาน แขวงทุ่งสองห้อง เขตหลักสี่ กรุงเทพมหานคร ขอเปลี่ยนแปลงบุคลากรของห้องปฏิบัติการวิเคราะห์
ความละเอียดแจ้งแล้ว นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว มีความเห็นดังนี้

๑. ให้ยกเลิกเจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๙ ราย

- | | |
|------------------------------|----------------------------|
| ๑) นายโสพล ป้อยแก้ว | ทะเบียนเลขที่ ว-๐๙๙-จ-๗๐๕๔ |
| ๒) นางสาวอชิรญาณัฐ อ่อนน้อม | ทะเบียนเลขที่ ว-๐๙๙-จ-๗๖๗๑ |
| ๓) นางสาวรัตนภรณ์ วงศ์ประโคน | ทะเบียนเลขที่ ว-๐๙๙-จ-๘๘๑๒ |
| ๔) นางสาวสรวรรณ พุ่มพินมาต | ทะเบียนเลขที่ ว-๐๙๙-จ-๘๘๑๙ |
| ๕) นางสาวปิยะดา จารุไชย | ทะเบียนเลขที่ ว-๐๙๙-จ-๘๘๒๖ |
| ๖) นางสาวฉวีวรรณ บุญจันทิก | ทะเบียนเลขที่ ว-๐๙๙-จ-๘๘๒๘ |
| ๗) นายศักรินทร์ นิภานันท์ | ทะเบียนเลขที่ ว-๐๙๙-จ-๘๘๓๗ |
| ๘) นายอภิเดช ยาสมดี | ทะเบียนเลขที่ ว-๐๙๙-จ-๘๘๓๘ |
| ๙) นางสาวพิไลวรรณ แปงทา | ทะเบียนเลขที่ ว-๐๙๙-จ-๙๕๒๑ |

๒. ให้เพิ่มเจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๑๕ ราย

- | | |
|--|----------------------------|
| ๑) ว่าที่ร้อยตรีหญิงภัทรนันท์ วิจิตรศักดิ์ | ทะเบียนเลขที่ ว-๐๙๙-จ-๐๐๐๑ |
| ๒) นางสาวณัฐธิดา ขาวสุทธิ | ทะเบียนเลขที่ ว-๐๙๙-จ-๐๐๐๒ |
| ๓) นางสาวเพชรภรณ์ พงษ์พันธ์ | ทะเบียนเลขที่ ว-๐๙๙-จ-๐๐๐๓ |
| ๔) นางสาวพัทธนันท์ คำยา | ทะเบียนเลขที่ ว-๐๙๙-จ-๐๐๐๔ |
| ๕) นางสาวสุธิดา ทองประภา | ทะเบียนเลขที่ ว-๐๙๙-จ-๐๐๐๕ |
| ๖) นางสาวรมย์ชลี เดือนแร่รัมย์ | ทะเบียนเลขที่ ว-๐๙๙-จ-๐๐๐๖ |
| ๗) นายจิรยุทธ์ สามารถ | ทะเบียนเลขที่ ว-๐๙๙-จ-๐๐๐๗ |
| ๘) นายอัษฎา ไชยวงศ์ | ทะเบียนเลขที่ ว-๐๙๙-จ-๐๐๐๘ |
| ๙) นางสาวณัฐริสา บุญหนัก | ทะเบียนเลขที่ ว-๐๙๙-จ-๐๐๐๙ |
| ๑๐) นางสาวสุพัตรา สุนทร | ทะเบียนเลขที่ ว-๐๙๙-จ-๐๐๑๐ |

๑๑) นายพงศ์ปวีร์ สัตระ

ทะเบียนเลขที่ ว-๐๙๙-จ-๐๐๑๑

๑๒) นายนฤตม์ โชติกาญจน์

ทะเบียนเลขที่ ว-๐๙๙-จ-๐๐๑๒

๑๓) นางสาวพรทิพย์ อัมภรัตน์

ทะเบียนเลขที่ ว-๐๙๙-จ-๐๐๑๓

๑๔) นางสาวจันทน์ ปิตพิทพงศ์

ทะเบียนเลขที่ ว-๐๙๙-จ-๐๐๑๔

๑๕) นายอัศวิน คชบก

ทะเบียนเลขที่ ว-๐๙๙-จ-๐๐๑๕

๓. ให้เปลี่ยนชื่อผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ จากเดิมนางสาววาสนา ชื่นเงิน ทะเบียนเลขที่ ว-๐๙๙-ค-๘๘๐๖ เป็น นางสาวถิรณัฐ ชื่นเงิน

๔. ให้เปลี่ยนชื่อ-สกุลเจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จากเดิมนางสาวเปรมวดี ปุริโสง ทะเบียนเลขที่ ว-๐๙๙-จ-๕๔๐๒ เป็น นางเตชินี สืบเสระ

อนึ่ง หนังสือฉบับนี้จะหุ้มด้วยพร้อมหนังสือต่ออายุรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ที่ อก-๐๓๑๐(๑)/๗๓๒๕ ลงวันที่ ๒๙ กรกฎาคม ๒๕๖๔ คือในวันที่ ๑๘ พฤษภาคม ๒๕๖๗ ทั้งนี้ สามารถยื่นคำขอผ่านระบบอิเล็กทรอนิกส์ได้ที่หน้าเว็บไซต์กรมโรงงานอุตสาหกรรม ตาม QR Code ท้ายหนังสือฉบับนี้

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ



(นางจันทา เดชะศรีนท)

ผู้อำนวยการกองวิจัยและเตือนภัยมลพิษโรงงาน

ปฏิบัติราชการแทนอธิบดีกรมโรงงานอุตสาหกรรม



ยื่นคำขอผ่านระบบอิเล็กทรอนิกส์

กองวิจัยและเตือนภัยมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๐๓-๕

โทรสาร ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๙๙

ไปรษณีย์อิเล็กทรอนิกส์ saraban@diw.mail.go.th



ภาคผนวกที่ 5

เอกสารสอบเทียบเครื่องมือตรวจวัด

Calibration Report

Customer Name : ICS Co., Ltd.
Address : 989 Siampiwat Tower, 20th Floor, Unit A1, Rama 1 Road, Pathum Wan, Bangkok 10330
Project Name : โครงการ RTWO
Sampling Date : July - November, 2022

Ambient

Item	Equipment	Manufacturer	Model	Serial Number	Calibration Date
1	TSP High-volume No. A18	Thermo Andersen	HIVOL-BBCBE	2014-03	July 23, 2022
2	TSP High-volume No. A10	Thermo Andersen	HIVOL-BBCBE	2012-04	July 23 and September 21, 2022
3	TSP High-volume No. A22	Thermo Scientific	HIVOL-BBCBE	2054	July 23, 2022
4	High-volume PM-10 No. 12	GRASEBY	HIVOL-BMBBE	B2012-10	July 23, 2022
5	High-volume PM-10 No. 19	Thermo Scientific	HIVOL-BMBBE	2133	July 23, 2022
6	High-volume PM-10 No. 21	Thermo Scientific	HIVOL-BMBBE	2132	July 23, 2022
7	TSP High-volume No. A7	Thermo Andersen	HIVOL-BBCBE	B-NO A.7	August 10, 2022
8	TSP High-volume No. A8	Thermo Andersen	HIVOL-BBCBE	BLA0902	August 10, 2022
9	TSP High-volume No. A9	Thermo Andersen	HIVOL-BBCBE	16119	August 10 and October 19, 2022
10	High-volume PM-10 No. 8	GRASEBY	HIVOL-BMBBE	BLA0901	August 10, 2022
11	High-volume PM-10 No. 5	Thermo Andersen	HIVOL-BMBBE	2015-5	August 10, 2022
12	High-volume PM-10 No. 12	GRASEBY	HIVOL-BMBBE	B2012-10	August 10, 2022

Calibration Report

Customer Name : ICS Co., Ltd.
Address : 989 Siampiwat Tower, 20th Floor, Unit A1, Rama 1 Road, Pathum Wan, Bangkok 10330
Project Name : โครงการ RTWO
Sampling Date : July - November, 2022

Ambient

Item	Equipment	Manufacturer	Model	Serial Number	Calibration Date
13	TSP High-volume No. A2	Thermo Andersen	HIVOL-BBCBE	TSP No.A2	September 21, 2022
14	TSP High-volume No. 8	Thermo Andersen	HIVOL-BMBBE	BLA0902	September 21, 2022
15	High-volume PM-10 No. 10	Thermo Andersen	HIVOL-BMBBE	610-047	September 21, 2022
16	High-volume PM-10 No. 12	GRASEBY	HIVOL-BMBBE	610-743	September 21, 2022
17	High-volume PM-10 No. 3	Thermo Andersen	HIVOL-BMBBE	2015-3	September 21, 2022
18	TSP High-volume No. A25	Thermo Scientific	HIVOL-BBCBE	2157	October 19, 2022
19	TSP High-volume No. A28	Thermo Scientific	HIVOL-BBCBE	2217	October 19, 2022
20	High-volume PM-10 No. 7	Thermo Scientific	HIVOL-BMBBE	B0411-001	October 19, 2022
21	High-volume PM-10 No. 12	GRASEBY	HIVOL-BMBBE	PM-10 No.12	October 19, 2022
22	High-volume PM-10 No. 19	Thermo Scientific	HIVOL-BMBBE	2133	October 19, 2022
23	TSP High-volume No. A26	Thermo Scientific	HIVOL-BBCBE	2141	November 28, 2022
24	TSP High-volume No. C28	Local	HIVOL-BBCBE	2022-1	November 28, 2022
25	TSP High-volume No. B3	Local	HIVOL-BBCBE	2022-02	November 28, 2022

Calibration Report

Customer Name : ICS Co., Ltd.
Address : 989 Siampiwat Tower, 20th Floor, Unit A1, Rama 1 Road, Pathum Wan, Bangkok 10330
Project Name : โครงการ RTWO
Sampling Date : July - November, 2022

Ambient

Item	Equipment	Manufacturer	Model	Serial Number	Calibration Date
26	High-volume PM-10 No. 28	Thermo Scientific	HIVOL-BMBBE	2206	November 28, 2022
27	High-volume PM-10 No. 18	Thermo Scientific	HIVOL-BMBBE	2139	November 28, 2022
28	Orifice	TISCH	TE-5025A	3362	January 29, 2021
29	Orifice	TISCH	TE-5025A	2913	March 8, 2022
30	Orifice	TISCH	TE-5025A	3141	May 2, 2022
31	Orifice	TISCH	TE-5025A	2915	January 18, 2021
32	Electronic Balance	METTLER TOLEDO	AB204-S	1123103723	January 4, 2022

Calibration Report

Customer Name : ICS Co., Ltd.
Address : 989 Siampiwat Tower, 20th Floor, Unit A1, Rama 1 Road, Pathum Wan, Bangkok 10330
Project Name : โครงการ RTWO
Sampling Date : July - November, 2022

Ambient

Item	Equipment	Manufacturer	Model	Serial Number	Calibration Date
33	NOx Chemiluminescence	Horiba	APNA-370	YCPL4HTM	July 23, 2022
34	NOx Chemiluminescence	Horiba	APNA-370	AX7HSMEO	July 23, 2022
35	NOx Chemiluminescence	API	200A	979	July 23, 2022
36	SO ₂ UV-Fluorescence	Horiba	APSA-370	A5VTX5AF	July 23, 2022
37	SO ₂ UV-Fluorescence	Horiba	APSA-370	JH9GS3FU	July 23, 2022
38	SO ₂ UV-Fluorescence	Thermo	43C	335804022	July 23, 2022
39	CO NDIR	Thermo	48C	0415406564	July 23, 2022
40	CO NDIR	Thermo	48C	0508011071	July 23 and October 19, 2022
41	CO NDIR	Horiba	48C	0415406563	July 23 and August 10, 2022
42	NOx Chemiluminescence	Horiba	APNA-370	KPACR8NA	August 10, 2022
43	NOx Chemiluminescence	Horiba	APNA-370	M4286P23	August 10, 2022
44	NOx Chemiluminescence	Horiba	APNA-370	XPWS7U3L	August 10, 2022

Calibration Report

Customer Name : ICS Co., Ltd.
Address : 989 Siampiwat Tower, 20th Floor, Unit A1, Rama 1 Road, Pathum Wan, Bangkok 10330
Project Name : โครงการ RTWO
Sampling Date : July - November, 2022

Ambient

Item	Equipment	Manufacturer	Model	Serial Number	Calibration Date
45	SO ₂ UV-Fluorescence	Thermo	43C	64390-343/2	August 10, 2022
46	SO ₂ UV-Fluorescence	Horiba	APSA-370	X7L602W6	August 10 and September 21, 2022
47	SO ₂ UV-Fluorescence	Horiba	APSA-370	G5XCFUN6	August 10, 2022
48	CO NDIR	Horiba	APMA-370	JHG8PWA8	August 10 and November 28, 2022
49	CO NDIR	Horiba	APMA-370	HXA8A4TG	August 10 and September 21, 2022
50	NOx Chemiluminescence	Horiba	APNA-370	R9G207JS	September 21, 2022
51	NOx Chemiluminescence	Horiba	APNA-370	KCDVY226	September 21, 2022
52	NOx Chemiluminescence	Horiba	APNA-360	8517870114	September 21, 2022
53	SO ₂ UV-Fluorescence	Horiba	APSA-370	4X01KWWU	September 21 and November 28, 2022
54	SO ₂ UV-Fluorescence	Thermo	43i	CM14430003	September 21, 2022
55	CO NDIR	Horiba	APMA-370	3VJ73T6X	September 21, 2022
56	CO NDIR	Horiba	APMA-370	Y05LRYAD	September 21, 2022

Calibration Report

Customer Name : ICS Co., Ltd.
Address : 989 Siampiwat Tower, 20th Floor, Unit A1, Rama 1 Road, Pathum Wan, Pathum Wan, Bangkok 10330
Project Name : โครงการ RTWO
Sampling Date : July - November, 2022

Ambient

Item	Equipment	Manufacturer	Model	Serial Number	Calibration Date
57	NOx Chemiluminescence	API	200A	1984	October 19, 2022
58	NOx Chemiluminescence	Horiba	APNA-360	EYC70000	October 19, 2022
59	NOx Chemiluminescence	API	200A	1051	October 19, 2022
60	SO ₂ UV-Fluorescence	Thermo	43i	CM14430005	October 19, 2022
61	SO ₂ UV-Fluorescence	Thermo	43C	0335804029	October 19, 2022
62	SO ₂ UV-Fluorescence	Thermo	43C	0335804022	October 19, 2022
63	CO NDIR	Horiba	APMA-360	0577274012	October 19, 2022
64	CO NDIR	Horiba	APMA-360	0576876075	October 19, 2022
65	NOx Chemiluminescence	API	200A	056	November 28, 2022
66	NOx Chemiluminescence	Horiba	APNA-370	A4LUUFHB	November 28, 2022
67	NOx Chemiluminescence	Horiba	APNA-370	VLR55LT4	November 28, 2022
68	SO ₂ UV-Fluorescence	Thermo	43i	CM14430004	November 28, 2022
69	SO ₂ UV-Fluorescence	Thermo	43i	CM14430002	November 28, 2022

Calibration Report

Customer Name : ICS Co., Ltd.
Address : 989 Siampiwat Tower, 20th Floor, Unit A1, Rama 1 Road, Pathum Wan, Bangkok 10330
Project Name : โครงการ RTWO
Sampling Date : July - November, 2022

Ambient

Item	Equipment	Manufacturer	Model	Serial Number	Calibration Date
70	CO NDIR	Horiba	APMA-370	SFB4TS99	November 28, 2022
71	CO NDIR	Horiba	APMA-370	RBBRWOL3	November 28, 2022
72	Gas Cal Analyzer	Airgas	E04NI99E15A0292	EB0123013	October 22, 2019
73	Sound Level Meter	RION	NL-42	00157458	July 23, 2022
74	Sound Level Meter	RION	NL-21	00410181	July 23, 2022
75	Sound Level Meter	BSWA TECH	BSWA 309	570119	July 23, August 10, September 21, October 19, November 28, 2022
76	Sound Level Meter	RION	NL-21	00609448	August 10, 2022
77	Sound Level Meter	RION	NL-42	00157509	August 10, 2022
78	Sound Level Meter	RION	NL-42	00157457	September 21, 2022
79	Sound Level Meter	RION	NL-42	01147309	September 21, 2022
80	Sound Level Meter	RION	NL-21	00610195	October 19, 2022
81	Sound Level Meter	RION	NL-42	01147317	October 19, 2022

Calibration Report

Customer Name : ICS Co., Ltd.
Address : 989 Siampiwat Tower, 20th Floor, Unit A1, Rama 1 Road, Pathum Wan, Bangkok 10330
Project Name : โครงการ RTWO
Sampling Date : July - November, 2022

Ambient

Item	Equipment	Manufacturer	Model	Serial Number	Calibration Date
82	Sound Level Meter	Scarlet Tech	ST-21D	820448	November 28, 2022
83	Sound Level Meter	Scarlet Tech	ST-21D	820469	November 28, 2022
84	Acoustic Calibrator	BSWA TECH	CA114	470160	September 6, 2022
85	Acoustic Calibrator	Larson Davis	CAL200	2903	March 30, 2022
86	Acoustic Calibrator	BSWA TECH	CA114	590043	March 29, 2022
87	Acoustic Calibrator	Larson Davis	CAL200	7004	January 6, 2022
88	Vibration Meter	Instantel	Micromate	UM10776	July 12, 2021
89	Vibration Meter	Instantel	Micromate	UM9252	November 12, 2021
90	Vibration Meter	Instantel	Micromate	UM7360	February 23, 2021
91	Vibration Meter	Instantel	Micromate	UM13389	November 20, 2020
92	Vibration Meter	Instantel	Micromate	UM10777	August 22, 2020
93	Vibration Meter	Instantel	Micromate	UM9258	August 11, 2021
94	Vibration Meter	Instantel	Micromate	UM10935	January 7, 2022

Calibration Report



Customer Name : ICS Co., Ltd.
Address : 989 Siampiwat Tower, 20th Floor, Unit A1, Rama 1 Road, Pathum Wan, Bangkok 10330
Project Name : โครงการ RTWO
Sampling Date : July - November, 2022

Water

Item	Equipment	Manufacturer	Model	Serial Number	Calibration Date
1	pH Meter	Eutech	pHTestr30	2561366	January 4, 2022
2	Incubator	Accuplus	Smart i250	2059-0218-0002	December 15, 2021
3	DO Meter	YSI	5000-115V	03C1280 AC	January 21, 2022
4	Incubator	Binder	ED 115	950433	January 6, 2022
5	Incubator	Ehret	BK 4106	22162	January 6, 2022
6	Hot Air Oven	Binder	FED 115 E2	11-22823	January 5, 2022
7	Electronic Balance	Mettler Toledo	MS204TS/00	B334691537	January 19, 2022
8	Hot Air Oven	Memmert	UF 110	B414.0652	January 21, 2022
9	Hot Air Oven	Binder	FED 115 E2	11-22823	January 5, 2022



(Ms. Supawan Suwannapa)
Environmental Scientist

(Ms. Panicha Promchai)
Laboratory Supervisor