

ภาคผนวกที่ 6

เอกสาร Detection Limit ของรายการทดสอบ

การตรวจวิเคราะห์คุณภาพอากาศ (Air Quality Analysis)

ประเภทตัวอย่าง : อากาศในบริเวณการทำงาน - Workplace Air Quality									
Items	Parameter	Sampling/Method	Reference Method / Analytical Technique	Air Volume	Sampling Rate / Period	LOQ / Range	Unit	Decimal point	Remark
	แผนภูมิวิธีการทดสอบ								
1	Illumination	Lux Meter	JIS C 1906 / Lux meter		-	0-5000	lux	-	
2	Sound (Leq, T.min, L.max, L.dn, L.p)	Integrated Sound Level Method	ISO 11202 / Sound Level Meter		-	40 - 140	dB (A)	1	
3	Noise Octave band	Integrated Sound Level Method	AS/NZS 4476 1997 / Sound Level Meter		-	40 - 140	dB (A)	1	1/3 Octave band หรือ 1/1 Octave band
4	Noise dose	Integrated Sound Level Method	BS6402 / Noise Dosimeter		-	0 - 9999	% Dose	2	
5	Carbon Monoxide (CO)	Non-Dispersive Infrared Photometric Method	U.S. EPA 10 (P.1-5) / Carbon Monoxide Analyzer		-	0.1 - 100	ppm	1	
6	Ozone (O ₃)	UV Fluorescence Method	U.S. EPA method / Ozone Analyzer		-	0.1 - 100	ppm	2	
7	Heat Stress	WBGT Method	ACGIH / Grove + DI + Thermometer / calculation	-	-	0 - 100	°C	2	
	ส่วนประกอบของตัวอย่าง								
1	Total Dust (TD)	Filtration, Gravimetric Method	NIOSH 0500 (P.1-3) / PS pump / Gravimetric	7-133 L	2 L/min (1 hr)	0.8	mg / m ³	1	SKC Cat No. 225-8-01
2	Respirable Dust (RD)	Cyclone - Filtration, Gravimetric Method	NIOSH 0600 (P.1-3) / PS pump cyclone / Gravimetric	20-400 L	1.70 L/min (1 hr)	0.6	mg / m ³	1	SKC Cat No. 225-8-01
3	Alkaline Dust (NaOH, KOH, LiOH)	Acid-Base Titrimetric Method	NIOSH 7401 (P.1-4) / PS pump / Titration	70-1000 L	1-4 L/min	0.4	mg / m ³	1	SKC Cat No. 225-17-01
	ส่วนประกอบของตัวอย่าง								
1	Ammonia	Impingement Absorption - Colorimetric Method	Modified NIOSH 6015 (P.1-7) / Spectrophotometer	0.1-96 L	1 L/min (1 hr)	0.01	mg / m ³	2	
2	Nitrogen Dioxide	Impingement Absorption, Spectrophotometer Method	APHA 817 (P.1-3) / Spectrophotometer	7.5 - 10 L	0.5 L/min (15-20 min)	0.01	mg / m ³	2	
3	Sulfur Dioxide	Impingement Absorption, Titrimetric Method	APHA 823 (P.1-3) / Titration	26 L	0.21 L/min (2 hrs)	0.30	mg / m ³	2	
4	P,P'-diphenylmethane diisocyanate (MDI) (MDI)	Impingement Absorption, Spectrophotometer Method	APHA 831 (P.1-3) / Spectrophotometer	20 L	1 L/min (20 min)	0.072	mg / m ³	2	
5	Aluminum (Al)	Filtration, ICP-OES Method	NIOSH 7300 (P.1-8) / PS pump / ICP-OES	5-100 L	2 L/min (1 hr)	0.01	mg / m ³	2	SKC Cat No. 225-5
6	Antimony (Sb)	Filtration, ICP-OES Method	NIOSH 7300 (P.1-8) / PS pump / ICP-OES	50-2000 L	2 L/min (1 hr)	0.05	mg / m ³	2	SKC Cat No. 225-5
7	Arsenic & Compound (as As)	Filtration, ICP-OES Method	NIOSH 7300 (P.1-8) / PS pump / ICP-OES	5-2000 L	2 L/min (1 hr)	0.05	mg / m ³	2	SKC Cat No. 225-5

การตรวจวิเคราะห์คุณภาพอากาศ (Air Quality Analysis)

ประเภทตัวอย่าง : อากาศในบริเวณการทำงาน - Workplace Air Quality									
Items	Parameter	Sampling/Method	Reference Method / Analytical Technique	Air Volume	Sampling Rate / Period	LOQ / Range	Unit	Decimal point	Remark
8	Barium (Ba)	Filtration, ICP-OES Method	NIOSH 7300 (P.1-8) / PS pump / ICP-OES	50-2000 L	2 L/min (1 hr)	0.01	mg / m ³	2	SKC Cat No. 225-5
9	Cadmium & Compounds (as Cd)	Filtration, ICP-OES Method	NIOSH 7300 (P.1-8) / PS pump / ICP-OES	25-1500 L	2 L/min (1 hr)	0.002	mg / m ³	2	SKC Cat No. 225-5
10	Calcium & Compounds (as Ca)	Filtration, ICP-OES Method	NIOSH 7300 (P.1-8) / PS pump / ICP-OES	20-400 L	2 L/min (1 hr)	0.50	mg / m ³	2	SKC Cat No. 225-5
12	Chromium & Compounds (as Cr)	Filtration, ICP-OES Method	NIOSH 7300 (P.1-8) / PS pump / ICP-OES	5-1000 L	2 L/min (1 hr)	0.01	mg / m ³	2	SKC Cat No. 225-5
13	Copper (Cu) (Dust & Fume)	Filtration, ICP-OES Method	NIOSH 7300 (P.1-8) / PS pump / ICP-OES	50-1500 L	2 L/min (1 hr)	0.01	mg / m ³	2	SKC Cat No. 225-5
14	Iron & Compounds (as Fe)	Filtration, ICP-OES Method	NIOSH 7300 (P.1-8) / PS pump / ICP-OES	5-1000 L	2 L/min (1 hr)	0.01	mg / m ³	2	SKC Cat No. 225-5
15	Lead (Pb)	Filtration, ICP-OES Method	NIOSH 7300 (P.1-8) / PS pump / ICP-OES	50-2000 L	2 L/min (1 hr)	0.01	mg / m ³	2	SKC Cat No. 225-5
16	Magnesium (Mg)	Filtration, ICP-OES Method	NIOSH 7300 (P.1-8) / PS pump / ICP-OES	6-67 L	2 L/min (1 hr)	0.50	mg / m ³	2	SKC Cat No. 225-5
17	Manganese (Mn)	Filtration, ICP-OES Method	NIOSH 7300 (P.1-8) / PS pump / ICP-OES	5-200 L	2 L/min (1 hr)	0.01	mg / m ³	2	SKC Cat No. 225-5
18	Mercury (Hg)	Filtration - AAS Method	NIOSH 6009 (P.1-5) / PS pump / AAS	2 - 100 L	0.2 L/min (1 hr)	0.0010	mg / m ³	2	SKC Cat No. 225-5
19	Nickel & Compounds (as Ni)	Filtration, ICP-OES Method	NIOSH 7300 (P.1-8) / PS pump / ICP-OES	5-1000 L	2 L/min (1 hr)	0.01	mg / m ³	2	SKC Cat No. 225-5
20	Selenium (Se)	Filtration, ICP-OES Method	NIOSH 7300 (P.1-8) / PS pump / ICP-OES	13-2000 L	2 L/min (1 hr)	0.05	mg / m ³	2	SKC Cat No. 225-5
21	Silver (Ag)	Filtration, ICP-OES Method	NIOSH 7300 (P.1-8) / PS pump / ICP-OES	250-2000 L	2 L/min (2-17 hr)	0.01	mg / m ³	2	SKC Cat No. 225-5
22	Sodium (Na)	Filtration, ICP-OES Method	NIOSH 7300 (P.1-8) / PS pump / ICP-OES	13-2000 L	2 L/min (1 hr)	0.50	mg / m ³	2	SKC Cat No. 225-5
23	Tin (Sn)	Filtration, ICP-OES Method	NIOSH 7300 (P.1-8) / PS pump / ICP-OES	5-1000 L	2 L/min (1 hr)	0.50	mg / m ³	2	SKC Cat No. 225-5

การตรวจวิเคราะห์คุณภาพอากาศ (Air Quality Analysis)

ประเภทตัวอย่าง : อากาศในบริเวณการทำงาน - Workplace Air Quality									
Items	Parameter	Sampling/Method	Reference Method / Analytical Technique	Air Volume	Sampling Rate / Period	LOQ / Range	Unit	Decimal point	Remark
24	Titanium (Ti)	Filtration, ICP-OES Method	NIOSH 7300(P, I-8) / PS pump / ICP-OES	5-1000 L	2 L/min (1 hr)	0.01	mg / m ³	2	SKC Cat No. 225-5
25	Vanadium (V)	Filtration, ICP-OES Method	NIOSH 7300(P, I-8) / PS pump / ICP-OES	5-2000 L	2 L/min (1 hr)	0.01	mg / m ³	2	SKC Cat No. 225-5
26	Zinc & Compounds (Zn)	Filtration, ICP-OES Method	NIOSH 7300(P, I-8) / PS pump / ICP-OES	5-2000 L	2 L/min (1 hr)	0.01	mg / m ³	2	SKC Cat No. 225-5
27	Acetone	Sorbent Adsorption, GC Method	NIOSH 1300 (P, I-5) / PS pump / GC-FID	0.5-3 L	0.10 L/min (30 min)	13.17 5.54	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
28	Benzene	Sorbent Adsorption, GC Method	NIOSH 1501 (P, I-7) / PS pump / GC-FID	5-30 L	0.10 L/min (1 hr)	2.93 0.92	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
29	Cyclohexanone	Sorbent Adsorption, GC Method	NIOSH 1300(P, I-5) / PS pump / GC-FID	1-10 L	0.10 L/min (1 hr)	3.96 0.99	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
30	Ethanol (Ethyl alcohol)	Sorbent Adsorption, GC Method	NIOSH 1400(P, I-4) / PS pump / GC-FID	12 L	0.10 L/min (1 hr)	3.29 1.75	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
31	Ethylacetate	Sorbent Adsorption, GC Method	NIOSH 1457 (P, I-4) / PS pump / GC-FID	0.1-10 L	0.10 L/min (1 hr)	7.21 2.00	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
32	Ethylbenzene	Sorbent Adsorption, GC Method	NIOSH 1501 (P, I-7) / PS pump / GC-FID	1-24 L	0.10 L/min (1 hr)	3.63 0.83	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
33	Hexane	Sorbent Adsorption, GC Method	NIOSH 1500(P, I-8) / PS pump / GC-FID	4 L	0.10 L/min (1 hr)	7.05 2.00	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
34	Isopropanol (Isopropyl alcohol) : IPA	Sorbent Adsorption, GC Method	NIOSH 1400(P, I-4) / PS pump / GC-FID	12 L	0.10 L/min (1 hr)	3.28 1.33	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
35	Methanol (Methyl alcohol)	Sorbent Adsorption, GC Method	OSHA 91(P, I-10) / PS pump / GC-FID	1-5 L	0.10 L/min (30 min)	3.96 3.02	mg / m ³ ppm	2	SKC Cat. No. ST 226-82
36	Methyl Ethyl Ketone (MEK)	Sorbent Adsorption, GC Method	OSHA 1004(P, I-27) / PS pump / GC-FID	0.25-12L	0.10 L/min (1 hr)	3.35 1.14	mg / m ³ ppm	2	SKC Cat. No. ST 226-81A
37	Methyl Isobutyl Ketone (MIBK)	Sorbent Adsorption, GC Method	OSHA 1004(P, I-27) / PS pump / GC-FID	0.25-12L	0.10 L/min (1 hr)	3.34 0.81	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
38	Styrene	Sorbent Adsorption, GC Method	NIOSH 1501 (P, I-7) / PS pump / GC-FID	1-24 L	0.10 L/min (1 hr)	3.78 0.89	mg / m ³ ppm	2	SKC Cat. No. ST 226-01

การตรวจวิเคราะห์คุณภาพอากาศ (Air Quality Analysis)

ประเภทตัวอย่าง : อากาศในบริเวณการทำงาน - Workplace Air Quality									
Items	Parameter	Sampling/Method	Reference Method / Analytical Technique	Air Volume	Sampling Rate / Period	LOQ / Range	Unit	Decimal point	Remark
39	Toluene	Sorbent Adsorption, GC Method	NIOSH 1501 (P, I-7) / PS pump / GC-FID	1-8 L	0.10 L/min (1 hr)	3.63 0.96	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
40	Xylene	Sorbent Adsorption, GC Method	NIOSH 1501 (P, I-7) / PS pump / GC-FID	2-23 L	0.10 L/min (1 hr)	3.58 0.83	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
41	Cumene	Sorbent Adsorption, GC Method	NIOSH 1501 (P, I-7) / PS pump / GC-FID	2-23 L	0.10 L/min (1 hr)	3.60 0.73	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
42	Methylcyclohexane	Sorbent Adsorption, GC Method	NIOSH 1500 (P, I-8) / PS pump / GC-FID	2-23 L	0.10 L/min (1 hr)	7.23 1.80	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
43	Diethyl Ether or Ethyl Ether	Sorbent Adsorption, GC Method	NIOSH 1610 (P, I-4) / PS pump / GC-FID	0.25-3 L	0.01-0.20 L/min (1 hr)	11.88 3.92	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
44	Methyl tert-Butyl Ether (MTBE)	Sorbent Adsorption, GC Method	NIOSH 1615 (P, I-4) / PS pump / GC-FID	2-96 L	0.01-0.20 L/min (1 hr)	3.08 0.86	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
45	Dichloromethane or Methylene chloride	Sorbent Adsorption, GC Method	NIOSH 1005 (P, I-4) / PS pump / GC-FID	0.5-2.5 L	0.01-0.20 L/min (1 hr)	22.1 6.36	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
46	1-Butanol /n-butyl alcohol	Sorbent Adsorption, GC Method	NIOSH 1401 (P, I-4) / PS pump / GC-FID	2-10 L	0.01-0.20 L/min (1 hr)	4.86 1.60	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
47	2-Butanol /sec-butyl alcohol	Sorbent Adsorption, GC Method	NIOSH 1401 (P, I-4) / PS pump / GC-FID	2-10 L	0.01-0.20 L/min (1 hr)	4.86 1.60	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
48	Isobutyl alcohol (IBA)	Sorbent Adsorption, GC Method	NIOSH 1401 (P, I-4) / PS pump / GC-FID	2-10 L	0.01-0.20 L/min (1 hr)	4.86 1.60	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
49	Beryllium (Be)	Filtration, ICP-OES Method	NIOSH 7300(P, I-8) / PS pump / ICP-OES	1250-2000 L	2 L/min (1 hr)	0.01	mg / m ³	2	SKC Cat No. 225-5
50	Cobalt (Co)	Filtration, ICP-OES Method	NIOSH 7300(P, I-8) / PS pump / ICP-OES	25-2000 L	2 L/min (1 hr)	0.01	mg / m ³	2	SKC Cat No. 225-5
51	Molybdenum (Mo)	Filtration, ICP-OES Method	NIOSH 7300(P, I-8) / PS pump / ICP-OES	5-67 L	2 L/min (1 hr)	0.01	mg / m ³	2	SKC Cat No. 225-5
52	Thallium (Tl)	Filtration, ICP-OES Method	NIOSH 7300(P, I-8) / PS pump / ICP-OES	25-2000 L	2 L/min (1 hr)	0.01	mg / m ³	2	SKC Cat No. 225-5
53	Silicon (Si)	Filtration, ICP-OES Method	NIOSH 7300(P, I-8) / PS pump / ICP-OES	5-1000 L	2 L/min (1 hr)	0.01	mg / m ³	2	SKC Cat No. 225-5
54	Potassium (K)	Filtration, ICP-OES Method	NIOSH 7300(P, I-8) / PS pump / ICP-OES	5-1000 L	2 L/min	0.01	mg / m ³	2	SKC Cat No. 225-5

การตรวจวิเคราะห์คุณภาพอากาศ (Air Quality Analysis)

(ประเภทตัวอย่าง : อากาศในบริเวณการทำงาน - Workplace Air Quality)									
Items	Parameter	Sampling/Method	Reference Method / Analytical Technique	Air Volume	Sampling Rate / Period	LOQ / Range	Unit	Decimal point	Remark
					(1 hr)				
55	Ketones	Sorbent Adsorption, GC Method	NIOSH 2555 (P.1-5) / PS pump / GC-FID	0.5-3.0 L	0.01-0.20 L/min (1 hr)	13.17 5.54	mg / m ³ ppm	2	SKC Cat. No. 226-01
56	n-Heptane	Sorbent Adsorption, GC Method	NIOSH 1500 (P.1-8) / PS pump / GC-FID	-	0.01-0.20 L/min (1 hr)	6.97 1.70	mg / m ³ ppm	2	SKC Cat. No. 226-01
57	n-Butyl acetate	Sorbent Adsorption, GC Method	NIOSH 1450(P.1-6) / PS pump / GC-FID	1-10 L	0.01-0.20 L/min (1 hr)	8.55 1.80	mg / m ³ ppm	2	SKC Cat. No. 226-01
58	n-Pentane	Sorbent Adsorption, GC Method	NIOSH 1500(P.1-8) / PS pump / GC-FID	-	0.01-0.20 L/min (1 hr)	2.63 0.89	mg / m ³ ppm	2	SKC Cat. No. 226-01
59	Chloroform	Sorbent Adsorption, GC Method	NIOSH 1003 (P.1-7) / PS pump / GC-FID	1-50 L	0.01-0.20 L/min (1 hr)	4.93 1.01	mg / m ³ ppm	2	SKC Cat. No. 226-01
60	Chlorobenzene	Sorbent Adsorption, GC Method	NIOSH 1003 (P.1-7) / PS pump / GC-FID	1.5-40 L	0.01-0.20 L/min (1 hr)	4.63 1.00	mg / m ³ ppm	2	SKC Cat. No. 226-01
61	Formaldehyde	Sorbent Adsorption, GC Method	NIOSH 2541 (P.1-5) / PS pump / GC-FID	1-36 L	0.01-0.10 L/min (1 hr)	0.43 0.35	mg / m ³ ppm	2	SKC Cat. No. 226-118
62	Hydrochloric acid	Sorbent Adsorption, IC Method	OSHA ID-174SG / PS pump / IC	100 L	500 L/min (15 min)	0.015 0.010	mg / m ³ ppm	3	SKC Cat. No. 226-10-03
63	Hydrogen Bromide	Sorbent Adsorption, IC Method	OSHA ID165SG / PS pump / IC	100 L	200 L/min (60min)	0.033 0.010	mg / m ³ ppm	3	SKC Cat. No. 226-10-03
64	Sulfuric Acid	Sorbent Adsorption, IC Method	OSHA ID165SG / PS pump / IC	100 L	200 L/min (60min)	0.033 0.010	mg / m ³ ppm	3	SKC Cat. No. 226-10-03
64	Phosphoric Acid	Sorbent Adsorption, IC Method	OSHA ID165SG / PS pump / IC	100 L	200 L/min (60min)	0.20 0.010	mg / m ³ ppm	3	SKC Cat. No. 226-10-03
65	Ammonia (NH ₃)	Sorbent Adsorption, IC Method	OSHA ID165SG / PS pump / IC	24 L	200 L/min (120min)	0.200 0.280	mg / m ³ ppm	3	SKC Cat. No. 226-10-03
67	Nitric	Sorbent Adsorption, IC Method	OSHA ID165SG / PS pump / IC	100 L	200 L/min (60min)	0.026 0.010	mg / m ³ ppm	3	SKC Cat. No. 226-10-03
68	Chlorine	Sorbent Adsorption, IC Method	OSHA ID165SG / PS pump / IC	60 L	200 L/min (60min)	0.029 0.010	mg / m ³ ppm	3	SKC Cat. No. 226-10-03

การตรวจวิเคราะห์คุณภาพอากาศ (Air Quality Analysis)

(ประเภทตัวอย่าง : อากาศในบริเวณการทำงาน - Workplace Air Quality)									
Items	Parameter	Sampling/Method	Reference Method / Analytical Technique	Air Volume	Sampling Rate / Period	LOQ / Range	Unit	Decimal point	Remark

เอกสารอ้างอิง

1. Method of Air Sampling and Analysis, APHA Intersociety Committee, 1997
2. NIOSH Manual of Analytical Method, 4th Edition, 1994
3. Code of Federal Regulation, U.S. EPA., 40 CFR Part 50, Part 60, 2000
4. OSHA Analytical Methods Manual, 2nd Edition, U.S. Department of Labor, 1992
5. International Standard Organization, ISO 11204:1995
6. Compendium of Methods for Determination of Inorganic Compound in Ambient Air, U.S. EPA., 1999
7. Annual Book of ASTM Standard, Section 11, 2001

การตรวจวิเคราะห์คุณภาพอากาศ (Air Quality Analysis)

ประเภทตัวอย่าง : อากาศในบรรยากาศโดยทั่วไป - Ambient Air Quality									
Items	Parameter	Method	Reference Method / Analytical Technique	Air Volume	Sampling Rate / Period	LOQ / Range	Unit	Decimal point	Remark
แบบปฏิบัติการภาคสนาม									
1	Sulfur Dioxide (SO ₂)	UV Fluorescence Method	U.S. EPA EQSA-0292-084 / Sulfur Dioxide Analyzer	288 L	24 hrs (1 hr avg.)	0.001 - 10	ppm	3	
2	Nitrogen Dioxide (NO ₂)	Chemiluminescence Method	U.S. EPA RFCA-0995-108 / Nitrogen Dioxide Analyzer	288 L	24 hrs (1 hr avg.)	0.001 - 10	ppm	3	
3	Carbon Monoxide (CO)	Non-Dispersive Infrared Photometric Method	U.S. EPA 40 CFR Part 50 Appendix C / Carbon Monoxide Analyzer	288 L	24 hrs (8 hr avg.)	0.1 - 100	ppm	1	
4	Ozone (O ₃)	UV Fluorescence Method	U.S. EPA 40 CFR Part 50 Appendix D / Ozone Analyzer	288 L	24 hrs (1 hr avg.)	0.001 - 10	ppm	3	
5	Sound (Leq, L min, L max, L dn, Lp)	Integrated Sound Level Method	ISO 1996-1 / Sound Level meter	288 L	24 hrs (1 hr avg.)	40 - 140	dB (A)	1	
6	Wind Speed & Wind Direction	Wind Speed & Wind Direction Sensor	ASTM D 4480-93 / WS/WD Equipment	288 L	24 hrs				Wind speed & Wind direction Diagram
ส่วนงานทดสอบพื้นฐาน									
ส่วนงานเครื่องมือทดสอบ									
1	Ammonia (NH ₃)	Impingement Absorption, Colorimetric Method	APHA 401 / Spectrophotometer	288 L	0.2 L/min (24 hrs)	0.01	mg / m ³	2	
2	Sulfur Dioxide (SO ₂)	Pararosaniline Method	U.S. EPA 40 CFR Part 50 Appendix A / Spectrophotometer	288 L	0.2 L/min (24 hrs)	0.01	mg / m ³	2	
3	Aluminium (Al)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.01	mg / m ³	2	Advantage MFS Cat. No. GA55 8 x 10 "
4	Antimony (Sb)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.01	mg / m ³	2	Advantage MFS Cat. No. GA55 8 x 10 "
6	Arsenic (As)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.05	mg / m ³	2	Advantage MFS Cat. No. GA55 8 x 10 "
7	Barium (Ba)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.01	mg / m ³	2	Advantage MFS Cat. No. GA55 8 x 10 "
8	Cadmium (Cd)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.01	mg / m ³	2	Advantage MFS Cat. No. GA55 8 x 10 "

Items	Parameter	Method	Reference Method / Analytical Technique	Air Volume	Sampling Rate / Period	LOQ / Range	Unit	Decimal point	Remark
9	Calcium (Ca)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.50	mg / m ³	2	Advantage MFS Cat. No. GA55 8 x 10 "
10	Chromium (Cr)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.01	mg / m ⁶	2	Advantage MFS Cat. No. GA55 8 x 10 "
11	Copper (Cu)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.01	mg / m ⁷	2	Advantage MFS Cat. No. GA55 8 x 10 "
12	Iron (Fe)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.01	mg / m ⁸	2	Advantage MFS Cat. No. GA55 8 x 10 "
13	Lead (Pb)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.01	mg / m ⁹	2	Advantage MFS Cat. No. GA55 8 x 10 "
14	Magnesium (Mg)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.05	mg / m ¹⁰	2	Advantage MFS Cat. No. GA55 8 x 10 "
15	Manganese (Mn)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.01	mg / m ¹¹	2	Advantage MFS Cat. No. GA55 8 x 10 "
16	Mercury (Hg)	Filtration, AAS Method	U.S. EPA Method IO-3.4 / High Volume - AAS	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.0010	mg / m ¹²	2	Advantage MFS Cat. No. GA55 8 x 10 "
17	Nickel (Ni)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.01	mg / m ¹³	2	Advantage MFS Cat. No. GA55 8 x 10 "
18	Potassium (K)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.25	mg / m ¹⁴	2	Advantage MFS Cat. No. GA55 8 x 10 "
19	Sodium (Na)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.50	mg / m ¹⁵	2	Advantage MFS Cat. No. GA55 8 x 10 "
20	Tin (Sn)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.05	mg / m ¹⁶	2	Advantage MFS Cat. No. GA55 8 x 10 "
21	Titanium (Ti)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.01	mg / m ¹⁷	2	Advantage MFS Cat. No. GA55 8 x 10 "
22	Vanadium (V)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.01	mg / m ¹⁸	2	Advantage MFS Cat. No. GA55 8 x 10 "
23	Zinc (Zn)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.01	mg / m ¹⁹	2	Advantage MFS Cat. No. GA55 8 x 10 "
24	Selenium (Se)	Filtration, ICP-OES Method	U.S. EPA Method IO-3.4 / High Volume - ICP-OES	1,590 - 2,447 m ³	39-60 ft ³ /min (24 hrs)	0.05	mg / m ²⁰	2	Advantage MFS Cat. No. GA55 8 x 10 "

Items	Parameter	Method	Reference Method / Analytical Technique	Air Volume	Sampling Rate / Period	LOQ / Range	Unit	Decimal point	Remark
25	Acetone	Sorbent Adsorption, GC Method	ASTM D 3687-95 / GC-FID	144 L	0.10 L/min (24 hrs)	0.14 0.06	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
26	Benzene	Sorbent Adsorption, GC Method	ASTM D 3687-95 / GC-FID	144 L	0.10 L/min (24 hrs)	0.12 0.04	mg / m ³ ppm	2	SKC Cat. No. ST 226-02
27	Cyclohexanone	Sorbent Adsorption, GC Method	ASTM D 3687-95 / GC-FID	144 L	0.10 L/min (24 hrs)	0.16 0.04	mg / m ³ ppm	2	SKC Cat. No. ST 226-04
28	Fahanol (Ethyl alcohol)	Sorbent Adsorption, GC Method	ASTM D 3687-95 / GC-FID	288 L	0.10 L/min (24 hrs)	0.14 0.07	mg / m ³ ppm	2	SKC Cat. No. ST 226-05
29	Ethylacetate	Sorbent Adsorption, GC Method	ASTM D 3687-95 / GC-FID	144 L	0.10 L/min (24 hrs)	0.61 0.20	mg / m ³ ppm	2	SKC Cat. No. ST 226-06
30	Ethylbenzene	Sorbent Adsorption, GC Method	ASTM D 3687-95 / GC-FID	144 L	0.10 L/min (24 hrs)	0.15 0.03	mg / m ³ ppm	2	SKC Cat. No. ST 226-07
31	Hexane	Sorbent Adsorption, GC Method	ASTM D 3687-95 / GC-FID	144 L	0.10 L/min (24 hrs)	0.32 0.09	mg / m ³ ppm	2	SKC Cat. No. ST 226-08
32	Isopropanol (Isopropyl alcohol) : IPA	Sorbent Adsorption, GC Method	ASTM D 3687-95 / GC-FID	288 L	0.10 L/min (24 hrs)	0.14 0.06	mg / m ³ ppm	2	SKC Cat. No. ST 226-09
33	Methanol (Methyl alcohol)	Sorbent Adsorption, GC Method	ASTM D 3687-95 / GC-FID	144 L	0.10 L/min (24 hrs)	0.07 0.05	mg / m ³ ppm	2	SKC Cat. No. ST 226-10
34	Methyl Ethyl Ketone (MEK)	Sorbent Adsorption, GC Method	ASTM D 3687-95 / GC-FID	144 L	0.10 L/min (24 hrs)	0.14 0.05	mg / m ³ ppm	2	SKC Cat. No. ST 226-11
35	Styrene	Sorbent Adsorption, GC Method	ASTM D 3687-95 / GC-FID	144 L	0.10 L/min (24 hrs)	0.16 0.04	mg / m ³ ppm	2	SKC Cat. No. ST 226-12
36	Toluene	Sorbent Adsorption, GC Method	ASTM D 3687-95 / GC-FID	144 L	0.10 L/min (24 hrs)	0.15 0.04	mg / m ³ ppm	2	SKC Cat. No. ST 226-13
37	Xylene	Sorbent Adsorption, GC Method	ASTM D 3687-95 / GC-FID	144 L	0.10 L/min (24 hrs)	0.15 0.03	mg / m ³ ppm	2	SKC Cat. No. ST 226-14
38	Methylcyclohexane	Sorbent Adsorption, GC Method	NIOSH 1500 (P,I-8) / PS pump / GC-FID	2-23 L	0.10 L/min (1 hr)	0.32 0.08	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
39	Diethyl Ether or Ethyl Ether	Sorbent Adsorption, GC Method	NIOSH 1610 (P,I-4) / PS pump / GC-FID	0.25-3 L	0.01-0.20 L/min (1 hr)	0.12 0.04	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
40	Methyl tert-Butyl Ether (MTBE)	Sorbent Adsorption, GC Method	NIOSH 1615 (P,I-4) / PS pump / GC-FID	2-96 L	0.01-0.20 L/min (1 hr)	0.13 0.04	mg / m ³ ppm	2	SKC Cat. No. ST 226-01

Items	Parameter	Method	Reference Method / Analytical Technique	Air Volume	Sampling Rate / Period	LOQ / Range	Unit	Decimal point	Remark
41	Dichloromethane	Sorbent Adsorption, GC Method	NIOSH 1005 (P,I-4) / PS pump / GC-FID	0.5-2.5 L	0.01-0.20 L/min (1 hr)	0.23 0.07	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
42	1-Butanol /n-butyl alcohol	Sorbent Adsorption, GC Method	NIOSH 1401 (P,I-4) / PS pump / GC-FID	2-10 L	0.01-0.20 L/min (1 hr)	0.17 0.06	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
43	2-Butanol /sec-butyl alcohol	Sorbent Adsorption, GC Method	NIOSH 1401 (P,I-4) / PS pump / GC-FID	2-10 L	0.01-0.20 L/min (1 hr)	0.17 0.06	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
44	Isobutyl alcohol (IBA)	Sorbent Adsorption, GC Method	NIOSH 1401 (P,I-4) / PS pump / GC-FID	2-10 L	0.01-0.20 L/min (1 hr)	0.17 0.06	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
45	Methyl Isobutyl Ketone (MIBK)	Sorbent Adsorption, GC Method	OSHA 1004(P,I-27) / PS pump / GC-FID	0.25-12L	0.10 L/min (1 hr)	0.14 0.03	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
46	Ketones	Sorbent Adsorption, GC Method	NIOSH 2555 (P,I-5) / PS pump / GC-FID	0.5-10L	0.01-0.20 L/min (1 hr)	0.14 0.06	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
47	n-Butyl acetate	Sorbent Adsorption, GC Method	NIOSH 1450 (P,I-6) / PS pump / GC-FID	1-10L	0.01-0.20 L/min (1 hr)	2.31 0.76	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
48	n-Pentane	Sorbent Adsorption, GC Method	NIOSH 1500 (P,I-8) / PS pump / GC-FID	1-10L	0.01-0.20 L/min (1 hr)	2.31 0.76	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
49	Chloroform	Sorbent Adsorption, GC Method	NIOSH 1003 (P,I-7) / PS pump / GC-FID	1-50L	0.01-0.20 L/min (1 hr)	2.31 0.76	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
50	Chlorobenzene	Sorbent Adsorption, GC Method	NIOSH 1003 (P,I-7) / PS pump / GC-FID	1.5-40L	0.01-0.20 L/min (1 hr)	2.31 0.76	mg / m ³ ppm	2	SKC Cat. No. ST 226-01
51	Formaldehyde	Sorbent Adsorption, GC Method	NIOSH 2541 (P,I-5) / PS pump / GC-FID	1-36L	0.01-0.10 L/min (1 hr)	0.01 0.01	mg / m ³ ppm	2	SKC Cat. No. 226-118
52	Hydrochloric acid	Sorbent Adsorption, IC Method	OSHA ID-1745G / PS pump / IC	1-7.5 L	0.20 L/min (24 hr)	0.015 0.010	mg / m ³ ppm	3	SKC Cat. No. 226-10-03
53	Hydrogen Bromide	Sorbent Adsorption, IC Method	OSHA ID1655G / PS pump / IC	1-96 L	0.20 L/min (24 hr)	0.033 0.010	mg / m ³ ppm	3	SKC Cat. No. 226-10-03
54	Sulfuric Acid	Sorbent Adsorption, IC Method	OSHA ID1655G / PS pump / IC	1-96 L	0.20 L/min (24 hr)	0.040 0.010	mg / m ³ ppm	3	Filter (PTFE)
55	Phosphoric Acid	Sorbent Adsorption, IC Method	OSHA ID1655G / PS pump / IC	1-96 L	0.20 L/min (24 hr)	0.04 0.010	mg / m ³ ppm	3	Filter (PTFE)
56	Nitric	Sorbent Adsorption, IC Method	OSHA ID1655G / PS pump / IC	1-96 L	0.20 L/min (24 hr)	0.026 0.010	mg / m ³ ppm	3	SKC Cat. No. 226-10-03

Items	Parameter	Method	Reference Method / Analytical Technique	Air Volume	Sampling Rate / Period	LOQ / Range	Unit	Decimal point	Remark
57	Chlorine	Sorbent Adsorption, IC Method	OSHA 1065SG / PS pump / IC	14 L	0.20 L/min (24 hr)	0.026 0.010	mg / m ³ ppm	3	SKC Cat. No. 226-10-03

- เอกสารอ้างอิง
1. Method of Air Sampling and Analysis, APHA Intersociety Committee, 2017
 2. NIOSH Manual of Analytical Methods (NMAM)
 3. Code of Federal Regulation, U.S. EPA., 40 CFR Part 50, Part 60, 2000
 4. Occupational Health and Safety Management System(OSHA) Analytical Methods Manual
 5. International Standard Organization, ISO 11204:1995
 6. Compendium of Methods for Determination of Inorganic Compound in Ambient Air, U.S. EPA., 1999
 7. Annual Book of ASTM Standard, Section 11, 2001

การตรวจวิเคราะห์คุณภาพอากาศ (Air Quality Analysis)

ประเภทตัวอย่าง : อากาศในปล่องระบาย - Stack Air Quality

ตารางที่ 1 สรุปขั้นตอนการเก็บตัวอย่างและคำแนะนำในการทดสอบตัวอย่างของโรงงาไฟฟ้าการ **ควบที่ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม**
(ประเภทตัวอย่าง : อากาศในปล่องระบาย - Stack Air Quality)

Items	Parameter	Method	Reference Method / Analytical Technique	Air Volume	Sampling Rate / Period	LOQ / Range	Unit	Decimal point	Remark
เกณฑ์ปฏิบัติการภาคสนาม									
1	Smoke density (Opacity)	Ringelmann's method	U.S. EPA Method 9 / Ringelmann's Chart	-	-	-	%	2	
2	Oxide of Nitrogen	Chemiluminescence Method	U.S. EPA Method 7E / Nitrogen dioxide Analyzer	±	±	0.1 - 100	ppm	1	ใช้ Dilution Probe ร่วมในการตรวจวัด
3	Sulfur Dioxide	UV Fluorescence Method	U.S. EPA Method 6C / Sulfur dioxide Analyzer	±	±	0.4 - 100	ppm	1	ใช้ Dilution Probe ร่วมในการตรวจวัด
4	Carbon Monoxide	Bag,Non-Dispersive Infrared Method	U.S. EPA method 10 / Carbon monoxide analyzer	-	-	0.1 - 100	ppm	1	ใช้ Dilution Probe ร่วมในการตรวจวัด
ส่วนงานทดสอบพื้นฐาน									
6	Hydrogen Sulfide (H ₂ S)	Absorption, Iodometric Method	U.S. EPA Method 11 / Iodometric			0.1	mg / m ³	1	
7	Sulfur Dioxide (SO ₂)	Absorption Barium Thorin Titrimetric Method	U.S. EPA Method 6 / Titration	0.03 m ³	Isokinetic (30 min)	1.3	mg / m ³	1	
8	Sulfuric acid (H ₂ SO ₄)	Isokinetic, Barium Thorin Titrimetric Method	U.S. EPA Method 8 / Titration	0.9 m ³	Isokinetic (30 min)	0.10	mg / m ³	2	
ส่วนงานเครื่องมือทดสอบ									
9	Oxide of Nitrogen (Nitrogen Dioxide ;	Chemical Absorption, Colorimetric Method	U.S. EPA Method 7 / Spectrophotometer	2.0 L	Non-Isokinetic (30 min)	1.0	mg / m ³	1	
10	Xylene	Sorbent Adsorption, Gas Chromatography Method	U.S. EPA Method 18 / GC-FID	0.21 m ³	0.7 L/min (30 min)	2.17 0.50	mg / m ³ ppm	2	SKC Cat. No. 226-09
11	Vanadium (V)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-OES	0.9 m ³	Isokinetic (30 min)	0.05	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
12	Tin (Sn)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-OES	0.9 m ³	Isokinetic (30 min)	1.00	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
13	Selenium (Se)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-OES	0.9 m ³	Isokinetic (30 min)	1.00	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
14	Antimony (Sb)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	1.00	mg / m ³	2	Advantage MFS Cat No. GC5090 MM

Items	Parameter	Method	Reference Method / Analytical Technique	Air Volume	Sampling Rate / Period	LOQ / Range	Unit	Decimal point	Remark
15	Arsenic (As)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	2.00	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
16	Cadmium (Cd)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	0.05	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
17	Chromium (Cr)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	0.01	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
18	Copper (Cu)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	0.05	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
19	Cobalt (Co)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	0.05	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
20	Lead and Inorganic Lead (Pb)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	0.05	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
21	Manganese (Mn)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	0.05	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
22	Nickel (Ni)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	0.05	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
23	Mercury (Hg)	Isokinetic, Sampling,Cold Vapor Technique-AAS Method	U.S. EPA Method 101 / AAS	0.053 m ³	Isokinetic (1.5 L/min)	0.0010	mg / m ³	2	Advantage MFS Cat No. GC5090 MM

การตรวจวิเคราะห์คุณภาพอากาศ (Air Quality Analysis)

ประเภทตัวอย่าง : อากาศในปล่องระบาย - Stack Air Quality

ตารางที่ 2 สรุปค่าการตรวจเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ ได้รับใบอนุญาตเป็นกรมโรงงานอุตสาหกรรม

(ประเภทตัวอย่าง : อากาศในปล่องระบาย - Stack Air Quality)

Items	Parameter	Method	Reference Method / Analytical Technique	Air Volume	Sampling Rate / Period	LOQ / Range	Unit	Decimal point	Remark
แผนปฏิบัติการภาคสนาม									
1	Sampling and Traverse point	U.S. EPA Recommend (Method 1)	U.S. EPA Method 1 / Calculation	-	-	-	-	-	
2	Velocity and Volumetric Flow rate		U.S. EPA Method 2 / Calculation	-	-	-	-	-	
3	Oxygen	Electrochemical Sensor	Modified U.S. EPA 3 / Electrochemical Sensor	-	-	0-20.9	%	1	
4	Moisture Content		U.S. EPA Method 4 / Calculation	-	-	-	-	2	
6	Carbon dioxide (CO ₂)	Electrochemical Sensor	Modified U.S. EPA 3 / Electrochemical Sensor	-	-	0-20.9	%	2	
ส่วนงานเครื่องมือทดสอบ									
7	Aluminium (Al)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	0.05	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
8	Antimony (Sb)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	1.00	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
9	Barium (Ba)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	0.05	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
10	Calcium (Ca)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	0.05	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
11	Iron (Fe)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	0.05	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
12	Magnesium (Mg)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	0.05	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
13	Nickel (Ni)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	0.05	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
14	Silver (Ag)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	0.05	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
15	Sodium (Na)	Isokinetic, Sampling,Digestion,ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	0.05	mg / m ³	2	Advantage MFS Cat No. GC5090 MM

Items	Parameter	Method	Reference Method / Analytical Technique	Air Volume	Sampling Rate / Period	LOQ / Range	Unit	Decimal point	Remark
16	Zinc (Zn)	Isokinetic, Sampling, Digestion, ICP-OES Method	U.S. EPA Method 29 / ICP-AES	0.9 m ³	Isokinetic (30 min)	0.05	mg / m ³	2	Advantage MFS Cat No. GC5090 MM
17	Acetone	Sorbent Adsorption, Gas Chromatography Method	US. EPA Method 18 / GC-FID	0.21 m ³	0.7 L/min (30 min)	1.88 0.79	mg / m ³ ppm	2	SKC Cat. No. 226-09
18	Benzene	Sorbent Adsorption, Gas Chromatography Method	US. EPA Method 18 / GC-FID	0.21 m ³	0.7 L/min (30 min)	0.64 0.20	mg / m ³ ppm	2	SKC Cat. No. 226-09
20	Cyclohexanone	Sorbent Adsorption, Gas Chromatography Method	US. EPA Method 18 / GC-FID	0.21 m ³	0.7 L/min (30 min)	2.00 0.50	mg / m ³ ppm	2	SKC Cat. No. 226-09
21	Ethanol (Ethyl alcohol)	Sorbent Adsorption, Gas Chromatography Method	US. EPA Method 18 / GC-FID	0.21 m ³	0.7 L/min (30 min)	1.88 1.00	mg / m ³ ppm	2	SKC Cat. No. 226-09
22	Ethylbenzene	Sorbent Adsorption, Gas Chromatography Method	US. EPA Method 18 / GC-FID	0.21 m ³	0.7 L/min (30 min)	2.17 0.50	mg / m ³ ppm	2	SKC Cat. No. 226-09
23	Ethylacetate	Sorbent Adsorption, Gas Chromatography Method	US. EPA Method 18 / GC-FID	0.21 m ³	0.7 L/min (30 min)	5.40 1.50	mg / m ³ ppm	2	SKC Cat. No. 226-09
24	Hexane	Sorbent Adsorption, Gas Chromatography Method	US. EPA Method 18 / GC-FID	0.21 m ³	0.7 L/min (30 min)	1.76 0.50	mg / m ³ ppm	2	SKC Cat. No. 226-09
25	Isopropanol (Isopropyl alcohol); IPA	Sorbent Adsorption, Gas Chromatography Method	US. EPA Method 18 / GC-FID	0.21 m ³	0.7 L/min (30 min)	2.46 1.00	mg / m ³ ppm	2	SKC Cat. No. 226-09
26	Methanol (Methyl alcohol)	Sorbent Adsorption, Gas Chromatography Method	US. EPA Method 18 / GC-FID	0.21 m ³	0.7 L/min (30 min)	2.62 2.00	mg / m ³ ppm	2	SKC Cat. No. 226-09
27	Methyl Ethyl Ketone (MEK)	Sorbent Adsorption, Gas Chromatography Method	US. EPA Method 18 / GC-FID	0.21 m ³	0.7 L/min (30 min)	2.95 1.00	mg / m ³ ppm	2	SKC Cat. No. 226-09
28	Styrene	Sorbent Adsorption, Gas Chromatography Method	US. EPA Method 18 / GC-FID	0.21 m ³	0.7 L/min (30 min)	2.13 0.50	mg / m ³ ppm	2	SKC Cat. No. 226-09
29	Toluene	Sorbent Adsorption, Gas Chromatography Method	US. EPA Method 18 / GC-FID	0.21 m ³	0.7 L/min (30 min)	1.88 0.50	mg / m ³ ppm	2	SKC Cat. No. 226-09
30	Methylcyclohexane	Sorbent Adsorption, Gas Chromatography Method	U.S.EPA Method 18/SKC/Guide/ GC-FID	2-23 L	0.10 L/min (1 hr)	0.08 0.02	mg / m ³ ppm	2	SKC Cat. No. ST 226-09

Items	Parameter	Method	Reference Method / Analytical Technique	Air Volume	Sampling Rate / Period	LOQ / Range	Unit	Decimal point	Remark
31	Ketones	Sorbent Adsorption, Gas Chromatography Method	NIOSH2555 (P.1-5) / PS pump / GC-FID	21 L	0.70 L/min (1 hr)	1.88 0.79	mg / m ³ ppm	2	SKC Cat. No. 226-09
32	n-Heptane	Sorbent Adsorption, Gas Chromatography Method	NIOSH1500 (P.1-8) / PS pump / GC-FID	21 L	0.70 L/min (1 hr)	3.89 0.95	mg / m ³ ppm	2	SKC Cat. No. 226-09
33	n-Butyl acetate	Sorbent Adsorption, Gas Chromatography Method	NIOSH 1450(P.1-6) / PS pump / GC-FID	21 L	0.70 L/min (1 hr)	4.75 1.00	mg / m ³ ppm	2	SKC Cat. No. 226-09
34	n-Pentane	Sorbent Adsorption, Gas Chromatography Method	NIOSH 1500(P.1-8) / PS pump / GC-FID	21 L	0.70 L/min (1 hr)	1.50 0.51	mg / m ³ ppm	2	SKC Cat. No. 226-09
35	Chloroform	Sorbent Adsorption, Gas Chromatography Method	NIOSH1003 (P.1-7) / PS pump / GC-FID	21 L	0.70 L/min (1 hr)	2.82 0.58	mg / m ³ ppm	2	SKC Cat. No. 226-09
36	Chlorobenzene	Sorbent Adsorption, Gas Chromatography Method	NIOSH1003 (P.1-7) / PS pump / GC-FID	21 L	0.70 L/min (1 hr)	2.64 0.57	mg / m ³ ppm	2	SKC Cat. No. 226-09
37	Formaldehyde	Sorbent Adsorption, Gas Chromatography Method	NIOSH2541 (P.1-5) / PS pump / GC-FID	21 L	0.70 L/min (1 hr)	0.31 0.25	mg / m ³ ppm	2	SKC Cat. No. 226-118
38	Hydrochloric acid	Sorbent Adsorption, IC Method	EPA Method 26A /IC	0.12 m ³	1 L/min (30 min)	0.015 0.010	mg / m ³ ppm	3	0.1 N H2SO4 / 0.1 N NaOH
39	Hydrofluoric	Sorbent Adsorption, IC Method	EPA Method 26A /IC	0.12 m ³	1 L/min (30 min)	0.012 0.015	mg / m ³ ppm	3	0.1 N H2SO4 / 0.1 N NaOH
40	Nitric	Sorbent Adsorption, IC Method	EPA Method 26A /IC	0.029 m ³	1 L/min (30 min)	0.029 0.010	mg / m ³ ppm	3	0.1 N H2SO4 / 0.1 N NaOH
41	Chlorine	Sorbent Adsorption, IC Method	EPA Method 26A /IC	0.12 m ³	1 L/min (30 min)	0.026 0.010	mg / m ³ ppm	3	Milli-Q Water

เอกสารอ้างอิง

1. Method of Air Sampling and Analysis, APHA Intersociety Committee, 2017
2. NIOSH Manual of Analytical Methods (NMAM)
3. Code of Federal Regulation, U.S. EPA., 40 CFR Part 50, Part 60, 2000
4. Occupational Health and Safety Management System(OSHA) Analytical Methods Manual
5. International Standard Organization, ISO 11204:1995
6. Compendium of Methods for Determination of Inorganic Compound in Ambient Air, U.S. EPA., 1999
7. Annual Book of ASTM Standard, Section 11, 2001

การตรวจวิเคราะห์คุณภาพน้ำ – ภาคตะกอน (Water – Solid wastes Quality Analysis)

ตารางที่ 1 สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ ตามที่ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง : น้ำเสียซึ่งเป็นกรมโรงงานฯ), น้ำเพื่ออุปโภค, น้ำประปา, น้ำผิวดิน, น้ำบาดาล และน้ำทะเล)

จำนวน : ส่วนงานทดสอบพื้นฐาน

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (ml)	MDL	LOQ	Unit	Decimal point	Remark
1.1	Biochemical Oxygen Demand (BOD ₅)	5-Day BOD Test, Membrane Electrode Method	Standard Method part 5210 B, 4500-O G / DO meter	Plastic	1000	~	2.0	mg/l	1	
1.2	Biochemical Oxygen Demand (BOD ₅)	5-Day BOD Test, Azide Modification Method	Standard Method part 5210 B, 4500-O C / Titration	Plastic	1000	~	2.0	mg/l	1	
2.1	Chemical Oxygen Demand (COD)	In-house Method	Standard Method part 5220 C / Titration	Plastic	100	~	40	mg/l as O ₂	0	
2.2	Chemical Oxygen Demand (COD)	Titrimetric, Closed Reflux Method	Standard Method part 5220 C / Titration	Plastic	100	~	40	mg/l as O ₂	0	
3	Free Chlorine	Iodometric Method	Standard Method part 4500-B / Titration	Plastic	100	~	0.50	mg/l	2	
4	Total Dissolved Solids (TDS)	Dried at 180 °C	Standard Method part 2540 C / Gravimetric	Plastic	200	~	25	mg/l	0	
5.1	Grease&Oil	In-house Method	Standard Method part 5520 B / Gravimetric	Glass	1000	~	3.0	mg/l	1	
5.2	Grease&Oil	Partition Gravimetric Method	Standard Method part 5520 B / Gravimetric	Glass	1001	~	3.0	mg/l	1	
6	Sulfide (S ₂)	ZnS Precipitation, Iodometric Method	Standard Method part 4500-S ² / Titration	BOD bottle	300	~	0.53	mg/l as H ₂ S	1	
7	pH	Electrometric Method	Standard Method part 4500 H / pH meter	Plastic	50	~	3.0-12.0	~	1	
8	Total Suspended Solids (TSS)	Dried at 103-105 °C	Standard Method part 2540 D / Gravimetric	Plastic	1000	~	5	mg/l	0	
9	Temperature	Laboratory and Field Method	Standard Method part 2550 B / Thermometer	at field		~	1.0	°C	0	
10	Total Kjeldahl Nitrogen (TKN)	Macro-Kjeldahl Method	Standard Method part 4500-N _{org} / Titration	Plastic	500	~	5	mg/l as NH ₃ -N	0	

การตรวจวิเคราะห์คุณภาพน้ำ – ภาคตะกอน (Water – Solid wastes Quality Analysis)

ตารางที่ 2 สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ ตามที่ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง : น้ำผิวดิน)

จำนวน : ส่วนงานทดสอบพื้นฐาน

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (ml)	MDL	LOQ	Unit	Decimal point	Remark
1	pH	Electrometric Method	Standard Method part 4500 H / pH meter	Plastic	50	~	3.0-12.0	~	1	

การตรวจวิเคราะห์คุณภาพน้ำ - ภาคตะกอน (Water – Solid wastes Quality Analysis)

ตารางที่ 3 สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ ที่ไม่ได้ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง : น้ำ, น้ำเสีย, น้ำเพื่ออุปโภค, น้ำประปา, น้ำโคลน, น้ำบาดาล และน้ำทะเล)

จำนวน : จำนวนทดสอบพื้นฐาน

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (ml)	MDL	LOQ	Unit	Decimal point	Remark
1	Acidity	Titration Method	Standard Method part 2310 B / Titration	Plastic	50	-	20.00	mg/l as CaCO ₃	1	
2	M-Alkalinity	Titration Method	Standard Method part 2320 B / Titration	Plastic	50	-	20.00	mg/l as CaCO ₃	1	
3	P-Alkalinity	Titration Method	Standard Method part 2320 B / Titration	Plastic	50	-	20.00	mg/l as CaCO ₃	1	
4	Ammonia Nitrogen (NH ₃ -N)	Distillation and Titrimetric Method	Standard Method part 4500-NH ₃ / Titration	Plastic	500		2	mg/l as NH ₃ -N	1	
5	Calcium Hardness	EDTA Titrimetric Method	Standard method part 3500-Ca B/ Titration	Plastic	100	-	3.0	mg/l as CaCO ₃	1	
6	Chloride (Cl ⁻)	Argentometric Method	Standard Method part 4500-Cl B / Titration	Plastic	50	-	5.0	mg/l as Cl ⁻	1	
7	Chlorine (Residual)	DPD Colorimetric Method	Standard Method part 4500-Cl G / Test kit	Plastic	500	-	0.1	mg/l as Cl ₂	1	
8	Chlorine (Total)	DPD Colorimetric Method	Modified Standard Method part 4500-Cl G / Test kit	Plastic	500	-	0.1	mg/l as Cl ₂	1	
9	Fixed Solids (FS)	Dried at 550 °C	Standard Method part 2540 E / Gravimetric	Plastic	200	-	30.0	mg/l	1	
10	Hardness	EDTA Titrimetric Method	Standard Method part 2340 C / Titration	Plastic	100	-	6.0	mg/l as CaCO ₃	1	
11	Magnesium (Mg)	Calculation Method	Standard Method part 3500-Mg / Calculation	Plastic	100	-	0.70	mg/l as Mg	1	
12	Magnesium Hardness	Calculation Method	Standard Method part 3500-Mg / Calculation	Plastic	100	-	3.0	mg/l as CaCO ₃	1	

การตรวจวิเคราะห์คุณภาพน้ำ - ภาคตะกอน (Water – Solid wastes Quality Analysis)

ตารางที่ 3 สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ ที่ไม่ได้ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง : น้ำ, น้ำเสีย, น้ำเพื่ออุปโภค, น้ำประปา, น้ำโคลน, น้ำบาดาล และน้ำทะเล)

จำนวน : จำนวนทดสอบพื้นฐาน

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (ml)	MDL	LOQ	Unit	Decimal point	Remark
13	Mix Liquor Suspended Solids (MLSS)	Dried at 103-105 °C	Standard Method part 2540 C / Gravimetric	Plastic	200	-	5	mg/l	1	
14	Mix Liquor Volatile Suspended Solids (MLVSS)	Dried at 550 °C	Standard Method part 2540 E / Gravimetric	Plastic	200	-	5	mg/l	1	
15	Organic Nitrogen	Macro-Kjeldahl Method	Standard Method part 4500-N _{org} / Titration	Plastic	500	-	5	mg/l as NH ₃ -N	1	Org-N =TKN-(Ammonia-N)
17	Conductivity	Laboratory Method	Standard Method part 2510 B	Plastic	200	-	0.1	us/cm		ค่าก่อน ค. 2 ส่วนก่อนวัดค่าอื่น (วัดก่อน)
18	Salinity	Electrical Conductivity Method	Standard Method part 2520 B / Conductivity meter	Plastic	100	-	0.01	ppt		ค่าก่อน ค. 2 ส่วนก่อนวัดค่าอื่น (วัดก่อน)
19	Sludge Volume Index (SV ₃₀)	Volumetric Method	Standard Method part 2540 F / Volumetric	Plastic	1000	-	0.1	ml/l	1	
20	Sulfite	Titrimetric Method	Standard Method part 4500-SO ₃ ²⁻ B / Titration	Plastic	200	-	2.00	mg/l as SO ₃ ²⁻	2	
21	Total Dissolved Solids (TDS)	Dried at 103-105 °C	Modified Standard Method part 2540 B / Gravimetric	Plastic	200	-	25	mg/l	0	
22	Turbidity	Nephelometric Method	Standard Method part 2130 B / Turbidity meter	Plastic	50	0.01	0.01	NTU		ค่าก่อน ค. 2 ส่วนก่อนวัดค่าอื่น (วัดก่อน) NTU=FTU=จลิตาเนต
23	Volatile Fatty Acid	Titrimetric Method	วิธีวิเคราะห์น้ำฉ่ำ เหมตกน้ำกรรมวิธีของกรมโรงงานอุตสาหกรรม / Titration	Plastic	200	-	1.00	mg/l	1	
24	Volatile Solids (VS)	Dried at 550 °C	Standard Method part 2540 E / Gravimetric	Plastic	200		3.0	mg/l	1	
25	Volatile Suspended Solids (VSS)	Dried at 550 °C	Standard Method part 2540 E / Gravimetric	Plastic	200		3.0	mg/l	1	

การตรวจวิเคราะห์คุณภาพน้ำ – ภาคตะกอน (Water – Solid wastes Quality Analysis)

ตารางที่ 3 สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ ที่ไม่ได้นับทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง : น้ำ, น้ำเสีย, น้ำทิ้งอุปโภค, น้ำประปา, น้ำผิวดิน, น้ำบาดาล และน้ำทะเล)

ส่วนงาน : ส่วนงานทดสอบพื้นฐาน

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (ml)	MDL	LOQ	Unit	Decimal point	Remark
26	Dissolved Oxygen(DO)	Azide Modification	Standard Method part 4500-O C/Titration	Plastic	300	-	0.3	mg/l	1	
	จำนวนจุลินทรีย์									
1	Benthos	Counting Chamber Method	Standard Method part 10500 B / Counting	ถุงดำ	-	-	-	ml/m ²	0	รายงานค่าสุญ –Not found
2	Escherichia Coli Bacteria (E.coli)	MPN Test	Standard Method part 9221 F / Fluorogenic Substrate , MPN	Glass	250	-	-	MPN/100 ml	ตามตาราง MPN-	รายงานค่าสุญ 1.1 (นับเต็ม) / 1.8 (นับ)
3	Total Coliform	MPN Test	Standard Method part 9221 B / Fermentation Technique , MPN	Glass	250	-	-	MPN/100 ml	ตามตาราง MPN-	รายงานค่าสุญ 1.1 (นับเต็ม) / 1.8 (นับ)
4	Thermotolerant coliforms (Fecal Coliform)	MPN Test	Standard Method part 9221 E /Thermotolerant Coliform ,MPN	Glass	250	-	-	MPN/100 ml	ตามตาราง MPN-	รายงานค่าสุญ 1.1 (นับเต็ม) / 1.8 (นับ)
5	Heterotrophic Bacteria (Total Bacteria)	Heterotrophic plate count (Standard Plate Count Method)	Standard Method part 9215 B / Pour plate	Glass	250	1	1	Colonies/cm ²	0	*Heterotrophic plate count Standard plate Count
6	Phytoplankton	Counting Chamber Method	Standard Method part 10200 F / Counting	Plastic	-	-	-	Cell / l	0	รายงานค่าสุญ –Not found
7	Zooplankton	Counting Chamber Method	Standard Method part 10200 G / Counting	Plastic	-	-	-	ml/l	0	รายงานค่าสุญ –Not found
8	S.Aureus	Enrichment	Standard Method part 9213 B	Glass	1000	-	-	-	รายงานพบ/ไม่พบ	รายงานค่าสุญ –Not found
9	Salmonella sp.	Membrane Filter	Standard Method part 9260 B	Glass	1000	-	-	-	รายงานพบ/ไม่พบ	รายงานค่าสุญ –Not found
10	Clostridium perfringens	Compendium 2003,Chapter 34	Compendium 2003,Chapter 34	Glass	1000	-	-	-	รายงานพบ/ไม่พบ	รายงานค่าสุญ –Not found

การตรวจวิเคราะห์คุณภาพน้ำ – ภาคตะกอน (Water – Solid wastes Quality Analysis)

ตารางที่ 4 สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ ตามที่ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง : น้ำเสียชุมชนเขตกรมโรงงานฯ), น้ำน้ำทิ้งอุปโภค, น้ำประปา, น้ำผิวดิน, น้ำบาดาล และน้ำทะเล)

ส่วนงาน : ส่วนงานเครื่องมือทดสอบ

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (ml)	MDL	LOQ	Unit	Decimal point	Remark
1	Arsenic (As)	Continuous Hydride Generation-AAS Method	APHA Method Part 3114 B / AAS	Plastic	500	0.0010	0.0020	mg/l as As	4	ไม่ทดสอบ MDL/LOQ = 1.00/2.00 ug/l
2	Barium (Ba)	Digestion,ICP-OES Method	APHA Method part3030F and 3120 B / ICP-OES	Plastic	500	0.02	0.03	mg/l as Ba	2	ไม่ทดสอบ MDL/LOQ = 20/30 ug/l
3	Cadmium (Cd)	Digestion,ICP-OES Method	APHA Method part3030F and 3120 B / ICP-OES	Plastic	500	0.02	0.03	mg/l as Cd	2	ไม่ทดสอบ MDL/LOQ = 20/30 ug/l ไม่เต็ม MDL/LOQ = 0.002/0.003 mg/l
4	Chromium (Cr)	Digestion,ICP-OES Method	APHA Method part3030F and 3120 B / ICP-OES	Plastic	500	0.02	0.03	mg/l as Cr	2	ไม่ทดสอบ MDL/LOQ = 20/30 ug/l
5	Color	ADMI Weighted-Ordinate Spectrophotometer Method	APHA Method part 2120 F / Spectrophotometer	Plastic	500	10	20.00	ADMI	0	
6	Chromium Hexavalence (Cr ⁶⁺)	Filtration,Colorimetric Method	APHA Method part 3500-Cr B / Spectrophotometer	Plastic	500	0.003	0.050	mg/l as Cr ⁶⁺	3	ไม่ทดสอบ MDL/LOQ = 3.00/50.0 ug/l
7	Copper (Cu)	Digestion,ICP-OES Method	APHA Method part3030F and 3120 B / ICP-OES	Plastic	500	0.02	0.03	mg/l as Cu	2	ไม่ทดสอบ MDL/LOQ = 20/30 ug/l
8	Cyanide (CN ⁻)	Distillation, Colorimetric Method	APHA Method part 4500 CN ⁻ C,E/ Spectrophotometer	Plastic	500	0.008	0.020	mg/l	3	ไม่ทดสอบ MDL/LOQ = 8/20 ug/l
9	Formaldehyde	Distillation, Colorimetric Method	คู่มือวิเคราะห์น้ำเสีย,สามารถวัดความเข้มข้นได้แม่นยำ ประเทศไทย	Plastic	500	0.20	0.50	mg/l	2	
10	Lead (Pb)	Digestion,ICP-OES Method	APHA Method part3030F and 3120 B / ICP-OES	Plastic	500	0.02	0.03	mg/l as Pb	2	ไม่ทดสอบ MDL/LOQ = 20/30 ug/l ไม่เต็ม MDL/LOQ = 0.005/0.010 mg/l
11	Manganese (Mn)	Digestion,ICP-OES Method	APHA Method part3030F and 3120 B / ICP-OES	Plastic	500	0.02	0.03	mg/l as Mn	2	ไม่ทดสอบ MDL/LOQ = 20/30 ug/l
12	Mercury (Hg)	In-house Method:APHA (3112B)	APHA Method part 3112 B / AAS	Plastic	500	0.0005	0.0010	mg/l as Hg	4	

การตรวจวิเคราะห์คุณภาพน้ำ – ภาคตะกอน (Water – Solid wastes Quality Analysis)
ฉบับที่ 4 สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ ตามที่ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง : น้ำเสีย(ขึ้นทะเบียนกรมโรงงาน), น้ำเพื่ออุปโภค, น้ำประปา, น้ำผิวดิน, น้ำบาดาล และน้ำทะเล)

ส่วนงาน : ส่วนงานเครื่องมือทดสอบ

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (ml)	MDL	LOQ	Unit	Decimal point	Remark
13	Nickel (Ni)	Digestion,ICP-OES Method	APHA Method part 3030F and 3120 B / ICP-OES	Plastic	500	0.02	0.03	mg/l as Ni	2	ไม่ระบุ MDL/LOQ = 20/30 ug/l
14	Phenols	Distillation, Direct Photometric Method	APHA Method part 5530 D / Spectrophotometer	Plastic	500	0.002	0.005	mg/l	3	
15	Trivalent Chromium (Cr ³⁺)	Digestion,Direct Aspiration-AAS Method; Filtration,Colorimetric Method;Calculation	APHA Method part 3500-Cr B & part 3120B /AAS	Plastic	500	0.05	0.10	mg/l	2	
16	Trivalent Chromium (Cr ³⁺)	Digestion,ICP-OES Method; Filtration,Colorimetric Method;Calculation	APHA Method part 3500-Cr B & part 3120B / ICP-OES	Plastic	500	0.02	0.05	mg/l	2	
17	Zinc (Zn)	Digestion,ICP-OES Method	APHA Method part 3030F and 3120 B / ICP-OES	Plastic	500	0.02	0.03	mg/l as Zn	2	ไม่ระบุ MDL/LOQ = 20/30 ug/l
18	Free Chlorine	DPD Colorimetric Method	APHA Method part 4500 Cl ₂ G/ Spectrophotometer	Plastic	500	0.03	0.050	mg/l	3	
19	Selenium (Se)	Continuous,Hydride Generation/AAS	APHA Method part 3030F , 3114 B and 3114C	Plastic	500	0.0010	0.0020	mg/l	4	
20	สารกำจัดวัชพืชอะโรมาติก (Pesticide) :	Liquid-Liquid Extraction Gas Chromatography	APHA Method part 6630B/GC	Glass	2500	0.03	0.05	ug/l	2	
	- alpha - BHC					0.03	0.05	ug/l	2	
	- beta - BHC					0.03	0.05	ug/l	2	
	- gamma - BHC					0.03	0.05	ug/l	2	
	- delta - BHC					0.03	0.05	ug/l	2	
	- Heptachlor					0.03	0.05	ug/l	2	
	- Aldrin					0.03	0.05	ug/l	2	
	- Heptachlor epoxide					0.03	0.05	ug/l	2	
	- Endosulfan I					0.03	0.05	ug/l	2	
	- p,p - DDE					0.03	0.05	ug/l	2	
	- Dieldrin					0.03	0.05	ug/l	2	
	- Endrin ketone					0.03	0.05	ug/l	2	

การตรวจวิเคราะห์คุณภาพน้ำ – ภาคตะกอน (Water – Solid wastes Quality Analysis)
ฉบับที่ 4 สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ ตามที่ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง : น้ำเสีย(ขึ้นทะเบียนกรมโรงงาน), น้ำเพื่ออุปโภค, น้ำประปา, น้ำผิวดิน, น้ำบาดาล และน้ำทะเล)

ส่วนงาน : ส่วนงานเครื่องมือทดสอบ

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (ml)	MDL	LOQ	Unit	Decimal point	Remark
	- Endosulfan II					0.03	0.05	ug/l	2	
	- p,p - DDD					0.03	0.05	ug/l	2	
	- Endrin Aldehyde					0.03	0.05	ug/l	2	
	- Endosulfan Sulfate					0.03	0.05	ug/l	2	
	- trans Chlordane					0.03	0.05	ug/l	2	
	- cis Chlordane					0.03	0.05	ug/l	2	

การตรวจวิเคราะห์คุณภาพน้ำ – ภาคตะกอน (Water – Solid wastes Quality Analysis)
กฎที่ ๕. สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ ตามที่ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง : น้ำใต้ดิน)
จำนวน : จำนวนเครื่องมือทดสอบ

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (ml)	MDL	LOQ	Unit	Decimal point	Remark
1	Antimony (Sb)	Digestion,ICP-OES Method	Standard Method part3030F and 3120 B / ICP-OES	Plastic	500	0.05	0.10	mg/l as Sb	2	
2	Arsenic (As)	Digestion,ICP-OES Method	Standard Method part 3120 B / ICP-OES	Plastic	500	0.0500	0.1000	mg/l as As	4	
3	Arsenic (As)	Continuous Hydride Generation-AAS Method	Standard Method Part 3114 B / AAS	Plastic	500	0.0005	0.0020	mg/l as As	4	
4	Barium (Ba)	Digestion,ICP-OES Method	Standard Method part3030F and 3120 B / ICP-OES	Plastic	500	0.02	0.03	mg/l as Ba	2	
5	Beryllium (Be)	Digestion,ICP-OES Method	Standard Method part3030F and 3120 B / ICP-OES	Plastic	500	0.005	0.01	mg/l as Be	2	
6	Cadmium (Cd)	Digestion,ICP-OES Method	Standard Method part3030F and 3120 B / ICP-OES	Plastic	500	0.02	0.03	mg/l as Cd	2	
8	Chromium (Cr)	Digestion,ICP-OES Method	Standard Method part3030F and 3120 B / ICP-OES	Plastic	500	0.02	0.03	mg/l as Cr	2	
9	Cyanide (CN)	Distillation, Colorimetric Method	Standard Method part 4500 CN C,E/ Spectrophotometer	Plastic	500	0.008	0.020	mg/l	3	
10	Chromium Hexavalence (Cr ⁶⁺)	Filtration,Colorimetric Method	Standard Method part 3500-Cr B / Spectrophotometer	Plastic	500	0.003	0.050	mg/l as Cr ⁶⁺	3	
12	Lead (Pb)	Digestion,ICP-OES Method	Standard Method part3030F and 3120 B / ICP-OES	Plastic	500	0.02	0.03	mg/l as Pb	2	
13	Manganese (Mn)	Digestion,ICP-OES Method	Standard Method part3030F and 3120 B / ICP-OES	Plastic	500	0.02	0.03	mg/l as Mn	2	
14	Mercury (Hg)	In-house Method :APHA2012 (3112B)	Standard Method part 3112 B / AAS	Plastic	500	0.0005	0.0010	mg/l as Hg	4	

การตรวจวิเคราะห์คุณภาพน้ำ – ภาคตะกอน (Water – Solid wastes Quality Analysis)
กฎที่ ๕. สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ ตามที่ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง : น้ำใต้ดิน)
จำนวน : จำนวนเครื่องมือทดสอบ

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (ml)	MDL	LOQ	Unit	Decimal point	Remark
16	Nickel (Ni)	Digestion,ICP-OES Method	Standard Method part3030F and 3120 B / ICP-OES	Plastic	500	0.02	0.03	mg/l as Ni	2	
17	Phenols	Distillation, Direct Photometric Method	Standard Method part 5530 D / Spectrophotometer	Plastic	500	0.002	0.005	mg/l	3	
18	Silver (Ag)	Digestion,ICP-OES Method	Standard Method part3030F and 3120 B / ICP-OES	Plastic	500	0.02	0.05	mg/l as Ag	2	
19	Trivalent Chromium (Cr ³⁺)	Digestion,Direct Aspiration-AAS Method; Filtration,Colorimetric Method;Calculation	Standard Method part 3500-Cr B & part 3120B /AAS	Plastic	500	0.05	0.10	mg/l	2	
20	Trivalent Chromium (Cr ³⁺)	Digestion,ICP-OES Method; Filtration,Colorimetric Method;Calculation	Standard Method part 3500-Cr B & part 3120B / ICP-OES	Plastic	500	0.05	0.05	mg/l	2	
21	Vanadium (V)	ICP-OES Method	Standard Method part3030F and 3120 B / ICP-OES	Plastic	500	0.01	0.02	mg/l as V	2	
22	Zinc (Zn)	Digestion,ICP-OES Method	Standard Method part3030F and 3120 B / ICP-OES	Plastic	500	0.02	0.03	mg/l as Zn	2	
23	Selenium (Se)	Continous,Hydride Generation/AAS	APHA Method part3030F , 3114 B and 3114C	Plastic	500	0.0020	0.0050	mg/l	4	จำกัดสอบ 1 ม.ก. 2565
24	Volatile organic compounds:VOC#1	Purge-and-Trap /GC-MS	APHA Method part 6200B	Glass	40 *4					
	- Benzene					0.00025	0.00050	mg/l	5	
	- Bromodichloromethane					0.00050	0.00050	mg/l	5	
	- Bromoform					0.00050	0.00050	mg/l	5	
	- Carbon tetrachloride					0.00025	0.00025	mg/l	5	
	- Chlorobenzene					0.00025	0.00050	mg/l	5	
	- Chlorodibromomethane					0.00050	0.00100	mg/l	5	
	- 1,2-Dichlorobenzene					0.00025	0.00050	mg/l	5	
	- 1,3-Dichlorobenzene					0.00025	0.00025	mg/l	5	
	- 1,4-Dichlorobenzene					0.00025	0.00025	mg/l	5	

การตรวจวิเคราะห์คุณภาพน้ำ - ภาคตะกอน (Water – Solid wastes Quality Analysis)
ฉบับที่ 5. สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ **ตามที่ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม**
(ประเภทตัวอย่าง : น้ำใต้ดิน)

จำนวน : ส่วนงานเครื่องมือทดสอบ

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (ml)	MDL	LOQ	Unit	Decimal point	Remark
10	• 1,1-Dichloroethane					0.00025	0.00025	mg/l	5	
11	• 1,2-Dichloroethane					0.00025	0.00050	mg/l	5	
12	• 1,1-Dichloroethylene					0.00025	0.00050	mg/l	5	
13	• cis-1,2-Dichloroethylene					0.00050	0.00050	mg/l	5	
14	• trans-1,2-Dichloroethylene					0.00025	0.00050	mg/l	5	
15	• 1,2-Dichloropropane					0.00025	0.00050	mg/l	5	
16	• 1,3-Dichloropropane					0.00025	0.00050	mg/l	5	
17	• Ethylbenzene					0.00025	0.00050	mg/l	5	
18	• Methyl tert-butyl ether					0.00025	0.00050	mg/l	5	
19	• Naphthalene					0.00025	0.00100	mg/l	5	
20	• Nitrobenzene					0.00025	0.00025	mg/l	5	
21	• Styrene					0.00050	0.00100	mg/l	5	
22	• 1,1,2,2-Tetrachloroethane					0.00050	0.00050	mg/l	5	
23	• Tetrachloroethylene					0.00025	0.00050	mg/l	5	
24	• Toluene					0.00025	0.00050	mg/l	5	
25	• 1,2,4-Trichlorobenzene					0.00025	0.00050	mg/l	5	
26	• 1,1,1-Trichloroethane					0.00025	0.00025	mg/l	5	
27	• 1,1,2-Trichloroethane					0.00025	0.00050	mg/l	5	
28	• Trichloroethylene					0.00025	0.00050	mg/l	5	
29	• 1,3,5-Trimethylbenzene					0.00025	0.00100	mg/l	5	
30	• Vinyl acetate					0.00050	0.00100	mg/l	5	
31	• Vinyl Chloride					0.00025	0.00025	mg/l	5	
32	• m-Xylene					0.00025	0.00100	mg/l	5	
33	• o-Xylene					0.00025	0.00100	mg/l	5	
34	• p-Xylene					0.00025	0.00100	mg/l	5	

การตรวจวิเคราะห์คุณภาพน้ำ - ภาคตะกอน (Water – Solid wastes Quality Analysis)
ฉบับที่ 5. สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ **ตามที่ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม**
(ประเภทตัวอย่าง : น้ำใต้ดิน)

จำนวน : ส่วนงานเครื่องมือทดสอบ

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (ml)	MDL	LOQ	Unit	Decimal point	Remark
25	• Xylene Total					0.00075	0.00100	mg/l	5	
25	Volatile organic compounds (VOC) #2	Purge-and-Trap / GC-MS Method	APHA Method part 6200B	Glass	40 *4					
26	• Acetone					0.00100	0.00100	mg/l	5	
26	• Butanol					0.00100	0.00100	mg/l	5	
26	• Carbon disulfide					0.00200	0.00500	mg/l	5	
26	• chloroform					0.00100	0.00200	mg/l	5	
26	• n-Hexane					0.00100	0.00200	mg/l	5	
26	• Dichloromethane					0.00200	0.00200	mg/l	5	
26	Semivolatile organic compounds #1	Liquid-Liquid Extraction / GC-MS (SM: 6410B)	APHA Method part 6410B	Glass	2500					
26	Acenaphthene					0.0005	0.0010	mg/l	4	
26	Anthracene					0.0005	0.0010	mg/l	4	
26	Benz[a]anthracene					0.0005	0.0010	mg/l	4	
26	Benz[b]fluoranthene					0.0005	0.0010	mg/l	4	
26	Benz[k]fluoranthene					0.0005	0.0010	mg/l	4	
26	Benz[a]pyrene					0.0005	0.0010	mg/l	4	
26	Benz[ghi]perylene					0.0005	0.0010	mg/l	4	
26	Bis(2-chloroethyl) ether					0.0005	0.0010	mg/l	4	
26	Bis(2-ethylhexyl) phthalate					0.0005	0.0010	mg/l	4	
26	Butyl benzyl phthalate					0.0005	0.0010	mg/l	4	
26	Carbazole					0.0005	0.0100	mg/l	4	
26	p-Chloroaniline					0.0005	0.0010	mg/l	4	
26	2-Chlorophenol					0.0005	0.0010	mg/l	4	
26	Chrysene					0.0005	0.0010	mg/l	4	
26	Dibenz[a,h]anthracene					0.0005	0.0100	mg/l	4	
26	Di-n-butyl phthalate					0.0005	0.0010	mg/l	4	

การตรวจวิเคราะห์คุณภาพน้ำ – ภาคตะกอน (Water – Solid wastes Quality Analysis)
ตารางที่ ๕ สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ ตามที่ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง: น้ำใต้ดิน)
ส่วนรวม : ส่วนงานเครื่องมือทดสอบ

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (ml)	MDL	LOQ	Unit	Decimal point	Remark
	2,4-Dichlorophenol					0.0005	0.0010	mg/l	4	
	Diethyl Phthalate					0.0005	0.0010	mg/l	4	
	2,4-Dimethylphenol					0.0005	0.0010	mg/l	4	
	2,4-Dinitrotoluene					0.0005	0.0010	mg/l	4	
	2,6-Dinitrotoluene					0.0005	0.0010	mg/l	4	
	Di-n-octyl phthalate					0.0005	0.0010	mg/l	4	
	Fluoranthene					0.0005	0.0010	mg/l	4	
	Fluorene					0.0005	0.0010	mg/l	4	
	Hexachlorobenzene					0.0005	0.0010	mg/l	4	
	Hexachloro-1,3-butadiene					0.0005	0.0100	mg/l	4	
	Hexachlorocyclopentadiene					0.0005	0.0010	mg/l	4	
	Hexachloroethane					0.0005	0.0010	mg/l	4	
	Indeno[1,2,3-cd]pyrene					0.0005	0.0010	mg/l	4	
	Isophorone					0.0005	0.0010	mg/l	4	
	2-Methylphenol (o-Cresol)					0.0005	0.0010	mg/l	4	
	2-Methylnaphthalene					0.0005	0.0010	mg/l	4	
	N-Nitrosodi-n-propylamine					0.0005	0.0010	mg/l	4	
	Phenanthrene					0.0005	0.0010	mg/l	4	
	Phenol					0.0005	0.0010	mg/l	4	
	Pyrene					0.0005	0.0010	mg/l	4	
	2,4,5-Trichlorophenol					0.0005	0.0010	mg/l	4	
	2,4,6-Trichlorophenol									

การตรวจวิเคราะห์คุณภาพน้ำ – ภาคตะกอน (Water – Solid wastes Quality Analysis)
ตารางที่ ๖ สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ ที่ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง: น้ำ, น้ำเสีย, น้ำใต้ดิน, น้ำเพื่ออุปโภค, น้ำประปา, น้ำดิน, น้ำบาดาล และน้ำทะเล)
ส่วนรวม : ส่วนงานเครื่องมือทดสอบ

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (ml)	MDL	LOQ	Unit	Decimal point	Remark
1	Antimony (Sb)	Digestion, ICP-OES Method	Standard Method part 3030F.3120 B / ICP-OES	Plastic	500	0.05	0.10	mg/l as Sb	2	
2	Aluminium (Al)	Digestion, ICP-OES Method	Standard Method part 3030F.3120 B / ICP-OES	Plastic	500	0.05	0.10	mg/l as Sb	2	
3	Boron (B)	Digestion, ICP-OES Method	Standard Method part 3030F.3120 B / ICP-OES	Plastic	500	0.01	0.02	mg/l as B	2	
4	Calcium (Ca)	Digestion, ICP-OES Method	Standard Method part 3030F.3120 B / ICP-OES	Plastic	500	0.50	1.00	mg/l as B	2	
5	Cadmium (Cd)	Digestion, ICP-OES Method	Standard Method part 3030F.3120 B / ICP-OES	Plastic	500	0.002	0.003	mg/l as Cd	3	น้ำดื่ม
6	Cobalt (Co)	Digestion, ICP-OES Method	Standard Method part 3030F.3120 B / ICP-OES	Plastic	500	0.01	0.02	mg/l as Co	2	
7	Color	Spectrophotometric Method	Standard Method part 2120 C / Spectrophotometer	Plastic	500	0.50	1.00	Pt-Co	2	
8	Iron (Fe)	Digestion, ICP-OES Method	Standard Method part 3030F.3120 B / ICP-OES	Plastic	500	0.02	0.03	mg/l as Cd	2	
10	Lead (Pb)	Digestion, ICP-OES Method	Standard Method part 3030F.3120 B / ICP-OES	Plastic	500	0.005	0.010	mg/l as Pb	3	น้ำดื่ม
11	Magnesium (Mg)	Digestion, ICP-OES Method	Standard Method part 3030F.3120 B / ICP-OES	Plastic	500	0.50	1.00	mg/l as Mg	2	
12	Molybdenum (Mo)	Digestion, ICP-OES Method	Standard Method part 3030F.3120 B / ICP-OES	Plastic	500	0.01	0.02	mg/l as Mo	2	
13	Nitrite (NO ₂ -)	Colorimetric Method	Standard Method part 4500-NO ₂ -B / Spectrophotometer	Plastic	500	0.010	0.030	mg/l as NO ₂ -	3	
14	Nitrite-Nitrogen (NO ₂ -N)	Colorimetric Method	Standard Method part 4500-NO ₂ -B / Spectrophotometer	Plastic	500	0.02	0.10	mg/l as NO ₂ -N	3	

การตรวจวิเคราะห์คุณภาพน้ำ – ภาคตะกอน (Water – Solid wastes Quality Analysis)
ฉบับที่ ๕ สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ ที่ไม่ได้ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง : น้ำ, น้ำเสีย,น้ำใต้ดิน, น้ำเพื่ออุปโภค, น้ำประปา, น้ำผิวดิน, น้ำบาดาล และน้ำทะเล)
ส่วนงาน : ส่วนงานเครื่องมือทดสอบ

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (ml)	MDL	LOQ	Unit	Decimal point	Remark
15	Nitrate (NO ₃ ⁻)	Colorimetric Method	Standard Method part 4500-NO ₃ ⁻ B / Spectrophotometer	Plastic	500	0.09	0.44	mg/l as NO ₃ ⁻	3	
16	Nitrate-Nitrogen (NO ₃ ⁻ -N)	Colorimetric Method	Standard Method part 4500-NO ₃ ⁻ B / Spectrophotometer	Plastic	500	0.02	0.10	mg/l as NO ₃ ⁻ -N	3	
17	Potassium (K)	Direct Aspiration-AAS Method	Standard Method part 3111 B / AAS	Plastic	500	0.008	0.025	mg/l as K	3	
18	Potassium (K)	Digestion,ICP-OES Method	Standard Method part 3030F,3120 B / ICP-OES	Plastic	500	0.5	1	mg/l as K	2	
19	Selenium (Se)	Digestion,ICP-OES Method	Standard Method part 3030F,3120 B / ICP-OES	Plastic	500	0.05	0.10	mg/l as Si	2	
20	Silica (SiO ₂)	Molybdosilicate Method	Standard Method part 4500-SiO ₂ C / Spectrophotometer	Plastic	500	0.20	0.40	mg/l as SiO ₂	2	
21	Silicon (Si)	Digestion,ICP-OES Method	Standard Method part 3030F,3120 B / ICP-OES	Plastic	500	0.02	0.05	mg/l as Si	2	
22	Silver (Ag)	Digestion,ICP-OES Method	Standard Method part 3030F,3120 B / ICP-OES	Plastic	500	0.02	0.05	mg/l as Ag	2	
23	Sodium (Na)	Direct Aspiration-AAS Method	Standard Method part 3111 B / AAS	Plastic	500	0.005	0.050	mg/l as Na	3	
24	Sodium (Na)	Digestion,ICP-OES Method	Standard Method part 3030F,3120 B / ICP-OES	Plastic	500	0.50	1.00	mg/l as Na	2	
25	Sodium Absorption Ratio (SAR)	Calculation,Digestion,ICP-OES Method	Standard Method part 3030F,3120 B / ICP-OES	Plastic	500	0.50	1.00	-	2	
26	Strontium (Sr)	Digestion,ICP-OES Method	Standard Method part 3030F,3120 B / ICP-OES	Plastic	500	0.01	0.02	mg/l as Sr	2	
27	Tin (Sn)	Digestion,ICP-OES Method	Standard Method part 3030F,3120 B / ICP-OES	Plastic	500	0.05	0.10	mg/l as Sn	2	

การตรวจวิเคราะห์คุณภาพน้ำ – ภาคตะกอน (Water – Solid wastes Quality Analysis)
ฉบับที่ ๕ สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ ที่ไม่ได้ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง : น้ำ, น้ำเสีย,น้ำใต้ดิน, น้ำเพื่ออุปโภค, น้ำประปา, น้ำผิวดิน, น้ำบาดาล และน้ำทะเล)
ส่วนงาน : ส่วนงานเครื่องมือทดสอบ

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (ml)	MDL	LOQ	Unit	Decimal point	Remark
28	Titanium (Ti)	Digestion,ICP-OES Method	Standard Method part 3030F,3120 B / ICP-OES	Plastic	500	0.01	0.02	mg/l as Ti	2	
29	Thallium (Tl)	Digestion,ICP-OES Method	Standard Method part 3030F,3120 B / ICP-OES	Plastic	500	0.05	0.1	mg/l as Ti	2	
30	Vanadium (V)	Digestion,ICP-OES Method	Standard Method part 3030F,3120 B / ICP-OES	Plastic	500	0.01	0.02	mg/l as V	2	
31	Phosphate (PO ₄ ⁻³)	Ascorbic Acid Method	Standard Method part 4500-PO ₄ ⁻³ B/ Spectrophotometer	Plastic	500	0.03	0.46	mg/l as P	2	
32	Phosphorus (P)	Ascorbic Acid Method	Standard Method part 4500-P B/ Spectrophotometer	Plastic	500	0.01	0.15	mg/l as PO ₄ ⁻³	2	
33	Sulfate (SO ₄ ⁻²)	Turbidimetric Method	Standard Method part 4500-SO ₄ ⁻² E/ Spectrophotometer	Plastic	500	1.50	5.00	mg/l as SO ₄ ⁻²	2	
34	Surfactant	Anionic Surfactants as MBAS	Standard Method Part 5540 C / Spectrophotometer	Plastic	500	0.35	0.40	mg/l as MBAS	2	
35	Surfactant (I AS)	Anionic Surfactants as MBAS	Standard Method Part 5540 C / Spectrophotometer	Plastic	1000	0.05	0.10	mg/l as MBAS	2	
36	Fluoride (F ⁻)	Ion-Selective Electrode Method	Standard Method part 4500-F C/ Spectrophotometer	Plastic	100	0.20	0.50	mg/l as F	2	
37	Gold (Au)	Digestion,ICP-OES Method	Standard Method part 3030F,3120 B / ICP-OES	Plastic	500	0.02	0.05	mg/l as Au	2	

การตรวจวิเคราะห์คุณภาพน้ำ – ภาคตะกอน (Water – Solid wastes Quality Analysis)
ตารางที่ 2 สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบด้วยห้องปฏิบัติการ ตามที่ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง : ภาคตะกอน ตามประกาศเรื่องสิ่งปฏิกูลที่ไม่ใช่สิ่งแฉะ และ ดิน)

จำนวน : ส่วนงานเครื่องมือทดสอบ

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (g)	MDL	LOQ	Unit	Decimal point	Remark
1	Antimony (Sb)	Waste Extraction , ICP-OES Method Digestion,ICP-OES Method	SW 846 Method 3050B / ICP-OES	Plastic	500	0.05 2.50	0.10 5.00	mg/l as Sb mg/kg as Sb	2	
2	Arsenic (As)	Waste Extraction , ICP-OES Method Digestion,ICP-OES Method	SW 846 Method 3050B / ICP-OES	Plastic	500	0.05 2.50	0.10 5.00	mg/l as As mg/kg as As	2	
3	Barium (Ba)	Waste Extraction , ICP-OES Method Digestion,ICP-OES Method	SW 846 Method 3050B / ICP-OES	Plastic	500	0.02 0.50	0.03 1.00	mg/l as Ba mg/kg as Ba	2	
4	Beryllium (Be)	Waste Extraction , ICP-OES Method Digestion,ICP-OES Method	SW 846 Method 3050B / ICP-OES	Plastic	500	0.005 0.50	0.01 1.00	mg/l as Be mg/kg as Be	2	
5	Cadmium (Cd)	Waste Extraction , ICP-OES Method Digestion,ICP-OES Method	SW 846 Method 3050B / ICP-OES	Plastic	500	0.02 0.50	0.03 1.00	mg/l as Cd mg/kg as Cd	2	
6	Chromium (Cr)	Waste Extraction , ICP-OES Method Digestion,ICP-OES Method	SW 846 Method 3050B / ICP-OES	Plastic	500	0.02 0.50	0.03 1.00	mg/l as Cr mg/kg as Cr	2	
7	Cobalt (Co)	Waste Extraction , ICP-OES Method Digestion,ICP-OES Method	SW 846 Method 3050B / ICP-OES	Plastic	500	0.01 0.50	0.02 1.00	mg/l as Co mg/kg as Co	2	
8	Copper (Cu)	Waste Extraction , ICP-OES Method Digestion,ICP-OES Method	SW 846 Method 3050B / ICP-OES	Plastic	500	0.02 0.50	0.03 1.00	mg/l as Cu mg/kg as Cu	2	
9	Hexavalent Chromium (Cr ⁶⁺)	Colorimetric Method/Spectrophotometer Alkaline Digestion,Colorimetric Method/ Spectrophotometer	SW 846 Method 3060A,7196A / Spectrophotometer	Plastic	500	0.003 0.40	0.050 2.00	mg/l as Cr mg/kg as Cr	3 2	
10	Lead (Pb)	Waste Extraction , ICP-OES Method Digestion,ICP-OES Method	SW 846 Method 3050B / ICP-OES	Plastic	500	0.02 0.50	0.03 1.50	mg/l as Pb mg/kg as Pb	2	
11	Mercury (Hg)	Waste Extraction , ICP-OES Method Digestion,Cold Vapor Technique-AAS Method	SW 846 Method 7471B / AAS	Plastic	500	0.0005 0.10	0.0010 0.20	mg/l as Hg mg/kg as Hg	4 2	
12	Molybdenum (Mo)	Waste Extraction , ICP-OES Method Digestion,ICP-OES Method	SW 846 Method 3050B / ICP-OES	Plastic	500	0.01 0.50	0.02 1.00	mg/l as Mo mg/kg as Mo	2	
13	Nickel (Ni)	Waste Extraction , ICP-OES Method	SW 846 Method 3050B / ICP-OES	Plastic	500	0.02	0.03	mg/l as Ni	2	

การตรวจวิเคราะห์คุณภาพน้ำ – ภาคตะกอน (Water – Solid wastes Quality Analysis)
ตารางที่ 3 สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบด้วยห้องปฏิบัติการ ตามที่ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง : ภาคตะกอน ตามประกาศเรื่องสิ่งปฏิกูลที่ไม่ใช่สิ่งแฉะ และ ดิน)

จำนวน : ส่วนงานเครื่องมือทดสอบ

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (g)	MDL	LOQ	Unit	Decimal point	Remark
		Digestion,ICP-OES Method				0.50	1.00	mg/kg as Ni		
14	Selenium (Se)	Waste Extraction , ICP-OES Method Digestion,ICP-OES Method	SW 846 Method 3050B / ICP-OES	Plastic	500	0.05 2.50	0.10 5.00	mg/l as Se mg/kg as Se	2	
15	Silver (Ag)	Waste Extraction , ICP-OES Method Digestion,ICP-OES Method	SW 846 Method 3050B / ICP-OES	Plastic	500	0.02 1.00	0.05 2.50	mg/l as Ag mg/kg as Ag	2	
16	Thallium (Tl)	Waste Extraction , ICP-OES Method Digestion,ICP-OES Method	SW 846 Method 3050B / ICP-OES	Plastic	500	0.05 2.50	0.10 5.00	mg/l as V mg/kg as V	2	
17	Vanadium (V)	Waste Extraction , ICP-OES Method Digestion,ICP-OES Method	SW 846 Method 3050B / ICP-OES	Plastic	500	0.01 0.50	0.02 1.00	mg/l as V mg/kg as V	2	
18	Zinc (Zn)	Waste Extraction , ICP-OES Method Digestion,ICP-OES Method	SW 846 Method 3050B / ICP-OES	Plastic	500	0.02 0.50	0.03 1.00	mg/l as Zn mg/kg as Zn	2	

การตรวจวิเคราะห์คุณภาพน้ำ – ภาคตะกอน (Water – Solid wastes Quality Analysis)

ตารางที่ 9 สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ ตามที่ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง : ดิน)

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (g)	MDL	LOQ	Unit	Decimal point	Remark
1	Arsenic (As)	Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	2.50	5.00	mg/kg as As	2	
2	Antimony (Sb)	Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	2.50	5.00	mg/kg as Sb	2	
3	Barium (Ba)	Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	0.50	1.00	mg/kg as Ba	2	
4	Beryllium (Be)	Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	0.50	1.50	mg/kg as Be	2	
5	Cadmium (Cd)	Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	1.00	1.50	mg/kg as Cd	2	
6	Chromium (Cr)	Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	1.00	1.50	mg/kg as Cr	2	
7	Hexavalent Chromium (Cr ⁶⁺)	Digestion, Colorimetric Method	US EPA SW 846 Method 3060A, 7196A / Spectrophotometer	Plastic	500	0.40	2.00	mg/kg as Cr	3	
8	Lead (Pb)	Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	0.50	1.00	mg/kg as Pb	2	
9	Manganese (Mn)	Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	0.50	1.00	mg/kg as Mn	2	
10	Mercury (Hg)	Digestion, Cold Vapor Technique-AAS Method	US EPA SW 846 Method 7471B / AAS	Plastic	500	0.10	0.20	mg/kg as Hg	4	
11	Nickel (Ni)	Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	1.00	1.50	mg/kg as Ni	2	
12	Selenium (Se)	Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	2.50	5.00	mg/kg as Se	2	
13	Silver (Ag)	Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	1.00	2.50	mg/kg as Ag	2	
14	Trivalent Chromium (Cr ³⁺)	Digestion, ICP-OES; Filtration, Colorimetric Method, Calculation/	US EPA SW 846 Method 3060A, 7196A / Spectrophotometer	Plastic	500	0.40	2.00	mg/kg as Cr	3	
15	Vanadium (V)	Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	0.50	1.00	mg/kg as V	2	
16	Zinc (Zn)	Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	0.50	1.00	mg/kg as Zn	2	
17	Volatile organic compounds (VOC)	Purge-and-Trap / GC-MS	US EPA SW 846 Method 5035A and 8260D	Glass	50					
	-Acetone	Purge-and-Trap / GC-MS	US EPA SW 846 Method 5035A and 8260D	Glass	50	0.005	0.010	mg/kg	3	
	-Benzene	Purge-and-Trap / GC-MS	US EPA SW 846 Method 5035A and 8260D	Glass	50	0.005	0.010	mg/kg	3	

การตรวจวิเคราะห์คุณภาพน้ำ – ภาคตะกอน (Water – Solid wastes Quality Analysis)

ตารางที่ 9 สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ ที่ไม่ได้ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง : ภาคตะกอน ตามประกาศเรื่องรังสีปฏิฤทธิ์ที่ไม่ใช่เชื้อ)

จำนวน : จำนวนเครื่องมือทดสอบ

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (ml)	MDL	LOQ	Unit	Decimal point	Remark
1	Aluminium (Al)	Waste Extraction , ICP-OES Method Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	0.05 2.50	0.10 5.00	mg/l as Al mg/kg as Al	2 2	
2	Boron (B)	Waste Extraction , ICP-OES Method Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	0.01 0.50	0.02 1.00	mg/l as B mg/kg as B	2 2	
3	Calcium (Ca)	Waste Extraction , ICP-OES Method Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	0.05 25.0	0.10 50.0	mg/l as Ca mg/kg as Ca	2 1	
4	Iron (Fe)	Waste Extraction , ICP-OES Method Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	0.02 1.00	0.03 1.50	mg/l as Fe mg/kg as Fe	2 2	
5	Magnesium (Mg)	Waste Extraction , ICP-OES Method Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	0.05 25.0	0.10 50.0	mg/l as Mg mg/kg as Mg	2 1	
6	Manganese (Mn)	Waste Extraction , ICP-OES Method Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	0.01 0.50	0.02 1.00	mg/l as Mn mg/kg as Mn	2 2	
7	Potassium (K)	Waste Extraction , ICP-OES Method Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	0.50 25.00	1.00 50.00	mg/l as K mg/kg as K	2 2	
8	Silicon (Si)	Waste Extraction , ICP-OES Method Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	0.02 1.00	0.05 2.50	mg/l as Si mg/kg as Si	2 2	
9	Sodium (Na)	Waste Extraction , ICP-OES Method Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	0.50 25.0	1.00 50.0	mg/l as Na mg/kg as Na	2 1	
10	Strontium (Sr)	Waste Extraction , ICP-OES Method Digestion, ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	0.01 0.50	0.02 1.00	mg/l as Sr mg/kg as Sr	2 2	

การตรวจวิเคราะห์คุณภาพน้ำ – ภาคตะกอน (Water – Solid wastes Quality Analysis)
ตารางที่ 9 สรุปข้อกำหนดการเก็บตัวอย่างและความสามารถในการทดสอบตัวอย่างของห้องปฏิบัติการ ที่ไม่ได้ขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม
(ประเภทตัวอย่าง : ภาคตะกอน ตามประกาศเรื่องสิ่งปฏิกูลที่ไม่ใช่เ็นว)

ส่วนงาน : ส่วนงานเครื่องมือทดสอบ

Items	Parameter	Method	Reference Method / Analytical Technique	Container	sample size (ml)	MDL	LOQ	Unit	Decimal point	Remark
11	Tin (Sn)	Waste Extraction , ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	0.05	0.10	mg/l as Sn	2	
		Digestion,ICP-OES Method				2.50	5.00	mg/kg as Sn	2	
12	Titanium (Ti)	Waste Extraction , ICP-OES Method	US EPA SW 846 Method 3050B / ICP-OES	Plastic	500	0.01	0.02	mg/l as Ti	2	
		Digestion,ICP-OES Method				0.50	1.00	mg/kg as Ti	2	

เอกสารอ้างอิง

- 1 Standard Methods for the Examination of Water and Wastewater 23rd Edition, APHA, AWWA, WEF, 2017
- 2 United States Environmental Protection Agency, Acid Digestion of Sediments Sludge and Solis. SW-846 Method 3050C,3060A,3510C,3620C,6010C,7000B,7196A,7471B
- 3 Methods of Sewater Analysis. 1976
- 4 ประกาศกระทรวงอุตสาหกรรม พ.ศ. 2548 เรื่อง การกำจัดสิ่งปฏิกูลหรือวัสดุที่ไม่ใช่เ็นว. ราชกิจจานุเบกษา.25 มกราคม 2549 เล่มที่ 123 ตอนพิเศษ 114
- 5 คู่มือวิเคราะห์คุณภาพน้ำเชื้อ ตามทศวรรรมสิ่งแวดล้อมแห่งประเทศไทย พิมพ์ครั้งที่ 3, 2540
- 6 แพลงค์คอนเคิซ มหาวิทยาลัยเกษตรศาสตร์ พิมพ์ครั้งที่ 2, 2544
- 7 แพลงค์คอนเคิซ มหาวิทยาลัยเกษตรศาสตร์ พิมพ์ครั้งที่ 2, 2545