

## ภาคผนวก ง

---

เอกสารการสอบเทียบเครื่องมือตรวจวิเคราะห์

รายการเครื่องมือที่ใช้ในการวิเคราะห์ / ทดสอบ

Sample Name	Parameter	Equipment Name	ID No.	Calibrated Date	Next Cal	Freq. Calibrate (Months)
Ambient	Nitrogen Dioxide	NO <sub>2</sub> Analyzer	RYG_FS0551	5-Jan-23	5-Jul-23	6
Ambient	Nitrogen Dioxide	NO <sub>2</sub> Analyzer	RYG_FS0459	5-Jan-23	5-Jul-23	6
Ambient	Sulfur Dioxide	SO <sub>2</sub> Analyzer	RYG_FS0257	4-Jan-23	4-Jul-23	6
Ambient	Sulfur Dioxide	SO <sub>2</sub> Analyzer	RYG_FS0458	4-Jan-23	4-Jul-23	6
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_FS0610	17-Nov-22	17-May-24	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_FS0530	19-Jan-23	19-Jul-24	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_FS0544	15-Sep-21	15-Mar-23	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_FS0329	31-Jan-22	29-Jul-23	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_FS0531	19-Jan-23	19-Jul-24	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_FS0545	14-Sep-21	15-Mar-23	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_FS0089	19-Jan-23	19-Jul-24	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_FS0087	19-Jan-23	19-Jul-24	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_FS0414	10-Feb-23	10-Aug-24	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	BKK_FS0141	5-Jan-23	5-Jul-24	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_FS0411	10-Feb-23	10-Aug-24	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	NKH_FS0053	11-Jan-23	11-Jul-24	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_FS0087	19-Jan-23	19-Jul-24	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_FS0089	19-Jan-23	19-Jul-24	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	BKK_FS0888	30-May-22	28-Nov-23	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_FS0435	31-Jan-22	27-Jul-23	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	BKK_FS0917	1-Nov-21	2-May-23	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	NKH_FS0053	11-Jan-23	11-Jul-24	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_FS0608	17-Nov-22	17-May-24	18
Ambient	Wind Speed / Wind Direction	Wind Speed / Wind Direction	RYG_FS0609	18-Nov-22	18-May-24	18
Ambient	Volatile Organic Compounds	GC-MSD	RYG_EN0136	7-Jul-22	7-Jan-24	18
Ambient	Non-Methane Hydrocarbon	Total Hydrocarbon Analyzer	RYG_EN0038	25-Jan-23	25-Jan-24	12
Sea Water	pH at 25 °C	pH meter	RYG_EN0183	27-Feb-23	27-Feb-24	12
Sea Water	Turbidity	Chamber (Cold Room)	RYG_EN0184	25-Jan-23	25-Jul-24	18
Sea Water	Dissolved Oxygen	Chamber (Cold Room)	RYG_EN0184	25-Jan-23	25-Jul-24	18
Sea Water	BOD	DO meter with Sensor	RYG_EN0032	14-Feb-22	15-Aug-23	18
Sea Water	BOD	Incubator	RYG_EN0154	22-Apr-22	21-Oct-23	18
Sea Water	BOD	Burette	243007	21-Sep-18	21-Sep-23	60
Sea Water	Total Suspended Solids	Electronic Balance	RYG_EN0002	1-Mar-23	1-Mar-24	12
Sea Water	Total Suspended Solids	Hot Air Oven	RYG_EN0010	20-Oct-22	20-Apr-24	18
Sea Water	Total Dissolved Solids 180°C	Electronic Balance	RYG_EN0002	1-Mar-23	1-Mar-24	12
Sea Water	Total Dissolved Solids 180°C	Hot Air Oven	RYG_EN0010	20-Oct-22	20-Apr-24	18
Sea Water	Conductivity	Conductivity meter	RYG_EN0029	23-Feb-22	24-Aug-23	18
Sea Water	Salinity	Conductivity meter	RYG_EN0029	23-Feb-22	24-Aug-23	18
Sea Water	Oil & Grease	Electronic Balance	RYG_EN0002	1-Mar-23	1-Mar-24	12
Sea Water	Oil & Grease	Hot Air Oven	RYG_EN0006	20-Oct-22	20-Apr-24	18
Sea Water	Oil & Grease	Water Bath	RYG_EN0061	20-Oct-22	20-Apr-24	18
Sea Water	Temperature	pH meter	RYG_FS0392	22-Dec-22	22-Dec-23	12
Sea Water	Volatile Organic Compound	Gas Chromatography (MSD)	BKK_EN0059	21-Jun-22	21-Dec-23	18
Sea Water	Methanol	Gas Chromatography	BKK_EN0041	25-Nov-21	25-May-23	18
Sea Water	Lead	ICP-MS	BKK_EL0026	14-Jun-22	14-Dec-23	18
Sea Water	Lead	Hot Block	BKK_EL0054	7-Apr-22	7-Oct-23	18
Sea Water	Lead	Chamber (Cold Room)	BKK_EN0167	30-Jun-22	30-Dec-23	18
Sea Water	Cadmium	ICP-MS	BKK_EL0026	14-Jun-22	14-Dec-23	18
Sea Water	Cadmium	Hot Block	BKK_EL0054	7-Apr-22	7-Oct-23	18
Sea Water	Cadmium	Chamber (Cold Room)	BKK_EN0167	30-Jun-22	30-Dec-23	18
Sea Water	Mercury	DUO-CVAFS / CVAAS	BKK_EL0023	6-Jun-22	5-Jun-23	12
Sea Water	Total Coliform	Autoclave	BKK_ML0041	20-May-22	20-Nov-23	18
Sea Water	Total Coliform	Incubator	BKK_ML0010	21-Jan-22	22-Jul-23	18
Sea Water	Total Coliform	Hot Air Oven	BKK_ML0013	21-Nov-22	21-May-24	18
Sea Water	Fecal Coliform	Autoclave	BKK_ML0041	20-May-22	20-Nov-23	18
Sea Water	Fecal Coliform	Incubator	BKK_ML0010	21-Jan-22	22-Jul-23	18
Sea Water	Fecal Coliform	Hot Air Oven	BKK_ML0013	21-Nov-22	21-May-24	18
Sea Water	Fecal Coliform	Water Bath	BKK_ML0056	20-May-22	20-May-23	12
Sea Water	Total Petroleum Hydrocarbon	Electronic Top-Loading Balance	BKK_EN0002	8-Feb-23	8-Feb-24	12

รายการเครื่องมือที่ใช้ในการวิเคราะห์ / ทดสอบ

Sample Name	Parameter	Equipment Name	ID No.	Calibrated Date	Next Cal	Freq. Calibrate (Months)
Rayong Lab	pH at 25 °C	pH meter	RYG_EN0183	27-Feb-23	27-Feb-24	12
Rayong Lab	Sulfide	Chamber (Cold Room)	RYG_EN0184	25-Jan-23	25-Jul-24	18
Rayong Lab	BOD	DO meter with Sensor	RYG_EN0032	14-Feb-22	15-Aug-23	18
Rayong Lab	BOD	Incubator	RYG_EN0154	22-Apr-22	21-Oct-23	18
Rayong Lab	BOD	Burette	243007	21-Sep-18	21-Sep-23	60
Rayong Lab	COD	Spectrophotometer	RYG_EN0037	27-Sep-22	27-Mar-24	18
Rayong Lab	Total Suspended Solids	Electronic Balance	RYG_EN0002	1-Mar-23	1-Mar-24	12
Rayong Lab	Total Suspended Solids	Hot Air Oven	RYG_EN0010	20-Oct-22	20-Apr-24	18
Rayong Lab	Total Dissolved Solids 180°C	Electronic Balance	RYG_EN0002	1-Mar-23	1-Mar-24	12
Rayong Lab	Total Dissolved Solids 180°C	Hot Air Oven	RYG_EN0010	20-Oct-22	20-Apr-24	18
Rayong Lab	Oil & Grease	Electronic Balance	RYG_EN0002	1-Mar-23	1-Mar-24	12
Rayong Lab	Oil & Grease	Hot Air Oven	RYG_EN0006	20-Oct-22	20-Apr-24	18
Rayong Lab	Oil & Grease	Water Bath	RYG_EN0061	20-Oct-22	20-Apr-24	18
Rayong Lab	Temperature	pH meter	RYG_FS0574	3-Apr-23	3-Apr-24	12
Rayong Lab	Total Kjeldahl Nitrogen	Block Digestion Unit	RYG_EN0188	15-Mar-23	15-Mar-24	12
Rayong Lab	Total Kjeldahl Nitrogen	pH Meter	RYG_EN0152	22-Dec-22	22-Dec-23	12
Rayong Lab	Total Petroleum Hydrocarbon	Electronic Balance	RYG_EN0002	1-Mar-23	1-Mar-24	12
Rayong Lab	Total Petroleum Hydrocarbon	Hot Air Oven	RYG_EN0006	20-Oct-22	20-Apr-24	18
Rayong Lab	Total Petroleum Hydrocarbon	Water Bath	RYG_EN0061	20-Oct-22	20-Apr-24	18
Water Lab	Volatile Organic Compound	Gas Chromatography (MSD)	BKK_EN0059	21-Jun-22	21-Dec-23	18
Water Lab	Methanol	Gas Chromatography (MSD)	BKK_EN0059	21-Jun-22	21-Dec-23	18
Water Lab	Lead	ICP-MS	BKK_EL0026	14-Jun-22	14-Dec-23	18
Water Lab	Lead	Hot Block	BKK_EL0054	7-Apr-22	7-Oct-23	18
Water Lab	Lead	Chamber (Cold Room)	BKK_EN0167	30-Jun-22	30-Dec-23	18
Water Lab	Arsenic	ICP-MS	BKK_EL0026	14-Jun-22	14-Dec-23	18
Water Lab	Arsenic	Hot Block	BKK_EL0054	7-Apr-22	7-Oct-23	18
Water Lab	Arsenic	Chamber (Cold Room)	BKK_EN0167	30-Jun-22	30-Dec-23	18
Water Lab	Zinc	ICP-MS	BKK_EL0026	14-Jun-22	14-Dec-23	18
Water Lab	Zinc	Hot Block	BKK_EL0054	7-Apr-22	7-Oct-23	18
Water Lab	Zinc	Chamber (Cold Room)	BKK_EN0167	30-Jun-22	30-Dec-23	18
Water Lab	Mercury	DUO-CVAFS / CVAAS	BKK_EL0023	24-May-23	24-May-24	12
Water Lab	Total Coliform	Autoclave	BKK_ML0041	20-May-22	20-Nov-23	18
Water Lab	Total Coliform	Incubator	BKK_ML0010	21-Jan-22	22-Jul-23	18
Water Lab	Total Coliform	Hot Air Oven	BKK_ML0013	21-Nov-22	21-May-24	18




## MULTIPOINT CALIBRATION REPORT

Calibration Date 5-Jan-23  
Manufacturer HORIBA  
Serial No. U8AOEAGK  
Calibrator Manufacturer Teledyne API  
Serial No. 947  
Std. Gas Concentration (PPM) 55.88  
Cylinder Pressure (psi) 1800  
Certified Date 9-Feb-22

Equipment Name NOx Analyzer  
Model APNA-370  
Equipment ID RYG\_FS0551  
Model 700  
Cylinder No. GN0027222  
Certified By Airgas Inc.  
Expired Date 9-Feb-30

Point	CALIBRATION RESULTS						
	Ideal	Actual NO	Error NO	%Error NO	Actual NOx	Error NOx	%Error NOx
ZERO	0.00	0.10	0.10	0.10	0.10	0.10	0.10
1	100.00	99.00	-1.00	-1.00	101.00	1.00	1.00
2	200.00	198.50	-1.50	-0.75	201.30	1.30	0.65
3	300.00	298.40	-1.60	-0.53	301.50	1.50	0.50
4	400.00	398.20	-1.80	-0.45	402.00	2.00	0.50
AVERAGE (%)				-0.53			0.55



Calibrated By

( Mr.Jirawut Sakarn )  
Field Environmental Scientist (3)

Approved By

( Mr.Sarayuth Jittranont )  
Assistant General Manager






## MULTIPOINT CALIBRATION REPORT

Calibration Date 5-Jan-23  
Manufacturer HORIBA  
Serial No. NV0ER3YH  
Calibrator Manufacturer Teledyne API  
Serial No. 947  
Std. Gas Concentration (PPM) 55.88  
Cylinder Pressure (psi) 1800  
Certified Date 9-Feb-22

Equipment Name NOx Analyzer  
Model APNA-370  
Equipment ID RYG\_FS0459  
Model 700  
Cylinder No. GN0027222  
Certified By Airgas Inc.  
Expired Date 9-Feb-30

Point	CALIBRATION RESULTS						
	Ideal	Actual NO	Error NO	%Error NO	Actual NOx	Error NOx	%Error NOx
ZERO	0.00	0.05	0.05	0.05	0.10	0.10	0.10
1	100.00	99.50	-0.50	-0.50	101.80	1.80	1.80
2	200.00	198.70	-1.30	-0.65	199.70	-0.30	-0.15
3	300.00	301.10	1.10	0.37	301.50	1.50	0.50
4	400.00	401.30	1.30	0.33	398.90	-1.10	-0.28
AVERAGE (%)				-0.08			0.39



Calibrated By

( Mr.Jirawut Sakarn )  
Field Environmental Scientist (3)

Approved By


( Mr.Sarayuth Jittranont )  
Assistant General Manager



## MULTIPOINT CALIBRATION REPORT

Calibration Date	4-Jan-23	Equipment Name	SO2 Analyzer
Manufacturer	HORIBA	Model	APSA-370
Serial No.	24PH0KNA	Equipment ID	RYG_FS0257
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	56.3	Cylinder No.	GN0027222
Cylinder Pressure (psi)	1800	Certified By	Airgas Inc.
Certified Date	9-Feb-22	Expired Date	9-Feb-30

Point	CALIBRATION RESULTS			
	Ideal	Autual	Error	%Error
ZERO	0.00	0.10	0.10	0.10
1	100.00	99.40	-0.60	-0.60
2	200.00	197.30	-2.70	-1.35
3	300.00	299.50	-0.50	-0.17
4	400.00	397.00	-3.00	-0.75
AVERAGE (%)				-0.55



Calibrated By

( Mr.Jirawut Sakarn )  
Field Environmental Scientist (3)

Approved By


( Mr.Sarayuth Jittranont )  
Assistant General Manager



## MULTIPOINT CALIBRATION REPORT

Calibration Date	4-Jan-23	Equipment Name	SO2 Analyzer
Manufacturer	HORIBA	Model	APSA-370
Serial No.	PAUY0T7A	Equipment ID	RYG_FS0458
Calibrator Manufacturer	Teledyne API	Model	700
Serial No.	947		
Std. Gas Concentration (PPM)	56.3	Cylinder No.	GN0027222
Cylinder Pressure (psi)	1800	Certified By	Airgas Inc.
Certified Date	9-Feb-22	Expired Date	9-Feb-30

Point	CALIBRATION RESULTS			
	Ideal	Autual	Error	%Error
ZERO	0.00	0.10	0.10	0.10
1	100.00	98.90	-1.10	-1.10
2	200.00	201.10	1.10	0.55
3	300.00	302.30	2.30	0.77
4	400.00	398.60	-1.40	-0.35
AVERAGE (%)				-0.01



Calibrated By

( Mr.Jirawut Sakarn )  
Field Environmental Scientist (3)

Approved By

( Mr.Sarayuth Jittranont )  
Assistant General Manager

Certificate Number

CL-003-65

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

**MEASUREMENT ITEM** : Cup anemometer  
**MANUFACTURER** : Novalynx  
**MODEL/TYPE** : Sensor: WS-02F  
Data logger: 110-WS-25DL-D  
**SERIAL NUMBER** : Sensor: WSD-013  
Data logger: A5911  
**ID NUMBER** : RYG\_FS0610  
**CONDITION AS-RECEIVED** : New item  
**CUSTOMER** : ALS laboratory group (Thailand) co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,  
Khet Suan Luang, Bangkok 10250 Thailand.

**RECEIVED DATE** : 09 Nov 2022  
**MEASUREMENT DATE** : 17 Nov 2022  
**ISSUE DATE** : 23 Nov 2022

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature :  $23.0 \pm 3.0$  °C  
Relative Humidity :  $55.0 \pm 15.0$  %RH  
Atmospheric Pressure :  $1010 \pm 10$  hPa

**PLACE OF CALIBRATION** : Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

**CALIBRATION CONDITIONS** : Wind tunnel cross-section area<sup>1</sup> 900 cm<sup>2</sup>  
Win direction frontal area<sup>2</sup> 100 cm<sup>2</sup>  
Diameter of mounting pipe<sup>3</sup> - mm  
Blockage ratio of test object<sup>4</sup> 0.111 [-]

**Preconditioning** : 24 hours at ambient conditions.  
**Measurement Condition** : The average values during measurement are (24.1) °C, (48.8) %RH and (1015.4) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol

### Remark:

- <sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>



Approved signatory:

*[Signature]*  
Mr. Parinya Booncharoen  
Calibration Department Manager

### Calibration procedure:

The cup anemometer was calibrated against Standard air velocity transducer model: 8455-12 and pitot tube with precision differential pressure meter model: DPM2500 in an close test-section of Eiffel-type wind tunnel with 900 cm<sup>2</sup> cross test section area. The WI-CL 007 based on IEC 61400-12-1, Wind energy generation systems – Part 12-1: Power performance measurements of electricity producing wind turbines, March 2017 was used as a calibration guideline.

### Traceability:

This certificate provides a traceability of The measurement to recognized the national standards, and to realization of the international system of units (SI) through the NIMT (National Metrology Institute of Thailand) via Certificate number: MW-0052-21 and MW-0066-22

### Uncertainty of Measurement:

The reported uncertainty of measurement is based on the standard uncertainty multiplied by a coverage factor  $k=2$ , Which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty has been determined in accordance with the GUM 'Evaluation of measurement data - Guide to the expression of uncertainty in measurement'

**MEASUREMENT RESULTS <sup>5</sup>**

The cup anemometer, Unit Under Calibration (UUC) was exercise at 10 m/s for 5 minutes prior to calibration being performed. The standard air velocity 0.5 m/s to 5 m/s was calculated by a standard air velocity transducer and above 5 m/s to 30 m/s was calculated by a pitot tube with precision differential pressure meter which was installed 40 mm and 300 mm respectively away from wind tunnel nozzle, UUC was installed at center of the test section. The calibration was carried out under both rising and falling air velocity in the range of 1 m/s to 16 m/s at calibration interval of 1 m/s. The results of calibration and associated measurement uncertainties are reported in the table below.

$V_{std}^6$ (m/s)	Temp. wind tunnel (°C)	Temp. room (°C)	$V_{UUC}^7$ (m/s)	Error (m/s)	$U (k=2)$ (m/s)
0.994	24.08	24.05	0.8	-0.2	0.17
2.036	24.10	24.05	1.8	-0.2	0.17
3.044	24.00	24.05	2.9	-0.2	0.17
4.217	24.10	24.05	3.9	-0.4	0.19
5.02	23.92	24.05	4.9	-0.1	0.17
6.00	24.24	24.05	5.9	-0.1	0.18
7.08	23.88	24.05	6.8	-0.2	0.20
8.20	24.12	24.05	7.9	-0.3	0.20
9.13	23.74	24.05	8.8	-0.3	0.19
10.11	24.04	24.05	9.8	-0.3	0.19
11.17	23.80	24.05	10.9	-0.3	0.20
12.15	23.98	24.05	11.8	-0.3	0.21
13.20	23.78	24.05	12.9	-0.3	0.26
14.25	23.80	24.05	14.0	-0.2	0.26
15.25	23.80	24.05	14.9	-0.3	0.23
16.30	23.80	24.05	16.0	-0.3	0.23

**Remark:**

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Velocity of standard

<sup>7</sup> Velocity of Unit Under Calibration

**PHOTO OF CALIBRATION SET-UP**

Calibration set up of the cup anemometer calibration in the wind tunnel of Jiranatee Associates Co., Ltd. The cup anemometer shown may differ from the calibrated one. Remark: The proportion of the set- up is not true to scale due to imaging geometry.

\*\*\*End of Certificate of Calibration\*\*\*





Jiranatee Associates Co., Ltd  
63/14-15, 67/35-36  
Petchkasem 7, 7/1, Rd. Watthapra, Bangkokyai,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Air speed measurement laboratory  
Calibration services department.

Certificate Number

CL-003-65

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

<b>MEASUREMENT ITEM</b>	: Cup anemometer
<b>MANUFACTURER</b>	: Novalynx
<b>MODEL/TYPE</b>	: Sensor: WS-02F Data logger: 110-WS-25DL-D
<b>SERIAL NUMBER</b>	: Sensor: WSD-013 Data logger: A5911
<b>ID NUMBER</b>	: RYG_FS0610
<b>CONDITION AS-RECEIVED</b>	: New item
<b>CUSTOMER</b>	: ALS laboratory group (Thailand) co., ltd. 104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand.

<b>RECEIVED DATE</b>	: 09 Nov 2022
<b>MEASUREMENT DATE</b>	: 17 Nov 2022
<b>ISSUE DATE</b>	: 23 Nov 2022

**ENVIRONMENTAL CONDITIONS:**

Ambient condition in the laboratory are as follow:

Temperature	: 23.0 ± 3.0	°C
Relative Humidity	: 55.0 ± 15.0	%RH
Atmospheric Pressure	: 1010 ± 10	hPa

<b>PLACE OF CALIBRATION</b>	: Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.
-----------------------------	---

<b>CALIBRATION CONDITIONS</b>	: Wind tunnel cross-section area <sup>1</sup>	900	cm <sup>2</sup>
	Win direction frontal area <sup>2</sup>	100	cm <sup>2</sup>
	Diameter of mounting pipe <sup>3</sup>	-	mm
	Blockage ratio of test object <sup>4</sup>	0.111	[-]

<b>Preconditioning</b>	: 24 hours at ambient conditions.
<b>Measurement Condition</b>	: The average values during measurement are (24.1) °C, (48.8) %RH and (1015.4) hPa.

**TABULATION OF RESULTS:**

The table on next page give the measured values.

**Calibrated by:**

☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol

**Remark:**

<sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>



Approved signatory:

Mr. Parinya Booncharoen  
Calibration Department Manager

**MEASUREMENT RESULTS<sup>5</sup>**

The cup anemometer, Unit Under Calibration (UUC) was exercised at 10 m/s for 5 minutes prior to calibration being performed. The standard air velocity 0.5 m/s to 5 m/s was calculated by a standard air velocity transducer and above 5 m/s to 30 m/s was calculated by a pitot tube with precision differential pressure meter which was installed 40 mm and 300 mm respectively away from wind tunnel nozzle, UUC was installed at center of the test section. The calibration was carried out under both rising and falling air velocity in the range of 1 m/s to 16 m/s at calibration interval of 1 m/s. The results of calibration and associated measurement uncertainties are reported in the table below.


$V_{std}^6$ (m/s)	Temp. wind tunnel (°C)	Temp. room (°C)	$V_{uuc}^7$ (m/s)	Error (m/s)	$U (k=2)$ (m/s)
0.994	24.08	24.05	0.8	-0.2	0.17
2.036	24.10	24.05	1.8	-0.2	0.17
3.044	24.00	24.05	2.9	-0.2	0.17
4.217	24.10	24.05	3.9	-0.4	0.19
5.02	23.92	24.05	4.9	-0.1	0.17
6.00	24.24	24.05	5.9	-0.1	0.18
7.08	23.88	24.05	6.8	-0.2	0.20
8.20	24.12	24.05	7.9	-0.3	0.20
9.13	23.74	24.05	8.8	-0.3	0.19
10.11	24.04	24.05	9.8	-0.3	0.19
11.17	23.80	24.05	10.9	-0.3	0.20
12.15	23.98	24.05	11.8	-0.3	0.21
13.20	23.78	24.05	12.9	-0.3	0.26
14.25	23.80	24.05	14.0	-0.2	0.26
15.25	23.80	24.05	14.9	-0.3	0.23
16.30	23.80	24.05	16.0	-0.3	0.23

**Remark:**

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Velocity of standard

<sup>7</sup> Velocity of Unit Under Calibration

**PHOTO OF CALIBRATION SET-UP**

Calibration set up of the cup anemometer calibration in the wind tunnel of Jiranatee Associates Co., Ltd. The cup anemometer shown may differ from the calibrated one. Remark: The proportion of the set- up is not true to scale due to imaging geometry.

\*\*\*End of Certificate of Calibration\*\*\*







JIRANATE ASSOCIATES CO., LTD.

Jiranatee Associates Co., Ltd.  
63/14 15, 67/35-36  
Petchkasem 7.7/1, Rd. Watthapra, Bangkokyai,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Pressure measurement laboratory  
Calibration services department.



NSC - TISI - TIS 17025  
CALIBRATION 0367

## CERTIFICATE OF CALIBRATION

Certificate No. : CL-016-65

Page 1 of 2 Pages

MEASUREMENT ITEM : Digital barometer  
MANUFACTURER : Novalynx  
MODEL/TYPE : 110-WS-25BP  
SERIAL NUMBER : A5911  
ID NUMBER : RYG\_FS0610  
CONDITION AS-RECEIVED : New item  
CUSTOMER : ALS laboratory group (Thailand) co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd,  
Khwaeng Suan Luang, Khet Suan Luang,  
Bangkok 10250 Thailand.  
RECEIVED DATE : 09 Nov 2022  
MEASUREMENT DATE : 22 Nov 2022  
ISSUE DATE : 23 Nov 2022

### Calibration procedure:

The pressure calibration was done by In-house calibration method as WI-CL-003 according to comparison method with Digital pressure calibrator based on DKD-R 6-1

### Traceability:

The measurement results are traceable to the international system of units (SI) through MENSOR which complies with the requirements of ISO/IEC17025:2017, ANSI/NCSL Z540-1 via Certificate number: 201479

The reported uncertainty is based on a standard uncertainty multiplied by a coverage  $k=2$ , providing a level of confidence of approximately 95%.

### CONDITION OF THIS RESULT OF CALIBRATION:

#### 1. Reference Standard Instrument:

Instrument	Model	Serial No.	Certificate No.	Due Date
Absolute Pressure Transducer	CPG2500	410018L1	201479	13 Sep 2022

#### 1. Calibration effort for calibration sequence A

2. The UUC\* was installed in vertical orientation above reference standard instrument and center of UUC\* was used as the reference level.

#### 3. Calibration conditions:

4. Condition : ☒ Normal ☐ Abnormal  
Pressure transmitting medium : Air  
 $\rho_{t1}$  (20°C, 1 bar) : 1.19 kg/m<sup>3</sup>  
 $H_{amb}$  : (55±15) %  
 $t_{amb}$  : (23±3) °C  
 $p_{amb}$  : (1010±10) mbar

5. The certificate is valid only to the item calibrated on date and place of calibration

### Calibrated by:

☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



Approved signatory:

Mr. Parinya Booncharoen  
Calibration Department Manager





JIRANATEE ASSOCIATES CO., LTD.

Jiranatee Associates Co., Ltd.  
63/14-15, 67/35-36  
Petchkasem 7, 7/1, Rd. Watthapra, Bangkokyai,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Pressure measurement laboratory  
Calibration services department.



NSC - TISI - TIS 17025  
CALIBRATION 0367

## CERTIFICATE OF CALIBRATION

Certificate No. : CL-016-65

Page 2 of 2 Pages

MEASUREMENT RESULTS : ☒ Without adjustment ☐ With adjustment

CALIBRATION IN THE RANGE OF : 950 – 1050 mbar

The results of calibration and associated measurement uncertainties are reported in the table below.

STD (mbar)	UUC* (mbar)	Error (mbar)	Uncertainty (k=2) (mbar)
950.00	950.0	0.0	0.37
970.00	969.8	-0.2	0.49
990.00	989.6	-0.4	0.62
1010.00	1009.5	-0.4	0.64
1030.00	1029.1	-0.9	1.1
1050.00	1049.0	-1.0	1.2

Note: UUC\* Unit Under Calibration

: To convert the result in report unit to Pa should be multiply by 100

\*End of certificate\*



## CERTIFICATE OF CALIBRATION

Certificate No.: CL-158-65  
 Page 1 of 2

Equipment Name: Data Logger with Temperature  
 Sensor

Manufacturer: Novalynx  
 Model: 110-WS-25DL-D  
 Serial No.: A5911  
 ID No.: RYG\_FS0610

### Customer

Name: ALS laboratory group (Thailand) Co.,Ltd.  
 Address: 104 Phatthanakan 40, Phatthanakan  
 Rd.,Khwaeng Suan Luang, Khet Suan Luang, Bangkok  
 10250 Thailand.

Received date: 09 Nov 2022  
 Calibration date: 18 Nov2022  
 Issue date: 23 Nov 2022

### Reference Used During Calibration

- 1.Standard Temperature Probe Model: STS-100 A500,  
 Serial No.: 667682-09, Due date: 23 Mar 2023
- 2.Digital Temperature Indicator Model: DTI-1000-A MK  
 II, Serial No.: 671407-00591 Due date: 22 July 2023

### Calibration Condition

Temperature: (23±3) °C  
 Relative Humidity: (55±15)%

### Calibration Procedure

The temperature calibration was done by In-House  
 calibration method as WI-CL-001 according to  
 comparison method with standard digital temperature  
 indicator and standard temperature probe. The  
 temperature scale use was based on ITS-90.

### Traceability

The measurement results are traceable to the  
 international system of units (SI) through National  
 Institute of Metrology Thailand (NIMT) Certificate  
 number: TT-0034-22, Certificate number: ER-0092-  
 22

### Calibrated by

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



### Approved Signatory:

*25/01/23*  
 Mr. Parinya Booncharoen  
 Calibration Department Manager

Certificate No.: CL-158 65  
Page 2 of 2

**Result of Calibration:-** ☒ Without Adjustment ☐ With Adjustment

**Calibration Range:** 20-40 °C

**Function:**

This equipment was connected with temperature sensor Model: HMP60 S/N: U3911245.

Dimension : Diameter 12 mm. Length 80 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
60	20.01	19.9	-0.1	0.30
60	25.02	24.9	-0.2	0.30
60	29.99	29.8	-0.2	0.30
60	35.00	34.6	-0.4	0.30
60	40.00	39.4	-0.6	0.30

UUC\*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

✱ End of Certificate ✱







63/14-15,67/35-36, Soi Petchkasem7,7/1, Petchkasem Rd,  
Watthapra, Bangkokyai,Bangkok 10600 Thailand.  
Tel.: (66) 02-8680812#13 Fax.: (66) 02-8680860 www.jiranalee.com

## CERTIFICATE OF CALIBRATION

Calibration No. : RH-03112022

Page 1 of 1 Pages

Measurement Item : Relative humidity with data logger  
Manufacturer : Novalynx  
Model/Type : 110-WS-25DL-D  
Serial Number : A5911  
ID No. : RYG\_FSO610  
Customer : ALS laboratory group (Thailand) Co., Ltd.  
: 104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Suan Luang, Khut Suan Luang, Bangkok  
10250 Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(25 \pm 3)^{\circ}\text{C}$ , and relative humidity of  $(50 \pm 15)\%$ .

### Measurement Method:

Unit Under Calibration (UUC) was calibrated by comparison method with standard thermo hygrometer in the humidity generator chamber to determine the errors.

### Traceability:

This instrument was calibrated using standard equipment whose accuracy is traceability through National Institute of Standards and Technology to the international system of units (SI) via MCS Calibration, Inc. Certificate number: 20314-101. Due date: Mar 14,2023.

Measurement Date : Nov 18, 2022

Issued Date : Nov 23, 2022

### Measurement Results:

This equipment was connected with Indoor air quality probe and Displayed (UR) on display. Model: HMP60, Serial number: U3911245

Calibration was performed in the range of 20%RH to 80%RH

The results of calibration are reported in table below.

Determined (%RH)	Standard (Reading) (%RH)	UUC (Reading) (%RH)	Error (%RH)	Uncertainty $\pm$ (%RH)
20	20.07	18.0	-2.0	0.56
50	50.29	48.2	-2.1	0.62
80	80.24	78.4	-1.8	0.55

Performed by

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



Approved Signatory: .....

Mr. Parinya Booncharoen.  
Calibration Department Manager

## CALIBRATION REPORT

Calibration Number. : RG-01112022

Page 1 of 2 Pages

Measurement Item : Rain gauge with data logger

Manufacturer : Data logger: Novalynx.  
: Rain gauge: Novalynx.

Model/Type : Data logger: 110-WS-25DL D  
: Rain gauge: 110 WS 25RG

Serial Number : Data logger: A5911  
: Rain gauge: RG-007

ID NO : RYG\_FSO610

Customer : ALS laboratory group (Thailand) co., ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang, Khet Suan Luang,  
Bangkok 10250, Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(25 \pm 3)^{\circ}\text{C}$ , and relative humidity of  $(50 \pm 15)\%$ .

### Measurement Method:

The Rain gauge, Unit Under Calibration (UUC) was calibrated by Precision reference bottle with flow adjuster at low rate 0.6 mm per minute or 1 tipping every 20 seconds. The tipping number was determined by procedures below.

1. Obtain rain gauge inlet area:  
Rain gauge precise diameter in cm = Diameter/2 = R (radius)  
Rain gauge area=  $R^2 \times 3.14$  (UUC diameter=20.3 cm, UUC radius=10.15 cm)  
Rain gauge area=  $323.6 \text{ cm}^2$ .
2. Obtain theoretical correct rain gauge answer (number of tipplings) using  $323.6 \text{ cm}^2$  inlet area and 0.5 L of rain.
  - a)  $10,000 \text{ cm}^3 / 323.6 \text{ cm}^2$  inlet area = 30.90 (rain gauge area = 1/30.90 of square meter)
  - b)  $30.90 \times 0.5 \text{ L volume} = 15.45 \text{ mm}$  (mm of rain over  $1 \text{ m}^2$  surface) 500 ml of rain volume on the rain gauge area = 15.45 mm of rain.
  - c) Number of tipping=  $15.45 / 0.25 \text{ mm} = 62$  tipplings.

*Note: Rain gauge is fully cleaned and leveling prior the calibration performed.*

Measurement Date : Nov 18, 2022

Issued Date : Nov 23, 2022

Performed by

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphot



Approved Signatory:.....

Mr. Parinya Booncharoen,  
Calibration Department Manager



63/14-15,67/35-36, Soi Petchkasem7,7/1, Petchkasem Rd,  
Wathapra, Bangkokyai,Bangkok 10600 Thailand.  
Tel.: (66) 02-8680812#13 Fax.: (66) 02-8680860 www.jiranalee.com

Continuation of Calibration of Calibration Number

Calibration Number: RG-01112022

Page 2 of 2 Pages

Result of Calibration: ☒ Without Adjustment ☐ With Adjustment.

The results of calibration are reported in table below.

Quantity of H <sub>2</sub> O (ml)	Determined Tipping	Tipping count	Acceptable Tipping count
500	62	60	60 - 64
500	62	61	60 - 64
500	62	60	60 - 64
500	62	60	60 - 64
500	62	61	60 - 64

*Remark: The procedure is made to verify the correct reading of the Unit under Calibration rain gauge when a precise volume of water falls into its cone. We suggest that the number of tipping should be within  $\pm 2\%$  different from the 62 tipping (correct range: 60-64 tipping) it means that the rain gauge meets the manufacturer acceptable limit.*

\*\*\*End of calibration report\*\*\*







JIRANATEE ASSOCIATES CO., LTD.

Jiranatee Associates Co., Ltd.  
63/14-15, 67/35-36  
Petchkasem 7, 7/1, Rd. Watthapra, Bangkokyai,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Air speed measurement laboratory  
Calibration services department.

REVIEW BY	Manakorn P.
APPROVED BY	[Signature]
NEXT CAL. DATE	19/1/24

Certificate Number

CL-012-66

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

**MEASUREMENT ITEM** : Wind Direction Sensor  
**MANUFACTURER** : Novalynx  
**MODEL/TYPE** : Sensor: WS-02F  
Data logger: 110-WS-25DL-D  
**SERIAL NUMBER** : Sensor: WSD-011  
Data logger: A5660  
**ID NUMBER** : RYG\_FS0530  
**CONDITION AS-RECEIVED** : Used item  
**CUSTOMER** : ALS laboratory group (Thailand) Co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,  
Khet Suan Luang, Bangkok 10250 Thailand.

**RECEIVED DATE** : 16 Jan 2023  
**MEASUREMENT DATE** : 19 Jan 2023  
**ISSUE DATE** : 20 Jan 2023

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature	: 23.0 ± 3.0	°C
Relative Humidity	: 55.0 ± 15.0	%RH
Atmospheric Pressure	: 1010 ± 10	hPa

**PLACE OF CALIBRATION** : Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

<b>CALIBRATION CONDITION</b>	: Wind tunnel cross-section area <sup>1</sup>	900	cm <sup>2</sup>
	Win direction frontal area <sup>2</sup>	129	cm <sup>2</sup>
	Diameter of mounting pipe <sup>3</sup>	-	mm
	Blockage ratio of test object <sup>4</sup>	0.143	[-]

**Preconditioning** : 24 hours at ambient conditions.

**Measurement Condition** : The average values during measurement are (23.7)°C, (44.2) %RH and (1015.2) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



### Approved signatory:

Mr. Parinya Booncharoen  
Calibration Department Manager

### Remark:

<sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>

MEASUREMENT RESULTS<sup>5</sup>

The wind direction sensor was calibrated against standard rotary encoder by comparison method. During calibration, the measurement was carried out at 45° intervals in clockwise and counterclockwise directions after offset adjustment has been made. The flow speed of wind tunnel (usually 5 m/s) is kept constant while the sensor is rotated around its vertical axis. The results of calibration and associated measurement uncertainties are reported in the table below.

Air speed m/s	$D^{\circ}_{std}$ Degree (°)	$D^{\circ}_{uuc}$ Degree (°)	Error Degree (°)	$U (k=2)$ Degree (°)
5.01	0.000	0	0	0.58
	45.000	42	-3	0.74
	90.000	88	-2	0.74
	135.000	133	-2	0.68
	180.000	179	-1	0.74
	225.000	226	1	0.74
	270.000	270	0	0.74
	315.000	316	1	0.74

## Remark:

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Direction of standard

<sup>7</sup> Direction of Unit Under Calibration

\*\*\*End of Certificate of Calibration\*\*\*







JIRANATEE ASSOCIATES CO.,LTD.

Jiranatee Associates Co.,Ltd  
63/14-15, 67/35-36  
Petchkasem 7,7/1, Rd.Watthapra, Bangkokyai,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Air speed measurement laboratory  
Calibration services department.

Certificate Number

CL-012-66

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

**MEASUREMENT ITEM** : Cup anemometer  
**MANUFACTURER** : Novalynx  
**MODEL/TYPE** : Sensor: WS-02F  
Data logger: 110-WS-25DL-D  
**SERIAL NUMBER** : Sensor: WSD-011  
Data logger: A5660  
**ID NUMBER** : RYG\_FS0530  
**CONDITION AS-RECEIVED** : Used item  
**CUSTOMER** : ALS laboratory group (Thailand) co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,  
Khet Suan Luang, Bangkok 10250 Thailand.

**RECEIVED DATE** : 16 Jan 2023  
**MEASUREMENT DATE** : 18 Jan 2023  
**ISSUE DATE** : 20 Jan 2023

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature :  $23.0 \pm 3.0$  °C  
Relative Humidity :  $55.0 \pm 15.0$  %RH  
Atmospheric Pressure :  $1010 \pm 10$  hPa

**PLACE OF CALIBRATION** : Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

**CALIBRATION CONDITIONS** : Wind tunnel cross-section area<sup>1</sup> 900 cm<sup>2</sup>  
Win direction frontal area<sup>2</sup> 100 cm<sup>2</sup>  
Diameter of mounting pipe<sup>3</sup> - mm  
Blockage ratio of test object<sup>4</sup> 0.111 [-]

**Preconditioning** : 24 hours at ambient conditions.  
**Measurement Condition** : The average values during measurement are (23.7) °C, (50.2) %RH and (1017.1) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



Approved signatory: .....

Mr. Parinya Booncharoen  
Calibration Department Manager

### Remark:

- <sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>

MEASUREMENT RESULTS<sup>5</sup>

The cup anemometer, Unit Under Calibration (UUC) was exercised at 10 m/s for 5 minutes prior to calibration being performed. The standard air velocity 0.5 m/s to 5 m/s was calculated by a standard air velocity transducer and above 5 m/s to 30 m/s was calculated by a pitot tube with precision differential pressure meter which was installed 40 mm and 300 mm respectively away from wind tunnel nozzle, UUC was installed at center of the test section. The calibration was carried out under both rising and falling air velocity in the range of 1 m/s to 16 m/s at calibration interval of 1 m/s. The results of calibration and associated measurement uncertainties are reported in the table below.

$V_{std}$ <sup>6</sup> (m/s)	Temp. wind tunnel (°C)	Temp. room (°C)	$V_{UUC}$ <sup>7</sup> (m/s)	Error (m/s)	$U$ (k=2) (m/s)
0.979	23.56	23.70	0.8	-0.2	0.16
2.025	23.80	23.70	1.8	-0.2	0.16
3.046	23.50	23.70	2.8	-0.2	0.20
4.120	23.64	23.70	3.9	-0.3	0.20
5.01	23.44	23.70	4.8	-0.2	0.18
5.98	23.60	23.70	5.8	-0.2	0.18
7.05	23.28	23.70	6.9	-0.1	0.19
8.17	23.60	23.70	8.0	-0.2	0.19
9.09	23.20	23.70	9.0	0.0	0.22
10.09	23.52	23.70	9.9	-0.2	0.20
11.13	23.20	23.70	10.9	-0.2	0.21
12.13	23.50	23.70	11.9	-0.2	0.21
13.19	23.20	23.70	13.0	-0.2	0.22
14.25	23.46	23.70	14.3	0.0	0.24
15.22	23.20	23.70	15.1	-0.1	0.34
16.31	23.30	23.70	16.1	-0.2	0.29


## Remark:

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Velocity of standard

<sup>7</sup> Velocity of Unit Under Calibration

## PHOTO OF CALIBRATION SET-UP



Calibration set-up of the cup anemometer calibration in the wind tunnel of Jiranatee Associates Co., Ltd. The cup anemometer shown may differ from the calibrated one. Remark: The proportion of the set-up is not true to scale due to imaging geometry.

\*\*\*End of Certificate of Calibration\*\*\*



## CERTIFICATE OF CALIBRATION

Certificate No.: CL-005-66

Page 1 of 2

Equipment Name: Data Logger with Temperature  
Sensor

Manufacturer: Novalynx

Model: 110-WS-25DL-D

Serial No.: A5660

ID No.: RYG\_FS0530

### Customer

Name: ALS laboratory group (Thailand) Co., Ltd.

Address: 104 Phatthanakan 40, Phatthanakan Rd.,  
Khwaeng Suan Luang, Khet Suan Luang, Bangkok  
10250 Thailand.

Received date: 16 Jan 2023

Calibration date: 18 Jan 2023

Issue date: 20 Jan 2023

### Reference Used During Calibration

1. Standard Temperature Probe Model: STS-100 A500,  
Serial No.: 667682-09, Due date: 23 Mar 2023

2. Digital Temperature Indicator Model: DTI-1000-A MK  
II, Serial No.: 671407-00591 Due date: 22 July 2023

### Calibration Condition

Temperature:  $(23 \pm 3)^{\circ}\text{C}$

Relative Humidity:  $(55 \pm 15)\%$

### Calibration Procedure

The temperature calibration was done by In-House  
calibration method as WI-CL-001 according to  
comparison method with standard digital temperature  
indicator and standard temperature probe. The  
temperature scale use was based on ITS-90.

### Traceability

The measurement results are traceable to the  
international system of units (SI) through National  
Institute of Metrology Thailand (NIMT) Certificate  
number: TT-0034-22, Certificate number: ER-0092-  
22

### Calibrated by

☐ Mr. Sorawit Thachalad

☒ Miss Jitraporn Lertsomphol



### Approved Signatory:

Mr. Parinya Booncharoen

Calibration Department Manager

**Result of Calibration:-** ☒ Without Adjustment ☐ With Adjustment

**Calibration Range:** 20-40 °C

**Function:**

This equipment was connected with temperature sensor Model: HMP60 S/N: S4620631.

Dimension : Diameter 12 mm. Length 80 mm.

<u>Immersion Depth (mm)</u>	<u>Standard Reading (°C)</u>	<u>UUC Reading (°C)</u>	<u>Error (°C)</u>	<u>Uncertainty (°C)</u>
60	20.066	19.8	-0.3	0.099
60	25.058	24.6	-0.5	0.14
60	30.052	29.5	-0.6	0.099
60	35.047	34.5	-0.5	0.099
60	40.038	39.4	-0.6	0.099

UUC\*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

**\* End of Certificate \***



## CERTIFICATE OF CALIBRATION

Calibration No. : RH-05012023

Page 1 of 1 Pages

Measurement Item : Relative humidity with data logger  
Manufacturer : Novalynx  
Model/Type : 110-WS-25DL-D  
Serial Number : A5660  
ID No. : RYG\_FS0530  
Customer : ALS laboratory group (Thailand) Co., Ltd.  
: 104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Suan Luang, Khet Suan Luang, Bangkok  
10250 Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(25 \pm 3)^{\circ}\text{C}$ , and relative humidity of  $(50 \pm 15)\%$ .

### Measurement Method:

Unit Under Calibration (UUC) was calibrated by comparison method with standard thermo hygrometer in the humidity generator chamber to determine the errors.

### Traceability:

This instrument was calibrated using standard equipment whose accuracy is traceability through National Institute of Standards and Technology to the international system of units (SI) via MCS Calibration, Inc. Certificate number: 20314-101. Due date: Mar 14, 2023.

Measurement Date : Jan 18, 2023

Issued Date : Jan 20, 2023

### Measurement Results:

This equipment was connected with Indoor air quality probe and Displayed (UR) on display. Model: HMP60, Serial number: S4620631.

Calibration was performed in the range of 20%RH to 80%RH

The results of calibration are reported in table below.

Determined (%RH)	Standard (Reading) (%RH)	UUC (Reading) (%RH)	Error (%RH)	Uncertainty $\pm$ (%RH)
20	20.03	17.8	-2.2	0.58
50	50.28	48.6	-1.7	0.57
80	80.29	79.8	-0.5	0.58

### Performed by

- ☐ Mr. Sorawit Thachalad  
☒ Miss Jitraporn Lertsomphol



Approved Signatory: 

Mr. Parinya Booncharoen.  
Calibration Department Manager



## CERTIFICATE OF CALIBRATION

Certificate No: WS-01092021

Page 1 of 2 pages

Measurement Item : Cup anemometer with data logger.

Manufacturer : Data logger: Novalynx.  
: Cup anemometer: Novalynx.

Model/Type : Data logger: 110-WS-25DL-D  
: Cup anemometer: WS-02F

Serial Number : Data logger: A5662  
: Cup anemometer: WSD-015

ID No : Data logger: -. 5167.55m2.0/1  
: Cup anemometer: -.

Customer : ALS laboratory group (Thailand) co., ltd.  
: 104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand.

Test Conditions : Wind tunnel cross test section area 900 cm<sup>2</sup>  
: Anemometer frontal area 100 cm<sup>2</sup>  
: Diameter of mounting pipe - mm  
: Blockage ratio of test object 0.111 [-]

Test Conditions : Air temperature 23.0 ±0.8 °C  
: Air pressure 1012.8 ±0.4 hPa  
: Relative air humidity 50.4 ±3.5 %RH

Calibration Procedure : Calibration was carried out base on;  
IEC 61400-12-1 ED.1: 2005-Power Performance Measurements of Electricity Producing Wind Turbines;  
MEASNET Anemometer Calibration Procedure – Version 2: 2009;

Traceability : This calibration documents the traceable to national standard, Which realize the unit of measurements according to the international system of units (SI) through National Institute of Metrology Thailand (NIMT).

Measurement Date : Sep 14, 2021.

Issued Date : Sep 15, 2021.



**Calibrated by**

- ☒ Mr. Sorawit Thachelad  
☐ Miss Orathai Wiwatwittaya



Approved Signatory: *Lempr*

Mr. Parinya Booncharoen  
Technical Support  
and Calibration Manager

Continuation of Certificate of Calibration Number

Certificate No: WS-01092021

Page 2 of 2 Pages

Result of calibration: ☒ Without adjustment ☐ With adjustment

Calibration in the range of 1 – 16 m/s at a calibration interval of 1 m/s.

The results of calibration and associated measurement uncertainties are reported in the table below.

V <sub>STD</sub> Reading m/s	V <sub>UUC</sub> Reading m/s	Error (m/s)	Uncertainty (%)
2.038	2.0	0.0	2.4
4.056	4.1	0.0	1.2
6.00	6.0	0.0	1.05
7.99	8.0	0.0	0.72
10.01	10.1	0.1	0.76
12.01	12.2	0.2	0.81
13.99	14.3	0.3	0.58
15.98	16.3	0.3	0.42
14.99	15.3	0.3	0.48
13.00	13.2	0.2	0.51
11.01	11.1	0.1	0.61
8.99	9.0	0.0	0.76
6.99	7.0	0.0	0.97
5.092	5.0	-0.1	1.1
2.992	3.0	0.0	1.7
1.008	1.0	0.0	4.8

UUC\*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

#### Appendix 1: Instrumentations

NO	Sensor	Manufacturer	Model/Type	Calibration Date	Certificate Report Number	Range
1	Pitot static	TESTO INC.	06352145	Aug 07, 2021	MW-0034-21	5 – 30 m/s
2	Precision Differential Pressure Meter	Zoglab	DPM2500	Aug 07, 2021	MW-0034-21	5 – 30 m/s
3	Air velocity transducer (hot wire)	TSI INC.	8455-12	Aug 08, 2021	MW-0035-21	0 – 5 m/s
4	Temperature	Zoglab	DSR-THP	March 30, 2021	CL-027-64	-30 – 70°C
5	Relative humidity	Zoglab	DSR-THP	March 30, 2021	RH-03032021	0 – 100 %RH
6	Atmospheric pressure	Zoglab	DSR-THP	March 30, 2021	BP-01032021	500 – 1100 hPa
7	Wind tunnel	BSSOM	MP3300	-	-	0 – 50 Hz

\*\*\*End of certificate of calibration\*\*\*



## CERTIFICATE OF CALIBRATION

Certificate No.: WD-01092021

Page 1 of 2 pages

Measurement Item : Wind direction sensor with data logger.

Manufacturer : Data logger: Novalynx.  
: Wind direction sensor: Novalynx.

Model/Type : Data logger: 110-WS-25DL-D  
: Wind direction sensor: WS-02F.

Serial Number : Data logger: A5662  
: Wind direction sensor: WSD-015.

ID No : Data logger: -  
: Wind direction sensor: -

Customer : ALS laboratory group (Thailand) Co.,Ltd.  
104 Phatthanakan 40, Phatthanakan Rd.,Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250  
Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(23\pm3)^{\circ}\text{C}$ , and relative humidity of  $(40\pm10)\%$ .

### Measurement Method:

The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and line laser is used for axis control. The measurement were taken at  $45^{\circ}$  intervals in clockwise and counterclockwise directions.

Note: The UUC was warmed up for 1 hour prior to the calibration being performed

### Traceability:

The measurement results are traceable to the international system of units (SI) through Certificate No.: CC563-07-0045, Certificate No.: KWS63/0044.

Measurement Date : Sep 15, 2021.

Issued Date : Sep 15, 2021.

### Performed by

- ☒ Mr. Sorawit Thachalad  
☐ Miss Orathai Wiwatwittaya



Approved Signatory.....

*Handwritten signature*

Mr. Parinya Booncharoen,  
Technical Support  
and Calibration Manager



Continuation of Certificate of Calibration Number

Certificate No: WD-01092021

Pages 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment.

Calibration in the range of 0 - 360 ° at a calibration interval of 45°.

The results of calibration and associated measurement uncertainties are reported in table below.

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC* Reading (°)	Error (°)	Uncertainty ±(°)
1	Clockwise	0/360	0	0	0	3.0
2		45	45	42	-3	3.0
3		90	90	88	-2	3.0
4		135	135	134	-1	3.0
5		180	180	181	1	3.0
6		225	225	226	1	3.0
7		270	270	273	3	3.0
8		315	315	318	3	3.0
9	Counter Clockwise	0/360	0	0	0	3.0
10		45	45	42	-3	3.0
11		90	90	88	-2	3.0
12		135	135	134	-1	3.0
13		180	180	181	1	3.0
14		225	225	226	1	3.0
15		270	270	273	3	3.0
16		315	315	318	3	3.0

UUC\*: Unit Under Calibration The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor  $k=2$  providing a level of confidence of approximately 95%

\*\*\*End of Certificate of Calibration\*\*\*



## CERTIFICATE OF CALIBRATION

Certificate No. : CL-067-64  
Page 1 of 2

Equipment Name : Data Logger with Temperature  
Sensor

Manufacturer : Novalynx  
Model : 110-WS-25 DL-D  
Serial No. : A5662  
ID No. : -

### Customer

Name : ALS laboratory group (thailand) Co.,Ltd.  
Address : 104 Phatthanakan 40, Phatthanakan  
Rd.,Khwaeng Suan Luang, Khet Suan Luang,Bangkok  
10250 Thailand.

Received date : 1 SEP 2021  
Calibration date : 13 SEP 2021  
Issue date : 15 SEP 2021

### Reference Used During Calibration

1.Standard Temperature Probe Model : STS-100 A500,  
Serial No. : 667682-09, Due date : 25 Mar 2022  
2.Digital Temperature Indicator Model : DTI-1000-A MK  
II, Serial No.: 671407-00591 Due date : 04 June 2022

### Calibration Condition

Temperature : (23±3)°C  
Relative Humidity : (55±15)%

### Calibration Procedure


The temperature calibration was done by In-House  
calibration method as WI-CL-001 according to  
comparison method with standard digital temperature  
indicator and standard temperature probe. The  
temperature scale use was based on ITS-90.

### Traceability

The measurement results are traceable to the  
international system of units (SI) through National  
Institute of Metrology Thailand (NIMT) Certificate  
number : TT-0036-21, Certificate number : ER-0032-  
21

### Calibrated by

- ☐ Mr. Sorawit Thachalad  
☒ Miss Orathai Wiwatwittaya



Approved Signatory: .....

Mr. Parinya Booncharoen  
Technical Support  
And Calibration Manager

Certificate No. : CL-067-64  
Page 2 of 2

Result of Calibration :- ☒ Without Adjustment ☐ With Adjustment

Calibration Range: 20 °C – 40 °C

Function:

This equipment was connected with temperature sensor Model : HMP60 S/N : T2320591

Dimension : Diameter 12mm. Length 80 mm.

<u>Immersion Depth (mm)</u>	<u>Standard Reading (°C)</u>	<u>UUC Reading (°C)</u>	<u>Error (°C)</u>	<u>Uncertainty (°C)</u>
60	20.049	19.8	-0.2	0.080
60	24.879	24.5	-0.4	0.16
60	29.864	29.4	-0.5	0.080
60	34.847	34.4	-0.5	0.13
60	39.835	39.3	-0.5	0.080

**UUC\*** : Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

\* End of Certificate \*



## CALIBRATION REPORT

Calibration No. : RH-01092021

Page 1 of 1 Pages

Measurement Item : Relative humidity with data logger.

Manufacturer : Data logger: Novalynx.  
: Relative humidity sensor: Novalynx.

Model/Type : Data logger: 110-WS-25 DL-D.  
: Relative humidity sensor: HMP60.

Serial Number : Data logger: A5662.  
: Relative humidity sensor: T2320691.

ID No : Data logger: -  
: Relative humidity sensor: -.

Customer : ALS laboratory group (Thailand) co., ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250  
Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(25 \pm 3)^{\circ}\text{C}$ , and relative humidity of  $(50 \pm 15)\%$ .

### Measurement Method:

The Relative humidity with data logger, Unit Under Calibration (UUC) was calibrated by comparison method with the equilibrium of standard salt solution  $\text{CH}_3\text{COOK}$ : Potassium Acetate,  $\text{Mg}(\text{NO}_3)_2$ : Magnesium Nitrate,  $\text{KCl}$ : Potassium Chloride to determine the errors.

Measurement Date : Sep 13, 2021

Issued Date : Sep 15, 2021

### Measurement Results:

The results of calibration are reported in table below.

Standard salt solution.	Standard (%RH)	UUC(Reading)	Error
$\text{CH}_3\text{COOK}$ : Potassium Acetate	22.51	23.5	1.0
$\text{Mg}(\text{NO}_3)_2$ : Magnesium Nitrate	52.89	53.1	0.2
$\text{KCl}$ : Potassium Chloride	84.34	83.9	-0.4

### Performed by

- ☐ Mr. Sorawit Thachalad  
☒ Miss Orathai Wawatwittaya



Signatory: \_\_\_\_\_

Mr. Parinya Booncharoen,  
Technical Support  
And Calibration Manager



## CERTIFICATE OF CALIBRATION

Certificate No: WS-05012022

Page 1 of 2 pages

Measurement Item : Cup anemometer with data logger.

Manufacturer : Data logger: Novalynx  
: Cup anemometer: Novalynx

Model/Type : Data logger: 200-WS-25LB  
: Cup anemometer: WS-02F

Serial Number : Data logger: A5190  
: Cup anemometer: -

ID No : Data logger: RYG\_FS0329  
: Cup anemometer: -

Customer : ALS laboratory group (Thailand) co., ltd.  
: 104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250  
Thailand.

Test Conditions : Wind tunnel cross test section area 900 cm<sup>2</sup>  
: Anemometer frontal area 100 cm<sup>2</sup>  
: Diameter of mounting pipe - mm  
: Blockage ratio of test object 0.111 [-]

Test Conditions : Air temperature 23.6 ±0.8 °C  
: Air pressure 1014.5 ±0.4 hPa  
: Relative air humidity 53.4 ±3.5 %RH

Calibration Procedure : Calibration was carried out base on;  
IEC 61400-12-1 ED.1: 2005-Power Performance Measurements of Electricity Producing Wind  
Turbines;  
MBASNET Anemometer Calibration Procedure – Version 2: 2009;

Traceability : This calibration documents the traceable to national standard, Which realize the unit of  
measurements according to the international system of units (SI) through National Institute of  
Metrology Thailand (NIMT).

Measurement Date : JAN 28, 2022.

Issued Date : JAN 31, 2022.

Calibrated by

- ☒ Mr. Sorawit Thachalad  
☐ Miss Orathai Wiwatwitaya



Approved Signatory: \_\_\_\_\_

Mr. Parinya Booncharoen  
Calibration Department Manager

Continuation of Certificate of Calibration Number

Certificate No: WS-05012022

Page 2 of 2 Pages

Result of calibration: ☒ Without adjustment ☐ With adjustment

Calibration in the range of 1 – 16 m/s at a calibration interval of 1 m/s.

The results of calibration and associated measurement uncertainties are reported in the table below.

V <sub>STD</sub> Reading m/s	V <sub>UUC</sub> Reading m/s	Error (m/s)	Uncertainty (%)
2.076	2.0	-0.1	2.4
4.101	4.1	0.0	1.2
5.99	6.0	0.0	0.95
8.01	8.0	0.0	0.83
10.01	10.1	0.1	0.79
12.01	12.1	0.1	0.57
13.99	14.1	0.1	0.70
15.99	16.4	0.4	0.43
15.00	15.2	0.2	0.79
13.01	13.0	0.0	0.83
11.02	11.0	0.0	0.76
9.03	9.0	0.0	0.81
7.02	7.0	0.0	0.82
5.130	5.1	0.0	0.96
2.991	3.0	0.0	1.6
1.036	0.9	-0.1	4.5

UUC\*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

#### Appendix 1: Instrumentations

NO	Sensor	Manufacturer	Model/Type	Calibration Date	Certificate Report Number	Range
1	Pitot static	TESTO INC.	06352145	Aug 07, 2021	MW-0034-21	5 – 30 m/s
2	Precision Differential Pressure Meter	Zoglab	DPM2500	Aug 07, 2021	MW-0034-21	5 – 30 m/s
3	Air velocity transducer (hot wire)	TSI INC.	8455-12	Aug 08, 2021	MW-0035-21	0 – 5 m/s
4	Temperature	Zoglab	DSR-THP	March 30, 2021	CL-027-64	-30 – 70°C
5	Relative humidity	Zoglab	DSR-THP	March 30, 2021	RH-03032021	0 – 100 %RH
6	Atmospheric pressure	Zoglab	DSR-THP	March 30, 2021	BP-01032021	500 – 1100 hPa
7	Wind tunnel	ESSOM	MP330D	-	-	0 – 50 Hz

\*\*\*End of certificate of calibration\*\*\*



## CERTIFICATE OF CALIBRATION

Certificate No.: WD-05012022

Page 1 of 2 pages

Measurement Item : Wind direction sensor with data logger.

Manufacturer : Data logger: Novalynx.  
: Wind direction sensor: Novalynx.

Model/Type : Data logger: 200-WS-25LB  
: Wind direction sensor: WS-02F

Serial Number : Data logger: A5190  
: Wind direction sensor: -

ID No : Data logger: RYG\_FS0329  
: Wind direction sensor: -

Customer : ALS laboratory group (Thailand) Co.,Ltd.  
104 Phatthanakan 40, Phatthanakan Rd.,Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250  
Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(23\pm3)$  °C, and relative humidity of  $(40\pm10)$  %.

### Measurement Method:

The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and line laser is used for axis control. The measurement were taken at 45° intervals in clockwise and counterclockwise directions.

Note: The UUC was warmed up for 1 hour prior to the calibration being performed

### Traceability:

The measurement results are traceable to the international system of units (SI) through Certificate No.: Q21086014, Certificate No.: KWS64/D025.

Measurement Date : JAN 26, 2022.

Issued Date : JAN 31, 2022.

### Performed by

- ☒ Mr. Sorawit Thachalad  
☐ Miss Orathai Wiwatwittaya



Approved Signatory:.....



Mr. Parinya Booncharoen.  
Calibration Department Manager



Continuation of Certificate of Calibration Number

Certificate No: WD-05012022

Pages 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment.

Calibration in the range of 0 – 360 ° at a calibration interval of 45°.

The results of calibration and associated measurement uncertainties are reported in table below.

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC* Reading (°)	Error (°)	Uncertainty ±(°)
1	Clockwise	0/360	0	0	0	3.0
2		45	45	43	-2	3.0
3		90	90	90	0	3.0
4		135	135	135	0	3.0
5		180	180	181	1	3.0
6		225	225	227	2	3.0
7		270	270	273	3	3.0
8		315	315	318	3	3.0
9	Counter Clockwise	0/360	0	0	0	3.0
10		45	45	43	-2	3.0
11		90	90	90	0	3.0
12		135	135	135	0	3.0
13		180	180	181	1	3.0
14		225	225	227	2	3.0
15		270	270	273	3	3.0
16		315	315	318	3	3.0

UUC\*: Unit Under Calibration The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

\*\*\*End of Certificate of Calibration\*\*\*







JIRANATEE ASSOCIATES CO., LTD.

Jiranatee Associates Co., Ltd.  
63/14-15, 67/35-36  
Petchkasem 7,7/1, Rd. Watthapra, Bangkokyal,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Air speed measurement laboratory  
Calibration services department.

REVIEW BY	<i>Naratorn P</i>
APPROVED BY	<i>Mr. P</i>
NEXT CAL. DATE	19/7/24

Certificate Number

CL-013-66

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

**MEASUREMENT ITEM** : Wind Direction Sensor  
**MANUFACTURER** : Novalynx  
**MODEL/TYPE** : Sensor: WS-02F  
Data logger: 110-WS-25DL-D  
**SERIAL NUMBER** : Sensor: WSD-014  
Data logger: A5789  
**ID NUMBER** : RYG\_FS0531  
**CONDITION AS-RECEIVED** : Used item  
**CUSTOMER** : ALS laboratory group (Thailand) Co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,  
Khet Suan Luang, Bangkok 10250 Thailand.

**RECEIVED DATE** : 16 Jan 2023  
**MEASUREMENT DATE** : 19 Jan 2023  
**ISSUE DATE** : 20 Jan 2023

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature	: 23.0 ± 3.0	°C
Relative Humidity	: 55.0 ± 15.0	%RH
Atmospheric Pressure	: 1010 ± 10	hPa

**PLACE OF CALIBRATION** : Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

<b>CALIBRATION CONDITION</b>	: Wind tunnel cross-section area <sup>1</sup>	900	cm <sup>2</sup>
	Win direction frontal area <sup>2</sup>	129	cm <sup>2</sup>
	Diameter of mounting pipe <sup>3</sup>	-	mm
	Blockage ratio of test object <sup>4</sup>	0.143	[-]

**Preconditioning** : 24 hours at ambient conditions.  
**Measurement Condition** : The average values during measurement are (23.6)°C, (46.6) %RH and (1014.9) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jittraporn Lertsomphol



### Approved signatory:

*Mr. Parinya Booncharoen*  
Calibration Department Manager

### Remark:

- <sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>

MEASUREMENT RESULTS<sup>5</sup>

The wind direction sensor was calibrated against standard rotary encoder by comparison method. During calibration, the measurement was carried out at 45° intervals in clockwise and counterclockwise directions after offset adjustment has been made. The flow speed of wind tunnel (usually 5 m/s) is kept constant while the sensor is rotated around its vertical axis. The results of calibration and associated measurement uncertainties are reported in the table below.

Air speed m/s	$D^{\circ}_{std}$ Degree (°)	$D^{\circ}_{uuc}$ Degree (°)	Error Degree (°)	$U (k=2)$ Degree (°)
4.99	0.000	0	0	0.58
	45.000	43	-2	0.74
	90.000	88	-2	0.74
	135.000	133	-2	0.74
	180.000	179	-1	0.74
	225.000	227	2	0.74
	270.000	272	2	0.74
	315.000	317	2	0.74

## Remark:

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Direction of standard

<sup>7</sup> Direction of Unit Under Calibration

\*\*\*End of Certificate of Calibration\*\*\*







JIRANATEE ASSOCIATES CO.,LTD.

Jiranatee Associates Co.,Ltd.  
63/14-15, 67/35-36  
Petchkasem 7,7/1, Rd.Watthapra, Bangkokyai,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Air speed measurement laboratory  
Calibration services department.

Certificate Number

CL-013-66

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

**MEASUREMENT ITEM** : Cup anemometer  
**MANUFACTURER** : Novalynx  
**MODEL/TYPE** : Sensor: WS-02F  
Data logger: 110-WS-25DL-D  
**SERIAL NUMBER** : Sensor: WSD-014  
Data logger: A5789  
**ID NUMBER** : RYG\_FS0531  
**CONDITION AS-RECEIVED** : Used item  
**CUSTOMER** : ALS laboratory group (Thailand) co., ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,  
Khet Suan Luang, Bangkok 10250 Thailand.

### Calibration procedure:

The cup anemometer was calibrated against Standard air velocity transducer model: 8455-12 and pitot tube with precision differential pressure meter model: DPM2500 in an close test-section of Eiffel-type wind tunnel with 900 cm<sup>2</sup> cross test section area. The WI-CL-007 based on IEC 61400-12-1, Wind energy generation systems – Part 12-1: Power performance measurements of electricity producing wind turbines, March 2017 was used as a calibration guideline.

### Traceability:

This certificate provides a traceability of The measurement to recognized the national standards, and to realization of the international system of units (SI) through the NIMT (National Metrology Institute of Thailand) via Certificate number: MW-0052-21 and MW-0066-22

### Uncertainty of Measurement:

The reported uncertainty of measurement is based on the standard uncertainty multiplied by a coverage factor k=2, Which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty has been determined in accordance with the GUM 'Evaluation of measurement data - Guide to the expression of uncertainty in measurement'

**RECEIVED DATE** : 16 Jan 2023  
**MEASUREMENT DATE** : 18 Jan 2023  
**ISSUE DATE** : 20 Jan 2023

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature : 23.0 ± 3.0 °C  
Relative Humidity : 55.0 ± 15.0 %RH  
Atmospheric Pressure : 1010 ± 10 hPa

**PLACE OF CALIBRATION** : Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

**CALIBRATION CONDITIONS** : Wind tunnel cross-section area<sup>1</sup> 900 cm<sup>2</sup>  
Win direction frontal area<sup>2</sup> 100 cm<sup>2</sup>  
Diameter of mounting pipe<sup>3</sup> - mm  
Blockage ratio of test object<sup>4</sup> 0.111 [-]

**Preconditioning** : 24 hours at ambient conditions.  
**Measurement Condition** : The average values during measurement are (23.7) °C, (44.5) %RH and (1018.3) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jittraporn Lertsomphol

Approved signatory: \_\_\_\_\_

Mr. Parinya Booncharoen  
Calibration Department Manager

### Remark:

- <sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>





**MEASUREMENT RESULTS<sup>5</sup>**

The cup anemometer, Unit Under Calibration (UUC) was exercise at 10 m/s for 5 minutes prior to calibration being performed. The standard air velocity 0.5 m/s to 5 m/s was calculated by a standard air velocity transducer and above 5 m/s to 30 m/s was calculated by a pitot tube with precision differential pressure meter which was installed 40 mm and 300 mm respectively away from wind tunnel nozzle, UUC was installed at center of the test section. The calibration was carried out under both rising and falling air velocity in the range of 1 m/s to 16 m/s at calibration interval of 1 m/s. The results of calibration and associated measurement uncertainties are reported in the table below.


$v_{std}^6$ (m/s)	Temp. wind tunnel (°C)	Temp. room (°C)	$v_{UUC}^7$ (m/s)	Error (m/s)	$U (k=2)$ (m/s)
0.983	23.60	23.70	0.7	-0.3	0.18
2.024	23.74	23.70	1.7	-0.3	0.16
3.044	23.50	23.70	2.9	-0.2	0.18
4.119	23.82	23.70	3.9	-0.2	0.19
5.02	23.50	23.70	4.9	-0.2	0.18
5.99	23.88	23.70	5.8	-0.2	0.18
7.08	23.50	23.70	6.9	-0.1	0.20
8.18	23.58	23.70	8.0	-0.2	0.18
9.11	23.50	23.70	9.0	-0.1	0.19
10.08	23.66	23.70	10.0	-0.1	0.25
11.15	23.32	23.70	11.0	-0.2	0.21
12.14	23.66	23.70	12.0	-0.2	0.20
13.20	23.32	23.70	13.2	0.0	0.25
14.25	23.50	23.70	14.1	-0.1	0.27
15.23	23.30	23.70	15.1	-0.2	0.27
16.29	23.40	23.70	16.2	-0.1	0.23

**Remark:**

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Velocity of standard

<sup>7</sup> Velocity of Unit Under Calibration

**PHOTO OF CALIBRATION SET-UP**

Calibration set-up of the cup anemometer calibration in the wind tunnel of Jiranatee Associates Co., Ltd. The cup anemometer shown may differ from the calibrated one. Remark: The proportion of the set-up is not true to scale due to imaging geometry.

## CERTIFICATE OF CALIBRATION

Certificate No.: CL-006-66  
Page 1 of 2

Equipment Name: Data Logger with Temperature  
Sensor

Manufacturer: Novalynx  
Model: 110-WS-25DL-D  
Serial No.: A5789  
ID No.: RYG\_FS0531

### Customer

Name: ALS laboratory group (Thailand) Co., Ltd.  
Address: 104 Phatthanakan 40, Phatthanakan Rd.,  
Khwaeng Suan Luang, Khet Suan Luang, Bangkok  
10250 Thailand.

Received date: 16 Jan 2023  
Calibration date: 18 Jan 2023  
Issue date: 20 Jan 2023

### Reference Used During Calibration

1. Standard Temperature Probe Model: STS-100 A500,  
Serial No.: 667682-09, Due date: 23 Mar 2023  
2. Digital Temperature Indicator Model: DTI-1000-A MK  
II, Serial No.: 671407-00591 Due date: 22 July 2023

### Calibration Condition

Temperature:  $(23 \pm 3)^{\circ}\text{C}$   
Relative Humidity:  $(55 \pm 15)\%$

### Calibration Procedure

The temperature calibration was done by In-House  
calibration method as WI-CL-001 according to  
comparison method with standard digital temperature  
indicator and standard temperature probe. The  
temperature scale use was based on ITS-90.

### Traceability

The measurement results are traceable to the  
international system of units (SI) through National  
Institute of Metrology Thailand (NIMT) Certificate  
number: TT-0034-22, Certificate number: ER-0092-  
22

### Calibrated by

☐ Mr. Sorawit Thachalad  
☒ Miss Jitraporn Lertsomphol



### Approved Signatory:

*25/Ans*

Mr. Parinya Booncharoen  
Calibration Department Manager

**Result of Calibration:-** ☒ Without Adjustment ☐ With Adjustment

**Calibration Range:** 20-40 °C

**Function:**

This equipment was connected with temperature sensor Model: HMP60 S/N: T0210901.

Dimension : Diameter 12 mm. Length 80 mm.

<u>Immersion Depth (mm)</u>	<u>Standard Reading (°C)</u>	<u>UUC Reading (°C)</u>	<u>Error (°C)</u>	<u>Uncertainty (°C)</u>
60	20.067	19.8	-0.3	0.099
60	25.058	24.6	-0.5	0.099
60	30.052	29.5	-0.6	0.099
60	35.047	34.5	-0.5	0.099
60	40.038	39.3	-0.7	0.099

UUC\*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

**\* End of Certificate \***





## CERTIFICATE OF CALIBRATION

Calibration No. : RH-06012023

Page 1 of 1 Pages

Measurement Item : Relative humidity with data logger  
Manufacturer : Novalynx  
Model/Type : 110-WS-25DL-D  
Serial Number : A5789  
ID No. : RY0\_FS0531  
Customer : ALS laboratory group (Thailand) Co., Ltd.  
: 104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Suan Luang, Khet Suan Luang, Bangkok  
10250 Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(25 \pm 3)^{\circ}\text{C}$ , and relative humidity of  $(50 \pm 15)\%$ .

### Measurement Method:

Unit Under Calibration (UUC) was calibrated by comparison method with standard thermo hygrometer in the humidity generator chamber to determine the errors.

### Traceability:

This instrument was calibrated using standard equipment whose accuracy is traceability through National Institute of Standards and Technology to the international system of units (SI) via MCS Calibration, Inc. Certificate number: 20314-101. Due date: Mar 14, 2023.

Measurement Date : Jan 18, 2023

Issued Date : Jan 20, 2023

### Measurement Results:

This equipment was connected with Indoor air quality probe and Displayed (UR) on display. Model: HMP60, Serial number: T0210901.

Calibration was performed in the range of 20%RH to 80%RH

The results of calibration are reported in table below.

Determined (%RH)	Standard (Reading) (%RH)	UUC (Reading) (%RH)	Error (%RH)	Uncertainty $\pm$ (%RH)
20	20.03	18.0	-2.0	0.51
50	50.24	47.8	-2.4	0.51
80	80.19	77.3	-2.9	0.51

Performed by

☐ Mr. Sorawit Thachalad

☒ Miss Jittraporn Lertsomphol



Approved Signatory: 

Mr. Parinya Booncharoen.  
Calibration Department Manager

## CERTIFICATE OF CALIBRATION

Certificate No: WS-02092021

Page 1 of 2 pages

Measurement Item : Cup anemometer with data logger.

Manufacturer : Data logger: Novalynx.  
: Cup anemometer: Novalynx.

Model/Type : Data logger: 110-WS-25DL-D  
: Cup anemometer: WS-02F

Serial Number : Data logger: A5816  
: Cup anemometer: WSD-016

ID No : Data logger: - R16 330445  
: Cup anemometer: -

Customer : ALS laboratory group (Thailand) co., ltd.  
: 104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand.

Test Conditions : Wind tunnel cross test section area 900 cm<sup>2</sup>  
: Anemometer frontal area 100 cm<sup>2</sup>  
: Diameter of mounting pipe - mm  
: Blockage ratio of test object 0.11 [-]

Test Conditions : Air temperature 23.6 ±0.8 °C  
: Air pressure 1012.9 ±0.4 hPa  
: Relative air humidity 57.8 ±3.5 %RH

Calibration Procedure : Calibration was carried out base on;  
ISO 61400-12-1 ED.1: 2005-Power Performance Measurements of Electricity Producing Wind Turbines;  
MCASNET Anemometer Calibration Procedure – Version 2: 2009;

Traceability : This calibration documents the traceable to national standard, Which realize the unit of measurements according to the international system of units (SI) through National Institute of Metrology Thailand (NIMT).

Measurement Date : Sep 14, 2021.  
Issued Date : Sep 15, 2021.

**Calibrated by**

- ☒ Mr. Sorawit Thachalad  
☐ Miss Orathai Wiwatwitaya



Approved Signatory:

*Signature of Mr. Parinya Booncharoen*

Mr. Parinya Booncharoen  
Technical Support  
and Calibration Manager

Continuation of Certificate of Calibration Number

Certificate No: WS-02092021

Page 2 of 2 Pages

Result of calibration: ☒ Without adjustment ☐ With adjustment

Calibration in the range of 1 – 16 m/s at a calibration interval of 1 m/s.

The results of calibration and associated measurement uncertainties are reported in the table below.

V <sub>STD</sub> Reading m/s	V <sub>UUC</sub> Reading m/s	Error (m/s)	Uncertainty (%)
2.045	2.0	0.0	2.4
4.087	4.0	-0.1	1.5
5.98	6.0	0.0	1.2
8.01	8.0	0.0	0.84
10.02	10.1	0.1	0.67
12.02	12.2	0.2	0.63
14.00	14.2	0.2	0.42
15.99	16.2	0.2	0.76
14.99	15.2	0.2	0.49
13.01	13.1	0.1	0.51
11.02	11.1	0.1	0.66
9.02	9.0	0.0	0.65
7.00	7.0	0.0	0.90
5.122	5.1	0.0	1.3
2.978	3.0	0.0	1.9
1.023	1.0	0.0	4.8

UUC\*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

#### Appendix 1: Instrumentations

NO	Sensor	Manufacturer	Model/Type	Calibration Date	Certificate Report Number	Range
1	Pitot static	TESTO INC.	Q6352145	Aug 07, 2021	MW-0034-21	5 – 30 m/s
2	Precision Differential Pressure Meter	Zoglab	DPM2500	Aug 07, 2021	MW-0034-21	5 – 30 m/s
3	Air velocity transducer (hot wire)	TSI INC.	8455-12	Aug 08, 2021	MW-0035-21	0 – 5 m/s
4	Temperature	Zoglab	DSR-THP	March 30, 2021	CL-027-64	-30 – 70°C
5	Relative humidity	Zoglab	DSR-THP	March 30, 2021	RH-03032021	0 – 100 %RH
6	Atmospheric pressure	Zoglab	DSR-THP	March 30, 2021	BP-01032021	500 – 1100 hPa
7	Wind tunnel	ESSOM	MP3300	-	-	0 – 50 Hz

\*\*\*End of certificate of calibration\*\*\*





## CERTIFICATE OF CALIBRATION

Certificate No.: WD-02092021

Page 1 of 2 pages

Measurement Item : Wind direction sensor with data logger.

Manufacturer : Data logger: Novalynx.  
: Wind direction sensor: Novalynx.

Model/Type : Data logger: 110-WS-25DL-D  
: Wind direction sensor: WS-02F.

Serial Number : Data logger: A5816.  
: Wind direction sensor: WSD-016.

ID No : Data logger: -  
: Wind direction sensor: -

Customer : ALS laboratory group (Thailand) Co.,Ltd.  
104 Phatthanakan 40, Phatthanakan Rd.,Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250  
Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(23\pm3)^{\circ}\text{C}$ , and relative humidity of  $(40\pm10)\%$ .

### Measurement Method:

The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and line laser is used for axis control. The measurement were taken at  $45^{\circ}$  intervals in clockwise and counterclockwise directions.

Note: The UUC was warmed up for 1 hour prior to the calibration being performed

### Traceability:

The measurement results are traceable to the International system of units (SI) through Certificate No.: CC563-07-0045, Certificate No.: KWS63/0044.

Measurement Date : Sep 15, 2021.  
Issued Date : Sep 15, 2021.

### Performed by

- ☒ Mr. Sorawit Thachalad  
☐ Miss Orathai Wiwatwittaya



Approved Signatory:.....



Mr. Parinya Booncharoen.  
Technical Support  
and Calibration Manager

Continuation of Certificate of Calibration Number

Certificate No: WD-02092021

Pages 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment.

Calibration in the range of 0 – 360 ° at a calibration interval of 45°.

The results of calibration and associated measurement uncertainties are reported in table below.

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC* Reading (°)	Error (°)	Uncertainty ±(°)
1	Clockwise	0/360	360	359	-1	3.0
2		45	45	44	-1	3.0
3		90	90	87	-3	3.0
4		135	135	133	-2	3.0
5		180	180	180	0	3.0
6		225	225	225	0	3.0
7		270	270	273	3	3.0
8		315	315	317	2	3.0
9	Counter Clockwise	0/360	360	359	-1	3.0
10		45	45	44	-1	3.0
11		90	90	87	-3	3.0
12		135	135	133	-2	3.0
13		180	180	180	0	3.0
14		225	225	225	0	3.0
15		270	270	273	3	3.0
16		315	315	317	2	3.0

UUC\*: Unit Under Calibration The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor  $k=2$  providing a level of confidence of approximately 95%

\*\*\*End of Certificate of Calibration\*\*\*

## CERTIFICATE OF CALIBRATION

Certificate No. : CL-068-64

Page 1 of 2

Equipment Name : Data Logger with Temperature  
Sensor

Manufacturer : Novalynx  
Model : 110-WS-25 DL-D  
Serial No. : A5816  
ID No. : -

### Customer

Name : ALS laboratory group (thailand) Co.,Ltd.  
Address : 104 Phatthanakan 40, Phatthanakan  
Rd., Khwaeng Suan Luang, Khet Suan Luang, Bangkok  
10250 Thailand.

Received date : 1 SEP 2021

Calibration date : 13 SEP 2021

Issue date : 15 SEP 2021

### Reference Used During Calibration

1. Standard Temperature Probe Model : STS-100 A500,  
Serial No. : 667682-09, Due date : 25 Mar 2022  
2. Digital Temperature Indicator Model : DTI-1000-A MK  
II, Serial No.: 671407-00591 Due date : 04 June 2022

### Calibration Condition

Temperature :  $(23 \pm 3)^{\circ}\text{C}$   
Relative Humidity :  $(55 \pm 15)\%$

### Calibration Procedure

The temperature calibration was done by In-House  
calibration method as WI-CL-001 according to  
comparison method with standard digital temperature  
indicator and standard temperature probe. The  
temperature scale use was based on ITS-90.

### Traceability

The measurement results are traceable to the  
international system of units (SI) through National  
Institute of Metrology Thailand (NIMT) Certificate  
number : TT-0036-21, Certificate number : ER-0032-  
21

### Calibrated by

- ☐ Mr. Sorawit Thachalad  
☒ Miss Orathai Wiwatwittaya



Approved Signatory: \_\_\_\_\_

*25/09/21*  
Mr. Parinya Booncharoen  
Technical Support  
And Calibration Manager



Result of Calibration :- ☒ Without Adjustment ☐ With Adjustment

Calibration Range: 20°C - 40 °C

Function:

This equipment was connected with temperature sensor Model : HMP60 S/N : T2320595

Dimension : Diameter 12mm. Length 80 mm.

<u>Immersion</u> <u>Depth</u> (mm)	<u>Standard</u> <u>Reading</u> (°C)	<u>UUC</u> <u>Reading</u> (°C)	<u>Error</u> (°C)	<u>Uncertainty</u> (°C)
60	20.050	19.6	-0.4	0.080
60	24.878	24.5	-0.4	0.080
60	29.857	29.4	-0.4	0.13
60	34.848	34.3	-0.5	0.080
60	39.842	39.3	-0.5	0.080

UUC\* : Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

\* End of Certificate \*



## CALIBRATION REPORT

Calibration No. : RH-02092021

Page 1 of 1 Pages

Measurement Item : Relative humidity with data logger.

Manufacturer : Data logger: Novalynx.  
: Relative humidity sensor: Novalynx.

Model/Type : Data logger: 110-WS-25 DL-D.  
: Relative humidity sensor: HMP60.

Serial Number : Data logger: A5816.  
: Relative humidity sensor: T2320595.

ID No : Data logger: -  
: Relative humidity sensor: -.

Customer : ALS laboratory group (Thailand) co., ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250  
Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(25 \pm 3)^{\circ}\text{C}$ , and relative humidity of  $(50 \pm 15)\%$ .

### Measurement Method:

The Relative humidity with data logger, Unit Under Calibration (UUC) was calibrated by comparison method with the equilibrium of standard salt solution  $\text{CH}_3\text{COOK}$ : Potassium Acetate,  $\text{Mg}(\text{NO}_3)_2$ : Magnesium Nitrate,  $\text{KCl}$ : Potassium Chloride to determine the errors.

Measurement Date : Sep 13, 2021

Issued Date : Sep 15, 2021

### Measurement Results:

The results of calibration are reported in table below.

Standard salt solution.	Standard (%RH)	UUC(Reading)	Error
$\text{CH}_3\text{COOK}$ : Potassium Acetate	22.51	23.6	1.1
$\text{Mg}(\text{NO}_3)_2$ : Magnesium Nitrate	52.89	53.1	0.2
$\text{KCl}$ : Potassium Chloride	84.34	85.4	1.0

### Performed by

- ☐ Mr. Sorawit Thachalad  
☒ Miss Orathai Wiwatwittaya



Signatory: \_\_\_\_\_

*Signature*  
Mr. Parinya Booncharoen,  
Technical Support  
and Calibration Manager



JIRANATEE ASSOCIATES CO.,LTD.

Jiranatee Associates Co.,Ltd.  
63/14-15, 67/35-36  
Petchkasem 7,7/1, Rd. Watthapra, Bangkokyal,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Air speed measurement laboratory  
Calibration services department.

REVIEW BY	Manakorn P.
APPROVED BY	[Signature]
NEXT CAL. DATE	19/7/24

Certificate Number

CL-011-66

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

**MEASUREMENT ITEM** : Wind Direction Sensor  
**MANUFACTURER** : Novalynx  
**MODEL/TYPE** : Sensor: WS-02F  
Data logger: 200-WS-25DL  
**SERIAL NUMBER** : Sensor: -  
Data logger: A4987  
**ID NUMBER** : RYG\_FS0089  
**CONDITION AS-RECEIVED** : Used item  
**CUSTOMER** : ALS laboratory group (Thailand) Co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,  
Khet Suan Luang, Bangkok 10250 Thailand.

**RECEIVED DATE** : 16 Jan 2023  
**MEASUREMENT DATE** : 19 Jan 2023  
**ISSUE DATE** : 20 Jan 2023

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature	: 23.0 ± 3.0	°C
Relative Humidity	: 55.0 ± 15.0	%RH
Atmospheric Pressure	: 1010 ± 10	hPa

**PLACE OF CALIBRATION** : Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

<b>CALIBRATION CONDITION</b>	: Wind tunnel cross-section area <sup>1</sup>	900	cm <sup>2</sup>
	Win direction frontal area <sup>2</sup>	129	cm <sup>2</sup>
	Diameter of mounting pipe <sup>3</sup>	-	mm
	Blockage ratio of test object <sup>4</sup>	0.143	[-]

**Preconditioning** : 24 hours at ambient conditions.  
**Measurement Condition** : The average values during measurement are (24.1)°C, (54.3) %RH and (1015.2) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



Approved signatory: .....

Mr. Parinya Booncharoen  
Calibration Department Manager

### Remark:

- <sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>



MEASUREMENT RESULTS <sup>5</sup>

The wind direction sensor was calibrated against standard rotary encoder by comparison method. During calibration, the measurement was carried out at 45° intervals in clockwise and counterclockwise directions after offset adjustment has been made. The flow speed of wind tunnel (usually 5 m/s) is kept constant while the sensor is rotated around its vertical axis. The results of calibration and associated measurement uncertainties are reported in the table below.

Air speed m/s	$D^{\circ}_{std}$ Degree (°)	$D^{\circ}_{uuc}$ Degree (°)	Error Degree (°)	$U (k=2)$ Degree (°)
5.03	0.000	0	0	0.58
	45.000	41	-4	0.68
	90.000	88	-2	0.74
	135.000	133	-2	0.58
	180.000	180	0	0.74
	225.000	228	3	0.74
	270.000	273	3	0.68
	315.000	316	1	0.74

## Remark:

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Direction of standard

<sup>7</sup> Direction of Unit Under Calibration

\*\*\*End of Certificate of Calibration\*\*\*





JIRANATEE ASSOCIATES CO.,LTD.

Jiranatee Associates Co.,Ltd  
63/14-15, 67/35-36  
Petchkasem 7,7/1, Rd. Watthapra, Bangkokyai,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Air speed measurement laboratory  
Calibration services department.

Certificate Number

CL-011-66

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

**MEASUREMENT ITEM** : Cup anemometer  
**MANUFACTURER** : Novalynx  
**MODEL/TYPE** : Sensor: WS-02F  
Data logger: 200-WS-25DL  
**SERIAL NUMBER** : Sensor: -  
Data logger: A4987  
**ID NUMBER** : RYG\_FS0089  
**CONDITION AS-RECEIVED** : Used item  
**CUSTOMER** : ALS laboratory group (Thailand) co., ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,  
Khet Suan Luang, Bangkok 10250 Thailand.

### Calibration procedure:

The cup anemometer was calibrated against Standard air velocity transducer model: 8455-12 and pitot tube with precision differential pressure meter model: DPM2500 in an close test-section of Eiffel-type wind tunnel with 900 cm<sup>2</sup> cross test section area. The WI-CL-007 based on IEC 61400-12-1, Wind energy generation systems – Part 12-1: Power performance measurements of electricity producing wind turbines, March 2017 was used as a calibration guideline.

### Traceability:

This certificate provides a traceability of The measurement to recognized the national standards, and to realization of the international system of units (SI) through the NIMT (National Metrology Institute of Thailand) via Certificate number: MW-0052-21 and MW-0066-22

### Uncertainty of Measurement:

The reported uncertainty of measurement is based on the standard uncertainty multiplied by a coverage factor k=2, Which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty has been determined in accordance with the GUM 'Evaluation of measurement data - Guide to the expression of uncertainty in measurement'

**RECEIVED DATE** : 16 Jan 2023  
**MEASUREMENT DATE** : 18 Jan 2023  
**ISSUE DATE** : 20 Jan 2023

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature : 23.0 ± 3.0 °C  
Relative Humidity : 55.0 ± 15.0 %RH  
Atmospheric Pressure : 1010 ± 10 hPa

**PLACE OF CALIBRATION** : Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

**CALIBRATION CONDITIONS** : Wind tunnel cross-section area<sup>1</sup> 900 cm<sup>2</sup>  
Win direction frontal area<sup>2</sup> 100 cm<sup>2</sup>  
Diameter of mounting pipe<sup>3</sup> - mm  
Blockage ratio of test object<sup>4</sup> 0.111 [-]

**Preconditioning** : 24 hours at ambient conditions.  
**Measurement Condition** : The average values during measurement are (23.5) °C, (52.8) %RH and (1014.1) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



### Approved signatory:

Mr. Parinya Booncharoen  
Calibration Department Manager

### Remark:

- <sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>

**MEASUREMENT RESULTS <sup>5</sup>**

The cup anemometer, Unit Under Calibration (UUC) was exercise at 10 m/s for 5 minutes prior to calibration being performed. The standard air velocity 0.5 m/s to 5 m/s was calculated by a standard air velocity transducer and above 5 m/s to 30 m/s was calculated by a pitot tube with precision differential pressure meter which was installed 40 mm and 300 mm respectively away from wind tunnel nozzle, UUC was installed at center of the test section. The calibration was carried out under both rising and falling air velocity in the range of 1 m/s to 16 m/s at calibration interval of 1 m/s. The results of calibration and associated measurement uncertainties are reported in the table below.

$v_{std}^6$ (m/s)	Temp. wind tunnel (°C)	Temp. room (°C)	$v_{uuc}^7$ (m/s)	Error (m/s)	$U (k=2)$ (m/s)
0.983	23.50	23.45	0.8	-0.2	0.17
2.035	23.44	23.45	1.9	-0.1	0.16
3.049	23.50	23.45	2.9	-0.2	0.19
4.136	23.50	23.45	3.9	-0.2	0.20
5.01	23.40	23.45	4.9	-0.1	0.18
6.00	23.50	23.45	5.9	-0.1	0.19
7.07	23.40	23.45	7.0	-0.1	0.19
8.18	23.50	23.45	8.0	-0.2	0.19
9.10	23.26	23.45	9.0	-0.1	0.20
10.09	23.44	23.45	9.9	-0.1	0.21
11.15	23.30	23.45	11.0	-0.1	0.21
12.14	23.42	23.45	12.0	-0.1	0.25
13.20	23.22	23.45	13.1	-0.1	0.26
14.25	23.34	23.45	14.1	-0.1	0.24
15.24	23.24	23.45	15.0	-0.3	0.26
16.31	23.24	23.45	16.1	-0.2	0.24

**Remark:**

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Velocity of standard

<sup>7</sup> Velocity of Unit Under Calibration

**PHOTO OF CALIBRATION SET-UP**

Calibration set-up of the cup anemometer calibration in the wind tunnel of Jiranatee Associates Co., Ltd. The cup anemometer shown may differ from the calibrated one. Remark: The proportion of the set-up is not true to scale due to imaging geometry.

\*\*\*End of Certificate of Calibration\*\*\*





JIRANATEE ASSOCIATES CO.,LTD.

Jiranatee Associates Co.,Ltd  
63/14-15, 67/35-36  
Petchkasem 7,7/1, Rd.Watthapra, Bangkokyai,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Air speed measurement laboratory  
Calibration services department.

REVIEW BY

Warakorn P.

APPROVED BY

Signature

NEXT CAL DATE

19/4/24

Certificate Number

CL-010-66

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

**MEASUREMENT ITEM** : Wind Direction Sensor  
**MANUFACTURER** : Novalynx  
**MODEL/TYPE** : Sensor: WS-02F  
Data logger: 200-WS-25DL  
**SERIAL NUMBER** : Sensor: -  
Data logger: A4986  
**ID NUMBER** : RYG\_FS0087  
**CONDITION AS-RECEIVED** : Used item  
**CUSTOMER** : ALS laboratory group (Thailand) Co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,  
Khet Suan Luang, Bangkok 10250 Thailand.

**RECEIVED DATE** : 16 Jan 2023  
**MEASUREMENT DATE** : 19 Jan 2023  
**ISSUE DATE** : 20 Jan 2023

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature	: 23.0 ± 3.0	°C
Relative Humidity	: 55.0 ± 15.0	%RH
Atmospheric Pressure	: 1010 ± 10	hPa

**PLACE OF CALIBRATION** : Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

<b>CALIBRATION CONDITION</b>	: Wind tunnel cross-section area <sup>1</sup>	900	cm <sup>2</sup>
	Win direction frontal area <sup>2</sup>	129	cm <sup>2</sup>
	Diameter of mounting pipe <sup>3</sup>	-	mm
	Blockage ratio of test object <sup>4</sup>	0.143	[-]

**Preconditioning** : 24 hours at ambient conditions.  
**Measurement Condition** : The average values during measurement are (23.5)°C, (47.4) %RH and (1015.6) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



Approved signatory:

Signature

Mr. Parinya Booncharoen  
Calibration Department Manager

### Remark:

<sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>

MEASUREMENT RESULTS <sup>5</sup>

The wind direction sensor was calibrated against standard rotary encoder by comparison method. During calibration, the measurement was carried out at 45° intervals in clockwise and counterclockwise directions after offset adjustment has been made. The flow speed of wind tunnel (usually 5 m/s) is kept constant while the sensor is rotated around its vertical axis. The results of calibration and associated measurement uncertainties are reported in the table below.

Air speed m/s	$D^{std}$ Degree (°)	$D^{uuc}$ Degree (°)	Error Degree (°)	$U (k=2)$ Degree (°)
5.02	0.000	0	0	0.58
	45.000	43	-2	0.74
	90.000	88	-2	0.74
	135.000	133	-2	0.74
	180.000	179	-1	0.74
	225.000	225	0	0.68
	270.000	273	3	0.58
	315.000	319	4	0.74

## Remark:

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Direction of standard

<sup>7</sup> Direction of Unit Under Calibration

\*\*\*End of Certificate of Calibration\*\*\*







JIRANATEE ASSOCIATES CO.,LTD.

Jiranatee Associates Co.,Ltd  
63/14-15, 67/35-36  
Petchkasem 7,7/1, Rd. Watthapra, Bangkokyai,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Air speed measurement laboratory  
Calibration services department.

Certificate Number

CL-010-66

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

**MEASUREMENT ITEM** : Cup anemometer  
**MANUFACTURER** : Novalynx  
**MODEL/TYPE** : Sensor: WS-02F  
Data logger: 200-WS-25DL  
**SERIAL NUMBER** : Sensor: -  
Data logger: A4986  
**ID NUMBER** : RYG\_FS0087  
**CONDITION AS-RECEIVED** : Used item  
**CUSTOMER** : ALS laboratory group (Thailand) co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,  
Khet Suan Luang, Bangkok 10250 Thailand.

**RECEIVED DATE** : 16 Jan 2023  
**MEASUREMENT DATE** : 18 Jan 2023  
**ISSUE DATE** : 20 Jan 2023

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature :  $23.0 \pm 3.0$  °C  
Relative Humidity :  $55.0 \pm 15.0$  %RH  
Atmospheric Pressure :  $1010 \pm 10$  hPa

**PLACE OF CALIBRATION** : Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

**CALIBRATION CONDITIONS** : Wind tunnel cross-section area<sup>1</sup> 900 cm<sup>2</sup>  
Win direction frontal area<sup>2</sup> 100 cm<sup>2</sup>  
Diameter of mounting pipe<sup>3</sup> - mm  
Blockage ratio of test object<sup>4</sup> 0.111 [-]

**Preconditioning** : 24 hours at ambient conditions.  
**Measurement Condition** : The average values during measurement are (23.6) °C, (55.3) %RH and (1013.5) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



### Approved signatory:

26/Jan

Mr. Parinya Booncharoen  
Calibration Department Manager

### Remark:

- <sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>



MEASUREMENT RESULTS<sup>5</sup>

The cup anemometer, Unit Under Calibration (UUC) was exercised at 10 m/s for 5 minutes prior to calibration being performed. The standard air velocity 0.5 m/s to 5 m/s was calculated by a standard air velocity transducer and above 5 m/s to 30 m/s was calculated by a pitot tube with precision differential pressure meter which was installed 40 mm and 300 mm respectively away from wind tunnel nozzle, UUC was installed at center of the test section. The calibration was carried out under both rising and falling air velocity in the range of 1 m/s to 16 m/s at calibration interval of 1 m/s. The results of calibration and associated measurement uncertainties are reported in the table below.

$v_{std}$ <sup>6</sup> (m/s)	Temp. wind tunnel (°C)	Temp. room (°C)	$v_{uuc}$ <sup>7</sup> (m/s)	Error (m/s)	$U$ ( $k=2$ ) (m/s)
0.985	23.68	23.60	0.8	-0.2	0.15
2.033	23.54	23.60	1.8	-0.2	0.16
3.046	23.68	23.60	2.9	-0.1	0.19
4.136	23.66	23.60	3.9	-0.2	0.20
5.03	23.50	23.60	4.9	-0.1	0.20
5.98	23.50	23.60	5.9	-0.1	0.18
7.05	23.36	23.60	7.0	-0.1	0.18
8.18	23.54	23.60	8.0	-0.2	0.20
9.10	23.30	23.60	8.9	-0.2	0.20
10.10	23.50	23.60	10.0	-0.1	0.19
11.14	23.28	23.60	11.1	-0.1	0.22
12.12	23.40	23.60	11.9	-0.2	0.21
13.19	23.10	23.60	13.0	-0.2	0.26
14.25	23.46	23.60	14.0	-0.2	0.32
15.26	23.10	23.60	15.0	-0.2	0.23
16.31	23.26	23.60	16.2	-0.1	0.29

## Remark:

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Velocity of standard

<sup>7</sup> Velocity of Unit Under Calibration

## PHOTO OF CALIBRATION SET-UP



Calibration set-up of the cup anemometer calibration in the wind tunnel of Jiranatee Associates Co., Ltd. The cup anemometer shown may differ from the calibrated one. Remark: The proportion of the set-up is not true to scale due to imaging geometry.

\*\*\*End of Certificate of Calibration\*\*\*



JIRANATEE ASSOCIATES CO., LTD.

Jiranatee Associates Co., Ltd.  
63/14-15, 67/35-36  
Petchkasem 7,7/1, Rd. Watthapra, Bangkokyai,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Air speed measurement laboratory  
Calibration services department.

REVIEW BY	Wanakorn P.
APPROVED BY	[Signature]
NEXT CAL. DATE	10/8/24

Certificate Number

CL-021-66

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

### MEASUREMENT ITEM

### MANUFACTURER

### MODEL/TYPE

### SERIAL NUMBER

### ID NUMBER

### CONDITION AS-RECEIVED

### CUSTOMER

: Cup anemometer  
: Novalynx  
: Sensor: WS-02F  
Data logger: 200-WS-25LB  
: Sensor: -  
Data logger: A5376  
: RYG\_FS0414  
: Used item  
: ALS laboratory group (Thailand) Co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,  
Khet Suan Luang, Bangkok 10250 Thailand.

### RECEIVED DATE

### MEASUREMENT DATE

### ISSUE DATE

: 27 Jan 2023  
: 10 Feb 2023  
: 10 Feb 2023

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature	: 23.0 ± 3.0	°C
Relative Humidity	: 55.0 ± 15.0	%RH
Atmospheric Pressure	: 1010 ± 10	hPa

### PLACE OF CALIBRATION

: Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

### CALIBRATION CONDITIONS

: Wind tunnel cross-section area <sup>1</sup>	900	cm <sup>2</sup>
: Win direction frontal area <sup>2</sup>	100	cm <sup>2</sup>
: Diameter of mounting pipe <sup>3</sup>	-	mm
: Blockage ratio of test object <sup>4</sup>	0.111	[-]

### Preconditioning

### Measurement Condition

: 24 hours at ambient conditions.  
: The average values during measurement are (24.1) °C, (47.6) %RH and (1014.7) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



### Approved signatory:

[Signature]

Mr. Parinya Booncharoen  
Calibration Department Manager

### Remark:

- <sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>



**MEASUREMENT RESULTS <sup>5</sup>**

The cup anemometer, Unit Under Calibration (UUC) was exercised at 10 m/s for 5 minutes prior to calibration being performed. The standard air velocity 0.5 m/s to 5 m/s was calculated by a standard air velocity transducer and above 5 m/s to 30 m/s was calculated by a pitot tube with precision differential pressure meter which was installed 40 mm and 300 mm respectively away from wind tunnel nozzle, UUC was installed at center of the test section. The calibration was carried out under both rising and falling air velocity in the range of 1 m/s to 16 m/s at calibration interval of 1 m/s. The results of calibration and associated measurement uncertainties are reported in the table below.


$V_{std}$ <sup>6</sup> (m/s)	Temp. wind tunnel (°C)	Temp. room (°C)	$V_{UUC}$ <sup>7</sup> (m/s)	Error (m/s)	$U (k=2)$ (m/s)
0.985	24.10	24.10	0.8	-0.2	0.36
2.033	24.10	24.10	1.9	-0.1	0.16
3.040	23.94	24.10	3.0	-0.1	0.23
4.134	24.10	24.10	4.0	-0.1	0.20
4.99	23.92	24.10	4.9	-0.1	0.44
5.98	24.10	24.10	6.0	0.0	0.18
7.05	23.90	24.10	7.0	-0.1	0.36
8.19	24.06	24.10	8.2	0.0	0.26
9.09	23.84	24.10	9.1	0.0	0.24
10.09	23.92	24.10	10.1	0.0	0.28
11.15	23.80	24.10	11.1	0.0	0.45
12.14	23.80	24.10	12.2	0.0	0.31
13.19	23.80	24.10	13.2	0.0	0.47
14.26	23.74	24.10	14.2	0.0	0.42
15.25	23.78	24.10	15.1	-0.1	0.66
16.28	23.70	24.10	16.3	0.0	0.56

**Remark:**

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Velocity of standard

<sup>7</sup> Velocity of Unit Under Calibration

**PHOTO OF CALIBRATION SET-UP**

Calibration set-up of the cup anemometer calibration in the wind tunnel of Jiranatee Associates Co., Ltd. The cup anemometer shown may differ from the calibrated one. Remark: The proportion of the set-up is not true to scale due to imaging geometry.

\*\*\*End of Certificate of Calibration\*\*\*

**NAC**  
JIRANATE ASSOCIATES CO., LTD.





JIRANATEE ASSOCIATES CO.,LTD.

Jiranatee Associates Co.,Ltd.  
63/14-15, 67/35-36  
Petchkasem 7,7/1, Rd. Watthapra, Bangkokyai,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Air speed measurement laboratory  
Calibration services department.

Certificate Number

CL-019-66

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

**MEASUREMENT ITEM** : Wind Direction Sensor  
**MANUFACTURER** : Novalynx  
**MODEL/TYPE** : Sensor: WS-02F  
Data logger: 200-WS-25LB  
**SERIAL NUMBER** : Sensor: -  
Data logger: A5376  
**ID NUMBER** : RYG\_FS0414  
**CONDITION AS-RECEIVED** : Used item  
**CUSTOMER** : ALS laboratory group (Thailand) Co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,  
Khet Suan Luang, Bangkok 10250 Thailand.

**RECEIVED DATE** : 27 Jan 2023  
**MEASUREMENT DATE** : 10 Feb 2023  
**ISSUE DATE** : 10 Feb 2023

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature :  $23.0 \pm 3.0$  °C  
Relative Humidity :  $55.0 \pm 15.0$  %RH  
Atmospheric Pressure :  $1010 \pm 10$  hPa

**PLACE OF CALIBRATION** : Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

**CALIBRATION CONDITION** : Wind tunnel cross-section area<sup>1</sup> 900 cm<sup>2</sup>  
Win direction frontal area<sup>2</sup> 129 cm<sup>2</sup>  
Diameter of mounting pipe<sup>3</sup> - mm  
Blockage ratio of test object<sup>4</sup> 0.143 [-]

**Preconditioning** : 24 hours at ambient conditions.  
**Measurement Condition** : The average values during measurement are (24.0)°C, (49.0) %RH and (1014.1) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



Approved signatory: .....

Mr. Parinya Booncharoen  
Calibration Department Manager

### Remark:

- <sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>

**MEASUREMENT RESULTS<sup>5</sup>**

The wind direction sensor was calibrated against standard rotary encoder by comparison method. During calibration, the measurement was carried out at 45° intervals in clockwise and counterclockwise directions after offset adjustment has been made. The flow speed of wind tunnel (usually 5 m/s) is kept constant while the sensor is rotated around its vertical axis. The results of calibration and associated measurement uncertainties are reported in the table below.

Air speed m/s	$D^{\circ}_{std}$ Degree (°)	$D^{\circ}_{uuc}$ Degree (°)	Error Degree (°)	$U (k=2)$ Degree (°)
4.99	0.000	0	0	0.58
	45.000	41	-4	0.58
	90.000	87	-3	0.58
	135.001	132	-3	0.68
	180.000	179	-1	0.74
	225.000	227	2	0.91
	270.001	273	3	0.58
	315.000	318	3	0.74

**Remark:**

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Direction of standard

<sup>7</sup> Direction of Unit Under Calibration



\*\*\*End of Certificate of Calibration\*\*\*



Certificate Number
CL-001-66

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

MEASUREMENT ITEM	: Cup anemometer
MANUFACTURER	: Novalynx
MODEL/TYPE	: Sensor: WS-02F Data logger: WS-25DL
SERIAL NUMBER	: Sensor: - Data logger: A4481
ID NUMBER	: BKK_FS0141
CONDITION AS-RECEIVED	: Used item
CUSTOMER	: ALS laboratory group (Thailand) Co., Ltd. 104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand.

RECEIVED DATE	: 28 Dec 2022
MEASUREMENT DATE	: 05 Jan 2023
ISSUE DATE	: 09 Jan 2023

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature	: 23.0 ± 3.0	°C
Relative Humidity	: 55.0 ± 15.0	%RH
Atmospheric Pressure	: 1010 ± 10	hPa

PLACE OF CALIBRATION	: Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.
----------------------	---

CALIBRATION CONDITIONS	: Wind tunnel cross-section area <sup>1</sup>	900	cm <sup>2</sup>
	Win direction frontal area <sup>2</sup>	100	cm <sup>2</sup>
	Diameter of mounting pipe <sup>3</sup>	-	mm
	Blockage ratio of test object <sup>4</sup>	0.111	[-]

Preconditioning	: 24 hours at ambient conditions.
Measurement Condition	: The average values during measurement are (23.9) °C, (47.3) %RH and (1015.0) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

☒ Mr. Sorawit Thachalad  
☐ Miss Jittraporn Lertsomphol



Approved signatory: .....

*[Signature]*  
Mr. Parinya Booncharoen  
Calibration Department Manager

### Remark:

<sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>



MEASUREMENT RESULTS<sup>5</sup>

The cup anemometer, Unit Under Calibration (UUC) was exercised at 10 m/s for 5 minutes prior to calibration being performed. The standard air velocity 0.5 m/s to 5 m/s was calculated by a standard air velocity transducer and above 5 m/s to 30 m/s was calculated by a pitot tube with precision differential pressure meter which was installed 40 mm and 300 mm respectively away from wind tunnel nozzle, UUC was installed at center of the test section. The calibration was carried out under both rising and falling air velocity in the range of 1 m/s to 16 m/s at calibration interval of 1 m/s. The results of calibration and associated measurement uncertainties are reported in the table below.

$V_{std}$ <sup>6</sup> (m/s)	Temp. wind tunnel (°C)	Temp. room (°C)	$V_{uuc}$ <sup>7</sup> (m/s)	Error (m/s)	$U (k=2)$ (m/s)
0.989	23.82	23.85	0.7	-0.3	0.16
2.031	23.90	23.85	1.7	-0.3	0.16
3.051	24.00	23.85	2.9	-0.2	0.20
4.132	23.84	23.85	3.9	-0.2	0.20
5.00	23.88	23.85	4.9	-0.1	0.24
5.98	23.94	23.85	5.8	-0.2	0.18
7.06	23.82	23.85	6.9	-0.2	0.19
8.17	23.90	23.85	8.0	-0.1	0.22
9.08	23.72	23.85	9.0	-0.1	0.21
10.09	23.86	23.85	9.9	-0.2	0.20
11.14	23.60	23.85	11.0	-0.1	0.26
12.14	23.74	23.85	12.1	-0.1	0.28
13.21	23.68	23.85	13.0	-0.2	0.21
14.28	23.70	23.85	14.1	-0.2	0.27
15.26	23.64	23.85	15.0	-0.3	0.26
16.30	23.60	23.85	16.1	-0.2	0.28

## Remark:

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Velocity of standard

<sup>7</sup> Velocity of Unit Under Calibration

## PHOTO OF CALIBRATION SET-UP



Calibration set-up of the cup anemometer calibration in the wind tunnel of Jiranatee Associates Co., Ltd. The cup anemometer shown may differ from the calibrated one. Remark: The proportion of the set-up is not true to scale due to imaging geometry.

\*\*\*End of Certificate of Calibration\*\*\*



## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

**MEASUREMENT ITEM** : Wind Direction Sensor  
**MANUFACTURER** : Novalynx  
**MODEL/TYPE** : Sensor: WS-02F  
Data logger: WS-25DL  
**SERIAL NUMBER** : Sensor: -  
Data logger: A4481  
**ID NUMBER** : BKK\_FS0141  
**CONDITION AS-RECEIVED** : Used item  
**CUSTOMER** : ALS laboratory group (Thailand) Co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,  
Khet Suan Luang, Bangkok 10250 Thailand.

**RECEIVED DATE** : 28 Dec 2022  
**MEASUREMENT DATE** : 06 Jan 2023  
**ISSUE DATE** : 09 Jan 2023

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature :  $23.0 \pm 3.0$  °C  
Relative Humidity :  $55.0 \pm 15.0$  %RH  
Atmospheric Pressure :  $1010 \pm 10$  hPa

**PLACE OF CALIBRATION** : Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

**CALIBRATION CONDITION** : Wind tunnel cross-section area<sup>1</sup> 900 cm<sup>2</sup>  
Win direction frontal area<sup>2</sup> 129 cm<sup>2</sup>  
Diameter of mounting pipe<sup>3</sup> - mm  
Blockage ratio of test object<sup>4</sup> 0.143 [-]

**Preconditioning** : 24 hours at ambient conditions.  
**Measurement Condition** : The average values during measurement are (23.5)°C, (48.8) %RH and (1015.8) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

☒ Mr. Sorawit Thachalad  
☐ Miss Jittraporn Lertsomphol



Approved signatory: .....

Mr. Parinya Booncharoen  
Calibration Department Manager

### Remark:

<sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>

MEASUREMENT RESULTS<sup>5</sup>

The wind direction sensor was calibrated against standard rotary encoder by comparison method. During calibration, the measurement was carried out at 45° intervals in clockwise and counterclockwise directions after offset adjustment has been made. The flow speed of wind tunnel (usually 5 m/s) is kept constant while the sensor is rotated around its vertical axis. The results of calibration and associated measurement uncertainties are reported in the table below.

Air speed m/s	$D_{std}^6$ Degree (°)	$D_{UUC}^7$ Degree (°)	Error Degree (°)	$U (k=2)$ Degree (°)
5.02	0.000	0	0	0.58
	45.000	41	-4	0.74
	90.000	87	-3	0.68
	135.000	134	-1	0.74
	180.001	181	1	0.74
	225.000	228	3	0.74
	270.001	273	3	0.74
	315.000	318	3	0.68

## Remark:

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Direction of standard

<sup>7</sup> Direction of Unit Under Calibration



\*\*\*End of Certificate of Calibration\*\*\*





JIRANATEE ASSOCIATES CO.,LTD.

Jiranatee Associates Co.,Ltd.  
63/14-15, 67/35-36  
Petchkasem 7,7/1, Rd. Watthapra, Bangkokyai,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Air speed measurement laboratory  
Calibration services department.

REVIEW BY	<i>Minakorn P.</i>
APPROVED BY	<i>[Signature]</i>
NEXT CAL. DATE	10/8/24

Certificate Number

CL-018-66

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

**MEASUREMENT ITEM** : Cup anemometer  
**MANUFACTURER** : Novalynx  
**MODEL/TYPE** : Sensor: WS-02F  
Data logger: 200-WS-25LB  
**SERIAL NUMBER** : Sensor: -  
Data logger: A5369  
**ID NUMBER** : RYG\_FS0411  
**CONDITION AS-RECEIVED** : Used item  
**CUSTOMER** : ALS laboratory group (Thailand) Co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,  
Khet Suan Luang, Bangkok 10250 Thailand.

**RECEIVED DATE** : 27 Jan 2023  
**MEASUREMENT DATE** : 10 Feb 2023  
**ISSUE DATE** : 10 Feb 2023

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature	: 23.0 ± 3.0	°C
Relative Humidity	: 55.0 ± 15.0	%RH
Atmospheric Pressure	: 1010 ± 10	hPa

**PLACE OF CALIBRATION** : Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

<b>CALIBRATION CONDITIONS</b>	: Wind tunnel cross-section area <sup>1</sup>	900	cm <sup>2</sup>
	Win direction frontal area <sup>2</sup>	100	cm <sup>2</sup>
	Diameter of mounting pipe <sup>3</sup>	-	mm
	Blockage ratio of test object <sup>4</sup>	0.111	[-]

**Preconditioning** : 24 hours at ambient conditions.  
**Measurement Condition** : The average values during measurement are (23.8) °C, (44.8) %RH and (1010.3) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



### Approved signatory:

*[Signature]*

Mr. Parinya Booncharoen  
Calibration Department Manager

### Remark:

- <sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>

**MEASUREMENT RESULTS <sup>5</sup>**

The cup anemometer, Unit Under Calibration (UUC) was exercise at 10 m/s for 5 minutes prior to calibration being performed. The standard air velocity 0.5 m/s to 5 m/s was calculated by a standard air velocity transducer and above 5 m/s to 30 m/s was calculated by a pitot tube with precision differential pressure meter which was installed 40 mm and 300 mm respectively away from wind tunnel nozzle, UUC was installed at center of the test section. The calibration was carried out under both rising and falling air velocity in the range of 1 m/s to 16 m/s at calibration interval of 1 m/s. The results of calibration and associated measurement uncertainties are reported in the table below.

$V_{std}$ <sup>6</sup> (m/s)	Temp. wind tunnel (°C)	Temp. room (°C)	$V_{UUC}$ <sup>7</sup> (m/s)	Error (m/s)	$U (k=2)$ (m/s)
0.983	23.60	23.55	0.8	-0.2	0.15
2.038	23.50	23.55	1.8	-0.2	0.16
3.044	23.50	23.55	2.9	-0.2	0.18
4.147	23.58	23.55	3.9	-0.3	0.19
5.00	23.50	23.55	4.9	-0.1	0.18
5.98	23.62	23.55	5.9	-0.1	0.18
7.04	23.28	23.55	7.0	-0.1	0.18
8.16	23.56	23.55	8.0	-0.2	0.19
9.10	23.26	23.55	9.0	-0.1	0.19
10.07	23.50	23.55	10.0	-0.1	0.19
11.13	23.10	23.55	11.0	-0.2	0.20
12.13	23.50	23.55	12.1	-0.1	0.30
13.21	23.12	23.55	13.1	-0.1	0.22
14.25	23.36	23.55	14.0	-0.2	0.27
15.24	23.10	23.55	15.1	-0.2	0.28
16.29	23.20	23.55	16.0	-0.3	0.24

**Remark:**

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Velocity of standard

<sup>7</sup> Velocity of Unit Under Calibration

**PHOTO OF CALIBRATION SET-UP**

Calibration set-up of the cup anemometer calibration in the wind tunnel of Jiranatee Associates Co., Ltd. The cup anemometer shown may differ from the calibrated one. Remark: The proportion of the set-up is not true to scale due to imaging geometry.

\*\*\*End of Certificate of Calibration\*\*\*  
JIRANATE ASSOCIATES CO., LTD.





JIRANATEE ASSOCIATES CO.,LTD.

Jiranatee Associates Co.,Ltd  
63/14-15, 67/35-36  
Petchkasem 7,7/1, Rd. Watthapra, Bangkokyai,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Air speed measurement laboratory  
Calibration services department.

Certificate Number

CL-016-66

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

**MEASUREMENT ITEM** : Wind Direction Sensor  
**MANUFACTURER** : Novalynx  
**MODEL/TYPE** : Sensor: WS-02F  
Data logger: 200-WS-25LB  
**SERIAL NUMBER** : Sensor: -  
Data logger: A5369  
**ID NUMBER** : RYG\_FS0411  
**CONDITION AS-RECEIVED** : Used item  
**CUSTOMER** : ALS laboratory group (Thailand) Co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,  
Khet Suan Luang, Bangkok 10250 Thailand.

**RECEIVED DATE** : 27 Jan 2023  
**MEASUREMENT DATE** : 10 Feb 2023  
**ISSUE DATE** : 10 Feb 2023

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature	: 23.0 ± 3.0	°C
Relative Humidity	: 55.0 ± 15.0	%RH
Atmospheric Pressure	: 1010 ± 10	hPa

**PLACE OF CALIBRATION** : Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

<b>CALIBRATION CONDITION</b>	Wind tunnel cross-section area <sup>1</sup>	900	cm <sup>2</sup>
	Win direction frontal area <sup>2</sup>	129	cm <sup>2</sup>
	Diameter of mounting pipe <sup>3</sup>	-	mm
	Blockage ratio of test object <sup>4</sup>	0.143	[-]

**Preconditioning** : 24 hours at ambient conditions.  
**Measurement Condition** : The average values during measurement are (23.9)°C, (48.3) %RH and (1011.4) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



Approved signatory: .....

Mr. Parinya Booncharoen  
Calibration Department Manager

### Remark:

- <sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>



MEASUREMENT RESULTS<sup>5</sup>

The wind direction sensor was calibrated against standard rotary encoder by comparison method. During calibration, the measurement was carried out at 45° intervals in clockwise and counterclockwise directions after offset adjustment has been made. The flow speed of wind tunnel (usually 5 m/s) is kept constant while the sensor is rotated around its vertical axis. The results of calibration and associated measurement uncertainties are reported in the table below.

Air speed m/s	$D^s_{std}$ Degree (°)	$D^r_{unc}$ Degree (°)	Error Degree (°)	$U (k=2)$ Degree (°)
5.00	0.000	0	0	0.58
	45.000	41	-4	0.58
	90.000	87	-3	0.58
	135.000	135	0	0.68
	180.000	182	2	0.74
	225.000	230	5	0.68
	270.001	275	5	0.58
	315.000	320	5	0.58

## Remark:

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Direction of standard

<sup>7</sup> Direction of Unit Under Calibration



\*\*\*End of Certificate of Calibration\*\*\*



JIRANATEE ASSOCIATES CO.,LTD.

Jiranatee Associates Co.,Ltd.  
63/14-15, 67/35-36  
Petchkasem 7,7/1, Rd. Watthapra, Bangkokyai,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Air speed measurement laboratory  
Calibration services department.

REVIEW BY	Narakorn P.
APPROVED BY	[Signature]
NEXT CAL. DATE	11/4/24

Certificate Number

CL-005-66

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

**MEASUREMENT ITEM** : Cup anemometer  
**MANUFACTURER** : Novalynx  
**MODEL/TYPE** : Sensor: WS-02F  
Data logger: 110-WS-25DL-N  
**SERIAL NUMBER** : Sensor: WSD-007  
Data logger: A5486  
**ID NUMBER** : NKH\_FS0053  
**CONDITION AS-RECEIVED** : Used item  
**CUSTOMER** : ALS laboratory group (Thailand) co., ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,  
Khet Suan Luang, Bangkok 10250 Thailand.

**RECEIVED DATE** : 09 Jan 2023  
**MEASUREMENT DATE** : 11 Jan 2023  
**ISSUE DATE** : 13 Jan 2023

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature	: 23.0 ± 3.0	°C
Relative Humidity	: 55.0 ± 15.0	%RH
Atmospheric Pressure	: 1010 ± 10	hPa

**PLACE OF CALIBRATION** : Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

<b>CALIBRATION CONDITIONS</b>	: Wind tunnel cross-section area <sup>1</sup>	900	cm <sup>2</sup>
	Win direction frontal area <sup>2</sup>	100	cm <sup>2</sup>
	Diameter of mounting pipe <sup>3</sup>	-	mm
	Blockage ratio of test object <sup>4</sup>	0.111	[-]

**Preconditioning** : 24 hours at ambient conditions.  
**Measurement Condition** : The average values during measurement are (23.6) °C, (48.9) %RH and (1015.0) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

☒ Mr. Sorawit Thachalad  
☐ Miss Jittrapon Lertsomphol



Approved signatory:

Mr. Parinya Booncharoen  
Calibration Department Manager

### Remark:

- <sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>

**MEASUREMENT RESULTS <sup>5</sup>**

The cup anemometer, Unit Under Calibration (UUC) was exercised at 10 m/s for 5 minutes prior to calibration being performed. The standard air velocity 0.5 m/s to 5 m/s was calculated by a standard air velocity transducer and above 5 m/s to 30 m/s was calculated by a pitot tube with precision differential pressure meter which was installed 40 mm and 300 mm respectively away from wind tunnel nozzle, UUC was installed at center of the test section. The calibration was carried out under both rising and falling air velocity in the range of 1 m/s to 16 m/s at calibration interval of 1 m/s. The results of calibration and associated measurement uncertainties are reported in the table below.


$V_{std}^6$ (m/s)	Temp. wind tunnel (°C)	Temp. room (°C)	$V_{uuc}^7$ (m/s)	Error (m/s)	$U (k=2)$ (m/s)
0.981	23.70	23.60	0.8	-0.2	0.15
2.030	23.56	23.60	1.9	-0.2	0.16
3.049	23.56	23.60	2.9	-0.2	0.18
4.127	23.50	23.60	4.0	-0.2	0.20
4.99	23.52	23.60	4.9	-0.1	0.20
5.98	23.64	23.60	5.9	-0.1	0.18
7.06	23.50	23.60	6.9	-0.2	0.18
8.18	23.54	23.60	8.1	-0.1	0.21
9.09	23.32	23.60	9.0	-0.1	0.20
10.08	23.60	23.60	10.0	-0.1	0.22
11.16	23.32	23.60	11.0	-0.1	0.22
12.15	23.60	23.60	12.0	-0.2	0.23
13.21	23.40	23.60	13.1	-0.1	0.23
14.26	23.60	23.60	14.2	-0.1	0.27
15.25	23.50	23.60	15.2	0.0	0.23
16.23	23.52	23.60	16.1	-0.1	0.55

**Remark:**

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Velocity of standard

<sup>7</sup> Velocity of Unit Under Calibration

**PHOTO OF CALIBRATION SET-UP**

Calibration set-up of the cup anemometer calibration in the wind tunnel of Jiranatee Associates Co., Ltd. The cup anemometer shown may differ from the calibrated one. Remark: The proportion of the set-up is not true to scale due to imaging geometry.





JIRANATEE ASSOCIATES CO.,LTD.

Jiranatee Associates Co.,Ltd.  
63/14-15, 67/35-36  
Petchkasem 7,7/1, Rd. Watthapra, Bangkokyai,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Air speed measurement laboratory  
Calibration services department.

Certificate Number

CL-005-66

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

**MEASUREMENT ITEM** : Wind Direction Sensor  
**MANUFACTURER** : Novalynx  
**MODEL/TYPE** : Sensor: WS-02F  
Data logger: 110-WS-25DL-N  
**SERIAL NUMBER** : Sensor: WSD-007  
Data logger: A5486  
**ID NUMBER** : NKH\_FS0053  
**CONDITION AS-RECEIVED** : Used item  
**CUSTOMER** : ALS laboratory group (Thailand) Co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,  
Khet Suan Luang, Bangkok 10250 Thailand.

**RECEIVED DATE** : 09 Jan 2023  
**MEASUREMENT DATE** : 11 Jan 2023  
**ISSUE DATE** : 13 Jan 2023

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature :  $23.0 \pm 3.0$  °C  
Relative Humidity :  $55.0 \pm 15.0$  %RH  
Atmospheric Pressure :  $1010 \pm 10$  hPa

**PLACE OF CALIBRATION** : Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

**CALIBRATION CONDITION** : Wind tunnel cross-section area<sup>1</sup> 900 cm<sup>2</sup>  
Win direction frontal area<sup>2</sup> 129 cm<sup>2</sup>  
Diameter of mounting pipe<sup>3</sup> - mm  
Blockage ratio of test object<sup>4</sup> 0.143 [-]

**Preconditioning** : 24 hours at ambient conditions.  
**Measurement Condition** : The average values during measurement are (23.3)°C, (53.3) %RH and (1011.2) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



Approved signatory: \_\_\_\_\_

*2s/aw*

Mr. Parinya Booncharoen  
Calibration Department Manager

### Remark:

- <sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>

MEASUREMENT RESULTS<sup>5</sup>

The wind direction sensor was calibrated against standard rotary encoder by comparison method. During calibration, the measurement was carried out at 45° intervals in clockwise and counterclockwise directions after offset adjustment has been made. The flow speed of wind tunnel (usually 5 m/s) is kept constant while the sensor is rotated around its vertical axis. The results of calibration and associated measurement uncertainties are reported in the table below.

Air speed m/s	$D^{\circ}_{std}$ Degree (°)	$D^{\circ}_{uuc}$ Degree (°)	Error Degree (°)	$U (k=2)$ Degree (°)
5.01	0.000	0	0	0.58
	45.000	42	-3	0.68
	90.000	88	-2	0.74
	135.000	133	-2	0.74
	180.000	180	-1	0.76
	225.000	225	0	0.74
	270.000	272	2	0.74
	315.000	317	2	0.68

## Remark:

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Direction of standard

<sup>7</sup> Direction of Unit Under Calibration

\*\*\*End of Certificate of Calibration\*\*\*





## CERTIFICATE OF CALIBRATION

Calibration No. : RH-03012023

Page 1 of 1 Pages

Measurement Item : Relative humidity with data logger  
Manufacturer : Novalynx  
Model/Type : 110-WS-25DL-N  
Serial Number : A5486  
ID No. : NKH\_FS0053  
Customer : ALS laboratory group (Thailand) Co., Ltd.  
: 104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Suan Luang, Khet Suan Luang, Bangkok  
10250 Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(25 \pm 3)^{\circ}\text{C}$ , and relative humidity of  $(50 \pm 15)\%$ .

### Measurement Method:

Unit Under Calibration (UUC) was calibrated by comparison method with standard thermo hygrometer in the humidity generator chamber to determine the errors.

### Traceability:

This instrument was calibrated using standard equipment whose accuracy is traceability through National Institute of Standards and Technology to the international system of units (SI) via MCS Calibration, Inc. Certificate number: 20314-101. Due date: Mar 14, 2023.

Measurement Date : Jan 13, 2023

Issued Date : Jan 13, 2023

### Measurement Results:

This equipment was connected with Indoor air quality probe and Displayed (UR) on display. Model: HMP60, Serial number: R3140638.

Calibration was performed in the range of 20%RH to 80%RH

The results of calibration are reported in table below.

Determined (%RH)	Standard (Reading) (%RH)	UUC (Reading) (%RH)	Error (%RH)	Uncertainty $\pm$ (%RH)
20	20.04	18.3	-1.7	0.69
50	50.29	48.5	-1.8	0.69
80	80.12	79.5	-0.6	0.71

### Performed by

- ☐ Mr. Sorawit Thachalad  
☒ Miss Jittraporn Lertsomphol



### Approved Signatory:

*215/MS*

Mr. Parinya Booncharoen.  
Calibration Department Manager



## CERTIFICATE OF CALIBRATION

Certificate No.: CL-002-66

Page 1 of 2

Equipment Name: Data Logger with Temperature  
Sensor

Manufacturer: Novalynx

Model: 110-WS-25DL-N

Serial No.: A5486

ID No.: NKH\_FS0053

### Customer

Name: ALS laboratory group (Thailand) Co., Ltd.  
Address: 104 Phatthanakan 40, Phatthanakan Rd.,  
Khwaeng Suan Luang, Khet Suan Luang, Bangkok  
10250 Thailand.

Received date: 09 Jan 2023

Calibration date: 13 Jan 2023

Issue date: 13 Jan 2023

### Reference Used During Calibration

1. Standard Temperature Probe Model: STS-100 A500,  
Serial No.: 667682-09, Due date: 23 Mar 2023

2. Digital Temperature Indicator Model: DTI-1000-A MK  
II, Serial No.: 671407-00591 Due date: 22 July 2023

### Calibration Condition

Temperature:  $(23 \pm 3)^{\circ}\text{C}$

Relative Humidity:  $(55 \pm 15)\%$

### Calibration Procedure

The temperature calibration was done by In-House  
calibration method as WI-CL-001 according to  
comparison method with standard digital temperature  
indicator and standard temperature probe. The  
temperature scale use was based on ITS-90.

### Traceability

The measurement results are traceable to the  
international system of units (SI) through National  
Institute of Metrology Thailand (NIMT) Certificate  
number: TT-0034-22, Certificate number: ER-0092-  
22

### Calibrated by

- ☐ Mr. Sorawit Thachalad  
☒ Miss Jitraporn Lertsomphol



### Approved Signatory: .....

Mr. Parinya Booncharoen  
Calibration Department Manager

**Result of Calibration:-** ☒ Without Adjustment ☐ With Adjustment

**Calibration Range:** 20-40 °C

**Function:**

This equipment was connected with temperature sensor Model: HMP60 S/N: R3140638.

Dimension : Diameter 12 mm. Length 80 mm.

<u>Immersion Depth (mm)</u>	<u>Standard Reading (°C)</u>	<u>UUC Reading (°C)</u>	<u>Error (°C)</u>	<u>Uncertainty (°C)</u>
60	20.067	19.9	-0.2	0.099
60	25.058	24.8	-0.3	0.099
60	30.052	29.7	-0.4	0.099
60	35.047	34.6	-0.4	0.099
60	40.038	39.5	-0.5	0.099

UUC\*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

\* End of Certificate \*



## CALIBRATION REPORT

Calibration Number. : RG-01012023

Page 1 of 2 Pages

Measurement Item : Rain gauge with data logger.

Manufacturer : Data logger: Novalynx.  
: Rain gauge: Novalynx.

Model/Type : Data logger: 110-WS-25DL-N  
: Rain gauge: 110-WS-25RG

Serial Number : Data logger: A5486  
: Rain gauge: RG-006

ID NO : NKH\_PS0053

Customer : ALS laboratory group (Thailand) co., ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang, Khet Suan Luang,  
Bangkok 10250, Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(25 \pm 3)^{\circ}\text{C}$ , and relative humidity of  $(50 \pm 15)\%$ .

### Measurement Method:

The Rain gauge, Unit Under Calibration (UUC) was calibrated by Precision reference bottle with flow adjuster at low rate 0.6 mm per minute or 1 tipping every 20 seconds. The tipping number was determined by procedures below.

1. Obtain rain gauge inlet area:  
Rain gauge precise diameter in cm = Diameter/2 = R (radius)  
Rain gauge area =  $R^2 \times 3.14$  (UUC diameter = 20.3 cm, UUC radius = 10.15 cm)  
Rain gauge area =  $323.6 \text{ cm}^2$ .
2. Obtain theoretical correct rain gauge answer (number of tipplings) using  $323.6 \text{ cm}^2$  inlet area and 0.5 L of rain.
  - a)  $10,000 \text{ cm}^3 / 323.6 \text{ cm}^2$  inlet area = 30.90 (rain gauge area = 1/30.90 of square meter)
  - b)  $30.90 \times 0.5 \text{ L volume} = 15.45 \text{ mm}$  (mm of rain over  $1 \text{ m}^2$  surface) 500 ml of rain volume on the rain gauge area = 15.45 mm of rain.
  - c) Number of tipping =  $15.45 / 0.25 \text{ mm} = 62$  tipplings.

Note: Rain gauge is fully cleaned and leveling prior the calibration performed.

Measurement Date : Jan 13, 2023  
Issued Date : Jan 13, 2023

Performed by

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



Approved Signatory: \_\_\_\_\_

Mr. Parinya Booncharoen.  
Calibration Department Manager



Continuation of Calibration of Calibration Number

Calibration Number: RG-01012023

Page 2 of 2 Pages

Result of Calibration: ☒ Without Adjustment ☐ With Adjustment.

The results of calibration are reported in table below.

Quantity of H <sub>2</sub> O (ml)	Determined Tipping	Tipping count	Acceptable Tipping count
500	62	62	60 - 64
500	62	61	60 - 64
500	62	62	60 - 64
500	62	62	60 - 64
500	62	61	60 - 64

*Remark: The procedure is made to verify the correct reading of the Unit under Calibration rain gauge when a precise volume of water falls into its cone. We suggest that the number of tipping should be within  $\pm 2\%$  different from the 62 tipping (correct range: 60-64 tipping) it means that the rain gauge meets the manufacturer acceptable limit.*

\*\*\*End of calibration report\*\*\*





JIRANATEE ASSOCIATES CO.,LTD.

Jiranatee Associates Co.,Ltd.  
63/14-15, 67/35-36  
Petchkasem 7,7/1, Rd. Watthapra, Bangkokyai,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Pressure measurement laboratory  
Calibration services department.



## CERTIFICATE OF CALIBRATION

Certificate No. : CL-003-66

Page 1 of 2 Pages

**MEASUREMENT ITEM** : Digital barometer  
**MANUFACTURER** : Novalynx  
**MODEL/TYPE** : 110-WS-25BP  
**SERIAL NUMBER** : A5486  
**ID NUMBER** : NKH\_FS0053  
**CONDITION AS-RECEIVED** : Used item  
**CUSTOMER** : ALS laboratory group (Thailand) Co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd,  
Khwaeng Suan Luang, Khet Suan Luang,  
Bangkok 10250 Thailand.  
**RECEIVED DATE** : 09 Jan 2023  
**MEASUREMENT DATE** : 12 Jan 2023  
**ISSUE DATE** : 13 Jan 2023

### Calibration procedure:

The pressure calibration was done by in-house calibration method as WI-CL-003 according to comparison method with Digital pressure calibrator based on DKD-R 6-1

### Traceability:

The measurement results are traceable to the international system of units (SI) through the NIMT (National Metrology Institute of Thailand) which complies with the requirements of ISO/IEC 17025:2017, ANSI/NCSL Z540-1 via Certificate number: MP-0205-22

The reported uncertainty is based on a standard uncertainty multiplied by a coverage  $k=2$ , providing a level of confidence of approximately 95%.

### CONDITION OF THIS RESULT OF CALIBRATION:

#### 1. Reference Standard Instrument:

Instrument	Model	Serial No.	Certificate No.	Due Date
Absolute Pressure Transducer	CPG2500	4100126P	MP-0205-22	02 Dec 2023

#### 1. Calibration effort for calibration sequence A

2. The UUC\* was installed in vertical orientation above reference standard instrument and center of UUC\* was used as the reference level.

#### 3. Calibration conditions:

4. Condition : ☒ Normal ☐ Abnormal  
Pressure transmitting medium : Air  
 $\rho_{Fl}$  (20°C, 1 bar) : 1.19 kg/m<sup>3</sup>  
 $H_{amb}$  : (55±15) %  
 $t_{amb}$  : (23±3) °C  
 $p_{amb}$  : (1010±10) mbar

5. The certificate is valid only to the item calibrated on date and place of calibration

### Calibrated by:

☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



Approved signatory: .....

Mr. Parinya Booncharoen  
Calibration Department Manager



JIRANATEE ASSOCIATES CO.,LTD.

Jiranatee Associates Co.,Ltd  
63/14-15, 67/35-36  
Petchkasem 7,7/1, Rd. Watthapra, Bangkokyai,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Pressure measurement laboratory  
Calibration services department.



## CERTIFICATE OF CALIBRATION

Certificate No. : CL-003-66

Page 2 of 2 Pages

MEASUREMENT RESULTS : ☒ Without adjustment ☐ With adjustment

CALIBRATION IN THE RANGE OF : 950 – 1050 mbar

The results of calibration and associated measurement uncertainties are reported in the table below.

STD (mbar)	UUC* (mbar)	Error (mbar)	Uncertainty (k=2) (mbar)
950.08	950.5	0.5	0.67
970.08	970.4	0.3	0.54
990.11	990.2	0.1	0.39
1010.07	1009.9	-0.2	0.42
1030.06	1029.7	-0.4	0.60
1050.08	1049.4	-0.7	0.88

Note: UUC\* Unit Under Calibration

: To convert the result in report unit to Pa should be multiply by 100

\*End of certificate\*





## CERTIFICATE OF CALIBRATION

Certificate No.: CL-077-65

Page 1 of 2

**Equipment Name:** Data Logger with Temperature  
Sensor

**Manufacturer:** Novalynx

**Model:** 200-WS-25LB

**Serial No.:** A5261

**ID No.:** BKK\_FS0888

**Customer**

**Name:** ALS laboratory group (Thailand) Co.,Ltd.  
**Address:** 104 Phatthanakan 40, Phatthanakan  
Rd.,Khwaeng Suan Luang, Khet Suan Luang, Bangkok  
10250Thailand.

**Received date:** 23 May 2022

**Calibration date:** 30 May 2022

**Issue date:** 02 Jun 2022

**Reference Used During Calibration**

- 1.Standard Temperature Probe Model: STS-100 A500,  
Serial No.: 667682-09, Due date: 23 Mar 2023
- 2.Digital Temperature Indicator Model: DTI-1000-A MK  
II, Serial No.: 671407-00591 Due date: 04 June 2022

**Calibration Condition**

Temperature:  $(23 \pm 3)^{\circ}\text{C}$   
Relative Humidity:  $(55 \pm 15)\%$

**Calibration Procedure**

The temperature calibration was done by In-House  
calibration method as WI-CL-001 according to  
comparison method with standard digital temperature  
indicator and standard temperature probe. The  
temperature scale use was based on ITS-90.

**Traceability**

The measurement results are traceable to the  
international system of units (SI) through National  
Institute of Metrology Thailand (NIMT) Certificate  
number: TT-0034-22, Certificate number: ER-0032-  
21

REVIEW BY	<i>Manakorn P.</i>
APPROVED BY	<i>[Signature]</i>
NEXT CAL. DATE	28/11/23

**Calibrated by**

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jittraporn Lertsomphol



**Approved Signatory:**

*[Signature]*  
Mr. Parinya Booncharoen  
Calibration Department Manager

**Result of Calibration:-** ☒ Without Adjustment ☐ With Adjustment

**Calibration Range:** 20-40 °C

**Function:**

This equipment was connected with temperature sensor Model : HMP60 S/N : N0330783

Dimension : Diameter 12mm. Length 80 mm.

<u>Immersion Depth (mm)</u>	<u>Standard Reading (°C)</u>	<u>UUC Reading (°C)</u>	<u>Error (°C)</u>	<u>Uncertainty (°C)</u>
60	19.98	20.0	0.1	0.30
60	24.98	24.7	-0.3	0.30
60	30.02	29.6	-0.4	0.30
60	35.01	34.5	-0.5	0.30
60	40.01	39.3	-0.7	0.30

**UUC\*:** Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

✱ End of Certificate ✱



## CALIBRATION REPORT

Calibration No. : RH-02062022

Page 1 of 1 Pages

Measurement Item : Relative humidity with data logger  
Manufacturer : Novalynx  
Model/Type : 200-WS-25LB  
Serial Number : A5261  
ID No. : BKK\_FSC888  
Customer : ALS laboratory group (Thailand) Co.,Ltd.  
: 104 Phatthanakan 40, Phatthanakan Rd.,Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250  
Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(25\pm3)^{\circ}\text{C}$ , and relative humidity of  $(50\pm15)\%$ .

### Measurement Method:

Unit Under Calibration (UUC) was calibrated by comparison method with standard thermo hygrometer in the humidity generator chamber to determine the errors.

### Traceability:

This instrument was calibrated using standard equipment whose accuracy is traceability through National Institute of Standards and Technology to the international system of units (SI) via MCS Calibration, Inc. Certificate number: 20314-101. Due date: Mar 14,2023.

Measurement Date : Jun 01, 2022

Issued Date : Jun 02, 2022

### Measurement Results:

This equipment was connected with Indoor air quality probe and Displayed (UR) on display. Model: HMP60, Serial number: N0330783

Calibration was performed in the range of 20%RH to 80%RH

The results of calibration are reported in table below.

Determined (%RH)	Standard (Reading) (%RH)	UUC (Reading) (%RH)	Error (%RH)	Uncertainty $\pm$ (%RH)
20	20.02	18.8	-1.2	0.61
50	50.22	49.4	-0.8	0.57
80	80.56	79.3	-1.3	0.69

Performed by

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jittraporn Lertsomphol



Approved Signatory: .....

Mr. Parinya Booncharoen.  
Calibration Department Manager



## CERTIFICATE OF CALIBRATION

Certificate No: WS-02062022

Page 1 of 2 pages

**Measurement Item** : Cup anemometer with data logger.

**Manufacturer** : Data logger: Novalynx  
: Cup anemometer: Novalynx

**Model/Type** : Data logger: 200-WS-25LB  
: Cup anemometer: WS-02P

**Serial Number** : Data logger: A5261  
: Cup anemometer: -

**ID No** : Data logger: BKK\_FS0888  
: Cup anemometer: -

**Customer** : ALS laboratory group (Thailand) co., ltd.  
: 104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10260 Thailand.

**Test Conditions** : Wind tunnel cross test section area 900 cm<sup>2</sup>  
: Anemometer frontal area 100 cm<sup>2</sup>  
: Diameter of mounting pipe - mm  
: Blockage ratio of test object 0.111 [-]

**Test Conditions** : Air temperature 24.7 ±0.8 °C  
: Air pressure 1005.2 ±0.4 hPa  
: Relative air humidity 46.1 ±3.5 %RH

**Calibration Procedure** Calibration was carried out base on;  
IEC 61400-12-1 ED.1: 2005-Power Performance Measurements of Electricity Producing Wind Turbines;  
MEASNET Anemometer Calibration Procedure – Version 2: 2009;

**Traceability** This calibration documents the traceable to national standard, Which realize the unit of measurements according to the international system of units (SI) through National Institute of Metrology Thailand (NIMT).

**Measurement Date** : Jun 01, 2022.

**Issued Date** : Jun 02, 2022.

**Calibrated by**

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



Approved Signatory: \_\_\_\_\_

Mr. Parinya Booncharoen  
Calibration Department Manager

Continuation of Certificate of Calibration Number

Certificate No: WS-02062022

Page 2 of 2 Pages

Result of calibration: ☒ Without adjustment ☐ With adjustment

Calibration in the range of 1 – 16 m/s at a calibration interval of 1 m/s.

The results of calibration and associated measurement uncertainties are reported in the table below.

V <sub>STD</sub> Reading m/s	V <sub>UUC</sub> Reading m/s	Error (m/s)	Uncertainty (%)
2.087	2.0	-0.1	2.4
4.140	4.2	0.1	1.0
6.02	6.0	0.0	0.88
8.00	8.0	0.0	0.74
10.00	10.0	0.0	0.59
11.99	12.1	0.1	0.55
14.02	14.3	0.3	0.42
16.00	16.4	0.4	0.63
15.01	15.4	0.4	0.39
12.97	13.1	0.1	0.59
11.00	11.0	0.0	0.52
9.01	9.0	0.0	0.66
7.01	7.0	0.0	0.85
5.186	5.2	0.0	0.96
3.003	3.1	0.1	1.6
1.053	0.8	-0.3	4.8

UUC\*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

#### Appendix 1: Instrumentations

NO	Sensor	Manufacturer	Model/Type	Calibration Date	Certificate Report Number	Range
1	Pitot static	TESTO INC.	06352145	Aug 07, 2021	MW-0034-21	5 – 30 m/s
2	Precision Differential Pressure Meter	Zoglab	DPM2600	Aug 07, 2021	MW-0034-21	5 – 30 m/s
3	Air velocity transducer (hot wire)	TSI INC.	8455-12	Aug 08, 2021	MW-0035-21	0 – 5 m/s
4	Temperature	Zoglab	DSR-THP	March 30, 2022	CL-027-65	-30 – 70°C
5	Relative humidity	Zoglab	DSR-THP	March 30, 2022	RH-03032022	0 – 100 %RH
6	Atmospheric pressure	Zoglab	DSR-THP	March 30, 2022	BP-01032022	500 – 1100 hPa
7	Wind tunnel	ESSOM	MP330D	-	-	0 – 50 Hz

\*\*\*End of certificate of calibration\*\*\*



## CERTIFICATE OF CALIBRATION

Certificate No.: WD-02062022

Page 1 of 2 pages

**Measurement Item** : Wind direction sensor with data logger.

**Manufacturer** : Data logger: Novalynx  
: Wind direction sensor: Novalynx

**Model/Type** : Data logger: 200-WS-25LB  
: Wind direction sensor: WS-02F

**Serial Number** : Data logger: A5261  
: Wind direction sensor: -

**ID No** : Data logger: BKK\_FS0888  
: Wind direction sensor: -

**Customer** : ALS laboratory group (Thailand) co., ltd.  
: 104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang, Khel Suan Luang, Bangkok 10250  
Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(23 \pm 3) ^\circ\text{C}$ , and relative humidity of  $(40 \pm 10) \%$ .

### Measurement Method:

The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and line laser is used for axis control. The measurement were taken at  $45^\circ$  intervals in clockwise and counterclockwise directions.

Note: The UUC was warmed up for 1 hour prior to the calibration being performed

### Traceability:

The measurement results are traceable to the international system of units (SI) through Certificate No.: Q21086014, Certificate No.: KWS64/0025.

**Measurement Date** : Jun 01, 2022.

**Issued Date** : Jun 02, 2022.

### Calibrated by

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



Approved Signatory: .....

Mr. Parinya Booncharoen.  
Calibration Department Manager



Continuation of Certificate of Calibration Number

Certificate No: WD-02062022

Pages 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment.

Calibration in the range of 0 – 360 ° at a calibration interval of 45°.

The results of calibration and associated measurement uncertainties are reported in table below.

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC* Reading (°)	Error (°)	Uncertainty ±(°)
1	Clockwise	0/360	360	359	-1	3.0
2		45	45	41	-4	3.0
3		90	90	87	-3	3.0
4		135	135	133	-2	3.0
5		180	180	181	1	3.0
6		225	225	229	4	3.0
7		270	270	274	4	3.0
8		315	315	319	4	3.0
9	Counter Clockwise	0/360	360	359	-1	3.0
10		45	45	41	-4	3.0
11		90	90	87	-3	3.0
12		135	135	133	-2	3.0
13		180	180	181	1	3.0
14		225	225	229	4	3.0
15		270	270	274	4	3.0
16		315	315	319	4	3.0

UUC\*: Unit Under Calibration The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

\*\*\*End of Certificate of Calibration\*\*\*



## CERTIFICATE OF CALIBRATION

Certificate No: WS-04012022

Page 1 of 2 pages

Measurement Item : Cup anemometer with data logger.

Manufacturer : Data logger: Novalynx  
: Cup anemometer: Novalynx

Model/Type : Data logger: 110-WS-25DL-D  
: Cup anemometer: WS-02F

Serial Number : Data logger: A5444  
: Cup anemometer: WSD-003

ID No : Data logger: RYG\_FS0435  
: Cup anemometer: -

Customer : ALS laboratory group (Thailand) co., ltd.  
: 104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250  
Thailand.

Test Conditions : Wind tunnel cross test section area 900 cm<sup>2</sup>  
: Anemometer frontal area 100 cm<sup>2</sup>  
: Diameter of mounting pipe - mm  
: Blockage ratio of test object 0.111 [-]

Test Conditions : Air temperature 24.4 ±0.8 °C  
: Air pressure 1011.2 ±0.4 hPa  
: Relative air humidity 55.6 ±3.5 %RH

Calibration Procedure : Calibration was carried out base on;  
: IEC 61400-12-1 GD.1: 2005-Power Performance Measurements of Electricity Producing Wind  
Turbines;  
: M&SNET Anemometer Calibration Procedure – Version 2: 2009;

Traceability : This calibration documents the traceable to national standard, Which realize the unit of  
measurements according to the international system of units (SI) through National Institute of  
Metrology Thailand (NIMT).

Measurement Date : JAN 26, 2022.  
Issued Date : JAN 31, 2022.

Calibrated by  
☒ Mr. Sorawit Thachalead  
☐ Miss Orathai Wiwatwittaya



Approved Signatory:

Mr. Parinya Booncharoen  
Calibration Department Manager

Continuation of Certificate of Calibration Number

Certificate No: WS-04012022

Page 2 of 2 Pages

Result of calibration: ☒ Without adjustment ☐ With adjustment

Calibration in the range of 1 – 16 m/s at a calibration interval of 1 m/s.

The results of calibration and associated measurement uncertainties are reported in the table below.

V <sub>STD</sub> Reading m/s	V <sub>UUC*</sub> Reading m/s	Error (m/s)	Uncertainty (%)
2.070	2.0	-0.1	2.4
4.105	4.1	0.0	1.2
6.01	6.0	0.0	0.99
8.01	8.0	0.0	0.71
10.01	10.1	0.1	1.1
12.01	12.2	0.2	0.65
13.98	14.3	0.3	0.61
15.94	16.1	0.2	1.4
14.98	15.1	0.1	1.0
13.00	13.1	0.1	0.76
11.02	11.1	0.1	0.63
9.02	9.0	0.0	0.97
7.03	7.0	0.0	0.84
5.166	5.1	-0.1	1.2
2.996	3.0	0.0	1.6
1.029	0.9	-0.1	4.5

UUC\*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

#### Appendix 1: Instrumentations

NO	Sensor	Manufacturer	Model/Type	Calibration Date	Certificate Report Number	Range
1	Pitot static	TCSTO INC.	06352145	Aug 07, 2021	MW-0034-21	5 – 30 m/s
2	Precision Differential Pressure Meter	Zoglab	DPM2500	Aug 07, 2021	MW-0034-21	5 – 30 m/s
3	Air velocity transducer (hot wire)	TSI INC.	8455-12	Aug 08, 2021	MW-0035-21	0 – 5 m/s
4	Temperature	Zoglab	DSR-THP	March 30, 2021	CL-027-64	-30 – 70°C
5	Relative humidity	Zoglab	DSR-THP	March 30, 2021	RH-03032021	0 – 100 %RH
6	Atmospheric pressure	Zoglab	DSR-THP	March 30, 2021	BP-01032021	500 – 1100 hPa
7	Wind tunnel	CSSOM	MP330D	-	-	0 – 50 Hz

\*\*\*End of certificate of calibration\*\*\*





## CERTIFICATE OF CALIBRATION

Certificate No.: WD-04012022

Page 1 of 2 pages

Measurement Item : Wind direction sensor with data logger.

Manufacturer : Data logger: Novalynx.  
: Wind direction sensor: Novalynx.

Model/Type : Data logger: 110-WS-25DL-D  
: Wind direction sensor: WS-02P

Serial Number : Data logger: A5444  
: Wind direction sensor: WSD-003

ID No : Data logger: RYG\_PSD435  
: Wind direction sensor: -

Customer : ALS laboratory group (Thailand) Co.,Ltd.  
104 Phatthanakan 40, Phatthanakan Rd.,Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250  
Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(23 \pm 3) ^\circ\text{C}$ , and relative humidity of  $(40 \pm 10) \%$ .

### Measurement Method:

The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and line laser is used for axis control. The measurement were taken at  $45^\circ$  intervals in clockwise and counterclockwise directions.

Note: The UUC was warmed up for 1 hour prior to the calibration being performed

### Traceability:

The measurement results are traceable to the international system of units (SI) through Certificate No.: Q21086014, Certificate No.: KWS64/0025.

Measurement Date : JAN 25, 2022.  
Issued Date : JAN 31, 2022.

### Performed by

- ☒ Mr. Sorawit Thachalad  
☐ Miss Orathai Wiwatwittaya



Approved Signatory:.....



Mr. Parinya Booncharoen.  
Calibration Department Manager

Continuation of Certificate of Calibration Number

Certificate No: WD-04012022

Pages 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment.

Calibration in the range of 0 – 360 ° at a calibration interval of 45°.

The results of calibration and associated measurement uncertainties are reported in table below.

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC* Reading (°)	Error (°)	Uncertainty ±(°)
1	Clockwise	0/360	0	0	0	3.0
2		45	45	41	-4	3.0
3		90	90	87	-3	3.0
4		135	135	133	-2	3.0
5		180	180	180	0	3.0
6		225	225	227	2	3.0
7		270	270	272	2	3.0
8		315	315	317	2	3.0
9	Counter Clockwise	0/360	0	0	0	3.0
10		45	45	41	-4	3.0
11		90	90	87	-3	3.0
12		135	135	133	-2	3.0
13		180	180	180	0	3.0
14		225	225	227	2	3.0
15		270	270	272	2	3.0
16		315	315	317	2	3.0

UUC\*: Unit Under Calibration The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

\*\*\*End of Certificate of Calibration\*\*\*



## CALIBRATION REPORT

Calibration No. : RH-04012022

Page 1 of 1 Pages

Measurement Item : Relative humidity with data logger.

Manufacturer : Data logger: Novalynx.  
: Relative humidity sensor: Novalynx.

Model/Type : Data logger: 110-WS-25DL-D  
: Relative humidity sensor: HMP60

Serial Number : Data logger: A5444  
: Relative humidity sensor: R1131112

ID No : Data logger: RYG\_FSD435  
: Relative humidity sensor: -

Customer : ALS laboratory group (Thailand) Co.,Ltd.  
: 104 Phatthanakan 40, Phatthanakan Rd.,Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250  
Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(25\pm3)^{\circ}\text{C}$ , and relative humidity of  $(50\pm15)\%$ .

### Measurement Method:

The Relative humidity with data logger, Unit Under Calibration (UUC) was calibrated by comparison method with the equilibrium of standard salt solution  $\text{CH}_3\text{COOK}$ : Potassium Acetate,  $\text{Mg}(\text{NO}_3)_2$ : Magnesium Nitrate,  $\text{KCl}$ : Potassium Chloride to determine the errors.

Measurement Date : JAN 24, 2022

Issued Date : JAN 25, 2022

### Measurement Results:

The results of calibration are reported in table below.

Standard salt solution.	Standard (%RH)	UUC <sub>(Reading)</sub>	Error
$\text{CH}_3\text{COOK}$ : Potassium Acetate	22.51	22.3	-0.2
$\text{Mg}(\text{NO}_3)_2$ : Magnesium Nitrate	52.89	52.5	-0.4
$\text{KCl}$ : Potassium Chloride	84.34	84.1	-0.2

### Performed by

- ☐ Mr. Sorawit Thachalad  
☒ Miss Orathai Wlwattittaya



Approved Signatory: \_\_\_\_\_



Mr. Parinya Booncharoen.  
Calibration Department Manager



## CALIBRATION REPORT

Calibration Number. : RG-04012022

Page 1 of 2 Pages

Measurement Item : Rain gauge with data logger.

Manufacturer : Data logger: Novalynx.  
: Rain gauge: Novalynx.

Model/Type : Data logger: 110-WS-25DL-D  
: Rain gauge: 110-WS-25RG

Serial Number : Data logger: A5444  
: Rain gauge: RG-003

ID NO : RYG\_FSO435

Customer : ALS laboratory group (Thailand) co., ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang, Khet Suan Luang,  
Bangkok 10250, Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(25 \pm 3)^{\circ}\text{C}$ , and relative humidity of  $(50 \pm 15)\%$ .

### Measurement Method:

The Rain gauge, Unit Under Calibration (UUC) was calibrated by Precision reference bottle with flow adjuster at low rate 0.6 mm per minute or 1 tipping every 20 seconds. The tipping number was determined by procedures below.

1. Obtain rain gauge inlet area:  
Rain gauge precise diameter in cm = Diameter/2 = R (radius)  
Rain gauge area =  $R^2 \times 3.14$  (UUC diameter=20.3 cm, UUC radius=10.15 cm)  
Rain gauge area =  $323.6 \text{ cm}^2$ .
2. Obtain theoretical correct rain gauge answer (number of tipplings) using  $323.6 \text{ cm}^2$  inlet area and 0.5 L of rain.
  - a)  $10,000 \text{ cm}^3 / 323.6 \text{ cm}^2$  inlet area = 30.90 (rain gauge area = 1/30.90 of square meter)
  - b)  $30.90 \times 0.5 \text{ L volume} = 15.45 \text{ mm}$  (mm of rain over  $1 \text{ m}^2$  surface) 500 ml of rain volume on the rain gauge area = 15.45 mm of rain.
  - c) Number of tipping =  $15.45 / 0.25 \text{ mm} = 62$  tipplings.

*Note: Rain gauge is fully cleaned and leveling prior the calibration performed.*

Measurement Date : JAN 28, 2022

Issued Date : JAN 31, 2022

### Performed by

- ☒ Mr. Sorawit Thachalad  
☐ Miss Orathai Wiwatwittaya



Approved Signatory: \_\_\_\_\_



Mr. Parinya Booncharoen,  
Calibration Department Manager

Continuation of Calibration of Calibration Number

Calibration Number: RG-04012022

Page 2 of 2 Pages

Result of Calibration: ☒ Without Adjustment ☐ With Adjustment.

The results of calibration are reported in table below.

Quantity of H <sub>2</sub> O (ml)	Determined Tipping	Tipping count	Acceptable Tipping count
500	62	64	60 - 64
500	62	63	60 - 64
500	62	62	60 - 64
500	62	63	60 - 64
500	62	62	60 - 64

*Remark: The procedure is made to verify the correct reading of the Unit under Calibration rain gauge when a precise volume of water falls into its cone. We suggest that the number of tipping should be within  $\pm 2\%$  different from the 62 tipping (correct range: 60-64 tipping) it means that the rain gauge meets the manufacturer acceptable limit.*

\*\*\*End of calibration report\*\*\*



## CERTIFICATE OF CALIBRATION

Certificate No. : CL-005-65

Page 1 of 2 Pages

MEASUREMENT ITEM : Digital barometer  
MANUFACTURER : Novalynx  
MODEL/TYPE : 110-WS-25BP  
SERIAL NUMBER : A5444  
ID NUMBER : RYG\_FS0435  
CUSTOMER : ALS laboratory group (Thailand) co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd,  
Khwaeng Suan Luang, Khet Suan Luang,  
Bangkok 10250 Thailand.  
RECEIVED DATE : 12 Jan 2022  
MEASUREMENT DATE : 29 Jan 2022  
ISSUE DATE : 31 Jan 2022

### Calibration procedure:

The pressure calibration was done by In-house calibration method as WI-CL-003 according to comparison method with Digital pressure calibrator based on DKD-R 6-1

### Traceability:

The measurement results are traceable to the international system of units (SI) through MENSOR which complies with the requirements of ISO/IEC17025:2017, ANSI/NCSL Z540-1 via Certificate number: 201479

The reported uncertainty is based on a standard uncertainty multiplied by a coverage  $k=2$ , providing a level of confidence of approximately 95%.

### CONDITION OF THIS RESULT OF CALIBRATION:

#### 1. Reference Standard Instrument:

Instrument	Model	Serial No.	Certificate No.	Due Date
Absolute Pressure Transducer	CPG2500	410018L1	201479	13 Sep 2022

2. The UUC\* was installed in vertical orientation above reference standard instrument and center of UUC\* was used as the reference level.

#### 3. Calibration conditions:

Pressure transmitting medium : Air  
 $\rho_{F1}(20^{\circ}\text{C}, 1\text{bar})$  :  $1.19 \text{ kg/m}^3$   
 $\Delta h$  :  $-0.080 \text{ m}$   
 $t_{amb}$  :  $(23 \pm 2)^{\circ}\text{C}$   
 $p_{amb}$  :  $1009.5 \text{ mbar}$

4. The certificate is valid only to the item calibrated on date and place of calibration.

### Calibrated by:

- ☒ Mr. Sorawit Thachalad  
☐ Miss Oratai Wiwatwittaya



### Approved signatory:

Mr. Parinya Booncharoen  
Calibration Department Manager



MEASUREMENT RESULTS : ☒ Without adjustment ☐ With adjustment

CALIBRATION IN THE RANGE OF : 950 – 1050 mbar

The results of calibration and associated measurement uncertainties are reported in the table below.

STD (mbar)	UUC* (mbar)	Error (mbar)	Uncertainty(k=2) (mbar)
950.32	951.181	0.856	1.3
970.14	970.682	0.538	0.70
990.05	990.524	0.470	0.58
1009.95	1010.106	0.157	0.34
1029.84	1029.946	0.107	0.25
1049.78	1049.594	-0.190	0.35

Note: UUC\* Unit Under Calibration

\*End of certificate\*



## CERTIFICATE OF CALIBRATION

Certificate No.: CL-004-65

Page 1 of 2

**Equipment Name:** Data Logger with Temperature  
Sensor

**Manufacturer:** Novalynx

**Model:** 110-WS-25DL-D

**Serial No.:** A5444

**ID No.:** RYG\_FS0435

**Customer**

**Name:** ALS laboratory group (Thailand) Co.,Ltd.

**Address:** 104 Phatthanakan 40, Phatthanakan  
Rd.,Khwaeng Suan Luang, Khet Suan Luang, Bangkok  
10250Thailand.

**Received date:** 12 JAN 2022

**Calibration date:** 24 JAN 2022

**Issue date:** 25 JAN 2022

**Reference Used During Calibration**

1.Standard Temperature Probe Model: STS-100 A500,  
Serial No.: 667682-09, Due date: 25 Mar 2022

2.Digital Temperature Indicator Model: DTI-1000-A MK  
II, Serial No.: 671407-00591 Due date: 04 June 2022

**Calibration Condition**

Temperature:  $(23 \pm 3)^\circ\text{C}$

Relative Humidity:  $(55 \pm 15)\%$

**Calibration Procedure**

The temperature calibration was done by In-House  
calibration method as WI-CL-001 according to  
comparison method with standard digital temperature  
indicator and standard temperature probe. The  
temperature scale use was based on ITS-90.

**Traceability**

The measurement results are traceable to the  
international system of units (SI) through National  
Institute of Metrology Thailand (NIMT) Certificate  
number: TT-0036-21, Certificate number: ER-0032-  
21.

**Calibrated by**

☐ Mr. Sorawit Thachalad

☒ Miss Orathai Wiwatwittaya



**Approved Signatory:** .....

*2kmf*  
Mr. Parinya Booncharoen  
Calibration Department Manager

Result of Calibration:- ☒ Without Adjustment ☐ With Adjustment

Calibration Range: 20-40 °C

Function:

This equipment was connected with temperature sensor Model : HMP60 S/N : R1131114

Dimension : Diameter 12mm. Length 80 mm.

<u>Immersion Depth (mm)</u>	<u>Standard Reading (°C)</u>	<u>UUC Reading (°C)</u>	<u>Error (°C)</u>	<u>Uncertainty (°C)</u>
60	20.053	19.8	-0.3	0.099
60	25.005	24.5	-0.5	0.099
60	29.995	29.5	-0.5	0.099
60	34.976	34.4	-0.6	0.099
60	39.957	39.3	-0.7	0.099

UUC\*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%





## CERTIFICATE OF CALIBRATION

Certificate No.: WD-01112021

Page 1 of 2 pages

Measurement Item : Wind direction sensor with data logger.

Manufacturer : Data logger: Novalynx.  
: Wind direction sensor: Novalynx.

Model/Type : Data logger: 200-WS-25LB  
: Wind direction sensor: WS-02P

Serial Number : Data logger: A5377  
: Wind direction sensor: -

ID No : Data logger: BKK\_PSD0917  
: Wind direction sensor: -

Customer : ALS laboratory group (Thailand) Co.,Ltd.  
104 Phatthanakan 40, Phatthanakan Rd.,Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250  
Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(23 \pm 3) ^\circ\text{C}$ , and relative humidity of  $(40 \pm 10) \%$ .

### Measurement Method:

The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and line laser is used for axis control. The measurement were taken at  $45^\circ$  intervals in clockwise and counterclockwise directions.

Note: The UUC was warmed up for 1 hour prior to the calibration being performed

### Traceability:

The measurement results are traceable to the international system of units (SI) through Certificate No.: Q21086014, Certificate No.: KWS64/0025.

Measurement Date : Nov 01, 2021.

Issued Date : Nov 01, 2021.

### Performed by

- ☒ Mr. Sorawit Thachalad  
☐ Miss Orathai Wiwatwittaya



Approved Signatory:.....

Mr. Parinya Booncharoen.  
Calibration Department Manager

Continuation of Certificate of Calibration Number

Certificate No: WD-01112021

Pages 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment.

Calibration in the range of 0 – 360 ° at a calibration interval of 45°.

The results of calibration and associated measurement uncertainties are reported in table below.

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC* Reading (°)	Error (°)	Uncertainty ±(°)
1	Clockwise	0/360	360	359	-1	3.0
2		45	45	41	-4	3.0
3		90	90	87	-3	3.0
4		135	135	135	0	3.0
5		180	180	183	3	3.0
6		225	225	229	4	3.0
7		270	270	274	4	3.0
8		315	315	320	5	3.0
9	Counter Clockwise	0/360	360	359	-1	3.0
10		45	45	41	-4	3.0
11		90	90	87	-3	3.0
12		135	135	135	0	3.0
13		180	180	183	3	3.0
14		225	225	229	4	3.0
15		270	270	274	4	3.0
16		315	315	320	5	3.0

UUC\*: Unit Under Calibration The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor  $k=2$  providing a level of confidence of approximately 95%.

\*\*\*End of Certificate of Calibration\*\*\*



## CERTIFICATE OF CALIBRATION

Certificate No: WS-01112021

Page 1 of 2 pages

Measurement Item : Cup anemometer with data logger.

Manufacturer : Data logger: Novalynx.  
: Cup anemometer: Novalynx.

Model/Type : Data logger: 200-WS-25LB  
: Cup anemometer: WS-02F

Serial Number : Data logger: A5377  
: Cup anemometer: -

ID No : Data logger: BKK\_FS0917  
: Cup anemometer: -

Customer : ALS laboratory group (Thailand) co., ltd.  
: 104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand.

Test Conditions : Wind tunnel cross test section area 900 cm<sup>2</sup>  
: Anemometer frontal area 100 cm<sup>2</sup>  
: Diameter of mounting pipe - mm  
: Blockage ratio of test object 0.111 [-]

Test Conditions : Air temperature 25.4 ±0.8 °C  
: Air pressure 1015.1 ±0.4 hPa  
: Relative air humidity 47.9 ±3.5 %RH

Calibration Procedure Calibration was carried out base on;  
IEC 61400-12-1 ED.1: 2005-Power Performance Measurements of Electricity Producing Wind Turbines;  
MEASNET Anemometer Calibration Procedure - Version 2: 2009;

Traceability This calibration documents the traceable to national standard, Which realize the unit of measurements according to the international system of units (SI) through National Institute of Metrology Thailand (NIMT).

Measurement Date : Nov 01, 2021.

Issued Date : Nov 01, 2021.

**Calibrated by**

- ☒ Mr. Sorawit Thachalee  
☐ Miss Orathai Wiwatwittaya



Approved Signatory: .....

*Mr. Parinya Booncharoen*

Mr. Parinya Booncharoen  
Calibration Department Manager



Continuation of Certificate of Calibration Number

Certificate No: WS-01112021

Page 2 of 2 Pages

Result of calibration: ☒ Without adjustment ☐ With adjustment

Calibration in the range of 1 – 16 m/s at a calibration interval of 1 m/s.

The results of calibration and associated measurement uncertainties are reported in the table below.

V <sub>STD</sub> Reading m/s	V <sub>UUC</sub> Reading m/s	Error (m/s)	Uncertainty (%)
1.994	1.9	-0.1	2.5
4.002	4.0	0.0	1.2
6.00	6.0	0.0	0.95
8.02	8.0	0.0	0.73
10.01	10.1	0.1	0.63
12.00	12.1	0.1	0.74
13.99	14.1	0.1	0.76
16.01	16.3	0.3	0.80
16.01	16.3	0.3	0.64
13.00	13.1	0.1	0.46
11.01	11.1	0.1	0.57
9.02	9.0	0.0	0.64
7.02	7.0	0.0	0.98
4.992	5.0	0.0	1.2
2.980	2.9	-0.1	1.5
0.996	0.9	-0.1	4.5

UUC\*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

#### Appendix 1: Instrumentations

NO	Sensor	Manufacturer	Model/Type	Calibration Date	Certificate Report Number	Range
1	Pitot static	TESTO INC.	06352145	Aug 07, 2021	MW-0034-21	5 – 30 m/s
2	Precision Differential Pressure Meter	Zoglab	DPM2500	Aug 07, 2021	MW-0034-21	5 – 30 m/s
3	Air velocity transducer (hot wire)	TSI INC.	8455-12	Aug 08, 2021	MW-0035-21	0 - 5 m/s
4	Temperature	Zoglab	DSR-THP	March 30, 2021	CL-027-64	-30 - 70°C
5	Relative humidity	Zoglab	DSR-THP	March 30, 2021	RH-03032021	0 – 100 %RH
6	Atmospheric pressure	Zoglab	DSR-THP	March 30, 2021	BP-01032021	500 – 1100 hPa
7	Wind tunnel	ESSOM	MP330D	-	-	0 – 60 Hz

\*\*\*End of certificate of calibration\*\*\*



Certificate Number

CL-002-65

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

**MEASUREMENT ITEM** : Wind Direction Sensor  
**MANUFACTURER** : Novalynx  
**MODEL/TYPE** : Sensor: WS-02F  
Data logger: 110-WS-25DL-D  
**SERIAL NUMBER** : Sensor: WSD-012  
Data logger: A5909  
**ID NUMBER** : RYG\_FS0608  
**CONDITION AS-RECEIVED** : New item  
**CUSTOMER** : ALS laboratory group (Thailand) co., ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,  
Khet Suan Luang, Bangkok 10250 Thailand.

**RECEIVED DATE** : 09 Nov 2022  
**MEASUREMENT DATE** : 17 Nov 2022  
**ISSUE DATE** : 23 Nov 2022

**ENVIRONMENTAL CONDITIONS:**

Ambient condition in the laboratory are as follow:

Temperature	: 23.0 ± 3.0	°C
Relative Humidity	: 55.0 ± 15.0	%RH
Atmospheric Pressure	: 1010 ± 10	hPa

**PLACE OF CALIBRATION** : Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

<b>CALIBRATION CONDITION</b>	: Wind tunnel cross-section area <sup>1</sup>	900	cm <sup>2</sup>
	Win direction frontal area <sup>2</sup>	129	cm <sup>2</sup>
	Diameter of mounting pipe <sup>3</sup>	-	mm
	Blockage ratio of test object <sup>4</sup>	0.143	[-]

**Preconditioning** : 24 hours at ambient conditions.  
**Measurement Condition** : The average values during measurement are (24.0)°C, (50.6) %RH and (1009.4) hPa.

**TABULATION OF RESULTS:**

The table on next page give the measured values.

**Calibrated by:**

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



Approved signatory: .....

*[Signature]*  
Mr. Parinya Booncharoen  
Calibration Department Manager

**Remark:**

- <sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>

MEASUREMENT RESULTS<sup>5</sup>

The wind direction sensor was calibrated against standard rotary encoder by comparison method. During calibration, the measurement was carried out at 45° intervals in clockwise and counterclockwise directions after offset adjustment has been made. The flow speed of wind tunnel (usually 5 m/s) is kept constant while the sensor is rotated around its vertical axis. The results of calibration and associated measurement uncertainties are reported in the table below.

Air speed m/s	$D_{std}^{\circ}$ Degree (°)	$D_{uuc}^{\circ}$ Degree (°)	Error Degree (°)	$U (k=2)$ Degree (°)
5.00	0.000	0	0	0.58
	45.000	42	-3	0.74
	89.999	88	-2	0.68
	135.001	133	-2	0.68
	180.001	179	-1	0.68
	225.000	225	0	0.68
	270.000	271	1	0.68
	315.000	318	3	0.74

## Remark:

<sup>3</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Direction of standard

<sup>7</sup> Direction of Unit Under Calibration

\*\*\*End of Certificate of Calibration\*\*\*





## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

<b>MEASUREMENT ITEM</b>	: Cup anemometer
<b>MANUFACTURER</b>	: Novalynx
<b>MODEL/TYPE</b>	: Sensor: WS-02F Data logger: 110-WS-25DL-D
<b>SERIAL NUMBER</b>	: Sensor: WSD-012 Data logger: A5909
<b>ID NUMBER</b>	: RYG_FS0608
<b>CONDITION AS-RECEIVED</b>	: New item
<b>CUSTOMER</b>	: ALS laboratory group (Thailand) co., Ltd. 104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand.

<b>RECEIVED DATE</b>	: 09 Nov 2022
<b>MEASUREMENT DATE</b>	: 17 Nov 2022
<b>ISSUE DATE</b>	: 23 Nov 2022

**ENVIRONMENTAL CONDITIONS:**

Ambient condition in the laboratory are as follow:

Temperature	: 23.0 ± 3.0	°C
Relative Humidity	: 55.0 ± 15.0	%RH
Atmospheric Pressure	: 1010 ± 10	hPa

<b>PLACE OF CALIBRATION</b>	: Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.
-----------------------------	---

<b>CALIBRATION CONDITIONS</b>	: Wind tunnel cross-section area <sup>1</sup>	900	cm <sup>2</sup>
	Win direction frontal area <sup>2</sup>	100	cm <sup>2</sup>
	Diameter of mounting pipe <sup>3</sup>	-	mm
	Blockage ratio of test object <sup>4</sup>	0.111	[-]

<b>Preconditioning</b>	: 24 hours at ambient conditions.
<b>Measurement Condition</b>	: The average values during measurement are (23.8) °C, (46.3) %RH and (1014.7) hPa.

**TABULATION OF RESULTS:**


The table on next page give the measured values.

**Calibrated by:**

☒ Mr. Sorawit Thachalad  
☐ Miss Jittraporn Lertsomphol



Approved signatory: .....

  
 Mr. Parinya Booncharoen  
 Calibration Department Manager

**Remark:**

<sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>

MEASUREMENT RESULTS <sup>5</sup>

The cup anemometer, Unit Under Calibration (UUC) was exercised at 10 m/s for 5 minutes prior to calibration being performed. The standard air velocity 0.5 m/s to 5 m/s was calculated by a standard air velocity transducer and above 5 m/s to 30 m/s was calculated by a pitot tube with precision differential pressure meter which was installed 40 mm and 300 mm respectively away from wind tunnel nozzle. UUC was installed at center of the test section. The calibration was carried out under both rising and falling air velocity in the range of 1 m/s to 16 m/s at calibration interval of 1 m/s. The results of calibration and associated measurement uncertainties are reported in the table below.

$V_{std}$ <sup>6</sup> (m/s)	Temp. wind tunnel (°C)	Temp. room (°C)	$V_{UUC}$ <sup>7</sup> (m/s)	Error (m/s)	$U$ ( $k=2$ ) (m/s)
0.988	23.90	23.80	0.8	-0.2	0.15
2.035	23.70	23.80	1.8	-0.2	0.16
3.040	23.90	23.80	2.8	-0.2	0.19
4.194	23.60	23.80	3.8	-0.4	0.20
5.01	23.70	23.80	4.8	-0.2	0.19
6.00	23.78	23.80	5.8	-0.2	0.17
7.08	23.80	23.80	6.8	-0.2	0.18
8.18	23.60	23.80	8.0	-0.2	0.20
9.10	23.80	23.80	8.9	-0.2	0.20
10.09	23.64	23.80	9.9	-0.2	0.21
11.15	23.56	23.80	10.9	-0.3	0.21
12.16	23.66	23.80	11.9	-0.3	0.21
13.20	23.52	23.80	12.9	-0.3	0.22
14.26	23.60	23.80	14.1	-0.2	0.22
15.25	23.58	23.80	15.0	-0.2	0.22
16.30	23.60	23.80	16.2	-0.1	0.24

## Remark:

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Velocity of standard

<sup>7</sup> Velocity of Unit Under Calibration

## PHOTO OF CALIBRATION SET-UP



Calibration set-up of the cup anemometer calibration in the wind tunnel of Jiranatee Associates Co., Ltd. The cup anemometer shown may differ from the calibrated one. Remark: The proportion of the set-up is not true to scale due to imaging geometry.



## CERTIFICATE OF CALIBRATION

Certificate No.: CL-157-65  
Page 1 of 2

Equipment Name: Data Logger with Temperature  
Sensor

Manufacturer: Novalyrix  
Model: 110-WS-25DL D  
Serial No.: A5909  
ID No.: RYG\_FS0608

### Customer

Name: ALS laboratory group (Thailand) Co.,Ltd.  
Address: 104 Phatthanakan 40, Phatthanakan  
Rd.,Khwaeng Suan Luang, Khet Suan Luang, Bangkok  
10250Thailand.

Received date: 09 Nov 2022  
Calibration date: 18 Nov2022  
Issue date: 23 Nov 2022

### Reference Used During Calibration

- 1.Standard Temperature Probe Model: STS-100 A500,  
Serial No.: 667682 09, Due date: 23 Mar 2023
- 2.Digital Temperature Indicator Model: DTI-1000 A MK  
II, Serial No.: 671407-00591 Due date: 22 July 2023

### Calibration Condition

Temperature: (23±3)°C  
Relative Humidity: (55±15)%

### Calibration Procedure

The temperature calibration was done by In-House  
calibration method as WI-CL-001 according to  
comparison method with standard digital temperature  
indicator and standard temperature probe. The  
temperature scale use was based on ITS-90.

### Traceability


The measurement results are traceable to the  
international system of units (SI) through National  
Institute of Metrology Thailand (NIMT) Certificate  
number: TT-0034-22, Certificate number: ER-0092-  
22

### Calibrated by

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



### Approved Signatory:

  
Mr. Parinya Booncharoen  
Calibration Department Manager



Result of Calibration:- ☒ Without Adjustment ☐ With Adjustment

Calibration Range: 20-40 °C

**Function:**

This equipment was connected with temperature sensor Model: HMP60 S/N: U3641220.

Dimension : Diameter 12 mm. Length 80 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
60	19.98	19.9	-0.1	0.30
60	25.00	24.8	-0.2	0.30
60	30.00	29.8	-0.2	0.30
60	35.01	34.7	-0.3	0.30
60	40.01	39.5	-0.5	0.30

UUC\*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

**\* End of Certificate \***



## CERTIFICATE OF CALIBRATION

Calibration No. : RH-02112022

Page 1 of 1 Pages

Measurement Item : Relative humidity with data logger  
Manufacturer : Novalynx  
Model/Type : 110-WS-25DL-D  
Serial Number : A5909  
ID No. : RYG\_FS0608  
Customer : ALS laboratory group (Thailand) Co., Ltd.  
: 104 Phalphanakan 40, Phalphanakan Rd., Khwaeng Suan Luang, Khet Suan Luang, Bangkok  
10250 Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(25 \pm 3)^{\circ}\text{C}$ , and relative humidity of  $(50 \pm 15)\%$ .

### Measurement Method:

Unit Under Calibration (UUC) was calibrated by comparison method with standard thermo hygrometer in the humidity generator chamber to determine the errors.

### Traceability:

This instrument was calibrated using standard equipment whose accuracy is traceability through National Institute of Standards and Technology to the international system of units (SI) via MCS Calibration, Inc. Certificate number: 20314-101. Due date: Mar 14,2023.

Measurement Date : Nov 18, 2022

Issued Date : Nov 23, 2022

### Measurement Results:

This equipment was connected with Indoor air quality probe and Displayed (UR) on display. Model: HMP60, Serial number: U3641220

Calibration was performed in the range of 20%RH to 80%RH

The results of calibration are reported in table below.

Determined (%RH)	Standard (Reading) (%RH)	UUC (Reading) (%RH)	Error (%RH)	Uncertainty $\pm$ (%RH)
20	19.94	17.4	-2.5	0.57
50	50.31	47.1	-3.3	0.55
80	80.30	77.4	-2.9	0.57

Performed by

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphot



Approved Signatory: \_\_\_\_\_

Mr. Parinya Booncharoen.  
Calibration Department Manager

Certificate Number
CL-006-65

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

MEASUREMENT ITEM	: Cup anemometer
MANUFACTURER	: Novalynx
MODEL/TYPE	: Sensor: WS-02F Data logger: 110-WS-25DL-D
SERIAL NUMBER	: Sensor: WSD-016 Data logger: AS910
ID NUMBER	: RYG_FS0609
CONDITION AS-RECEIVED	: New item
CUSTOMER	: ALS laboratory group (Thailand) co., ltd. 104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250 Thailand.

RECEIVED DATE	: 09 Nov 2022
MEASUREMENT DATE	: 18 Nov 2022
ISSUE DATE	: 23 Nov 2022

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature	: 23.0 ± 3.0	°C
Relative Humidity	: 55.0 ± 15.0	%RH
Atmospheric Pressure	: 1010 ± 10	hPa

PLACE OF CALIBRATION	: Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.
----------------------	---

CALIBRATION CONDITIONS	: Wind tunnel cross-section area <sup>1</sup>	900	cm <sup>2</sup>
	Win direction frontal area <sup>2</sup>	100	cm <sup>2</sup>
	Diameter of mounting pipe <sup>3</sup>	-	mm
	Blockage ratio of test object <sup>4</sup>	0.111	[-]

Preconditioning	: 24 hours at ambient conditions.
Measurement Condition	: The average values during measurement are (23.7) °C, (49.5) %RH and (1011.2) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



Approved signatory: *[Signature]*

Mr. Parinya Booncharoen  
Calibration Department Manager

### Remark:

- <sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>



**MEASUREMENT RESULTS <sup>5</sup>**

The cup anemometer, Unit Under Calibration (UUC) was exercised at 10 m/s for 5 minutes prior to calibration being performed. The standard air velocity 0.5 m/s to 5 m/s was calculated by a standard air velocity transducer and above 5 m/s to 30 m/s was calculated by a pitot tube with precision differential pressure meter which was installed 40 mm and 300 mm respectively away from wind tunnel nozzle, UUC was installed at center of the test section. The calibration was carried out under both rising and falling air velocity in the range of 1 m/s to 16 m/s at calibration interval of 1 m/s. The results of calibration and associated measurement uncertainties are reported in the table below.

$V_{std}$ <sup>6</sup> (m/s)	Temp. wind tunnel (°C)	Temp. room (°C)	$V_{uuc}$ <sup>7</sup> (m/s)	Error (m/s)	$U$ ( $k=2$ ) (m/s)
0.971	23.98	23.70	0.8	-0.2	0.15
2.042	23.46	23.70	1.8	-0.2	0.16
3.069	23.90	23.70	2.8	-0.2	0.21
4.221	23.60	23.70	3.9	-0.4	0.20
5.02	23.80	23.70	4.8	-0.2	0.20
5.99	23.84	23.70	5.8	-0.2	0.19
7.06	23.66	23.70	6.8	-0.2	0.18
8.17	23.78	23.70	7.9	-0.2	0.18
9.10	23.60	23.70	8.8	-0.3	0.22
10.09	23.74	23.70	9.9	-0.2	0.19
11.16	23.68	23.70	10.9	-0.2	0.22
12.14	23.94	23.70	11.8	-0.3	0.22
13.19	23.70	23.70	13.0	-0.2	0.22
14.27	23.94	23.70	13.8	-0.4	0.22
15.25	23.78	23.70	15.1	-0.2	0.26
16.30	23.84	23.70	15.9	-0.4	0.25

**Remark:**

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Velocity of standard

<sup>7</sup> Velocity of Unit Under Calibration

**PHOTO OF CALIBRATION SET-UP**

Calibration set-up of the cup anemometer calibration in the wind tunnel of Jiranatee Associates Co., Ltd. The cup anemometer shown may differ from the calibrated one. Remark: The proportion of the set-up is not true to scale due to imaging geometry.

\*\*\*End of Certificate of Calibration\*\*\*



Certificate Number

CL-006-65

## CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

**MEASUREMENT ITEM** : Wind Direction Sensor  
**MANUFACTURER** : Novalynx  
**MODEL/TYPE** : Sensor: WS-02F  
Data logger: 110-WS-25DL-D  
**SERIAL NUMBER** : Sensor: WSD-016  
Data logger: A5910  
**ID NUMBER** : RYG\_FS0609  
**CONDITION AS-RECEIVED** : New item  
**CUSTOMER** : ALS laboratory group (Thailand) co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,  
Khet Suan Luang, Bangkok 10250 Thailand.

**RECEIVED DATE** : 09 Nov 2022  
**MEASUREMENT DATE** : 18 Nov 2022  
**ISSUE DATE** : 23 Nov 2022

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature :  $23.0 \pm 3.0$  °C  
Relative Humidity :  $55.0 \pm 15.0$  %RH  
Atmospheric Pressure :  $1010 \pm 10$  hPa

**PLACE OF CALIBRATION** : Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

**CALIBRATION CONDITION** : Wind tunnel cross-section area<sup>1</sup> 900 cm<sup>2</sup>  
Win direction frontal area<sup>2</sup> 129 cm<sup>2</sup>  
Diameter of mounting pipe<sup>3</sup> - mm  
Blockage ratio of test object<sup>4</sup> 0.143 [-]

**Preconditioning** : 24 hours at ambient conditions.  
**Measurement Condition** : The average values during measurement are (24.1)°C, (45.1) %RH and (1012.5) hPa.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibrated by:

☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



Approved signatory: \_\_\_\_\_



Mr. Parinya Booncharoen  
Calibration Department Manager

### Remark:

- <sup>1</sup> Nozzle cross-section area of the wind tunnel  
<sup>2</sup> Projected cross-section area of the tested object include mounting pipe  
<sup>3</sup> Diameter of mounting pipe  
<sup>4</sup> Ratio <sup>2</sup> to <sup>1</sup>

MEASUREMENT RESULTS<sup>5</sup>

The wind direction sensor was calibrated against standard rotary encoder by comparison method. During calibration, the measurement was carried out at 45° intervals in clockwise and counterclockwise directions after offset adjustment has been made. The flow speed of wind tunnel (usually 5 m/s) is kept constant while the sensor is rotated around its vertical axis. The results of calibration and associated measurement uncertainties are reported in the table below.

Air speed	$D'_{std}$	$D'_{uuc}$	Error	$U (k=2)$
m/s	Degree (°)	Degree (°)	Degree (°)	Degree (°)
5.01	0.000	0	0	0.58
	45.000	44	-2	0.76
	90.000	87	-3	0.58
	135.000	132	-3	0.68
	180.000	177	-3	0.68
	225.000	222	-3	0.58
	270.001	270	0	0.74
	315.000	318	3	0.58

## Remark:

<sup>5</sup> Calibration results only count for the tested circumstances and environmental conditions during which calibration took place

<sup>6</sup> Direction of standard

<sup>7</sup> Direction of Unit Under Calibration

\*\*\*End of Certificate of Calibration\*\*\*







63/14-15,67/35-36, Soi Petchkasem7,7/1, Petchkasem Rd,  
Walthapra, Bangkokyai,Bangkok 10600 Thailand.  
Tel.: (66) 02-8680812#13 Fax.: (66) 02-8680860 www.jiranalee.com

## CERTIFICATE OF CALIBRATION

Calibration No. : RH-06112022

Page 1 of 1 Pages

Measurement Item : Relative humidity with data logger  
Manufacturer : Novalynx  
Model/Type : 110-WS-25DL-D  
Serial Number : A5910  
ID No. : RYG\_PS0609  
Customer : ALS laboratory group (Thailand) Co., Ltd.  
: 104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Suan Luang, Khel Suan Luang, Bangkok  
10250 Thailand.

### Environmental Condition:

The measurement was carried out in an ambient temperature of  $(25\pm3)^{\circ}\text{C}$ , and relative humidity of  $(50\pm15)\%$ .

### Measurement Method:

Unit Under Calibration (UUC) was calibrated by comparison method with standard thermo hygrometer in the humidity generator chamber to determine the errors.

### Traceability:

This instrument was calibrated using standard equipment whose accuracy is traceability through National Institute of Standards and Technology to the international system of units (SI) via MCS Calibration, Inc. Certificate number: 20314-101. Due date: Mar 14,2023.

Measurement Date : Nov 18, 2022

Issued Date : Nov 23, 2022

### Measurement Results:

This equipment was connected with Indoor air quality probe and Displayed (UR) on display. Model: HMP60, Serial number: U3641223

Calibration was performed in the range of 20%RH to 80%RH

The results of calibration are reported in table below.

Determined (%RH)	Standard (Reading) (%RH)	UUC (Reading) (%RH)	Error (%RH)	Uncertainty $\pm$ (%RH)
20	19.98	17.6	-2.4	0.56
50	50.28	47.3	-3.0	0.51
80	80.30	77.6	-2.7	0.52

Performed by

- ☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



Approved Signatory: .....

Mr. Parinya Booncharoen.  
Calibration Department Manager

## CERTIFICATE OF CALIBRATION

Certificate No.: CL-161-65  
Page 1 of 2

**Equipment Name:** Data Logger with Temperature  
Sensor

**Manufacturer:** Novalynx  
**Model:** 110-WS-25DL-D  
**Serial No.:** A5910  
**ID No.:** RYG\_FS0609

**Customer**

**Name:** ALS laboratory group (Thailand) Co., Ltd.  
**Address:** 104 Phatthanakan 40, Phatthanakan  
Rd., Khwaeng Suan Luang, Khet Suan Luang, Bangkok  
10250 Thailand.

**Received date:** 09 Nov 2022  
**Calibration date:** 18 Nov 2022  
**Issue date:** 23 Nov 2022

**Reference Used During Calibration**

1. Standard Temperature Probe Model: STS-100 A500,  
Serial No.: 667682-09, Due date: 23 Mar 2023  
2. Digital Temperature Indicator Model: DTI-1000-A MK  
II, Serial No.: 671407-00591 Due date: 22 July 2023

**Calibration Condition**

Temperature:  $(23 \pm 3)^\circ\text{C}$   
Relative Humidity:  $(55 \pm 15)\%$

**Calibration Procedure**

The temperature calibration was done by In-House  
calibration method as WI CL 001 according to  
comparison method with standard digital temperature  
indicator and standard temperature probe. The  
temperature scale use was based on ITS-90.

**Traceability**

The measurement results are traceable to the  
international system of units (SI) through National  
Institute of Metrology Thailand (NIMT) Certificate  
number: TT-0034-22, Certificate number: ER-0092-  
22

**Calibrated by**

☒ Mr. Sorawit Thachalad  
☐ Miss Jitraporn Lertsomphol



**Approved Signatory:**



Mr. Parinya Booncharoen  
Calibration Department Manager

Result of Calibration:- ☒ Without Adjustment ☐ With Adjustment

Calibration Range: 20 - 40 °C

**Function:**

This equipment was connected with temperature sensor Model: HMP60 S/N: U3641223.

Dimension : Diameter 12 mm. Length 80 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
60	20.03	19.8	-0.2	0.30
60	25.02	24.8	-0.2	0.30
60	30.00	29.7	-0.3	0.30
60	35.00	34.6	-0.4	0.30
60	40.00	39.5	-0.5	0.30

UUC\*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor  $k=2$  providing a level of confidence of approximately 95%





# Certificate of System Qualification

GC-OQ + GCMS-OQ

System ID: RYG\_EN0136  
Organization Name: ALS Laboratory Group ( Thailand ) Co Ltd.  
Organization Location: 616/10 Moo 5, Tambol Mae Nam Koo, A.Pluakdaeng, Rayong, 21140, Thailand  
Date: July 7, 2022 11:27:53 AM  
EQP Name: AgilentRecommended , AgilentRecommended  
EQP Revision: GC.02.52, GCMS.02.52  
Overall Qualification Status: Pass

REVIEW BY N. Banniy  
APPROVED BY [Signature]  
NEXT CAL. DATE 07/01/24

## CDS Logon Verification - GC

Logon: dej.changchon

## Overall CDS Logon Verification - GC Test Status

Pass

## System Inspection and Basic Safety and Operation

Name: 7890

Setpoint Status: Pass

## Overall System Inspection and Basic Safety and Operation Test Status

Pass

## Inlet Pressure Accuracy

Name: 7890

Front SSL

Setpoint Status: Pass

	Setpoint	Actual
Inlet Pressure:	25.0 psi	25.1 psi
Accuracy:		0.1 psi
Agilent Recommended:		<= 1.2

Date: July 7, 2022 11:27:53 AM  
System ID: RYG\_EN0136

## Overall Inlet Pressure Accuracy Test Status

Pass

## GC Oven Temperature Accuracy

Name: 7890

Setpoint Status: Pass

Zone: Oven

Setpoint/Actual

Temperature: 230.0 230.6 °C

Accuracy: 0.6 °C

Agilent Recommended:  $\geq -1.0$  % setpoint in K ( -5.0 °C )  
 $\leq 1.0$  % setpoint in K ( 5.0 °C )

Setpoint Status: Pass

Zone: Oven

Setpoint/Actual

Temperature: 100.0 99.9 °C

Accuracy: -0.1 °C

Agilent Recommended:  $\geq -1.0$  % setpoint in K ( -3.7 °C )  
 $\leq 1.0$  % setpoint in K ( 3.7 °C )

## Overall GC Oven Temperature Accuracy Test Status

Pass

## GC Oven Temperature Stability

Name: 7890

Setpoint Status: Pass

Setpoint/Average

Temperature: 100.0 99.91667 °C

Stability: 0.1 °C

Agilent Recommended:  $\leq 0.5$ 

## Overall GC Oven Temperature Stability Test Status

Pass

## Log Amp

Tested Combination1	Front	SSL	/ External	SQ
Name:	5977B			
Setpoint Status:	Pass			
Overall Log Amp Test Status				
Pass				

## RFPA

Tested Combination1	Front	SSL	/ External	SQ
Name:	5977B			
Setpoint Status:	Pass			
Amu:	1050	m/z	Drift After Five Minutes:	RFPA Voltage:
			-1 mV	479 mV
Agilent Recommended:	>= -100 and <= 100		<= 1100	
Overall RFPA Test Status				
Pass				

## Tune EI

Tested Combination1	Front	SSL	/ External	SQ
Name:	5977B			
Setpoint Status:	Pass			
Filament:	1			
Setpoint Status:	Pass			
Filament:	2			
Overall Tune EI Test Status				
Pass				

## Signal to Noise EI

Date: July 7, 2022 11:27:53 AM  
System ID: RYG\_EN0136



---

Tested Combination1	Front	SSL	/ External	SQ
Name:	5977B			

---

Source:	EI - Extractor	Filament:	1
---------	----------------	-----------	---

Setpoint Status:	Pass
------------------	------

Signal to Noise:	7485
------------------	------

Agilent Recommended:	>= 1200
----------------------	---------

---

Source:	EI - Extractor	Filament:	2
---------	----------------	-----------	---

Setpoint Status:	Pass
------------------	------

Signal to Noise:	2097
------------------	------

Agilent Recommended:	>= 1200
----------------------	---------

---

This test's 2 comment(s) and 7 deviation(s) are available in the Attachments section.

**Overall Signal to Noise EI Test Status**

Pass
------

## Instrument Details

### Purpose

This section describes the as found system configuration.

### Details

#### System

System ID	RYG_EN0136
Manufacturer	Agilent Technologies
Name	7890
Flow Data Input	Manual Data
Temperature Data Input	Manual Data or Other Data Logging

#### Tested Combination1

Injection Technique	Manual Injection
Inlet	Front
Detector	External
LTM Included?	No

#### Sampler 1

Manufacturer	Agilent Technologies
Type	Manual Injection
Usage	Sample Injection
Syringe Volume (µL)	10

#### Mainframe 1

Manufacturer	Agilent Technologies
Name	7890
Model Number	G3442B
Serial Number	CN16463238
Firmware Revision	B.02.04.3
Component ID/Asset No.	081117000236
Oven Type	Standard

## Inlet 1

Manufacturer	Agilent Technologies
Name	7890
Type	SSL
Location	Front
Carrier Gas	Helium
Control Type	Electronic Pressure Control (EPC)
Purged Inlet	Yes

## Detector 1

Manufacturer	Agilent Technologies
Name	Mass Spectrometer
Type	Mass Spectrometer
Location	External

## Mass Spectrometer 1

Manufacturer	Agilent Technologies
Type	SQ
Name	5977B
Serial Number	US1701M008
Firmware Revision	5977 6.00.34
High Vacuum System	Turbo Pump
Scouting Run Standard	OFN Std
Component ID/Asset No.	081117000236

## MS EI Source 1

Manufacturer	Agilent Technologies
Source Type	EI - Extractor
Number of filaments	2



# Electronic Signature

## Purpose

This signature page was created and published because the ACE sign-off action was executed, which is valid for the entire document, including attachments. The ACE sign-off is an electronic signature that requires two distinct identification components: unique username and personal password. The Agilent representative who has delivered this service understands the meaning and legal status of an electronic signature. As a trained official operator, the Agilent representative has a unique password and logon to access ACE and electronically sign this document. (Other e-signatures can be applied to this document using a Document Content Management or other suitable method defined in your data access and control procedures.)

## Details

Full Name of Signer:	Eaknarin Puangsopa
Logged On User Name:	eaknarin_puangsopa@agilent.com
Signature Creation Date:	July 7, 2022
Reason for Signature:	Executed protocol and published this original version of document

## Regulatory Disclaimer

This document provides a protocol to verify and record instrument configuration and evidence of proper operation. It has been prepared from our interpretation of applicable regulations as well as industry best practices. The document is designed to provide an important component of a complete compliance package. Validation depends upon many factors and use of this protocol alone does not assure compliance. Agilent Technologies makes no promises or representations as to its sufficiency for any specific regulatory program.

## Warranty

Agilent Technologies makes no warranty of any kind to this material, including but not limited to, the implied warranties or merchantability and fitness for a particular purpose. Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

---

Date:	July 7, 2022 11:27:53 AM
System ID:	RYG_EN0136

User Name: eaknarin\_puangsoa  
 Hostname: ASRYGW7002

System Id: RYG\_EN0136  
 Print Date: July 7, 2022 11:27:56 AM

ALS\_RYG\_EN0136 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
July 6, 2022 1:11:54 PM	Audit	SessionCreated	Session	None
July 6, 2022 1:11:54 PM	Start	Configuration	Session	None
July 6, 2022 1:11:54 PM	Audit	Entitlement	Licensing	User is Nonpaying and does not require an unlock code
July 6, 2022 1:17:19 PM	Audit	EqpLoaded	Session	EQP details for primary technique [Gc] - File path: [ProtocolPacks/Gc/Configurations/02.52/Gc.02.52.eqp], EQP File Name: [Gc.02.52.eqp], EQP Name: [AgilentRecommended] EQP details for hyphenated technique [GcMs] - File path: [ProtocolPacks/GcMs/Configurations/02.52/GcMs.02.52.eqp], EQP File Name: [GcMs.02.52.eqp], EQP Name: [AgilentRecommended]
July 6, 2022 1:17:25 PM	End	Configuration	Session	None
July 6, 2022 1:17:29 PM	Start	Qualification	Session	OQ
July 6, 2022 1:17:30 PM	Start	Execution	CDS Logon Verification - GC : - Qualitative test	None
July 6, 2022 1:19:43 PM	End	Execution	CDS Logon Verification - GC : - Qualitative test	Run Count : 1
July 6, 2022 1:19:46 PM	Start	Execution	System Inspection and Basic Safety and Operation - 7890: - Qualitative Test - No setpoints associated	None

User Name: eaknarin\_puangsoa  
 Hostname: ASRYGW7002

System Id: RYG\_EN0136  
 Print Date: July 7, 2022 11:27:56 AM

ALS\_RYG\_EN0136 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
July 6, 2022 1:19:59 PM	End	Execution	System Inspection and Basic Safety and Operation - 7890: - Qualitative Test - No setpoints associated	Run Count : 1
July 6, 2022 1:20:15 PM	Start	Execution	Inlet Pressure Accuracy - Front SSL: - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	None
July 6, 2022 1:21:43 PM	End	Execution	Inlet Pressure Accuracy - Front SSL: - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	Run Count : 1
July 6, 2022 1:21:45 PM	Start	Execution	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	None
July 6, 2022 1:25:12 PM	Audit	Data	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Manual Data Entry
July 6, 2022 1:25:15 PM	End	Execution	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Run Count : 1
July 6, 2022 1:25:17 PM	Start	Execution	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	None
July 6, 2022 1:25:32 PM	Start	Execution	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	None
July 6, 2022 1:33:42 PM	Audit	Data	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Manual Data Entry

Page 2 / 10



User Name: eaknarin\_puangsoa  
 Hostname: ASRYGW7002

System Id: RYG\_EN0136  
 Print Date: July 7, 2022 11:27:56 AM

ALS\_RYG\_EN0136 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
July 6, 2022 1:33:43 PM	End	Execution	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Run Count : 1
July 6, 2022 1:33:45 PM	Start	Execution	GC Oven Temperature Stability - 7890: - Temperature : Oven - S: 100.0°C - L: <= 0.5°C	None
July 6, 2022 1:53:05 PM	Audit	Data	GC Oven Temperature Stability - 7890: - Temperature : Oven - S: 100.0°C - L: <= 0.5°C	Manual Data Entry
July 6, 2022 1:53:07 PM	End	Execution	GC Oven Temperature Stability - 7890: - Temperature : Oven - S: 100.0°C - L: <= 0.5°C	Run Count : 1
July 6, 2022 1:53:11 PM	Start	Execution	Log Amp - 5977B SQ: - Source: None EI - Extractor	
July 6, 2022 1:57:10 PM	End	Execution	Log Amp - 5977B SQ: - Source: EI EI - Extractor	Run Count : 1
July 6, 2022 1:57:24 PM	Start	Execution	RFPA - 5977B SQ: - Source: EI - Extractor	None
July 6, 2022 2:09:24 PM	End	Execution	RFPA - 5977B SQ: - Source: EI - Extractor	Run Count : 1
July 6, 2022 2:09:28 PM	Start	Execution	Tune EI - 5977B SQ: - Source: - None EI - Extractor Filament 1 (Qualitative - No setpoints associated)	
July 6, 2022 2:24:46 PM	End	Qualification	Session	OQ
July 6, 2022 2:24:46 PM	Start	Reporting	Session	None
July 6, 2022 2:41:39 PM	End	Reporting	Session	None
July 6, 2022 2:41:39 PM	Start	Configuration	Session	None
July 6, 2022 2:41:40 PM	End	Configuration	Session	None

User Name: eaknarin\_puangsoa  
 Hostname: ASRYGW7002

System Id: RYG\_EN0136  
 Print Date: July 7, 2022 11:27:56 AM

ALS\_RYG\_EN0136 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
July 6, 2022 2:41:40 PM	Start	Qualification	Session	OQ
July 6, 2022 2:41:40 PM	Start	Execution	Tune EI - 5977B SQ: - Source: - None EI - Extractor Filament 1 (Qualitative - No setpoints associated)	
July 6, 2022 2:41:56 PM	End	Execution	Tune EI - 5977B SQ: - Source: - Run Count : 1 EI - Extractor Filament 1 (Qualitative - No setpoints associated)	
July 6, 2022 2:41:58 PM	Start	Execution	Tune EI - 5977B SQ: - Source: - None EI - Extractor Filament 2 (Qualitative - No setpoints associated)	
July 6, 2022 2:42:48 PM	End	Qualification	Session	OQ
July 6, 2022 2:42:48 PM	Start	Reporting	Session	None
July 6, 2022 2:50:52 PM	End	Reporting	Session	None
July 6, 2022 2:50:52 PM	Start	Qualification	Session	OQ
July 6, 2022 2:50:52 PM	Start	Execution	Tune EI - 5977B SQ: - Source: - None EI - Extractor Filament 2 (Qualitative - No setpoints associated)	
July 6, 2022 2:51:12 PM	End	Qualification	Session	OQ
July 6, 2022 2:51:12 PM	Start	Reporting	Session	None
July 6, 2022 2:55:29 PM	End	Reporting	Session	None
July 6, 2022 2:55:29 PM	Start	Qualification	Session	OQ
July 6, 2022 2:55:29 PM	Start	Execution	Tune EI - 5977B SQ: - Source: - None EI - Extractor Filament 2 (Qualitative - No setpoints associated)	

User Name: eaknarin\_puangsoapa  
 Hostname: ASRYGW7002

System Id: RYG\_EN0136  
 Print Date: July 7, 2022 11:27:56 AM

ALS\_RYG\_EN0136 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
July 6, 2022 2:55:40 PM	End	Execution	Tune EI - 5977B SQ: - Source: - Run Count : 1 EI - Extractor Filament 2 (Qualitative - No setpoints associated)	
July 6, 2022 2:55:45 PM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None
July 6, 2022 3:21:52 PM	End	Qualification	Session	QQ
July 6, 2022 3:21:52 PM	Start	Reporting	Session	None
July 6, 2022 3:25:04 PM	End	Reporting	Session	None
July 6, 2022 3:25:04 PM	Start	Qualification	Session	QQ
July 6, 2022 3:25:04 PM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None
July 6, 2022 4:06:40 PM	Audit	AceClosed	Session	None
July 7, 2022 9:13:47 AM	Audit	AceRestarted	Session	None
July 7, 2022 9:13:49 AM	Audit	SessionReloaded	Session	None
July 7, 2022 9:13:54 AM	Start	Qualification	Session	QQ
July 7, 2022 9:13:54 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None
July 7, 2022 9:58:06 AM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data files Path : D:\OQ2022\OQFN_SN_F01.D



User Name: eaknarin\_puangsoa  
 Hostname: ASRYGW7002

System Id: RYG\_EN0136  
 Print Date: July 7, 2022 11:27:56 AM

ALS\_RYG\_EN0136 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
July 7, 2022 9:59:53 AM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Run Count : 1
July 7, 2022 10:01:46 AM	Audit	TestUnlocked	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Deviation filed for Run Count : 1
July 7, 2022 10:01:46 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None
July 7, 2022 10:02:00 AM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data files Path : D:\OQ2022\OFN_SN_F01.D
July 7, 2022 10:04:55 AM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Run Count : 2
July 7, 2022 10:07:30 AM	Audit	TestUnlocked	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Deviation filed for Run Count : 2
July 7, 2022 10:07:30 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None
July 7, 2022 10:07:44 AM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data files Path : D:\OQ2022\OFN_SN_F01.D

User Name: eaknarin\_puangsoa  
 Hostname: ASRYGW7002

System Id: RYG\_EN0136  
 Print Date: July 7, 2022 11:27:56 AM

ALS\_RYG\_EN0136 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
July 7, 2022 10:08:18 AM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Run Count : 3
July 7, 2022 10:10:28 AM	Audit	TestUnlocked	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Deviation filed for Run Count : 3
July 7, 2022 10:10:28 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None
July 7, 2022 10:10:55 AM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data files Path : D:\OQ2022\OFN_SN_F01.D
July 7, 2022 10:14:03 AM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Run Count : 4
July 7, 2022 10:14:54 AM	Audit	TestUnlocked	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Deviation filed for Run Count : 4
July 7, 2022 10:14:54 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None
July 7, 2022 10:15:15 AM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data files Path : D:\OQ2022\OFN_SN_F01.D

User Name: eaknarin\_puangsoa  
 Hostname: ASRYGW7002

System Id: RYG\_EN0136  
 Print Date: July 7, 2022 11:27:56 AM

ALS\_RYG\_EN0136 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
July 7, 2022 10:15:27 AM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Run Count : 5
July 7, 2022 10:16:48 AM	Audit	TestUnlocked	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Deviation filed for Run Count : 5
July 7, 2022 10:16:48 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None
July 7, 2022 10:17:05 AM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data files Path : D:\OQ2022\OFN_SN_F01.D
July 7, 2022 10:17:14 AM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Run Count : 6
July 7, 2022 10:18:40 AM	End	Qualification	Session	OQ
July 7, 2022 10:18:40 AM	Start	Reporting	Session	None
July 7, 2022 10:21:10 AM	End	Reporting	Session	None
July 7, 2022 10:21:10 AM	Start	Qualification	Session	OQ
July 7, 2022 10:21:17 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	None
July 7, 2022 10:56:49 AM	End	Qualification	Session	OQ
July 7, 2022 10:56:49 AM	Start	Reporting	Session	None
July 7, 2022 10:57:38 AM	End	Reporting	Session	None

Page 8 / 10

Date: July 7, 2022 11:27:53 AM  
 System ID: RYG\_EN0136



User Name: eaknarin\_puangsoa

System id: RYG\_EN0136

Hostname: ASRYGW7002

Print Date: July 7, 2022 11:27:56 AM

## ALS\_RYG\_EN0136 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
July 7, 2022 10:57:38 AM	Start	Qualification	Session	OQ
July 7, 2022 10:57:38 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	None
July 7, 2022 11:06:50 AM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data files Path : D:\OQ2022\OFN_SN_F021.D
July 7, 2022 11:11:47 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	None
July 7, 2022 11:13:13 AM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Run Count : 1
July 7, 2022 11:14:29 AM	Audit	TestUnlocked	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Deviation filed for Run Count : 1
July 7, 2022 11:14:29 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	None
July 7, 2022 11:14:47 AM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data files Path : D:\OQ2022\OFN_SN_F021.D
July 7, 2022 11:16:34 AM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Run Count : 2

User Name: eaknarin\_puangsoa

System Id: RYG\_EN0136

Hostname: ASRYGW7002

Print Date: July 7, 2022 11:27:56 AM


## ALS\_RYG\_EN0136 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
July 7, 2022 11:19:56 AM	Audit	TestUnlocked	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Deviation filed for Run Count : 2
July 7, 2022 11:19:56 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	None
July 7, 2022 11:20:13 AM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data files Path : D:\OQ2022\OFN_SN_F021.D
July 7, 2022 11:21:52 AM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Run Count : 3
July 7, 2022 11:22:49 AM	End	Qualification	Session	OQ
July 7, 2022 11:22:49 AM	Start	Reporting	Session	None
July 7, 2022 11:26:46 AM	Audit	Reporting	Session	Report Generated : Certificate

CUSTOMER NAME	: ALS Laboratory Group (Thailand) Co., Ltd. [บริษัท เอแอลเอส แลบบอราทอรี กรุ๊ป (ประเทศไทย) จำกัด]						
EQUIPMENT NAME	: THC Analyzer						
MANUFACTURER	: HORIBA	MODEL	: APHA-370	SERIAL NO	: UA3NG4TH		
STANDARD GAS CONCENTRATION (PPM)	: 506.1 PPM			CYLINDER NO	: CC734373		
CYLINDER PRESSURE (psig)	: 1,600 PSI			CERTIFIED DATE	: 12/05/2020		
CERTIFIED BY	: AIRGAS			EXPIRED DATE	: 12/05/2028		

TEST RESULTS

POINT NO	TEST RESULTS						
	IDEAL	ACTUAL CH4	ERROR CH4	%ERROR CH4	ACTUAL THC	ERROR THC	%ERROR THC
ZERO	0.000	0.210	0.210	-	0.200	0.200	-
1	10.000	10.050	0.050	0.50	10.050	0.050	0.50
2	20.000	20.120	0.120	0.60	20.150	0.150	0.75
3	30.000	30.110	0.110	0.37	30.050	0.050	0.17
4	40.000	40.030	0.030	0.08	40.030	0.030	0.08
AVERAGE (%)				0.39			0.37



CALIBRATED BY : [Signature] DATE : 25/1/16  
 CHECKED BY : [Signature] DATE : 25/1/16

**J**  
**NAC**  
JIRANATE ASSOCIATES CO., LTD.

ต้องการข้อมูลทางด้านเทคนิคเพิ่มเติม : เจ้าหน้าที่ฝ่ายบริการหลังการขาย , โทร 02-868-0812 # 15,16 , E-Mail : Engineer@jiranatee.com  
 เลขที่ 63/14-15,67/35-36 ถนนเพชรเกษม 7,7/1 แขวงวัดท่าพระ เขตบางกอกใหญ่ กรุงเทพฯ 10600 โทร 02-8680812-13 โทรสาร 02-868-1889



CUSTOMER NAME	: ALS Laboratory Group (Thailand) Co., Ltd. [บริษัท เอแอลเอส แลบบอราทอรี กรุ๊ป (ประเทศไทย) จำกัด]		
EQUIPMENT NAME	: THC Analyzer		
MANUFACTURER	: HORIBA	MODEL : APHA-370	SERIAL NO. : U430GTHB

TEST VALUES				
NO.	THC Analyzer ( APHA - 370 )	UNIT	BEFORE	AFTER
1	Signal ( CH4 )	mV	4.300	42.400
2	Signal ( THC )	mV	3.200	64.400
3	Detector	Temp °C , Standard Value : Ambient temp+(5°Cto15°C)	46.700	50.000
		Pressure kPa , Standard Value : (Ambient/1013x100-20)±4kPa	70.000	70.100
4	Ambient	kPa current atmospheric pressure	101.000	101.100
5	Purifire	°C , Standard Value : 390 °C to 430 °C	420.400	421.200
		kPa , Normal value : 8 kPa to 25 kPa	9.800	9.800
6	NMHC	°C , Standard Value : 230 °C to 260 °C	244.800	245.100
7	DC 24 V	V , Standard Value : 24 V ± 0.5 V	23.900	23.900
8	DC 5 V	V , Standard Value : 5 V ± 0.5 V	5.000	5.000
9	Bypass (Optional)	L/min, Normal value : 0.9 L/min ± 0.3 L/min	-	-
10	Over Flow (Optional)	L/min, Standard Value : 0.8 L/min or More	-	-
11	CH4 Sampling Reading	PPM	3.530	2.330
12	NMHC Sampling Reading	PPM	4.280	1.150
13	THC Sampling Reading	PPM	8.810	3.480
14	Zero Gas CH4/THC	PPM	0.21/0.20	0.00/0.00
15	Span Gas	PPM	54.87/55.78	40.03/40.03
G	Gas H2 ...../.....	20 PSI	20	20

Remark : Reference EX-EN-017-56 , Ambient HC Monitor APHA-370 Operation Manual Page #81

Remark : ( Ambient temperature = 5°C to 40°C )

อาการที่ตรวจพบ

- Service Maintenance

รายละเอียดการดำเนินการ

- ทำ Calibration Zero/Span , Multipoint

ผลการดำเนินการ

- เรียบร้อย เครื่องสามารถดำเนินการตรวจวัดได้ตามปกติ

CALIBRATED BY :                     

CHECKED BY :                     



DATE :                     

DATE :                     

ต้องการข้อมูลทางด้านเทคนิคเพิ่มเติม : เจ้าหน้าที่ฝ่ายบริการหลังการขาย , โทร 02-868-0812 # 15-16 , E-Mail : Engineer@jiranatee.com

เลขที่ 63/14-15,67/35-36 ซอยเพชรเกษม 7,7/1 ถนนเพชรเกษม แขวงวัดท่าพระ เขตบางกอกใหญ่ กรุงเทพฯ 10600 โทร 02-868-0812-13 โทรสาร 02-868-1889



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL. 0-2717-3000-29 FAX. 0-2719-9484



Cert.No.: 23CH275

Page.: 1 of 3

## Certificate of Calibration

Equipment : pH Meter  
Manufacturer : Mettler Toledo  
Model : SevenCompact S220  
Serial No. : C104059460  
ID No. : RYG\_EN0183  
Condition As-Received: Used Item  
Received Date : 24 February 2023  
Calibration Date : 27 February 2023  
Reference : 2302-0886DSC-2  
Submitted by : ALS Laboratory Group (Thailand) Co.,Ltd.  
(Rayong Branch)  
616/10 Moo 5, T.Maenam Khu, A.Pluakdaeng,  
Rayong 21140, Thailand

REVIEW BY N. Banit  
APPROVED BY D. Sathip  
NEXT CAL. DATE 27/2/24

Ambient Temperature : (25 ± 2.5) °C  
Relative Humidity : (50 ± 15) %  
Calibration Procedure : In - house method :  
- CP-CH5 by direct measurement with standard  
voltage calibrator and direct measurement with  
certified reference material (CRM)  
- CP-CH8 by comparison with standard thermometer

Calibrated by : Walalak Sirithean

Approved by :

Saithip

Approved Signatory

- ( ) Malee Butkruea  
(✓) Saithip Meangmai  
( ) Warakorn Lernagtrakul

Issue Date : 28 February 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

A 0051538



Cert.No.: 23CH275

Page.: 2 of 3

**Condition of this calibration result**

1. Reference Standard Instrument : -

<u>Instrument</u>	<u>Serial No.</u>	<u>ID No.</u>	<u>Cert. No.</u>	<u>Due Date</u>
1) Document Process Calibrator	54030049	130RC116	22E2769	24 Aug 2023
2) Ref. Standard Thermometer	4982054	110RC044	22I1306	27 Oct 2023

This certification is traceable to the International System of Unit maintained at:-

- Traceable to National Institute of Metrology (Thailand), NIMT

2. Certified Reference Materials : The measurement results are traceable to SI through CPA chem Ltd.,  
ANSI-ASQ National Accreditation Board, Accredited No. AR-1835

<u>Buffer Solution</u>	<u>Manufacturer</u>	<u>Lot No.</u>	<u>Exp. date</u>
pH 4.008	CPA chem	826588	09 July 2024
pH 6.987	CPA chem	826589	09 July 2023
pH 10.010	CPA chem	863835	28 Dec 2023

3. This certificate is valid only to the item calibrated on date and place of calibration.

**Calibration Results****Function : mV Measurement****Performing standard curve by Fluke at pH (4,7,10)**

Unit Under Calibration	Nominal Value	Standard Voltage Input	Actual Reading		Uncertainty of Measurement ( $\pm$ mV)	Coverage factor <i>k</i>
	pH	mV	mV	pH		
pH Meter S/N.: C104059460	4.000	177.48	177.4	4.000	0.058	2.00
	7.000	0.00	-0.1	7.000	0.058	2.00
	10.000	-177.48	-177.5	10.000	0.058	2.00

*Saethip*





Cert.No.: 23CH275

Page.: 3 of 3

**Calibration Results****Function : pH Measurement**

Performing three buffers standard curve by using buffer nominal pH (4,7,10)

Unit Under Calibration	Standard pH Buffer Solution	Actual pH Reading	Actual mV Reading ( mV )	Uncertainty of pH measurement ( $\pm$ )	Coverage factor $k$
pH Electrode S/N.: 1453404	4.008	4.008	179.1	0.0046	2.00
	6.987	6.988	4.7	0.0084	2.00
	10.010	10.013	-172.4	0.0069	2.00

**Function : Temperature Measurement****( \* ) Without adjustment**

This equipment was connected with Temperature Probe;

- Model : InLabExpert Pro-ISM

- Serial No. : 1453404

Dimension of probe;

- Length : 120 mm.

- Diameter : 12 mm.

- Immersion Depth : 100 mm.

Calibration Point ( °C )	Standard Temperature ( °C )	UUC* Reading ( °C )	Error ( °C )	Uncertainty of measurement ( $\pm$ °C )	Coverage factor $k$
25.0	25.001	24.8	-0.201	0.13	2.00

**Remark : - UUC\* = Unit Under Calibration**

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor  $k$ , providing a level of confidence of approximately 95 %.

-o0o-

*Saitip*

a 1149924



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG, BANGKOK 10250  
TEL. 0-2717-3000-24 FAX. 0-2719-9484



## Certificate of Calibration

Certificate No. : 23E753

Page : 1 of 2

Equipment : pH Meter  
Manufacturer: Mettler Toledo  
Model : SevenCompact S220  
Serial No.: C104059460  
ID No.: RYG\_EN0183

This certificate may not be reproduced other than in full,  
except with the prior written approval of the head of  
Corporate Services 3: Equipment Calibration and Testing Services.

Condition As-Received: Used Item  
Received Date: 24 February 2023  
Calibration Date: 28 February 2023

Reference: 2302-0886DSC  
Ambient Temperature: ( 23  $\pm$  2 ) °C  
Relative Humidity: ( 50  $\pm$  10 ) %

Submitted by: ALS Laboratory Group (Thailand) Co.,Ltd. (Rayong Branch)

616/10 Moo 5, T.Maenam Khu, A.Pluakdaeng,  
Rayong 21140, Thailand

Procedure used: Calibration were conducted using In-house calibration Procedure CP-E17 According to direct measurement method with Multi-Product Calibrator.

### Condition of this result of calibration

1.Reference standards instruments :

<u>Instrument</u>	<u>Model</u>	<u>Serial No.</u>	<u>Certificate No.</u>	<u>Due Date</u>
1) Multi-Product Calibrator	5500A	6440007	22E1670	18 May 2023

2.This result of calibration was made on requested at the point specified by customer.


3.The certificate is valid only to the item calibrated on date and place of calibration.

4.This Certification is traceable to the International System of Unit maintained at:-

-National Institute of Metrology Thailand (NIMT)

Calibrated by : Wutchareeporn Wongchutikrane  
Issue Date : 02 March 2023

Approved Signatory :

  
[ ] Phalinee Prabpaipal  
[x] Nuntawat Khamchai  
[ ] Pornthippa Tameyakul

B 0309672





Cert. No.: 23E753

Page.: 2 of 2

**Result of calibration :-** (\*) Without adjustment ( ) After adjustment

<b>Function:</b>	DC voltage measuremer	<b>Range:</b>	2000	mV	
<u>Standard Value</u>	<u>UUC* Reading</u>	<u>Error</u>	<u>Uncertainty</u>		
( mV )	( mV )	( mV )	( $\pm \mu V$ )		
-200.0000	-200.0	0.0	72		
-150.0000	-150.0	0.0	69		
-100.0000	-100.0	0.0	65		
-50.0000	-50.0	0.0	62		
0.0000	0.0	0.0	58		
50.0000	50.0	0.0	62		
100.0000	99.9	-0.1	65		
150.0000	149.9	-0.1	69		
200.0000	199.9	-0.1	72		

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95 %

UUC\* = Unit Under Calibration.

-o0o-





# Metrological Center

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhoi, Saraburi 18110, Thailand.

Saraburi Tel : +66 3627 3096 Fax : +66 3627 3100

Bangkok Tel : +668 9205 6851 , +669 8247 2360

Website : www.scieco.co.th E-Mail : calibrate@scg.com



Certificate No. T230116

Page 1 of 4

## Certificate of Calibration

Equipment : Chamber ( Cooling Room )

Manufacturer : MODULAR

Model : IREVCOHCOO

Serial No. : C00351459

Customer Code : RYG\_EN0184

ID No. : T1939A5

Customer : ALS Laboratory Group (Thailand) Co.,Ltd. ( Rayong Branch)

616/10 Moo 5 T.Maenam Khu,

A.Pluakdaeng, Rayong 21140


Customer Location : Laboratory

Date of Receipt : 23 January 2023

Calibrated By : Atiphong Rongrat ( Technician )

Approved By : Boonchai / Boonchai Suriyawong (Site Calibration Manager)

Date of Issue : 07 FEB 2023



The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standard laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Metrological Center.

Certificate No. T230116

Page 2 of 4

## Calibration Report

**Equipment** : Chamber ( Cooling Room )  
**Date of Calibration** : 25 January 2023  
**Environment** : Temperature : 23.4-24.9 °C  
                               Line Voltage : 221.4-230.2 V  
                               Relative Humidity : 55 - 65 %RH

### Condition of this results of calibration :

1. This equipment was calibrated by insert 16 standard thermocouples type T into its chamber , the other one standard thermocouples type T use for ambient temperature measurement . The calibration was done in according to WI-T20 ( based on ASTM E145-94 ( Reapproved 2001) and AS2853-1986 ).

All data show below were final values and the initial data from customer request . The temperature scale used was based on ITS - 90 .

2. Reference Standard Instrument :

Instrument	Model	Instrument No.	Certificate No.	Due Date
TC	TYPE T	TN141-TN150	T222123	5 October 2023
TC	TYPE T	TN151-TN160	T222123	5 October 2023
DATA LOGGER	34970A	T150	T222123	5 October 2023

3. This certificate is traceable to :

National Institute of Metrology ( Thailand ) through Metrological Center ( NSC-TISI-TIS 17025 CALIBRATION 0244.)

4. Condition of calibrated item : good

Equipment Description :

Time Constant    1 Hour                      -                      Minute At    3 °C  
 Fresh Air Damper   ☐ Open                      ☐ Min                      ☐ Medium                      ☐ Max  
                                  ☐ Close  
                                  ☒ Not Available

5. Adjustment :

( X ) without adjustment

(   ) after adjustment


Approved By. \_\_\_\_\_



Certificate No. T230116

Page 3 of 4

## Calibration Report




C = Centre , F = Centre of Face , A = Corner , E = Centre of Edge

1C	=	TN141
2A	=	TN142
3A	=	TN143
4F	=	TN144
5A	=	TN145
6A	=	TN146
7F	=	TN147
8F	=	TN148
9A	=	TN149
10A	=	TN150
11F	=	TN151

12F	=	TN152
13A	=	TN153
14A	=	TN154
15F	=	TN155
16E	=	TN156

Approved By. \_\_\_\_\_





Certificate No. T230116

Page 4 of 4

## Calibration Report

### Measurement Results

Calibration Point	Average Standard Reading at each position (°C)											
	TN141	TN142	TN143	TN144	TN145	TN146	TN147	TN148	TN149	TN150	TN151	TN152
3.0	3.03	3.16	3.15	3.19	3.45	3.47	3.21	3.35	3.54	3.45	3.24	3.34
	TN153	TN154	TN155	TN156								
	3.28	3.22	3.28	3.21								

Chamber ( Cooling Room )			Temperature Distribution			
Setting (°C)	Reading (°C)		Stability (+°C)	Uniformity (°C)	Uncertainty (±°C)	Coverage Factor <i>k</i>
	Min , Max	Average				
3.0	2.8 , 4.1	3.5	1.20	1.20	1.90	2.07

The calibration result apply only the above calibrated item.

The result of test was found accurate as shown on date and place of test only.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor *k* which for a t-distribution, providing a level of confidence of approximately 95 % .

Approved By. 



**TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)**  
**CORPORATE SERVICES 3 : EQUIPMENT CALIBRATION AND TESTING SERVICES**


534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250

TEL. 0-2717-3000 FAX. 0-2719-9484

**Cert.No.:** 22TW34

**Page.:** 1 of 2

## Certificate of Testing

<b>Equipment :</b>	DO Meter
<b>Manufacturer :</b>	YSI
<b>Model :</b>	5000-115V
<b>Serial No. :</b>	15E102796
<b>ID No. :</b>	RYG_EN0032
<b>Received Date :</b>	11 February 2022
<b>Test Date :</b>	14 February 2022
<b>Reference :</b>	2202-0404DSC-4
<b>Submitted by :</b>	ALS Laboratory Group (Thailand) Co.,Ltd. (Rayong Branch) 616/10 Moo 5 T.Maenam Khu, A.Pluakdaeng, Rayong 21140, Thailand
<b>Laboratory Condition :</b>	Temperature ( $25 \pm 5$ ) °C Humidity ( $50 \pm 20$ ) %
<b>Test Procedure :</b>	In - house method : CP-CH9 by Comparison Technique with Azide Modification Method
<b>Tested by :</b>	Walalak Sirithean
<b>Approved by :</b>	 Approved Signatory
( ) Malee Butkruea	
( <input checked="" type="checkbox"/> ) Saithip Meangmai	
( ) Warakorn Lernagtrakul	

**Issue Date :** 18 February 2022

REVIEW BY	<u>N. Bannit</u>
APPROVED BY	<u>D. [Signature]</u>
NEXT CAL. DATE	<u>15/8/23</u>



Cert.No.: 22TW34

Page.: 2 of 2

**Result :** Dissolved Oxygen Meter Adjustment With Air 100 %

Dissolved Oxygen Probe No.: 15E100464

Titration Method (Azide Modification Method) (mg/L)	DO Meter Reading (mg/L)	Standard Deviation (mg/L)
8.02	8.02	0.0084

This report was certified only for the instrument we tested. It is allowable to use for study the system efficiency, The environmental impact control and present to organization it may concerned. Intend to use for advertising and referral purpose is prohibited. This report may not be reproduced other in full, without written approval of the laboratory

-o0o-

*Saithip*





TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert. No.: 22LM12

Page.: 1 of 2

## Certificate of Calibration

**Equipment :** DO Meter with Sensor

**Manufacturer :** YSI

**Model :** 5000-115V

**Serial No. :** 15E102796

**ID No. :** RYG\_EN0032

**Submitted by :** ALS Laboratory Group (Thailand) Co.,Ltd. (Rayong Branch)  
616/10 Moo 5 T.Maenam Khu, A.Pluakdaeng,  
Rayong 21140, Thailand

**Location :** TPA On Site Calibration Laboratory

**Received Order :** 11 February 2022


**Calibrated Date :** 21 February 2022

**Ambient Temperature :** ( 26 ± 10 ) °C

**Relative Humidity :** ( 50 ± 30 ) %

**AC Line Voltage :** ( 220 ± 22 ) V

**Calibrated by :** Kunchit Promprat

**Approved by :**   
Approved Signatory

( ) Pornthippa Tameyakul

(✓) Malee Butkruea

( ) Suwit Imjai

**Issue Date :** 21 February 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

A 0038008



Equipment : DO Meter with Sensor

Condition As-Received : Used Item

Reference : 2202-0404DSC-5

Cert. No.: 22LM12

Page.: 2 of 2

**Procedure Used :-**

Calibration were conducted using in-house calibration procedure CP-OT01 according to comparison with Industrial Platinum Resistance Thermometer ( IPRT ) into Temperature Bath.

The temperature scale used was based on ITS-90.

**Condition of this result of calibration**

1. Reference standard instrument:-

<u>Instrument</u>	<u>Model</u>	<u>Serial No.</u>	<u>Cert. No.</u>	<u>Due Date</u>
1) Digital Thermometer	1523	2188080	2111273	22 Nov 2022

2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

**Result of Calibration :-** ( \* ) Without Adjustment

**Function :** Temperature measurement.

This instrument was connected with temperature sensor, S/N.: 15E100464

<u>Calibration Point</u> ( °C )	<u>Immersion Depth</u> ( mm )	<u>Standard Temperature</u> ( °C )	<u>UUC* Reading</u> ( °C )	<u>Error</u> ( °C )	<u>Uncertainty</u> ( ± °C )	<u>Coverage Factor</u> <i>k</i>
20.00	45	20.001	19.88	-0.121	0.15	2.00

**UUC\* :** Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-o0o-

*Mahu*





TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert. No.: 22TM317

Page.: 1 of 3

## Certificate of Calibration

**Equipment :** Low Temp. Incubator

**Manufacturer :** Memmert

**Model :** IPP750

**Serial No. :** V818.0084

**ID No. :** RYG\_EN0154

**Submitted by :** ALS Laboratory Group (Thailand) Co.,Ltd.  
(Rayong Branch)  
616/10 Moo 5 T.Maenam Khu,  
A.Pluakdaeng, Rayong 21140, Thailand

**Location :** BOD Room

**Received Order :** 22 April 2022

**Calibration Date :** 22 April 2022

**Ambient Temperature :** ( 26 ± 10 ) °C

**Relative Humidity :** ( 50 ± 30 ) %

**Calibrated by :** Man Pattanapongpaiboon

REVIEW BY	<i>N. Banvit</i>
APPROVED BY	<i>D. [Signature]</i>
NEXT CAL. DATE	29/10/23

**Approved by :**

*Manu*  
Approved Signatory

- ( ) Pornthippa Tameyakul  
( / ) Malee Butkruea  
( ) Suwit Imjai

**Issue Date :**

3 May 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

A 0040735





Equipment : Low Temp. Incubator

Condition As-Received : Used Item

Reference : 2204-0146OC-1

Cert. No.: 22TM317

Page.: 2 of 3

**Procedure Used :-**

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement

The temperature scale used was based on ITS-90.

**Condition of this result of calibration**

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1 ) Data Acquisition	34970A	MY44031769	21LM12	02 Sep 2022

2. This certificate is valid only to the item calibrated on date and place of calibration.


3. This certification is traceable to the International System of Unit.

**Result of Calibration :-** ( \* ) Without Adjustment

**Function of UUC\* :** Temperature Source

**Fresh air setting :** Close

Environment during calibration		
	Beginning	Finished
Temp. ( °C )	25	25
REL.Humid. ( % )	54	58
AC Supply ( Volt )	221	223



Position :	Ref. Std. ID No.:
1	9RTD-2/1
2	9RTD-2/2
3	9RTD-2/3
4	9RTD-2/4
5	9RTD-2/5
6	9RTD-2/6
7	9RTD-2/7
8	9RTD-2/8
9 (ref.)	9RTD-2/9

**Probe Installation Details :**

a = 10 cm  
b = 10 cm  
c = 10 cm

**Dimension of Chamber :**

D = 0.60 m  
W = 1.0 m  
H = 1.2 m  
Capacity = 0.75 m<sup>3</sup>

*Malu*



Equipment : Low Temp. Incubator  
Condition As-Received : Used Item  
Reference : 2204-0146OC-1  
Result of Calibration :- ( \* ) Without Adjustment  
Function of UUC\* : Temperature Source  
Fresh air setting : Close

Cert. No.: 22TM317

Page.: 3 of 3

Calibration Point ( °C )	UUC* Setting ( °C )	UUC* Reading ( °C )	Temperature stability ( ± °C )	Temperature uniformity ( °C )	Overall Variation ( °C )	Uncertainty ( ± °C )	Coverage Factor <i>k</i>
20.0	20.0	20.0	0.022	0.20	0.22	0.30	2

Calibration Point ( °C )	Measured Temperature ( °C )								
	Position								
	1	2	3	4	5	6	7	8	9 (ref.)
20.0	20.209	20.174	20.199	20.110	20.075	20.062	20.027	20.069	20.030

**Average\*** : The average of 30 values in each position.

**Temperature stability** : One-half of the greatest maximum difference of measured temperature at any one sensor.

**Temperature uniformity** : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

**Overall Variation** : The Difference of the maximum and minimum measured temperatures throughout observation.

**UUC\*** : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-o0o-

Malu





TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3 : EQUIPMENT CALIBRATION AND TESTING SERVICES

534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250

TEL. 0-2717-3000-24 FAX. 0-2719-9484




Cert.No.: 18CG4595

Page.: 1 of 2

## Certificate of Calibration

Equipment :	Burette
Capacity :	50 mL
Serial No. :	-
ID. No. :	243007
Manufacturer :	Witeg
Made in :	Germany
Submitted by :	ALS Laboratory Group (Thailand) Co.,Ltd. Eastern Seaboard Industrial Estate (Rayong) 64/77 Moo 4, Building No.B1, Highway 331, km 91.5 T.Pluakdaeng, A.Pluakdaeng, Rayong 21140
Ambient Temperature :	(22 ± 2.5) °C
Relative Humidity :	(50 ± 10) %
Barometric Pressure :	757 mmHg
Calibration Procedure :	ASTM E 542 - 01
Calibrated by :	Natcha Chayingcheiw



Approved by :

*malee*

Approved Signatory

- ( ☒ ) Pornthippa Tameyakul  
( ☒ ) Malee Butkruea  
( ☐ ) Ponpan Paipim  
( ☐ ) Srisuda Khamtha

Issue Date :

27 September 2018

The Uncertainties are for a confidence probability of approximately 95%.

This certificate may not be reproduced other than in full, except with the prior written approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

A 0087224





Equipment : Burette  
Capacity : 50 mL  
Serial No. : -  
ID. No. : 243007  
Manufacturer : Witeg  
Received Date : 10 September 2018  
Condition As-Received : Used Item  
Calibration Date : 21 September 2018  
Reference : 1809-0411DPC

Cert.No.: 18CG4595

Page.: 2 of 2

**Condition of this result of calibration**

1. Reference Standard Instruments :

<u>Instruments</u>	<u>Model</u>	<u>Serial No.</u>	<u>ID. No.</u>	<u>Certificate No.</u>	<u>Traceability</u>	<u>Due date</u>
1) Balance	XP205DR	1126143764	140RC004	18MM1	NIMT	2 Jan 2019

This certification is traceable to SI Unit

2. This certificate was certified only for the measuring instrument we calibrated.
3. This result of calibration was found accurate as shown on date and place of calibration only.
4. True value is converted to true volume at the standard temperature of 20 °C

**Calibration result :**

Nominal capacity ( mL )	Reading ( mL )	Uncertainty ( $\pm$ mL )	k Factor
50	49.9901	0.010	2.00

**Remark** mL = cm<sup>3</sup>

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor  $k$ , providing a level of confidence of approximately 95 %.

-o0o-

malu

a 0901034

**Sartorius (Thailand) Co., Ltd.**

129 Rama 9 Road, Huaykwang, Huaykwang, Bangkok 10310

Tel: +66 2643 8361-6, e-mail: service.thailand@sartorius.com



NSC-TISI-TIS 17025

CALIBRATION 0426

**SARTORIUS**REVIEW BY Thavitall.APPROVED BY D.NEXT CAL. DATE 01/03/24

# Certificate of Calibration

Model Number : MSE224S-100-DUCertificate No. : 23BCI0112Description : Analytical BalanceIssued Date : Friday, March 03, 2023Serial Number : 0026207038Reference No. : 204833ID No. : RYG\_EN0002Manufacturer : SartoriusPage No. : 1 of 2Customer Name : ALS Laboratory Group (Thailand) Co., Ltd. (Rayong Branch)616/10 Moo 5 T.Maenam Khu, A.Pluak Daeng, Rayong 21140, Thailand.Calibrated Place : ALS Laboratory Group (Thailand) Co., Ltd.(Balance Room)616/10 Moo 5 T.Maenam Khu, A.Pluakdaeng, Rayong.21140, Thailand.Calibrated By : Mr.Chonchai Inthana

Calibration

Calibration Date : Wednesday, March 01, 2023Procedure No. : This calibration was conducted byUsing in-house calibration procedure number (WI-003)Based on UKAS LAB 14 : 2019**Metrological data :**Capacity : 220 g Readability : 0.0001 g**Ambients Conditions:**Temperature : 23.6 °C ± 5.0 °CHumidity : 60.0 % RH ± 10.0 % RHPressure : — ± —**Reasons for calibration**☐ New Installation ☐ Service / Repaired ☒ Re-calibration/ MaintenanceEquipment Condition: ☒ Good Operate ☐ Fair**Measurement Method UKAS Publication Ref :Lab 14**

The measurement uncertainty stated is the expended uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM). The calibration certificate documents the traceability to National Standards, which realise the unit of measurement according to the International Standard System of Units (SI). Report of Tolerance came form list of Sartorius Metrological Specifications.

**Traceability:**

Model Number	Description	Traceability	Certificate No.	Due Date
YCS011-522-00	Sartorius weight set 1mg - 5000g E2,YCS011-522-00	SPC-RT	C02212565	14-Sep-2023
MHB-382SD	Humidity/Barometer/Temp Lutron MHB-382SD	DKSH	C19220444	5-Sep-2023

This certificate relate and apply this equipment only.

This certificate may not be reproduced other than in full except with the prior written approval of the Verification Operation Division Sartorius (Thailand) Co., Ltd.

Mr.chonchai Inthana(Technical Manager)

S  
T  
A  
M  
P



# Certificate of Calibration

Model Number : MSE224S-100-DU

Certificate No. : 23BCI0112

Description : Analytical Balance

Issued Date : Friday, March 03, 2023

Serial Number : 0026207038

Reference No. : 204833

ID No. : RYG\_EN0002

Manufacturer : Sartorius

Page No. : 2 of 2

## Calibration Results : Without Adjustment

### Repeatability



The reproducibility is the ability of a weighing instrument to display nearly identical readouts under constant test conditions when the same load within a measurement series is placed repeatedly on the weighing pan in the same manner. The standard deviation is used to express reproducibility quantitatively.

Nominal Value : (Low Load)	20.0000	199.9999
20 g	20.0000	200.0000
Tolerance	20.0000	199.9999
0.0001 g	20.0000	200.0000
	20.0000	199.9999
Nominal Value : (High Load)	20.0000	199.9999
200 g	19.9999	200.0000
Tolerance	20.0000	200.0000
0.0001 g	20.0000	199.9999
	20.0000	200.0000
Standard Deviation	0.00003	0.00005

### Eccentricity (Off-center loading error)

The off-center loading error is yielded by the difference between the readout of the load, i.e. 1/3 or 1/4 of maximum capacity, placed in the middle of the weighing pan and between each of four additional measurement points (positions defined according to OIML R76).

Nominal value : 100 g  
Tolerance 0.0004 g



	Difference
1	—
2	-0.0001
3	-0.0001
4	0.0001
5	0.0002
6	-

### Linearity

The linearity, also called linearity error. Describes the deviation of the characteristic curve of a weighing instrument from the linear slope.

Tolerance 0.0002 g

Nominal Value (g)	Conventional Mass Value (g)	Displayed Value (g)	Deviation (g)	Uncertainty (g)
0.01	0.0100	0.0100	0.0000	0.00014
0.05	0.0500	0.0500	0.0000	0.00014
0.1	0.1000	0.1000	0.0000	0.00014
0.5	0.5000	0.5000	0.0000	0.00014
1	1.0000	1.0000	0.0000	0.00014
5	5.0000	5.0000	0.0000	0.00014
10	10.0000	10.0001	0.0001	0.00014
20	20.0000	20.0000	0.0000	0.00024
50	50.0000	50.0000	0.0000	0.00015
100	100.0000	99.9999	-0.0001	0.00019
200	200.0000	200.0000	0.0000	0.00032

End of Report.





TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert. No.: 22TM1517

Page : 1 of 3

## Certificate of Calibration

<b>Equipment :</b>	Hot Air Oven
<b>Manufacturer :</b>	Memmert
<b>Model :</b>	UFE 500
<b>Serial No. :</b>	G511.1572
<b>ID No. :</b>	RYG_EN0010
<b>Submitted by :</b>	ALS Laboratory Group (Thailand) Co.,Ltd. (Rayong Branch) 616/10 Moo 5 T. Maenam Khu, A. Pluakdaeng, Rayong 21140 Thailand
<b>Location :</b>	Oven Room
<b>Received Order :</b>	20 October 2022
<b>Calibration Date :</b>	20 October 2022
<b>Ambient Temperature :</b>	( 26 ± 10 ) °C
<b>Relative Humidity :</b>	( 50 ± 30 ) %
<b>Calibrated by :</b>	Man Pattanapongpaiboon

REVIEW BY	Thanitall.
APPROVED BY	D. [Signature]
NEXT CAL. DATE	30/04/24

Approved by :

*Manu.*

Approved Signatory

- ( / ) Pornthippa Tameyakul  
( / ) Malee Butkruea  
( ) Suwit Imjai

Issue Date :

2 November 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.



Equipment : Hot Air Oven  
 Condition As-Received : Used Item  
 Reference : 2210-0376OC-2

Cert. No.: 22TM1517  
 Page : 2 of 3

**Procedure Used :-**

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector ( RTD ) and Thermocouple Type T.

The temperature scale used was based on ITS-90.

**Condition of this result of calibration**

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1 ) Data Acquisition	34972A	MY49023932	22LM97	29 Jul 2023


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

**Result of Calibration :-** ( \* ) Without Adjustment

**Function of UUC\* :** Temperature Source

**Fresh air setting :** Close



Environment during calibration		
	Beginning	Finished
Temp. ( °C )	25	25
REL.Humid. ( % )	54	59
AC Supply ( Volt )	223	225

**Ref. Std. ID No.: @ Calibration Point**

Position :	( 180 ) °C	( 104 ) °C
1	21-16TC-01	20-16RTD-01
2	21-16TC-02	20-16RTD-02
3	21-16TC-03	20-16RTD-03
4	21-16TC-04	20-16RTD-04
5	21-16TC-05	22-16RTD-05
6	21-16TC-06	20-16RTD-06
7	21-16TC-07	20-16RTD-07
8	21-16TC-08	22-16RTD-08
9 (ref.)	21-16TC-09	22-16RTD-09

**Probe Installation Details :**

**Dimension of Chamber :**

a =	5.0	cm	D =	0.40	m
b =	5.0	cm	W =	0.56	m
c =	5.0	cm	H =	0.48	m
			Capacity =	0.11	m <sup>3</sup>

*Malu*





Equipment : Hot Air Oven  
Condition As-Received : Used Item  
Reference : 2210-0376OC-2  
Result of Calibration :- ( \* ) Without Adjustment  
Function of UUC\* : Temperature Source  
Fresh air setting : Close

Cert. No.: 22TM1517

Page : 3 of 3

Calibration Point ( °C )	UUC* Setting ( °C )	UUC* Reading ( °C )	Temperature stability ( ± °C )	Temperature uniformity ( °C )	Overall Variation ( °C )	Uncertainty ( ± °C )	Coverage Factor <i>k</i>
104.0	104.0	104.0	0.076	0.52	0.60	0.42	2
180.0	180.0	180.0	0.13	0.88	1.2	1.1	2

Calibration Point ( °C )	Measured Temperature ( °C )								
	Position								
	1	2	3	4	5	6	7	8	9 (ref.)
104.0	103.768	103.734	103.723	103.800	104.215	104.131	104.132	103.740	103.747
180.0	179.723	179.359	179.439	179.489	180.361	180.114	180.131	180.243	179.605

**Average\*** : The average of 30 values in each position.

**Temperature stability** : One-half of the greatest maximum difference of measured temperature at any one sensor

**Temperature uniformity** : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

**Overall Variation** : The Difference of the maximum and minimum measured temperatures throughout observation

**UUC\*** : Unit Under Calibration

**Note** : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-o0o-

*Maka .*





TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert.No.: 22CH283

Page.: 1 of 2

## Certificate of Calibration

<b>Equipment :</b>	Conductivity Meter
<b>Manufacturer :</b>	Mettler Toledo
<b>Model :</b>	S230
<b>Serial No. :</b>	B241407147
<b>ID No. :</b>	RYG_EN0029
<b>Condition As-Received:</b>	Used Item
<b>Received Date :</b>	22 February 2022
<b>Calibration Date :</b>	23 February 2022
<b>Reference :</b>	2202-0732DSC-1
<b>Submitted by :</b>	ALS Laboratory Group (Thailand) Co.,Ltd. (Rayong Branch) 616/10 Moo 5 T.Maenam Khu, A.Pluakdaeng, Rayong 21140, Thailand
<b>Ambient Temperature :</b>	(25 $\pm$ 2.5) °C
<b>Relative Humidity :</b>	(50 $\pm$ 15) %
<b>Calibration Procedure:</b>	In -house method : - CP-CH6 : based on direct measurement by using certified reference material (CRM)
<b>Calibrated by :</b>	Walalak Sirithean
<b>Approved by :</b>	<u>Malee Butkruea</u> Approved Signatory
	( <input checked="" type="checkbox"/> ) Malee Butkruea ( <input type="checkbox"/> ) Saithip Meangmai ( <input type="checkbox"/> ) Warakorn Lernagtrakul
<b>Issue Date :</b>	25 February 2022



The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.



Cert.No.: 22CH283

Page.: 2 of 2

**Condition of this result of calibration**

## 1. Reference Standard Instrument :-

<u>Instrument</u>	<u>Serial No.</u>	<u>ID No.</u>	<u>Certificate No.</u>	<u>Due date</u>
1) Thermometer	9549224	130RC003	211451	15 Apr 2022

This certification is traceable to the International System of Unit maintained at:-

- Traceable to National Institute of Metrology (Thailand), NIMT

## 2. Certified Reference Materials :-

- Conductivity calibration solution, CPA chem Ltd., The measurement results are traceable to SI through CPA chem Ltd., ANSI-ASQ National Accreditation Board, Accredited No. AR-1835

<u>Conductivity Solution</u>	<u>Manufacturer</u>	<u>Lot No.</u>	<u>Exp. date</u>
84.000 $\mu\text{S/cm}$	CPA Chem	754034	28 June 2022
1413.0 $\mu\text{S/cm}$	CPA Chem	766815	04 Sep 2022
12.880 mS/cm	CPA Chem	761022	02 Aug 2022

- Control Conductivity calibration solution temperature by Water bath ( $25 \pm 0.1$ )  $^{\circ}\text{C}$

## 3. This certificate is valid only to the item calibrated on date and place of calibration.

**Calibration results****Function : Conductivity Measurement**(\*) After Adjustment at 1413.0  $\mu\text{S/cm}$ 

Conductivity Electrode Serial No.: 5821441030

Standard Conductivity Solution	Before Adjustment UUC* Reading	After Adjustment UUC* Reading	Uncertainty of Measurement ( $\pm$ )	Coverage factor k
84.000 $\mu\text{S/cm}$	82.4 $\mu\text{S/cm}$	84.4 $\mu\text{S/cm}$	0.62 $\mu\text{S/cm}$	2.00
1413.0 $\mu\text{S/cm}$	1375 $\mu\text{S/cm}$	1413 $\mu\text{S/cm}$	9.2 $\mu\text{S/cm}$	2.00
12.880 mS/cm	12.54 mS/cm	12.81 mS/cm	0.086 mS/cm	2.00

**Remark**

- UUC\* = Unit Under Calibration

- Cell constant =  $0.555236 \text{ cm}^{-1}$ 

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor  $k$ , providing a level of confidence of approximately 95 %.

-o0o-

Malu

a 1090534






TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert. No.: 22TM1492

Page : 1 of 3

## Certificate of Calibration

<b>Equipment :</b>	Hot Air Oven
<b>Manufacturer :</b>	Memmert
<b>Model :</b>	UM 400
<b>Serial No. :</b>	b495.0899
<b>ID No. :</b>	RYG_EN0006
<b>Submitted by :</b>	ALS Laboratory Group (Thailand) Co.,Ltd. (Rayong Branch) 616/10 Moo 5, T. Maenam Khu, A. Pluakdaeng, Rayong 21140, Thailand
<b>Location :</b>	Oven Room
<b>Received Order :</b>	20 October 2022
<b>Calibration Date :</b>	20 October 2022
<b>Ambient Temperature :</b>	( 26 ± 10 ) °C
<b>Relative Humidity :</b>	( 50 ± 30 ) %
<b>Calibrated by :</b>	Preecha Hlahib
<b>Approved by :</b>	<div style="text-align: center;">             _____            Approved Signatory         </div> <div style="margin-top: 10px;"> <input checked="" type="checkbox"/> Pornthippa Tameyakul  <input checked="" type="checkbox"/> Malee Butkruea  <input type="checkbox"/> Suwit Imjai         </div>



The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.





Equipment : Hot Air Oven  
 Condition As-Received : Used Item  
 Reference : 2210-0376OC-1

Cert. No.: 22TM1492  
 Page : 2 of 3

**Procedure Used :-**

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector ( RTD ).

The temperature scale used was based on ITS-90.

**Condition of this result of calibration**

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1 ) Data Acquisition	34970A	MY44035217	21LM30	23 Dec 2022

2. This certificate is valid only to the item calibrated on date and place of calibration.


3. This certification is traceable to the International System of Unit.

**Result of Calibration :-** ( \* ) Without Adjustment

**Function of UUC\* :** Temperature Source

**Fresh air setting :** Close

Environment during calibration		
	Beginning	Finished
Temp. ( °C )	28	29
REL.Humid. ( % )	43	47
AC Supply ( Volt )	220	221



Position :	Ref. Std. ID No.:
1	18-10RTD-01
2	18-10RTD-02
3	18-10RTD-03
4	18-10RTD-04
5	18-10RTD-05
6	18-10RTD-06
7	18-10RTD-07
8	18-10RTD-08
9 (ref.)	18-10RTD-09

**Probe Installation Details :**

a = 5.0 cm  
 b = 5.0 cm  
 c = 5.0 cm

**Dimension of Chamber :**

D = 0.33 m  
 W = 0.40 m  
 H = 0.40 m  
 Capacity = 0.053 m<sup>3</sup>

*Malu.*



Equipment : Hot Air Oven  
Condition As-Received : Used Item  
Reference : 2210-0376OC-1  
Result of Calibration :- ( \* ) Without Adjustment  
Function of UUC\* : Temperature Source  
Fresh air setting : Close

Cert. No.: 22TM1492  
Page : 3 of 3

Calibration Point ( °C )	UUC* Setting ( °C )	UUC* Reading ( °C )	Temperature stability ( ± °C )	Temperature uniformity ( °C )	Overall Variation ( °C )	Uncertainty ( ± °C )	Coverage Factor <i>k</i>
70.0	70.0	70.0	0.079	0.47	0.77	0.42	2

Calibration Point ( °C )	Measured Temperature ( °C )								
	Position								
	1	2	3	4	5	6	7	8	9 (ref.)
70.0	70.262	69.995	70.079	70.177	70.664	70.039	70.688	70.149	70.328

**Average\*** : The average of 30 values in each position.

**Temperature stability** : One-half of the greatest maximum difference of measured temperature at any one sensor

**Temperature uniformity** : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

**Overall Variation** : The Difference of the maximum and minimum measured temperatures throughout observation

**UUC\*** : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-o0o-

Malu.





TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert. No.: 22TM1491

Page : 1 of 3

## Certificate of Calibration

<b>Equipment :</b>	Water Bath
<b>Manufacturer :</b>	Memmert
<b>Model :</b>	WNB22
<b>Serial No. :</b>	L513.0648
<b>ID No. :</b>	RYG_EN0061
<b>Submitted by :</b>	ALS Laboratory Group (Thailand) Co.,Ltd. (Rayong Branch) 616/10 Moo 5, T. Maenam Khu, A. Pluakdaeng, Rayong 21140, Thailand
<b>Location :</b>	Wet Chemistry Lab
<b>Received Order :</b>	20 October 2022
<b>Calibration Date :</b>	20 October 2022
<b>Ambient Temperature :</b>	( 26 ± 10 ) °C
<b>Relative Humidity :</b>	( 50 ± 30 ) %
<b>Calibrated by :</b>	Preecha Hlahib



Approved by :

Malee

Approved Signatory

- ( ) Pornthippa Tameyakul  
( ☒ ) Malee Butkruea  
( ) Suwit Imjai

Issue Date :

2 November 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.





**Equipment :** Water Bath  
**Condition As-Received :** Used Item  
**Reference :** 2210-0376OC-4  
**Procedure Used :-**

**Cert. No.:** 22TM1491  
**Page :** 2 of 3

Calibration were conducted using in-house calibration procedure CP-OT04 according to direct measurement method with Data Acquisition which connected with Industrial Platinum Resistance Thermometer ( IPRT ).

The temperature scale used was based on ITS-90.

**Condition of this result of calibration**

1. Reference standard instrument:-

<u>Instrument</u>	<u>Model</u>	<u>Serial No.</u>	<u>Cert. No.</u>	<u>Due Date</u>
1 ) Data Acquisition	34970A	MY44035217	21LM30	23 Dec 2022


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

**Result of Calibration :-** ( \* ) Without Adjustment

**Function of UUC\* :** Temperature Source

	Environmental		AC Voltage Supply
	( °C )	( %R.H. )	( Volt )
Beginning of Calibration	24	53	222
Finished of Calibration	24	50	221



Front

Position :	Ref. Std. S/N.:
1	N37P300726
2	N37P300727
3	N37P300728
4	N37P300729
5(ref.)	N37P300730

*Malu*



Equipment : Water Bath  
Condition As-Received : Used Item  
Reference : 2210-0376OC-4  
Result of Calibration :- ( \* ) Without Adjustment  
Function of UUC\* : Temperature Source

Cert. No.: 22TM1491  
Page : 3 of 3

Calibration point ( °C )	UUC* Setting ( °C )	UUC* Reading ( °C )	Average* Standard Reading ( °C )				
			Position				
			1	2	3	4	5 (ref.)
85.0	85.0	85.0	84.527	84.563	84.628	84.516	84.580

Calibration point ( °C )	Uniformity ( °C )	Stability ( ± °C )	Uncertainty ( ± °C )	Coverage Factor <i>k</i>
85.0	0.12	0.081	0.18	2

**Average\*** : The average of 30 values in each position.

**Uniformity** : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

**Stability** : One-half of the greatest maximum difference of measured temperature at any one probe.

**UUC\*** : Unit Under Calibration

**Note** : The reported uncertainty of measurement was included stability and excluded uniformity.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-o0o-

Malu.





TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert.No.: 22CH1734

Page.: 1 of 3

## Certificate of Calibration

Equipment :	pH Meter
Manufacturer :	Mettler Toledo
Model :	Seven2Go
Serial No. :	B628755984
ID No. :	RYG_FS0392
Condition As-Received:	Used Item
Received Date :	21 December 2022
Calibration Date :	22 December 2022
Reference :	2212-0602DSC-3
Submitted by :	ALS Laboratory Group (Thailand) Co.,Ltd. Rayong Branch 616/10 Moo 5 T.Maenam Khu, A.Pluakdaeng, Rayong 21140, Thailand
Ambient Temperature :	(25 ± 2.5) °C
Relative Humidity :	(50 ± 15) %
Calibration Procedure :	In - house method : - CP-CH5 by direct measurement with standard voltage calibrator and direct measurement with certified reference material (CRM) - CP-CH8 by comparison with standard thermometer



Calibrated by : Warakorn Lernagatrakul

Approved by :

*Malee*

Approved Signatory

- (☒) Malee Butkruea  
( ) Saithip Meangmai  
( ) Warakorn Lernagatrakul

Issue Date : 26 December 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

A 0048759





Cert.No.: 22CH1734

Page.: 2 of 3

**Condition of this calibration result**

1. Reference Standard Instrument : -

<u>Instrument</u>	<u>Serial No.</u>	<u>ID No.</u>	<u>Cert. No.</u>	<u>Due Date</u>
1) Document Process Calibrator	54030049	130RC116	22E2769	24 Aug 2023
2) Ref. Standard Thermometer	4982054	110RC044	22I1306	27 Oct 2023

This certification is traceable to the International System of Unit maintained at:-

- Traceable to National Institute of Metrology (Thailand), NIMT

2. Certified Reference Materials : The measurement results are traceable to SI through CPA chem Ltd.,  
ANSI-ASQ National Accreditation Board, Accredited No. AR-1835

<u>Buffer Solution</u>	<u>Manufacturer</u>	<u>Lot No.</u>	<u>Exp. date</u>
pH 4.008	CPA chem	826588	09 July 2024
pH 6.987	CPA chem	823322	20 June 2023
pH 10.008	CPA chem	826590	09 July 2023

3. This certificate is valid only to the item calibrated on date and place of calibration.

**Calibration Results**

**Function : mV Measurement**

**Performing standard curve by Fluke at pH (4,7,10)**

Unit Under Calibration	Nominal Value	Standard Voltage Input	Actual Reading		Uncertainty of Measurement ( ±mV )	Coverage factor k
	pH	mV	mV	pH		
pH Meter S/N.: B628755984	4.00	177.48	178	4.00	0.58	2.00
	7.00	0.00	0	7.00	0.58	2.00
	10.00	-177.48	-177	10.00	0.58	2.00

*Malu*



Cert.No.: 22CH1734

Page.: 3 of 3

**Calibration Results****Function : pH Measurement**

Performing three buffers standard curve by using buffer nominal pH (4,7,10)

Unit Under Calibration	Standard pH Buffer Solution	Actual pH Reading	Actual mV Reading ( mV )	Uncertainty of pH measurement ( $\pm$ )	Coverage factor $k$
pH Electrode	4.008	4.01	165	0.0079	2.00
S/N.: 0281238	6.987	6.99	-6	0.011	2.00
	10.008	10.01	-180	0.0097	2.00

**Function : Temperature Measurement****( \* ) Without adjustment**

This equipment was connected with Temperature Probe;

- Model : InLab Expert Go-ISM  
- Serial No. : 0281238

Dimension of probe;

- Length : 120 mm.  
- Diameter : 12 mm.  
- Immersion Depth : 100 mm.

Calibration Point ( $^{\circ}\text{C}$ )	Standard Temperature ( $^{\circ}\text{C}$ )	UUC* Reading ( $^{\circ}\text{C}$ )	Error ( $^{\circ}\text{C}$ )	Uncertainty of measurement ( $\pm$ $^{\circ}\text{C}$ )	Coverage factor $k$
25.0	25.002	25.2	0.198	0.13	2.00
30.0	30.001	30.2	0.199	0.13	2.00
40.0	40.003	40.2	0.197	0.13	2.00

**Remark : - UUC\* = Unit Under Calibration**

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor  $k$ , providing a level of confidence of approximately 95 %.

-o0o-

Mahu

a 1141164



REVIEW BY

Nont Somb

APPROVED BY

KL AL

NEXT CAL. DATE

21/12/23

## Certificate of System Qualification

GC-OQ + GCMS-OQ

System ID: GM-7  
Organization Name: ALS Laboratory Group (Thailand) Co., Ltd.  
Organization Location: 104 Patthanakarn 40, Patthanakarn rd., Khwang Suan Luang, Khet Suan Luang, Bangkok 10250  
Date: June 21, 2022 2:04:12 PM  
EQP Name: AgilentRecommended , AgilentRecommended  
EQP Revision: GC.02.50, GCMS.02.50  
Overall Qualification Status: Pass

## System Inspection and Basic Safety and Operation

Name: 7890

Setpoint Status: Pass

## Overall System Inspection and Basic Safety and Operation Test Status

Pass

## Inlet Pressure Accuracy

Name: 7890

Front SSL

Setpoint Status: Pass

	Setpoint		Actual	
Inlet Pressure:	25.0	psi	25.0	psi
Accuracy:			0.0	psi
Agilent Recommended:			<= 1.2	

## Overall Inlet Pressure Accuracy Test Status

Pass

## GC Oven Temperature Accuracy

Name: 7890

Date: June 21, 2022 2:04:12 PM  
System ID: GM-7



**Setpoint Status:**

Pass

Zone:

Oven

Setpoint/Actual

Temperature:

230.0

230.0

°C

Accuracy:

0.0

°C

Agilent Recommended:

&gt;=

-1.0

% setpoint in K

(

-5.0

°C

)

&lt;=

1.0

% setpoint in K

(

5.0

°C

)

**Setpoint Status:**

Pass

Zone:

Oven

Setpoint/Actual

Temperature:

100.0

100.4

°C

Accuracy:

0.4

°C

Agilent Recommended:

&gt;=

-1.0

% setpoint in K

(

-3.7

°C

)

&lt;=

1.0

% setpoint in K

(

3.7

°C

)

**Overall GC Oven Temperature Accuracy Test Status**

Pass

**GC Oven Temperature Stability**

Name:

7890

**Setpoint Status:**

Pass

Setpoint/Average

Temperature:

100.0

100.0333

°C

Stability:

0.1

°C

Agilent Recommended:

&lt;=

0.5

**Overall GC Oven Temperature Stability Test Status**

Pass

**Log Amp**

Tested Combination1

Front

SSL

/ External

SQ

Name:

5977A

**Setpoint Status:**

Pass

Date:

June 21, 2022 2:04:12 PM

System ID:

GM-7

## Overall Log Amp Test Status

Pass

## RFPA

Tested Combination1

Front

SSL

/ External

SQ

Name:

5977A

Setpoint Status:

Pass

Amu:

1050

m/z

Drift After Five Minutes:

22

mV

RFPA Voltage:

568

mV

Agilent Recommended:

>=

-100

and

<=

100

<=

1100

## Overall RFPA Test Status

Pass

## Tune EI

Tested Combination1

Front

SSL

/ External

SQ

Name:

5977A

Setpoint Status:

Pass

Filament:

1

Setpoint Status:

Pass

Filament:

2

## Overall Tune EI Test Status

Pass

## Signal to Noise EI

Tested Combination1

Front

SSL

/ External

SQ

Name:

5977A

Date:

June 21, 2022 2:04:12 PM

System ID:

GM-7

Source:  Filament:

Setpoint Status:

Signal to Noise:

Agilent Recommended:

Source:  Filament:

Setpoint Status:

Signal to Noise:

Agilent Recommended:

This test's 0 comment(s) and 1 deviation(s) are available in the Attachments section.

**Overall Signal to Noise EI Test Status**



## Instrument Details

### Purpose

This section describes the as found system configuration.

### Details

#### System

System ID	GM-7
Manufacturer	Agilent Technologies
Name	7890

#### Tested Combination1

Injection Technique	Manual Injection
Inlet	Front
Detector	External
LTM Included?	No

#### Sampler 1

Manufacturer	Agilent Technologies
Type	Manual Injection
Usage	Sample Injection
Syringe Volume (µL)	10

#### Mainframe 1

Manufacturer	Agilent Technologies
Name	7890
Model Number	G3442B
Serial Number	CN14133181
Firmware Revision	B.02.03
Oven Type	Standard

## Inlet 1

Manufacturer	Agilent Technologies
Name	7890
Type	SSL
Location	Front
Carrier Gas	Helium
Control Type	Electronic Pressure Control (EPC)
Purged Inlet	Yes

## Detector 1

Manufacturer	Agilent Technologies
Name	Mass Spectrometer
Type	Mass Spectrometer
Location	External

## Mass Spectrometer 1

Manufacturer	Agilent Technologies
Type	SQ
Name	5977A
Serial Number	US1415M209
Firmware Revision	5977 6.00.21
High Vacuum System	Turbo Pump
Scouting Run Standard	OFN Std

## MS EI Source 1

Manufacturer	Agilent Technologies
Source Type	EI - Extractor
Number of filaments	2

## Electronic Signature

### Purpose

This signature page was created and published because the ACE sign-off action was executed, which is valid for the entire document, including attachments. The ACE sign-off is an electronic signature that requires two distinct identification components: unique username and personal password. The Agilent representative who has delivered this service understands the meaning and legal status of an electronic signature. As a trained official operator, the Agilent representative has a unique password and logon to access ACE and electronically sign this document. (Other e-signatures can be applied to this document using a Document Content Management or other suitable method defined in your data access and control procedures.)

### Details

Full Name of Signer:	Supasak Nimsongtham
Logged On User Name:	supasak.nimsongtham@agilent.com
Signature Creation Date:	June 21, 2022
Reason for Signature:	Executed protocol and published this original version of document

### Regulatory Disclaimer

This document provides a protocol to verify and record instrument configuration and evidence of proper operation. It has been prepared from our interpretation of applicable regulations as well as industry best practices. The document is designed to provide an important component of a complete compliance package. Validation depends upon many factors and use of this protocol alone does not assure compliance. Agilent Technologies makes no promises or representations as to its sufficiency for any specific regulatory program.

### Warranty

Agilent Technologies makes no warranty of any kind to this material, including but not limited to, the implied warranties or merchantability and fitness for a particular purpose. Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

---

Date:	June 21, 2022 2:04:12 PM
System ID:	GM-7



User Name: supasak.nimsongtham  
 Hostname: 5CG1115HKC

System Id: GM-7  
 Print Date: June 21, 2022 2:04:17 PM

ALS-GM7-2022 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 21, 2022 10:25:05 AM	Audit	SessionCreated	Session	None
June 21, 2022 10:25:05 AM	Start	Configuration	Session	None
June 21, 2022 10:25:05 AM	Audit	Entitlement	Licensing	User is FieldEngineer and does not require an unlock code
June 21, 2022 10:25:26 AM	Audit	EqpLoaded	Session	EQP details for primary technique [Gc] - File path: [ProtocolPacks/Gc/Configurations/02.50/Gc.02.50.eqp], EQP File Name: [Gc.02.50.eqp], EQP Name: [AgilentRecommended] EQP details for hyphenated technique [GcMs] - File path: [ProtocolPacks/GcMs/Configurations/02.50/GcMs.02.50.eqp], EQP File Name: [GcMs.02.50.eqp], EQP Name: [AgilentRecommended]
June 21, 2022 10:25:39 AM	End	Configuration	Session	None
June 21, 2022 10:25:43 AM	Start	Qualification	Session	OQ
June 21, 2022 10:25:43 AM	Start	Execution	System Inspection and Basic Safety and Operation - 7890: - Qualitative Test - No setpoints associated	None
June 21, 2022 10:25:54 AM	End	Execution	System Inspection and Basic Safety and Operation - 7890: - Qualitative Test - No setpoints associated	Run Count : 1

User Name: supasak.nimsongtham  
 Hostname: 5CG1115HKC

System Id: GM-7  
 Print Date: June 21, 2022 2:04:17 PM

## ALS-GM7-2022 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 21, 2022 10:26:00 AM	Start	Execution	Inlet Pressure Accuracy - Front SSL: - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	None
June 21, 2022 10:26:10 AM	End	Execution	Inlet Pressure Accuracy - Front SSL: - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	Run Count : 1
June 21, 2022 10:26:12 AM	Start	Execution	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	None
June 21, 2022 10:34:09 AM	Audit	Data	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Manual Data Entry
June 21, 2022 10:34:10 AM	End	Execution	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Run Count : 1
June 21, 2022 10:34:11 AM	Start	Execution	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	None
June 21, 2022 10:38:42 AM	Audit	Data	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Manual Data Entry
June 21, 2022 10:38:44 AM	End	Execution	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Run Count : 1
June 21, 2022 10:38:46 AM	Start	Execution	GC Oven Temperature Stability - 7890: - Temperature : Oven - S: 100.0°C - L: <= 0.5°C	None

Page 2 / 8

Date: June 21, 2022 2:04:12 PM  
 System ID: GM-7

User Name: supasak.nimsongtham  
 Hostname: 5CG1115HKC

System Id: GM-7  
 Print Date: June 21, 2022 2:04:17 PM

## ALS-GM7-2022 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 21, 2022 11:01:00 AM	Audit	AceClosed	Session	None
June 21, 2022 11:01:47 AM	Audit	AceRestarted	Session	None
June 21, 2022 11:01:48 AM	Audit	SessionReloaded	Session	None
June 21, 2022 11:01:51 AM	Start	Qualification	Session	OQ
June 21, 2022 11:01:51 AM	Start	Execution	GC Oven Temperature Stability	None
			- 7890: - Temperature : Oven -	
			S: 100.0°C - L: <= 0.5°C	
June 21, 2022 11:03:14 AM	Audit	Data	DataManager	DataManager was in a data verification state but the user chose to start over.
June 21, 2022 11:04:19 AM	Audit	Data	GC Oven Temperature Stability	Manual Data Entry
			- 7890: - Temperature : Oven -	
			S: 100.0°C - L: <= 0.5°C	
June 21, 2022 11:04:22 AM	End	Execution	GC Oven Temperature Stability	Run Count : 1
			- 7890: - Temperature : Oven -	
			S: 100.0°C - L: <= 0.5°C	
June 21, 2022 11:04:24 AM	Start	Execution	Log Amp - 5977A SQ: - Source:	None
			EI - Extractor	
June 21, 2022 11:04:34 AM	End	Execution	Log Amp - 5977A SQ: - Source:	Run Count : 1
			EI - Extractor	
June 21, 2022 11:04:37 AM	Start	Execution	RFP A - 5977A SQ: - Source:	EI None
			- Extractor	
June 21, 2022 11:07:49 AM	End	Execution	RFP A - 5977A SQ: - Source:	EI Run Count : 1
			- Extractor	
June 21, 2022 11:07:52 AM	Start	Execution	Tune EI - 5977A SQ: - Source:	None
			EI - Extractor Filament 1	
			(Qualitative - No setpoints associated)	



User Name: supasak.nimsongtham  
 Hostname: SCG1115HKC

System Id: GM-7  
 Print Date: June 21, 2022 2:04:17 PM

## ALS-GM7-2022 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 21, 2022 11:08:35 AM	End	Execution	Tune EI - 5977A SQ: - Source: - Run Count : 1 EI - Extractor Filament 1 (Qualitative - No setpoints associated)	
June 21, 2022 11:14:59 AM	Start	Execution	Tune EI - 5977A SQ: - Source: - None EI - Extractor Filament 2 (Qualitative - No setpoints associated)	
June 21, 2022 11:16:48 AM	End	Execution	Tune EI - 5977A SQ: - Source: - Run Count : 1 EI - Extractor Filament 2 (Qualitative - No setpoints associated)	
June 21, 2022 11:16:49 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None
June 21, 2022 11:17:05 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	None
June 21, 2022 11:17:10 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None
June 21, 2022 11:26:09 AM	Audit	AceClosed	Session	None
June 21, 2022 12:36:20 PM	Audit	AceRestarted	Session	None
June 21, 2022 12:36:22 PM	Audit	SessionReloaded	Session	None
June 21, 2022 12:36:26 PM	Start	Qualification	Session	OQ
June 21, 2022 12:36:26 PM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None

User Name: supasak.nimsongtham  
 Hostname: SCG1115HKC

System Id: GM-7  
 Print Date: June 21, 2022 2:04:17 PM

## ALS-GM7-2022 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 21, 2022 12:37:07 PM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	None
June 21, 2022 12:37:08 PM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None
June 21, 2022 12:38:54 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data files Path : H:\ALSGM7_2022\SNF1_001.D
June 21, 2022 12:39:24 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data files Path : H:\ALSGM7_2022\SNF1_001.D
June 21, 2022 12:40:09 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data files Path : H:\ALSGM7_2022\SNF1_001.D
June 21, 2022 12:42:04 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data files Path : H:\ALSGM7_2022\SNF1_001.D
June 21, 2022 12:42:17 PM	Audit	AceClosed	Session	None
June 21, 2022 12:33:31 PM	Audit	AceRestarted	Session	None
June 21, 2022 12:33:33 PM	Audit	SessionReloaded	Session	None
June 21, 2022 12:33:37 PM	Start	Qualification	Session	OQ
June 21, 2022 12:33:37 PM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None

User Name: supasak.nimsongtham  
 Hostname: 5CG1115HKC

System Id: GM-7  
 Print Date: June 21, 2022 2:04:17 PM

ALS-GM7-2022 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 21, 2022 12:34:44 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data files Path : E:\ALSGM7_2022\SNF1_001.D
June 21, 2022 12:36:26 PM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Run Count : 1
June 21, 2022 12:37:11 PM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	None
June 21, 2022 12:38:15 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data files Path : E:\ALSGM7_2022\SNF2_001.D
June 21, 2022 12:38:30 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data files Path : E:\ALSGM7_2022\SNF2_001.D
June 21, 2022 12:38:45 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data files Path : E:\ALSGM7_2022\SNF2_001.D
June 21, 2022 12:39:00 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data files Path : E:\ALSGM7_2022\SNF2_001.D
June 21, 2022 12:39:14 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data files Path : E:\ALSGM7_2022\SNF2_001.D



User Name: supasak.nimsongtham  
 Hostname: 5CG1115HKC

System Id: GM-7  
 Print Date: June 21, 2022 2:04:17 PM

## ALS-GM7-2022 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 21, 2022 12:39:45 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data files Path : E:\ALSGM7_2022\SNF2_001.D
June 21, 2022 12:40:16 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data files Path : E:\ALSGM7_2022\SNF2_001.D
June 21, 2022 12:40:40 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data files Path : E:\ALSGM7_2022\SNF2_001.D
June 21, 2022 12:41:09 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data files Path : E:\ALSGM7_2022\SNF2_001.D
June 21, 2022 12:41:29 PM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Run Count : 1
June 21, 2022 12:42:30 PM	Audit	TestUnlocked	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Deviation filed for Run Count : 1
June 21, 2022 12:42:30 PM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	None
June 21, 2022 12:42:35 PM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Data files Path : E:\ALSGM7_2022\SNF2_001.D

User Name: supasak.nimsongtham  
Hostname: 5CG1115HKC

System Id: GM-7  
Print Date: June 21, 2022 2:04:17 PM

## ALS-GM7-2022 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 21, 2022 12:42:45 PM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	Run Count : 2
June 21, 2022 12:42:50 PM	End	Qualification	Session	OQ
June 21, 2022 12:42:50 PM	Start	Reporting	Session	None
June 21, 2022 12:45:17 PM	Audit	AceClosed	Session	None
June 21, 2022 1:57:47 PM	Audit	AceRestarted	Session	None
June 21, 2022 1:57:50 PM	Audit	SessionReloaded	Session	None
June 21, 2022 1:57:56 PM	Start	Qualification	Session	OQ
June 21, 2022 2:02:42 PM	Audit	Reporting	Session	Report Generated : Certificate

REVIEW BY	Nant Sot
APPROVED BY	LL AL
NEXT CAL. DATE	25/05/23

# Certificate of System Qualification

GC-OQ + GCMS-OQ

System ID: GM-6  
Organization Name: ALS Laboratory Group(Thailand) Co., Ltd.  
Organization Location: 104 Patthanakarn 40, Patthanakarn Rd., Kwang Suan Luang< Khet Suan Luang, Bangkok 10250  
Date: November 25, 2021 5:20:10 PM  
EQP Name: AgilentRecommended , AgilentRecommended  
EQP Revision: GC.02.52, GCMS.02.51  
Overall Qualification Status: Pass

## CDS Logon Verification - GC

Logon: Nanthawadee.Somboon

## Overall CDS Logon Verification - GC Test Status

Pass

## System Inspection and Basic Safety and Operation

Name: 7890

Setpoint Status: Pass

## Overall System Inspection and Basic Safety and Operation Test Status

Pass

## Inlet Pressure Accuracy

Name: 7890

Front SSL

Setpoint Status: Pass

	Setpoint	Actual
Inlet Pressure:	25.0 psi	25.1 psi
Accuracy:		0.1 psi
Agilent Recommended:	<=	1.2

Date: November 25, 2021 5:20:10 PM  
System ID: GM-6



## Overall Inlet Pressure Accuracy Test Status

Pass

## Headspace Leak

Name:

7697A with Tray

Sampler 1

Setpoint Status:

Pass

## Overall Headspace Leak Test Status

Pass

## Headspace Heated Zones Temperature Accuracy

Name:

7697A with Tray

Sampler 1

Setpoint Status:

Pass

Zone:

Transferline

Temperature:

Setpoint

115.0

°C

Actual

114.9

Accuracy:

-0.1

°C

Agilent Recommended:

&gt;=

-1.8

% setpoint

(

-2.1

°C

)

&lt;=

5.2

% setpoint

(

6.0

°C

)

Setpoint Status:

Pass

Zone:

Sample Loop

Temperature:

Setpoint

110.0

°C

Actual

109.8

Accuracy:

-0.2

°C

Agilent Recommended:

&gt;=

-4.0

&lt;=

4.0

Setpoint Status:

Pass

Zone:

Oven

Temperature:

Setpoint	100.0	°C
Actual	99.9	

Accuracy:

-0.1 °C

Agilent Recommended:

>=	-4.0
<=	4.0

## Overall Headspace Heated Zones Temperature Accuracy Test

Pass

## GC Oven Temperature Accuracy

Name:

7890

Setpoint Status:

Pass

Zone:

Oven

Setpoint/Actual

Temperature:	230.0	229.8	°C
--------------	-------	-------	----

Accuracy:

-0.2 °C

Agilent Recommended:

>=	-1.0	% setpoint in K	( -5.0 °C )
<=	1.0	% setpoint in K	( 5.0 °C )

Setpoint Status:

Pass

Zone:

Oven

Setpoint/Actual

Temperature:	100.0	99.8	°C
--------------	-------	------	----

Accuracy:

-0.2 °C

Agilent Recommended:

>=	-1.0	% setpoint in K	( -3.7 °C )
<=	1.0	% setpoint in K	( 3.7 °C )

## Overall GC Oven Temperature Accuracy Test Status

Pass

## GC Oven Temperature Stability

Name:

7890

Date:

November 25, 2021 5:20:10 PM

System ID:

GM-6

Setpoint Status:

Pass

Setpoint/Average

Temperature:

100.0 99.8 °C

Stability:

0.2 °C

Agilent Recommended:

&lt;= 0.5

## Overall GC Oven Temperature Stability Test Status

Pass

## Log Amp

Tested Combination1

Front

SSL

/ External

SQ

Name:

5975C inert XL with TAD

Setpoint Status:

Pass

## Overall Log Amp Test Status

Pass

## RFPA

Tested Combination1

Front

SSL

/ External

SQ

Name:

5975C inert XL with TAD

Setpoint Status:

Pass

Amu: 1050 m/z

Drift After Five Minutes:

RFPA Voltage:

18 mV

519 mV

Agilent Recommended:

&gt;= -100 and &lt;= 100

&lt;= 1100

## Overall RFPA Test Status

Pass

## Tune EI

Tested Combination1

Front

SSL

/ External

SQ

Name:

5975C inert XL with TAD

Setpoint Status:

Pass

Filament:

1

Date: November 25, 2021 5:20:10 PM

System ID: GM-6



Setpoint Status: Pass

Filament: 2

This test's 0 comment(s) and 1 deviation(s) are available in the Attachments section.

## Overall Tune EI Test Status

Pass

## Scouting Run

Tested Combination1 Front SSL / External SQ

Headspace

Name: 7697A with Tray

Source: EI - Inert

Setpoint Status: Completed

Injection Volume on Column: 1000 µL

## Overall Scouting Run Status

Completed

## Injection Precision

Tested Combination1 Front SSL / External SQ

Name: 7697A with Tray

Source: EI - Inert

Setpoint Status: Pass

Injection Volume on Column: 1000 µL

Area RSD: 1.61 %

Retention Time RSD: 0.01 %

Agilent Recommended: ≤ 5.00

≤ 1.00

## Overall Injection Precision Test Status

Pass

## Mass Ratio Precision

Date: November 25, 2021 5:20:10 PM

System ID: GM-6

Tested Combination1	Front	SSL	/ External	SQ
Headspace				
Name:	7697A with Tray			
Source:	EI - Inert			
Setpoint Status:	Pass			
Injection Volume on Column:	1000	uL		
	Area Mass 1		Mass Ratio	
	Abundance*s			
RSD:	1.61	%	0.25	%
Agilent Recommended:	<= 5.00		<= 5.00	
	Pass		Pass	

## Overall Mass Ratio Precision Test Status

Pass

## Injection Carry Over

Tested Combination1	Front	SSL	/ External	SQ
Name:	7697A with Tray			
Source:	EI - Inert			
Setpoint Status:	Pass			
Injection Volume on Column:	1000	uL		
Area Carry Over:	0.00	%		
Agilent Recommended:	<= 1.00			

This test's 0 comment(s) and 2 deviation(s) are available in the Attachments section.

## Overall Injection Carry Over Test Status

Pass

**Agilent CrossLab Compliance**

Qualification Type: ICPMS-OQ

System ID: JP12091612

EQP Name: AgilentRecommended

EQP Revision: ICPMS.02.50

EQP Publish Date: March 2020

Date: June 14, 2022 10:32:16 AM

Report Type: Report

Org. Name: ALS Laboratory Group (Thailand) Co.,Ltd.

Org. Location: 104 Phatthanakarn 40, Suan Luang, Bangkok 10250  
Thailand.REVIEW BY Tattaporn C.APPROVED BY Santana N.NEXT CAL. DATE 14/12/23



## Table of Contents

Section	Page
Cover	1
Table of Contents	2
Test Summary	3
Service Details	4
Instrument Details	5
Calculation Formulas	6
Protocol Details	7
Tests	8
Autosampler Check : ASX-520	8
Integrated Sample Introduction System (ISIS) Check : ISIS2	9
Autotune : G3281A	10
Background (No Gas Mode) : G3281A	12
Background (Gas Modes) : G3281A	13
20-Minute Stability (No Gas Mode) : G3281A	14
Declaration of Change Control	15
Attachments	16
Electronic Signature	28
Transaction Logs	29

## Test Summary

### Purpose

This section includes a status for each scheduled test and the overall qualification. For each test that is run, (1) the status is automatically determined based on pre-defined limits, and (2) the total number of times the test was run is displayed. For detailed results and specifications for a test, refer to the test results in this EQR.

### Details

Test	Status	Runs
Autosampler Check : ASX-520	Pass	1
Integrated Sample Introduction System (ISIS) Check : ISIS2	Pass	1
Autotune : G3281A	Pass	1
Background (No Gas Mode) : G3281A	Pass	1
Background (Gas Modes) : G3281A	Pass	1
20-Minute Stability (No Gas Mode) : G3281A	Pass	1

### Overall Qualification Status

Pass

## Service Details

### Purpose

This section includes local contact and delivery details for this service.

### General Details

Service Order No./Request: 6005218484  
EQP Name: AgilentRecommended  
EQP Revision: ICPMS.02.50  
Report Type: Report

### Organization Details

Name: ALS Laboratory Group (Thailand) Co.,Ltd.  
Location: 104 Phatthanakarn 40, Suan Luang, Bangkok 10250 Thailand.

### Local Contact Details

Name: Khan Chatchanai  
Job Title: Lab Manager  
Qualification Location: Spectro Room

### Operator Details

Name: Panthep Kurasathain  
Job Title: Field Service Engineer

### Data Acquisition Details

Acquisition Software Name: MassHunter  
Acquisition Software Revision: D.01.01

Customer Data System (CDS): IcpMs: MassHunter



## Instrument Details

### Purpose

This section describes the as found system configuration.

### Details

#### ICP-MS 1

Manufacturer	Agilent Technologies
Name	7700x
Model Number	G3281A
Detector Type	SQ
Nebulizer	Mira Mist (G3161)
Spray Chamber	Quartz
Torch	Quartz
Sampling Cone	Ni
Skimmer Cone	Ni
Serial Number	JP12091612
Firmware Revision	D.01.01

#### ISIS 1

Manufacturer	Agilent Technologies
Name	ISIS2
Model Number	G4911A
Installed Options	#003: 2 pumps, 1 valve, auto dilution and discrete sampling
Type	Peristaltic pump system

#### Autosampler 1

Manufacturer	Agilent Technologies
Name	ASX-520
Model Number	G3286A
Serial Number	031403A520

#### Chiller 1

Manufacturer	Agilent Technologies
Name	Chiller
Model Number	G3292A
Serial Number	4N1220700

## Calculation Formulas

### Purpose

This section includes calculation formulas for all available tests. Depending upon which tests are scheduled, all or some apply to your qualification.

For a description of calculations for ICP-MS tests performed by the MassHunter software, refer to the MassHunter application and documentation.

## Protocol Details

### Purpose

This section lists the revisions for all test units used in this report. For complete test-specific and high-level change details, refer to the Revision History document.

Test Revision	Test
ICPMS.02.50	20-Minute Stability (No Gas Mode)
ICPMS.02.50	Autosampler Check
ICPMS.02.50	Autotune
ICPMS.02.50	Background (Gas Modes)
ICPMS.02.50	Background (No Gas Mode)
ICPMS.02.50	Integrated Sample Introduction System (ISIS) Check



# Autosampler Check

## Purpose

This test demonstrates that the autosampler module is correctly installed and connected. It does not test module performance.

---

## Setpoint

### Results

Criteria	Observed Result	Expected Result	Status
After the self test, is probe in the home position?	Yes	Yes	Pass
As commanded, is the probe positioned at vial 2?	Yes	Yes	Pass

Setpoint Status: Pass Runs: 1

### Overall Autosampler Check Test Status

Pass

---

# Integrated Sample Introduction System (ISIS) Check

Purpose

This test demonstrates that the ISIS module is correctly installed and connected. It does not test module performance.

Setpoint

Results

Criteria	Observed Result	Expected Result	Status
As commanded, does the pump rotate?	Yes	Yes	Pass
As commanded, do the valves load and inject?	Yes	Yes	Pass
Setpoint Status:	Pass		Runs: 1

Overall Integrated Sample Introduction System (ISIS) Check Test Status

Pass

## Autotune

### Purpose

This test uses traceable checkout standards to run a software-executed autotune in all modes. The tune report provides values for peak width, mass axis, sensitivity, oxide species, and doubly-charged species tests.

---

### Setpoint

### Results

Peakwidth Mass 7

Agilent Recommended:

Status:

	0.735	AMU
>=	0.65	
<=	0.80	
Pass		

Peakwidth Mass 89

Agilent Recommended:

Status:

	0.732	AMU
>=	0.65	
<=	0.80	
Pass		

Peakwidth Mass 205

Agilent Recommended:

Status:

	0.746	AMU
>=	0.65	
<=	0.80	
Pass		

Mass Axis 7

Agilent Recommended:

Status:

	7.00	AMU
>=	6.9	
<=	7.1	
Pass		

Mass Axis 89

Agilent Recommended:

Status:

	89.00	AMU
>=	88.9	
<=	89.1	
Pass		

Mass Axis 205

Agilent Recommended:

Status:

	205.00	AMU
>=	204.9	
<=	205.1	
Pass		

Mass 7 Sensitivity No Gas

81.18

Mcps/ppm

Agilent Recommended:

&gt;= 25.5

Status:

Pass

Mass 89 Sensitivity No Gas

247.81

Mcps/ppm

Agilent Recommended:

&gt;= 85

Status:

Pass

Mass 205 Sensitivity No Gas

184.87

Mcps/ppm

Agilent Recommended:

&gt;= 51

Status:

Pass

Mass 59 Sensitivity He

84.86

Mcps/ppm

Agilent Recommended:

&gt;= 20.4

Status:

Pass

Oxide Ratio 156/140

1.119

%

Agilent Recommended:

&lt;= 1.38

Status:

Pass

Doubly Charged Species Ratio 70/140

1.140

%

Agilent Recommended:

&lt;= 2.3

Status:

Pass

Setpoint Status:

Pass

Runs: 1

Overall Autotune Test Status

Pass



## Background (No Gas Mode)

### Purpose

This test examines the background of the ICP-MS in no gas mode by monitoring ions during a blank run.

### Setpoint

#### Conditions

Masses:	7	AMU
	89	AMU
	205	AMU

#### Measurements and Results

Masses (AMU):	7	89	205	
Measured Value:	4.900	7.100	18.400	cps
Agilent Recommended:	<= 10	<= 10	<= 30	
Status:	Pass	Pass	Pass	

Setpoint Status: **Pass**

Runs: 1

#### Overall Background (No Gas Mode) Test Status

Pass

## Background (Gas Mode)

### Purpose

This test examines the background of the ICP-MS in the various gas modes by monitoring ions during a blank run.

Setpoint Gas Mode: Helium

### Conditions

Mass: 78 AMU  
Integration Time: 1.0 sec  
Cycles: 20

### Measurements and Results

Mass (AMU): 78  
Measured Value: 21.1000 cps  
Agilent Recommended: ≤ 460  
Status: Pass

Setpoint Status: Pass

Runs: 1

### Overall Background (Gas Mode) Test Status

Pass

## 20-Minute Stability (No Gas Mode)

### Purpose

This test monitors the abundance of ions present in the checkout standard over a 20-minute period to verify that the signal is stable. The %RSD of the abundance of given ions is calculated internally by the software and compared to the limit.

### Setpoint

#### Conditions

Mode:	Spectrum	
Masses:	7, 9, 59, 89, 140, 205	
Integration Time:	9.99	sec
Peak Pattern:	3	points/peak
Repetitions:	20	
Sweeps/Replicates:	100	

#### Measurements and Results

Masses (AMU):	7	89	205	
Stability RSD:	0.2	0.6	0.6	%
Agilent Recommended:	<= 3.45	<= 3.45	<= 3.45	
Status:	Pass	Pass	Pass	

Setpoint Status: Pass Runs: 1

#### Overall 20-Minute Stability (No Gas Mode) Test Status

Pass

## Declaration of Change Control

This document is under change control. Revision history is maintained and printed on each document. Access to the master documents is limited to process owners. Documents receive periodic review and cannot be assigned an evergreen status. The qualification performed according to this document refers only to the hardware/software configuration in place at the time of the qualification. Agilent Technologies recommends that instrument configuration change management procedures be in place in order to maintain the validation process. Any changes to the analytical or computer hardware or software must be clearly specified. A change management system provides a means for determining the degree of requalification required according to the extent of the changes made. All details of the changes must be thoroughly recorded and documented, together with details of completed tests and their results. Note: Hardware/software configuration management is the customer's responsibility.



## Attachments

Training requirements note: The delivery engineer attaches an ACE technique-specific training certificate to the Equipment Qualification Report (EQR). Obtaining ACE technique-specific certification includes pre-requisite trainings for Data Integrity, General Compliance topics (GMP, GLP, ALCOA, etc.), instrument hardware and software components, and the ACE technique itself. The one certificate encompasses all pre-requisite trainings as documented in the Agilent Learning Management System called Success Factors.

Location	Category	Document Name	Page
EQR	General	Certificate of System Qualification	17
EQR	General	Operator's training certificate and qualifications	18
EQR	General	Certificate of Qualification for ACE	19
EQR	General	Certificate of Qualification for ACE	20
EQR	General	Tune reports	21
EQR	General	Test Report	24
EQR	General	Test Report	26

## General

Document Name: Certificate of System Qualification



## Agilent Compliance Engine Self Qualification

Date: September 14, 2021 4:59:15 PM

Drive Serial #: ACA025C9

Platform Revision:

ACE 3.11

Individual self-qualification reports for each specific technique installed are also available upon request. They provide additional details on the general report from the concise summary and are structured by the actual algorithms challenged during the process. There is not a one-to-one relationship between algorithms and OQ program tests because some algorithms are used by several tests and across multiple similar hardware components of the qualified systems.

Technique Type	Tests Completed	Result
Atomic Absorption	7	Conforms
Capillary Electrophoresis	10	Conforms
Dissolution	6	Conforms
Emission Spectroscopy	3	Conforms
Gas Chromatography - GCMS	17	Conforms
Gas Chromatography	29	Conforms
Gel Permeation Chromatography	9	Conforms
ICP-MS	6	Conforms
Infrared Spectroscopy	7	Conforms
Liquid Chromatography	17	Conforms
Liquid Chromatography - LCMS	8	Conforms
Microfluidics	18	Conforms
Sample Preparation - Gas Chromatography	9	Conforms
Sample Preparation - Liquid Chromatography	8	Conforms
Supercritical Fluid Chromatography	15	Conforms
Software	6	Conforms
UV-Vis Spectrophotometer	13	Conforms

## Overall Qualification Status

Conforms

## General

Document Name: Operator's training certificate and qualifications



## Certificate of Completion

Learner Name:	Panthep Kurasathain
Title Of Course:	AN-CE-ICPMS-2-017-B:7700x/7700s ICP-MS Intro. -Oper.H/W.S/W & OQ/PV
Completion Date:	November 22, 2012
Certified By Company:	Learning at Agilent

All Service and Support training certificates have the following specific limitations.

A certificate for Service and Support training is only valid while employed by Agilent Technologies or while working as an Agilent-authorized service provider, through which the service employee has ongoing access to Agilent's: Safety Alerts, Service Notes, internal technical updates, update training, current documentation, technical support, current parts, and parts updates. Completion of training alone, without being employed by Agilent Technologies, does not qualify an individual to safely install, service or maintain Agilent products.

General

Document Name:

Certificate of Qualification for ACE



## Certificate of Completion

Learner Name:	Panthep Kurasathain
Title Of Course:	AN-CE-SS-II-030-A: ACE 3.X User Update Training
Completion Date:	July 7, 2020
Certified By Company:	Learning at Agilent

All Service and Support training certificates have the following specific limitations.

A certificate for Service and Support training is only valid while employed by Agilent Technologies or while working as an Agilent-authorized service provider, through which the service employee has ongoing access to Agilent's: Safety Alerts, Service Notes, internal technical updates, update training, current documentation, technical support, current parts, and parts updates. Completion of training alone, without being employed by Agilent Technologies, does not qualify an individual to safely install, service or maintain Agilent products.



General

Document Name: Certificate of Qualification for ACE



## Certificate of Completion

Learner Name: Panthep Kurasuthain

Title Of Course: AN-CE-ICPMS-2-035-B: CrossLab Compliance Hardware Specific Delivery for Agilent ICP-MS Systems

Completion Date: October 31, 2020

Certified By Company: Learning at Agilent

All Service and Support training certificates have the following specific limitations.

A certificate for Service and Support training is only valid while employed by Agilent Technologies or while working as an Agilent-authorized service provider, through which the service employee has ongoing access to Agilent's: Safety Alerts, Service Notes, internal technical updates, update training, current documentation, technical support, current parts, and parts updates. Completion of training alone, without being employed by Agilent Technologies, does not qualify an individual to safely install, service or maintain Agilent products.

General


Document Name: Tune reports

## Tune Report

Operator Name Supakwan Mak  
 Acq/Data Batch C:\Agilent\ICPMH11\User\tune.b  
 Acq. Date-Time 6/14/2022 9:03:15 AM  
 Report Comment PMOQ 14 June 2022  
 Instrument Name G3281A.JP12091612

[No Gas]

Sensitivity




Mass	Range	Count	RSD%	Background
7	10000	6118	3.685	4.900
89	50000	24781	3.128	7.100
205	60000	18487	3.808	18.400

Sampling Period (sec) 0.311  
 Integration Time (sec) 0.1

Odder/Doubly Charged Ratio

Oxide 156 / 140 1.119 %  
 Doubly Charged 70 / 140 1.140 %

Resolution/Axis



Mass	Peak Height	Axis	W-50%	W-10%
7	8050.47	7.00	0.65	0.735
89	24725.81	89.00	0.57	0.732
205	18589.81	205.00	0.49	0.746

Integration Time (sec) 0.1  
 Acquisition Time (sec) 22.74  
 Y Axis Linear

## Tune Parameters

## Plasma Parameters

Plasma Mode	---	Nebulizer Gas	1.05 L/min	Makeup Gas	0.00 L/min
RF Power	1550 W	Option Gas	---	Auxiliary Gas	0.90 L/min
RF Matching	1.80 V	Nebulizer Pump	0.10 rps	Plasma Gas	15.0 L/min
Sample Depth	8.0 mm	S/C Temp	2 °C		

## Lens Parameters

Extract 1	0.0 V	Omega Lens	8.6 V	Deflect	12.0 V
Extract 2	-180.0 V	Cell Entrance	-30 V	Plate Bias	-40 V
Omega Bias	-90 V	Cell Exit	-50 V		

## Cell Parameters

Use Gas	No	3rd Gas Flow	---	Energy Discrimination	5.0 V
He Flow	0.0 mL/min	OctP Bias	-8.0 V		

1 of 3

6/14/2022 9:03 AM

Date: June 14, 2022 10:32:16 AM  
 System ID: JP12091612

Document Name: Tune reports

## Tune Report

H2 Flow — OctP RF 200 V

**QP Parameters**

Mass Gain	152	Axis Gain	1.0032	QP Bias	-3.0 V
Mass Offset	123	Axis Offset	0.12		

**Hardware Settings**

**Torch**


Torch H	-0.4 mm	Torch V	0.0 mm
---------	---------	---------	--------

**EM**

Discriminator	4.5 mV	Analog HV	1708 V	Pulse HV	1956 V
---------------	--------	-----------	--------	----------	--------

[He]

## Sensitivity



Mass	Range	Count	RSD%	Background
58	10000	8486	3.392	15.700
89	20000	10724	3.135	13.900
205	50000	20200	3.166	22.100

Sampling Period [sec] 0.31  
Integration Time [sec] 0.1

## Oxide/Doubly Charged Ratio

Oxide 158 / 140 1.122 %  
Doubly Charged 70 / 140 1.288 %

## Tune Parameters

## Plasma Parameters

Plasma Mode	—	Nebulizer Gas	1.05 L/min	Makeup Gas	0.00 L/min
RF Power	1550 W	Option Gas	—	Auxiliary Gas	0.90 L/min
RF Matching	1.80 V	Nebulizer Pump	0.10 rps	Plasma Gas	15.0 L/min
Sample Depth	8.0 mm	S/C Temp	2 °C		

## Lens Parameters

Extract 1	0.0 V	Omega Lens	7.4 V	Deflect	3.6 V
Extract 2	-200.0 V	Cell Entrance	-90 V	Plate Bias	+115 V
Omega Bias	-80 V	Cell Exit	-70 V		

## Cell Parameters

Use Gas	Yes	3rd Gas Flow	—	Energy Discrimination	3.0 V
He Flow	4.5 mL/min	OctP Bias	-21.0 V		
H2 Flow	—	OctP RF	190 V		

## QP Parameters

Mass Gain	152	Axis Gain	1.0032	QP Bias	-18.0 V
Mass Offset	123	Axis Offset	0.12		

## Hardware Settings

**Torch**

Torch H	-0.4 mm	Torch V	0.0 mm
---------	---------	---------	--------

2 of 3

6/14/2022 9:03 AM

Document Name: Tune reports

## Tune Report

### EM

Discriminator

4.5 mV

Analog HV

1708 V

Pulse HV

1356 V

3 of 3

6/14/2022 9:03 AM

Date: June 14, 2022 10:32:16 AM  
System ID: JP12091612



## General

Document Name: Test Report

## Batch Summary Report

Batch Folder: D:\Agilent Service\PMOQ 13-6-22\BG He.b\A  
Analysis File: BG He.batch.bin  
Tune Step: #1 He

	Rkt	Acq. Date-Time	Data File	Sample Name	Type	Level	Dilution
1		6/14/2022 10:03:39 AM	001SMPL.d	BG He	Sample		1.0000

Document Name: Test Report

## Batch Summary Report

Analyte Table

		78 (He1)	
	Sample Name	CPS	CPS RSD
1	BG He	21.1000	38.0

Page 2 / 2

6/14/2022 10:09:04 AM

Date: June 14, 2022 10:32:16 AM  
System ID: JP12091612

## General

Document Name: Test Report

## Batch Summary Report

Batch Folder: D:\Agilent Service\PMOQ 13-6-22\ OQ 20 Min.b\

Analysis File: OQ 20 Min batch.bin

Tune Step: #1 No Gas

	Ret	Acq. Date-Time	Data File	Sample Name	Type	Level	Dilution
1		6/14/2022 9:29:27 AM	001SMPL.d	20 min	Sample		1.0000

Document Name:

Test Report

## Batch Summary Report

Analyte Table

		7 [No Gas]		9 [No Gas]		59 [No Gas]		89 [No Gas]	
	Sample Name	CPS	CPS RSD	CPS	CPS RSD	CPS	CPS RSD	CPS	CPS RSD
1	20 min	82477.8975	0.2	285.3875	5.8	162011.0035	0.7	241211.0790	0.6

		140 [No Gas]		205 [No Gas]	
	Sample Name	CPS	CPS RSD	CPS	CPS RSD
1	20 min	252452.6990	0.7	181154.8205	0.6



## Electronic Signature

### Purpose

This signature page was created and published because the ACE sign-off action was executed, which is valid for the entire document, including attachments. The ACE sign-off is an electronic signature that requires two distinct identification components: unique username and personal password. The Agilent representative who has delivered this service understands the meaning and legal status of an electronic signature. As a trained official operator, the Agilent representative has a unique password and logon to access ACE and electronically sign this document. (Other e-signatures can be applied to this document using a Document Content Management or other suitable method defined in your data access and control procedures.)

### Details

Full Name of Signer:	Panthep Kurasathain
Logged On User Name:	panthep_kurasathain@agilent.com
Signature Creation Date:	June 14, 2022
Reason for Signature:	Executed protocol and published this original version of document

### Regulatory Disclaimer

This document provides a protocol to verify and record instrument configuration and evidence of proper operation. It has been prepared from our interpretation of applicable regulations as well as industry best practices. The document is designed to provide an important component of a complete compliance package. Validation depends upon many factors and use of this protocol alone does not assure compliance. Agilent Technologies makes no promises or representations as to its sufficiency for any specific regulatory program.

### Warranty

Agilent Technologies makes no warranty of any kind to this material, including but not limited to, the implied warranties or merchantability and fitness for a particular purpose. Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

User Name: panthep\_kurasathain  
 Hostname: ASBKKWX313

System Id: JP12091612  
 Print Date: June 14, 2022 10:32:20 AM

ALS OQHW 7700 14Jun2022 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 14, 2022 10:14:43 AM	Audit	SessionCreated	Session	None
June 14, 2022 10:14:43 AM	Start	Configuration	Session	None
June 14, 2022 10:14:43 AM	Audit	Entitlement	Licensing	User is FieldEngineer and does not require an unlock code
June 14, 2022 10:19:18 AM	Audit	EqpLoaded	Session	EQP details for primary technique [lcpMs] - File path: [Protocol]Packs/lcpMs/Configurations/02.50/lcpMs.02.50.eqp], EQP File Name: [lcpMs.02.50.eqp], EQP Name: [AgilentRecommended]
June 14, 2022 10:19:20 AM	End	Configuration	Session	None
June 14, 2022 10:19:24 AM	Start	Qualification	Session	OQ
June 14, 2022 10:19:24 AM	Start	Execution	Autosampler Check : ASX-520: Autosampler Check	None
June 14, 2022 10:19:42 AM	End	Execution	Autosampler Check : ASX-520: Autosampler Check	Run Count : 1
June 14, 2022 10:19:43 AM	Start	Execution	Integrated Sample Introduction System (ISIS) Check : ISIS2: Integrated Sample Introduction System (ISIS) Check	None
June 14, 2022 10:19:47 AM	End	Execution	Integrated Sample Introduction System (ISIS) Check : ISIS2: Integrated Sample Introduction System (ISIS) Check	Run Count : 1
June 14, 2022 10:19:50 AM	Start	Execution	Autotune : G3281A: Autotune 1	None
June 14, 2022 10:22:22 AM	End	Execution	Autotune : G3281A: Autotune 1	Run Count : 1

User Name: panthep\_kurasathain  
 Hostname: ASBKKWX313

System Id: JP12091612  
 Print Date: June 14, 2022 10:32:20 AM

## ALS OQHW 7700 14Jun2022 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 14, 2022 10:22:24 AM	Start	Execution	Background (No Gas Mode) : G3281A: No Gas Mode Background 1	None
June 14, 2022 10:22:48 AM	End	Execution	Background (No Gas Mode) : G3281A: No Gas Mode Background 1	Run Count : 1
June 14, 2022 10:22:49 AM	Start	Execution	Background (Gas Modes) : G3281A: Gas Mode Background :Helium	None
June 14, 2022 10:23:35 AM	End	Execution	Background (Gas Modes) : G3281A: Gas Mode Background :Helium	Run Count : 1
June 14, 2022 10:23:37 AM	Start	Execution	20-Minute Stability (No Gas Mode) : G3281A: 20-Minute Stability (No Gas Mode) 1	None
June 14, 2022 10:24:06 AM	End	Execution	20-Minute Stability (No Gas Mode) : G3281A: 20-Minute Stability (No Gas Mode) 1	Run Count : 1
June 14, 2022 10:24:08 AM	End	Qualification	Session	OQ
June 14, 2022 10:24:08 AM	Start	Reporting	Session	None
June 14, 2022 10:30:26 AM	Audit	Reporting	Session	Report Generated : Certificate
June 14, 2022 10:30:39 AM	Audit	Reporting	Session	Report Generated : Report

# Certificate of System Qualification

ICPMS-OQ

System ID: JP12091612  
Organization Name: ALS Laboratory Group (Thailand) Co.,Ltd.  
Organization Location: 104 Phatthanakarn 40, Suan Luang, Bangkok 10250 Thailand.

Date: June 14, 2022 10:32:51 AM  
EQP Name: AgilentRecommended  
EQP Revision: ICPMS.02.50  
Overall Qualification Status: Pass

## Autosampler Check

### Overall Autosampler Check Test Status

Pass

## Integrated Sample Introduction System (ISIS) Check

### Overall Integrated Sample Introduction System (ISIS) Check Test Status

Pass

## Autotune

Peakwidth Mass 7	Pass
Peakwidth Mass 89	Pass
Peakwidth Mass 205	Pass
Mass Axis 7	Pass
Mass Axis 89	Pass
Mass Axis 205	Pass
Mass 7 Sensitivity No Gas	Pass
Mass 89 Sensitivity No Gas	Pass
Mass 205 Sensitivity No Gas	Pass
Mass 59 Sensitivity He	Pass
Oxide Ratio 156/140	Pass
Doubly Charged Species Ratio 70/140	Pass

### Overall Autotune Test Status

Pass

Date: June 14, 2022 10:32:51 AM  
System ID: JP12091612



## Background (No Gas Mode)

Setpoint Status: Pass

Masses (AMU):	7	89	205	
Measured Value:	4.900	7.100	18.400	cps
Agilent Recommended:	<= 10	<= 10	<= 30	
Status:	Pass	Pass	Pass	

## Overall Background (No Gas Mode) Test Status

Pass

## Background (Gas Mode)

Gas Mode: Helium

Setpoint Status: Pass

Mass (AMU):	78	
Measured Value:	21.1000	cps
Agilent Recommended:	<= 460	
Status:	Pass	

## Overall Background (Gas Mode) Test Status

Pass

## 20-Minute Stability (No Gas Mode)

Masses (AMU):	7	89	205	
Stability RSD:	0.2	0.6	0.6	%
Agilent Recommended:	<= 3.45	<= 3.45	<= 3.45	
Status:	Pass	Pass	Pass	

## Overall 20-Minute Stability (No Gas Mode) Test Status

Pass

## Instrument Details

### Purpose

This section describes the as found system configuration.

### Details

#### ICP-MS 1

Manufacturer	Agilent Technologies
Name	7700x
Model Number	G3281A
Detector Type	SQ
Nebulizer	Mira Mist (G3161)
Spray Chamber	Quartz
Torch	Quartz
Sampling Cone	Ni
Skimmer Cone	Ni
Serial Number	JP12091612
Firmware Revision	D.01.01

#### ISIS 1

Manufacturer	Agilent Technologies
Name	ISIS2
Model Number	G4911A
Installed Options	#003: 2 pumps, 1 valve, auto dilution and discrete sampling
Type	Peristaltic pump system

#### Autosampler 1

Manufacturer	Agilent Technologies
Name	ASX-520
Model Number	G3286A
Serial Number	031403A520

#### Chiller 1

Manufacturer	Agilent Technologies
Name	Chiller
Model Number	G3292A
Serial Number	4N1220700

## Electronic Signature

### Purpose

This signature page was created and published because the ACE sign-off action was executed, which is valid for the entire document, including attachments. The ACE sign-off is an electronic signature that requires two distinct identification components: unique username and personal password. The Agilent representative who has delivered this service understands the meaning and legal status of an electronic signature. As a trained official operator, the Agilent representative has a unique password and logon to access ACE and electronically sign this document. (Other e-signatures can be applied to this document using a Document Content Management or other suitable method defined in your data access and control procedures.)

### Details

Full Name of Signer:	Panthep Kurasathain
Logged On User Name:	panthep_kurasathain@agilent.com
Signature Creation Date:	June 14, 2022
Reason for Signature:	Executed protocol and published this original version of document

### Regulatory Disclaimer

This document provides a protocol to verify and record instrument configuration and evidence of proper operation. It has been prepared from our interpretation of applicable regulations as well as industry best practices. The document is designed to provide an important component of a complete compliance package. Validation depends upon many factors and use of this protocol alone does not assure compliance. Agilent Technologies makes no promises or representations as to its sufficiency for any specific regulatory program.

### Warranty

Agilent Technologies makes no warranty of any kind to this material, including but not limited to, the implied warranties or merchantability and fitness for a particular purpose. Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

User Name: panthep\_kurasathain  
 Hostname: ASBKKWX313

System Id: JP12091612  
 Print Date: June 14, 2022 10:32:52 AM

ALS OQHW 7700 14Jun2022 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 14, 2022 10:14:43 AM	Audit	SessionCreated	Session	None
June 14, 2022 10:14:43 AM	Start	Configuration	Session	None
June 14, 2022 10:14:43 AM	Audit	Entitlement	Licensing	User is FieldEngineer and does not require an unlock code
June 14, 2022 10:19:18 AM	Audit	EqpLoaded	Session	EQP details for primary technique [lcpMs] - File path: [ProtocolPacks/lcpMs/Configurations/02.50/lcpMs.02.50.eqp], EQP File Name: [lcpMs.02.50.eqp], EQP Name: [AgilentRecommended]
June 14, 2022 10:19:20 AM	End	Configuration	Session	None
June 14, 2022 10:19:24 AM	Start	Qualification	Session	OO
June 14, 2022 10:19:24 AM	Start	Execution	Autosampler Check : ASX-520: Autosampler Check	None
June 14, 2022 10:19:42 AM	End	Execution	Autosampler Check : ASX-520: Autosampler Check	Run Count : 1
June 14, 2022 10:19:43 AM	Start	Execution	Integrated Sample Introduction System (ISIS) Check : ISIS2: Integrated Sample Introduction System (ISIS) Check	None
June 14, 2022 10:19:47 AM	End	Execution	Integrated Sample Introduction System (ISIS) Check : ISIS2: Integrated Sample Introduction System (ISIS) Check	Run Count : 1
June 14, 2022 10:19:50 AM	Start	Execution	Autotune : G3281A: Autotune 1	None
June 14, 2022 10:22:22 AM	End	Execution	Autotune : G3281A: Autotune 1	Run Count : 1



User Name: panthep\_kurasathaln  
 Hostname: ASBKKWX313

System Id: JP12091612  
 Print Date: June 14, 2022 10:32:52 AM

ALS OQHW 7700 14Jun2022 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 14, 2022 10:22:24 AM	Start	Execution	Background (No Gas Mode) : G3281A: No Gas Mode Background 1	None
June 14, 2022 10:22:48 AM	End	Execution	Background (No Gas Mode) : G3281A: No Gas Mode Background 1	Run Count : 1
June 14, 2022 10:22:49 AM	Start	Execution	Background (Gas Modes) : G3281A: Gas Mode Background :Helium	None
June 14, 2022 10:23:35 AM	End	Execution	Background (Gas Modes) : G3281A: Gas Mode Background :Helium	Run Count : 1
June 14, 2022 10:23:37 AM	Start	Execution	20-Minute Stability (No Gas Mode) : G3281A: 20-Minute Stability (No Gas Mode) 1	None
June 14, 2022 10:24:06 AM	End	Execution	20-Minute Stability (No Gas Mode) : G3281A: 20-Minute Stability (No Gas Mode) 1	Run Count : 1
June 14, 2022 10:24:08 AM	End	Qualification	Session	OQ
June 14, 2022 10:24:08 AM	Start	Reporting	Session	None
June 14, 2022 10:30:26 AM	Audit	Reporting	Session	Report Generated : Certificate
June 14, 2022 10:30:39 AM	Audit	Reporting	Session	Report Generated : Report

User Name: panthep\_kurasathain  
Hostname: ASBKKWX313

System Id: JP12091612  
Print Date: June 14, 2022 10:32:52 AM

ALS OQHW 7700 14Jun2022 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
June 14, 2022 10:32:26 AM	Audit	Reporting	Session	Report Signed : Report PDF Name: ALS OQHW 7700 14Jun2022_20220614_OQ Report_1.pdf User Name: panthep_kurasathain@agilen t.com Full Name of Signer: Panthep Kuresathain Reason for signature: Executed protocol and published this original version of document



# Metrological Center

## SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhoi, Saraburi 18110

Telephone : +66 2 586 5792-4 Fax : +66 2 586 5109

Website : www.scieco.co.th

E-Mail : calibrate@scg.co.th

Certificate No. T220730

Page 1 of 6

## Certificate of Calibration

Equipment : HEATING BLOCK

Manufacturer : Environmental Express

Model : SC 196

Serial No. : 6974CECW3285

Customer Code : BKK\_EL0054

ID No. : T5306A3

Customer : ALS Laboratory Group (Thailand) Co.,Ltd.


104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Phatthanakan,

Khet Suan Luang, Bangkok 10250

Customer Location : Acid Digestion Lab

Date of Receipt : 30 March 2022

Calibrated By : Watcharapon Sangtong (Technician )

Approved By :  / Sujjar Naknakred ( Site Calibration Manager )

Date of Issue : 12 APR 2022

REVIEW BY	Tattaporn C.
APPROVED BY	Santun N.
NEXT CAL. DATE	7/10/23

The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standard laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Metrological Center.

Certificate No. T220730

Page 2 of 6

## Calibration Report

**Equipment** : HEATING BLOCK  
**Date of Calibration** : 7 April 2022  
**Environment** : Temperature : 21.8-23.1 °C  
 Line Voltage : 221.6-226.3 V  
 Relative Humidity : 55 - 65 %RH

### Condition of this results of calibration :

1. This equipment was calibrated by insert nine standard thermocouples type T into its chamber , the other one standard thermocouples type T use for ambient temperature measurement . The calibration was done in according to WI-T20.

All data show below were final values and the initial data from customer request . The temperature scale used was based on ITS - 90 .

2. Reference Standard Instrument :

Instrument	Model	Instrument No.	Certificate No.	Due Date
TC	TYPE T	TN221-TN230	T210008	08 June 2022
TC	TYPE T	TN231-TN240	T210008	08 June 2022
DATA LOGGER	34970A	T149	T210008	08 June 2022

3. This certificate is traceable to :

National Institute of Metrology ( Thailand ) through Metrological Center ( NSC-TISI-TIS 17025 CALIBRATION 0244.)

4. Condition of calibrated item : good

Equipment Description :

Time Constant 2 Hour 25 Minute At 95 °C  
 Fresh Air Damper ☐ Open ☐ Min ☐ Medium ☐ Max  
☐ Close  
☒ Not Available

5. Adjustment :

( ) without adjustment

( X ) after adjustment

Approved By.







# Metrological Center

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhoi, Saraburi 18110

Telephone : +66 2 586 5792-4 Fax : +66 2 586 5109


Website : [www.scieco.co.th](http://www.scieco.co.th)

E-Mail : [calibrate@scg.co.th](mailto:calibrate@scg.co.th)

Certificate No. T220730

Page 3 of 6

## Calibration Report



FRONT CONTROL

Approved By. \_\_\_\_\_

## Calibration Report

### Measurement Results

Calibration Point		Average Standard Reading at each position ( ° C )					
<b>R1 Hole1-Hole6</b>		<b>TN221</b>	<b>TN222</b>	<b>TN223</b>	<b>TN224</b>	<b>TN225</b>	<b>TN226</b>
CAL POINT	Max	93.60	93.82	94.05	94.20	94.36	94.26
95	Min	93.07	93.26	93.51	93.66	93.82	93.71
	Average	93.33	93.54	93.78	93.93	94.09	93.98
<b>R2 Hole7-Hole12</b>		<b>TN227</b>	<b>TN228</b>	<b>TN229</b>	<b>TN230</b>	<b>TN231</b>	<b>TN232</b>
	Max	94.59	94.79	94.63	94.55	94.82	95.00
	Min	94.05	94.25	94.08	93.97	94.26	94.44
	Average	94.32	94.52	94.36	94.26	94.54	94.72
<b>R3 Hole13-Hole18</b>		<b>TN233</b>	<b>TN234</b>	<b>TN235</b>	<b>TN236</b>	<b>TN237</b>	<b>TN238</b>
	Max	95.03	94.54	94.78	94.84	95.06	94.73
	Min	94.46	93.98	94.20	94.28	94.49	94.18
	Average	94.74	94.26	94.49	94.56	94.78	94.45
<b>R4 Hole19-Hole24</b>		<b>TN239</b>	<b>TN240</b>	<b>TN221</b>	<b>TN222</b>	<b>TN223</b>	<b>TN224</b>
	Max	94.89	94.82	95.73	95.85	95.73	96.10
	Min	94.33	94.26	95.51	95.62	95.51	95.85
	Average	94.61	94.54	95.62	95.73	95.62	95.97
<b>R5 Hole25-Hole30</b>		<b>TN225</b>	<b>TN226</b>	<b>TN227</b>	<b>TN228</b>	<b>TN229</b>	<b>TN230</b>
	Max	96.28	96.39	96.37	96.54	96.19	96.04
	Min	96.01	96.10	96.02	96.20	95.89	95.71
	Average	96.15	96.24	96.20	96.37	96.04	95.88
<b>R6 Hole31-Hole36</b>		<b>TN231</b>	<b>TN232</b>	<b>TN233</b>	<b>TN234</b>	<b>TN235</b>	<b>TN236</b>
	Max	96.84	96.97	97.03	96.48	96.33	95.76
	Min	96.53	96.65	96.71	96.08	95.98	95.43
	Average	96.68	96.81	96.87	96.28	96.16	95.60
<b>R7 Hole37-Hole42</b>		<b>TN237</b>	<b>TN238</b>	<b>TN239</b>	<b>TN240</b>	<b>TN221</b>	<b>TN222</b>
	Max	96.46	96.15	96.19	96.06	96.95	97.09
	Min	96.13	95.84	95.85	95.72	96.64	96.78
	Average	96.30	95.99	96.02	95.89	96.80	96.93
<b>R8 Hole43-Hole48</b>		<b>TN223</b>	<b>TN224</b>	<b>TN225</b>	<b>TN226</b>	<b>TN227</b>	<b>TN228</b>
	Max	96.91	96.58	96.13	96.19	96.34	96.19
	Min	96.55	96.21	95.80	95.87	96.03	95.88
	Average	96.73	96.40	95.96	96.03	96.18	96.03

Approved By.





## Calibration Report

### Measurement Results

Calibration Point		Average Standard Reading at each position (°C)					
<b>R1 Hole1-Hole6</b>		<b>TN221</b>	<b>TN222</b>	<b>TN223</b>	<b>TN224</b>	<b>TN225</b>	<b>TN226</b>
CAL POINT	Max	104.47	104.65	104.79	105.31	105.47	105.46
105	Min	104.15	104.27	104.45	104.98	105.14	105.20
	Average	104.31	104.46	104.62	105.15	105.31	105.33
<b>R2 Hole7-Hole12</b>		<b>TN227</b>	<b>TN228</b>	<b>TN229</b>	<b>TN230</b>	<b>TN231</b>	<b>TN232</b>
	Max	105.55	105.73	105.65	105.84	105.97	106.07
	Min	105.28	105.43	105.35	105.52	105.68	105.83
	Average	105.42	105.58	105.50	105.68	105.82	105.95
<b>R3 Hole13-Hole18</b>		<b>TN233</b>	<b>TN234</b>	<b>TN235</b>	<b>TN236</b>	<b>TN237</b>	<b>TN238</b>
	Max	106.14	106.06	105.81	106.05	105.81	105.87
	Min	105.85	105.81	105.55	105.80	105.53	105.64
	Average	106.00	105.94	105.68	105.92	105.67	105.75
<b>R4 Hole19-Hole24</b>		<b>TN239</b>	<b>TN240</b>	<b>TN221</b>	<b>TN222</b>	<b>TN223</b>	<b>TN224</b>
	Max	105.86	105.60	104.44	104.51	104.28	104.78
	Min	105.61	105.37	104.27	104.35	104.12	104.61
	Average	105.74	105.48	104.35	104.43	104.20	104.69
<b>R5 Hole25-Hole30</b>		<b>TN225</b>	<b>TN226</b>	<b>TN227</b>	<b>TN228</b>	<b>TN229</b>	<b>TN230</b>
	Max	104.94	104.93	104.97	105.08	104.68	104.69
	Min	104.77	104.75	104.76	104.90	104.51	104.49
	Average	104.85	104.84	104.86	104.99	104.60	104.59
<b>R6 Hole31-Hole36</b>		<b>TN231</b>	<b>TN232</b>	<b>TN233</b>	<b>TN234</b>	<b>TN235</b>	<b>TN236</b>
	Max	105.44	105.45	105.61	104.95	104.84	104.42
	Min	105.27	105.27	105.44	104.76	104.66	104.25
	Average	105.36	105.36	105.53	104.86	104.75	104.33
<b>R7 Hole37-Hole42</b>		<b>TN237</b>	<b>TN238</b>	<b>TN239</b>	<b>TN240</b>	<b>TN221</b>	<b>TN222</b>
	Max	105.17	104.70	104.59	104.51	105.22	105.53
	Min	105.00	104.53	104.41	104.35	105.04	105.37
	Average	105.08	104.62	104.50	104.43	105.13	105.45
<b>R8 Hole43-Hole48</b>		<b>TN223</b>	<b>TN224</b>	<b>TN225</b>	<b>TN226</b>	<b>TN227</b>	<b>TN228</b>
	Max	105.61	105.45	105.10	104.77	104.87	105.02
	Min	105.44	105.28	104.92	104.60	104.70	104.85
	Average	105.53	105.37	105.01	104.69	104.79	104.93

Approved By.



Certificate No. T220730

Page 5 of 6

## Calibration Report

### Measurement Results:

HEATING BLOCK			Temperature Distribution	
Setting (°C)	Reading (°C)		Stability ( $\pm$ °C)	Uncertainty ( $\pm$ °C)
	Min , Max	Average		
100.0	100.0 , 100.4	100.1	0.29	0.83
105.0	105.0 , 105.4	105.1	0.20	0.79

\* The quoted uncertainty exclude " uniformity "

The calibration result apply only the above calibrated item.

The result of test was found accurate as shown on date and place of test only.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k$  which for a t-distribution, providing a level of confidence of approximately 95 % .

 Approved By. 





# Metrological Center

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhoi, Saraburi 18110, Thailand.

Saraburi Tel : +66 3627 3096 Fax : +66 3627 3100

Bangkok Tel : +668 9205 6851 , +669 8247 2360

Website : www.scieco.co.th

E-Mail : calibrate@scg.co.th



Certificate No. T221644

Page 1 of 4

## Certificate of Calibration

**Equipment** : Chamber ( Cold Room )

**Manufacturer** : KOLDTECH

**Model** : KM 320

**Serial No.** : TBN-1012061/05

**Customer Code** : BKK\_EN0167

**ID No.** : T2463A3

**Customer** : ALS Laboratory Group (Thailand) Co.,Ltd.

104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Phatthanakan,

Khet Suan Luang, Bangkok 10250



**Customer Location** : Environmental Laboratory

**Date of Receipt** : 27 June 2022

**Calibrated By** : Sujjar Naknakred ( Site Calibration Manager )

**Approved By** :  / Boonchai Suriyawong (Site Calibration Manager)

**Date of Issue** : 04 JUL 2022

REVIEW BY	
APPROVED BY	
NEXT CAL. DATE	30/12/23

The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standard laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Metrological Center.

Certificate No. T221644

Page 2 of 4

## Calibration Report

**Equipment** : Chamber ( Cold Room )  
**Date of Calibration** : 30 June - 1 July 2022  
**Environment** : Temperature : 18.9-23.7 °C  
Line Voltage : 222.9-226.5 V  
Relative Humidity : 55 - 65 %RH

### Condition of this results of calibration :

1. This equipment was calibrated by insert nine standard thermocouples type T into its chamber , the other one standard thermocouples type T use for ambient temperature measurement . The calibration was done in according to WI-T20 ( based on ASTM E145-94 ( Reapproved 2001) and AS2853-1986 ).  
All data show below were final values and the initial data from customer request . The temperature scale used was based on ITS - 90 .

2. Reference Standard Instrument :

Instrument	Model	Instrument No.	Certificate No.	Due Date
TC	TYPE T	TN161-TN170	T210009	30 July 2022
TC	TYPE T	TN171-TN180	T210009	30 July 2022
DATA LOGGER	34970A	T149	T210009	30 July 2022

3. This certificate is traceable to :

National Institute of Metrology ( Thailand ) through Metrological Center ( NSC-TISI-TIS 17025 CALIBRATION 0244.)

4. Condition of calibrated item : good

Equipment Description :

Time Constant 3 Hour - Minute At 3 °C  
Fresh Air Damper ☐ Open ☐ Min ☐ Medium ☐ Max  
☐ Close  
☒ Not Available

5. Adjustment :

( ) without adjustment


( X ) after adjustment

Approved By. 

Certificate No. T221644

Page 3 of 4

## Calibration Report



C = Centre , F = Centre of Face , A = Corner , E = Centre of Edge

1C	=	TN161
2A	=	TN162
3A	=	TN163
4F	=	TN164
5A	=	TN165
6A	=	TN166
7F	=	TN167
8F	=	TN168
9A	=	TN169
10A	=	TN170

11F	=	TN171
12F	=	TN172
13A	=	TN173
14A	=	TN174
15F	=	TN175
16E	=	TN176

Approved By 

Certificate No. T221644

Page 4 of 4

## Calibration Report

**Measurement Results:**

Average Standard Reading at each position (°C)										
Calibration Point	TN161	TN162	TN163	TN164	TN165	TN166	TN167	TN168	TN169	TN170
3	2.71	2.82	2.75	2.89	2.95	3.68	3.02	2.96	3.03	2.85
	TN171	TN172	TN173	TN174	TN175	TN176				
	2.97	3.02	2.89	3.04	2.97	3.33				

Chamber ( Cold Room )			Temperature Distribution				
Setting (°C)	Reading (°C)		Average (°C)	Stability (± °C)	Uniformity (°C)	Uncertainty (± °C)	Coverage
	Min , Max	Average					Factor <i>k</i>
3.0	2.9 , 4.0	3.2	2.99	1.05	1.30	1.66	2.00

\* The quoted uncertainty exclude " uniformity "

The calibration result apply only the above calibrated item.

The result of test was found accurate as shown on date and place of test only.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor *k* which for a t-distribution, providing a level of confidence of approximately 95 % .

Approved By 



REVIEW BY	Sudarot N.
APPROVED BY	Samir N.
NEXT CAL. DATE	5/06/2023

---

# Maintenance Protocol

---

Atomic Fluorescence Spectrometer  
**mercur DUO /**  
**mercur DUO plus**

---

Serial-No.: R170A0143 Customer-No.: C104-002  
Date: 6/06/2022. Carried out by: Mr. Srichai Fah-on.

Maintenance with following Operational Qualification (OQ)  
(requires a separate OQ protocol)



Company	บริษัท 1010105 จำกัด (มหาชน) บริษัท
User	สุวิทย์ นันทิวัฒน์
Department	Lab
Street	104 ซอยพหลโยธิน 40 ถนนพหลโยธิน แขวงจตุจักร
Zip Code, City	กรุงเทพมหานคร 10250
Country	ประเทศไทย
Phone	
Fax	
E-mail	

### Maintenance works basic unit

tightness visual check inside the Mercur	<input checked="" type="checkbox"/>
visual check if gold-traps are broken	<input checked="" type="checkbox"/>
visual check if spectrometer is contaminated	<input checked="" type="checkbox"/>
visual check of the fluorescence cell	<input checked="" type="checkbox"/>
visual check of the absorption cell, incl. window	<input checked="" type="checkbox"/>
reactor cleaning	<input checked="" type="checkbox"/>
check pump-hose, if necessary change it	<input checked="" type="checkbox"/>
check swivel drive (SEV)	<input checked="" type="checkbox"/>
check drying-hose, output gas-liquid-separator	<input checked="" type="checkbox"/>
test Bubble-Sensor	<input checked="" type="checkbox"/>
check gas flows	<input checked="" type="checkbox"/>
check volume flows, reagents	<input checked="" type="checkbox"/>
recording stray light values	<input checked="" type="checkbox"/>
measurement with 30 ng/l	<input checked="" type="checkbox"/>

### Maintenance works Autosampler

Serial No.: 52 1102 250

lubricate the dosing-winding (Teflon-grease-spray)	<input checked="" type="checkbox"/>
clean the dosing cylinder, if necessary exchange it	<input checked="" type="checkbox"/>
lubricate the winding system of the height drive with some drops of oil	<input checked="" type="checkbox"/>
check the toothed belt	<input checked="" type="checkbox"/>
check the position of the mechanical stopper (height: 13mm )	<input checked="" type="checkbox"/>
check the pump rate of mixing pump (<14s AS52, typ.7s/<20s AS52S, typ.10s)	<input checked="" type="checkbox"/>
check the pump rate of washing cup	<input checked="" type="checkbox"/>
check the electrical hose connections for good contact	<input checked="" type="checkbox"/>
check the connectors of the magnetic valves	<input checked="" type="checkbox"/>
check the dosing hose for buckling, if necessary exchange it	<input checked="" type="checkbox"/>

Device parameter	nominal value	actual value
visual check general tightness inside the Mercur	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
visual check Goldtraps <i>(Goldtraps 2 / NG)</i>	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
visual check spectrometer		
Fluorescence cell	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
Absorption cell, incl. window	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
lens	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
Swivel drive (SEV)	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check pump hoses	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check hoses and hose connectors	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check and clean reactor	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check drying hose output Gas-liquid-separator	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check bubble-sensor	o.k.: <input checked="" type="checkbox"/>	not o.k.: <input type="checkbox"/>
<b>Check gasflow</b>		
Argon pressure valve 4	1.2 – 1.5 bar	<i>1.5 bar</i>
Valve 1	10 NI/h or 0.166 NL/min	<i>0.166</i>
Valve 2	50 NI/h or 0.833 NL/min	<i>0.831</i>
Valve 3	5 NI/h or 0.083 NL/min	<i>0.084</i>
Valve 4	10 NI/h or 0.166 NL/min	<i>0.167</i>
<b>Check liquidflow</b>		
Acid	2.5ml/min ± 1 ml	<i>2.5 ml/min.</i>
Red.-agent	2.5ml/min ± 1 ml	<i>2.5 ml/min.</i>
Sample	10ml/min ± 2 ml	<i>10 ml/min.</i>
<b>Adventitious light - values</b>	<b>(V)</b>	<b>from file</b>
100	<i>0</i>	<i>0</i>
200	<i>0</i>	<i>0</i>
300	<i>0</i>	<i>0</i>
350	<i>0</i>	<i>1</i>
400	<i>1</i>	<i>3</i>
450	<i>4</i>	<i>7</i>
500	<i>9</i>	<i>17</i>
550	<i>19</i>	<i>36</i>
575	<i>26</i>	<i>51</i>
600	<i>36</i>	<i>91.</i>



Device parameter	nominal value	actual value
<b>Analytical parameters Fluorescence cell</b>		
Conditions.: max.conc.: 10µg/L PMT-voltage: ..... <u>369</u> .....V		
Blank-solution		Int ..... <u>0.0003</u>
without enrichment / FBR 30 ng/L	Int > 0.0015 RSD < 3 %	Int <sub>1</sub> ..... <u>0.0058</u> RSD ..... <u>1.07</u> %
Conditions.: max.conc.: 1.7µg/L PMT-voltage: ..... <u>352</u> .....V		
Blank-solution		Int ..... <u>0.0040</u>
with enrichment / FBR 30 ng/L	Int > 0.008 RSD < 3 %	Int <sub>2</sub> ..... <u>0.0244</u> RSD ..... <u>0.87</u> %
Fok.- factor ( Int <sub>2</sub> / Int <sub>1</sub> )	> 3.5	<u>4.206</u>
<b>Analytical parameters Absorption cell</b>		
Blank-solution		Ext ..... <u>0.0010</u>
without enrichment / FBR 100 ng/L	Ext. > 0.0012 RSD < 5 %	Ext ..... <u>0.0048</u> RSD ..... <u>3.92</u> %
<b>Comments</b>		

Mr. Srirachai Pak-on  
Signature Technician

Bangkok, 6/06/2022.  
Place, Date (DD/MM/YYYY)

สตีเฟน งามปรีชา  
Signature Customer

6/06/2022  
Place, Date (DD/MM/YYYY)



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert. No.: 22TM676

Page.: 1 of 3

## Certificate of Calibration

Equipment : Autoclave  
Manufacturer : TOMY  
Model : SX-700  
Serial No. : 48134190  
ID No. : BKK\_ML0041  
Submitted by : ALS Laboratory Group (Thailand) Co.,Ltd.  
104 Phatthanakan 40, Phatthanakan Rd.,  
Khwaeng Phatthanakan, Khet Suan Luang,  
Bangkok 10250 Thailand  
Location : Media Preparation Room  
Received Order : 20 May 2022  
Calibration Date : 20 May 2022  
Ambient Temperature : (  $26 \pm 10$  ) °C  
Relative Humidity : (  $50 \pm 30$  ) %

Calibrated by : Preecha Hlahib

Approved by :

Approved Signatory

- ( ) Pornthippa Tameyakul  
( ) Malee Butkruea  
(✓) Suwit Imjai

Issue Date : 24 May 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

REVIEW BY	Sithichok
APPROVED BY	
NEXT CAL. DATE	20/n/23





Equipment : Autoclave  
Condition As-Received : Used Item  
Reference : 2205-0404OC-2

Cert. No.: 22TM676

Page.: 2 of 3

**Procedure Used :-**

Calibration were conducted using in-house calibration procedure CP-OT03 according to direct measurement method with Data Acquisition which connected with Thermocouple Type T

The temperature scale used was based on ITS-90.

**Condition of this result of calibration**

1. Reference standard instrument:-

<u>Instrument</u>	<u>Model</u>	<u>Serial No.</u>	<u>Cert. No.</u>	<u>Due Date</u>
1 ) Data Acquisition	34972A	MY57013823	22LM24	26 Feb 2023

2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

4. This result of calibration covers laboratory autoclaves for the sterilization of goods and material which could be infected with organisms categorized as Hazard Group 1, 2 and 3\*\*


(\*\* = Categorization of pathogens according to hazard and categories of containment, second edition, 1990 )

It does not cover autoclaves for use with material infect with organisms in Hazard Group 4, for which complete containment and sterilization of infected condensate is considered to be essential.

This result of calibration does not apply to sterilizers or disinfectors used for medical, dental, pharmaceutical or veterinary purposes which are directly concerned with patient care, or those used for fabrics subjected to sterilization which are required to be dry at the end of cycle.

**Result of Calibration :-** ( \* ) Without Adjustment

**Function of UUC\* :** Temperature Source



	Environmental		
	( °C )	( %R.H. )	( Volt )
Beginning of Calibration	24	55	220
Finished of Calibration	26	57	221

<u>Position</u>	<u>Description</u>	<u>Ref. Std. ID No.:</u>
1 =	Center of chamber	19-17TC-11
2 =	Temperature sensor	19-17TC-12
3 =	Exhaust port	19-17TC-13

*Signature*



Equipment : Autoclave  
Condition As-Received : Used Item  
Reference : 2205-0404OC-2

Cert. No.: 22TM676

Page.: 3 of 3

**Result of Calibration :-** ( \* ) Without Adjustment

Operating parameter Set : Temperature = 108 °C  
Sterilization period = 10 minute

UUC* Setting ( °C )	UUC* Reading ( °C )	Position	Average* Standard Reading ( °C )	Stability ( ± °C )	Pressure Reading ( MPa )	Uncertainty ( ± °C )	Coverage Factor <i>k</i>
108	108	1	107.536	0.19	0.04	0.91	2
		2	107.542				
		3	107.471				

Operating parameter Set : Temperature = 115 °C  
Sterilization period = 20 minute

UUC* Setting ( °C )	UUC* Reading ( °C )	Position	Average* Standard Reading ( °C )	Stability ( ± °C )	Pressure Reading ( MPa )	Uncertainty ( ± °C )	Coverage Factor <i>k</i>
115	115	1	114.502	0.15	0.08	0.89	2
		2	114.582				
		3	114.539				

Operating parameter Set : Temperature = 118 °C  
Sterilization period = 10 minute

UUC* Setting ( °C )	UUC* Reading ( °C )	Position	Average* Standard Reading ( °C )	Stability ( ± °C )	Pressure Reading ( MPa )	Uncertainty ( ± °C )	Coverage Factor <i>k</i>
118	118	1	117.517	0.094	0.09	0.88	2
		2	117.616				
		3	117.530				

**Result of Calibration :-** ( \* ) Without Adjustment

Operating parameter Set : Temperature = 121 °C  
Sterilization period = 30 minute

UUC* Setting ( °C )	UUC* Reading ( °C )	Position	Average* Standard Reading ( °C )	Stability ( ± °C )	Pressure Reading ( MPa )	Uncertainty ( ± °C )	Coverage Factor <i>k</i>
121	121	1	120.400	0.18	1.1	0.90	2
		2	120.511				
		3	120.465				

**Average\*** : The average of 30 values in each position.

**Stability** : One-half of the greatest maximum difference of measured temperature at any one probe.

**UUC\*** : Unit Under Calibration

**Note** : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-oOo-

*Signature*

a 1109669





TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert. No.: 22TM102

Page.: 1 of 3

## Certificate of Calibration

Equipment : Incubator  
Manufacturer : SHEL-LAB  
Model : 1915A  
Serial No. : 0200599  
ID No. : BKK\_ML0010

REVIEW BY	Sithichok T.
APPROVED BY	[Signature]
NEXT CAL. DATE	22/07/23

Submitted by : ALS Laboratory Group (Thailand) Co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd.,  
Khwaeng Phatthanakan, Khet Suan Luang,  
Bangkok 10250 Thailand

Location : Incubation & Micrological Reading

Received Order : 21 January 2022

Calibration Date : 21 January 2022

Ambient Temperature :  $(26 \pm 10) ^\circ\text{C}$

Relative Humidity :  $(50 \pm 30) \%$

Calibrated by : Krisda Malee

Approved by :

Malee  
Approved Signatory

( ) Pornthippa Tameyakul

( ) Malee Butkruea

( ) Suwit Imjai

Issue Date : 3 February 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

A 0037377



Equipment : Incubator  
 Condition As-Received : Used Item  
 Reference : 2201-0616OC-1

Cert. No.: 22TM102

Page.: 2 of 3

**Procedure Used :-**

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector ( RTD ).

The temperature scale used was based on ITS-90.

**Condition of this result of calibration**

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1 ) Data Acquisition	34972A	MY57013711	21LM7	16 Jun 2022


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

**Result of Calibration :-** ( \* ) Without Adjustment

**Function of UUC\* :** Temperature Source

**Fresh air setting :** Close



Environment during calibration		
	Beginning	Finished
Temp. ( °C )	26	25
REL.Humid. ( % )	53	54
AC Supply ( Volt )	220	221

Position :	Ref. Std. ID No.:
1	18-18RTD-01
2	18-18RTD-02
3	18-18RTD-03
4	18-18RTD-04
5	18-18RTD-05
6	18-18RTD-06
7	18-18RTD-07
8	18-18RTD-08
9 (ref.)	18-18RTD-09

**Probe Installation Details :**

a = 10 cm  
 b = 10 cm  
 c = 10 cm

**Dimension of Chamber :**

D = 0.90 m  
 W = 0.75 m  
 H = 1.2 m  
 Capacity = 0.81 m<sup>3</sup>

*Malu .*





Equipment : Incubator  
Condition As-Received : Used Item  
Reference : 2201-0616OC-1  
**Result of Calibration :-** ( \* ) Without Adjustment  
Function of UUC\* : Temperature Source  
Fresh air setting : Close

Cert. No.: 22TM102

Page.: 3 of 3

Calibration Point ( °C )	UUC* Setting ( °C )	UUC* Reading ( °C )	Temperature stability ( ± °C )	Temperature uniformity ( °C )	Overall Variation ( °C )	Uncertainty ( ± °C )	Coverage Factor <i>k</i>
35.0	35.0	35.0	0.043	0.41	0.42	0.30	2

Calibration Point ( °C )	Measured Temperature ( °C )								
	Position								
	1	2	3	4	5	6	7	8	9 (ref.)
35.0	34.801	34.868	34.862	35.012	35.040	35.010	35.084	35.040	35.178

**Average\*** : The average of 30 values in each position.

**Temperature stability** : One-half of the greatest maximum difference of measured temperature at any one sensor.

**Temperature uniformity** : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

**Overall Variation** : The Difference of the maximum and minimum measured temperatures throughout observation.

**UUC\*** : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-o0o-

*Mah.*



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert. No.: 22TM1571

Page : 1 of 3

## Certificate of Calibration

Equipment : Hot Air Oven

Manufacturer : Binder

Model : ED 240/E2

Serial No. : 00-15533

ID No. : BKK\_ML0013

Submitted by : ALS Laboratory Group (Thailand) Co., Ltd.  
104 Phatthanakan 40, Phatthanakan Rd.,  
Khwaeng Phatthanakan, Khet Suan Luang,  
Bangkok 10250 Thailand

Location : Media Preparation Room

Received Order : 21 November 2022

Calibration Date : 21 November 2022

Ambient Temperature : ( 26 ± 10 ) °C

Relative Humidity : ( 50 ± 30 ) %

Calibrated by : Krisda Malee

Approved by :

*Malee*

Approved Signatory

( ) Pornthippa Tameyakul

(✓) Malee Butkruea

( ) Suwit Imjai

Issue Date : 29 November 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.





Equipment : Hot Air Oven  
 Condition As-Received : Used Item  
 Reference : 2211-0623OC-1

Cert. No.: 22TM1571

Page : 2 of 3

**Procedure Used :-**

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Thermocouple Type T.

The temperature scale used was based on ITS-90.

**Condition of this result of calibration**

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1 ) Data Acquisition	34970A	MY44067817	22LM121	22 Aug 2023

2. This certificate is valid only to the item calibrated on date and place of calibration.


3. This certification is traceable to the International System of Unit.

**Result of Calibration :-** ( \* ) After Adjustment

**Function of UUC\* :** Temperature Source

**Fresh air setting :** Not Available

Environment during calibration		
	Beginning	Finished
Temp. ( °C )	26	26
REL.Humid. ( % )	53	55
AC Supply ( Volt )	219	220



Position :	Ref. Std. ID No.:
1	21-15TC-01
2	21-15TC-02
3	21-15TC-03
4	21-15TC-04
5	21-15TC-05
6	21-15TC-06
7	21-15TC-07
8	21-15TC-08
9 (ref.)	21-15TC-09

**Probe Installation Details :**

a = 5.0 cm  
 b = 5.0 cm  
 c = 5.0 cm

**Dimension of Chamber :**

D = 0.50 m  
 W = 0.80 m  
 H = 0.60 m  
 Capacity = 0.24 m<sup>3</sup>

*Malu.*



Equipment : Hot Air Oven  
Condition As-Received : Used Item  
Reference : 2211-0623OC-1  
Result of Calibration :- ( \* ) After Adjustment  
Function of UUC\* : Temperature Source  
Fresh air setting : Not Available

Cert. No.: 22TM1571

Page : 3 of 3

Calibration Point ( °C )	UUC* Setting ( °C )	UUC* Reading ( °C )	Temperature stability ( ± °C )	Temperature uniformity ( °C )	Overall Variation ( °C )	Uncertainty ( ± °C )	Coverage Factor <i>k</i>
180	180	180	0.70	1.5	2.9	1.4	2

Calibration Point ( °C )	Measured Temperature ( °C )								
	Position								
	1	2	3	4	5	6	7	8	9 (ref.)
180	179.520	180.585	178.855	179.482	178.827	179.938	179.074	180.199	180.068

**Average\*** : The average of 30 values in each position.

**Temperature stability** : One-half of the greatest maximum difference of measured temperature at any one sensor.

**Temperature uniformity** : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

**Overall Variation** : The Difference of the maximum and minimum measured temperatures throughout observation.

**UUC\*** : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-o0o-

*Malu*





TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert. No.: 22TM677

Page.: 1 of 3

## Certificate of Calibration

Equipment : Water Bath  
Manufacturer : Memmert  
Model : WNE 45  
Serial No. : L712.0429  
ID No. : BKK\_ML0056  
Submitted by : ALS Laboratory Group (Thailand) Co.,Ltd.  
104 Phatthanakan 40, Phatthanakan Rd.,  
Khwaeng Phatthanakan, Khet Suan Luang ,  
Bangkok 10250 Thailand  
Location : Incubator & Microbiological Reading  
Received Order : 20 May 2022  
Calibration Date : 20 May 2022  
Ambient Temperature : ( 26 ± 10 ) °C  
Relative Humidity : ( 50 ± 30 ) %

Calibrated by : Preecha Hlahib

Approved by :

  
Approved Signatory

- ( ) Pornthippa Tameyakul  
( ) Malee Butkruea  
( ☒ ) Suwit Imjai

Issue Date :

24 May 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

A 0041433



Equipment : Water Bath  
Condition As-Received : Used Item  
Reference : 2205-0404OC-1

Cert. No.: 22TM677

Page.: 2 of 3

**Procedure Used :-**

Calibration were conducted using in-house calibration procedure CP-OT04 according to direct measurement method with Data Acquisition which connected with Industrial Platinum Resistance Thermometer ( IPRT ).

The temperature scale used was based on ITS-90.

**Condition of this result of calibration**

1. Reference standard instrument:-

**Instrument**

**Model**

**Serial No.**

**Cert. No.**

**Due Date**

1 ) Data Acquisition

34972A

MY57013823

22LM24

26 Feb 2023


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

**Result of Calibration :-** ( \* ) Without Adjustment

**Function of UUC\* :** Temperature Source

	Environmental		AC Voltage Supply ( Volt )
	( °C )	( %R.H. )	
Beginning of Calibration	24	47	220
Finished of Calibration	24	52	221



Front

Position :	Ref. Std. S/N.:
1	4804539-006
2	4804539-007
3	4804539-008
4	4804539-009
5(ref.)	4804539-010

*Signature*





Equipment : Water Bath  
Condition As-Received : Used Item  
Reference : 2205-0404OC-1  
Result of Calibration :- ( \* ) Without Adjustment  
Function of UUC\* : Temperature Source

Cert. No.: 22TM677

Page.: 3 of 3

Calibration point ( °C )	UUC* Setting ( °C )	UUC* Reading ( °C )	Average* Standard Reading ( °C )				
			Position				
			1	2	3	4	5 (ref.)
44.5	44.4	44.4	44.539	44.497	44.476	44.506	44.507

Calibration point ( °C )	Uniformity ( °C )	Stability ( ± °C )	Uncertainty ( ± °C )	Coverage Factor <i>k</i>
44.5	0.068	0.030	0.15	2

**Average\*** : The average of 30 values in each position.

**Uniformity** : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

**Stability** : One-half of the greatest maximum difference of measured temperature at any one probe.

**UUC\*** : Unit Under Calibration

**Note** : The reported uncertainty of measurement was included stability and excluded uniformity.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-o0o-

*Amir*

**Sartorius (Thailand) Co., Ltd.**

129 Rama 9 Road, Huaykwang, Huaykwang, Bangkok 10310

Tel: +66 2643 8361-6, e-mail: service.thailand@sartorius.com

**SARTORIUS**

# Certificate

## of Calibration

REVIEW BY	<u>Sirilut P.</u>
APPROVED BY	<u>LL AL</u>
NEXT CAL. DATE	<u>8/2/24</u>

Model Number : MSE224S-100-DUDescription : Analytical BalanceSerial Number : 26207042ID No. : BKK\_EN0002Manufacturer : SartoriusCertificate No. : 23BCI0072Issued Date : Monday, February 13, 2023Reference No. : 203245Page No. : 1 of 2Customer Name : ALS Laboratory Group (Thailand) Co., Ltd.104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Phatthanakan, Khet Suan Luang, Bangkok 10250.Calibrated Place : Balance RoomCalibrated By : Mr. Chonchai InthanaCalibration Date : Wednesday, February 08, 2023**Calibration**Procedure No. : This calibration was conducted by  
Using in-house calibration procedure number (WI-003)

Based on UKAS LAB 14 : 2019

**Metrological data :**Capacity : 220 g Readability : 0.0001 g**Ambients Conditions:**Temperature : 23.2 °C ± 5.0 °CHumidity : 60.0 % RH ± 10.0 % RHPressure :                      ±                     **Reasons for calibration**☐ New Installation ☐ Service / Repaired ☒ Re-calibration/ MaintenanceEquipment Condition: ☒ Good Operate ☐ Fair**Measurement Method UKAS Publication Ref :Lab 14**

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor ( $k=2$ ) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM). The calibration certificate documents the traceability to National Standards, which realise the unit of measurement according to the International Standard System of Units (SI). Report of Tolerance came from list of Sartorius Metrological Specifications.

**Traceability:**

Model Number	Description	Traceability	Certificate No.	Due Date
YCS011-522-00	Sartorius weight set 1mg - 5000g E2, YCS011-522-00	SPC-RT	C02212565	14-Sep-2023
MHB-382SD	Humidity/Barometer/Temp Lutron MHB-382SD	DKSH	C19220444	5-Sep-2023

This certificate relate and apply this equipment only.

This certificate may not be reproduced other than in full except with the prior written approval of the Verification Operation Division Sartorius (Thailand) Co., Ltd.

SOP FM 33 03 February 2022

Choi  
Mr. Chonchai Inthana (Technical Manager)

S  
T  
A  
M  
P

# Certificate of Calibration

Model Number : MSE224S-100-DU  
 Description : Analytical Balance  
 Serial Number : 26207042  
 ID No. : BKK\_EN0002  
 Manufacturer : Sartorius

Certificate No. : 23BCI0072  
 Issued Date : Monday, February 13, 2023  
 Reference No. : 203245  
 Page No. : 2 of 2

## Calibration Results : Without Adjustment

### Repeatability



The reproducibility is the ability of a weighing instrument to display nearly identical readouts under constant test conditions when the same load within a measurement series is placed repeatedly on the weighing pan in the same manner. The standard deviation is used to express reproducibility quantitatively.

Nominal Value : (Low Load)	20.0000	200.0000
20 g	20.0000	199.9999
Tolerance	20.0000	200.0000
0.0001 g	20.0000	199.9999
	20.0001	200.0000
Nominal Value : (High Load)	20.0000	200.0000
200 g	20.0000	199.9999
Tolerance	20.0000	199.9999
0.0001 g	20.0000	200.0000
	20.0001	199.9999
Standard Deviation	0.00004	0.00005

### Eccentricity (Off-center loading error)

The off-center loading error is yielded by the difference between the readout of the load, i.e. 1/3 or 1/4 of maximum capacity, placed in the middle of the weighing pan and between each of four additional measurement points (positions defined according to OIML R76).

Nominal value : 50 g  
 Tolerance 0.0004 g



Difference	
1	—
2	-0.0001
3	0.0000
4	0.0001
5	0.0000
6	—

### Linearity

The linearity, also called linearity error. Describes the deviation of the characteristic curve of a weighing instrument from the linear slope.

Tolerance 0.0002 g

Nominal Value (g)	Conventional Mass Value (g)	Displayed Value (g)	Deviation (g)	Uncertainty (g)
0.01	0.0100	0.0100	0.0000	0.00014
0.1	0.1000	0.1000	0.0000	0.00014
1	1.0000	1.0000	0.0000	0.00014
2	2.0000	2.0000	0.0000	0.00014
5	5.0000	5.0000	0.0000	0.00014
10	10.0000	10.0000	0.0000	0.00014
20	20.0000	20.0000	0.0000	0.00014
50	50.0000	50.0000	0.0000	0.00015
100	100.0000	100.0000	0.0000	0.00019
200	200.0000	199.9999	-0.0001	0.00030

End of Report.



# Certificate of Calibration

<b>Equipment:</b>	SPECTROPHOTOMETER	<b>Certificate No.:</b>	C06220464
<b>Model:</b>	DR6000	<b>Issued Date:</b>	27 September 2022
<b>Serial No. (or ID.):</b>	1627845 (RYG_EN0037)	<b>Job No.:</b>	KSPR2212224
<b>Manufacturer:</b>	HACH	<b>Page:</b>	1 of 3
<b>Condition:</b>	In Condition		

**Customer:** ALS Laboratory Group (Thailand) Co.,Ltd. (Rayong Branch)  
616/10 Moo 5 T.Maenam Khu,  
A.Pluakdaeng, Rayong 21140, Thailand.

REVIEW BY *N. Banat*  
APPROVED BY *D. ...*  
NEXT CAL. DATE *27/13/24*

**Environment Condition:**

Temperature	23.1	°C	±
Humidity	65.4	%RH	±

**Calibration Place:** ALS Laboratory Group (Thailand) Co.,Ltd. (Rayong Branch) ( Wet Chemistry )  
616/10 Moo 5 T.Maenam Khu,  
A.Pluakdaeng, Rayong 21140, Thailand.

**Calibration By:** Mr. Chattuphon Foithong

**Calibration Date:** 27 September 2022

**The Method used:** In house method, CAL-WI-24, base on ASTM E 275-08 and ASTM E 387-04

**Traceability:** This certificate is traceable to the CRM maintained by National Institute of Standards and Technology (NIST) through Starna Scientific Limited.

The standard for Wavelength Certificate No. 91418 and 91435  
The standard for Photometric Certificate No. 91441 and 101088  
The standard for Stray light Certificate No. 101041 and 101040  
The standard for Spectral resolution Certificate No. 101037

  
(Mr. Chattuphon Foithong)  
Person in charge

  
(Mr. Thalerngkeat POUNGNGAM)  
Authorized signatory

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor ( $k=2$ ) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).

These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of DKSH Technology Limited.

บริษัท ดีเคเอส อีเซีย เทคโนโลยี จำกัด  
DKSH Technology Limited  
2533 ถนนสุขุมวิท แขวงคลองเตย เขตคลองเตย กรุงเทพมหานคร 10260  
2533 Sukhumvit Road, Bangchak, Phrakhanong, Bangkok 10260  
Phone: +66 2639 7000 Email: info.calibration@dksh.com Website: www.dksh.com/scientific-thailand



**Calibration Results:**
**Without Adjustment**

Wavelength Accuracy (nm), The spectral bandwidth of Std at 2 nm and UUC at 2 nm

Standard Wavelength	Unit Under Calibration	Correction	Uncertainty
418.61	418.4	0.21	0.14
536.66	536.7	-0.04	0.14
637.98	638.3	-0.32	0.14
748.48	748.8	-0.32	0.14
807.03	807.4	-0.37	0.13

**Photometric Accuracy (Absorbance)**

Wavelength	Standard absorbance	Unit Under Calibration	Correction	Uncertainty
420 nm	0.0000	0.000	0.0000	0.0045
	0.5605	0.563	-0.0025	0.0045
	0.7334	0.737	-0.0036	0.0045
	1.0534	1.057	-0.0036	0.0045
440 nm	0.0000	0.000	0.0000	0.0045
	0.5503	0.553	-0.0027	0.0045
	0.7179	0.720	-0.0021	0.0045
	1.0312	1.034	-0.0028	0.0045
465 nm	0.0000	0.000	0.0000	0.0045
	0.5024	0.506	-0.0036	0.0045
	0.6693	0.672	-0.0027	0.0045
	0.9604	0.964	-0.0036	0.0045
546.1 nm	0.0000	0.000	0.0000	0.0045
	0.5168	0.519	-0.0022	0.0045
	0.6903	0.691	-0.0007	0.0045
	0.9904	0.992	-0.0016	0.0045
590 nm	0.0000	0.000	0.0000	0.0045
	0.5525	0.554	-0.0015	0.0045
	0.7175	0.718	-0.0005	0.0045
	1.0301	1.031	-0.0009	0.0045
635 nm	0.0000	0.000	0.0000	0.0045
	0.5367	0.538	-0.0013	0.0045
	0.6847	0.685	-0.0003	0.0046
	0.9823	0.983	-0.0007	0.0045

**Calibration Results:**
**Without Adjustment**
**Photometric Accuracy (Absorbance)**

Wavelength	Standard absorbance	Unit Under Calibration	Correction	Uncertainty
235 nm	0.0000	0.000	0.0000	0.0080
	0.7423	0.744	-0.0017	0.0083
257 nm	0.0000	0.000	0.0000	0.0080
	0.8609	0.861	-0.0001	0.0084
313 nm	0.0000	0.000	0.0000	0.0080
	0.2895	0.292	-0.0025	0.0080
350 nm	0.0000	0.000	0.0000	0.0080
	0.6381	0.638	0.0001	0.0080

**Stray light \***

Standard: cut-off	UUC: Wavelength (nm)	UUC: Transmission (%T)	Absorbance (A)
260.67 +/- 0.11 nm	260.7	2.1	1.678
391.94 +/- 0.11 nm	391.9	1.7	1.770

**Spectral Resolution \***

Nominal Concentration 0.02 % v/v	Peak	Trough	Ratio	SBW
Standard Wavelength ( nm )	268.60	266.63	1.39	2.00
UUC: Wavelength (nm)	268.2	266.1		
Std Absorbance ( A )	0.4810	0.3176		
Absorbance ( A )	0.373	0.268		

\* Calibration Marked " Not TISI Accredited " in this Certificate have been included for completeness.

**The End of Certificate**



## ใบตรวจสอบสภาพเครื่องวัดสิ่งแวดล้อม

เลขที่ใบงาน: KSPR2212224

ชนิดเครื่องมือ: SPECTROPHOTOMETER

รุ่น: DR6000

หมายเลขเครื่อง: 1627845

ตรวจสอบ (รับ)		รายการตรวจเช็ค	ตรวจสอบ (ส่ง)		หมายเหตุ
27 Sep 2022			27 Sep 2022		
ปกติ	ไม่ปกติ		ปกติ	ไม่ปกติ	
		General			
<input checked="" type="checkbox"/>	<input type="checkbox"/>	1. ความสมบูรณ์เครื่อง	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	2. ความสะอาด ( ช่องใส่ตัวอย่าง, ภายใน-นอกเครื่อง)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	3. สวิตซ์ ปิด – เปิด เครื่อง (On-Off Swicth)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	4. ปุ่มกด (Keypad)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	5. หน้าจอ (Display, Screen Contrast)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
		Spectrophotometer			
<input type="checkbox"/>	<input type="checkbox"/>	6. แรงดันไฟฟ้า (Battery Backup) >= 2.5 VDC	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	7. ตัวหมุนเลือกความยาวคลื่น (Wavelength Control)	<input type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	8. ความยาวคลื่น (Wavelength Check)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	656.1 ได้ 656.1 nm
<input checked="" type="checkbox"/>	<input type="checkbox"/>	9. แหล่งกำเนิดแสง (UV < 3,000 hour)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	10. แหล่งกำเนิดแสง (Visible < 5,000 hour)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	11. ช่องวัดหลายตัวอย่าง (Carousel Module)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
		pH Meter and Conductivity Meter			
<input type="checkbox"/>	<input type="checkbox"/>	12. อิเล็กโทรด ( Electrode and Connection Cable )	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	13. ระดับสารละลายใน Electrode (Level KCl )	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	14. ฝาปิดกันปลาย Electrode (Dust Protection Hood)	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	15. ขาจับอิเล็กโทรด (Stand)	<input type="checkbox"/>	<input type="checkbox"/>	
		Turbidimeter			
<input type="checkbox"/>	<input type="checkbox"/>	16. ค่าความขุ่นที่ต่ำสุด (No Sample)	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	17. ระดับการส่องสว่างของแสง (>= 2.5 ไม่นเกิน 3.0)	<input type="checkbox"/>	<input type="checkbox"/>	
		Automatic titrator			
<input type="checkbox"/>	<input type="checkbox"/>	18. สภาพ Piston Burettes	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	19. Function Rinsing and Dosing	<input type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	20. ระบบท่อสายยางและอุปกรณ์ประกอบ	<input type="checkbox"/>	<input type="checkbox"/>	

เพิ่มเติม/ข้อแนะนำ :

Mr. Chattuphon Foithong

Service Engineer



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL. 0-2717-3000-29 FAX. 0-2719-9484



Cert.No.: 23CH442  
Page.: 1 of 2

## Certificate of Calibration

Equipment :	pH Meter
Manufacturer :	Mettler Toledo
Model :	Seven2Go TM pH/mV S2
Serial No. :	C202355606
ID No. :	RYG_FS0574
Condition As-Received:	Used Item
Received Date :	31 March 2023
Calibration Date :	03 April 2023
Reference :	2303-1133DSC-3
Submitted by :	ALS Laboratory Group (Thailand) Co.,Ltd. Rayong Branch 616/10 Moo 5, T.Maenam Khu, A.Pluakdaeng, Rayong 21140, Thailand
Ambient Temperature :	(25 ± 2.5) °C
Relative Humidity :	(50 ± 15) %
Calibration Procedure :	In - house method : - CP-CH5 by direct measurement with standard voltage calibrator and direct measurement with certified reference material (CRM)



Calibrated by : Warakorn Lernagtrakul

Approved by :

*Malee*

Approved Signatory

- ( ☒ ) Malee Butkruea  
( ☐ ) Saithip Meangmai  
( ☐ ) Warakorn Lernagtrakul

Issue Date : 5 April 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

A 0052954





Cert. No.: 23CH442

Page.: 2 of 2

**Condition of this calibration result**

1. Reference Standard Instrument : -

<u>Instrument</u>	<u>Serial No.</u>	<u>ID No.</u>	<u>Cert. No.</u>	<u>Due Date</u>
1) Document Process Calibrator	54030049	130RC116	22E2769	24 Aug 2023

This certification is traceable to the International System of Unit maintained at:-

- Traceable to National Institute of Metrology (Thailand), NIMT

2. Certified Reference Materials : The measurement results are traceable to SI through CPA chem Ltd., ANSI-ASQ National Accreditation Board, Accredited No. AR-1835

<u>Buffer Solution</u>	<u>Manufacturer</u>	<u>Lot No.</u>	<u>Exp. date</u>
pH 4.008	CPA chem	863832	28 Dec 2024
pH 6.987	CPA chem	826589	09 July 2023
pH 10.010	CPA chem	863835	28 Dec 2023

3. This certificate is valid only to the item calibrated on date and place of calibration.

**Calibration Results****Function : mV Measurement**

Performing standard curve by Fluke at pH (4,7,10)

Unit Under Calibration	Nominal Value	Standard Voltage Input	Actual Reading		Uncertainty of Measurement ( ±mV )	Coverage factor k
	pH	mV	mV	pH		
pH Meter S/N.: C202355606	4.00	177.48	177	4.00	0.58	2.00
	7.00	0.00	0	7.00	0.58	2.00
	10.00	-177.48	-178	10.00	0.58	2.00

**Function : pH Measurement**

Performing three buffers standard curve by using buffer nominal pH (4,7,10)

Unit Under Calibration	Standard pH Buffer Solution	Actual pH Reading	Actual mV Reading ( mV )	Uncertainty of pH measurement ( ± )	Coverage factor k
pH Electrode S/N.: 2015870	4.008	4.01	170	0.0071	2.00
	6.987	7.00	-5	0.011	2.00
	10.010	10.01	-181	0.0095	2.00

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor  $k$ , providing a level of confidence of approximately 95 %.

-o0o-

Malu.



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL. 0-2717-3000-29 FAX. 0-2719-9484



Cert. No.: 23LM86

Page.: 1 of 2

## Certificate of Calibration

**Equipment :** pH Meter with Sensor

**Manufacturer :** Mettler Toledo

**Model :** Seven2GoTM pH/mV S2

**Serial No. :** C202355606

**ID No. :** RYG\_FS0574

**Submitted by :** ALS Laboratory Group (Thailand) Co.,Ltd.  
Rayong Branch  
616/10 Moo 5, T.Maenam Khu, A.Pluakdaeng,  
Rayong 21140, Thailand

**Location :** TPA On Site Calibration Laboratory

**Received Order :** 31 March 2023

**Calibrated Date :** 05 April 2023

**Ambient Temperature :** (  $26 \pm 10$  ) °C

**Relative Humidity :** (  $50 \pm 30$  ) %

**AC Line Voltage :** (  $220 \pm 22$  ) V

**Calibrated by :** Preecha Hlahib

**Approved by :**

Approved Signatory

- ( ) Pornthippa Tameyakul  
( / ) Malee Butkruea  
( ) Suwit Imjai

**Issue Date :**

21 April 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

A 0053338





Equipment : pH Meter with Sensor

Condition As-Received : Used Item

Reference : 2303-1133DSC-4

Cert. No.: 23LM86

Page.: 2 of 2

**Procedure Used :-**

Calibration were conducted using in-house calibration procedure CP-OT01 according to comparison with Industrial Platinum Resistance Thermometer ( IPRT ) into Temperature Bath.

The temperature scale used was based on ITS-90.

**Condition of this result of calibration**

1. Reference standard instrument:-

<u>Instrument</u>	<u>Model</u>	<u>Serial No.</u>	<u>Cert. No.</u>	<u>Due Date</u>
1) Digital Thermometer	1502A	A52847	22I1325	31 Oct 2023

2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

**Result of Calibration :-** ( \* ) Without Adjustment

**Function :** Temperature measurement.

This instrument was connected with temperature sensor, S/N.: 2015870

<u>Calibration Point</u> ( °C )	<u>Immersion Depth</u> ( mm )	<u>Standard Temperature</u> ( °C )	<u>UUC* Reading</u> ( °C )	<u>Error</u> ( °C )	<u>Uncertainty</u> ( ± °C )	<u>Coverage Factor</u> <i>k</i>
25.0	100	25.002	25.1	0.098	0.16	2.00
40.0	100	40.001	40.2	0.199	0.16	2.00
60.0	100	60.005	60.5	0.495	0.16	2.00

**UUC\* :** Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-o0o-

*Malu.*

# Certificate of Calibration

<b>Equipment:</b>	Block Digestion Unit	<b>Certificate No.:</b>	C29230010
<b>Model:</b>	KT-20s	<b>Issued Date:</b>	18 March 2023
<b>Serial No. (or ID.):</b>	5720210009/5770200073	<b>Job No.:</b>	KSPR2304362
<b>Manufacturer:</b>	Gerhardt	<b>Page:</b>	1 of 4
<b>Condition:</b>	In Condition	<b>Digestion Block:</b>	20 holes.

**Customer:** ALS Laboratory Group (Thailand) Co.,Ltd. (Rayong Branch)  
616/10 Moo 5 T.Maenam Khu, A.Pluakdaeng,  
Rayong 21140, Thailand.

**Environment Condition:**

Temperature:	25 °C	±	0.5 °C
Humidity:	65 %RH	±	3.7 %RH
Voltage:	231 VAC	±	3.1 VAC

REVIEW BY	<i>N. Benjitt</i>
APPROVED BY	<i>D. Srichana</i>
NEXT CAL. DATE	15/03/24

**Calibration Place:** ALS Laboratory Group (Thailand) Co.,Ltd. (Rayong Branch)  
( Wet Chemistry Lab )  
616/10 Moo 5 T.Maenam Khu, A.Pluakdaeng,  
Rayong 21140, Thailand.

**Calibration By:** Mr. Nakarin Ruenros

**Calibration Date:** 15 March 2023

**The Method used:** In house method, base on by comparison with standard

**Traceability:** This certificate is traceable to the SI Units maintained by National Institute of Metrology (NIMT), Thailand through N.M. Technical Center Laboratory (NTL)  
Certificate No.: TC22/0080



(Mr. Nakarin Ruenros)

Person in charge



(Mr. Udon Srichana)

Authorized signatory

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).

These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of DKSH Technology Limited.







Fig. 1.: Front view



Location of standard

Fig. 2.: Digestion block

## Definitions

**Indicating Temperature:** The average reading of indicating device which forms the integral part of the Digestion block.

**Measured Temperature:** The average reading of working standard at any positions or location.

**Calibration Results:**
**Before adjustment**

Locations	Desired (°C)	Setting (°C)	Indicating (°C)	Measured Temperature (°C)	Correction of UUC. (°C)	Uncertainty (± °C)
A1	380	380	380	375.1	-4.9	1.5
A2				374.3	-5.7	1.5
A3				374.6	-5.4	1.5
A4				376.3	-3.7	1.5
A5				373.2	-6.8	1.5
B1				374.4	-5.6	1.5
B2				374.3	-5.7	1.5
B3				374.6	-5.4	1.5
B4				375.2	-4.8	1.5
B5				375.1	-4.9	1.5
C1				373.5	-6.5	1.5
C2				372.8	-7.2	1.5
C3				372.1	-7.9	1.5
C4				372.2	-7.8	1.5
C5				374.5	-5.5	1.5
D1				374.7	-5.3	1.5
D2				375.3	-4.7	1.5
D3				375.5	-4.5	1.5
D4				375.8	-4.2	1.5
D5				375.1	-4.9	1.5

**Calibration Results:**  
**After adjustment**

Locations	Desired (°C)	Setting (°C)	Indicating (°C)	Measured Temperature (°C)	Correction of UUC. (°C)	Uncertainty (± °C)
A1	380	380	380	379.0	-1.0	1.5
A2				378.7	-1.3	1.5
A3				379.4	-0.6	1.5
A4				379.2	-0.8	1.5
A5				379.2	-0.8	1.5
B1				379.8	-0.2	1.5
B2				379.2	-0.8	1.5
B3				379.5	-0.5	1.5
B4				378.9	-1.1	1.5
B5				379.1	-0.9	1.5
C1				379.1	-0.9	1.5
C2				377.7	-2.3	1.5
C3				378.4	-1.6	1.5
C4				378.2	-1.8	1.5
C5				378.0	-2.0	1.5
D1				379.5	-0.5	1.5
D2				378.7	-1.3	1.5
D3				379.7	-0.3	1.5
D4				379.5	-0.5	1.5
D5				379.4	-0.6	1.5

**The End of Certificate**

## ใบตรวจสอบสภาพเครื่องควบคุมอุณหภูมิ

เลขที่ใบงาน: KSPR2304362

ชนิดเครื่องมือ: Block Digestion Unit

รุ่น: KT-20s

หมายเลขเครื่อง: 5720210009/5770200073

ตรวจสอบ (รับ)		รายการตรวจเช็ค	ตรวจสอบ (ส่ง)		หมายเหตุ
15 Mar 2023			15 Mar 2023		
ปกติ	ไม่ปกติ		ปกติ	ไม่ปกติ	
		General			
<input checked="" type="checkbox"/>	<input type="checkbox"/>	1. สายไฟ	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	2. การทำงาน Main Switch	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	3. การทำงาน Selector Key	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	4. การแสดงผล Display	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	5. สภาพ Hole	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	6. สภาพฝาปิด	<input type="checkbox"/>	<input type="checkbox"/>	ไม่มี
<input checked="" type="checkbox"/>	<input type="checkbox"/>	7. สภาพตัวเครื่อง	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	8. สภาพแวดล้อม ณ สถานที่ตั้งเครื่อง	<input checked="" type="checkbox"/>	<input type="checkbox"/>	

ข้อแนะนำ :

---



---



---

Mr. Nakarin Ruenros

Service Engineer





TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert.No.: 22CH1733

Page.: 1 of 3

## Certificate of Calibration

Equipment :	pH Meter
Manufacturer :	Mettler Toledo
Model :	SevenExcellence
Serial No. :	B834291445
ID No. :	RYG_EN0152
Condition As-Received:	Used Item
Received Date :	21 December 2022
Calibration Date :	22 December 2022
Reference :	2212-0602DSC-1
Submitted by :	ALS Laboratory Group (Thailand) Co.,Ltd. Rayong Branch 616/10 Moo 5 T.Maenam Khu, A.Pluakdaeng, Rayong 21140, Thailand
Ambient Temperature :	(25 ± 2.5) °C
Relative Humidity :	(50 ± 15) %
Calibration Procedure :	In - house method : - CP-CH5 by direct measurement with standard voltage calibrator and direct measurement with certified reference material (CRM) - CP-CH8 by comparison with standard thermometer

REVIEW BY	<i>N. Bannia</i>
APPROVED BY	<i>D. [Signature]</i>
NEXT CAL. DATE	22/12/23 22/06/24

Calibrated by : Warakorn Lerngatrakul

Approved by :

*Malee*

Approved Signatory

- (☒) Malee Butkruea  
( ) Saithip Meangmai  
( ) Warakorn Lerngatrakul

Issue Date : 26 December 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.



Cert.No.: 22CH1733

Page.: 2 of 3

**Condition of this calibration result**

1. Reference Standard Instrument : -

<u>Instrument</u>	<u>Serial No.</u>	<u>ID No.</u>	<u>Cert. No.</u>	<u>Due Date</u>
1) Document Process Calibrator	54030049	130RC116	22E2769	24 Aug 2023
2) Ref. Standard Thermometer	4982054	110RC044	22I1306	27 Oct 2023

This certification is traceable to the International System of Unit maintained at:-

- Traceable to National Institute of Metrology (Thailand), NIMT

2. Certified Reference Materials : The measurement results are traceable to SI through CPA chem Ltd.,  
ANSI-ASQ National Accreditation Board, Accredited No. AR-1835

<u>Buffer Solution</u>	<u>Manufacturer</u>	<u>Lot No.</u>	<u>Exp. date</u>
pH 4.008	CPA chem	826588	09 July 2024
pH 6.987	CPA chem	823322	20 June 2023
pH 10.008	CPA chem	826590	09 July 2023

3. This certificate is valid only to the item calibrated on date and place of calibration.

**Calibration Results**

**Function :** mV Measurement

**Performing standard curve by Fluke at pH (4,7,10)**

Unit Under Calibration	Nominal Value	Standard Voltage Input	Actual Reading		Uncertainty of Measurement ( ±mV )	Coverage factor <i>k</i>
	pH	mV	mV	pH		
pH Meter S/N.: B834291445	4.000	177.48	177.3	4.000	0.058	2.00
	7.000	0.00	-0.1	7.000	0.058	2.00
	10.000	-177.48	-177.5	10.000	0.058	2.00

*Malu*





Cert.No.: 22CH1733

Page.: 3 of 3

**Calibration Results****Function :** pH Measurement

Performing three buffers standard curve by using buffer nominal pH (4,7,10)

Unit Under Calibration	Standard pH Buffer Solution	Actual pH Reading	Actual mV Reading ( mV )	Uncertainty of pH measurement ( $\pm$ )	Coverage factor $k$
pH Electrode	4.008	4.011	185.2	0.0052	2.06
S/N.: 1475518	6.987	6.990	10.4	0.0088	2.00
	10.008	10.014	-166.5	0.0072	2.00

**Function : Temperature Measurement****( \* ) Without adjustment**

This equipment was connected with Temperature Probe;

- Model : InLab Expert Pro-ISM

- Serial No. : 1475518

Dimension of probe;

- Length : 120 mm.

- Diameter : 12 mm.

- Immersion Depth : 100 mm.

Calibration Point ( $^{\circ}\text{C}$ )	Standard Temperature ( $^{\circ}\text{C}$ )	UUC* Reading ( $^{\circ}\text{C}$ )	Error ( $^{\circ}\text{C}$ )	Uncertainty of measurement ( $\pm$ $^{\circ}\text{C}$ )	Coverage factor $k$
25.0	25.001	24.9	-0.101	0.13	2.00

**Remark : - UUC\* = Unit Under Calibration**

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor  $k$ , providing a level of confidence of approximately 95 %.

-o0o-

Malu



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG, BANGKOK 10250  
TEL. 0-2717-3000-24 FAX. 0-2719-9484



## Certificate of Calibration

Certificate No. : 22E4098

Page : 1 of 2

Equipment : pH Meter  
Manufacturer: Mettler Toledo  
Model : SevenExcellence  
Serial No.: B834291445  
ID No.: RYG\_EN0152

This certificate may not be reproduced other than in full,  
except with the prior written approval of the head of  
Corporate Services 3: Equipment Calibration and Testing Services.

Condition As-Received: Used Item  
Received Date: 21 December 2022  
Calibration Date: 23 December 2022

Reference: 2212-0602DSC  
Ambient Temperature: ( 23 ± 2 ) °C  
Relative Humidity: ( 50 ± 10 ) %

Submitted by: ALS Laboratory Group (Thailand) Co.,Ltd. Rayong Branch

616/10 Moo 5, T.Maenam Khu, A.Pluakdaeng,  
Rayong 21140, Thailand

Procedure used: Calibration were conducted using In-house calibration Procedure CP-E17 According to direct measurement method with Multi-Product Calibrator.

### Condition of this result of calibration

1.Reference standards instruments :

<u>Instrument</u>	<u>Model</u>	<u>Serial No.</u>	<u>Certificate No.</u>	<u>Due Date</u>
1) Multi-Product Calibrator	5500A	6315011	22E1431	05 May 2023

2.This result of calibration was made on requested at the point specified by customer.

3.The certificate is valid only to the item calibrated on date and place of calibration.

4.This Certification is traceable to the International System of Unit maintained at:-

-National Institute of Metrology Thailand (NIMT)

Calibrated by : Wutchareeporn Wongchutikrane  
Issue Date : 26 December 2022

Approved Signatory :

☒ Phalinee Prabpaipal

☐ Nuntawat Khamchai

☐ Pornthippa Tameyakul

B 0304803





Cert. No.: 22E4098

Page.: 2 of 2

**Result of calibration :-** (\*) Without adjustment ( ) After adjustment

<b>Function:</b>	DC voltage measuremer		<b>Range:</b>	2000	mV	
<u>Standard Value</u>		<u>UUC* Reading</u>		<u>Error</u>		<u>Uncertainty</u>
( mV )		( mV )		( mV )		( $\pm \mu V$ )
-200.0000		-200.0		0.0		72
-150.0000		-150.0		0.0		69
-100.0000		-100.0		0.0		65
-50.0000		-50.0		0.0		62
0.0000		0.0		0.0		58
50.0000		50.0		0.0		62
100.0000		100.0		0.0		65
150.0000		150.0		0.0		69
200.0000		199.9		-0.1		72

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95 %

\*UUC= Unit Under Calibration.

-o0o-

Ag ✓

a 1140616

REVIEW BY	Orawan T.
APPROVED BY	Savitree N.
NEXT CAL. DATE	24.10.24

---

# Maintenance Protocol

---

Atomic Fluorescence Spectrometer  
**mercur DUO /**  
**mercur DUO plus**

---

Serial-No.: K170A0143 Customer-No.: \_\_\_\_\_  
Date: 24 May 2023 Carried out by: Srichai Fak-on

**Maintenance with following Operational Qualification (OQ)**



(requires a separate OQ protocol)

Company	บริษัท เอแอลเอส แลборาทอรี กรุ๊ป (ประเทศไทย) จำกัด
User	
Department	ห้องแลปปฏิบัติการ
Street	104 ซอย 40 ถนนพัฒนาการ แขวงสวนหลวง เขตสวนหลวง
Zip Code, City	กรุงเทพมหานคร 10250
Country	ประเทศไทย
Phone	
Fax	
E-mail	

### Maintenance works basic unit

tightness visual check inside the Mercur	<input checked="" type="checkbox"/>
visual check if gold-traps are broken	<input checked="" type="checkbox"/>
visual check if spectrometer is contaminated	<input checked="" type="checkbox"/>
visual check of the fluorescence cell	<input checked="" type="checkbox"/>
visual check of the absorption cell, incl. window	<input checked="" type="checkbox"/>
reactor cleaning	<input checked="" type="checkbox"/>
check pump-hose, if necessary change it	<input checked="" type="checkbox"/>
check swivel drive (SEV)	<input checked="" type="checkbox"/>
check drying-hose, output gas-liquid-separator	<input checked="" type="checkbox"/>
test Bubble-Sensor	<input checked="" type="checkbox"/>
check gas flows	<input checked="" type="checkbox"/>
check volume flows, reagents	<input checked="" type="checkbox"/>
recording stray light values	<input checked="" type="checkbox"/>
measurement with 30 ng/l	<input checked="" type="checkbox"/>

### Maintenance works Autosampler

Serial No.: **701 739**

lubricate the dosing-winding (Teflon-grease-spray)	<input checked="" type="checkbox"/>
clean the dosing cylinder, if necessary exchange it	<input checked="" type="checkbox"/>
lubricate the winding system of the height drive with some drops of oil	<input checked="" type="checkbox"/>
check the toothed belt	<input checked="" type="checkbox"/>
check the position of the mechanical stopper (height: 13mm )	<input checked="" type="checkbox"/>
check the pump rate of mixing pump (<14s AS52, typ.7s/<20s AS52S, typ.10s)	<input checked="" type="checkbox"/>
check the pump rate of washing cup	<input checked="" type="checkbox"/>
check the electrical hose connections for good contact	<input checked="" type="checkbox"/>
check the connectors of the magnetic valves	<input checked="" type="checkbox"/>
check the dosing hose for buckling, if necessary exchange it	<input checked="" type="checkbox"/>



Device parameter	nominal value	actual value
visual check general tightness inside the Mercur	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
visual check Goldtraps	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
visual check spectrometer		
Fluorescence cell	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
Absorption cell, incl. window	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
lens	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
Swivel drive (SEV)	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check pump hoses	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check hoses and hose connectors	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check and clean reactor	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check drying hose output Gas-liquid-seperator	o.k.: <input checked="" type="checkbox"/>	changed: <input type="checkbox"/>
check bubble-sensor	o.k.: <input checked="" type="checkbox"/>	not o.k.: <input type="checkbox"/>
<b>Check gasflow</b>		
Argon pressure valve 4	1.2 – 1.5 bar	1.5 bar
Valve 1	10 NI/h or 0.166 NL/min	0.163 NL/min
Valve 2	50 NI/h or 0.833 NL/min	0.403 NL/min
Valve 3	5 NI/h or 0.083 NL/min	0.140 NL/min
Valve 4	10 NI/h or 0.166 NL/min	0.108 NL/min
<b>Check liquidflow</b>		
Acid	2.5ml/min ± 1 ml	2.5 ml/min
Red.-agent	2.5ml/min ± 1 ml	2.5 ml/min
Sample	10ml/min ± 2 ml	10 ml/min
<b>Adventitious light - values</b>	<b>(V)</b>	<b>from file</b>
100	0	0
200	0	0
300	0	0
350	0	0
400	0	0
450	2	2
500	5	5
550	10	10
575	15	14
600	20	20

Device parameter	nominal value	actual value
<b>Analytical parameters Fluorescence cell</b>		
Conditions.: max.conc.: 10µg/L PMT-voltage: ...360.....V		
Blank-solution		Int ...0.00024...
without enrichment / FBR 30 ng/L	Int > 0.0015 RSD < 3 %	Int <sub>1</sub> ...0.00172... RSD 0.45.....%
Conditions.: max.conc.: 1.7µg/L PMT-voltage: ...352.....V		
Blank-solution		Int...0.00370...
with enrichment / FBR 30 ng/L	Int > 0.008 RSD < 3 %	Int <sub>2</sub> ...0.01060... RSD 2.38.....%
Fok.- factor ( Int <sub>2</sub> / Int <sub>1</sub> )	> 3.5	6.16
<b>Analytical parameters Absorption cell</b>		
Blank-solution		Ext. 0.00093...
without enrichment / FBR 100 ng/L	Ext. > 0.0012 RSD < 5 %	Ext. 0.00449... RSD 2.58.....%
<b>Comments</b>		

*Sorchari Fakhar.*

Signature Technician

24 May 2023

Place, Date (DD/MM/YYYY)





*Orawan T.*

Signature Customer

24 May 2023

Place, Date (DD/MM/YYYY)

## Service Report

Customer's address :		Customer's Ref. No.	
บริษัท 10110105 101010105 101010105 (101010105) 101010105			
104 100 40 101010105 101010105 101010105			
101010105 10250.			
E-mail :		Phone :	Fax :
Job No. 2305282 PM.	User :	Service Engineer : <i>ASR</i>	Date : 24/5/2023 Page : 1/1
Instrument model : Mercur	Serial No. K170A0143	Software Version No. 4.7.10.0	
<input type="checkbox"/> Repair (RE) <input checked="" type="checkbox"/> Maintenance (PM) <input type="checkbox"/> Installation (IN) <input type="checkbox"/> Warranty <input type="checkbox"/> Application (AP) <input type="checkbox"/> Site Prep.(SP) <input type="checkbox"/> Visit(VI)			
Fault / Claim : - 101010105 101010105 101010105 Po. No. 23001112 / (INV2305-037)			<input type="checkbox"/> Error Code
- 101010105 PM Contract Year 2023 (1 Time / Year 2023)			
Action taken : - Maintenance work Basic Unit		 	
- Check Device parameter.			
- Check gas flow.			
- Check liquid flow.			
- Check Adventitious light - valves			
# Test run Analytical parameter Fluorescence cell			
Test run Analytical parameter Absorption cell			
Action Pending / Recommendation : 101010105 101010105 101010105		 	
		101010105 101010105 101010105	
		Bubler sensor.	
<input type="checkbox"/> Spare Part <input type="checkbox"/> Instrument Configuration :			
Item No.	Name	Quantity	Unit Price
1.			
2.			
3.			
4.			
5.			
6.			
7.			
8.			
Herewith the undersigned confirm the time devoted, the work performed, the perfect function of the device, and the receipt/delivery of the specified spare parts. *Traveled hours and kilometers can only be entered after the return of the service engineer.		Date / Signature of Customer <i>Orawan T.</i>	Date / Signature of Service Engineer <i>ASR</i> Work completed? <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No



**Mercur**

Report file: C:\WinAAS\TMP\2023\May\Pro\_032  
 Program version: 4.7.10.0 Printed on: 5/24/2023 12:46  
 Recording started on 5/24/2023 12:35 GMT+7.0  
 Operator: PSU,OTA  
 Laboratory: ALS-BKK  
 Code: II\_Hg095\_2023

Remarks:  
 Food,water

**Method parameters****Hg**

Method Without enrichment / FBR 30ng/L\_PM24052023  
 Created on 5/24/2023 Time 12:27  
 Program ---

**Parameters Mercur Technique: Hg fluorescence**

Line	253.7 nm		
Lamp type	Hg-LP		
Integr. mode	Peak height	Integr. time	30 s
PMT	360 V		
AZ time	5 s	Peak smoothing	8/5
Delay	0 s		
	---		
Working mode	w/o enrich.	System cleaning	Acid
FBR technique	on	Wash time acid	10 s
Pump speed	3	Soaking time	20 s
Sample load time	10 s	Gas load time	5 NL/h
Reaction time	10 s		
Waiting time AZ	5 s		
Delay	0 s		
Purge time1	28 s		
Purge time2	15 s	Gas wash time2	10 NL/h

**Autosampler**

Autosampler	AS51S/F	Tray type	87/139
Working mode	continuous		

Dilution

---



**QC parameters**

QC type	Conc. check		
QC check samp. 1	---	QC check samp. 2	---
Conc.	---	Conc.	---
Error limit	---	Error limit	---
Rep. measurement	off	Reaction	flag + continue
QC std.1 no.	1(30.000 ng/L)	QC std.2 no.	1(30.000 ng/L)
QC std.1 limit	± 50.00%	QC std.2 limit	± 50.00%
QC std. act.	flag + continue		
Expect. blank abs.	0.0100± 0.0100	Reaction	flag + continue
QC precision	off		
		Reaction	off
		QC Recal.factor	Off

**Calibration settings**

Calib. meth	Standard calib.	Calibr. unit	ng/L
No. standards	1	Conversion fac.	1000000
Type of standards	---	Standard prep.	Premixed
		Blank correct.	---
		Recalib. std. no.	---
Output unit	µg/L	Conversion fac.	1000
Calib. stat.	Mean	Meas. cycles	3
		Blind cycles	1
Stock sol. 1	---	Stock sol. 2	---
Stock sol. 3	---	Stock sol. 4	---
Type of cal. curve	linear	Intercept	calculated
Weighted cal.	off	Grubbs stat.	off
Check of cal. curve	no outlier test		

**Sample statistics**

Stat. mode	Mean	Meas. cycles	2
Confid. level	95.4 %	Blind cycles	1
Grubbs stat.	---		

**Calibration standards****Hg**

No	Name	State	Pos	Conc./ ng/L	Ints	SD	RSD/%
1	Cal-Zero	(--)	79	0.000	H: 0.000249 A: 0.004274	0.000132 0.001698	53.13 39.72
2	Cal-Std1	(--)	80	30.000	H: 0.001720 A: 0.02172	0.000007 0.000023	0.459 0.107

**Calibration function 1 5/24/2023 12:44 Calibration (Peak height)**

Ints=k1+k2\*conc


k1=0.000249

k2=0.000049

Recal. factor:


---

Slope	0.00005 Ints/(ng/L)	R2-adjusted	1.0000
sc0	1.00000 ng/L		
Lower limit	0 ng/L	Upper limit	33.0 ng/L
Detection limit	---	Deter. limit	---

**Measurements and events (sorted by time)**

Hg	Without enrichment / FBR 30ng/L_PM 24052023					5/24/2023	12:35
ID	Conc.	Ints	BG	SD	RSD/%	Int. type	Time
Cal-Zero		0.000143				PKH	12:37
		0.000397					12:38
		0.000207					12:40
	0ng/L	0.000249		0.0001324	53.13		12:40
Cal-Std1		0.001720				PKH	12:42
		0.001712					12:43
		0.001728					12:44
	30.00ng/L	0.001720		0.000007897	0.459		12:44
Calibration	Calibration function: 01						12:44

Peak plots Hg



**Mercur**

Report file: C:\WinAAS\TMP\2023\May\Pro\_033  
 Program version: 4.7.10.0 Printed on: 5/24/2023 14:01  
 Recording started on 5/24/2023 13:37 GMT+7.0  
 Operator: PSU,OTA  
 Laboratory: ALS-BKK  
 Code: II\_Hg095\_2023

Remarks:  
 Food,water

**Method parameters****Hg**

Method Enrichment / FER 30ng/L PM\_24052023  
 Created on 5/24/2023 Time 13:36  
 Program ---

**Parameters Mercur Technique: Hg fluorescence**

Line	253.7 nm		
Lamp type	Hg-LP		
Integr. mode	Peak height	Integr. time	40 s
PMT	352 V		
AZ time	5 s	Peak smoothing	12/11
Delay	0 s		
	---		
Working mode	Enr. w/o reload.	System cleaning	Off
FBR technique	off	Wash time acid	10 s
Pump speed	3	Soaking time	20 s
Sample load time	10 s	Gas load time	10 NL/h
Reaction time	10 s		
Waiting time AZ	10 s	Gas AZ wait	10 NL/h
Purge time1	30 s		
Purge time2	15 s	Gas wash time2	5 NL/h
Purge time3	20 s		
Heat.time coll.1	20 s	Cool. time coll.1	30 s



**QC parameters**

QC type	Conc. check		
QC check samp. 1	---	QC check samp. 2	---
Conc.	---	Conc.	---
Error limit	---	Error limit	---
Rep. measurement	off	Reaction	flag + continue
QC std.1 no.	1(30.000 ng/L)	QC std.2 no.	1(30.000 ng/L)
QC std.1 limit	± 50.00%	QC std.2 limit	± 50.00%
QC std. act.	flag + continue		
Expect. blank abs.	0.0100± 0.0100	Reaction	flag + continue
QC precision	off		
		Reaction	off
		QC Recal.factor	Off

**Calibration settings**

Calib. meth	Standard calib.	Calibr. unit	ng/L
No. standards	1	Conversion fac.	1000000
Type of standards	---	Standard prep.	Premixed
		Blank correct.	---
		Recalib. std. no.	---
Output unit	µg/L	Conversion fac.	1000
Calib. stat.	Mean	Meas. cycles	3
		Blind cycles	1
Stock sol. 1	---	Stock sol. 2	---
Stock sol. 3	---	Stock sol. 4	---
Type of cal. curve	linear	Intercept	calculated
Weighted cal.	off	Grubbs stat.	off
Check of cal. curve	no outlier test		

**Sample statistics**

Stat. mode	off	Meas. cycles	1
Confid. level	95.4 %	Blind cycles	1
Grubbs stat.	---		

**Calibration standards****Hg**

No	Name	State	Pos	Conc./ ng/L	Ints	SD	RSD/%
1	Cal-Zero	(--)	##	0.000	H: 0.003700 A: 0.02531	0.000081 0.000153	2.192 0.607
2	Cal-Std1	(--)	##	30.000	H: 0.01060 A: 0.06689	0.000253 0.002766	2.386 4.136

**Calibration function 1 5/24/2023 14:00 Calibration (Peak height)**

Ints=k1+k2\*conc


k1=0.003700

k2=0.000230

Recal. factor:


---

Slope	0.00023 Ints/(ng/L)	R2-adjusted	1.0000
sc0	1.00000 ng/L		
Lower limit	0 ng/L	Upper limit	33.0 ng/L
Detection limit	---	Deter. limit	---

**Measurements and events (sorted by time)**

Hg	Enrichment / FER 30ng/L PM_24052023					5/24/2023	13:37
ID	Conc.	Ints	BG	SD	RSD/%	Int. type	Time
Cal-Zero	0ng/L	0.003792				PkH	13:41
		0.003666					13:43
		0.003640					13:44
		0.003700		0.000081090	2.192		13:44
Cal-Std1	30.00ng/L	0.009498				PkH	13:49
		0.008333					13:50
		0.008961					13:52
		0.008931		0.0005830	6.528		13:52
Cal-Std1	30.00ng/L	0.01031				PkH	13:57
		0.01074					13:58
		0.01076					14:00
		0.01060		0.0002530	2.386		14:00
Calibration	Calibration function: 01						14:00

Peak plots Hg



**Mercur**

Report file: C:\WinAAS\TMP\2023\May\Pro\_034  
 Program version: 4.7.10.0 Printed on: 5/24/2023 14:33  
 Recording started on 5/24/2023 14:19 GMT+7.0  
 Operator: PSU,OTA  
 Laboratory: ALS-BKK  
 Code: II\_Hg095\_2023

Remarks:  
 Food,water

**Method parameters****Hg**

Method Without enrichment / Abs / FBR 100ng/L\_PM 24052023  
 Created on 5/24/2023 Time 14:18  
 Program ---

**Parameters Mercur Technique: Hg absorption**

Line	253.7 nm		
Lamp type	Hg-LP		
Integr. mode	Peak height	Integr. time	55 s
PMT	225 V		
AZ time	5 s	Peak smoothing	2/5
Delay	8 s		
	---		
Working mode	w/o enrich.	System cleaning	Acid
FBR technique	on	Wash time acid	15 s
Pump speed	4	Soaking time	20 s
Sample load time	8 s	Gas load time	5 NL/h
Reaction time	12 s		
Waiting time AZ	15 s		
Delay	10 s		
Purge time1	50 s		
Purge time2	10 s	Gas wash time2	10 NL/h



**QC parameters**

QC type	Conc. check		
QC check samp. 1	---	QC check samp. 2	---
Conc.	---	Conc.	---
Error limit	---	Error limit	---
Rep. measurement	off	Reaction	flag + continue
QC std.1 no.	1(100.00 ng/L)	QC std.2 no.	1(100.00 ng/L)
QC std.1 limit	± 50.00%	QC std.2 limit	± 0.00%
QC std. act.	flag + continue		
Expect. blank abs.	0.0100± 0.0100	Reaction	flag + continue
QC precision	off		
		Reaction	off
		QC Recal.factor	Off

**Calibration settings**

Calib. meth	Standard calib.	Calibr. unit	ng/L
No. standards	1	Conversion fac.	1000000
Type of standards	---	Standard prep.	Premixed
		Blank correct.	---
		Recalib. std. no.	---
Output unit	µg/L	Conversion fac.	1000
Calib. stat.	Mean	Meas. cycles	3
		Blind cycles	1
Stock sol. 1	---	Stock sol. 2	---
Stock sol. 3	---	Stock sol. 4	---
Type of cal. curve	linear	Intercept	calculated
Weighted cal.	off	Grubbs stat.	off
Check of cal. curve	no outlier test		

**Sample statistics**

Stat. mode	Mean	Meas. cycles	2
Confid. level	95.4 %	Blind cycles	1
Grubbs stat.	---		

**Calibration standards****Hg**

No	Name	State	Pos	Conc./ ng/L	Abs	SD	RSD/%
1	Cal-Zero	(--)	##	0.00	H: 0.000932 A: 0.035926	0.000138 0.006208	14.88 17.28
2	Cal-Std1	(--)	##	100.00	H: 0.004494 A: 0.061286	0.000116 0.001275	2.586 2.082

**Calibration function 1****5/24/2023 14:33 Calibration (Peak height)**

$$\text{Abs} = k_1 + k_2 \cdot \text{conc}$$


k1=0.000932

k2=0.000036

Recal. factor:

---

Slope	0.00004 Abs/(ng/L)	R2-adjusted	1.0000
sc0	1.00000 ng/L	Charact. conc.	122.411 (ng/L)/1%I
Lower limit	0 ng/L	Upper limit	110. ng/L
Detection limit	---	Deter. limit	---

**Measurements and events (sorted by time)**

Hg	Without enrichment / Abs / FBR 100ng/L_PM 24052023					5/24/2023	14:19
ID	Conc.	Abs	BG	SD	RSD/%	Int. type	Time
Cal-Zero		0.001039				PkH	14:22
		0.000775					14:23
		0.000981					14:25
	0ng/L	0.000932		0.00013872	14.88		14:25
Cal-Std1		0.004528				PkH	14:29
		0.004364					14:31
		0.004589					14:33
	100.ng/L	0.004494		0.00011623	2.586		14:33
Calibration	Calibration function: 01						14:33

## Peak plots

Hg

