



# เอกสารแนบ 6


เอกสารสอบเทียบ  
เครื่องมือที่ใช้ในการวิเคราะห์



**CLC**  
Accredited  
ISO/IEC 17025



**ilac-MRA**  
MEMBER



**NIST**  
NIST-STAR 17025  
CALIBRATION 6899

**CALIBRATION LABORATORY Co., LTD.**  
2/10-11/4, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10220  
Tel. 02-578-0353-4 Fax. 02-578-2672 www.calibration.com Email: info@calibration.com

# REPORT OF CALIBRATION

**FOR**

**NOMENCLATURE** : BALANCE

**MANUFACTURER** : SHIMADZU

**MODEL / TYPE** : AP255MD

**SERIAL NO.** : D376300692[LA-001]

**LOCATION SITE** : LABORATORY - BALANCE ROOM

**DATE OF CALIBRATION** : 23 January 2021

**ENVIRONMENT CONDITIONS** :  
Temperature : 24 °C to 25 °C  
Relative Humidity : 52 % to 55 %


**PROCEDURE USED** :  
This instrument was calibrated under procedure No. WI-305-46 based on EURAMET/cg-18V version 4.0 (11/2015).  
The calibration was performed by using Weight Set which maintained by the Calibration Laboratory Co., Ltd.

**REFERENCE STANDARD USED** :  
Weight Set Phoenix Class E2 S/N. WBS-SET-E2-02.

**TRACEABILITY** :  
The measurements are traceable to International System of Units (SI), through National Institute of Metrology (Thailand).  
Certificate No. MM-0032-20, Due Date 19 May 2022.


**UNCERTAINTY** :  
The reported expanded uncertainty of measurement is stated as the standard uncertainty multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95%. It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-402 M:2013)"

**Certificate No. Q21006482**  
**F3-011-04/01-12**




g-clc-calibration


page 2 of 3



**CLC**  
Accredited  
ISO/IEC 17025



**ilac-MRA**  
MEMBER



**NIST**  
NIST-STAR 17025  
CALIBRATION 6899

**CALIBRATION LABORATORY Co., LTD.**  
2/10-11/4, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10220  
Tel. 02-578-0353-4 Fax. 02-578-2672 www.calibration.com Email: info@calibration.com

# CERTIFICATE OF CALIBRATION

**FOR**

**NOMENCLATURE** : BALANCE

**MANUFACTURER** : SHIMADZU

**MODEL / TYPE** : AP255MD

**SERIAL NO.** : D376300692[LA-001]

**CLID. NO.** : 362100172

**JOB CONTROL NO.** : 210122006482


**CUSTOMER** : ENVIRONMENTAL MEASUREMENTS CO., LTD.  
5/45 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BON SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

**DATE OF RECEIVED** : 22 January 2021  
**DATE OF ISSUED** : 28 January 2021

**Calibrated By** : Mangkornchai Lungkratok

**Approved By** :  
Authorized Signatory  
28 January 2021

**Certificate No. Q21006482**  
**F3-011-04/01-12**



g-clc-calibration

page 1 of 3

u.6/1



**CALIBRATION LABORATORY CO., LTD.**  
210-11, 14, 55 Soi Prasert Manukit 29 Yeak 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax. 02-578-2672 www.cal-lab.com E-mail: sale@cal-lab.com



**CLC**  
Accredited  
ISO/IEC 17025  
NSC-TISI-TIS 1725  
CALIBRATION 0659  
CLC

CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment ( ) adjustment

### CALIBRATION DATA



#### 1. Error of indications

Nominal Test Value (g)	Conventional mass (g)	Display Value (g)	Error of Balance (g)	Uncertainty $\pm$ (mg)	Coverage factor k
0.0000	0.0000	0.0000	0.0000	-	-
5.0000	5.0000	5.0000	0.0000	0.09	2.00
10.0000	10.0000	10.0000	0.0000	0.09	2.00
20.0000	20.0000	20.0000	0.0000	0.09	2.00
40.0000	40.0000	40.0000	0.0000	0.11	2.00
60.0000	60.0000	60.0000	0.0000	0.14	2.00
80.0000	80.0000	80.0000	0.0000	0.15	2.00
100.0000	100.0000	99.9999	-0.0001	0.15	2.00
120.0000	120.0000	119.9999	-0.0001	0.22	2.00
140.0000	140.0000	139.9999	-0.0001	0.22	2.00
160.0000	160.0000	159.9999	-0.0001	0.22	2.00
180.0000	180.0000	179.9999	-0.0001	0.25	2.00
200.0000	200.0001	199.9999	-0.0002	0.25	2.00

#### 2. Repeatability of indications

Nominal Test Value (g)	Standard Deviation of Reading (g)
200.0000	0.0000

#### 3. Effect of eccentric application of a load on the indication

5. Effect of eccentric application of a load on the indication						
<div><div></div><div><input type="checkbox"/></div></div>	<div><div></div><div><input checked="" type="checkbox"/></div></div>					
Nominal Test Value ( g )	Display Value ( g )					Maximum Difference of Center Value ( g )
	Position 1	Position 2	Position 3	Position 4	Position 5	
	50.0000	50.0000	50.0000	50.0000	50.0000	
50.0000	50.0000	50.0000	50.0000	50.0000	50.0000	0.0000

Note: The Scope of Accredited TISI Certificate No. 19C08710655 Issue 1 Page 88 of 111

This report is valid for the above stated instrument/s only.

### End of Certificate ###

Certificate No. Q21006482

F3-011-04/01-12

page 3 of 3



@clcalibration



**CALIBRATION LABORATORY CO., LTD.**  
210-11, 14, 55 Soi Prasert Manukit 29 Yeak 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax. 02-578-2672 www.cal-lab.com E-mail: sale@cal-lab.com



**CLC**  
Accredited  
ISO/IEC 17025  
NSC-TISI-TIS 1725  
CALIBRATION 0659  
CLC

## CERTIFICATE OF CALIBRATION

### FOR

NOMENCLATURE : HOT AIR OVEN  
MANUFACTURER : MEMMERT  
MODEL / TYPE : UF30  
SERIAL NO. : B119.1030[LA-005]  
CLID. NO. : 332100152  
JOB CONTROL NO. : 210122006483  
CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.  
5/45 BAAN KIANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 461 (PRAMOTE),  
NONG BON SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 22 January 2021 DATE OF ISSUED : 28 January 2021

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Wenick Inchai Sri  
Calibration Engineer

Approved By :  
28 January 2021

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q21006482

F3-011-04/01-12

page 1 of 4



@clcalibration





## CALIBRATION LABORATORY CO., LTD.

210-11, 14, 55 Soi Prasert Manulit 29 Yaek 4, Prasert Manulit Rd., Ladphrao, Bangkok 10230  
Tel: 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com Email: sale@cal-laboratory.com



ISO/IEC 17025  
CALIBRATION 0659  
CLC

### REPORT OF CALIBRATION

#### FOR

NOMENCLATURE : HOT AIR OVEN  
MANUFACTURER : MEMMERT  
MODEL / TYPE : UF30  
SERIAL NO. : B119.1030[LA-005]  
LOCATION SITE : LABORATORY-HOT ZONE  
DATE OF CALIBRATION : 23 January 2021

#### ENVIRONMENT CONDITIONS :

Temperature : 24 °C to 26 °C

Relative Humidity : 50 % to 55 %

#### PROCEDURE USED :

This instrument was calibrated under procedure No. WI-305-165 based on TLAS G-20 as calibration guidelines.

The calibration was performed by using Hydra Data Logger and Hydra Data Bucket which maintained by the Calibration Laboratory Co., Ltd.

#### REFERENCE STANDARD USED :

1. Hydra Data Logger, Fluke Model 2620 S/N: 5592550.
2. Hydra Data Bucket, Fluke Model 2635A S/N: 6496317.

#### TRACEABILITY :

1. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q20067187, Due Date 13 August 2021.
2. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q20093772, Due Date 22 October 2021.

#### UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.

It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2013)"

Certificate No. Q21006483

F3-011-04/01-12

page 2 of 4



@cccalibration



## CALIBRATION LABORATORY CO., LTD.

210-11, 14, 55 Soi Prasert Manulit 29 Yaek 4, Prasert Manulit Rd., Ladphrao, Bangkok 10230  
Tel: 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com Email: sale@cal-laboratory.com



ISO/IEC 17025  
CALIBRATION 0659  
CLC

#### CONDITION OF CALIBRATION ITEM : GOOD

#### MEASUREMENT RESULTS : (X) without adjustment ( ) adjustment

The table in the following gives the calibration results and associated measurement uncertainties of the measuring hot air oven.

#### CALIBRATION DATA

##### 1. HOT AIR OVEN PERFORMANCE

Setting ( °C )	DUC		Measured Uniformity ( °C )	Measured Stability ( °C )	Measured Overall Variation ( °C )
	Indicating ( °C )				
104.0	104.0		0.94	0.11	1.57
180.0	180.0		1.34	0.17	2.08

Certificate No. Q21006483

F3-011-04/01-12

page 3 of 4



@cccalibration



CALIBRATION LABORATORY Co., LTD.

210-11, 14, 55 Soi Prasert Manukit 20 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0333-4 Fax. 02-578-2872 www.cal-laboratory.com Email: sale@cal-laboratory.com



CLC  
Accredited  
ISO/IEC 17025


## CALIBRATION DATA

### 2. TEMPERATURE DISTRIBUTION

DUC	Measured Temperature ( °C ) @ Probe No. 9 is Ref.									Uncertainty ± ( °C )	Coverage factor k
Setting ( °C )	Indicating ( °C )	1	2	3	4	5	6	7	8	9	
104.0	104.0	105.60	105.45	104.78	104.41	104.78	105.20	105.82	104.41	104.95	2.00
180.0	180.0	181.54	181.58	180.51	180.21	180.62	181.09	181.99	180.72	180.74	2.00

Technical Note : W = 40 cm, D = 25 cm, H = 32 cm.

The Scope of Accredited TISI Certificate No. 19C087/0655 Issue 1 Page 105 of 111



This report is valid for the above stated instrument's only.

### End of Certificate ###

Certificate No. Q21006483  
F3-011-04/01-12



@ccalibration



CALIBRATION LABORATORY Co., LTD.

210-11, 14, 55 Soi Prasert Manukit 20 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0333-4 Fax. 02-578-2872 www.cal-laboratory.com Email: sale@cal-laboratory.com



CLC  
Accredited  
ISO/IEC 17025

## CERTIFICATE OF CALIBRATION

### FOR

NOMENCLATURE : WATER BATH  
MANUFACTURER : MEMMERT  
MODEL / TYPE : WNB14  
SERIAL NO. : L418.0758[LA-004]  
CLID. NO. : 332100157  
JOB CONTROL NO. : 210122006486

CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.  
5/45 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BOK SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 22 January 2021 DATE OF ISSUED : 28 January 2021

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Wenick Inchaisri  
Calibration Engineer

Approved By : [Redacted]  
Authorized Signatory

28 January 2021  
This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q21006486  
F3-011-04/01-12



@ccalibration





**CALIBRATION LABORATORY Co., LTD.**  
2/10-11/14, 55 Soi Prasert Manukit 29 Yeak 4, Prasert Manukit Rd, Ladprao, Bangkok 10230  
Tel. 02-578-0353-4 Fax. 02-578-2672 www.cal-laboratory.com E-mail: sale@cal-laboratory.com



NSC-TS-175 TS-175  
CALIBRATION 0859  
CLC

## REPORT OF CALIBRATION

### FOR

NOMENCLATURE : WATER BATH  
MANUFACTURER : MEMMERT  
MODEL / TYPE : WNB14  
SERIAL NO. : L418.0758[LA-004]  
LOCATION SITE : LABORATORY-HOT ZONE  
DATE OF CALIBRATION : 23 January 2021

#### ENVIRONMENT CONDITIONS :

Temperature : 24 °C to 26 °C

Relative Humidity : 50% to 55%

#### PROCEDURE USED :

This instrument was calibrated under procedure No. WI-305-135 based on ASTM E 715-80 as calibration guidelines.  
The calibration was performed by using Hydra Data Logger which maintained by the Calibration Laboratory Co., Ltd.

#### REFERENCE STANDARD USED :

Hydra Data Logger; Fluke Model 2620 S/N: 5592550.

#### TRACEABILITY :

The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd.  
Certificate No. Q20102430, Due Date 18 May 2021.

#### UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k = 2.00$  which for a normal distribution corresponds to a coverage probability of approximately 95 %.  
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2013)"

Certificate No. Q21006486

F3-011-04/01-12

page 2 of 4



CLC Calibration



**CALIBRATION LABORATORY Co., LTD.**  
2/10-11/14, 55 Soi Prasert Manukit 29 Yeak 4, Prasert Manukit Rd, Ladprao, Bangkok 10230  
Tel. 02-578-0353-4 Fax. 02-578-2672 www.cal-laboratory.com E-mail: sale@cal-laboratory.com



NSC-TS-175 TS-175  
CALIBRATION 0859  
CLC

#### CONDITION OF CALIBRATION ITEM : GOOD

#### MEASUREMENT RESULTS : ( X ) without adjustment ( ) adjustment

The table in the following gives the calibration results and associated measurement uncertainties of the measuring water bath.

#### CALIBRATION DATA

##### 1. WATER BATH PERFORMANCE

Test Point ( °C )	DUC Reading ( °C )	Uniformity ( °C )	Stability ( °C )
95.0	95.0	0.49	0.17

Certificate No. Q21006486

F3-011-04/01-12

page 3 of 4



CLC Calibration



**CALIBRATION LABORATORY Co., LTD.**

210-11-14, 55 Soi Prasert Manukit 29 Yeak 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax 02-578-2672 www.cal-laboratory.com Email: sale@cal-laboratory.com



## CERTIFICATE OF CALIBRATION

### FOR

NOMENCLATURE : pH METER  
MANUFACTURER : APERA  
MODEL / TYPE : PH700/201T-F  
SERIAL NO. : PH700X1019061009/N/A[PHM-01]  
CLID. NO. : 272100152  
JOB CONTROL NO. : 210121006418

CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.  
545 BAAN KIANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BOY SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 21 January 2021 ON LABORATORY DATE OF ISSUED : 25 January 2021

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By :

Sulgasem Sechanart  
Tanawan Secnam-Ngoen  
Calibration Engineer

Approved By :

Authorized Signatory  
25 January 2021

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q21006418  
F3-011-04/01-12

page 1 of 3



@calibration



**CALIBRATION LABORATORY Co., LTD.**

210-11-14, 55 Soi Prasert Manukit 29 Yeak 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax 02-578-2672 www.cal-laboratory.com Email: sale@cal-laboratory.com



## REPORT OF CALIBRATION

### FOR

NOMENCLATURE : pH METER  
MANUFACTURER : APERA  
MODEL / TYPE : PH700/201T-F  
SERIAL NO. : PH700X1019061009/N/A[PHM-01]  
DATE OF CALIBRATION : 22 January 2021

#### ENVIRONMENT CONDITIONS :

Temperature :  $(25 \pm 2.5)$  °C Relative Humidity :  $(50 \pm 15)$  % RH

#### PROCEDURE USED :

This instrument was calibrated under procedure No. WI-305-128, 238 by direct measurement with Certified Reference Material (CRM) and comparison with Calibration Bath, Precision Thermometer and IPRT which maintained by the Calibration Laboratory Co., Ltd.

#### REFERENCE STANDARD USED :

- pH Standard Solution, TRM CODE TRM-S-2003, TRM CODE TRM-S-2005, TRM CODE TRM-S-2007.
- Calibration Bath, Kambie Model OB-222 ULT S/N. 17115653.
- Precision Thermometer, ASL Model F200-A-8 S/N. 014433/03 with IPRT, ASL Model T100-250-ID S/N. PO106346-1-13.

#### TRACEABILITY :

- The measurements are traceable to International System of Units (SI), through National Institute of Metrology (Thailand), Lot Number, 280319, 280119, 080719, Due Date 16 June 2021.
- The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q21008277, Due Date 05 February 2021.
- The measurements are traceable to International System of Units (SI), through National Institute of Metrology (Thailand), Certificate No. ER-0032-20, TT-0011-20, Due Date 09 April 2021, 06 February 2021.

#### UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %. It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4:02 M:2013)"

Certificate No. Q21006418  
F3-011-04/01-12

page 2 of 3



@calibration



MEASUREMENT RESULTS: (X) without adjustment ( ) adjustment

# CERTIFICATE OF CALIBRATION

FOR

NOMENCLATURE : DO METER

MANUFACTURER : YSI

MODEL / TYPE	:	5000-230V/5010
--------------	---	----------------

SERIAL NO. : 16D101626/19D100367[DOM-01]

CLID. NO. : 272100329

JOB CONTROL NO. : 210204010947

CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.

545 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BON SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

Immersion depth (mm)	Actual Temperature (°C)	DUC Reading (°C)	Correction (°C)	Uncertainty ± (°C)
100	25.00	25.0	0.00	0.07

Note. \* means Calibrations marked "Not TISI Accredited" in this Certificate have been included for completeness.

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor of  $k = 2.00$ .

DATE OF RECEIVED : 04 February 2021

DATE OF ISSUED : 08 February 2021

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

**Calibrated By :**

**Sukgasm Seehanat**  
**Calibration Engineer**

Approved By :

Authorized Signatory

08 February 2021

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q21006418  
F3-011-04/01-12

Certificate No. Q21010947  
E3-011-04/01-12

page 3 of 3



**Subclassification**

page 1 of 3



@clccalibration





CALIBRATION LABORATORY Co., LTD.

210-11, 14, 55 Soi Prasert Manukit 29 Yeark 4, Prasert Manukit Rd, Ladphrno, Bangkok 10230  
Tel. 02-578-0353-4 Fax. 02-578-2672 www.cal-laboratory.com Email: sale@cal-laboratory.com



ISO/IEC 17025

## REPORT OF CALIBRATION

### FOR

NOMENCLATURE : DO METER  
MANUFACTURER : YSI  
MODEL / TYPE : 5000-230V/5010  
SERIAL NO. : 16D101626/19D100367IDOM-01]  
DATE OF CALIBRATION : 05 February 2021

#### ENVIRONMENT CONDITIONS :

Temperature :  $(25 \pm 2.5) ^\circ\text{C}$

Relative Humidity :  $(50 \pm 15) \% \text{RH}$

#### PROCEDURE USED :

This instrument was calibrated under procedure No. CLC-CPCH-06 by direct measurement with Dissolved Oxygen which maintained by the Calibration Laboratory Co., Ltd.

#### REFERENCE STANDARD USED :

Dissolved Oxygen, Sigma-Aldrich Product ID QC3077-500ML.

#### TRACEABILITY :

The measurements are traceable to International System of Units (SI), through Sigma-Aldrich, Lot LRAC4478, Due Date January 2022.

#### UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k = 2.00$  which for a normal distribution corresponds to a coverage probability of approximately 95 %. It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2013)"

Certificate No. Q21010947

F3-011-04/01-12

page 2 of 3



@ccalibration



CALIBRATION LABORATORY Co., LTD.

210-11, 14, 55 Soi Prasert Manukit 29 Yeark 4, Prasert Manukit Rd, Ladphrno, Bangkok 10230  
Tel. 02-578-0353-4 Fax. 02-578-2672 www.cal-laboratory.com Email: sale@cal-laboratory.com



ISO/IEC 17025

#### CONDITION OF CALIBRATION ITEM : GOOD

#### MEASUREMENT RESULTS : (X) without adjustment ( ) adjustment

The table in the following gives the calibration results and associated measurement uncertainties of Do Meter.

#### CALIBRATION DATA

Nominal Value (mg/L)	DUC Reading (mg/L)	Correction (mg/L)	Uncertainty (mg/L)
8.49	8.49	0.00	$\pm 0.31$

Note: The Scope of Accredited ANAB Certificate No. ACDM-2814 Version 006 Page 4 of 57

This report is valid for the above stated instrument/s only.

### End of Certificate ###

Certificate No. Q21010947

F3-011-04/01-12

page 3 of 3



@ccalibration



**CALIBRATION LABORATORY Co., LTD.**  
2/10-11, 14, 55 Soi Prasert Manulit 29 Yeak 4, Prasert Manulit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax. 02-578-2672 www.cal-laboratory.com E-mail: sale@cal-laboratory.com



NSC-TIS-TIS 17025  
CALIBRATION 0859  
CLC

## CERTIFICATE OF CALIBRATION

### FOR

NOMENCLATURE : INCUBATOR  
MANUFACTURER : ACCUPLUS  
MODEL / TYPE : SMART 1250  
SERIAL NO. : 2059-0718-0010[LA-002]  
CLID. NO. : 332100155  
JOB CONTROL NO. : 210122006484

CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.  
545 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTTE),  
NONG BOH SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 22 January 2021 DATE OF ISSUED : 28 January 2021

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By :

Wenick Inchaistri  
Calibration Engineer

Approved By :

Authorized Signatory

28 January 2021

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q21006484  
F3-011-04/01-12

page 1 of 4



@cccalibration



**CALIBRATION LABORATORY Co., LTD.**  
2/10-11, 14, 55 Soi Prasert Manulit 29 Yeak 4, Prasert Manulit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax. 02-578-2672 www.cal-laboratory.com E-mail: sale@cal-laboratory.com



NSC-TIS-TIS 17025  
CALIBRATION 0859  
CLC

## REPORT OF CALIBRATION

### FOR

NOMENCLATURE : INCUBATOR  
MANUFACTURER : ACCUPLUS  
MODEL / TYPE : SMART 1250  
SERIAL NO. : 2059-0718-0010[LA-002]  
LOCATION SITE : LABORATORY  
DATE OF CALIBRATION : 23 January 2021

ENVIRONMENT CONDITIONS :  
Temperature : 24 °C to 26 °C  
Relative Humidity : 50 % to 55 %

PROCEDURE USED :  
This instrument was calibrated under procedure No. WI-305-165 based on TIAS G-20 as calibration guidelines.  
The calibration was performed by using Hydra Data Logger which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

Hydra Data Logger: Fluke Model 2620 SN: 5592550.

TRACEABILITY :

The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd.  
Certificate No. Q20067187, Due Date 13 August 2021.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.  
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-402 M:2013)"

Certificate No. Q21006484  
F3-011-04/01-12

page 2 of 4



@cccalibration





CALIBRATION LABORATORY CO., LTD.

270-11, 14, 55 Soi Prasert Manukit 29 Yeak 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cali-laboratory.com E-mail: sale@cali-laboratory.com



NSC-TISI-TIS 17025  
CALIBRATION 0859  
CLC

## CONDITION OF CALIBRATION ITEM : GOOD

### MEASUREMENT RESULTS : ( X ) without adjustment ( ) adjustment

The table in the following gives the calibration results and associated measurement uncertainties of the measuring incubator.

#### CALIBRATION DATA

##### 1. INCUBATOR PERFORMANCE

DUC		Measured Uniformity	Measured Stability	Measured Overall
Setting ( °C )	Indicating ( °C )	( °C )	( °C )	Variation ( °C )
20.0	20.0	0.34	0.44	1.02



CALIBRATION LABORATORY CO., LTD.

270-11, 14, 55 Soi Prasert Manukit 29 Yeak 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cali-laboratory.com E-mail: sale@cali-laboratory.com



NSC-TISI-TIS 17025  
CALIBRATION 0859  
CLC


## CALIBRATION DATA

### 2. TEMPERATURE DISTRIBUTION

DUC		Measured Temperature ( °C ) ( @Probe No.9 is Ref.									Uncertainty	Coverage
Setting ( °C )	Indicating ( °C )	1	2	3	4	5	6	7	8	9	± ( °C )	factor k
20.0	20.0	20.49	20.42	20.27	20.33	20.41	20.34	20.19	20.16	20.28	0.67	2.00

Technical Note : W = 50 cm, D = 48 cm, H = 110 cm.

The Scope of Accredited TISI Certificate No. 19C087/0655 Issue 1 Page 105 of 111



This report is valid for the above stated instrument/s only.

Certificate No. Q21006484  
F3-011-04/01-12

page 3 of 4



@cccalibration

### End of Certificate ###

Certificate No. Q21006484  
F3-011-04/01-12

page 4 of 4



@cccalibration



**CALIBRATION LABORATORY Co., LTD.**  
210-11-14, 55 Soi Praent Manukit 20 Yaek 4, Praent Manukit Rd., Ladphrao, Bangkok 10230  
Tel: 02-578-0353-4 Fax: 02-578-2872 www.cal-laboratory.com Email:sale@cal-laboratory.com



## CERTIFICATE OF CALIBRATION

### FOR

NOMENCLATURE : DIGITAL THERMOHYGRO METER  
MANUFACTURER : DIGICON  
MODEL / TYPE : TH-02A  
SERIAL NO. : 1919E0284991[DTH-01]  
CLID. NO. : 232100200  
JOB CONTROL NO. : 210121006416

CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.  
545 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BON SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 21 January 2021 DATE OF ISSUED : 26 January 2021

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By :

Oranut Kamchatphai  
Calibration Engineer

Approved By :

Authorized Signatory

26 January 2021

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q21006416  
F3-011-04/01-12

page 1 of 3



@calibration



**CALIBRATION LABORATORY Co., LTD.**  
210-11-14, 55 Soi Praent Manukit 20 Yaek 4, Praent Manukit Rd., Ladphrao, Bangkok 10230  
Tel: 02-578-0353-4 Fax: 02-578-2872 www.cal-laboratory.com Email:sale@cal-laboratory.com



## REPORT OF CALIBRATION

### FOR

NOMENCLATURE : DIGITAL THERMOHYGRO METER  
MANUFACTURER : DIGICON  
MODEL / TYPE : TH-02A  
SERIAL NO. : 1919E0284991[DTH-01]  
DATE OF CALIBRATION : 25 January 2021

#### ENVIRONMENT CONDITIONS :

Temperature :  $(23 \pm 2) ^\circ\text{C}$  Relative Humidity :  $(55 \pm 10) \% \text{RH}$

#### PROCEDURE USED :

This instrument was calibrated under procedure No. WI-305-74 as calibration guidelines.

The calibration was performed by using Chilled Mirror Hygrometer and Temperature & Humidity Chamber which maintained by the Calibration Laboratory Co., Ltd.

#### REFERENCE STANDARD USED :

Chilled Mirror Hygrometer, EdgeTech Model Dew Master SN. 36151.  
Temperature & Humidity Chamber, PGC Model 9141-5114 SN.0802282.

#### TRACEABILITY :

The measurements are traceable to International System of Units (SI), through Thunder Scientific Corporation.  
Certificate No. 18263, Due Date 29 April 2021.

#### UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k = 2,00$  which for a normal distribution corresponds to a coverage probability of approximately 95 %.  
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4:02 M:2013)"

Certificate No. Q21006416  
F3-011-04/01-12

page 2 of 3



@calibration





CALIBRATION LABORATORY CO., LTD.

2/10-11, 14, 55 Soi Prasert Manukit 29 Yeark 4, Prasert Manukit Rd., Ladphrao, Bangkok, 10230  
Tel. 02-578-0353-4 Fax. 02-578-2672 www.cal-lab.com E-mail: sale@cal-lab.com



NSC-TISIR TISIR  
CALIBRATION 0899



Accredited  
ISO/IEC 17025

#### CONDITION OF CALIBRATION ITEM : GOOD

#### MEASUREMENT RESULTS : ( X ) without adjustment ( ) adjustment

The table in the following gives the calibration results and associated measurement uncertainties of the measuring digital thermohygro meter.

#### CALIBRATION DATA

##### 1. CORRECTION OF TEMPERATURE

Test point ( ° C )	Actual Temperature ( ° C )	DUC Reading ( ° C )	Correction ( ° C )	Uncertainty ± ( ° C )
20.0	20.03	19.4	+0.63	0.40
25.0	25.04	24.4	+0.64	
30.0	30.13	29.5	+0.63	

##### 2. CORRECTION OF HUMIDITY

STD Temperature ( ° C )	STD Reading ( %RH )	DUC Reading ( %RH )	Correction ( %RH )	Uncertainty ± ( %RH )
25	40.0	36	+4.0	1.3
25	60.0	56	+4.0	1.5

Note: The Scope of Accredited TISIR Certificate No. 19C087/0655 Issue 1 Page 36 of 111

This report is valid for the above stated instrument/s only.

### End of Certificate ###

Certificate No. Q21006416

F3-011-04/01-12

page 3 of 3



@cccalibration



CALIBRATION LABORATORY CO., LTD.

2/10-11, 14, 55 Soi Prasert Manukit 29 Yeark 4, Prasert Manukit Rd., Ladphrao, Bangkok, 10230  
Tel. 02-578-0353-4 Fax. 02-578-2672 www.cal-lab.com E-mail: sale@cal-lab.com



NSC-TISIR TISIR  
CALIBRATION 0899



Accredited  
ISO/IEC 17025

## CERTIFICATE OF CALIBRATION

### FOR

NOMENCLATURE : REFRIGERATOR  
MANUFACTURER : MEDICOOL  
MODEL / TYPE : BB-117  
SERIAL NO. : BB117-190725001[LA-003]  
CLID. NO. : 332100156  
JOB CONTROL. NO. : 210122006485

CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.

545 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BON SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 22 January 2021 DATE OF ISSUED : 28 January 2021

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By :

Wenick Inchaisri

Approved By :

Authorized Signatory

28 January 2021

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q21006485

F3-011-04/01-12

page 1 of 4



@cccalibration



page 3 of 4







# CALIBRATION LABORATORY CO., LTD.

210-11,14, 55 Soi Prasert Manulit 29 Yaek 4, Prasert Manulit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com




## CALIBRATION DATA

### 2. TEMPERATURE DISTRIBUTION

DUC	Measured Temperature ( °C )@Probe No.9 is Ref.										Uncertainty ± ( °C )	Coverage factor k
Setting ( °C )	1	2	3	4	5	6	7	8	9			
4.0	3.47	3.83	3.21	3.80	3.41	3.32	3.12	3.13	3.43		1.15	2.00

Technical Note : W = 50 cm, D = 50 cm, H = 120 cm.

The Scope of Accredited TISI Certificate No. 19C087/0655 Issue 1 Page 105 of 111



This report is valid for the above stated instrument/s only.

### End of Certificate ###

Certificate No. Q21006485  
F3-011-04/01-12

page 4 of 4



etccalibration



# CALIBRATION LABORATORY CO., LTD.

210-11,14, 55 Soi Prasert Manulit 29 Yaek 4, Prasert Manulit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



## CERTIFICATE OF CALIBRATION

### FOR

NOMENCLATURE : DIGITAL THERMOHYGRO METER  
MANUFACTURER : DIGICON  
MODEL / TYPE : TH-02A  
SERIAL NO. : 1919E0284980[DTH-02]  
CLID. NO. : 232100201  
JOB CONTROL NO. : 210121006417

CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.  
545 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BON SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 21 January 2021 DATE OF ISSUED : 26 January 2021

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By :

Oranut Kamchatphai  
Calibration Engineer

Approved By :

26 January 2021

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units ( SI )

Certificate No. Q21006417  
F3-011-04/01-12

page 1 of 3



etccalibration



# CALIBRATION LABORATORY Co., LTD.

2/10-11/14, 55 Soi Prasert Manukit 29 Year 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail: sale@cal-laboratory.com



NSC-TIS-17925  
CALIBRATION 0059  
CLC

## REPORT OF CALIBRATION

### FOR

NOMENCLATURE : DIGITAL THERMOHYGRO METER  
MANUFACTURER : DIGICON  
MODEL / TYPE : TH-02A  
SERIAL NO. : 1919E0284980[DTH-02]  
DATE OF CALIBRATION : 25 January 2021

#### ENVIRONMENT CONDITIONS :

Temperature :  $(23 \pm 2) ^\circ\text{C}$

Relative Humidity :  $(55 \pm 10) \% \text{RH}$

#### PROCEDURE USED :

This instrument was calibrated under procedure No. WI-305-74 as calibration guidelines.

The calibration was performed by using Chilled Mirror Hygrometer and Temperature & Humidity Chamber which maintained by the Calibration Laboratory Co., Ltd.

#### REFERENCE STANDARD USED :

Chilled Mirror Hygrometer, Edgetech Model Dew Master S/N. 36151.  
Temperature & Humidity Chamber, PGC Model 9141-5114 S/N.0802282.

#### TRACEABILITY :

The measurements are traceable to International System of Units (SI), through Thunder Scientific Corporation. Certificate No. 18263, Due Date 29 April 2021.

#### UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k = 2.00$  which for a normal distribution corresponds to a coverage probability of approximately 95 %. It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2013)"

Certificate No. Q21006417

F3-011-04/01-12

page 2 of 3



edcalibration



# CALIBRATION LABORATORY Co., LTD.

2/10-11/14, 55 Soi Prasert Manukit 29 Year 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail: sale@cal-laboratory.com



NSC-TIS-17925  
CALIBRATION 0059  
CLC

## REPORT OF CALIBRATION

### FOR

NOMENCLATURE : DIGITAL THERMOHYGRO METER  
MANUFACTURER : DIGICON  
MODEL / TYPE : TH-02A  
SERIAL NO. : 1919E0284980[DTH-02]  
DATE OF CALIBRATION : 25 January 2021

#### ENVIRONMENT CONDITIONS :

Temperature :  $(23 \pm 2) ^\circ\text{C}$

Relative Humidity :  $(55 \pm 10) \% \text{RH}$

#### PROCEDURE USED :

This instrument was calibrated under procedure No. WI-305-74 as calibration guidelines.

The calibration was performed by using Chilled Mirror Hygrometer and Temperature & Humidity Chamber which maintained by the Calibration Laboratory Co., Ltd.

#### REFERENCE STANDARD USED :

Chilled Mirror Hygrometer, Edgetech Model Dew Master S/N. 36151.  
Temperature & Humidity Chamber, PGC Model 9141-5114 S/N.0802282.

#### TRACEABILITY :

The measurements are traceable to International System of Units (SI), through Thunder Scientific Corporation. Certificate No. 18263, Due Date 29 April 2021.

#### UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k = 2.00$  which for a normal distribution corresponds to a coverage probability of approximately 95 %. It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2013)"

Certificate No. Q21006417

F3-011-04/01-12

page 2 of 3



edcalibration