

## ภาคผนวก 6

เอกสารขึ้นทะเบียนบริษัทวิเคราะห์คุณภาพน้ำ

สำเนาใบอนุญาตการขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน



ที่ อก ๐๓๑๐(๑)/ ๑๘๗ ๙

กรมโรงงานอุตสาหกรรม  
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท  
เขตราชเทวี กรุงเทพฯ ๑๐๔๐๐

๐ ๙ กุมภาพันธ์ ๒๕๖๕

เรื่อง ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

เรียน กรรมการผู้จัดการ บริษัท ยูโนเด็ค แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด

อ้างถึง คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และชนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน  
ลงวันที่ ๒๗ ธันวาคม ๒๕๖๔

- สิ่งที่ส่งมาด้วย ๑. รายชื่อผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ จำนวน ๔๐ ราย  
๒. รายชื่อเจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๑๐๖ ราย  
๓. ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม

ตามหนังสือที่อ้างถึง บริษัท ยูโนเด็ค แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด ขอต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน เลขทะเบียน ว-๑๔๕ สถานที่ตั้งเลขที่ ๓ ซอยอุดมสุข ๔๑ ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง กรุงเทพมหานคร ต่อกรมโรงงานอุตสาหกรรม นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว ให้บริษัท ยูโนเด็ค แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน โดยมีองค์ประกอบดังนี้  
ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ จำนวน ๔๐ ราย ตามสิ่งที่ส่งมาด้วย ๑  
ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๑๐๖ ราย ตามสิ่งที่ส่งมาด้วย ๒  
ค. ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนให้วิเคราะห์ในน้ำเสีย น้ำใต้ดิน อากาศเสีย สิ่งปฏิกูล หรือวัสดุที่ไม่ใช้แล้ว และดิน ตามสิ่งที่ส่งมาด้วย ๓

หนังสือฉบับนี้จะหมดอายุในวันที่ ๒ กุมภาพันธ์ ๒๕๖๘ หากประสงค์จะต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ให้ยื่นคำขอต่ออายุพร้อมเอกสารประกอบคำขอต่อม กรมโรงงานอุตสาหกรรมภายใน ๓๐ วัน ก่อนวันสิ้นอายุของหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ทั้งนี้ สามารถยื่นคำขอผ่านระบบอิเล็กทรอนิกส์ได้ที่หน้าเว็บไซต์กรมโรงงานอุตสาหกรรม ตาม QR Code ท้ายหนังสือฉบับนี้

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ



ผู้อำนวยการกองวิจัยและเตือนภัยมลพิษโรงงาน  
ปฏิบัติการตามมาตรฐานโรงงานอุตสาหกรรม  
**IAE**  
CONSULTANT COMPANY LIMITED

ยื่นคำขอผ่านระบบอิเล็กทรอนิกส์

**สำเนาถูกต้อง**

กองวิจัยและเตือนภัยมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๐๓-๕

โทรสาร ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๙๙

ไปรษณีย์อิเล็กทรอนิกส์ saraban@diw.gmail.go.th

ที่ อก ๐๓๑๐(๑)/ ๔๗๘ ๗



กรมโรงงานอุตสาหกรรม  
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท  
เขตราชเทวี กรุงเทพฯ ๑๐๔๐๐

๒ ๑ เมษายน ๒๕๖๕

เรื่อง เปลี่ยนแปลงบุคลากรของห้องปฏิบัติการวิเคราะห์

เรียน กรรมการผู้จัดการ บริษัท ยูไนเต็ต แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด

อ้างถึง คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และชนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน  
ลงวันที่ ๓๐ มีนาคม ๒๕๖๕

ตามหนังสือที่อ้างถึง บริษัท ยูไนเต็ต แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด  
ห้องปฏิบัติการวิเคราะห์เอกชน เลขทะเบียน ว-๑๔๕ สถานที่ตั้งเลขที่ ๓ ซอยอุดมสุข ถนนสุขุมวิท แขวงบางจาก  
เขตพระโขนง กรุงเทพมหานคร ขอเปลี่ยนแปลงบุคลากรของห้องปฏิบัติการวิเคราะห์ ความละเอียดแจ้งแล้ว นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว มีความเห็นดังนี้

๑. ให้ยกเลิกผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ จำนวน ๒ ราย

๑) นางมานิดา แยมไย ทะเบียนเลขที่ ว-๑๔๕-ค-๐๐๐๕

๒) นางสาวนภสรณ คงขำ ทะเบียนเลขที่ ว-๑๔๕-ค-๐๐๓๒

๒. ให้ยกเลิกเจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๒ ราย

๑) นางสาวศิริพร อภิการัตน์ ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๖๔

๒) นางสาวพรนัชชา กลิ่นฉุน ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๘๔

๓. ให้เพิ่มเจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๒ ราย


๑) นางสาวธัญลักษณ์ ธนโชติกาญจนการ ทะเบียนเลขที่ ว-๑๔๕-จ-๐๑๐๗

๒) นางสาวจันทร์จิรา ประกอบทรัพย์ ทะเบียนเลขที่ ว-๑๔๕-จ-๐๑๐๘

อนึ่ง หนังสือฉบับนี้จะหมดอายุพร้อมหนังสือต่ออายุรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน  
ที่ อก ๐๓๑๐(๑)/๑๘๗๙ ลงวันที่ ๙ กุมภาพันธ์ ๒๕๖๕ คือในวันที่ ๒ กุมภาพันธ์ ๒๕๖๘ ทั้งนี้ สามารถยื่นคำขอ  
ผ่านระบบอิเล็กทรอนิกส์ได้ที่หน้าเว็บไซต์กรมโรงงานอุตสาหกรรม ตาม QR Code ท้ายหนังสือฉบับนี้

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ



ปฏิบัติราชการแทนอธิบดีกรมโรงงานอุตสาหกรรม

ยื่นคำขอผ่านระบบอิเล็กทรอนิกส์

กองวิจัยและเตือนภัยมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๐๓-๕ โทรสาร ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๔๙

ไปรษณีย์อิเล็กทรอนิกส์ saraban@diw.mail.go.th

UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำเนาถูกต้อง



“อุตสาหกรรมก้าวไกล ประเทศไทยก้าวหน้า ร่วมกันพัฒนา อุตสาหกรรมสีเขียว”



เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท ยูไนเต็ต แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด เลขทะเบียน ว-๑๔๕

ที่ อก ๐๓๑๐(๑)/

๑ ๘ ๗ ๙

ลงวันที่ ๐ ๙ กุมภาพันธ์ ๒๕๖๕

ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ จำนวน ๔๐ ราย

๑) นางสาวกฤษวรรณ ภัทรธีรกุล	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๐๑
๒) นายณรงค์ ฉิมพาลี	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๐๒
๓) นางสาวนันทิดา บุญไสย	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๐๓
๔) นางปิยะพัชร สุทมนัสวงษ์	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๐๔
๕) นางมานิดา แย้มใย	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๐๕
๖) นางสาวเบญจวรรณ วิริโยทัย	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๐๖
๗) นายณพรัตน์ วงศ์อนุรักษชัย	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๐๗
๘) นางสาวฉวีวรรณ บุญลา	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๐๘
๙) นายสุวิทย์ จอดนอก	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๐๙
๑๐) นางสาวโชติภา สมบรรณ	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๑๐
๑๑) นางสาวบุษกร เลิศภานุมาศ	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๑๑
๑๒) นางสาววิไลลักษณ์ ศรีสุข	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๑๒
๑๓) นางสาวปวีณา จรัสโชติพินิต	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๑๓
๑๔) นายศिला บรรจงใจรักษ์	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๑๔
๑๕) นายปฏิกรณ์ คณะนา	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๑๕
๑๖) นายธีรวัฒน์ ชมมิ่ง	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๑๖
๑๗) นางสาวศิริพร ศรีประดิษฐ์	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๑๗
๑๘) นางสาวสาวิตรี รุ่ง	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๑๘
๑๙) นางสาวนพวรรณ อูรารักษ์	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๑๙
๒๐) นายภูซงค์ พานิชย์เลิศอำไพ	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๒๐
๒๑) นายณัฐวัฒน์ แดงสวัสดิ์	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๒๑
๒๒) นายเอกรัตน์ ปละคามินทร์	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๒๒
๒๓) นางสาวนิศารัตน์ ศรีสกุลสิทธิโชค	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๒๓
๒๔) นางสาวเจตจรินทร์ ทำสะอาด	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๒๔
๒๕) นางสาวสุพรรณ คงทอง	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๒๕
๒๖) นางสาววรกร พัดสองชั้น	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๒๖
๒๗) นายวิรัช โมกแก้ว	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๒๗
๒๘) นายวัชรพงษ์ เทพดนตรี	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๒๘
๒๙) นายอนุศาสน์ สวดี	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๒๙
๓๐) นายกรวิทย์ เจียศิริสกุล	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๓๐
๓๑) นางสาวอริกา รงค์สวัสดิ์	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๓๑
๓๒) นางสาวนภสวรรณ คงข้า	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๓๒
๓๓) นายสุทธิระ อรุณจันทร์	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๓๓
๓๔) นางสาวทัศนีย์ อ่อนคำ	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๓๔
๓๕) นางสาวพริ้มพรรณ สมบูรณ์	ทะเบียนเลขที่	ว-๑๔๕-ค-๐๐๓๕

**UAE**  
UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

**ผู้ควบคุมดูแลห้องปฏิบัติการ**

- ๓๖) นายสุภณัฐ คุณชนกาญจน์
- ๓๗) นางสาวศิริภาพร เหมือนแร่
- ๓๘) นางสาวนัส ข่านิล
- ๓๙) นางสาวพรนิกา อีระจินดาชล
- ๔๐) นายนาคินทร์ พันธุ์ชาติกุล


ทะเบียนเลขที่ ว-๑๔๕-ค-๐๐๓๖

ทะเบียนเลขที่ ว-๑๔๕-ค-๐๐๓๗

ทะเบียนเลขที่ ว-๑๔๕-ค-๐๐๓๘

ทะเบียนเลขที่ ว-๑๔๕-ค-๐๐๓๙

ทะเบียนเลขที่ ว-๑๔๕-ค-๐๐๔๐



ผู้อำนวยการกองวิจัยและเตือนภัยมลพิษโรงงาน  
ปฏิบัติราชการแทนอธิบดีกรมโรงงานอุตสาหกรรม



  
สำเนาถูกต้อง

เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท ยูไนเต็ท แอนนาไลส์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด เลขทะเบียน ว-๑๔๕

ที่ อก ๐๓๑๐(๑)/ ๑๘๗๙ ลงวันที่ ๐๙ กุมภาพันธ์ ๒๕๖๕


ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๑๐๖ ราย

๑) นายสุชนันต์ พันสิงห์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๐๑
๒) นางสาวสุธรรมา แก้วชื่อนอก	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๐๒
๓) นายพีรณัฐ เจริญผล	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๐๓
๔) นางสาววิไลลักษณ์ เกื้อสง	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๐๔
๕) นายสมชาติ อุทุมรัตน์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๐๕
๖) นางสาวปรมาภรณ์ ทองแก้ว	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๐๖
๗) นางสาวกัลยา สมพงษ์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๐๗
๘) นายอรรถพร เทพทอง	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๐๘
๙) นางสาวอมรรัตน์ พุทธาธิ	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๐๙
๑๐) นางสาววรรณิ์ สายบุญเรือน	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๑๐
๑๑) นายกฤษณพงษ์ นามทิพย์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๑๑
๑๒) นางสาวอาภรณ์ อ่อนคง	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๑๒
๑๓) นายกิตติศักดิ์ ทรงจำรัส	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๑๓
๑๔) นางสาวอักษรินทร์ บุญคง	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๑๔
๑๕) นางสาวพรพิมล แวนทอง	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๑๕
๑๖) นายวิชญ์ สุวรรณราช	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๑๖
๑๗) นายอภิวิชญ์ ท่วงที	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๑๗
๑๘) นายมานิตย์ ปานโชติ	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๑๘
๑๙) นายทศพร ธนะพิรุฬห์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๑๙
๒๐) นางสาวกัลยาณี โยธา	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๒๐
๒๑) นางสาวเกวลี สุขรี	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๒๑
๒๒) นางสาวชมชนัญ อภิพัทธ์ปภา	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๒๒
๒๓) นายศิริพัชร จงผดุงเกียรติ	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๒๓
๒๔) นางสาวสุภาวดี อินยาศรี	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๒๔
๒๕) นายพงศ์เทพ เหล่าขจร	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๒๕
๒๖) นายขวัญชัย พันทุกข์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๒๖
๒๗) นางสาวพัชจิรา คติพิศาล	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๒๗
๒๘) นางสาวเมวิกา เสือคำจันทร์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๒๘
๒๙) นายกานต์พงศ์ บุญพวง	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๒๙
๓๐) นางสาวพริดา เจริญชัยสมบัติ	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๓๐
๓๑) นายณพรัตน์ จะโต	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๓๑
๓๒) นายพีระพัฒน์ บัญญัติศิลป์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๓๒
๓๓) นายปรีดา ไชยภูมิสกุล	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๓๓
๓๔) นายชัชวาลย์ เลื่อนล่อง	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๓๔
๓๕) นายปิยะณัฐ ศรีภูโรจน์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๓๕



นางสาวกมลทิพย์...  
นางสาวกมลทิพย์...  
นางสาวกมลทิพย์...  
นางสาวกมลทิพย์...

๓๖) นายณภสินธุ์ ธนุธรรมรัตน์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๓๖
๓๗) นายกันนิกร ระโส	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๓๗
๓๘) นายจักรพันธ์ ภูมิรินทร์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๓๘
๓๙) นายปริญา กลมเกลียว	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๓๙
๔๐) นายธีรวัฒน์ มาตรโพธิ์ศรี	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๔๐
๔๑) นายธีรเมธ สุขศรี	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๔๑
๔๒) นายบุญญฤทธิ์ ก้อนสิน	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๔๒
๔๓) นายพรชวุฒิ โฉวสกุล	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๔๓
๔๔) นายอชิตะ แสงจันทร์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๔๔
๔๕) นายณัฐพงศ์ เมืองชัย	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๔๕
๔๖) นายธนัท เลิศประเสริฐ	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๔๖
๔๗) นางสาวนิภาพร จันทเขตต์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๔๗
๔๘) นายยุทธพงษ์ อิศระสุข	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๔๘
๔๙) นายรณภพ ภูตระกูลพัฒนา	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๔๙
๕๐) นางสาวศิริวรรณ ชอนพา	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๕๐
๕๑) นายสมพงศ์ สกฤไทย	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๕๑
๕๒) นายสุริยัน นิธิเชิดชูวงศ์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๕๒
๕๓) นายอัษฎาวุธ ยนศิริ	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๕๓
๕๔) นายเอกวุฒิ เสนอใจ	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๕๔
๕๕) นายสุขสันต์ บุญเลี้ยง	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๕๕
๕๖) นายธนเดช หวานเสนาะ	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๕๖
๕๗) นายพิพัฒน์ ตันธนกุล	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๕๗
๕๘) นายอภิสิทธิ์ ศรีคงแก้ว	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๕๘
๕๙) นายภูวดล มงคลสูง	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๕๙
๖๐) นายอุทัย แก้วรามุข	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๖๐
๖๑) นางสาววนารินทร์ สานนท์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๖๑
๖๒) นายศุภกร รินวงศ์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๖๒
๖๓) นายศักดิ์สิทธิ์ เกิดขัง	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๖๓
๖๔) นางสาวศิริพร อภิการัตน์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๖๔
๖๕) นางสาวจินตสุภา เปลี่ยนศรี	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๖๕
๖๖) นางสาวเนตรนภา กมลบูรณ์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๖๖
๖๗) นางสาวอารียา ทารมย์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๖๗
๖๘) นายจิรวัฒน์ สุขเกษม	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๖๘
๖๙) นายกิตติพงษ์ สอนชัยภูมิ	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๖๙
๗๐) นายจุมพล สวนเพชร	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๗๐
๗๑) นางสาวพัชรภรณ์ แสงฟ้า	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๗๑
๗๒) นายรัตนชัย เหล่ามา	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๗๒



๗๓) นายอิทธิพงษ์...

๗๓) นายอิทธิพงษ์ ศรีวิเศษ	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๗๓
๗๔) นางสาวกรรณิการ์ สำลีทา	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๗๔
๗๕) นายธราปกรณ์ พิมพ์ศรี	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๗๕
๗๖) นายพรชัย คุ่มม่วง	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๗๖
๗๗) นางสาวทัศนีย์ ไชยหาร	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๗๗
๗๘) นายธีรพงษ์ ศรีคำแหง	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๗๘
๗๙) นางสาวณัฐชา พรหมศิริ	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๗๙
๘๐) นางสาวลัดดาวัลย์ โพธิ์พันธ์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๘๐
๘๑) นางสาวกมลวรรณ เจริญจันทร์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๘๑
๘๒) นายนพรัตน์ จันทะคุณ	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๘๒
๘๓) นายปิยวัฒน์ ไหมชู	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๘๓
๘๔) นางสาวพรนัชชา กลิ่นฉุน	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๘๔
๘๕) นายนภสัทธิ ศรีพิมพ์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๘๕
๘๖) นางสาวลักขิกา จันทรสุข	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๘๖
๘๗) นายสงกรานต์ มาลัยทอง	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๘๗
๘๘) นางสาวสาธิตา แซ่เตียว	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๘๘
๘๙) นายศักดิ์ศิรินทร์ นุ่มนัม	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๘๙
๙๐) นายวรพงษ์ นนทจันทร์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๙๐
๙๑) นางสาวชนาภา มาคะมาตร	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๙๑
๙๒) นางสาวธนธรณ์ คุณานุพันธ์ชัย	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๙๒
๙๓) นายวีระยุทธ สาระภักดี	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๙๓
๙๔) นางสาวธิดิยา วีระพันธุ์วัฒน์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๙๔
๙๕) นายกฤตพล พงศ์สถาพร	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๙๕
๙๖) นายณัฐชัย พรหมอารักษ์	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๙๖
๙๗) นายชนินทร์ พานแก้ว	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๙๗
๙๘) นายปรัชชาพล โสภา	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๙๘
๙๙) นายวัชรินทร์ แสนงาม	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๐๙๙
๑๐๐) นางสาวธนภรณ์ ลาพรม	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๑๐๐
๑๐๑) นายอาทิตย์ อุดมผล	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๑๐๑
๑๐๒) นายปรวร บุณนาค	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๑๐๒
๑๐๓) นายอิทธิเดช ใจบุญ	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๑๐๓
๑๐๔) นายคณิติน พงษ์อิศรานุพร	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๑๐๔
๑๐๕) นางสาวสุภารัตน์ จันทร์ประທတ်	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๑๐๕
๑๐๖) นายเสกฐวุฒิ เอมกลิ่นบัว	ทะเบียนเลขที่ ว-๑๔๕-จ-๐๑๐๖

เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท ยูไนเต็ด แอนาไลสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด เลขทะเบียน ว-๑๔๕

ที่ ออ ๐๓๑๐(๑)/ ๑๘๗๙ ลงวันที่ ๐๙ กุมภาพันธ์ ๒๕๖๕

ขอขยายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๓๕๗ รายการ

น้ำเสีย จำนวน 46 รายการ

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
1	Aldrin	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
2	Arsenic	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method <sup>[4]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
3	Barium	Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
4	α-BHC	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
5	β-BHC	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
6	δ-BHC	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
7	γ-BHC	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
8	Biochemical Oxygen Demand	1) 5-Day BOD Test, Azide Modification Method <sup>[4]</sup> 2) 5-Day BOD Test, Membrane Electrode Method <sup>[4]</sup>
9	Cadmium	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method <sup>[4]</sup> 3) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
10	Chemical Oxygen Demand	1) Closed Reflux, Titrimetric Method <sup>[4]</sup> 2) Closed Reflux, Colorimetric Method <sup>[4]</sup> 3) Open Reflux, Titrimetric Method <sup>[4]</sup>
11	Chlordane	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
12	Chromium	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method <sup>[4]</sup> 3) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
13	Color	ADMI Weighted-Ordinate Spectrophotometric Method <sup>[4]</sup>
14	Copper	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method <sup>[4]</sup> 3) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
15	Cyanide	1) Distillation, Colorimetric Method <sup>[4]</sup> 2) Flow Injection Analysis Method <sup>[4]</sup>

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
16	o,p'-DDT	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
17	4,4'-DDD	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
18	4,4'-DDE	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
19	4,4'-DDT	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
20	Dieldrin	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
21	Endosulfan I	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
22	Endosulfan II	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
23	Endosulfan sulfate	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
24	Endrin	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
25	Endrin aldehyde	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
26	Formaldehyde	Distillation, Colorimetric Method <sup>[3]</sup>
27	Free Chlorine	1) Iodometric Method <sup>[4]</sup> 2) DPD Ferrous Titrimetric Method <sup>[4]</sup>
28	Heptachlor	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
29	Heptachlor Epoxide	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
30	Hexavalent Chromium	1) Colorimetric Method <sup>[4]</sup> 2) Extraction, Direct Air-Acetylene Flame Method <sup>[4]</sup>
31	Lead	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method <sup>[4]</sup> 3) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
32	Manganese	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method <sup>[4]</sup> 3) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
33	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method <sup>[4]</sup>
34	Methoxychlor	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
35	Nickel	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method <sup>[4]</sup> 3) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
36	Oil & Grease	1) Liquid-Liquid, Partition-Gravimetric Method <sup>[4]</sup> 2) Soxhlet Extraction Method <sup>[4]</sup>
37	pH	Electrometric Method <sup>[4]</sup>
38	Phenols	1) Distillation, Chloroform Extraction Method <sup>[4]</sup> 2) Distillation, Direct Photometric Method <sup>[4]</sup>
39	Selenium	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method <sup>[4]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
40	Sulfide	1) Iodometric Method <sup>[4]</sup> 2) Methylene Blue Method <sup>[4]</sup>
41	Temperature	Laboratory and Field Methods <sup>[4]</sup>
42	Total Dissolved Solids	Dried at 180 °C <sup>[4]</sup>
43	Total Kjeldahl Nitrogen	Semi-Micro-Kjeldahl Method <sup>[4]</sup>
44	Total Suspended Solids	Dried at 103-105 °C <sup>[4]</sup>
45	Trivalent Chromium	1) Digestion, Direct Air-Acetylene Flame Method; Colorimetric Method; Calculation <sup>[4]</sup> 2) Digestion, Inductively Coupled Plasma Method; Colorimetric Method; Calculation <sup>[4]</sup>
46	Zinc	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method <sup>[4]</sup> 3) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>

น้ำใต้ดิน จำนวน 126 รายการ

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
1	Acenaphthene	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
2	Acetone	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
3	Aldrin	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
4	Anthracene	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
5	Antimony	Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
6	Arsenic	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method <sup>[4]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
7	Atrazine	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
8	Barium	1) Digestion, Electrothermal Atomic Absorption Spectrometric Method <sup>[4]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
9	Benz(a)anthracene	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
10	Benzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
11	Benzo(b)fluoranthene	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
12	Benzo(k)fluoranthene	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
13	Benzoic acid	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
14	Benzo(a)pyrene	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>

UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำนักงานผู้ตรวจการ  
แผ่นดิน

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
15	Benzo(g,h,i)perylene	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
16	Beryllium	Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
17	Bis(2-chloroethyl)ether	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
18	Bis(2-ethylhexyl)phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
19	Bromodichloromethane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
20	Bromoform	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
21	Butanol	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
22	Butyl benzyl phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
23	Cadmium	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method <sup>[4]</sup> 3) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
24	Carbazole	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
25	Carbon disulfide	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
26	Carbon tetrachloride	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
27	Chlordane	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
28	p-Chloroaniline	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
29	Chlorobenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
30	Chlorodibromomethane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
31	Chloroform	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
32	2-Chlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
33	Chromium	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method <sup>[4]</sup> 3) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
34	Chromium (III)	1) Digestion, Direct Air-Acetylene Flame Method; Colorimetric Method; Calculation <sup>[4]</sup> 2) Digestion, Inductively Coupled Plasma Method; Colorimetric Method; Calculation <sup>[4]</sup>
35	Chromium (VI)	1) Colorimetric Method <sup>[4]</sup> 2) Extraction, Air-Acetylene Flame Method <sup>[4]</sup>
36	Chrysene	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
37	Cyanide	Distillation, Colorimetric Method <sup>[4]</sup>
38	2,4-D	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
39	DDD	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
40	DDE	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
41	DDT	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>

UAE  
UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

ตามถูกต้อง

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
42	Dibenz(a,h)anthracene	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
43	Di-n-butyl phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
44	1,2-Dichlorobenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
45	1,3-Dichlorobenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
46	1,4-Dichlorobenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
47	3,3'-Dichlorobenzidine	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
48	1,1-Dichloroethane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
49	1,2-Dichloroethane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
50	1,1-Dichloroethylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
51	cis-1,2-Dichloroethylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
52	trans-1,2-Dichloroethylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
53	2,4-Dichlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
54	1,2-Dichloropropane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
55	1,3-Dichloropropane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
56	1,3-Dichloropropene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
57	Dieldrin	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
58	Diethyl phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
59	2,4-Dimethylphenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
60	2,4-Dinitrophenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
61	2,4-Dinitrotoluene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
62	2,6-Dinitrotoluene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
63	Di-n-Octyl phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
64	Endosulfan	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
65	Endrin	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
66	Ethylbenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
67	Fluoranthene	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
68	Fluorene	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
69	Heptachlor	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
70	Heptachlor epoxide	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
71	Hexachlorobenzene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
72	Hexachloro-1,3-butadiene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
73	n-Hexane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
74	$\alpha$ -HCH	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
75	$\beta$ -HCH	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
76	$\gamma$ -HCH	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
77	Hexachlorocyclopentadiene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
78	Hexachloroethane	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
79	Indeno(1,2,3-cd)pyrene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
80	Isophorone	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
81	Lead	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method <sup>[4]</sup> 3) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>

CONSULTANT COMPANY LIMITED

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
82	Manganese	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method <sup>[4]</sup> 3) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
83	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method <sup>[4]</sup>
84	Methanol	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
85	Methoxychlor	Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup>
86	Methyl bromide	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
87	Methylene chloride	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
88	2-Methylphenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
89	2-Methylnaphthalene	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
90	Methyl tert-butyl ether	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
91	Naphthalene	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
92	Nickel	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method <sup>[4]</sup> 3) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
93	Nitrobenzene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
94	N-Nitrosodiphenylamine	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
95	N-Nitrosodi-n-propylamine	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
96	Polychlorinated Biphenyls - PCB 1016 - PCB 1221 - PCB 1232 - PCB-1242 - PCB-1248 - PCB-1254 - PCB-1260	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
97	Pentachlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
98	pH	Electrometric Method <sup>[4]</sup>
99	Phenanthrene	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
100	Phenol	1) Distillation, Chloroform Extraction Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
101	Pyrene	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
102	Selenium	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method <sup>[4]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
103	Silver	Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
104	Styrene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
105	1,1,2,2-Tetrachloroethane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
106	Tetrachloroethylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
107	Toluene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>

WAE  
UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำเนาถูกต้อง

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
108	Toxaphene	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
109	TPH (C <sub>5</sub> - C <sub>8</sub> )	1) Purge and Trap, Gas Chromatographic Method <sup>[11,21]</sup> 2) Purge and Trap, Gas Chromatographic/Mass spectrometric Method <sup>[11,25]</sup>
110	TPH (C <sub>8</sub> - C <sub>16</sub> )	Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[9,21]</sup>
111	TPH (C <sub>16</sub> - C <sub>35</sub> )	Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[9,21]</sup>
112	1,2,4-Trichlorobenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
113	1,1,1-Trichloroethane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
114	1,1,2-Trichloroethane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
115	Trichloroethylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
116	2,4,5-Trichlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
117	2,4,6-Trichlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
118	1,3,5-Trimethylbenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
119	Vanadium	Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
120	Vinyl acetate	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
121	Vinyl chloride	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
122	m-Xylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
123	o-Xylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
124	p-Xylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
125	Xylene (Total)	Purge and Trap Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
126	Zinc	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method <sup>[4]</sup> 3) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>

อากาศเสีย (ปล่อยระบาย) จำนวน 25 รายการ

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
1	Antimony	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
2	Arsenic	1) Isokinetic Sampling, Digestion, Hydride Generation/Atomic Absorption Spectrometric Method <sup>[5]</sup> 2) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
3	Cadmium	1) Isokinetic Sampling, Digestion, Direct Air-Acetylene Flame Method <sup>[5]</sup> 2) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
4	Carbon Monoxide	Instrumental Analyzer Method <sup>[5]</sup>
5	Chlorine	Isokinetic Sampling, Ion Chromatographic Method <sup>[5]</sup>
6	Chromium	1) Isokinetic Sampling, Digestion, Direct Air-Acetylene Flame Method <sup>[5]</sup> 2) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
7	Cobalt	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
8	Copper	1) Isokinetic Sampling, Digestion, Direct Air-Acetylene Flame Method <sup>[5]</sup> 2) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
9	Cresol	Absorption Sampling, Gas Chromatographic Method <sup>[5]</sup>

**WAE**  
WATANA ENGINEERING  
CONSULTANT COMPANY LIMITED

**สำเนาถูกต้อง**

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
10	Dioxins/Furans	Isokinetic Sampling <sup>[5]</sup>
11	Hydrogen Chloride	Isokinetic Sampling, Ion Chromatographic Method <sup>[5]</sup>
12	Hydrogen Fluoride	Isokinetic Sampling, Ion Chromatographic Method <sup>[5]</sup>
13	Hydrogen Sulfide	Absorption Sampling, Iodometric Method <sup>[5]</sup>
14	Lead	1) Isokinetic Sampling, Digestion, Direct Air-Acetylene Flame Method <sup>[5]</sup> 2) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
15	Manganese	1) Isokinetic Sampling, Digestion, Direct Air-Acetylene Flame Method <sup>[5]</sup> 2) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
16	Mercury	Isokinetic Sampling, Digestion, Cold-Vapor Atomic Absorption Spectrometric Method <sup>[5]</sup>
17	Nickel	1) Isokinetic Sampling, Digestion, Direct Air-Acetylene Flame Method <sup>[5]</sup> 2) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
18	Opacity	Ringelmann's Method <sup>[1]</sup>
19	Oxides of Nitrogen	1) Absorption Sampling, Phenoldisulfonic acid Method <sup>[5]</sup> 2) Instrumental Analyzer Method <sup>[5]</sup>
20	Selenium	1) Isokinetic Sampling, Digestion, Hydride Generation/Atomic Absorption Spectrometric Method <sup>[5]</sup> 2) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
21	Sulfur Dioxide	1) Absorption Sampling, Barium-Thorin Titrimetric Method <sup>[5]</sup> 2) Instrumental Analyzer Method <sup>[5]</sup>
22	Sulfuric Acid	Isokinetic Sampling, Barium-Thorin Titrimetric Method <sup>[5]</sup>
23	Total Suspended Particulate	Isokinetic Sampling, Gravimetric Method <sup>[5]</sup>
24	Vanadium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
25	Xylene	1) Bag Sampling, Gas Chromatographic Method <sup>[5]</sup> 2) Adsorption Sampling, Gas Chromatographic Method <sup>[5]</sup>




สิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว จำนวน 35 รายการ

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
1	Aldrin	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[2,9,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup>
2	Antimony	Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
3	Arsenic	1) Waste Extraction, Digestion, Hydride Generation/Atomic Absorption Spectrometric Method <sup>[2,6,15]</sup> 2) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,6,13]</sup> 3) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method <sup>[7,15]</sup> 4) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
4	Barium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,6,13]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
5	Beryllium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,6,13]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
6	Cadmium	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method <sup>[2,6,14]</sup> 2) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,6,13]</sup> 3) Digestion, Flame Atomic Absorption Spectrometric Method <sup>[7,14]</sup> 4) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
7	Chlordane	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[2,9,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup>
8	Chromium	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method <sup>[2,6,14]</sup> 2) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,6,13]</sup>

3) Digestion,...

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
9	Chromium (III)	<p>3) Digestion, Flame Atomic Absorption Spectrometric Method<sup>[7,14]</sup></p> <p>4) Digestion, Inductively Coupled Plasma Method<sup>[7,13]</sup></p> <p>1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method; Waste Extraction, Colorimetric Method; Calculation<sup>[2,6,14,16]</sup></p> <p>2) Waste Extraction, Digestion, Inductively Coupled Plasma Method; Waste Extraction, Colorimetric Method; Calculation<sup>[2,6,13,16]</sup></p> <p>3) Digestion, Flame Atomic Absorption Spectrometric Method; Alkaline Digestion, Colorimetric Method; Calculation<sup>[7,8,14,16]</sup></p> <p>4) Digestion, Inductively Coupled Plasma Method; Alkaline Digestion, Colorimetric Method; Calculation<sup>[7,8,13,16]</sup></p>
10	Chromium (VI)	<p>1) Waste Extraction, Colorimetric Method<sup>[2,16]</sup></p> <p>2) Alkaline Digestion, Colorimetric Method<sup>[8,16]</sup></p>
11	Cobalt	<p>1) Waste Extraction, Digestion, Inductively Coupled Plasma Method<sup>[2,6,13]</sup></p> <p>2) Digestion, Inductively Coupled Plasma Method<sup>[7,13]</sup></p>
12	Copper	<p>1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method<sup>[2,6,14]</sup></p> <p>2) Waste Extraction, Digestion, Inductively Coupled Plasma Method<sup>[2,6,13]</sup></p> <p>3) Digestion, Flame Atomic Absorption Spectrometric Method<sup>[7,14]</sup></p> <p>4) Digestion, Inductively Coupled Plasma Method<sup>[7,13]</sup></p>
13	2,4-D	<p>1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method<sup>[2,9,22]</sup></p> <p>2) Ultrasonic Extraction, Gas Chromatographic Method<sup>[10,22]</sup></p>
14	DDD	<p>1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method<sup>[2,9,22]</sup></p> <p>2) Ultrasonic Extraction, Gas Chromatographic Method<sup>[10,22]</sup></p>

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
15	DDE	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[2,9,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup>
16	DDT	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[2,9,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup>
17	Dieldrin	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[2,9,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup>
18	Endrin	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[2,9,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup>
19	Heptachlor	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[2,9,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup>
20	Lead	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method <sup>[2,6,14]</sup> 2) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,6,13]</sup> 3) Digestion, Flame Atomic Absorption Spectrometric Method <sup>[7,14]</sup> 4) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
21	Lindane	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[2,9,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup>
22	Mercury	1) Waste Extraction, Digestion, Cold-Vapor Atomic Absorption Spectrometric Method <sup>[2,17]</sup> 2) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,6,13]</sup>

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
23	Methoxychlor	3) Digestion, Cold-Vapor Atomic Absorption Spectrometric Method <sup>[18]</sup> 4) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup> 5) Thermal Decomposition Amalgamation and Atomic Absorption Spectrometric Method <sup>[19]</sup> 1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[2,9,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup>
24	Molybdenum	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,6,13]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
25	Nickel	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method <sup>[2,6,14]</sup> 2) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,6,13]</sup> 3) Digestion, Flame Atomic Absorption Spectrometric Method <sup>[7,14]</sup> 4) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
26	Polychlorinated Biphenyls - Aroclor 1016 - Aroclor 1221 - Aroclor 1232 - Aroclor 1242 - Aroclor 1248 - Aroclor 1254 - Aroclor 1260 - 2-Chlorobiphenyl - 2,3-Dichlorobiphenyl - 2,2',5-Trichlorobiphenyl - 2,4',5-Trichlorobiphenyl - 2,2',3,5'-Tetrachlorobiphenyl - 2,2',5,5'-Tetrachlorobiphenyl - 2,3',4,4'-Tetrachlorobiphenyl - 2,2',3,4,5'-Pentachlorobiphenyl	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[2,9,23]</sup> 2) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,23]</sup>    สำเนาถูกต้อง

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
27	- 2,2',4,5,5'- Pentachlorobiphenyl - 2,3,3',4',6- Pentachlorobiphenyl - 2,2',3,4,4',5'- Hexachlorobiphenyl - 2,2',3,4,5,5'- Hexachlorobiphenyl - 2,2',3,5,5',6- Hexachlorobiphenyl - 2,2',4,4',5,5'- Hexachlorobiphenyl - 2,2',3,3',4,4',5- Heptachlorobiphenyl - 2,2',3,4,4',5,5'- Heptachlorobiphenyl - 2,2',3,4,4',5,6- Heptachlorobiphenyl - 2,2',3,4',5,5',6- Heptachlorobiphenyl - 2,2',3,3',4,4',5,5',6- Nonachlorobiphenyl Pentachlorophenol	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[2,9,28]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup> Electrometric Method <sup>[31,32]</sup> 1) Waste Extraction, Digestion, Hydride Generation/Atomic Absorption Spectrometric Method <sup>[2,6,20]</sup> 2) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,6,13]</sup> 3) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method <sup>[7,13]</sup> 4) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
28	pH	
29	Selenium	

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
30	Silver	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,6,13]</sup>
31	Thallium	2) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup> 1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,6,13]</sup>
32	Toxaphene	2) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup> 1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[2,9,22]</sup>
33	Trichloroethylene	2) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup> 1) Waste Extraction, Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[2,12,25]</sup>
34	Vanadium	2) Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup> 1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,6,13]</sup>
35	Zinc	2) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup> 1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method <sup>[2,6,14]</sup> 2) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,6,13]</sup> 3) Digestion, Flame Atomic Absorption Spectrometric Method <sup>[7,14]</sup> 4) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>

ดิน จำนวน 125 รายการ

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
1	Acenaphthene	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,24]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
2	Acetone	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>

UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำเนาถูกต้อง

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
3	Aldrin	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
4	Anthracene	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,26]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
5	Antimony	Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
6	Arsenic	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method <sup>[7,15]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
7	Atrazine	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
8	Barium	Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
9	Benz(a)anthracene	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,24]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
10	Benzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
11	Benzo(b)fluoranthene	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,24]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
12	Benzo(k)fluoranthene	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,24]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
13	Benzoic acid	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
14	Benzo(a)pyrene	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,24]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
15	Benzo(g,h,i)perylene	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
16	Beryllium	Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
17	Bis(2-chloroethyl)ether	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
18	Bis(2-ethylhexyl)phthalate	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
19	Bromodichloromethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
20	Bromoform	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
21	Butanol	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
22	Butyl benzyl phthalate	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
23	Cadmium	1) Digestion, Flame Atomic Absorption Spectrometric Method <sup>[7,14]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
24	Carbazole	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
25	Carbon disulfide	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
26	Carbon tetrachloride	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
27	Chlordane	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
28	p-Chloroaniline	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
29	Chlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
30	Chlorodibromomethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
31	Chloroform	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
32	2-Chlorophenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
33	Chromium	1) Digestion, Flame Atomic Absorption Spectrometric Method <sup>[7,14]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
34	Chromium (III)	1) Digestion, Flame Atomic Absorption Spectrometric Method; Alkaline Digestion, Colorimetric Method; Calculation <sup>[7,8,14,16]</sup> 2) Digestion, Inductively Coupled Plasma Method; Alkaline Digestion, Colorimetric Method; Calculation <sup>[7,8,13,16]</sup>
35	Chromium (VI)	Alkaline Digestion, Colorimetric Method <sup>[8,16]</sup>
36	Chrysene	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,24]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
37	Cyanide	Extraction, Distillation, Colorimetric Method <sup>[28,29,30]</sup>
38	2,4-D	Ultrasonic Extraction, Gas Chromatographic Method <sup>[27]</sup>
39	DDD	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
40	DDE	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
41	DDT	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
42	Dibenz(a,h)anthracene	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,24]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
43	Di-n-butyl phthalate	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
44	1,2-Dichlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
45	1,3-Dichlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
46	1,4-Dichlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
47	3,3'-Dichlorobenzidine	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
48	1,1-Dichloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
49	1,2-Dichloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
50	1,1-Dichloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
51	cis-1,2-Dichloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
52	trans-1,2-Dichloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
53	2,4-Dichlorophenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
54	1,2-Dichloropropane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
55	1,3-Dichloropropane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
56	1,3-Dichloropropene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
57	Dieldrin	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
58	Diethyl phthalate	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
59	2,4-Dimethylphenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>

WAE  
WATANA ENGINEERING  
CONSULTANT COMPANY LIMITED

ดำเนินการ  
วันที่ ๒๕/๑๐/๖๕

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
60	2,4-Dinitrophenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
61	2,4-Dinitrotoluene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
62	2,6-Dinitrotoluene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
63	Di-n-Octyl phthalate	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
64	Endosulfan	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
65	Endrin	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
66	Ethylbenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
67	Fluoranthene	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,24]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
68	Fluorene	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,24]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
69	Heptachlor	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
70	Heptachlor epoxide	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
71	Hexachlorobenzene	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
72	Hexachloro-1,3-butadiene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
73	n-Hexane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
74	$\alpha$ -HCH	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
75	$\beta$ -HCH	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
76	$\gamma$ -HCH	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
77	Hexachlorocyclopentadiene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
78	Hexachloroethane	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
79	Indeno(1,2,3-cd)pyrene	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,24]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
80	Isophorone	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
81	Lead	1) Digestion, Flame Atomic Absorption Spectrometric Method <sup>[7,14]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
82	Manganese	1) Digestion, Flame Atomic Absorption Spectrometric Method <sup>[7,14]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
83	Mercury	1) Digestion, Cold-Vapor Atomic Absorption Spectrometric Method <sup>[18]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup> 3) Thermal Decomposition Amalgamation and Atomic Absorption Spectrometric Method <sup>[19]</sup>
84	Methanol	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
85	Methoxychlor	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
86	Methyl bromide	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
87	Methylene chloride	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
88	2-Methylphenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
89	2-Methylnaphthalene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
90	Methyl tert-butyl ether	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
91	Naphthalene	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,24]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
92	Nickel	1) Digestion, Flame Atomic Absorption Spectrometric Method <sup>[7,14]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
93	Nitrobenzene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
94	N-Nitrosodiphenylamine	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
95	N-Nitrosodi-n-propylamine	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>

UNAE  
UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

ลงนามถูกต้อง



**LIAE**  
UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำเนาถูกต้อง

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
97	- 2,2',3,4',5,5',6- Heptachlorobiphenyl - 2,2',3,3',4,4',5,5',6- Nonachlorobiphenyl Pentachlorophenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
98	Phenanthrene	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,24]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
99	Phenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
100	Pyrene	1) Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,24]</sup> 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
101	Selenium	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method <sup>[7,22]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
102	Silver	Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
103	Styrene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
104	1,1,2,2-Tetrachloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
105	Tetrachloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
106	Toluene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
107	Toxaphene	Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,22]</sup>
108	TPH (C <sub>5</sub> -C <sub>8</sub> )	1) Purge and Trap, Gas Chromatographic Method <sup>[12,21]</sup> 2) Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
109	TPH (C <sub>8</sub> -C <sub>16</sub> )	Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,21]</sup>
110	TPH (C <sub>16</sub> -C <sub>35</sub> )	Ultrasonic Extraction, Gas Chromatographic Method <sup>[10,21]</sup>
111	1,2,4-Trichlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>

ลำดับ	สารมลพิษ	วิธีวิเคราะห์
112	1,1,1-Trichloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
113	1,1,2-Trichloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
114	Trichloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
115	2,4,5-Trichlorophenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
116	2,4,6-Trichlorophenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[10,26]</sup>
117	1,3,5-Trimethylbenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
118	Vanadium	Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>
119	Vinyl acetate	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
120	Vinyl chloride	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
121	m-Xylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
122	o-Xylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
123	p-Xylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
124	Xylene (Total)	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[12,25]</sup>
125	Zinc	1) Digestion, Flame Atomic Absorption Spectrometric Method <sup>[7,14]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[7,13]</sup>

#### เอกสารอ้างอิง

- กระทรวงอุตสาหกรรม. ประกาศกระทรวงอุตสาหกรรม, พ.ศ. 2549. เรื่อง กำหนดค่าปริมาณเขม่าควันที่เจือปนในอากาศที่ระบายออกจากปล่องของหม้อไอน้ำโรงสีข้าวที่ใช้ถ่านเป็นเชื้อเพลิง. ราชกิจจานุเบกษา. 4 ธันวาคม 2549. เล่มที่ 123 ตอนพิเศษ 125 ง.
- กระทรวงอุตสาหกรรม. ประกาศกระทรวงอุตสาหกรรม, พ.ศ. 2548. เรื่อง การกำจัดสิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว. ราชกิจจานุเบกษา. 25 มกราคม 2549. เล่มที่ 123 ตอนพิเศษ 11 ง.



3. สมาคมวิศวกรรมสิ่งแวดล้อมแห่งประเทศไทย. คู่มือวิเคราะห์น้ำเสีย. พิมพ์ครั้งที่ 4. กรุงเทพฯ: เรือนแก้วการพิมพ์, 2547.

4. APHA, AWWA, WEF. **Standard Methods for the Examination of Water and Wastewater**. 23<sup>rd</sup> ed. Washington, DC: APHA, 2017.

5. United States Environmental Protection Agency. **Standards of Performance for New Stationary Sources**. 40 CFR 60. Appendix A, 2019.

6. United States Environmental Protection Agency. **Test Methods for Evaluation Solid Waste Physical/Chemical Methods**. SW-846, 1997.

7. United States Environmental Protection Agency. **Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Acid Digestion of Sediments, Sludges, and Soils**. SW-846 Method 3050B, 1996.

8. United States Environmental Protection Agency. **Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Alkaline Digestion for Hexavalent Chromium**. SW-846 Method 3060A, 1996.

9. United States Environmental Protection Agency. **Test Methods for Evaluation Solid Waste 3. Physical/Chemical Methods. Separatory Funnel Liquid-Liquid Extraction**. SW-846 Method 3510C, 1996.

10. United States Environmental Protection Agency. **Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Ultrasonic Extraction**. SW-846 Method 3550C, 2007.


11. United States Environmental Protection Agency. **Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Purge and Trap for Aqueous Samples**. SW-846 Method 5030C, 2003.

12. United States Environmental Protection Agency. **Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Closed System Purge and Trap and Extraction for Volatile Organics in Soil and Waste Sample**. SW-846 Method 5035A, 2000.

13. United States Environmental Protection Agency. **Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Inductively Coupled Plasma-Optical Emission Spectrometry**. SW-846 Method 6010D, 2014.

14. United States Environmental Protection Agency. **Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Flame Atomic Absorption Spectrophotometry**. SW-846 Method 7000B, 2007.

15. United States Environmental Protection Agency. **Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Arsenic (Atomic Absorption, Gaseous Hydride)**. SW-846 Method 7061A, 1992.



สำนักมาตรฐาน

16. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Chromium, Hexavalent (Colorimetric). SW-846 Method 7196A, 1992.

17. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Mercury in Liquid Waste (Manual Cold Vapor Technique). SW-846 Method 7470A, 1994.

18. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Mercury in Solid or Semisolid Waste (Manual Cold-Vapor Technique). SW-846 Method 7471B, 1998.

19. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Mercury in Solids and Solutions by Thermal Decomposition, Amalgamation, and Atomic Absorption Spectrophotometry. SW-846 Method 7473, 2007.

20. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Selenium (Atomic Absorption, Borohydride Reduction). SW-846 Method 7742, 1994.

21. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Nonhalogenated Organics Using GC/FID. SW-846 Method 8015D, 2003.

22. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Organochlorine Pesticides by Gas Chromatography. SW-846 Method 8081B, 2007.


23. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Polychlorinated Biphenyls (PCBs) by Gas Chromatography. SW-846 Method 8082A, 2007.

24. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Polynuclear Aromatic Hydrocarbons. SW-846 Method 8100, 1980.

25. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Volatile Organic Compounds by Gas Chromatography/Mass Spectrometry. SW-846 Method 8260D, 2018.

26. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Semivolatile Organic Compounds by Gas Chromatography/Mass Spectrometry. SW-846 Method 8270E, 2018.

27. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Chlorinated Herbicides by GC Using Methylation or Pentafluorobenzoylation Derivatization. SW-846 Method 8151A, 1996.



28. United States...

28. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Total and Amenable Cyanide : Distillation. SW-846 Method 9010C, 2004.

29. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Cyanide Extraction Procedure for Solids and Oils. SW-846 Method 9013A, 2014.

30. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Cyanide in Waters and Extracts using Titrimetric and Manual Spectrophotometric Procedures. SW-846 Method 9014, 2014.

31. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. pH Electrometric Measurement. SW-846 Method 9040C, 2004.

32. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Soil and Waste pH. SW-846 Method 9045D, 2004. *สมชาย*



*สมชาย*  
ดำเนินถูกต้อง

สำเนาใบรับรองห้องปฏิบัติการทดสอบ ISO-IEC 17025-2017\_สมอ.



ใบรับรองเลขที่ 21-LB0022  
(Certificate No.)

## ใบรับรองระบบงาน (Certificate of Accreditation)

อาศัยอำนาจตามความในพระราชบัญญัติการมาตรฐานแห่งชาติ พ.ศ. ๒๕๕๑  
(By Virtue of National Standardization Act B.E. 2551 (2008))

เลขาธิการสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม  
(Secretary-General, Thai Industrial Standards Institute)

ออกใบรับรองฉบับนี้ให้  
(Issues this certificate to)

บริษัท ยูไนเต็ด แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด  
(United Analyst and Engineering Consultant Company Limited)

ตั้งอยู่เลขที่  
(Address)

๓ ซอยอุดมสุข ๔๑ ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง กรุงเทพมหานคร  
(3, Soi Udomsuk 41, Sukhumvit Road, Bangchak, Phrakhanong, Bangkok)

ได้รับการรับรองความสามารถ  
(Certificate of competence)

ตามมาตรฐานเลขที่ มอก. ๑๗๐๒๕ - ๒๕๖๑  
(Standard No. TIS 17025-2561 (2018) (ISO/IEC 17025: 2017))

ข้อกำหนดทั่วไปว่าด้วยความสามารถของ ห้องปฏิบัติการทดสอบและห้องปฏิบัติการสอบเทียบ  
(General requirements for the competence of testing and calibration laboratories)

หมายเลขการรับรองที่ ทดสอบ ๐๒๐๗  
(Accreditation No. Testing 0207)

โดยมีรายละเอียดสาขาและขอบข่ายที่ได้ใบรับรอง แสดงไว้ใน QR CODE และ [www.tisi.go.th](http://www.tisi.go.th)  
(Details of the scheme and scope of the certificate are shown in QR CODE and [www.tisi.go.th](http://www.tisi.go.th))

ออกให้ ณ วันที่ ๑๑ ตุลาคม พ.ศ. ๒๕๖๔  
(Issue date : 11 October B.E. 2564 (2021))



United Analyst and Engineering  
Consultant Company Limited

สำเนาถูกต้อง



รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ชื่อห้องปฏิบัติการ  
(Laboratory Name)

หมายเลขการรับรองที่  
(Accreditation No.)

ฉบับที่ 04  
(Issue No. 04)

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

ห้องปฏิบัติการทดสอบ บริษัท ยูไนटेด แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด  
(United Analyst and Engineering Consultant Company Limited)

ทดสอบ 0207  
(Testing 0207)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565  
(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566  
(Until) (17 May B.E.2566(2023))

☒ ถาวร  
(Permanent)

☐ นอกสถานที่  
(Site)

☐ชั่วคราว  
(Temporary)

☐เคลื่อนที่  
(Mobile)

☐หลายสถานที่  
(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาส่งแวดล้อม (Environmental field)</p> <p>1. น้ำ (water)</p> <p>- น้ำผิวดิน (surface water)</p> <p>- น้ำใต้ดิน (ground water)</p>	<p>- Heavy metals</p> <ul style="list-style-type: none"> <li>• Copper 0.025 mg/l to 20.0 mg/l</li> <li>• Nickel 0.050 mg/l to 20.0 mg/l</li> <li>• Zinc 0.025 mg/l to 20.0 mg/l</li> <li>• Chromium 0.050 mg/l to 20.0 mg/l</li> <li>• Cadmium 0.010 mg/l to 20.0 mg/l</li> <li>• Lead 0.100 mg/l to 20.0 mg/l</li> <li>• Manganese 0.025 mg/l to 20.0 mg/l</li> <li>• Iron 0.050 mg/l to 20.0mg/l</li> </ul>	<p>- UAE.TP.HEM.005, UAE.TP.HEM.003 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23<sup>rd</sup> edition, 2017, part 3030 E and part 3111 B</p>

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)



หน้า 1/27

สำเนาถูกต้อง

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

☒ ถาวร

(Permanent)

☐ นอกสถานที่

(Site)

☐ ชั่วคราว

(Temporary)

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

☐ เคลื่อนที่

(Mobile)

☐ หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสังแวดล้อม (Environmental field)</p> <p>1. น้ำ(ต่อ) (water) (cont.)</p> <p>- น้ำผิวดิน (surface water)</p> <p>- น้ำใต้ดิน (ground water)</p> <p>- น้ำผิวดิน (surface water)</p>	<p>- Chloride 2.0 mg/l to 1 000 mg/l</p> <p>- Total hardness 4.0 mg/l to 1 000 mg/l</p> <p>- Total suspended solids 5.0 mg/l to 500 mg/l</p>	<p>- Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23<sup>rd</sup> edition, 2017, part 4500-Cl<sup>-</sup> B</p> <p>- Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23<sup>rd</sup> edition, 2017, part 2340 C</p> <p>- Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23<sup>rd</sup> edition, 2017, part 2540 D</p>



กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม  
(Ministry of Industry, Thai Industrial Standards Institute)

หน้า 2/27

สำเนาถูกต้อง

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☒ ถาวร

(Permanent)

☐ นอกสถานที่

(Site)

☐ชั่วคราว

(Temporary)

☐เคลื่อนที่

(Mobile)

☐หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสิ่งแวดล้อม (Environmental field)</p> <p>1. น้ำ (ต่อ) (water) (cont.)</p> <p>- น้ำใต้ดิน (ground water)</p>	<p>- Volatile organic compounds (VOCs)</p> <ul style="list-style-type: none"> <li>• Benzene 0.20 µg/l to 1 000 µg/l</li> <li>• Carbon Tetrachloride 0.20 µg/l to 1 000 µg/l</li> <li>• 1,2-Dichloroethane 0.20 µg/l to 1 000 µg/l</li> <li>• 1,1-Dichloroethylene (1,1-Dichloroethene) 0.20 µg/l to 1 000 µg/l</li> <li>• cis-1,2-Dichloroethylene (cis-1,2-Dichloroethene) 0.20 µg/l to 1 000 µg/l</li> <li>• trans-1,2-Dichloroethylene (trans-1,2-Dichloroethene) 0.20 µg/l to 1 000 µg/l</li> <li>• Dichloromethane (Methylene Chloride) 0.20 µg/l to 1 000 µg/l</li> <li>• Ethylbenzene 0.20 µg/l to 1 000 µg/l</li> </ul>	<p>- Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23<sup>rd</sup> edition, 2017, part 6200 B</p> <p><i>Handwritten signature</i></p>

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)



*Handwritten signature*  
สำเนาถูกต้อง

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☒ ถาวร

(Permanent)

☐ นอกสถานที่

(Site)

☐ชั่วคราว

(Temporary)

☐เคลื่อนที่

(Mobile)

☐หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสังแวดล้อม (Environmental field)</p> <p>1.น้ำ (ต่อ) (water) (cont.)</p> <p>- น้ำใต้ดิน (ground water)</p>	<p>- Volatile organic compounds (VOCs) (cont.)</p> <ul style="list-style-type: none"> <li>Styrene 0.20 µg/l to 1000 µg/l</li> <li>Tetrachloroethylene (Tetrachloroethene) 0.20 µg/l to 1 000 µg/l</li> <li>Toluene 0.20 µg/l to 1 000 µg/l</li> <li>Trichloroethylene (Trichloroethene) 0.20 µg/l to 1 000 µg/l</li> <li>1,1,1-Trichloroethane 0.20 µg/l to 1 000 µg/l</li> <li>1,1,2-Trichloroethane 0.20 µg/l to 1 000 µg/l</li> <li>Total Xylenes(o,m,p-Xylene) (Xylene (total)) 0.60 µg/l to 3000 µg/l</li> </ul>	<p>- Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23<sup>rd</sup> edition, 2017, part 6200 B</p>



UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)

สำเนาถูกต้อง

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ

(Laboratory status)

☒ ถาวร

(Permanent)

☐ นอกสถานที่

(Site)

☐ชั่วคราว

(Temporary)

☐เคลื่อนที่

(Mobile)

☐หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสิ่งแวดล้อม (Environmental field)</p> <p>2. น้ำเสีย (wastewater)</p>	<p>- Heavy metals</p> <ul style="list-style-type: none"> <li>• Copper 0.050 mg/l to 50.0 mg/l</li> <li>• Nickel 0.100 mg/l to 50.0 mg/l</li> <li>• Zinc 0.050 mg/l to 50.0 mg/l</li> <li>• Chromium 0.100 mg/l to 50.0 mg/l</li> <li>• Cadmium 0.020 mg/l to 50.0 mg/l</li> <li>• Lead 0.200 mg/l to 50.0 mg/l</li> <li>• Manganese 0.050 mg/l to 50.0 mg/l</li> <li>• Iron 0.100 mg/l to 50.0 mg/l</li> </ul>	<p>- .UAE.TP.HEM.004 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23<sup>rd</sup> edition, 2017, part 3030 and part 3111 B</p>

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)



ดำเนินการถูกต้อง

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☒ ถาวร

(Permanent)

☐ นอกสถานที่

(Site)

☐ชั่วคราว

(Temporary)

☐เคลื่อนที่

(Mobile)

☐หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสังแวดล้อม (Environmental field)</p> <p>2.น้ำเสีย (ต่อ) (wastewater) (cont.)</p>	<p>- Heavy metals</p> <ul style="list-style-type: none"> <li>• Copper 0.010 mg/l to 50.0 mg/l</li> <li>• Nickel 0.010 mg/l to 50.0 mg/l</li> <li>• Zinc 0.010 mg/l to 50.0 mg/l</li> <li>• Chromium 0.010 mg/l to 50.0 mg/l</li> <li>• Cadmium 0.010 mg/l to 50.0 mg/l</li> <li>• Lead 0.010 mg/l to 50.0 mg/l</li> <li>• Manganese 0.010 mg/l to 50.0 mg/l</li> <li>• Iron 0.010 mg/l to 50.0 mg/l</li> </ul> <p>- Heavy metals</p> <ul style="list-style-type: none"> <li>• Copper 0.010 mg/l to 50.0 mg/l</li> <li>• Cadmium 0.010 mg/l to 50.0 mg/l</li> <li>• Lead 0.010 mg/l to 50.0 mg/l</li> </ul>	<p>- UAE.TP.HEM.008 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23<sup>rd</sup> edition, 2017, part 3030 F and part 3120 B</p> <p>- Standard Methods for the Examination of Water and Wastewater, APHA, 'AWWA, WEF, 23<sup>rd</sup> edition, 2017, part 3030 K and 3120 B</p>

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)



สำเนาถูกต้อง

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☒ ถาวร

(Permanent)

☐ นอกสถานที่

(Site)

☐ ชั่วคราว

(Temporary)

☐ เคลื่อนที่

(Mobile)

☐ หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาส่งแวดล้อม (Environmental field)</p> <p>2. น้ำเสีย (ต่อ) (wastewater) (cont.)</p>	<p>- COD 25.0 mg/l to 20 000 mg/l</p> <p>- COD 40.0 mg/l to 2 000 mg/l</p> <p>- Total suspended solids 5.0 mg/l to 5 000 mg/l</p> <p>- BOD 2.0 mg/l to 10 000 mg/l</p> <p>- Oil and Grease 3 mg/l to 200 mg/l</p>	<p>- Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23<sup>rd</sup> edition, 2017, part 5220 D</p> <p>- Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23<sup>rd</sup> edition, 2017, part 5220 C</p> <p>- Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23<sup>rd</sup> edition, 2017, part 2540 D</p> <p>- Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23<sup>rd</sup> edition, 2017, part 5210 B</p> <p>- Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23<sup>rd</sup> edition, 2017, part 5520B</p>

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)



UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำเนาถูกต้อง

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☒ ถาวร

(Permanent)

☐ นอกสถานที่

(Site)

☐ชั่วคราว

(Temporary)

☐เคลื่อนที่

(Mobile)

☐หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสังแวดล้อม (Environmental field)</p> <p>3. น้ำและน้ำเสีย (water and wastewater)</p> <p>4. น้ำทะเล (seawater)</p>	<p>- pH 2.0 to 12.0</p> <p>- Total mercury 0.020 µg/l to 3.50 µg/l</p> <p>- Total mercury 0.010 µg/l to 0.100 µg/l</p>	<p>- Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23<sup>rd</sup> edition, 2017, part 4500-H<sup>+</sup> B</p> <p>- US EPA Method 245.7, Revision 2.0, February 2005</p> <p>- US EPA Method 1631, Revision E, August 2002</p>



UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำเนาถูกต้อง

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☒ ถาวร

(Permanent)

☐ นอกสถานที่

(Site)

☐ชั่วคราว

(Temporary)

☐เคลื่อนที่

(Mobile)

☐หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสังแวดล้อม (Environmental field)</p> <p>5. กากตะกอน (sludge)</p>	<p>- Heavy metals</p> <ul style="list-style-type: none"> <li>• Barium 5.00 mg/kg to 10000 mg/kg</li> <li>• Cadmium 5.00 mg/kg to 10000 mg/kg</li> <li>• Chromium 5.00 mg/kg to 10000 mg/kg</li> <li>• Cobalt 5.00 mg/kg to 10000 mg/kg</li> <li>• Copper 5.00 mg/kg to 10000 mg/kg</li> <li>• Nickel 5.00 mg/kg to 10000 mg/kg</li> <li>• Lead 5.00 mg/kg to 10000 mg/kg</li> <li>• Zinc 5.00 mg/kg to 10000 mg/kg</li> </ul>	<p>- US EPA Method 3050 B, Revision 2 :1996 and US EPA Method 6010D, Revision 5 : 2018</p> <p><i>Handwritten signature</i></p>

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)

**UAE**

UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

หน้า 9/27

*Handwritten signature*  
**สำเนาถูกต้อง**

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☒ ถาวร

(Permanent)

☐ นอกสถานที่

(Site)

☐ชั่วคราว

(Temporary)

☐เคลื่อนที่

(Mobile)

☐หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสังแวดล้อม (Environmental field)</p> <p>6. บรรยากาศ (ambient)</p>	<p>- Total suspended particulate matter <math>\leq 100 \mu\text{m}</math> <math>2.0 \mu\text{g}/\text{m}^3</math> to <math>750 \mu\text{g}/\text{m}^3</math></p> <p>- Particulate matter <math>\leq 10 \mu\text{m}</math> <math>2.7 \mu\text{g}/\text{m}^3</math> to <math>300 \mu\text{g}/\text{m}^3</math></p> <p>- Volatile organic compounds (VOCs)</p> <ul style="list-style-type: none"> <li>Benzene <math>0.08 \text{ ppbv}</math> to <math>25 \text{ ppbv}</math> <math>(0.26 \mu\text{g}/\text{m}^3 \text{ to } 79.9 \mu\text{g}/\text{m}^3)</math></li> <li>Bromodichloromethane <math>0.08 \text{ ppbv}</math> to <math>25 \text{ ppbv}</math> <math>(0.53 \mu\text{g}/\text{m}^3 \text{ to } 166 \mu\text{g}/\text{m}^3)</math></li> </ul>	<p>- US EPA, Code of Federal Regulations, 40 CFR chapter I-part 50 appendix B, revised as of July 1, 2012 (High-Volume method)</p> <p>- US EPA, Code of Federal Regulations, 40 CFR chapter I-part 50 appendix J, revised as of July 1, 2012 (High-Volume method)</p> <p>- UAE.TP.TOX.003 based on U.S.EPA, Compendium Method TO-15, 2<sup>nd</sup> edition, January 1999</p>



UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำเนาถูกต้อง

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ

(Laboratory status)

☒ถาวร

(Permanent)

☐นอกสถานที่

(Site)

☐ชั่วคราว


(Temporary)

☐เคลื่อนที่

(Mobile)

☐หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสิ่งแวดล้อม (Environmental field)</p> <p>6. บรรยากาศ(ต่อ) (ambient) (cont.)</p>	<p>- Volatile organic compounds (VOCs) (cont.)</p> <ul style="list-style-type: none"> <li>Bromoform 0.08 ppbv to 25 ppbv (0.82 <math>\mu\text{g}/\text{m}^3</math> to 256 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Bromomethane 0.08 ppbv to 25 ppbv (0.31 <math>\mu\text{g}/\text{m}^3</math> to 96.1 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Carbon Disulfide 0.08 ppbv to 25 ppbv (0.25 <math>\mu\text{g}/\text{m}^3</math> to 77.7 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Carbon Tetrachloride 0.08 ppbv to 25 ppbv (0.50 <math>\mu\text{g}/\text{m}^3</math> to 155 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Chlorobenzene 0.08 ppbv to 25 ppbv (0.37 <math>\mu\text{g}/\text{m}^3</math> to 115 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Chloroform 0.08 ppbv to 25 ppbv (0.39 <math>\mu\text{g}/\text{m}^3</math> to 121 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>1,2-Dichlorobenzene 0.08 ppbv to 25 ppbv (0.48 <math>\mu\text{g}/\text{m}^3</math> to 149 <math>\mu\text{g}/\text{m}^3</math>)</li> </ul>	<p>- UAE.TP.TOX.003 based on U.S.EPA, Compendium Method TO-15, 2<sup>nd</sup> edition, January 1999</p> 

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)



UNITED ANALYST AND ENGINEERING

COMPANY LIMITED

หน้า 11/27

สำเนาถูกต้อง

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☒ ถาวร

(Permanent)

☐ นอกสถานที่

(Site)

☐ชั่วคราว


(Temporary)

☐เคลื่อนที่

(Mobile)

☐หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสิ่งแวดล้อม (Environmental field)</p> <p>6. บรรยากาศ(ต่อ) (ambient) (cont.)</p>	<p>- Volatile organic compounds (VOCs) (cont.)</p> <ul style="list-style-type: none"> <li>1,3-Dichlorobenzene 0.08 ppbv to 25 ppbv (0.48 <math>\mu\text{g}/\text{m}^3</math> to 149 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>1,1-Dichloroethane 0.08 ppbv to 25 ppbv (0.32 <math>\mu\text{g}/\text{m}^3</math> to 100 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>1,2-Dichloroethane 0.08 ppbv to 25 ppbv (0.32 <math>\mu\text{g}/\text{m}^3</math> to 100 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>1,2-Dibromoethane 0.08 ppbv to 25 ppbv (0.61 <math>\mu\text{g}/\text{m}^3</math> to 190 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Freon-11 (Trichloro monofluoromethane) 0.08 ppbv to 25 ppbv (0.44 <math>\mu\text{g}/\text{m}^3</math> to 139 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Freon-113 (1,1,2-Trichloro- 1,2,2-Trifluoroethane) 0.08 ppbv to 25 ppbv (0.61 <math>\mu\text{g}/\text{m}^3</math> to 190 <math>\mu\text{g}/\text{m}^3</math>)</li> </ul>	<p>- UAE.TP.TOX.003 based on U.S.EPA, Compendium Method TO-15, 2<sup>nd</sup> edition, January 1999</p> 



UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำเนาถูกต้อง

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ

(Laboratory status)

☒ ถาวร

(Permanent)

☐ นอกสถานที่

(Site)

☐ชั่วคราว

(Temporary)

☐เคลื่อนที่

(Mobile)

☐หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสิ่งแวดล้อม (Environmental field)</p> <p>6. บรรยากาศ(ต่อ) (ambient) (cont.)</p>	<p>- Volatile organic compounds (VOCs) (cont.)</p> <ul style="list-style-type: none"> <li>Freon-114(1,2-Dichloro tetrafluoroethane) 0.08 ppbv to 25 ppbv (0.56 <math>\mu\text{g}/\text{m}^3</math> to 174 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Pentane 0.08 ppbv to 25 ppbv (0.24 <math>\mu\text{g}/\text{m}^3</math> to 73.6 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>1,1,2,2-Tetrachloroethane 0.08 ppbv to 25 ppbv (0.54 <math>\mu\text{g}/\text{m}^3</math> to 170 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Toluene 0.08 ppbv to 25 ppbv (0.30 <math>\mu\text{g}/\text{m}^3</math> to 94.1 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Tetrachloroethylene 0.08 ppbv to 25 ppbv (0.54 <math>\mu\text{g}/\text{m}^3</math> to 168 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Trichloroethylene 0.08 ppbv to 25 ppbv (0.43 <math>\mu\text{g}/\text{m}^3</math> to 133 <math>\mu\text{g}/\text{m}^3</math>)</li> </ul>	<p>- UAE.TP.TOX.003 based on U.S.EPA, Compendium Method TO-15, 2<sup>nd</sup> edition, January 1999</p>

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)

UAE

UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

ดำเนินการถูกต้อง

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☒ถาวร

(Permanent)

☐นอกสถานที่

(Site)

☐ชั่วคราว

(Temporary)

☐เคลื่อนที่

(Mobile)

☐หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสังแวดล้อม (Environmental field)</p> <p>6. บรรยากาศ (ต่อ) (ambient) (cont.)</p>	<p>- Volatile organic compounds (VOCs)(cont.)</p> <ul style="list-style-type: none"> <li>1,1,1-Trichloroethane 0.08 ppbv to 25 ppbv (0.43 <math>\mu\text{g}/\text{m}^3</math> to 135 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Chloromethane 0.08 ppbv to 25 ppbv (0.16 <math>\mu\text{g}/\text{m}^3</math> to 51.1 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Isobutene 0.08 ppbv to 25 ppbv (0.18 <math>\mu\text{g}/\text{m}^3</math> to 57.3 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Vinyl Chloride 0.08 ppbv to 25 ppbv (0.20 <math>\mu\text{g}/\text{m}^3</math> to 63.4 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>1,3-Butadiene 0.08 ppbv to 25 ppbv (0.18 <math>\mu\text{g}/\text{m}^3</math> to 55.2 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Acetaldehyde 0.08 ppbv to 25 ppbv (0.14 <math>\mu\text{g}/\text{m}^3</math> to 45.0 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Chloroethane 0.08 ppbv to 25 ppbv (0.21 <math>\mu\text{g}/\text{m}^3</math> to 65.4 <math>\mu\text{g}/\text{m}^3</math>)</li> </ul>	<p>- UAE.TP.TOX.003 based on U.S.EPA, Compendium Method TO-15, 2<sup>nd</sup> edition, January 1999</p>

UAE

UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำเนาถูกต้อง

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☒ ถาวร

(Permanent)

☐ นอกสถานที่

(Site)

☐ชั่วคราว

(Temporary)

☐เคลื่อนที่

(Mobile)

☐หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสีสิ่งแวดล้อม (Environmental field)</p> <p>6. บรรยากาศ (ต่อ) (ambient) (cont.)</p>	<p>- Volatile organic compounds (VOCs) (cont.)</p> <ul style="list-style-type: none"> <li>Acrolein 0.08 ppbv to 25 ppbv (0.18 <math>\mu\text{g}/\text{m}^3</math> to 57.3 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>1,1-Dichloroethene(1,1-Dichloroethylene) 0.08 ppbv to 25 ppbv (0.31 <math>\mu\text{g}/\text{m}^3</math> to 98.2 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Acetone 0.08 ppbv to 25 ppbv (0.19 <math>\mu\text{g}/\text{m}^3</math> to 59.4 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Methyl Iodide 0.08 ppbv to 25 ppbv (0.46 <math>\mu\text{g}/\text{m}^3</math> to 145 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Acetonitrile 0.08 ppbv to 25 ppbv (0.13 <math>\mu\text{g}/\text{m}^3</math> to 41.9 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Methylene Chloride (Dichloromethane) 0.08 ppbv to 25 ppbv (0.27 <math>\mu\text{g}/\text{m}^3</math> to 85.9 <math>\mu\text{g}/\text{m}^3</math>)</li> </ul>	<p>- UAE.TP.TOX.003 based on U.S.EPA, Compendium Method TO-15, 2<sup>nd</sup> edition, January 1999</p>

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)

**UAE**

UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

อ.สม  
สำเนาถูกต้อง

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ  
(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022  
(Certification No. 21-LB0022)



ฉบับที่ 04  
(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565  
(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566  
(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☒ถาวร  
(Permanent)

☐นอกสถานที่  
(Site)

☐ชั่วคราว  
(Temporary)

☐เคลื่อนที่  
(Mobile)

☐หลายสถานที่  
(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาส่งแวดล้อม (Environmental field)</p> <p>6. บรรยากาศ (ต่อ) (ambient) (cont.)</p>	<p>- Volatile organic compounds (VOCs) (cont.)</p> <ul style="list-style-type: none"> <li>Acrylonitrile 0.08 ppbv to 25 ppbv (0.17 <math>\mu\text{g}/\text{m}^3</math> to 54.2 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Hexane 0.08 ppbv to 25 ppbv (0.28 <math>\mu\text{g}/\text{m}^3</math> to 87.9 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>cis-1,2-Dichloroethene(cis-1,2-Dichloroethylene) 0.08 ppbv to 25 ppbv (0.31 <math>\mu\text{g}/\text{m}^3</math> to 98.2 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Methyl Ethyl Ketone (MEK) 0.08 ppbv to 25 ppbv (0.24 <math>\mu\text{g}/\text{m}^3</math> to 73.6 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>Cyclohexane 0.08 ppbv to 25 ppbv (0.27 <math>\mu\text{g}/\text{m}^3</math> to 85.9 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>2-Pentanone 0.08 ppbv to 25 ppbv (0.28 <math>\mu\text{g}/\text{m}^3</math> to 87.9 <math>\mu\text{g}/\text{m}^3</math>)</li> </ul>	<p>- UAE.TP.TOX.003 based on U.S.EPA, Compendium Method TO-15, 2<sup>nd</sup> edition, January 1999</p>



*[Signature]*

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม  
(Ministry of Industry, Thai Industrial Standards Institute)

ดำเนินาถูกต้อง

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ

(Laboratory status)

☒ ถาวร

(Permanent)

☐ นอกสถานที่

(Site)

☐ชั่วคราว

(Temporary)

☐เคลื่อนที่

(Mobile)

☐หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสีสิ่งแวดล้อม (Environmental field)</p> <p>6. บรรยากาศ (ต่อ) (ambient) (cont.)</p>	<p>- Volatile organic compounds (VOCs) (cont.)</p> <ul style="list-style-type: none"> <li>• 1,2-Dichloropropane 0.08 ppbv to 25 ppbv (0.37 <math>\mu\text{g}/\text{m}^3</math> to 115 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>• 3-Pentanone 0.08 ppbv to 25 ppbv (0.28 <math>\mu\text{g}/\text{m}^3</math> to 87.9 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>• 1,4-Dioxane 0.08 ppbv to 25 ppbv (0.29 <math>\mu\text{g}/\text{m}^3</math> to 90.0 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>• trans-1,3-Dichloropropene 0.08 ppbv to 25 ppbv (0.36 <math>\mu\text{g}/\text{m}^3</math> to 112 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>• 1,1,2-Trichloroethane 0.08 ppbv to 25 ppbv (0.43 <math>\mu\text{g}/\text{m}^3</math> to 135 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>• 3-Hexanone 0.08 ppbv to 25 ppbv (0.33 <math>\mu\text{g}/\text{m}^3</math> to 102 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>• Ethylbenzene 0.08 ppbv to 25 ppbv (0.35 <math>\mu\text{g}/\text{m}^3</math> to 108 <math>\mu\text{g}/\text{m}^3</math>)</li> </ul>	<p>- UAE.TP.TOX.003 based on U.S.EPA, Compendium Method TO-15, 2<sup>nd</sup> edition, January 1999</p>

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)

**UAE**

UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำเนาถูกต้อง

หน้า 17/27

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ  
(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☒ถาวร

(Permanent)

☐นอกสถานที่

(Site)

☐ชั่วคราว

(Temporary)

☐เคลื่อนที่

(Mobile)

☐หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาส่งแวดล้อม (Environmental field)</p> <p>6. บรรยากาศ(ต่อ) (ambient) (cont.)</p>	<p>- Volatile organic compounds (VOCs) (cont.)</p> <ul style="list-style-type: none"> <li>• m,p-Xylene 0.16 ppbv to 50 ppbv (0.70 <math>\mu\text{g}/\text{m}^3</math> to 217 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>• o-Xylene 0.08 ppbv to 25 ppbv (0.35 <math>\mu\text{g}/\text{m}^3</math> to 108 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>• 1,4-Dichlorobenzene 0.08 ppbv to 25 ppbv (0.48 <math>\mu\text{g}/\text{m}^3</math> to 149 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>• 1,2,3-Trimethylbenzene 0.08 ppbv to 25 ppbv (0.39 <math>\mu\text{g}/\text{m}^3</math> to 123 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>• Benzyl Chloride 0.08 ppbv to 25 ppbv (0.41 <math>\mu\text{g}/\text{m}^3</math> to 129 <math>\mu\text{g}/\text{m}^3</math>)</li> <li>• Propanal 0.08 ppbv to 25 ppbv (0.19 <math>\mu\text{g}/\text{m}^3</math> to 59.3 <math>\mu\text{g}/\text{m}^3</math>)</li> </ul>	<p>- UAE.TP.TOX.003 based on U.S.EPA, Compendium Method TO-15, 2<sup>nd</sup> edition, January 1999</p>



UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำเนาถูกต้อง

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม  
(Ministry of Industry, Thai Industrial Standards Institute)

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ

(Laboratory status)

☒ ถาวร

(Permanent)

☐ นอกสถานที่

(Site)

☐ชั่วคราว

(Temporary)

☐เคลื่อนที่

(Mobile)

☐หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาโภคภัณฑ์ (Consumer products field)</p> <p>1. น้ำสำหรับบริโภคและน้ำประปา (drinkingwater and tap water)</p>	<p>- Chloride 2.0 mg/l to 500 mg/l</p> <p>- Total hardness 4.0 mg/l to 500 mg/l</p> <p>- Fluoride 0.08 mg/l to 5.20 mg/l</p>	<p>- Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23<sup>rd</sup> edition, 2017, part 4500-Cl<sup>-</sup> B</p> <p>- Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23<sup>rd</sup> edition, 2017, part 2340 C</p> <p>- Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23<sup>rd</sup> Edition , 2017, part 4500-F<sup>-</sup> D</p>



กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)



UAE CONSULTANT AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำเนาถูกต้อง

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022.)



ฉบับที่ 03

(Issue No.)

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

ออกให้ตั้งแต่วันที่ 16 สิงหาคม พ.ศ. 2564

(Valid from) (16 August B.E.2564 (2021))

☐ ถาวร

(Permanent)

☒ นอกสถานที่

(Site)

☐ ชั่วคราว

(Temporary)

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))


☐ เคลื่อนที่

(Mobile)

☐ หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสีสิ่งแวดล้อม (Environmental field)</p> <p>1. บรรยากาศ (ambient)</p>	<ul style="list-style-type: none"> <li>- ระดับเสียง (sound level)</li> <li>• ระดับเสียงเฉลี่ย (equivalent continuous sound pressure level; <math>L_{Aeq T}</math>) 30 dB(A) to 120 dB(A)</li> <li>• ระดับเสียงสูงสุด (maximum sound level; <math>L_{Amax}</math>) 30 dB(A) to 120 dB(A)</li> <li>• ระดับเสียงต่ำสุด (minimum sound level; <math>L_{Amin}</math>) 30 dB(A) to 120 dB(A)</li> <li>• ระดับเสียงเปอร์เซ็นต์ไทล์ที่ N (percentile sound level; <math>L_{AN}</math>) 30 dB(A) to 120 dB(A)</li> </ul>	<ul style="list-style-type: none"> <li>- ISO 1996-1: 2016</li> <li>- ประกาศคณะกรรมการสิ่งแวดล้อมแห่งชาติ ฉบับที่ 15 (พ.ศ.2540) เรื่อง กำหนดมาตรฐานระดับเสียงโดยทั่วไป ลงวันที่ 12 มีนาคม พ.ศ.2540, ประกาศกรมควบคุมมลพิษ (พ.ศ. 2540) เรื่องวิธีการคำนวณค่าระดับเสียง ลงวันที่ 11 สิงหาคม พ.ศ.2540 และประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อมเรื่องกำหนดมาตรฐานควบคุมระดับเสียงและความสั่นสะเทือน ลงวันที่ 7 พฤศจิกายน พ.ศ.2548</li> </ul>



UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำเนาถูกต้อง

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☐ ถาวร  
(Permanent)

☒ นอกสถานที่  
(Site)

☐ชั่วคราว  
(Temporary)

☐เคลื่อนที่  
(Mobile)

☐หลายสถานที่  
(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสิ่งแวดล้อม (Environmental field)</p> <p>1. บรรยากาศ (ต่อ) (ambient) (cont.)</p>	<ul style="list-style-type: none"> <li>- ระดับเสียงรบกวน <ul style="list-style-type: none"> <li>• ระดับเสียงพื้นฐานหรือระดับเสียง เปอร์เซ็นต์ไทล์ที่ 90 (background noise level ; <math>L_{A90}</math>) 30 dB(A) to 120 dB(A)</li> <li>• ระดับเสียงขณะไม่มีการรบกวน (residual noise level; <math>L_{Aeq,T}</math>) 30 dB(A) to 120 dB(A)</li> <li>• ระดับเสียงขณะมีการรบกวน (specific noise level; <math>L_{Aeq,T}</math>) 30 dB(A) to 120 dB(A)</li> <li>• ระดับการรบกวน 2 dB(A) to 40 dB(A)</li> </ul> </li> </ul>	<ul style="list-style-type: none"> <li>- ISO 1996-1: 2016</li> <li>- ประกาศคณะกรรมการสิ่งแวดล้อม แห่งชาติ ฉบับที่ 29 (พ.ศ.2550) เรื่อง ค่าระดับเสียงรบกวน ลงวันที่ 29 มิถุนายน พ.ศ.2550, ประกาศ คณะกรรมการควบคุมมลพิษ เรื่อง วิธีการตรวจวัดระดับเสียงพื้นฐาน ระดับเสียงขณะไม่มีการรบกวน การ ตรวจวัดและคำนวณระดับเสียงขณะมี การรบกวน และการคำนวณค่าระดับ การรบกวน และแบบบันทึกการ ตรวจวัดเสียงรบกวน ลงวันที่ 31 สิงหาคม พ.ศ.2550, ประกาศ กระทรวงอุตสาหกรรม (พ.ศ.2548) เรื่อง กำหนดค่าระดับเสียงการรบกวน และระดับเสียงที่เกิดจากการประกอบ กิจการโรงงาน พ.ศ.2548 ลงวันที่ 27 ธันวาคม พ.ศ.2548และประกาศกรม โรงงานอุตสาหกรรม เรื่อง วิธีการ ตรวจวัดระดับเสียงการรบกวน ระดับ เสียงเฉลี่ย 24 ชั่วโมง และระดับเสียง สูงสุดที่เกิดจากการประกอบกิจการ โรงงาน พ.ศ.2553 ลงวันที่ 20 ธันวาคม พ.ศ.2553</li> </ul>

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)



สำเนาถูกต้อง

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☐ ถาวร

(Permanent)

☒ นอกสถานที่

(Site)

☐ชั่วคราว

(Temporary)

☐เคลื่อนที่

(Mobile)

☐หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสังแวดล้อม (Environmental field)</p> <p>1. บรรยากาศ (ต่อ) (ambient) (cont.)</p>	<ul style="list-style-type: none"> <li>- ความสั่นสะเทือน (Vibration) <ul style="list-style-type: none"> <li>• ความเร็วอนุภาคสูงสุด(Velocity) 10mm/s to 30 mm/s (ทั้งแกน X,Y,Z)</li> <li>• ความถี่ (Frequency) 50Hz to 160 Hz (ทั้งแกน X,Y,Z)</li> </ul> </li> <li>- Fine Particulate Matter as PM<sub>2.5</sub> 2.00 µg/m<sup>3</sup> to 200µg/m<sup>3</sup></li> </ul>	<ul style="list-style-type: none"> <li>- ประกาศคณะกรรมการ สิ่งแวดล้อมแห่งชาติ ฉบับที่ 37 (พ.ศ. 2553)เรื่องกำหนด มาตรฐานความสั่นสะเทือนเพื่อ ป้องกันผลกระทบต่ออาคาร ลง วันที่ 26 เมษายน พ.ศ. 2553</li> <li>- ประกาศกระทรวง ทรัพยากรธรรมชาติและ สิ่งแวดล้อม เรื่อง กำหนด มาตรฐานควบคุมระดับเสียงและ ความสั่นสะเทือนจากการทำ เหมืองหิน ลงวันที่ 7 พฤศจิกายน พ.ศ.2548</li> <li>- DIN 45669-1:2010</li> <li>- DIN 45669-2:2005</li> <li>- DIN 4150-3:1999</li> <li>- US EPA, Code of Federal Regulation, 40 CFR Chapter I -Part 50, Appendix L, Reference Method for the Determination of Fine Particulate Matter As PM<sub>2.5</sub> in the Atmosphere, 2021</li> </ul>

**UAE**

UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำเนาถูกต้อง

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04  
(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565  
(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566  
(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☐ ถาวร

(Permanent)

☒ นอกสถานที่

(Site)

☐ ชั่วคราว

(Temporary)

☐ เคลื่อนที่

(Mobile)

☐ หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสิ่งแวดล้อม (Environmental field)</p> <p>2. พื้นที่ชุมชนโดยรอบสนามบิน (community areas in vicinity of airport)</p>	<p>- ระดับเสียงอากาศยาน (aircraft sound)</p> <p>• ระดับเสียงเฉลี่ยกลางวันและกลางคืน (day-night average sound level; <math>L_{Adn}</math>)</p> <p>30 dB(A) to 120 dB(A)</p>	<p>- ประกาศกรมควบคุมมลพิษ (พ.ศ. 2556) เรื่อง วิธีการตรวจวัดระดับเสียงอากาศยานในพื้นที่ชุมชน ข้อ 2 วิธีการตรวจวัดระดับเสียงอากาศยานสำหรับจุดตรวจวัดชั่วคราวในพื้นที่ชุมชน ลงวันที่ 4 กันยายน พ.ศ.2556 และประกาศกรมควบคุมมลพิษ (พ.ศ.2540) เรื่องการคำนวณระดับเสียง ลงวันที่ 11 สิงหาคม พ.ศ.2540</p>

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)



ดำเนินาถูกต้อง

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☐ ถาวร

(Permanent)

☒ นอกสถานที่

(Site)

☐ ชั่วคราว

(Temporary)

☐ เคลื่อนที่

(Mobile)

☐ หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสังแวดล้อม (Environmental field)</p> <p>3. สถานประกอบการ (workplace)</p>	<p>- ระดับเสียง (sound level)</p> <ul style="list-style-type: none"> <li>ระดับเสียงเฉลี่ย (equivalent continuous sound pressure level; <math>L_{Aeq T}</math>) 30 dB(A) to 120 dB(A)</li> <li>ระดับเสียงสูงสุด (maximum sound level; <math>L_{Amax}</math>) 30 dB(A) to 120 dB(A)</li> <li>ระดับเสียงต่ำสุด (minimum sound level ; <math>L_{Amin}</math>) 30 dB(A) to 120 dB(A)</li> <li>ระดับเสียงเปอร์เซ็นต์ไทล์ที่ N(percentile sound level; <math>L_{AN}</math>) 30 dB(A) to 120 dB(A)</li> </ul>	<p>- ประกาศกรมสวัสดิการและคุ้มครองแรงงาน เรื่อง หลักเกณฑ์วิธีการตรวจวัด และการวิเคราะห์สภาวะการทำงานเกี่ยวกับระดับความร้อน แสงสว่าง หรือเสียง รวมทั้งระยะเวลาและประเภทกิจการที่ต้องดำเนินการ ลงวันที่ 8 กุมภาพันธ์ 2561, กฎกระทรวง (กระทรวงแรงงาน) กำหนดมาตรฐานในการบริหารจัดการและดำเนินการด้านความปลอดภัย อาชีวอนามัยและสภาพแวดล้อมในการทำงานเกี่ยวกับความร้อน แสงสว่าง และเสียง พ.ศ. 2559 ลงวันที่ 7 ตุลาคม พ.ศ.2559 และประกาศกระทรวงอุตสาหกรรม เรื่อง มาตรการคุ้มครองความปลอดภัยในการประกอบกิจการโรงงานเกี่ยวกับสภาวะแวดล้อมในการทำงาน พ.ศ.2546 ลงวันที่ 6 พฤศจิกายน พ.ศ.2546</p>

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)

UAE

UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำเนาถูกต้อง

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☐ ถาวร

(Permanent)

☒ นอกสถานที่

(Site)

☐ชั่วคราว

(Temporary)

☐เคลื่อนที่

(Mobile)


☐หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสิ่งแวดล้อม (Environmental field)</p> <p>3. สถานประกอบการ (ต่อ) (workplace) (cont.)</p>	<ul style="list-style-type: none"> <li>- ความเข้มของแสงสว่าง (light Intensity) 0 Lux to 20000 Lux</li> <li>- ระดับเสียงแบบติดตัวบุคคล (noise dose) <ul style="list-style-type: none"> <li>• ระดับเสียงเฉลี่ยตลอดเวลาการทำงาน (time weighted average) 40 dB(A) to 140 dB(A)</li> <li>• ระดับเสียงสูงสุด (peak) 115 dB(A) to 143 dB(A)</li> </ul> </li> <li>- ระดับความร้อน (heat stress) <ul style="list-style-type: none"> <li>• อุณหภูมิเวทบัลโบลบ (wet bulb globe temperature) 20 °C to 40 °C</li> </ul> </li> </ul>	<ul style="list-style-type: none"> <li>- กฎกระทรวง (กระทรวงแรงงาน) เรื่อง กำหนดมาตรฐานในการ บริหารจัดการ และดำเนินการด้าน ความปลอดภัย อาชีวอนามัยและ สภาพแวดล้อมในการทำงาน เกี่ยวกับความร้อน แสงสว่าง และเสียง พ.ศ. 2559ลงวันที่ 7 ตุลาคม พ.ศ.2559</li> <li>- ประกาศกระทรวงอุตสาหกรรม เรื่อง มาตรการคุ้มครองความ ปลอดภัยในการประกอบ กิจการโรงงานเกี่ยวกับสภาวะ แวดล้อมในการทำงาน พ.ศ. 2546 ลงวันที่ 6 พฤศจิกายน พ.ศ.2546</li> <li>- ประกาศกระทรวงอุตสาหกรรม เรื่อง มาตรการคุ้มครองความ ปลอดภัยในการประกอบ กิจการโรงงานเกี่ยวกับสภาวะ แวดล้อมในการทำงาน พ.ศ. 2546 ลงวันที่ 6 พฤศจิกายน พ.ศ.2546</li> </ul>

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)



UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

ดำเนินถูกต้อง

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☐ ถาวร

(Permanent)

☒ นอกสถานที่

(Site)

☐ชั่วคราว

(Temporary)

☐เคลื่อนที่

(Mobile)

☐หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาสังแวดล้อม (Environmental field)</p> <p>3. สถานประกอบการ (ต่อ) (workplace) (cont.)</p> <p>4. ปล่องระบายอากาศเสีย (Stack)</p>	<p>- Total Dust 0.200 mg/m<sup>3</sup> to 15.0 mg/m<sup>3</sup></p> <p>- Respirable Dust 0.010 mg/m<sup>3</sup> to 5.00 mg/m<sup>3</sup></p> <p>- Sulfur dioxide 45 ppm to 1 000 ppm</p> <p>- Nitrogen oxide 45 ppm to 700 ppm</p> <p>- Carbon monoxide 45 ppm to 5000 ppm</p>	<p>- NIOSH manual of analytical method (NMAM), method 0500, fourth edition, 15<sup>th</sup> Aug, 1994</p> <p>- NIOSH manual of analytical method (NMAM), method 0600, fourth edition, 15<sup>th</sup> Aug, 1994</p> <p>- U.S. EPA, Code of Federal Regulations, 40 CFR Part 60 Appendix A, Method 6C, July 2018</p> <p>- U.S. EPA, Code of Federal Regulations, 40 CFR Part 60 Appendix A, Method 7E, July 2018</p> <p>- U.S. EPA, Code of Federal Regulations, 40 CFR Part 60 Appendix A, Method 10, July 2018</p>

UAE

UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำเนาถูกต้อง

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม

(Ministry of Industry, Thai Industrial Standards Institute)

รายละเอียดสาขาและขอบข่ายใบรับรองห้องปฏิบัติการ

(Scope of Accreditation for Testing)

ใบรับรองเลขที่ 21-LB0022

(Certification No. 21-LB0022)



ฉบับที่ 04

(Issue No. 04)

ออกให้ตั้งแต่วันที่ 14 กุมภาพันธ์ 2565

(Valid from) (14 February B.E.2565 (2022))

ถึงวันที่ 17 พฤษภาคม พ.ศ. 2566

(Until) (17 May B.E.2566(2023))

สถานภาพห้องปฏิบัติการ  
(Laboratory status)

☐ ถาวร

(Permanent)

☒ นอกสถานที่

(Site)

☐ ชั่วคราว

(Temporary)

☐ เคลื่อนที่

(Mobile)

☐ หลายสถานที่

(Multisite)

สาขาการทดสอบ (Field of Testing)	รายการทดสอบ (Parameter)	วิธีทดสอบ (Test Method)
<p>สาขาส่งแวดล้อม (Environmental field)</p> <p>5. น้ำ/น้ำเสีย/น้ำผิวดิน/น้ำทะเล (Water/Wastewater/ Surface Water/Seawater)</p>	<p>- pH 4.0 – 10.0</p>	<p>- Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23<sup>rd</sup> edition, 2017, Part 4500-H<sup>+</sup> B (Include sampling)</p>

กระทรวงอุตสาหกรรมสำนักงานมาตรฐานผลิตภัณฑ์อุตสาหกรรม  
(Ministry of Industry, Thai Industrial Standards Institute)

**UAE**  
UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำเนาถูกต้อง

สำเนาใบรับรองความสามารถ ISO-IEC 17025-2017\_กรมวิทย์ฯ



ที่ อว 0303/5029

## ใบรับรองความสามารถห้องปฏิบัติการทดสอบ

ใบรับรองฉบับนี้ให้ไว้เพื่อแสดงว่า

บริษัท ยูไนเต็ด แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด  
เลขที่ 3 ซอยอุดมสุข 41 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง  
กรุงเทพมหานคร 10260

ได้ผ่านการประเมินความสามารถห้องปฏิบัติการทดสอบตามมาตรฐาน ISO/IEC 17025 : 2017  
และข้อกำหนด กฎระเบียบ และเงื่อนไขการรับรองความสามารถห้องปฏิบัติการทดสอบ  
ของกองบริหารและรับรองห้องปฏิบัติการ กรมวิทยาศาสตร์บริการ

หมายเลขการรับรองระบบงานที่ ทดสอบ - 0063


รายละเอียดการรับรองดังข้อบ่งชี้การรับรองแนบท้าย

ออกให้ ณ วันที่ : 29 มีนาคม 2565

หมดอายุ วันที่ : 28 มีนาคม 2569

ลงชื่อ

:



ผู้อำนวยการกองบริหารและรับรองห้องปฏิบัติการ  
**UAE**  
UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED  
สำเนาถูกต้อง

กองบริหารและรับรองห้องปฏิบัติการ กรมวิทยาศาสตร์บริการ  
กระทรวงการอุดมศึกษา วิทยาศาสตร์ วิจัย และนวัตกรรม

ขอข่ายการรับรองความสามารถห้องปฏิบัติการทดสอบ

ชื่อห้องปฏิบัติการ : บริษัท ยูไนเต็ด แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด

สถานที่ตั้ง : เลขที่ 3 ซอยอุดมสุข 41 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง  
กรุงเทพมหานคร 10260

หมายเลขการรับรองระบบงานที่ : ทดสอบ - 0063

สถานะของห้องปฏิบัติการ : ☒ ถาวร ☐ นอกสถานที่ ☐ชั่วคราว ☐เคลื่อนที่

ลำดับ ที่	วัสดุ / ผลิตภัณฑ์ที่ทดสอบ	รายการที่ทดสอบ / ช่วงของการทดสอบ	วิธีทดสอบ / เทคนิคที่ใช้
1	น้ำ	<p>- Coliforms MPN/100 mL</p> <p>- Fecal coliforms MPN/100 mL</p> <p>- <i>E. coli</i> MPN/100 mL</p>	<p>Standard Methods for the Examination of Water and Wastewater, APHA, AWWA &amp; WEF, 23<sup>rd</sup> ed., 2017, part 9221 B</p> <p>Standard Methods for the Examination of Water and Wastewater, APHA, AWWA &amp; WEF, 23<sup>rd</sup> ed., 2017, part 9221 B, E</p> <p>Standard Methods for the Examination of Water and Wastewater, APHA, AWWA &amp; WEF, 23<sup>rd</sup> ed., 2017, part 9221 B, E, F</p>

ออกครั้งแรก ณ วันที่ 21 กันยายน 2553



สัญญาถูกต้อง

กองบริหารและรับรองห้องปฏิบัติการ กรมวิทยาศาสตร์บริการ กระทรวงการอุดมศึกษา วิทยาศาสตร์ วิจัย และนวัตกรรม

ขอข่ายการรับรองความสามารถห้องปฏิบัติการทดสอบ

ชื่อห้องปฏิบัติการ : บริษัท ยูไนเต็ด แอนาไลสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด

สถานที่ตั้ง : เลขที่ 3 ซอยอุดมสุข 41 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง  
กรุงเทพมหานคร 10260

หมายเลขการรับรองระบบงานที่ : ทดสอบ - 0063

สถานะของห้องปฏิบัติการ : ☒ ถาวร ☐ นอกสถานที่ ☐ชั่วคราว ☐เคลื่อนที่

ลำดับ ที่	วัสดุ / ผลิตภัณฑ์ที่ทดสอบ	รายการที่ทดสอบ / ช่วงของการทดสอบ	วิธีทดสอบ / เทคนิคที่ใช้
1 (ต่อ)	น้ำ	- Standard plate count cfu/mL  - <i>E. coli</i> Detected or not detected  - <i>Salmonella</i> spp. Detected or not detected	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 9215 B  Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 9221 D, F  ISO 19250 : 2010

ออกครั้งแรก ณ วันที่ 21 กันยายน 2553



สำเนาถูกต้อง  
ฉบับที่ 8

กองบริหารและรับรองห้องปฏิบัติการ กรมวิทยาศาสตร์บริการ กระทรวงการอุดมศึกษา วิทยาศาสตร์ วิจัย และนวัตกรรม

**ขอข่ายการรับรองความสามารถห้องปฏิบัติการทดสอบ**

ชื่อห้องปฏิบัติการ : บริษัท ยูไนเต็ด แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด

สถานที่ตั้ง : เลขที่ 3 ซอยอุดมสุข 41 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง  
กรุงเทพมหานคร 10260

หมายเลขการรับรองระบบงานที่ : ทดสอบ - 0063

สถานะของห้องปฏิบัติการ : ☒ ถาวร ☐ นอกสถานที่ ☐ชั่วคราว ☐เคลื่อนที่

ลำดับ ที่	วัสดุ / ผลิตภัณฑ์ที่ทดสอบ	รายการที่ทดสอบ / ช่วงของการทดสอบ	วิธีทดสอบ / เทคนิคที่ใช้
1 (ต่อ)	น้ำ	<ul style="list-style-type: none"> <li>- สารที่ละลายได้ทั้งหมด ที่อุณหภูมิ 180 °C 25 mg/L ถึง 1 000 mg/L</li> <li>- สารทั้งหมด ที่อุณหภูมิ 103 °C ถึง 105 °C 25 mg/L ถึง 1 000 mg/L</li> <li>- อินทรีย์คาร์บอนทั้งหมด 0.50 mg/L ถึง 100 mg/L</li> </ul>	<p>Standard Methods for the Examination of Water and Wastewater, APHA, AWWA &amp; WEF, 23<sup>rd</sup> ed., 2017, part 2540 C</p> <p>Standard Methods for the Examination of Water and Wastewater, APHA, AWWA &amp; WEF, 23<sup>rd</sup> ed., 2017, part 2540 B</p> <p>Standard Methods for the Examination of Water and Wastewater, APHA, AWWA &amp; WEF, 23<sup>rd</sup> ed., 2017, part 5310 B</p>

ออกครั้งแรก ณ วันที่ 21 กันยายน 2553

  
UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

  
ฉบับที่ 89  
**สำเนาถูกต้อง**

กองบริหารและรับรองห้องปฏิบัติการ กรมวิทยาศาสตร์บริการ กระทรวงการอุดมศึกษา วิทยาศาสตร์ วิจัย และนวัตกรรม

**ขอข่ายการรับรองความสามารถห้องปฏิบัติการทดสอบ**

ชื่อห้องปฏิบัติการ : บริษัท ยูไนเต็ด แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด

สถานที่ตั้ง : เลขที่ 3 ซอยอุดมสุข 41 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง  
กรุงเทพมหานคร 10260

หมายเลขการรับรองระบบงานที่ : ทดสอบ - 0063

สถานะของห้องปฏิบัติการ : ☒ ถาวร ☐ นอกสถานที่ ☐ชั่วคราว ☐เคลื่อนที่

ลำดับ ที่	วัสดุ / ผลิตภัณฑ์ที่ทดสอบ	รายการที่ทดสอบ / ช่วงของการทดสอบ	วิธีทดสอบ / เทคนิคที่ใช้
1 (ต่อ)	น้ำ	- ฟีนอล 0.005 mg/L ถึง 0.100 mg/L  - ปรีท 0.500 µg/L ถึง 2 000 µg/L  - แพลงก์ตอนพืช (สกุล) <i>Scenedesmus</i> spp. <i>Pediastrum</i> spp. <i>Euglena</i> spp. <i>Phacus</i> spp. <i>Coelastrum</i> spp. Natural unit/mL	In - house method : UAE.TP.WAS.009 based on ISO 14402: 1999  In - house method : UAE.TP.HEM.002 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 3112 B  Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 10200 F

ออกครั้งแรก ณ วันที่ 21 กันยายน 2553


**สำเนาถูกต้อง**  
ฉบับที่ 8

กองบริหารและรับรองห้องปฏิบัติการ กรมวิทยาศาสตร์บริการ กระทรวงการอุดมศึกษา วิทยาศาสตร์ วิจัย และนวัตกรรม

ขอข่ายการรับรองความสามารถห้องปฏิบัติการทดสอบ

ชื่อห้องปฏิบัติการ : บริษัท ยูไนเต็ด แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด

สถานที่ตั้ง : เลขที่ 3 ซอยอุดมสุข 41 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง  
กรุงเทพมหานคร 10260

หมายเลขการรับรองระบบงานที่ : ทดสอบ - 0063

สถานะของห้องปฏิบัติการ : ☒ ถาวร ☐ นอกสถานที่ ☐ชั่วคราว ☐เคลื่อนที่

ลำดับ ที่	วัสดุ / ผลิตภัณฑ์ที่ทดสอบ	รายการที่ทดสอบ / ช่วงของการทดสอบ	วิธีทดสอบ / เทคนิคที่ใช้
2	น้ำเสีย	<p>- Coliforms MPN/100 mL</p> <p>- Fecal coliforms MPN/100 mL</p> <p>- <i>E. coli</i> MPN/100 mL</p>	<p>Standard Methods for the Examination of Water and Wastewater, APHA, AWWA &amp; WEF, 23<sup>rd</sup> ed., 2017, part 9221 B</p> <p>Standard Methods for the Examination of Water and Wastewater, APHA, AWWA &amp; WEF, 23<sup>rd</sup> ed., 2017, part 9221 B, E</p> <p>Standard Methods for the Examination of Water and Wastewater, APHA, AWWA &amp; WEF, 23<sup>rd</sup> ed., 2017, part 9221 B, E, F</p>

ออกครั้งแรก ณ วันที่ 21 กันยายน 2553



ผู้มีอำนาจต้อง  
ลงนาม

กองบริหารและรับรองห้องปฏิบัติการ กรมวิทยาศาสตร์บริการ กระทรวงการอุดมศึกษา วิทยาศาสตร์ วิจัย และนวัตกรรม

**ขอข่ายการรับรองความสามารถห้องปฏิบัติการทดสอบ**

ชื่อห้องปฏิบัติการ : บริษัท ยูไนเต็ด แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด

สถานที่ตั้ง : เลขที่ 3 ซอยอุดมสุข 41 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง  
กรุงเทพมหานคร 10260

หมายเลขการรับรองระบบงานที่ : ทดสอบ - 0063

สถานะของห้องปฏิบัติการ : ☒ ถาวร ☐ นอกสถานที่ ☐ชั่วคราว ☐เคลื่อนที่

ลำดับ ที่	วัสดุ / ผลิตภัณฑ์ที่ทดสอบ	รายการที่ทดสอบ / ช่วงของการทดสอบ	วิธีทดสอบ / เทคนิคที่ใช้
2 (ต่อ)	น้ำเสีย	<p>- สารที่ละลายได้ทั้งหมด ที่อุณหภูมิ 103 °C ถึง 105 °C 25 mg/L ถึง 6 000 mg/L</p> <p>- สารที่ละลายได้ทั้งหมด ที่อุณหภูมิ 180 °C 25 mg/L ถึง 6 000 mg/L</p> <p>- ไนโตรเจน ในรูป ที เค เอ็น 5.0 mg/L ถึง 500 mg/L</p>	<p>In - house method : UAE.TP.WAO.007 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA &amp; WEF, 23<sup>rd</sup> ed., 2017, part 2540 C</p> <p>Standard Methods for the Examination of Water and Wastewater, APHA, AWWA &amp; WEF, 23<sup>rd</sup> ed., 2017, part 2540 C</p> <p>In - house method : UAE.TP.WAS.001 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA &amp; WEF, 23<sup>rd</sup> ed., 2017, part 4500 - N<sub>org</sub> C</p>

ออกครั้งแรก ณ วันที่ 21 กันยายน 2553

  
UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

  
**สำเนาถูกต้อง**  
ฉบับที่ 8

กองบริหารและรับรองห้องปฏิบัติการ กรมวิทยาศาสตร์บริการ กระทรวงการอุดมศึกษา วิทยาศาสตร์ วิจัย และนวัตกรรม

**ขอข่ายการรับรองความสามารถห้องปฏิบัติการทดสอบ**

ชื่อห้องปฏิบัติการ : บริษัท ยูไนเต็ด แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด

สถานที่ตั้ง : เลขที่ 3 ซอยอุดมสุข 41 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง  
กรุงเทพมหานคร 10260

หมายเลขการรับรองระบบงานที่ : ทดสอบ - 0063

สถานะของห้องปฏิบัติการ : ☒ ถาวร ☐ นอกสถานที่ ☐ชั่วคราว ☐เคลื่อนที่

ลำดับ ที่	วัสดุ / ผลิตภัณฑ์ที่ทดสอบ	รายการที่ทดสอบ / ช่วงของการทดสอบ	วิธีทดสอบ / เทคนิคที่ใช้
2 (ต่อ)	น้ำเสีย	- สี 10 ADMI ถึง 300 ADMI  - ไซยาไนต์ 0.005 mg/L ถึง 0.100 mg/L  - เบนซีน 0.20 µg/L ถึง 500 µg/L - เอทิลเบนซีน 0.20 µg/L ถึง 500 µg/L - โทลูอีน 0.20 µg/L ถึง 500 µg/L - ออโร-ไซลีน 0.20 µg/L ถึง 500 µg/L	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 2120 F  ISO 14403-2 : 2012  Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 6200 B

ออกครั้งแรก ณ วันที่ 21 กันยายน 2553

  
UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

  
อำนาจถูกต้อง  
ฉบับที่ 8

กองบริหารและรับรองห้องปฏิบัติการ กรมวิทยาศาสตร์บริการ กระทรวงการอุดมศึกษา วิทยาศาสตร์ วิจัย และนวัตกรรม

## ขอข่ายการรับรองความสามารถห้องปฏิบัติการทดสอบ

ชื่อห้องปฏิบัติการ : บริษัท ยูโนเต็ด แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด


สถานที่ตั้ง : เลขที่ 3 ซอยอุดมสุข 41 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง  
กรุงเทพมหานคร 10260

หมายเลขการรับรองระบบงานที่ : ทดสอบ - 0063

สถานะของห้องปฏิบัติการ : ☒ ถาวร ☐ นอกสถานที่ ☐ชั่วคราว ☐เคลื่อนที่

ลำดับ ที่	วัสดุ / ผลิตภัณฑ์ที่ทดสอบ	รายการที่ทดสอบ / ช่วงของการทดสอบ	วิธีทดสอบ / เทคนิคที่ใช้
2 (ต่อ)	น้ำเสีย	- เมตา,พารา-ไซลีน 0.40 µg/L ถึง 1 000 µg/L - ไซลีนทั้งหมด 0.60 µg/L ถึง 1 500 µg/L  - แพลงก์ตอนพืช (สกุล) <i>Scenedesmus</i> spp. <i>Pediastrum</i> spp. <i>Euglena</i> spp. <i>Phacus</i> spp. <i>Coelastrum</i> spp. Natural unit/mL  - แอมโมเนีย - ไนโตรเจน 5.0 mg/L ถึง 500 mg/L	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 6200 B  Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 10200 F  Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 4500 NH <sub>3</sub> C

ออกครั้งแรก ณ วันที่ 21 กันยายน 2553



A blue ink signature is written over a blue rectangular stamp. The stamp contains the Thai text 'อำนาจถูกต้อง' (Authority is correct) and the number 'ฉบับที่ 8' (Version 8) below it.

กองบริหารและรับรองห้องปฏิบัติการ กรมวิทยาศาสตร์บริการ กระทรวงการอุดมศึกษา วิทยาศาสตร์ วิจัย และนวัตกรรม

**ขอข่ายการรับรองความสามารถห้องปฏิบัติการทดสอบ**

ชื่อห้องปฏิบัติการ : บริษัท ยูไนเต็ด แอนาไลสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด

สถานที่ตั้ง : เลขที่ 3 ซอยอุดมสุข 41 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง  
กรุงเทพมหานคร 10260

หมายเลขการรับรองระบบงานที่ : ทดสอบ - 0063

สถานะของห้องปฏิบัติการ : ☒ ถาวร ☐ นอกสถานที่ ☐ชั่วคราว ☐เคลื่อนที่

ลำดับ ที่	วัสดุ / ผลิตภัณฑ์ที่ทดสอบ	รายการที่ทดสอบ / ช่วงของการทดสอบ	วิธีทดสอบ / เทคนิคที่ใช้
2 (ต่อ)	น้ำเสีย	- ซัลไฟต์ 0.50 mg/L ถึง 3.0 mg/L	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 4500 S <sup>2-</sup> F
3	น้ำทะเล	- Coliforms MPN/100 mL  - บีโตรเลียมไฮโดรคาร์บอนทั้งหมด 0.05 µg/L ถึง 3.00 µg/L	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 9221 B  Intergovernmental Oceanographic Commission, Manual for Monitoring Oil and Dissolved/ Dispersed Petroleum Hydrocarbons in Marine Waters and on Beaches, 1984

ออกครั้งแรก ณ วันที่ 21 กันยายน 2553

  
UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

  
ฉบับที่ 8  
สำเนาถูกต้อง

กองบริหารและรับรองห้องปฏิบัติการ กรมวิทยาศาสตร์บริการ กระทรวงการอุดมศึกษา วิทยาศาสตร์ วิจัย และนวัตกรรม

ขอข่ายการรับรองความสามารถห้องปฏิบัติการทดสอบ

ชื่อห้องปฏิบัติการ : บริษัท ยูไนเต็ด แอนาไลสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด


สถานที่ตั้ง : เลขที่ 3 ซอยอุดมสุข 41 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง  
กรุงเทพมหานคร 10260

หมายเลขการรับรองระบบงานที่ : ทดสอบ - 0063

สถานะของห้องปฏิบัติการ : ☒ ถาวร ☐ นอกสถานที่ ☐ชั่วคราว ☐เคลื่อนที่

ลำดับ ที่	วัสดุ / ผลิตภัณฑ์ที่ทดสอบ	รายการที่ทดสอบ / ช่วงของการทดสอบ	วิธีทดสอบ / เทคนิคที่ใช้
3 (ต่อ)	น้ำทะเล	- ฟอสเฟต-ฟอสฟอรัส 1.5 µg/L ถึง 150 µg/L  - แอมโมเนีย-ไนโตรเจน 50.0 µg/L ถึง 1 000 µg/L	In - house method : UAE.TP.WAT.002 based on Practical Handbook of Seawater Analysis Strickland and Parson, 1972  In - house method : UAE.TP.WAT.001 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 4500 NH <sub>3</sub> H
4	น้ำแข็ง	- Coliforms MPN/100 mL	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 9221 B

ออกครั้งแรก ณ วันที่ 21 กันยายน 2553



สำเนาถูกต้อง  
ฉบับที่ 8

กองบริหารและรับรองห้องปฏิบัติการ กรมวิทยาศาสตร์บริการ กระทรวงการอุดมศึกษา วิทยาศาสตร์ วิจัย และนวัตกรรม

**ขอข่ายการรับรองความสามารถห้องปฏิบัติการทดสอบ**

ชื่อห้องปฏิบัติการ : บริษัท ยูไนเต็ด แอนาไลสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด

สถานที่ตั้ง : เลขที่ 3 ซอยอุดมสุข 41 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง  
กรุงเทพมหานคร 10260

หมายเลขการรับรองระบบงานที่ : ทดสอบ - 0063

สถานะของห้องปฏิบัติการ : ☒ ภายใน ☐ นอกสถานที่ ☐ ชั่วคราว ☐ เคลื่อนที่

ลำดับ ที่	วัสดุ / ผลิตภัณฑ์ที่ทดสอบ	รายการที่ทดสอบ / ช่วงของการทดสอบ	วิธีทดสอบ / เทคนิคที่ใช้
4 (ต่อ)	น้ำแข็ง	- Fecal coliforms MPN/100 mL  - <i>E. coli</i> MPN/100 mL  - Standard plate count cfu/mL	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 9221 B, E  Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 9221 B, E, F  Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 9215 B

ออกครั้งแรก ณ วันที่ 21 กันยายน 2553



UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

  
ฉบับที่ 8  
ถ้าไม่ถูกต้อง

กองบริหารและรับรองห้องปฏิบัติการ กรมวิทยาศาสตร์บริการ กระทรวงการอุดมศึกษา วิทยาศาสตร์ วิจัย และนวัตกรรม

**ขอข่ายการรับรองความสามารถห้องปฏิบัติการทดสอบ**

ชื่อห้องปฏิบัติการ : บริษัท ยูไนเต็ด แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด

สถานที่ตั้ง : เลขที่ 3 ซอยอุดมสุข 41 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง  
กรุงเทพมหานคร 10260

หมายเลขการรับรองระบบงานที่ : ทดสอบ - 0063

สถานะของห้องปฏิบัติการ : ☒ ถาวร ☐ นอกสถานที่ ☐ชั่วคราว ☐เคลื่อนที่

ลำดับ ที่	วัสดุ / ผลิตภัณฑ์ที่ทดสอบ	รายการที่ทดสอบ / ช่วงของการทดสอบ	วิธีทดสอบ / เทคนิคที่ใช้
4 (ต่อ)	น้ำแข็ง	- <i>E. coli</i>  Detected or not detected    - <i>Salmonella</i> spp.  Detected or not detected	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 9221 D, F    ISO 19250 : 2010
5	น้ำสระว่ายน้ำ	- Coliforms  MPN/100 mL    - Fecal coliforms  MPN/100 mL	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 9221 B    Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 9221 B, E



UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

**สำเนาถูกต้อง**  
ฉบับที่ 8

ออกครั้งแรก ณ วันที่ 21 กันยายน 2553

กองบริหารและรับรองห้องปฏิบัติการ กรมวิทยาศาสตร์บริการ กระทรวงการอุดมศึกษา วิทยาศาสตร์ วิจัย และนวัตกรรม

**ขอข่ายการรับรองความสามารถห้องปฏิบัติการทดสอบ**

ชื่อห้องปฏิบัติการ : บริษัท ยูโนเด็ค แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด

สถานที่ตั้ง : เลขที่ 3 ซอยอุดมสุข 41 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง  
กรุงเทพมหานคร 10260

หมายเลขการรับรองระบบงานที่ : ทดสอบ - 0063

สถานะของห้องปฏิบัติการ : ☒ ถาวร ☐ นอกสถานที่ ☐ชั่วคราว ☐เคลื่อนที่

ลำดับ ที่	วัสดุ / ผลิตภัณฑ์ที่ทดสอบ	รายการที่ทดสอบ / ช่วงของการทดสอบ	วิธีทดสอบ / เทคนิคที่ใช้
5 (ต่อ)	น้ำสระว่ายน้ำ	- <i>E. coli</i> MPN/100 mL	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 9221 B, E, F
		- Standard plate count cfu/mL	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 9215 B
		- <i>E. coli</i> Detected or not detected	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 9221 D, F
		- <i>Salmonella</i> spp. Detected or not detected	ISO 19250 : 2010

ออกครั้งแรก ณ วันที่ 21 กันยายน 2553



ฉบับที่ 8  
สำเนาถูกต้อง

กองบริหารและรับรองห้องปฏิบัติการ กรมวิทยาศาสตร์บริการ กระทรวงการอุดมศึกษา วิทยาศาสตร์ วิจัย และนวัตกรรม

## ขอข่วยการรับรองความสามารถห้องปฏิบัติการทดสอบ

ชื่อห้องปฏิบัติการ : บริษัท ยูไนเต็ด แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด

สถานที่ตั้ง : เลขที่ 3 ซอยอุดมสุข 41 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง  
กรุงเทพมหานคร 10260

หมายเลขการรับรองระบบงานที่ : ทดสอบ - 0063

สถานะของห้องปฏิบัติการ : ☒ ถาวร ☐ นอกสถานที่ ☐ชั่วคราว ☐เคลื่อนที่

ลำดับ ที่	วัสดุ / ผลิตภัณฑ์ที่ทดสอบ	รายการที่ทดสอบ / ช่วงของการทดสอบ	วิธีทดสอบ / เทคนิคที่ใช้
6	น้ำบริโภคในภาชนะ บรรจุที่ปิดสนิท	- <i>E. coli</i>  Detected or not detected	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 <sup>rd</sup> ed., 2017, part 9221 D, F
7	ดิน	- ความเป็นกรด-ด่าง  2.0 ถึง 9.0	United States Environmental Protection Agency, 2004, EPA Method 9045 D, Revision 4
8	กากตะกอน	- ความเป็นกรด-ด่าง  2.0 ถึง 9.0	United States Environmental Protection Agency, 2004, EPA Method 9045 D, Revision 4


  
UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED


  
สำเนาถูกต้อง  
ฉบับที่ 8

ออกครั้งแรก ณ วันที่ 21 กันยายน 2553

กองบริหารและรับรองห้องปฏิบัติการ กรมวิทยาศาสตร์บริการ กระทรวงการอุดมศึกษา วิทยาศาสตร์ วิจัย และนวัตกรรม

ขอข่ายการรับรองความสามารถห้องปฏิบัติการทดสอบ

ชื่อห้องปฏิบัติการ : บริษัท ยูไนเต็ด แอนนาลิสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด

สถานที่ตั้ง : เลขที่ 3 ซอยอุดมสุข 41 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง  
กรุงเทพมหานคร 10260

หมายเลขการรับรองระบบงานที่ : ทดสอบ - 0063

สถานะของห้องปฏิบัติการ : ☒ ถาวร ☐ นอกสถานที่ ☐ชั่วคราว ☐ เคลื่อนที่

ลำดับ ที่	วัสดุ / ผลิตภัณฑ์ที่ทดสอบ	รายการที่ทดสอบ / ช่วงของการทดสอบ	วิธีทดสอบ / เทคนิคที่ใช้
9	น้ำปราศจากไอออน	- อินทรีย์คาร์บอนทั้งหมด 250 µg/L ถึง 2 000 µg/L	Standard Methods for the Examination of Water and Wastewater, APHA AWWA & WEF, 23rd ed., 2017, part 5310 C

ออกให้ ณ วันที่ : 29 มีนาคม 2565

ลงชื่อ :



(นางพจมาน ทวีจีน)

ผู้อำนวยการกองบริหารและรับรองห้องปฏิบัติการ



UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

สำเนาถูกต้อง

ฉบับที่ 8

ออกครั้งแรก ณ วันที่ 21 กันยายน 2553

กองบริหารและรับรองห้องปฏิบัติการ กรมวิทยาศาสตร์บริการ กระทรวงการอุดมศึกษา วิทยาศาสตร์ วิจัย และนวัตกรรม

ใบรับรองฉบับภาษาไทย\_กรมวิทยาศาสตร์การแพทย์



สำนักมาตรฐานห้องปฏิบัติการ  
กระทรวงสาธารณสุข


หนังสือฉบับนี้ให้ไว้เพื่อแสดงว่า

ห้องปฏิบัติการ

บริษัทยูไนเต็ดแอนาไลสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด  
เลขที่ 3 ซอย อุดมสุข 41 ถนนสุขุมวิท แขวงบางจาก  
เขตพระโขนง กรุงเทพมหานคร 10260

ได้รับการขึ้นทะเบียนเป็นห้องปฏิบัติการที่ผ่านการรับรองความสามารถ  
ตามมาตรฐาน ISO/IEC 17025 : 2017 และข้อกำหนดและเงื่อนไขการรับรองความสามารถ  
ห้องปฏิบัติการทดสอบด้านการแพทย์และสาธารณสุขของสำนักมาตรฐานห้องปฏิบัติการ  
ตามรายการและวิธีทดสอบที่กำหนดในเอกสารแนบท้ายในด้าน

การทดสอบอาหาร



ผู้อำนวยการสำนักมาตรฐานห้องปฏิบัติการ

ให้ไว้ ณ วันที่ 22 เมษายน 2565

ถึงวันที่ 21 เมษายน 2569

**UAE**  
UNITED ANALYST AND ENGINEERING  
CONSULTANT COMPANY LIMITED

  
ดำเนินถูกต้อง

หมายเลขทะเบียน 1349/65

ห้องปฏิบัติการบริษัท ยูไนเต็ด แอนาליสต์ แอนด์ เอ็นจิเนียริง คอนซัลแตนท์ จำกัด  
 ได้รับการรับรองความสามารถในการทดสอบอาหาร ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
1.	<ul style="list-style-type: none"> <li>● น้ำบริโภค                             <ul style="list-style-type: none"> <li>- น้ำที่ใช้ในกระบวนการผลิต</li> <li>- น้ำบริโภคในภาชนะบรรจุปิดสนิท</li> <li>- น้ำดื่ม</li> <li>- น้ำประปา</li> <li>- น้ำกรอง</li> </ul> </li> <li>● น้ำอุปโภค                             <ul style="list-style-type: none"> <li>- น้ำประปา</li> <li>- น้ำที่ผ่านการกรอง</li> <li>- น้ำสระว่ายน้ำ</li> <li>- น้ำฟิวคิน</li> <li>- น้ำคิบ</li> <li>- น้ำใต้ดิน</li> </ul> </li> <li>● น้ำแข็ง</li> <li>● น้ำหล่อเย็น</li> </ul>	1. <i>Legionella</i> spp. (CFU)  2. <i>Staphylococcus aureus</i> (Detected or not detected/ CFU)	ISO 11731:2017  Standard Methods for the Examination of Water and Wastewater. APHA, AWWA, WEF. 23 <sup>rd</sup> Edition, 2017. Part 9213 B

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 1 ของทั้งหมด 1 หน้า

แก้ไขครั้งที่ 00

วันที่ 22 เมษายน 2565

**UAE**  
 UNITED ANALYST AND ENGINEERING  
 CONSULTANT COMPANY LIMITED

หมายเลขทะเบียน 1349/65

ให้ไว้ ณ วันที่ 22 เมษายน 2565

ถึงวันที่ 21 เมษายน 2569

**สำเนาถูกต้อง**

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติ

(สุรศักดิ์ หมั่นเพล)