

ภาคผนวกที่ 5-4

เอกสารสอบเทียบเครื่องมือตรวจวัดระดับความดังเสียง

Certificate No.: CP20230151EA
Operation No.: CP2023030003

Certificate of Calibration

Equipment: Sound Calibrator
Manufacturer: TENMARS
Model/Type: TM-100
Serial No.: 220501964
ID No.: -
Customer: Pacific Laboratory Co., Ltd.
Address: 14/5358 Moo 14 T. Bang Bua Thong
A. Bang Bua Thong, Nonthaburi 11110
Received Date: 3 March 2023
Calibrated Date: 14 March 2023
Issued Date: 16 March 2023
Calibrated by: Ms. Juntaporn Kunhakom



This report was prepared electronically using applicable electronic signature. Printing or copy of file are considered as a copy of the document.

The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor (k) providing a level of confidence of approximately 95%. This certificate may not be reproduced other than in full except with the prior written approval of the Electrical and Electronics Institute, Foundation for Industrial Development.

Certificate No.: CP20230151EA

Calibration Report

Equipment: Sound Calibrator
Manufacturer: TENMARS
Model/Type: TM-100
Serial No.: 220501964
ID No.: -
Ambient Temperature: (23 ± 2) °C
Relative Humidity: (50 ± 15) %
Pressure: (101.3 ± 1.5) kPa

Method of Calibration :-

IEC 60942:2017

Condition of this result of calibration

1. Reference standards instrument :-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Standard microphone	4180	2661000	AA-1020-22	14 June 2023
2) Waveform Generator	33511B	MY52302264	CK20220058EA	19 June 2023
3) Audio Analyzing DMM	2015-P	4079144	E1U221042	16 March 2023
4) Pressure humidity and Temperature Transmitter	PTU301	F0640002	CL1-P220024 CD20220164EA	17 March 2023 24 July 2023

2. This result of calibration was found accurate as shown on date and place of calibration only.

3. This certification is traceable to the international system of unit maintained at :-

Reference standards instrument for Acoustic function

- National Institute of Metrology (Thailand)

Reference standards instrument for Electrical function

- Electrical and Electronics Institute; NSC Accredited Calibration No.0119

Result of Calibration:-

1. Function : Sound pressure level

Norminal Frequency (Hz)	Specified Sound Pressure level (dB)	Measured value (dB)	Deviated value ^[1] (dB)	Acceptance limit ^[3] (dB)
1000	94	94.25	0.25	±0.40
1000	114	114.39	0.39	±0.40

2. Function : Frequency

Norminal Sound Pressure level (dB)	Specified Frequency (Hz)	Measured value (Hz)	Deviated value ^[2] (%)	Acceptance limit ^[3] (%)
94	1000	970.1	-2.99	±1.7
114	1000	967.4	-3.26	±1.7

Certificate No.: CP20230151EA

Calibration Report

3. Function : Total distortion + noise

Norminal Sound Pressure level (dB)	Norminal Frequency (Hz)	Measured value ^[4] (%)	Acceptance limit ^[5] (%)
94	1000	1.7	3.0
114	1000	1.4	3.0


Uncertainty of measurement

Function	Uncertainty	Maximum-permitted uncertainty of measurement
Sound pressure level	0.10 dB	0.35 dB
Frequency	0.10 %	0.20 %
Total distortion + noise	0.40 %	1.00 %

- Note:
- [1] The deviated value is the absolute value of the difference between the measured value and the corresponding specified sound pressure level.
 - [2] The deviated value is the absolute value of the difference in percent between the measured value and the corresponding specified frequency.
 - [3] The acceptance limit is for the deviated value.
 - [4] The measured value is the total distortion + noise, measured over the frequency range from 20 Hz to 20 kHz.
 - [5] The acceptance limit is for the Measured value.

Remarks: 1. Acceptance limit was IEC 60942:2017 Class 2.
2. The coverage factor $k = 2.00$

-- End of Report --




APTITECH CALIBRATION CO., LTD.

50/40 Moo 5 T. Lat Sawai, A. Lamlukka, Pathumthani 12150

Tel. +66 2103-6290 Fax. +66 2103-6291

Email. sales@aptitech-cal.com



CERTIFICATE OF CALIBRATION

Certificate Number : SC230103

Customer : Pacific Laboratory Co.,Ltd.
Address : 14/5358 Moo 14 Tambol Bang Bua Thong, Amphoe Bang Thong,
Nonthaburi 11110

Description	: Sound Level Meter	W/O Number	: SC230103
Manufacturer	: ACO	Calibration Location	: Laboratory
Model	: 6236	Ambient Temperature	: 22 ± 2 °C
Serial Number	: 222166	Ambient Humidity	: 55 ± 15 %RH
ID. Number	: N/A	Received Date	: 24-Feb-23

This certifies that the above instrument was calibrated in compliance with the Calibration Systems Requirement of ISO/IEC 17025:2017 in accordance with referenced procedures. Standards used to perform this calibration are certified by or traceable to National Institute of Metrology (Thailand) and/or other recognized national measurement institutes which realizes the units of measurement according to the International System of Units (SI Unit).

Measurement uncertainties at the time of test are given where applicable. They are calculated in accordance with the method described in The Expression of Uncertainty and Confidence in Measurement (M3003).

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor $k=2$ such that the coverage probability corresponds to approximately 95%. This result of calibration was found accurate as shown on date and place of calibration only.

Standard Equipments

Description	Serial No.	Certificate No.	Traceability	Due Date
Sound Level Calibrator	141011576	CP20220161EA	TISI: 22-LB0125	13-May-24

APTITECH

Authority of Calibration

☐ Mr. [Signature]
☒ Mr. [Signature]

Calibration Date : 27-Feb-2023
Issued Date : 2-Mar-2023
Calibrated By : Mr. Sataporn Petnoi

Calibration certificates without signatures are not valid. This certificate applied to only the item identified and shall not be reproduced other than in full, without the specific written approval by APTITECH CALIBRATION CO., LTD.

CALIBRATION REPORT

Certificate Number : SC230103

Calibration Method

The Unit Under Calibration (UUC) was calibrated by comparison measurement with sound level calibrator. The calibration has been accomplished in an ambient environment controlled, base on the in-house calibration procedure. The identification of the laboratory's calibration procedure employed are CP-7.2-01-107

Calibration Results

Appearance and function of use : Good
Results of Calibration : Without any adjustment


Sound Level Measurement (Slow Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.9 dB	0.04 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.9 dB	0.04 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB

Sound Level Measurement (Fast Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.8 dB	-0.06 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.8 dB	-0.06 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB

--- End of Certificate ---




APTITECH CALIBRATION CO., LTD.

50/40 Moo 5 T. Lat Sawai, A. Lamlukka, Pathumthani 12150

Tel. +66 2103-6290 Fax. +66 2103-6291

Email. sales@aptitech-cal.com



CERTIFICATE OF CALIBRATION

Certificate Number : SC230020

Customer : Pacific Laboratory Co.,Ltd.
Address : 14/5358 Moo 14 Tambol Bang Bua Thong, Amphoe Bang Thong,
Nonthaburi 11110

Description	: Sound Level Meter	W/O Number	: SC230020
Manufacturer	: ACO	Calibration Location	: Laboratory
Model	: 6236	Ambient Temperature	: 22 ± 2 °C
Serial Number	: 222019	Ambient Humidity	: 55 ± 15 %RH
ID. Number	: N/A	Received Date	: 04-Jan-23

This certifies that the above instrument was calibrated in compliance with the Calibration Systems Requirement of ISO/IEC 17025:2017 in accordance with referenced procedures. Standards used to perform this calibration are certified by or traceable to National Institute of Metrology (Thailand) and/or other recognized national measurement institutes which realizes the units of measurement according to the International System of Units (SI Unit).

Measurement uncertainties at the time of test are given where applicable. They are calculated in accordance with the method described in The Expression of Uncertainty and Confidence in Measurement (M3003).

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor $k=2$ such that the coverage probability corresponds to approximately 95%. This result of calibration was found accurate as shown on date and place of calibration only.

Standard Equipments

Description	Serial No.	Certificate No.	Traceability	Due Date
Sound Level Calibrator	141011576	CP20220161EA	TISI: 22-LB0125	13-May-24




Authority of Calibration

Calibration Date : 05-Jan-2023
Issued Date : 10-Jan-2023
Calibrated By : Mr. Rattapong Janpanya

☐ Mr. Anuwat Simsiriwat [Laboratory Manager]
☒ Mr. Sompoch Srisunart [Technical Manager]

Calibration certificates without signatures are not valid. This certificate applied to only the item identified and shall not be reproduced other than in full, without the specific written approval by APTITECH CALIBRATION CO., LTD.




APTITECH CALIBRATION CO., LTD.

50/40 Moo 5 T. Lat Sawai, A. Lamlukka, Pathumthani 12150

Tel. +66 2103-6290 Fax. +66 2103-6291

Email. sales@aptitech-cal.com



CALIBRATION REPORT

Certificate Number : SC230020

Calibration Method

The Unit Under Calibration (UUC) was calibrated by comparison measurement with sound level calibrator. The calibration has been accomplished in an ambient environment controlled, base on the in-house calibration procedure. The identification of the laboratory's calibration procedure employed are CP-7.2-01-107

Calibration Results

Appearance and function of use : Good
Results of Calibration : Without any adjustment


Sound Level Measurement (Slow Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.7 dB	-0.16 dB	0.60 dB
		113.87 dB	113.7 dB	-0.17 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.7 dB	-0.16 dB	0.60 dB
		113.87 dB	113.6 dB	-0.27 dB	0.60 dB

Sound Level Measurement (Fast Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.6 dB	-0.26 dB	0.60 dB
		113.87 dB	113.6 dB	-0.27 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.6 dB	-0.26 dB	0.60 dB
		113.87 dB	113.6 dB	-0.27 dB	0.60 dB

--- End of Certificate ---




APTITECH CALIBRATION CO., LTD.

50/40 Moo 5 T. Lat Sawai, A. Lamlukka, Pathumthani 12150

Tel. +66 2103-6290 Fax. +66 2103-6291

Email. sales@aptitech-cal.com



CERTIFICATE OF CALIBRATION

Certificate Number : SC230016

Customer : Pacific Laboratory Co.,Ltd.
Address : 14/5358 Moo 14 Tambol Bang Bua Thong, Amphoe Bang Thong,
Nonthaburi 11110

Description	: Sound Level Meter	W/O Number	: SC230016
Manufacturer	: ACO	Calibration Location	: Laboratory
Model	: 6236	Ambient Temperature	: 22 ± 2 °C
Serial Number	: 222015	Ambient Humidity	: 55 ± 15 %RH
ID. Number	: N/A	Received Date	: 04-Jan-23

This certifies that the above instrument was calibrated in compliance with the Calibration Systems Requirement of ISO/IEC 17025:2017 in accordance with referenced procedures. Standards used to perform this calibration are certified by or traceable to National Institute of Metrology (Thailand) and/or other recognized national measurement institutes which realizes the units of measurement according to the International System of Units (SI Unit).

Measurement uncertainties at the time of test are given where applicable. They are calculated in accordance with the method described in The Expression of Uncertainty and Confidence in Measurement (M3003).

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor $k=2$ such that the coverage probability corresponds to approximately 95%. This result of calibration was found accurate as shown on date and place of calibration only.

Standard Equipments

Description	Serial No.	Certificate No.	Traceability	Due Date
Sound Level Calibrator	141011576	CP20220161EA	TISI: 22-LB0125	13-May-24


APTITECH

Authority of Calibration




Calibration Date : 05-Jan-2023
Issued Date : 10-Jan-2023
Calibrated By : Mr. Rattapong Janpanya

Calibration certificates without signatures are not valid. This certificate applied to only the item identified and shall not be reproduced other than in full, without the specific written approval by APTITECH CALIBRATION CO., LTD.



APTITECH CALIBRATION CO., LTD.

50/40 Moo 5 T. Lat Sawai, A. Lamlukka, Pathumthani 12150
Tel. +66 2103-6290 Fax. +66 2103-6291
Email. sales@aptitech-cal.com



CALIBRATION REPORT

Certificate Number : SC230016

Calibration Method

The Unit Under Calibration (UUC) was calibrated by comparison measurement with sound level calibrator. The calibration has been accomplished in an ambient environment controlled, base on the in-house calibration procedure. The identification of the laboratory's calibration procedure employed are CP-7.2-01-107

Calibration Results

Appearance and function of use : Good
Results of Calibration : Without any adjustment


Sound Level Measurement (Slow Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.6 dB	-0.26 dB	0.60 dB
		113.87 dB	113.7 dB	-0.17 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.6 dB	-0.26 dB	0.60 dB
		113.87 dB	113.6 dB	-0.27 dB	0.60 dB

Sound Level Measurement (Fast Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.7 dB	-0.16 dB	0.60 dB
		113.87 dB	113.7 dB	-0.17 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.7 dB	-0.16 dB	0.60 dB
		113.87 dB	113.7 dB	-0.17 dB	0.60 dB

--- End of Certificate ---




APTITECH CALIBRATION CO., LTD.

50/40 Moo 5 T. Lat Sawai, A. Lamlukka, Pathumthani 12150

Tel. +66 2103-6290 Fax. +66 2103-6291

Email. sales@aptitech-cal.com



CERTIFICATE OF CALIBRATION

Certificate Number : SC230024

Customer : Pacific Laboratory Co.,Ltd.
Address : 14/5358 Moo 14 Tambol Bang Bua Thong, Amphoe Bang Thong,
Nonthaburi 11110

Description	: Sound Level Meter	W/O Number	: SC230024
Manufacturer	: ACO	Calibration Location	: Laboratory
Model	: 6236	Ambient Temperature	: 22 ± 2 °C
Serial Number	: 222041	Ambient Humidity	: 55 ± 15 %RH
ID. Number	: N/A	Received Date	: 04-Jan-23

This certifies that the above instrument was calibrated in compliance with the Calibration Systems Requirement of ISO/IEC 17025:2017 in accordance with referenced procedures. Standards used to perform this calibration are certified by or traceable to National Institute of Metrology (Thailand) and/or other recognized national measurement institutes which realizes the units of measurement according to the International System of Units (SI Unit).

Measurement uncertainties at the time of test are given where applicable. They are calculated in accordance with the method described in The Expression of Uncertainty and Confidence in Measurement (M3003).

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor $k=2$ such that the coverage probability corresponds to approximately 95%. This result of calibration was found accurate as shown on date and place of calibration only.

Standard Equipments

Description	Serial No.	Certificate No.	Traceability	Due Date
Sound Level Calibrator	141011576	CP20220161EA	TISI: 22-LB0125	13-May-24

APTITECH


Authority of Calibration



☒ Mr. Sompoch Srisunart [Technical Manager]


Calibration Date : 05-Jan-2023
Issued Date : 10-Jan-2023
Calibrated By : Mr. Rattapong Janpanya

Calibration certificates without signatures are not valid. This certificate applied to only the item identified and shall not be reproduced other than in full, without the specific written approval by APTITECH CALIBRATION CO., LTD.



APTITECH CALIBRATION CO., LTD.

50/40 Moo 5 T. Lat Sawai, A. Lamlukka, Pathumthani 12150
Tel. +66 2103-6290 Fax. +66 2103-6291
Email. sales@aptitech-cal.com



CALIBRATION REPORT

Certificate Number : SC230024

Calibration Method

The Unit Under Calibration (UUC) was calibrated by comparison measurement with sound level calibrator. The calibration has been accomplished in an ambient environment controlled, base on the in-house calibration procedure. The identification of the laboratory's calibration procedure employed are CP-7.2-01-107

Calibration Results

Appearance and function of use : Good
Results of Calibration : Without any adjustment


Sound Level Measurement (Slow Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.7 dB	-0.16 dB	0.60 dB
		113.87 dB	113.6 dB	-0.27 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.7 dB	-0.16 dB	0.60 dB
		113.87 dB	113.6 dB	-0.27 dB	0.60 dB

Sound Level Measurement (Fast Mode)


Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.6 dB	-0.26 dB	0.60 dB
		113.87 dB	113.6 dB	-0.27 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.6 dB	-0.26 dB	0.60 dB
		113.87 dB	113.6 dB	-0.27 dB	0.60 dB

--- End of Certificate ---



APTITECH CALIBRATION CO., LTD.

50/40 Moo 5 T. Lat Sawai, A. Lamlukka, Pathumthani 12150
Tel. +66 2103-6290 Fax. +66 2103-6291
Email. sales@aptitech-cal.com



CERTIFICATE OF CALIBRATION

Certificate Number : SC230112

Customer : Pacific Laboratory Co.,Ltd.
Address : 14/5358 Moo 14 Tambol Bang Bua Thong, Amphoe Bang Thong,
Nonthaburi 11110

Description	: Sound Level Meter	W/O Number	: SC230112
Manufacturer	: ACO	Calibration Location	: Laboratory
Model	: 6236	Ambient Temperature	: 22 ± 2 °C
Serial Number	: 222177	Ambient Humidity	: 55 ± 15 %RH
ID. Number	: N/A	Received Date	: 24-Feb-23

This certifies that the above instrument was calibrated in compliance with the Calibration Systems Requirement of ISO/IEC 17025:2017 in accordance with referenced procedures. Standards used to perform this calibration are certified by or traceable to National Institute of Metrology (Thailand) and/or other recognized national measurement institutes which realizes the units of measurement according to the International System of Units (SI Unit).

Measurement uncertainties at the time of test are given where applicable. They are calculated in accordance with the method described in The Expression of Uncertainty and Confidence in Measurement (M3003).

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor $k=2$ such that the coverage probability corresponds to approximately 95%. This result of calibration was found accurate as shown on date and place of calibration only.

Standard Equipments

Description	Serial No.	Certificate No.	Traceability	Due Date
Sound Level Calibrator	141011576	CP20220161EA	TISI: 22-LB0125	13-May-24



Authority of Calibration



☒ Mr. Sompoch Srisunart [Technical Manager]

Calibration Date : 27-Feb-2023
Issued Date : 2-Mar-2023
Calibrated By : Mr. Sataporn Petnoi

Calibration certificates without signatures are not valid. This certificate applied to only the item identified and shall not be reproduced other than in full, without the specific written approval by APTITECH CALIBRATION CO., LTD.

CALIBRATION REPORT

Certificate Number : SC230112

Calibration Method

The Unit Under Calibration (UUC) was calibrated by comparison measurement with sound level calibrator. The calibration has been accomplished in an ambient environment controlled, base on the in-house calibration procedure. The identification of the laboratory's calibration procedure employed are CP-7.2-01-107

Calibration Results

Appearance and function of use : Good
Results of Calibration : Without any adjustment


Sound Level Measurement (Slow Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.8 dB	-0.06 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.8 dB	-0.06 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB

Sound Level Measurement (Fast Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.8 dB	-0.06 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.8 dB	-0.06 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB

--- End of Certificate ---




APTITECH CALIBRATION CO., LTD.

50/40 Moo 5 T. Lat Sawai, A. Lamlukka, Pathumthani 12150

Tel. +66 2103-6290 Fax. +66 2103-6291

Email. sales@aptitech-cal.com



CERTIFICATE OF CALIBRATION

Certificate Number : SC230106

Customer : Pacific Laboratory Co.,Ltd.
Address : 14/5358 Moo 14 Tambol Bang Bua Thong, Amphoe Bang Thong,
Nonthaburi 11110

Description	: Sound Level Meter	W/O Number	: SC230106
Manufacturer	: ACO	Calibration Location	: Laboratory
Model	: 6236	Ambient Temperature	: 22 ± 2 °C
Serial Number	: 222169	Ambient Humidity	: 55 ± 15 %RH
ID. Number	: N/A	Received Date	: 24-Feb-23

This certifies that the above instrument was calibrated in compliance with the Calibration Systems Requirement of ISO/IEC 17025:2017 in accordance with referenced procedures. Standards used to perform this calibration are certified by or traceable to National Institute of Metrology (Thailand) and/or other recognized national measurement institutes which realizes the units of measurement according to the International System of Units (SI Unit).

Measurement uncertainties at the time of test are given where applicable. They are calculated in accordance with the method described in The Expression of Uncertainty and Confidence in Measurement (M3003).

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor $k=2$ such that the coverage probability corresponds to approximately 95%. This result of calibration was found accurate as shown on date and place of calibration only.

Standard Equipments

Description	Serial No.	Certificate No.	Traceability	Due Date
Sound Level Calibrator	141011576	CP20220161EA	TISI: 22-LB0125	13-May-24

APTITECH

Authority of Calibration



Calibration Date : 27-Feb-2023
Issued Date : 2-Mar-2023
Calibrated By : Mr. Sataporn Petnoi

- ☐ Mr. Anuwat Simsiriwat [Laboratory Manager]
☒ Mr. Sompoch Srisunart [Technical Manager]

Calibration certificates without signatures are not valid. This certificate applied to only the item identified and shall not be reproduced other than in full, without the specific written approval by APTITECH CALIBRATION CO., LTD.

CALIBRATION REPORT

Certificate Number : SC230106

Calibration Method

The Unit Under Calibration (UUC) was calibrated by comparison measurement with sound level calibrator. The calibration has been accomplished in an ambient environment controlled, base on the in-house calibration procedure. The identification of the laboratory's calibration procedure employed are CP-7.2-01-107

Calibration Results

Appearance and function of use : Good
Results of Calibration : Without any adjustment


Sound Level Measurement (Slow Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.8 dB	-0.06 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.8 dB	-0.06 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB

Sound Level Measurement (Fast Mode)


Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.7 dB	-0.16 dB	0.60 dB
		113.87 dB	113.7 dB	-0.17 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.7 dB	-0.16 dB	0.60 dB
		113.87 dB	113.7 dB	-0.17 dB	0.60 dB

--- End of Certificate ---



APTITECH CALIBRATION CO., LTD.

50/40 Moo 5 T. Lat Sawai, A. Lamlukka, Pathumthani 12150
Tel. +66 2103-6290 Fax. +66 2103-6291
Email. sales@aptitech-cal.com



CERTIFICATE OF CALIBRATION

Certificate Number : SC230034

Customer : Pacific Laboratory Co.,Ltd.
Address : 14/5358 Moo 14 Tambol Bang Bua Thong, Amphoe Bang Thong,
Nonthaburi 11110

Description	: Sound Level Meter	W/O Number	: SC230034
Manufacturer	: ACO	Calibration Location	: Laboratory
Model	: 6236	Ambient Temperature	: 22 ± 2 °C
Serial Number	: 222042	Ambient Humidity	: 55 ± 15 %RH
ID. Number	: N/A	Received Date	: 13-Jan-23

This certifies that the above instrument was calibrated in compliance with the Calibration Systems Requirement of ISO/IEC 17025:2017 in accordance with referenced procedures. Standards used to perform this calibration are certified by or traceable to National Institute of Metrology (Thailand) and/or other recognized national measurement institutes which realizes the units of measurement according to the International System of Units (SI Unit).

Measurement uncertainties at the time of test are given where applicable. They are calculated in accordance with the method described in The Expression of Uncertainty and Confidence in Measurement (M3003).

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor $k=2$ such that the coverage probability corresponds to approximately 95%. This result of calibration was found accurate as shown on date and place of calibration only.

Standard Equipments

Description	Serial No.	Certificate No.	Traceability	Due Date
Sound Level Calibrator	141011576	CP20220161EA	TISI: 22-LB0125	13-May-24



Authority of Calibration



☐ [Signature]
☒ Mr. Sompoch Srisunart [Technical Manager]

Calibration Date : 15-Jan-2023
Issued Date : 20-Jan-2023
Calibrated By : Mr. Sataporn Petnoi

Calibration certificates without signatures are not valid. This certificate applied to only the item identified and shall not be reproduced other than in full, without the specific written approval by APTITECH CALIBRATION CO., LTD.

CALIBRATION REPORT

Certificate Number : SC230034

Calibration Method

The Unit Under Calibration (UUC) was calibrated by comparison measurement with sound level calibrator. The calibration has been accomplished in an ambient environment controlled, base on the in-house calibration procedure. The identification of the laboratory's calibration procedure employed are CP-7.2-01-107

Calibration Results

Appearance and function of use : Good
Results of Calibration : Without any adjustment


Sound Level Measurement (Slow Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	94.00 dB	93.8 dB	-0.20 dB	0.60 dB
		114.00 dB	113.8 dB	-0.20 dB	0.60 dB
LC	40-130 dB	94.00 dB	93.8 dB	-0.20 dB	0.60 dB
		114.00 dB	113.8 dB	-0.20 dB	0.60 dB

Sound Level Measurement (Fast Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	94.00 dB	93.8 dB	-0.20 dB	0.60 dB
		114.00 dB	113.8 dB	-0.20 dB	0.60 dB
LC	40-130 dB	94.00 dB	93.7 dB	-0.30 dB	0.60 dB
		114.00 dB	113.7 dB	-0.30 dB	0.60 dB

--- End of Certificate ---




APTITECH CALIBRATION CO., LTD.

50/40 Moo 5 T. Lat Sawai, A. Lamlukka, Pathumthani 12150

Tel. +66 2103-6290 Fax. +66 2103-6291

Email: sales@aptitech-cal.com



CERTIFICATE OF CALIBRATION

Certificate Number : SC230107

Customer : Pacific Laboratory Co.,Ltd.
Address : 14/5358 Moo 14 Tambol Bang Bua Thong, Amphoe Bang Thong,
Nonthaburi 11110

Description	: Sound Level Meter	W/O Number	: SC230107
Manufacturer	: ACO	Calibration Location	: Laboratory
Model	: 6236	Ambient Temperature	: 22 ± 2 °C
Serial Number	: 222170	Ambient Humidity	: 55 ± 15 %RH
ID. Number	: N/A	Received Date	: 24-Feb-23

This certifies that the above instrument was calibrated in compliance with the Calibration Systems Requirement of ISO/IEC 17025:2017 in accordance with referenced procedures. Standards used to perform this calibration are certified by or traceable to National Institute of Metrology (Thailand) and/or other recognized national measurement institutes which realizes the units of measurement according to the International System of Units (SI Unit).

Measurement uncertainties at the time of test are given where applicable. They are calculated in accordance with the method described in The Expression of Uncertainty and Confidence in Measurement (M3003).

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor $k=2$ such that the coverage probability corresponds to approximately 95%. This result of calibration was found accurate as shown on date and place of calibration only.

Standard Equipments

Description	Serial No.	Certificate No.	Traceability	Due Date
Sound Level Calibrator	141011576	CP20220161EA	TISI: 22-LB0125	13-May-24

APTITECH

Authority of Calibration



☒ Mr. Sompoch Srisunart [Technical Manager]

Calibration Date : 27-Feb-2023
Issued Date : 2-Mar-2023
Calibrated By : Mr. Sataporn Petnoi

Calibration certificates without signatures are not valid. This certificate applied to only the item identified and shall not be reproduced other than in full, without the specific written approval by APTITECH CALIBRATION CO., LTD.

CALIBRATION REPORT

Certificate Number : SC230107

Calibration Method

The Unit Under Calibration (UUC) was calibrated by comparison measurement with sound level calibrator. The calibration has been accomplished in an ambient environment controlled, base on the in-house calibration procedure. The identification of the laboratory's calibration procedure employed are CP-7.2-01-107

Calibration Results

Appearance and function of use : Good
Results of Calibration : Without any adjustment


Sound Level Measurement (Slow Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.8 dB	-0.06 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.8 dB	-0.06 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB

Sound Level Measurement (Fast Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.8 dB	-0.06 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.8 dB	-0.06 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB

--- End of Certificate ---




APTITECH CALIBRATION CO., LTD.

50/40 Moo 5 T. Lat Sawai, A. Lamlukka, Pathumthani 12150

Tel. +66 2103-6290 Fax. +66 2103-6291

Email. sales@aptitech-cal.com



CERTIFICATE OF CALIBRATION

Certificate Number : SC230021

Customer : Pacific Laboratory Co.,Ltd.
Address : 14/5358 Moo 14 Tambol Bang Bua Thong, Amphoe Bang Thong,
Nonthaburi 11110

Description	: Sound Level Meter	W/O Number	: SC230021
Manufacturer	: ACO	Calibration Location	: Laboratory
Model	: 6236	Ambient Temperature	: 22 ± 2 °C
Serial Number	: 222020	Ambient Humidity	: 55 ± 15 %RH
ID. Number	: N/A	Received Date	: 04-Jan-23

This certifies that the above instrument was calibrated in compliance with the Calibration Systems Requirement of ISO/IEC 17025:2017 in accordance with referenced procedures. Standards used to perform this calibration are certified by or traceable to National Institute of Metrology (Thailand) and/or other recognized national measurement institutes which realizes the units of measurement according to the International System of Units (SI Unit).

Measurement uncertainties at the time of test are given where applicable. They are calculated in accordance with the method described in The Expression of Uncertainty and Confidence in Measurement (M3003).

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor $k=2$ such that the coverage probability corresponds to approximately 95%. This result of calibration was found accurate as shown on date and place of calibration only.

Standard Equipments

Description	Serial No.	Certificate No.	Traceability	Due Date
Sound Level Calibrator	141011576	CP20220161EA	TISI: 22-LB0125	13-May-24




Authority of Calibration



☒ Mr. Sompoch Srisunart [Technical Manager]


Calibration Date : 05-Jan-2023
Issued Date : 10-Jan-2023
Calibrated By : Mr. Rattapong Janpanya

Calibration certificates without signatures are not valid. This certificate applied to only the item identified and shall not be reproduced other than in full, without the specific written approval by APTITECH CALIBRATION CO., LTD.



APTITECH CALIBRATION CO., LTD.

50/40 Moo 5 T. Lat Sawai, A. Lamlukka, Pathumthani 12150
Tel. +66 2103-6290 Fax. +66 2103-6291
Email. sales@aptitech-cal.com



CALIBRATION REPORT

Certificate Number : SC230021

Calibration Method

The Unit Under Calibration (UUC) was calibrated by comparison measurement with sound level calibrator. The calibration has been accomplished in an ambient environment controlled, base on the in-house calibration procedure. The identification of the laboratory's calibration procedure employed are CP-7.2-01-107

Calibration Results

Appearance and function of use : Good
Results of Calibration : Without any adjustment


Sound Level Measurement (Slow Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.7 dB	-0.16 dB	0.60 dB
		113.87 dB	113.6 dB	-0.27 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.6 dB	-0.26 dB	0.60 dB
		113.87 dB	113.6 dB	-0.27 dB	0.60 dB

Sound Level Measurement (Fast Mode)


Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.6 dB	-0.26 dB	0.60 dB
		113.87 dB	113.6 dB	-0.27 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.6 dB	-0.26 dB	0.60 dB
		113.87 dB	113.5 dB	-0.37 dB	0.60 dB

--- End of Certificate ---



APTITECH CALIBRATION CO., LTD.

50/40 Moo 5 T. Lat Sawai, A. Lamlukka, Pathumthani 12150
Tel. +66 2103-6290 Fax. +66 2103-6291
Email. sales@aptitech-cal.com



CERTIFICATE OF CALIBRATION

Certificate Number : SC230109

Customer : Pacific Laboratory Co.,Ltd.
Address : 14/5358 Moo 14 Tambol Bang Bua Thong, Amphoe Bang Thong,
Nonthaburi 11110

Description	: Sound Level Meter	W/O Number	: SC230109
Manufacturer	: ACO	Calibration Location	: Laboratory
Model	: 6236	Ambient Temperature	: 22 ± 2 °C
Serial Number	: 222173	Ambient Humidity	: 55 ± 15 %RH
ID. Number	: N/A	Received Date	: 24-Feb-23

This certifies that the above instrument was calibrated in compliance with the Calibration Systems Requirement of ISO/IEC 17025:2017 in accordance with referenced procedures. Standards used to perform this calibration are certified by or traceable to National Institute of Metrology (Thailand) and/or other recognized national measurement institutes which realizes the units of measurement according to the International System of Units (SI Unit).

Measurement uncertainties at the time of test are given where applicable. They are calculated in accordance with the method described in The Expression of Uncertainty and Confidence in Measurement (M3003).

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor $k=2$ such that the coverage probability corresponds to approximately 95%. This result of calibration was found accurate as shown on date and place of calibration only.

Standard Equipments

Description	Serial No.	Certificate No.	Traceability	Due Date
Sound Level Calibrator	141011576	CP20220161EA	TISI: 22-LB0125	13-May-24



Authority of Calibration

Approved Signatory

Calibration Date : 27-Feb-2023

Issued Date : 2-Mar-2023

Calibrated By : Mr. Sataporn Petnoi

- ☐ Mr. Anuwat Simsiriwat [Laboratory Manager]
☒ Mr. Sompoch Srisunart [Technical Manager]

Calibration certificates without signatures are not valid. This certificate applied to only the item identified and shall not be reproduced other than in full, without the specific written approval by APTITECH CALIBRATION CO., LTD.

CALIBRATION REPORT

Certificate Number : SC230109

Calibration Method

The Unit Under Calibration (UUC) was calibrated by comparison measurement with sound level calibrator. The calibration has been accomplished in an ambient environment controlled, base on the in-house calibration procedure. The identification of the laboratory's calibration procedure employed are CP-7.2-01-107

Calibration Results

Appearance and function of use : Good
Results of Calibration : Without any adjustment


Sound Level Measurement (Slow Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.8 dB	-0.06 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.8 dB	-0.06 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB

Sound Level Measurement (Fast Mode)


Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.8 dB	-0.06 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.8 dB	-0.06 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB

--- End of Certificate ---



APTITECH CALIBRATION CO., LTD.

50/40 Moo 5 T. Lat Sawai, A. Lamlukka, Pathumthani 12150
Tel. +66 2103-6290 Fax. +66 2103-6291
Email. sales@aptitech-cal.com



CERTIFICATE OF CALIBRATION

Certificate Number : SC230036

Customer : Pacific Laboratory Co.,Ltd.
Address : 14/5358 Moo 14 Tambol Bang Bua Thong, Amphoe Bang Thong,
Nonthaburi 11110

Description	: Sound Level Meter	W/O Number	: SC230036
Manufacturer	: ACO	Calibration Location	: Laboratory
Model	: 6236	Ambient Temperature	: 22 ± 2 °C
Serial Number	: 222045	Ambient Humidity	: 55 ± 15 %RH
ID. Number	: N/A	Received Date	: 13-Jan-23

This certifies that the above instrument was calibrated in compliance with the Calibration Systems Requirement of ISO/IEC 17025:2017 in accordance with referenced procedures. Standards used to perform this calibration are certified by or traceable to National Institute of Metrology (Thailand) and/or other recognized national measurement institutes which realizes the units of measurement according to the International System of Units (SI Unit).

Measurement uncertainties at the time of test are given where applicable. They are calculated in accordance with the method described in The Expression of Uncertainty and Confidence in Measurement (M3003).

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor $k=2$ such that the coverage probability corresponds to approximately 95%. This result of calibration was found accurate as shown on date and place of calibration only.

Standard Equipments

Description	Serial No.	Certificate No.	Traceability	Due Date
Sound Level Calibrator	141011576	CP20220161EA	TISI: 22-LB0125	13-May-24


APTITECH

Authority of Calibration

☐ [Signature] [Name]
☒ Mr. Sompoch Srisunart [Technical Manager]

Calibration Date : 15-Jan-2023
Issued Date : 20-Jan-2023
Calibrated By : Mr. Sataporn Petnoi

Calibration certificates without signatures are not valid. This certificate applied to only the item identified and shall not be reproduced other than in full, without the specific written approval by APTITECH CALIBRATION CO., LTD.




APTITECH CALIBRATION CO., LTD.

50/40 Moo 5 T. Lat Sawai, A. Lamlukka, Pathumthani 12150

Tel. +66 2103-6290 Fax. +66 2103-6291

Email. sales@aptitech-cal.com



CALIBRATION REPORT

Certificate Number : SC230036

Calibration Method

The Unit Under Calibration (UUC) was calibrated by comparison measurement with sound level calibrator. The calibration has been accomplished in an ambient environment controlled, base on the in-house calibration procedure. The identification of the laboratory's calibration procedure employed are CP-7.2-01-107

Calibration Results

Appearance and function of use : Good
Results of Calibration : Without any adjustment


Sound Level Measurement (Slow Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	94.00 dB	93.7 dB	-0.30 dB	0.60 dB
		114.00 dB	113.6 dB	-0.40 dB	0.60 dB
LC	40-130 dB	94.00 dB	93.7 dB	-0.30 dB	0.60 dB
		114.00 dB	113.8 dB	-0.20 dB	0.60 dB

Sound Level Measurement (Fast Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	94.00 dB	93.7 dB	-0.30 dB	0.60 dB
		114.00 dB	113.7 dB	-0.30 dB	0.60 dB
LC	40-130 dB	94.00 dB	93.7 dB	-0.30 dB	0.60 dB
		114.00 dB	113.7 dB	-0.30 dB	0.60 dB

--- End of Certificate ---




APTITECH CALIBRATION CO., LTD.

50/40 Moo 5 T. Lat Sawai, A. Lamlukka, Pathumthani 12150

Tel. +66 2103-6290 Fax. +66 2103-6291

Email. sales@aptitech-cal.com



CERTIFICATE OF CALIBRATION

Certificate Number : SC230099

Customer : Pacific Laboratory Co.,Ltd.
Address : 14/5358 Moo 14 Tambol Bang Bua Thong, Amphoe Bang Thong,
Nonthaburi 11110

Description	: Sound Level Meter	W/O Number	: SC230099
Manufacturer	: ACO	Calibration Location	: Laboratory
Model	: 6236	Ambient Temperature	: 22 ± 2 °C
Serial Number	: 222135	Ambient Humidity	: 55 ± 15 %RH
ID. Number	: N/A	Received Date	: 24-Feb-23

This certifies that the above instrument was calibrated in compliance with the Calibration Systems Requirement of ISO/IEC 17025:2017 in accordance with referenced procedures. Standards used to perform this calibration are certified by or traceable to National Institute of Metrology (Thailand) and/or other recognized national measurement institutes which realizes the units of measurement according to the International System of Units (SI Unit).

Measurement uncertainties at the time of test are given where applicable. They are calculated in accordance with the method described in The Expression of Uncertainty and Confidence in Measurement (M3003).

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor $k=2$ such that the coverage probability corresponds to approximately 95%. This result of calibration was found accurate as shown on date and place of calibration only.

Standard Equipments

Description	Serial No.	Certificate No.	Traceability	Due Date
Sound Level Calibrator	141011576	CP20220161EA	TISI: 22-LB0125	13-May-24

APTITECH

Authority of Calibration



☒ Mr. Sompoch Srisunart [Technical Manager]

Calibration Date : 27-Feb-2023

Issued Date : 2-Mar-2023

Calibrated By : Mr. Sataporn Petnoi

Calibration certificates without signatures are not valid. This certificate applied to only the item identified and shall not be reproduced other than in full, without the specific written approval by APTITECH CALIBRATION CO., LTD.

CALIBRATION REPORT

Certificate Number : SC230099

Calibration Method

The Unit Under Calibration (UUC) was calibrated by comparison measurement with sound level calibrator. The calibration has been accomplished in an ambient environment controlled, base on the in-house calibration procedure. The identification of the laboratory's calibration procedure employed are CP-7.2-01-107

Calibration Results

Appearance and function of use : Good
Results of Calibration : Without any adjustment

Sound Level Measurement (Slow Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.8 dB	-0.06 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.8 dB	-0.06 dB	0.60 dB
		113.87 dB	113.8 dB	-0.07 dB	0.60 dB

Sound Level Measurement (Fast Mode)

Function	UUC Range	Standard Value	UUC Reading	UUC Error	(±) Uncertainty
LA	40-130 dB	93.86 dB	93.7 dB	-0.16 dB	0.60 dB
		113.87 dB	113.7 dB	-0.17 dB	0.60 dB
LC	40-130 dB	93.86 dB	93.7 dB	-0.16 dB	0.60 dB
		113.87 dB	113.7 dB	-0.17 dB	0.60 dB

--- End of Certificate ---