



บริษัท ปตท. น้ำมันและการค้าปลีก จำกัด (มหาชน)

รายงานผลการปฏิบัติตามมาตรการป้องกันและแก้ไขผลกระทบสิ่งแวดล้อม
และมาตรการติดตามตรวจสอบผลกระทบสิ่งแวดล้อม

โครงการทำเทียบเรือขนถ่ายน้ำมันและก๊าซปิโตรเลียมเหลวสุราษฎร์ธานี แห่งที่ 2

ระหว่างเดือนกรกฎาคมถึงธันวาคม พ.ศ. 2566

ภาคผนวก จ

ใบรับรองการสอบเทียบเครื่องมือ



Southern Calibration Service Co., Ltd.
669/35 Karmjanavanit Rd., Banpru, Hatyai, Songkla 90250 Thailand
Tel: 08 1599 0417 Fax: 0 7480 5133 Email: s.calibration@gmail.com www.scal-lab.com



CALIBRATION CERTIFICATE

Issued Date : 1-May-2023

Certificate No. : 23CH0203

CSR No. : A088/04367

Page : 1 of 2

Customer : ALS Laboratory Group (Thailand) Co., Ltd
114/1 Moo 8, Kamchanawanich Rd. Tambon, Ban Phru,
Amphoe Hat Yai, Songkhla, 90250

Calibration Place : Chemical Laboratory
Instrument Name : pH meter
Manufacturer : Mettler Toledo
Model : S220
Serial No. : B825631849
ID No. : SGK_CL0030
Electrode No. : 1204613
Received Date : 28-Apr-2023
Calibrated Date : 28-Apr-2023
Ambient Temperature : (25 ± 3) °C
Relative Humidity : (55 ± 15) %

REVIEW BY : Ananta B.
APPROVED BY : Kantha H.
NEXT CAL. DATE : 28 / 10 / 2024

Calibration Method Used :

This instrument was calibrated using the Calibration In - house method : SCAL.WI.008 based on direct measurement by using certified reference Material (CRM)

The Southern Calibration Service Co., Ltd. calibration control system complies with requirement of ISO/IEC 17025:2017

Traceability of measurement :

This Certificate is traceable to the International and/or national standards which realize the units of measurement according to the International System of Unit (SI) through :

- HACH : HACH LANGE GmbH
- WK : WK Electric Co., Ltd.

- SCAL : Southern Calibration Service Co., Ltd.

Calibrated by : Alisara Ma

Approved by :

Imon Rattanyalum / Technical Manager

The uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written approval of Southern Calibration Service Co., Ltd.



SCAL

Details of Calibration

1. Reference Standard Equipment Used:

Equipment	Model	Serial No.	Cert. no.	Due Date
Standard Solution	4.005	C02994	1777	5-Sep-2024
Standard Solution	7.000	C03007	1787	17-Oct-2024
Standard Solution	10.012	C02953	1735	29-Apr-2024
Temperature/Electrical Calibrator	MC2-TE	14987	WK2106-299-223	5-Jun-2024
Digital Thermometer With Sensor	DP-77	1360896	22SDTH005	8-Aug-2023

2. The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the longterm stability of instrument.

3. This certificate is not certified any commercial transaction

4. Condition of Item : normal condition , no indication for any damage or malfunction

Result of Calibration :

1. Electrical Measurement

Applied Voltage (mV)	pH meter Reading		Correction (mV)	Uncertainty (± mV)
	(mV)	(pH)		
177.48	177.5	3.70	-0.02	0.17
0.00	0.0	6.70	0.00	0.13
-177.48	-177.4	9.80	-0.08	0.17

2. Before Sample Test Measurement

Standard Buffer Solutions (pH)	pH meter Reading		Correction (pH)	Uncertainty (± pH)
	(pH)	(mV)		
4.005	3.97	159.2	0.035	0.0090
6.999	6.98	-15.4	0.019	0.013
10.012	9.95	-188.0	0.062	0.036

3. After Sample Test Measurement

Standard Buffer Solutions (pH)	pH meter Reading		Correction (pH)	Uncertainty (± pH)
	(pH)	(mV)		
4.005	3.97	158.9	0.035	0.0090
6.999	7.01	-17.5	-0.011	0.013
10.012	9.98	-187.6	0.032	0.036

4. Temperature Measurement

Cal Point (°C)	Standard Temperature (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
25	25.032	25.1	-0.068	0.11

The report uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

... End ...

Southern Calibration Service Co., Ltd.
669/35 Karmjanavanit Rd., Banpru, Hatyai, Songkla 90250 Thailand
Tel: 08 1599 0417 Fax: 0 7480 5133 Email: s.calibration@gmail.com www.scal-lab.com



CALIBRATION CERTIFICATE

Issued Date : 1-May-2023

Certificate No. : 23CH0203

CSR No. : A088/04367

Page : 1 of 2

Customer : ALS Laboratory Group (Thailand) Co., Ltd
114/1 Moo 8, Kamchanawanich Rd. Tambon, Ban Phru,
Amphoe Hat Yai, Songkhla, 90250

Calibration Place : Chemical Laboratory
Instrument Name : pH meter
Manufacturer : Mettler Toledo
Model : S220
Serial No. : B825631849
ID No. : SGK_CL0030
Electrode No. : 1204613
Received Date : 28-Apr-2023
Calibrated Date : 28-Apr-2023
Ambient Temperature : (25 ± 3) °C
Relative Humidity : (55 ± 15) %

REVIEW BY : Ananta B.
APPROVED BY : Kantha H.
NEXT CAL. DATE : 28 / 10 / 2024

Calibration Method Used :

This instrument was calibrated using the Calibration In - house method : SCAL.WI.008 based on direct measurement by using certified reference Material (CRM)

The Southern Calibration Service Co., Ltd. calibration control system complies with requirement of ISO/IEC 17025:2017

Traceability of measurement :

This Certificate is traceable to the International and/or national standards which realize the units of measurement according to the International System of Unit (SI) through :

- HACH : HACH LANGE GmbH
- WK : WK Electric Co., Ltd.

- SCAL : Southern Calibration Service Co., Ltd.

Calibrated by : Alisara Ma

Approved by :

Imon Rattanyalum / Technical Manager

The uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written approval of Southern Calibration Service Co., Ltd.



REVIEW BY Ananta B.
APPROVED BY Kanitta H.
NEXT CAL. DATE 35/11/24

NSC-TIS-TIS 17025
CALIBRATION 0426

Certificate of Calibration

Model Number : MSE224S-100-DU
Description : Analytical Balance
Serial Number : 0034705158
ID No. : SGK_CL0045
Manufacturer : Sartorius

Certificate No. : 23BCI0044
Issued Date : Friday, January 27, 2023
Reference No. : 202361
Page No. : 2 of 2

Calibration Results : Without Adjustment

Repeatability
The reproducibility is the ability of a weighing instrument to display nearly identical readouts under constant test conditions when the same load within a measurement series is placed repeatedly on the weighing pan in the same manner. The standard deviation is used to express reproducibility quantitatively.

Nominal Value : (Low Load)	20.0000	200.0000
20 g	20.0001	200.0000
Tolerance	20.0000	200.0000
0.0001 g	20.0000	200.0000
Nominal Value : (High Load)	20.0000	200.0001
200 g	20.0000	200.0001
Tolerance	20.0000	200.0000
0.0001 g	20.0000	200.0001
Standard Deviation	0.00003	0.00005

Eccentricity (Off-center loading error)
The off-center loading error is yielded by the difference between the readout of the load, i.e. 1/3 or 1/4 of maximum capacity, placed in the middle of the weighing pan and between each of four additional measurement points (positions defined according to OIML R76).

Nominal value : 50 g	9 g
Tolerance 0.0004 g	9 g

Difference

1	2	3	4	5	6
—	0.0001	0.0000	0.0000	0.0000	—

Linearity
The linearity, also called linearity error, describes the deviation of the characteristic curve of a weighing instrument from the linear slope.

Tolerance	0.0002 g	9 g
Nominal Value (g)	0.01	0.1
Conventional Mass Value (g)	0.0100	0.1000
Displayed Value (g)	0.0100	0.1000
Deviation (g)	0.0000	0.0000
Uncertainty (g)	0.00013	0.00013
Nominal Value (g)	1	2
Conventional Mass Value (g)	1.0000	2.0000
Displayed Value (g)	1.0000	2.0000
Deviation (g)	0.0000	0.0000
Uncertainty (g)	0.00014	0.00014
Nominal Value (g)	5	10
Conventional Mass Value (g)	5.0000	10.0000
Displayed Value (g)	5.0000	10.0000
Deviation (g)	0.0000	0.0000
Uncertainty (g)	0.00014	0.00014
Nominal Value (g)	20	50
Conventional Mass Value (g)	20.0000	50.0000
Displayed Value (g)	20.0000	50.0000
Deviation (g)	0.0000	0.0000
Uncertainty (g)	0.00015	0.00015
Nominal Value (g)	100	200
Conventional Mass Value (g)	100.0000	200.0000
Displayed Value (g)	100.0000	200.0001
Deviation (g)	0.0000	0.0001
Uncertainty (g)	0.00019	0.00030

End of Report.

Customer Name : ALS Laboratory Group (Thailand) Co., Ltd.
Songkhla Branch: 114/1 Moo 8 Kamchanawanich Rd., T. Ban Phru, A. Hat Yai, Songkhla. 90250.

Calibrated Place : Balance Room.

Calibrated By : Mr. Chonchai Inthana
Calibration Date : Wednesday, January 25, 2023

Using in-house calibration procedure number (WI-003)
Based on UKAS LAB 14 : 2019

Ambients Conditions:
Temperature : 22.4 °C ± 3.0 °C
Humidity : 65.0 % RH ± 5.0 % RH
Pressure : ±

Equipment Condition:
☒ New Installation ☐ Service / Repaired ☒ Re-calibration/ Maintenance ☐ Good Operate ☐ Fair

Measurement Method UKAS Publication Ref : Lab 14
The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM). The calibration certificate documents the traceability to National Standards, which realise the unit of measurement according to the International Standard System of Units (SI). Report of Tolerance came form list of Sartorius Metrological Specifications.

Model Number	Description	Traceability	Certificate No.	Due Date
YCS011-522-00	Sartorius weight set 1mg - 1kg E2 s/n 37929119	SPC-RT	C02212565	14-Sep-2023
MHB-382SD	Humidity/Barometer/Temp Lutron MHB-382SD	DKSH	C19220444	5-Sep-2023

This certificate relate and apply this equipment only.
This certificate may not be reproduced other than in full except with the prior written approval of the Verification Operation Division
Sartorius (Thailand) Co., Ltd.

Mr. Chonchai Inthana (Technical Manager)

S T A M P



Southern Calibration Service Co., Ltd.

669/35 Karmjanavanit Rd., Banpru, Hatyai, Songkla 90250 Thailand

Tel: 08 1599 0417 Fax: 0 7480 5133 Email: s.calibration@gmail.com www.scal-lab.com



CALIBRATION CERTIFICATE

Issued Date : 1-May-2023

Certificate No. : 23TH1728

CSR No. : A088/04367

Page : 1 of 3

Customer : ALS Laboratory Group (Thailand) Co., Ltd
114/1 Moo 8, Kamchanawanich Rd. Tambon, Ban Phru,
Amphoe Hat Yai, Songkhla, 90250

Calibration Place : Chemical Laboratory

Instrument Name : Hot Air Oven

Manufacturer : Memmert

Model : UF110

Serial No. : B416.3387

ID No. : SGK_CL0024

Resolution : 0.1 °C

Received Date : 28-Apr-2023

Calibrated Date : 28-Apr-2023

Ambient Temperature : (30 ± 10) °C

Relative Humidity : (50 ± 30) %

REVIEW BY	Ananta B.
APPROVED BY	Kawitha H.
NEXT CAL DATE	28/11/2024

Calibration Method Used :

This instrument was calibrated using the Calibration In - house method : SCAL.WI.012 based on GLA - 20

The Southern Calibration Service Co., Ltd. calibration control system complies with requirement of ISO/IEC 17025:2017

Traceability of measurement :

This Certificate is traceable to the International and for national standards which realize the units of measurement

according to the International System of Unit (SI) through :

- SCAL : Southern Calibration Service Co., Ltd.,

Calibrated by : Ibrohim Saleemin

Approved by :

Imron Rattanayum / Technical Manager

The uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written approval of Southern Calibration Service Co., Ltd.



Certificate No. : 23TH1728

CSR No. : A088/04367

Page : 2 of 3

Details of Calibration

1. Reference Standard Equipment Used:

Equipment	Model	Serial No.	Cert. no.	Due Date
Data Acquisition/Switch Unit	34970A	MY58009813	22SDAT004	24-May-2023

2. The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the longterm stability of instrument.

3. This certificate is not certified any commercial transaction


4. Condition of Item : normal condition , no indication for any damage or malfunction

Result of Calibration :

(✓) Without Adjustment

() After Adjustment

1. Sensor Installation Diagram



Sensor Installation Details

a = 5.0 cm
b = 5.0 cm
c = 5.0 cm

Dimension of the chamber

W = 40.0 cm
H = 40.0 cm
D = 33.0 cm



Southern Calibration Service Co., Ltd.
669/35 Karnjanavanit Rd., Banpru, Hatyai, Songkla 90250 Thailand
Tel : 08 1599 0417 Fax : 0 7480 5133 Email : s.calibration@gmail.com www.scal-hk.com



Certificate No. : 23TH1728
CSR No. : A088/04387
Page : 3 of 3



Result of Calibration :

2. Temperature Measurement Accuracy Test

The measurement results of the Hot Air Oven and associates are reported in the manner as shown below

Cal point (°C)	Measured Standard Temperature At Spread Locations (°C)									Uncertainty (±°C)
	#1	#2	#3	#4	#5	#6	#7	#8	Ref.9	
40	40.48	40.28	40.28	39.91	40.17	40.09	39.93	40.27	39.89	0.36
70	70.36	70.23	70.58	69.74	69.99	69.92	69.86	70.13	70.04	0.36
103	103.19	103.12	103.46	103.37	103.10	103.54	103.43	103.06	103.40	0.36
104	104.31	104.23	104.62	103.77	104.12	104.06	103.90	104.20	104.56	0.36
105	105.07	105.03	105.48	105.27	105.12	105.01	105.01	105.00	104.96	0.36
180	180.31	180.00	180.00	180.07	180.18	180.05	180.01	180.10	180.24	0.41

3. Performance Result

The performance of the Hot Air Oven are reported as shown below

Cal point (°C)	UUC Setting (°C)	UUC Reading (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
40	40.0	40.0	0.20	0.70	0.72
70	70.0	70.0	0.20	0.60	0.94
103	103.0	103.0	0.20	0.43	0.54
104	104.0	104.0	0.10	0.79	0.88
105	105.0	105.0	0.10	0.59	0.69
180	180.0	180.0	0.10	0.38	0.38

- UUC = Unit Under Calibration

The report uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

... End ...

CALIBRATION CERTIFICATE

Issued Date : 16-Jul-2023

Certificate No. : 23TH3096

CSR No. : A095/04743

Page : 1 of 3

Customer : ALS Laboratory Group (Thailand) Co., Ltd
114/1 Moo 8, Kamchanwanich Rd. Tambon, Ban Phru,
Amphoe Hat Yai, Songkhla, 90250

Calibration Place : Chemical Laboratory
Instrument Name : Incubator
Manufacturer : Memmert
Model : ICP750
Serial No. : F816.0063
ID No. : SGK_CL0028
Resolution : 0.1 °C
Received Date : 13-Jul-2023
Calibrated Date : 13-Jul-2023
Ambient Temperature : (30 ± 10) °C
Relative Humidity : (50 ± 30) %

REVIEW BY Ananta B.....
APPROVED BY Kantha H.....
NEXT CAL. DATE : 13/01/25

Calibration Method Used :

This instrument was calibrated using the Calibration In - house method : SCAL.WI.012 based on GLA - 20
The Southern Calibration Service Co., Ltd. calibration control system complies with requirement of ISO/IEC 17025:2017

Traceability of measurement :

This Certificate is traceable to the International and/or national standards which realize the units of measurement

according to the International System of Unit (SI) through :

- SCAL : Southern Calibration Service Co., Ltd.,

Calibrated by : Ibrorhin Saleemhin

Approved by :

Imron Rattanaylum / Technical Manager

The uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written approval of Southern Calibration Service Co., Ltd.



Details of Calibration

1. Reference Standard Equipment Used:

Equipment	Model	Serial No.	Cert. no.	Due Date
Data Acquisition/Switch Unit	34970A	MY58009813	23SDAT004	23-May-2024

2. The results reported in this certificate refer to the condition of the instrument on the date of calibration

and carry no implication regarding the longterm stability of instrument.


3. This certificate is not certified any commercial transaction

4. Condition of Item : normal condition , no indication for any damage or malfunction

Result of Calibration :

(✓) Without Adjustment () After Adjustment

1. Sensor Installation Diagram



Sensor Installation Details

a = 5.0 cm
b = 5.0 cm
c = 5.0 cm

Dimension of the chamber

W = 40.0 cm
H = 40.0 cm
D = 33.0 cm



Result of Calibration :

2. Temperature Measurement Accuracy Test

The measurement results of the Incubator and associates are reported in the manner as shown below

Cal point (°C)	Measured Standard Temperature At Spread Locations (°C)									Uncertainty (±°C)
	#1	#2	#3	#4	#5	#6	#7	#8	Ref. 9	
20	20.10	20.04	20.03	19.97	20.08	20.23	20.10	19.94	20.07	0.38

3. Performance Result

The performance of the Incubator are reported as shown below

Cal point (°C)	UUC Setting (°C)	UUC Reading (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
20	20.0	20.0	0.14	0.17	0.32

- UUC = Unit Under Calibration

The report uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

... End ...



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert. No.: 22LM162
Page.: 1 of 2

Certificate of Calibration

Equipment : DO Meter with Sensor
Manufacturer : YSI
Model : 5000
Serial No. : 17B101473
ID No. : SGK_CL0073

Submitted by : ALS Laboratory Group (Thailand) Co., Ltd.
Songkhla Branch.
114/1 Moo 8, Karjanavanij Rd., Banphru,

Location : TPA Chemistry Calibration Lab.2

Received Order : 18 November 2022
Calibrated Date : 21 November 2022
Ambient Temperature : $(26 \pm 10) ^\circ\text{C}$
Relative Humidity : $(50 \pm 30) \%$
AC Line Voltage : $(220 \pm 22) \text{ V}$

Calibrated by : Warakorn Lemgagtrakul

Approved by :
Approved Signatory

() Ponthippa Tameyakul
(x) Malee Bukruea
() Suwit Imjai

Issue Date : 22 November 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.



Equipment : DO Meter with Sensor
Condition As-Received : Used Item
Reference : 2111-0663DSC-2
Procedure Used :-

Calibration were conducted using in-house calibration procedure CP-OT01 according to comparison with
Industrial Platinum Resistance Thermometer (IPRT) into Temperature Bath.
The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument : Model : Serial No. : Cert. No. : Due Date :
1) Digital Thermometer : 1523 : 3240076 : 221249 : 02 Mar 2023
2. This certificate is valid only to the item calibrated on date and place of calibration.
3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function : Temperature measurement.

This instrument was connected with temperature sensor, S/N.: 17B100103

Calibration Point ($^{\circ}\text{C}$)	Immersion Depth (mm)	Standard Temperature ($^{\circ}\text{C}$)	UUC* Reading ($^{\circ}\text{C}$)	Error ($^{\circ}\text{C}$)	Uncertainty ($\pm ^{\circ}\text{C}$)	Coverage Factor k
20.00	60	20.001	19.88	-0.121	0.15	2.00

UUC* : Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

-o-o-

Walu-



Certificate of Testing

Equipment : DO Meter
Manufacturer : YSI
Model : 5000
Serial No. : 17B101473
ID No. : SGK_CL0073
Received Date : 18 November 2022
Test Date : 21 November 2022
Reference : 2211-0663DSC-1
Submitted by : ALS Laboratory Group (Thailand) Co.,Ltd.
Songkhla Branch.
114/1 Moo 8, Kanjanavanij Rd., Banphru,
Hatyai, Songkhla 90250, Thailand
Laboratory Condition : Temperature (25 ± 5) °C
Humidity (50 ± 20) %
Test Procedure : In - house method : CP-CH9
by Comparison Technique with Azide Modification Method
Tested by : Walalak Sirihean

Approved by : Wala.
Approved Signatory

(/) Malee Butkruea
() Saithip Meangmai
() Warakorn Lengagtrakul

Issue Date : 22 November 2022

Condition of this result of calibration

1. Reference Standard Instruments :
This certification is traceable to the International System of Unit through the reference standards laboratory of Industrial Calibration Center, Technology Promotion Association (Thailand-Japan).

Instruments :
1) Burette
2) Balance
Material :
Sodium Thiosulfate pentahydrate
Serial No. : 1126143764
ID No. : 130BU10
Certificate No. : 21CG1389
Due Date : 25 Mar 2023
Manufacturer : Merck
Lot.No. : AM1763316
Assay : 100.2%

Result : Dissolved Oxygen Meter Adjustment With Air 100 %
Dissolved Oxygen Probe No.: 17B100103

Titration Method (Azide Modification Method) (mg/L)	DO Meter Reading (mg/L)	Standard Deviation (mg/L)
8.12	8.12	0.0045

This report was certified only for the instrument we tested. It is allowable to use for study the system efficiency. The environmental impact control and present to organization it may concerned. Intend to use for advertising and referral purpose is prohibited. This report may not be reproduced other in full, without written approval of the laboratory.

-o0o-

Wala.



Southern Calibration Service Co., Ltd.

669/35 Kamjanavanit Rd., Banpru, Hatyai, Songkla 90250 Thailand

Tel : 08 1599 0417 Fax : 0 7480 5133 Email : sscalibration@gmail.com www.sscal-lab.com



MSC-TS-TESTING
CALIBRATION 02/04

CALIBRATION CERTIFICATE

Issued Date : 16-Jul-2023

Certificate No. : 23TH3097

CSR No. : A095/04743

Page : 1 of 3

Customer : ALS Laboratory Group (Thailand) Co., Ltd

114/1 Moo 8, Kamchanawanch Rd. Tambon, Ban Phru,

Amphoe Hat Yai, Songkhla, 90250

Calibration Place

Instrument Name

Manufacturer

Model

Serial No.

ID No.

Resolution

Received Date

Calibrated Date

Ambient Temperature

Relative Humidity

REVIEW BY : Ananta B.

APPROVED BY : Kanitra B.

NEXT CAL DATE : 13/01/25

Calibration Method Used :

This instrument was calibrated using the Calibration In - house method : SCAL WI.014 based on ASTM E 715 : 1980 (reapproved 2001)

The Southern Calibration Service Co., Ltd calibration control system complies with requirement of ISO/IEC 17025:2017

Traceability of measurement :

This Certificate is traceable to the International and/or national standards which realize the units of measurement according to the International System of Unit (SI) through :

- SCAL : Southern Calibration Service Co., Ltd.,

Calibrated by : Ibrohim Saleem

Approved by :

Imron Rattanayum / Technical Manager

The uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written approval of Southern Calibration Service Co., Ltd.



Certificate No. : 23TH3097

CSR No. : A095/04743

Page : 2 of 3

Details of Calibration

1. Reference Standard Equipment Used:

Equipment	Model	Serial No.	Cert. no.	Due Date
Data Acquisition/Switch Unit	34970A	MY58009813	23SDAT004	23-May-2024

2. The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the longterm stability of instrument.

3. This certificate is not certified any commercial transaction


4. Condition of Item : normal condition , no indication for any damage or malfunction

Result of Calibration :

(✓) Without Adjustment

() After Adjustment

1. Sensor Installation Diagram



Sensor Installation Details

a =	5	cm
b =	5	cm
c =	5	cm

Dimension of the chamber

W =	45	cm
H =	30	cm
D =	35	cm



Certificate No. : 23TH3097
CSR No. : A095/04743
Page. : 3 of 3

Result of Calibration :-

2. Temperature Measurement Accuracy Test

The measurement results of the Water Bath and associates are reported in the manner as shown below

Cal point (°C)	Measured Standard Temperature At Spread Locations (°C)				Uncertainty (±°C)
	#1	#2	#3	#4	
80	79.17	79.47	79.43	79.25	0.14

3. Performance Result

The performance of the Water Bath are reported as shown below

Cal point (°C)	UUC Setting (°C)	UUC Reading (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
80	80.0	80.0	0.24	0.38	0.38

- UUC = Unit Under Calibration

The report uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

... End ...



Agilent Technologies

Agilent Technologies (Thailand) Limited
U CHU LIANG BLDG. 22/F UNIT A.D
968 RAMA 4 ROAD, SILOM, BANGRAK
Bangkok 10500 Thailand
Tel. +662 637 6363
Fax: +662 637 4334
Email: ccc-smt@agilent.com
Website: www.agilent.com/chem

SERVICE REPORT

Customer Purchase Order Number:	Customer Number: 70579367
Service Request:	Service Request Date:
Service Order: 6005939422	Service Confirmation: 6905074481

Customer Contact:
ALS Laboratory Group (Thailand) Co
Ltd
Branch Number 0002
114/1 Moo8 Banplu Subdistrict Hat
Yai District
TAX ID : 0105540004689
kanitta.hemprasatpor@alsglobal.com
0811721334

Invoice To:
ALS Laboratory Group (Thailand) Co
Ltd
Branch Number 0002
114/1 Moo8 Banplu Subdistrict Hat
Yai District SONGKHLA 90250

Payer:
ALS Laboratory Group (Thailand) Co
Ltd Head Office
104 Phatthanakan 40 Phatthanakan Rd
Khaeng Phatthanakan Khet Suan
Luang
BANGKOK 10250

Delivery Site:
ALS Laboratory Group (Thailand) Co
Ltd
Branch Number 0002
114/1 Moo8 Banplu Subdistrict Hat
Yai District

Location:
Room
Bldg
Lab
Dept

Direct Inquiries to:
Contact Name:
Contact E-mail:
Contact Telephone:
Contact Fax:

Customer Contact Center
ccc-smt@agilent.com
+662 637 6363
+662 632 4334

REVIEW BY ...Kwinnia...
APPROVED BY Kanitta H.
NEXT CAL. DATE 2/01/25

products | applications | software | services

Agilent Technologies (Thailand) Limited, Head Office
U Chu Liang Bldg. 22/F Unit A.D
968 Rama 4 Road, Silom, Bangkok,
Bangkok 10500 Thailand
Tax ID : 0105542068218

Citibank N.A. Bangkok Branch
389 Interchange 21 Building, Sukhumvit Road, Klongtoey Neu
Sub-district, Watana District, Bangkok 10110 Thailand
Acc. No: 012-4452-087,
THBKrung Thai Bank PCL
Siam Square Bx. 418/1-2 Rama 1 Rd., Pathumwan, BKK 10330
Thailand

Learn more about Agilent's Special Offers, Products, Services and our full range of laboratory productivity solutions optimized for your applications and workflows. Visit us at www.agilent.com/chem

ORIGINAL

Service Instrument:



Model Number	Model Description	Serial Number	System Handle	Parent Asset
SYS-IM-7900	ICPMS 7900 System			
G7201C	ICP-MS MassHunter SW only (excludes PC)	USH3799575	ICP MS 7900	SYS-IM-7900
G8403A	Agilent 7900 ICP-MS	JP16511669	ICP MS 7900	SYS-IM-7900
G8411A	ISIS 3 for Agilent 7850/7900/8900	JP16510379	ICP MS 7900	SYS-IM-7900

Service Items:

Item	Service/Part #	Description	Qty	Entitlement	Service Start	Service End
1000	EQ	Enterprise Operational Qualification	1.00	Agreement Entitlement - 100 % covered	02.08.2023	02.08.2023
1010	5185-5850	ICP-MS Checkout Solutions	1.00	Agreement Entitlement - 100 % covered		

Additional Information:

Service Information:

Problem Description: WU-S-QQ-IM-7900-5001093854	
Service Provided: -Perform OQ hardware. -Test OQ of instrument (CP-MS = SGK_CL0048. All tests Passed.	
Service Overview Code: Reason Code: Scheduled Service Diagnosis Code: Scheduled Service Resolution Code: Scheduled Service	
Reported Hours: 9.0	Travel Hours: 3.0
Customer Field Service Representative Name: Uthai Ngamlertsirichai	Customer Field Service Representative Signature: 
Customer Name: KANITTA HEMPRASATPORN	Customer Signature: 
Date: 02 Aug 2023	
Date: 02 Aug 2023	
Additional Comments:	



Southern Calibration Service Co., Ltd.

669/35 Karmjanavanit Rd., Banpru, Hatyai, Songkla 90250 Thailand

Tel : 08 1599 0417 Fax : 0 7480 5133 Email : sculibration@gmail.com www.scul-lab.com



ISO/IEC 17025

CALIBRATION CERTIFICATE

Issued Date : 3-Feb-2023

Certificate No. : 23TH0527

CSR No. : A073/03634

Page. : 1 of 3

Customer : ALS Laboratory Group (Thailand) Co., Ltd
114/1 Moo 8 Kamchanawanich Rd. T.Ban Phru,
A. Hat Yai, Songkhla 90250 TH

Calibration Place : Chemical Laboratory
Instrument Name : Cold Room Water
Manufacturer : MODULAR
Model : N/A
Serial No. : N/A
ID No. : SGK_CL0065
Resolution : 0.1 °C
Received Date : 31-Jan-2023
Calibrated Date : 31-Jan-2023
Ambient Temperature : (30 ± 10) °C
Relative Humidity : (50 ± 30) %

REVIEW BY Ananta B.
APPROVED BY Kampha H.
NEXT CAL. DATE 31/07/24

Calibration Method Used :

This instrument was calibrated using the Calibration In - house method : SCAL.WI.012 based on G-20

The Southern Calibration Service Co., Ltd. calibration control system complies with requirement of ISO/IEC 17025:2017

Traceability of measurement :

This Certificate is traceable to the International and for national standards which realize the units of measurement

according to the International System of Unit (SI) through :

- SCAL : Southern Calibration Service Co., Ltd.,

Calibrated by : Iborhim Saleem

Approved by :

Imron Rattanyalum / Technical Manager

The uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written approval of Southern Calibration Service Co., Ltd.



Certificate No. : 23TH0527

CSR No. : A073/03634

Page. : 2 of 3

Details of Calibration

1. Reference Standard Equipment Used:

Equipment	Model	Serial No.	Cert. no.	Due Date
Data Acquisition/Switch Unit	34970A	MY5800813	22SDAT004	24-May-2023

2. The results reported in this certificate refer to the condition of the instrument on the date of calibration and carry no implication regarding the longterm stability of instrument.

3. This certificate is not certified any commercial transaction


4. Condition of Item : normal condition , no indication for any damage or multifunction

Result of Calibration :

(✓) Without Adjustment

() After Adjustment

1. Sensor Installation Diagram



Sensor Installation Details

a = 5.0 cm
b = 5.0 cm
c = 5.0 cm

Dimension of the chamber

W = 40.0 cm
H = 40.0 cm
D = 33.0 cm



Certificate No. : 23TH0527
CSR No. : A073/03634
Page. : 3 of 3

Result of Calibration :

2. Temperature Measurement Accuracy Test

The measurement results of the Cold Room Water and associates are reported in the manner as shown below

Cal point (°C)	Measured Standard Temperature At Spread Locations (°C)														
	#1	#2	#3	#4	#5	#6	#7	#8	#9	Ref.10	#11	#12	#13	#14	#15
4	3.18	3.39	3.54	3.77	3.99	3.86	3.85	3.92	4.02	3.86	3.78	3.84	3.85	4.09	3.91

The uncertainty of measurement was ± 0.38 °C

3. Performance Result

The performance of the Cold Room Water are reported as shown below

Cal point (°C)	UUC Setting (°C)	UUC Reading (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
4	4.0	4.0	1.23	0.50	0.84

- UUC = Unit Under Calibration

The report uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

... End ...



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-29 FAX. 0-2719-9484



MSC-TS17157025
CALIBRATION 0008

Cert. No.: 23TM1408
Page : 1 of 4

Certificate of Calibration

Equipment : Autoclave

Manufacturer : TOMY

Model : SX-700

Serial No. : 48134190

ID No. : BKK_ML0041

Submitted by : ALS Laboratory Group (Thailand) Co.,Ltd.
104 Phatthanakan 40, Phatthanakan Rd.,
Khwaeng Phatthanakan, Khet Suan Luang,
Bangkok 10250 Thailand

Location : Media Preparation Room

Received Order : 03 October 2023

Calibration Date : 04 October 2023

Ambient Temperature : (26 ± 10) °C

Relative Humidity : (50 ± 30) %

Calibrated by : Khit Ruitanaprapachai

Approved by :

() Pornthippa Tameyakul
(✓) Ponpan Paipim
() Suwit Imjai


Approved Signatory

Issue Date : 11 October 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written

Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.



A 0059272



Equipment : Autoclave
Condition As-Received : Used Item
Reference : 2310-0006OC-6
Cert. No.: 23TM1408
Page : 2 of 4

Procedure Used :-

Calibration were conducted using in-house calibration procedure CP-OT03 according to direct measurement method with Data Acquisition which connected with Thermocouple Type T

The temperature scale used was based on ITS-90.

Condition of this result of calibration


1. Reference standard instrument:-
- | Instrument | Serial No. | Cert. No. | Traceable | Due Date |
|----------------------|------------|-----------|-----------|-------------|
| 1) Data Acquisition | MY57013823 | 23LM66 | TPA | 25 Mar 2024 |
2. This certificate is valid only to the item calibrated on date and place of calibration.
3. This certification is traceable to the International System of Unit.
4. This result of calibration covers laboratory autoclaves for the sterilization of goods and material which could be infected with organisms categorized as Hazard Group 1, 2 and 3**

(** = Categorization of pathogens according to hazard and categories of containment, second edition, 1990)
It does not cover autoclaves for use with material infect with organisms in Hazard Group 4, for which complete containment and sterilization of infected condensate is considered to be essential.
This result of calibration does not apply to sterilizers or disinfectors used for medical, dental, pharmaceutical or veterinary purposes which are directly concerned with patient care, or those used for fabrics subjected to sterilization which are required to be dry at the end of cycle.

Remark : TPA : Technology Promotion Association (Thailand - Japan)

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source



Environmental	
(°C)	(%R.H.) (Volt)
Beginning of Calibration	26 64 221
Finished of Calibration	27 67 222

Position	Description	Ref. Std. ID No.:
1 =	Center of chamber	19-177C-08
2 =	Temperature sensor	19-177C-09
3 =	Exhaust port	19-177C-10

Signature



Equipment : Autoclave
Condition As-Received : Used Item
Reference : 2310-0006OC-6
Cert. No.: 23TM1408
Page : 3 of 4

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Operating parameter Set : Temperature = 108 °C

Sterilization period = 10 minute

UUC* Setting (°C)	UUC* Reading (°C)	Position	Average* Standard Reading (°C)	Stability (± °C)	Pressure Reading (MPa)	Uncertainty (± °C)	Coverage Factor k
108	108	1	108.352	0.12	0.04	0.90	2
		2	108.263				
		3	108.140				

Operating parameter Set : Temperature = 115 °C

Sterilization period = 20 minute

UUC* Setting (°C)	UUC* Reading (°C)	Position	Average* Standard Reading (°C)	Stability (± °C)	Pressure Reading (MPa)	Uncertainty (± °C)	Coverage Factor k
115	115	1	115.376	0.13	0.08	0.90	2
		2	115.297				
		3	115.157				

Operating parameter Set : Temperature = 118 °C

Sterilization period = 10 minute

UUC* Setting (°C)	UUC* Reading (°C)	Position	Average* Standard Reading (°C)	Stability (± °C)	Pressure Reading (MPa)	Uncertainty (± °C)	Coverage Factor k
118	118	1	118.083	0.11	0.09	0.90	2
		2	118.037				
		3	117.954				

Average* : The average of 30 values in each position.

Stability : One-half of the greatest maximum difference of measured temperature at any one probe.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

Signature



Equipment : Autoclave
Condition As-Received : Used Item
Reference : 2310-0006OC-6
Result of Calibration :- (*) Without Adjustment
Function of UUC* : Temperature Source
Operating parameter Set : Temperature = 121 °C
Sterilization period = 30 minute

UUC* Setting (°C)	UUC* Reading (°C)	Position	Average* Standard Reading (°C)	Stability (± °C)	Pressure Reading (MPa)	Uncertainty (± °C)	Coverage Factor k
121	121	1	121.186	0.17	0.11	0.91	2
		2	121.082				
		3	120.980				

Average* : The average of 30 values in each position.
Stability : One-half of the greatest maximum difference of measured temperature at any one probe.
UUC* : Unit Under Calibration
Note : The reported uncertainty of measurement was excluded stability.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor **k**, providing a level of confidence of approximately 95 %.

-o0o-

PR

a 1184531



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-27 FAX. 0-2719-9484



NSC-TS1-TS17025
CALIBRATION 0008

Cert. No.: 22TM1571
Page : 1 of 3

Certificate of Calibration

Equipment : Hot Air Oven

Manufacturer : Binder

Model : ED 240/E2

Serial No. : 00-15533

ID No. : BKK_ML0013

Submitted by : ALS Laboratory Group (Thailand) Co.,Ltd.
104 Phatthanakan 40, Phatthanakan Rd.,
Khwaeng Phatthanakan, Khet Suan Luang,
Bangkok 10250 Thailand
Location : Media Preparation Room

Received Order : 21 November 2022

Calibration Date : 21 November 2022

Ambient Temperature : (26 ± 10) °C

Relative Humidity : (50 ± 30) %

Calibrated by : Krisda Malee

Approved by : 
Approved Signatory

() Ponthippa Tameyakul
(☒) Malee Butkruea
() Suwit Imjai

Issue Date : 29 November 2022

The Uncertainties are for a confidence probability of approximately 95 %

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

A 0048150



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-29 FAX. 0-2719-9484



Cert. No.: 23TM1364
Page : 1 of 3

Certificate of Calibration

Equipment : Incubator
Manufacturer : Memmert
Model : INE 800
Serial No. : E805.0063
ID No. : BKK_ML0018
Submitted by : ALS Laboratory Group (Thailand) Co.,Ltd.
104 Phatthanakan 40, Phatthanakan Rd.,
Khwaeng Phatthanakan, Khet Suan Luang,
Bangkok 10250 Thailand
Location : Incubation & Microbiological Reading
Received Order : 03 October 2023
Calibration Date : 03 October 2023
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %
Calibrated by : Man Pattanapongpaiboon

REVIEW BY	<i>S. Wicks</i>
APPROVED BY	<i>[Signature]</i>
NEXT CAL. DATE	03/10/24

Approved by : *[Signature]* Approved Signatory

() Pornthippa Tameyakul
(✓) Ponpan Paipim
() Suwit Injai

Issue Date : 11 October 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

A 0053270



Equipment : Incubator
Condition As-Received : Used Item
Reference : 2310-0006OC-4

Cert. No.: 23TM1364
Page : 2 of 3

Procedure Used :-

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD).
The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument Serial No. Cert. No. Traceable Due Date
1) Data Acquisition MY49023932 23LM122 TPA 26 Jul 2024

2. This certificate is valid only to the item calibrated on date and place of calibration.
3. This certification is traceable to the International System of Unit.


Remark : TPA : Technology Promotion Association (Thailand - Japan)

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Close

Environment during calibration	
Beginning	Finished
Temp. (°C)	24
REL.Humid. (%)	54
AC Supply (Volt)	222
	224



Probe Installation Details :
a = 10 cm
b = 10 cm
c = 10 cm
Dimension of Chamber :
D = 0.60 m
W = 1.0 m
H = 1.2 m
Capacity = 0.75 m³

Position :	Ref. Std. ID No.:
1	20-16RTD-01
2	20-16RTD-02
3	20-16RTD-03
4	23-16RTD-04
5	22-16RTD-05
6	20-16RTD-06
7	20-16RTD-07
8	22-16RTD-08
9 (ref.)	22-16RTD-09

PR

a 1184537



Equipment : Incubator
Condition As-Received : Used Item
Reference : 2310-006OC-4
Result of Calibration :- (*) Without Adjustment
Function of UUC* : Temperature Source
Fresh air setting : Close

Calibration Point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Temperature stability (± °C)	Temperature uniformity (°C)	Overall Variation (°C)	Coverage Factor k
35.0	35.0	35.0	0.077	0.51	0.68	2

Calibration Point (°C)	Measured Temperature (°C)									Uncertainty (± °C)
	Position									
	1	2	3	4	5	6	7	8	9 (ref.)	
35.0	35.361	35.313	35.282	35.403	34.871	34.826	34.807	34.964	34.937	0.30

Average* : The average of 30 values in each position.

Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor.

Temperature uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-o0o-



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-29 FAX. 0-2719-9484



Cert. No.: 23TM637
Page : 1 of 3

Certificate of Calibration

Equipment : Water Bath

Manufacturer : Memmert

Model : WNE 45

Serial No. : L712.0429

ID No. : BKK_ML0056

Submitted by : ALS Laboratory Group (Thailand) Co., Ltd.
104 Phatthanakan 40, Phatthanakan Rd.,
Khwaeng Phatthanakan, Khet Suan Luang,
Bangkok 10250 Thailand

Location : Incubator & Microbiological Reading

Received Order : 20 April 2023

Calibration Date : 20 April 2023

Ambient Temperature : (26 ± 10) °C

Relative Humidity : (50 ± 30) %

Calibrated by : Kunchit Promprat

Approved by :  Approved Signatory

() Ponthippa Tameyakul

(✓) Malee Butkruea

() Suwit Imjai

Issue Date : 24 April 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written

Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.



Equipment : Water Bath
Condition As-Received : Used Item
Reference : 2304-0253OC-1
Procedure Used :-

Cert. No.: 23TM637
Page : 2 of 3

Calibration were conducted using in-house calibration procedure CP-OT04 according to direct measurement method with Data Acquisition which connected with Industrial Platinum Resistance Thermometer (IPRT).

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument Model Serial No. Cert. No. Due Date
1) Data Acquisition 34970A MY44073381 22LM78/1 12 May 2023
2. This certificate is valid only to the item calibrated on date and place of calibration.


3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Heat transfer medium used : Water

	Environmental		AC Voltage Supply
	(°C)	(%R.H.)	(Volt)
Beginning of Calibration	25	45	223
Finished of Calibration	25	43	223



Front

Position :	Ref. Std. S/N.:
1	4803988-006
2	4803988-007
3	4804539-014
4	4804539-015
5(ref.)	4804539-016

Wala

a 1158265



Equipment : Water Bath
Condition As-Received : Used Item
Reference : 2304-0253OC-1
Result of Calibration :- (*) Without Adjustment
Function of UUC* : Temperature Source

Cert. No.: 23TM637
Page : 3 of 3

Calibration point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Average* Standard Reading (°C)					Uncertainty (± °C)
			1	2	3	4	5 (ref.)	
44.5	44.5	44.5	44.492	44.463	44.475	44.510	44.491	0.15
45.0	45.0	45.0	45.005	44.962	44.979	45.016	44.986	0.15

Calibration point (°C)	Uniformity (°C)	Stability (± °C)	Coverage Factor k
44.5	0.051	0.022	2
45.0	0.080	0.026	2

Average* : The average of 30 values in each position.

Uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Stability : One-half of the greatest maximum difference of measured temperature at any one probe.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

-o-o-

Wala

a 1158264


CALIBRATION REPORT

REVIEW BY : Vithan N
APPROVED BY : Sirarat M.
NEXT CAL. DATE : 9/2/67

CUSTOMER NAME :	ALS Laboratory Group (Thailand) Co., Ltd.
EQUIPMENT NAME :	Total Hydrocarbon Analyzer
MANUFACTURER :	Baseline
SERIAL NO :	0314DRO170
STANDARD GAS CONCENTRATION (PPM) :	100 PPM (Methane)
CYLINDER NO :	ND55981
CYLINDER PRESSURE (psig) :	900 PSI
CERTIFIED DATE :	12/02/2022
EXPIRED DATE :	12/02/2025

CALIBRATION RESULTS

POINT NO	CALIBRATION RESULTS			
	IDEAL	ACTUAL	ERROR	%ERROR
ZERO	0.00	0.00	0.00	-
1	100.00	101.00	1.0	1.00
AVERAGE (%)				0.25



CALIBRATED BY : วราภรณ์ อภิชาติธรรม DATE : 9/8/65
CHECKED BY : ศิริรัตน์ อภิชาติธรรม DATE : 9/8/65


ต้องการข้อมูลทางด้านเทคนิคเพิ่มเติม : เจ้าหน้าที่ฝ่ายปฏิบัติการบริการ , โทร 02-868-0812 # 31 , E-Mail : Engineer@jirante.com
เลขที่ 03/14-15.67/35-36 ถนนเพชรเกษม 7/1 แขวง วัดท่าพระ เขต บางกอกใหญ่ กรุงเทพฯ 10600 โทร 02-868-0812-13 โทรสาร 02868-1889

CALIBRATION REPORT

CUSTOMER NAME :	ALS Laboratory Group (Thailand) Co., Ltd.
EQUIPMENT NAME :	Total Hydrocarbon Analyzer
MANUFACTURER :	Baseline
SERIAL NO :	0314DRO170
STANDARD GAS CONCENTRATION (PPM) :	100 PPM (Propane)
CYLINDER NO :	ND55981
CYLINDER PRESSURE (psig) :	900 PSI
CERTIFIED DATE :	12/02/2022
EXPIRED DATE :	12/02/2025

CALIBRATION RESULTS

POINT NO	CALIBRATION RESULTS			
	IDEAL	ACTUAL	ERROR	%ERROR
ZERO	0.00	0.00	0.00	-
1	300.00	301.00	1.0	0.33
AVERAGE (%)				0.08



CALIBRATED BY : วราภรณ์ อภิชาติธรรม DATE : 9/8/65
CHECKED BY : ศิริรัตน์ อภิชาติธรรม DATE : 9/8/65

ต้องการข้อมูลทางด้านเทคนิคเพิ่มเติม : เจ้าหน้าที่ฝ่ายปฏิบัติการบริการ , โทร 02-868-0812 # 31 , E-Mail : Engineer@jirante.com
เลขที่ 03/14-15.67/35-36 ถนนเพชรเกษม 7/1 แขวง วัดท่าพระ เขต บางกอกใหญ่ กรุงเทพฯ 10600 โทร 02-868-0812-13 โทรสาร 02868-1889

FLOW CALIBRATE

CUSTOMER NAME	: ALS Laboratory Group (Thailand) Co., Ltd.
EQUIPMENT NAME	: Flow Calibrator
MANUFACTURER	: Bios
MODEL	: 510 L
SERIAL NO	: 129549

Flow Parameter	Step	Set	Display	Flow Meter
Sample	Before	40	38	15 cc/min
	After	40	40	39.7 cc/min
Air	Before	175	175	190 cc/min
	After	175	175	176 cc/min
Fuel	Before	35	32	36 cc/min
	After	35	35	35 cc/min

CALIBRATED BY : วราภรณ์ คงอยู่ DATE : 9/8/65
 CHECKED BY : ธนากร อึ้งอานันท์ DATE : 9/8/65



ผลการสอบเทียบตามข้อกำหนด : ใช้เพื่อใช้ในการวัดการไหล , โทร 02-868-0812 # 15-16 , E-Mail : Engineer@jnanatee.com
 เลขที่ 63/14-15,67/35-36 ถนน เพชรเกษม 7/1 แขวง วัฒนาพรบง 7/1 เขต วัฒนา กรุงเทพมหานคร 10600 โทร 02-868-0812-13 โทรสาร 02868-1889

Certificate of System Qualification

GC-OQ + GCMS-OQ

System ID: RYG_EN0136
 Organization Name: ALS Laboratory Group (Thailand) Co Ltd.
 Organization Location: 616/10 Moo 5, Tambol Mae Nam Koo, A.Pluakdaeng, Rayong,21140, Thailand
 Date: July 7, 2022 11:27:53 AM
 EQP Name: AgilentRecommended , AgilentRecommended
 EQP Revision: GC.02.52, GCMS.02.52
 Overall Qualification Status: Pass

REVIEW BY N. B. S. S.
 APPROVED BY D. S.
 NEXT CAL. DATE 09/01/24

CDS Logon Verification - GC

Logon: dej.changchon

Overall CDS Logon Verification - GC Test Status

Pass

System Inspection and Basic Safety and Operation

Name: 7890

Setpoint Status:

Pass

Overall System Inspection and Basic Safety and Operation Test Status

Pass

Inlet Pressure Accuracy

Name: 7890

Front

SSL

Setpoint Status:

Pass

Setpoint

Inlet Pressure: 25.0 psi

Actual 25.1 psi

Accuracy: 0.1 psi

Agilent Recommended: <= 1.2

Date: July 7, 2022 11:27:53 AM
 System ID: RYG_EN0136

Log Amp

Tested Combination1FrontSSLExternalSQ

Name:5977B

Setpoint Status:Pass

Overall Log Amp Test Status

Pass

RFPA

Tested Combination1FrontSSLExternalSQ

Name:5977B

Setpoint Status:Pass

Amu:1050m/z

Drift After Five Minutes:

>=1mVand<=100mV

Agilent Recommended:

>=479mVand<=1100mV

Overall RFPA Test Status

Pass

Tune EI

Tested Combination1FrontSSLExternalSQ

Name:5977B

Setpoint Status:Pass

Filament:1

Setpoint Status:Pass

Filament:2

Overall Tune EI Test Status

Pass

Signal to Noise EI

Overall Inlet Pressure Accuracy Test Status

Pass

GC Oven Temperature Accuracy

Name:7890

Setpoint Status:Pass

Zone:Oven

Setpoint/Actual

Temperature:230.0230.6°C

Accuracy:0.6°C

Agilent Recommended:

>=-1.0% setpoint in K<=1.0% setpoint in K

(-5.0°C)(5.0°C)

Setpoint Status:Pass

Zone:Oven

Setpoint/Actual

Temperature:100.099.9°C

Accuracy:-0.1°C

Agilent Recommended:

>=-1.0% setpoint in K<=1.0% setpoint in K

(-3.7°C)(3.7°C)

Overall GC Oven Temperature Accuracy Test Status

Pass

GC Oven Temperature Stability

Name:7890

Setpoint Status:Pass

Setpoint/Average

Temperature:100.099.91667°C

Stability:0.1°C

Agilent Recommended:

<=0.5

Overall GC Oven Temperature Stability Test Status

Pass

Instrument Details

Purpose

This section describes the as found system configuration.

Details

System	
System ID	RYG_EN0136
Manufacturer	Agilent Technologies
Name	7890
Flow Data Input	Manual Data
Temperature Data Input	Manual Data or Other Data Logging
Tested Combination1	
Injection Technique	Manual Injection
Inlet	Front
Detector	External
LTM Included?	No
Sampler 1	
Manufacturer	Agilent Technologies
Type	Manual Injection
Usage	Sample Injection
Syringe Volume (µL)	10
Mainframe 1	
Manufacturer	Agilent Technologies
Name	7890
Model Number	G3442B
Serial Number	CN16463238
Firmware Revision	B.02.04.3
Component ID/Asset No.	081117000236
Oven Type	Standard

Tested Combination1

Name: 5977B

Front

SSL

/ External

SQ

Source: EI - Extractor

Filament: 1

Setpoint Status: Pass

Signal to Noise: 7485

Agilent Recommended: >= 1200

Source: EI - Extractor

Filament: 2

Setpoint Status: Pass

Signal to Noise: 2097

Agilent Recommended: >= 1200

This test's 2 comment(s) and 7 deviation(s) are available in the Attachments section.

Overall Signal to Noise EI Test Status

Pass

Electronic Signature

Purpose

This signature page was created and published because the ACE sign-off action was executed, which is valid for the entire document, including attachments. The ACE sign-off is an electronic signature that requires two distinct identification components: unique username and personal password. The Agilent representative who has delivered this service understands the meaning and legal status of an electronic signature. As a trained official operator, the Agilent representative has a unique password and login to access ACE and electronically sign this document. (Other e-signatures can be applied to this document using a Document Content Management or other suitable method defined in your data access and control procedures.)

Details

Full Name of Signer: Eaknarin Puangsopa
Logged On User Name: eaknarin_puangsoa@agilent.com
Signature Creation Date: July 7, 2022
Reason for Signature: Executed protocol and published this original version of document

Regulatory Disclaimer

This document provides a protocol to verify and record instrument configuration and evidence of proper operation. It has been prepared from our interpretation of applicable regulations as well as industry best practices. The document is designed to provide an important component of a complete compliance package. Validation depends upon many factors and use of this protocol alone does not assure compliance. Agilent Technologies makes no promises or representations as to its sufficiency for any specific regulatory program.

Warranty

Agilent Technologies makes no warranty of any kind to this material, including but not limited to, the implied warranties or merchantability and fitness for a particular purpose. Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Inlet 1	Agilent Technologies
Manufacturer	
Name	7890
Type	SSL
Location	Front
Carrier Gas	Helium
Control Type	Electronic Pressure Control (EPC)
Purged Inlet	Yes
Detector 1	Agilent Technologies
Manufacturer	
Name	Mass Spectrometer
Type	Mass Spectrometer
Location	External
Mass Spectrometer 1	Agilent Technologies
Manufacturer	
Type	SQ
Name	5977B
Serial Number	US1701M008
Firmware Revision	5977 6.00.34
High Vacuum System	Turbo Pump
Scouting Run Standard	OFN Std
Component ID/Asset No.	081117000236
MS EI Source 1	Agilent Technologies
Manufacturer	
Source Type	EI - Extractor
Number of filaments	2

User Name: eaknarin_puangsoa
Hostname: ASRYGW7002

User Name: eaknarin_puangsoa
Hostname: ASRYGW7002

System Id: RYG_EN0136
Print Date: July 7, 2022 11:27:56 AM

ALS_RYG_EN0136 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
July 6, 2022 1:19:59 PM	End	Execution	System Inspection and Basic Safety and Operation - 7890: - Qualitative Test - No setpoints associated	Run Count : 1
July 6, 2022 1:20:15 PM	Start	Execution	Inlet Pressure Accuracy - Front SSL - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	None
July 6, 2022 1:21:43 PM	End	Execution	Inlet Pressure Accuracy - Front SSL - Pressure Controlled Inlet - S: 25.0 psi - L: <= 1.2 psi	Run Count : 1
July 6, 2022 1:21:45 PM	Start	Execution	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	None
July 6, 2022 1:25:12 PM	Audit	Data	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Manual Data Entry
July 6, 2022 1:25:15 PM	End	Execution	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 230.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Run Count : 1
July 6, 2022 1:25:17 PM	Start	Execution	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	None
July 6, 2022 1:25:32 PM	Start	Execution	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	None
July 6, 2022 1:33:42 PM	Audit	Data	GC Oven Temperature Accuracy - 7890: - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Manual Data Entry

User Name: eaknarin_puangsoa
Hostname: ASRYGW7002

User Name: eaknarin_puangsoa
Hostname: ASRYGW7002

System Id: RYG_EN0136
Print Date: July 7, 2022 11:27:56 AM

ALS_RYG_EN0136 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
July 6, 2022 1:11:54 PM	Audit	SessionCreated	Session	None
July 6, 2022 1:11:54 PM	Start	Configuration	Session	None
July 6, 2022 1:11:54 PM	Audit	Entitlement	Licensing	User is Nonpaying and does not require an unlock code
July 6, 2022 1:17:19 PM	Audit	ExpLoaded	Session	EQP details for primary technique [GC] - File path: [ProtocolPacks/Gc/Configurations/02.52/Gc/02.52.exp], EQP File Name: [Gc.02.52.exp], EQP Name: [AgilentRecommended] EQP details for hyphenated technique [GoMs] - File path: [ProtocolPacks/GoMs/Configurations/02.52/GoMs.02.52.eqp], EQP File Name: [GoMs.02.52.exp], EQP Name: [AgilentRecommended]
July 6, 2022 1:17:25 PM	End	Configuration	Session	None
July 6, 2022 1:17:28 PM	Start	Qualification	Session	OQ
July 6, 2022 1:17:30 PM	Start	Execution	CDS Logon Verification - GC : - Qualitative test	None
July 6, 2022 1:19:43 PM	End	Execution	CDS Logon Verification - GC : - Qualitative test	Run Count : 1
July 6, 2022 1:19:45 PM	Start	Execution	System Inspection and Basic Safety and Operation - 7890: - Qualitative Test - No setpoints associated	None

User Name: eaknarin_puangsoopa
Hostname: ASRYGW7002

System Id: RYG_EN0136
Print Date: July 7, 2022 11:27:56 AM

ALS_RYG_EN0136 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
July 6, 2022 1:33:43 PM	End	Execution	GC Oven Temperature Accuracy - 7890 - Temperature : Oven - S: 100.0°C - L: >= -1.0 AND <= 1.0 % setpoint in K	Run Count : 1
July 6, 2022 1:33:45 PM	Start	Execution	GC Oven Temperature Stability - 7890 - Temperature : Oven - S: 100.0°C - L: <= 0.5°C	None
July 6, 2022 1:55:05 PM	Audit	Data	GC Oven Temperature Stability Manual Data Entry - 7890 - Temperature : Oven - S: 100.0°C - L: <= 0.5°C	
July 6, 2022 1:53:07 PM	End	Execution	GC Oven Temperature Stability - 7890 - Temperature : Oven - S: 100.0°C - L: <= 0.5°C	Run Count : 1
July 6, 2022 1:53:11 PM	Start	Execution	Log Amp - 5977B SQ - Source: None	
July 6, 2022 1:57:10 PM	End	Execution	Log Amp - 5977B SQ - Source: None	
July 6, 2022 1:57:24 PM	Start	Execution	Log Amp - 5977B SQ - Source: None	
July 6, 2022 2:09:24 PM	End	Execution	Log Amp - 5977B SQ - Source: None	
July 6, 2022 2:09:28 PM	Start	Execution	Log Amp - 5977B SQ - Source: None	
July 6, 2022 2:24:46 PM	End	Execution	Log Amp - 5977B SQ - Source: None	
July 6, 2022 2:24:46 PM	Start	Execution	Log Amp - 5977B SQ - Source: None	
July 6, 2022 2:41:39 PM	End	Execution	Log Amp - 5977B SQ - Source: None	
July 6, 2022 2:41:39 PM	Start	Execution	Log Amp - 5977B SQ - Source: None	
July 6, 2022 2:41:40 PM	End	Execution	Log Amp - 5977B SQ - Source: None	

User Name: eaknarin_puangsoopa
Hostname: ASRYGW7002

System Id: RYG_EN0136
Print Date: July 7, 2022 11:27:56 AM

ALS_RYG_EN0136 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
July 6, 2022 2:41:40 PM	Start	Qualification	Session	OQ
July 6, 2022 2:41:40 PM	Start	Execution	Tune EI - 5977B SQ - Source: None	None
July 6, 2022 2:41:56 PM	End	Execution	EI - Extractor Filament 1 (Qualitative - No setpoints associated)	
July 6, 2022 2:41:58 PM	Start	Execution	Tune EI - 5977B SQ - Source: None	Run Count : 1
July 6, 2022 2:41:58 PM	Start	Execution	EI - Extractor Filament 1 (Qualitative - No setpoints associated)	
July 6, 2022 2:42:48 PM	End	Execution	Tune EI - 5977B SQ - Source: None	None
July 6, 2022 2:42:48 PM	Start	Qualification	EI - Extractor Filament 2 (Qualitative - No setpoints associated)	OQ
July 6, 2022 2:50:52 PM	End	Reporting	Session	None
July 6, 2022 2:50:52 PM	End	Reporting	Session	None
July 6, 2022 2:50:52 PM	Start	Qualification	Session	OQ
July 6, 2022 2:50:52 PM	Start	Execution	Tune EI - 5977B SQ - Source: None	None
July 6, 2022 2:51:12 PM	End	Qualification	EI - Extractor Filament 2 (Qualitative - No setpoints associated)	OQ
July 6, 2022 2:51:12 PM	Start	Reporting	Session	None
July 6, 2022 2:55:29 PM	End	Reporting	Session	None
July 6, 2022 2:55:29 PM	Start	Qualification	Session	OQ
July 6, 2022 2:55:29 PM	Start	Execution	Tune EI - 5977B SQ - Source: None	None

User Name: eaknarin_puangsoopa
Hostname: ASRYGW7002

System Id: RYG_EN0136
Print Date: July 7, 2022 11:27:56 AM

ALS_RYG_EN0136 Transaction log :
ALS_RYG_EN0136 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
July 7, 2022 10:08:18 AM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Run Count: 3
July 7, 2022 10:10:28 AM	Audit	TestUnlocked	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Deviation filed for Run Count : 3
July 7, 2022 10:10:28 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None
July 7, 2022 10:10:55 AM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data files Path : D:\OQ2022\OFN_SN_F01.D
July 7, 2022 10:14:03 AM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Run Count: 4
July 7, 2022 10:14:54 AM	Audit	TestUnlocked	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Deviation filed for Run Count : 4
July 7, 2022 10:14:54 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None
July 7, 2022 10:15:15 AM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data files Path : D:\OQ2022\OFN_SN_F01.D

User Name: eaknarin_puangsoopa
Hostname: ASRYGW7002

System Id: RYG_EN0136
Print Date: July 7, 2022 11:27:56 AM

ALS_RYG_EN0136 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
July 7, 2022 10:15:27 AM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Run Count: 5
July 7, 2022 10:16:48 AM	Audit	TestUnlocked	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Deviation filed for Run Count : 5
July 7, 2022 10:16:48 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	None
July 7, 2022 10:17:05 AM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Data files Path : D:\OQ2022\OFN_SN_F01.D
July 7, 2022 10:17:14 AM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 1 - L: >= 1200	Run Count: 6
July 7, 2022 10:18:40 AM	End	Qualification	Session	OQ
July 7, 2022 10:18:40 AM	Start	Reporting	Session	None
July 7, 2022 10:21:10 AM	End	Reporting	Session	None
July 7, 2022 10:21:10 AM	Start	Qualification	Session	OQ
July 7, 2022 10:21:17 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ: - Source: EI - Extractor using Filament 2 - L: >= 1200	None
July 7, 2022 10:56:49 AM	End	Qualification	Session	OQ
July 7, 2022 10:56:49 AM	Start	Reporting	Session	None
July 7, 2022 10:57:38 AM	End	Reporting	Session	None

User Name: eaknarin_puangsoa

Hostname: ASRYGW7002

System id: RYG_EN0136

Print Date: July 7, 2022 11:27:56 AM

ALS_RYG_EN0136 Transaction log :

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
July 7, 2022 11:19:56 AM	Audit	TestUnlocked	Signal to Noise EI - Liquid Injection, Front SSL, SQ - Source: EI - Extractor using Filament 2 - L: >= 1200	Deviation filed for Run Count : 2
July 7, 2022 11:19:56 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ - Source: EI - Extractor using Filament 2 - L: >= 1200	None
July 7, 2022 11:20:13 AM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ - Source: EI - Extractor using Filament 2 - L: >= 1200	Data files Path : D:\OQ2022\O\FN_SN_F021.D
July 7, 2022 11:21:52 AM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ - Source: EI - Extractor using Filament 2 - L: >= 1200	Run Count : 3
July 7, 2022 11:22:49 AM	End	Qualification	Session	OQ
July 7, 2022 11:22:49 AM	Start	Reporting	Session	None
July 7, 2022 11:26:46 AM	Audit	Reporting	Session	Report Generated : Certificate

Page 10 / 10

Page 10 / 10

User Name: eaknarin_puangsoa

Hostname: ASRYGW7002

ALS_RYG_EN0136 Transaction log :

System id: RYG_EN0136

Print Date: July 7, 2022 11:27:56 AM

Time	Transaction State	Activity Performed	Type of Transaction	Optional Information
July 7, 2022 10:57:38 AM	Start	Qualification	Session	OQ
July 7, 2022 10:57:38 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ - Source: EI - Extractor using Filament 2 - L: >= 1200	None
July 7, 2022 11:06:50 AM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ - Source: EI - Extractor using Filament 2 - L: >= 1200	Data files Path : D:\OQ2022\O\FN_SN_F021.D
July 7, 2022 11:11:47 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ - Source: EI - Extractor using Filament 2 - L: >= 1200	None
July 7, 2022 11:13:13 AM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ - Source: EI - Extractor using Filament 2 - L: >= 1200	Run Count: 1
July 7, 2022 11:14:29 AM	Audit	TestUnlocked	Signal to Noise EI - Liquid Injection, Front SSL, SQ - Source: EI - Extractor using Filament 2 - L: >= 1200	Deviation filed for Run Count : 1
July 7, 2022 11:14:29 AM	Start	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ - Source: EI - Extractor using Filament 2 - L: >= 1200	None
July 7, 2022 11:14:47 AM	Audit	Data	Signal to Noise EI - Liquid Injection, Front SSL, SQ - Source: EI - Extractor using Filament 2 - L: >= 1200	Data files Path : D:\OQ2022\O\FN_SN_F021.D
July 7, 2022 11:16:34 AM	End	Execution	Signal to Noise EI - Liquid Injection, Front SSL, SQ - Source: EI - Extractor using Filament 2 - L: >= 1200	Run Count: 2

Page 9 / 10

Page 9 / 10



JIRANATEE ASSOCIATES CO., LTD.

Accredited calibration laboratory
ISO/IEC 17025:2017
NSC-TIS-17025
CALIBRATION 0367

Air speed measurement laboratory
Calibration services department.

Jiranatee Associates Co., Ltd.
63/14-15, 67/35-36
Petchkasem 7/71, Rd. Watthapra, Bangkokkylai,
Bangkok 10500 (Thailand)
Tel: +6608680812
Mobile: +66863599453
E-mail: jnac-calibration@jiranatee.com
Web site: www.jiranatee.com

REVIEW BY	Mr. Parinya P.
APPROVED BY	Mr. Parinya P.
NEXT CAL. DATE	19/12/24

Certificate Number

CC-009-66

CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

MEASUREMENT ITEM

MANUFACTURER : Cup anemometer

MODEL/TYPE : Novalyx

SERIAL NUMBER : Sensor: WS-02FA

ID NUMBER : Data logger: 110-WS-25DL-D

CONDITION AS-RECEIVED : Sensor: WSD-A5966

CUSTOMER : Data logger: A5966

: BKK_F51371

: New item

: ALS laboratory group (Thailand) Co., Ltd.

: 104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,

: Khet Suan Luang, Bangkok 10250 Thailand.

RECEIVED DATE

MEASUREMENT DATE : 16 Jun 2023

ISSUE DATE : 19 Jun 2023

: 19 Jun 2023

ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature : 23.0 ± 3.0 °C

Relative Humidity : 55.0 ± 15.0 %RH

Atmospheric Pressure : 1010 ± 10 hPa

PLACE OF CALIBRATION

: Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

CALIBRATION CONDITIONS

: Wind tunnel cross-section area¹ 900 cm²

Win direction frontal area² 100 cm²

Diameter of mounting pipe³ - mm

Blockage ratio of test object⁴ 0.111 [-]

Preconditioning

Measurement Condition

: 24 hours at ambient conditions.

: The average values during measurement are (23.9) °C, (45.2) %RH and (1008.6) hPa.

TABULATION OF RESULTS:

The table on next page give the measured values.

Calibrated by:

☒ Mr. Sorawit Thachalad

☐ Miss Jitraporn Lertsomphol

Approved signatory:

Mr. Parinya Booncharoen

Calibration Department Manager



Remark:

¹ Nozzle cross-section area of the wind tunnel

² Projected cross-section area of the tested object include mounting pipe

³ Diameter of mounting pipe

⁴ Ratio ² to ¹

THIS CERTIFICATE OF CALIBRATION MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY



JIRANATEE ASSOCIATES CO., LTD.

Accredited calibration laboratory
ISO/IEC 17025:2017
NSC-TIS-17025
CALIBRATION 0367

Air speed measurement laboratory
Calibration services department.

Jiranatee Associates Co., Ltd.
63/14-15, 67/35-36
Petchkasem 7/71, Rd. Watthapra, Bangkokkylai,
Bangkok 10500 (Thailand)
Tel: +6608680812
Mobile: +66863599453
E-mail: jnac-calibration@jiranatee.com
Web site: www.jiranatee.com

CERTIFICATE OF CALIBRATION

Page 1 of 2 Pages

MEASUREMENT ITEM

MANUFACTURER : Wind Direction Sensor

MODEL/TYPE : Novalyx

SERIAL NUMBER : Data logger: 110-WS-25DL-D

ID NUMBER : Sensor: WSD-A5966

CONDITION AS-RECEIVED : Data logger: A5966

CUSTOMER : BKK_F51371

: New item

: ALS laboratory group (Thailand) Co., Ltd.

: 104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang,

: Khet Suan Luang, Bangkok 10250 Thailand.

RECEIVED DATE

MEASUREMENT DATE : 16 Jun 2023

ISSUE DATE : 19 Jun 2023

: 19 Jun 2023

ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature : 23.0 ± 3.0 °C

Relative Humidity : 55.0 ± 15.0 %RH

Atmospheric Pressure : 1010 ± 10 hPa

PLACE OF CALIBRATION

: Eiffel-type wind tunnel of Jiranatee Associates Co., Ltd.

CALIBRATION CONDITION

: Wind tunnel cross-section area¹ 900 cm²

Win direction frontal area² 129 cm²

Diameter of mounting pipe³ - mm

Blockage ratio of test object⁴ 0.143 [-]

Preconditioning

Measurement Condition

: 24 hours at ambient conditions.

: The average values during measurement are (24.3) °C, (44.7) %RH and (1010.1) hPa.

TABULATION OF RESULTS:

The table on next page give the measured values.

Calibrated by:

☒ Mr. Sorawit Thachalad

☐ Miss Jitraporn Lertsomphol

Approved signatory:

Mr. Parinya Booncharoen

Calibration Department Manager



Remark:

¹ Nozzle cross-section area of the wind tunnel

² Projected cross-section area of the tested object include mounting pipe

³ Diameter of mounting pipe

⁴ Ratio ² to ¹

THIS CERTIFICATE OF CALIBRATION MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY

Air speed m/s	D^*_{sid} Degree (°)	D^*_{acc} Degree (°)	Error Degree (°)	U ($k=2$) Degree (°)
	0.000	0	0	1.0
	45.000	43	-2	1.0
	90.000	87	-3	1.0
	135.000	132	-3	1.0
	180.000	182	2	1.0
5.01	225.000	228	3	1.0
	270.001	273	3	1.0
	315.000	318	3	1.0

⁷ Direction of Unit Under Calibration

**J
NAC**
JIRANATE ASSOCIATES CO.

JNAC
HIRANATE ASSOCIATES CO.,

Approved Signatory: 
Mr. Parinya Booncharoen
Calibration Department Manager

THIS CERTIFICATE MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.

Result of Calibration: ☒ Without Adjustment ☐ With Adjustment
Calibration Range: 20-40 °C

Function: This equipment was connected with temperature sensor Model: HMP60 S/N: V1920207.
Dimension : Diameter 12 mm. Length 80 mm.

Immersion Depth (mm)	Standard Reading (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)
70	20.058	20.0	-0.1	0.099
70	25.051	24.9	-0.2	0.099
70	30.044	29.9	-0.1	0.099
70	35.040	34.9	-0.1	0.099
70	40.034	39.8	-0.2	0.099

UUC* : Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

* End of Certificate *



CERTIFICATE OF CALIBRATION

Calibration No. : RH-02062023
Page 1 of 1 Pages

Measurement Item : Relative humidity with data logger
Manufacturer : Novalynx
Model/Type : 110-WS-25DL-D
Serial Number : A5966
ID No. : BKK_FS1371
Customer : ALS laboratory group (Thailand) Co., Ltd.
: 104 Phatthanakan 40, Phatthanakan Rd., Khwaeng Suan Luang, Khel Suan Luang, Bangkok
10250 Thailand.

Environmental Condition:

The measurement was carried out in an ambient temperature of (25±3)°C, and relative humidity of (50±15)%.

Measurement Method:

Unit Under Calibration (UUC) was calibrated by comparison method with standard chilled mirror hygrometer model: 1860-3 in the humidity generator chamber to determine the errors.

Traceability:

This instrument was calibrated using standard equipment whose accuracy is traceability through National Institute of Standards and Technology to the international system of units (SI) via MCS Calibration, Inc. Certificate number: 20926-601. Due date: Sep 26, 2024.

Measurement Date : Jun 19, 2023
Issued Date : Jun 22, 2023

Measurement Results:

This equipment was connected with Indoor air quality probe and Displayed (UR) on display. Model: HMP60, Serial number: V1920207.

Calibration was performed in the range of 20%RH to 80%RH

The results of calibration are reported in table below.

Determined (%RH)	Standard (%RH)	UUC (Reading) (%RH)	Error (%RH)	Uncertainty ±(%RH)
20	20.06	19.5	-0.6	0.53
50	50.22	50.4	0.2	0.53
80	80.21	81.5	1.3	0.53

Performed by
☐ Mr. Sorawit Thachalad
☒ Miss Jitraporn Lertsomphol
☐ Miss Ruangrumpai Phoommit



Approved Signatory:
Mr. Parinya Booncharoen.
Calibration Department Manager

CERTIFICATE OF CALIBRATION

Certificate No. : CP-003-66

Page 1 of 2 Pages

MEASUREMENT ITEM
MANUFACTURER
MODEL/TYPE

SERIAL NUMBER
ID NUMBER

CONDITION AS-RECEIVED
CUSTOMER

: Digital barometer
: Novolyne
: Sensor: 110-WS-25BP
Data logger: 110-WS-25DL-D
: Sensor: BP-A5966
Data logger: A5966
: BKK, F51371

: New item
: ALS laboratory group (Thailand) Co., Ltd.
104 Phatthanakan 40, Phatthanakan Rd,
Khwaeng Suan Luang, Khet Suan Luang,
Bangkok 10250 Thailand.

RECEIVED DATE
MEASUREMENT DATE
ISSUE DATE

: 16 Jun 2023
: 19 Jun 2023
: 19 Jun 2023

Calibration procedure:

The pressure calibration was done by in-house calibration method as WI-CL-003 according to comparison method with Digital pressure calibrator based on DKD-R 6-1

Traceability:

The measurement results are traceable to the international system of units (SI) through the NIMT (National Metrology Institute of Thailand) via Certificate number: MP-0205-22

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%.

CONDITION OF THIS RESULT OF CALIBRATION:

1. Reference Standard Instrument:

Instrument Model Serial No. Certificate No. Due Date
Absolute Pressure Transducer CPG2500 4100126P MP-0205-22 02 Dec 2023

1. Calibration effort for calibration sequence C

2. The UUC* was installed in vertical orientation above reference standard instrument and center of UUC* was used as the reference level.

3. Calibration conditions:

4. Condition
: ☒ Normal
: Air
Pressure transmitting medium
A₁ (20°C, 1 bar)
: 1.19 kg/m³
H_{amb} (55±15) %
: (23±3) °C
t_{amb}
p_{amb} (1010±10) mbar

5. The certificate is valid only to the item calibrated on date and place of calibration

Calibrated by:

☒ Mr. Sorawit Thachalad
☐ Miss Jitraporn Lertsomphol

Approved signatory:

Mr. Parinya Booncharoen
Calibration Department Manager



Parinya Booncharoen

End of certificate



CERTIFICATE OF CALIBRATION

Certificate No. : CP-003-66

Page 2 of 2 Pages

MEASUREMENT RESULTS

: ☒ Without adjustment ☐ With adjustment

CALIBRATION IN THE RANGE OF : 950 mbar to 1050 mbar

The results of calibration and associated measurement uncertainties are reported in the table below.

STD (mbar)	UUC* (mbar)	Error (mbar)	Uncertainty (k=2) (mbar)
950.03	950.6	0.6	0.77
970.14	970.4	0.3	0.51
990.04	990.1	0.1	0.39
1010.12	1010.1	-0.1	0.38
1030.09	1029.8	-0.3	0.50
1050.07	1049.6	-0.5	0.70

Note: UUC* Unit Under Calibration

: To convert the result in report unit to Pa should be multiply by 100

CERTIFICATE OF CALIBRATION

Certificate No: WD-03072022
Page 1 of 2 pages

Measurement Item : Wind direction sensor with data logger.

Manufacturer : Data logger: Novallux.
: Wind direction sensor: Novallux.

Model/Type : Data logger: 200-WS-25LB
: Wind direction sensor: WS-02F

Serial Number : Data logger: A5193
: Wind direction sensor: -

ID No : Data logger: S0K_F50036
: Wind direction sensor: -

Customer : ALS laboratory group (Thailand) co., ltd.
: 104 Phatthanakan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10250
Thailand.

Environmental Condition:

The measurement was carried out in an ambient temperature of (23±3) °C, and relative humidity of (40±10) %.

Measurement Method:

The wind direction sensor calibration according to comparison method with reference angle measurement electronic theodolite and line laser is used for axis control. The measurement were taken at 45° intervals in clockwise and counterclockwise directions.

Note: The UUC was warned up for 1 hour prior to the calibration being performed

Traceability:

The measurement results are traceable to the international system of units (SI) through Certificate No: Q21066014, Certificate No: KWS64/0025.

Measurement Date : Jul 13, 2022.
Issued Date : Jul 14, 2022.

Calibrated by

☒ Mr. Sorawit Thachalad
☐ Miss Ullraporn Lertsomphol

Approved Signatory:

Mr. Parinya Booncharoen,
Calibration Department Manager



Continuation of Certificate of Calibration Number

Certificate No: WD-03072022
Pages 2 of 2 pages

Result of calibration: ☐ Without adjustment ☒ With adjustment.

Calibration in the range of 0 – 360 ° at a calibration interval of 45°.

The results of calibration and associated measurement uncertainties are reported in table below.

NO	Turning Direction	Nominal Angle (°)	Standard Reading (°)	UUC* Reading (°)	Error (°)	Uncertainty ±(°)
1	Clockwise	0/360	360	359	-1	3.0
2		45	45	42	-3	3.0
3		90	90	87	-3	3.0
4		135	135	131	-4	3.0
5		180	180	178	-2	3.0
6		225	225	226	1	3.0
7		270	270	272	2	3.0
8		315	315	317	2	3.0
9	Counter Clockwise	0/360	360	359	-1	3.0
10		45	45	42	-3	3.0
11		90	90	87	-3	3.0
12		135	135	131	-4	3.0
13		180	180	178	-2	3.0
14		225	225	226	1	3.0
15		270	270	272	2	3.0
16		315	315	317	2	3.0

UUC*: Unit Under Calibration The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

End of Certificate of Calibration





63/14-15,67/35-36, Soi Petchkasem7,7/1, Petchkasem Rd,
Wathapra, Bangkokkayai,Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranatee.com

JIRANATE ASSOCIATES CO., LTD.

CERTIFICATE OF CALIBRATION

Certificate No: WS-03072022
Page 1 of 2 pages

Measurement Item : Cup anemometer with data logger.
Manufacturer : Data logger: Novallnx
: Cup anemometer: Novallnx
Model/Type : Data logger: 200-WS-25LB
: Cup anemometer: WS-02F
Serial Number : Data logger: A5193
: Cup anemometer: -
ID No : Data logger: SSK-FS0036
: Cup anemometer: -
Customer : ALS laboratory group (Thailand) co. ltd.
: 104 Phatthanakan 40, Phatthanakan Rd, Khwaeng Suan Luang, Khet Suan Luang, Bangkok 10260 Thailand.
Test Conditions : Wind tunnel cross test section area 900 cm²
: Anemometer frontal area 100 cm²
: Diameter of mounting pipe mm
: Blockage ratio of test object 0.111 [-]
Test Conditions : Air temperature 23.5 ±0.8 °C
: Air pressure 1004.2 ±0.4 hPa
: Relative air humidity 44.5 ±3.5 %RH

Calibration Procedure : Calibration was carried out base on:
ISO 61400-12-1 6D.1: 2005-Power Performance Measurements of Electricity Producing Wind Turbines;
IEC61400-12-1 6D.1: 2005-Power Performance Measurements of Electricity Producing Wind Turbines;
IEC61400-12-1 6D.1: 2005-Power Performance Measurements of Electricity Producing Wind Turbines;

Traceability : This calibration documents the traceable to national standard, which realize the unit of measurements according to the international system of units (SI) through National Institute of Metrology Thailand (NIMT).

Measurement Date : Jul 13, 2022.
Issued Date : Jul 14, 2022.

Calibrated by
☒ Mr. Sravit Thachalad
☐ Miss Jitraporn Lertsonphol



Approved Signatory:
Mr. Parinya Booncharoen
Calibration Department Manager



63/14-15,67/35-36, Soi Petchkasem7,7/1, Petchkasem Rd,
Wathapra, Bangkokkayai,Bangkok 10600 Thailand.
Tel: (66) 02-8680812#13 Fax: (66) 02-8680860 www.jiranatee.com

JIRANATE ASSOCIATES CO., LTD.

Continuation of Certificate of Calibration Number

Certificate No: WS-03072022
Page 2 of 2 Pages

Result of calibration: ☒ Without adjustment ☐ With adjustment
Calibration in the range of 1 - 16 m/s at a calibration interval of 1 m/s.
The results of calibration and associated measurement uncertainties are reported in the table below.

V _{iso} Reading m/s	V _{unc} Reading m/s	Error (m/s)	Uncertainty (%)
2.061	1.9	-0.2	2.6
4.123	3.9	-0.2	1.3
6.02	5.9	-0.1	1.3
7.99	8.0	0.0	0.86
9.98	10.0	0.0	0.87
12.01	12.1	0.1	0.87
14.00	14.1	0.1	0.83
15.98	16.3	0.3	0.61
15.00	16.1	0.1	0.56
13.01	13.2	0.2	0.74
10.98	11.1	0.1	0.63
8.99	9.0	0.0	0.70
6.98	7.0	0.0	1.1
5.162	5.0	-0.2	1.0
3.012	2.8	-0.2	1.7
1.044	0.8	-0.2	4.8

UUC*: Unit Under Calibration

The reported expanded uncertainty is based on standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%

Appendix 1: Instrumentations

NO	Sensor	Manufacturer	Model/Type	Calibration Date	Certificate Report Number	Range
1	Pitot static	TESTO INC.	D0352145	Aug 07, 2021	MW-0034-21	5 - 30 m/s
2	Precision Differential Pressure Meter	Zoglab	DPM2600	Aug 07, 2021	MW-0034-21	5 - 30 m/s
3	Air velocity transducer (hot wire)	TSI INC.	B456-12	Aug 08, 2021	MW-0035-21	0 - 5 m/s
4	Temperature	Zoglab	DSR-THP	March 30, 2022	CL-027-65	-30 - 70 °C
5	Relative humidity	Zoglab	DSR-THP	March 30, 2022	RH-03032022	0 - 100 %RH
6	Atmospheric pressure	Zoglab	DSR-THP	March 30, 2022	BP-01032022	500 - 1100 hPa
7	Wind tunnel	ESSOM	MP330D	-	-	0 - 50 Hz

End of certificate of calibration



THIS CERTIFICATE MAY NOT BE REPRODUCED EXCEPT IN FULL UNLESS PERMISSION FOR REPRODUCTION HAS BEEN OBTAINED IN WRITING FROM THE LABORATORY.

SITHIPORN ASSOCIATES CO.,LTD. CALIBRATION LABORATORY

451-451/1 Sirinthorn Rd, Bangbunru, Bangplud Bangkok 10700 THAILAND.
Tel:0-2435-8800 Fax:0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.com



Cert. No. : ACC22038
Pages : 1 of 3

Calibration Certificate

Equipment : SOUND CALIBRATOR

Manufacturer : RION

Model : NC-75

Serial No.: 35024429

ID No.: - S&V-FS 0114

Condition As Found : GOOD

Customer :
ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWANG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location :

Ambient Temperature :

Pressure :

Relative Humidity :

Received Date :

Calibration Date :

Date of Issue :

Calibrated by :

Nathakorn Pisutpaisan

Approved by :

(Thanakul Petchurai)

T. Petchurai

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Continuation of Calibration Certificate

Cert. No. : ACC22038
Job No. : VC66AC0006
Pages : 2 of 3

Calibration Procedure : CP-AC-03

Calibration Method :

This equipment was calibrated by based on IEC-60942-2003 Standard.

The sound pressure level, frequency and total distortion of the sound calibrator was measured using the reference microphone.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33511B	MY52302742	EF-0008-22	04-Feb-23
Digital Multimeter	33461A	MY53220104	EEL.BP. 04/0265	09-Feb-23
Digital Multimeter	33461A	MY53220076	EEL.BP. 03/0265	09-Feb-23
Digital Multimeter	33461A	MY60024273	EEL.BP. 05/0265	09-Feb-23
Programmable Attenuator	MAT-1070	62100114	EF-0009-22	07-Feb-23
Condenser Microphone	4180	2977900	AA-1013-22	24-Feb-23
Measuring Amplifier	NA-42KAI	34560495	AA-3005-22	22-Feb-23
Audio Analyzer	AVR-3360A	V744B6069	EF-0010-22	07-Feb-23

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

T. Petchurai

Continuation of Calibration Certificate

Cert. No. : ACC22038

Job No. : VC66AC0006

Pages : 3 of 3

Result of calibration :

1. Sound pressure level

Specified sound pressure level (dB)	Measured value (dB)	Deviated value (dB)	Uncertainty (dB)	Tolerance limit (dB)
94	93.94	-0.06	0.14	0.40

2. Frequency

Specified Frequency (Hz)	Measured value (Hz)	Deviated value (%)	Uncertainty (%)	Tolerance limit (%)
1000	1000.0	0.0	0.1	1.0

3. Total distortion

Measured value (%)	Uncertainty (%)	Tolerance limit (%)
0.33	0.10	3.0

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

— End of Calibration Certificate —

451-451/1 Sittithorn Rd.,Bangbunmru, Bangplud Bangkok 10700 THAILAND

Tel.0-2435-8800 Fax.0-2433-1679 e-mail:cal-center@sithiporn.com http://www.sithiporn.com

Cert. No. : ACL23106

Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
Manufacturer : RION
Model : NL-52A / Microphone UC-59 / Preamplifier NH-25
Serial No. : 00331095 / 22927 / 22570
ID No. : -

Condition As Found : GOOD

Customer : ALS LABORATORY GROUP (THAILAND) CO., LTD.
104 PHATTHANAKAN 40, PHATTHANAKAN ROAD,
KHWAEANG PHATTHANAKAN, KHET SUAN LUANG,
BANGKOK, 10250 THAILAND.

Location : -
Ambient Temperature : (23.0 ± 3) °C
Pressure : (101.3 ± 3) kPa
Relative Humidity : (50.0 ± 20) %

Received Date : 17 MARCH 2023
Calibration Date : 27-28 MARCH 2023
Date of Issue : 31 MARCH 2023

Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchurai
(Thanakul Petchurai)



This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Continuation of Calibration Certificate

Cert. No. : ACL23106
Job No. : VC66AC0042
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.

For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0009-23	07-FEB-24
Waveform Generator	33511B	MY52302742	EF-0010-23	07-FEB-24
Digital Multimeter	33461A	MY53220104	EEL-BP 30/0266	13-FEB-24
Digital Multimeter	33461A	MY53220076	EEL-BP 29/0266	13-FEB-24
Digital Multimeter	34461A	MY60024273	EEL-BP 31/0266	14-FEB-24
Programmable Attenuator	MAT-1070	62100114	EF-0011-23	08-FEB-24
Condenser Microphone	4180	2977900	AA-1001-23	14-FEB-24
Measuring Amplifier	NA-42KAI	34560495	AA-3002-23	14-FEB-24

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

- 3.1 National Institute of Metrology (Thailand).
- 3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL23106
Job No. : VC66AC0042
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.2	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.4	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	✓	-	0.3	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

Note : Pass/Fail evaluation for each parameter, will be considered together from the acceptance limit and the Maximum-permitted uncertainty of measurement.

Continuation of Calibration Certificate

Cert. No. : ACL23106

Job No. : VC66AC0042

Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93.9 (93.98)	94.0	0.0	±0.3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
13.3

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	8.7
C - weight	14.3
Flat	19.8

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight
125	0.3	0.3	0.3
1000	0.2	0.2	0.2
8000	0.0	0.0	0.1
+ 1.5, - 2.5			

Continuation of Calibration Certificate

Cert. No. : ACL23106

Job No. : VC66AC0042

Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)		
	Flat	C-weight	A-weight
63	0.0	0.0	0.0
125	0.0	0.1	0.0
250	0.0	0.0	0.0
500	0.0	0.0	0.0
1000	0.0	0.0	0.0
2000	0.0	0.0	0.0
4000	0.0	0.0	0.0
8000	0.0	0.1	0.1
16000	0.0	-1.2	-1.2
+ 1.5, - 2.5			
+ 2.5, -16.0			

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.2
C - weight	94.0	94.0	0.0	± 0.2
Flat	94.0	94.0	0.0	± 0.2

5.2 Time weighting at 1 kHz

Frequency Weighting	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94.0	94.0	0.0	± 0.1
Slow	94.0	94.0	0.0	± 0.1
Leq	94.0	94.0	0.0	± 0.1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94.0	94.0	0.0	± 0.1

Continuation of Calibration Certificate

Cert. No. : ACL23106
Job No. : VC66AC0042
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.0	0.0	±0.8
136.0	136.0	0.0	±0.8
135.0	135.0	0.0	±0.8
134.0	134.0	0.0	±0.8
133.0	133.0	0.0	±0.8
132.0	132.0	0.0	±0.8
131.0	131.0	0.0	±0.8
129.0	129.0	0.0	±0.8
124.0	124.0	0.0	±0.8
119.0	119.0	0.0	±0.8
114.0	114.0	0.0	±0.8
109.0	109.0	0.0	±0.8
104.0	104.0	0.0	±0.8
99.0	99.0	0.0	±0.8
94.0	94.0	0.0	±0.8
89.0	89.0	0.0	±0.8
84.0	84.0	0.0	±0.8
79.0	79.0	0.0	±0.8
74.0	74.1	0.1	±0.8
69.0	69.1	0.1	±0.8
64.0	64.0	0.0	±0.8
59.0	59.1	0.1	±0.8
54.0	54.0	0.0	±0.8
49.0	49.0	0.0	±0.8
44.0	44.0	0.0	±0.8
39.0	39.0	0.0	±0.8
34.0	34.0	0.0	±0.8
30.0	30.0	0.0	±0.8
29.0	29.0	0.0	±0.8
28.0	28.0	0.0	±0.8
27.0	27.0	0.0	±0.8
26.0	26.0	0.0	±0.8
25.0	24.9	-0.1	±0.8

Continuation of Calibration Certificate

Cert. No. : ACL23106
Job No. : VC66AC0042
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±0.8

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.5 ; -5.0
	2	8	117.0	117.0	0.0	1.0 ; -2.5
	200	800	134.0	134.0	0.0	±1.0
Slow	2	8	108.0	108.0	0.0	1.5 ; -5.0
	200	800	127.6	127.6	0.0	±1.0
SEL	0.25	1	99.0	98.9	-0.1	1.5 ; -5.0
	2	8	108.0	108.0	0.0	1.0 ; -2.5
	200	800	128.0	128.0	0.0	±1.0

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	±2.0
One	136.4	135.6	-0.8	±2.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	±1.0
Positive half cycle	135.4	135.2	-0.2	±1.0
Negative half cycle	135.4	135.2	-0.2	±1.0

Continuation of Calibration Certificate

Cert. No. : ACL23106
Job No. : VC66AC0042
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle	-0.1	±1.5
89.6	89.5		

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.1

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$ or any value following calculation, providing a level of confidence of approximately 95 %

_____ End of Calibration Certificate _____