

ภาคผนวกที่ 4

เอกสารสอบเทียบความถูกต้องของเครื่องมือตรวจวัดคุณภาพสิ่งแวดล้อม

ลำดับที่ 1 คุณภาพอากาศในบรรยากาศ

ลำดับที่ 2 ระดับเสียงในบรรยากาศ

ตารางสรุปรายการเอกสารการสอบเทียบความถูกต้องของเครื่องมือเก็บตัวอย่าง
และเครื่องมือตรวจวิเคราะห์คุณภาพสิ่งแวดล้อม

รายการตรวจวัด	เครื่องมือเก็บตัวอย่าง	เครื่องมือตรวจวิเคราะห์
	ชื่อเครื่องมือ	ชื่อเครื่องมือ
1. คุณภาพอากาศ		
TSP	High Volume Air Sampler No. Ro๑, Ro๓	Digital Balance
PM _{๑๐}	High Volume PM-๑๐ Air Sampler No. Ro๒, Ro๓, Ro๔	Digital Balance
2. ระดับเสียง		
L _{max}	Acoustic Calibrator Sound Level Meter No. ACO- Ro๓, Ro๔, Ro๕, Ro๖, R๒๑, R๒๒, R๒๓, R๒๔, R๒๕, R๒๖, R๒๗, R๒๘, R๒๙, R๒๑, R๒๒, R๒๓, R๒๔, R๒๕, R๒๖, R๒๗, CR- Bo๓, Bo๔, Bo๕, Bo๖	-

ลำดับที่ 1

คุณภาพอากาศในบรรยากาศ



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด

S.P.S. CONSULTING SERVICE CO., LTD.

7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900

7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900

Tel : (662) 939-4370-72 Fax : (662) 513-4221 E-mail : sale@spscon.com, www.spscon.com

High Volume Air Sampler Calibration Report

Calibration Method : Multipoint Orifice Flow Transfer Standard

Model : TE 5025A

S/N : 3611


Calibration Data

High Volume Air Sampler Data		Calibration Data		
Recorder No.	Blower No.	Date	Actual Flowrate (ft ³ /min)	R ²
B35	B35	03/08/2023	y = 1.221x-4.116	0.995
B36	B36	03/08/2023	y = 1.247x-6.537	0.999
B37	B37	03/08/2023	y = 1.313x-8.352	0.997
B38	B38	03/08/2023	y = 1.279x-8.340	0.998
B39	B39	03/08/2023	y = 1.286x-6.520	0.999
B40	B40	03/08/2023	y = 1.241x-6.104	1.000
B41	B41	03/08/2023	y = 1.203x-4.249	0.999
B42	B42	03/08/2023	y = 1.296x-8.828	0.999
B43	B43	04/08/2023	y = 1.245x-5.710	0.997
B44	B44	04/08/2023	y = 1.262x-5.417	0.999
R01	R01	04/08/2023	y = 1.285x-8.953	0.999
R02	R02	04/08/2023	y = 1.268x-8.283	0.998
R03	R03	04/08/2023	y = 1.283x-9.563	0.999
R04	R04	04/08/2023	y = 1.234x-5.231	0.999
R05	R05	04/08/2023	y = 1.303x-10.505	0.999
R06	R06	04/08/2023	y = 1.287x-7.927	0.997
R07	R07	04/08/2023	y = 1.084x+0.577	0.999
R08	R08	04/08/2023	y = 1.304x-9.687	0.998
R09	R09	04/08/2023	y = 1.286x-8.387	0.998
R10	R10	03/08/2023	y = 1.241x-6.099	0.996
R11	R11	03/08/2023	y = 1.112x-1.473	0.998
R12	R12	03/08/2023	y = 1.250x-6.933	0.997
R13	R13	02/08/2023	y = 1.142x-2.480	0.998
R14	R14	02/08/2023	y = 1.205x-3.813	0.998
R15	R15	01/08/2023	y = 1.160x-3.518	0.999
R16	R16	01/08/2023	y = 1.229x-7.416	0.998
R17	R17	01/08/2023	y = 1.209x-4.808	0.998
R18	R18	01/08/2023	y = 1.257x-6.979	0.999
R19	R19	01/08/2023	y = 1.256x-7.676	0.998
R20	R20	01/08/2023	y = 1.279x-8.603	0.998

Calibrated by :



Approved by :





บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
 7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
 7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
 Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

High Volume Air Sampler Calibration Report

Calibration Method : Multipoint Orifice Flow Transfer Standard


Model : TE 5025A

S/N : 3611

Calibration Data

High Volume Air Sampler Data		Calibration Data		
Recorder No.	Blower No.	Date	Actual Flowrate (ft ³ /min)	R ²
B35	B35	01/11/2023	y = 1.247x-5.373	0.999
B36	B36	02/11/2023	y = 1.190x-2.630	0.995
B37	B37	01/11/2023	y = 1.188x-2.249	0.999
B38	B38	01/11/2023	y = 1.191x-5.051	0.995
B39	B39	01/11/2023	y = 1.230x-3.335	0.995
B40	B40	02/11/2023	y = 1.172x-2.695	0.998
B41	B41	02/11/2023	y = 1.169x-2.206	0.998
B42	B42	02/11/2023	y = 1.212x-5.591	0.998
B43	B43	03/11/2023	y = 1.223x-3.058	0.997
B44	B44	01/11/2023	y = 1.194x-2.207	0.996
R01	R01	08/11/2023	y = 1.199x-4.374	0.998
R02	R02	06/11/2023	y = 1.229x-6.243	0.999
R03	R03	08/11/2023	y = 1.239x-7.264	0.998
R04	R04	09/11/2023	y = 1.182x-3.161	0.998
R05	R05	09/11/2023	y = 1.141x-2.095	0.997
R06	R06	03/11/2023	y = 1.155x-2.543	0.997
R07	R07	09/11/2023	y = 1.057x+1.380	0.999
R08	R08	02/11/2023	y = 1.230x-6.615	0.997
R09	R09	04/11/2023	y = 1.188x-1.331	0.997
R10	R10	04/11/2023	y = 1.213x-3.571	0.998
R11	R11	01/11/2023	y = 1.136x-2.259	0.999
R12	R12	01/11/2023	y = 1.145x-3.404	0.998
R13	R13	02/11/2023	y = 1.076x-0.153	0.999
R14	R14	02/11/2023	y = 1.166x+1.197	0.996
R15	R15	09/11/2023	y = 1.171x-4.139	0.997
R16	R16	09/11/2023	y = 1.142x-3.462	0.999
R17	R17	01/11/2023	y = 1.169x-3.932	0.998
R18	R18	01/11/2023	y = 1.192x-4.280	0.998
R19	R19	01/11/2023	y = 1.158x-4.004	0.996
R20	R20	02/11/2023	y = 1.191x-4.426	0.997

Calibrated by :



Approved by :





บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด

S.P.S. CONSULTING SERVICE CO., LTD.

7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900

7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900

Tel : (662) 939-4370-72 Fax : (662) 513-4221 E-mail : sale@spscon.com, www.spscon.com

High Volume PM-10 Air Sampler Calibration Report

Calibration Method : Multipoint Orifice Flow Transfer Standard


Model : TE 5025A

S/N : 3611


Calibration Data

High Volume PM-10 Data		Calibration Data		
Recorder No.	Blower No.	Date	Actual Flowrate (ft ³ /min)	R ²
R01	R01	01/08/2023	y = 1.244x-6.482	0.998
R02	R02	01/08/2023	y = 1.274x-6.620	0.998
R03	R03	02/08/2023	y = 1.268x-7.028	0.999
R04	R04	02/08/2023	y = 1.259x-8.726	0.998
R05	R05	02/08/2023	y = 1.210x-5.918	0.999
R06	R06	02/08/2023	y = 1.246x-5.062	0.998
R07	R07	02/08/2023	y = 1.222x-4.950	0.999
R08	R08	02/08/2023	y = 1.287x-8.890	0.998
R09	R09	02/08/2023	y = 1.245x-8.340	0.998
R10	R10	02/08/2023	y = 1.228x-6.133	0.999
R11	R11	04/08/2023	y = 1.282x-6.014	0.997
R12	R12	04/08/2023	y = 1.303x-9.748	0.998
R13	R13	04/08/2023	y = 1.305x-8.462	0.997
R14	R14	04/08/2023	y = 1.299x-7.936	0.997
R15	R15	02/08/2023	y = 1.291x-7.250	0.999
R16	R16	02/08/2023	y = 1.275x-7.402	0.995
R17	R17	02/08/2023	y = 1.292x-8.739	0.999
R18	R18	02/08/2023	y = 1.215x-5.881	0.998
R19	R19	02/08/2023	y = 1.273x-7.472	0.999
R20	R20	02/08/2023	y = 1.288x-10.306	0.997

Calibrated by :



Approved by :





บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
 7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
 7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
 Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com., www.spscon.com

High Volume PM-10 Air Sampler Calibration Report

Calibration Method : Multipoint Orifice Flow Transfer Standard


Model : TE 5025A

S/N : 3611


Calibration Data

High Volume PM-10 Data		Calibration Data		
Recorder No.	Blower No.	Date	Actual Flowrate (ft ³ /min)	R ²
R01	R01	01/11/2023	y = 1.166x-2.827	0.999
R02	R02	01/11/2023	y = 1.140x+0.079	0.998
R03	R03	02/11/2023	y = 1.176x-2.983	0.999
R04	R04	01/11/2023	y = 1.155x-3.417	0.996
R05	R05	01/11/2023	y = 1.193x-5.553	0.998
R06	R06	08/11/2023	y = 1.173x-2.245	0.998
R07	R07	08/11/2023	y = 1.183x-3.521	0.999
R08	R08	06/11/2023	y = 1.195x-5.732	0.997
R09	R09	01/11/2023	y = 1.161x-3.857	0.995
R10	R10	09/11/2023	y = 1.181x-4.105	0.998
R11	R11	09/11/2023	y = 1.155x-0.430	0.996
R12	R12	09/11/2023	y = 1.196x-4.411	0.999
R13	R13	09/11/2023	y = 1.188x-3.031	0.996
R14	R14	01/11/2023	y = 1.167x-2.357	0.998
R15	R15	02/11/2023	y = 1.186x-3.860	0.998
R16	R16	02/11/2023	y = 1.187x-2.636	0.999
R17	R17	02/11/2023	y = 1.169x-3.318	0.996
R18	R18	03/11/2023	y = 1.190x-4.826	0.997
R19	R19	01/11/2023	y = 1.152x-2.131	0.996
R20	R20	01/11/2023	y = 1.183x-4.568	0.997

Calibrated by :



Approved by :



**QUALITY CALIBRATION CO.,LTD.**

235 Petchkasem 63/2 Road, Laksong, Bangkae, Bangkok 10160

Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4584

www.qcalibration.com

CERTIFICATE No : 23M2441

REFERENCE No : 68471-1

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : DIGITAL BALANCE

MANUFACTURER : METTLER TOLEDO

MODEL : XS105DU

SERIAL No : 1126422905

ID No : BA 05/50

CONDITION AS RECEIVED : USED ITEM

SUBMITTED BY : S.P.S. CONSULTING SERVICE CO., LTD.
7 SOI PHAHOLYOTHIN 24, PHAHOLYOTHIN RD.,
JOMPOL, CHATUCHAK, BANGKOK 10900

CALIBRATED BY : ATSAWIN Y.

CALIBRATION DATE : 10-Mar-23

APPROVED BY : [REDACTED]
PONGSAK J.

ISSUED DATE : 16-Mar-23

RECEIVED DATE : 10-Mar-23

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.



CERTIFICATE No : 23M2441

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : DIGITAL BALANCE MODEL : XS105DU
MANUFACTURER : METTLER TOLEDO S/N : 1126422905
ID No : BA 05/50 RECEIVED DATE : 10-Mar-23
AIR PRESSURE : 1010mbar \pm 1mbar CALIBRATION DATE : 10-Mar-23
AMBIENT TEMPERATURE : 23°C \pm 1°C RELATIVE HUMIDITY : 49 %RH \pm 10 % RH

CONDITION OF THIS RESULTS OF CALIBRATION

1. THIS INSTRUMENT WAS CALIBRATED BY ACCORDING TO UKAS LAB 14 EDITION 6:2019 BY USING KNOWN WEIGHT STANDARD WEIGHT. THE BALANCE WAS NOT ADJUSTED BEFORE CALIBRATION. THE BALANCE HAS NO ZERO TRACKING FUNCTION. REPEATABILITY WAS MEASURED BY USING 10 REPEATED MEASUREMENTS. LINEARITY WAS MEASURED COVERING 10 POINTS, EVENLY SPREAD OVER THE RANGE. THE INSTRUMENT WAS SET ZERO BEFORE PERFORMING THE LINEARITY TEST. OFF-CENTER LOADING WAS MEASURED BY USING STANDARD WEIGHTS PLACED ON THE PAN AND MOVED TO VARIOUS POSITIONS ON THE PAN.

2. REFERENCE STANDARD INSTRUMENTS :-

INSTRUMENT	MODEL	SERIAL No	CERTIFICATE No	DUE DATE
1) STANDARD WEIGHT SET	E2	QK-I-151	M2302013S	02-Feb-25
2) STANDARD WEIGHT	E2	15843	M2302014S	02-Feb-25

3. THIS RESULT WAS FOUND ACCURATE AS SHOWN ON DATE AND PLACE OF CALIBRATION ONLY.

4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.

5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-

- NATIONAL INSTITUTE OF METROLOGY (THAILAND) THROUGH CENTRAL BUREAU OF WEIGHTS&MEASURES

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT

1. ZERO SETTING FUNCTION : NORMAL

2. TARE FUNCTION : NORMAL

3. REPEATABILITY OF READING AT 200 g WAS 0 g

4. DEPARTURE FROM NOMINAL VALUE/ LINEARITY

NOMINAL VALUE (g)	BALANCE READING (g)	CORRECTION (g)	UNCERTAINTY (\pm g)
0.00	0.00000	0.00000	0.000039
0.02	0.02000	0.00000	0.000039
0.10	0.10000	0.00000	0.000039
0.20	0.20001	-0.00001	0.000040
0.50	0.50001	-0.00001	0.000040
1.00	1.00000	0.00000	0.000041
2.00	2.00003	-0.00003	0.000042
5.00	5.00001	-0.00001	0.000046
10.00	10.00003	-0.00003	0.000053
20.00	20.00005	-0.00005	0.000067
50.00	50.00001	-0.00001	0.00011
100.00	100.00001	-0.00001	0.00019
200.00	200.00001	-0.00001	0.00032

5. OFF CENTER LOADING ERROR



POINT	READING (g)
1	50.0000
2	50.0001
3	50.0000
4	50.0000
5	49.9999
OFF-CENTER LOADING	0.0001

NOTE: THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA

THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR $k=2$, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.

END OF CALIBRATION REPORT

ลำดับที่ 2

ระดับเสียงในบรรยากาศ

THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-65/0455

MTC No. EEL. BP. 41/0465

CALIBRATION CERTIFICATE

Submitted by : S.P.S. Consulting Service Co.,Ltd.

Address : 7 Soi Phaholyothin 24, Phaholyothin Road, Jompol, Chatuchak, Bangkok 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.
: Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

Instrument Calibrated :

Description : Sound Calibrator

Manufacturer : ACO

Model : 2127

Serial No. : 130006

Ambient Environment

Temperature : $(23 \pm 3) ^\circ\text{C}$

Relative Humidity : $(50 \pm 15) \%$

Ambient Pressure : $(101.325 \pm 1.500) \text{ kPa}$

Standards used : 1. Digital Function Synthesizer NF Electronic DF-193A S/N 122037.

2. Measuring Amplifier Bruel&Kjaer 2636 S/N 1537484.

3. Programmable Attenuator Tamagawa TPA-303A S/N OF 2214.

4. Digital Multimeter Agilent 34401A S/N MY44005560.

5. Pressure Transmitter Vaisala PTB202AD S/N T0650001.

6. Audio Analyzer Keithley 2015-P S/N 4106495.

7. Condenser Microphone Bruel&Kjaer 4180 S/N 2889871.

Calibration Procedure: CP-102-04 based on IEC 60942-2003. The sound pressure level of instrument was measured by standard microphone using an insert voltage technique.

This instrument has been calibrated against standards maintained at Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

Date of Receipt : 22 Apr. 2022

Date of Calibration : 28 Apr. 2022

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.4

Head Office
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website:www.tistr.or.th

Office/Laboratory
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office
196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-65/0455

MTC No. EEL. BP. 41/0465

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.

Nominal Output of Unit Under Test = 94 dB re 20 μ Pa at 1000 Hz

Acoustic Output in dB re 20 μ Pa, Corrected to Reference Conditions: 101.325 kPa, 23.0 °C and 50 %RH.

1. Sound Pressure Level

Standard Microphone Type	Measured Sound Pressure Level (dB)	Deviated value (dB)	Uncertainty (dB)	Tolerance limit IEC60942:2003 Class 1
1/2 inch Bruel&Kjaer 4180	93.93	-0.07	± 0.10	± 0.40 dB

2. Frequency


Standard Microphone Type	Measured Frequency (Hz)	Deviated value (Hz)	Uncertainty (Hz)	Tolerance limit IEC60942:2003 Class 1
1/2 inch Bruel&Kjaer 4180	999.9	-0.1	± 1.5	$\pm 1.0\%$

3. Total Distortion

Standard Microphone Type	Measured Total Distortion (%)	Uncertainty (%)	Tolerance limit IEC60942:2003 Class 1
1/2 inch Bruel&Kjaer 4180	1.44	± 0.50	$\pm 3.0\%$

- Note :
1. No adjustment.
 2. The calibrator pressure correction was not included.
 3. The microphone volume correction was not included.

Calibrated by :



Approved by :



Director

Electrical and Electronic Standards Laboratory
Industrial Metrology and Testing Service Centre

Date of Calibration : 28 Apr. 2022

Date of Issue : 28 Apr. 2022

Ref : 2011265042601787001

2 / 2

End of Certificate

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.4



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-66/0358

MTC No. EEL. BP. 22/0366

CALIBRATION CERTIFICATE

Submitted by : S.P.S. Consulting Service Co., Ltd.

Address : 7 Soi Phaholyothin 24, Phaholyothin Road, Jompol, Chatuchak, Bangkok 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

Instrument Calibrated :

Description : Sound Calibrator

Manufacturer : Cirrus

Model : CR:515

Serial No. : 92002

Ambient Environment

Temperature : $(23 \pm 3) ^\circ\text{C}$

Relative Humidity : $(50 \pm 15) \%$

Ambient Pressure : $(101.325 \pm 1.500) \text{ kPa}$

Standards used : 1. Digital Function Synthesizer NF Electronic DF-193A S/N 122037.

2. Measuring Amplifier Bruel&Kjaer 2636 S/N 1537484.

3. Programmable Attenuator Tamagawa TPA-303A S/N OF 2214.

4. Digital Multimeter Agilent 34401A S/N MY44005560.

5. Pressure Transmitter Vaisala PTB202AD S/N T0650001.

6. Audio Analyzer Keithley 2015-P S/N 4106495.

7. Condenser Microphone Bruel&Kjaer 4180 S/N 2889871.

Calibration Procedure: CP-102-04 based on IEC 60942-2003. The sound pressure level of instrument was measured by standard microphone using an insert voltage technique.

This instrument has been calibrated against standards maintained at Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

Date of Receipt : 3 Mar. 2023

Date of Calibration : 13 Mar. 2023

1 / 2

The results relate only to the items tested/calibrated or value assigned.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.4

Head Office

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website:www.tistr.or.th

Office/Laboratory

Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office

196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th

THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-66/0358

MTC No. EEL. BP. 22/0366

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.

Nominal Output of Unit Under Test = 94 dB re 20 μ Pa at 1000 Hz

Acoustic Output in dB re 20 μ Pa, Corrected to Reference Conditions: 101.325 kPa, 23.0 °C and 50 %RH.

1. Sound Pressure Level

Standard Microphone Type	Measured Sound Pressure Level (dB)	Deviated value (dB)	Uncertainty (dB)	Tolerance limit IEC60942:2003 Class 1
1/2 inch Bruel&Kjaer 4180	93.99	-0.01	± 0.10	± 0.40 dB

2. Frequency

Standard Microphone Type	Measured Frequency (Hz)	Deviated value (Hz)	Uncertainty (Hz)	Tolerance limit IEC60942:2003 Class 1
1/2 inch Bruel&Kjaer 4180	1000.3	0.3	± 1.5	$\pm 1.0\%$

3. Total Distortion

Standard Microphone Type	Measured Total Distortion (%)	Uncertainty (%)	Tolerance limit IEC60942:2003 Class 1
1/2 inch Bruel&Kjaer 4180	1.39	± 0.50	$\pm 3.0\%$

- Note :
1. No adjustment.
 2. The calibrator pressure correction was not included.
 3. The microphone volume correction was not included.

Calibrated by :



Approved by :



Director

Electrical and Electronic Standards Laboratory

Industrial Metrology and Testing Service Centre

Date of Calibration : 13 Mar. 2023

Date of Issue : 14 Mar. 2023

Ref : 2011266030300928001

2 / 2

End of Certificate

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.4

Head Office

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9000
Fax. (66) 0 2577 9009
E-mail : rumpai@tistr.or.th Website:www.tistr.or.th

Office/Laboratory

Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,
Amphoe Muang, Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
Fax. (66) 0 2323 9165
E-mail : mtc@tistr.or.th

Office

196 Phahonyothin Road, Chatuchak, Bangkok 10900,
Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
Fax. (66) 0 2579 8592
E-mail : sumalee@tistr.or.th



Noise R_439/23

Sound Level Meter Calibration Report


Acoustic Calibrator Data

Brand	ACO	Number	AC 03/56
Model	2127	Serial No.	130006
Calibration Range	94 dB, 1000 Hz	Last Calibration	29 March 2023
		Due Date	29 March 2024


Calibration Data

Sound Level Meter Data				Calibration Data		
SLM No.	Brand	Model	Serial No.	Date	Actual Reading [dB]	
					Before Adjustment	After Adjustment
ACO-R๐๙	ACO	๖๒๓๖	๐๐๑๗๒๐๓๕	๒๑ August ๒๐๒๓	๙๕.๑	๙๕.๐
ACO-R๓๑	ACO	๖๒๓๖	๐๐๑๙๒๐๔๓	๒๑ August ๒๐๒๓	๙๕.๐	๙๕.๐
ACO-R๔๖	ACO	๖๒๓๖	๐๐๑๙๒๐๕๘	๒๑ August ๒๐๒๓	๙๕.๐	๙๕.๐
ACO-R๔๘	ACO	๖๒๓๖	๐๐๑๙๒๐๖๐	๒๑ August ๒๐๒๓	๙๕.๐	๙๕.๐
Acoustic Certified Value : Thailand Institute of Scientific and Technological Research (TISTR)					93.94 ± 0.10 dB	

Calibrated by :



Approved by :





บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com., www.spscon.com

Noise R_473/23

Sound Level Meter Calibration Report


Acoustic Calibrator Data

Brand	CIRRUS	Number	AC-CR01/63
Model	CR515	Serial No.	92002
Calibration Range	94 dB, 1000 Hz	Last Calibration	13 March 2023
		Due Date	13 March 2024


Calibration Data

Sound Level Meter Data				Calibration Data		
SLM No.	Brand	Model	Serial No.	Date	Actual Reading [dB]	
					Before Adjustment	After Adjustment
CR-B๐๓	Cirrus	CR๑๖๓B	G๓๐๑๑๕๕	๑๐ September ๒๕๖๓	๙๕.๐	๙๕.๐
CR-B๐๕	Cirrus	CR๑๖๓B	G๓๐๑๑๓๔	๑๐ September ๒๕๖๓	๙๕.๑	๙๕.๐
CR-B๐๖	Cirrus	CR๑๖๓B	G๓๐๑๑๕๑	๑๐ September ๒๕๖๓	๙๕.๐	๙๕.๐
CR-B๐๙	Cirrus	CR๑๖๓B	G๓๐๑๑๐๑	๑๐ September ๒๕๖๓	๙๕.๐	๙๕.๐
Acoustic Certified Value : Thailand Institute of Scientific and Technological Research (TISTR)					93.99 ± 0.10 dB	

Calibrated by :



Approved by :





Noise R_440/23

Sound Level Meter Calibration Report


Acoustic Calibrator Data

Brand	ACO	Number	AC 03/56
Model	2127	Serial No.	130006
Calibration Range	94 dB, 1000 Hz	Last Calibration	29 March 2023
		Due Date	29 March 2024


Calibration Data

Sound Level Meter Data				Calibration Data		
SLM No.	Brand	Model	Serial No.	Date	Actual Reading [dB]	
					Before Adjustment	After Adjustment
ACO-R๐๗	ACO	๖๒๓๖	๐๐๑๕๒๐๘๐	๒๑ August ๒๐๒๓	๙๕.๐	๙๕.๐
ACO-R๒๒	ACO	๖๒๓๖	๐๐๑๘๒๐๑๐	๒๑ August ๒๐๒๓	๙๕.๑	๙๕.๐
ACO-R๓๙	ACO	๖๒๓๖	๐๐๑๙๒๐๕๑	๒๑ August ๒๐๒๓	๙๕.๐	๙๕.๐
ACO-R๔๔	ACO	๖๒๓๖	๐๐๑๙๒๐๕๖	๒๑ August ๒๐๒๓	๙๕.๐	๙๕.๐
Acoustic Certified Value : Thailand Institute of Scientific and Technological Research (TISTR)					93.94 ± 0.10 dB	

Calibrated by :



Approved by :





Noise R_536/23

Sound Level Meter Calibration Report


Acoustic Calibrator Data

Brand	ACO	Number	AC 03/56
Model	2127	Serial No.	130006
Calibration Range	94 dB, 1000 Hz	Last Calibration	29 March 2023
		Due Date	29 March 2024

Calibration Data

Sound Level Meter Data				Calibration Data		
SLM No.	Brand	Model	Serial No.	Date	Actual Reading [dB]	
					Before Adjustment	After Adjustment
ACO-R๑๙	ACO	๖๒๓๖	๐๐๑๘๒๐๐๑	๑๐ October ๒๐๒๓	๙๕.๐	๙๕.๐
ACO-R๒๑	ACO	๖๒๓๖	๐๐๑๘๒๐๐๔	๑๐ October ๒๐๒๓	๙๕.๑	๙๕.๐
ACO-R๓๐	ACO	๖๒๓๖	๐๐๑๙๒๐๔๒	๑๐ October ๒๐๒๓	๙๕.๐	๙๕.๐
ACO-R๓๓	ACO	๖๒๓๖	๐๐๑๙๒๐๔๕	๑๐ October ๒๐๒๓	๙๕.๐	๙๕.๐
Acoustic Certified Value : Thailand Institute of Scientific and Technological Research (TISTR)					93.94 ± 0.10 dB	

Calibrated by :



Approved by :





Noise R_496/23

Sound Level Meter Calibration Report


Acoustic Calibrator Data

Brand	ACO	Number	AC 03/56
Model	2127	Serial No.	130006
Calibration Range	94 dB, 1000 Hz	Last Calibration	29 March 2023
		Due Date	29 March 2024


Calibration Data

Sound Level Meter Data				Calibration Data		
SLM No.	Brand	Model	Serial No.	Date	Actual Reading [dB]	
					Before Adjustment	After Adjustment
ACO-R๒๑	ACO	๖๒๓๖	๐๐๑๘๒๐๐๔	๒๔ September ๒๐๒๓	๙๔.๐	๙๔.๐
ACO-R๒๔	ACO	๖๒๓๖	๐๐๑๘๒๐๐๑	๒๔ September ๒๐๒๓	๙๔.๑	๙๔.๐
ACO-R๓๓	ACO	๖๒๓๖	๐๐๑๘๒๐๐๕	๒๔ September ๒๐๒๓	๙๔.๐	๙๔.๐
ACO-R๔๔	ACO	๖๒๓๖	๐๐๑๘๒๐๐๖	๒๔ September ๒๐๒๓	๙๔.๐	๙๔.๐
Acoustic Certified Value : Thailand Institute of Scientific and Technological Research (TISTR)					93.94 ± 0.10 dB	

Calibrated by :



Approved by :





บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com., www.spscon.com

Noise R_679/23

Sound Level Meter Calibration Report


Acoustic Calibrator Data

Brand	ACO	Number	AC 03/56
Model	2127	Serial No.	130006
Calibration Range	94 dB, 1000 Hz	Last Calibration	29 March 2023
		Due Date	29 March 2024

Calibration Data

Sound Level Meter Data				Calibration Data		
SLM No.	Brand	Model	Serial No.	Date	Actual Reading [dB]	
					Before Adjustment	After Adjustment
ACO-R๑๘	ACO	๖๒๓๖	๐๐๑๗๒๐๖๕	๒๐ November ๒๐๒๓	๙๕.๐	๙๕.๐
ACO-R๒๑	ACO	๖๒๓๖	๐๐๑๘๒๐๐๕	๒๐ November ๒๐๒๓	๙๕.๑	๙๕.๐
ACO-R๓๐	ACO	๖๒๓๖	๐๐๑๙๒๐๔๒	๒๐ November ๒๐๒๓	๙๕.๐	๙๕.๐
ACO-R๓๓	ACO	๖๒๓๖	๐๐๑๙๒๐๔๕	๒๐ November ๒๐๒๓	๙๕.๐	๙๕.๐
Acoustic Certified Value : Thailand Institute of Scientific and Technological Research (TISTR)					93.94 ± 0.10 dB	

Calibrated by :



Approved by :

