

สำเนาหนังสือรับรองห้องปฏิบัติการวิเคราะห์เอกชน



ที่ อก ๐๓๑๐/(๑) ๑ ๒ ๓ ๗ ๐

กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ เขตราชเทวี
กรุงเทพมหานคร ๑๐๔๐๐

๐๙ กันยายน ๒๕๖๒

เรื่อง ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

เรียน กรรมการผู้จัดการ บริษัท ศูนย์วิเคราะห์น้ำ จำกัด

อ้างถึง ๑. คำขอขึ้นทะเบียน/ต่ออายุหนังสืออนุญาตขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

ลงวันที่ ๑๐ พฤษภาคม ๒๕๖๒

๒. หนังสือบริษัท ศูนย์วิเคราะห์น้ำ จำกัด ลงวันที่ ๕ สิงหาคม ๒๕๖๒

สิ่งที่ส่งมาด้วย เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท ศูนย์วิเคราะห์น้ำ จำกัด จำนวน ๖ แผ่น

ตามหนังสือที่อ้างถึง บริษัท ศูนย์วิเคราะห์น้ำ จำกัด ขอต่ออายุหนังสือรับขึ้นทะเบียน
ห้องปฏิบัติการวิเคราะห์เอกชน เลขทะเบียน ว-๑๙๐ สถานที่ตั้งเลขที่ ๑/๙๔ หมู่ที่ ๕ ตำบลคานหาม อำเภออุทัย
จังหวัดพระนครศรีอยุธยา ต่อกรมโรงงานอุตสาหกรรม นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว ให้บริษัท ศูนย์วิเคราะห์น้ำ จำกัด ต่ออายุหนังสือ
รับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน โดยมีองค์ประกอบดังนี้

ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์

๑) นางนীরมล ผดุงสงฆ์

ทะเบียนเลขที่ ว-๑๙๐-ก-๔๑๒๘

๒) นางสาวเปรมฤดี ชิวเศรษฐ์

ทะเบียนเลขที่ ว-๑๙๐-ก-๔๖๒๘

๓) นางสาวนิตยา ชันธบุตร

ทะเบียนเลขที่ ว-๑๙๐-ก-๕๓๐๗

๔) นางสาวจุฑารัตน์ ภูผ่าน

ทะเบียนเลขที่ ว-๑๙๐-ก-๗๗๖๐

ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์

๑) นางสาวอนุสรฯ แพ่งดวงแก้ว

ทะเบียนเลขที่ ว-๑๙๐-จ-๔๖๒๙

๒) นายรังศศิกร โกสุมภ์

ทะเบียนเลขที่ ว-๑๙๐-จ-๔๖๓๐

๓) นางสาวธัญญ์สิตา ชัยนันท์นทีกุล

ทะเบียนเลขที่ ว-๑๙๐-จ-๕๓๐๘

๔) นางสาวสุวลี บังแสงอ่อน

ทะเบียนเลขที่ ว-๑๙๐-จ-๕๗๕๔

๕) นางสาววราพร วันวิเศษ

ทะเบียนเลขที่ ว-๑๙๐-จ-๖๗๖๒

๖) นางสุนันทา แจ่มมิน

ทะเบียนเลขที่ ว-๑๙๐-จ-๖๗๖๓

๗) นายพุดพิงศ์ วรสุมนต์

ทะเบียนเลขที่ ว-๑๙๐-จ-๖๗๖๔

๘) นางสาวอรรพรณ สี่ใต้

ทะเบียนเลขที่ ว-๑๙๐-จ-๖๗๖๖

๙) นางสาวอรรพรณ จันทรวิชิต

ทะเบียนเลขที่ ว-๑๙๐-จ-๗๐๑๗

๑๐) นายวิชาวุฒิ...

- ๑๐) นายวิชาวุฒิ อุไรวรรณ
- ๑๑) นางสาวคณิตศรา สร้อยจิตร
- ๑๒) นางสาวรณกร ผดุงเวียง
- ๑๓) นายมานพ สลามซอ
- ๑๔) นายจตุเมธ อินทรโณภาส
- ๑๕) นางสาวแคทรียา มีแก้ว
- ๑๖) นายภาณุเดช เพชรอุด
- ๑๗) นางสาวอัญชิสา แผลงศรี
- ๑๘) นางสาวเกล้านภา ทองมี
- ๑๙) นายรัตพล ใบไกร
- ๒๐) นางสาวสมมาต อยู่สา
- ๒๑) นายภูเบศร์ สารยศ


- ทะเบียนเลขที่ ว-๑๙๐-จ-๗๐๑๘
ทะเบียนเลขที่ ว-๑๙๐-จ-๗๐๑๙
ทะเบียนเลขที่ ว-๑๙๐-จ-๗๐๒๐
ทะเบียนเลขที่ ว-๑๙๐-จ-๗๕๘๕
ทะเบียนเลขที่ ว-๑๙๐-จ-๗๕๘๖
ทะเบียนเลขที่ ว-๑๙๐-จ-๗๗๖๒
ทะเบียนเลขที่ ว-๑๙๐-จ-๗๙๐๙
ทะเบียนเลขที่ ว-๑๙๐-จ-๗๙๑๐
ทะเบียนเลขที่ ว-๑๙๐-จ-๗๙๑๑
ทะเบียนเลขที่ ว-๑๙๐-จ-๘๒๓๔
ทะเบียนเลขที่ ว-๑๙๐-จ-๘๒๓๕
ทะเบียนเลขที่ ว-๑๙๐-จ-๘๒๓๖

ค. ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนให้วิเคราะห์ในน้ำเสีย จำนวน ๒๗ รายการ
น้ำใต้ดิน จำนวน ๑๙ รายการ สิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว จำนวน ๑๖ รายการ และดิน จำนวน ๑๗ รายการ
รวมทั้งสิ้นจำนวน ๗๙ รายการ ตามสิ่งที่ส่งมาด้วย

หนังสือฉบับนี้จะหมดอายุในวันที่ ๑๖ มิถุนายน ๒๕๖๕ หากประสงค์จะต่ออายุหนังสือ
รับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ให้ยื่นคำขอต่ออายุพร้อมเอกสารประกอบคำขอต่อ
กรมโรงงานอุตสาหกรรมภายใน ๓๐ วัน ก่อนวันสิ้นอายุของหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
ซึ่งคำขอต่ออายุดังกล่าวขอรับได้ที่กรมโรงงานอุตสาหกรรม

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ



(นางจินดา เตชะศรีรินทร์)

ผู้อำนวยการกองวิจัยและเตือนภัยมลพิษโรงงาน
ปฏิบัติราชการแทนอธิบดีกรมโรงงานอุตสาหกรรม

กองวิจัยและเตือนภัยมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๒๐๒ ๔๑๔๖-๗ ๐ ๒๒๐๒ ๔๐๐๒

โทรสาร ๐ ๒๓๕๔ ๓๒๐๘ ๐ ๒๓๕๔ ๓๔๑๕

เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท ศูนย์วิเคราะห์น้ำ จำกัด

เลขทะเบียน ว-๑๙๐

ที่ ออก ๐๓๑๐/(๑) ๑ ๒ ๓ ๗ ๐

ลงวันที่ ๐๙ กันยายน ๒๕๖๒

ขอขยายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๗๙ รายการ

น้ำเสีย จำนวน 27 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Arsenic	Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[3]
2	Barium	Digestion, Direct Nitrous Oxide-Acetylene Flame Method ^[3]
3	Biochemical Oxygen Demand	1) 5-Day BOD Test, Azide Modification Method ^[3] 2) 5-Day BOD Test, Membrane Electrode Method ^[3]
4	Cadmium	1) Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[3]
5	Chemical Oxygen Demand	Closed Reflux, Titrimetric Method ^[3]
6	Chromium	Digestion, Direct Air-Acetylene Flame Method ^[3]
7	Color	ADMI Weighted-Ordinate Spectrophotometric Method ^[3]
8	Copper	Digestion, Direct Air-Acetylene Flame Method ^[3]
9	Cyanide	Distillation, Colorimetric Method ^[3]
10	Formaldehyde	Distillation, Colorimetric Method ^[2]
11	Free Chlorine	Iodometric Method ^[3]
12	Hexavalent Chromium	Filtration, Colorimetric Method ^[3]
13	Lead	1) Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[3]
14	Manganese	Digestion, Direct Air-Acetylene Flame Method ^[3]
15	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[3]
16	Nickel	Digestion, Direct Air-Acetylene Flame Method ^[3]
17	Oil & Grease	Soxhlet Extraction Method ^[3]
18	pH	Electrometric Method ^[3]

(นางริภาญจน์ ฉัตรสกุลวิไล)

ผู้อำนวยการกลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ

และทะเบียนห้องปฏิบัติการ

ภาคผนวก จ-3

19 Phenols...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
19	Phenols	Distillation, Direct Photometric Method ^[3]
20	Selenium	Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[3]
21	Sulfide	ZnS Precipitation, Iodometric Method ^[3]
22	Temperature	Laboratory and Field Methods ^[3]
23	Total Dissolved Solids	Dried at 180 °C ^[3]
24	Total Kjeldahl Nitrogen	Macro Kjeldahl, Titrimetric Method ^[3]
25	Total Suspended Solids	Dried at 103-105 °C ^[3]
26	Trivalent Chromium	Digestion, Direct Air-Acetylene Flame Method; Filtration, Colorimetric Method; Calculation ^[3]
27	Zinc	Digestion, Direct Air-Acetylene Flame Method ^[3]

น้ำใต้ดิน จำนวน 19 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Antimony	Digestion, Direct Air-Acetylene Flame Method ^[3]
2	Arsenic	Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[3]
3	Barium	Digestion, Direct Nitrous Oxide-Acetylene Flame Method ^[3]
4	Beryllium	Digestion, Direct Nitrous Oxide-Acetylene Flame Method ^[3]
5	Cadmium	1) Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[3]
6	Chromium	Digestion, Direct Air-Acetylene Flame Method ^[3]
7	Chromium (III)	Digestion, Direct Air-Acetylene Flame Method; Filtration, Colorimetric Method; Calculation ^[3]
8	Chromium (VI)	Filtration, Colorimetric Method ^[3]
9	Cyanide	Distillation, Colorimetric Method ^[3]
10	Lead	1) Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[3]
11	Manganese	Digestion, Direct Air-Acetylene Flame Method ^[3]

(นางริกาญจน์ ฉัตรสกุลวิไล)
ผู้อำนวยการศูนย์มาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
ภาคผนวก จ-4
และทะเบียนห้องปฏิบัติการ

12 Mercury...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
12	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[3]
13	Nickel	Digestion, Direct Air-Acetylene Flame Method ^[3]
14	pH	Electrometric Method ^[3]
15	Phenols	Distillation, Direct Photometric Method ^[3]
16	Selenium	Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[3]
17	Silver	Digestion, Direct Air-Acetylene Flame Method ^[3]
18	Vanadium	Digestion, Direct Nitrous Oxide-Acetylene Flame Method ^[3]
19	Zinc	Digestion, Direct Air-Acetylene Flame Method ^[3]

สิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว จำนวน 16 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Antimony	1) Waste Extraction, Digestion, Direct Air-Acetylene Flame Method ^[1,6]
2	Arsenic	2) Digestion, Direct Air-Acetylene Flame Method ^[4,6] 1) Waste Extraction, Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[1,7] 2) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[4,7]
3	Barium	1) Waste Extraction, Digestion, Direct Nitrous Oxide-Acetylene Flame Method ^[1,6] 2) Digestion, Direct Nitrous Oxide-Acetylene Flame Method ^[4,6]
4	Beryllium	1) Waste Extraction, Digestion, Direct Nitrous Oxide-Acetylene Flame Method ^[1,6] 2) Digestion, Direct Nitrous Oxide-Acetylene Flame Method ^[4,6]
5	Cadmium	1) Waste Extraction, Digestion, Direct Air-Acetylene Flame Method ^[1,6] 2) Digestion, Direct Air-Acetylene Flame Method ^[4,6]

Signature

(นางริกาญจน์ ฉัตรสกุลวิไล)

ผู้อำนวยการกลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
ภาคผนวก จ-5
และทะเบียนห้องปฏิบัติการ

6 Chromium...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
6	Chromium	1) Waste Extraction, Digestion, Direct Air-Acetylene Flame Method ^[1,6] 2) Digestion, Direct Air-Acetylene Flame Method ^[4,6]
7	Chromium (VI)	1) Waste Extraction, Colorimetric Method ^[1,8] 2) Digestion, Colorimetric Method ^[5,8]
8	Copper	1) Waste Extraction, Digestion, Direct Air-Acetylene Flame Method ^[1,6] 2) Digestion, Direct Air-Acetylene Flame Method ^[4,6]
9	Lead	1) Waste Extraction, Digestion, Direct Air-Acetylene Flame Method ^[1,6] 2) Digestion, Direct Air-Acetylene Flame Method ^[4,6]
10	Mercury	1) Waste Extraction, Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[1,9] 2) Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[4,10]
11	Nickel	1) Waste Extraction, Digestion, Direct Air-Acetylene Flame Method ^[1,6] 2) Digestion, Direct Air-Acetylene Flame Method ^[4,6]
12	pH	Electrometric Method ^[13]
13	Selenium	1) Waste Extraction, Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[1,11] 2) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[4,11]
14	Silver	1) Waste Extraction, Digestion, Direct Air-Acetylene Flame Method ^[1,6] 2) Digestion, Direct Air-Acetylene Flame Method ^[4,6]
15	Vanadium	1) Waste Extraction, Digestion, Direct Air-Acetylene Flame Method ^[1,6] 2) Digestion, Direct Air-Acetylene Flame Method ^[4,6]
16	Zinc	1) Waste Extraction, Digestion, Direct Air-Acetylene Flame Method ^[1,6] 2) Digestion, Direct Air-Acetylene Flame Method ^[4,6]

วิธีตรวจ

(นางวิภาดา ชื่นชูศรี)

ผู้อำนวยการกองมาตรฐานวิธีตรวจวิเคราะห์ของเสีย
ภาคผนวก จ-6
และทะเบียนห้องปฏิบัติการ


ดิน จำนวน...

ดิน จำนวน 17 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Antimony	Digestion, Direct Air-Acetylene Flame Method ^[4,6]
2	Arsenic	Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[4,7]
3	Barium	Digestion, Direct Air-Acetylene Flame Method ^[4,6]
4	Beryllium	Digestion, Direct Air-Acetylene Flame Method ^[4,6]
5	Cadmium	Digestion, Direct Air-Acetylene Flame Method ^[4,6]
6	Chromium	Digestion, Direct Air-Acetylene Flame Method ^[4,6]
7	Chromium (III)	Digestion, Direct Air-Acetylene Flame, Colorimetric Method; Calculation ^[4,5,6,8]
8	Chromium (VI)	Digestion, Colorimetric Method ^[5,8]
9	Cyanide	Cyanide Extraction Method ^[12]
10	Lead	Digestion, Direct Air-Acetylene Flame Method ^[4,6]
11	Manganese	Digestion, Direct Air-Acetylene Flame Method ^[4,6]
12	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[4,10]
13	Nickel	Digestion, Direct Air-Acetylene Flame Method ^[4,6]
14	Selenium	Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[4,11]
15	Silver	Digestion, Direct Air-Acetylene Flame Method ^[4,6]
16	Vanadium	Digestion, Direct Air-Acetylene Flame Method ^[4,6]
17	Zinc	Digestion, Direct Air-Acetylene Flame Method ^[4,6]

เอกสารอ้างอิง

- กระทรวงอุตสาหกรรม. ประกาศกระทรวงอุตสาหกรรม, พ.ศ. 2548. เรื่องการกำจัดสิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว. ราชกิจจานุเบกษา. 25 มกราคม 2549. เล่มที่ 123 ตอนพิเศษ 11ง.
- สมาคมวิศวกรรมสิ่งแวดล้อมแห่งประเทศไทย. คู่มือวิเคราะห์น้ำเสีย. พิมพ์ครั้งที่ 4 กรุงเทพฯ : เรือนแก้วการพิมพ์, 2547.
- APHA, AWWA, WEF. Standard Methods for the Examination of Water and Wastewater. 23rd ed. Washington, DC: APHA, 2017.
- United States Environmental Protection Agency. Acid Digestion of Sediments, Sludges, and Soils. SW-846 Method 3050B, 1996.




(นางริภาณูจน์ ฉัตรสกุลวิไล)

ผู้อำนวยการกลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
ภาคแผนกพิษวิทยาห้องปฏิบัติการ

5. United States...

5. United States Environmental Protection Agency. Alkaline Digestion for Hexavalent Chromium. SW-846 Method 3060A, 1996.
6. United States Environmental Protection Agency. Flame Atomic Absorption Spectrophotometry. SW-846 Method 7000B, 2007.
7. United States Environment Protection Agency. Antimony and Arsenic (Atomic Absorption, Borohydride Reduction). SW-846 Method 7062, 1994.
8. United States Environmental Protection Agency. Chromium, Hexavalent (Colorimetric). SW-846 Method 7196A, 1992.
9. United States Environmental Protection Agency. Mercury in Liquid Waste (Manual Cold-Vapor Technique). SW-846 Method 7470A, 1994.
10. United States Environmental Protection Agency. Mercury in Solid or Semisolid Waste (Manual Cold-Vapor Technique). SW-846 Method 7471B, 2007.
11. United States Environmental Protection Agency. Selenium (Atomic Absorption, Borohydride Reduction). SW-846 Method 7742, 1994.
12. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Cyanide Extraction Procedure for Solids and Oils. SW-846 Method 9013A, 2014.
13. United States Environmental Protection Agency. Soil and Waste pH. SW-846 Method 9045D, 2004.



(นางริกาญจน์ ฉัตรสกุลวิไล)

ผู้อำนวยการกลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ

และทะเบียนห้องปฏิบัติการ



ที่ อก ๐๓๑๐(๑)/ ๙๐๙๖

กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ เขตราชเทวี
กรุงเทพมหานคร ๑๐๔๐๐

๐๕ สิงหาคม ๒๕๖๓

เรื่อง เปลี่ยนแปลงบุคลากรและสารมลพิษที่วิเคราะห์

เรียน กรรมการผู้จัดการ บริษัท ศูนย์วิเคราะห์น้ำ จำกัด

อ้างถึง ๑. คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และชนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน
ลงวันที่ ๒๖ มีนาคม ๒๕๖๓

๒. หนังสือบริษัท ศูนย์วิเคราะห์น้ำ จำกัด ลงวันที่ ๓๐ กรกฎาคม ๒๕๖๓

สิ่งที่ส่งมาด้วย เอกสารแนบท้ายหนังสือเปลี่ยนแปลงบุคลากรและสารมลพิษที่วิเคราะห์
บริษัท ศูนย์วิเคราะห์น้ำ จำกัด จำนวน ๕ แผ่น

ตามหนังสือที่อ้างถึง บริษัท ศูนย์วิเคราะห์น้ำ จำกัด ห้องปฏิบัติการวิเคราะห์เอกชน
เลขทะเบียน ว-๑๙๐ สถานที่ตั้งเลขที่ ๑/๙๔ หมู่ที่ ๕ ตำบลคานหาม อำเภออุทัย จังหวัดพระนครศรีอยุธยา
ขอเปลี่ยนแปลงบุคลากรและสารมลพิษที่วิเคราะห์ ความละเอียดแจ้งแล้ว นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว มีความเห็นดังนี้

๑. ให้เพิ่มเจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๒ ราย

๑) นางสาวกันชญา อาจโยธา

ทะเบียนเลขที่ ว-๑๙๐-จ-๘๗๘๘

๒) นายสุทิวส ใจธีรภาพกุล

ทะเบียนเลขที่ ว-๑๙๐-จ-๘๗๘๙

๒. ให้เพิ่มขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนให้วิเคราะห์ในน้ำเสีย จำนวน ๑๗ รายการ
น้ำใต้ดิน จำนวน ๑๒ รายการ สิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว จำนวน ๙ รายการ และดิน จำนวน ๑๒ รายการ
รวมจำนวนทั้งสิ้น ๕๐ รายการ ตามสิ่งที่ส่งมาด้วย

อนึ่ง หนังสือฉบับนี้จะหมดอายุพร้อมหนังสือต่ออายุรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
ที่ อก ๐๓๑๐(๑) ๑๒๓๗๐ ลงวันที่ ๙ กันยายน ๒๕๖๒ คือในวันที่ ๑๖ มิถุนายน ๒๕๖๕

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ

(นางจินดา เคชะครินทร์)

ผู้อำนวยการกองวิจัยและเตือนภัยมลพิษโรงงาน
ปฏิบัติราชการแทนอธิบดีกรมโรงงานอุตสาหกรรม

กองวิจัยและเตือนภัยมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๒๐๒ ๔๑๔๖ ๐ ๒๒๐๒ ๔๐๐๒

โทรสาร ๐ ๒๓๕๔ ๓๒๐๘ ๐ ๒๓๕๔ ๓๔๑๕

เอกสารแนบท้ายหนังสือเปลี่ยนแปลงบุคลากรและสารมลพิษที่วิเคราะห์

บริษัท ศูนย์วิเคราะห์น้ำ จำกัด เลขทะเบียน ว-๑๙๐

ที่ อก ๐๓๑๐(๑)/ ๕๐๕๖

ลงวันที่ ๐๕ สิงหาคม ๒๕๖๓

ขอข่ายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๕๐ รายการ

น้ำเสีย จำนวน 17 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Aldrin	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
2	α -BHC	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
3	β -BHC	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
4	δ -BHC	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
5	γ -BHC	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
6	4,4'-DDD	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
7	4,4'-DDE	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
8	4,4'-DDT	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
9	Dieldrin	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
10	Endosulfan I	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
11	Endosulfan II	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
12	Endosulfan Sulfate	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
13	Endrin	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
14	Endrin Aldehyde	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]

วิมล

15 Heptachlor...

(นางริกาญจน์ อัครสกุลวิไล)

ผู้อำนวยการกลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
ภาคผนวก จ-10
และทะเบียนห้องปฏิบัติการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
15	Heptachlor	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
16	Heptachlor Epoxide	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
17	Methoxychlor	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]

น้ำใต้ดิน จำนวน 12 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Aldrin	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
2	DDD	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
3	DDE	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
4	DDT	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
5	Dieldrin	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
6	Endrin	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
7	Heptachlor	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
8	Heptachlor epoxide	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
9	α -HCH	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
10	β -HCH	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
11	γ -HCH	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]
12	Methoxychlor	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[2]

วิมล

(นางริกาญจน์ ฉัตรสกุลวิไล)

ผู้อำนวยการกลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
ภาคพื้นทวีป จ-11
และทะเบียนห้องปฏิบัติการ

สิ่งปลูก...

สิ่งปลูกหรือวัสดุที่ไม่ใช้แล้ว จำนวน 9 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Aldrin	1) Waste Extraction, Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[1,3,4] 2) Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^[3,5]
2	γ -BHC	1) Waste Extraction, Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[1,3,4] 2) Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^[3,5]
3	DDD	1) Waste Extraction, Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[1,3,4] 2) Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^[3,5]
4	DDE	1) Waste Extraction, Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[1,3,4] 2) Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^[3,5]
5	DDT	1) Waste Extraction, Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[1,3,4] 2) Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^[3,5]
6	Dieldrin	1) Waste Extraction, Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[1,3,4] 2) Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^[3,5]
7	Endrin	1) Waste Extraction, Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[1,3,4] 2) Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^[3,5]
8	Heptachlor	1) Waste Extraction, Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[1,3,4] 2) Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^[3,5]

Signature

9 Methoxychlor...

(นางริยาญจน์ วัชรสกุลวิไล)

ผู้อำนวยการกลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
ภาคผนวก จ-12
และระเบียบห้องปฏิบัติการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
9	Methoxychlor	1) Waste Extraction, Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[1,3,4] 2) Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3,5]

ดิน จำนวน 12 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Aldrin	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3,5]
2	DDD	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3,5]
3	DDE	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3,5]
4	DDT	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3,5]
5	Dieldrin	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3,5]
6	Endrin	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3,5]
7	Heptachlor	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3,5]
8	Heptachlor epoxide	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3,5]
9	α -HCH	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3,5]
10	β -HCH	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3,5]
11	γ -HCH	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3,5]
12	Methoxychlor	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3,5]

วิมล

เอกสารอ้างอิง...

(นางริกาญจน์ ฉัตรสกุลวิไล)

ผู้อำนวยการกลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ

และทะเบียนห้องปฏิบัติการ

เอกสารอ้างอิง

1. กระทรวงอุตสาหกรรม. ประกาศกระทรวงอุตสาหกรรม, พ.ศ. 2548. เรื่อง การกำจัดสิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว. ราชกิจจานุเบกษา. 25 มกราคม 2549. เล่มที่ 123 ตอนพิเศษ 11ง.
2. APHA, AWWA, WEF. Standard Methods for the Examination of Water and Wastewater. 23rd ed. Washington, DC: APHA, 2017.
3. United States Environmental Protection Agency. Semivolatile Organic Compounds by Gas Chromatography Mass Spectrometry (GC/MS). SW-846 Method 8270D, 2014.
4. United States Environmental Protection Agency. Separatory Funnel Liquid-Liquid Extraction. SW-846 Method 3510C, 1996.
5. United States Environmental Protection Agency. Soxhlet Extraction. SW-846 Method 3540C, 1996.



(นางริภาณูญ์ จิตรสกุลไธ)

ผู้อำนวยการกลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
และทะเบียนห้องปฏิบัติการ

ที่ อก ๐๓๑๐(๑)/ ๑๒๐๗๒



กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท
เขตราชเทวี กรุงเทพฯ ๑๐๔๐๐

๐๒ ธันวาคม ๒๕๖๕

เรื่อง เปลี่ยนแปลงบุคลากรของห้องปฏิบัติการวิเคราะห์

เรียน กรรมการผู้จัดการ บริษัท ศูนย์วิเคราะห์น้ำ จำกัด

อ้างถึง ๑. คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากรและชนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์
เอกชน ลงวันที่ ๑๖ พฤศจิกายน ๒๕๖๔

๒. หนังสือบริษัท ศูนย์วิเคราะห์น้ำ จำกัด ลงวันที่ ๑๖ พฤศจิกายน ๒๕๖๔

ตามหนังสือที่อ้างถึง ๑ และ ๒ บริษัท ศูนย์วิเคราะห์น้ำ จำกัด ห้องปฏิบัติการวิเคราะห์เอกชน
เลขทะเบียน ว-๑๙๐ สถานที่ตั้งเลขที่ ๑/๙๔ หมู่ที่ ๕ ตำบลคานหาม อำเภอดุสิต จังหวัดพระนครศรีอยุธยา
ขอเปลี่ยนแปลงบุคลากรของห้องปฏิบัติการวิเคราะห์ ความละเอียดแจ้งแล้ว นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว มีความเห็นดังนี้

๑. ให้ยกเลิกเจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๓ ราย

๑) นางสาวธัญยลิตา ชัยนันท์นทีกุล ทะเบียนเลขที่ ว-๑๙๐-จ-๕๓๐๘

๒) นางสาวอรพรรณ จันทรวิจิต ทะเบียนเลขที่ ว-๑๙๐-จ-๗๐๑๗

๓) นางสาวเกล้านภา ทองมี ทะเบียนเลขที่ ว-๑๙๐-จ-๗๙๑๑

๒. ให้เพิ่มเจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๖ ราย

๑) นายธนกฤต สุจริต ทะเบียนเลขที่ ว-๑๙๐-จ-๙๖๙๙

๒) นายกฤษณพล ปัญญาวงค์ ทะเบียนเลขที่ ว-๑๙๐-จ-๙๗๐๐

๓) นางสาวกนกพร หลวงประมุข ทะเบียนเลขที่ ว-๑๙๐-จ-๙๗๐๑

๔) นางสาววณิชยา แก้วรุ่งฟ้า ทะเบียนเลขที่ ว-๑๙๐-จ-๙๗๐๒

๕) นางสาวสุธาสินี หอมสวาท ทะเบียนเลขที่ ว-๑๙๐-จ-๙๗๐๓

๖) นางสาวเครือวัลลี สมภิงษ์ ทะเบียนเลขที่ ว-๑๙๐-จ-๙๗๐๔

อนึ่ง หนังสือฉบับนี้จะหมดอายุพร้อมหนังสือต่ออายุรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์
เอกชน ที่ อก ๐๓๑๐(๑) ๑๒๓๗๐ ลงวันที่ ๙ กันยายน ๒๕๖๖ คือในวันที่ ๑๖ มิถุนายน ๒๕๖๕

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ

(นายศิระ จันทรเจ็ด)

นักวิทยาศาสตร์ชำนาญการพิเศษ

รักษาการนักวิทยาศาสตร์เชี่ยวชาญ รักษาราชการแทน

ผู้อำนวยการกองวิจัยและเตือนภัยมลพิษโรงงาน

กองวิจัยและเตือนภัยมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๐๓-๕

โทรสาร ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๙๙

ไปรษณีย์อิเล็กทรอนิกส์ saraban@diw.mail.go.th

ภาคผนวก จ-15



Ref No. : 0303/16046

CERTIFICATE OF TESTING LABORATORY ACCREDITATION

This is to certify that

*Laboratory of Water Analysis Center Co., Ltd.
1/94 Moo 5, Tambon Kanharm, Amphoe U-Thai,
Changwat Phra Nakhon Si Ayutthaya 13210*

has successfully undergone assessment according to ISO/IEC 17025 : 2017
and under the Bureau of Laboratory Accreditation, Department of Science Service
for the requirements, regulations and criteria for the competence of testing laboratories

Accreditation Number TESTING - 0029

The scope of accreditation is as annexed hereto

Issue date : 7th November 2019

Expired date : 6th November 2022

Signature : 

(Mrs. Pochaman Tagheen)

Director of Bureau of Laboratory Accreditation

Bureau of Laboratory Accreditation, Department of Science Service,
Ministry of Higher Education, Science, Research and Innovation

Scope of Testing Laboratory Accreditation

Laboratory Name : Laboratory of Water Analysis Center Co., Ltd.

Address : 1/94 Moo 5, Tambon Kanharm, Amphoe U-Thai,

Changwat Phra Nakhon Si Ayutthaya 13210

Accreditation Number : Testing - 0029

Laboratory Status : ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Item Number	Test Material / Product	Test Item / Range of Testing	Test Method / Technique Used
1	Bottled drinking water	- Chloride 6 mg/dm³ to 1 000 mg/dm³ - Total Hardness (Calculated as CaCO₃) 5 mg/dm³ to 2 000 mg/dm³ - Total Solids at 103 °C to 105 °C 25 mg/dm³ to 4 000 mg/dm³	In - house method : TM 008 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 4500 - Cl⁻ B In - house method : TM 009 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 2340 C In - house method : TM 015 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 2540 B

Initial Issue Date 23rd September 2008

Issue Number 12

Bureau of Laboratory Accreditation, Department of Science Service, Ministry of Higher Education, Science, Research and Innovation

Scope of Testing Laboratory Accreditation

Laboratory Name : Laboratory of Water Analysis Center Co., Ltd.

Address : 1/94 Moo 5, Tambon Kanharm, Amphoe U-Thai,
Changwat Phra Nakhon Si Ayutthaya 13210

Accreditation Number : Testing - 0029

Laboratory Status : ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Item Number	Test Material / Product	Test Item / Range of Testing	Test Method / Technique Used
1 (cont.)	Bottled drinking water	- Manganese 0.05 mg/dm ³ to 5 mg/dm ³ - Iron 0.10 mg/dm ³ to 5 mg/dm ³	In - house method : TM 040 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 rd ed., 2017, part 3111 B
2	Water	- pH 6.0 to 10.0 - Total Suspended Solids at 103 °C to 105 °C 10 mg/dm ³ to 1 000 mg/dm ³	In - house method : TM 001 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 rd ed., 2017, part 4500 - H ⁺ B In - house method : TM 016 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 rd ed., 2017, part 2540 D

Initial Issue Date 23rd September 2008

Issue Number 12

Bureau of Laboratory Accreditation, Department of Science Service, Ministry of Higher Education, Science, Research and Innovation

Scope of Testing Laboratory Accreditation

Laboratory Name : Laboratory of Water Analysis Center Co., Ltd.

Address : 1/94 Moo 5, Tambon Kanharm, Amphoe U-Thai,

Changwat Phra Nakhon Si Ayutthaya 13210

Accreditation Number : Testing - 0029

Laboratory Status : ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Item Number	Test Material / Product	Test Item / Range of Testing	Test Method / Technique Used
2 (cont.)	Water	- Total Dissolved Solids at 180 °C 25 mg/dm ³ to 4 000 mg/dm ³ - Cadmium 0.05 mg/dm ³ to 0.9 mg/dm ³ - Copper 0.05 mg/dm ³ to 5 mg/dm ³ - Zinc 0.05 mg/dm ³ to 5 mg/dm ³ - Chromium 0.05 mg/dm ³ to 5 mg/dm ³	In - house method : TM 017 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 rd ed., 2017, part 2540 C In - house method : TM 040 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 rd ed., 2017, part 3111 B

Initial Issue Date 23rd September 2008

Issue Number 12

Bureau of Laboratory Accreditation, Department of Science Service, Ministry of Higher Education, Science, Research and Innovation

Scope of Testing Laboratory Accreditation

Laboratory Name : Laboratory of Water Analysis Center Co., Ltd.

Address : 1/94 Moo 5, Tambon Kanharm, Amphoe U-Thai,
Changwat Phra Nakhon Si Ayutthaya 13210

Accreditation Number : Testing - 0029

Laboratory Status : ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Item Number	Test Material / Product	Test Item / Range of Testing	Test Method / Technique Used
2 (cont.)	Water	- Nickel 0.10 mg/dm³ to 4 mg/dm³ - Manganese 0.05 mg/dm³ to 5 mg/dm³ - Lead 0.10 mg/dm³ to 2 mg/dm³ - Iron 0.10 mg/dm³ to 5 mg/dm³ - Water soluble silica (Calculated as SiO₂) 1.1 mg/dm³ to 26 mg/dm³	In - house method : TM 040 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 3111 B In - house method : TM 030 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 4500-SiO₂ C

Initial Issue Date 23rd September 2008

Issue Number 12

Bureau of Laboratory Accreditation, Department of Science Service, Ministry of Higher Education, Science, Research and Innovation

Scope of Testing Laboratory Accreditation

Laboratory Name : Laboratory of Water Analysis Center Co., Ltd.

Address : 1/94 Moo 5, Tambon Kanharm, Amphoe U-Thai,
Changwat Phra Nakhon Si Ayutthaya 13210

Accreditation Number : Testing - 0029

Laboratory Status : ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Item Number	Test Material / Product	Test Item / Range of Testing	Test Method / Technique Used
2 (cont.)	Water	- Chloride 6 mg/dm³ to 1 000 mg/dm³ - Total Hardness (Calculated as CaCO₃) 5 mg/dm³ to 2 000 mg/dm³ - BOD 2 mg/dm³ to 500 mg/dm³	In - house method: TM 008 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 4500-Cl⁻ B In - house method : TM 009 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 2340 C In - house method : TM 041 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 5210 B

Initial Issue Date 23rd September 2008

Issue Number 12

Bureau of Laboratory Accreditation, Department of Science Service, Ministry of Higher Education, Science, Research and Innovation

Scope of Testing Laboratory Accreditation

Laboratory Name : Laboratory of Water Analysis Center Co., Ltd.

Address : 1/94 Moo 5, Tambon Kanharm, Amphoe U-Thai,

Changwat Phra Nakhon Si Ayutthaya 13210

Accreditation Number : Testing - 0029

Laboratory Status : ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Item Number	Test Material / Product	Test Item / Range of Testing	Test Method / Technique Used
2 (cont.)	Water	- COD 40 mg/dm³ to 200 mg/dm³ - Total Kjeldahl Nitrogen 5 mg/dm³ to 200 mg/dm³ - Oil and grease 2 mg/dm³ to 100 mg/dm³	In - house method : TM 014 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 5220 C In - house method : TM 023 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 4500-NH₃ C and 4500-N_{org} B In - house method : TM 020 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 5520 D

Initial Issue Date 23rd September 2008

Issue Number 12

Bureau of Laboratory Accreditation, Department of Science Service, Ministry of Higher Education, Science, Research and Innovation

Scope of Testing Laboratory Accreditation

Laboratory Name : Laboratory of Water Analysis Center Co., Ltd.

Address : 1/94 Moo 5, Tambon Kanharm, Amphoe U-Thai,

Changwat Phra Nakhon Si Ayutthaya 13210

Accreditation Number : Testing - 0029

Laboratory Status : ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Item Number	Test Material / Product	Test Item / Range of Testing	Test Method / Technique Used
2 (cont.)	Water	- Total Solids at 103 °C to 105 °C 25 mg/dm ³ to 4 000 mg/dm ³ - Selenium 5 µg/dm ³ to 50 µg/dm ³	In - house method : TM 015 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 rd ed., 2017, part 2540 B In - house method : TM 038 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 rd ed., 2017, part 3114 C
3	Wastewater	- pH 4.0 to 10.0	In - house method : TM 001 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 rd ed., 2017, part 4500 - H ⁺ B

Initial Issue Date 23rd September 2008

Issue Number 12

Bureau of Laboratory Accreditation, Department of Science Service, Ministry of Higher Education, Science, Research and Innovation

Scope of Testing Laboratory Accreditation

Laboratory Name : Laboratory of Water Analysis Center Co., Ltd.

Address : 1/94 Moo 5, Tambon Kanharm, Amphoe U-Thai,

Changwat Phra Nakhon Si Ayutthaya 13210

Accreditation Number : Testing - 0029

Laboratory Status : ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Item Number	Test Material / Product	Test Item / Range of Testing	Test Method / Technique Used
3 (cont.)	Wastewater	- Total Suspended Solids at 103 °C to 105 °C 10 mg/dm³ to 1 000 mg/dm³ - Total Dissolved Solids at 180 °C 50 mg/dm³ to 4 000 mg/dm³ - Cadmium 0.05 mg/dm³ to 0.9 mg/dm³ - Copper 0.05 mg/dm³ to 5 mg/dm³	In - house method : TM 016 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 2540 D In - house method : TM 017 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 2540 C In - house method : TM 040 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 3111 B

Initial Issue Date 23rd September 2008

Issue Number 12

Bureau of Laboratory Accreditation, Department of Science Service, Ministry of Higher Education, Science, Research and Innovation

Scope of Testing Laboratory Accreditation

Laboratory Name : Laboratory of Water Analysis Center Co., Ltd.

Address : 1/94 Moo 5, Tambon Kanharm, Amphoe U-Thai,
Changwat Phra Nakhon Si Ayutthaya 13210

Accreditation Number : Testing - 0029

Laboratory Status : ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Item Number	Test Material / Product	Test Item / Range of Testing	Test Method / Technique Used
3 (cont.)	Wastewater	- Zinc 0.05 mg/dm^3 to 5 mg/dm^3 - Chromium 0.05 mg/dm^3 to 5 mg/dm^3 - Nickel 0.10 mg/dm^3 to 4 mg/dm^3 - Manganese 0.05 mg/dm^3 to 5 mg/dm^3 - Lead 0.10 mg/dm^3 to 2 mg/dm^3 - Iron 0.10 mg/dm^3 to 5 mg/dm^3	In - house method : TM 040 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23 rd ed., 2017, part 3111 B

Initial Issue Date 23rd September 2008

Issue Number 12

Bureau of Laboratory Accreditation, Department of Science Service, Ministry of Higher Education, Science, Research and Innovation

Scope of Testing Laboratory Accreditation

Laboratory Name : Laboratory of Water Analysis Center Co., Ltd.

Address : 1/94 Moo 5, Tambon Kanharm, Amphoe U-Thai,
Changwat Phra Nakhon Si Ayutthaya 13210

Accreditation Number : Testing - 0029

Laboratory Status : ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Item Number	Test Material / Product	Test Item / Range of Testing	Test Method / Technique Used
3 (cont.)	Wastewater	- Total Hardness (Calculated as CaCO_3) 5 mg/dm^3 to $2\,000 \text{ mg/dm}^3$ - BOD 4 mg/dm^3 to $7\,000 \text{ mg/dm}^3$ - COD 40 mg/dm^3 to $3\,000 \text{ mg/dm}^3$	In - house method : TM 009 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 2340 C In - house method : TM 041 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 5210 B In - house method : TM 014 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 5220 C

Initial Issue Date 23rd September 2008

Issue Number 12

Bureau of Laboratory Accreditation, Department of Science Service, Ministry of Higher Education, Science, Research and Innovation

Scope of Testing Laboratory Accreditation

Laboratory Name : Laboratory of Water Analysis Center Co., Ltd.

Address : 1/94 Moo 5, Tambon Kanharm, Amphoe U-Thai,
Changwat Phra Nakhon Si Ayutthaya 13210

Accreditation Number : Testing - 0029

Laboratory Status : ☒ Permanent ☐ Site ☐ Temporary ☐ Mobile

Item Number	Test Material / Product	Test Item / Range of Testing	Test Method / Technique Used
3 (cont.)	Wastewater	- Total Kjeldahl Nitrogen 5 mg/dm³ to 200 mg/dm³ - Oil and Grease 2 mg/dm³ to 1 000 mg/dm³ - Total Solids at 103 °C to 105 °C 25 mg/dm³ to 4 000 mg/dm³	In - house method : TM 023 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 4500-NH₃ C and 4500-N_{org} B In - house method : TM 020 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 5520 D In - house method : TM 015 based on Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 23rd ed., 2017, part 2540 B

Initial Issue Date 23rd September 2008

Issue Number 12

Bureau of Laboratory Accreditation, Department of Science Service, Ministry of Higher Education, Science, Research and Innovation