

ภาคผนวก 1

หนังสือเห็นชอบ มาตรการป้องกัน แก้ไขผลกระทบสิ่งแวดล้อม
และมาตรการติดตามตรวจสอบผลกระทบสิ่งแวดล้อม



ที่ ทส 1009/ 12170

สำนักงานนโยบายและแผน

ทรัพยากรธรรมชาติและสิ่งแวดล้อม

60/1 ซอยพิบูลวัฒนา 7 ถนนพระรามที่ 6

กรุงเทพฯ 10400

- 2 ส.ป. 2547
พฤษภาคม 2547

เรื่อง รับทราบมติการพิจารณารายงานการวิเคราะห์ผลกระทบสิ่งแวดล้อมโครงการพงษ์พรรณ แมนชั่น

เรียน ผู้ว่าราชการกรุงเทพมหานคร

อ้างถึง หนังสือกรุงเทพมหานคร ที่ กท 0410/5828 ลงวันที่ 25 พฤศจิกายน 2547

สิ่งที่ส่งมาด้วย มาตราการป้องกันและแก้ไขผลกระทบสิ่งแวดล้อมที่โครงการพงษ์พรรณ แมนชั่น
ของ บริษัท พงษ์พรรณ แมนชั่น จำกัด ต้องยึดถือปฏิบัติอย่างเคร่งครัด

ตามหนังสือที่อ้างถึง กรุงเทพมหานครแจ้งมติคณะกรรมการผู้ชำนาญการพิจารณา
รายงานการวิเคราะห์ผลกระทบสิ่งแวดล้อมโครงการในเขตพื้นที่กรุงเทพมหานคร ในการประชุมครั้งที่
5/2547 เมื่อวันที่ 25 มิถุนายน 2547 เห็นชอบรายงานการวิเคราะห์ผลกระทบสิ่งแวดล้อมโครงการ
พงษ์พรรณ แมนชั่น ของ บริษัท พงษ์พรรณ แมนชั่น จำกัด ตั้งอยู่ที่ ซอยประดิพัทธ์ 23 ถนนประดิพัทธ์
แขวงสามเสนใน เขตพญาไท กรุงเทพมหานคร มีพื้นที่ 1 งาน 89.6 ตารางวา ประกอบด้วยอาคารสูง
9 ชั้น (22.96 เมตร) จำนวน 1 อาคาร จำนวนห้องพัก 118 ห้อง จัดทำรายงานโดยบริษัท เอเซีย แลป
แอนด์ คอนซัลแตนท์ จำกัด และเสนอสำนักงานนโยบายและแผนทรัพยากรธรรมชาติและสิ่งแวดล้อมเพื่อ
ดำเนินการ ความละเอียดแจ้งแล้ว นั้น

สำนักงานนโยบายและแผนทรัพยากรธรรมชาติและสิ่งแวดล้อม รับทราบมติคณะกรรมการ
ผู้ชำนาญการพิจารณารายงานการวิเคราะห์ผลกระทบสิ่งแวดล้อมโครงการในเขตพื้นที่กรุงเทพมหานคร
ดังกล่าว โดยให้โครงการพงษ์พรรณ แมนชั่น ของ บริษัท พงษ์พรรณ แมนชั่น จำกัด ปฏิบัติตามมาตรการ


2ป้องกัน...

- 2 -

ป้องกันและแก้ไขผลกระทบสิ่งแวดล้อม ดังรายละเอียดตามสิ่งที่ส่งมาด้วยอย่างเคร่งครัดต่อไป อนึ่ง ตามมาตรา 50 วรรคท้าย ของพระราชบัญญัติส่งเสริมและรักษาคุณภาพสิ่งแวดล้อมแห่งชาติ พ.ศ. 2535 ได้กำหนดให้เจ้าหน้าที่ซึ่งมีอำนาจตามกฎหมายในการพิจารณาสั่งอนุญาตหรือต่ออายุใบอนุญาต นำมาตรวจการตามที่เสนอไว้ในรายงานการวิเคราะห์ผลกระทบสิ่งแวดล้อมไปกำหนดเป็นเงื่อนไขในการสั่งอนุญาตหรือต่ออายุใบอนุญาต โดยให้ถือว่าเป็นเงื่อนไขที่กำหนดตามกฎหมายในเรื่องนั้นด้วย ทั้งนี้ สำนักงานได้สำเนาหนังสือแจ้งให้บริษัท พงษ์พรรณ แมนชัย จำกัด และบริษัท เอเซีย แล็บ แอนด์ คอนซัลแตนท์ จำกัด เพื่อดำเนินการด้วยแล้ว

จึงเรียนมาเพื่อโปรดพิจารณาดำเนินการต่อไป

ขอแสดงความนับถือ



(นางอรอนงค์ วงษ์สวัสดิ์)

รองเลขาธิการ ปฏิบัติราชการแทน

เลขาธิการสำนักงานนโยบายและแผนทรัพยากรธรรมชาติและสิ่งแวดล้อม

สำนักวิเคราะห์ผลกระทบสิ่งแวดล้อม

โทรศัพท์ 0-2298-6157, 0-2271-4232 – 8 ต่อ 245

โทรสาร 0-2279-2792

ภาคผนวกที่ไม่ต้องแนบ ได้แก่ ตารางมาตรการ ,

ภาคผนวก 2

หนังสือสำคัญนิติบุคคล

อ1 , ลำเนาหนังสือรับรองบริษัท

ภาคผนวก 3

ผลการตรวจวิเคราะห์คุณภาพน้ำทิ้ง



บริษัท วิศวกรรมเคมี จำกัด

THAI CHEMICAL & ENGINEERING CO., LTD.

(048/2 ซ.สุขุมวิท 66/1 ถ.สุขุมวิท (พวงมณี) ถนนสุขุมวิท กรุงเทพฯ 10110 โทร. 0-2744-9911 แฟกซ์ 0-2393-6165

(048/2 ซ. Sukhumvit 66/1 Sukhumvit Rd., Prakanong Tai, Prakanong, Bangkok 10260 TEL. 0-2744-9911 FAX 0-2393-6165

No. 1809/66

WASTE WATER ANALYSIS REPORT

Date : 02/08/66 Analysis Date : 25/07/66-31/07/66
 Customer : บริษัท พงษ์พรหม แมนชั่น จำกัด สำนักงานใหญ่ Sampling Date : 25/07/66
 Address : เลขที่ 24 ซอยประดิพัทธ์ 23 ถนนประดิพัทธ์ แขวงสามเสนใต้
 เขตพญาไท กรุงเทพฯ Sampling Time : 14.00
 Received Date : 25/07/66
 Tel : 084-703-9129

Reference Number	WP/NM 3310/66			Method of Analysis
Parameter	Unit	Effluent	Standard	
Appearance		เหลืองใสมีตะกอนเล็กน้อย		
pH		@ 25.2 °C = 7.8	5-9	Electrometric (SM 2017.4500-H ⁺ B.)
Biochemical Oxygen Demand	(mg/l)	22	≤ 30	5-Day BOD Test, Azide Modification (SM 2017.5210 B.)
Total Suspended Solids	(mg/l)	18	≤ 40	Dried at 103-105 °C (SM 2017.2540 D.)
Total Dissolved Solids	(mg/l)	356	≤ 500	Dried at 180 °C (SM 2017.2540 C.)
Oil & Grease	(mg/l)	< 5.00	≤ 20	Soxhlet Extraction (SM 2017.5520 D.)
Total Kjeldahl Nitrogen	(mg/l)	40.22	≤ 35	Macro-Kjeldahl, Titrimetric (SM 2017.4500-N(org) B.)
Sulfide	(mg/l)	Not detected	≤ 1.0	ZnS Precipitation, Iodometric (SM 2017.4500-S ²⁻ F.)
Settleable Solids	(ml)	< 0.5	≤ 0.5	Imhoff Cone, Volumetric (SM 2017.2540 F.)

SM : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017.

❖ The results relate only to the samples tested and apply to customer's self-drawn samples only

❖ This analysis report may not be reproduced other than in full, except with the prior written approval of the technical manager.

Approved by

 ชัยณัฐรัตน์ พงษ์พรหมแมนชั่น

 (ในฐานะ วิศวกร)



บริษัท วิศวกรรมเคมี จำกัด

THAI CHEMICAL & ENGINEERING CO., LTD.

เลขที่ ๖ ซอยสุขุมวิท ๖๖/๑ อ.สุขุมวิท เขตวัฒนา กรุงเทพมหานคร 10260 โทร. 0-2744-9911 แฟกซ์ 0-2393-0165

(948/2 So Sukhumvit 66/1, Sukhumvit Rd., Prakanong Tai, Prakanong, Bangkok 10260 TEL. 0-2744-9911 FAX 0-2393-0165)

No. 2047/66

WASTE WATER ANALYSIS REPORT

Date: 29/08/66 Analysis Date: 22/08/66-28/08/66
 Customer: บริษัท พงษ์พรณ แมนชั่น จำกัด สำนักงานใหญ่ Sampling Date: 22/08/66
 Address: เลขที่ 24 ซอยประดิพัทธ์ 23 ถนนประดิพัทธ์ แขวงสามเสนใต้ Sampling Time: 13.30
 เขตพญาไท กรุงเทพฯ Received Date: 22/08/66
 Tel: 084-703-9129

Reference Number	WP/NM 3718/66			
Parameter	Unit	Effluent	Standard	Method of Analysis
Appearance		เหลืองใสมีตะกอน		
pH		@ 25.0 °C = 7.5	5-9	Electrometric (SM 2017.4500-H ⁺ B.)
Biochemical Oxygen Demand	(mg/l)	57	≤ 30	5-Day BOD Test, Azide Modification (SM 2017.5210 B.)
Total Suspended Solids	(mg/l)	30	≤ 40	Dried at 103-105 °C (SM 2017.2540 D.)
Total Dissolved Solids	(mg/l)	306	≤ 500	Dried at 180 °C (SM 2017.2540 C.)
Oil & Grease	(mg/l)	< 5.00	≤ 20	Soxhlet Extraction (SM 2017.5520 D.)
Total Kjeldahl Nitrogen	(mg/l)	27.37	≤ 35	Macro-Kjeldahl, Titrimetric (SM 2017.4500-N(org) B.)
Sulfide	(mg/l)	Not detected	≤ 1.0	ZnS Precipitation, Iodometric (SM 2017.4500-S ²⁻ F.)
Settleable Solids	(ml/l)	< 0.5	≤ 0.5	Imhoff Cone, Volumetric (SM 2017.2540 F.)

SM = Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23rd Edition, 2017.

❖ The results relate only to the samples tested and apply to customer's self-drawn samples only.

❖ This analysis report may not be reproduced other than in full, except with the prior written approval of the technical manager.

Approved by

ศุภณัฐ วัฒนวิเศษกุล
 วิศวกรเคมี พงษ์พรณแมนชั่น จำกัด



WASTE WATER ANALYSIS REPORT

SM: Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23rd Edition, 2017.

- Approved by ผอ.กิตติพงษ์ นพรัตน์
* * * * *



บริษัท วิศวกรรมเคมี จำกัด

THAI CHEMICAL & ENGINEERING CO., LTD.

106/2 ซ.สุขุมวิท 66/1 อ.สุขุมวิท แขวงพระโขนงใต้ เขตพระโขนง กรุงเทพฯ 10260 โทร. 0-2544-9911 แฟกซ์ 0-2593-0187

106/2 Soi Sukhumvit 66/1 Sukhumvit Rd., Prakhong Tai, Prakhong, Bangkok 10260 TEL: 0-2544-9911 FAX: 0-2593-0187

No. 2539/66

WASTE WATER ANALYSIS REPORT

Date: 26/10/66

Analysis Date: 19/10/66-25/10/66

Customer: บริษัท พงษ์พรหม แมนชั่น จำกัด สำนักงานใหญ่

Sampling Date: 19/10/66

Address: เลขที่ 24 ซอปประดิพัทธ์ 23 ถนนประดิพัทธ์ แขวงสามเสนใต้

Sampling Time: 11.20

เขตพญาไท กรุงเทพฯ

Received Date: 19/10/66

Tel: 084-703-9129

Reference Number	WP/NM 4711/66			Method of Analysis
Parameter	Unit	Effluent	Standard	
Appearance		เหลืองขุ่น		
pH		@ 24.3 °C = 7.9	5-9	Electrometric (SM 2017.4500-H ⁺ B.)
Biochemical Oxygen Demand	(mg/l)	22	< 30	5-Day BOD Test, Azide Modification (SM 2017.5210 B.)
Total Suspended Solids	(mg/l)	65	< 40	Dried at 103-105 °C (SM 2017.2540 D.)
Total Dissolved Solids	(mg/l)	245	< 500	Dried at 180 °C (SM 2017.2540 C.)
Oil & Grease	(mg/l)	< 5.00	< 20	Soxhlet Extraction (SM 2017.5520 D.)
Total Kjeldahl Nitrogen	(mg/l)	45.74	< 35	Macro-Kjeldahl, Timetric (SM 2017.4500-Niorg) B.)
Sulfide	(mg/l)	Not detected	< 1.0	ZnS Precipitation, Iodometric (SM 2017.4500-S ²⁻ F.)
Settleable Solids	(ml/l)	< 0.5	< 0.5	Inhoff Cone, Volumetric (SM 2017.2540 F.)

SM : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017.

❖ The results relate only to the samples tested and apply to customer's self-drawn samples only.

❖ This analysis report may not be reproduced other than in full, except with the prior written approval of the technical manager.

Approved by ศุภณัฐ วัฒนวิจิตร พญณภากร คุ้ม
 ๙/๙ ธันวาคม ๒๕๖๖



บริษัท วิศวกรรมเคมี จำกัด

THAI CHEMICAL & ENGINEERING CO., LTD.

1048/2 ซ.สุขุมวิท 66/1 อ.สุขุมวิท แขวงพระโขนงใต้ เขตพระโขนง กรุงเทพฯ 10260 โทร. 0-2744-9911 แฟกซ์ 0-2393-0165

1048/2 Soi Sukhumvit 66/1, Prakanong Tai, Prakanong, Bangkok 10260 TEL. 0-2744-9911 FAX 0-2393-0165

No. 2853/66

WASTE WATER ANALYSIS REPORT

Date : 29/11/66

Analysis Date : 22/11/66-28/11/66

Customer : บริษัท พงษ์พรหม แมนชั่น จำกัด สำนักงานใหญ่

Sampling Date : 22/11/66

Address : เลขที่ 24 ซอยประดิพัทธ์ 23 ถนนประดิพัทธ์ แขวงสามเสนใต้

Sampling Time : 14.20

เขตพญาไท กรุงเทพฯ

Received Date : 22/11/66

Tel : 084-703-9129

Reference Number	WP/WS 5238/66			Method of Analysis
Parameter	Unit	Effluent	Standard	
Appearance		เหลืองค่อนข้างขุ่น		
pH		@ 25.6 °C = 7.6	5-9	Electrometric (SM 2017:4500-H ⁺ B.)
Biochemical Oxygen Demand	(mg/l)	20	≤ 30	5-Day BOD Test, Azide Modification (SM 2017:5210 B.)
Total Suspended Solids	(mg/l)	43	≤ 40	Dried at 103-105°C (SM 2017:2540 D.)
Total Dissolved Solids	(mg/l)	316	≤ 500	Dried at 180°C (SM 2017:2540 C.)
Oil & Grease	(mg/l)	< 5.00	≤ 20	Soxhlet Extraction (SM 2017:5520 D.)
Total Kjeldahl Nitrogen	(mg/l)	49.64	≤ 35	Macro-Kjeldahl, Titrimetric (SM 2017:4500-N(org) B.)
Sulfide	(mg/l)	Not detected	≤ 1.0	ZnS Precipitation, Iodometric (SM 2017:4500-S ²⁻ F.)
Settleable Solids	(ml/l)	< 0.5	≤ 0.5	Imhoff Cone, Volumetric (SM 2017:2540 F.)

SM : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WFF, 23 rd Edition, 2017

❖ The results relate only to the samples tested and apply to customer's self-drawn samples only.

❖ This analysis report may not be reproduced other than in full, except with the prior written approval of the technical manager.

Approved by

พงษ์พรหมแมนชั่น
 วิศวกรเคมี
 (ธีรยุทธ วัฒนกิจ)



บริษัท วิศวกรรมเคมี จำกัด

THAI CHEMICAL & ENGINEERING CO., LTD.

1048/2 ซ.สุขุมวิท 66/1 ถ.สุขุมวิท แขวงพระโขนงใต้ เขตพระโขนง กรุงเทพฯ 10260 โทร. 0-2744-9911 แฟกซ์ 0-2393-0165

1048/2 Soi Sukhumvit 66/1, Sukhumvit Rd., Prakanong Tai, Prakanong, Bangkok 10260 TEL: 0-2744-9911 FAX: 0-2393-0165

No. 3082/66

WASTE WATER ANALYSIS REPORT

Date : 25/12/66

Analysis Date : 13/12/66-22/12/66

Customer : บริษัท พงษ์พรหม เนนชั่น จำกัด สำนักงานใหญ่

Sampling Date : 13/12/66

Address : เลขที่ 24 ซอยประดิพัทธ์ 23 ถนนประดิพัทธ์ แขวงสามเสนใต้

Sampling Time : 14.00

เขตพญาไท กรุงเทพฯ

Received Date : 13/12/66

Tel : 084-703-9129

Reference Number	WP/WS 5649/66			
Parameter	Unit	Effluent	Standard	Method of Analysis
Appearance		เหลืองขุ่นมีตะกอน		
pH		@ 25.2 °C = 7.7	5-9	Electrometric (SM 2017:4500-H ⁺ B.)
Biochemical Oxygen Demand	(mg/l)	36	≤ 30	5-Day BOD Test, Azide Modification (SM 2017:5210 B.)
Total Suspended Solids	(mg/l)	63	≤ 40	Dried at 103-105°C (SM 2017:2540 D.)
Total Dissolved Solids	(mg/l)	332	≤ 500	Dried at 180°C (SM 2017:2540 C.)
Oil & Grease	(mg/l)	7.18	≤ 20	Soxhlet Extraction (SM 2017:5520 D.)
Total Kjeldahl Nitrogen	(mg/l)	49.08	≤ 35	Macro-Kjeldahl, Titrimetric (SM 2017:4500-N(org) B.)
Sulfide	(mg/l)	Not detected	≤ 1.0	ZnS Precipitation, Iodometric (SM 2017:4500-S ²⁻ F.)
Settleable Solids	(ml/l)	< 0.5	≤ 0.5	Imhoff Cone, Volumetric (SM 2017:2540 F.)

SM : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23rd Edition, 2017.

◆ The results relate only to the samples tested and apply to customer's self-drawn samples only.

◆ This analysis report may not be reproduced other than in full, except with the prior written approval of the technical manager.

Approved by ศุภณัฐวณิช พลสมกระต่าย
 ๙/๙ ธีรณัฐวณิช พลสมกระต่าย

ภาคผนวก 4

หนังสือรับรองห้องปฏิบัติการวิเคราะห์เอกชน



ที่ อก ๐๓๑๐(๑)/ ๕๕๓ ๓

กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท
เขตราชเทวี กรุงเทพฯ ๑๐๔๐๐

๙ พฤษภาคม ๒๕๖๕

เรื่อง ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

เรียน กรรมการผู้จัดการ บริษัท วิศวกรรมเคมี จำกัด

อ้างถึง คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และชนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน
ลงวันที่ ๒ มีนาคม ๒๕๖๕สิ่งที่ส่งมาด้วย เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
บริษัท วิศวกรรมเคมี จำกัด จำนวน ๑ แผ่น

ตามหนังสือที่อ้างถึง บริษัท วิศวกรรมเคมี จำกัด ขอต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน เลขทะเบียน ๖-๐๐๑ สถานที่ตั้งเลขที่ ๑๐๔๘/๒ ซอยสุขุมวิท ๖๖/๑ ถนนสุขุมวิท แขวงพระโขนงใต้ เขตพระโขนง กรุงเทพมหานคร ต่อกรมโรงงานอุตสาหกรรม นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว ให้บริษัท วิศวกรรมเคมี จำกัด ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน โดยมีองค์ประกอบดังนี้

ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์

- ๑) นางสาวธัญญรัตน์ พลอยกระจำจ
- ๒) นางสาวกรานตนา สว่างรุ่งรัตน์ะ

ทะเบียนเลขที่ ๖-๐๐๑-ค-๐๐๐๑

ทะเบียนเลขที่ ๖-๐๐๑-ค-๐๐๐๒

ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์

- ๑) นางสาวรัชนิษฐ์ วนิชกุลวิริยะ
- ๒) นางสาวกมลชนก วงศ์พนาไกร
- ๓) นางสาววรลักษณ์ เทียนกระจำจ
- ๔) นางสาวเกวรินทร์ ศิริวัฒนสกุล
- ๕) นางสาวจิราพร เบญจจริยาภรณ์
- ๖) นางสาวจารวรรณ ต้นสกุล
- ๗) นางสาวกาญจนา ลาชุมเหล็ก
- ๘) นางสาวกิตติยา นารี

ทะเบียนเลขที่ ๖-๐๐๑-จ-๐๐๐๑

ทะเบียนเลขที่ ๖-๐๐๑-จ-๐๐๐๒

ทะเบียนเลขที่ ๖-๐๐๑-จ-๐๐๐๓

ทะเบียนเลขที่ ๖-๐๐๑-จ-๐๐๐๔

ทะเบียนเลขที่ ๖-๐๐๑-จ-๐๐๐๕

ทะเบียนเลขที่ ๖-๐๐๑-จ-๐๐๐๖

ทะเบียนเลขที่ ๖-๐๐๑-จ-๐๐๐๗

ทะเบียนเลขที่ ๖-๐๐๑-จ-๐๐๐๘

ค. ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนให้วิเคราะห์ในน้ำเสีย ตามสิ่งที่ส่งมาด้วย

หนังสือฉบับนี้...



- ๒ -

หนังสือฉบับนี้จะหมดอายุในวันที่ ๘ เมษายน ๒๕๖๘ หากประสงค์จะต่ออายุหนังสือ
รับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ให้ยื่นคำขอต่ออายุพร้อมเอกสารประกอบคำขอ
ต่อกรมโรงงานอุตสาหกรรม ภายใน ๓๐ วัน ก่อนวันสิ้นอายุของหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์
เอกชน ทั้งนี้ สามารถยื่นคำขอผ่านระบบอิเล็กทรอนิกส์ได้ที่หน้าเว็บไซต์กรมโรงงานอุตสาหกรรม ตาม QR Code
ท้ายหนังสือฉบับนี้

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ



(นางจันทา เดชะศรีรินทร์)

ผู้อำนวยการกองวิจัยและเคอีนกัมลพิษโรงงาน
ปฏิบัติราชการแทนอธิบดีกรมโรงงานอุตสาหกรรม

ยื่นคำขอผ่านระบบอิเล็กทรอนิกส์

กองวิจัยและเคอีนกัมลพิษโรงงาน
กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ
โทร. ๐ ๒๕๓๐ ๖๓๑๒ ต่อ ๒๑๐๓-๕
โทรสาร ๐ ๒๕๓๐ ๖๓๑๒ ต่อ ๒๑๔๔
ไปรษณีย์อิเล็กทรอนิกส์ saraban@diw.mail.go.th



เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท วิศวกรรมเคมี จำกัด

เลขทะเบียน ๖-๐๐๑

ที่ อก ๐๓๒๐(๑)/ ๕๕๓๓

ถึงวันที่ ๙ พฤษภาคม ๒๕๖๕

ขอขยายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๒๐ รายการ

น้ำเสีย จำนวน 20 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Arsenic	Digestion, Inductively Coupled Plasma Method ^[2]
2	Barium	Digestion, Inductively Coupled Plasma Method ^[2]
3	Biochemical Oxygen Demand	5-Day BOD Test, Azide Modification Method ^[2]
4	Cadmium	Digestion, Inductively Coupled Plasma Method ^[2]
5	Chemical Oxygen Demand	1) Closed Reflux, Colorimetric Method ^[2] 2) Closed Reflux, Titrimetric Method ^[2]
6	Chromium	Digestion, Inductively Coupled Plasma Method ^[2]
7	Color	ADMI Weighted-Ordinate Spectrophotometric Method ^[2]
8	Copper	Digestion, Inductively Coupled Plasma Method ^[2]
9	Formaldehyde	Distillation, Colorimetric Method ^[1]
10	Lead	Digestion, Inductively Coupled Plasma Method ^[2]
11	Manganese	Digestion, Inductively Coupled Plasma Method ^[2]
12	Nickel	Digestion, Inductively Coupled Plasma Method ^[2]
13	Oil & Grease	Soxhlet Extraction Method ^[2]
14	pH	Electrometric Method ^[2]
15	Selenium	Digestion, Inductively Coupled Plasma Method ^[2]
16	Sulfide	Iodometric Method ^[2]
17	Total Dissolved Solids	Dried at 180 °C ^[2]
18	Total Kjeldahl Nitrogen	Macro Kjeldahl Method ^[2]
19	Total Suspended Solids	Dried at 103-105 °C ^[2]
20	Zinc	Digestion, Inductively Coupled Plasma Method ^[2]

เอกสารอ้างอิง

1. สมาคมวิศวกรรมสิ่งแวดล้อมแห่งประเทศไทย. คู่มือวิเคราะห์น้ำเสีย. พิมพ์ครั้งที่ 4. กรุงเทพฯ: เรือนแก้วการพิมพ์, 2547.
2. APHA, AWWA, WEF. Standard Methods for the Examination of Water and Wastewater. 23rd ed. Washington, DC: APHA, 2017.

ภาคผนวก 5

เอกสารสอบเทียบเครื่องมือที่ใช้ในการวิเคราะห์



Certificate of Calibration

Equipment:	Balance	Certificate No.:	C01212659
Model:	BSA224S-CW	Issued Date:	13 September 2021
Serial No. (or ID.):	28092544 (INS/LB-109)	Job No.:	KSPR2112116
Manufacturer:	Sartorius	Page:	1 of 2
Condition:	In condition		

Customer: THAI CHEMICAL & ENGINEERING CO., LTD.
1048/2 Soi Sukhumvit 66/1, Sukhumvit Road,
Bangchak, Phrakhanong, Bangkok 10260 Thailand

Environment Condition: Temperature 28 °C ± 0.2 °C
Humidity 50 %RH ± 1.8 %RH

Calibration Place: THAI CHEMICAL & ENGINEERING CO., LTD. (Laboratory)
1048/2 Soi Sukhumvit 66/1, Sukhumvit Road,
Bangchak, Phrakhanong, Bangkok 10260 Thailand

Calibration By: Mr. Adinan Ninviboon

Calibration Date: 13 September 2021

The Method used: In house method, SPCC-WI-47, base on UKAS Lab 14

Traceability: This certificate is traceable to the SI Units maintained by National Institute of Metrology (NIMT), Thailand through SPC RT Co., Ltd. Certificate No. C02210017

(Mr. Adinan Ninviboon)

Person in charge

SERT
บริษัท เอสพีซี อาร์ที จำกัด
SPC RT Co., Ltd.

(Mr. Rungrod Jenkittrakulchai)

Authorized signatory


This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).

These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of SPC RT Co., Ltd.

Calibration Results:**Without Adjustment**

Eccentric Error: Weight to be 1/4 or 1/3 of Maximum capacity, taken from the center of the pan as a zero reference.



Nominal Test Value 50 (g)

Reference Points (g)				
A	B	C	D	E
-	0.0001	0.0001	0.0000	-0.0001

Repeatability: Determination of the standard deviation of weighing balance., Readability 0.0001 (g)

Nominal test value (g)	Standard Deviation
20	0.00005
200	0.00006

Departure of indication from nominal value., Readability 0.0001 (g)

Nominal Value (g)	Conventional Mass (g)	Displayed Value (g)	Correction of Balance (g)	Uncertainty (g)	k
1	1.00001	1.0000	0.0000	0.00011	2.05
2	2.00002	2.0000	0.0000	0.00011	2.05
5	4.99999	5.0000	0.0000	0.00012	2.05
10	10.00000	10.0000	0.0000	0.00012	2.05
20	19.99999	20.0000	0.0000	0.00012	2.04
50	49.99997	50.0000	0.0000	0.00013	2.03
100	100.00000	100.0000	0.0000	0.00018	2.01
120	119.99999	120.0000	0.0000	0.00021	2.00
150	149.99997	150.0000	0.0000	0.00024	2.00
200	199.99990	200.0000	-0.0001	0.00030	2.00

The End of Certificate

Statements of conformity:

This conformity certificate documents the validity of the following statements of conformity based on the measurement results of corresponding calibration certificate:

The correction of indication determined during calibration are under given measurement and environmental conditions and considering the expanded measurement uncertainty (coverage probability 95%) within the specification. The given measurement uncertainty already includes other all effects by according to the standard method, UKAS Lab14. Therefore, those parameters have not been assessed separately.

Tolerance and Decision rules:

Assessment of the conformity of the measurement device are done based on direct comparison of the relevant measurement results with the tolerances and decision rule are prescribed by the customer.

- Decision rule :** ☐ Choice A Binary Statement for Simple Acceptance Rule ($w = 0$), Specific Risk < 50% PFA
☒ Choice B Non-binary statement with guard band ($w = 1 U$), Specific Risk < 2.5% PFA
☐ Choice C Customer defined, Customers may define arbitrary multiple of r to have applied as guard band ($w = r U$) .
; PFA – Probability of False Accept

Without Adjustment**Readability; 0.0001 g****Tolerances : 0.0005 g**

Nominal Value g	Correction of Balance g	Guard band (w) g	Tolerance ($\pm$) g	Conformity
1	0.0000	0.00011	0.0005	Pass
2	0.0000	0.00011	0.0005	Pass
5	0.0000	0.00012	0.0005	Pass
10	0.0000	0.00012	0.0005	Pass
20	0.0000	0.00012	0.0005	Pass
50	0.0000	0.00013	0.0005	Pass
100	0.0000	0.00018	0.0005	Pass
120	0.0000	0.00021	0.0005	Pass
150	0.0000	0.00024	0.0005	Pass
200	-0.0001	0.00030	0.0005	Pass

The validity of the statements of conformity cannot be guaranteed for different places of use, environmental conditions or improper use.

The End of Statements of Certificate



Certificate of Calibration

Equipment:	Hot Air Oven	Certificate No.:	C31211853
Model:	UF 55	Issued Date:	16 September 2021
Serial No.(or ID):	B218.3817 (IN-LB-134)	Job No.:	KSPR2112118
Manufacturer:	Memmert	Page:	1 of 4
Condition:	In Condition	Ventilation Valve:	Closed
Shelves(pc.):	2		

Customer: THAI CHEMICAL & ENGINEERING CO., LTD.
1048/2 Soi Sukhumvit 66/1, Sukhumvit Road,
Bangchak, Phrakhanong, Bangkok 10260 Thailand

Environment Condition:

Temperature:	26 °C	±	0.3 °C
Humidity:	54 %RH	±	2.2 %RH
Voltage:	225 VAC	±	1.9 VAC

Calibration Place: THAI CHEMICAL & ENGINEERING CO., LTD. (Laboratory)
1048/2 Soi Sukhumvit 66/1, Sukhumvit Road,
Bangchak, Phrakhanong, Bangkok 10260 Thailand

Calibration By: Mr. Tharanid Fasawang TR
Calibration Date: 13 September 2021
The Method used: In house method, SPCC-WI-16, base on TLAS-G20
Traceability: This certificate is traceable to the SI Units maintained by National Institute of Metrology (NIMT), Thailand through SPC RT Co., Ltd. Certificate No. C10210008



(Mr. Tharanid Fasawang)

Person in charge



บริษัท เอสพีซี อาร์ที จำกัด
SPC RT Co., Ltd.




(Mr. Udon Srichana)

Authorized signatory

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor ($k=2$) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).

These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of SPC RT Co., Ltd.



Standard Installation Locations

Volume (Calibration Zone)= 16 (Liters)

Inside chamber: $W = 40$ (cm) $D = 33$ (cm) $H = 40$ (cm)

Standard Locations (#1, #2, #3, #4): $w = 5$ (cm) $d = 5$ (cm) $h = 5$ (cm)

Standard Locations (#5, #6, #7, #8): $w = 5$ (cm) $d = 5$ (cm) $h = 12$ (cm)

#9: Geometric center of the chamber

Position of Std	#1	#2	#3	#4	#5	#6	#7	#8	#9
Channel of Logger	1	2	3	4	5	6	7	8	9

Definitions

Indicating Temperature: The average reading of indicating device which forms the integral part of the enclosure.

Measured Temperature: The average reading of standards at any positions or location.

Measured Uniformity: The maximum difference of measured temperatures between of any probes and the measured temperature at the reference location which are observed at same time or at close observation time as possible to determine the temperature pattern or homogeneity with the chamber at steady-state. The reference probe is preferably located in the geometric center of the chamber.

Measured Stability: The one-half of greatest maximum difference of measured temperatures at any one probe.

Overall Variation: The difference of maximum and minimum measured temperatures throughout observation time.

Certificate No.: C31211853

Page: 3 of 4

Calibration Results:
Without adjustment

Measurement Temperature at Spread Locations, Indicating of Unit Under Calibration: 104.0 °C

Locations	Measured Temperature (°C)	Correction of UUC. (°C)	Uncertainty (± °C)
#1	104.06	0.06	0.40
#2	103.89	-0.11	0.40
#3	103.99	-0.01	0.40
#4	103.87	-0.13	0.40
#5	104.23	0.23	0.41
#6	104.26	0.26	0.40
#7	105.03	1.03	0.40
#8	104.06	0.06	0.40
#9	104.07	0.07	0.40

Temperature Distribution

Desired (°C)	Setting (°C)	Indicating (°C)	Measured Temperature at Spread Locations (°C)									Uncertainty (± °C)*
			#1	#2	#3	#4	#5	#6	#7	#8	#9	
104.0	104.0	104.0	104.06	103.89	103.99	103.87	104.23	104.26	105.03	104.06	104.07	0.41

Chamber Characterization

Indicating (°C)	Measured Uniformity (°C)	Measured Stability (± °C)	Overall Variation (°C)
104.0	1.03	0.11	1.33

Note: * Maximum uncertainty of the each position

Certificate No.: C31211853

Page: 4 of 4

Without adjustment (Cont.)

Measurement Temperature at Spread Locations, Indicating of Unit Under Calibration: 180.0 °C

Locations	Measured Temperature (°C)	Correction of UUC. (°C)	Uncertainty (± °C)
#1	179.55	-0.45	0.50
#2	179.36	-0.64	0.50
#3	179.45	-0.55	0.50
#4	179.04	-0.96	0.50
#5	180.27	0.27	0.52
#6	180.38	0.38	0.50
#7	181.33	1.33	0.52
#8	179.69	-0.31	0.50
#9	179.77	-0.23	0.50

Temperature Distribution

Desired (°C)	Setting (°C)	Indicating (°C)	Measured Temperature at Spread Locations (°C)									Uncertainty (± °C)*
			#1	#2	#3	#4	#5	#6	#7	#8	#9	
180.0	180.0	180.0	179.55	179.36	179.45	179.04	180.27	180.38	181.33	179.69	179.77	0.52

Chamber Characterization

Indicating (°C)	Measured Uniformity (°C)	Measured Stability (± °C)	Overall Variation (°C)
180.0	1.68	0.13	2.50

Note: * Maximum uncertainty of the each position

The End of Certificate

Certificate No.: C31211853 Page: 1 of 2

Statements of conformity:

This conformity certificate documents the validity of the following statements of conformity based on the measurement results of corresponding calibration certificate:

The correction of indication determined during calibration are under given measurement and environmental conditions and considering the expanded measurement uncertainty (coverage probability 95%) within the specification. The given measurement uncertainty already includes other all effects by according to the standard method, TLAS-G20. Therefore, those parameters have not been assessed separately.

Tolerance and Decision rules:

Assessment of the conformity of the measurement device are done based on direct comparison of the relevant measurement results with the tolerances and decision rule are prescribed by the customer.

- Decision rule :** ☐ Choice A Binary Statement for Simple Acceptance Rule ($w = 0$), Specific Risk < 50% PFA
☒ Choice B Non-binary statement with guard band ($w = 1$ U), Specific Risk < 2.5% PFA
☐ Choice C Customer defined, Customers may define arbitrary multiple of r to have applied as guard band ($w = r$ U) .
; PFA – Probability of False Accept

Without adjustment**Desired Temperature : 104.0°C Tolerances : 1.0 °C**

Measurement Temperature at Spread Locations, Indicating of Unit Under Calibration: 104.0 °C

Locations	Measured (°C)	Correction of UUC. (°C)	Guard band (W) (± °C)	Tolerance (± °C)	Conformity
#1	104.06	0.06	0.40	1.0	Pass
#2	103.89	-0.11	0.40	1.0	Pass
#3	103.99	-0.01	0.40	1.0	Pass
#4	103.87	-0.13	0.40	1.0	Pass
#5	104.23	0.23	0.41	1.0	Pass
#6	104.26	0.26	0.40	1.0	Pass
#7	105.03	1.03	0.40	1.0	Condition Fail
#8	104.06	0.06	0.40	1.0	Pass
#9	104.07	0.07	0.40	1.0	Pass

Correction of UUC.* = Measured Temperature - Desired Temperature

The validity of the statements of conformity cannot be guaranteed for different places of use, environmental conditions or improper use.

Certificate No.: C31211853 Page: 2 of 2

Statements of conformity:(Cont.)**Without adjustment (Cont.)**

Desired Temperature : 180.0°C Tolerances : 2.0 °C

Measurement Temperature at Spread Locations, Indicating of Unit Under Calibration: 180.0 °C

Locations	Measured (°C)	Correction of UUC.* (°C)	Guard band (W) (± °C)	Tolerance (± °C)	Conformity
#1	179.55	-0.45	0.50	2.0	Pass
#2	179.36	-0.64	0.50	2.0	Pass
#3	179.45	-0.55	0.50	2.0	Pass
#4	179.04	-0.96	0.50	2.0	Pass
#5	180.27	0.27	0.52	2.0	Pass
#6	180.38	0.38	0.50	2.0	Pass
#7	181.33	1.33	0.52	2.0	Pass
#8	179.69	-0.31	0.50	2.0	Pass
#9	179.77	-0.23	0.50	2.0	Pass

Correction of UUC.* = Measured Temperature - Desired Temperature

The validity of the statements of conformity cannot be guaranteed for different places of use, environmental conditions or improper use.

The End of Statements of Conformity



CALIBRATION CERTIFICATE

Date of Issue Jul 12, 2021 **Cert No.** 21/2561
Site Calibration **Order No.** 21070314

Customer Thai Chemical & Engineering Co., Ltd.
1048/2 Soi Sukhumvit 66/1, Sukhumvit Road, Prakanong Tai, Prakanong, Bangkok 10260

Place of Calibration 1350, 1352 Sutthisarnwinitchai Rd, Dindaeng, Bangkok 10400. (Calibration Room)

Description Cooled Incubator
Model IPP750eco
Serial No. V821.0094
ID.No. -
Date of Receipt Jul 07, 2021
Date of Calibration Jul 07, 2021

Environment

Temperature	(Min)	23.2	°C	(Max)	26.6	°C
Relative Humidity	(Min)	51.3	%RH	(Max)	76.7	%RH

TR

Calibration Method

WI-17 : The reference thermometer was placed into the chamber and measurement was performed based on AS-2853.
The temperature scale in use at this laboratory is the International Temperature Scale of 1990.

Standard

1) Data Acquisition with Sensor Model 34972A S/N. MY49013906, Certificate No. QR21-0136, Calibrated by Quality Reborn Co., Ltd., ONAC Calibration No. 0292.

This certificate is traceable to SI unit.

0.00



CALIBRATION CERTIFICATE


Date of Issue Jul 12, 2021

Cert No. 21/2561

Site Calibration

Order No. 21070314

Results (without adjustment)



Position of reference thermometers were placed

Note.

- 1). Dimension (W x L x H) is 104 x 60 x 120 cm.
- 2). Stability - greatest one half of difference between max peak and min peak of each reference probe measured temperature obtained during the calibration interval.
- 3). Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.



CALIBRATION CERTIFICATE

Date of Issue Jul 12, 2021

Cert No. 21/2561

Site Calibration

Order No. 21070314

Results (without adjustment)

UUC Setting (°C)	UUC Reading (°C)	Reference Thermometer (°C)		Stability $\pm$ (°C)	Uniformity (°C)	Uncertainty $\pm$ (°C)
20.0	20.0	Position 1	20.230	0.031	0.143	0.30
		Position 2	20.130			
		Position 3	20.137			
		Position 4	20.167			
		Position 5	20.234			
		Position 6	20.212			
		Position 7	20.192			
		Position 8	20.214			
		Position 9	20.129			

The stability and uniformity was taken into account in the measurement uncertainty stated.

The above results are valid exclusively for calibration samples as mentioned in the report.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with ONAC requirements.

APPROVED SIGNATORY :

(MR. DAMRONG MULSING)



Certificate of Calibration

Equipment:	Incubator	Certificate No.:	C31211845
Model:	IN 55	Issued Date:	14 September 2021
Serial No.(or ID):	D213.0681 (INS/LB-121)	Job No.:	KSPR2112121
Manufacturer:	Memmert	Page:	1 of 3
Condition:	In Condition	Ventilation Valve:	Closed
Shelves(pc.):	1		

Customer: THAI CHEMICAL & ENGINEERING CO., LTD.
1048/2 Soi Sukhumvit 66/1, Sukhumvit Road,
Bangchak, Phrakhanong, Bangkok 10260 Thailand

Environment Condition:

Temperature:	27 °C	±	0.6 °C
Humidity:	59 %RH	±	3.8 %RH
Voltage:	225 VAC	±	0.8 VAC

Calibration Place: THAI CHEMICAL & ENGINEERING CO., LTD. (Laboratory)
1048/2 Soi Sukhumvit 66/1, Sukhumvit Road,
Bangchak, Phrakhanong, Bangkok 10260 Thailand

Calibration By: Mr. Chanachol Moohammudrosol

Calibration Date: 13 September 2021

The Method used: In house method, SPCC-WI-16, base on TLAS-G20

Traceability: This certificate is traceable to the SI Units maintained by National Institute of Metrology (NIMT), Thailand through SPC RT Co., Ltd. Certificate No. C10210006

(Mr. Chanachol Moohammudrosol)

Person in charge

บริษัท เอสพีซี อาร์ที จำกัด
SPC RT Co., Ltd.


(Mr. Udon Srichana)

Authorized signatory

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).

These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of SPC RT Co., Ltd.



Standard Installation Locations

Volume (Calibration Zone)= 21 (Liters)

Inside chamber: W = 40 (cm) D = 33 (cm) H = 40 (cm)

Standard Locations (#1, #2, #3, #4): w = 5 (cm) d = 5 (cm) h = 5 (cm)

Standard Locations (#5, #6, #7, #8): w = 5 (cm) d = 5 (cm) h = 5 (cm)

#9: Geometric center of the chamber

Position of Std	#1	#2	#3	#4	#5	#6	#7	#8	#9
Channel of Logger	1	2	3	4	5	6	7	8	9

Definitions

Indicating Temperature: The average reading of indicating device which forms the integral part of the enclosure.

Measured Temperature: The average reading of standards at any positions or location.

Measured Uniformity: The maximum difference of measured temperatures between of any probes and the measured temperature at the reference location which are observed at same time or at close observation time as possible to determine the temperature pattern or homogeneity with the chamber at steady-state. The reference probe is preferably located in the geometric center of the chamber.

Measured Stability: The one-half of greatest maximum difference of measured temperatures at any one probe.

Overall Variation: The difference of maximum and minimum measured temperatures throughout observation time.

Certificate No.: C31211845

Page: 3 of 3

Calibration Results:**Without adjustment**

Measurement Temperature at Spread Locations, Indicating of Unit Under Calibration: 35.0 °C

Locations	Measured Temperature (°C)	Correction of UUC. (°C)	Uncertainty (± °C)
#1	34.92	-0.08	0.23
#2	34.96	-0.04	0.23
#3	34.93	-0.07	0.23
#4	34.82	-0.18	0.23
#5	34.78	-0.22	0.23
#6	35.10	0.10	0.23
#7	34.76	-0.24	0.25
#8	34.89	-0.11	0.23
#9	34.92	-0.08	0.23

Temperature Distribution

Desired (°C)	Setting (°C)	Indicating (°C)	Measured Temperature at Spread Locations (°C)									Uncertainty
			#1	#2	#3	#4	#5	#6	#7	#8	#9	(± °C)*
35.0	35.0	35.0	34.92	34.96	34.93	34.82	34.78	35.10	34.76	34.89	34.92	0.25

Chamber Characterization

Indicating (°C)	Measured Uniformity (°C)	Measured Stability (± °C)	Overall Variation (°C)
35.0	0.27	0.13	0.49

Note: * Maximum uncertainty of the each position

The End of Certificate

Certificate No.: C31211845 Page: 1 of 1

Statements of conformity:

This conformity certificate documents the validity of the following statements of conformity based on the measurement results of corresponding calibration certificate:

The correction of indication determined during calibration are under given measurement and environmental conditions and considering the expanded measurement uncertainty (coverage probability 95%) within the specification. The given measurement uncertainty already includes other all effects by according to the standard method, TLAS-G20. Therefore, those parameters have not been assessed separately.

Tolerance and Decision rules:

Assessment of the conformity of the measurement device are done based on direct comparison of the relevant measurement results with the tolerances and decision rule are prescribed by the customer.

- Decision rule :**
- ☐ Choice A Binary Statement for Simple Acceptance Rule ($w = 0$), Specific Risk < 50% PFA
 - ☒ Choice B Non-binary statement with guard band ($w = 1 U$), Specific Risk < 2.5% PFA
 - ☐ Choice C Customer defined, Customers may define arbitrary multiple of r to have applied as guard band ($w = r U$).
; PFA – Probability of False Accept

Without adjustment

Desired Temperature : 35.0°C Tolerances : 1.0 °C

Measurement Temperature at Spread Locations, Indicating of Unit Under Calibration: 35.0 °C

Locations	Measured (°C)	Correction of UUC. (°C)	Guard band (W) (± °C)	Tolerance (± °C)	Conformity
#1	34.92	-0.08	0.23	1.0	Pass
#2	34.96	-0.04	0.23	1.0	Pass
#3	34.93	-0.07	0.23	1.0	Pass
#4	34.82	-0.18	0.23	1.0	Pass
#5	34.78	-0.22	0.23	1.0	Pass
#6	35.10	0.10	0.23	1.0	Pass
#7	34.76	-0.24	0.25	1.0	Pass
#8	34.89	-0.11	0.23	1.0	Pass
#9	34.92	-0.08	0.23	1.0	Pass

Correction of UUC.* = Measured Temperature - Desired Temperature

The validity of the statements of conformity cannot be guaranteed for different places of use, environmental conditions or improper use.

The End of Statements of Conformity



Certificate of Calibration

Equipment:	Incubator	Certificate No.:	C31211846
Model:	IN 55	Issued Date:	14 September 2021
Serial No.(or ID):	D215.1344 (INS/LB-022)	Job No.:	KSPR2112123
Manufacturer:	Memmert	Page:	1 of 3
Condition:	In Condition	Ventilation Valve:	Closed
Shelves(pc.):	2		

Customer: THAI CHEMICAL & ENGINEERING CO., LTD.
1048/2 Soi Sukhumvit 66/1, Sukhumvit Road,
Bangchak, Phrakhanong, Bangkok 10260 Thailand

Environment Condition:

Temperature:	25 °C	±	0.8 °C
Humidity:	60 %RH	±	3.5 %RH
Voltage:	225 VAC	±	1.6 VAC

Calibration Place: THAI CHEMICAL & ENGINEERING CO., LTD. (Laboratory)
1048/2 Soi Sukhumvit 66/1, Sukhumvit Road,
Bangchak, Phrakhanong, Bangkok 10260 Thailand

Calibration By: Mr. Chanachol Moohammudrosol

Calibration Date: 13 September 2021

The Method used: In house method, SPCC-WI-16, base on TLAS-G20

Traceability: This certificate is traceable to the SI Units maintained by National Institute of Metrology (NIMT), Thailand through SPC RT Co., Ltd. Certificate No. C10210006



(Mr. Chanachol Moohammudrosol)

Person in charge



บริษัท เอสพีซี อาร์ที จำกัด
SPC RT Co., Ltd.




(Mr. Udon Srichana)

Authorized signatory

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).

These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of SPC RT Co., Ltd.



Standard Installation Locations

Volume (Calibration Zone)= 21 (Liters)

Inside chamber: W = 40 (cm) D = 33 (cm) H = 40 (cm)

Standard Locations (#1, #2, #3, #4): w = 5 (cm) d = 5 (cm) h = 5 (cm)

Standard Locations (#5, #6, #7, #8): w = 5 (cm) d = 5 (cm) h = 5 (cm)

#9: Geometric center of the chamber

Position of Std	#1	#2	#3	#4	#5	#6	#7	#8	#9
Channel of Logger	1	2	3	4	5	6	7	8	9

Definitions

Indicating Temperature: The average reading of indicating device which forms the integral part of the enclosure.

Measured Temperature: The average reading of standards at any positions or location.

Measured Uniformity: The maximum difference of measured temperatures between of any probes and the measured temperature at the reference location which are observed at same time or at close observation time as possible to determine the temperature pattern or homogeneity with the chamber at steady-state. The reference probe is preferably located in the geometric center of the chamber.

Measured Stability: The one-half of greatest maximum difference of measured temperatures at any one probe.

Overall Variation: The difference of maximum and minimum measured temperatures throughout observation time.

Certificate No.: C31211846

Page: 3 of 3

Calibration Results:

Before adjustment

Setting: Indicating: #1: #2: #3: #4: #5: #6: #7: #8: #9:
45.0 45.0 44.93 44.85 45.03 44.97 44.48 44.41 44.57 44.80 44.85

After adjustment

Measurement Temperature at Spread Locations, Indicating of Unit Under Calibration: 45.0 °C

Locations	Measured Temperature (°C)	Correction of UUC. (°C)	Uncertainty (± °C)
#1	45.33	0.33	0.23
#2	45.21	0.21	0.23
#3	45.41	0.41	0.23
#4	45.31	0.31	0.23
#5	44.90	-0.10	0.23
#6	44.76	-0.24	0.23
#7	44.99	-0.01	0.23
#8	45.18	0.18	0.23
#9	45.23	0.23	0.24

Temperature Distribution

Desired (°C)	Setting (°C)	Indicating (°C)	Measured Temperature at Spread Locations (°C)									Uncertainty (± °C)*
			#1	#2	#3	#4	#5	#6	#7	#8	#9	
45.0	45.0	45.0	45.33	45.21	45.41	45.31	44.90	44.76	44.99	45.18	45.23	0.24

Chamber Characterization

Indicating (°C)	Measured Uniformity (°C)	Measured Stability (± °C)	Overall Variation (°C)
45.0	0.53	0.06	0.72

Note: * Maximum uncertainty of the each position

The End of Certificate

Statements of conformity:

This conformity certificate documents the validity of the following statements of conformity based on the measurement results of corresponding calibration certificate:

The correction of indication determined during calibration are under given measurement and environmental conditions and considering the expanded measurement uncertainty (coverage probability 95%) within the specification. The given measurement uncertainty already includes other all effects by according to the standard method, TLAS-G20. Therefore, those parameters have not been assessed separately.

Tolerance and Decision rules:

Assessment of the conformity of the measurement device are done based on direct comparison of the relevant measurement results with the tolerances and decision rule are prescribed by the customer.

- Decision rule :** ☐ Choice A Binary Statement for Simple Acceptance Rule ($w = 0$), Specific Risk < 50% PFA
☒ Choice B Non-binary statement with guard band ($w = 1 U$), Specific Risk < 2.5% PFA
☐ Choice C Customer defined, Customers may define arbitrary multiple of r to have applied as guard band ($w = r U$) .
; PFA – Probability of False Accept

After adjustment

Desired Temperature : 45.0°C Tolerances : 1.0 °C

Measurement Temperature at Spread Locations, Indicating of Unit Under Calibration: 45.0 °C

Locations	Measured (°C)	Correction of UUC. (°C)	Guard band (W) (± °C)	Tolerance (± °C)	Conformity
#1	45.33	0.33	0.23	1.0	Pass
#2	45.21	0.21	0.23	1.0	Pass
#3	45.41	0.41	0.23	1.0	Pass
#4	45.31	0.31	0.23	1.0	Pass
#5	44.90	-0.10	0.23	1.0	Pass
#6	44.76	-0.24	0.23	1.0	Pass
#7	44.99	-0.01	0.23	1.0	Pass
#8	45.18	0.18	0.23	1.0	Pass
#9	45.23	0.23	0.24	1.0	Pass

Correction of UUC.* = Measured Temperature - Desired Temperature

The validity of the statements of conformity cannot be guaranteed for different places of use, environmental conditions or improper use.

The End of Statements of Conformity



Certificate of Calibration

Equipment: pH METER
Model: Seven Compact S220
Serial No. (or ID.): B914466655
Manufacturer: Mettler Toledo
Electrode Serial No.: 8473213
Condition: In Condition

Certificate No.: C07210227
Issued Date: 6 May 2021
Job No.: KSPR2106247
Page: 1 of 3
Model: InLab Expert Pro-ISM **Brand:** Mettler Toledo

Customer: THAI CHEMICAL & ENGINEERING CO., LTD.
 1048/2 Soi Sukhumvit 66/1, Sukhumvit Road,
 Bangchak, Phrakhanong, Bangkok 10260 Thailand

Environment Condition: Temperature 23 °C ± 2 °C
 Humidity 50 %RH ± 15 %RH

Calibration Place: Environment Laboratory, SPC RT Co., Ltd.
 1194 Soi Wachirathamsathit 57, Sukhumvit 101/1 Rd.,
 Bangchak, Phrakhanong, Bangkok 10260 Thailand

Calibration By: Mr. Vanasapol Lerksanthia

Calibration Date: 6 May 2021

The Method used: In house method, SPCC-WI-58, base on ASTM E 70-07

Traceability: This certificate is traceable to SI Units, Sample Test is assured through primary measurement method Harned cell, through CPAchem Ltd. (ISO/IEC 17034) Certificate No. 724708, 724716, 724715 And pH Scale traceable to the SI Units maintained by National Institute of Metrology (NIMT), Thailand through Industrial Foundation Electrical and Electronics Institute Certificate No. 0627EL20

Vanasapol

(Mr. Vanasapol Lerksanthia)

Person in charge

SERT
 บริษัท เอสพีซี อาร์ที จำกัด
 SPC RT Co., Ltd.

[Signature]

(Mr. Dumrong Boonsopon)

Authorized signatory

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor ($k=2$) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).

These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of SPC RT Co., Ltd.

Calibration Results:

pH Scale

Input	pH Meter Reading			Uncertainty of Measurement (mV)	Coverage Factor (k)
	(mV)	Error (mV)	(pH)		
414.12	414.1	-0.02	0.000	0.065	2.00
354.96	355.0	0.04	1.000	0.065	2.00
295.80	295.8	0.00	2.000	0.065	2.00
236.64	236.7	0.06	3.000	0.065	2.00
177.48	177.5	0.02	4.000	0.065	2.00
118.32	118.4	0.08	5.000	0.065	2.00
59.16	59.2	0.04	6.000	0.065	2.00
0.00	0.1	0.10	7.000	0.065	2.00
-59.16	-59.1	0.06	8.000	0.065	2.00
-118.32	-118.3	0.02	9.000	0.065	2.00
-177.48	-177.4	0.08	10.000	0.065	2.00
-236.64	-236.5	0.14	11.000	0.065	2.00
-295.80	-295.7	0.10	12.000	0.065	2.00
-354.96	-354.8	0.16	13.000	0.065	2.00
-414.12	-414.0	0.12	14.000	0.065	2.00

Electrode Test Results*

The three-point calibration using three standard buffer solutions; pH 4.008 , pH 6.985 and pH 10.012

The practical slope of the pH electrode; 57.90 (mV/pH), 97.87%

The zero point of the pH electrode; 6.94 (pH)

Sample Test Results

Standard Buffer Solution (pH)	Unit Under Calibration (pH)	Difference (pH)	Uncertainty of Measurement (pH)	Coverage Factor (k)
4.008	4.008	0.000	0.0070	2.00
6.985	6.986	0.001	0.0078	2.00
10.012	10.013	0.001	0.0070	2.00

* Calibration Marked " Not TISI Accredited " in this Certificate have been included for completeness.

The End of Certificate



Certificate of Calibration

Equipment:	Digital Thermometer	Certificate No.:	C15210165
Model:	Seven Compact S220	Issued Date:	07 May 2021
Serial No.(or ID):	B914466655	Job No.:	KSPR2106245
Manufacturer:	METTLER TOLEDO	Page:	1 of 2
Condition:	In Condition		

Customer: THAI CHEMICAL & ENGINEERING CO., LTD.
1048/2 Soi Sukhumvit 66/1, Sukhumvit Road,
Bangchak, Phrakhanong, Bangkok 10260 Thailand

Environment Condition: Temperature: 22 °C ± 3.0 °C
Humidity: 50 %RH ± 15.0 %RH
Voltage: 230 VAC ± 11.0 VAC

Calibration Place: Sensor Laboratory, SPC RT Co., Ltd.
1194 Soi Wachirathamsathit 57, Sukhumvit 101/1 Rd.,
Bangchak, Prakhanong, Bangkok 10260 Thailand

Calibration By: Mr. Anat Karapitak

Calibration Date: 07 May 2021

The Method used: In house method, SPCC-WI-19, by comparison with standard thermometer

Traceability: This certificate is traceable to the SI Units maintained by National Institute of Metrology (NIMT), Thailand through Quality Reborn Co.,Ltd. (QR) Certificate No. QR21-0415

(Mr. Anat Karapitak)

Person in charge

SPC RT
บริษัท เอสพีซี อาร์ที จำกัด
SPC RT Co., Ltd.

(Mr. Udon Srichana)

Authorized signatory

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).

These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of SPC RT Co., Ltd.

Calibration Results:

Without Adjustment

Sensor Type: Thermistor

Channel: -

Diameter (mm) 12

Length (mm): 100

Immersion (mm): 110

Desired Temp.(°C)	STD. Reading (°C)	UUC. Reading (°C)	Correction of UUC (°C)	Uncertainty (± °C)
20.0	20.004	19.8	0.204	0.10
25.0	25.000	24.9	0.100	0.10
30.0	30.004	29.8	0.204	0.10

The End of Certificate

Preventive Maintenance Kjeldahl

Service No. PM21-S08-141

1. Customer Information

Customer Name	Instrument	Serial Number	Service Date
บริษัท วิศวกรรมเคมี จำกัด 1048/2 ซ.สุขุมวิท 66/1 ถ.สุขุมวิท แขวงพระโขนงใต้ เขตพระโขนง กรุงเทพมหานคร 10260 คุณธัญญารัตน์ Tel: Fax:	K-350	0700000546	09 Sep 2021 PM 2/2

2. Instrument

2.1 Cooling water (If it connected)	OK	NOT OK	Remark
- Temperature 15 – 20 °C	/		
- Cooling water inlet	/		
- Cooling water outlet	/		
- Control Temperature	/		

2.2 Cleaning	DONE	NOT DONE	Remark
- Outside Instrument	/		
- Inside Instrument	/		
- Splash protector	/		
- Condenser	/		

Preventive Maintenance Kjeldahl

2.3 Visual Test

	OK	NOT OK	Remark
- Screw Coupling (between splash protector and condenser)	/		
- Condenser	/		
- Splash protector	/		
- Hypalon connection (connection tube)	/		
- Rubber bung	/		
- Ventilation valve	/		
- PTFE tube	/		
- Cooling water inlet	/		
- Cooling water outlet	/		
- Magnetic valve	/		

2.4 System control

	OK	NOT OK	Remark
- Key board	/		
- Display	/		
- Program	/		
- Adding H ₂ O	-	-	Do not have
- Adding NaOH	/		
- Adding H ₃ BO ₃	-	-	Do not have
- Aspiration	-	-	Do not have

Preventive Maintenance Kjeldahl



2.5 System Distillation	OK	NOT OK	Remark
- Boiler	/		
- Water level sensor	/		
- One way valve	/		
- Pressure switch	/		
- Thermostat	/		
- Steam valve1 (Y4)	/		
- Steam valve2 (Y5)	/		
- Drain valve (Y3)	-	-	Do not have
- Water 3/2 way valve (Y1)	-	-	Do not have

2.6 Hose	OK	NOT OK	Remark
- Unisil hose	/		
- Hypalon hose	/		
- Drain hose	-	-	Do not have
- Viton hose	/		
- Silicone hose	/		

2.7 Diaphragm pump	OK	NOT OK	Remark
- Diaphragm pump for H ₂ O	-	-	Do not have
- Diaphragm pump for NaOH	/		
- Diaphragm pump for H ₃ BO ₃	-	-	Do not have

2.8 Program test	OK	NOT OK	Remark
- Distillation	/		
- Aspiration	-	-	Do not have
- Preheating	/		
- Cleaning	/		



Preventive Maintenance Kjeldahl

3. Function Test

Addition H ₂ O	0 ml	Reaction time	0 min
Addition NaOH	0 ml	Distillation time	5 min
Addition H ₃ BO ₃	0 ml	Steam capacity	100%
		Aspiration	SAM

Result: Water in receiving vessel now approximately 161 ml, 166 ml

4. Summary



All specifications OK	Specification not OK
OK	

Comments

PM K-350 In PM contract 2/2:ok

Signature BUCHI

- Service by Tada S.

Date 09 Sep 2021

- Approve by Suphan C.

Date 13 Sep 2021



Buchi (Thailand) Limited

Preventive Maintenance IR Digestion

Service No. PM21-S08-141

1. Customer Information

Customer Name	Instrument	Serial Number	Service Date
บริษัท วิศวกรรมเคมี จำกัด 1048/2 ซ.สุขุมวิท 66/1 ถ.สุขุมวิท แขวงพระโขนงใต้ เขตพระโขนง กรุงเทพมหานคร 10260	K-424	0600004943	09 Sep 2021 PM 2/2
คุณธัญญารัตน์ Tel: Fax:			

2. Instrument

2.1 Housing	OK	NOT OK	Remark
- Clean the housing	/		
- Visual check	/		
- Check for defects (e.g. cracks)	/		

2.2 Heating	OK	NOT OK	Remark
- Clean insulation plate	/		
- Visual check	/		
- Check heating element	/		

Preventive Maintenance IR Digestion

2.3 Visual Check	OK	NOT OK	Remark
- Connection to suction	/		
- PTFE seal	/		
- O-ring	/		
- Glass holder set	/		
- Suction module	/		

2.4 System control (for K-439 only)	OK	NOT OK	Remark
- Keyboard	/		
- Display	/		
- Program	/		

BUCHI

Preventive Maintenance IR Digestion

3. Summary

All specifications OK	Specification not OK
OK	

Comments

PM K-424 In contract 2/2 : OK



Signature BUCHI

- Service by Tada S.

Date 09 Sep 2021

- Approve by Suraphan C.

Date 13 Sep 2021



Buchi (Thailand) Limited

Preventive Maintenance Scrubber

Service No. PM21-S08-141

1. Customer Information

Customer Name	Instrument	Serial Number	Service Date
บริษัท วิศวกรรมเคมี จำกัด 1048/2 ซ.สุขุมวิท 66/1 ถ.สุขุมวิท แขวงพระโขนงใต้ เขตพระโขนง กรุงเทพมหานคร 10260 คุณธัญญารัตน์ Tel: Fax:	B-414	0700002972	09 Sep 2021 PM 2/2

2. Instrument

2.1 Cooling water (If it connected)	OK	NOT OK	Remark
- Temperature 10 – 20 °C	/		
- Cooling water inlet	/		
- Cooling water outlet	/		

2.2 Cleaning	DONE	NOT DONE	Remark
- Housing	/		
- Condenser	/		
- Swirl disc	/		

Preventive Maintenance Scrubber

2.3 Visual Check	OK	NOT OK	Remark
- Hose connection to suction	/		
- Glassware	/		
- Lip gasket	/		
- GL-14 connector	/		
- Activated charcoal	/		

2.4 Flush Pump



Make sure, the bypass valve is closed completely (for maximum suction power).

- Disconnect the silencer, move it down (or take it away from the instrument), and flush out the pump with at least 500 mL of distilled water through the pump inlet, until the collected washing water is clean.
- Switch on the instrument and collect the waste water from the pump output in a suitable vessel.

Flush pump

☒ OK

☐ NOT OK

2.5 Washing Solution

- Sodium hydroxide 8-10 %, max. 20 %
- Sodium carbonate
 - dissolve 600 g Na_2CO_3 in 3 L distilled warm water, or
 - dissolve 1.7 kg Na_2CO_3 in 10 H_2O in 3 L distilled warm water

Washing solution

☒ OK

☐ NOT OK

Preventive Maintenance Scrubber

3. Summary

All specifications OK	Specification not OK
OK	

Comments

PM B-414 In PM contract 2/2:ok



Signature BUCHI



Buchi (Thailand) Limited

- Service by Tada S.

Date 09 Sep 2021

- Approve by Suvaphan C.

Date 13 Sep 2021

ภาคผนวก 6

เอกสารและรูปภาพประกอบการปฏิบัติ

ตามมาตรการ


ป้ายชื่อโครงการ



ถังน้ำรดน้ำต้นไม้



รูปถ่าย



บริเวณจอดรถ





ที่พักขยะ



ขนย้ายขยะมูลฝอย



อุปกรณ์ดับเพลิง ทางหนีไฟ



ห้ามใช้ลิฟท์ขณะเกิดเพลิงไหม้



พื้นที่สีเขียว

