

ภาคผนวก

- | | |
|-----------|---|
| ภาคผนวก ก | มาตรฐานน้ำทิ้งจากอาคาร |
| ภาคผนวก ข | เอกสารการป้องกันและระงับอัคคีภัย |
| ภาคผนวก ค | อุปกรณ์เก็บตัวอย่างและเครื่องมือตรวจวัด |
| ภาคผนวก ง | ใบรายงานผลการตรวจวิเคราะห์ |
| ภาคผนวก จ | เอกสารสอบเทียบเครื่องมือ |
| ภาคผนวก ฉ | หนังสืออนุญาตขึ้นทะเบียน
ห้องปฏิบัติการวิเคราะห์ออกซน |
| ภาคผนวก ช | หลักฐานการส่งรายงานผลการปฏิบัติตามมาตรการฯ
ฉบับที่ผ่านมา |

ภาคผนวก ก
มาตรฐานนำทึ้งจากอาคาร



ประกาศกรุงเทพมหานครเรื่อง กำหนดมาตรฐานความคุ้มครองระบบนำ้ทั่วไป
จากอาชญากรรมประเภทและบางขนาด

โดยต้องได้รับการอนุมัติจากกระทรวงมหาดไทยให้มีการจัดตั้ง理事会ท้องที่เพื่อจัดการศิริสุขและสืบสานประเพณีด้วยความร่วมมือของชาวบ้านและชุมชนท้องที่ในพื้นที่ ทั้งนี้ ต้องได้รับการอนุมัติจากกระทรวงมหาดไทยและจังหวัดท้องที่ ตามกฎหมายว่าด้วยการปกครองท้องที่ พ.ศ.๒๕๖๙ ไม่เป็นของ理事会ท้องที่ที่ต้องจัดตั้ง ประกอบกันทั้งเป็นการสมควรให้คณะกรรมการการคุณภาพดินที่มีผู้ดูแลอย่างดีของบ้านท้องที่เข้าร่วมเป็นกรรมการ理事会ท้องที่ นักหนែนอาจใช้การที่กำหนดไว้ เกี่ยวกับการคุณภาพดิน จึงสามารถแก้ไขได้โดยไม่ต้องขออนุมัติจากกระทรวงมหาดไทย

อาทิตย์อ่อนหวานตามความในมาตรา ๔๕ ให้ทรงพระราชนูญด้วยส่วนเชิญและรักษาความพำเพี้ยนเดล้อตน แห่งชาติ พ.ศ. ๒๕๓๕ แก่ไขโดยมหาตราช ๑๐๙ ให้ทรงพระราชนูกุณรัฐบาลนูญด้วยให้ห้องดีดองบัน การโภณอ่อนหวานนี้ที่ของส่วนพระราชการ ให้เป็นไปตามพระราชนูญด้วยเริ่มใช้ราชการทั่วไป ทบวง กรม พ.ศ. ๒๕๓๕ พ.ศ. ๒๕๓๕ อัยนั้นพระราชนูญด้วยที่มีนาบานูญด้วยส่วนประกาศเดือนกันกราเดือนธันวา และเดือนพฤษภาคมปุกคุ ซึ่งมาตรา ๒๖ ประกอบกับมาตรา ๓๙ มาตรา ๔๕ มาตรา ๕๐ และมาตรา ๕๑ ของรัฐธรรมนูญแห่งราชอาณาจักรไทยนูญด้วยให้ทรงทำได้ ให้ออกหัวข้อตามนาบานูญด้วยกฤษฎหมาย รัฐมนตรีว่าการกระทรวงพาณิชย์และรัฐมนตรีว่าการกระทรวงมหาดไทยและส่วนราชการอื่นๆ โดยที่แนะนำที่จะออกกฎหมายดังที่ได้เสนอไว้ จึงออกประกาศไว้ ดังต่อไปนี้

ข้อ ๑ ให้เข้าคึกประการกระทรงวิทยาศาสตร์ เทคนิคโน้มีเปลี่ยนสั่นเวลล่อน ทึ่ง กำหนดน้ำครรภูน ความคุณการบรรเทาท้อทึงจากอาการรำงน้ำประกายและนานาชนิด ลงมาที่ ๑๐ มกราคม พ.ศ. ๒๕๓๗

๑๒ ในประการนี้
“จากฯ” นายคุณวานา ใจกลางที่รักษาสิ่งเหล่านี้ ไม่ว่าจะมีลักษณะเป็นอาชญากรรมดีชรา หรือเป็นกุญแจของอาชญากรรมที่ซ่อนอยู่ในไฟฟ้าที่เชื่อมโยงกับเครื่องหัก กัน และไม่ว่าจะมีที่อยู่ขนาดใดก็ตาม หรือมีหลักทรัพย์ที่ซ่อนอยู่ในบ้านทุกวัววะของตนหรือไม่ก็ตาม ต้องได้รับ

(๑) อาคารชุด ตามกฎหมายว่าด้วยอาคารชุด
 (๒) ปั๊มน้ำ รวมอุปกรณ์ที่ใช้ปั๊มน้ำ

(๔) จ่าก้าวไวร์เริ่มเดอกชน ไวร์เริ่มของงานราชการ สถาบันอุดมศึกษานามของเดอกชน หรือสถาบันอุดมศึกษาของงานราชการที่มีพื้นที่สำหรับอบรมกัณฑุกัชชันของอาจารย์เรือกุ่มของอาจารย์ทึ่งเตี้ย เดือด๙๐๐ ดาวรุณเดชชูปัน

(๔) จักษุที่ทำการของทางราชการ รัฐวิสาหกิจ องค์กรระหว่างประเทศ หรือของเอกชน ที่มีเจ้าหน้าที่ดูแลศูนย์เรียนภาษาต่างๆ ของภาครัฐหรือคู่บุญช่วยภาครัฐอื่นๆ ๕๕๐๐ ตามเงื่อนไขข้างไป

(๖) จักษุของศูนย์บริการท้าวหรือห้างสรรพสินค้าที่มีพื้นที่ใช้สอยรวมกันทุกชั้นของอาคารหรือ
อันดับของห้องครัวอีก ๑ ห้อง ๒๐๐ ตารางเมตรหรือต่ำกว่า

(๔) ผลิตภัณฑ์ที่มีหัวเข็มทิ่มสำหรับรับกินทุกๆ ขั้นของอาหารหรืออุปกรณ์ของอาหารตัวต่อตัว ๒,๕๐๐ ตารางเมตร

(๙) ภักดิการหรือร้านอาหารที่มีพื้นที่ให้บริการรวมกันทุกชั้นของอาคารหรือกลุ่มของอาคาร

ເລື່ອງການສະໜັບສະໜູນ ແລ້ວ ພະຍາຍອວວະເມືອນ ອາວລາວຕີ້ວ່າໄປ

(๑) อาจารชุดที่มีจำนวนห้องเรียนสำหรับใช้เป็นต้นที่อยู่อาศัยรวมกันทุกห้องของอาคาร หรือกลุ่มของอาคาร

(๒) โรงเรียนที่มีจำนวนห้องสำหรับไปใช้เป็นห้องทั่วไปรวมกันทุกชั้นของอาคาร หรืออุปกรณ์ของอาคาร

(๑) ให้พักก่อนเข้าห้องสำหรับใช้ปืนที่อยู่อาศัยรวมกันทุกชั้นของอาคาร หรืออุปกรณ์ของอาคาร

(๔) สถานบันบริการที่มีพื้นที่ใช้สอยรวมกันทุกชั้นของอาคาร หรือคุ่มของอาคารตั้งแต่ ๕,๐๐๐ ตร.ม.

(๔) โรงพยาบาลพิษณุโลก รัฐวิสาหกิจ หรือสถานพยาบาล ตามกฎหมายฯ ทั่วทุกส่วนราชการ ที่มีอำนาจสำนักงานบัญชีป้ายไว้ลักษณะเดียวกันรวมถึงทุกหน่วยงานของภาครัฐอื่นๆ ของภาคต้องแต่งตั้ง เมื่อเดือน มกราคม พ.ศ.๒๕๖๐ ให้เป็นผู้ดูแลและรับผิดชอบดูแลการบัญชี ให้ได้โดยทันท่วงที

(๖) ภาคการโรงเรียนนอกชน โรงเรียนของการราชการ สถาบันอุดมศึกษาของกองทัพ หรือสถาบันอุดมศึกษาของทางราชการ ที่มีที่ตั้งที่ใช้สืบรวมทั้งถูกขั้นของอาจารย์อีกปุ่มของชาติแต่เดิม

๕,๐๐๐ ตารางเมตร แล้วมีอีก ๒๕,๐๐๐ ตารางเมตร

- (๑) หอดูก ตามกฎหมายว่าด้วยหอดัก
(๔) สถานบริการประปาส่วนภูมิภาค นำหัวเรื่องดังนี้ ซึ่งมีผู้ให้บริการแก่กรุงศรีฯ ตามกฎหมายว่าด้วยสถานบริการ

(๕) โรงเรียนตามของทางราชการหรือสถานพยาบาล ตามกฎหมายว่าด้วยโรงเรียนของทางราชการ
(๖) อาจาริย์โรงเรียนเอกชน ตามกฎหมายว่าด้วยโรงเรียนเอกชน โรงเรียนของทางราชการ อาจาริย์สถาบันอุดมศึกษากิจกรรมของสอน ตามกฎหมายว่าด้วยสถาบันอุดมศึกษากิจกรรมของสอนและสถาบันอุดมศึกษาของทางราชการ

(๗) อาจาริย์ที่ทำการของทางราชการ รัฐวิสาหกิจ หรือองค์กรระหว่างประเทศและของเอกชน
(๘) อาจาริย์ของศูนย์การศึกษาเรื่องห้ามรับสิ่งของศรีสินค้า
(๙) คลินิก ตามกฎหมายว่าด้วยการสาธารณสุข แต่ไม่วรรณลึง ท่าที่เข็นเรื่องประมง สะพานปลื้มหรือกิจกรรมเพลปล่าน

(๑๐) กัตตาการหรือข้าวอาหาร
“น้ำดื่ม” หมายความว่า น้ำสีเทาที่ผ่านกระบวนการบำบัดน้ำสีขาวแล้วจนเป็นໄ:inline>ความสะอาดตามกฎหมายน้ำที่คนเด็กได้รับไปในประเทศไทย
กระบวนการน้ำที่คนเด็กได้รับไปในประเทศไทย

ข้อ ๑ ให้แบ่งประเภทของอาหารตามข้อ ๒ ออกเป็น ๔ ประเภท คือ

(๑) อาหารประเภท ก.
(๒) อาหารประเภท ข.
(๓) อาหารประเภท ค.
(๔) อาหารประเภท ง.
(๕) อาหารประเภท จ.
ข้อ ๒ อาหารประเภท ก หมายความอ้าง อาหารตัดต่อໄ:inline>
(๑) อาหารชุดที่มีน้ำหนึ่งห้องเชิงหรือใช้เป็นห้องทั่วไปอื่นๆ ที่รวมกันหลักห้องของอาหาร หรือกลุ่มน่องอาหาร ตั้งแต่ ๔๐๐ ห้องลงมาจนถึง

(๒) โรงเรียนที่มีน้ำหนึ่งห้องเชิงหรือใช้เป็นห้องทั่วไปอื่นๆ ที่รวมกันหลักห้องของอาหาร หรือกลุ่มน่องอาหาร ตั้งแต่ ๒๐๐ ห้องลงมาจนถึง

(๓) โรงเรียนตามของทางราชการ รัฐวิสาหกิจหรือสถานพยาบาล ตามกฎหมายว่าด้วยสถาบันพยาบาล ที่มีเด็กเข้าเรียนได้ไว้ใช้ห้องเรียนรวมกันทักษิณของอาหารหรือกลุ่มน่องอาหาร ตั้งแต่ ๑๐๐ ห้องลงมาจนถึง

- (๒) ภาคการที่ก้าว่างของงานราชการ รัฐวิสาหกิจ องค์กรระหว่างประเทศ หรือของอุตสาหกรรมที่มีเป้าหมายที่ใช้สืบรวมกันทุกขั้นของอาคารหรืออุปกรณ์ของอาคารที่ตั้งแต่ ๑๐,๐๐๐ ตารางเมตร แล้วไปถึง ๔๕,๐๐๐ ตารางเมตร

(๓) ภาคของอุปกรณ์การศึกษาที่ห้องสมุดหรือห้องเรียนที่ใช้สืบรวมกันทุกขั้นของอาคารหรืออุปกรณ์ของอาคารที่ตั้งแต่ ๕,๐๐๐ ตารางเมตร แล้วไปถึง ๒๕,๐๐๐ ตารางเมตร

(๔) ผลิตภัณฑ์ที่มีเป้าหมายที่ใช้สืบรวมกันทุกขั้นของอาคารหรืออุปกรณ์ของอาคารที่ตั้งแต่ ๑,๕๐๐ ตารางเมตร แล้วไปถึง ๑,๘๐๐ ตารางเมตร

(๕) ภาคการที่มีเป้าหมายที่ใช้สืบรวมกันทุกขั้นของอาคารหรืออุปกรณ์ของอาคารที่ตั้งแต่ ๑,๕๐๐ ตารางเมตร แล้วไปถึง ๒,๕๐๐ ตารางเมตร

(๖) ภาคการที่มีเป้าหมายที่ใช้สืบรวมกันทุกขั้นของอาคารหรืออุปกรณ์ของอาคารที่ตั้งแต่ ๕๐๐ ตารางเมตร แล้วไปถึง ๒,๕๐๐ ตารางเมตร

๙๖ ๖ อาคารประเภท ก หมายความว่า อาคารตั้งต่อไปนี้

(๑) อาคารชุดที่มีจำนวนห้องสำหรับใช้เป็นที่อยู่อาศัยรวมกันทุกขั้นของอาคาร หรืออุปกรณ์ของอาคารที่ไม่เกิน ๑๐๐ ห้องนอน

(๒) โรงเรียนที่มีจำนวนห้องสำหรับใช้เป็นห้องพักร่วมกันทุกขั้นของอาคาร หรืออุปกรณ์ของอาคารที่ไม่เกิน ๖๐ ห้อง

(๓) หอพักที่มีจำนวนห้องสำหรับใช้เป็นห้องพักร่วมกันทุกขั้นของอาคาร หรืออุปกรณ์ของอาคารที่ตั้งแต่ ๕๐ ห้อง แล้วไปถึง ๒๕๐ ห้อง

(๔) สถานบริการที่มีเป้าหมายที่ใช้สืบรวมกันทุกขั้นของอาคาร หรืออุปกรณ์ของอาคารที่ตั้งแต่ ๑,๐๐๐ ตารางเมตร แล้วไปถึง ๕,๐๐๐ ตารางเมตร

(๕) ภาคการที่ก้าว่างของงานราชการ รัฐวิสาหกิจ องค์กรระหว่างประเทศ หรือของอุตสาหกรรมที่มีเป้าหมายที่ใช้สืบรวมกันทุกขั้นของอาคารหรืออุปกรณ์ของอาคารที่ตั้งแต่ ๕,๐๐๐ ตารางเมตร แล้วไปถึง ๑๐,๐๐๐ ตารางเมตร

(๖) ผลิตภัณฑ์ที่มีเป้าหมายที่ใช้สืบรวมกันทุกขั้นของอาคารหรืออุปกรณ์ของอาคารที่ตั้งแต่ ๑,๐๐๐ ตารางเมตร แล้วไปถึง ๑,๘๐๐ ตารางเมตร

(๗) ภาคการที่มีเป้าหมายที่ใช้สืบรวมกันทุกขั้นของอาคารหรืออุปกรณ์ของอาคารที่ตั้งแต่ ๒๕๐ ตารางเมตร แล้วไปถึง ๕๐๐ ตารางเมตร

๙๗ ๗ อาคารประเภท ก หมายความว่า ภาครัฐตั้งต่อไปนี้

(๑) หาดทรายที่อิฐขนาดห้องส่วนรำขี้เป็นที่อยู่อาศัยรวมกันทุกชั้นของอาคาร หรือกุ่มของอาคาร ตึ้งตื้อ ๑๐ ห้อง แต่ไม่มีเส้น ๕๐ ห้อง

(๒) คลอกที่มีพื้นที่ใช้สอยรวมกันทุกชั้นของอาคาร หรือกุ่มของอาคารตั้งตื้อ ๔๐๐ ตารางเมตร แต่ไม่ถึง ๑,๐๐๐ ตารางเมตร

(๓) บ้านเดียวหรือร้านอาหารที่มีพื้นที่ให้บริการรวมกันทุกชั้นของอาคาร หรือกุ่มของอาคาร ตึ้งตื้อ ๑๐๐ ตารางเมตร แต่ไม่มีเส้น ๒๕๐ ตารางเมตร

ข้อ ๙ อาคารประปาฯ ๑. หมายความว่า บ้านเดียวหรือร้านอาหารที่มีพื้นที่ให้บริการรวมกันทุกชั้นไม่ถึง ๑๐๐ ตารางเมตร

ข้อ ๑๐ มาตรฐานควบคุมการระบายน้ำที่ใช้จากอาคาร ประจำทศ. ก. ต้องมีค่าดังต่อไปนี้

(๑) ความเป็นกรดและด่าง (pH) ค่ามีค่าระหว่าง ๕-๘

(๒) บีโอลีดี (BOD) ต้องมีค่าไม่เกิน ๒๐ มิลลิกรัมต่อลิตร

(๓) สารเคมี固溶物 (Suspended Solids) ต้องมีค่าไม่เกิน ๓๐ มิลลิกรัมต่อลิตร

(๔) สูบไฟค์ (Sulfide) ต้องมีค่าไม่เกิน ๑.๐ มิลลิกรัมต่อลิตร

(๕) สารที่ละลายได้ทั้งหมด (Total Dissolved Solids) ต้องมีค่าเพิ่มขึ้นจากปัจจุบันตามด้วย ในน้ำใช้ตามปกติไม่เกิน ๔๐ มิลลิกรัมต่อลิตร

(๖) ตะกอนน้ำแข็ง (Settleable Solids) ต้องมีค่าไม่เกิน ๐.๕ มิลลิกรัมต่อลิตร

(๗) น้ำมันและไขมัน (Fat Oil and Grease) ต้องมีค่าไม่เกิน ๒.๐ มิลลิกรัมต่อลิตร

(๘) บีโอลีน (TKN) ต้องมีค่าไม่เกิน ๓๐ มิลลิกรัมต่อลิตร

ข้อ ๑๑ มาตรฐานควบคุมการระบายน้ำที่ใช้จากอาคาร ประจำทศ. ข. ต้องเป็นไปตามข้อ ๘ ดังนี้

(๑) บีโอลีดี ต้องมีค่าไม่เกิน ๓๐ มิลลิกรัมต่อลิตร

(๒) สารเคมี固溶物 ต้องมีค่าไม่เกิน ๔๐ มิลลิกรัมต่อลิตร

ข้อ ๑๒ มาตรฐานควบคุมการระบายน้ำที่ใช้จากอาคาร ประจำทศ. ค. ต้องเป็นไปตามข้อ ๙ ดังนี้

(๑) บีโอลีดี ต้องมีค่าไม่เกิน ๔๐ มิลลิกรัมต่อลิตร

(๒) สารเคมี固溶物 ต้องมีค่าไม่เกิน ๔๐ มิลลิกรัมต่อลิตร

ข้อ ๑๓ มาตรฐานควบคุมการระบายน้ำที่ใช้จากอาคาร ประจำทศ. ด. ต้องมีค่าดังต่อไปนี้

(๑) ชัลไฟต์ คัมมิค่าไม่เกิน ๓.๐ มิลลิกรัมต่อลิตร

(๔) ตีบีโอลีน คัมมิค่าไม่เกิน ๔.๐ มิลลิกรัมต่อลิตร

ข้อ ๑๔ มาตรฐานควบคุมการระบายน้ำที่ใช้จากอาคาร ประจำทศ. จ. ต้องเป็นไปตามข้อ ๘ ดังนี้

(๑) บีโอลีดี ต้องมีค่าไม่เกิน ๕๐ มิลลิกรัมต่อลิตร

(๒) สารเคมี固溶物 ต้องมีค่าไม่เกิน ๕๐ มิลลิกรัมต่อลิตร

(๓) ชัลไฟต์ คัมมิค่าไม่เกิน ๕.๐ มิลลิกรัมต่อลิตร

(๔) ตีบีโอลีน คัมมิค่าไม่เกิน ๕.๐ มิลลิกรัมต่อลิตร

ข้อ ๑๕ มาตรฐานควบคุมการระบายน้ำที่ใช้จากอาคาร ประจำทศ. จ. ต้องมีค่าดังต่อไปนี้

(๑) ความเป็นกรดและด่างของน้ำร่างหวาย ๕-๘

(๒) บีโอลีดี ต้องมีค่าไม่เกิน ๑๐๐ มิลลิกรัมต่อลิตร

(๓) สารเคมี固溶物 ต้องมีค่าไม่เกิน ๑๐๐ มิลลิกรัมต่อลิตร

(๔) น้ำมันแลดูไขมัน ต้องมีค่าไม่เกิน ๑๐๐ มิลลิกรัมต่อลิตร

ข้อ ๑๖ การตรวจสอบมาตรฐานการระบายน้ำที่ใช้จากอาคาร ให้ใช้วิธีการดังต่อไปนี้

(๑) การตรวจสอบค่าสารเคมี固溶物โดยใช้กรอบที่ได้ให้กระทำให้ไว้ก่อนจะใช้ดูดโมลิฟิคชัน (Azide Modification)

ที่อุณหภูมิ ๒๐ องศาเซลเซียส เป็นเวลา ๕ วัน ดิบดักกับเครื่องวัดค่าที่ดูดกระบวนการควบคุมผลพิษที่กาวน้ำที่น้ำของ

(๒) การตรวจสอบค่าสารเคมีด้วย GLASS FIBRE FILTER DISC

(๓) การตรวจสอบค่าไฟฟ้าให้กระทำให้ไว้ก่อนจะใช้ดูดโมลิฟิคชัน (Titration)

(๔) การตรวจสอบค่าสารที่จะละลายให้กับน้ำดักกับเครื่องวัดค่าที่ดูดกระบวนการควบคุมผลพิษ

๑๐๑ องศาเซลเซียส ที่อุณหภูมิ ๑๐๕ องศาเซลเซียส ในเวลา ๑ ชั่วโมง

(๕) การตรวจสอบค่าสารเคมีด้วย IMBOFF CONE ขนาดบรรจุ ๐.๐๐๐ ลูกบาศก์เซนติเมตร ในเวลา ๑ ชั่วโมง

(๑) การตรวจสอบค่าน้ำมันและไขมันให้กระทำให้ไว้ก่อนจะใช้ดูดโมลิฟิคชันกับตัวตัวที่ละลาย แพร่เศษท่าน้ำกับน้ำของน้ำมันและไขมัน

(๒) การตรวจสอบค่าให้ก่อเรื้อนให้ไว้ก่อนจะใช้ดูดโมลิฟิคชัน (Kjeldahl)

ข้อ ๑๕ การคิดค่าน้ำมันที่ต้องซื้อ จำนวนการและจำนวนที่ออกของอาคาร หรือกุ่มของอาคาร ให้เป็นไปตามวิธีการที่คณะกรรมการควบคุมและพิจารณาด โดยประกาศในราชกิจจานุเบกษา

ข้อ ๑๖ วิธีการเก็บตัวอย่างน้ำ ความถี่ และระยะเวลาในการเก็บตัวอย่างน้ำ ให้เก็บในไวนิลที่คณาจารย์คุมการคุ้มครองพื้นที่ทาง ให้ประปาฯในราชกิจจานุเบกษา

ข้อ ๑๗ ประกาศนี้ให้ใช้บังคับตั้งแต่วันถัดจากวันประกาศในราชกิจจานุเบกษาเป็นต้นไป

ประภาฯ ๘ วันที่ ๑ พฤษภาคม พ.ศ. ๒๕๖๘


ลงนาม ดังนี้

รัฐมนตรีว่าการกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม

ภาคผนวก ข

เอกสารการป้องกันและระงับอัคคีภัย





พ.ศ.๒๕๖๔ วันที่

๐๑ กันยายน ๒๕๖๔

บัญชีรายรับ ๘๖๗๙๖
บัญชีรายจ่าย ๘๖๗๙๖

ผู้รับผิดชอบ นางสาว จำเนา พานิช
ผู้รับผิดชอบ นางสาว ณัฐา ใจดี

ตามกฎหมายห้ามตามตัวลงนามในเอกสารนี้โดยเด็ดขาด แต่ถ้ามีการติดต่อทางโทรศัพท์ โทร. ๐๘๑-๒๓๔๕๖๗๘๙ ให้รับรองว่าเป็นการดำเนินการจริงๆ ไม่ใช่การหลอกลวง

ได้ดำเนินการฝึกซ้อมดับเพลิงและฝึกซ้อมพยพหนีไฟ

๕๙๙ ซอยวัดสุวนาราม ถนนเจริญกรุง แขวงบางรัก เขตบางรัก กรุงเทพมหานคร ๑๐๐๐๐

เอกสารนี้

อาคารจ้าวพระยา膨衙อร์

ได้รับใบอนุญาตจากกรมสังคธิการและมาตรฐานคุณภาพ ใบอนุญาตเลขที่ ดพม.-๓ ๒๐๑๗

กรุงเทพมหานคร



บัญชีรายรับ-จ่ายประจำเดือน



ที่ กท ๑๙๐๕/๔๕๗

สำนักป้องกันและบรรเทาสาธารณภัย
๗๗/๑ ถนนพระรามที่ ๖ กทม. ๑๐๔๐๐

๒๖ กันยายน ๒๕๖๖

เรื่อง รายงานผลการฝึกซ้อมดับเพลิงและฝึกซ้อมอพยพหนีไฟ


เรียน กรรมการผู้จัดการ อาคารเจ้าพระยาทาวเวอร์

- สิ่งที่ส่งมาด้วย ๑. รายงานผลการฝึกซ้อมดับเพลิงและฝึกซ้อมอพยพหนีไฟ จำนวน ๑ ฉบับ
๒. วุฒิบัตรการฝึกซ้อมดับเพลิงและฝึกซ้อมอพยพหนีไฟ จำนวน ๑ ฉบับ

ตามที่ อาคารเจ้าพระยาทาวเวอร์ ขอรับการสนับสนุนวิทยากรดำเนินการฝึกอบรมฝึกซ้อมดับเพลิงและฝึกซ้อมอพยพหนีไฟ ให้กับพนักงานของอาคารเจ้าพระยาทาวเวอร์ เมื่อวันที่ ๔ กันยายน ๒๕๖๖ ความละเอียดแจ้งแล้ว นั้น

สำนักป้องกันและบรรเทาสาธารณภัย กรุงเทพมหานคร ในฐานะหน่วยงานฝึกอบรมฯ ของกรุงเทพมหานคร ได้ดำเนินการฝึกซ้อมดับเพลิงและฝึกซ้อมอพยพหนีไฟให้กับพนักงานอาคารเจ้าพระยาทาวเวอร์ เรียบร้อยแล้ว รายละเอียดตามสิ่งที่ส่งมาด้วย

จึงเรียนมาเพื่อโปรดทราบ



ผู้อำนวยการสำนักป้องกันและบรรเทาสาธารณภัย

กองปฏิบัติการดับเพลิงและกู้ภัย ๒

โทร. ๐ ๒๓๕๕ ๖๘๕๖

โทรสาร ๐ ๒๓๕๕ ๖๘๕๘ ต่อ ๓๒๑




ที่ กท ๑๙๐๕/๒๔๖๖

สำนักป้องกันและบรรเทาสาธารณภัย
๗๗/๑ ถนนพระรามที่ ๖ กทม. ๑๐๔๐๐

หนังสือฉบับนี้ให้ไว้เพื่อรับรองว่า อาคารเจ้าพระยาทาวเวอร์ ตั้งอยู่ที่ ๘๙ ซอยสุวนพลุ ถนน
เจริญกรุง แขวงบางรัก เขตบางรัก กรุงเทพมหานคร ได้ดำเนินการฝึกซ้อมดับเพลิงและฝึกซ้อมอพยพหนีไฟ
เมื่อวันที่ ๕ กันยายน ๒๕๖๖ มีผู้เข้ารับการฝึกซ้อม จำนวน ๓๑๕ คน (ตามบัญชีรายชื่อที่แนบ)

ให้ไว้ ณ วันที่ ๒๙ กันยายน พ.ศ. ๒๕๖๖

ผู้อำนวยการสำนักป้องกันและบรรเทาสาธารณภัย



ມະນາຄ ດັບອະນຸຍາກ

ມະນາຄ ດັບອະນຸຍາກ

ໃຫຍ່ ພົມ ດັບອະນຸຍາກ ດັບອະນຸຍາກ ດັບອະນຸຍາກ ດັບອະນຸຍາກ ດັບອະນຸຍາກ ດັບອະນຸຍາກ ດັບອະນຸຍາກ

ເມືອງ ສາວນະເລດ ສາວນະເລດ ສາວນະເລດ ສາວນະເລດ ສາວນະເລດ ສາວນະເລດ ສາວນະເລດ

ມະນາຄ ດັບອະນຸຍາກ

ຕູ້ອາຍ່ານທີ່ ແລະ ເອຍວັດສະນພູ ດານແຮຮູກ ແລະ ແຈວັງບາງຮັກ ເບີນບາງຮັກ ກຽມທ່ານໝານໂຮງ ០៩៣០០

ໂຮງຮຽນ ແລະ ລາ ກຽມທ່ານ

ຊອກປະຊາຊົນ

ໃຊ້ຮັບໃນບ່ອນຫຼຸງເຈົ້າການສະວັດທິການແລະລູ່ນົມຂອງແກຣມ ໂນຍຸ້ນາຜູ້ເລີຍ ດັວຍ - ៥ ໂດຍ

ກຽມທ່ານ



ສັນຕະພາບ ສປປ.ລາວ
ເມືອງ ສາວນະເລດ



ที่ กท ๑๙๐๕/๒๕๖๖

สำนักป้องกันและบรรเทาสาธารณภัย
๗๗/๑ ถนนพระรามที่ ๖ กทม. ๑๐๔๐๐

๒๗ กันยายน ๒๕๖๖

เรื่อง รายงานผลการฝึกซ้อมดับเพลิงและฝึกซ้อมอพยพหนีไฟ


เรียน กรรมการผู้จัดการ โรงเรม แซงกรี-ลา กรุงเทพ

- สิ่งที่ส่งมาด้วย ๑. รายงานผลการฝึกซ้อมดับเพลิงและฝึกซ้อมอพยพหนีไฟ จำนวน ๑ ฉบับ
๒. วุฒิบัตรการฝึกซ้อมดับเพลิงและฝึกซ้อมอพยพหนีไฟ จำนวน ๑ ฉบับ

ตามที่ โรงเรม แซงกรี-ลา กรุงเทพ ขอรับการสนับสนุนวิทยาการดำเนินการฝึกอบรมฝึกซ้อมดับเพลิงและฝึกซ้อมอพยพหนีไฟ ให้กับพนักงานโรงเรม แซงกรี-ลา กรุงเทพ เมื่อวันที่ ๔ กันยายน ๒๕๖๖ ความละเอียดแจ้งแล้ว นั้น

สำนักป้องกันและบรรเทาสาธารณภัย กรุงเทพมหานคร ในฐานะหน่วยงานฝึกอบรมฯ ของ กรุงเทพมหานคร ได้ดำเนินการฝึกซ้อมดับเพลิงและฝึกซ้อมอพยพหนีไฟให้กับพนักงานโรงเรม แซงกรี-ลา กรุงเทพ เรียบร้อยแล้ว รายละเอียดตามสิ่งที่ส่งมาด้วย

จึงเรียนมาเพื่อโปรดทราบ



กองปฏิบัติการดับเพลิงและกู้ภัย ๒

โทร. ๐ ๒๓๕๕ ๖๘๕๖

โทรสาร ๐ ๒๓๕๕ ๖๘๕๘ ต่อ ๓๒๑



ที่ กท ๑๘๐๕/๒๕๖๗

สำนักป้องกันและบรรเทาสาธารณภัย
๓๗/๑ ถนนพระรามที่ ๖ กทม. ๑๐๔๐๐

หนังสือฉบับนี้ให้ไว้เพื่อรับรองว่า โรงเรม แซงกรี-ลา กรุงเทพ ตั้งอยู่ที่ ๘๙ ซอยสุวนพูล
ถนนเจริญกรุง แขวงบางรัก เขตบางรัก กรุงเทพมหานคร ได้ดำเนินการฝึกซ้อมดับเพลิงและฝึกซ้อมอพยพหนีไฟ
เมื่อวันที่ ๕ กันยายน ๒๕๖๖ มีผู้เข้ารับการฝึกซ้อม จำนวน ๓๕๕ คน (ตามบัญชีรายชื่อที่แนบ)

ให้ไว้ ณ วันที่ ๕ กันยายน พ.ศ. ๒๕๖๖



ຮ່ວມທີ່
ຕະຫຼາດ
ກົດ

ຮ່າຍງານພຸດທະນາຄາຊອບປິເສັນ
ຕາມໃບປັບປອດທຽບຈະສອບປະລັບ
ລັດທີ່ ៣០ ຕຸລາມ ແລ້ວ
ເລີ້ນທີ່ ຄະແນລ/ໜັກຕົວ ລັດວັນທີ ៣០ ພຶກສົກພາຍໃນລາວ

ໃບປັບປອດການຕຽບຈະສອບອາຄານ

ໃບປັບປອດຈະນີ້ອອກໃຫ້ເພື່ອແສດງວ່າ

ວາງ ໂຮງຮ່າງພະນັກງານພຸດທະນາກົດສອບປະລັບ ປະເທດ ປະເທດ ປະເທດ ປະເທດ
ລັດທີ່ ៣០ ຕຸລາມ ແລ້ວ ເລີ້ນທີ່ ຄະແນລ/ໜັກຕົວ ລັດວັນທີ ៣០ ພຶກສົກພາຍໃນລາວ
ໄດ້ປັບປອດຈະນີ້ອອກໃຫ້ເພື່ອແສດງວ່າ ອາຄານນີ້ສະບັບກຳນົດກົດຢູ່ໃນການປິເສັນ

ແລ້ວ
ແລ້ວ

ເລີ້ນທີ່ ៣០ ຕຸລາມ ແລ້ວ ເລີ້ນທີ່ ៣០ ດັນ ດັນ ດັນ ດັນ ດັນ ດັນ
ໄດ້ປັບປອດຈະນີ້ອອກໃຫ້ເພື່ອແສດງວ່າ ອາຄານນີ້ສະບັບກຳນົດກົດຢູ່ໃນການປິເສັນ

ວັນ
ວັນ
ວັນ

ໃບປັບປອດຈະນີ້ຫຼັງຈາກຈົດກົດວ່າ

ຄຳເຫຼືອນ

1. ໃປ່ງປ່ອງຈະນີ້ປິເສັນການຕຽບຈະສອບປະລັບທະນາຄານ
ນີ້ຕໍ່ປັບປອດກົດສອບປະລັບທີ່ອາການກອອກສ້າງອາຄານ
ດັດເປັນກາງອາຄານ ທີ່ຮັບເຫຼືອນໝາຍອາຄານແຕ່ຍ່າງໃຈ

2. ໃປ່ງປ່ອງຈະນີ້ປິເສັນການຕຽບຈະສອບປະລັບອາຄານ (ປິເສັນ ຮຸ. ៣) ຈຳນວ
ກອນໃບປັບປອດການຕຽບຈະສອບປະລັບອາຄານ ໂດຍ ປຸນ ៣០ ວັນ
ຮະນະວັດທີ ៣ ປີ

ແບບ ຮ. ៣

ຕາມໃບປັບປອດການຕຽບຈະສອບປິເສັນທີ່ ២០១៩/ສັດວ
ລັດວັນທີ ៣០ ຕຸລາມ ແລ້ວ
ເລີ້ນທີ່ ៣០ ດັນ

ໃບປັບປອດການຕຽບຈະສອບອາຄານ

ໃບປັບປອດຈະນີ້ອອກໃຫ້ເພື່ອແສດງວ່າ

ວາງ ໂຮງຮ່າງພະນັກງານພຸດທະນາກົດສອບປະລັບ ປະເທດ ປະເທດ ປະເທດ
ລັດທີ່ ៣០ ຕຸລາມ ແລ້ວ ເລີ້ນທີ່ ៣០ ດັນ ດັນ ດັນ ດັນ ດັນ
ໄດ້ປັບປອດຈະນີ້ອອກໃຫ້ເພື່ອແສດງວ່າ ອາຄານນີ້ສະບັບກຳນົດກົດຢູ່ໃນການປິເສັນ

ແລ້ວ
ແລ້ວ

ເລີ້ນທີ່ ៣០ ຕຸລາມ ແລ້ວ ເລີ້ນທີ່ ៣០ ດັນ ດັນ ດັນ ດັນ
ໄດ້ປັບປອດຈະນີ້ອອກໃຫ້ເພື່ອແສດງວ່າ ອາຄານນີ້ສະບັບກຳນົດກົດຢູ່ໃນການປິເສັນ


	<p>ຜູ້ອໍານວຍດັບກົດ ຕະຫຼາດ ກົດ</p>
<p>ລັດທີ່ ៣០ ຕຸລາມ ແລ້ວ ເລີ້ນທີ່ ៣០ ດັນ</p>	
<p>ໃບປັບປອດຈະນີ້ປິເສັນ ວາງ ໂຮງຮ່າງພະນັກງານພຸດທະນາກົດ ລັດທີ່ ៣០ ຕຸລາມ ແລ້ວ ເລີ້ນທີ່ ៣០ ດັນ</p>	

ພູ້ອໍານວຍດັບກົດ
ຕະຫຼາດ
ກົດ

ລັດທີ່ ៣០ ຕຸລາມ ແລ້ວ
ເລີ້ນທີ່ ៣០ ດັນ

ເລີຍທີ່ ຕະຫຼາດ / ແກ້ວມ

ຮາຍຂານແຄດກາຣຕຣາຈສອນບໍໃໝ່
ຕາມໃປປະໂອກຕຣາຈສອນປະຈຳປັກ ດຽວທີ່ (ລ່າສຸດ)
ເລີຍທີ່ ຕະຫຼາດ/ໜັດວິວ ຕະວັດທີ່ ເອເມືອນ ແມ່ນິກາຍ ໄກສອນ



ເລີຍທີ່ ຕະຫຼາດ/ໜັດວິວ ຕະວັດທີ່ ເອເມືອນ ແມ່ນິກາຍ ໄກສອນ

ໃບຮັງຮອງກາຣຕຣາຈສອນການອາຄານ

ໃບຮັງຮອງກາຣຕຣາຈສອນບໍໃໝ່ວ່ານີ້

ອາຄານ ໂຮມແຮມແຊງກີຣີ ລາກ ກຣູມເທິພາ (ອາຄານ KRUNG THEP WING) ໃຫຍ້ ບຽນທີ່ ແລະ ສັນຕະ ໄກສອນ ຈຳກັດ (ຂໍ້ທາງໆ) ຕ້ອງຢູ່ເລື່ອງ ຕ້ອງຢູ່ເລື່ອງ ຕ້ອງຢູ່ເລື່ອງ ຕ້ອງຢູ່ເລື່ອງ ຕ້ອງຢູ່ເລື່ອງ ຕ້ອງຢູ່ເລື່ອງ . ວິຊາສົວນຸພາບ
(ເຈົ້າຮັກກັງ ຊະເລີຍ/ຮ.) ກໍາເນີນ ເຈົ້າຮັກກັງ ຊະເລີຍ/ຮ. ພົມ ຊະເລີຍ/ຮ. - ຕໍ່ປະບົບ/ເຫັນຊື່ ນາງຮັກກັງ ອຳເນົາ/ອຳເນົາ ນາງຮັກກັງ ລົງທຶນ ນາງຮັກກັງ ລົງທຶນ ກໍາຮັກກັງ ລົງທຶນ ດີເລີ່ມການເປັນພະນັກງານ
ພະນະການບໍ່ມີຕົກລາຄາຕຸລາ ແລ້ວ ເພື່ອຫຼັງຫວ່າມອນນັບພະນັກງານ ສູ່ລົງທຶນ ຖະແຫຼງຕຣາຈສອນບໍໃໝ່ ປັບປຸງຕຣາຈສອນບໍໃໝ່ ຈຳກັດ ເສັ້ນ

ເລັກທີ່ກ່າຍການທີ່ກ່າຍການ ແລ້ວ ເພື່ອຫຼັງຫວ່າມອນນັບພະນັກງານ ສູ່ລົງທຶນ ຖະແຫຼງຕຣາຈສອນບໍໃໝ່ ຈຳກັດ ເສັ້ນ

ນີ້ມີຄວາມເປັນພົນຕົ້ນຕົ້ນ ເພື່ອຫຼັງຫວ່າມອນນັບພະນັກງານ ສູ່ລົງທຶນ ຖະແຫຼງຕຣາຈສອນບໍໃໝ່ ຈຳກັດ ເສັ້ນ
ກ່າຍການທີ່ກ່າຍການ ແລ້ວ ເພື່ອຫຼັງຫວ່າມອນນັບພະນັກງານ ສູ່ລົງທຶນ ຖະແຫຼງຕຣາຈສອນບໍໃໝ່ ຈຳກັດ ເສັ້ນ

ກໍາເລີຍ

- ໄປປະໂອກຮອງອະນຸບໍ່ເປັນກາຣຮັບຮອງພະຍາການກາຣຕຣາຈສອນການອາຄານ
ນີ້ມີຄວາມເປັນພົນຕົ້ນຕົ້ນ ເພື່ອຫຼັງຫວ່າມອນນັບພະນັກງານກາຣຕຣາຈສອນການອາຄານ
ຕົດປະຕົງອາຄານ ທີ່ຮ່ວມເຄື່ອນຍ່າຍອາຄານແຕ່ອ່າງໃດ
໒. ໃຫ້ຜູ້ສັນຕິພາບ ແລ້ວ ເພື່ອຫຼັງຫວ່າມອນນັບພະນັກງານກາຣຕຣາຈສອນບໍໃໝ່ ຈຳກັດ
ກ່າຍການທີ່ກ່າຍການ ແລ້ວ ເພື່ອຫຼັງຫວ່າມອນນັບພະນັກງານກາຣຕຣາຈສອນບໍໃໝ່ (ແບ່ນ ຮ. ๓) ຈະນີ້
ຮະບອບວິຕະການບັນດາ ๑ ປີ
- ໄປປະໂອກຮອງກາຣຕຣາຈສອນບໍໃໝ່ (ແບ່ນ ຮ. ๓) ຈະນີ້
ຮະບອບວິຕະການບັນດາ ๑ ປີ

BID 995F1D14EF3B

ແບ່ນ ຮ. ๓

ຕາມໃປປະໂອກຕຣາຈສອນບໍໃໝ່ເພື່ອເລີຍທີ່
ລົງວັນທີ ۲ ຕຸລາດັບ ໄກສອນ



ປະກິບປົກຕະຫຼາດກາແນ່ງໝູ້ກ່າຍການກາຣຕຣາຈສອນການອາຄານ
ເຖິງພັນງານທີ່ກ່າຍການ
ເຖິງພັນງານທີ່ກ່າຍການ

ภาคผนวก ค

อุปกรณ์เก็บตัวอย่างและเครื่องมือตรวจวัด





(ก) เครื่องวัดค่าความเป็นกรด-ด่าง (pH Meter)

(ข) อุปกรณ์เก็บตัวอย่างแบบเชือก



(ค) ถุงมือยาง

(ง) ภาชนะบรรจุตัวอย่าง

อุปกรณ์เก็บตัวอย่าง และเครื่องมือตรวจวัด

ภาคผนวก ๙
ใบรายงานผลการตรวจวิเคราะห์



ANALYSIS REPORT

CUSTOMER NAME	: SHANGRI-LA HOTEL		
ADDRESS	: 89 SOI WATSUANPLU CHAROENKRUNG BANG RAK BANG RAK BANGKOK 10500		
CONTACT INFORMATION	: TEL : 0 2236 7777 e-mail : boonjung.kullayut@shangri-la.com		
SAMPLING SOURCE	: WASTEWATER TREATMENT PLANT		
SAMPLE TYPE	: WASTEWATER	RECEIVED DATE	: JULY 26, 2023
SAMPLING DATE	: JULY 26, 2023	ANALYTICAL DATE	: JULY 26-AUGUST 6, 2023
SAMPLING TIME	: 1/	REPORT NO.	: 2023-U064773
SAMPLING METHOD^c	: GRAB, GRAB AND STERILE TECHNIQUE	WORK NO.	: 2023-004346
SAMPLING BY^c		ANALYSIS NO.	: T23AO429-0003, T23AO429-0005
ANALYZED BY			

PARAMETER	UNIT	METHOD OF ANALYSIS	RESULT		REGULATORY STANDARD	DETECTION LIMIT
			1 15:10 HOUR 1/ T23AO429-0003	2 15:00 HOUR 1/ T23AO429-0005		
pH ^c	-	ELECTROMETRIC METHOD AT SITE (SM: PART 4500-H ^a B)	6.8 (31°C)	7.0 (32°C)	5-9	-
TEMPERATURE ^c	°C	THERMOMETER AT SITE (SM: PART 2550 B)	-	32	-	-
BIOCHEMICAL OXYGEN DEMAND ^c	mg/L	MEMBRANE ELECTRODE METHOD (SM: PART 5210 B AND PART 4500-O G)	327	ND	≤ 20	2.0
CHEMICAL OXYGEN DEMAND ^c	mg/L	CLOSED REFLUX, COLOURIMETRIC METHOD (SM: PART 5220 D)	-	ND	-	25.0
SUSPENDED SOLIDS ^c	mg/L	SUSPENDED SOLIDS DRIED AT 103-105 °C (SM: PART 2540 D)	116	ND	≤ 30	5.0
TOTAL DISSOLVED SOLIDS ^b	mg/L	IN-HOUSE METHOD: UAE.TP.WAO.007 (TOTAL DISSOLVED SOLIDS DRIED AT 103-105 °C); SM: PART 2540 C	482	370	500*	25
SETTLEABLE SOLIDS ^c	mL/L	IMHOFF CONE (SM: PART 2540 F)	11.0	< 0.1	≤ 0.5	0.1
RESIDUAL CHLORINE ^c	mg/L Cl ₂	MODIFIED DPD COLOURIMETRIC METHOD (AT SITE)	-	0.2	-	0.1
SULPHIDE ^b	mg/L	IODOMETRIC METHOD (SM: PART 4500-S ^a F)	2.0	< 0.50	≤ 1.0	0.50
TOTAL KJELDAHL NITROGEN ^c	mg/L	IN-HOUSE METHOD: UAE.TP.WAS.001 (KJELDAHL METHOD); SM: PART 4500-Norg C	< LOQ	ND	≤ 35	1.5
FAT, OIL AND GREASE ^c	mg/L	LIQUID-LIQUID, PARTITION-GRAVIMETRIC METHOD (SM: PART 5520 B)	9	ND	≤ 20	3



- PROHIBITED TO PARTIALLY COPY ANALYSIS REPORT PRIOR TO WRITTEN PERMISSION BY THE LABORATORY.
- THIS ANALYSIS REPORT APPROVES ONLY FOR SUBMITTED SAMPLES.



PARAMETER	UNIT	METHOD OF ANALYSIS	RESULT		REGULATORY STANDARD	DETECTION LIMIT
			1 15:10 HOUR 1/ T23AO429-0003	2 15:00 HOUR 1/ T23AO429-0005		
MICROBIOLOGY						
FAECAL COLIFORM BACTERIA ^b	MPN/100 mL	MULTIPLE-TUBE FERMENTATION TECHNIQUE (SM: PART 9221E)	-	22,000	-	1.8
SAMPLE CONDITION						
WATER'S COLOUR/TURBID			YELLOW/TURBID	YELLOW/CLEAR		
SEDIMENT			BROWN	BROWN		

a : ISO/IEC 17025 ACCREDITED BY THAI INDUSTRIAL STANDARDS INSTITUTE (TISI)

b : ISO/IEC 17025 ACCREDITED BY DEPARTMENT OF SCIENCE SERVICE (DSS)

c : VERIFIED BY OWN LABORATORY QUALITY SYSTEM, BUT STILL NOT ACCREDITED

IN-HOUSE : BASED ON STANDARD METHODS FOR THE EXAMINATION OF WATER AND WASTEWATER, APHA, AWWA, WEF, 23rd EDITION, 2017.

SM : STANDARD METHODS FOR THE EXAMINATION OF WATER AND WASTEWATER, APHA, AWWA, WEF, 23rd EDITION, 2017.

RESULT 1 : INFLUENT

RESULT 2 : EFFLUENT

REGULATORY STANDARD : RANGE OR MAXIMUM PERMITTED VALUE FOR BUILDING EFFLUENT STANDARDS CLASS A, NOTIFICATION OF THE MINISTRY OF RESOURCES AND ENVIRONMENT, PUBLISHED IN THE ROYAL GOVERNMENT GAZETTE, VOL 122, PART 125 D, DATED DECEMBER 29, 2005.

500* : PERMITTED EXCESS VALUE TO TOTAL DISSOLVED SOLIDS OF ACTUAL TAP WATER USED.

ND : NON-DETECTABLE.

< LOQ : < LIMIT OF QUANTITATION (TOTAL KJELDAHL NITROGEN \geq 1.5 AND < 5.0 mg/L).

COMMENT : ALL TESTED PARAMETERS OF THE EFFLUENT ARE COMPLIED WITH REGULATORY STANDARD.



* PROHIBITED TO PARTIALLY COPY ANALYSIS REPORT PRIOR TO WRITTEN PERMISSION BY THE LABORATORY.

* THIS ANALYSIS REPORT APPROVES ONLY FOR SUBMITTED SAMPLES.

ANALYSIS REPORT

CUSTOMER NAME : SHANGRI-LA HOTEL PUBLIC COMPANY LIMITED
ADDRESS : 89 SOI WATSUANPLU, BANG RAK BANG RAK BANGKOK 10500
CONTACT INFORMATION : TEL : 0 2236 7777 e-mail : boonjung.kullayut@shangri-la.com
SAMPLING SOURCE : WASTEWATER TREATMENT PLANT
SAMPLE TYPE : WASTEWATER **RECEIVED DATE** : AUGUST 31, 2023
SAMPLING DATE : AUGUST 31, 2023 **ANALYTICAL DATE** : AUGUST 31-SEPTEMBER 7, 2023
SAMPLING TIME : 1/ **REPORT NO.** : 2023-U075459
SAMPLING METHOD^c : GRAB, GRAB AND STERILE TECHNIQUE **WORK NO.** : 2023-004346
SAMPLING BY^c : **ANALYSIS NO.** : T23AR047-0003, T23AR047-0005
ANALYZED BY :

PARAMETER	UNIT	METHOD OF ANALYSIS	RESULT		REGULATORY STANDARD	DETECTION LIMIT
			1 13:35 HOUR 1/ T23AR047-0003	2 13:30 HOUR 1/ T23AR047-0005		
pH ^c	-	ELECTROMETRIC METHOD AT SITE (SM: PART 4500-H ^c B)	6.9 (31°C)	7.1 (31°C)	5-9	-
TEMPERATURE ^c	°C	THERMOMETER AT SITE (SM: PART 2550 B)	-	31	-	-
BIOCHEMICAL OXYGEN DEMAND ^c	mg/L	MEMBRANE ELECTRODE METHOD (SM: PART 5210 B AND PART 4500-O G)	254	ND	≤ 20	2.0
CHEMICAL OXYGEN DEMAND ^c	mg/L	CLOSED REFLUX, COLOURIMETRIC METHOD (SM: PART 5220 D)	-	ND	-	25.0
SUSPENDED SOLIDS ^c	mg/L	SUSPENDED SOLIDS DRIED AT 103-105 °C (SM: PART 2540 D)	80.9	ND	≤ 30	5.0
TOTAL DISSOLVED SOLIDS ^b	mg/L	IN-HOUSE METHOD: UAE.TP.WAO.007 (TOTAL DISSOLVED SOLIDS DRIED AT 103-105 °C); SM: PART 2540 C	483	338	500*	25
SETTLEABLE SOLIDS ^c	mL/L	IMHOFF CONE (SM: PART 2540 F)	2.0	< 0.1	≤ 0.5	0.1
RESIDUAL CHLORINE ^c	mg/L Cl ₂	MODIFIED DPD COLOURIMETRIC METHOD (AT SITE)	-	0.2	-	0.1
SULPHIDE ^b	mg/L	IODOMETRIC METHOD (SM: PART 4500-S ^c F)	0.54	< 0.50	≤ 1.0	0.50
TOTAL KJELDAHL NITROGEN ^c	mg/L	IN-HOUSE METHOD: UAE.TP.WAS.001 (KJELDAHL METHOD); SM: PART 4500-Norg C	10.2	ND	≤ 35	1.5
FAT, OIL AND GREASE ^c	mg/L	LIQUID-LIQUID, PARTITION-GRAVIMETRIC METHOD (SM: PART 5520 B)	15	ND	≤ 20	3



PARAMETER	UNIT	METHOD OF ANALYSIS	RESULT		REGULATORY STANDARD	DETECTION LIMIT
			1 13:35 HOUR 1/ T23AR047-0003	2 13:30 HOUR 1/ T23AR047-0005		
MICROBIOLOGY						
FAECAL COLIFORM BACTERIA b	MPN/100 mL	MULTIPLE-TUBE FERMENTATION TECHNIQUE (SM: PART 9221E)	-	49	-	1.8
SAMPLE CONDITION						
WATER'S COLOUR/TURBID			YELLOW/TURBID	YELLOW/CLEAR		
SEDIMENT			BROWN	BROWN		

a : ISO/IEC 17025 ACCREDITED BY THAI INDUSTRIAL STANDARDS INSTITUTE (TISI)

b : ISO/IEC 17025 ACCREDITED BY DEPARTMENT OF SCIENCE SERVICE (DSS)

c : VERIFIED BY OWN LABORATORY QUALITY SYSTEM, BUT STILL NOT ACCREDITED

IN-HOUSE : BASED ON STANDARD METHODS FOR THE EXAMINATION OF WATER AND WASTEWATER, APHA, AWWA, WEF, 23rd EDITION, 2017.

SM : STANDARD METHODS FOR THE EXAMINATION OF WATER AND WASTEWATER, APHA, AWWA, WEF, 23rd EDITION, 2017.

RESULT 1 : INFLUENT

RESULT 2 : EFFLUENT

REGULATORY STANDARD : RANGE OR MAXIMUM PERMITTED VALUE FOR BUILDING EFFLUENT STANDARDS CLASS A, NOTIFICATION OF THE MINISTRY OF RESOURCES AND ENVIRONMENT, PUBLISHED IN THE ROYAL GOVERNMENT GAZETTE, VOL 122, PART 125 D, DATED DECEMBER 29, 2005.

500* : PERMITTED EXCESS VALUE TO TOTAL DISSOLVED SOLIDS OF ACTUAL TAP WATER USED.

ND : NON-DETECTABLE.

COMMENT : ALL TESTED PARAMETERS OF THE EFFLUENT ARE COMPLIED WITH REGULATORY STANDARD.



- PROHIBITED TO PARTIALLY COPY ANALYSIS REPORT PRIOR TO WRITTEN PERMISSION BY THE LABORATORY.
- THIS ANALYSIS REPORT APPROVES ONLY FOR SUBMITTED SAMPLES.

ANALYSIS REPORT

CUSTOMER NAME : SHANGRI-LA HOTEL PUBLIC COMPANY LIMITED
ADDRESS : 89 SOI WATSUANPLU, BANG RAK BANG RAK BANGKOK 10500
CONTACT INFORMATION : TEL : 0 2236 7777 e-mail : boonjung.kullayut@shangri-la.com
SAMPLING SOURCE : WASTEWATER TREATMENT PLANT
SAMPLE TYPE : WASTEWATER **RECEIVED DATE** : SEPTEMBER 27, 2023
SAMPLING DATE : SEPTEMBER 27, 2023 **ANALYTICAL DATE** : SEPTEMBER 27-OCTOBER 4, 2023
SAMPLING TIME : 1/ **REPORT NO.** : 2023-U085279
SAMPLING METHOD^c : GRAB, GRAB AND STERILE TECHNIQUE **WORK NO.** : 2023-001278
SAMPLING BY^c : **ANALYSIS NO.** : T23AT254-0003, T23AT254-0005
ANALYZED BY :

PARAMETER	UNIT	METHOD OF ANALYSIS	RESULT		REGULATORY STANDARD	DETECTION LIMIT
			1 14:35 HOUR 1/ T23AT254-0003	2 14:30 HOUR 1/ T23AT254-0005		
pH ^c	-	ELECTROMETRIC METHOD AT SITE (SM: PART 4500-H ^a B)	6.8 (30°C)	7.2 (30°C)	5-9	-
TEMPERATURE ^c	°C	THERMOMETER AT SITE (SM: PART 2550 B)	-	30	-	-
BIOCHEMICAL OXYGEN DEMAND ^c	mg/L	AZIDE MODIFICATION METHOD (SM: PART 5210 B AND PART 4500-O C)	319	4.0	≤ 20	2.0
CHEMICAL OXYGEN DEMAND ^a	mg/L	CLOSED REFLUX, COLOURIMETRIC METHOD (SM: PART 5220 D)	-	42.0	-	25.0
SUSPENDED SOLIDS ^a	mg/L	SUSPENDED SOLIDS DRIED AT 103-105 °C (SM: PART 2540 D)	83.0	19.2	≤ 30	5.0
TOTAL DISSOLVED SOLIDS ^b	mg/L	IN-HOUSE METHOD: UAE.TP.WAO.007 (TOTAL DISSOLVED SOLIDS DRIED AT 103-105 °C); SM: PART 2540 C	446	308	500*	25
SETTLEABLE SOLIDS ^c	mL/L	IMHOFF CONE (SM: PART 2540 F)	< 0.1	< 0.1	≤ 0.5	0.1
RESIDUAL CHLORINE ^c	mg/L Cl ₂	MODIFIED DPD COLOURIMETRIC METHOD (AT SITE)	-	0.1	-	0.1
SULPHIDE ^b	mg/L	IODOMETRIC METHOD (SM: PART 4500-S ^a F)	0.65	< 0.50	≤ 1.0	0.50
TOTAL KJELDAHL NITROGEN ^b	mg/L	IN-HOUSE METHOD: UAE.TP.WAS.001 (KJELDAHL METHOD); SM: PART 4500-Norg C	31.4	8.4	≤ 35	1.5
FAT, OIL AND GREASE ^c	mg/L	LIQUID-LIQUID, PARTITION-GRAVIMETRIC METHOD (SM: PART 5520 B)	5	ND	≤ 20	3



PARAMETER	UNIT	METHOD OF ANALYSIS	RESULT		REGULATORY STANDARD	DETECTION LIMIT
			1 14:35 HOUR 1/ T23AT254-0003	2 14:30 HOUR 1/ T23AT254-0005		
MICROBIOLOGY						
FAECAL COLIFORM BACTERIA b	MPN/100 mL	MULTIPLE-TUBE FERMENTATION TECHNIQUE (SM: PART 9221E)	-	>160,000	-	1.8
SAMPLE CONDITION						
WATER'S COLOUR/TURBID SEDIMENT			YELLOW/TURBID BROWN	YELLOW/TURBID BROWN		

a : ISO/IEC 17025 ACCREDITED BY THAI INDUSTRIAL STANDARDS INSTITUTE (TISI)

b : ISO/IEC 17025 ACCREDITED BY DEPARTMENT OF SCIENCE SERVICE (DSS)

c : VERIFIED BY OWN LABORATORY QUALITY SYSTEM, BUT STILL NOT ACCREDITED

IN-HOUSE : BASED ON STANDARD METHODS FOR THE EXAMINATION OF WATER AND WASTEWATER, APHA, AWWA, WEF, 23rd EDITION, 2017.

SM : STANDARD METHODS FOR THE EXAMINATION OF WATER AND WASTEWATER, APHA, AWWA, WEF, 23rd EDITION, 2017.

RESULT 1 : INFLUENT

RESULT 2 : EFFLUENT

REGULATORY STANDARD : RANGE OR MAXIMUM PERMITTED VALUE FOR BUILDING EFFLUENT STANDARDS CLASS A, NOTIFICATION OF THE MINISTRY OF RESOURCES AND ENVIRONMENT, PUBLISHED IN THE ROYAL GOVERNMENT GAZETTE, VOL 122, PART 125 D, DATED DECEMBER 29, 2005.

500* : PERMITTED EXCESS VALUE TO TOTAL DISSOLVED SOLIDS OF ACTUAL TAP WATER USED.

ND : NON-DETECTABLE.

COMMENT : ALL TESTED PARAMETERS OF THE EFFLUENT ARE COMPLIED WITH REGULATORY STANDARD.

ANALYSIS REPORT

CUSTOMER NAME : SHANGRI-LA HOTEL PUBLIC COMPANY LIMITED
ADDRESS : 89 SOI WATSUANPLU, BANG RAK BANG RAK BANGKOK 10500
CONTACT INFORMATION : TEL : 0 2236 7777 e-mail : boonjung.kullayut@shangri-la.com
SAMPLING SOURCE : WASTEWATER TREATMENT PLANT
SAMPLE TYPE : WASTEWATER **RECEIVED DATE** : NOVEMBER 1, 2023
SAMPLING DATE : OCTOBER 31, 2023 **ANALYTICAL DATE** : NOVEMBER 1-9, 2023
SAMPLING TIME : 1/ **REPORT NO.** : 2023-U096512
SAMPLING METHOD^c : GRAB, GRAB AND STERILE TECHNIQUE **WORK NO.** : 2023-001278
SAMPLING BY^c : **ANALYSIS NO.** : T23AV650-0003, T23AV650-0005
ANALYZED BY :

PARAMETER	UNIT	METHOD OF ANALYSIS	RESULT		REGULATORY STANDARD	DETECTION LIMIT
			1 17:15 HOUR 1/ T23AV650-0003	2 17:10 HOUR 1/ T23AV650-0005		
pH ^c	-	ELECTROMETRIC METHOD (AT SITE) SM: PART 4500-H ^a B AND 1060 B	6.9 (30°C)	7.2 (30°C)	5-9	-
TEMPERATURE ^c	°C	THERMOMETER AT SITE (SM: PART 2550 B)	-	30	-	-
BIOCHEMICAL OXYGEN DEMAND ^c	mg/L	AZIDE MODIFICATION METHOD (SM: PART 5210 B AND PART 4500-O C)	205	ND	≤ 20	2.0
CHEMICAL OXYGEN DEMAND ^a	mg/L	CLOSED REFLUX, COLOURIMETRIC METHOD (SM: PART 5220 D)	-	29.0	-	25.0
SUSPENDED SOLIDS ^a	mg/L	SUSPENDED SOLIDS DRIED AT 103-105 °C (SM: PART 2540 D)	75.0	7.7	≤ 30	5.0
TOTAL DISSOLVED SOLIDS ^b	mg/L	IN-HOUSE METHOD: UAE.TP.WAO.007 (TOTAL DISSOLVED SOLIDS DRIED AT 103-105 °C); SM: PART 2540 C	317	270	500*	25
SETTLEABLE SOLIDS ^c	mL/L	IMHOFF CONE (SM: PART 2540 F)	< 0.1	< 0.1	≤ 0.5	0.1
RESIDUAL CHLORINE ^c	mg/L Cl ₂	MODIFIED DPD COLOURIMETRIC METHOD (AT SITE)	-	0.3	-	0.1
SULPHIDE ^b	mg/L	IODOMETRIC METHOD (SM: PART 4500-S ^a F)	1.7	< 0.50	≤ 1.0	0.50
TOTAL KJELDAHL NITROGEN ^b	mg/L	IN-HOUSE METHOD: UAE.TP.WAS.001 (KJELDAHL METHOD); SM: PART 4500-Norg C	17.0	7.0	≤ 35	1.5
FAT, OIL AND GREASE ^c	mg/L	LIQUID-LIQUID, PARTITION-GRAVIMETRIC METHOD (SM: PART 5520 B)	6	ND	≤ 20	3



PARAMETER	UNIT	METHOD OF ANALYSIS	RESULT		REGULATORY STANDARD	DETECTION LIMIT
			1 17:15 HOUR 1/ T23AV650-0003	2 17:10 HOUR 1/ T23AV650-0005		
MICROBIOLOGY						
FAECAL COLIFORM BACTERIA b	MPN/100 mL	MULTIPLE-TUBE FERMENTATION TECHNIQUE (SM: PART 9221E)	-	35,000	-	1.8
SAMPLE CONDITION						
WATER'S COLOUR/TURBID			YELLOW/TURBID	YELLOW/CLEAR		
SEDIMENT			BROWN	BROWN		

a : ISO/IEC 17025 ACCREDITED BY THAI INDUSTRIAL STANDARDS INSTITUTE (TISI)

b : ISO/IEC 17025 ACCREDITED BY DEPARTMENT OF SCIENCE SERVICE (DSS)

c : VERIFIED BY OWN LABORATORY QUALITY SYSTEM, BUT STILL NOT ACCREDITED

IN-HOUSE : BASED ON STANDARD METHODS FOR THE EXAMINATION OF WATER AND WASTEWATER, APHA, AWWA, WEF, 23rd EDITION, 2017.

SM : STANDARD METHODS FOR THE EXAMINATION OF WATER AND WASTEWATER, APHA, AWWA, WEF, 23rd EDITION, 2017.

RESULT 1 : INFLUENT

RESULT 2 : EFFLUENT

REGULATORY STANDARD : RANGE OR MAXIMUM PERMITTED VALUE FOR BUILDING EFFLUENT STANDARDS CLASS A, NOTIFICATION OF THE MINISTRY OF RESOURCES AND ENVIRONMENT, PUBLISHED IN THE ROYAL GOVERNMENT GAZETTE, VOL 122, PART 125 D, DATED DECEMBER 29, 2005.

500* : PERMITTED EXCESS VALUE TO TOTAL DISSOLVED SOLIDS OF ACTUAL TAP WATER USED.

ND : NON-DETECTABLE.

COMMENT : ALL TESTED PARAMETERS OF THE EFFLUENT ARE COMPLIED WITH REGULATORY STANDARD.



* PROHIBITED TO

PERMISSION BY THE LABORATORY.

ANALYSIS REPORT

CUSTOMER NAME	: SHANGRI-LA HOTEL PUBLIC COMPANY LIMITED		
ADDRESS	: 89 SOI WATSUANPLU, BANG RAK BANG RAK BANGKOK 10500		
CONTACT INFORMATION	: TEL : 0 2236 7777 e-mail : boonjung.kullayut@shangri-la.com		
SAMPLING SOURCE	: WASTEWATER TREATMENT PLANT		
SAMPLE TYPE	: WASTEWATER	RECEIVED DATE	: DECEMBER 1, 2023
SAMPLING DATE	: NOVEMBER 30, 2023	ANALYTICAL DATE	: DECEMBER 1-7, 2023
SAMPLING TIME	: 1/	REPORT NO.	: 2023-U107265
SAMPLING METHOD ^c	: GRAB, GRAB AND STERILE TECHNIQUE	WORK NO.	: 2023-001278
SAMPLING BY		ANALYSIS NO.	: T23AX968-0003, T23AX968-0005
ANALYZED BY			

PARAMETER		ANALYSIS	RESULT		REGULATORY STANDARD	DETECTION LIMIT
			1 14:35 HOUR 1/ T23AX968-0003	2 14:30 HOUR 1/ T23AX968-0005		
pH ^c	-	ELECTROMETRIC METHOD (AT SITE) SM: PART 4500-H ^a B AND 1060 B	6.9 (31°C)	6.9 (31°C)	5-9	-
TEMPERATURE ^c	°C	THERMOMETER AT SITE (SM: PART 2550 B)	-	31	-	-
BIOCHEMICAL OXYGEN DEMAND ^c	mg/L	AZIDE MODIFICATION METHOD (SM: PART 5210 B AND PART 4500-O C)	268	ND	≤ 20	2.0
CHEMICAL OXYGEN DEMAND ^c	mg/L	CLOSED REFLUX, COLOURIMETRIC METHOD (SM: PART 5220 D)	-	ND	-	25.0
SUSPENDED SOLIDS ^c	mg/L	SUSPENDED SOLIDS DRIED AT 103-105 °C (SM: PART 2540 D)	89.8	ND	≤ 30	5.0
TOTAL DISSOLVED SOLIDS ^b	mg/L	IN-HOUSE METHOD: UAE.TP.WAO.007 (TOTAL DISSOLVED SOLIDS DRIED AT 103-105 °C); SM: PART 2540 C	393	320	500*	25
SETTLEABLE SOLIDS ^c	mL/L	IMHOFF CONE (SM: PART 2540 F)	2.0	< 0.1	≤ 0.5	0.1
RESIDUAL CHLORINE ^c	mg/L Cl ₂	MODIFIED DPD COLOURIMETRIC METHOD (AT SITE)	-	0.2	-	0.1
SULPHIDE ^b	mg/L	IODOMETRIC METHOD (SM: PART 4500-S ^a F)	1.4	< 0.50	≤ 1.0	0.50
TOTAL KJELDAHL NITROGEN ^b	mg/L	IN-HOUSE METHOD: UAE.TP.WAS.001 (KJELDAHL METHOD); SM: PART 4500-Norg C	25.0	< LOQ	≤ 35	1.5
FAT, OIL AND GREASE ^c	mg/L	LIQUID-LIQUID, PARTITION-GRAVIMETRIC METHOD (SM: PART 5520 B)	17	ND	≤ 20	3



PARAMETER	UNIT	METHOD OF ANALYSIS	RESULT		REGULATORY STANDARD	DETECTION LIMIT
			1 14:35 HOUR 1/ T23AX968-0003	2 14:30 HOUR 1/ T23AX968-0005		
MICROBIOLOGY						
FAECAL COLIFORM BACTERIA b	MPN/100 mL	MULTIPLE-TUBE FERMENTATION TECHNIQUE (SM: PART 9221E)	-	790	-	1.8
SAMPLE CONDITION						
WATER'S COLOUR/TURBID SEDIMENT			YELLOW/TURBID BROWN	YELLOW/CLEAR BROWN		

a : ISO/IEC 17025 ACCREDITED BY THAI INDUSTRIAL STANDARDS INSTITUTE (TISI)

b : ISO/IEC 17025 ACCREDITED BY DEPARTMENT OF SCIENCE SERVICE (DSS)

c : VERIFIED BY OWN LABORATORY QUALITY SYSTEM, BUT STILL NOT ACCREDITED

IN-HOUSE : BASED ON STANDARD METHODS FOR THE EXAMINATION OF WATER AND WASTEWATER, APHA, AWWA, WEF, 23rd EDITION, 2017.

SM : STANDARD METHODS FOR THE EXAMINATION OF WATER AND WASTEWATER, APHA, AWWA, WEF, 23rd EDITION, 2017.

RESULT 1 : INFLUENT

RESULT 2 : EFFLUENT

REGULATORY STANDARD : RANGE OR MAXIMUM PERMITTED VALUE FOR BUILDING EFFLUENT STANDARDS CLASS A, NOTIFICATION OF THE MINISTRY OF RESOURCES AND ENVIRONMENT, PUBLISHED IN THE ROYAL GOVERNMENT GAZETTE, VOL 122, PART 125 D, DATED DECEMBER 29, 2005.

500* : PERMITTED EXCESS VALUE TO TOTAL DISSOLVED SOLIDS OF ACTUAL TAP WATER USED.

ND : NON-DETECTABLE.

< LOQ : < LIMIT OF QUANTITATION (TOTAL KJELDAHL NITROGEN \geq 1.5 AND < 5.0 mg/L).

BIOCHEMICAL OXYGEN DEMAND (T23AX968-0005) : ADDITION OF NITRIFICATION INHIBITION (TCMP) FOLLOW TO THE SM: 5210B,5(e).

COMMENT : ALL TESTED PARAMETERS OF THE EFFLUENT ARE COMPLIED WITH REGULATORY STANDARD.

ANALYSIS REPORT

CUSTOMER NAME : SHANGRI-LA HOTEL PUBLIC COMPANY LIMITED
ADDRESS : 89 SOI WATSUANPLU, BANG RAK BANG RAK BANGKOK 10500
CONTACT INFORMATION : TEL : 0 2236 7777 e-mail : boonjung.kullayut@shangri-la.com
SAMPLING SOURCE : WASTEWATER TREATMENT PLANT
SAMPLE TYPE : WASTEWATER **RECEIVED DATE** : DECEMBER 22, 2023
SAMPLING DATE : DECEMBER 22, 2023 **ANALYTICAL DATE** : DECEMBER 22, 2023 - JANUARY 3, 2024
SAMPLING TIME : 1/
SAMPLING METHOD^c : GRAB, GRAB AND STERILE TECHNIQUE **REPORT NO.** : 2024-U002286
SAMPLING BY^c : **WORK NO.** : 2023-001278
ANALYZED BY : **ANALYSIS NO.** : T23AZ870-0003, T23AZ870-0005

PARAMETER	UNIT	METHOD OF ANALYSIS	RESULT		REGULATORY STANDARD	DETECTION LIMIT
			1 14:05 HOUR 1/ T23AZ870-0003	2 14:00 HOUR 1/ T23AZ870-0005		
pH ^c	-	ELECTROMETRIC METHOD (AT SITE) SM: PART 4500-H ^a B AND 1060 B	7.2 (30°C)	7.0 (30°C)	5-9	-
TEMPERATURE ^c	°C	THERMOMETER AT SITE (SM: PART 2550 B)	-	30	-	-
BIOCHEMICAL OXYGEN DEMAND ^c	mg/L	AZIDE MODIFICATION METHOD (SM: PART 5210 B AND PART 4500-O C)	242	ND	≤ 20	2.0
CHEMICAL OXYGEN DEMAND ^c	mg/L	CLOSED REFLUX, COLOURIMETRIC METHOD (SM: PART 5220 D)	-	ND	-	25.0
SUSPENDED SOLIDS ^c	mg/L	SUSPENDED SOLIDS DRIED AT 103-105 °C (SM: PART 2540 D)	122	ND	≤ 30	5.0
TOTAL DISSOLVED SOLIDS ^b	mg/L	IN-HOUSE METHOD: UAE.TP.WAO.007 (TOTAL DISSOLVED SOLIDS DRIED AT 103-105 °C); SM: PART 2540 C	474	394	500*	25
SETTLEABLE SOLIDS ^c	mL/L	IMHOFF CONE (SM: PART 2540 F)	4.0	< 0.1	≤ 0.5	0.1
RESIDUAL CHLORINE ^c	mg/L Cl ₂	MODIFIED DPD COLOURIMETRIC METHOD (AT SITE)	-	0.2	-	0.1
SULPHIDE ^b	mg/L	IODOMETRIC METHOD (SM: PART 4500-S ²⁻ F)	< 0.50	< 0.50	≤ 1.0	0.50
TOTAL KJELDAHL NITROGEN ^b	mg/L	IN-HOUSE METHOD: UAE.TP.WAS.001 (KJELDAHL METHOD); SM: PART 4500-Norg C	30.4	< LOQ	≤ 35	1.5
FAT, OIL AND GREASE ^c	mg/L	LIQUID-LIQUID, PARTITION-GRAVIMETRIC METHOD (SM: PART 5520 B)	12	ND	≤ 20	3



PARAMETER	UNIT	METHOD OF ANALYSIS	RESULT		REGULATORY STANDARD	DETECTION LIMIT
			1 14:05 HOUR 1/ T23AZ870-0003	2 14:00 HOUR 1/ T23AZ870-0005		
MICROBIOLOGY						
FAECAL COLIFORM BACTERIA b	MPN/100 mL	MULTIPLE-TUBE FERMENTATION TECHNIQUE (SM: PART 9221 E)	-	790	-	1.8
SAMPLE CONDITION						
WATER'S COLOUR/TURBID			YELLOW/TURBID	YELLOW/CLEAR		
SEDIMENT			BROWN	BROWN		

a : ISO/IEC 17025 ACCREDITED BY THAI INDUSTRIAL STANDARDS INSTITUTE (TISI)

b : ISO/IEC 17025 ACCREDITED BY DEPARTMENT OF SCIENCE SERVICE (DSS)

c : VERIFIED BY OWN LABORATORY QUALITY SYSTEM, BUT STILL NOT ACCREDITED

IN-HOUSE : BASED ON STANDARD METHODS FOR THE EXAMINATION OF WATER AND WASTEWATER, APHA, AWWA, WEF, 23rd EDITION, 2017.

SM : STANDARD METHODS FOR THE EXAMINATION OF WATER AND WASTEWATER, APHA, AWWA, WEF, 23rd EDITION, 2017.

RESULT 1 : INFLUENT

RESULT 2 : EFFLUENT

REGULATORY STANDARD : RANGE OR MAXIMUM PERMITTED VALUE FOR BUILDING EFFLUENT STANDARDS CLASS A, NOTIFICATION OF THE MINISTRY OF RESOURCES AND ENVIRONMENT, PUBLISHED IN THE ROYAL GOVERNMENT GAZETTE, VOL 122, PART 125 D, DATED DECEMBER 29, 2005.

500* : PERMITTED EXCESS VALUE TO TOTAL DISSOLVED SOLIDS OF ACTUAL TAP WATER USED.

ND : NON-DETECTABLE.

< LOQ : < LIMIT OF QUANTITATION (TOTAL KJELDAHL NITROGEN \geq 1.5 AND < 5.0 mg/L).

COMMENT : ALL TESTED PARAMETERS OF THE EFFLUENT ARE COMPLIED WITH REGULATORY STANDARD.



EN PERMISSION BY THE LABORATORY.

• PROHIBITED TO

• THIS ANALYSIS REPORT APPROVES ONLY FOR SUBMITTED SAMPLES.

ANALYSIS REPORT

CUSTOMER NAME	: SHANGRI-LA HOTEL		
ADDRESS	: 89 SOI WATSUANPLU CHAROENKRUNG BANG RAK BANG RAK BANGKOK 10500		
CONTACT INFORMATION	: TEL : 0 2236 7777 e-mail : boonjung.kullayut@shangri-la.com		
SAMPLING SOURCE	: AERATION TANK		
SAMPLE TYPE	: SLURRY IN AERATION TANK	RECEIVED DATE	: JULY 26, 2023
SAMPLING DATE	: JULY 26, 2023	ANALYTICAL DATE	: JULY 26-28, 2023
SAMPLING TIME	: 15:05 HOUR	REPORT NO.	: 2023-U065432
SAMPLING METHOD	: GRAB	WORK NO.	: 2023-004346
SAMPLING BY		ANALYSIS NO.	: T23AO429-0004
ANALYZED BY			

PARAMETER	UNIT	METHOD OF ANALYSIS	RESULT	DETECTION LIMIT
			AERATION TANK T23AO429-0004	
pH	-	ELECTROMETRIC METHOD AT SITE (SM: PART 4500-H ^a B)	7.2 (31°C)	-
DISSOLVED OXYGEN	mg/L	MEMBRANE ELECTRODE METHOD AT SITE (SM: PART 4500-O G)	1.0	0.5
MIXED LIQUOR SUSPENDED SOLIDS	mg/L	MIXED LIQUOR SUSPENDED SOLIDS DRIED AT 103-105 °C (SM: PART 2540 D)	218	5.0

SAMPLE CONDITION

WATER'S COLOUR/TURBID

SEDIMENT

BROWN/TURBID

BROWN

SM : STANDARD METHODS FOR THE EXAMINATION OF WATER AND WASTEWATER, APHA, AWWA, WEF, 23rd EDITION, 2017.



ANALYSIS REPORT

CUSTOMER NAME	: SHANGRI-LA HOTEL PUBLIC COMPANY LIMITED		
ADDRESS	: 89 SOI WATSUANPLU, BANG RAK BANG RAK BANGKOK 10500		
CONTACT INFORMATION	: TEL : 0 2236 7777 e-mail : boonjung.kullayut@shangri-la.com		
SAMPLING SOURCE	: AERATION TANK		
SAMPLE TYPE	: SLURRY IN AERATION TANK	RECEIVED DATE	: AUGUST 31, 2023
SAMPLING DATE	: AUGUST 31, 2023	ANALYTICAL DATE	:
SAMPLING TIME	: 13:40 HOUR	REPORT NO.	: 2023-U075460
SAMPLING METHOD	: CRAP	WORK NO.	: 2023-004346
SAMPLING BY			
ANALYZED BY			

PARAMETER	UNIT	METHOD OF ANALYSIS	RESULT	DETECTION LIMIT
			AERATION TANK T23AR047-0004	
pH	-	ELECTROMETRIC METHOD AT SITE (SM: PART 4500-H ^a B)	7.0 (32°C)	-
DISSOLVED OXYGEN	mg/L	MEMBRANE ELECTRODE METHOD AT SITE (SM: PART 4500-O G)	2.5	0.5
MIXED LIQUOR SUSPENDED SOLIDS	mg/L	MIXED LIQUOR SUSPENDED SOLIDS DRIED AT 103-105 °C (SM: PART 2540 D)	409	5.0
SAMPLE CONDITION				
WATER'S COLOUR/TURBID			BROWN/TURBID	
SEDIMENT			BROWN	

SM : STANDARD METHODS FOR THE EXAMINATION OF WATER AND WASTEWATER, APHA, AWWA, WEF, 23rd EDITION, 2017.



ANALYSIS REPORT

CUSTOMER NAME	: SHANGRI-LA HOTEL PUBLIC COMPANY LIMITED		
ADDRESS	: 89 SOI WATSUANPLU, BANG RAK BANG RAK BANGKOK 10500		
CONTACT INFORMATION	: TEL : 0 2236 7777 e-mail : boonjung.kullayut@shangri-la.com		
SAMPLING SOURCE	: AERATION TANK		
SAMPLE TYPE	: SLURRY IN AERATION TANK	RECEIVED DATE	: SEPTEMBER 27, 2023
SAMPLING DATE	: SEPTEMBER 27, 2023	ANALYTICAL DATE	: SEPTEMBER 27-OCTOBER 2, 2023
SAMPLING TIME	: 14:40 HOUR	REPORT NO.	: 2023-U085281
SAMPLING METHOD	: GRAB	WORK NO.	: 2023-001278
SAMPLING BY		ANALYSIS NO.	: T23AT254-0004
ANALYZED BY			

PARAMETER	UNIT	METHOD OF ANALYSIS	RESULT	DETECTION LIMIT
			AERATION TANK T23AT254-0004	
pH	-	ELECTROMETRIC METHOD AT SITE (SM: PART 4500-H ^a B)	6.8 (30°C)	-
DISSOLVED OXYGEN	mg/L	AZIDE MODIFICATION METHOD AT SITE (SM: PART 4500-O C)	2.1	0.5
MIXED LIQUOR SUSPENDED SOLIDS	mg/L	MIXED LIQUOR SUSPENDED SOLIDS DRIED AT 103-105 °C (SM: PART 2540 D)	1,060	5.0
SAMPLE CONDITION			BROWN/TURBID - BROWN	
WATER'S COLOUR/TURBID				
SEDIMENT				

SM : STANDARD METHODS FOR THE EXAMINATION OF WATER AND WASTEWATER, APHA, AWWA, WEF, 23rd EDITION, 2017.



ANALYSIS REPORT

CUSTOMER NAME	: SHANGRI-LA HOTEL PUBLIC COMPANY LIMITED			
ADDRESS	: 89 SOI WATSUANPLU, BANG RAK BANG RAK BANGKOK 10500			
CONTACT INFORMATION	: TEL : 0 2236 7777 e-mail : boonjung.kullayut@shangri-la.com			
SAMPLING SOURCE	: AERATION TANK			
SAMPLE TYPE	: SLURRY IN AERATION TANK	RECEIVED DATE	: NOVEMBER 1, 2023	
SAMPLING DATE	: OCTOBER 31, 2023	ANALYTICAL DATE	: NOVEMBER 1-7, 2023	
SAMPLING TIME	: 17:20 HOUR	REPORT NO.	: 2023-U096677	
SAMPLING METHOD	: GRAB	WORK NO.	: 2023-001278	
SAMPLING BY		ANALYSIS NO.	: T23AV650-0004	
ANALYZED BY				


PARAMETER	UNIT	METHOD OF ANALYSIS	RESULT	DETECTION LIMIT
			AERATION TANK T23AV650-0004	
pH	-	ELECTROMETRIC METHOD (AT SITE) SM: PART 4500 -H' B AND 1060 B	7.0 (30°C)	-
DISSOLVED OXYGEN	mg/L	AZIDE MODIFICATION METHOD AT SITE (SM: PART 4500-O C)	1.8	0.5
MIXED LIQUOR SUSPENDED SOLIDS	mg/L	MIXED LIQUOR SUSPENDED SOLIDS DRIED AT 103-105 °C (SM: PART 2540 D)	955	5.0

SAMPLE CONDITION	
WATER'S COLOUR/TURBID	BROWN/TURBID
SEDIMENT	BROWN

SM : STANDARD METHODS FOR THE EXAMINATION OF WATER AND WASTEWATER, APHA, AWWA, WEF, 23rd EDITION, 2017.



MISSION BY THE LABORATORY.



- PROHIBITED TO PAR
- THIS ANALYSIS REPORT APPROVES ONLY FOR SUBMITTED SAMPLES.




ANALYSIS REPORT

CUSTOMER NAME	: SHANGRI-LA HOTEL PUBLIC COMPANY LIMITED		
ADDRESS	: 89 SOI WATSUANPLU, BANG RAK BANG RAK BANGKOK 10500		
CONTACT INFORMATION	: TEL : 0 2236 7777 e-mail : boonjung.kullayut@shangri-la.com		
SAMPLING SOURCE	: AERATION TANK		
SAMPLE TYPE	: SLURRY IN AERATION TANK	RECEIVED DATE	: DECEMBER 1, 2023
SAMPLING DATE	: NOVEMBER 30, 2023	ANALYTICAL DATE	: DECEMBER 1-2, 2023
SAMPLING TIME	: 14:40 HOUR	REPORT NO.	: 2023-U107266
SAMPLING METHOD	: GRAB	WORK NO.	: 2023-001278
SAMPLING BY			
ANALYZED BY			

PARAMETER	UNIT	METHOD OF ANALYSIS	RESULT	DETECTION LIMIT
			AERATION TANK T23AX968-0004	
pH	-	ELECTROMETRIC METHOD (AT SITE) SM: PART 4500 -H' B AND 1060 B	6.9 (31°C)	-
DISSOLVED OXYGEN	mg/L	AZIDE MODIFICATION METHOD AT SITE (SM: PART 4500-O C)	1.9	0.5
MIXED LIQUOR SUSPENDED SOLIDS	mg/L	MIXED LIQUOR SUSPENDED SOLIDS DRIED AT 103-105 °C (SM: PART 2540 D)	1,338	5.0
SAMPLE CONDITION				
WATER'S COLOUR/TURBID			BROWN/TURBID	
SEDIMENT			BROWN	

SM : STANDARD METHODS FOR THE EXAMINATION OF WATER AND WASTEWATER, APHA, AWWA, WEF, 23rd EDITION, 2017.



- PROHIBITED TO PARTIALLY COPY ANALYSIS REPORT PRIOR TO WRITTEN PERMISSION BY THE LABORATORY.
- THIS ANALYSIS REPORT APPROVES ONLY FOR SUBMITTED SAMPLES.



ANALYSIS REPORT

CUSTOMER NAME	: SHANGRI-LA HOTEL PUBLIC COMPANY LIMITED			
ADDRESS	: 89 SOI WATSUANPLU, BANG RAK BANG RAK BANGKOK 10500			
CONTACT INFORMATION	: TEL : 0 2236 7777 e-mail : boonjung.kullayut@shangri-la.com			
SAMPLING SOURCE	: AERATION TANK			
SAMPLE TYPE	: SLURRY IN AERATION TANK	RECEIVED DATE	: DECEMBER 22, 2023	
SAMPLING DATE	: DECEMBER 22, 2023	ANALYTICAL DATE	: DECEMBER 22, 2023 - JANUARY 3, 2024	
SAMPLING TIME	: 14:10 HOUR	REPORT NO.	: 2024-U002288	
SAMPLING METHOD	: GRAB	WORK NO.	: 2023-001278	
SAMPLING BY		ANALYSIS NO.	: T23AZ870-0004	
ANALYZED BY				

PARAMETER	UNIT	METHOD OF ANALYSIS	RESULT	DETECTION LIMIT
			AERATION TANK T23AZ870-0004	
pH	-	ELECTROMETRIC METHOD (AT SITE) SM: PART 4500 -H ⁺ B AND 1060 B	7.0 (31°C)	-
DISSOLVED OXYGEN	mg/L	AZIDE MODIFICATION METHOD AT SITE (SM: PART 4500-O C)	2.0	0.5
MIXED LIQUOR SUSPENDED SOLIDS	mg/L	MIXED LIQUOR SUSPENDED SOLIDS DRIED AT 103-105 °C (SM: PART 2540 D)	1,275	5.0
SAMPLE CONDITION				
WATER'S COLOUR/TURBID			BROWN/TURBID	
SEDIMENT			BROWN	

SM : STANDARD METHODS FOR THE EXAMINATION OF WATER AND WASTEWATER, APHA, AWWA, WEF, 23rd EDITION, 2017.



ภาคผนวก จ
เอกสารสอบเทียบเครื่องมือ



ภาคผนวก จ-1
เอกสารสอบเทียบเครื่องมือวิเคราะห์



รายงานใบรับรองสอบเทียบ หวานสอน เครื่องมือแล็บประกันสำหรับปฏิบัติการวิเคราะห์ สำหรับคุณภาพน้ำ

No.	Instrument/Equipment	Parameter	Manufacturer	Model/Serial No.	Calibrator	Certification No.	Date of Calibration	Due date of Calibration*	Remark
ตารางที่แสดงรายการเครื่องมือที่ใช้ในการวิเคราะห์ค่าคุณภาพน้ำ									
1	pH Meter	อุณหภูมิ (Temperature) ค่าเป็นกรดและด่าง (pH)	Mettler-Toledo	Seven Easy TM S20 pH 1231155210	National Food Institute, Ministry of Industry, Thailand	2301846-001-01	17 Feb 23	17 Feb 24	-
2	pH Meter		Mettler-Toledo	Seven Easy pH 1230525212	National Food Institute, Ministry of Industry, Thailand	2302181-001-01	14 Mar 23	13 Mar 24	-
3	UV-VIS Spectrophotometer	ซัลฟิด (Sulphide), ซูโคตี (COD)	Hitachi	U-1900 / 2021-064	DQE Services Co.,Ltd.	SP23-007	3 Jan 23	2 Jan 24	-
4	UV-VIS Spectrophotometer		Hitachi	U-2900 / 21E22-009	DQE Services Co.,Ltd.	SP22-007	6 Jan 23	6 Jan 24	-
5	Analytical Balance (Repeatability 0.01 mg)	สารแขวนลอย (Total Suspended Solids) สารทึบแสงได้ร้อยละ (TDS)	Mettler-Toledo	XSR205DU / C009071872	Technology Promotion Association (Thailand-Japan)	23TM211	26 Apr 23	24 Apr 24	-
6	Hot Air Oven		Memmert	XSR205DU / C009071872	Technology Promotion Association (Thailand-Japan)	23TM112	26 Apr 23	24 Apr 24	-
7	Analytical Balance (Repeatability 0.1 mg)	ไขมันและไขมัน (Fat, Oil & Grease)	Mettler-Toledo	AB-204S/FACT / 1129361010	National Food Institute, Ministry of Industry, Thailand	2203120-001-01	1 Jun 22	31 May 23	-
8	BOD Incubator	บีโอดี (BOD)	Aico	UR-1320 / (UAE)W0.006/2553)	Technology Promotion Association (Thailand-Japan)	23TM372	11 Apr 23	9 Apr 24	-
9	COD Reactor (Heating Block)	ซูโคตี (COD)	Hanna	H1839800-02 / H0185001	Hanna Instruments (Thailand) Ltd. (Thailand-Japan)	H1-2312-0342	10 Mar 23	9 Mar 24	-
10	Digestor Unit	ซีโนเจน (Total Kjeldahl Nitrogen)	FOSS TECATOR	2520auto / 91794469	National Food Institute, Ministry of Industry, Thailand	2302413-001-01	28 Mar 23	27 Mar 24	-
11	Digestor Unit		Velp	DKL20 / 213517	National Food Institute, Ministry of Industry, Thailand	2304455-001-01	28 Aug 23	26 Aug 24	-
12	Distillation Unit (Kjeldahl Method)		FOSS TECATOR	KT8100 / 91889052	FOSS South East Asia	8411	29 May 23	27 May 24	-
13	Incubator (Cooled Incubator)	แบคทีเรียคั่งพิเศษเชิงเคมี (Fecal Coliform Bacteria)	Memmert	IPP 260 / V616.0066	Technology Promotion Association (Thailand-Japan)	23TM728	27 Apr 23	25 Apr 24	-

บริษัท ทีบีเอ็ม จำกัด ผู้ผลิต ผู้นำเข้า ผู้จัดจำหน่าย อุปกรณ์วิทยาศาสตร์ คณิตศาสตร์ ชีววิทยา

ผลิตภัณฑ์ภาคอุตสาหกรรม ISO/IEC 17025:2017 by TS-L 17025:2017 by DSS
ไฟร์บาร์โค้ด GS 9001:2015 และ ISO 14001:2015 มาตรฐานสากลที่ยอมรับ

รายการใบรับรองสอบเทียบ หวานสอน เครื่องมือลักษณะจำพวกวิเคราะห์ เช่น บีบีพาร์คิวเรตเตอร์ สำหรับคุณภาพน้ำดื่ม

No.	Instrument/Equipment	Parameter	Manufacturer	Model/Serial No.	Calibrator	Certification No.	Date of Calibration	Due date of Calibration*	Remark
เครื่องมือทดสอบคุณภาพน้ำดื่ม บีบีพาร์คิวเรตเตอร์ สำหรับคุณภาพน้ำดื่ม									
14	Incubator (Cooled Incubator)		Memmert	IPP 260 / V618.0033	Technology Promotion Association (Thailand-Japan)	23TM729	11 Mar 23	9 Mar 24	-
15	Incubator		Memmert	IF 75 / D317.0305	Technology Promotion Association (Thailand-Japan)	23TM727	27 Apr 23	26 Apr 24	-
16	Incubator		Memmert	IN 75 / D317.0307	Technology Promotion Association (Thailand-Japan)	23TM765	27 Apr 23	26 Apr 24	-
17	Water Bath		Memmert	VNE 14 / L416.0606	Technology Promotion Association (Thailand-Japan)	23TM193	15 Feb 23	15 Feb 24	-
18	Water Bath		Memmert	VNE 14 / L416.0612	Technology Promotion Association (Thailand-Japan)	23TM194	15 Feb 23	15 Feb 24	-
19	Water Bath	แยกเชื้อแบคทีเรียคอลิฟอร์ม (Fecal Coliform Bacteria)	Memmert	VNE 14 / L414.1407	Technology Promotion Association (Thailand-Japan)	23TM374	11 Apr 23	10 Apr 24	-
20	Water Bath		Memmert	VNE 14 / L414.1410	Technology Promotion Association (Thailand-Japan)	23TM377	11 Apr 23	10 Apr 24	-
21	Water Bath		Memmert	VNE 14 / L416.0614	Technology Promotion Association (Thailand-Japan)	23TM250	16 Feb 23	16 Feb 24	-
22	Analytical Balance		Mettler-Toledo	MS603S / B0070110311	Technology Promotion Association (Thailand-Japan)	23MM150	7 Apr 23	5 Apr 24	-
23	Autoclave		ALP	CL-40L / 807298	Technology Promotion Association (Thailand-Japan)	2304203-001-01	10 Aug 23	8 Aug 24	-
24	Autoclave		ALP	CL-40L / 808763	Technology Promotion Association (Thailand-Japan)	23TM763	27 Apr 23	25 Apr 24	-
25	Refrigerator	เก็บรักษาตัวอย่างสำหรับวิเคราะห์ทางคุณภาพ	Sanyo	SBC-337KD(GYN) / 71100607	Technology Promotion Association (Thailand-Japan)	22TM562	7 Apr 22	7 Apr 23	-
26	Refrigerator	เก็บรักษาตัวอย่างสำหรับวิเคราะห์ทางคุณภาพ	Sanyo	SBC-337KD(GYN) / 71100607	Technology Promotion Association (Thailand-Japan)	23TM376	11 Apr 23	10 Apr 24	-

บริษัท ไทยแลนด์ เอนจิニアริ่ง จำกัด ผู้รับผิดชอบที่ปรึกษาด้านคุณภาพ

หลักฐานการทดสอบมาตรฐาน ISO/IEC 17025:2017 by TSL 17025:2017 by DSS
ไฟล์มาตรฐาน ISO 9001:2015 และ ISO 14001:2015 จัดทำโดยบ้านคุณภาพไทย

รายงานฉบับนี้เป็นฉบับที่ชี้แจง ทราบเสนอ เตรียมอีกประจําปีของปฏิบัติการวิเคราะห์ สำหรับคุณภาพน้ำ

No.	Instrument/Equipment	Parameter	Manufacturer	Model/Serial No.	Calibrator	Certification No.	Date of Calibration	Due date of Calibration*	Remark
ตรวจสอบผลการดำเนินโครงการวิเคราะห์สำหรับคุณภาพน้ำ									

Due Date of Calibration* : Based on the annual calibration plan. At least 1 time per year.



Calibration Certificate

Certificate No.: 2301846-001-01
Client name: UNITED ANALYST AND ENGINEERING CONSULTANT CO.,LTD.
Address: 3 Soi Udomsuk 41, Sukhumvit Road, Bangkok, Prakanong, Bangkok 10260

Page 1 of 5

Equipment: pH Meter

Manufacturer: Mettler Toledo

Model: SevenEasy TM S20 pH

Serial No.: 1231155210

ID No.: UAE.WAT.010/2553

Order No.: 2301846

Operation No.: 2301846-001

Date of Receipt: 17 February 2023

Date of Calibration: 24 February 2023

Calibrated by Mr.Worapob Sooktong
Scientist

Approved by N. Nuttapol
(Mr.Nuttapol Nyomchart)
Specialist, Division of Calibration Laboratory

Date of Issue: 24 February 2023 **Responsible for the Technical Management Team**

The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full, except with the prior written approval of the National Food Institute.

F-CS-009 Revision: 01 Date: 20-04-65

2008 ถนนสุขุมวิท 36 แขวงคลองเตย เขตคลองเตย กรุงเทพมหานคร
กุศลบริการและศูนย์บริการมาตรฐานภาคกลางและภาคใต้
2008 Soi 36, Aun Amarn Road, Bang Yai Khan Subdistrict, Bang Phlat District, Bangkok 10700, Thailand
Tel: +66(0) 2422 8688, Fax: +66(0) 2422 8545
[nfitdh.org](http://nfitdh.nfitdh.org)



Calibration Report

Certificate No.: 2301846-001-01
Equipment: pH Meter **Resolution:** 0.01 pH : 1 mV
Manufacturer: Mettler Toledo **Model:** SevenEasy TM S20 pH
Serial No.: 1231155210 **Type:** Bench top
ID No.: UAE.WAT.010/2553

Date of Calibration: 24 February 2023 **Page 3 of 5**

Calibration Results:

1. Calibration of pH Meter (Manual Temperature Compensation at 25 °C)

Nominal pH	DC Voltage Standard (mV)	Average Indicator Reading	Uncertainty (±mV)	Coverage Factor (k)
0	414.120	414	0.00	2.00
2	295.814	298	2.00	2.00
4	177.484	178	4.00	2.00
6	58.160	59	6.00	2.00
7	0.000	0	7.00	2.00
8	-59.158	-59	8.00	2.00
10	-177.480	-177	10.00	2.00
12	-295.811	-298	12.00	2.00
14	-414.117	-414	14.00	2.00

2. Calibration of pH Meter with Electrode (Manual Temperature Compensation at 25 °C)

Equipment: pH Electrode **Type:** Combined Electrode
Manufacturer: Mettler Toledo **Model:** InLab Solids
Serial No.: 9018311 **ID No.:** N/A

Performance of Electrode system (Three-Point Calibration at pH 4, pH 7 and pH 10)

Certified Value @25 °C (pH)	Average Indicator Reading	Relative Slope (%)	Uncertainty (± pH)	Coverage Factor (k)
4.008	4.01	180	-	2.00
6.865	6.80	19	97.68	2.00
10.008	10.01	-160	97.29	2.00
6.985	6.99	15	-	2.00

F-CS-012 Revision: 01 Date: 20-04-65

2008 ถนนสุขุมวิท 36 แขวงคลองเตย เขตคลองเตย กรุงเทพมหานคร
กุศลบริการและศูนย์บริการมาตรฐานภาคกลางและภาคใต้
2008 Soi 36, Aun Amarn Road, Bang Yai Khan Subdistrict, Bang Phlat District, Bangkok 10700, Thailand
Tel: +66(0) 2422 8688, Fax: +66(0) 2422 8545
[nfitdh.org](http://nfitdh.nfitdh.org)



Calibration Report

Certificate No.: 2301846-001-01
Equipment: pH Meter **Resolution:** 0.01 pH : 1 mV
Manufacturer: Mettler Toledo **Model:** SevenEasy TM S20 pH
Serial No.: 1231155210 **Type:** Bench top
ID No.: UAE.WAT.010/2553

Date of Calibration: 24 February 2023

Page 2 of 5

Equipment: pH Meter
Manufacturer: Mettler Toledo
Model: SevenEasy TM S20 pH
Serial No.: 1231155210
ID No.: UAE.WAT.010/2553
Order No.: 2301846
Operation No.: 2301846-001
Date of Receipt: 17 February 2023
Date of Calibration: 24 February 2023

Calibrated by Mr.Worapob Sooktong
Scientist
Approved by N. Nuttapol
(Mr.Nuttapol Nyomchart)
Specialist, Division of Calibration Laboratory

Date of Issue: 24 February 2023 **Responsible for the Technical Management Team**

The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full, except with the prior written approval of the National Food Institute.

F-CS-009 Revision: 01 Date: 20-04-65

2008 ถนนสุขุมวิท 36 แขวงคลองเตย เขตคลองเตย กรุงเทพมหานคร
กุศลบริการและศูนย์บริการมาตรฐานภาคกลางและภาคใต้
2008 Soi 36, Aun Amarn Road, Bang Yai Khan Subdistrict, Bang Phlat District, Bangkok 10700, Thailand
Tel: +66(0) 2422 8688, Fax: +66(0) 2422 8545
[nfitdh.org](http://nfitdh.nfitdh.org)



นิติบัญญัติแห่งประเทศไทย
สถาบันพัฒนาอุตสาหกรรมอาหาร
ศูนย์บริการมาตรฐานภาคกลางและภาคใต้
Foundation for Industrial Development National Food Institute
Food Industrial Laboratory Service Center

Calibration Report

Certificate No.: 2301846-001-01
Equipment: Digital Thermometer with RTD
Resolution: 0.1 °C **Model:** SevenEasy TM S20 pH
Serial No.: 1231155210 **ID No.:** UAE.WAT.010/2553
Manufacturer: Mettler Toledo

Date of Calibration: 24 February 2023

Page 4 of 5

Calibration Results:

1. Calibration of pH Meter (Manual Temperature Compensation at 25 °C)

Nominal pH	DC Voltage Standard (mV)	Average Indicator Reading	Uncertainty (±mV)	Coverage Factor (k)
0	414.120	414	0.00	2.00
2	295.814	298	2.00	2.00
4	177.484	178	4.00	2.00
6	58.160	59	6.00	2.00
7	0.000	0	7.00	2.00
8	-59.158	-59	8.00	2.00
10	-177.480	-177	10.00	2.00
12	-295.811	-298	12.00	2.00
14	-414.117	-414	14.00	2.00

2. Calibration of pH Meter with Electrode (Manual Temperature Compensation at 25 °C)

Equipment: pH Electrode **Type:** Combined Electrode
Manufacturer: Mettler Toledo **Model:** InLab Solids
Serial No.: 9018311 **ID No.:** N/A

Performance of Electrode system (Three-Point Calibration at pH 4, pH 7 and pH 10)

Certified Value @25 °C (pH)	Average Indicator Reading	Relative Slope (%)	Uncertainty (± pH)	Coverage Factor (k)
4.008	4.01	180	-	2.00
6.865	6.80	19	97.68	2.00
10.008	10.01	-160	97.29	2.00
6.985	6.99	15	-	2.00

F-CS-012 Revision: 01 Date: 20-04-65

2008 ถนนสุขุมวิท 36 แขวงคลองเตย เขตคลองเตย กรุงเทพมหานคร
กุศลบริการและศูนย์บริการมาตรฐานภาคกลางและภาคใต้
2008 Soi 36, Aun Amarn Road, Bang Yai Khan Subdistrict, Bang Phlat District, Bangkok 10700, Thailand
Tel: +66(0) 2422 8688, Fax: +66(0) 2422 8545
[nfitdh.org](http://nfitdh.nfitdh.org)

F-CS-012 Revision: 01 Date: 20-04-65

Calibration Report

Certificate No.: 2301846-001-01
Equipment: Digital Thermometer with RTD
 Resolution: 0.1 °C Model: SevenEasy TM S20 pH
 Serial No.: 1231155210 ID No.: UAE.WAT.010/2553
 Manufacturer: Mettler Toledo

Date of Calibration: 24 February 2023

Page 5 of 5

Calibration point: 15.0, 25.0 and 35.0 °C

Calibration result:

- The probe was immersed in liquid bath or dry bath to a minimum depth of 120 mm.
- Description of probe, model : - S/N : -
- Dimension of probe : Diameter 9 mm, Length 120 mm,
- Sheath material : Stainless Steel

UUC* Reading (°C)	Standard Temperature (°C)	Correction Value (°C)	Uncertainty ± (°C)
15.1	15.015	+ 0.1	0.11
25.0	25.014	0.0	0.11
35.1	35.016	+ 0.1	0.11

Note:

- UUC* : Unit Under Calibration

The report uncertainty of measurement was based on standard uncertainty multiplied by coverage factor k= 2, providing a level of confidence of approximately 95 %.

----- End -----

N. niyomchart

F-CS-012 Revision: 01 Date: 20-04-65

2005 Soi 36, Arun Amarn Road, Bang Yai Khan Subdistrict, Bang Phlat District, Bangkok 10700, Thailand.
Tel: +66(0) 2422 8668, Fax: +66(0) 2422 8545

nfi.or.th

Calibration Report

Certificate No.: 2302181-001-01
Equipment: pH Meter Resolution: 0.01 pH : 1 mV
 Manufacturer: METTLER TOLEDO Model: SevenEasy pH
 Serial No.: 1236525212 Type: Bench top
 ID No.: UAE.WAS.003/2553

Date of Calibration: 24 March 2023

Page 2 of 5

Location: Chemical Calibration Laboratory, National Food Institute

Environment Condition: Ambient Temperature: (23.4 ± 1.5) °C Relative Humidity: (52 ± 3) %

Condition of Equipment: Good Condition

Condition of This Results of Calibration:

1. Calibration Method In house method : W-CC-002 based on direct measurement by using standard voltage calibrator and certified reference material (CRM)

2. Reference Standards / Certified Reference Material

Instruments	Serial / ID No.	Manufacturer	Certificate No.	Due Date
2.1 DC Voltage Calibrator	2709007	Fluke	22E1959	17 June 2023
2.2 Digital Thermometer	2709007	Fluke	CC-650567-01	30 October 2023
2.3 Thermo-Hygro Meter	NFLBTH003/17	PONPE	TE 6050555-01	21 September 2023
Certified Reference Material	Lot. No.	Manufacturer	Ref N	Expire Date
2.4 pH buffer 4.008 (Primary pH buffer Solution)	873608	CPAchem	PH216L5	16 February 2025
2.5 pH buffer 6.865 (Primary pH buffer Solution)	873609	CPAchem	PH217L5	16 February 2025
2.6 pH buffer 10.01 (Primary pH buffer Solution)	873611	CPAchem	PH220L5	16 February 2024
2.7 pH buffer 7.00 (Standard pH buffer Solution)	873612	CPAchem	PH107.L5	16 February 2024

3. This certification is traceable to The International System of Unit (SI Unit).

3.1 Instruments No.2.1	through	NSC-TIS-TIS 17025 Laboratory Accreditation of Calibration No.0008
3.2 Instruments No.2.2	through	NSC-TIS-TIS 17025 Laboratory Accreditation of Calibration No.0061
3.3 Instruments No.2.3	through	NSC-TIS-TIS 17025 Laboratory Accreditation of Calibration No.0061
3.4 Certified Reference Material No. 2.4 to 2.6	traceable to	Primary measurement method -衡温电极使用校准的温度计、气压计和毫伏表。The Standard Solution preparation and certified by CPAchem Ltd is accredited to ISO 17034 and ISO/IEC 17025
3.5 Certified Reference Material No.2.7	traceable to	BIM ReNh Hi-13 LoN 25.05.2022; BIM ReNh Hi-16 LoN 02.06.2022; BIM ReNh Hi-13 LoN 25.05.2022; BIM ReNh Hi-16 LoN 02.06.2022, the Standard Solution preparation and certified by CPAchem Ltd is accredited to ISO 17034 and ISO/IEC 17025

4. This certificate was certified only for the instrument we calibrated.

5. This result of calibration was found accurate as shown on date and place of calibration only.

N. niyomchart

F-CS-012 Revision: 01 Date: 20-04-65

2005 Soi 36, Arun Amarn Road, Bang Yai Khan Subdistrict, Bang Phlat District, Bangkok 10700, Thailand.
Tel: +66(0) 2422 8668, Fax: +66(0) 2422 8545

nfi.or.th

Calibration Certificate

Certificate No.: 2302181-001-01

Client name: UNITED ANALYST AND ENGINEERING CONSULTANT CO.,LTD.

Address: 3 Soi Udomsuk 41, Sukhumvit Road,

Bangchak, Prakanong, Bangkok 10260

Page 1 of 5

Equipment: pH Meter

Manufacturer: METTLER TOLEDO

Model: SevenEasy pH

Serial No.: 1230525212

ID No.: UAE.WAS.003/2553

Order No.: 2302181

Operation No.: 2302181-001

Date of Receipt: 14 March 2023

Date of Calibration: 24 March 2023

Calibrated by: Mr.Pheraphat Tuanjit

Scientist

Approved by: *N. niyomchart*

(Mr.Nuttapol Niyomchart)

Specialist, Division of Calibration Laboratory

Date of Issue: 24 March 2023

Responsible for the Technical Management Team

The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its inability to recognize national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the National Food Institute.

F-CS-009 Revision: 01 Date: 20-04-65

2005 Soi 36, Arun Amarn Road, Bang Yai Khan Subdistrict, Bang Phlat District, Bangkok 10700, Thailand.

lac-MRA
NSC-TIS-TIS 17025
CALIBRATION 0061

nfi.or.th

Calibration Report

Certificate No.: 2302181-001-01

Equipment: pH Meter Resolution: 0.01 pH : 1 mV

Manufacturer: METTLER TOLEDO Model: SevenEasy pH

Serial No.: 1230525212 Type: Bench top

ID No.: UAE.WAS.003/2553

Date of Calibration: 24 March 2023

Page 3 of 5

Calibration Results:

1. Calibration of pH Meter (Manual Temperature Compensation at 25 °C)

2. Calibration of pH Meter with Electrode (Manual Temperature Compensation at 25 °C)

Equipment: pH Electrode Type: Combined Electrode

Manufacturer: METTLER TOLEDO Model: InLab Solids

Serial No.: 1156883 **ID.No.:** N/A

Performance of Electrode system (Three-Point Calibration at pH 4, pH 7 and pH 10)

Certified Value @25 °C (pH)	Average Indicator Reading		Relative Slope (%)	Uncertainty (± pH)	Coverage Factor (k)
	pH	mV			
4.008	4.01	187	-	0.0071	2.00
6.865	6.86	22	97.86	0.0075	2.00
10.010	10.01	-160	97.86	0.0086	2.00
6.865	6.99	14	-	0.0093	2.00

2. Calibration of pH Meter with Electrode (Manual Temperature Compensation at 25 °C)

Equipment: pH Electrode Type: Combined Electrode

Manufacturer: METTLER TOLEDO Model: InLab Solids

Serial No.: 1156883 **ID.No.:** N/A

Performance of Electrode system (Three-Point Calibration at pH 4, pH 7 and pH 10)

Certified Value @25 °C (pH)	Average Indicator Reading		Relative Slope (%)	Uncertainty (± pH)	Coverage Factor (k)
	pH	mV			
4.008	4.01	187	-	0.0071	2.00
6.865	6.86	22	97.86	0.0075	2.00
10.010	10.01	-160	97.86	0.0086	2.00
6.865	6.99	14	-	0.0093	2.00

F-CS-012 Revision: 01 Date: 20-04-65

N. niyomchart

nfi.or.th

Calibration Report

Certificate No.: 2302181-001-01
Equipment: Digital Thermometer with RTD (pH Meter)
 Resolution: 0.1 °C Model: SevenEasy pH
 Serial No.: 1230525212 ID No.: UAE.WAS.003/2553
 Manufacturer: METTLER TOLEDO

Date of Calibration: 24 March 2023

Page 4 of 5

Location: Chemical Calibration Laboratory, National Food Institute
Environment Condition: Ambient Temperature 25 °C ± 1 °C
 Relative Humidity 55 % ± 5 %

Condition of this results of Calibration:

1. Calibration Method : - In house method: W-TE-025 by comparison with standard thermometer.
- The Calibration is determined by comparing with a known temperature from a standard resistance thermometer.
- The temperature scale in use at this laboratory is the International Temperature scale of 1990 (ITS-90).

2. Reference Standard Instrument :

Instrument	Model	Serial No.	Certificate No.	Due Date	Through
HANDHELD THERMOMETER	1521	A85997	TE 660039-01	10-Dec-23	NATIONAL FOOD INSTITUTE
Platinum Resistance Thermometer (PRT)	385	509201			

Support Equipment : - Low Temperature Bath (ISOCAL-6), Model: Europa-6 Plus Basic, S/N: 341592/2

3. This certificate is traceable to International System of Units (SI Units).


4. This certificate was certified only for the instrument we calibrated.

5. This result of calibration was found accurate as shown on date and place of calibration only.


6. Condition of Calibrated item : Good

7. Result of Calibration : Without adjustment After adjustment

F-CS-012 Revision: 01 Date: 20-04-65



nfi.th



CERTIFICATE OF CALIBRATION

Certificate No.: SP23-007 Page 1 of 5

Customer : United Analyst and Engineering Consultant Co.,Ltd. (Head Office)

Address : 3 Soi Udomsuk 41, Sukhumvit Road, Bangchak, Phrakanong, Bangkok 10260

Location of calibration : Laboratory 315

Equipment : UV-Vis Spectrophotometer

Manufacturer : Hitachi

Model : U-1900

Serial No. : 2021-064

ID No. : UAE.WAS.006/2552

Received Date : 6 January 2023

Calibration Date : 6 January 2023

Issue Date : 10 January 2023

Condition Instrument : Used

Calibrated by : 
(Mr. Tanawut Rittidach)

Approved by : 
(Ms. Chonthicha Sangngern)

Technical Manager

Quality Manager

The calibration result is applied only to the above calibrated item and was found accurate as shown on date and place of calibration only.

The measurement capability of the laboratory and its traceability to recognized national standards and to the unit of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the DQE Services Co., Ltd.

เอกสารไม่ควบคุม

FM-708-02 Rev 1/11/2021

Calibration Report

Certificate No.: 2302181-001-01

Equipment: Digital Thermometer with RTD (pH Meter)

Resolution: 0.1 °C Model: SevenEasy pH

Serial No.: 1230525212 ID No.: UAE.WAS.003/2553

Manufacturer: METTLER TOLEDO

Date of Calibration: 24 March 2023

Page 5 of 5

Calibration point: 15.0, 25.0 and 30.0 °C

Calibration result:

- The probe was immersed in liquid bath or dry bath to a minimum depth of 120 mm.
- Description of probe, model : N/A S/N : N/A
- Dimension of probe : Diameter 3 mm., Length 120 mm., Sheath material : N/A


UUC* Reading (°C)	Standard Temperature (°C)	Correction Value (°C)	Uncertainty ± (°C)
15.2	14.999	- 0.2	0.12
25.2	24.999	- 0.2	0.12
30.2	29.999	- 0.2	0.12

Note


- UUC* : Unit Under Calibration

The report uncertainty of measurement was based on standard uncertainty multiplied by coverage factor k= 2, providing a level of confidence of approximately 95 %.


----- End -----



F-CS-012 Revision: 01 Date: 20-04-65



nfi.th



REPORT OF CALIBRATION

Certificate No.: SP23-007 Page 2 of 5

Environment Condition : Ambient Temperature 25 ± 5 °C

Relative humidity 55 ± 20 %RH

Calibration method : In-house method CP-01 Based on ASTM E275-08

Certified Reference Materials :

Material	Serial No.	Certificate No.	Due date
Absobance Standard set	25760	95935	22 October 2023
Absobance Standard set	25757	95929	22 October 2023
Wavelength Standard set	25806	95916	22 October 2023
Wavelength Standard set	25758	95915	22 October 2023

Traceability This certification is traceable to the International System of Unit maintained at National -

Institute of Standards and Technology (NIST) through Starna Scientific Limited

Spectral Band Width of UUC : 4.0 nm

Scan Speed of UUC : 200 nm/min

Scan Interval of UUC : 0.1 nm

Resolution of UUC : Photometric 0.001 Abs.

Wavelength 0.1 nm

เอกสารไม่ควบคุม

FM-708-02 Rev 1/11/2021



REPORT OF CALIBRATION

Certificate No. : SP23-007

Page 3 of 5

Calibration Results : Without adjustment

Photometric Accuracy :

Wavelength (nm.)	CRMs Values (Abs)	UUC Reading (Abs)	Correction (Abs)	Uncertainty (Abs)	Coverage factor k
420	0.0000	0.000	0.0000	0.0028	2.00
	0.5787	0.575	0.0037	0.0031	2.00
	1.0490	1.044	0.0050	0.0029	2.00
440	2.1900	2.181	0.0090	0.0080	2.00
	0.0000	0.000	0.0000	0.0028	2.00
	0.5607	0.558	0.0027	0.0034	2.00
465	1.0247	1.021	0.0037	0.0035	2.00
	2.1229	2.115	0.0079	0.0081	2.00
	0.0000	0.000	0.0000	0.0028	2.00
546.1	0.5236	0.520	0.0036	0.0030	2.00
	0.9634	0.961	0.0024	0.0029	2.00
	1.9763	1.968	0.0083	0.0070	2.00
590	0.0000	0.000	0.0000	0.0028	2.00
	0.5523	0.552	0.0003	0.0030	2.00
	1.0809	1.082	-0.0011	0.0030	2.00
635	2.0391	2.031	0.0081	0.0080	2.00
	0.0000	0.000	0.0000	0.0028	2.00
	0.5601	0.562	-0.0019	0.0032	2.00
635	1.0512	1.052	-0.0008	0.0030	2.00
	1.9294	1.923	0.0064	0.0079	2.00

เอกสารไม่ควบคุม

FM-708-02 R01 1/11/2021



REPORT OF CALIBRATION

Certificate No. : SP23-007

Page 5 of 5

Wavelength Accuracy :

CRMs Values (nm.)	UUC Reading (nm.)	Correction (nm.)	Uncertainty (nm.)	Coverage factor k
241.54	240.8	0.74	0.18	2.00
279.40	278.5	0.90	0.18	2.00
288.70	288.0	0.70	0.18	2.00
334.22	333.5	0.72	0.18	2.00
361.26	360.5	0.76	0.18	2.00
418.48	417.8	0.68	0.21	2.00
446.70	445.9	0.80	0.18	2.00
453.20	452.5	0.70	0.18	2.00
460.06	459.5	0.56	0.18	2.00
536.90	536.0	0.90	0.18	2.00
637.94	637.1	0.84	0.18	2.00
440.74	440.0	0.74	0.18	2.00
472.22	471.5	0.72	0.18	2.00
513.70	513.0	0.70	0.18	2.00
528.72	528.0	0.72	0.18	2.00
574.60	574.0	0.60	0.18	2.00
585.48	584.6	0.88	0.20	2.00
684.63	684.0	0.63	0.18	2.00
740.27	740.0	0.27	0.20	2.00
748.28	747.5	0.78	0.18	2.00
807.16	806.5	0.66	0.18	2.00
879.70	879.0	0.70	0.18	2.00

Remark : - UUC = Unit Under Calibration

- N/A = Not Available

- The result expanded uncertainty of measurement U is stated as the standard uncertainty of measurement multiplied by the coverage factor k.

which for a normal distribution corresponds to a coverage probability of approximately 99%

- * Indicates non TISI accredited

- End of Certificate -

เอกสารไม่ควบคุม

FM-708-02 R01 1/11/2021



REPORT OF CALIBRATION

Certificate No. : SP23-007

Page 4 of 5

Photometric Accuracy :

Wavelength (nm.)	CRMs Values (Abs)	UUC Reading (Abs)	Correction (Abs)	Uncertainty (Abs)	Coverage factor k
235	0.0000	0.000	0.0000	0.0050	2.00
	0.7478	0.743	0.0048	0.0057	2.00
257	0.0000	0.000	0.0000	0.0050	2.00
	0.8686	0.861	0.0076	0.0059	2.00
313	0.0000	0.000	0.0000	0.0050	2.00
	0.2912	0.291	0.0002	0.0051	2.00
350	0.0000	0.000	0.0000	0.0050	2.00
	0.6448	0.639	0.0058	0.0055	2.00

เอกสารไม่ควบคุม

FM-708-02 R01 1/11/2021



CERTIFICATE OF CALIBRATION

Certificate No. : SP23-007

Page 1 of 5

Customer : United Analyst and Engineering Consultant Co.,Ltd. (Head Office)

Address : 3 Soi Udomsuk 41, Sukhumvit Road, Bangchak, Phrakanong, Bangkok 10260

Location of calibration : Laboratory 315

Equipment : UV-Vis Spectrophotometer

Manufacturer : Hitachi

Model : U-1900

Serial No. : 2021-064

ID No. : UAE.WAS.006/2552

Received Date : 6 January 2023

Calibration Date : 6 January 2023

Issue Date : 10 January 2023

Condition Instrument : Used

Calibrated by :

(Mr.Tanawut Ritidach)

Technical Manager

Approved by :

(Ms. Chonthicha Sangnern)

Quality Manager

The calibration result is applied only to the above calibrated item and was found accurate as shown on date and place of calibration only.

The measurement capability of the laboratory and its traceability to recognized national standards and to the unit of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the DQE Services Co.,Ltd.

FM-708-02 R01 1/11/2021



REPORT OF CALIBRATION

Certificate No. : SP23-007

Page 2 of 5

Environment Condition : Ambient Temperature $25 \pm 5^{\circ}\text{C}$ Relative humidity $55 \pm 20\%$ RH

Calibration method : In-house method CP-01 Based on ASTM E275-08

Certified Reference Materials :

Material	Serial No.	Certificate No.	Due date
Absorbance Standard set	25760	95935	22 October 2023
Absorbance Standard set	25757	95929	22 October 2023
Wavelength Standard set	25806	95916	22 October 2023
Wavelength Standard set	25758	95915	22 October 2023

Traceability : This certification is traceable to the International System of Unit maintained at National -

Institute of Standards and Technology (NIST) through Starna Scientific Limited

Spectral Band Width of UUC : 4.0 nm.

Scan Speed of UUC : 200 nm/min

Scan Interval of UUC : 0.1 nm.

Resolution of UUC : Photometric 0.001 Abs.

Wavelength 0.1 nm.



REPORT OF CALIBRATION

Certificate No. : SP23-007

Page 3 of 5

Calibration Results : Without adjustment

Photometric Accuracy :

Wavelength (nm.)	CRMs Values (Abs)	UUC Reading (Abs)	Correction (Abs)	Uncertainty (Abs)	Coverage factor k
420	0.0000	0.000	0.0000	0.0028	2.00
	0.5787	0.575	0.0037	0.0031	2.00
	1.0490	1.044	0.0050	0.0029	2.00
440	2.1900	2.181	0.0090	0.0080	2.00
	0.0000	0.000	0.0000	0.0028	2.00
	0.5607	0.558	0.0027	0.0034	2.00
465	1.0247	1.021	0.0037	0.0035	2.00
	2.1229	2.115	0.0079	0.0081	2.00
	0.0000	0.000	0.0000	0.0028	2.00
546.1	0.5236	0.520	0.0036	0.0030	2.00
	0.9634	0.961	0.0024	0.0029	2.00
	1.9763	1.968	0.0083	0.0070	2.00
590	0.0000	0.000	0.0000	0.0028	2.00
	0.5191	0.518	0.0011	0.0031	2.00
	1.0003	1.000	0.0003	0.0033	2.00
635	1.9987	1.993	0.0057	0.0084	2.00
	0.0000	0.000	0.0000	0.0028	2.00
	0.5523	0.552	0.0003	0.0030	2.00
635	1.0809	1.082	-0.0011	0.0030	2.00
	2.0391	2.031	0.0081	0.0080	2.00
	0.0000	0.000	0.0000	0.0028	2.00
635	0.5601	0.562	-0.0019	0.0032	2.00
	1.0512	1.052	-0.0008	0.0030	2.00
	1.9294	1.923	0.0064	0.0079	2.00

REPORT OF CALIBRATION

Certificate No. : SP23-007

Page 4 of 5

Photometric Accuracy :

Wavelength (nm.)	CRMs Values (Abs)	UUC Reading (Abs)	Correction (Abs)	Uncertainty (Abs)	Coverage factor k
235	0.0000	0.000	0.0000	0.0050	2.00
	0.7478	0.743	0.0048	0.0057	2.00
257	0.0000	0.000	0.0000	0.0050	2.00
	0.8686	0.861	0.0076	0.0059	2.00
313	0.0000	0.000	0.0000	0.0050	2.00
	0.2912	0.291	0.0002	0.0051	2.00
350	0.0000	0.000	0.0000	0.0050	2.00
	0.6448	0.639	0.0058	0.0055	2.00



REPORT OF CALIBRATION

Certificate No. : SP23-007

Page 5 of 5

Wavelength Accuracy :

CRMs Values (nm.)	UUC Reading (nm.)	Correction (nm.)	Uncertainty (nm.)	Coverage factor k
241.54	240.8	0.74	0.18	2.00
279.40	278.5	0.90	0.18	2.00
288.70	288.0	0.70	0.18	2.00
334.22	333.5	0.72	0.18	2.00
361.26	360.5	0.76	0.18	2.00
418.48	417.8	0.68	0.21	2.00
446.70	445.9	0.80	0.18	2.00
453.20	452.5	0.70	0.18	2.00
460.06	459.5	0.56	0.18	2.00
536.90	536.0	0.90	0.18	2.00
637.94	637.1	0.84	0.18	2.00
440.74	440.0	0.74	0.18	2.00
472.22	471.5	0.72	0.18	2.00
513.70	513.0	0.70	0.18	2.00
528.72	528.0	0.72	0.18	2.00
574.60	574.0	0.60	0.18	2.00
585.48	584.6	0.88	0.20	2.00
684.63	684.0	0.63	0.18	2.00
740.27	740.0	0.27	0.20	2.00
748.28	747.5	0.78	0.18	2.00
807.16	806.5	0.66	0.18	2.00
879.70	879.0	0.70	0.18	2.00

Remark : - UUC = Unit Under Calibration

- N/A = Not Available

- The result expanded uncertainty of measurement U is stated as the standard uncertainty of measurement multiplied by the coverage factor k, which for a normal distribution corresponds to a coverage probability of approximately 95%*

- * Indicates non TISI accredited

- End of Certificate -



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-29 FAX. 0-2719-9484



Cert.No.: 23MM112
Page.: 1 of 3

Certificate of Calibration

Equipment : Electronic Balance
Manufacturer : Mettler Toledo
Model : XSR205
Serial No. : C009071872
ID No. : UAE.WAO.012/2563
Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
 3 Soi Udomsuk 41, Sukhumvit Road,
 Bangkok, Phakhanong,
 Bangkok 10260
Location : Balance Room
Received order : 26 April 2023
Calibration Date : 26 April 2023
Ambient Temperature : 15 °C to 40 °C
Relative Humidity : 30 % to 90 %
Calibrated by : Man Pattanapongpaiboon
Approved by :
 Approved Signatory
 () Ponthippa Tameyakul
 () Malee Buikrua
 (✓) Suwit Imjai
Issue Date : 2 May 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

เอกสารไม่มีความ



Equipment : Electronic Balance
Condition As-Received : Used Item
Reference : 2304-0459OC-1

Result of calibration

2. Effect of off center loading

A mass of 100 g was placed to various position on the pan.
The weighing machine reading error obtained is given in the table

Position 1	Position 2	Position 3	Position 4	Position 5	Maximum difference between off-center and central loading
(g)	(g)	(g)	(g)	(g)	(g)
-0.0001	-0.0001	0.0000	-0.0001	-0.0001	0.0001

3. Departure from nominal value

Applied Weight	Balance Reading	Correction	Measurement Uncertainty	Coverage Factor
(g)	(g)	(g)	(± mg)	(k)
Unload	0.00000	0.00000	0.014	2.13
0.05	0.05001	-0.00001	0.015	2.09
0.1	0.10001	-0.00001	0.015	2.09
1	1.00001	-0.00001	0.018	2.04
5	5.00003	-0.00003	0.026	2.00
20	20.00006	-0.00006	0.045	2.00
50	50.00006	-0.00006	0.080	2.00
80	80.00004	-0.00004	0.15	2.00
100	100.00000	0.00000	0.16	2.00
150	150.00000	0.00000	0.29	2.00
200	200.00000	0.00000	0.29	2.00

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-00-

เอกสารไม่มีความ

Equipment : Electronic Balance
Condition As-Received : Used Item
Reference : 2304-0459OC-1
Procedure used :-

Calibration were conducted using in-house calibration procedure CP-OB01 according to direct measurement method against standard weight.

Condition of this result of calibration

1. Reference standard instruments:-

Instruments	Model	Serial No.	ID No.	Test report No.	Due date
1) Standard Weight Set (E2)	15884	24053	70RC007	MM-0010-22	20 Jan 2024

2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This result of calibration was made on requested at the point specified by customer.

4. This certificate is not certified for any commercial transaction.

5. This certification is traceable to the International System of Unit.

Result of calibration () Without Adjustment (*) After Adjustment by Internal Calibration

Range capacity : 0 g to 81 g Resolution 0.00001 g

81 g to 220 g Resolution 0.0001 g

Before Adjustment :

Applied Weight (g)	Balance Reading (g)	Correction (g)	Measurement Uncertainty (± mg)	Coverage Factor (k)
80	80.00005	-0.00005	0.15	2.00
200	199.9999	+0.0001	0.29	2.00

After Adjustment :

1. Determination of the standard deviation of weighing machine (n = 10)

Applied Weight (g)	Standard Deviation of Reading (g)
80	0.000007
200	0.000000

เอกสารไม่มีความ

Jan



Cert.No.: 23MM112

Page.: 1 of 3

Certificate of Calibration

Equipment : Electronic Balance

Manufacturer : Mettler Toledo

Model : XSR205

Serial No. : C009071872

ID No. : UAE.WAO.012/2563

Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
3 Soi Udomsuk 41, Sukhumvit Road,
Bangkok, Phakhanong,
Bangkok 10260

Location : Balance Room

Received order : 26 April 2023

Calibration Date : 26 April 2023

Ambient Temperature : 15 °C to 40 °C

Relative Humidity : 30 % to 90 %

Calibrated by : Man Pattanapongpaiboon

Approved by :

() Ponthippa Tameyakul
 () Malee Buikrua
 (✓) Suwit Imjai

Issue Date : 2 May 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written

Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

เอกสารไม่มีความ

Jan



Equipment : Electronic Balance
 Condition As-Received : Used Item
 Reference : 2304-0459OC-1
 Procedure used :-

Calibration were conducted using in-house calibration procedure CP-OB01 according to direct measurement method against standard weight.

Condition of this result of calibration

1. Reference standard instruments:-

Instruments	Model	Serial No.	ID No.	Test report No.	Due date
1) Standard Weight Set (E2)	15884	24053	70RC007	MM-0010-22	20 Jan 2024

2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This result of calibration was made on requested at the point specified by customer.

4. This certificate is not certified for any commercial transaction.

5. This certification is traceable to the International System of Unit.

Result of calibration () Without Adjustment (*) After Adjustment by Internal Calibration

Range capacity : 0 g to 81 g Resolution 0.00001 g

81 g to 220 g Resolution 0.0001 g

Before Adjustment :

Applied Weight	Balance Reading	Correction	Measurement Uncertainty	Coverage Factor
(g)	(g)	(g)	(± mg)	(k)
80	80.00005	-0.00005	0.15	2.00
200	199.9999	+0.0001	0.29	2.00

After Adjustment :

1. Determination of the standard deviation of weighing machine (n = 10)

Applied Weight	Standard Deviation of Reading (g)
(g)	0.00007
200	0.00000

Cert.No.: 23MM112
 Page: 2 of 3




Equipment : Electronic Balance
 Condition As-Received : Used Item
 Reference : 2304-0459OC-1
 Result of calibration

2. Effect of off center loading

A mass of 100 g was placed to various position on the pan. The weighing machine reading error obtained is given in the table

Cert.No.: 23MM112

Page: 3 of 3



Maximum difference between off-center and central loading (g)

3. Departure from nominal value

Applied Weight	Balance Reading	Correction	Measurement Uncertainty	Coverage Factor
(g)	(g)	(g)	(± mg)	(k)
Unload	0.00000	0.00000	0.014	2.13
0.05	0.05001	-0.00001	0.015	2.09
0.1	0.10001	-0.00001	0.015	2.09
1	1.00001	-0.00001	0.018	2.04
5	5.00003	-0.00003	0.026	2.00
20	20.00006	-0.00006	0.045	2.00
50	50.00006	-0.00006	0.080	2.00
80	80.00004	-0.00004	0.15	2.00
100	100.0000	0.0000	0.16	2.00
150	150.0000	0.0000	0.29	2.00
200	200.0000	0.0000	0.29	2.00

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95%.

-00-

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม



Calibration Certificate

Certificate No.: 2203120-001-01
 Client name: UNITED ANALYST AND ENGINEERING CONSULTANT CO., LTD.
 Address: 3 Soi Udomsuk 41, Sukhumvit Road, Bangkok, Prakanong, Bangkok 10260

Page 1 of 3

Equipment: Electronic Balance

Manufacturer: METTLER TOLEDO

Model: AB204-S/FACT

Serial No.: 1129361010

ID No.: UAE.WAS.002/2552

Order No.: 2203120

Operation No.: 2203120-001

Date of Receipt: 1 June 2022

Date of Calibration: 1 June 2022

Calibrated by Mr.Taveesak Seilee
 Scientist

Approved by

 Manager, Division of Calibration Laboratory

Date of Issue: 7 June 2022

Responsible for the Technical Management Team

The uncertainties are for a confidence probability of approximately 95%

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the National Food Institute.

F-CS-009 Revision: 01 Date: 20-04-65

เอกสารไม่ควบคุม

Certificate No.: 2203120-001-01

Equipment: Electronic Balance

Manufacturer: METTLER TOLEDO

Model: AB204-S/FACT

Resolution: 0.0001 g

Serial No.: 1129361010

ID No.: UAE.WAS.002/2552

Capacity: 220 g

Date of Calibration: 1 June 2022

Environment Condition: Ambient Temperature: 19.9 ± 0.3 °C Relative Humidity: 45 ± 1.5 %

Place of Calibration: 108, Balance Room, UNITED ANALYST AND ENGINEERING CONSULTANT CO., LTD.

Condition of Equipment: Good Condition

Condition of This Results of Calibration:

1. Calibration Method: NFI Method W-MA-001 In-House Method based on UKAS Lab 14 : 2019

2. Reference Standards:

Reference Standard	Model	Serial No.	Calibrated By	Certificate No.	Due Date
Standard Weight Class E2	J-500mg	B308068554	TCS	M22010205	6 January 2023
Standard Weight Class E2	J-1000mg	B308068128	TCS	M22010215	6 January 2023

Instrument	Model	Serial No.	Calibrated By	Certificate No.	Due Date
Thermo-Hydro Meter	PONPE 490	NFLBTH 010/18	Quality Reborn	QR22-0350	18 February 2023

3. This certification is traceable to SI UNIT

4. This certificate was certified only for the instrument we calibrated.

5. This result of calibration was found accurate as shown on date and place of calibration only.

Calibration Results:

1. Repeatability of Reading:

Nominal Value (g)	Standard Deviation of Reading (g)
100	0.00048
200	0.00052

2. Off-Center Error:

A mass of 50 g was placed and moved to various position on pan.

The balance reading obtained is given in the table.

1	2	3	4	5	6	(Maximum Difference)
(g)	(g)	(g)	(g)	(g)	(g)	(g)

F-CS-012 Revision: 01 Date: 20-04-65

Calibration Report

Certificate No.: 2203120-001-01
Equipment: Electronic Balance
 Model: AB204-S/FACT
 Serial No.: 1129361010
 Capacity: 220 g

Date of Calibration: 1 June 2022

Page 3 of 3

Calibration Results: (Continued)

Calibration Range: 0 - 200 g

Calibration Adjustment: Internal Calibration

3. Departure from Nominal Value:

Nominal Value (g)	Standard Value (g)	Average Reading (g)	Correction (g)	Uncertainty (± g)	Coverage Factor k
Unload	0.00000	0.0000	0.0000	0.000088	2.00
0.01	0.01000	0.0100	0.0000	0.000088	2.00
0.05	0.05000	0.0499	0.0001	0.000088	2.00
0.1	0.10000	0.1000	0.0000	0.000088	2.00
0.2	0.20000	0.2000	0.0000	0.000088	2.00
0.5	0.50000	0.5000	0.0000	0.000088	2.00
1	1.00000	0.9999	0.0001	0.000088	2.00
2	2.00000	1.9999	0.0001	0.000089	2.00
5	5.00000	5.0000	0.0000	0.000099	2.00
10	9.9998	9.9999	0.0001	0.000092	2.00
20	19.9999	19.9999	0.0001	0.000094	2.00
50	49.9999	49.9999	0.0000	0.00012	2.00
70	69.9999	69.9998	0.0001	0.00014	2.00
100	100.0001	99.9999	0.0001	0.00017	2.00
150	149.9991	149.9997	0.0002	0.00022	2.00
200	200.0007	199.9998	0.0003	0.00030	2.00

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95%.

----- End -----

F-CS-012 Revision: 01 Date: 20-04-65

เอกสารไม่มีความลับ



infodit

2008 ถนนสุขุมวิท 36 แขวงคลองเตย เขตคลองเตย กรุงเทพฯ 10110
 ประเทศไทย 10110
 โทร. +66 2 242 2222 โทร. +66 2 242 233545

Cert. No.: 23TM249
 Page : 1 of 3

Certificate of Calibration

Equipment : BOD Incubator
Manufacturer : Arco
Model : UC4-1320
Serial No. : 13URC4S013201
ID No. : UAE.WAO.015/2561
Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
 3 Soi Udomsuk 41, Sukhumvit Road,
 Bangchak, Phrakanong,
 Bangkok 10260
Location : Lab Floor 2
Received Order : 15 February 2023
Calibration Date : 15 February 2023
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %
Calibrated by : Preecha Hiahib
Approved by :
 Approved Signatory
 Pornthippa Tameyakul
 Malee Butkruea
 Suwit Imjai

Issue Date : 24 February 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
 Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

เอกสารไม่มีความลับ

A 0051476



Equipment : BOD Incubator
Condition As-Received : Used Item
Reference : 2302-0297OC-1
Procedure Used :-

Cert. No.: 23TM249
 Page : 2 of 3

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD).

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1.) Data Acquisition	34972A	MY57013711	22LM93	02 Jul 2023


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Not Available



Probe Installation Details : Dimension of Chamber :

a = 10 cm
 b = 10 cm
 c = 10 cm

Capacity = 0.89 m³

Environment during calibration		
	Beginning	Finished
Temp. (°C)	29	31
REL.Humid. (%)	63	67
AC Supply (Volt)	220	220

Position :	Ref. Std. ID No.:
1	22-18RTD-2/1
2	18RTD-2/2
3	18RTD-2/3
4	18RTD-2/4
5	18RTD-2/5
6	18RTD-2/6
7	18RTD-2/7
8	18RTD-2/8
9 (ref.)	18RTD-2/9



Equipment : BOD Incubator
Condition As-Received : Used Item
Reference : 2302-0297OC-1
Result of Calibration :- (*) Without Adjustment
Function of UUC* : Temperature Source
Fresh air setting : Not Available

Cert. No.: 23TM249
 Page : 3 of 3

Calibration Point (°C)	Measured Temperature (°C) Position								
	1	2	3	4	5	6	7	8	9 (ref.)
20.0	20.086	19.916	20.386	19.976	19.973	19.838	19.837	19.821	19.949

Average*: The average of 30 values in each position.

Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor.

Temperature uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-00-

เอกสารไม่มีความลับ

เอกสารไม่มีความลับ

a 1149512



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
334/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL: 0-2717-3000-29 FAX: 0-2719-9484



Cert. No.: 23TM372
Page : 1 of 3

Certificate of Calibration

Equipment : BOD Incubator
Manufacturer : ARCO
Model : UR-1320
Serial No. : -
ID No. : UAE.WAO.006/2553
Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
3 Soi Udomsuk 41, Sukhumvit Road,
Bangchak, Phrakanong,
Bangkok 10260
Location : Lab Floor 2
Received Order : 11 April 2023
Calibration Date : 11 April 2023
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %
Calibrated by : Krisda Malee

Approved by :
Approved Signatory
() Pornthipha Tameyakul
(✓) Malee Butkrua
() Suwit Imjai

Issue Date : 24 April 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม

A 0053361



Equipment : BOD Incubator
Condition As-Received : Used Item
Reference : 2304-0156OC-3
Result of Calibration :- (*) Without Adjustment
Function of UUC* : Temperature Source
Fresh air setting : Not Available

Calibration Point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Temperature stability (± °C)	Temperature uniformity (°C)	Overall Variation (°C)	Coverage Factor k
20.0	20.0	19.9	0.40	0.72	0.97	2

Calibration Point (°C)	Measured Temperature (°C)									Uncertainty (± °C)
	Position									
20.0	1	2	3	4	5	6	7	8	9 (ref.)	0.59

Average* : The average of 30 values in each position.

Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor.
Temperature uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location, which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-00-

เอกสารไม่ควบคุม
a 1151821



Equipment : BOD Incubator
Condition As-Received : Used Item
Reference : 2304-0156OC-3
Procedure Used :-

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD).
The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34972A	MY59003411	22LM165	26 Nov 2023

2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Not Available

Environment during calibration		
	Beginning	Finished
Temp. (°C)	27	28
REL.Humid. (%)	44	41
AC Supply (Volt)	221	220

Position :	Ref. Std. ID No.:
1	20RTD-2/1
2	20RTD-2/2
3	20RTD-2/3
4	20RTD-2/4
5	20RTD-2/5
6	20RTD-2/6
7	20RTD-2/7
8	20RTD-2/8
9 (ref.)	20RTD-2/9

Probe Installation Details : Dimension of Chamber :
a = 10 cm D = 0.62 m
b = 10 cm W = 1.2 m
c = 10 cm H = 1.2 m
Capacity = 0.89 m³



Hanna Instruments (Thailand) Ltd.

410/67-68 Soi Ratchadapisek 24, Ratchadapisek Rd., Samson-nok,
Huaykwang, Bangkok 10310 Tel: 0-2541-4199 Fax: 0-2541-4198



Certificate No. : HIT-2312-0342

Page : 1 of 2

CERTIFICATE OF CALIBRATION

Equipment : COD Test Tube Heater
Meter Model : HI839800-02 Serial No. : H018500I
Tube Heater : 25 Vial Capacity Accuracy : ± 2 °C
Temperature Range : -10 °C to 160 °C Temperature of Reaction : 150 °C
Ambient Temperature : (25 ± 2) °C Relative Humidity : (50 ± 15) % RH
Manufacturer : Hanna Instruments Made in : Romania
Condition As-Received : Used Product Reference : RE230392
Customer name : United Analyst and Engineering Consultant Co., Ltd.
3 Soi Udomsuk 41, Sukhumvit Rd., Bangchak,
Phrakanong, Bangkok 10260
Received date : 8 March 2023
Calibrate date : 10 March 2023
Issue date : 20 March 2023
Calibrated Location : Hanna Instruments (Thailand) Ltd.
Calibration Procedure : This calibrator was conducted by using in-house: calibration procedure CP-04 by using certified reference material.

Calibrated by : Mr. Pichit Petpong
 Mr. Jakkapob Pentisan Approved by : Mr. Anan Suwancharaiskul
 Mr. Channarong Soinak
Authorized Signatory



This certificate was certified only for the instrument we calibrated.

This result of calibration was found accurate on date and place of calibration only.

** This certificate may not be reproduced other than in full, except with the prior written **
approval of the head of Hanna Instrument (Thailand).

เอกสารไม่ควบคุม

Condition of this calibration result

Reference Standard Instruments:

Instruments	Model	Serial No.	Certificate No.	Traceable
Data Acquisition Switch Unit	34970A	MY44065265	WK2207-065-I	WK Electric Co., Ltd.

Calibration Result:

Measurement Temperature Source Accuracy for COD Reactor

Capacity (Vial)	Nominal Value (°C)	Average Value (°C)	± Uncertainty (°C)	± Tolerance of UUC (°C)	Acceptance Criteria
25 Vial	150.0	150.3	0.59	2	Pass

Figure: Shows the location of the temperature source.

(1A)	(2A)	(3A)	(4A)	(5A)
149.78°C	150.31°C	150.63°C	149.93°C	150.31°C
(1B)	(2B)	(3B)	(4B)	(5B)
150.35°C	150.18°C	149.93°C	150.18°C	150.21°C
(1C)	(2C)	(3C)	(4C)	(5C)
150.24°C	151.10°C	150.80°C	150.36°C	150.86°C
(1D)	(2D)	(3D)	(4D)	(5D)
150.16°C	149.77°C	150.22°C	150.67°C	150.43°C
(1E)	(2E)	(3E)	(4E)	(5E)
149.94°C	150.44°C	150.06°C	150.63°C	149.29°C

Remark: The Acceptance criteria is the error value plus or minus the Measurement Uncertainty, and then Not More than the Tolerance value of UUC, therefore concluded that pass.

The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%.

** End of certificate **

เอกสารไม่ควบคุม

Verification Certificate

Certificate No.:

2302413-001-01

Client name:

UNITED ANALYST AND ENGINEERING CONSULTANT CO., LTD.

Address:

3 Soi Udomsuk 41, Sukhumvit Road,
Bangchak, Prakanong, Bangkok 10260

Page 1 of 4

Equipment: HEATING BLOCK DIGESTION

Manufacturer: FOSS

Model: 2520

Serial No.: 91794469

ID No.: UAE.WAS.011/2560

Order No.: 2302413

Operation No.: 2302413-001

Date of Receipt: 28 March 2023

Date of Calibration: 30-31 March 2023

Calibrated by Mr.Nuttapol Niyomchat
SpecialistApproved by (Mr.Pheraphat Tuanjit)
Manager, Division of Calibration Laboratory

Date of Issue: 10 April 2023

Responsible for the Technical Management Team

The uncertainties are for a confidence probability of approximately 95 %.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the National Food Institute.

F-CS-009 Revision: 01 Date: 20-04-65

2008 ถนนกรุงศรีฯ 35 หมู่บ้านจัดสรร หนองบัวใหญ่ ตำบลหนองบัวใหญ่ อำเภอเมือง จังหวัดนนทบุรี 10700
โทรศัพท์ 089-2 86668 Fax. 089-2 84222 8545
e-mail: nfit@nfit.or.th

Verification Report

Certificate No.: 2302413-001-01

Equipment: HEATING BLOCK DIGESTION

Model: 2520 Serial No.: 91794469

Resolution: 1 °C ID No.: UAE.WAS.011/2560

Manufacturer: FOSS

Date of Calibration: 30-31 March 2023

Page 2 of 4

Location: Laboratory Room, NATIONAL FOOD INSTITUTE

Environment Condition: Ambient Temperature (25 ± 3) °C
Relative Humidity (55 ± 15) %
Line Voltage (220 ± 10) Volt

Condition of this results of Calibration:

- This instrument was calibrated by insert standard thermocouples type R into its heating block digestion and compared to temperature obtained from reference standards thermometer at calibrated point.
- The temperature scale used was based on ITS - 90 .
- All data show below were final values and the initial data may be obtained upon request.

2. Reference Standard Instrument :

Instrument	Model	Serial No.	Certificate No.	Due Date	Through
Digital Thermometer with Thermocouple	34970A	MY44065265/MY44065265	TC22/0044	5-May-2023	N.I.M. Technical Center Laboratory
Type R	TC#101-103 / CH#101-103				

3. This certificate is traceable to international system of units (SI Units).

4. This certificate was certified only for the instrument we calibrated.

5. This result of calibration was found accurate as shown on date and place of calibration only.

6. Condition of Calibrated item : Good

UUC* Description

Time of Record - Hour 30 Minute At 380 °C

7. Result of Calibration : Without adjustment After adjustment

Certificate No.: 2302413-001-01

Equipment: HEATING BLOCK DIGESTION

Model: 2520 Serial No.: 91794469

Resolution: 1 °C ID No.: UAE.WAS.011/2560

Manufacturer: FOSS

Date of Calibration: 30-31 March 2023

Page 3 of 4

Calibration point:

Calibration result:

Reporting of Temperature

Block No.	UUC* Setting (°C)	UUC* Reading (°C)	Stability (±°C)	Standard Thermometer (°C)	Uncertainty (±°C)
1	380	380	0.96	377.74	2.1
2	380	380	0.40	377.28	2.1
3	380	380	1.18	377.82	2.1
4	380	380	0.44	377.19	1.6
5	380	380	0.11	377.30	1.6
6	380	380	0.14	377.90	1.6
7	380	380	1.17	373.85	2.1
8	380	380	0.33	376.96	2.1
9	380	380	0.14	374.18	2.1
10	380	380	0.96	378.56	2.0
11	380	380	1.04	378.34	2.0
12	380	380	0.35	378.06	2.0
13	380	380	0.48	377.05	1.6
14	380	380	0.38	379.19	1.6
15	380	380	0.50	377.48	1.6
16	380	380	0.48	378.33	1.7
17	380	380	0.71	377.60	1.7
18	380	380	0.35	376.77	1.7
19	380	380	0.84	377.06	1.8
20	380	380	0.41	378.58	1.8

Note:

- UUC* = Unit Under Calibration


- Immersion depth of standard thermometer in tube level high of sand is equal heater plate of UUC.

- Stability = One-half of the greatest maximum difference of measured temperatures at one sensors, for at least half an hour after reaching steady state.

Verification Report

Certificate No.: 2302413-001-01
Equipment: HEATING BLOCK DIGESTION
 Model: 2520 Serial No.: 91794469
 Resolution: 1 °C ID No.: UAE.WAS.011/2560
 Manufacturer: FOSS
Date of Calibration: 30-31 March 2023 Page 4 of 4
Calibration point: 380 °C
Calibration result: Continued

Figure 1. Location of Reference Standard and Block Diagram of Digestion Unit




- Note:
 - UUC* = Unit Under Calibration
 - Immersion depth of standard thermometer in tube level high of sand is equal heater plate of UUC.
 - Stability = One-half of the greatest maximum difference of measured temperatures at one sensors, for at least half an hour after reaching steady state.

The report uncertainty of measurement was based on standard uncertainty multiplied by coverage factor $k=2$, providing a level of confidence of approximately 95 %. *H2*

----- End -----

FCS-009 Revision: 01 Date: 20-04-65



Verification Certificate

Certificate No.: 2304455-001-01
Client name: UNITED ANALYST AND ENGINEERING CONSULTANT CO., LTD.
Address: 3 Soi Udomsuk 41, Sukhumvit Road, Bangkok, Prakanong, Bangkok 10260

Page 1 of 4

Equipment: Digestion Unit (Heating Block)
Manufacturer: VELP SCIENTIFICA
Model: DKL20
Serial No.: 213517
ID No.: UAE.WAS.005/2555
Order No.: 2304455
Operation No.: 2304455-001
Date of Receipt: 28 August 2023
Date of Calibration: 28-29 August 2023


Calibrated by Mr. Manas Somsak Specialist **Approved by** *H2*
(Mr.Pheraphat Tuanjitt)

Date of Issue: 1 September 2023 **Manager, Division of Calibration Laboratory**
Responsible for the Technical Management Team

The uncertainties are for a confidence probability of approximately 95 %.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the National Food Institute.

FCS-009 Revision: 01 Date: 20-04-65



Verification Report

Certificate No.: 2304455-001-01
Equipment: Digestion Unit (Heating Block)
 Model: DKL20 Serial No.: 213517
 Resolution: 1 °C ID No.: UAE.WAS.005/2555
 Manufacturer: VELP SCIENTIFICA
Date of Calibration: 28-29 August 2023 Page 2 of 4

Location: Dry Laboratory (312) , UNITED ANALYST AND ENGINEERING CONSULTANT CO., LTD.
Environment Condition: Ambient Temperature (28 ± 1) °C
 Relative Humidity (56 ± 3) %
 Line Voltage (224 ± 2) Volt

Condition of this results of Calibration:

- This instrument was calibrated by insert standard thermocouples type R into its Digestion blocks and Calibration according to NFI Method W-TE-02 based on BS 4309 : 1968
- The temperature scale used was based on ITS - 90 .
- All data show below were final values and the initial data may be obtained upon request.

2. Reference Standard Instrument :

Instrument	Model	Serial No.	Certificate No.	Due Date	Through
Digital Thermometer with Thermocouple	34970A Type R	MV4049576/MV4119453	TC23/0048	2-Jun-2024	N.M. Technical Center Laboratory

3. This certificate is traceable to International System of Units (SI Units).

4. This certificate was certified only for the instrument we calibrated.

5. This result of calibration was found accurate as shown on date and place of calibration only.


6. Condition of Calibrated item : Good

UUC* Description

Time of Record 1 Hour 6 Minute At 380 °C

7. Result of Calibration : Without adjustment After adjustment

F-CS-012 Revision: 01 Date: 20-04-65



Verification Report

Certificate No.: 2304455-001-01
Equipment: Digestion Unit (Heating Block)
 Model: DKL20 Serial No.: 213517
 Resolution: 1 °C ID No.: UAE.WAS.005/2555
 Manufacturer: VELP SCIENTIFICA

Date of Calibration: 28-29 August 2023 Page 3 of 4

Calibration point: 380 °C

Calibration result:

Table1 : Reporting of Temperature


Block No.	UUC* Setting (°C)	UUC* Reading (°C)	Stability (±°C)	Standard Thermometer (°C)	Uncertainty (±°C)
1	380	380	0.16	378.59	2.0
2	380	380	0.18	378.65	2.0
3	380	380	0.18	381.62	2.0
4	380	380	0.24	380.23	2.0
5	380	380	0.26	379.86	2.0
6	380	380	0.26	380.93	2.0
7	380	380	0.25	381.11	2.0
8	380	380	0.19	382.35	2.0
9	380	380	0.26	381.55	2.0
10	380	380	0.25	380.20	2.0
11	380	380	0.29	382.08	2.0
12	380	380	0.19	382.26	2.0
13	380	380	0.19	382.26	2.0
14	380	380	0.21	382.15	2.0
15	380	380	0.12	382.15	2.0
16	380	380	0.20	381.91	2.0
17	380	380	0.15	381.09	2.0
18	380	380	0.13	381.42	2.0
19	380	380	0.13	381.77	2.0
20	380	380	0.29	382.08	2.0

Note:

- UUC* = Unit Under Calibration
- Immersion depth of standard thermometer in tube level high of sand is equal heater plate of UUC.
- Stability = One-half of the greatest maximum difference of measured temperatures at one sensors, for at least half an hour after reaching steady state.

The report uncertainty of measurement was based on standard uncertainty multiplied by coverage factor $k=2$, providing a level of confidence of approximately 95 %.

F-CS-012 Revision: 01 Date: 20-04-65



Kjeltec™ 8100 Distillation Unit

This IQ applies to Kjeltec™ 8100 Distillation Unit manufactured by FOSS Analytical. The installation is performed by FOSS trained service personnel.

1 Intended Use

Kjeltec 8100 is intended for laboratory use analyzing parameters as specified in FOSS Analytical AB's Application Notes.

2 Purpose

This installation Qualification is designed to assure that:

- The Kjeltec instrument is received complete, with all required parts in good condition.
- The location of the instrument is environmentally and ergonomically suitable
- The instrument is assembled and configured correctly
- Suitable electricity and water are supplied to the instrument, see table 2 for requirements.

3 Identification

Description	Serial Number
Kjeltec 8100 Distillation Unit	91% 69052

4 Control of Received Equipment**4.1 Verify that the correct instrument type and accessory kit items are received and in proper condition**

The packing list (shipped with the instrument) specifies all the items. The installer will verify that all items are received as shipped on the packing list. For each item listed, verify that the acceptance criteria are met. If so, write "Y" in the right column of the table immediately following.

Packing List Item	Acceptance Criteria	Pass/Y/N
Kjeltec 8100 Distillation Unit	No visible damage, received in undamaged FOSS Analytical's standard shipping container	Y
Accessory kit, according to packing list	Included. No visible damage, received in undamaged FOSS Analytical's standard shipping container	Y
Handling device for digestion tube	Included. No visible damage.	Y
Tanks with level sensors for Waste, Alkali and Water	Included. No visible damage.	Y
Receiver flask	Included. No visible damage.	Y
One digestion tube 250ml	Included. No visible damage.	Y
One digestion tube 100 ml	Included. No visible damage.	Y
Tube adapter	Included. No visible damage.	Y
User manual	Kjeltec 8100 Distillation Unit	Y
Owners guide	Kjeltec 8100 Distillation Unit	Y
Quick guide	Kjeltec 8100 Distillation Unit	Y
Spare parts manual	Kjeltec 8100 Distillation Unit	Y
Application notes	AN 300 included AN 303 included	Y

Dedicated Analytical Solutions

FOSS Analytical A/S
69 Slangerupsgade
DK-3400 Hillerød
Denmark
Tel +45 7010 3370
Fax +45 7010 3371
E-mail support@foss.dk
Web www.foss.dk

FOSS Analytical AB
Box 70
SE-263 21 Höganäs
Sweden
Tel +46 42 361500
Fax +46 42 340349
E-mail support@foss.dk
Web www.foss.dk

Customer Support, 6003 7242 / Rev. 1

ເອກສານໄໝຄວບຄຸມ

Customer Support, 6003 7242 / Rev. 1

ເອກສານໄໝຄວບຄຸມ

2(6)

5 Installation**5.1 The equipment must be installed in a suitable location with power, water and draining available**

Verify that the instrument installation site meets the acceptance criteria given in the table below. If so, write "Y" in the right column of the table immediately following.

Location Requirements	Acceptance Criteria	Pass (Y/N)
Adequate space for instrument	Dimensions 48x58x69 cm	Y
AC supply available for instrument	200-240 V 50/60Hz	Y
Current	10 A	Y
Cold water supply available	2 l/min at 30°C	Y
Drain	For cooling water and waste (depending on local waste disposal legislation)	Y
Ambient temperature	Max. 40°C	Y
Ambient humidity	Max. 80% relative	Y
Internal fuses	T10A AH	Y

5.2 The instrument must be assembled correctly

Verify that all tubes are correctly connected. If so, write "Y" in the right column of the table immediately following.

Instrument Tubing Connections	Acceptance Criteria	Pass (Y/N)
<p>Visual verification by installer: Y</p> <ol style="list-style-type: none"> Deionised water in (separ. generator) Deionised water in (dilution water) Receiver solution in Alkali in Power No. used External titration module Level sensors Cooling water in (tap water) Waste water out (tube drain vessel) Drain Cooling water out (tap water) Only on Kjeltec 8200 		

5.3 The instrument should be assembled and powered up

Connect the distilling unit to the power supply. Perform the start up procedure and check that the expected response is obtained. If so, write "Y" in the right column of the table immediately following.

Action	Expected Response	Pass (Y/N)
Switch on the power	The instruments start up and the self test will run. The sample counter shows the number of analysed samples since first power and the Software Version shows the version of the instruments software.	Y
	After start-up, Program 1 is loaded and the Analyse menu is displayed.	Y
Turn on the cold water tap	No visible reaction	Y
Press the "Manual" view	The Manual menu is opened	Y
Open the door with the handle, place the test tube and receiver flask in position. Close the door.		Y
Select Dilution and press Start	Water is added to the tube	Y
Select Alkali and press Start	Alkali is added to the tube	Y
Select Steam and press start	After heating up, steam is entering the tube	Y
Select Drain and press Start	The tube is drained	Y

6 Summary of Deviations/Comments

Deviations from above requirements are specified below and any corrective actions are noted.

Deviation	Action	Comment

7 IQ Documentation

Upon successful completion and recording of all instructions above, sign and date this sheet below. If required by customer, leave one signed copy with instrument.

If customer's internal procedures require further reporting or witnessing of results, execute those procedures as required.

Installed By: Pannipa Onnum
 Company: Foss SEA
 Customer Name: United Analyst and Engineering
 Company: United Analyst and Engineering
 Date completed: July 25, 2022

Customer Support, 6003 7242 / Rev. 1

เอกสารไม่ควบคุม

Customer Support, 6003 7242 / Rev. 1

เอกสารไม่ควบคุม

FOSS

Operation Qualification

Kjeltec™ 8100 Distillation Unit

This OQ applies to Kjeltec 8100 Distillation Unit manufactured by FOSS Analytical. The operation qualification is performed by FOSS trained service personnel.

1 Intended Use

Kjeltec 8100 is intended for laboratory use analyzing parameters as specified in FOSS Analytical Application Notes.

2 Purpose

This procedure is designed to test the function of the instrument according to factory test specifications:

- Alkali volume
- Distillation Accuracy
- Distillation Repeatability

3 Identification

Description	Serial Number
Kjeltec 8100 Distillation Unit, 200-240 V 50/60 Hz	Q1 769052

Dedicated Analytical Solutions

FOSS Analytical A/S
 Tel +45 7010 3370
 Fax +45 7010 3371
 E-mail support@foss.dk
 Web www.foss.dk
 DK-3400 Hillerød
 Denmark

Customer Support, 6003 7246 / Rev. 1

FOSS Analytical AB
 Box 300
 SE-263 21 Höganäs
 Sweden

Tel +46 42 361500
 Fax +46 42 361549
 E-mail support@foss.se
 Web www.foss.se

เอกสารไม่ควบคุม

4 Performance

4.1 Verify the dispensed volumes of reagents

Note! To verify the dispensed volumes of reagents a triple test should be done to be statistic correct. Then calculate a mean value.

1. Choose "Manual" in the menu. (When starting up the instrument Program 1 is loaded)
2. Open the safety door by pressing Open and place a tube in the instrument. Close the safety door.

Water

1. Press Dilution and then press Start. 80 ml of water will be filled into the tube.
2. Measure the collected water in a graduated measuring glass and note the result in table 1 below.
3. Check acceptance criteria in the table and make the judgment if passed or not.

Note! If the water volume needs to be calibrated, go to 4.8.5 Dilution Pump Calibration in the User Manual.

Alkali

1. Press Alkali and then press Start. 50 ml of alkali will be filled into the tube.
2. Measure the collected alkali in a graduated measuring glass and note the result in table 1 below.
3. Check acceptance criteria in the table and make the judgment if passed or not.

Table 1 Volume control

Test	Result	Expected result	Passed (Y/N)
Water volume	<u>75</u> ml <u>83</u> ml <u>72</u> ml Mean <u>72.67</u> ml	76- 84 ml	Y
Alkali volume	<u>52</u> ml <u>52</u> ml <u>53</u> ml Mean <u>52.22</u> ml	47- 54 ml	Y

Customer Support, 6003 7246 / Rev. 1

เอกสารไม่ควบคุม

4.2 Verify the distillation procedure, accuracy and precision

The distillation principle is to convert ammonium (NH_4^+) into ammonia (NH_3) by using an alkali (NaOH) and thereafter steam distil it into a receiver flask containing boric acid and titrate with standard acid solution using colorimetric end-point detection. Ammonium sulphate, a substance with known ammonia content, can be used to check the accuracy of the distillation. The recovery is calculated from obtained result.

The way to perform this test will be described in the following.

Chemical Check

Use ammonium sulphate ($(\text{NH}_4)_2\text{SO}_4$, purity > 99.5 % *)

Mol. weight = 132.14 g/mol, Nitrogen content in ammonium sulphate (99.5 %) = 21.09% *)

Analysis conditions according to AN 300

Water	80 ml
Alkali	50 ml NaOH (40% w/w)
Receiver solution	30 ml boric acid (4%)
Distillation time	5 minutes
SAFE	5 seconds
Titrant	0.2N HCl

For reagent preparation see Appendix A

- Start the instrument and run two blanks without chemicals according to above analysis conditions, distil into a receiver flask containing boric acid. Titrate with a standard acid solution using colorimetric end-point detection. If the blanks are less than 0.2 ml continue with the recovery tests:
- Weigh 0.15 g ammonium sulphate into a tube. Prepare 6 samples (tubes).
- Run the six samples according to above analysis conditions. Titrate with a standard acid solution using colorimetric end-point detection.
- Calculate the recovery according to below equations. Expected results of recovery should be $100\% \pm 1\%$.

Recovery test	Result	Expected result	Passed (Y/N)
Blank value (water blank)	1. <u>0.09</u> ml 2. <u>0.10</u> ml	0.05-0.20 ml	Y
Recovery	1. <u>100.30</u> % 2. <u>100.30</u> % 3. <u>100.55</u> % 4. <u>99.91</u> % 5. <u>99.97</u> % 6. <u>100.01</u> %		
Accuracy	Mean Value: <u>100.0%</u>	99-101%	Y
Precision	SD: <u>0.057</u>	SD <1%	Y

*) Note! Please also note that the below calculations must be adjusted if other purity levels of ammonium salts are used. A certificate for the chemical supplier should be available

Purity	Nitrogen content
99.5%	21.09% ✓
99.6%	21.12%
99.7%	21.14%
99.8%	21.16%
99.9%	21.18%

$$\% \text{ Nitrogen} = \frac{(ml_{\text{sample}} - ml_{\text{blank}}) \times N \times 14,007 \times 100}{mg_{\text{sample}}} \quad 2.1-7.2$$

N = Normality of titrant to 4 places of decimal.

$$\% \text{ Recovery} = \frac{\% \text{ Nitrogen}}{21.09} \times 100$$

mg sample

- ① 0.1592 21.5%
- ②
- ③
- ④
- ⑤
- ⑥

Customer Support, 6003 7246 / Rev. 1

เอกสารไม่ควบคุม

Customer Support, 6003 7246 / Rev. 1

เอกสารไม่ควบคุม

5 Summary of Deviations/Comments

Deviations from above requirements are specified below and any corrective actions are noted.

Deviation	Action	Comment

6 CQ Documentation

Upon successful completion of tests above, sign and date this sheet below. If required by customer, leave one signed copy with instrument.

If customer's internal procedures require further reporting or witnessing of results, execute those procedures as required.

Performed By: _____

Company: _____

Customer Name: _____

Company: _____

Date completed: _____

7 Appendix A

7.1 Preparation of Reagents

7.1.1 Alkali

To convert ammonium into ammonia an excess of sodium hydroxide is necessary.

Use 400 g NaOH per litre of solution. Commercially available in concentrations up to 50 %. Do not use concentrations above 40 % as this will lead to crystal formation impairing the function of the pumps. If you can only buy concentrations > 40 %, dilute it before use.

7.1.2 Titrant acid, determination of concentration

To be able to achieve accurate nitrogen / protein results, one must be quite sure that the HCl (hydrochloric acid) concentration is what it is supposed to be. A titration against a predetermined solution of sodium carbonate as described below is thus necessary. Incorrect HCl concentration can otherwise cause substantial errors.

• Standard substance

Weight approx. 0.4 g of anhydrous sodium carbonate (Na_2CO_3). Use a mortar to make a fine powder. Dry it for 1 h at 265 °C or 2 h at 200 °C. After cooling in a desiccator, transfer the sodium carbonate to a beaker with a tight lid. Store it in a desiccator.

• Indicator solutions

Dissolve 0.1 g methyl red in 100 ml methanol. Dissolve 0.1 g bromocresol green in 100 ml methanol.

• Procedure

Weight approx. 0.4 g of the standard substance, using an analytical balance, note the weight (W_1). Transfer the sodium carbonate to a receiver flask and add 40 ml of Tl_2O (distilled or deionized). Add 8 drops from each of the indicator solutions. Titrate to pink. Note the amount in ml used (A_1). Boil this solution for a few minutes. The solution will turn green. Cool rapidly to room temperature under running water. Continue the titration until the next pink colour change occurs. Note also this volume (A_2).

(A₂). Boil the solution for a few minutes. Cool rapidly to room temperature under running water. Continue the titration until the next pink colour occurs. Note also this volume (A_3).

Note! Temperature changes will influence the volume and the concentration of the titrant solution. The working temperature of the titrant should approximate that of its temperature during standardization. If temperature corrections are necessary, sufficient accuracy may be obtained by use of a correction table. (AGAC 912.25)

Customer Support, 6003 7246 / Rev. 1

เอกสารไม่ควบคุม

Customer Support, 6003 7246 / Rev. 1

เอกสารไม่ควบคุม

7.2 Calculation

$$\text{Solarity (d)} = \frac{18,870 \times W}{(A_1 + A_2 + A_3)}$$

Note! Concentration must be accurate to four digits, i.e. 0.2000 M.

Note! The colour change of this official procedure (AOAC 936.5) may be difficult to see, therefore a pH meter or a mixed indicator (e.g. 0.1 g Methyl red and 0.1 g Bromoresol green in 100 ml methanol) will make it much easier to perform.

7.3 Receiver Solution

Boric acid 4% with bromoresol green / methyl red indicator solution

In order to obtain accurate results the receiver solution is adjusted so that a small (0.05-0.20 ml) positive blank is obtained when running a blank sample. The 4% boric acid receiver solution is prepared by dissolving 400 g of boric acid in about 5-6 l very hot deionized water. Mix and add more hot deionized water to a volume of about 9 l. Cool the solution to room temperature and add 100 ml of bromoresol green solution (100 mg in 100 ml methanol) and 70 ml of methyl red solution (100 mg in 100 ml of methanol). Dilute to 10 l with deionized water and mix carefully.

Note! The addition of alkali is to achieve a positive blank value. This should, however, be kept between 0.05 - 0.20 ml titrant, to obtain good repeatability when testing blanks.

Adjustment of the boric acid is made by the following procedure:

- Transfer 25 ml boric acid solution to a receiver flask and add 100 ml of distilled water. If the solution in the flask is still red, titrate with 0.1 M sodium hydroxide solution until a neutral grey colour is obtained. Calculate the amount of sodium hydroxide solution necessary to adjust the boric acid solution in the 10 l flask with the formula: ml 1.0 M alkali / ml titrant $\times 7.0$
- Add the calculated amount of 1.0 M alkali solution to the boric acid solution. Mix.
- To check proceed as follows using 25 ml of the boric acid solution. Run a blank. If the value of this blank is high (0.5 ml of 0.2 M HCl) the boric acid is incorrectly adjusted. This might create irregular blanks. For correction add HCl directly into the boric acid tank, mix it carefully and repeat until a reading of 0.05 - 0.20 ml HCl is obtained. If a positive blank is not achieved, add further small quantities of 1 M NaOH and repeat the check until a satisfactory value is achieved.

© FOSS Analytical A/S 2002 - Rev. 1

ເອກສານໄໝຄວບຄຸມ

4.2 Ongoing Qualification Tests

Block Temperature

The temperature for the digestion is limited by the boiling point for the sulphuric acid, this can be increased by adding a salt (K_2SO_4) to the digestion mixture. It's important that the optimal ratio between acid and salt is kept; please follow recommendation in AN 300 or suggested procedures for a specific kind of sample material.

The block temperature itself can be controlled external by inserting a temperature probe in the intended hole at the aluminium block (front row of holes).

Use the reagents and method procedure specified in AN 300. Use only reagents of recognized analytical grade, unless otherwise specified and distilled or demineralised water or water of equivalent purity.

Suggested standard material for internal quality control:

Ammonium sulphate ($(NH_4)_2SO_4$), min. 99.5 % (mass fraction), with certified purity.

Note: The above chemical is usually readily available with a certificate specifying the purity.

Alternatively ammonium iron(II) sulphate, $(NH_4)_2Fe(SO_4)_2 \times 6H_2O$, with certified purity may be used.

Tryptophan ($C_{11}H_{14}N_2O$), minimum assay 99 % (mass fraction). Nitrogen content: 37.2 g/kg. Do not dry in an oven before use.

Ascanilide ($C_6H_{11}NO$), minimum assay 99 % (mass fraction). Nitrogen content: 103.6 g/kg. Do not dry in an oven before use.

Shurese, ($C_{12}H_{17}O_2$), with a nitrogen content of not more than 41.002 % (mass fraction). Do not dry in an oven before use.

Carry out a blank test following the currently used procedure for digestion, distillation and titration taking 2 ml of water and about 0.7 g of sucrose instead of the test portion. Keep a record of blank values. If blank values change, identify the cause.

Note: The amount of titrant used in the blank test should always be greater than 0.8 ml. Blanks within the same laboratory should be consistent across time.

4.3 Recovery Tests

Regularly run recovery studies to check the accuracy of procedures and equipment:

- Nitrogen loss** - Use 0.12 g ammonium sulphate and 0.67 g sucrose per flask weighed to the nearest 0.1 mg. Add all other reagents as stated in the method currently used (Kjeldahl, H_2SO_4 etc.). Digest and distill under same conditions as for sample. Recoveries shall be >99 %.
- Digestion efficiency** - Use a test portion of minimum 0.15 g of tryptophan or ascanilide and 0.67 g sucrose per flask weighed to the nearest 0.1 mg. Determine the nitrogen content according to the current procedure in use. The recoveries of tryptophan shall be >98.5 %; the recoveries of ascanilide shall be >99.5 %.
- Distillation and titration efficiency** - Distill 0.19 - 0.15 g ammonium sulphate, omitting the digestion step. The recoveries should be >99.5 %.

Note: Results less than 98.5 % or more than 101.0 % in either of the recovery tests indicate failures in the procedure and/or inaccurate concentration of the standard volumetric hydrochloric acid solution (should be adjusted to four decimals accuracy according to procedure in AN 300)

Kjeltec™ 8100 Distillation Unit Tecator™ 2508/2520 Digestor

1 Scope

This PQ applies to the Digestion system 2508/2520 (including exhaust and scrubber unit) and Kjeltec 8100 Distillation Unit manufactured by FOSS Analytical. The user of the instrument performs the PQ.

2 Intended Use

The Digestion system (including exhaust and scrubber) and Kjeltec 8100 Distillation Unit are intended for laboratory use analyzing parameters as specified in FOSS Application Notes.

3 Purpose

The guidelines are intended to assist the user in successfully developing Performance Qualifications for the specific application(s) to which the instrument is applied.

The Performance Qualification (PQ) includes the process of demonstrating that the Digestion system 2508/2520 (including exhaust and scrubber unit) and the Kjeltec 8100 Distillation unit consistently perform according to a specification appropriate for its routine use. Main activities in the PQ phase are:

- Preventive maintenance
- On-going verification tests

This document suggests routines to fulfill the requirements for an acceptable PQ but the final procedure should be adapted to local routines for similar equipment.

4 Definition of Test Procedures

4.1 Preventive Maintenance

Maintenance of the Kjeltec 8100 should be performed according to the instructions in manual, see User Manual Kjeltec 8100/8200 Distillation Unit, chapter 5. Maintenance. A yearly service is recommended (service agreement).

Maintenance of the Digestion block (including exhaust and scrubber) should be performed according to instruction in the user manual, see User Manual Tecator Digestor, chapter 5. Maintenance.

Dedicated Analytical Solutions

FOSS Analytical A/S
69 Stængerejalgade
DK-3400 Hillerød
Denmark

Tel +45 7010 3710
Fax +45 7010 3371
E-mail support@foss.dk
Web www.foss.dk

FOSS Analytical AB
Box 55
SE-263 21 Höganäs
Sweden

Tel +46 42 361500
Fax +46 42 340349
E-mail support@foss.se
Web www.foss.se

Customer Support, 6003 7363 / Rev. 2

ເອກສານໄໝຄວບຄຸມ

External Quality Control Program

It is recommended to participate in an external quality control program, such a proficiency program, or ring test, with equivalent sample material as analyzed within the laboratory.

Calculation and Expression of Results

$$\nu_{\text{N}} = \frac{(F_0 - F_1) \cdot V_0}{m} \times 100\%$$

Where:
 ν_{N} is the nitrogen content of the sample, expressed as a percentage by mass.

V_0 is the numerical value of the volume of the hydrochloric acid standard volumetric solution used in the sample test, in milliliters, expressed to the nearest 0.05 ml

V_1 is the numerical value of the volume of the hydrochloric acid standard volumetric solution used in the blank test, in milliliters, expressed to the nearest 0.05 ml.

N is the numerical value of the exact normality of the hydrochloric acid standard volumetric solution, expressed to four decimal places.

m is the numerical value of the mass of the test portion, in milligrams, expressed to the nearest 1 mg for sample weights >1 g or to the nearest 0.1 mg for sample weights <1 g.

5 Maintenance

5.1 Maintenance Kjeltec™ 8100

See instructions in User Manual - Kjeltec 8100/8200, chapter 5 Maintenance.

5.2 Maintenance Tecator™ Digestor

See instructions in User Manual - Tecator Digestor, chapter 5 Maintenance.

6 The Maintenance Record Charts

This record charts are provided to assist you in keeping your system in good working order. Please make copies and use them regularly as they can often help us to help you in the unlikely event a system malfunction.

Customer Support, 6003 7363 / Rev. 2

ເອກສານໄໝຄວບຄຸມ

ເອກສານໄໝຄວບຄຸມ



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-29 FAX. 0-2719-9484



Cert. No.: 23TM729
Page : 1 of 3

Certificate of Calibration

Equipment : Incubator
Manufacturer : Memmert
Model : IPP 260
Serial No. : V618.0033
ID No. : UAE.MIC.021/2561
Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
 3 Soi Udomsuk 41, Sukhumvit Road,
 Bangchak, Phrakanong,
 Bangkok 10260
Location : Microbiology Laboratory (302)
Received Order : 27 April 2023
Calibration Date : 27 April 2023
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %
Calibrated by : Tawatchai Pama

Approved by : *Mala.*
 Approved Signatory
 Pornthippe Tameyakul
 Malee Butkrua
 Suwit Imjai

Issue Date : 11 May 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
 Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services

เอกสารไม่ควบคุม

Equipment : Incubator
Condition As-Received : Used Item
Reference : 2304-0461OC-7
Procedure Used :-

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD).

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument-
Instrument Model Serial No. Cert. No. Due Date
 1) Data Acquisition 34972A MY57013711 22LM93 02 Jul 2023

2. This certificate is valid only to the item calibrated on date and place of calibration.


3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Close

Environment during calibration		
	Beginning	Finished
Temp. (°C)	20	21
REL.Humid. (%)	72	77
AC Supply (Volt)	230	231



Probe Installation Details :
 a = 10 cm D = 0.50 m
 b = 10 cm W = 0.64 m
 c = 10 cm H = 0.80 m
 Capacity = 0.26 m³

Position :	Ref. Std. ID No.:
1	18-18RTD-01
2	18-18RTD-02
3	18-18RTD-03
4	18-18RTD-04
5	18-18RTD-05
6	18-18RTD-10
7	18-18RTD-07
8	22-18RTD-08
9 (ref.)	18-18RTD-09

เอกสารไม่ควบคุม



Equipment : Incubator
Condition As-Received : Used Item
Reference : 2304-0461OC-7
Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Close

Calibration Point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Temperature stability (± °C)	Temperature uniformity (°C)	Overall Variation (°C)	Coverage Factor k
22.0	22.0	22.0	0.058	0.11	0.19	2
44.0	44.0	44.0	0.066	0.50	0.87	2

Calibration Point (°C)	Measured Temperature (°C)								Uncertainty (± °C)
	Position								
1	2	3	4	5	6	7	8	9 (ref.)	(± °C)
22.0	22.009	22.038	21.971	22.005	22.004	22.009	21.941	21.959	0.30
44.0	44.393	44.447	44.029	44.204	43.899	43.895	43.637	43.923	0.30

Average* : The average of 30 values in each position.

Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor.

Temperature uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-00-

เอกสารไม่ควบคุม

TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
 CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
 534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
 TEL. 0-2717-3000-29 FAX. 0-2719-9484



Cert. No.: 23TM727

Page : 1 of 3

Certificate of Calibration

Equipment : Incubator
Manufacturer : Memmert
Model : IF 75
Serial No. : D317.0305
ID No. : UAE.MIC.022/2561
Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
 3 Soi Udomsuk 41, Sukhumvit Road,
 Bangchak, Phrakanong,
 Bangkok 10260
Location : Microbiology Laboratory (302)
Received Order : 27 April 2023
Calibration Date : 27 - 28 April 2023
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %
Calibrated by : Tawatchai Pama

Approved by : *Mala.*
 Approved Signatory
 Pornthippe Tameyakul
 Malee Butkrua
 Suwit Imjai

Issue Date : 11 May 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
 Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services

เอกสารไม่ควบคุม



Equipment : Incubator
Condition As-Received : Used Item
Reference : 2304-0461OC-4
Result of Calibration :- (*) Without Adjustment

Cert. No.: 23TM727
Page : 3 of 3

Calibration Point	UUC* Setting (°C)	UUC* Reading (°C)	Temperature stability (± °C)	Temperature uniformity (°C)	Overall Variation (°C)	Coverage Factor k
44.0	44.0	44.0	0.055	0.42	0.52	2

Calibration Point (°C)	Measured Temperature (°C)									Uncertainty (± °C)
	Position									
44.0	1	2	3	4	5	6	7	8	9 (ref.)	(± °C)
44.0	43.993	44.061	44.107	44.073	44.067	44.067	43.938	43.687	44.060	0.30

Average* : The average of 30 values in each position.

Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor.

Temperature uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-00-

เอกสารไม่ควบคุม



Equipment : Incubator
Condition As-Received : Used Item
Reference : 2304-0461OC-4
Procedure Used :-

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD).

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-
Instrument Model Serial No. Cert. No. Due Date
1) Data Acquisition 34972A MY57013711 22LM93 02 Jul 2023

2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Close

Environment during calibration		
	Beginning	Finished
Temp. (°C)	25	24
REL.Humid. (%)	76	80
AC Supply (Volt)	231	231

Position :	Ref. Std. ID No.:
1	18-18RTD-01
2	18-18RTD-02
3	18-18RTD-03
4	18-18RTD-04
5	18-18RTD-05
6	18-18RTD-10
7	18-18RTD-07
8	22-18RTD-08
9 (ref.)	18-18RTD-09

Probe Installation Details : Dimension of Chamber :
a = 5.0 cm D = 0.33 m
b = 5.0 cm W = 0.40 m
c = 5.0 cm H = 0.56 m
Capacity = 0.074 m³



Equipment : Incubator
Condition As-Received : Used Item
Reference : 2304-0461OC-5
Result of Calibration :- (*) Without Adjustment
Function of UUC* : Temperature Source
Fresh air setting : Close

Cert. No.: 23TM765
Page : 3 of 3

Calibration Point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Temperature stability (± °C)	Temperature uniformity (°C)	Overall Variation (°C)	Coverage Factor k
37.0	37.0	37.0	0.070	0.28	0.39	2

Calibration Point (°C)	Measured Temperature (°C)									Uncertainty (± °C)
	Position									
37.0	1	2	3	4	5	6	7	8	9 (ref.)	(± °C)
37.0	37.164	37.118	37.079	37.121	36.852	37.039	36.822	36.923	36.905	0.30

Average* : The average of 30 values in each position.

Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor.

Temperature uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-00-

เอกสารไม่ควบคุม



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
53/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL: 0-2717-3000-29 FAX: 0-2719-9434



Cert. No.: 23TM765
Page : 1 of 3

Certificate of Calibration

Equipment : Incubator
Manufacturer : Memmert
Model : IN 75
Serial No. : D317.0307
ID No. : UAE.MIC.023/2561
Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
3 Soi Udomsuk 41, Sukhumvit Road,
Bangchak, Phrakanong,
Bangkok 10260
Location : Microbiology Laboratory (302)
Received Order : 27 April 2023
Calibration Date : 27 April 2023
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %
Calibrated by : Preecha Hlahib

Approved by : Maliee Approved Signatory
(✓) Pormtipha Tameyakul
(✓) Maliee Butkrea
(✓) Suwit Imjai

Issue Date : 11 May 2023

The Uncertainties are for a confidence probability of approximately 95%
This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services

เอกสารไม่ควบคุม



Equipment : Incubator
Condition As-Received : Used Item
Reference : 2304-0461OC-5
Procedure Used :-

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD).

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34972A	MY59003411	22LM165	26 Nov 2023


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Close



Probe Installation Details : Dimension of Chamber :

a = 5.0 cm	D = 0.32 m
b = 5.0 cm	W = 0.42 m
c = 5.0 cm	H = 0.56 m
Capacity = 0.075 m³	

Environment during calibration		
	Beginning	Finished
Temp. (°C)	23	22
REL.Humid. (%)	69	73
AC Supply (Volt)	220	221

Position :	Ref. Std. ID No.:
1	20RTD-2/1
2	20RTD-2/2
3	20RTD-2/3
4	20RTD-2/4
5	20RTD-2/5
6	20RTD-2/6
7	20RTD-2/7
8	20RTD-2/8
9 (ref.)	20RTD-2/9

เอกสารไม่ควบคุม

- ๔๔๕๕๒๐๒๐ -



Equipment : Water Bath
Condition As-Received : Used Item
Reference : 2302-0295OC-2
Procedure Used :-

Calibration were conducted using in-house calibration procedure CP-OT04 according to direct measurement method with Data Acquisition which connected with Industrial Platinum Resistance Thermometer (IPRT).

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34972A	MY59003411	22LM165	26 Nov 2023


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

	Environmental		AC Voltage Supply (Volt)
	(°C)	(%R.H.)	
Beginning of Calibration	22	65	231
Finished of Calibration	23	61	231



Front

เอกสารไม่ควบคุม



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
3344 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10230
TEL: 0-2717-3090-29 FAX: 0-2719-9484



Cert. No.: 23TM193
Page : 1 of 3

Certificate of Calibration

Equipment : Water Bath
Manufacturer : Memmert
Model : WNE 14
Serial No. : L416.0606
ID No. : UAE.MIC.002/2580
Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
3 Soi Udomsuk 41, Sukhumvit Road,
Bangchak, Phrakanong,
Bangkok 10260
Location : Microbiology Laboratory
Received Order : 15 February 2023
Calibration Date : 15 February 2023
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %
Calibrated by : Suwit Imjai

Approved by : Approved Signatory
 Pornthippa Tameyakul
 Malee Butkrua

Issue Date : 24 February 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written approval of the head of Corporate Services 3: Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม



Equipment : Water Bath
Condition As-Received : Used Item
Reference : 2302-0295OC-2
Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Cert. No.: 23TM193
Page : 3 of 3

Calibration point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Average* Standard Reading (°C)				
			1	2	3	4	5 (ref.)
44.5	44.5	44.5	44.453	44.437	44.428	44.477	44.459
44.5	0.079	0.038	0.15	2			
Calibration point (°C)			Uniformity	Stability	Uncertainty	Coverage Factor	
			(± °C)	(± °C)	(± °C)	k	

Average* : The average of 30 values in each position.

Uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Stability : One-half of the greatest maximum difference of measured temperature at any one probe.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-00-

เอกสารไม่ควบคุม



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
334/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-29 FAX. 0-2719-9484



Certificate of Calibration

Equipment : Water Bath
Manufacturer : Memmert
Model : WNE 14
Serial No. : L416.0612
ID No. : UAE.MIC.003/2560
Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
 3 Soi Udomsuk 41, Sukhumvit Road,
 Bangkok, Phrakanong,
 Bangkok 10260
Location : Microbiology Laboratory
Received Order : 15 February 2023
Calibration Date : 15 February 2023
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %
Calibrated by : Suwit Imjai

Approved by :
 Approved Signatory
 () Pornthippa Tameyakul
 (✓) Malee Butkrua

Issue Date : 24 February 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
 Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม



Equipment : Water Bath
Condition As-Received : Used Item
Reference : 2302-0295OC-3
Result of Calibration :- (*) Without Adjustment
Function of UUC* : Temperature Source

Calibration point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Average* Standard Reading (°C)				
			Position				
			1	2	3	4	5 (ref.)
44.5	44.5	44.6	44.520	44.509	44.498	44.552	44.530
Calibration point (°C)	Uniformity (°C)	Stability (± °C)	Uncertainty (± °C)	Coverage Factor <i>k</i>			
44.5	0.077	0.037	0.15	2			

Average* : The average of 30 values in each position.

Uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Stability : One-half of the greatest maximum difference of measured temperature at any one probe.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-00-

เอกสารไม่ควบคุม



Equipment : Water Bath
Condition As-Received : Used Item
Reference : 2302-0295OC-3
Procedure Used :-

Calibration were conducted using in-house calibration procedure CP-OT04 according to direct measurement method with Data Acquisition which connected with Industrial Platinum Resistance Thermometer (IRT).

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34972A	MY59003411	22LM165	26 Nov 2023


2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

	Environmental		AC Voltage Supply
	(°C)	(%R.H.)	(Volt)
Beginning of Calibration	22	65	231
Finished of Calibration	22	63	230



Front

Position :	Ref. Std. ID No.:
1	4804539-001
2	4804539-002
3	4804539-003
4	4804539-004
5(ref.)	4804539-005

เอกสารไม่ควบคุม



Equipment : Water Bath
Condition As-Received : Used Item
Reference : 2302-0295OC-3
Result of Calibration :- (*) Without Adjustment
Function of UUC* : Temperature Source

Calibration point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Average* Standard Reading (°C)				
			Position				
			1	2	3	4	5 (ref.)
44.5	44.5	44.6	44.520	44.509	44.498	44.552	44.530
Calibration point (°C)	Uniformity (°C)	Stability (± °C)	Uncertainty (± °C)	Coverage Factor <i>k</i>			
44.5	0.077	0.037	0.15	2			

Average* : The average of 30 values in each position.

Uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Stability : One-half of the greatest maximum difference of measured temperature at any one probe.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-00-



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
334/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-29 FAX. 0-2719-9484



Cert. No.: 23TM374
Page: 1 of 3

Certificate of Calibration

Equipment : Water Bath

Manufacturer : Memmert

Model : WNE 14

Serial No. : L414.1407

ID No. : UAE.MIC.006/2558

Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
3 Soi Udomsuk 41, Sukhumvit Road,
Bangchak, Phrakanong,
Bangkok 10260

Location : Microbiology Laboratory

Received Order : 11 April 2023

Calibration Date : 11 April 2023

Ambient Temperature : (26 ± 10) °C

Relative Humidity : (50 ± 30) %

Calibrated by : Krisda Malee

Approved by :
Approved Signatory

(✓) Pornthippa Tameyakul
 (✓) Malee Butkrua
 () Suwit Imjai

Issue Date : 24 April 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written

Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม



Equipment : Water Bath
 Condition As-Received : Used Item
 Reference : 2304-0155OC-3
 Result of Calibration :- (*) Without Adjustment
 Function of UUC* : Temperature Source

Cert. No.: 23TM374
 Page : 3 of 3

Calibration point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Average* Standard Reading (°C)					Uncertainty (± °C)
			Position					
44.5	44.5	44.508	44.466	44.456	44.478	44.483	0.15	

Calibration point (°C)	Uniformity (°C)	Stability (°C)	Coverage Factor k
44.5	0.065	0.031	2

Average* : The average of 30 values in each position.

Uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Stability : One-half of the greatest maximum difference of measured temperature at any one probe.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-00-

เอกสารไม่ควบคุม

a 1158268



Equipment : Water Bath
 Condition As-Received : Used Item
 Reference : 2304-0155OC-3
 Procedure Used :-

Calibration were conducted using in-house calibration procedure CP-OT04 according to direct measurement method with Data Acquisition which connected with Industrial Platinum Resistance Thermometer (IPRT).

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument
 1) Data Acquisition
 34972A
 MY59003411

2. This certificate is valid only to the item calibrated on date and place of calibration.


3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Heat transfer medium used : Water

Beginning of Calibration	Environmental		AC Voltage Supply (Volt)
	(°C)	(%R.H.)	
Finished of Calibration	26	55	220
	26	56	221



Front

เอกสารไม่ควบคุม

a 1158268



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
 CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
 354/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10230
 TEL. 0-2717-3000-29 FAX. 0-2719-9484



Cert. No.: 23TM377
 Page : 1 of 3

Certificate of Calibration

Equipment : Water Bath
 Manufacturer : Memmert
 Model : WNE 14
 Serial No. : L414,1410
 ID No. : UAE.MIC.007/2558
 Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
 3 Soi Udomsuk 41, Sukhumvit Road,
 Bangkok, Phra Khanong,
 Bangkok 10260
 Location : Microbiology Laboratory (302)
 Received Order : 11 April 2023
 Calibration Date : 12 April 2023
 Ambient Temperature : (26 ± 10) °C
 Relative Humidity : (50 ± 30) %
 Calibrated by : Preecha Hlahib

Approved by : Malu.
 Approved Signatory
 Pornthippa Tameyakul
 Malee Butkrua
 Suwit Imjai

Issue Date : 24 April 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม

A 0053356

Cert. No.: 23TM377
 Page : 3 of 3

Equipment : Water Bath

Condition As-Received : Used Item

Reference : 2304-0155OC-4

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Calibration point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Average* Standard Reading (°C)					Uncertainty (± °C)
			Position					
44.5	44.5	44.5	44.494	44.459	44.477	44.507	44.498	0.15

Calibration point Uniformity Stability Coverage Factor k

Calibration point (°C)	Uniformity (°C)	Stability (°C)	Coverage Factor k
44.5	0.13	0.056	2

Average* : The average of 30 values in each position.

Uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Stability : One-half of the greatest maximum difference of measured temperature at any one probe.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-00-

Malu.

เอกสารไม่ควบคุม

a 1158266



Equipment : Water Bath
 Condition As-Received : Used Item
 Reference : 2304-0155OC-4

Cert. No.: 23TM377
 Page : 2 of 3

Procedure Used :-
 Calibration were conducted using in-house calibration procedure CP-OT04 according to direct measurement method with Data Acquisition which connected with Industrial Platinum Resistance Thermometer (IPRT).

The temperature scale used was based on ITS-90.


Condition of this result of calibration

- Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34972A	MY49001451	23LM27	25 Feb 2024
- This certificate is valid only to the item calibrated on date and place of calibration.
- This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment
Function of UUC* : Temperature Source
Heat transfer medium used : Water

	Environmental		AC Voltage Supply
	(°C)	(%R.H.)	(Volt)
Beginning of Calibration	27	65	220
Finished of Calibration	30	70	221



Front

Position :	Ref. Std. ID No.:
1	N37P301419
2	N37P300732
3	N37P301420
4	N37P301421
5(ref.)	N37P301425

Malu.
 เอกสารไม่ควบคุม
 a 1158267



Equipment : Electronic Balance
 Condition As-Received : Used Item
 Reference : 2304-0016OC-1

Cert. No.: 23MM150
 Page : 2 of 3

Procedure used :-
 Calibration were conducted using in-house calibration procedure CP-OB01 according to direct measurement method against standard weight.

Condition of this result of calibration

- Reference standard instruments:-

Instruments	Model	Serial No.	ID No.	Test report No.	Due date
1) Standard Weight Set (E2)	15884	24053	70RC007	MM-0010-22	20 Jan 2024
2) Standard Weight (E2)	158471	-	70RC197	MM-0059-22	25 Apr 2024
- This certificate is valid only to the item calibrated on date and place of calibration.
- This result of calibration was made on requested at the point specified by customer.
- This certificate is not certified for any commercial transaction.
- This certification is traceable to the International System of Unit.

Result of calibration () Without Adjustment (*) After Adjustment by Internal Calibration

Range capacity : 0 g to 620 g Resolution 0.001 g

Before Adjustment :

Applied Weight	Balance Reading	Correction	Measurement Uncertainty	Coverage Factor
(g)	(g)	(g)	(± mg)	(k)
300	299.999	+0.001	1.4	2.07
600	599.997	+0.003	1.7	2.09

After Adjustment :

1. Determination of the standard deviation of weighing machine (n = 10)

Applied Weight	Standard Deviation of Reading (g)
300	0.0006
600	0.0008

เอกสารไม่ควบคุม



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
 CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
 334/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
 TEL: 0-2717-3006-29 FAX: 0-2719-9484



Cert. No.: 23MM150
 Page.: 1 of 3

Certificate of Calibration

Equipment :	Electronic Balance
Manufacturer :	Mettler Toledo
Model :	MS603S /01
Serial No. :	B007010311
ID No. :	UAE.TOX.008/2553
Submitted by :	United Analyst and Engineering Consultant Co.,Ltd. 3 Soi Udomsuk 41, Sukhumvit Road, Bangchak, Phrakanong, Bangkok 10260
Location :	Balance Room 2
Received order :	07 April 2023
Calibration Date :	07 April 2023
Ambient Temperature :	15 °C to 40 °C
Relative Humidity :	30 % to 90 %
Calibrated by :	Suwit Imjai
Approved by :	Malu. Malee Butkrua

Issue Date :


10 April 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written approval of the head of Corporate Services 3: Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม

Cert. No.: 23MM150
 Page : 3 of 3



Maximum difference between off-center and central loading (g) 0.002

3. Departure from nominal value					
Applied Weight	Balance Reading	Correction	Measurement Uncertainty	Coverage Factor	
(g)	(g)	(g)	(± mg)	(k)	
Unload	0.000	0.000	1.3	2.09	
0.5	0.499	+0.001	1.3	2.09	
1	0.999	+0.001	1.3	2.09	
50	50.000	0.000	1.3	2.09	
100	99.999	+0.001	1.3	2.09	
150	149.999	+0.001	1.3	2.09	
200	199.999	+0.001	1.3	2.09	
300	299.999	+0.001	1.4	2.07	
400	399.999	+0.001	1.6	2.11	
500	500.000	0.000	1.7	2.11	
600	599.999	+0.001	1.7	2.09	

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

-00-

เอกสารไม่ควบคุม

Calibration Certificate

Certificate No.: 2304203-001-01
Client name: UNITED ANALYST AND ENGINEERING CONSULTANT CO., LTD.
Address: 3 Soi Udomsuk 41, Sukhumvit Road, Bangchak, Prakanong, Bangkok 10260

Page 1 of 3

Equipment: Autoclave
Manufacturer: ALP
Model: CL-40L
Serial No.: 807298
ID No.: UAE.MIC.019/2560
Order No.: 2304203
Operation No.: 2304203-001
Date of Receipt: 10 August 2023
Date of Calibration: 10 August 2023


Calibrated by Mr.Worapob Sooktong **Approved by** 
 Scientist (Mr.Pheraphat Tuanjit)
 Manager, Division of Calibration Laboratory

Date of Issue: 15 August 2023 **Responsible for the Technical Management Team**

The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the National Food Institute.

F-CS-009 Revision: 01 Date: 20-04-65



Calibration Report

Certificate No.: 2304203-001-01
Equipment: Autoclave
 Model: CL-40L Serial No.: 807298
 Resolution: 1 °C ID No.: UAE.MIC.019/2560
 Manufacturer: ALP


Date of Calibration: 10 August 2023

Page 3 of 3

Calibration point: 121 °C

Calibration result:

Calibration Condition	Temperature (°C)	Relative Humidity (%)	Line Voltage (Volt)
Min	27.0	63.5	223.3
Max	28.3	67.3	225.9



Standardized Position:
 Sensor 1 = Attached to the base temperature probe.
 Sensor 2 = In the upper half of the chamber.
 Sensor 3 = In the chamber drain, within 200 mm.

Table 1 : Reporting of Temperature

Calibration Point (°C)	Measured Temperature (°C) @ Sensor No. (Sensor No.2 is REF)			Uncertainty ± (°C)
	Std. # 1	Std. # 2 (Ref)	Std. # 3	
121	121.68	121.70	121.66	0.66

Table 2 : Reporting of Characterization Result

UUC* Setting (°C)	UUC* Reading				Stability ± (°C)	Uniformity (°C)	Overall Variation (°C)
	Min (°C)	Max (°C)	Average (°C)	MPa			
121	121	121	121	0.10	0.11	0.12	0.23

Note

The quoted uncertainty include " Stability " and " Loading effect (20% of Uniformity)"

UUC* = Unit Under Calibration

Stability = One-half of the greatest maximum difference of measured temperatures at any one sensors,

for at least half an hour after reaching steady state.

Uniformity = The maximum difference of measured temperatures at any sensors and the measured

temperature at the reference location which are observed at the same time.


Overall Variation = The difference of the maximum and minimum measured temperatures throughout observation time.

The report uncertainty of measurement was based on standard uncertainty multiplied by coverage factor k= 2, providing a level of confidence of approximately 95 %.

----- End -----



F-CS-012 Revision: 01 Date: 20-04-65



Calibration Report

Certificate No.: 2304203-001-01
Equipment: Autoclave
 Model: CL-40L Serial No.: 807298
 Resolution: 1 °C ID No.: UAE.MIC.019/2560
 Manufacturer: ALP

Date of Calibration: 10 August 2023

Page 2 of 3

Location: 301, UNITED ANALYST AND ENGINEERING CONSULTANT CO., LTD.
Environment Condition: Ambient Temperature (28 ± 1) °C
 Relative Humidity (65 ± 2) %
 Line Voltage (225 ± 1) Volt

Condition of this results of Calibration:

- This instrument was calibrated by insert 3 standard temperature recorder with RTD into its autoclave and calibration according to W-TE-01B based on BS 2646-1(2021) : Autoclaves for sterilization in laboratories Design, construction, safety and performance Specification.
- The temperature scale used was based on ITS - 90.
- All data show below were final values and the initial data may be obtained upon request.

2. Reference Standard Instrument :

Instrument	Model	Serial No.	Certificate No.	Due Date	Through
Digital Thermometer with RTD (Data Logger)	HiTemp140-2	S25601	NC-22-11-22-176	9-Nov-23	MADGETECH INC.
	HiTemp140-2	S25602	NC-22-11-22-175	9-Nov-23	MADGETECH INC.
	HiTemp140-2	R54918	TE 660383-01	8-Apr-24	NATIONAL FOOD INSTITUTE

3. This certificate is traceable to International System of Units (SI Units).

4. This certificate was certified only for the instrument we calibrated.

5. The result of calibration was found accurate as shown on date and place of calibration only.

6. This standard does not apply to sterilizers or disinfectors used for medical, dental, pharmaceutical.


7. Condition of Calibrated item : Good

UUC Description : Setting program function sterilization : STERILIZE/NORMAL

Time of sterilization 15 Minute At 121 °C

8. Result of Calibration :
- | |
|--|
| <input checked="" type="checkbox"/> Without adjustment |
| <input type="checkbox"/> After adjustment |

F-CS-012 Revision: 01 Date: 20-04-65



Cert. No.: 22TM1121
Page.: 1 of 3

Certificate of Calibration

Equipment : Autoclave
Manufacturer : ALP
Model : CL-40L
Serial No. : 807298
ID No. : UAE.MIC.019/2560

Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
 3 Soi Udomsuk 41, Sukhumvit Road, Bangkok 10260

Location : 301 Room

Received Order : 11 July 2022

Calibration Date : 11 July 2022

Ambient Temperature : (26 ± 10) °C

Relative Humidity : (50 ± 30) %

Calibrated by : Preecha Hihib

Approved by : 
 Approved Signatory

(/) Pornthippa Tameyakul

(/) Malee Butkrua


(/) Suwit Imjai

Issue Date : 18 July 2022

The Uncertainties are for a confidence probability of approximately 95%.

This certificate may not be reproduced other than in full, except with the prior written approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม





Equipment : Autoclave
Condition As-Received : Used Item
Reference : 2207-0245OC-7

Procedure Used :-

Calibration were conducted using in-house calibration procedure CP-OT03 according to direct measurement method with Data Acquisition which connected with Thermocouple Type T

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34970A	MY44060450	22LM46	28 Mar 2023

2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.


4. This result of calibration covers laboratory autoclaves for the sterilization of goods and material which could be infected with organisms categorized as Hazard Group 1, 2 and 3**

(**) Categorization of pathogens according to hazard and categories of containment, second edition, 1990) It does not cover autoclaves for use with material infect with organisms in Hazard Group 4, for which complete containment and sterilization of infected condensate is considered to be essential.

This result of calibration does not apply to sterilizers or disinfectors used for medical, dental, pharmaceutical or veterinary purposes which are directly concerned with patient care, or those used for fabrics subjected to sterilization which are required to be dry at the end of cycle.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source



Environmental		
(°C)	(%R.H.)	(Volt)
Beginning of Calibration	29	49
Finished of Calibration	32	48
		220

Position	Description	Ref. Std. ID No.:
1 =	Center of chamber	22-14TC-01
2 =	Temperature sensor	22-14TC-02
3 =	Exhaust port	22-14TC-03

เอกสารไม่ควบคุม

Equipment : Autoclave
Condition As-Received : Used Item
Reference : 2207-0245OC-7

Result of Calibration :- (*) Without Adjustment

Operating parameter Set : Temperature = 115 °C
Sterilization period = 15 minute

UUC* Setting (°C)	UUC* Reading (°C)	Position	Average* Standard Reading (°C)	Stability (± °C)	Pressure Reading (MPa)	Uncertainty (± °C)	Coverage Factor k
116	116	1	116.523				
		2	116.566	0.14	0.08	0.90	2
		3	116.440				

Operating parameter Set : Temperature = 121 °C
Sterilization period = 30 minute

UUC* Setting (°C)	UUC* Reading (°C)	Position	Average* Standard Reading (°C)	Stability (± °C)	Pressure Reading (MPa)	Uncertainty (± °C)	Coverage Factor k
122	122	1	122.503				
		2	122.637	0.19	0.12	0.91	2
		3	122.558				

Average* : The average of 30 values in each position.

Stability : One-half of the greatest maximum difference of measured temperature at any one probe.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-00-

เอกสารไม่ควบคุม



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
334/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL: 0-2717-3000-29 FAX: 0-2719-9484



Cert. No.: 23TM763
Page : 1 of 3

Certificate of Calibration

Equipment : Autoclave

Manufacturer : ALP

Model : CL-40L

Serial No. : 808763

ID No. : UAE.MIC.026/2563

Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
3 Soi Udomsuk 41, Sukhumvit Road,
Bangchak, Phrakanong,
Bangkok 10260

Location : Microbiology Laboratory (301)

Received Order : 27 April 2023

Calibration Date : 27 April 2023

Ambient Temperature : (26 ± 10) °C

Relative Humidity : (50 ± 30) %

Calibrated by : Preecha Hiahib

Approved by : 
Approved Signatory

() Pornthippa Tameyakul
(✓) Malee Bulkruea
() Suwit Imjai

Issue Date : 11 May 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3: Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม

A 0053944

Equipment : Autoclave
Condition As-Received : Used Item
Reference : 2304-0461OC-2

Procedure Used :-

Calibration were conducted using in-house calibration procedure CP-OT03 according to direct measurement method with Data Acquisition which connected with Thermocouple Type T

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34972A	MY59003411	22LM165	26 Nov 2023

2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.


4. This result of calibration covers laboratory autoclaves for the sterilization of goods and material which could be infected with organisms categorized as Hazard Group 1, 2 and 3**

(**) Categorization of pathogens according to hazard and categories of containment, second edition, 1990) It does not cover autoclaves for use with material infect with organisms in Hazard Group 4, for which complete containment and sterilization of infected condensate is considered to be essential.

This result of calibration does not apply to sterilizers or disinfectors used for medical, dental, pharmaceutical or veterinary purposes which are directly concerned with patient care, or those used for fabrics subjected to sterilization which are required to be dry at the end of cycle.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source



Environmental		
(°C)	(%R.H.)	(Volt)
Beginning of Calibration	27	60
Finished of Calibration	27	58
		220

Position	Description	Ref. Std. ID No.:
1 =	Center of chamber	18-20TC-04
2 =	Temperature sensor	18-20TC-05
3 =	Exhaust port	18-20TC-06

เอกสารไม่ควบคุม

a 1159968



Equipment : Autoclave
 Condition As-Received : Used Item
 Reference : 2304-0461OC-2
 Result of Calibration :- (*) Without Adjustment
 Function of UUC* : Temperature Source

Cert. No.: 23TM763
 Page : 3 of 3

Operating parameter Set : Temperature = 115.0 °C
 Sterilization period = 15 minute

UUC* Setting (°C)	UUC* Reading (°C)	Position	Average* (°C)	Stability (± °C)	Pressure Reading (MPa)	Uncertainty (± °C)	Coverage Factor k
115.0	115.0	1	115.213	0.22	0.08	0.75	2
		2	115.166				
		3	115.260				

Operating parameter Set : Temperature = 121.0 °C
 Sterilization period = 30 minute

UUC* Setting (°C)	UUC* Reading (°C)	Position	Average* (°C)	Stability (± °C)	Pressure Reading (MPa)	Uncertainty (± °C)	Coverage Factor k
121.0	121.0	1	121.260	0.29	1.1	0.75	2
		2	121.224				
		3	121.284				

Average* : The average of 30 values in each position.

Stability : One-half of the greatest maximum difference of measured temperature at any one probe.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-00-

Malu.

เอกสารไม่ควบคุม

a 1159967



Equipment : Refrigerator
 Condition As-Received : Used Item
 Reference : 2204-0016OC-2

Cert. No.: 22TM562
 Page.: 2 of 3

Procedure Used :-

Calibration were conducted using calibration procedure CP-OT02 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD).

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Data Acquisition	34970A	MY44067817	21LM10	20 Jul 2022

2. This certificate is valid only to the item calibrated on date and place of calibration.


3. This certification is traceable to the International System of Unit.

Result of Calibration :- (*) Without Adjustment

Function of UUC* : Temperature Source

Fresh air setting : Not Available

Environment during calibration		
	Beginning	Finished
Temp. (°C)	25	25
REL.Humid. (%)	56	60
AC Supply (Volt)	220	220



Probe Installation Details : Dimension of Chamber :
 a = 10 cm D = 0.45 m
 b = 10 cm W = 0.45 m
 c = 10 cm H = 1.0 m
 Capacity = 0.20 m³

Position :	Ref. Std. ID No.:
1	15RTD2/11
2	15RTD2/12
3	15RTD2/13
4	15RTD2/14
5	15RTD2/15
6	15RTD2/16
7	15RTD2/17
8	15RTD2/18
9 (ref.)	15RTD2/19

เอกสารไม่ควบคุม



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
 CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
 534/8 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
 TEL. 0-2717-3000-27 FAX. 0-2719-6484



HAC-MRA
 NSC-TIS-T101/T20
 CALIBRATION 99%

Cert. No.: 22TM562
 Page.: 1 of 3

Certificate of Calibration

Equipment : Refrigerator

Manufacturer : Sanyo

Model : SBC-337KD(GYN)

Serial No. : 71100607

ID No. : UAE.MIC.003/2021

Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
 3 Soi Udomsuk 41, Sukhumvit Road,
 Bangchak, Phrakanong,
 Bangkok 10260

Location : Microbiology Laboratory

Received Order : 7 April 2022

Calibration Date : 7 April 2022

Ambient Temperature : (26 ± 10) °C

Relative Humidity : (50 ± 30) %

Calibrated by : Prawit Sodaivitchit

Approved by : Malu.

Approved Signatory

() Pornthippe Tamayakul

(✓) Malie Butkrea

() Suwit Imjai

Issue Date : 18 April 2022

The Uncertainties are for a confidence probability of approximately 95%.

This certificate may not be reproduced other than in full, except with the prior written approval of the head of Corporate Services 3: Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม



Equipment : Refrigerator
 Condition As-Received : Used Item
 Reference : 2204-0016OC-2
 Result of Calibration :- (*) Without Adjustment
 Function of UUC* : Temperature Source
 Fresh air setting : Not Available

Cert. No.: 22TM562
 Page.: 3 of 3

Calibration Point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Temperature stability (± °C)	Temperature uniformity (°C)	Overall Variation (°C)	Uncertainty (± °C)	Coverage Factor k	
4	4	-	3.1	1.8	7.5	3.7	2	
Measured Temperature (°C)								
			Position					
4	6.043	5.013	4.454	4.485	4.549	4.837	4.268	4.121

Average* : The average of 30 values in each position.

Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor.

Temperature uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-00-

เอกสารไม่ควบคุม

เอกสารไม่ควบคุม

ภาคผนวก จ-2
เอกสารสอบเทียบเครื่องมือตรวจวัด





TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3 : EQUIPMENT CALIBRATION AND TESTING SERVICES

534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250

TEL. 0-2717-3000 FAX. 0-2719-9484

Cert.No.: 23TW153

Page.: 1 of 2

Certificate of Testing

Equipment :	DO Meter
Manufacturer :	YSI
Model :	Pro 20i
Serial No. :	18K104053
ID No. :	UAE.EFM.066/2562 (ENV.DO.01/62)
Received Date :	26 June 2023
Test Date :	27 June 2023
Reference :	2306-0827WSC-5
Submitted by :	United Analyst and Engineering Consultant Co.,Ltd. 3 Soi Udomsuk 41, Sukhumvit Road, Bangchak, Phrakanong, Bangkok 10260
Laboratory Condition :	Temperature (25 ± 5) °C Humidity (50 ± 20) %
Test Procedure :	In - house method : CP-CH9 by Comparison Technique with Azide Modification Method
Tested by :	Walalak Sirithean

Approved by :

Malee.

Approved Signatory

- (✓) Malee Butkruea
() Saithip Meangmai
() Warakorn Lerngagtrakul

Issue Date :

28 June 2023

เอกสารไม่ควบคุม



Cert.No.: 23TW153

Page.: 2 of 2

Condition of this result of calibration

1. Reference Standard Instruments :

This certification is traceable to the International System of Unit through the reference standards laboratory of Industrial Calibration Center,Technology Promotion Association (Thailand-Japan).

<u>Instruments</u>	<u>Serial No.</u>	<u>ID No.</u>	<u>Certificate No.</u>	<u>Due Date</u>
1) Burette	-	130BU10	23CG1172	22 Mar 2025
2) Balance	1126143764	140RC004	22MM50	20 Sep 2023

2. Standard Material :-

<u>Material</u>	<u>Manufacturer</u>	<u>Lot.No.</u>	<u>Assay</u>
Sodium Thiosulfate pentahydrate	Merck	AM1763316	100.2%

Result : Dissolved Oxygen Meter Adjustment With Air 100 %

Dissolved Oxygen Probe No.: 18K100663

Titration Method (Azide Modification Method) (mg/L)	DO Meter Reading (mg/L)	Standard Deviation (mg/L)
8.20	8.19	0.0055

This report was certified only for the instrument we tested. It is allowable to use for study the system efficiency, The environmental impact control and present to organization it may concerned. Intend to use for advertising and referral purpose is prohibited. This report may not be reproduced other in full, without written approval of the laboratory

-o0o-

เอกสารมาความ
Mal.



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-29 FAX. 0-2719-9484



Cert. No.: 23LM112
Page.: 1 of 2

Certificate of Calibration

Equipment :	DO Meter with Sensor
Manufacturer :	YSI
Model :	Pro 20i
Serial No. :	18K104053
ID No. :	UAE.EFM.066/2562(ENV.DO.01/62)
Submitted by :	United Analyst and Engineering Consultant Co.,Ltd. 3 Soi Udomsuk 41, Sukhumvit Road, Bangchak, Phrakanong, Bangkok 10260
Location :	TPA On Site Calibration Laboratory
Received Order :	26 June 2023
Calibrated Date :	29 June 2023
Ambient Temperature :	(26 ± 10) °C
Relative Humidity :	(50 ± 30) %
AC Line Voltage :	(220 ± 22) V
Calibrated by :	Tawatchai Parna
Approved by :	 Malee Butkruea Approved Signatory
(<input type="checkbox"/>) Pornthippa Tameyakul (<input checked="" type="checkbox"/>) Malee Butkruea (<input type="checkbox"/>) Suwit Imjai	
Issue Date :	30 June 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม



Equipment : DO Meter with Sensor

Cert. No.: 23LM112

Condition As-Received : Used Item

Page.: 2 of 2

Reference : 2306-0827WSC-6

Procedure Used :-

Calibration were conducted using in-house calibration procedure CP-OT01 according to comparison with Industrial Platinum Resistance Thermometer (IPRT) into Temperature Bath.

The temperature scale used was based on ITS-90.

Condition of this result of calibration

1. Reference standard instrument:-

<u>Instrument</u>	<u>Serial No.</u>	<u>Cert. No.</u>	<u>Traceable</u>	<u>Due Date</u>
1) Digital Thermometer	2188080	22I1285	TPA	21 Oct 2023

2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.

Remark : TPA : Technology Promotion Association (Thailand - Japan)

Result of Calibration :- (*) Without Adjustment

Function : Temperature measurement.

This instrument was connected with temperature sensor, S/N.: 18K100663

<u>Calibration Point</u> (°C)	<u>Immersion Depth</u> (mm)	<u>Standard Temperature</u> (°C)	<u>UUC* Reading</u> (°C)	<u>Error</u> (°C)	<u>Uncertainty</u> (± °C)	<u>Coverage Factor</u> k
25.0	120	24.998	24.7	-0.298	0.16	2.00
30.0	120	30.007	29.7	-0.307	0.16	2.00
35.0	120	35.005	34.7	-0.305	0.16	2.00

UUC* : Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-00-

เอกสารมีความถูกต้อง^{มี}
เอกสารนี้ควบคุม



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-27 FAX. 0-2719-9484



Cert.No.: 22CH1728

Page.: 1 of 3

Certificate of Calibration

Equipment : pH Meter
Manufacturer : EcoSense
Model : pH100A
Serial No. : JC03354
ID No. : UAE.EFM.063/2562(ENV.pH03/62)
Condition As-Received: Used Item
Received Date : 20 December 2022
Calibration Date : 21 December 2022
Reference : 2212-0568WSC-1
Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
3 Soi Udomsuk 41, Sukhumvit Road,
Bangchak, Phrakanong, Bangkok 10260
Ambient Temperature : (25 ± 2.5) °C
Relative Humidity : (50 ± 15) %
Calibration Procedure :
In - house method :
- CP-CH5 by direct measurement with standard voltage calibrator and direct measurement with certified reference material (CRM)
- CP-CH8 by comparison with standard thermometer
Calibrated by : Warakorn Lerngagtrakul

Approved by :

Malee

Approved Signatory

- (✓) Malee Butkruea
() Saithip Meangmai
() Warakorn Lerngagtrakul

Issue Date :

23 December 2022

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม



Cert.No.: 22CH1728

Page.: 2 of 3

Condition of this calibration result

1. Reference Standard Instrument : -

Instrument	Serial No.	ID No.	Cert. No.	Due Date
1) Document Process Calibrator	54030049	130RC116	22E2769	24 Aug 2023
2) Ref. Standard Thermometer	2188080	130RC044	22I1285	20 Oct 2023

This certification is traceable to the International System of Unit maintained at:-

- Traceable to National Institute of Metrology (Thailand), NIMT

2. Certified Reference Materials : The measurement results are traceable to SI through CPA chem Ltd., ANSI-ASQ National Accreditation Board, Accredited No. AR-1835

Buffer Solution	Manufacturer	Lot No.	Exp. date
pH 4.008	CPA chem	826588	09 July 2024
pH 6.987	CPA chem	823322	20 June 2023
pH 10.008	CPA chem	826590	09 July 2023

3. This certificate is valid only to the item calibrated on date and place of calibration.

Calibration Results

Function : mV Measurement

Performing standard curve by Fluke at pH (4,7)(7,10)

Unit Under Calibration	Nominal Value	Standard Voltage Input	Actual Reading		Uncertainty of Measurement (±mV)	Coverage factor <i>k</i>
			pH	mV		
pH Meter S/N.: JC03354	4.00	177.48	177	4.01	0.58	2.00
	7.00	0.00	0	7.00	0.58	2.00
	7.00	0.00	0	7.00	0.58	2.00
	10.00	-177.48	-178	10.01	0.58	2.00

เอกสารไม่คงที่



Cert.No.: 22CH1728

Page.: 3 of 3

Calibration Results**Function : pH Measurement**

Performing three buffers standard curve by using buffer nominal pH (4,7)(7,10)

Unit Under Calibration	Standard pH Buffer Solution	Actual pH Reading	Actual mV Reading (mV)	Uncertainty of pH measurement (\pm)	Coverage factor k
pH Electrode S/N.: 220202SIA605377	4.008	4.01	159	0.0079	2.00
	6.987	6.99	-14	0.011	2.00
	6.987	7.00	-14	0.011	2.00
	10.008	10.01	-190	0.0095	2.00

Function : Temperature Measurement

(*) Without adjustment

This equipment was connected with Temperature Probe;

- Model : -
 - Serial No. : 220202SIA605377
 Dimension of probe;
 - Length : 112 mm.
 - Diameter : 12 mm.
 - Immersion Depth : 100 mm.

Calibration Point (°C)	Standard Temperature (°C)	UUC* Reading (°C)	Error (°C)	Uncertainty of measurement (\pm °C)	Coverage factor k
25.0	25.001	25.1	0.099	0.13	2.00
30.0	30.002	30.1	0.098	0.13	2.00
35.0	35.003	35.1	0.097	0.13	2.00

Remark : - UUC* = Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-00-

เอกสารไม่ครบถ้วน

ภาคผนวก ฉบับที่ ๑

หนังสืออนุญาตขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกสาร





ที่ ยก ๑๒๙๙๙/ ล ๖ ๕ ๑ ๘

กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๒ แขวงทุ่งสองห้อง
เขตราชเทวี กรุงเทพฯ ๑๐๔๐๐

๑๓ มีนาคม ๒๕๖๒

เรื่อง เบี้ยเลี้ยงเบี้ยชดเชยค่าใช้จ่ายของพนักงานที่เดินทางไปราชการต่างประเทศ

เรียน กรมการสู่จัดการ บริษัท บูรพา จำกัด และนายสิริ แอนด์ เอ็นจิเนียริ่ง คอมเพล็กซ์ จำกัด
ผู้ขอเบี้ยชดเชยค่าใช้จ่ายเดินทางไปราชการต่างประเทศ ตามที่ได้ระบุไว้ในแบบฟอร์มที่ ๑ ของกฎบัญชี ๑๐ ถนนสุขุมวิท
แขวงคลองเตย เขตคลองเตย กรุงเทพมหานคร ขออนุมัติเบี้ยชดเชยค่าใช้จ่ายของพนักงานที่เดินทางไปราชการต่างประเทศ

ตามหนังสือที่ ๑๒๙๙๙/ บ ๖ ๕ ๑ ๘ ลงวันที่ ๗ มกราคม ๒๕๖๒ สถานที่ที่เมืองที่ ๑ ของอุซเบกิสถาน จำนวนสี่วัน
ห้องปฏิบัติการวิเคราะห์ออกบิน เดินทางไป ๒-๓๕๕ สถานที่ที่เมืองที่ ๑ ของอุซเบกิสถาน จำนวนสี่วัน



ที่ ยก ๑๒๙๙๙/ ล ๖ ๕ ๑ ๘

กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๒ แขวงทุ่งสองห้อง
เขตราชเทวี กรุงเทพฯ ๑๐๔๐๐

๒๕ พฤศจิกายน ๒๕๖๒

เรื่อง เบี้ยเลี้ยงเบี้ยชดเชยค่าใช้จ่ายของพนักงานที่เดินทางไปราชการต่างประเทศ

เรียน กรมการสู่จัดการ บริษัท บูรพา จำกัด และนายสิริ แอนด์ เอ็นจิเนียริ่ง คอมเพล็กซ์ จำกัด
ผู้ขอเบี้ยชดเชยค่าใช้จ่ายเดินทางไปราชการต่างประเทศ ตามที่ได้ระบุไว้ในแบบฟอร์มที่ ๑ ของอุซเบกิสถาน จำนวนสี่วัน

ตามหนังสือที่ ๑๒๙๙๙/ บ ๖ ๕ ๑ ๘ ลงวันที่ ๗ มกราคม ๒๕๖๒ สถานที่ที่เมืองที่ ๑ ของอุซเบกิสถาน จำนวนสี่วัน



“อุตสาหกรรมก้าวไกล ประเพณีไทยก้าวหน้า ร่วมกันพัฒนา อุตสาหกรรมที่ดี”



ที่ ยก ๑๒๙๙๙/ ๖ ๐ ๒ ๙



กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๒ แขวงทุ่งสองห้อง
เขตราชเทวี กรุงเทพฯ ๑๐๔๐๐

๒๕ มีนาคม ๒๕๖๒

เรื่อง เบี้ยเลี้ยงเบี้ยชดเชยค่าใช้จ่ายและสามารถพำนักที่ต่างประเทศ

เรียน กรมการสู่จัดการ บริษัท บูรพา จำกัด และนายสิริ แอนด์ เอ็นจิเนียริ่ง คอมเพล็กซ์ จำกัด

ผู้ขอเบี้ยชดเชยค่าใช้จ่ายเดินทางไปราชการต่างประเทศ ตามที่ได้ระบุไว้ในแบบฟอร์มที่ ๑ ของอุซเบกิสถาน จำนวนสี่วัน

ลงวันที่ ๒๙ มกราคม ๒๕๖๒

สังเขปเรื่องดังนี้ เอกสารแนบท้ายนี้ขอเบี้ยชดเชยค่าใช้จ่ายและสามารถพำนักที่ต่างประเทศ

บริษัท บูรพา จำกัด และนายสิริ แอนด์ เอ็นจิเนียริ่ง คอมเพล็กซ์ จำกัด จำนวน ๒ แผ่น


ตามหนังสือที่ ๑๒๙๙๙/ บ ๖ ๕ ๑ ๘ ลงวันที่ ๗ มกราคม ๒๕๖๒ สถานที่ที่เมืองที่ ๑ ของอุซเบกิสถาน จำนวนสี่วัน

ห้องปฏิบัติการวิเคราะห์ออกบิน เดินทางไป ๒-๓๕๕ สถานที่ที่เมืองที่ ๑ ของอุซเบกิสถาน จำนวนสี่วัน
เดินทางจากกรุงเทพมหานคร ขออนุมัติเบี้ยชดเชยค่าใช้จ่ายและสามารถพำนักที่ต่างประเทศ

ความละเอียดเจ้า บัน

- ๑๖ -

ฉบับนี้ หนังสือด้านล่างนี้จะถือว่าเป็นเอกสารที่ออกให้แก่บุคคลที่ได้รับอนุญาตให้เข้าประเทศนี้ ห้องปฏิบัติการวิเคราะห์ออกบิน
ที่ ยก ๑๒๙๙๙/ ๖ ๐ ๒ ๙ ถูกกำหนดให้ใช้ในวันที่ ๒ พฤษภาคม ๒๕๖๓ คือในวันที่ ๒ พฤษภาคม ๒๕๖๓ ที่นี่ สำหรับการเดินทาง
ท่องเที่ยวและธุรกิจ สำหรับบุคคลที่ได้รับอนุญาตให้เดินทางไปประเทศไทย ตาม QR Code ที่แนบมาด้านล่าง



กองวิจัยและพัฒนาภารกิจโครงสร้าง
กลุ่มมาตรฐานวิธีการวิเคราะห์ต่อต้านยาเสพติดและยาเสพติด ห้องปฏิบัติการ
ไทร. ๑๙๙๓๐ ๒๙๙๒๒ ต.๑ ๑๙๙๓๐๔๔ โทรสาร ๐ ๑๙๙๓๐ ๒๙๙๒๒ ต.๑ ๑๙๙๓๐๔๔
ไปรษณีย์ตัวอย่างนี้ส่งให้ sara@oia.go.th

๙๙

“อุตสาหกรรมก้าวไกล ประเพณีไทยก้าวหน้า ร่วมกันพัฒนา อุตสาหกรรมที่ดี”



ที่ อ ก ๐๑๑๐(๑)/ ๔ ๒ ๕ ๗ ๙

๐ ๑ ๕๘๖๘๙ ๒๕๖๕

กรมโรงงานอุตสาหกรรม
ถนนพหลโยธินที่ ๒ แขวงกรุงเทพฯ
เขตราชเทวี กรุงเทพฯ ๑๐๑๐๐

เรื่อง เบี้ยนแปลงบุคลากรของห้องปฏิบัติการวิเคราะห์ที่

เรียน กรมการสุ่มตัด บริษัท ภูมิเพ็ชร และนาโนลิฟต์ และน้ำอุบลรัตน์ เอ็นจิเนียร์ิง ศบค. จำกัด
ข้างต้น ค่าขออนุญาตเบิกบุคลากรของห้องปฏิบัติการวิเคราะห์ที่อยู่กัน
ลงวันที่ ๒๒ สิงหาคม ๒๕๖๕

ตามหนังสือที่อ้างอิง บริษัท ภูมิเพ็ชร ภูมิในเดือน แผนบ้านเดือน แผนต่อ เอ็นจิเนียร์ิง ศบค. จำกัด
ห้องปฏิบัติการวิเคราะห์ที่อยู่กัน เผยแพร่เมื่อ ๒๕๖๕ สถานที่ตั้งเลขที่ ๓ ซอยอุดมสุข ๑๗ ถนนอุดมสุข แขวงบางจาก
เขตพญาไท กรุงเทพมหานคร ขอบคุณแปลงบุคลากรของห้องปฏิบัติการวิเคราะห์ ความละเอียดแจ้งแล้ว นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว มีความเห็นดังนี้

๑. ให้ยกเลิกข้อห้ามที่ประชารัฐห้องปฏิบัติการวิเคราะห์ จำนวน ๔ ราย
 - (๑) นายบริษัท ไชยรุจิร์สุด หอบเป็นเลขที่ ๑-๙๙๕-๔-๐๐๘๘
 - (๒) นายปิยะพันธุ์ ศรีไชยวัฒน์ หอบเป็นเลขที่ ๒-๙๙๕-๔-๐๐๘๘
 - (๓) นายธีรัตน์ ศรีศรี หอบเป็นเลขที่ ๒-๙๙๕-๔-๐๐๘๘
 - (๔) นางสาวกานิรุณ อะนันดา หอบเป็นเลขที่ ๒-๙๙๕-๔-๐๐๘๘
 - (๕) นายตักลิศกิจวิชัย ใจดี หอบเป็นเลขที่ ๒-๙๙๕-๔-๐๐๘๘
 - (๖) นางสาวลีดัคกี้วาย โพธิ์พันธ์ หอบเป็นเลขที่ ๒-๙๙๕-๔-๐๐๘๘
 - (๗) นางสาววนิดา แม่จันทร์ หอบเป็นเลขที่ ๒-๙๙๕-๔-๐๐๘๘
 - (๘) นางสาวชาติราษฎร์ ประกอบบริรักษ์ หอบเป็นเลขที่ ๒-๙๙๕-๔-๐๐๘๘
 - (๙) นางสาวอรุณารักษ์ แนะนำ ๗๐ ราย หอบเป็นเลขที่ ๑-๙๙๕-๔-๐๐๘๘
 - (๑๐) นายวิวัฒน์ พิมพ์ธรรม หอบเป็นเลขที่ ๑-๙๙๕-๔-๐๐๘๘
 - (๑๑) นายวิวัฒน์พันธุ์วงศ์ วงศ์คำ หอบเป็นเลขที่ ๑-๙๙๕-๔-๐๐๘๘
๒. ให้เพิ่มข้อห้ามที่ประชารัฐห้องปฏิบัติการวิเคราะห์ จำนวน ๑๐ ราย
 - (๑) นางสาวอรุณารักษ์ แนะนำ ๗๐ ราย หอบเป็นเลขที่ ๑-๙๙๕-๔-๐๐๘๘
 - (๒) นางสาวอรุณารักษ์ แนะนำ ๗๐ ราย หอบเป็นเลขที่ ๑-๙๙๕-๔-๐๐๘๘
 - (๓) นายวิวัฒน์ พิมพ์ธรรม หอบเป็นเลขที่ ๑-๙๙๕-๔-๐๐๘๘
 - (๔) นายวิวัฒน์พันธุ์วงศ์ วงศ์คำ หอบเป็นเลขที่ ๑-๙๙๕-๔-๐๐๘๘
 - (๕) นายประพันธ์พันธุ์วงศ์ นีติกานต์ หอบเป็นเลขที่ ๒-๙๙๕-๔-๐๐๘๘
 - (๖) นางสาวอรุณารักษ์ วงศ์คำ หอบเป็นเลขที่ ๒-๙๙๕-๔-๐๐๘๘
 - (๗)
 - (๘)
 - (๙)
 - (๑๐)
 - (๑๑)

- ๒ -

ดังนั้น หนังสือฉบับนี้จะยกเว้นให้ใช้อำยุบันทึกและบันทึกปฏิบัติการวิเคราะห์ที่ออก
ที่ อ ก ๐๑๑๐(๑)/๔๒๕ ลงวันที่ ๕ กุมภาพันธ์ ๒๕๖๕ หรือในวันที่ ๖ กุมภาพันธ์ ๒๕๖๕ ทั้งนี้ สามารถอ่านค่าขอ
ผู้บุคคลที่ใช้ก่อนได้ที่หน้าเว็บไซต์กรมโรงงานอุตสาหกรรม ตาม QR Code ท้ายหนังสือฉบับนี้

ภาค

G Green Industry “มุ่งสู่การบรรจุภัณฑ์เชิงรุก ประเทศไทย” ชั้นนำแห่งประเทศไทย ด้วยความตั้งใจที่จะเป็นผู้นำ ด้วยการพัฒนา ดูแลความยั่งยืน



ที่ อ ก ๐๑๑๐(๑)/ ๔ ๒ ๕ ๗ ๙

๒ ๑ ๕๘๖๘๙ ๒๕๖๕

กรมโรงงานอุตสาหกรรม
ถนนพหลโยธินที่ ๒ แขวงกรุงเทพฯ
เขตราชเทวี กรุงเทพฯ ๑๐๑๐๐

เรื่อง เบี้ยนแปลงบุคลากรของห้องปฏิบัติการวิเคราะห์ที่

เรียน กรมการสุ่มตัด บริษัท ภูมิเพ็ชร และนาโนลิฟต์ และน้ำอุบลรัตน์ เอ็นจิเนียร์ิง ศบค. จำกัด
ข้างต้น ค่าขออนุญาตเบิกบุคลากรของห้องปฏิบัติการวิเคราะห์ที่อยู่กัน
ลงวันที่ ๓๐ มีนาคม ๒๕๖๕

ตามหนังสือที่อ้างอิง บริษัท ภูมิเพ็ชร ภูมิในเดือน แผนบ้านเดือน แผนต่อ เอ็นจิเนียร์ิง ศบค. จำกัด
ห้องปฏิบัติการวิเคราะห์ที่อยู่กัน เผยแพร่เมื่อ ๒๕๖๕ สถานที่ตั้งเลขที่ ๓ ซอยอุดมสุข ๑๗ ถนนอุดมสุข แขวงบางจาก
เขตพญาไท กรุงเทพมหานคร ขอบคุณแปลงบุคลากรของห้องปฏิบัติการวิเคราะห์ ความละเอียดแจ้งแล้ว นั้น

โดยวันที่ ๓๐ มีนาคม ๒๕๖๕

ตามหนังสือที่อ้างอิง บริษัท ภูมิเพ็ชร ภูมิในเดือน แผนบ้านเดือน แผนต่อ เอ็นจิเนียร์ิง ศบค. จำกัด
ห้องปฏิบัติการวิเคราะห์ที่อยู่กัน เผยแพร่เมื่อ ๒๕๖๕ สถานที่ตั้งเลขที่ ๓ ซอยอุดมสุข ๑๗ ถนนอุดมสุข แขวงบางจาก
เขตพญาไท กรุงเทพมหานคร ขอบคุณแปลงบุคลากรของห้องปฏิบัติการวิเคราะห์ ความละเอียดแจ้งแล้ว นั้น

โดยวันที่ ๓๐ มีนาคม ๒๕๖๕



ທີ ອກ ០២០០(១)/ ន.ដ. ៥

ກມໄງຮ່ານຊຸດສາກຽມ
ຕະຫະພະນາຖ. ៦ ແລະຖາງໝາງຢາໄ
ເມືອງໄກເຈົ້າ, ຖະແຫຼງ ຂອງລາວ

° ៩ ກຸມພາພັນທຶນ ៥

ເຊື່ອ: ດ້ວຍຫຸ້ນສີຂໍ້ຕ້ອນທະເບູນຫ້ອງນິ້ຕິກາວວິເຄາະທີ່ເອກະນຸ

ເຮັດວຽກ ກໍານະກິດການ ບ້ານເຕີຕະກິ ແລະນາສີສັດ ແລະນິຕ່ ເປົ້ນຈີນເຍົ່າງ ຄອນຊື່ແຕນທຶນທີ່ຈ້າກັດ
ຫ້າງໃໝ່ ດ້ວຍເຫັນຫຸ້ນທະເບູນຫ້ອງນິ້ຕິກາວວິເຄາະທີ່ເອກະນຸ

ໂລກວິທີ ៨៧ ວິນາຄານ ៦៥០៥

ສື່ເກີນສາມາດວຍ ១. ຮາມທີ່ຫຼູ້ຮັບຜູ້ປະທິກາວໃກ່ໄກເຈົ້າ ៥ໜັນນ ៤០ ວາຍ
២. ຮາມທີ່ເຫັນຫຸ້ນທະເບູນຫ້ອງນິ້ຕິກາວວິເຄາະທີ່ຫຼູ້ຮັບຜູ້ປະທິກາວໃກ່ໄກເຈົ້າ ៣០ ວາຍ
៣. ຂອບໃຈ່ສາມາດພື້ນທີ່ໄດ້ໃຫ້ໃຫ້ນ໌ຜູ້ປະທິກາວການໄງຮ່ານຊຸດສາກຽມ

ຄວນຫົວໜ້າສີ່ວິທີ່ກາງໃກ່ ប້ານເຕີຕະກິ ໂຢີມຕິດ ແລະນາສີສັດ ແລະນິຕ່ ເປົ້ນຈີນເຍົ່າງ ຄອນຊື່ແຕນທຶນທີ່ຈ້າກັດ
ດ້ວຍຫຸ້ນສີ່ວິທີ່ກາງໃກ່ ប້ານເຕີຕະກິ ໂຢີມຕິດ ແລະນາສີສັດ ແລະນິຕ່ ເປົ້ນຈີນເຍົ່າງ ຄອນຊື່ແຕນທຶນທີ່ຈ້າກັດ
ຂອງທຸກຄົງ ៥៩ ເມນຸມອຸນຫະກາດ ແລະພຣະໄຈແລ ກ່ຽວທະການຕາມ ຕ່ອງການໄງຮ່ານຊຸດສາກຽມ ນັ້ນ

ກມໄງຮ່ານຊຸດສາກຽມທີ່ຈ້າກັດ ດ້ວຍຫຸ້ນທະເບູນຫ້ອງນິ້ຕິກາວວິເຄາະທີ່ໃຫ້ໃຫ້ນ໌ຜູ້ປະທິກາວໃກ່ໄກເຈົ້າ
ຄວນຫົວໜ້າສີ່ວິທີ່ກາງໃກ່ ៥ໜັນນ ៤០ ວາຍ ເປົ້ນຈີນທີ່ໄດ້ກ່າວໂໝ່ ໃຫ້ໃຫ້ນ໌ຜູ້ປະທິກາວໃກ່ໄກເຈົ້າ
ກ. ສູ່ຄວາມມູນຄຸມທີ່ສ່ວນໃຫ້ທ່ານທີ່ໄດ້ກ່າວໂໝ່ ພຳການ ຖື່ນສິນທີ່ໄດ້ກ່າວໂໝ່
ក. ເຫັນຫຸ້ນທີ່ປະ່ວິທີ່ທີ່ໄດ້ກ່າວໂໝ່ ພຳການ ພຳການ ເປົ້ນຈີນທີ່ໄດ້ກ່າວໂໝ່
៣. ຂອບໃຈ່ສາມາດພື້ນທີ່ໄດ້ໃຫ້ໃຫ້ນ໌ຜູ້ປະທິກາວໃກ່ໄກເຈົ້າ ເປົ້ນຈີນ ພຳການ ເປົ້ນຈີນ ສີ່ວິທີ່
ທີ່ໄດ້ກ່າວໂໝ່ ແລະຕົນ ຄວາມສິ່ງທີ່ສ່ວນກັດວັນ ៣

ພັນດີຂອງບັນຫຼາດແຫ່ງດ້ວຍການເປົ້ນຈີນທີ່ໄດ້ກ່າວໂໝ່ ໂຢີມຕິດ ៥៥០៥ ອາກປະສົງສະກະດ້ວຍຫຸ້ນທະເບູນຫ້ອງນິ້ຕິກາວວິເຄາະທີ່ໄດ້ກ່າວໂໝ່
ໄຫຉທີ່ທະຍາບີນທີ່ໄດ້ກ່າວໂໝ່ໄດ້ກ່າວໂໝ່ ໃຫ້ໃຫ້ນ໌ຜູ້ປະທິກາວໃກ່ໄກເຈົ້າ ໂຢີມຕິດ ແລະນາສີສັດ
ກມໄງຮ່ານຊຸດສາກຽມມາໃຫຍ່ໃນ ៣០ ວານ ກໍຍົນວ້າເລີ່ມຂໍອະຫວາງຫຸ້ນທີ່ໄດ້ກ່າວໂໝ່ ທີ່ໄດ້ກ່າວໂໝ່
ຕ່ອງການໃຫ້ສາມາດມີການດໍາລັບນອກທີ່ໄດ້ກ່າວໂໝ່ ຫຼື ໃຫ້ໃຫ້ນ໌ຜູ້ປະທິກາວໃກ່ໄກເຈົ້າ
ການເປົ້ນຈີນ ຢ່າງເປົ້ນຈີນ ຢ່າງຫຼັງມາດໍາລັບນອກທີ່ໄດ້ກ່າວໂໝ່ ໃຫ້ໃຫ້ນ໌ຜູ້ປະທິກາວໃກ່ໄກເຈົ້າ
ຫຼັງຈາກນີ້ ສາມາດໃຫ້ກ່າວໂໝ່ໄດ້ກ່າວໂໝ່ໄດ້ກ່າວໂໝ່ ໃຫ້ໃຫ້ນ໌ຜູ້ປະທິກາວໃກ່ໄກເຈົ້າ
ການເປົ້ນຈີນ ລັງທຶນສື່ວິທີ່ກາງໃກ່

ເອກສາມະນີນສີ່ວິທີ່ກາງທັງໝົດທີ່ໄດ້ກ່າວໂໝ່ ໂຢີມຕິດ ແລະນາສີສັດ
ບ້ານເຕີຕະກິ ແລະນິຕ່ ແລະນິຕ່ ເປົ້ນຈີນເຍົ່າງ ຄອນຊື່ແຕນທຶນທີ່ຈ້າກັດ ແລະທະເບູນ ၇-၅၄
ທີ ອກ ០២០០(១)/ ន.ດ. ៥



ສັນຕະພິເກດ ໨. ຕົກມາພັນທຶນ ៥



ກມໄງຮ່ານຊຸດສາກຽມ ມາຮັດວຽກ

ເອກສາມະນີນສີ່ວິທີ່ກາງທັງໝົດທີ່ໄດ້ກ່າວໂໝ່ ໂຢີມຕິດ ແລະນາສີສັດ
ບ້ານເຕີຕະກິ ແລະນິຕ່ ແລະນິຕ່ ເປົ້ນຈີນເຍົ່າງ ຄອນຊື່ແຕນທຶນທີ່ຈ້າກັດ ແລະທະເບູນ ၇-၅၄
ທີ ອກ ០២០០(១)/ ន.ດ. ៥

ສັນຕະພິເກດ ໨. ຕົກມາພັນທຶນ ៥



ກມໄງຮ່ານຊຸດສາກຽມ ມາຮັດວຽກ

บัญชีรายการน้ำเสียที่ได้รับไปเมืองท่องเที่ยว

สิ่งที่ต้องดูแล ๓
ตรวจสอบห้ามห้ามสิ่งของที่อยู่ในน้ำเสียที่มีปฏิกิริยาที่อาจเป็นอันตราย
บริษัท จำกัด แผนกสิ่งแวดล้อม แผนกที่ ๑ เอ็นจิเนียร์ ศูนย์ซัพพลาย ชั้น ๕
ที่ อ. ก. ๑๒๓๐(๑) / ๑๔๘๙ ถนนที่ ๑๙ หมู่บ้านที่ ๑๙ รายการ
ขอบข่ายการผลิตที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๑๕๗ รายการ
น้ำเสีย จำนวน ๔๖ รายการ

ลำดับ	สารเคมี	วิธีการตรวจ
1	Aldrin	Liquid-Liquid Extraction, Gas Chromatographic Method ^[6]
2	Arsenic	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[6] 2) Digestion, Inductively Coupled Plasma Method ^[6]
3	Barium	Digestion, Inductively Coupled Plasma Method ^[6]
4	α -BHC	Liquid-Liquid Extraction, Gas Chromatographic Method ^[6]
5	β -BHC	Liquid-Liquid Extraction, Gas Chromatographic Method ^[6]
6	δ -BHC	Liquid-Liquid Extraction, Gas Chromatographic Method ^[6]
7	γ -BHC	Liquid-Liquid Extraction, Gas Chromatographic Method ^[6]
8	Biochemical Oxygen Demand	1) 5-Day BOD Test, Azide Modification Method ^[6] 2) 5-Day BOD Test, Membrane Electrode Method ^[6] 1) Digestion, Direct Air-Acetylene Flame Method ^[6] 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[6] 3) Digestion, Inductively Coupled Plasma Method ^[6]
9	Cadmium	1) Digestion, Direct Air-Acetylene Flame Method ^[6] 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[6] 3) Digestion, Inductively Coupled Plasma Method ^[6]
10	Chemical Oxygen Demand	1) Closed Reflux, Titrimetric Method ^[6] 2) Closed Reflux, Colorimetric Method ^[6] 3) Open Reflux, Titrimetric Method ^[6]
11	Chlordane	Liquid-Liquid Extraction, Gas Chromatographic Method ^[6]
12	Chromium	1) Digestion, Direct Air-Acetylene Flame Method ^[6] 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[6]
13	Color	
14	Copper	
15	Cyanide	

16 o,p'-DDT...

ลำดับ	สารเคมี	วิธีการตรวจ
16	o,p'-DDT	Liquid-Liquid Extraction, Gas Chromatographic Method ^[6]
17	4,4'-DDD	Liquid-Liquid Extraction, Gas Chromatographic Method ^[6]
18	4,4'-DDE	Liquid-Liquid Extraction, Gas Chromatographic Method ^[6]
19	4,4'-DDT	Liquid-Liquid Extraction, Gas Chromatographic Method ^[6]
20	Dieleadrin	Liquid-Liquid Extraction, Gas Chromatographic Method ^[6]
21	Endosulfan I	Liquid-Liquid Extraction, Gas Chromatographic Method ^[6]
22	Endosulfan II	Liquid-Liquid Extraction, Gas Chromatographic Method ^[6]
23	Endosulfan sulfate	Liquid-Liquid Extraction, Gas Chromatographic Method ^[6]
24	Endrin	Liquid-Liquid Extraction, Gas Chromatographic Method ^[6]
25	Endrin aldehyde	Liquid-Liquid Extraction, Gas Chromatographic Method ^[6]
26	Formaldehyde	Distillation, Colorimetric Method ^[3]
27	Free Chlorine	1) Iodometric Method ^[6] 2) DPD Ferrous Titrimetric Method ^[6]
28	Heptachlor	Liquid-Liquid Extraction, Gas Chromatographic Method ^[6]
29	Heptachlor Epoxide	Liquid-Liquid Extraction, Gas Chromatographic Method ^[6]
30	Hexavalent Chromium	1) Colorimetric Method ^[6] 2) Extraction, Direct Air-Acetylene Flame Method ^[6] 1) Digestion, Direct Air-Acetylene Flame Method ^[6] 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[6] 3) Digestion, Inductively Coupled Plasma Method ^[6]
31	Lead	1) Digestion, Direct Air-Acetylene Flame Method ^[6] 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[6] 3) Digestion, Inductively Coupled Plasma Method ^[6]
32	Manganese	1) Digestion, Direct Air-Acetylene Flame Method ^[6] 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[6] 3) Digestion, Inductively Coupled Plasma Method ^[6]
33	Mercury	
34	Methoxychlor	
35	Nickel	

36 Oil & Grease...

ลำดับ	สารเคมี	วิธีการตรวจ
36	Oil & Grease	1) Liquid-Liquid, Partition-Gravimetric Method ^[4] 2) Soxhlet Extraction Method ^[4]
37	pH	Electrometric Method ^[4]
38	Phenols	1) Distillation, Chloroform Extraction Method ^[4] 2) Distillation, Direct Photometric Method ^[4]
39	Selenium	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[4] 2) Digestion, Inductively Coupled Plasma Method ^[4]
40	Sulfide	1) Iodometric Method ^[4] 2) Methylene Blue Method ^[4]
41	Temperature	Laboratory and Field Methods ^[4]
42	Total Dissolved Solids	Dried at 180 °C ^[4]
43	Total Kjeldahl Nitrogen	Semi-Micro-Kjeldahl Method ^[4]
44	Total Suspended Solids	Dried at 103-105 °C ^[4]
45	Trivalent Chromium	1) Digestion, Direct Air-Acetylene Flame Method; Colorimetric Method; Calculation ^[4] 2) Digestion, Inductively Coupled Plasma Method; Colorimetric Method; Calculation ^[4]
46	Zinc	1) Digestion, Direct Air-Acetylene Flame Method ^[4] 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[4] 3) Digestion, Inductively Coupled Plasma Method ^[4]

น้ำได้คืน จำนวน 126 รายการ

ลำดับ	สารเคมี	วิธีการตรวจ
1	Acenaphthene	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
2	Acetone	
3	Aldrin	

ลำดับ	สารเคมี	วิธีการตรวจ
4	Anthracene	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
5	Antimony	Digestion, Inductively Coupled Plasma Method ^[4]
6	Arsenic	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[4] 2) Digestion, Inductively Coupled Plasma Method ^[4]
7	Atrazine	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
8	Barium	1) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[4] 2) Digestion, Inductively Coupled Plasma Method ^[4]
9	Benz(a)anthracene	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
10	Benzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
11	Benz(b)fluoranthene	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
12	Benz(k)fluoranthene	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
13	Benzoic acid	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
14	Benz(a)pyrene	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4]

15 Benzo(g,h,i)perylene...

ลำดับ	สารเคมี	วิธีการตรวจ
15	Benzo(g,h,i)perylene	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
16	Beryllium	Digestion, Inductively Coupled Plasma Method ^[4]
17	Bis(2-chloroethyl)ether	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
18	Bis(2-ethylhexyl)phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
19	Bromodichloromethane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
20	Bromoform	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
21	Butanol	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
22	Butyl benzyl phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
23	Cadmium	1) Digestion, Direct Air-Acetylene Flame Method ^[4] 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[4] 3) Digestion, Inductively Coupled Plasma Method ^[4]
24	Carbazole	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
25	Carbon disulfide	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
26	Carbon tetrachloride	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
27	Chlordane	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
28	p-Chloraniline	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
29	Chlorobenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]

30 Chlorodibromomethane...

ลำดับ	สารเคมี	วิธีการตรวจ
30	Chlorodibromomethane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
31	Chloroform	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
32	2-Chlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
33	Chromium	1) Digestion, Direct Air-Acetylene Flame Method ^[4] 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[4] 3) Digestion, Inductively Coupled Plasma Method ^[4]
34	Chromium (III)	1) Digestion, Direct Air-Acetylene Flame Method; Colorimetric Method; Calculation ^[4] 2) Digestion, Inductively Coupled Plasma Method; Colorimetric Method; Calculation ^[4]
35	Chromium (VI)	1) Colorimetric Method ^[4] 2) Extraction, Air-Acetylene Flame Method ^[4]
36	Chrysene	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
37	Cyanide	Distillation, Colorimetric Method ^[4]
38	2,4-D	Liquid-Liquid Extraction, Gas Chromatographic Method ^[4]
39	DDD	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
40	DDE	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
41	DDT	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]

42 Dibenz(a,h)anthracene...

ลำดับ	สารเคมี	วิธีการตรวจ
42	Dibenz[a,h]anthracene	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
43	Di-n-butyl phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
44	1,2-Dichlorobenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
45	1,3-Dichlorobenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
46	1,4-Dichlorobenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
47	3,3'-Dichlorobenzidine	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
48	1,1-Dichloroethane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
49	1,2-Dichloroethane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
50	1,1-Dichloroethylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
51	cis-1,2-Dichloroethylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
52	trans-1,2-Dichloroethylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
53	2,4-Dichlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
54	1,2-Dichloropropane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
55	1,3-Dichloropropane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
56	1,3-Dichloropropene	
57	Diehrin	

58 Diethyl phthalate...

ลำดับ	สารเคมี	วิธีการตรวจ
58	Diethyl phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
59	2,4-Dimethylphenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
60	2,4-Dinitrophenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
61	2,4-Dinitrotoluene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
62	2,6-Dinitrotoluene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
63	Di-n-Octyl phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
64	Endosulfan	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
65	Endrin	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
66	Ethylbenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
67	Fluoranthene	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
68	Fluorene	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
69	Heptachlor	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]

UNITED ANALYST AND ENGINEER CONSULTANT COMPANY LIMITED

58 Diethyl phthalate...

ลำดับ	สารเคมี	วิธีการตรวจ
70	Heptachlor epoxide	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
71	Hexachlorobenzene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
72	Hexachloro-1,3-butadiene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
73	n-Hexane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
74	α -HCH	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
75	β -HCH	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
76	γ -HCH	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
77	Hexachlorocyclopentadiene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
78	Hexachloroethane	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
79	Indeno(1,2,3-cd)pyrene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
80	Isophorone	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
81	Lead	1) Digestion, Direct 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[4] 3) Digestion, Inductively Coupled Plasma Method ^[4]

UNITED ANALYST AND ENGINEER CONSULTANT COMPANY LIMITED

82 Manganese...

ลำดับ	สารเคมี	วิธีการตรวจ
82	Manganese	1) Digestion, Direct Air-Acetylene Flame Method ^[4] 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[4] 3) Digestion, Inductively Coupled Plasma Method ^[4]
83	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[4]
84	Methanol	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
85	Methoxychlor	Liquid-Liquid Extraction, Gas Chromatographic Method ^[4]
86	Methyl bromide	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
87	Methylene chloride	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
88	2-Methylphenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
89	2-Methylnaphthalene	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
90	Methyl tert-butyl ether	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
91	Naphthalene	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
92	Nickel	1) Digestion, Direct Air-Acetylene Flame Method ^[4] 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[4] 3) Digestion, Inductively Coupled Plasma Method ^[4]
93	Nitrobenzene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
94	N-Nitrosodiphenylamine	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
95	N-Nitrosodi-n-propylamine	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]

82 Manganese...

96 Polychlorinated Biphenyls...

ลำดับ	สารเคมี	วิธีการตรวจ
96	Polychlorinated Biphenyls - PCB 1016 - PCB 1221 - PCB 1232 - PCB-1242 - PCB-1248 - PCB-1254 - PCB-1260	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
97	Pentachlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
98	pH	Electrometric Method ^[4]
99	Phenanthrene	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
100	Phenol	1) Distillation, Chloroform Extraction Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
101	Pyrene	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
102	Selenium	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[4] 2) Digestion, Inductively Coupled Plasma Method ^[4]
103	Silver	Digestion, Inductively Coupled Plasma Method ^[4]
104	Styrene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
105	1,1,2,2-Tetrachloroethane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
106	Tetrachloroethylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
107	Toluene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]

108 Toxaphene...

ลำดับ	สารเคมี	วิธีการตรวจ
108	Toxaphene	1) Liquid-Liquid Extraction, Gas Chromatographic Method ^[4] 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
109	TPH (C ₅ - C ₈)	1) Purge and Trap, Gas Chromatographic Method ^[11,21] 2) Purge and Trap, Gas Chromatographic/Mass spectrometric Method ^[11,21]
110	TPH (C _{9,16} - C ₃₅)	Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^[9,21] Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^[9,21]
111	1,2,4-Trichlorobenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
112	1,1,1-Trichloroethane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
113	1,1,2-Trichloroethane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
114	1,2,4-Trichloroethylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
115	Trichloroethylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
116	2,4,5-Trichlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
117	2,4,6-Trichlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[4]
118	1,3,5-Trimethylbenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
119	Vanadium	Digestion, Inductively Coupled Plasma Method ^[4]
120	Vinyl acetate	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
121	Vinyl chloride	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
122	m-Xylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
123	o-Xylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]

UNIVERSITY ANALYST AND ENGINEER CONSULTANT COMPANY LIMITED

124 p-Xylene...

ลำดับ	สารเคมี	วิธีการตรวจ
124	p-Xylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
125	Xylene (Total)	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ^[4]
126	Zinc	1) Digestion, Direct Air-Acetylene Flame Method ^[4] 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^[4] 3) Digestion, Inductively Coupled Plasma Method ^[4]

รายการที่ (ป้องกันน้ำยา) ร้านค้า 25 รายการ

ลำดับ	สารเคมี	วิธีการตรวจ
1	Antimony	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[3]
2	Arsenic	1) Isokinetic Sampling, Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[3] 2) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[3]
3	Cadmium	1) Isokinetic Sampling, Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[3]
4	Carbon Monoxide	Instrumental Analyzer Method ^[3]
5	Chlorine	Isokinetic Sampling, Ion Chromatographic Method ^[3]
6	Chromium	1) Isokinetic Sampling, Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[3]
7	Cobalt	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[3]
8	Copper	1) Isokinetic Sampling, Digestion, Flame Method ^[3] 2) Isokinetic Sampling, Digestion, Plasma Method ^[3]
9	Cresol	Isokinetic Sampling, Digestion, Absorption Sampling, Flame Method ^[3]

10 Dioxins/Furans...

ลำดับ	สารเคมี	วิธีการตรวจ
10	Dioxins/Furans	Isokinetic Sampling ^[3]
11	Hydrogen Chloride	Isokinetic Sampling, Ion Chromatographic Method ^[3]
12	Hydrogen Fluoride	Isokinetic Sampling, Ion Chromatographic Method ^[3]
13	Hydrogen Sulfide	Absorption Sampling, Iodometric Method ^[3]
14	Lead	1) Isokinetic Sampling, Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[3]
15	Manganese	1) Isokinetic Sampling, Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[3]
16	Mercury	Isokinetic Sampling, Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[3]
17	Nickel	1) Isokinetic Sampling, Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[3]
18	Opacity	Ringelmann's Method ^[3]
19	Oxides of Nitrogen	1) Absorption Sampling, Phenoldisulfonic acid Method ^[3] 2) Instrumental Analyzer Method ^[3]
20	Selenium	1) Isokinetic Sampling, Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[3] 2) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ^[3]
21	Sulfur Dioxide	1) Absorption Sampling, Barium-Thorin Titrimetric Method ^[3] 2) Instrumental Analyzer Method ^[3]
22	Sulfuric Acid	Isokinetic Sampling, B
23	Total Suspended Particulate	Isokinetic Sampling, G
24	Vanadium	Isokinetic Sampling, D Plasma Method ^[3]
25	Xylene	1) Gas Sampling, Gas Chromatographic Method ^[3] 2) Absorption Sampling

สืบต่อ...

สื่อเบื้องต้นเรื่องวัสดุที่มีไว้แล้ว จำนวน 35 รายการ


ลำดับ	สารเคมี	วิธีวิเคราะห์
1	Aldrin	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^[2,9,21] 2) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22]
2	Antimony	Digestion, Inductively Coupled Plasma Method ^[7,13]
3	Arsenic	1) Waste Extraction, Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[2,6,15] 2) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[2,6,13] 3) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[7,15] 4) Digestion, Inductively Coupled Plasma Method ^[7,13]
4	Barium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[2,6,13] 2) Digestion, Inductively Coupled Plasma Method ^[7,13]
5	Beryllium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[2,6,13] 2) Digestion, Inductively Coupled Plasma Method ^[7,13]
6	Cadmium	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[2,6,14] 2) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[2,6,13] 3) Digestion, Flame Atomic Absorption Spectrometric Method ^[7,14] 4) Digestion, Inductively Coupled Plasma Method ^[7,13]
7	Chlordane	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^[2,9,21] 2) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22]
8	Chromium	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[2,6,14] 2) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[2,6,13] 3) Digestion, Flame Atomic Absorption Spectrometric Method ^[7,14] 4) Digestion, Inductively Coupled Plasma Method ^[7,13]

CONSULTANT COMPANY LIMITED

3) Digestion,...

ลำดับ	สารเคมี	วิธีวิเคราะห์
9	Chromium (II)	3) Digestion, Flame Atomic Absorption Spectrometric Method ^[7,14] 4) Digestion, Inductively Coupled Plasma Method ^[7,13] 1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method; Waste Extraction, Colorimetric Method; Calculation ^[2,6,13,16] 2) Waste Extraction, Digestion, Inductively Coupled Plasma Method; Waste Extraction, Colorimetric Method; Calculation ^[2,6,13,16]
10	Chromium (VI)	3) Digestion, Flame Atomic Absorption Spectrometric Method; Alkaline Digestion, Colorimetric Method; Calculation ^[1,2,14,16] 4) Digestion, Inductively Coupled Plasma Method; Alkaline Digestion, Colorimetric Method; Calculation ^[7,8,13,16]
11	Cobalt	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[2,6,13] 2) Digestion, Inductively Coupled Plasma Method ^[7,13]
12	Copper	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[2,6,14] 2) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[2,6,13] 3) Digestion, Flame Atomic Absorption Spectrometric Method ^[7,14] 4) Digestion, Inductively Coupled Plasma Method ^[7,13]
13	2,4-D	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^[2,9,21] 2) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22]
14	DDD	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[2,6,14] 2) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22]

CONSULTANT COMPANY LIMITED



ลำดับ	สารเคมี	วิธีวิเคราะห์
15	DOE	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^[2,9,21] 2) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22]
16	DDT	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^[2,9,21] 2) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22]
17	Dielein	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^[2,9,21] 2) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22]
18	Endrin	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^[2,9,21] 2) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22]
19	Heptachlor	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^[2,9,21] 2) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22]
20	Lead	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[2,6,14] 2) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[2,6,13] 3) Digestion, Flame Atomic Absorption Spectrometric Method ^[7,14] 4) Digestion, Inductively Coupled Plasma Method ^[7,13]
21	Lindane	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^[2,9,21] 2) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22]
22	Mercury	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[2,6,14] 2) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[2,6,13]

CONSULTANT COMPANY LIMITED

3) Digestion,...

ลำดับ	สารเคมี	วิธีวิเคราะห์
23	Methoxychlor	3) Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[14] 4) Digestion, Inductively Coupled Plasma Method ^[7,13] 5) Thermal Decomposition Amalgamation and Atomic Absorption Spectrometric Method ^[14]
24	Molybdenum	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^[2,9,21] 2) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22]
25	Nickel	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[2,6,13] 2) Digestion, Inductively Coupled Plasma Method ^[7,13] 1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[2,6,14] 2) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[2,6,13]
26	Polychlorinated Biphenyls	3) Digestion, Flame Atomic Absorption Spectrometric Method ^[7,14] 4) Digestion, Inductively Coupled Plasma Method ^[7,13] 1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^[2,9,21] 2) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22]



ลำดับ	สารเคมี	วิธีการตรวจ
27.	- 2,2',4,5,5'-Pentachlorobiphenyl - 2,3,3',4',6'-Pentachlorobiphenyl - 2,2',3,4,4',5'-Hexachlorobiphenyl - 2,2',3,4,5,5'-Hexachlorobiphenyl - 2,2',3,5,5',6'-Hexachlorobiphenyl - 2,2',4,4,5,5'-Hexachlorobiphenyl - 2,2',3,3,4,4',5'-Heptachlorobiphenyl - 2,2',3,4,4',5,5'-Heptachlorobiphenyl - 2,2',3,4,4',5,6'-Heptachlorobiphenyl - 2,2',3,4,5,5'-Heptachlorobiphenyl - 2,2',3,3',4,4',5,5'-Nonachlorobiphenyl Pentachlorophenol	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[7,28] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26] Electrometric Method ^[11,34]
28.	pH	1) Waste Extraction, Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[2,20]
29.	Selenium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[7,14] 2) Digestion, Hydride Generation Spectrometric Method ^[2,20] 3) Digestion, Hydride Generation Spectrometric Method ^[2,20] 4) Digestion, Inductively Coupled Plasma Method ^[7,14]


ลำดับ	สารเคมี	วิธีการตรวจ
3.	Aldrin	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
4.	Anthracene	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,26] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
5.	Antimony	Digestion, Inductively Coupled Plasma Method ^[7,14]
6.	Arsenic	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[7,15] 2) Digestion, Inductively Coupled Plasma Method ^[7,13]
7.	Atrazine	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
8.	Barium	Digestion, Inductively Coupled Plasma Method ^[7,14]
9.	Benz(a)anthracene	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,26] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
10.	Benzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,25]
11.	Benz(b)fluoranthene	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,24] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
12.	Benz(k)fluoranthene	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,24] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
13.	Benzoic acid	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
14.	Benz(a)pyrene	1) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]

ลำดับ	สารเคมี	วิธีการตรวจ
30.	Silver	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[8,13] 2) Digestion, Inductively Coupled Plasma Method ^[7,11] 1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[8,13] 2) Digestion, Inductively Coupled Plasma Method ^[7,11]
31.	Thallium	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^[12,22] 2) Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,25]
32.	Toxaphene	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^[12,22] 2) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22]
33.	Trichloroethylene	1) Waste Extraction, Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[2,12,25] 2) Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,25]
34.	Vanadium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[8,13] 2) Digestion, Inductively Coupled Plasma Method ^[7,13]
35.	Zinc	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[7,14] 2) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[8,13] 3) Digestion, Flame Atomic Absorption Spectrometric Method ^[7,14] 4) Digestion, Inductively Coupled Plasma Method ^[7,13]

พิธี จำนวน 125 รายการ

ลำดับ	สารเคมี	วิธีการตรวจ
1.	Acenaphthene	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,26] 2) Ul Spec Purge Spec
2.	Acetone	

ลำดับ	สารเคมี	วิธีการตรวจ
15.	Benzo(g,h,i)perylene	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,26] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
16.	Beryllium	Digestion, Inductively Coupled Plasma Method ^[7,13]
17.	Bis(2-chloroethyl)ether	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
18.	Bis(2-ethylhexyl)phthalate	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
19.	Bromodichloromethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,25]
20.	Bromoform	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,25]
21.	Butanol	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,25]
22.	Butyl benzyl phthalate	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
23.	Cadmium	1) Digestion, Flame Atomic Absorption Spectrometric Method ^[7,14] 2) Digestion, Inductively Coupled Plasma Method ^[7,13]
24.	Carbazole	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
25.	Carbon disulfide	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,25]
26.	Carbon tetrachloride	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,25]
27.	Chlordane	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,26] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
28.	p-Chloroaniline	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
29.	Chlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,25]
30.	Chlorodibromomethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,25]



ลำดับ	สารเคมี	วิธีการตรวจ
31	Chloroform	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,25]
32	2-Chlorophenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
33	Chromium	1) Digestion, Flame Atomic Absorption Spectrometric Method ^[7,14] 2) Digestion, Inductively Coupled Plasma Method ^[7,13]
34	Chromium (III)	1) Digestion, Flame Atomic Absorption Spectrometric Method; Alkaline Digestion, Colorimetric Method; Calculation ^[7,14,16] 2) Digestion, Inductively Coupled Plasma Method; Alkaline Digestion, Colorimetric Method; Calculation ^[7,13,16]
35	Chromium (VI)	Alkaline Digestion, Colorimetric Method ^[8,16]
36	Chrysene	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,28] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,28]
37	Cyanide	Extraction, Distillation, Colorimetric Method ^[28,29,30]
38	2,4-D	Ultrasonic Extraction, Gas Chromatographic Method ^[10]
39	DDD	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
40	DDE	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
41	DOT	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
42	Dibenz(a,h)anthracene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]

43 Di-n-butyl phthalate...

ลำดับ	สารเคมี	วิธีการตรวจ
43	Di-n-butyl phthalate	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
44	1,2-Dichlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
45	1,3-Dichlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
46	1,4-Dichlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
47	3,3'-Dichlorobenzidine	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
48	1,1-Dichloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
49	1,2-Dichloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
50	1,1-Dichloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
51	cis-1,2-Dichloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
52	trans-1,2-Dichloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
53	2,4-Dichlorophenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
54	1,2-Dichloropropane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
55	1,3-Dichloropropane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
56	1,3-Dichloropropene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
57	Dieldrin	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,26] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
58	Diethyl phthalate	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
59	2,4-Dimethylphenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]

60 2,4-Dinitrophenol...

ลำดับ	สารเคมี	วิธีการตรวจ
60	2,4-Dinitrophenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
61	2,4-Dinitrotoluene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
62	2,6-Dinitrotoluene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
63	Di-n-Octyl phthalate	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
64	Endosulfan	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
65	Endrin	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
66	Ethylbenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
67	Fluoranthene	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,26] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
68	Fluorene	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,26] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
69	Heptachlor	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
70	Heptachlor epoxide	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]

UNITED ANALYST AND ENGINEERING CONSULTANT COMPANY LIMITED

71 Hexachlorobenzene...

ลำดับ	สารเคมี	วิธีการตรวจ
71	Hexachlorobenzene	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,26] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
72	Hexachloro-1,3-butadiene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
73	n-Hexane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
74	α -HCH	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,26] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
75	β -HCH	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,26] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
76	γ -HCH	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,26] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
77	Hexachlorocyclopentadiene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
78	Hexachloroethane	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
79	Indeno(1,2,3-cd)pyrene	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,26] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
80	Isophorone	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
81	Lead	1) Digestion, Flame Atomic Method ^[7,14] 2) Digestion, Inductively Coupled Plasma Method ^[7,14]
82	Manganese	1) Digestion, Flame Atomic Method ^[7,14] 2) Digestion, Inductively Coupled Plasma Method ^[7,14]

UNITED ANALYST AND ENGINEERING CONSULTANT COMPANY LIMITED

-iso-

ลำดับ	สารเคมี	วิธีการตรวจ
83	Mercury	1) Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[14] 2) Digestion, Inductively Coupled Plasma Method ^[7,13] 3) Thermal Decomposition Amalgamation and Atomic Absorption Spectrometric Method ^[19]
84	Methanol	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
85	Methoxychlor	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,22] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
86	Methyl bromide	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
87	Methylene chloride	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
88	2-Methylphenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
89	2-Methylnaphthalene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
90	Methyl tert-butyl ether	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
91	Naphthalene	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,26] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
92	Nickel	1) Digestion, Flame Atomic Absorption Spectrometric Method ^[7,14] 2) Digestion, Inductively Coupled Plasma Method ^[7,13]
93	Nitrobenzene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
94	N-Nitrosodiphenylamine	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
95	N-Nitrosodi-n-propylamine	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]

UNITED ANALYST AND INDUSTRIAL CONSULTANT COMPANY LIMITED



96 Polychlorinated Biphenyls...

-iso-

ลำดับ	สารเคมี	วิธีการตรวจ
-	- 2,2',3,4',5,5'- Heptachlorobiphenyl - 2,2',3,3',4,4',5,5'- Nonachlorobiphenyl	
97	Pentachlorophenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
98	Phenanthrene	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,26] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
99	Phenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
100	Pyrene	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,26] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26]
101	Selenium	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[7,22] 2) Digestion, Inductively Coupled Plasma Method ^[7,13]
102	Silver	Digestion, Inductively Coupled Plasma Method ^[7,13]
103	Styrene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
104	1,1,2,2-Tetrachloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
105	Tetrachloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
106	Toluene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
107	Toxaphene	Ultrasonic Extraction, Gas Chromatographic Method ^[10,26]
108	TPH (C ₈ -C ₁₆)	1) Purge and Trap, Gas Chromatographic Method ^[12,21] 2) Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
109	TPH (C ₈ -C ₁₆)	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
110	TPH (C ₁₆ -C ₃₀)	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
111	1,2,4-Trichlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]

UNITED ANALYST AND INDUSTRIAL CONSULTANT COMPANY LIMITED



112 1,1,1-Trichloroethane...

-iso-

ลำดับ	สารเคมี	วิธีการตรวจ
96	Polychlorinated Biphenyls - Aroclor 1016 - Aroclor 1221 - Aroclor 1232 - Aroclor 1242 - Aroclor 1248 - Aroclor 1254 - Aroclor 1260 Polychlorinated Biphenyls - 2-Chlorobiphenyl - 2,3-Dichlorobiphenyl - 2,2',5-Trichlorobiphenyl - 2,4',5-Trichlorobiphenyl - 2,2',3,5-Tetrachlorobiphenyl - 2,2,5,5-Tetrachlorobiphenyl - 2,3,4,4'-Tetrachlorobiphenyl - 2,2,3,4,5'-Pentachlorobiphenyl - 2,2',4,5,5'-Pentachlorobiphenyl - 2,3,3',4,6'-Pentachlorobiphenyl - 2,2',3,4,4',5'-Hexachlorobiphenyl - 2,2,3,4,5,5'-Hexachlorobiphenyl - 2,2,3,5,5',6'-Hexachlorobiphenyl - 2,2',4,4',5,5'-Hexachlorobiphenyl - 2,2,3,3',4,4',5'-Heptachlorobiphenyl - 2,2,3,4,4',5'-Heptachlorobiphenyl - 2,2,3,4,4',5'-Heptachlorobiphenyl	1) Ultrasonic Extraction, Gas Chromatographic Method ^[10,23] 2) Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,26] Ultrasonic Extraction, Gas Chromatographic Method ^[10,23]



- 2,2',3,4,5,5',6...


-iso-

ลำดับ	สารเคมี	วิธีการตรวจ
112	1,1,1-Trichloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
113	1,1,2-Trichloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
114	Trichloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
115	2,4,5-Trichlorophenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,24]
116	2,4,6-Trichlorophenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method ^[10,24]
117	1,3,5-Trimethylbenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
118	Vanadium	Digestion, Inductively Coupled Plasma Method ^[7,13]
119	Vinyl acetate	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
120	Vinyl chloride	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
121	m-Xylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
122	o-Xylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
123	p-Xylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
124	Xylene (Total)	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[12,21]
125	Zinc	1) Digestion, Flame Atomic Absorption Spectrometric Method ^[7,14] 2) Digestion, Inductively Coupled Plasma Method ^[7,13]

เอกสารแนบท้าย


1. กระทรวงอุตสาหกรรม, ประกาศกระทรวงว่าด้วยการออกใบอนุญาตประกอบกิจการชั่วคราว
ค้าหรือใช้เป็นในรายการศรีที่บัญชีของกฎหมายเรื่องของพิษภัย พ.ศ.๒๕๔๗ ประจำ
ราชกิจจานุเบกษา, ๔ ธันวาคม ๒๕๔๙, เล่มที่ ๑๒๓ พ.๘๖๗

2. กระทรวงอุตสาหกรรม, ประกาศกระทรวงว่าด้วยการออกใบอนุญาตประกอบกิจการชั่วคราว
ห้องรักษาความปลอดภัย ราชกิจจานุเบกษา, ๒๕ มกราคม ๒๕๔๙, พ.๘๖๗



3. สมาคมวิศวกรรมฯ...

3. สมาคมวิศวกรรมสิ่งแวดล้อมแห่งประเทศไทย. คู่มือวิเคราะห์น้ำเสีย. คอมพิวเตอร์ที่ ๔. กรุงเทพฯ: เรียนแมริสาพิมพ์, ๒๕๔๗.
4. APHA, AWWA, WEF. Standard Methods for the Examination of Water and Wastewater. 23rd ed. Washington, DC: APHA, 2017.
5. United States Environmental Protection Agency. Standards of Performance for New Stationary Sources. 40 CFR 60. Appendix A, 2019.
6. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. SW-846, 1997.
7. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Acid Digestion of Sediments, Sludges, and Soils. SW-846 Method 3050B, 1996.
8. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Alkaline Digestion for Hexavalent Chromium. SW-846 Method 3060A, 1996.
9. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste 3. Physical/Chemical Methods. Separatory Funnel Liquid-Liquid Extraction. SW-846 Method 3510C, 1996.
10. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Ultrasonic Extraction. SW-846 Method 3550C, 2007.
11. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Purge and Trap for Aqueous Samples. SW-846 Method 5030C, 2003.
12. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Closed System Purge and Trap and Extraction for Volatile Organics in Soil and Waste Sample. SW-846 Method 5035A, 2000.
13. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Inductively Coupled Plasma-Optical Emission Spectrometry. SW-846 Method 6010D, 2014.
14. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Flame Atomic Absorption Spectrometry. SW-846 Method 7000B, 2007.



16. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Chromium, Hexavalent (Colorimetric). SW-846 Method 7196A, 1992.
17. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Mercury In Liquid Waste (Manual Cold Vapor Technique). SW-846 Method 7470A, 1994.
18. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Mercury in Solid or Semisolid Waste (Manual Cold-Vapor Technique). SW-846 Method 7471B, 1998.
19. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Mercury in Solids and Solutions by Thermal Decomposition, Amalgamation, and Atomic Absorption Spectrophotometry. SW-846 Method 7473, 2007.
20. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Selenium (Atomic Absorption, Borohydride Reduction). SW-846 Method 7742, 1994.
21. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Nonhalogenated Organics Using GC/FID. SW-846 Method 8015D, 2003.
22. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Organochlorine Pesticides by Gas Chromatography. SW-846 Method 8082A, 2007.
23. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Polychlorinated Biphenyls (PCBs) by Gas Chromatography. SW-846 Method 8082B, 2007.
24. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Polynuclear Aromatic Hydrocarbons. SW-846 Method 8100, 1980.
25. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Volatile Mass Spectrometry. SW-846 Method 8260, 1992.
26. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Semivolatile Chromatography/Mass Spectrometry. SW-846 Method 8270, 1992.
27. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Chlorination-Pentafluorobenzylation Derivatization. SW-846 Method 8270A, 1992.



28. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Total and Amenable Cyanide : Distillation. SW-846 Method 9010C, 2004.
29. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Cyanide Extraction Procedure for Solids and Oils. SW-846 Method 9013A, 2014.
30. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Cyanide in Waters and Extracts using Titrimetric and Manual Spectrophotometric Procedures. SW-846 Method 9014, 2014.
31. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. pH Electrometric Measurement. SW-846 Method 9040C, 2004.
32. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Soil and Waste pH. SW-846 Method 9045D, 2004.



bsi.



By Royal Charter

Certificate of Registration

QUALITY MANAGEMENT SYSTEM - ISO 9001:2015

This is to certify that:

United Analyst and Engineering
Consultant Co., Ltd.
3 and 81 Soi Udomsuk 41,
Sukhumvit Rd., Bangkok,
Phrakanong, Bangkok
10260
Thailand

Holds Certificate Number:

FS 712681

and operates a Quality Management System which complies with the requirements of ISO 9001:2015 for the following scope:

The provision of environmental, social and health-related consultancy and services on laboratory, field sampling, safety, monitoring and audit, project study, and impact assessment.

For and on behalf of BSI

Original Registration Date:
Latest Revision Date: 7



This certificate was issued electronically and remains the property of BSI and is bound by the conditions of contract.
An electronic certificate can be authenticated online.

Printed copies can be validated at www.bsi-global.com/ClientDirectory or telephone +44(2) 294489-92.

Further clarifications regarding the scope of this certificate and the applicability of ISO 9001:2015 requirements may be obtained by consulting the organization.
This certificate is valid only if provided original copies are in complete sets.

Information and Contact: BSI, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, Tel: +44 345 680 9000
BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.

A Member of the BSI Group of Companies.

bsi.



By Royal Charter

Certificate of Registration

ENVIRONMENTAL MANAGEMENT SYSTEM - ISO 14001:2015

This is to certify that:

United Analyst and Engineering
Consultant Co., Ltd.
3 and 81 Soi Udomsuk 41,
Sukhumvit Rd., Bangkok,
Phrakanong, Bangkok
10260
Thailand

Holds Certificate Number:

EMS 724006

and operates an Environmental Management System which complies with the requirements of ISO 14001:2015 for the following scope:

The provision of environmental, social and health-related consultancy and services on laboratory, field sampling, safety, monitoring and audit, project study, and impact assessment.

For and

Original
Latest R



This certificate was issued electronically and remains the property of BSI and is bound by the conditions of contract.
An electronic certificate can be authenticated online.

Printed copies can be validated at www.bsi-global.com/ClientDirectory or telephone +44(2) 294489-92.

Further clarifications regarding the scope of this certificate and the applicability of ISO 14001:2015 requirements may be obtained by consulting the organization.
This certificate is valid only if provided original copies are in complete sets.

Information and Contact: BSI, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, Tel: +44 345 680 9000
BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.

A Member of the BSI Group of Companies.

ภาคผนวก ช

หลักฐานการส่งรายงานผลการปฏิบัติตามมาตรการฯ

ฉบับที่ผ่านมา



เลขที่รับ..
ลงวันที่ ๑

ที่ 66/2/21

วันที่ 25 กรกฎาคม พ.ศ. 2566

เรื่อง ขอ拿来ส่งรายงานผลการปฏิบัติตามมาตรการป้องกันและแก้ไขผลกระทบสิ่งแวดล้อม และมาตรการติดตามตรวจสอบผลกระทบสิ่งแวดล้อม โครงการโรงแรมชังกรี-ลา กรุงเทพฯ ระหว่างเดือนมกราคม-มิถุนายน พ.ศ. 2566

เรียน อธิบดีกรมการปกครอง / นายทะเบียนโรงเรม กรมการปกครอง

- อ้างถึง 1. พระราชบัญญัติส่งเสริมสิ่งแวดล้อม ราชบัญญัติส่งเสริมและรักษาคุณภาพสิ่งแวดล้อม (ฉบับที่ 2) พ.ศ. 2561
2. ประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง หลักเกณฑ์ และวิธีการจัดทำรายงานผลการปฏิบัติตามมาตรการที่กำหนดไว้ในรายงานการประเมินผลกระทบสิ่งแวดล้อม ซึ่งผู้ดำเนินการหรือผู้ขออนุญาตจะต้องจัดทำเมื่อได้รับอนุญาตให้ดำเนินโครงการหรือกิจการแล้ว พ.ศ. 2561 ลงวันที่ 19 พฤษภาคม พ.ศ. 2561
3. หนังสือสำนักงานนโยบายและแผนทรัพยากรธรรมชาติและสิ่งแวดล้อม ที่ วพ 0507/7138 ลงวันที่ 17 สิงหาคม พ.ศ. 2532

- สิ่งที่ส่งมาด้วย 1. รายงานผลการปฏิบัติตามมาตรการฯ โครงการโรงแรมชังกรี-ลา กรุงเทพฯ ระยะดำเนินการ ระหว่างเดือนมกราคม-มิถุนายน พ.ศ. 2566 จำนวน 2 ชุด
2. ไฟล์อิเล็กทรอนิกส์บันทึกลงแฟ้มชีต จำนวน 2 แผ่น

ตามหนังสือที่อ้างถึง โครงการโรงแรมชังกรี-ลา กรุงเทพฯ ของบริษัท ชังกรี-ลา โฮเต็ล จำกัด (มหาชน) ตั้งอยู่เลขที่ 89 ซอยวัดสวนพลู ถนนเจริญกรุง เขตบางรัก กรุงเทพมหานคร ได้รับความเห็นชอบรายงานการประเมินผลกระทบสิ่งแวดล้อม โดยมีเงื่อนไขให้บริษัทฯ เจ้าของโครงการปฏิบัติตามมาตรการป้องกันและแก้ไขผลกระทบสิ่งแวดล้อม และมาตรการติดตามตรวจสอบผลกระทบสิ่งแวดล้อมที่เสนอไว้ในรายงานฯ อย่างเคร่งครัด ประกอบกับพระราชบัญญัติส่งเสริมและรักษาคุณภาพสิ่งแวดล้อมแห่งชาติ (ฉบับที่ 2) พ.ศ. 2561 มาตรา 51/5 กำหนดให้เจ้าของโครงการจัดทำรายงานผลการปฏิบัติตามมาตรการฯ ระยะดำเนินการ เสนอต่ออธิบดีกรมการปกครอง / นายทะเบียนโรงเรม กรมการปกครอง เพื่อรับรวมส่งให้สำนักงานนโยบายและแผนทรัพยากรธรรมชาติและสิ่งแวดล้อมสำหรับในเขตกรุงเทพมหานคร นั้น

บริษัท ชังกรี-ลา โฮเต็ล จำกัด (มหาชน) ได้จัดทำรายงานผลการปฏิบัติตามมาตรการฯ เสร็จเรียบร้อยแล้ว ดังสิ่งที่ส่งมาด้วย 1. และ 2. จึงขอส่งให้กรมการปกครองในฐานะหน่วยงานอนุญาตดำเนินการตามอำนาจหน้าที่ต่อไป

ลงวันที่