

เอกสารแนบ 6

เอกสารสอบเทียบเครื่องมือที่ใช้ในการวิเคราะห์

Certificate of Calibration

Certificate No. : 65-400508-3 **Page : 1 of 2**

Submitted by : Ecotech Water Systems Co., Ltd.
20 Soi Kheharomklao 74 Yak 6, Ratphatthana, Saphansung, Bangkok 10240

Equipment : Autoclave
Manufacturer : LABTECH **Model :** LAC-5060S
Range : N/A °C **Resolution** 0.1 °C
Serial No. : 090414007 **ID No. :** INS008

Environment : On site calibration was carried out at the Laboratory, Ecotech Water Systems Co., Ltd.

Ambient Temperature : (26.0 to 27.0) °C

Relative Humidity : (48 to 52) %

Line Voltage : (225.0 to 226.4) V

Date of Received : 30 September 2022

Date of Calibration : 30 September 2022

Date of Issue : 01 October 2022

Calibrated by : Permpon Chanpu

Calibration Method : This instrument was calibrated by In-house method CAL-M4007 based on
BS 2646 Part5 : 1993

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Temperature Data Logger with RTD pt 100

<u>ID No.</u>	<u>Cert. No.</u>	<u>Due Date</u>	<u>Traceability</u>
400039	65-400420-1	15 Feb 2023	National Institute of Metrology Thailand (NIMT)
400040	65-400420-2	15 Feb 2023	National Institute of Metrology Thailand (NIMT)
400041	65-400420-3	15 Feb 2023	National Institute of Metrology Thailand (NIMT)

Approved by

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration


Certificate No. 65-400508-3

Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement



Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.			Uncertainty (± °C)	Measured Uniformity (°C)	Measured Stability (°C)	Sterilizing Time (minute)	Pressure Gauge Reading (kgf/cm²)
			1	2	3					
121.0	121.0	121.0	121.5	121.0	121.1	0.71	0.6	0.1	15	1.2


Remark

1. UUC : Unit Under Calibration
2. Pressure Gauge reading are out of accreditation's scope.

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- o0o -



Certificate of Calibration

Certificate No. : 65-200308-1

Page : 1 of 2

Submitted by : Ecotech Water Systems Co., Ltd.
20 Soi Kheharomklao 74 Yak 6, Ratphatthana, Saphansung, Bangkok 10240

Equipment : Electronic Balance
Manufacturer : OHAUS Model : PA214
Serial No. : 8328380168 ID No. : INS013
Capacity : 210 g Resolution : 0.0001 g

Environment : On site calibration was carried out at the Laboratory, Ecotech Water Systems Co., Ltd.
Ambient Temperature : (24.6 to 24.7) °C
Relative Humidity : (59.9 to 61.3) %
Air Pressure : 1005.0 mbar

Date of Received : 30 September 2022

Date of Calibration : 30 September 2022

Date of Issue : 01 October 2022

Calibrated by : Satja Sangkhum


Calibration Method : In-house method CAL-M2001 based on UKAS Publication ref : LAB 14
Edition 5, July 2015

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Weights

ID No.	Cert. No.	Due Date	Traceability
E261-E2624	C02213103	18 Nov 2022	National Institute of Metrology (Thailand), (NIMT)

Approved by



Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 65-200308-1

Page : 2 of 2

Result of Calibration : After Adjustment

UUC Condition As-Received : Good

Departure of indication from nominal value

Nominal Value (g)	Correction (g)	Uncertainty \pm (g)	Error before Adjustment (g)
0.01	0.0000	0.00011	0.0000
0.1	0.0000	0.00011	0.0000
1	0.0000	0.00011	-0.0001
5	0.0000	0.00011	0.0000
10	0.0000	0.00011	-0.0003
20	-0.0001	0.00011	-0.0004
50	0.0001	0.00012	-0.0014
100	0.0000	0.00020	-0.0027
150	0.0000	0.00038	-0.0042
200	0.0000	0.00038	-0.0055


This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2.11$, providing a level of confidence of approximately 95%

Eccentric error

Load test : 50 g

A B C D E
0.0002 0.0001 -0.0003 -0.0001 0.0000 g



Repeatability

Load test : 200 g

Stdev. : 0.00005 g

- o0o -



Certificate of Calibration

Certificate No. : 65-400523-1

Page : 1 of 2

Submitted by : Ecotech Water Systems Co., Ltd.

20 Soi Kheharomklao 74 Yak 6, Ratphatthana, Saphansung, Bangkok 10240

Equipment : COD Reactor

Manufacturer : Hanna

Model : HI839800

Range : N/A °C

Resolution : 0.1 °C

Serial No. : 05220009101

ID No. : N/A

Environment : Ambient Temperature : (23 ± 2) °C

Relative Humidity : (50 ± 15) %

Date of Received : 30 September 2022

Date of Calibration : 03 October 2022

Date of Issue : 03 October 2022

Calibration Method : This instrument was calibrated by In-house method direct measurement with Standard Digital Thermometer with TC Type T probe

The temperature scale used was based on ITS-90

Reference Standard Instruments :

Standard Digital Thermometer with TC Probe

<u>ID No.</u>	<u>Cert. No.</u>	<u>Due Date</u>	<u>Traceability</u>
400029 & 400030	65-400272-1	24 Nov 2022	National Institute of Metrology Thailand (NIMT)
400029 & 400032	65-400274-1	25 Nov 2022	National Institute of Metrology Thailand (NIMT)

Approved by

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.




Certificate of Calibration

Certificate No. : 65-400523-1

Page : 2 of 2

Result of Calibration : Without Adjustment

Function : Temperature measurement



Controller

Test Point (° C)	UUC Setting (° C)	UUC Reading (° C)	Standard Reading at Position (° C)									
			1	2	3	4	5	6	7	8	9	10
150.0	150.0	150.0	150.6	150.1	152.0	150.1	149.7	149.3	151.3	152.0	150.9	150.4

Test Point (° C)	UUC Setting (° C)	UUC Reading (° C)	Standard Reading at Position (° C)									
			11	12	13	14	15	16	17	18	19	20
150.0	150.0	150.0	149.4	152.0	151.4	151.0	150.1	150.3	150.6	151.8	151.7	149.9

Test Point (° C)	UUC Setting (° C)	UUC Reading (° C)	Standard Reading at Position (° C)					Uncertainty (± ° C)
			21	22	23	24	25	
150.0	150.0	150.0	150.3	149.9	151.4	150.2	150.1	0.73

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- o0o -



Certificate of Calibration

Certificate No. : 65-400508-2 **Page : 1 of 2**

Submitted by : Ecotech Water Systems Co., Ltd.
20 Soi Kheharomklao 74 Yak 6, Ratphatthana, Saphansung, Bangkok 10240

Equipment : Air Chamber (Oven)
Manufacturer : LABTECH **Model :** LDO-080F
Range : N/A °C **Resolution :** 0.1 °C
Serial No. : 081029024 **ID No. :** INS007

Environment : On site calibration was carried out at the Laboratory, Ecotech Water Systems Co., Ltd.
Ambient Temperature : (26.0 to 27.0) °C
Relative Humidity : (48 to 52) %
Line Voltage : (225.0 to 226.4) V

Date of Received : 30 September 2022

Date of Calibration : 30 September 2022

Date of Issue : 01 October 2022

Calibrated by : Permpon Chanpu

Calibration Method : CAL-M4004, TLAS G-20

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units
Standard Digital Thermometer with Thermocouple probe

<u>ID No.</u>	<u>Cert. No.</u>	<u>Due Date</u>	<u>Traceability</u>
400029 & 400030	65-400272-1	24 Nov 2022	National Institute of Metrology Thailand (NIMT)

Approved by

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 65-400508-2


Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber

W = 0.50 m

D = 0.40 m

H = 0.40 m

Capacity = 0.08 m³

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (± °C)
			1	2	3	4	5	6	7	8	9	
104.0	104.0	104.0	104.3	103.4	104.6	104.2	103.2	103.5	104.1	105.3	105.3	1.7
180.0	180.0	180.0	179.1	178.7	180.0	179.1	178.6	178.9	178.9	181.6	181.8	2.9

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Uniformity (°C)	Measured Stability (°C)	Overall Variation (°C)
104.0	104.0	104.0	2.5	1.3	4.6
180.0	180.0	180.0	4.3	2.2	7.5

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- oOo -



Certificate of Calibration

Certificate No. : 65-400508-6

Page : 1 of 2

Submitted by : Ecotech Water Systems Co., Ltd.
20 Soi Kheharomklao 74 Yak 6, Ratphatthana, Saphansung, Bangkok 10240

Equipment : Air Chamber (Incubator)
Manufacturer : Biobase Model : BJPX-B400II
Range : N/A °C Resolution : 0.1 °C
Serial No. : KYP400II2010002 ID No. : N/A

Environment : On site calibration was carried out at the Laboratory, Ecotech Water Systems Co., Ltd.
Ambient Temperature : (24.6 to 26.8) °C
Relative Humidity : (58 to 60) %
Line Voltage : (225.0 to 226.5) V

Date of Received : 30 September 2022

Date of Calibration : 30 September 2022

Date of Issue : 01 October 2022

Calibrated by : Bunjerd Masri


Calibration Method : CAL-M4004, TLAS G-20

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units
Standard Digital Thermometer with RTD Probe

ID No.	Cert. No.	Due Date	Traceability
400046 & 400047	65-400419-3	03 Feb 2023	National Institute of Metrology Thailand (NIMT)

Approved by



Supervisor

Certificate of Calibration

Certificate No. : 65-400508-6


Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber

W = 0.58 m

D = 0.55 m

H = 1.28 m

Capacity = 0.41 m³

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (± °C)
			1	2	3	4	5	6	7	8	9	
20.0	20.0	20.0	20.18	20.15	20.21	20.13	20.07	20.07	20.11	20.07	20.02	0.44

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Uniformity (°C)	Measured Stability (°C)	Overall Variation (°C)
20.0	20.0	20.0	0.22	0.20	0.5

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2 , providing a level of confidence of approximately 95%

- o0o -



Certificate of Calibration

Certificate No. : 65-420082-2

Page : 1 of 2

Submitted by : Ecotech Water Systems Co., Ltd.

20 Soi Kheharomklao 74 Yak 6, Ratphatthana, Saphansung, Bangkok 10240

Equipment : pH Meter with electrode

pH meter

Manufacturer : Eutech

Model : PC 450

Range : N/A

pH

Resolution : 0.01 pH

Serial No. : 2535550

ID No. : N/A

Electrode

Model : N/A

Serial No. : 01X099323

Environment : On site calibration was carried out at the Laboratory, Ecotech Water Systems Co., Ltd.

Ambient Temperature : (25.0 to 26.0)° C

Relative Humidity : (55 to 60) %

Date of Received : 30 September 2022

Date of Calibration : 30 September 2022

Date of Issue : 01 October 2022

Calibrated by : Permpon Chanpu

Calibration Method : In-house method CAL-M4201 direct measurement by using standard voltage calibrator and using certified reference material (CRM)

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Multiproduct Calibrator

<u>ID No.</u>	<u>Cert. No.</u>	<u>Due Date</u>	<u>Traceability</u>
400005	SG-E-00473/64	27 Aug 2023	National Institute of Metrology Thailand (NIMT)

2. Standard Buffer Solution

<u>pH</u>	<u>Cert. No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>	<u>Traceability</u>
4.008	61235182	833447	19 Aug 2024	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025
6.985	61255708	833449	19 Aug 2023	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025
10.008	61244986	833448	19 Aug 2023	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025

Approved by

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 65-420082-2

Page : 2 of 2

Result of Calibration :

UUC Condition As-Received : Good

Function : Electrical measurement

pH meter

Performing standard curve by Multiproduct Calibrator at pH (4,7,10)

Adjustment Curve at nominal pH	Applied Voltage (mV)	Nominal Value (pH)	UUC Reading		Correction (mV)	Uncertainty (± mV)
			(pH)	(mV)		
4, 7, 10	177.4800	4	4.00	177.6	-0.1	0.12
	0.0000	7	7.02	0.0	0.0	0.086
	-177.4800	10	10.00	-177.5	0.0	0.12

Function : pH meter with electrode

Performing a three - buffer standard curve using buffer nominal pH (4,7,10)

Adjustment Curve at nominal pH	Standard Buffer (pH)	UUC Reading (pH)	Correction (pH)	Uncertainty (± pH)
4, 7, 10	4.008	4.01	0.00	0.0097
	6.985	7.00	-0.01	0.011
	10.008	10.00	0.00	0.014


Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- o0o -



Certificate of Calibration

Certificate No. : 65-400509-2

Page : 1 of 2

Submitted by : Ecotech Water Systems Co., Ltd.
20 Soi Kheharomklao 74 yeak 6, Ratphatthana, Saphansung, Bangkok 10240

Equipment : Digital Thermometer with Thermistor probe
Temperature Indicator

Manufacturer : Eutech Model : PC 450

Range : N/A °C Resolution : 0.1 °C

Serial No. : 2535550 ID No. : N/A

Thermistor probe

Model : N/A Sheath Material : Stainless

Diameter : 3.5 mm. Length : 100 mm.

Serial No. : CONSEN91W 141 ID No. : N/A

Environment : On site calibration was carried out at the Laboratory, Ecotech Water Systems Co., Ltd.

Ambient Temperature : (25.0 to 26.0) °C

Relative Humidity : (55 to 60) %

Line Voltage : (225.0 to 225.9) VAC

Date of Received : 30 September 2022

Date of Calibration : 30 September 2022

Date of Issue : 01 October 2022

Calibrated by : Permpon Chanpu

Calibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4003 by compared with PRT in the liquid bath at the constant controlled temperature.

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Platinum Resistance Thermometer (PRT)

ID No.	Cert. No.	Due Date	Traceability
400002	TT-0074-22	20 Jun 2024	National Institute of Metrology Thailand (NIMT)

2. Standard Digital Thermometer

ID No.	Cert. No.	Due Date	Traceability
400033	22E569	22 Feb 2024	National Institute of Metrology Thailand (NIMT)

Approved

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 65-400509-2

Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

Immersion Depth (mm.)	Standard Reading (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
100	20.005	20.1	-0.1	0.19
100	25.006	25.1	-0.1	0.19
100	30.005	30.1	-0.1	0.19

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- o0o -



Certificate of Calibration

Certificate No. : 65-420082-1

Page : 1 of 2

Submitted by : Ecotech Water Systems Co., Ltd.

20 Soi Kheharomklao 74 Yak 6, Ratphatthana, Saphansung, Bangkok 10240

Equipment : pH Meter with electrode

pH meter

Manufacturer : Eutech

Model : PC 700

Range : N/A pH

Resolution : 0.01 pH

Serial No. : 2728583

ID No. : N/A

Electrode

Model : N/A

Serial No. : 01X099323

Environment : On site calibration was carried out at the Laboratory, Ecotech Water Systems Co., Ltd.

Ambient Temperature : (25.0 to 26.0)° C

Relative Humidity : (55 to 60) %

Date of Received : 30 September 2022

Date of Calibration : 30 September 2022

Date of Issue : 01 October 2022

Calibrated by : Permpon Chanpu

Calibration Method : In-house method CAL-M4201 direct measurement by using standard voltage calibrator and using certified reference material (CRM)

Reference Standard Instruments : This certification is traceable to the International System of Units


1. Multiproduct Calibrator

<u>ID No.</u>	<u>Cert. No.</u>	<u>Due Date</u>	<u>Traceability</u>
400005	SG-E-00473/64	27 Aug 2023	National Institute of Metrology Thailand (NIMT)

2. Standard Buffer Solution

<u>pH</u>	<u>Cert. No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>	<u>Traceability</u>
4.008	61235182	833447	19 Aug 2024	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025
6.985	61255708	833449	19 Aug 2023	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025
10.008	61244986	833448	19 Aug 2023	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025

Approved by



Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 65-420082-1

Page : 2 of 2

Result of Calibration :

UUC Condition As-Received : Good

Function : Electrical measurement

pH meter

Performing standard curve by Multiproduct Calibrator at pH (4,7,10)

Adjustment Curve at nominal pH	Applied Voltage (mV)	Nominal Value (pH)	UUC Reading		Correction (mV)	Uncertainty (± mV)
			(pH)	(mV)		
4, 7, 10	177.4800	4	4.00	166.7	10.8	0.12
	0.0000	7	7.00	-8.4	8.4	0.086
	-177.4800	10	10.00	-183.5	6.0	0.12

Function : pH meter with electrode

Performing a three - buffer standard curve using buffer nominal pH (4,7,10)

Adjustment Curve at nominal pH	Standard Buffer (pH)	UUC Reading (pH)	Correction (pH)	Uncertainty (± pH)
4, 7, 10	4.008	4.01	0.00	0.0097
	6.985	7.00	-0.01	0.011
	10.008	10.01	0.00	0.014

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- o0o -



Certificate of Calibration

Certificate No. : 65-400509-1

Page : 1 of 2

Submitted by : Ecotech Water Systems Co., Ltd.
20 Soi Kheharomklao 74 yeak 6, Ratphatthana, Saphansung, Bangkok 10240

Equipment : Digital Thermometer with Thermistor probe
Temperature Indicator

Manufacturer : Eutech Model : PC 700

Range : N/A °C Resolution : 0.1 °C

Serial No. : 2728583 ID No. : N/A

Thermistor probe

Model : N/A Sheath Material : Stainless

Diameter : 3.5 mm. Length : 100 mm.

Serial No. : CONSEN9501D 028 ID No. : N/A

Environment : On site calibration was carried out at the Laboratory, Ecotech Water Systems Co., Ltd.

Ambient Temperature : (25.0 to 26.0) °C

Relative Humidity : (55 to 60) %

Line Voltage : (225.0 to 225.9) VAC

Date of Received : 30 September 2022

Date of Calibration : 30 September 2022

Date of Issue : 01 October 2022

Calibrated by : Permpoon Chanpu

Calibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4003 by compared with PRT in the liquid bath at the constant controlled temperature.

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Platinum Resistance Thermometer (PRT)

ID No.	Cert. No.	Due Date	Traceability
400002	TT-0074-22	20 Jun 2024	National Institute of Metrology Thailand (NIMT)

2. Standard Digital Thermometer

ID No.	Cert. No.	Due Date	Traceability
400033	22E569	22 Feb 2024	National Institute of Metrology Thailand (NIMT)

Approved by

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 65-400509-1

Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

Immersion Depth (mm.)	Standard Reading (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
100	20.003	20.3	-0.3	0.19
100	25.002	25.3	-0.3	0.19
100	30.004	30.3	-0.3	0.19

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- o0o -



Certificate of Calibration

Certificate No. : 65-420082-3

Page : 1 of 2

Submitted by : Ecotech Water Systems Co., Ltd.
20 Soi Kheharomklao 74 Yak 6, Ratphatthana, Saphansung, Bangkok 10240

Equipment : pH Meter with electrode
pH meter
Manufacturer : Eutech Model : PC 700
Range : N/A pH Resolution : 0.01 pH
Serial No. : 3082600 ID No. : N/A
Electrode
Model : N/A Serial No. : 01X099320

Environment : On site calibration was carried out at the Laboratory, Ecotech Water Systems Co., Ltd.
Ambient Temperature : (25.0 to 26.0)° C
Relative Humidity : (55 to 60) %

Date of Received : 30 September 2022

Date of Calibration : 30 September 2022

Date of Issue : 01 October 2022

Calibrated by : Permpon Chanpu

Calibration Method : In-house method CAL-M4201 direct measurement by using standard voltage calibrator and using certified reference material (CRM)

Reference Standard Instruments : This certification is traceable to the International System of Units


1. Multiproduct Calibrator

<u>ID No.</u>	<u>Cert. No.</u>	<u>Due Date</u>	<u>Traceability</u>
400005	SG-E-00473/64	27 Aug 2023	National Institute of Metrology Thailand (NIMT)

2. Standard Buffer Solution

<u>pH</u>	<u>Cert. No.</u>	<u>Lot No.</u>	<u>Exp. Date</u>	<u>Traceability</u>
4.008	61235182	833447	19 Aug 2024	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025
6.985	61255708	833449	19 Aug 2023	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025
10.008	61244986	833448	19 Aug 2023	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025

Approved by



The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 65-420082-3

Page : 2 of 2

Result of Calibration :

UUC Condition As-Received : Good

Function : Electrical measurement

pH meter

Performing standard curve by Multiproduct Calibrator at pH (4,7,10)

Adjustment Curve at nominal pH	Applied Voltage (mV)	Nominal Value (pH)	UUC Reading		Correction (mV)	Uncertainty (± mV)
			(pH)	(mV)		
4, 7, 10	177.4800	4	4.00	177.4	0.1	0.12
	0.0000	7	7.00	0.0	0.0	0.086
	-177.4800	10	10.00	-177.4	-0.1	0.12

Function : pH meter with electrode

Performing a three - buffer standard curve using buffer nominal pH (4,7,10)

Adjustment Curve at nominal pH	Standard Buffer (pH)	UUC Reading (pH)	Correction (pH)	Uncertainty (± pH)
4, 7, 10	4.008	4.01	0.00	0.0097
	6.985	7.00	-0.01	0.011
	10.008	10.01	0.00	0.014

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurment was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- o0o -



Certificate of Calibration

Certificate No. : 65-400509-3

Page : 1 of 2

Submitted by : Ecotech Water Systems Co., Ltd.
20 Soi Kheharomklao 74 yeak 6, Ratphatthana, Saphansung, Bangkok 10240

Equipment : Digital Thermometer with Thermistor probe

Temperature Indicator

Manufacturer : Eutech Model : PC 700

Range : N/A °C Resolution : 0.1 °C

Serial No. : 3082600 ID No. : N/A

Thermistor probe

Model : N/A Sheath Material : Stainless

Diameter : 3.5 mm. Length : 100 mm.

Serial No. : CONSEN9501D 102 ID No. : N/A

Environment : On site calibration was carried out at the Laboratory, Ecotech Water Systems Co., Ltd.

Ambient Temperature : (25.0 to 26.0) °C

Relative Humidity : (55 to 60) %

Line Voltage : (225.0 to 225.9) VAC

Date of Received : 30 September 2022

Date of Calibration : 30 September 2022

Date of Issue : 01 October 2022

Calibrated by : Permpon Chanpu

Calibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4003 by compared with PRT in the liquid bath at the constant controlled temperature.

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units


1. Platinum Resistance Thermometer (PRT)

ID No.	Cert. No.	Due Date	Traceability
400002	TT-0074-22	20 Jun 2024	National Institute of Metrology Thailand (NIMT)

2. Standard Digital Thermometer

ID No.	Cert. No.	Due Date	Traceability
400033	22E569	22 Feb 2024	National Institute of Metrology Thailand (NIMT)

Approved



The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 65-400509-3

Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

Immersion Depth (mm.)	Standard Reading (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
100	20.002	20.2	-0.2	0.19
100	25.005	25.2	-0.2	0.19
100	30.003	30.2	-0.2	0.19

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$,
providing a level of confidence of approximately 95%

- o0o -



Certificate of Calibration

Certificate No. : 65-400508-1

Page : 1 of 2

Submitted by : Ecotech Water Systems Co., Ltd.
20 Soi Kheharomklao 74 Yak 6, Ratphatthana, Saphansung, Bangkok 10240

Equipment : Air Chamber (Refrigatorator)
Manufacturer : Every Digital Model : N/A
Range : N/A °C Resolution : 0.1 °C
Serial No. : ASS1001 ID No. : INS005

Environment : On site calibration was carried out at the Laboratory, Ecotech Water Systems Co., Ltd.
Ambient Temperature : (24.6 to 26.8) °C
Relative Humidity : (58 to 60) %
Line Voltage : (225.0 to 226.5) V

Date of Received : 30 September 2022

Date of Calibration : 30 September 2022

Date of Issue : 01 October 2022

Calibrated by : Bunjerd Masri


Calibration Method : CAL-M4004, TLAS G-20

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units
Standard Digital Thermometer with Thermocouple probe

ID No.	Cert. No.	Due Date	Traceability
400046 & 400023	65-400157-1	02 Oct 2022	National Institute of Metrology Thailand (NIMT)

Approved by



Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 65-400508-1


Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber

W = 1.00 m

D = 0.50 m

H = 1.35 m

Capacity = 0.68 m³

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (± °C)
			1	2	3	4	5	6	7	8	9	
4.0	3.0	3.0	4.3	4.1	3.9	4.2	4.1	4.2	4.1	4.3	4.0	0.63

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Uniformity (°C)	Measured Stability (°C)	Overall Variation (°C)
4.0	3.0	3.0	0.6	0.3	0.8

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- o0o -



Certificate of Calibration

Certificate No. : 65-400508-4 **Page : 1 of 2**

Submitted by : Ecotech Water Systems Co., Ltd.
20 Soi Kheharomklao 74 Yak 6, Ratphatthana, Saphansung, Bangkok 10240

Equipment : Air Chamber (Refrigator)

Manufacturer : S-Cool Model : N/A
Range : N/A °C Resolution : 1 °C
Serial No. : Eco-Ins14 ID No. : N/A

Environment : On site calibration was carried out at the Laboratory, Ecotech Water Systems Co., Ltd.

Ambient Temperature : (24.6 to 26.8) °C
Relative Humidity : (58 to 60) %
Line Voltage : (225.0 to 226.5) V

Date of Received : 30 September 2022
Date of Calibration : 30 September 2022
Date of Issue : 01 October 2022
Calibrated by : Bunjerd Masri
Calibration Method : CAL-M4004, TLAS G-20

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units
Standard Digital Thermometer with RTD Probe

ID No.	Cert. No.	Due Date	Traceability
400046 & 400043	65-400419-2	02 Feb 2023	National Institute of Metrology Thailand (NIMT)

Approved by

Supervisor



Certificate of Calibration

Certificate No. : 65-400508-4


Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber

W = 1.02 m

D = 0.44 m

H = 1.30 m

Capacity = 0.58 m³

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (± °C)
			1	2	3	4	5	6	7	8	9	
4	4	4	5.18	4.99	4.44	3.88	5.05	5.20	4.41	3.66	4.04	1.3

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Uniformity (°C)	Measured Stability (°C)	Overall Variation (°C)
4.0	4.0	4.0	1.30	0.62	2.3

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- oOo -



Certificate of Calibration

Certificate No. : 65-410116-1

Page : 1 of 2

Submitted by : Ecotech Water Systems Co., Ltd.

20 Soi Kheharomklao 74 Yak 6, Ratphatthana, Saphansung, Bangkok 10240

Equipment : Digital Thermo-Hygrometer

Manufacturer : Digicon

Model : TH-03A

Range Temperature : -10 °C to 50 °C

Resolution : 0.1 °C

Range Humidity : 20 %R.H. to 99 %R.H.

Resolution : 1 %R.H.

Serial No. : 365052106

ID No. : N/A

Environment : Ambient Temperature : (23 ± 2) °C

Relative Humidity : (50 ± 15) %

Date of Received : 30 September 2022

Date of Calibration : 03 October to 05 October 2022

Date of Issue : 05 October 2022

Calibrated by : Chortip Samchusri

Calibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4013 by compared with standard probe sensor humidity/temperature into humidity/temperature chamber.

Reference Standard Instruments : This certification is traceable to the International System of Units

Digital Indicator with Standard Probe Temp&Hum

ID No.	Cert. No.	Due Date	Traceability
400034 & 400035	SG-H-00713/65	07 Jan 2023	Success Gateway Co., Ltd., Accredited by TISI Calibration No.0268

Approved by

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 65-410116-1

Page : 2 of 2

UUC Condition As-Received : Good

Result of Calibration : Without Adjustment

Function : Temperature measurement (Mode : In)

Reference Humidity @ 50 %R.H.

Standard Temperature (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
20.00	20.5	-0.5	0.46
24.97	25.4	-0.4	0.46
30.01	29.9	0.1	0.46

Result of Calibration : Without Adjustment

Function : Humidity measurement

Reference Temperature @ 25 °C

Standard Humidity (%R.H.)	UUC Reading (%R.H.)	Correction (%R.H.)	Uncertainty (± %R.H.)
40.01	38	2	2.2
49.99	48	2	2.2
60.02	58	2	2.3


Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- o0o -



Certificate of Calibration

Certificate No. : 65-410116-2

Page : 1 of 2

Submitted by : Ecotech Water Systems Co., Ltd.

20 Soi Kheharomklao 74 Yak 6, Ratphatthana, Saphansung, Bangkok 10240

Equipment : Digital Thermo-Hygrometer

Manufacturer : Digicon

Model : TH-03A

Range Temperature : -10 °C to 50 °C Resolution : 0.1 °C

Range Humidity : 20 %R.H. to 99 %R.H. Resolution : 1 %R.H.

Serial No. : 365051554

ID No. : N/A

Environment : Ambient Temperature : (23 ± 2) °C

Relative Humidity : (50 ± 15) %

Date of Received : 30 September 2022

Date of Calibration : 03 October to 05 October 2022

Date of Issue : 05 October 2022

Calibrated by : Chortip Samchusri

Calibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4013 by compared with standard probe sensor humidity/temperature into humidity/temperature chamber.

Reference Standard Instruments : This certification is traceable to the International System of Units

Digital Indicator with Standard Probe Temp&Hum

ID No.	Cert. No.	Due Date	Traceability
400034 & 400035	SG-H-00713/65	07 Jan 2023	Success Gateway Co., Ltd., Accredited by TISI Calibration No.0268

Approved by

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 65-410116-2

Page : 2 of 2

UUC Condition As-Received : Good

Result of Calibration : Without Adjustment

Function : Temperature measurement (Mode : In)

Reference Humidity @ 50 %R.H.

Standard Temperature (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
19.99	20.3	-0.3	0.46
25.00	25.2	-0.2	0.46
30.00	29.8	0.2	0.46

Result of Calibration : Without Adjustment

Function : Humidity measurement

Reference Temperature @ 25 °C

Standard Humidity (%R.H.)	UUC Reading (%R.H.)	Correction (%R.H.)	Uncertainty (± %R.H)
39.98	39	1	2.2
50.00	49	1	2.2
59.99	59	1	2.3

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- o0o -

