



ID LINE : IEC17025



## Certificate of Calibration

Certificate Number : SPR24030285-10 Page : 1 of 3

Customer : S.P.S. CONSULTING SERVICE CO., LTD.  
7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,  
Bangkok 10900

Equipment Name : Sound Level Meter  
Manufacturer : ACO  
Model : 6236  
Serial Number : 182011  
ID. Number : ACO-B29

### Environmental Conditions

Ambient Temperature :  $23^{\circ}\text{C} \pm 3^{\circ}\text{C}$  Received Date : 19 Mar 2024  
Relative Humidity :  $50\% \pm 15\%$  Calibration Date : 23 Mar 2024  
Location of Calibration : In-Lab Recommend Due Date : 23 Mar 2025  
Calibration Procedure : SP-CPE-04-01 Date of Issue : 24 Mar 2024

### Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs.

The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Chumpon Dokpikul

Calibration Officer

Approved by :

( Mr.Prayoon Topart )

Authorized Signatory



ID LINE : IEC17025



## Calibration Report

Certificate Number : SPR24030285-10 Page : 2 of 3

### Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Sound Level Calibrator	ST-120	211203773	EEL.BP. 140/0167	26 Jan 2025

### Traceability

This certification is traceable to the International System of Unit maintained at :  
TISTR - Thailand Institute of Scientific and Technological Research



ID LINE : IEC17025



## Result of Calibration

Certificate Number : SPR24030285-10

Page : 3 of 3

Range : 94 to 114 dB

Function : @1kHz

Select A

Unit : dB

Standard Setting	UUC Reading		Error		Uncertainty ( ± )
	Fast	Slow	Fast	Slow	
94	94.1	94.1	0.1	0.1	0.15
114	114.0	114.0	0.0	0.0	0.15

Select C

Unit : dB

Standard Setting	UUC Reading		Error		Uncertainty ( ± )
	Fast	Slow	Fast	Slow	
94	94.1	94.1	0.1	0.1	0.15
114	114.0	114.0	0.0	0.0	0.15

### Note :

The result of calibration was found accurate as show on date and place of calibration only.  
This Certificate is not certified for any commercial transaction.

### Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2.00$ , providing a level of confidence approximately 95%.

- End of Certificate -



ID LINE : IEC17025



## Certificate of Calibration

Certificate Number : SPR24030285-11

Page : 1 of 3

Customer

: S.P.S. CONSULTING SERVICE CO., LTD.

7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,  
Bangkok 10900

Equipment Name : Sound Level Meter

Manufacturer : ACO

Model : 6236

Serial Number : 192027

ID. Number : ACO-B36

### Environmental Conditions

Ambient Temperature :  $23^{\circ}\text{C} \pm 3^{\circ}\text{C}$

Received Date : 19 Mar 2024

Relative Humidity :  $50\% \pm 15\%$

Calibration Date : 23 Mar 2024

Location of Calibration : In-Lab

Recommend Due Date : 23 Mar 2025

Calibration Procedure : SP-CPE-04-01

Date of Issue : 24 Mar 2024

### Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs. The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Chumpon Dokpikul

Calibration Officer

Approved by :

( Mr.Prayoon Topart )

Authorized Signatory



ID LINE : IEC17025



## Calibration Report

Certificate Number : SPR24030285-11

Page : 2 of 3

### Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Sound Level Calibrator	ST-120	211203773	EEL.BP. 140/0167	26 Jan 2025

### Traceability

This certification is traceable to the International System of Unit maintained at :  
TISTR - Thailand Institute of Scientific and Technological Research



ID LINE : IEC17025



## Result of Calibration

Certificate Number : SPR24030285-11

Page : 3 of 3

Range : 94 to 114 dB

Function : @1kHz

Select A

Unit : dB

Standard Setting	UUC Reading		Error		Uncertainty ( ± )
	Fast	Slow	Fast	Slow	
94	93.9	93.9	-0.1	-0.1	0.15
114	114.1	114.1	0.1	0.1	0.15

Select C

Unit : dB

Standard Setting	UUC Reading		Error		Uncertainty ( ± )
	Fast	Slow	Fast	Slow	
94	93.9	93.9	-0.1	-0.1	0.15
114	114.1	113.9	0.1	-0.1	0.15

### Note :

The result of calibration was found accurate as show on date and place of calibration only.  
This Certificate is not certified for any commercial transaction.

### Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2.00$ , providing a level of confidence approximately 95%.

- End of Certificate -



ID LINE : IEC17025



## Certificate of Calibration

Certificate Number : SPR24030285-12 Page : 1 of 3

Customer : S.P.S. CONSULTING SERVICE CO., LTD.  
7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,  
Bangkok 10900

Equipment Name : Sound Level Meter  
Manufacturer : ACO  
Model : 6236  
Serial Number : 192032  
ID. Number : ACO-B41

### Environmental Conditions

Ambient Temperature : 23 °C ± 3 °C Received Date : 19 Mar 2024  
Relative Humidity : 50 % ± 15 % Calibration Date : 23 Mar 2024  
Location of Calibration : In-Lab Recommend Due Date : 23 Mar 2025  
Calibration Procedure : SP-CPE-04-01 Date of Issue : 24 Mar 2024

### Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs.

The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Chumpon Dokpikul

Calibration Officer

Approved by :

( Mr.Prayoon Topart )

Authorized Signatory



ID LINE : IEC17025



## Calibration Report

Certificate Number : SPR24030285-12 Page : 2 of 3

### Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Sound Level Calibrator	ST-120	211203773	EEL.BP. 140/0167	26 Jan 2025

### Traceability

This certification is traceable to the International System of Unit maintained at :  
TISTR - Thailand Institute of Scientific and Technological Research



ID LINE : IEC17025



## Result of Calibration

Certificate Number : SPR24030285-12

Page : 3 of 3

Range : 94 to 114 dB

Function : @1kHz

Select A

Unit : dB

Standard Setting	UUC Reading		Error		Uncertainty ( ± )
	Fast	Slow	Fast	Slow	
94	94.0	94.0	0.0	0.0	0.15
114	114.0	114.0	0.0	0.0	0.15

Select C

Unit : dB

Standard Setting	UUC Reading		Error		Uncertainty ( ± )
	Fast	Slow	Fast	Slow	
94	94.0	94.0	0.0	0.0	0.15
114	114.0	114.0	0.0	0.0	0.15

### Note :

The result of calibration was found accurate as show on date and place of calibration only.  
This Certificate is not certified for any commercial transaction.

### Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2.00$ , providing a level of confidence approximately 95%.

- End of Certificate -



ID LINE : IEC17025



## Certificate of Calibration

Certificate Number : SPR24030285-13

Page : 1 of 3

Customer

: S.P.S. CONSULTING SERVICE CO., LTD.

7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,  
Bangkok 10900

Equipment Name : Sound Level Meter

Manufacturer : ACO

Model : 6236

Serial Number : 192034

ID. Number : ACO-B43

### Environmental Conditions

Ambient Temperature :  $23^{\circ}\text{C} \pm 3^{\circ}\text{C}$

Received Date : 19 Mar 2024

Relative Humidity :  $50\% \pm 15\%$

Calibration Date : 23 Mar 2024

Location of Calibration : In-Lab

Recommend Due Date : 23 Mar 2025

Calibration Procedure : SP-CPE-04-01

Date of Issue : 24 Mar 2024

### Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs. The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Chumpon Dokpikul

Calibration Officer

Approved by :

( Mr.Prayoon Topart )

Authorized Signatory



# Calibration Report

Certificate Number : SPR24030285-13      Page : 2 of 3

## Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Sound Level Calibrator	ST-120	211203773	EEL.BP. 140/0167	26 Jan 2025

### Traceability

This certification is traceable to the International System of Unit maintained at :  
TISTR - Thailand Institute of Scientific and Technological Research

69/29 Moo 1 Klongsi Klonguang Pathumthani 12120 ( Thailand ) Tel: (662) 193-2220 5 ผู้ขาย www.สอบเทียบเครื่องมือวัด.com



# Result of Calibration

Certificate Number : SPR24030285-13      Page : 3 of 3

Range : 94 to 114 dB      Function : @1kHz

Select A      Unit : dB

Standard Setting	UUC Reading		Error		Uncertainty ( ± )
	Fast	Slow	Fast	Slow	
94	94.0	94.0	0.0	0.0	0.15
114	113.9	113.9	-0.1	-0.1	0.15

Select C      Unit : dB

Standard Setting	UUC Reading		Error		Uncertainty ( ± )
	Fast	Slow	Fast	Slow	
94	94.0	94.0	0.0	0.0	0.15
114	113.9	113.9	-0.1	-0.1	0.15

### Note :

The result of calibration was found accurate as show on date and place of calibration only.  
This Certificate is not certified for any commercial transaction.

### Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor k = 2.00, providing a level of confidence approximately 95%.  
- End of Certificate -

69/29 Moo 1 Klongsi Klonguang Pathumthani 12120 ( Thailand ) Tel: (662) 193-2220 5 ผู้ขาย www.สอบเทียบเครื่องมือวัด.com



## Certificate of Calibration

Certificate Number : SPR23090267-7 Page : 1 of 3

Customer : S.P.S. CONSULTING SERVICE CO., LTD.  
7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,  
Bangkok 10900

Equipment Name : Sound Level Meter  
Manufacturer : ACO  
Model : 6236  
Serial Number : 192052  
ID. Number : R-40

### Environmental Conditions

Ambient Temperature :  $23^{\circ}\text{C} \pm 3^{\circ}\text{C}$  Received Date : 18 Sep 2023  
Relative Humidity :  $50\% \pm 15\%$  Calibration Date : 18 Sep 2023  
Location of Calibration : In-Lab Recommend Due Date : 18 Sep 2024  
Calibration Procedure : SP-CPE-04-01 Date of Issue : 19 Sep 2023

### Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs.

The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Karoon Pengsalung

Calibration Officer

Approved by :

( Mr.Nirut Loha )

Authorized Signatory



## Calibration Report

Certificate Number : SPR23090267-7 Page : 2 of 3

### Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Sound Level Calibrator	ST-120	211203773	EEL.BP. 114/0166	17 Jan 2024

### Traceability

This certification is traceable to the International System of Unit maintained at :  
TISTR - Thailand Institute of Scientific and Technological Research



## Result of Calibration

Certificate No. : SPR23090267-7

Page : 3 of 3

Range : 94 to 114 dB

Function : @1kHz

Select A Unit : dB

Standard Setting	UUC Reading		Error		Uncertainty ( ± )
	Fast	Slow	Fast	Slow	
94	94.0	94.0	0.0	0.0	0.15
114	113.8	113.8	-0.2	-0.2	0.15

Select C Unit : dB

Standard Setting	UUC Reading		Error		Uncertainty ( ± )
	Fast	Slow	Fast	Slow	
94	93.9	93.9	-0.1	-0.1	0.15
114	113.7	113.7	-0.3	-0.3	0.15

### Note:

The result of calibration was found accurate as show on date and place of calibration only.  
This Certificate is not certified for any commercial transaction.

### Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2.00$ , providing a level of confidence approximately 95%.

- End of Certificate -



## Certificate of Calibration

Certificate Number : SPR23090267-8

Page : 1 of 3

Customer : S.P.S. CONSULTING SERVICE CO., LTD.

7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,  
Bangkok 10900

Equipment Name : Sound Level Meter

Manufacturer : ACO

Model : 6236

Serial Number : 192053

ID. Number : R-41

### Environmental Conditions

Ambient Temperature : $23^{\circ}\text{C} \pm 3^{\circ}\text{C}$	Received Date : 18 Sep 2023
Relative Humidity : $50\% \pm 15\%$	Calibration Date : 18 Sep 2023
Location of Calibration : In-Lab	Recommend Due Date : 18 Sep 2024
Calibration Procedure : SP-CPE-04-01	Date of Issue : 19 Sep 2023

### Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs. The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Karoon Pengsalung

Calibration Officer

Approved by :

( Mr.Nirut Loha )

Authorized Signatory





## Certificate of Calibration

Certificate Number : SPR23090267-9

Page : 1 of 3

Customer : S.P.S. CONSULTING SERVICE CO., LTD.

7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,  
Bangkok 10900

Equipment Name : Sound Level Meter

Manufacturer : ACO

Model : 6236

Serial Number : 192062

ID, Number : R-50

### Environmental Conditions

Ambient Temperature :  $23^{\circ}\text{C} \pm 3^{\circ}\text{C}$

Relative Humidity :  $50\% \pm 15\%$

Location of Calibration : In-Lab

Calibration Procedure : SP-CPE-04-01

Received Date : 18 Sep 2023

Calibration Date : 18 Sep 2023

Recommend Due Date : 18 Sep 2024

Date of Issue : 19 Sep 2023

### Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs.

The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Karoon Pengsalung

Calibration Officer

Approved by :

( Mr.Nirut Loha )

Authorized Signatory



## Calibration Report

Certificate Number : SPR23090267-9

Page : 2 of 3

### Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Sound Level Calibrator	ST-120	211203773	EEL.BP. 114/0166	17 Jan 2024

### Traceability

This certification is traceable to the International System of Unit maintained at :  
TISTR - Thailand Institute of Scientific and Technological Research



## Result of Calibration

Certificate No. : SPR23090267-9

Page : 3 of 3

Range : 94 to 114 dB

Function : @1kHz

Select A Unit : dB

Standard Setting	UUC Reading		Error		Uncertainty ( ± )
	Fast	Slow	Fast	Slow	
94	94.0	94.0	0.0	0.0	0.15
114	113.9	113.9	-0.1	-0.1	0.15

Select C Unit : dB

Standard Setting	UUC Reading		Error		Uncertainty ( ± )
	Fast	Slow	Fast	Slow	
94	94.0	93.9	0.0	-0.1	0.15
114	113.9	113.9	-0.1	-0.1	0.15

### Note:

The result of calibration was found accurate as show on date and place of calibration only.  
This Certificate is not certified for any commercial transaction.

### Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2.00$ , providing a level of confidence approximately 95%.

- End of Certificate -



## Certificate of Calibration

Certificate Number : SPR23090267-10

Page : 1 of 3

Customer : S.P.S. CONSULTING SERVICE CO., LTD.

7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,  
Bangkok 10900

Equipment Name : Sound Level Meter

Manufacturer : ACO

Model : 6236

Serial Number : 192063

ID. Number : R-51

### Environmental Conditions

Ambient Temperature : $23^{\circ}\text{C} \pm 3^{\circ}\text{C}$	Received Date : 18 Sep 2023
Relative Humidity : $50\% \pm 15\%$	Calibration Date : 18 Sep 2023
Location of Calibration : In-Lab	Recommend Due Date : 18 Sep 2024
Calibration Procedure : SP-CPE-04-01	Date of Issue : 19 Sep 2023

### Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs. The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Karoon Pengsalung

Calibration Officer

Approved by :

( Mr.Nirut Loha )

Authorized Signatory



## Calibration Report

Certificate Number : SPR23090267-10

Page : 2 of 3

### Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Sound Level Calibrator	ST-120	211203773	EEL.BP. 114/0166	17 Jan 2024

### Traceability

This certification is traceable to the International System of Unit maintained at :  
TISTR - Thailand Institute of Scientific and Technological Research



## Result of Calibration

Certificate No. : SPR23090267-10

Page : 3 of 3

Range : 94 to 114 dB

Function : @1kHz

Select A

Unit : dB

Standard Setting	UUC Reading		Error		Uncertainty ( ± )
	Fast	Slow	Fast	Slow	
94	94.0	94.0	0.0	0.0	0.15
114	113.8	113.8	-0.2	-0.2	0.15

Select C

Unit : dB

Standard Setting	UUC Reading		Error		Uncertainty ( ± )
	Fast	Slow	Fast	Slow	
94	93.9	93.9	-0.1	-0.1	0.15
114	113.8	113.8	-0.2	-0.2	0.15

### Note:

The result of calibration was found accurate as show on date and place of calibration only.  
This Certificate is not certified for any commercial transaction.

### Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2.00$ , providing a level of confidence approximately 95%.

- End of Certificate -



## Certificate of Calibration

Certificate Number : SPR23090267-11

Page : 1 of 3

Customer : S.P.S. CONSULTING SERVICE CO., LTD.

7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,  
Bangkok 10900

Equipment Name : Sound Level Meter

Manufacturer : ACO

Model : 6236

Serial Number : 192064

ID. Number : R-52

### Environmental Conditions

Ambient Temperature :  $23^{\circ}\text{C} \pm 3^{\circ}\text{C}$

Relative Humidity :  $50\% \pm 15\%$

Location of Calibration : In-Lab

Calibration Procedure : SP-CPE-04-01

Received Date : 18 Sep 2023

Calibration Date : 18 Sep 2023

Recommend Due Date : 18 Sep 2024

Date of Issue : 19 Sep 2023

### Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs.

The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Karoon Pengsalung

Calibration Officer

Approved by :

( Mr.Nirut Loha )

Authorized Signatory



## Calibration Report

Certificate Number : SPR23090267-11

Page : 2 of 3

### Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Sound Level Calibrator	ST-120	211203773	EEL.BP. 114/0166	17 Jan 2024

### Traceability

This certification is traceable to the International System of Unit maintained at :

TISTR - Thailand Institute of Scientific and Technological Research











THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-67/0349

MTC No. EEL. BP. 61/0367

### CALIBRATION CERTIFICATE

Submitted by : S.P.S.Consulting Service Co., Ltd.

Address : 7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok, 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

#### Instrument Calibrated :

Description : Noise Dosimeter

Manufacturer : Svantek

Model : SV-104IS

Serial No. : 80854

#### Ambient Environment

Temperature :  $(23 \pm 3) ^\circ\text{C}$

Relative Humidity :  $(50 \pm 15) \%$

Ambient Pressure :  $(101.325 \pm 1.5) \text{ kPa}$

#### Standards used :

Multifunction Acoustic Calibrator Brüel&Kjær 4226 S/N 2810358 with Coupler UA0915 S/N 2810358.

#### Calibration Procedure :

This instrument was calibrated by using calibration procedure no CP-102-01, which was based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). This calibration procedure was related to the acoustical signal test of frequency weightings using a multifunction acoustic calibrator.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%.

Date of Receipt : 15 Mar. 2024

Date of Calibration : 26 Mar. 2024

1 / 2

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.5



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-67/0349

MTC No. EEL. BP. 61/0367

### Acoustic signal test of frequency weightings

Frequency (Hz)	Deviation from response curve		Uncertainty ( $\pm\text{dB}$ )	Tolerance Limits Class 2 ( $\pm\text{dB}$ )
	A-weighting (dB)	C-weighting (dB)		
125	-0.6	-0.6	0.25	2.0
1 000	0.0	-0.1	0.25	1.4
4 000	-0.4	-0.4	0.25	3.6

Note : 1) There was no adjustment.

2) The calibration was performed at a sound pressure level of 114 dB.

3) The measured values did not include the correction of microphone of UUT.

4) The deviation was produced from the absolute difference between the measured values and the responding sound pressure levels in IEC 61672-1 (2002).

Calibrated by :

*G. Sanaey*

(Mr. Sanaey Grajang)

Approved by :

*Prawate Kluaypa*

(Mr. Prawate Kluaypa)

Director

Electrical and Electronic Standards Laboratory

Industrial Metrology and Testing Service Centre

Ref : 2011267031501086007

Date of Calibration : 26 Mar. 2024

Date of Issue : 27 Mar. 2024

2 / 2

End of Certificate

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.5

#### Head Office

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9036  
Fax. (66) 0 2577 9009

#### Office/Laboratory

668 Mu 2 Tambon Bangpoochai, Amphoe Muang Samutprakan,  
Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
(66) 08 3219 9440  
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

#### Office

196 Phahonyothin Road, Ladyao, Chatuchak,  
Bangkok 10900, Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
(66) 08 1889 6827

#### Head Office

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9036  
Fax. (66) 0 2577 9009

#### Office/Laboratory

668 Mu 2 Tambon Bangpoochai, Amphoe Muang Samutprakan,  
Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
(66) 08 3219 9440  
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

#### Office

196 Phahonyothin Road, Ladyao, Chatuchak,  
Bangkok 10900, Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
(66) 08 1889 6827



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-67/0349

MTC No. EEL. BP. 62/0367

### CALIBRATION CERTIFICATE

Submitted by : S.P.S.Consulting Service Co., Ltd.

Address : 7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok, 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

**Instrument Calibrated :**

**Ambient Environment**

Description : Noise Dosimeter

Temperature :  $(23 \pm 3) ^\circ\text{C}$

Manufacturer : Svantek

Relative Humidity :  $(50 \pm 15) \%$

Model : SV-104IS

Ambient Pressure :  $(101.325 \pm 1.5) \text{ kPa}$

Serial No. : 80856

**Standards used :**

Multifunction Acoustic Calibrator Brüel&Kjær 4226 S/N 2810358 with Coupler UA0915 S/N 2810358.

**Calibration Procedure :**

This instrument was calibrated by using calibration procedure no CP-102-01, which was based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). This calibration procedure was related to the acoustical signal test of frequency weightings using a multifunction acoustic calibrator.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%.

Date of Receipt : 15 Mar. 2024

Date of Calibration : 26 Mar. 2024

1 / 2

G. Sany

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-67/0349

MTC No. EEL. BP. 62/0367

### Acoustic signal test of frequency weightings

Frequency (Hz)	Deviation from response curve		Uncertainty ( $\pm\text{dB}$ )	Tolerance Limits Class 2 ( $\pm\text{dB}$ )
	A-weighting (dB)	C-weighting (dB)		
125	-0.4	-0.4	0.25	2.0
1 000	0.0	-0.1	0.25	1.4
4 000	0.5	0.3	0.25	3.6

Note : 1) There was no adjustment.

2) The calibration was performed at a sound pressure level of 114 dB.

3) The measured values did not include the correction of microphone of UUT.

4) The deviation was produced from the absolute difference between the measured values and the responding sound pressure levels in IEC 61672-1 (2002).

Calibrated by :

G. Sany

(Mr. Sanaey Grajang)

Approved by :

Prawate Kluaypa

(Mr. Prawate Kluaypa)

Director

Electrical and Electronic Standards Laboratory

Industrial Metrology and Testing Service Centre

Ref : 2011267031501086008

Date of Calibration : 26 Mar. 2024

Date of Issue : 27 Mar. 2024

2 / 2

End of Certificate

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9036  
Fax. (66) 0 2577 9009

Office/Laboratory  
668 Mu 2 Tambon Bangpoomai, Amphoe Muang Samutprakan,  
Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
(66) 08 3219 9440  
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office  
196 Phahonyothin Road, Ladyao, Chatuchak,  
Bangkok 10900, Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
(66) 08 1889 6827

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9036  
Fax. (66) 0 2577 9009

Office/Laboratory  
668 Mu 2 Tambon Bangpoomai, Amphoe Muang Samutprakan,  
Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
(66) 08 3219 9440  
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office  
196 Phahonyothin Road, Ladyao, Chatuchak,  
Bangkok 10900, Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
(66) 08 1889 6827

Request No. 21-66/0475

MTC No. EEL. BP. 39/0566

### CALIBRATION CERTIFICATE

Submitted by : S.P.S.Consulting Service Co., Ltd.

Address : 7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok, 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

#### Instrument Calibrated :

Description : Noise Dosimeter

Manufacturer : Svantek

Model : SV-104IS

Serial No. : 80816

#### Ambient Environment

Temperature :  $(23 \pm 3) ^\circ\text{C}$ 

Relative Humidity :  $(50 \pm 15) \%$ 

Ambient Pressure :  $(101.325 \pm 1.5) \text{ kPa}$ 

#### Standards used :

Multifunction Acoustic Calibrator Brüel&amp;Kjær 4226 S/N 2810358 with Coupler UA0915 S/N 2810358.

#### Calibration Procedure :

This instrument was calibrated by using calibration procedure no CP-102-01, which was based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). This calibration procedure was related to the acoustical signal test of frequency weightings using a multifunction acoustic calibrator.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%.

Date of Receipt : 11 May 2023

Date of Calibration : 23 May 2023

1 / 2

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9000  
Fax. (66) 0 2577 9009  
E-mail : rumpai@tistr.or.th Website:www.tistr.or.th

Office/Laboratory  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,  
Amphoe Muang, Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
Fax. (66) 0 2323 9165  
E-mail : mtc@tistr.or.th

Office  
196 Phahonyothin Road, Chatuchak, Bangkok 10900,  
Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
Fax. (66) 0 2579 8592  
E-mail : sumalee@tistr.or.th

Request No. 21-66/0475

MTC No. EEL. BP. 39/0566

### Acoustic signal test of frequency weightings

Frequency (Hz)	Deviation from response curve		Uncertainty ( $\pm$ dB)	Tolerance Limits Class 2 ( $\pm$ dB)
	A-weighting (dB)	C-weighting (dB)		
125	-0.1	-0.3	0.25	2.0
1 000	0.0	-0.1	0.25	1.4
4 000	-0.1	-0.1	0.25	3.6

- Note :
- 1) There was no adjustment.
  - 2) The calibration was performed at a sound pressure level of 114 dB.
  - 3) The measured values did not include the correction of microphone of UUT.
  - 4) The deviation was produced from the absolute difference between the measured values and the responding sound pressure levels in IEC 61672-1 (2002).

Calibrated by :

*G. Sanaey*  
.....  
(Mr. Sanaey Grajang)

Approved by :

*Prawate Kluyap*  
.....  
(Mr. Prawate Kluyap)  
Director

Electrical and Electronic Standards Laboratory  
Industrial Metrology and Testing Service Centre

Ref : 2011266051101784001

Date of Calibration : 23 May 2023

Date of Issue : 24 May 2023

2 / 2

End of Certificate

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9000  
Fax. (66) 0 2577 9009  
E-mail : rumpai@tistr.or.th Website:www.tistr.or.th

Office/Laboratory  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,  
Amphoe Muang, Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
Fax. (66) 0 2323 9165  
E-mail : mtc@tistr.or.th

Office  
196 Phahonyothin Road, Chatuchak, Bangkok 10900,  
Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
Fax. (66) 0 2579 8592  
E-mail : sumalee@tistr.or.th



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-66/0475

MTC No. EEL. BP. 40/0566

### CALIBRATION CERTIFICATE

Submitted by : S.P.S.Consulting Service Co., Ltd.

Address : 7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok, 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

Instrument Calibrated :

Ambient Environment

Description : Noise Dosimeter

Temperature :  $(23 \pm 3) ^\circ\text{C}$

Manufacturer : Svantek

Relative Humidity :  $(50 \pm 15) \%$

Model : SV-104IS

Ambient Pressure :  $(101.325 \pm 1.5) \text{ kPa}$

Serial No. : 80817

Standards used :

Multifunction Acoustic Calibrator Brüel&Kjær 4226 S/N 2810358 with Coupler UA0915 S/N 2810358.

Calibration Procedure :

This instrument was calibrated by using calibration procedure no CP-102-01, which was based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). This calibration procedure was related to the acoustical signal test of frequency weightings using a multifunction acoustic calibrator.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%.

Date of Receipt : 11 May 2023

Date of Calibration : 23 May 2023

1 / 2

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-66/0475

MTC No. EEL. BP. 40/0566

### Acoustic signal test of frequency weightings

Frequency (Hz)	Deviation from response curve		Uncertainty ( $\pm$ dB)	Tolerance Limits Class 2 ( $\pm$ dB)
	A-weighting (dB)	C-weighting (dB)		
125	-0.9	-0.9	0.25	2.0
1 000	0.0	0.0	0.25	1.4
4 000	-1.0	-1.1	0.25	3.6

Note : 1) There was no adjustment.

2) The calibration was performed at a sound pressure level of 114 dB.

3) The measured values did not include the correction of microphone of UUT.

4) The deviation was produced from the absolute difference between the measured values and the responding sound pressure levels in IEC 61672-1 (2002).

Calibrated by :

*G. Sanaey*  
(Mr. Sanaey Grajang)

Approved by :

*Prawate Kluaypa*  
(Mr. Prawate Kluaypa)  
Director

Electrical and Electronic Standards Laboratory  
Industrial Metrology and Testing Service Centre

Ref : 2011266051101784002

Date of Calibration : 23 May 2023

Date of Issue : 24 May 2023

2 / 2

End of Certificate

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9000  
Fax. (66) 0 2577 9009  
E-mail : rumpai@tistr.or.th Website:www.tistr.or.th

Office/Laboratory  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,  
Amphoe Muang, Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
Fax. (66) 0 2323 9165  
E-mail : mtc@tistr.or.th

Office  
196 Phahonyothin Road, Chatuchak, Bangkok 10900,  
Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
Fax. (66) 0 2579 8592  
E-mail : sumalee@tistr.or.th

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9000  
Fax. (66) 0 2577 9009  
E-mail : rumpai@tistr.or.th Website:www.tistr.or.th

Office/Laboratory  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,  
Amphoe Muang, Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
Fax. (66) 0 2323 9165  
E-mail : mtc@tistr.or.th

Office  
196 Phahonyothin Road, Chatuchak, Bangkok 10900,  
Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
Fax. (66) 0 2579 8592  
E-mail : sumalee@tistr.or.th















THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-67/0114

MTC No. EEL. BP. 71/1166

## CALIBRATION CERTIFICATE

Submitted by : S.P.S.Consulting Service Co., Ltd.

Address : 7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok, 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

### Instrument Calibrated :

Description : Noise Dosimeter

Manufacturer : Svantek

Model : SV-104IS

Serial No. : 80875

### Ambient Environment

Temperature :  $(23 \pm 3) ^\circ\text{C}$

Relative Humidity :  $(50 \pm 15) \%$

Ambient Pressure :  $(101.325 \pm 1.5) \text{ kPa}$

### Standards used :

Multifunction Acoustic Calibrator Brüel&Kjær 4226 S/N 2810358 with Coupler UA0915 S/N 2810358.

### Calibration Procedure :

This instrument was calibrated by using calibration procedure no CP-102-01, which was based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). This calibration procedure was related to the acoustical signal test of frequency weightings using a multifunction acoustic calibrator.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%.

Date of Receipt : 21 Nov. 2023

Date of Calibration : 27 Nov. 2023

1/2  
G. Sany

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

### Head Office

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9000  
Fax. (66) 0 2577 9009  
E-mail : rumpai@tistr.or.th Website:www.tistr.or.th

### Office/Laboratory

Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,  
Amphoe Muang, Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
Fax. (66) 0 2323 9165  
E-mail : mtc@tistr.or.th

### Office

196 Phahonyothin Road, Chatuchak, Bangkok 10900,  
Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
Fax. (66) 0 2579 8592  
E-mail : sumalee@tistr.or.th



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-67/0114

MTC No. EEL. BP. 71/1166

## Acoustic signal test of frequency weightings

Frequency (Hz)	Deviation from response curve		Uncertainty ( $\pm\text{dB}$ )	Tolerance Limits Class 2 ( $\pm\text{dB}$ )
	A-weighting (dB)	C-weighting (dB)		
125	0.1	0.1	0.25	2.0
1 000	0.0	-0.1	0.25	1.4
4 000	0.2	0.2	0.25	3.6

Note : 1) There was no adjustment.

2) The calibration was performed at a sound pressure level of 114 dB.

3) The measured values did not include the correction of microphone of UUT.

4) The deviation was produced from the absolute difference between the measured values and the responding sound pressure levels in IEC 61672-1 (2002).

Calibrated by :

*G. Sany*  
.....  
(Mr. Sanaey Grajang)

Approved by :

*Prawate Kluaypa*  
.....  
(Mr. Prawate Kluaypa)

**TISTR**

Electrical and Electronic Standards Laboratory  
Industrial Metrology and Testing Service Centre

Ref : 2011266112104685004

Date of Calibration : 27 Nov. 2023

Date of Issue : 28 Nov. 2023

2 / 2

End of Certificate

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

### Head Office

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9000  
Fax. (66) 0 2577 9009  
E-mail : rumpai@tistr.or.th Website:www.tistr.or.th

### Office/Laboratory

Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,  
Amphoe Muang, Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
Fax. (66) 0 2323 9165  
E-mail : mtc@tistr.or.th

### Office

196 Phahonyothin Road, Chatuchak, Bangkok 10900,  
Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
Fax. (66) 0 2579 8592  
E-mail : sumalee@tistr.or.th



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-67/0114

MTC No. EEL. BP. 72/1166

## CALIBRATION CERTIFICATE

Submitted by : S.P.S.Consulting Service Co., Ltd.

Address : 7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok, 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

### Instrument Calibrated :

Description : Noise Dosimeter

Manufacturer : Svantek

Model : SV-104IS

Serial No. : 80880

### Standards used :

Multifunction Acoustic Calibrator Brüel&Kjær 4226 S/N 2810358 with Coupler UA0915 S/N 2810358.

### Calibration Procedure :

This instrument was calibrated by using calibration procedure no CP-102-01, which was based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). This calibration procedure was related to the acoustical signal test of frequency weightings using a multifunction acoustic calibrator.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%.

Date of Receipt : 21 Nov. 2023

Date of Calibration : 27 Nov. 2023

1 / 2  
G. Samy

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9000  
Fax. (66) 0 2577 9009  
E-mail : rumpai@tistr.or.th Website:www.tistr.or.th

Office/Laboratory  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,  
Amphoe Muang, Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
Fax. (66) 0 2323 9165  
E-mail : mtc@tistr.or.th

Office  
196 Phahonyothin Road, Chatuchak, Bangkok 10900,  
Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
Fax. (66) 0 2579 8592  
E-mail : sumalee@tistr.or.th



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-67/0114

MTC No. EEL. BP. 72/1166

### Acoustic signal test of frequency weightings

Frequency (Hz)	Deviation from response curve		Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)		
125	-0.2	-0.6	0.25	2.0
1 000	0.0	-0.1	0.25	1.4
4 000	0.0	0.1	0.25	3.6

Note : 1) There was no adjustment.

2) The calibration was performed at a sound pressure level of 114 dB.

3) The measured values did not include the correction of microphone of UUT.

4) The deviation was produced from the absolute difference between the measured values and the responding sound pressure levels in IEC 61672-1 (2002).

Calibrated by :

*G. Samy*  
.....  
(Mr. Sanaey Grajang)

Approved by :

*Prawate Klauyapa*  
.....  
(Mr. Prawate Klauyapa)

Electrical and Electronic Standards Laboratory  
Industrial Metrology and Testing Service Centre

Ref : 2011266112104685005

Date of Calibration : 27 Nov. 2023

Date of Issue : 28 Nov. 2023

2 / 2

End of Certificate

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9000  
Fax. (66) 0 2577 9009  
E-mail : rumpai@tistr.or.th Website:www.tistr.or.th

Office/Laboratory  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,  
Amphoe Muang, Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
Fax. (66) 0 2323 9165  
E-mail : mtc@tistr.or.th

Office  
196 Phahonyothin Road, Chatuchak, Bangkok 10900,  
Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
Fax. (66) 0 2579 8592  
E-mail : sumalee@tistr.or.th



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-67/0380

MTC No. EEL. BP. 13/0467

### CALIBRATION CERTIFICATE

Submitted by : S.P.S.Consulting Service Co., Ltd.

Address : 7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok, 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

**Instrument Calibrated :**

Description : Noise Dosimeter

Manufacturer : Syntek

Model : SV-104IS

Serial No. : 106120

**Ambient Environment**

Temperature :  $(23 \pm 3) ^\circ\text{C}$

Relative Humidity :  $(50 \pm 15) \%$

Ambient Pressure :  $(101.325 \pm 1.5) \text{ kPa}$

**Standards used :**

Multifunction Acoustic Calibrator Brüel&Kjær 4226 S/N 2810358 with Coupler UA0915 S/N 2810358.

**Calibration Procedure :**

This instrument was calibrated by using calibration procedure no CP-102-01, which was based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). This calibration procedure was related to the acoustical signal test of frequency weightings using a multifunction acoustic calibrator.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%.

Date of Receipt : 1 Apr. 2024

Date of Calibration : 10 Apr. 2024

1/2  
G. Sany

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-67/0380

MTC No. EEL. BP. 13/0467

### Acoustic signal test of frequency weightings

Frequency (Hz)	Deviation from response curve		Uncertainty ( $\pm$ dB)	Tolerance Limits Class 2 ( $\pm$ dB)
	A-weighting (dB)	C-weighting (dB)		
125	0.5	0.5	0.25	2.0
1 000	0.2	-0.1	0.25	1.4
4 000	0.2	-0.3	0.25	3.6

Note : 1) There was no adjustment.

2) The calibration was performed at a sound pressure level of 114 dB.

3) The measured values did not include the correction of microphone of UUT.

4) The deviation was produced from the absolute difference between the measured values and the responding sound pressure levels in IEC 61672-1 (2002).

Calibrated by :

G. Sany  
(Mr. Sanaey Grajang)

Approved by :

Prawate Kluyap  
(Mr. Prawate Kluyap)  
Director

Electrical and Electronic Standards Laboratory

Industrial Metrology and Testing Service Centre

Ref : 2011267040101274006

Date of Calibration : 10 Apr. 2024

Date of Issue : 11 Apr. 2024

2 / 2

End of Certificate

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9036  
Fax. (66) 0 2577 9009

Office/Laboratory  
668 Mu 2 Tambon Bangpoomai, Amphoe Muang Samutprakan,  
Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
(66) 08 3219 9440  
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office  
196 Phahonyothin Road, Ladyao, Chatuchak,  
Bangkok 10900, Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
(66) 08 1889 6827

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9036  
Fax. (66) 0 2577 9009

Office/Laboratory  
668 Mu 2 Tambon Bangpoomai, Amphoe Muang Samutprakan,  
Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
(66) 08 3219 9440  
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office  
196 Phahonyothin Road, Ladyao, Chatuchak,  
Bangkok 10900, Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
(66) 08 1889 6827



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-67/0380

MTC No. EEL. BP. 14/0467

### CALIBRATION CERTIFICATE

Submitted by : S.P.S.Consulting Service Co., Ltd.

Address : 7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok, 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

**Instrument Calibrated :**

Description : Noise Dosimeter

Manufacturer : Syvantek

Model : SV-104IS

Serial No. : 106122

**Standards used :**

Multifunction Acoustic Calibrator Brüel&Kjær 4226 S/N 2810358 with Coupler UA0915 S/N 2810358.

**Calibration Procedure :**

This instrument was calibrated by using calibration procedure no CP-102-01, which was based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). This calibration procedure was related to the acoustical signal test of frequency weightings using a multifunction acoustic calibrator.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%.

Date of Receipt : 1 Apr. 2024

Date of Calibration : 10 Apr. 2024

1/2  
G. Samy

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.5



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-67/0380

MTC No. EEL. BP. 14/0467

### Acoustic signal test of frequency weightings

Frequency (Hz)	Deviation from response curve		Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)		
125	0.1	0.2	0.25	2.0
1 000	0.2	0.4	0.25	1.4
4 000	-0.1	0.3	0.25	3.6

Note : 1) There was no adjustment.

2) The calibration was performed at a sound pressure level of 114 dB.

3) The measured values did not include the correction of microphone of UUT.

4) The deviation was produced from the absolute difference between the measured values and the responding sound pressure levels in IEC 61672-1 (2002).

Calibrated by :

G. Samy  
(Mr. Sanaey Grajang)

Approved by :

Prawate Klauyapa  
(Mr. Prawate Klauyapa)  
Director

Electrical and Electronic Standards Laboratory  
Industrial Metrology and Testing Service Centre

Ref : 2011267040101274007

Date of Calibration : 10 Apr. 2024

Date of Issue : 11 Apr. 2024

2 / 2

End of Certificate

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.5

**Head Office**

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9036  
Fax. (66) 0 2577 9009

**Office/Laboratory**

668 Mu 2 Tambon Bangpoomai, Amphoe Muang Samutprakan,  
Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
(66) 08 3219 9440  
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

**Office**

196 Phahonyothin Road, Ladyao, Chatuchak,  
Bangkok 10900, Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
(66) 08 1889 6827

**Head Office**

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9036  
Fax. (66) 0 2577 9009

**Office/Laboratory**

668 Mu 2 Tambon Bangpoomai, Amphoe Muang Samutprakan,  
Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
(66) 08 3219 9440  
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

**Office**

196 Phahonyothin Road, Ladyao, Chatuchak,  
Bangkok 10900, Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
(66) 08 1889 6827



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-67/0380

MTC No. EEL. BP. 15/0467

## CALIBRATION CERTIFICATE

Submitted by : S.P.S.Consulting Service Co., Ltd.

Address : 7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok, 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

### Instrument Calibrated :

Description : Noise Dosimeter

Manufacturer : SvanTek

Model : SV-104IS

Serial No. : 106123

### Ambient Environment

Temperature :  $(23 \pm 3) ^\circ\text{C}$

Relative Humidity :  $(50 \pm 15) \%$

Ambient Pressure :  $(101.325 \pm 1.5) \text{ kPa}$

### Standards used :

Multifunction Acoustic Calibrator Brüel&Kjær 4226 S/N 2810358 with Coupler UA0915 S/N 2810358.

### Calibration Procedure :

This instrument was calibrated by using calibration procedure no CP-102-01, which was based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). This calibration procedure was related to the acoustical signal test of frequency weightings using a multifunction acoustic calibrator.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%.

Date of Receipt : 1 Apr. 2024

Date of Calibration : 10 Apr. 2024

1/2  
G. Samy

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-67/0380

MTC No. EEL. BP. 15/0467

### Acoustic signal test of frequency weightings

Frequency (Hz)	Deviation from response curve		Uncertainty ( $\pm$ dB)	Tolerance Limits Class 2 ( $\pm$ dB)
	A-weighting (dB)	C-weighting (dB)		
125	-0.4	-0.3	0.25	2.0
1 000	-0.2	-0.1	0.25	1.4
4 000	-0.5	-0.5	0.25	3.6

Note : 1) There was no adjustment.

2) The calibration was performed at a sound pressure level of 114 dB.

3) The measured values did not include the correction of microphone of UUT.

4) The deviation was produced from the absolute difference between the measured values and the responding sound pressure levels in IEC 61672-1 (2002).

Calibrated by :

*G. Samy*  
.....  
(Mr. Sanaey Grajang)

Approved by :

*Prawate Kluaypa*  
.....  
(Mr. Prawate Kluaypa)

**Director**  
**Electrical and Electronic Standards Laboratory**  
**Industrial Metrology and Testing Service Centre**

Ref : 2011267040101274008

Date of Calibration : 10 Apr. 2024

Date of Issue : 11 Apr. 2024

2 / 2

End of Certificate

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9036  
Fax. (66) 0 2577 9009

Office/Laboratory  
668 Mu 2 Tambon Bangpoomai, Amphoe Muang Samutprakan,  
Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
(66) 08 3219 9440  
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office  
196 Phahonyothin Road, Ladyao, Chatuchak,  
Bangkok 10900, Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
(66) 08 1889 6827

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9036  
Fax. (66) 0 2577 9009

Office/Laboratory  
668 Mu 2 Tambon Bangpoomai, Amphoe Muang Samutprakan,  
Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
(66) 08 3219 9440  
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office  
196 Phahonyothin Road, Ladyao, Chatuchak,  
Bangkok 10900, Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
(66) 08 1889 6827









THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-67/0380

MTC No. EEL. BP. 8/0467

### CALIBRATION CERTIFICATE

Submitted by : S.P.S.Consulting Service Co., Ltd.

Address : 7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok, 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

#### Instrument Calibrated :

Description : Noise Dosimeter

Manufacturer : Svantek

Model : SV-104IS

Serial No. : 60146

#### Standards used :

Multifunction Acoustic Calibrator Brüel&Kjær 4226 S/N 2810358 with Coupler UA0915 S/N 2810358.

#### Calibration Procedure :

This instrument was calibrated by using calibration procedure no CP-102-01, which was based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). This calibration procedure was related to the acoustical signal test of frequency weightings using a multifunction acoustic calibrator.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%.

Date of Receipt : 1 Apr. 2024

Date of Calibration : 10 Apr. 2024

1/2  
G. Samy

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-67/0380

MTC No. EEL. BP. 8/0467

### Acoustic signal test of frequency weightings

Frequency (Hz)	Deviation from response curve		Uncertainty (±dB)	Tolerance Limits Class 2 (±dB)
	A-weighting (dB)	C-weighting (dB)		
125	0.3	0.3	0.25	2.0
1 000	0.1	-0.1	0.25	1.4
4 000	-1.3	-1.2	0.25	3.6

Note : 1) There was no adjustment.

2) The calibration was performed at a sound pressure level of 114 dB.

3) The measured values did not include the correction of microphone of UUT.

4) The deviation was produced from the absolute difference between the measured values and the responding sound pressure levels in IEC 61672-1 (2002).

Calibrated by :

G. Samy

(Mr. Sanaey Grajang)

Approved by :

P. Kiatyapa

(Mr. Prawate Kiatyapa)

Director

Electrical and Electronic Standards Laboratory

Industrial Metrology and Testing Service Centre

Ref : 2011267040101274001

Date of Calibration : 10 Apr. 2024

Date of Issue : 11 Apr. 2024

2 / 2

End of Certificate

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9036  
Fax. (66) 0 2577 9009

Office/Laboratory  
668 Mu 2 Tambon Bangpoomai, Amphoe Muang Samutprakan,  
Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
(66) 08 3219 9440  
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office  
196 Phahonyothin Road, Ladyao, Chatuchak,  
Bangkok 10900, Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
(66) 08 1889 6827

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9036  
Fax. (66) 0 2577 9009

Office/Laboratory  
668 Mu 2 Tambon Bangpoomai, Amphoe Muang Samutprakan,  
Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
(66) 08 3219 9440  
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office  
196 Phahonyothin Road, Ladyao, Chatuchak,  
Bangkok 10900, Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
(66) 08 1889 6827



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-67/0380

MTC No. EEL. BP. 12/0467

### CALIBRATION CERTIFICATE

Submitted by : S.P.S.Consulting Service Co., Ltd.

Address : 7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok, 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

**Instrument Calibrated :**

Description : Noise Dosimeter

Manufacturer : Svantek

Model : SV-104IS

Serial No. : 63438

**Ambient Environment**

Temperature :  $(23 \pm 3) ^\circ\text{C}$

Relative Humidity :  $(50 \pm 15) \%$

Ambient Pressure :  $(101.325 \pm 1.5) \text{ kPa}$

**Standards used :**

Multifunction Acoustic Calibrator Brüel&Kjær 4226 S/N 2810358 with Coupler UA0915 S/N 2810358.

**Calibration Procedure :**

This instrument was calibrated by using calibration procedure no CP-102-01, which was based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). This calibration procedure was related to the acoustical signal test of frequency weightings using a multifunction acoustic calibrator.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%.

Date of Receipt : 1 Apr. 2024

Date of Calibration : 10 Apr. 2024

1/2  
G. Sany

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-67/0380

MTC No. EEL. BP. 12/0467

### Acoustic signal test of frequency weightings

Frequency (Hz)	Deviation from response curve		Uncertainty ( $\pm$ dB)	Tolerance Limits Class 2 ( $\pm$ dB)
	A-weighting (dB)	C-weighting (dB)		
125	0.1	0.2	0.25	2.0
1 000	0.0	-0.1	0.25	1.4
4 000	-0.4	0.4	0.25	3.6

Note : 1) There was no adjustment.

2) The calibration was performed at a sound pressure level of 114 dB.

3) The measured values did not include the correction of microphone of UUT.

4) The deviation was produced from the absolute difference between the measured values and the responding sound pressure levels in IEC 61672-1 (2002).

Calibrated by :

G. Sany  
(Mr. Sanaey Grajang)

Approved by :

P. Kluaypa  
(Mr. Prawate Kluaypa)  
Director

Electrical and Electronic Standards Laboratory  
Industrial Metrology and Testing Service Centre

Ref : 2011267040101274005

Date of Calibration : 10 Apr. 2024

Date of Issue : 11 Apr. 2024

2 / 2

End of Certificate

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9036  
Fax. (66) 0 2577 9009

Office/Laboratory  
668 Mu 2 Tambon Bangpoomai, Amphoe Muang Samutprakan,  
Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
(66) 08 3219 9440  
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office  
196 Phahonyothin Road, Ladyao, Chatuchak,  
Bangkok 10900, Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
(66) 08 1889 6827

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9036  
Fax. (66) 0 2577 9009

Office/Laboratory  
668 Mu 2 Tambon Bangpoomai, Amphoe Muang Samutprakan,  
Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
(66) 08 3219 9440  
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office  
196 Phahonyothin Road, Ladyao, Chatuchak,  
Bangkok 10900, Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
(66) 08 1889 6827



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-66/0475

MTC No. EEL. BP. 41/0566

### CALIBRATION CERTIFICATE

Submitted by : S.P.S.Consulting Service Co., Ltd.

Address : 7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok, 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

#### Instrument Calibrated :

Description : Noise Dosimeter

Manufacturer : Svanek

Model : SV-104IS

Serial No. : 70035

#### Ambient Environment

Temperature :  $(23 \pm 3) ^\circ\text{C}$

Relative Humidity :  $(50 \pm 15) \%$

Ambient Pressure :  $(101.325 \pm 1.5) \text{ kPa}$

#### Standards used :

Multifunction Acoustic Calibrator Brüel&Kjær 4226 S/N 2810358 with Coupler UA0915 S/N 2810358.

#### Calibration Procedure :

This instrument was calibrated by using calibration procedure no CP-102-01, which was based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). This calibration procedure was related to the acoustical signal test of frequency weightings using a multifunction acoustic calibrator.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%.

Date of Receipt : 11 May 2023

Date of Calibration : 23 May 2023

1 / 2

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.4

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9000  
Fax. (66) 0 2577 9009  
E-mail : rumpa@tistr.or.th Website:www.tistr.or.th

Office/Laboratory  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,  
Amphoe Muang, Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
Fax. (66) 0 2323 9165  
E-mail : mtc@tistr.or.th

Office  
196 Phaholyothin Road, Chatuchak, Bangkok 10900,  
Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
Fax. (66) 0 2579 8592  
E-mail : sumalee@tistr.or.th



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-66/0475

MTC No. EEL. BP. 41/0566

### Acoustic signal test of frequency weightings

Frequency (Hz)	Deviation from response curve		Uncertainty ( $\pm$ dB)	Tolerance Limits Class 2 ( $\pm$ dB)
	A-weighting (dB)	C-weighting (dB)		
125	0.3	0.2	0.25	2.0
1 000	0.1	-0.1	0.25	1.4
4 000	-1.3	-1.4	0.25	3.6

Note : 1) There was no adjustment.

2) The calibration was performed at a sound pressure level of 114 dB.

3) The measured values did not include the correction of microphone of UUT.

4) The deviation was produced from the absolute difference between the measured values and the responding sound pressure levels in IEC 61672-1 (2002).

Calibrated by :

*G. Sany*  
.....  
(Mr. Sanaey Grajang)

Approved by :

*Prawate Kluaypa*  
.....  
(Mr. Prawate Kluaypa)  
Director

Electrical and Electronic Standards Laboratory  
Industrial Metrology and Testing Service Centre

Ref : 2011266051101784003

Date of Calibration : 23 May 2023

Date of Issue : 24 May 2023

2 / 2

End of Certificate

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.4

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9000  
Fax. (66) 0 2577 9009  
E-mail : rumpa@tistr.or.th Website:www.tistr.or.th

Office/Laboratory  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,  
Amphoe Muang, Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
Fax. (66) 0 2323 9165  
E-mail : mtc@tistr.or.th

Office  
196 Phaholyothin Road, Chatuchak, Bangkok 10900,  
Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
Fax. (66) 0 2579 8592  
E-mail : sumalee@tistr.or.th



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-66/0452

MTC No. EEL. BP. 2/0566

## CALIBRATION CERTIFICATE

Submitted by : S.P.S.Consulting Service Co., Ltd.

Address : 7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok, 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

### Instrument Calibrated :

Description : Noise Dosimeter

Manufacturer : Svantek

Model : SV-104IS

Serial No. : 80801

### Standards used :

Multifunction Acoustic Calibrator Brüel&Kjær 4226 S/N 2810358 with Coupler UA0915 S/N 2810358.

### Calibration Procedure :

This instrument was calibrated by using calibration procedure no CP-102-01, which was based on IEC 61672-3 Electroacoustics - Sound Level Meters - Part 3 : Periodic tests (2006). This calibration procedure was related to the acoustical signal test of frequency weightings using a multifunction acoustic calibrator.

This instrument has been calibrated against standards maintained at the Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a level of confidence of approximately 95%.

Date of Receipt : 2 May 2023

Date of Calibration : 9 May 2023

1 / 2  
G. Sany

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9000  
Fax. (66) 0 2577 9009  
E-mail : rumpai@tistr.or.th Website:www.tistr.or.th

Office/Laboratory  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,  
Amphoe Muang, Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
Fax. (66) 0 2323 9165  
E-mail : mtc@tistr.or.th

Office  
196 Phahonyothin Road, Chatuchak, Bangkok 10900,  
Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
Fax. (66) 0 2579 8592  
E-mail : sumalee@tistr.or.th



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-66/0452

MTC No. EEL. BP. 2/0566

### Acoustic signal test of frequency weightings

Frequency (Hz)	Deviation from response curve		Uncertainty (+dB)	Tolerance Limits Class 2 (+dB)
	A-weighting (dB)	C-weighting (dB)		
125	-0.1	-0.1	0.25	2.0
1 000	0.0	-0.1	0.25	1.4
4 000	-0.1	-0.1	0.25	3.6

Note : 1) There was no adjustment.

2) The calibration was performed at a sound pressure level of 114 dB.

3) The measured values did not include the correction of microphone of UUT.

4) The deviation was produced from the absolute difference between the measured values and the responding sound pressure levels in IEC 61672-1 (2002).

Calibrated by :

*G. Sany*  
.....  
(Mr. Sanaey Grajang)

Approved by :

*Mr. P. Kluaypa*  
.....  
(Mr. Prawate Kluaypa)  
Director

Electrical and Electronic Standards Laboratory  
Industrial Metrology and Testing Service Centre

Ref : 2011266050201661001

Date of Calibration : 9 May 2023

Date of Issue : 9 May 2023

2 / 2

End of Certificate

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BLMTC.002 Rev.4

Head Office  
35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,  
Changwat Pathumthani 12120, Thailand  
Tel. (66) 0 2577 9000  
Fax. (66) 0 2577 9009  
E-mail : rumpai@tistr.or.th Website:www.tistr.or.th

Office/Laboratory  
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Road,  
Amphoe Muang, Changwat Samutprakan 10280, Thailand  
Tel. (66) 0 2323 1672-80 ext. 115, 116  
Fax. (66) 0 2323 9165  
E-mail : mtc@tistr.or.th

Office  
196 Phahonyothin Road, Chatuchak, Bangkok 10900,  
Thailand  
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217  
Fax. (66) 0 2579 8592  
E-mail : sumalee@tistr.or.th









บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด  
S.P.S. CONSULTING SERVICE CO., LTD.  
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900  
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900  
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Heat 116\_2

Heat Stress WBGT Meter Verification Report			
Verification Data			
Heat Stress WBGT Meter No.	: B22	Verification Date	: 22 April 2024
Brand	: Quest Technologies	Ambient Temp.	: 24.5 °C
Model	: QUESTemp <sup>o</sup> 32	Barometric Pressure	: 1011 mmbar
Serial No.	: TPK040059	Relative Humidity	: 49 %
Verification Module (Electronic Sensor Check) :			
Verification Module No. : 21 WB = 12.5 °C, DB = 47.1 °C, G = 69.3 °C			
Result of Verification : Without Adjustment			
Wet Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
12.5	12.6	-0.1	± 0.5
Dry Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
47.1	47.3	-0.2	± 0.5
Globe Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
69.3	69.1	0.2	± 0.5
UUC* = UNIT UNDER CALIBRATION			

Verified by : Adul Dangklom  
(Mr.Adul Dangklom)

Approved by : Peera Detudom  
(Mr. Peera Detudom)



69/29 Moo 1 Klongsi Klongluang Pathumthani 12120 (Thailand) Tel: (662) 193-2220 5 คู่มือสาย www.สอบเทียบอุปกรณ์วัด.com

METROLOGY SYSTEM ( THAILAND ) CO.,LTD.



## Certificate of Calibration

Certificate Number : SPR23110155-4

Page : 1 of 3

Customer : S.P.S. CONSULTING SERVICE CO., LTD.

7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,  
Bangkok 10900

Equipment Name : Area Heat Stress Monitor

Manufacturer : Quest Technologies

Model : QUESTemp 32

Serial Number : TPK040059

ID. Number : B22

Environmental Conditions

Ambient Temperature : 23 °C ± 2 °C

Received Date : 10 Nov 2023

Relative Humidity : 50 % ± 15 %

Calibration Date : 11 Nov 2023

Location of Calibration : In-Lab

Recommend Due Date : 11 Nov 2024

Calibration Procedure : SP-CPT-04-13

Date of Issue : 12 Nov 2023

### Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs.

The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Navaporn Uengseng

Calibration Officer

Approved by : Sombut Srikampa

( Mr. Sombut Srikampa )

Authorized Signatory





บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด  
S.P.S. CONSULTING SERVICE CO., LTD.  
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900  
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900  
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscor.com, www.spscor.com

Heat 116\_3

Heat Stress WBGT Meter Verification Report			
Verification Data			
Heat Stress WBGT Meter No. :	B25	Verification Date :	22 April 2024
Brand :	Quest Technologies	Ambient Temp. :	24.5 °C
Model :	QUESTemp <sup>o</sup> 32	Barometric Pressure :	1011 mmbar
Serial No. :	TPH050019	Relative Humidity :	49 %
Verification Module (Electronic Sensor Check) :			
Verification Module No. : 21 WB = 12.5 °C, DB = 47.1 °C, G = 69.3 °C			
Result of Verification : Without Adjustment			
Wet Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
12.5	12.7	-0.2	± 0.5
Dry Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
47.1	47.2	-0.1	± 0.5
Globe Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
69.3	69.2	0.1	± 0.5
UUC* = UNIT UNDER CALIBRATION			

Verified by :

Adul Dangklom  
(Mr. Adul Dangklom)

Approved by :

(Mr. Peera Detudom)



METROLOGY SYSTEM ( THAILAND ) CO., LTD.



## Certificate of Calibration

Certificate Number : SPR23110155-5

Page : 1 of 3

Customer : S.P.S. CONSULTING SERVICE CO., LTD.

7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,  
Bangkok 10900

Equipment Name : Area Heat Stress Monitor

Manufacturer : Quest Technologies

Model : QUESTemp 32

Serial Number : TPH050019

ID. Number : B25

Environmental Conditions

Ambient Temperature : 23 °C ± 2 °C Received Date : 10 Nov 2023

Relative Humidity : 50 % ± 15 % Calibration Date : 11 Nov 2023

Location of Calibration : In-Lab Recommend Due Date : 11 Nov 2024

Calibration Procedure : SP-CPT-04-13 Date of Issue : 12 Nov 2023

### Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs.

The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr. Navaporn Uengseng

Calibration Officer

Approved by :

(Mr. Sombut Srikampa)

Authorized Signatory



## Calibration Report

Certificate Number : SPR23110155-5

Page : 2 of 3

## Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Humidity Chamber	TH-80S	N/A	SPR23010480-5	22 Feb 2024
THERMO-HYGROMETER	5020A	A47046	QR23-0176	26 Jan 2024

## Traceability

This certification is traceable to the International System of Unit maintained at :  
SP Metrology - SP Metrology system (Thailand) Co.Ltd.  
Quality Reborn Co., Ltd



## Result of Calibration

Certificate No. : SPR23110155-5

Page : 3 of 3

Temperature Accuracy in the Measurement. (WET)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
30.0	30.011	30.2	0.189	0.20
35.0	35.014	35.2	0.186	0.20
40.0	40.017	40.2	0.183	0.20

Temperature Accuracy in the Measurement. (DRY)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
30.0	30.011	30.2	0.189	0.20
35.0	35.014	35.2	0.186	0.20
40.0	40.017	40.2	0.183	0.20

Temperature Accuracy in the Measurement. (GLOBE)

Unit : °C

Humidity Setting	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
30.0	30.011	30.5	0.489	0.20
35.0	35.014	35.5	0.486	0.20
40.0	40.017	40.5	0.483	0.20

## Note:

The result of calibration was found accurate as show on date and place of calibration only.  
This Certificate is not certified for any commercial transaction.

## Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2$ , providing a level of confidence approximately 95%.

- End of Certificate -



Heat Stress WBGT Meter Verification Report			
Verification Data			
Heat Stress WBGT Meter No.	: B28	Verification Date	: 22 April 2024
Brand	: Quest Technologies	Ambient Temp.	: 24.5 °C
Model	: QUESTemp <sup>®</sup> 32	Barometric Pressure	: 1011 mmbar
Serial No.	: TPH050046	Relative Humidity	: 49 %
Verification Module (Electronic Sensor Check) :			
Verification Module No. :	21	WB = 12.5 °C,	DB = 47.1 °C, G = 69.3 °C
Result of Verification : Without Adjustment			
Wet Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
12.5	12.7	-0.2	± 0.5
Dry Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
47.1	47.2	-0.1	± 0.5
Globe Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
69.3	69.5	-0.2	± 0.5
UUC* = UNIT UNDER CALIBRATION			

Approved by :   
(Mr. Peera Detudom)



69/29 Moo 1 Klongsri Klongluang Pathumthani 12120 ( Thailand ) Tel: (662) 193-2220 5 คู่สาย [www.สอบเทียบเครื่องมือวัด.com](http://www.สอบเทียบเครื่องมือวัด.com)

**ANAB**  
ANSI National Accreditation Board  
**ACCREDITED**  
ISO 9001  
CALIBRATION AND  
DIMENSIONAL MEASUREMENTS  
ACT – 2010



## Calibration Report

Certificate Number : SPR23110155-6

Page : 2 of 3

## Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Humidity Chamber	TH-80S	N/A	SPR23010480-5	22 Feb 2024
THERMO-HYGROMETER	5020A	A47046	QR23-0176	26 Jan 2024

## Traceability

This certification is traceable to the International System of Unit maintained at :

SP Metrology - SP Metrology system (Thailand) Co.Ltd.

Quality Reborn Co., Ltd



## Result of Calibration

Certificate No. : SPR23110155-6

Page : 3 of 3

Temperature Accuracy in the Measurement. (WET)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
30.0	30.011	30.3	0.289	0.20
35.0	35.014	35.3	0.286	0.20
40.0	40.017	40.3	0.283	0.20

Temperature Accuracy in the Measurement. (DRY)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
30.0	30.011	30.3	0.289	0.20
35.0	35.014	35.3	0.286	0.20
40.0	40.017	40.3	0.283	0.20

Temperature Accuracy in the Measurement. (GLOBE)

Unit : °C

Humidity Setting	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
30.0	30.011	30.5	0.489	0.20
35.0	35.014	35.5	0.486	0.20
40.0	40.017	40.5	0.483	0.20

## Note:

The result of calibration was found accurate as show on date and place of calibration only.  
This Certificate is not certified for any commercial transaction.

## Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2$ , providing a level of confidence approximately 95%.

- End of Certificate -



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด  
S.P.S. CONSULTING SERVICE CO., LTD.  
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900  
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900  
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com., www.spscon.com

Heat 116\_5

Heat Stress WBGT Meter Verification Report			
Verification Data			
Heat Stress WBGT Meter No.	: B30	Verification Date	: 22 April 2024
Brand	: Quest Technologies	Ambient Temp.	: 24.5 °C
Model	: QUESTemp <sup>o</sup> 32	Barometric Pressure	: 1011 mmbar
Serial No.	: TPH050057	Relative Humidity	: 49 %
Verification Module (Electronic Sensor Check) :			
Verification Module No. : 21 WB = 12.5 °C, DB = 47.1 °C, G = 69.3 °C			
Result of Verification : Without Adjustment			
Wet Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
12.5	12.4	0.1	± 0.5
Dry Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
47.1	47.3	-0.2	± 0.5
Globe Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
69.3	69.4	-0.1	± 0.5
UUC* = UNIT UNDER CALIBRATION			

Verified by : Adul Dangklom  
(Mr.Adul Dangklom)

Approved by : (Signature)  
(Mr. Peera Detudom)



METROLOGY SYSTEM ( THAILAND ) CO.,LTD.



## Certificate of Calibration

Certificate Number : SPR23110050-3 Page : 1 of 3  
Customer : S.P.S. CONSULTING SERVICE CO., LTD.  
7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,  
Bangkok 10900

Equipment Name : Area Heat Stress Monitor  
Manufacturer : Quest Technologies  
Model : QUESTemp 32  
Serial Number : TPH050057  
ID. Number : B30  
Environmental Conditions  
Ambient Temperature : 23 °C ± 2 °C Received Date : 03 Nov 2023  
Relative Humidity : 50 % ± 15 % Calibration Date : 03 Nov 2023  
Location of Calibration : In-Lab Recommend Due Date : 03 Nov 2024  
Calibration Procedure : SP-CPT-04-13 Date of Issue : 04 Nov 2023

### Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs. The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Pitak Srisutam  
Calibration Officer

Approved by : (Signature)  
( Mr.Prayoon Topart )  
Authorized Signatory



## Calibration Report

Certificate Number : SPR23110050-3

Page : 2 of 3

## Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Humidity Chamber	TH-80S	N/A	SPR23010480-5	22 Feb 2024
THERMO-HYGROMETER	5020A	A47046	QR23-0176	26 Jan 2024

## Traceability

This certification is traceable to the International System of Unit maintained at :

SP Metrology - SP Metrology system (Thailand) Co.Ltd.

Quality Reborn Co., Ltd



## Result of Calibration

Certificate No. : SPR23110050-3

Page : 3 of 3

## Temperature Accuracy in the Measurement. (WET)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
30.0	30.014	30.2	0.186	0.20
35.0	35.012	35.2	0.188	0.20
40.0	40.017	40.2	0.183	0.20

## Temperature Accuracy in the Measurement. (DRY)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
30.0	30.014	30.1	0.086	0.20
35.0	35.012	35.1	0.088	0.20
40.0	40.017	40.1	0.083	0.20

## Temperature Accuracy in the Measurement. (GLOBE)

Unit : °C

Humidity Setting	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
30.0	30.014	30.2	0.186	0.20
35.0	35.012	35.2	0.188	0.20
40.0	40.017	40.2	0.183	0.20

## Note :

The result of calibration was found accurate as show on date and place of calibration only.  
This Certificate is not certified for any commercial transaction.

## Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2$ , providing a level of confidence approximately 95%.

- End of Certificate -



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด  
S.P.S. CONSULTING SERVICE CO., LTD.  
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900  
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900  
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com., www.spscon.com


Heat 116\_6

Heat Stress WBGT Meter Verification Report			
Verification Data			
Heat Stress WBGT Meter No. :	B31	Verification Date :	22 April 2024
Brand :	Quest Technologies	Ambient Temp. :	24.5 °C
Model :	QUESTemp <sup>o</sup> 32	Barometric Pressure :	1011 mmbar
Serial No. :	TPH050047	Relative Humidity :	49 %
Verification Module (Electronic Sensor Check) :			
Verification Module No. : 21 WB = 12.5 °C, DB = 47.1 °C, G = 69.3 °C			
Result of Verification : Without Adjustment			
Wet Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
12.5	12.6	-0.1	± 0.5
Dry Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
47.1	47.3	-0.2	± 0.5
Globe Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
69.3	69.2	0.1	± 0.5
UUC* = UNIT UNDER CALIBRATION			

Verified by :

Adul Dangklom  
(Mr.Adul Dangklom)

Approved by :

  
(Mr. Peera Detudom)



69/29 Moo 1, Klongsi Klonguang Pathumthani 12120 ( Thailand ) Tel: (662) 193-2220 5 คู่มือราย WWW.สออบีซบเคื่องมือวัด.com

METROLOGY SYSTEM ( THAILAND ) CO.,LTD.



## Certificate of Calibration

Certificate Number : SPR23110050-4

Page : 1 of 3

Customer : S.P.S. CONSULTING SERVICE CO., LTD.

7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,  
Bangkok 10900

Equipment Name : Area Heat Stress Monitor

Manufacturer : Quest Technologies

Model : QUESTemp 32

Serial Number : TPH050047

ID. Number : B31

Environmental Conditions

Ambient Temperature : 23 °C ± 2 °C

Received Date : 03 Nov 2023

Relative Humidity : 50 % ± 15 %

Calibration Date : 03 Nov 2023

Location of Calibration : In-Lab

Recommend Due Date : 03 Nov 2024

Calibration Procedure : SP-CPT-04-13

Date of Issue : 04 Nov 2023

Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs.

The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Pitak Srisutam

Calibration Officer

Approved by :

( Mr.Prayoon Topart )

Authorized Signatory



## Calibration Report

Certificate Number : SPR23110050-4

Page : 2 of 3

### Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Humidity Chamber	TH-80S	N/A	SPR23010480-5	22 Feb 2024
THERMO-HYGROMETER	5020A	A47046	QR23-0176	26 Jan 2024

### Traceability

This certification is traceable to the International System of Unit maintained at :

SP Metrology - SP Metrology system (Thailand) Co.Ltd.

Quality Reborn Co., Ltd



## Result of Calibration

Certificate No. : SPR23110050-4

Page : 3 of 3

### Temperature Accuracy in the Measurement. (WET)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
30.0	30.014	30.1	0.086	0.20
35.0	35.012	35.1	0.088	0.20
40.0	40.017	40.1	0.083	0.20

### Temperature Accuracy in the Measurement. (DRY)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
30.0	30.014	30.1	0.086	0.20
35.0	35.012	35.1	0.088	0.20
40.0	40.017	40.1	0.083	0.20

### Temperature Accuracy in the Measurement. (GLOBE)

Unit : °C

Humidity Setting	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
30.0	30.014	30.2	0.186	0.20
35.0	35.012	35.2	0.188	0.20
40.0	40.017	40.2	0.183	0.20

### Note :

The result of calibration was found accurate as show on date and place of calibration only.  
This Certificate is not certified for any commercial transaction.

### Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2$ , providing a level of confidence approximately 95%.

- End of Certificate -



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด  
S.P.S. CONSULTING SERVICE CO., LTD.  
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900  
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900  
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Heat 116\_7

Heat Stress WBGT Meter Verification Report			
Verification Data			
Heat Stress WBGT Meter No. :	B32	Verification Date :	22 April 2024
Brand :	Quest Technologies	Ambient Temp. :	24.5 °C
Model :	QUESTemp <sup>32</sup>	Barometric Pressure :	1011 mmbar
Serial No. :	TPH050015	Relative Humidity :	49 %
Verification Module (Electronic Sensor Check) :			
Verification Module No. : 21    WB = 12.5 °C,    DB = 47.1 °C,    G = 69.3 °C			
Result of Verification : Without Adjustment			
Wet Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
12.5	12.6	-0.1	± 0.5
Dry Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
47.1	47.3	-0.2	± 0.5
Globe Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
69.3	69.1	0.2	± 0.5
UUC* = UNIT UNDER CALIBRATION			

Verified by : Adul Dangklorn  
(Mr.Adul Dangklorn)

Approved by : Peera Detudom  
(Mr. Peera Detudom)



69/29 Moo 1 Klongsi Klonguang Pathumthani 12120 (Thailand) Tel: (662) 193-2220 5 ตู้สาย www.สอบเทียบเครื่องวัด.com

**METROLOGY SYSTEM ( THAILAND ) CO.,LTD.**



## Certificate of Calibration

Certificate Number : SPR24030285-2 Page : 1 of 3  
Customer : S.P.S. CONSULTING SERVICE CO., LTD.  
7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,  
Bangkok 10900

Equipment Name : Area Heat Stress Monitor  
Manufacturer : Quest Technologies  
Model : QUESTemp 32  
Serial Number : TPH050015  
ID. Number : B32  
Environmental Conditions  
Ambient Temperature : 23 °C ± 2 °C Received Date : 19 Mar 2024  
Relative Humidity : 50 % ± 15 % Calibration Date : 20 Mar 2024  
Location of Calibration : In-Lab Recommend Due Date : 20 Mar 2025  
Calibration Procedure : SP-CPT-04-13 Date of Issue : 21 Mar 2024

### Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs. The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Navaporn Uengseng  
Calibration Officer

Approved by : Bussakorn Chaikaew  
( Ms.Bussakorn Chaikaew )  
Authorized Signatory



## Calibration Report

Certificate Number : SPR24030285-2

Page : 2 of 3

### Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Humidity Chamber	TH-80S	N/A	SPR24020149-7	23 Feb 2025
THERMO-HYGROMETER	5020A	A47046	QR24-0167	26 Jan 2025

### Traceability

This certification is traceable to the International System of Unit maintained at :

SP Metrology - SP Metrology system (Thailand) Co.Ltd.

Quality Reborn Co., Ltd



## Result of Calibration

Certificate No. : SPR24030285-2

Page : 3 of 3

Temperature Accuracy in the Measurement. (WET)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
30.0	30.012	30.0	-0.012	0.20
35.0	35.010	35.0	-0.010	0.20
40.0	40.015	40.1	0.085	0.20

Temperature Accuracy in the Measurement. (DRY)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
30.0	30.012	30.0	-0.012	0.20
35.0	35.010	35.0	-0.010	0.20
40.0	40.015	40.1	0.085	0.20

Temperature Accuracy in the Measurement. (GLOBE)

Unit : °C

Humidity Setting	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
30.0	30.012	30.0	-0.012	0.20
35.0	35.010	35.0	-0.010	0.20
40.0	40.015	40.1	0.085	0.20

### Note :

The result of calibration was found accurate as show on date and place of calibration only.  
This Certificate is not certified for any commercial transaction.

### Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2$ , providing a level of confidence approximately 95%.

- End of Certificate -



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด  
S.P.S. CONSULTING SERVICE CO., LTD.  
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900  
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900  
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscn.com, www.spscn.com

Heat 116\_8

Heat Stress WBGT Meter Verification Report			
Verification Data			
Heat Stress WBGT Meter No.	: B33	Verification Date	: 22 April 2024
Brand	: Quest Technologies	Ambient Temp.	: 24.5 °C
Model	: QUESTemp <sup>32</sup>	Barometric Pressure	: 1011 mmbar
Serial No.	: TPK120034	Relative Humidity	: 49 %
Verification Module (Electronic Sensor Check) :			
Verification Module No. : 21 WB = 12.5 °C, DB = 47.1 °C, G = 69.3 °C			
Result of Verification : Without Adjustment			
Wet Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
12.5	12.7	-0.2	± 0.5
Dry Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
47.1	47.2	-0.1	± 0.5
Globe Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
69.3	69.2	0.1	± 0.5
UUC* = UNIT UNDER CALIBRATION			

Verified by : \_\_\_\_\_  
(Mr.Adul Dangklorn)

Approved by : \_\_\_\_\_  
(Mr. Peera Detudom)



METROLOGY SYSTEM ( THAILAND ) CO.,LTD.



ANAB  
ASQ Business Accreditation Board  
ACCREDITED  
CALIBRATION AND  
DIMENSIONAL MEASUREMENT  
ACT - 2010

## Certificate of Calibration

Certificate Number : SPR23110050-5 Page : 1 of 3  
Customer : S.P.S. CONSULTING SERVICE CO., LTD.  
7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,  
Bangkok 10900

Equipment Name : Area Heat Stress Monitor  
Manufacturer : Quest Technologies  
Model : QUESTemp 32  
Serial Number : TPK120034  
ID. Number : B33  
Environmental Conditions  
Ambient Temperature : 23 °C ± 2 °C Received Date : 03 Nov 2023  
Relative Humidity : 50 % ± 15 % Calibration Date : 03 Nov 2023  
Location of Calibration : In-Lab Recommend Due Date : 03 Nov 2024  
Calibration Procedure : SP-CPT-04-13 Date of Issue : 04 Nov 2023

### Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs. The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Pitak Srisutam  
Calibration Officer

Approved by : \_\_\_\_\_  
( Mr.Prayoon Topart )  
Authorized Signatory



## Calibration Report

Certificate Number : SPR23110050-5

Page : 2 of 3

### Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Humidity Chamber	TH-80S	N/A	SPR23010480-5	22 Feb 2024
THERMO-HYGROMETER	5020A	A47046	QR23-0176	26 Jan 2024

### Traceability

This certification is traceable to the International System of Unit maintained at :

SP Metrology - SP Metrology system (Thailand) Co.Ltd.

Quality Reborn Co., Ltd



## Result of Calibration

Certificate No. : SPR23110050-5

Page : 3 of 3

### Temperature Accuracy in the Measurement. (WET)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
30.0	30.014	30.2	0.186	0.20
35.0	35.012	35.2	0.188	0.20
40.0	40.017	40.2	0.183	0.20

### Temperature Accuracy in the Measurement. (DRY)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
30.0	30.014	30.0	-0.014	0.20
35.0	35.012	35.0	-0.012	0.20
40.0	40.017	40.0	-0.017	0.20

### Temperature Accuracy in the Measurement. (GLOBE)

Unit : °C

Humidity Setting	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
30.0	30.014	30.1	0.086	0.20
35.0	35.012	35.1	0.088	0.20
40.0	40.017	40.1	0.083	0.20

### Note :

The result of calibration was found accurate as show on date and place of calibration only.  
This Certificate is not certified for any commercial transaction.

### Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2$ , providing a level of confidence approximately 95%.

- End of Certificate -



# CALIBRATION LABORATORY Co.,LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



## CERTIFICATE OF CALIBRATION

### FOR

NOMENCLATURE : LUX METER  
MANUFACTURER : EXTECH INSTRUMENTS  
MODEL / TYPE : 407026  
SERIAL NO. : A.052156 [B08]  
CLID. NO. : 252201678  
JOB CONTROL NO. : 230725081566

CUSTOMER : S.P.S. CONSULTING SERVICE CO., LTD.  
7 SOI PHAHOLYOTHIN 24 ROAD., JOMPOL,  
CHATUCHAK, BANGKOK 10900

DATE OF RECEIVED : 25 July 2023

DATE OF ISSUED : 31 July 2023

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Suwit Phuanbusabong  
Calibration Engineer

Approved By : Mongkol Yotsoontorn  
Authorized Signatory  
31 July 2023



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q23081566

F3-011-04/01-12

page 1 of 3



@clccalibration



# CALIBRATION LABORATORY Co.,LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



## REPORT OF CALIBRATION

### FOR

NOMENCLATURE : LUX METER  
MANUFACTURER : EXTECH INSTRUMENTS  
MODEL / TYPE : 407026  
SERIAL NO. : A.052156 [B08]  
DATE OF CALIBRATION : 26 July 2023  
DUE DATE OF CALIBRATION : 26 July 2024

### ENVIRONMENT CONDITIONS :

Temperature :  $(23 \pm 2) ^\circ\text{C}$

Relative Humidity :  $(55 \pm 15) \% \text{RH}$

### PROCEDURE USED :

This instrument was calibrated under procedure No. CLC-CPEE-18 by comparison with Photometer/Radiometer & Illuminance Sensor which maintained by the Calibration Laboratory Co., Ltd.

### REFERENCE STANDARD USED :

Photometer/Radiometer & Illuminance Sensor, Bentham Model ORM400/DH400VL S/N. 27710,27585/3.

### TRACEABILITY :

The measurements are traceable to International System of Units (SI), through Optical Test and Calibration Ltd. Certificate No. 144408/ABU, Due Date 03 April 2024.

### UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k = 2.00$  which for a normal distribution corresponds to a coverage probability of approximately 95 %. It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2022)"

Certificate No. Q23081566

F3-011-04/01-12

page 2 of 3



@clccalibration



# CALIBRATION LABORATORY CO.,LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



# METROLOGY SYSTEM ( THAILAND ) CO.,LTD.



## Certificate of Calibration

CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : ( X ) without adjustment ( ) adjustment

### CALIBRATION DATA

#### LUX METER

STD Applied ( lux )	DUC Reading ( lux )	Correction ( lux )	Uncertainty $\pm$ ( % of rdg. )
100	100	0	2.6
200	198	+2	2.6
300	295	+5	2.6
1000	984	+16	2.6
2000	1932	+68	2.6
3000	2890	+110	3.6

Note. The Scope of Accredited ANAB Certificate No. ACDM-2814 Version 008 Page 46 of 54

This report is valid for the above stated instrument/s only.

### End of Certificate ###

Certificate No. Q23081566

F3-011-04/01-12

page 3 of 3



@clccalibration

Certificate Number : SPR24010295-3

Page : 1 of 3

Customer : S.P.S. CONSULTING SERVICE CO., LTD.

7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,  
Bangkok 10900

Equipment Name : Light Meter

Manufacturer : Exttech

Model : 407026

Serial Number : A.055543

ID. Number : LUX-B10

#### Environmental Conditions

Ambient Temperature :  $23^{\circ}\text{C} \pm 3^{\circ}\text{C}$

Received Date : 19 Jan 2024

Relative Humidity :  $50\% \pm 15\%$

Calibration Date : 23 Jan 2024

Location of Calibration : In-Lab

Recommend Due Date : 23 Jan 2025

Calibration Procedure : SP-CPE-04-32

Date of Issue : 24 Jan 2024

#### Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs.

The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Karoon Pengsalung

Calibration Officer

Approved by :

( Mr.Yodyaim Chansang )

Authorized Signatory



ID LINE : IEC17025



## Calibration Report

Certificate Number : SPR24010295-3

Page : 2 of 3

### Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Digital Light Meter	LX-73	Q842777	23PH462	05 Sep 2024

### Traceability

This certification is traceable to the International System of Unit maintained at :  
TPA - Technology Promotion Association (Thailand-Japan)



ID LINE : IEC17025



## Result of Calibration

Certificate No. : SPR24010295-3

Page : 3 of 3

Function: Illumination Measurement

Unit : Lux

Calibration Point	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
100	100.0	98	-2	1.7
200	200	198	-2	6.6
300	300	300	0	7
1000	1000	1004	4	13
2000	2000	1978	-22	26
3000	3000	2980	-20	39

### Note:

The result of calibration was found accurate as show on date and place of calibration only.  
This Certificate is not certified for any commercial transaction.

### Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2.00$ , providing a level of confidence approximately 95 %  
- End of Certificate -



# Certificate of Calibration

Certificate Number : SPR24010295-4      Page : 1 of 3


Customer : S.P.S. CONSULTING SERVICE CO., LTD.  
7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,  
Bangkok 10900

Equipment Name : Light Meter  
Manufacturer : Extech  
Model : 407026  
Serial Number : A.055615  
ID. Number : LUX-B11

Environmental Conditions  
Ambient Temperature : 23 °C ± 3 °C      Received Date : 19 Jan 2024  
Relative Humidity : 50 % ± 15 %      Calibration Date : 23 Jan 2024  
Location of Calibration : In-Lab      Recommend Due Date : 23 Jan 2025  
Calibration Procedure : SP-CPE-04-32      Date of Issue : 24 Jan 2024

**Method of Calibration**  
This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received.Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs.  
The calibration certificate shall not be reproduced except in full,without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Karoon Pengsalung  
Calibration Officer

Approved by :   
( Mr.Yodyaim Chansang )  
Authorized Signatory

69/29 Moo 1 Klongsi Klongluang Pathumthani 12120 ( Thailand ) Tel: (662) 193-2220 5 ตู้สาย www.sevitechnology.co.th



# Calibration Report

Certificate Number : SPR24010295-4      Page : 2 of 3

## Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Digital Light Meter	LX-73	Q842777	23PH462	05 Sep 2024

**Traceability**  
This certification is traceable to the International System of Unit maintained at :  
TPA - Technology Promotion Association (Thailand-Japan)



ID LINE : IEC17025



## Result of Calibration

Certificate No. : SPR24010295-4

Page : 3 of 3

Function: Illumination Measurement

Unit : Lux

Calibration Point	Standard Reading	UUC Reading	Error	Uncertainty ( ± )
100	100.0	99	-1	1.7
200	200	199	-1	6.6
300	300	300	0	7
1000	1000	1009	9	13
2000	2000	1980	-20	26
3000	3000	2990	-10	39

### Note:

The result of calibration was found accurate as show on date and place of calibration only.  
This Certificate is not certified for any commercial transaction.

### Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2.00$ , providing a level of confidence approximately 95 %

- End of Certificate -



### Personal Pump Calibration Report

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

### Environmental Conditions

Temperature : 25 ± 3 °C  
Pressure : 1010 ± 15 mmbar

Personal Pump Data					Calibration Data							
No.	Brand	Model	Serial No.	Date	Flow Rate (ml/min)						Value From Calibration Curve	
					Setting			Actual (Q std.)				
					1	2	3	1	2	3	y	R <sup>2</sup>
B01	SKC	224-PCXR4	262101	02/10/2023	1,000	1,500	2,000	997	1,494	1,995	0.998x - 5.198	1.000
B02	SKC	224-PCXR4	626166	02/10/2023	1,000	1,500	2,000	995	1,491	1,987	0.995x - 0.239	1.000
B03	SKC	224-PCXR4	612968	02/10/2023	1,000	1,500	2,000	994	1,498	1,996	1.004x - 17.211	0.999
B04	SKC	224-PCXR4	602804	03/10/2023	1,000	1,500	2,000	1,001	1,502	1,997	0.999x - 3.961	1.000
B05	SKC	224-PCXR4	612693	03/10/2023	1,000	1,500	2,000	1,000	1,500	1,998	1.008x - 19.564	0.999
B06	SKC	224-PCXR4	262188	04/10/2023	1,000	1,500	2,000	999	1,497	1,998	1.005x - 13.275	1.000
B07	SKC	224-PCXR4	626262	04/10/2023	1,000	1,500	2,000	997	1,491	1,992	0.995x + 0.103	1.000
B08	SKC	224-PCXR4	626100	03/10/2023	1,000	1,500	2,000	995	1,490	1,994	0.999x - 3.162	1.000
B09	SKC	224-PCXR4	626479	04/10/2023	1,000	1,500	2,000	1,012	1,500	2,001	0.998x + 1.604	0.999
B10	SKC	224-PCXR4	091950	05/10/2023	1,000	1,500	2,000	992	1,486	1,994	1.002x - 11.842	1.000
B11	SKC	224-PCXR4	564315	05/10/2023	1,000	1,500	2,000	993	1,501	1,996	1.010x - 26.335	0.999
B12	SKC	224-PCXR4	034656	05/10/2023	1,000	1,500	2,000	1,000	1,496	1,998	1.007x - 17.721	0.999
B13	SKC	224-PCXR4	602073	04/10/2023	1,000	1,500	2,000	1,000	1,488	1,987	0.986x + 13.398	1.000
B14	SKC	224-PCXR4	626313	04/10/2023	1,000	1,500	2,000	996	1,493	1,996	0.999x - 2.380	1.000
B15	SKC	224-PCXR4	626474	06/10/2023	1,000	1,500	2,000	1,000	1,498	1,998	1.007x - 16.567	0.999
B16	SKC	224-PCXR4	626477	06/10/2023	1,000	1,500	2,000	1,001	1,498	1,999	1.010x - 21.673	0.999
B17	SKC	224-PCXR4	626860	06/10/2023	1,000	1,500	2,000	1,000	1,492	1,998	0.997x - 1.859	1.000
B18	SKC	224-PCXR4	691484	03/10/2023	1,000	1,500	2,000	995	1,494	1,992	1.000x - 5.493	1.000
B19	SKC	224-PCXR4	691599	02/10/2023	1,000	1,500	2,000	991	1,500	1,998	1.015x - 32.922	0.999
B20	SKC	224-PCXR4	691587	02/10/2023	1,000	1,500	2,000	1,001	1,496	1,999	1.010x - 23.222	0.999
B21	SKC	224-PCXR4	691531	03/10/2023	1,000	1,500	2,000	994	1,491	1,997	1.004x - 12.881	1.000
B22	SKC	224-PCXR4	691654	03/10/2023	1,000	1,500	2,000	991	1,492	1,994	1.002x - 9.860	1.000
B23	SKC	224-PCXR4	798393	02/10/2023	1,000	1,500	2,000	991	1,498	1,997	1.014x - 33.810	0.999
B24	SKC	224-PCXR4	626363	02/10/2023	1,000	1,500	2,000	1,001	1,499	2,001	1.011x - 23.676	0.999
B25	SKC	224-PCXR4	798489	04/10/2023	1,000	1,500	2,000	996	1,497	1,989	0.991x + 6.619	1.000
B26	SKC	224-PCXR4	798479	05/10/2023	1,000	1,500	2,000	996	1,492	1,990	0.996x - 1.146	1.000
B27	SKC	224-PCXR4	691673	09/10/2023	1,000	1,500	2,000	989	1,506	1,998	1.016x - 34.646	0.999
B28	SKC	224-PCXR4	691570	09/10/2023	1,000	1,500	2,000	992	1,487	1,996	1.006x - 16.996	1.000
B29	SKC	224-PCXR4	626472	09/10/2023	1,000	1,500	2,000	998	1,495	1,992	0.997x - 0.693	1.000
B30	SKC	224-PCXR4	691489	03/10/2023	1,000	1,500	2,000	993	1,490	1,990	0.999x - 7.320	1.000
B31	SKC	224-PCXR4	691509	03/10/2023	1,000	1,500	2,000	1,001	1,497	1,997	1.007x - 18.788	0.999
B32	SKC	224-PCXR4	091567	04/10/2023	1,000	1,500	2,000	998	1,499	1,996	1.009x - 22.780	0.999
B33	SKC	224-PCXR4	091756	05/10/2023	1,000	1,500	2,000	1,000	1,489	1,994	0.995x - 0.223	1.000
B34	SKC	224-PCXR4	612962	05/10/2023	1,000	1,500	2,000	992	1,501	1,997	1.013x - 31.362	0.999
B35	SKC	224-PCXR4	602682	03/10/2023	1,000	1,500	2,000	998	1,496	1,998	0.998x - 7.157	0.999
B36	SKC	224-PCXR4	626164	07/10/2023	1,000	1,500	2,000	995	1,487	1,990	0.991x + 3.901	1.000
B37	SKC	224-PCXR4	626256	02/10/2023	1,000	1,500	2,000	990	1,500	1,993	1.000x - 6.520	1.000
B38	SKC	224-PCXR4	626167	03/10/2023	1,000	1,500	2,000	989	1,498	1,995	1.015x - 35.470	0.999
B39	SKC	224-PCXR4	034637	09/10/2023	1,000	1,500	2,000	991	1,495	1,994	1.004x - 14.572	1.000
B40	SKC	224-PCXR4	798349	07/10/2023	1,000	1,500	2,000	999	1,497	1,996	1.008x - 21.525	0.999

Calibrated by :

Adul Dangklom  
(Mr. Adul Dangklom)

Approved by :

Peera Detudom  
(Mr. Peera Detudom)





บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด  
S.P.S. CONSULTING SERVICE CO., LTD.  
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจตุจักร เขตจตุจักร กรุงเทพฯ 10900  
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900  
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Personal Pump Calibration Report		
----------------------------------	--	--

Calibration Method : Dry Cal Primary Flowmeter	Model : Defender 510-H	S/N : 136164
--	------------------------	--------------

Environmental Conditions		
Temperature	25 ± 3 °C	
Pressure	1010 ± 15 mmbar	

Personal Pump Data				Calibration Data								
No.	Brand	Model	Serial No.	Date	Flow Rate (ml/min)						Value From Calibration Curve	
					Setting			Actual (Q std.)				
					1	2	3	1	2	3	y	R <sup>2</sup>
B41	SKC	224-PCXR4	612669	10/10/2023	1,000	1,500	2,000	999	1,491	1,993	0.994x + 2.802	1.000
B42	SKC	224-PCXR4	626041	10/10/2023	1,000	1,500	2,000	994	1,490	1,989	0.995x - 1.759	1.000
B43	SKC	224-PCXR4	034636	07/10/2023	1,000	1,500	2,000	995	1,488	1,989	0.991x + 2.866	1.000
B44	SKC	224-PCXR8	529341	07/10/2023	1,000	1,500	2,000	992	1,503	1,998	1.009x - 23.051	0.999
B45	SKC	224-PCXR8	529594	10/10/2023	1,000	1,500	2,000	1,000	1,495	1,989	0.989x + 10.094	1.000
B46	SKC	224-PCXR8	566743	02/10/2023	1,000	1,500	2,000	1,000	1,500	1,998	1.008x - 19.564	0.999
B47	SKC	224-PCXR8	566747	02/10/2023	1,000	1,500	2,000	994	1,502	1,996	1.011x - 27.787	0.999
B48	SKC	224-PCXR8	566753	03/10/2023	1,000	1,500	2,000	1,000	1,495	2,000	1.005x - 13.577	1.000
B49	SKC	224-PCXR8	566780	02/10/2023	1,000	1,500	2,000	998	1,498	2,000	1.010x - 21.853	0.999
B50	SKC	224-PCXR8	500400	07/10/2023	1,000	1,500	2,000	999	1,495	1,989	0.991x + 5.640	1.000
B51	SKC	224-PCXR8	500363	07/10/2023	1,000	1,500	2,000	993	1,501	1,996	1.009x - 24.941	0.999
B52	SKC	224-PCXR8	093186	07/10/2023	1,000	1,500	2,000	994	1,500	1,991	0.996x + 2.910	1.000
B53	SKC	224-PCXR8	707670	06/10/2023	1,000	1,500	2,000	990	1,498	1,996	1.014x - 33.838	0.999
B54	SKC	224-PCXR3	509821	05/10/2023	1,000	1,500	2,000	991	1,499	1,995	1.012x - 30.494	0.999
B55	SKC	224-PCXR3	510710	05/10/2023	1,000	1,500	2,000	996	1,493	1,996	0.999x - 2.301	1.000
B56	SKC	224-PCXR3	511450	05/10/2023	1,000	1,500	2,000	992	1,487	1,996	1.006x - 16.797	1.000
B57	SKC	224-PCXR3	510798	04/10/2023	1,000	1,500	2,000	989	1,493	1,994	1.001x - 9.175	1.000
B58	SKC	224-PCXR3	509852	04/10/2023	1,000	1,500	2,000	1,000	1,497	1,997	1.009x - 21.172	0.999
B59	SKC	224-PCXR3	509862	04/10/2023	1,000	1,500	2,000	995	1,495	1,988	0.993x + 2.723	1.000
B60	SKC	224-PCXR3	512655	07/10/2023	1,000	1,500	2,000	992	1,498	1,997	1.013x - 31.979	0.999
B61	SKC	224-PCXR3	503915	07/10/2023	1,000	1,500	2,000	1,000	1,502	1,997	1.007x - 20.065	0.999
B62	SKC	224-PCXR3	505975	07/10/2023	1,000	1,500	2,000	996	1,489	1,991	0.990x + 6.791	1.000
B63	SKC	224-PCXR3	511432	07/10/2023	1,000	1,500	2,000	993	1,500	1,995	1.003x - 8.208	1.000
B64	SKC	224-PCXR3	508302	05/10/2023	1,000	1,500	2,000	991	1,496	1,988	0.998x - 5.262	1.000
B65	SKC	224-PCXR3	508310	05/10/2023	1,000	1,500	2,000	993	1,492	1,991	0.999x - 4.884	1.000
B66	SKC	224-PCXR3	509861	06/10/2023	1,000	1,500	2,000	996	1,493	1,985	0.992x + 2.675	1.000
B67	SKC	224-PCXR3	506295	04/10/2023	1,000	1,500	2,000	1,000	1,498	1,998	1.009x - 21.534	0.999
B68	SKC	224-PCXR3	505872	04/10/2023	1,000	1,500	2,000	994	1,493	1,987	0.993x + 3.176	1.000
B69	SKC	224-PCXR3	508375	07/10/2023	1,000	1,500	2,000	999	1,495	1,996	1.005x - 19.592	0.999
B70	SKC	224-PCXR3	510623	04/10/2023	1,000	1,500	2,000	992	1,486	1,995	1.002x - 11.762	1.000
B71	SKC	224-PCXR3	508367	05/10/2023	1,000	1,500	2,000	999	1,497	1,996	1.008x - 21.646	0.999
B72	SKC	224-PCXR3	505977	03/10/2023	1,000	1,500	2,000	993	1,490	1,990	0.997x - 4.295	1.000
B73	SKC	224-PCXR3	512606	05/10/2023	1,000	1,500	2,000	995	1,495	1,989	0.994x + 1.210	1.000
B74	SKC	224-PCXR3	505993	05/10/2023	1,000	1,500	2,000	997	1,496	1,986	0.987x + 12.602	1.000
B75	SKC	224-PCXR3	509820	05/10/2023	1,000	1,500	2,000	994	1,490	1,991	0.998x - 5.143	1.000
B76	SKC	224-PCXR3	509811	06/10/2023	1,000	1,500	2,000	1,000	1,497	1,999	1.010x - 23.063	0.999
B77	SKC	224-PCXR3	508301	06/10/2023	1,000	1,500	2,000	992	1,501	1,996	1.013x - 32.023	0.999
B78	SKC	224-PCXR3	510677	05/10/2023	1,000	1,500	2,000	1,001	1,498	1,997	1.007x - 18.549	0.999
B79	SKC	224-PCXR3	510920	03/10/2023	1,000	1,500	2,000	999	1,509	1,997	0.996x + 4.999	1.000

Calibrated by :  (Mr. Adul Dangklom)	Approved by :  (Mr. Peera Detudom)
--	--





บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด  
S.P.S. CONSULTING SERVICE CO., LTD.  
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจตุจักร เขตจตุจักร กรุงเทพฯ 10900  
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900  
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Personal Pump Calibration Report		
----------------------------------	--	--

Calibration Method : Dry Cal Primary Flowmeter	Model : Defender 510-H	S/N : 136164
--	------------------------	--------------

Environmental Conditions		
Temperature	25 ± 3 °C	
Pressure	1010 ± 15 mmbar	

Personal Pump Data				Calibration Data									
No.	Brand	Model	Serial No.	Date	Flow Rate (ml/min)						Value From Calibration Curve		
					Setting			Actual (Q std.)					
					1	2	3	1	2	3	y	R <sup>2</sup>	
B80	SKC	224-PCXR3	504569	07/10/2023	1,000	1,500	2,000	1,001	1,497	2,000	1.007x - 19.202	0.999	
B81	SKC	224-PCXR3	503480	07/10/2023	1,000	1,500	2,000	994	1,496	1,996	1.006x - 17.526	1.000	
B82	SKC	224-PCXR3	505673	05/10/2023	1,000	1,500	2,000	992	1,495	1,992	1.002x - 11.742	1.000	
B83	SKC	224-PCXR3	510785	05/10/2023	1,000	1,500	2,000	994	1,497	1,998	1.005x - 15.177	1.000	
B84	SKC	224-PCXR3	508333	04/10/2023	1,000	1,500	2,000	1,000	1,498	1,999	1.003x - 16.041	0.999	
B85	SKC	224-PCXR3	505757	04/10/2023	1,000	1,500	2,000	999	1,498	1,999	1.010x - 23.715	0.999	
B86	SKC	224-PCXR3	512625	02/10/2023	1,000	1,500	2,000	1,000	1,493	1,989	0.994x + 1.568	1.000	
B87	SKC	224-PCXR3	504324	09/10/2023	1,000	1,500	2,000	1,001	1,500	1,996	1.006x - 16.049	0.999	
B88	SKC	224-PCXR3	508307	09/10/2023	1,000	1,500	2,000	1,000	1,496	1,990	0.990x + 9.617	1.000	
B89	SKC	224-PCXR3	509860	05/10/2023	1,000	1,500	2,000	999	1,496	1,998	0.997x - 5.214	0.999	
B90	SKC	224-PCXR3	508366	03/10/2023	1,000	1,500	2,000	992	1,495	1,992	0.999x - 5.095	1.000	
B91	SKC	224-PCXR3	510919	10/10/2023	1,000	1,500	2,000	998	1,496	1,993	0.995x - 1.847	1.000	
B92	SKC	224-PCXR3	510987	04/10/2023	1,000	1,500	2,000	1,001	1,496	1,997	1.005x - 22.028	0.999	
B93	SKC	224-PCXR3	509845	04/10/2023	1,000	1,500	2,000	997	1,491	1,990	0.995x + 2.516	1.000	
B94	SKC	224-PCXR8	A127871	04/10/2023	1,000	1,500	2,000	998	1,495	1,995	1.005x - 19.074	0.999	
B95	SKC	224-PCXR8	A127921	09/10/2023	1,000	1,500	2,000	998	1,498	1,999	1.012x - 25.793	0.999	
B96	SKC	224-PCXR8	A127942	09/10/2023	1,000	1,500	2,000	999	1,495	1,989	0.991x + 5.720	1.000	
B97	SKC	224-PCXR8	A127955	09/10/2023	1,000	1,500	2,000	994	1,496	1,996	1.011x - 28.512	0.999	
B98	SKC	224-PCXR8	A127956	10/10/2023	1,000	1,500	2,000	995	1,489	1,987	0.994x + 1.218	1.000	

Calibrated by :  (Mr. Adul Dangklom)	Approved by :  (Mr. Peera Detudom)
--	--



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด  
S.P.S. CONSULTING SERVICE CO., LTD.  
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจตุจักร เขตจตุจักร กรุงเทพฯ 10000  
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900  
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

### Personal Pump Calibration Report

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

#### Environmental Conditions

Temperature : 25 ± 3 °C  
Pressure : 1010 ± 15 mmbar

Personal Pump Data				Calibration Data									
No.	Brand	Model	Serial No.	Date	Flow Rate (ml/min)						Value From Calibration Curve		
					Setting			Actual (Q std.)					
					1	2	3	1	2	3	y	R <sup>2</sup>	
R01	SKC	224-PCXR4	602467	02/10/2023	1,000	1,500	2,000	1,001	1,499	1,999	1.010x - 22.581	0.999	
R02	SKC	224-PCXR4	626450	06/10/2023	1,000	2,000	3,000	996	1,493	1,986	0.992x + 3.742	1.000	
R03	SKC	224-PCXR4	691592	06/10/2023	1,000	1,500	2,000	989	1,495	1,994	0.999x - 6.866	1.000	
R04	SKC	224-PCXR4	691672	06/10/2023	1,000	1,500	2,000	998	1,491	1,989	0.991x + 5.421	1.000	
R05	SKC	224-PCXR4	798470	06/10/2023	1,000	1,500	2,000	999	1,495	1,995	1.005x - 18.995	0.999	
R06	SKC	224-PCXR4	798456	02/10/2023	1,000	1,500	2,000	1,000	1,488	1,987	0.986x + 13.398	1.000	
R07	SKC	224-PCXR4	798480	02/10/2023	1,000	1,500	2,000	1,000	1,497	1,998	1.009x - 21.689	0.999	
R08	SKC	224-PCXR4	883215	05/10/2023	1,000	1,500	2,000	994	1,500	1,990	0.995x + 3.109	1.000	
R09	SKC	224-PCXR4	034650	05/10/2023	1,000	1,500	2,000	999	1,497	1,996	1.008x - 21.526	0.999	
R10	SKC	224-PCXR4	091765	05/10/2023	1,000	1,500	2,000	996	1,493	1,994	1.000x - 6.596	1.000	
R11	SKC	224-PCXR4	091763	04/10/2023	1,000	1,500	2,000	998	1,496	1,983	0.998x - 9.346	0.999	
R12	SKC	224-PCXR4	091568	04/10/2023	1,000	1,500	2,000	1,000	1,497	1,999	1.009x - 21.948	0.999	
R13	SKC	224-PCXR4	091638	02/10/2023	1,000	1,500	2,000	994	1,495	1,986	0.993x + 2.981	1.000	
R14	SKC	224-PCXR4	091764	06/10/2023	1,000	1,500	2,000	998	1,498	2,000	1.012x - 26.788	0.999	
R15	SKC	224-PCXR8	529457	06/10/2023	1,000	1,500	2,000	995	1,492	1,987	0.994x + 1.457	1.000	
R16	SKC	224-PCXR8	529643	04/10/2023	1,000	1,500	2,000	1,000	1,498	1,997	1.007x - 17.908	0.999	
R17	SKC	224-PCXR8	529645	07/10/2023	1,000	1,500	2,000	998	1,496	1,998	1.011x - 25.546	0.999	
R18	SKC	224-PCXR8	566756	03/10/2023	1,000	1,500	2,000	994	1,490	1,989	0.995x - 1.759	1.000	
R19	SKC	224-PCXR8	566802	02/10/2023	1,000	1,500	2,000	1,000	1,496	1,999	1.010x - 22.864	0.999	
R20	SKC	224-PCXR8	529089	06/10/2023	1,000	1,500	2,000	992	1,506	1,996	1.008x - 22.151	0.999	
R21	SKC	224-PCXR8	665728	02/10/2023	1,000	1,500	2,000	992	1,486	1,994	1.002x - 11.842	1.000	
R22	SKC	224-PCXR8	707444	03/10/2023	1,000	1,500	2,000	1,001	1,500	1,999	1.007x - 18.171	0.999	
R23	SKC	224-PCXR8	761067	06/10/2023	1,000	1,500	2,000	1,000	1,488	1,993	0.992x + 5.744	1.000	
R24	SKC	224-PCXR8	707893	05/10/2023	1,000	1,500	2,000	994	1,505	1,996	1.005x - 15.010	0.999	
R25	SKC	224-PCXR8	761052	06/10/2023	1,000	1,500	2,000	999	1,495	1,989	0.991x + 5.640	1.000	
R26	SKC	224-PCXR8	707956	07/10/2023	1,000	1,500	2,000	1,010	1,497	2,002	0.999x - 2.874	0.999	
R27	SKC	224-PCXR8	707398	07/10/2023	1,000	1,500	2,000	1,001	1,496	1,997	1.008x - 20.237	0.999	
R28	SKC	224-PCXR8	707481	07/10/2023	1,000	1,500	2,000	993	1,506	1,995	1.002x - 10.719	1.000	
R29	SKC	224-PCXR8	707402	04/10/2023	1,000	1,500	2,000	995	1,495	1,989	0.995x + 1.091	1.000	
R30	SKC	224-PCXR8	093811	04/10/2023	1,000	1,500	2,000	998	1,495	1,992	0.997x - 0.693	1.000	
R31	SKC	224-PCXR8	093183	06/10/2023	1,000	1,500	2,000	999	1,502	1,997	0.988x + 9.127	0.999	
R32	SKC	224-PCXR8	671950	07/10/2023	1,000	1,500	2,000	998	1,495	1,994	0.998x - 3.451	1.000	
R33	SKC	224-PCXR4	626254	07/10/2023	1,000	1,500	2,000	992	1,503	1,995	1.011x - 30.016	0.999	
R34	SKC	224-PCXR4	626131	03/10/2023	1,000	1,500	2,000	990	1,499	1,997	1.014x - 32.986	0.999	
R35	SKC	224-PCXR8	707460	07/10/2023	1,000	1,500	2,000	990	1,501	1,997	1.005x - 15.898	1.000	
R36	SKC	224-PCXR8	707446	05/10/2023	1,000	1,500	2,000	1,000	1,497	1,997	1.002x - 7.547	1.000	
R37	SKC	224-PCXR8	707432	02/10/2023	1,000	1,500	2,000	995	1,498	1,995	0.999x - 4.856	1.000	
R38	SKC	224-PCXR8	707349	02/10/2023	1,000	1,500	2,000	991	1,496	1,992	1.000x - 7.364	1.000	
R39	SKC	224-PCXR8	761095	06/10/2023	1,000	1,500	2,000	995	1,489	1,985	0.990x + 6.253	1.000	

Calibrated by :

Adul Dangklom  
(Mr. Adul Dangklom)

Approved by :

Peera Detudom  
(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด  
S.P.S. CONSULTING SERVICE CO., LTD.  
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจตุจักร เขตจตุจักร กรุงเทพฯ 10000  
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900  
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

### Personal Pump Calibration Report

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

#### Environmental Conditions

Temperature : 25 ± 3 °C  
Pressure : 1010 ± 15 mmbar

Personal Pump Data				Calibration Data									
No.	Brand	Model	Serial No.	Date	Flow Rate (ml/min)						Value From Calibration Curve		
					Setting			Actual (Q std.)					
					1	2	3	1	2	3	y	R <sup>2</sup>	
R40	SKC	224-PCXR4	612753	07/10/2023	1,000	1,500	2,000	999	1,496	1,996	1.008x - 21.287	0.999	
R41	SKC	224-PCXR4	626140	03/10/2023	1,000	1,500	2,000	990	1,499	1,996	1.013x - 31.991	0.999	
R42	SKC	224-PCXR4	626463	07/10/2023	1,000	1,500	2,000	998	1,493	1,994	0.998x - 4.088	1.000	
R43	SKC	224-PCXR4	626129	07/10/2023	1,000	1,500	2,000	1,001	1,498	1,999	1.010x - 21.673	0.999	
R44	SKC	224-PCXR4	602753	07/10/2023	1,000	1,500	2,000	994	1,492	1,990	0.997x - 4.275	1.000	
R45	SKC	224-PCXR4	626137	06/10/2023	1,000	1,500	2,000	992	1,487	1,996	1.006x - 16.996	1.000	

Calibrated by :

Adul Dangklom  
(Mr. Adul Dangklom)

Approved by :

Peera Detudom  
(Mr. Peera Detudom)

Rotameter Calibration Report (For Personal Pump High Flow Adjust)

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164


Calibration Data

Rotameter Data			Calibration Data								
No.	Brand	Model	Date	Flow Rate (ml/min)						Value From Calibration Curve	
				Flow Rate (Reading)			Actual (Q std.)				
				1	2	3	1	2	3	y	R <sup>2</sup>
H-B01	Dwyer	VFB-65	02/10/2023	500	1,000	2,000	500.1	992.7	1979.6	0.993x + 4.560	1.000
H-B02	Dwyer	VFB-65	03/10/2023	500	1,000	2,000	503.6	989.1	1983.1	0.991x + 6.131	1.000
H-B03	Dwyer	VFB-65	02/10/2023	500	1,000	2,000	498.7	991.6	2006.8	0.999x - 9.214	0.999
H-B04	Dwyer	VFB-65	04/10/2023	500	1,000	2,000	502.3	988.1	2003.7	1.000x - 2.013	1.000
H-B05	Dwyer	VFB-65	03/10/2023	500	1,000	2,000	497.8	989.1	1971.6	0.981x + 16.401	0.999
H-B06	Dwyer	VFB-65	05/10/2023	500	1,000	2,000	499.3	995.9	1979.0	0.988x + 11.304	1.000
H-B07	Dwyer	VFB-65	04/10/2023	500	1,000	2,000	495.1	995.8	1991.3	0.997x - 1.222	1.000
H-B08	Dwyer	VFB-65	05/10/2023	500	1,000	2,000	500.7	998.7	1975.8	0.990x + 5.555	0.999
H-B09	Dwyer	VFB-65	03/10/2023	500	1,000	2,000	496.9	998.5	1979.0	0.989x + 10.735	1.000
H-B10	Dwyer	VFB-65	07/10/2023	500	1,000	2,000	499.4	997.5	2004.6	0.998x - 1.062	1.000

Calibrated by :

  
 (Mr. Peera Detudom)

Approved by :

  
 (Mr. Peera Detudom)

Rotameter Calibration Report (For Personal Pump Low Flow Adjust)

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164


Calibration Data

Rotameter Data				Calibration Data							
No.	Brand	Model	Date	Flow Rate (mL/min)						Value From Calibration Curve	
				Flow Rate (Reading)			Actual (Q std.)				
				1	2	3	1	2	3	y	R <sup>2</sup>
L-B01	Dwyer	VFA-21	02/10/2023	50	100	200	50.2	100.3	202.6	0.997x + 0.475	0.999
L-B02	Dwyer	VFA-21	03/10/2023	50	100	200	50.5	98.9	201.1	1.001x - 0.121	1.000
L-B03	Dwyer	VFA-21	02/10/2023	50	100	200	50.1	100.7	200.2	1.007x - 1.206	0.999
L-B04	Dwyer	VFA-21	04/10/2023	50	100	200	50.4	99.6	201.9	1.006x - 0.142	1.000
L-B05	Dwyer	VFA-21	03/10/2023	50	100	200	49.7	101.1	197.7	0.997x - 0.218	1.000
L-B06	Dwyer	VFA-21	05/10/2023	50	100	200	50.3	101.5	200.1	1.003x - 0.332	0.999
L-B07	Dwyer	VFA-21	04/10/2023	50	100	200	50.9	100.4	202.4	0.990x + 2.441	1.000
L-B08	Dwyer	VFA-21	05/10/2023	50	100	200	50.7	99.8	197.9	1.005x - 1.343	0.999
L-B09	Dwyer	VFA-21	03/10/2023	50	100	200	50.2	100.3	203.0	1.007x + 0.375	1.000
L-B10	Dwyer	VFA-21	07/10/2023	50	100	200	49.5	99.4	200.3	1.009x - 1.182	1.000

Calibrated by :

  
 (Mr. Peera Detudom)

Approved by :

  
 (Mr. Peera Detudom)



## Certificate of Calibration

**Aquion: Anion (ID#894)**

This certificate is to verify that instrument below are calibrated

by Archemica Lab Co.,Ltd.

AQUION S/N : 190840059

AS-DV S/N : 190915235

for

**S.P.S. Consulting Service Co., Ltd.**



Operator Signature: บริษัท อาร์เคมีกา แล็บ จำกัด Date: June 24, 2024  
ARCHEMICA LAB CO.,LTD.

(Mr. Ponwut Kornthongnimit)

Test Engineer