

ภาคผนวก จ

เอกสารสอบเทียบเครื่องมือ

List of Instruments Certification for Air & Noise Quality Analysis

No.	Instrument/Equipment	Parameter	Manufacturer	Model/Serial No.	Calibrator	Certification No.	Date of Calibration	Due date of Calibration	Remark
Stack									
1	Pre-Test Console	Total Suspended Particulate BTEX Styrene Hydrogen Chloride	Apex Instruments, USA.	XC-572-V 0803018	Envi Equipment Service Co., Ltd.	E23-12097	24 Dec 23	23 Dec 24	-
2	Pre-Test Console	Total Suspended Particulate BTEX Styrene Hydrogen Chloride	Apex Instruments, USA.	XC-572-V 1701018	Envi Equipment Service Co., Ltd.	E23-12095	12 Sep 23	11 Sep 24	-
3	Flue gas Analyzer	Sulphur Dioxide Oxide of Nitrogen as Nitrogen Dioxide Carbon Monoxide	Testo	Testo 350 2376344	Entech Industrial Sulation Co., Ltd.	G 660610	4 Oct 23	3 Oct 24	-
4	Flue gas Analyzer	Sulphur Dioxide Oxide of Nitrogen as Nitrogen Dioxide Carbon Monoxide	Testo	Testo 350 60723967	Entech Industrial Sulation Co., Ltd.	G 660472	10 Aug 23	9 Aug 24	-
5	Total Hydrocarbons Analyzer	Total Hydrocarbons	Thermo Scientific	55i 1182920025	UAE Consultant Co.,Ltd.	25012023	25 Dec 23	24 Dec 24	-
6	Standard Gas	Total Hydrocarbons	Linde	D824432	Linde	09042013	4 Aug 20	4 Aug 28	-

List of Opacity Training Certification for Opacity Mesurement

No.	Name	Training Couse	Train	Date	Remark
1	Mr.Rattanachai Laoma	Opacity	Pollution Control Department	28-29 March 2019	-
2	Mr.Kitipong Sonchayaphum	Opacity	Pollution Control Department	28-29 March 2019	-

List of Instruments Certification for Air & Noise Quality Analysis

No.	Instrument/Equipment	Parameter	Manufacturer	Model/Serial No.	Calibrator	Certification No.	Date of Calibration	Due date of Calibration	Remark
Ambient									
1	Orifice Transfer Standard Calibrator	Total Suspended Particulate (TSP) Particulate Matter < 10 µm (PM ₁₀)	Tisch Environmental,Inc.	TE-5025A 3383	Jiranatee Associates Co., Ltd.	CL-003-65	26 Jul 22	25 Jul 24	-
2	U-Tube Manometer	Total Suspended Particulate (TSP) Particulate Matter < 10 µm (PM ₁₀)	Dwyer	1221-36-W/M -	Technology Promotion Association (Thailand-Japan)	24P1250	10 Apr 24	9 Apr 25	-
3	Aneroid Barometer	Total Suspended Particulate (TSP) Particulate Matter < 10 µm (PM ₁₀)	Barijo, Germany	-	Technology Promotion Association (Thailand-Japan)	24P1367	22 Apr 24	21 Apr 25	-
4	Dial Thermo-Hygrometer	Total Suspended Particulate (TSP) Particulate Matter < 10 µm (PM ₁₀)	Barijo, Germany	-	Technology Promotion Association (Thailand-Japan)	24H756	10 Apr 24	9 Apr 25	-
5	Nitrogen Dioxide Analyzer	Nitrogen Dioxide	Thermo Scientific	42i CM08130002	UAE Consultant Co.,Ltd.	01112023	1 Nov 23	31 Oct 24	-
6	Nitrogen Dioxide Analyzer	Nitrogen Dioxide	Thermo Scientific	42i CM19050148	UAE Consultant Co.,Ltd.	13112023	13 Nov 23	12 Nov 24	-
7	Nitrogen Dioxide Analyzer	Nitrogen Dioxide	Thermo Scientific	42i CM19050149	UAE Consultant Co.,Ltd.	01112023	1 Nov 23	31 Oct 24	-
8	Standard Gases (Mixture)	Nitrogen Dioxide	Airgas	EB0143262 2015PSIG	Airgas an Air Liquide company	E04N99E15A01D3	21 Jun 21	21 Jun 24	-
9	Wind Speed/Wind Direction	WS/WD	Scarlet Tech Ltd.	WL-21 2205DT0114	Thai Meteorological Department	099/24	22 Feb 24	21 Feb 25	-
10	Wind Speed/Wind Direction	WS/WD	Scarlet Tech Ltd.	WL-21 2205DT0116	Thai Meteorological Department	100/24	22 Feb 24	21 Feb 25	-
11	Wind Speed/Wind Direction	WS/WD	Scarlet Tech Ltd.	WL-21 2301DR00024	Thai Meteorological Department	096/24	22 Feb 24	21 Feb 25	-
12	Sound Level Calibrator (Acoustic Calibrator)	Calibrate Sound Level Meter	Larson Davis	CAL150 6171	Innovative Instrument Co.,Ltd.	23-ACT-118	4 Aug 23	3 Aug 24	-
13	Sound Level Meter	L _{Aeq} 24 hours* L _{Aeq} 1 hour* L _{Amax} * L _{A90} * L _{Ach}	Rion, Japan	NL-62 00130355	Sithiporn Associates Co., Ltd.	ACT23199	27 Jun 23	26 Jun 24	-
14	Sound Level Meter	L _{Aeq} 24 hours* L _{Aeq} 1 hour* L _{Amax} * L _{A90} * L _{Ach}	Rion, Japan	NL-62 00130356	Innovative Instrument Co.,Ltd.	CP20230290EA	3 Aug 23	2 Aug 24	-

List of Instruments Certification for Air & Noise Quality Analysis

No.	Instrument/Equipment	Parameter	Manufacturer	Model/Serial No.	Calibrator	Certification No.	Date of Calibration	Due date of Calibration	Remark
Ambient									
15	Sound Level Meter	$L_{Aeq} 24\text{ hours}$, $L_{Aeq} 1\text{ hour}$, L_{Amax} , L_{A90} , L_{A10}	Rion, Japan	NL-62 00130357	Innovative Instrument Co.,Ltd.	CP20230291EA	3 Aug 23	2 Aug 24	-
16	Sound Level Meter	$L_{Aeq} 24\text{ hours}$, $L_{Aeq} 1\text{ hour}$, L_{Amax} , L_{A90} , L_{A10}	Cube, France	Cube 11070	Electrical And Electronics Institute Foundation For Industrial Development	CP20230293EA	4 Aug 23	3 Aug 24	-
17	Sound Level Meter	$L_{Aeq} 24\text{ hours}$, $L_{Aeq} 1\text{ hour}$, L_{Amax} , L_{A90} , L_{A10}	Cube, France	Cube 11073	Electrical And Electronics Institute Foundation For Industrial Development	CP20230294EA	4 Aug 23	3 Aug 24	-
18	Sound Level Meter	$L_{Aeq} 24\text{ hours}$, $L_{Aeq} 1\text{ hour}$, L_{Amax} , L_{A90} , L_{A10}	Cube, France	Cube 11113	Electrical And Electronics Institute Foundation For Industrial Development	CP20230295EA	4 Aug 23	3 Aug 24	-
19	Sound Level Meter	$L_{Aeq} 24\text{ hours}$, $L_{Aeq} 1\text{ hour}$, L_{Amax} , L_{A90} , L_{A10}	Cube, France	Cube 11135	Electrical And Electronics Institute Foundation For Industrial Development	CP20230296EA	5 Aug 23	4 Aug 24	-

List of Instruments Certification for Air & Noise Quality Analysis

No.	Instrument/Equipment	Parameter	Manufacturer	Model/Serial No.	Calibrator	Certification No.	Date of Calibration	Due date of Calibration	Remark
Workplace									
1	Sound Level Calibrator (Acoustic Calibrator)	Calibrate Sound Level Meter	Quest Technologies, Inc	QC-20 QOF110030	Innovative Instrument Co.,Ltd.	23-ACT-116	4 Aug 23	3 Aug 24	-
2	Sound Level Calibrator (Acoustic Calibrator)	Calibrate Sound Level Meter	Svantek	SV35A 73249	Innovative Instrument Co.,Ltd.	23-ACT-111	27 Jun 23	26 Jun 24	-
3	Noise Dosimeter	Noise Dosimeter	Svantek	SV 104 143229	Innovative Instrument Co.,Ltd.	23-NDM-183	7 Aug 23	6 Aug 24	-
4	Noise Dosimeter	Noise Dosimeter	Svantek	SV 104IS 106069	Innovative Instrument Co.,Ltd.	24-NDM-018	25 Jan 24	24 Jan 25	-
5	Noise Dosimeter	Noise Dosimeter	Svantek	SV 104IS 128372	Innovative Instrument Co.,Ltd.	24-NDM-073	21 Mar 24	20 Mar 25	-
6	Sound Level Meter	Octave Band	Rion, Japan	NL-62 00130357	Innovative Instrument Co.,Ltd.	CP20230291EA	3 Aug 23	2 Aug 24	-
7	Air Sampling Pump	Total Dust Respirable Dust	Sensidyne	GiAir 5 20120301021	Innovative Instrument Co., Ltd.	23-ASP-073	9 Mar 23	8 Mar 24	-
8	Air Sampling Pump	Total Dust Respirable Dust	Sensidyne	GiAir 5 20170701010	Innovative Instrument Co., Ltd.	23-ASP-075	9 Mar 23	8 Mar 24	-

List of Instruments Certification for Water Quality Analysis

No.	Instrument/Equipment	Parameter	Manufacturer	Model/Serial No.	Calibrator	Certification No.	Date of Calibration	Due date of Calibration	Remark
Water									
1	pH Meter	pH	YSI	pH100A JC03354	Technology Promotion Association (Thailand-Japan)	23CH1487	22 Dec 23	21 Dec 24	-
2	Conductivity Meter	Conductivity	YSI	Pro30 18K100976	Technology Promotion Association (Thailand-Japan)	23CH1488	22 Dec 23	21 Dec 24	-

Envi Equipment Service Co., Ltd.

110/254 Moo 3, Tumbon Bang Rak Phatthana, Amphur Bang Bua Thong, Nonthaburi 11110

Tel. 098 362 9152, 089 478 7885

E-mail: sales@envi-ees.com

Certificate No. : E23-12097

Page : 1 of 6

CERTIFICATE OF CALIBRATION

Customer : United Analyst and Engineering Consultant Co., Ltd.

Address : 81 Soi Udomsuk 41, Sukhumvit Road, Bangchak, Phrakhanong, Bangkok 10260

Description of Equipment : Console meter

Manufacturer : Apex Instrument

Model Number : XC-572-V

Serial Number : 0803018

ID./Control No. : UAE.ANV.193/2551 S/N 0803018

Environment Conditions : Temperature (25 ± 2) °C
Humidity (50 ± 15) % RH

Cal. Date : 24/12/2023

Issue Date : 24/12/2023

Calibration Method or Calibration Procedure Used

US EPA Method (United State Environmental Protection Agency)

This certificate is traceable to national standard, which realize the units of measurement according to the International System of Units (IS).

Result of Calibration

This certificate may not be reproduced other than in full except with prior Written approval of the Technical Manager, Envi Equipment Service Company Limited.

These reported uncertainties of measurement are expanded by a coverage factor of k=2, providing a 95% confidence level

Calibrated by : Mr. Sanya Sangnil

Approved by : 

(Mr. Mana Fuekhud)

Technical Manager

เอกสารไม่ควบคุม

Certificate No. : E23-12097

Page : 2 of 6

METHOD 5 CONSOLE CALIBRATION USING REFERENCE WET GAS METER W-NK-2.5-B-Z No.547425 5-POINT METRIC UNIT

Meter Console Information		Calibration Conditions				Factors/Conversions		
Console Model Number	XC-572-V	Date	Time	24/12/2023	10:45 AM	Std Temp	293	K
Console Serial Number	0803018	Calibration Reference No.		SER23-12040		Std Press	760	mm Hg
DGM Model Number	SK25EX	Barometric Pressure		765.81	mm Hg	K ₁	0.386	
DGM Serial Number	00009766	Calibration Meter Gamma		0.999		Console Leak Check	PASS	

Calibration Data									
Run Time	Metering Console					Calibration Meter			
Elapsed	DGM Orifice DH	Volume Initial	Volume Final	Outlet Temp Initial	Outlet Temp Final	Volume Initial	Volume Final	Outlet Temp Initial	Outlet Temp Final
(Q)	(P _m)	(V _{mi})	(V _{mf})	(t _{mi})	(t _{mf})	(V _{wi})	(V _{wf})	(t _{wi})	(t _{wf})
min	mm H ₂ O	m ³	m ³	°C	°C	m ³	m ³	°C	°C
12.30	13.0	0.1650	0.3050	27	27	186.76006	186.89848	25	25
12.27	13.0	0.3050	0.4450	27	27	186.89848	187.03686	25	25
8.55	26.0	0.4520	0.5920	27	27	187.04374	187.18152	25	25
8.55	26.0	0.5920	0.7320	28	28	187.18152	187.31926	25	25
14.05	40.0	0.7430	1.0230	28	28	187.32982	187.60284	25	25
14.02	40.0	1.0230	1.3030	28	28	187.60284	187.87292	25	25
10.33	70.0	1.3160	1.5960	28	28	187.88578	188.15852	25	25
10.30	70.0	1.5960	1.8760	28	28	188.15852	188.43032	25	25
9.03	90.0	1.8870	2.1670	29	29	188.44084	188.71052	24	24
9.02	90.0	2.1670	2.4470	29	29	188.71052	188.97928	24	24



เอกสารไม่ควบคุม

METHOD 5 CONSOLE CALIBRATION
USING REFERENCE WET GAS METER W-NK-2.5-B-Z No.547425
5-POINT METRIC UNIT

Meter Console Information		Calibration Conditions				Factors/Conversions		
Console Model Number	XC-572-V	Date	Time	24/12/2023	10:45 AM	Std Temp	293	K
Console Serial Number	0803018	Calibration Reference No.	SER23-12040			Std Press	760	mm Hg
DGM Model Number	SK25EX	Barometric Pressure	765.81			K ₁	0.386	
DGM Serial Number	00009766	Calibration Meter Gamma	0.999			Console Leak Check	PASS	

Calibration Data								
Results								
Standardized Data				Dry Gas Meter				
Dry Gas Meter		Calibration Meter		Calibration Factor		Flowrate		
(V _{m(std)})	(Q _{m(std)})	(V _{w(std)})	(Q _{w(std)})	Value	Variation	Std & Corr	.0212 m ³ /min	Variation
m ³	m ³ /min	m ³	m ³ /min	(Y)	(ΔY)	(Q _{m(std)(corr)})	(ΔH _g)	(ΔH _g)
0.139	0.011	0.137	0.011	0.986	0.017	0.011	46.929	-0.077
0.139	0.011	0.137	0.011	0.986	0.016	0.011	46.702	-0.304
0.139	0.016	0.136	0.016	0.981	0.011	0.016	45.888	-1.118
0.139	0.016	0.136	0.016	0.980	0.011	0.016	45.915	-1.091
0.278	0.020	0.270	0.019	0.970	0.001	0.019	48.680	1.674
0.278	0.020	0.267	0.019	0.960	-0.010	0.019	49.510	2.504
0.279	0.027	0.270	0.026	0.967	-0.003	0.026	46.441	-0.565
0.279	0.027	0.269	0.026	0.963	-0.007	0.026	46.462	-0.545
0.281	0.031	0.268	0.030	0.954	-0.016	0.030	46.694	-0.313
0.281	0.031	0.267	0.030	0.951	-0.019	0.030	46.840	-0.166
				0.970	Y Average			ΔH _g Average
								47.006


Note: For Calibration Factor Y, the ratio of the reading of the calibration meter to the dry gas meter, acceptable tolerance of individual values from the average is ±0.02.

For ΔH_g, orifice pressure differential that equates to 0.75 cfm (0.0212 m³/min) at standard temperature and pressure, acceptable tolerance of individual values from the average is ±0.2 inches (5.1mm) H₂O.



เอกสารไม่ควบคุม

Meter Console Information		Calibration Conditions				Factors/Conversions		
Console Model Number	XC-572-V	Date	Time	24/12/2023	10:45 AM	Std Temp	293	K
Console Serial Number	0803018	Calibration Reference No.	SER23-12040			Std Press	760	mm Hg
DGM Model Number	SK25EX	Barometric Pressure	765.81			K ₁	0.386	
DGM Serial Number	00009766	Calibration Meter Gamma	0.999			Console Leak Check	PASS	



เอกสารไม่ควบคุม

Meter Console Information	
Console Model Number	XC-572-V
Console Serial Number	0803018
DGM Model Number	SK25EX
DGM Serial Number	00009766


Calibration Conditions			
Date	Time	24/12/2023	10:45 AM
Calibration Reference No.	SER23-12040		
Barometric Pressure	765.81	mmHg	
Calibration Meter Gamma	0.999		

Factors/Conversions		
Std Temp	293	K
Std Press	760	mm Hg
K ₁	0.386	
Console Leak Check	PASS	

Calibration Date: 24-12-2023

Calibration Reference No: SER23-12040

Meter Pressure vs Flowrate



Console Serial: 0803018

Console Model: XC-572-V



เอกสารไม่ควบคุม

THERMOCOUPLES SYSTEM CALIBRATION

Sampling System Equipment Information	
Console Model Number	XC-572-V
Console Serial Number	0803018
DGM Model Number	SK25EX
DGM Serial Number	00009766
Meter Box Model Number	JENCO 765 KF
Meter Box Serial Number	JC 16095

Calibration Conditions			
Date	Time	24/12/2023	00:45 PM
Calibration Reference No.	SER23-12040		
Reference Thermometer	DIGICON		
Serial Number	183169105		

Results

Console Thermocouple Simulator											
Channel and test point	Meter Box Channel Temperature Reading (°C)										
	-18.0	25.0	38.0	93.0	149.0	260.0	371.0	482.0	593.0	816.0	1038.0
Stack	-17.0	25.0	38.0	94.0	150.0	259.0	371.0	482.0	593.0	816.0	1038.0
Aux	-17.0	25.0	38.0	94.0	150.0						
Probe	-17.0	25.0	38.0	94.0	150.0						
Filter	-17.0	26.0	38.0	94.0	150.0						
Oven	-	-	-	-	-						
Exit	-17.0	26.0	38.0								

Tolerance Range

Stack ± 1.50% Absolute
Probe ± 3.0 °C
Filter ± 3.0 °C

Meter ± 3.0 °C
Exit ± 2.0 °C



เอกสารไม่ควบคุม

Envi Equipment Service Co., Ltd.

110/254 Moo 3, Tumbon Bang Rak Phatthana, Amphur Bang Bua Thong, Nonthaburi 11110

Tel. 098 362 9152, 089 478 7885

E-mail: sales@envi-ees.com

Certificate No. : E23-12095

Page : 1 of 6

CERTIFICATE OF CALIBRATION

Customer : United Analyst and Engineering Consultant Co., Ltd.

Address : 81 Soi Udomsuk 41, Sukhumvit Road, Bangchak, Phrakhanong, Bangkok 10260

Description of Equipment : Console meter

Manufacturer : Apex Instrument

Model Number : XC-572-V

Serial Number : 1701018

ID./Control No. : -

Environment Conditions : Temperature (25 ± 2) °C

: Humidity (50 ± 15) % RH

Cal. Date : 09/12/2023

Issue Date : 09/12/2023

Calibration Method or Calibration Procedure Used

US EPA Method (United State Environmental Protection Agency)

This certificate is traceable to national standard, which realize the units of measurement according to the International System of Units (IS).

Result of Calibration

This certificate may not be reproduced other than in full except with prior Written approval of the Technical Manager, Envi Equipment Service Company Limited.

These reported uncertainties of measurement are expanded by a coverage factor of k=2, providing a 95% confidence level

Calibrated by : Mr. Sanya Sangnil

Approved by : 

(Mr. Mana Fuekhud)

Technical Manger

เอกสารไม่ควบคุม

Certificate No. : E23-12095

Page : 2 of 6

METHOD 5 CONSOLE CALIBRATION USING REFERENCE WET GAS METER W-NK-2.5-B-Z No.547425 5-POINT METRIC UNIT

Meter Console Information		Calibration Conditions				Factors/Conversions	
Console Model Number	XC-572-V	Date	Time	09/12/2023	11:40 AM	Std Temp	293 K
Console Serial Number	1701018	Calibration Reference No.		SER23-12095		Std Press	760 mm Hg
DGM Model Number	SK25EX	Barometric Pressure		759.66	mmHg	K ₁	0.386
DGM Serial Number	00002030	Calibration Meter Gamma		0.999		Console Leak Check	PASS

Calibration Data									
Run Time	Metering Console					Calibration Meter			
Elapsed	DGM Orifice DH	Volume Initial	Volume Final	Outlet Temp Initial	Outlet Temp Final	Volume Initial	Volume Final	Outlet Temp Initial	Outlet Temp Final
(Q)	(P _m)	(V _{mi})	(V _{mf})	(t _{mi})	(t _{mf})	(V _{wi})	(V _{wf})	(t _{wi})	(t _{wf})
min	mm H ₂ O	m ³	m ³	°C	°C	m ³	m ³	°C	°C
12.48	13.0	2395.552	2395.692	29	29	184.31872	184.46072	28	28
12.48	13.0	2395.692	2395.832	29	29	184.46072	184.60178	28	28
8.70	26.0	2395.839	2395.979	30	30	184.60924	184.75154	28	28
8.72	26.0	2395.979	2396.119	30	30	184.75154	184.89358	28	28
14.52	40.0	2396.133	2396.413	31	31	184.90784	185.19162	28	28
14.47	40.0	2396.413	2396.693	32	32	185.19162	185.47356	27	27
10.77	70.0	2396.701	2396.981	32	32	185.48168	185.76272	27	27
10.73	70.0	2396.981	2397.261	32	32	185.76272	186.04256	27	27
9.43	90.0	2397.272	2397.552	33	33	186.05334	186.33194	27	27
9.42	90.0	2397.552	2397.832	33	33	186.33194	186.60946	27	27



เอกสารไม่ควบคุม

METHOD 5 CONSOLE CALIBRATION
USING REFERENCE WET GAS METER W-NK-2.5-B-Z No.547425
5-POINT METRIC UNIT

Meter Console Information		Calibration Conditions				Factors/Conversions		
Console Model Number	XC-572-V	Date	Time	09/12/2023	11:40 AM	Std Temp	293	K
Console Serial Number	1701018	Calibration Reference No.	SER23-12095			Std Press	760	mm Hg
DGM Model Number	SK25EX	Barometric Pressure	759.66	mmHg		K ₁	0.386	
DGM Serial Number	00002030	Calibration Meter Gamma	0.999			Console Leak Check	PASS	

Calibration Data								
Results								
Standardized Data				Dry Gas Meter				
Dry Gas Meter		Calibration Meter		Calibration Factor		Flowrate		Variation
(V _{m(std)})	(Q _{m(std)})	(V _{w(std)})	(Q _{w(std)})	Value	Variation	Std & Corr	.0212 m ³ /min	
m ³	m ³ /min	m ³	m ³ /min	(Y)	(ΔY)	(Q _{m(std)(corr)})	(ΔH _@)	(ΔH _@)
m ³	m ³ /min	m ³	m ³ /min			m ³ /min	mm H ₂ O	
0.136	0.011	0.138	0.011	1.012	0.011	0.011	46.771	-0.951
0.136	0.011	0.137	0.011	1.005	0.005	0.011	47.396	-0.325
0.137	0.016	0.138	0.016	1.013	0.012	0.016	45.356	-2.365
0.137	0.016	0.138	0.016	1.011	0.010	0.016	45.697	-2.024
0.273	0.019	0.276	0.019	1.009	0.008	0.019	48.982	1.261
0.274	0.019	0.275	0.019	1.002	0.001	0.019	49.119	1.398
0.275	0.026	0.274	0.025	0.996	-0.005	0.025	48.195	0.473
0.275	0.026	0.273	0.025	0.992	-0.009	0.025	48.308	0.587
0.276	0.029	0.272	0.029	0.985	-0.015	0.029	48.591	0.869
0.276	0.029	0.271	0.029	0.982	-0.019	0.029	48.797	1.076
				1.001	Y Average		47.721	ΔH _@ Average


Note: For Calibration Factor Y, the ratio of the reading of the calibration meter to the dry gas meter, acceptable tolerance of individual values from the average is ±0.02.

For ΔH_@, orifice pressure differential that equates to 0.75 cfm (0.0212 m³/min) at standard temperature and pressure, acceptable tolerance of individual values from the average is ±0.2 inches (5.1mm) H₂O.



เอกสารไม่ควบคุม

Meter Console Information		Calibration Conditions				Factors/Conversions		
Console Model Number	XC-572-V	Date	Time	09/12/2023	11:40 AM	Std Temp	293	K
Console Serial Number	1701018	Calibration Reference No.	SER23-12095			Std Press	760	mm Hg
DGM Model Number	SK25EX	Barometric Pressure	759.66	mmHg		K ₁	0.386	
DGM Serial Number	00002030	Calibration Meter Gamma	0.999			Console Leak Check	PASS	



เอกสารไม่ควบคุม

Certificate No. : E23-12095

Page : 5 of 6

Meter Console Information	
Console Model Number	XC-572-V
Console Serial Number	1701018
DGM Model Number	SK25EX
DGM Serial Number	00002030


Calibration Conditions			
Date	Time	09/12/2023	11:40 AM
Calibration Reference No.	SER23-12095		
Barometric Pressure	759.66	mmHg	
Calibration Meter Gamma	0.999		

Factors/Conversions		
Std Temp	293	K
Std Press	760	mm Hg
K ₁	0.386	
Console Leak Check	PASS	

Calibration Date: 9-12-2023

Calibration Reference No: SER23-12038

Meter Pressure vs Flowrate



Console Serial: 1701018

Console Model: XC-572-V



เอกสารไม่ควบคุม

Certificate No. : E23-12095

Page : 6 of 6

THERMOCOUPLES SYSTEM CALIBRATION

Sampling System Equipment Information	
Console Model Number	XC-572-V
Console Serial Number	1701018
DGM Model Number	SK25EX
DGM Serial Number	00002030
Meter Box Model Number	JENCO 765 KF
Meter Box Serial Number	JC 16103

Calibration Conditions			
Date	Time	09/12/2023	01:45 PM
Calibration Reference No.	SER23-12095		
Reference Thermometer	DIGICON		
Serial Number	183169105		

Results

Console Thermocouple Simulator											
Channel and test point	Meter Box Channel Temperature Reading (°C)										
	-18.0	25.0	38.0	93.0	149.0	260.0	371.0	482.0	593.0	816.0	1038.0
Stack	-17.0	25.0	38.0	94.0	150.0	259.0	372.0	482.0	593.0	815.0	1037.0
Aux	-17.0	25.0	38.0	94.0	150.0						
Probe	-17.0	25.0	38.0	94.0	150.0						
Filter	-17.0	25.0	39.0	94.0	150.0						
Oven	-17.0	25.0	38.0	94.0	150.0						
Exit	-17.0	26.0	39.0								

Tolerance Range

Stack ± 1.50% Absolute
Probe ± 3.0 °C
Filter ± 3.0 °C

Meter ± 3.0 °C
Exit ± 2.0 °C



เอกสารไม่ควบคุม

Certificate No: G 660610
Date of issue : 05-Oct-23

Instrument description : Flue Gas Analyzer
Instrument model : Testo 350 New
Control unit serial no. : 02398589/208
Instrument serial no. : 02376344/208
ID no. or control no. : UAE.EMA2.113/2555
Manufacturer : Testo SE & Co. KGaA
Probe description : -
Probe model : -
Probe serial no. : -
Customer name : UNITED ANALYST AND ENGINEERING CONSULTANT CO.,LTD.
Customer address : 81 SOI UDOMSUK41,SUKHUMVIT ROAD,BANGCHAK PRAKANONG BANGKOK 10260

Total pages of certificate : 2 Pages
Receiving no. : L-233244
Receiving date. : 28-Sep-23
Parameter of calibration : Gas Calibration(Oxygen 2.498,10.04,21.02 %vol, Carbon Monoxide 80.14,302,1003 ppm, Nitrogen Dioxide 30.34,80.96,201.9 ppm, Nitric Oxide 30.01,151.5,322.5 ppm, Sulphur Dioxide 50.36,100.8,600.8 ppm)
Condition of UUC. : Used
Ambient condition : All of the Measurment were caried out the stabilized labotary
Temperature : 23 ±5 °C
Humidity : 55 ± 15 %RH
Calibration place : 17/121 Soi Ngamwongwan 47 Yaek 48, Toongsonghong, Laksi, Bangkok 10210
Calibration procedure no. : This instrument was calibrated by comparison with Standard gas mixture according to calibration Work Instruction no. WI-CL-28-C

*The calibration certificate expanded uncertainty of measurement is stated as the standard uncertainty of measurent Multiplied by coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%. This certificate is applied only to item under test Environmental condition.
This Calibration Certificate may not be reporduced other than in full except with the permission of the issuing laboratory. Calibration certificates without signature and seal not valid and The results relate only to the items tested/calibrated.
This calibration certificate documents are tracebility to national standards, which realize measurement according to the International System of Units (SI).*

Date of calibration : 04-Oct-23

Kwanchai
Mr. Kwanchai Khamdoun
Calibration Technician

Wongsettee
Mrs. Nongluck Wongsettee
Technical Manager

Certificate No.: G 660610

Standard References (Table 1)

Standard	Certificate No.	Vendor	Due date
Oxygen (O2) 2.498 % Vol	4219/21	Linde	30-Sep-25
Oxygen (O2) 10.04 % Vol	CG-0153-21	Nimt	18-Nov-26
Oxygen (O2) 21.02 % Vol	CG-0041-22	Nimt	10-Feb-27
Carbon monoxide (CO) 80.14 ppm	CG-0040-22	Nimt	14-Feb-27
Carbon monoxide (CO) 302 ppm	1915/23	Linde	16-Jun-25
Carbon monoxide (CO) 1003 ppm	2584/23	Linde	10-Sep-25
Nitrogen Dioxide (NO2) 30.34 ppm	2703/22	Linde	22-Aug-24
Nitrogen Dioxide (NO2) 80.96 ppm	3240/21	Linde	26-Jun-24
Nitrogen Dioxide (NO2) 201.9 ppm	1975/23	Linde	17-Jul-25
Nitric Oxide (NO) 30.01 ppm	CG-0014-23	Nimt	19-Feb-25
Nitric Oxide (NO) 151.5 ppm	0161/23	Linde	22-Jan-25
Nitric Oxide (NO) 322.5 ppm	1974/23	Linde	17-Jul-25
Sulphur Dioxide (SO2) 50.36 ppm	2004/23	Linde	17-Jul-25
Sulphur Dioxide (SO2) 100.8 ppm	3507/22	Linde	09-Nov-24
Sulphur Dioxide (SO2) 600.8 ppm	2003/23	Linde	17-Jul-25

Measured room conditions

Temperature : 22.5 °C Humidity : 66.7 %RH Pressure : 1007.5 mbar

Calibration conditions

Gas Temperature : 23 °C Flow rate : 1,100 ml/min Gas pressure : 1019.4 mbar

Calibration Results (Without adjustment) (Table 2)

Parameter of Standard	Standard Values	Mean of UUC	Error	Uncertainty (±)
O2 (%Vol)	2.498	2.53	0.032	0.15
O2 (%Vol)	10.04	10.09	0.05	0.20
O2 (%Vol)	21.02	21.11	0.09	0.30
CO (ppm)	80.14	81	0.86	3.0
CO (ppm)	302	302	0	6.0
CO (ppm)	1003	1000	-3	12
NO2 (ppm)	30.34	28.2	-2.14	8.0
NO2 (ppm)	80.96	79.4	-1.56	8.0
NO2 (ppm)	201.9	200.8	-1.1	12
NO (ppm)	30.01	28	-2.01	8.0
NO (ppm)	151.5	153	1.5	8.0
NO (ppm)	322.5	325	2.5	12
SO2 (ppm)	50.36	51	0.64	6.0
SO2 (ppm)	100.8	101	0.2	6.0
SO2 (ppm)	600.8	600	-0.8	13

Remark : 1 cmol/mol = 1 %vol. 1 µmol/mol = 1 ppm.

End of Report

Certificate No: G 660472

Date of issue : 11-Aug-23

Instrument description : Flue Gas Analyzer
Instrument model : Testo 350 New
Control unit serial no. : 03064673/609
Instrument serial no. : 60723967/609
ID no. or control no. : UAE.EFM.027/2559
Manufacturer : Testo SE & Co. KGaA
Probe description : -
Probe model : -
Probe serial no. : -
Customer name : UNITED ANALYST AND ENGINEERING CONSULTANT CO., LTD.
Customer address : 81 SOI UDOMSUK41, SUKHUMVIT ROAD, BANGCHAK, PRAKANONG BANGKOK 10260

Total pages of certificate : 2 Pages
Receiving no. : L-232591
Receiving date. : 04-Aug-23
Parameter of calibration : Gas Calibration(Oxygen 2.498,10.04,21.02 %vol, Carbon Monoxide 80.14,302,1003 ppm, Nitrogen Dioxide 30.34, 80.96, 201.9 ppm, Nitric Oxide 30.01, 151.5, 322.5 ppm, Sulphur Dioxide 50.36, 100.8, 600.8 ppm)
Condition of UUC. : Used
Ambient condition : All of the Measurment were caried out the stabilized labotary
Temperature : 23 ± 5 °C
Humidity : 55 ± 15 %RH
Calibration place : 17/121 Soi Ngamwongwan 47 Yaek 48, Toongsonghong, Laksi, Bangkok 10210
Calibration procedure no. : This instrument was calibrated by comparison with Standard gas mixture according to calibration work instration no. WI-CL-28-C

The calibration certificate expanded uncertainty of measurement is stated as the standard uncertainty of measurent Multiplied by coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.
This certificate is applied only to item under test Environmental condition.
This Calibration Certificate may not be reporduced other than in full except with the permission of the issuing laboratory.
Calibration certificates without signature and seal not valid and The results relate only to the items tested/calibrated.
This calibration certificate documents are tracebility to national standards, which realize measurement according to the International System of Units (SI).

Date of calibration : 10-Aug-23

Kwanchoi K.

Mr. Kwanchai Khamdoun

Calibration Tecnician

D. Wongsatee

Mrs. Nongluck Wongsettee

Technical Manager

Certificate No.: G 660472

Standard References (Table 1)

Standard	Certificate No.	Vendor	Due date
Oxygen (O2) 2.498 % Vol	4219/21	Linde	30-Sep-25
Oxygen (O2) 10.04 % Vol	CG-0153-21	Nimt	18-Nov-26
Oxygen (O2) 21.02 % Vol	CG-0041-22	Nimt	10-Feb-27
Carbon monoxide (CO) 80.14 ppm	CG-0040-22	Nimt	14-Feb-27
Carbon monoxide (CO) 302 ppm	1915/23	Linde	15-Feb-27
Carbon monoxide (CO) 1003 ppm	2583/22	Linde	16-Feb-27
Nitrogen Dioxide (NO2) 30.34 ppm	2703/22	Linde	22-Aug-24
Nitrogen Dioxide (NO2) 80.96 ppm	3240/21	Linde	26-Jun-24
Nitrogen Dioxide (NO2) 201.9 ppm	1975/23	Linde	17-Jul-25
Nitric Oxide (NO) 30.01 ppm	CG-0014-23	Nimt	19-Feb-25
Nitric Oxide (NO) 151.5 ppm	0161/23	Linde	22-Jan-25
Nitric Oxide (NO) 322.5 ppm	1974/23	Linde	17-Jul-25
Sulphur Dioxide (SO2) 50.36 ppm	2004/23	Linde	17-Jul-25
Sulphur Dioxide (SO2) 100.8 ppm	3507/22	Linde	09-Nov-24
Sulphur Dioxide (SO2) 600.8 ppm	2003/23	Linde	17-Jul-25

Measured room conditions

Temperature : 22.6 °C Humidity : 68.1 %RH Pressure : 1007.1 mbar

Calibration conditions

Gas Temperature : 23 °C Flow rate : 1,200 ml/min Gas pressure : 1020.3 mbar

Calibration Results (Without adjustment) (Table 2)

Parameter of Standard	Standard Values	Mean of UUC	Error	Uncertainty (±)
O2 (%Vol)	2.498	2.55	0.052	0.15
O2 (%Vol)	10.04	10.08	0.04	0.20
O2 (%Vol)	21.02	21.13	0.11	0.30
CO (ppm)	80.14	80	-0.14	3.0
CO (ppm)	302	301	-1	6.0
CO (ppm)	1003	999	-4	12
NO2 (ppm)	30.34	26.5	-3.84	8.0
NO2 (ppm)	80.96	78.4	-2.56	8.0
NO2 (ppm)	201.9	204.5	2.6	12
NO (ppm)	30.01	28	-2.01	8.0
NO (ppm)	151.5	152	0.5	8.0
NO (ppm)	322.5	318	-4.5	12
SO2 (ppm)	50.36	49	-1.36	6.0
SO2 (ppm)	100.8	100	-0.8	6.0
SO2 (ppm)	600.8	599	-1.8	13

Remark : 1 cmol/mol = 1 %vol, 1 µmol/mol = 1 ppm.

End of Report

MULTI-POINT GAS TEST REPORT

Test Date : Dec 15,2023

Equipment :	Hydrocarbon Analyzer	Model :	55i
Manufacturer :	Thermo SCIENTIFIC	Serial Number :	1182920025

Standard Gas Concentration

Sulphur Dioxide (SO ₂)	-	PPM
Nitric Oxide (NO)	-	PPM
Methane (CH ₄)	39.8	PPM
Carbon Monoxide (CO)	-	PPM
Cylinder No. :	D824432	
Expiration Date :	Aug 4, 2028	

Dilutor Detail

Manufacturer :
Model :
Serial Number :


Multi-point gas test data

Reference Value (ppm)			Analyzer Display (ppm)	Difference Error	Percent Error	[% Error]
Level 1	Zero	0.00	0.45	0.45	0.45	0.45
Level 2	80.00%	40.00	39.50	-0.50	-1.27	1.27
Remark : Measuring Range			50.00 ppm	Average Difference (%)		0.86

Remark : Measuring Range 50.00 ppm

:Acceptable Limit + 5%

Multi-Point Gas Test Chart




Calculate by

15 / 12 / 2023

Approve by _____

.....16...../.....Dec...../.....2023.....



Certificate Of Analysis
Special Gases Mixture

Customer Details

Name:	Address:	Customer Tag No.
United Analyst & Engineering Co., Ltd.	3 Soi Udomsuk 41, Sukhumvit Rd., Bang Chak, Khet Phra Khanong, Bangkok 10260	

Certificate Details

Number:	3384/20	Date of Issue:	4-Aug-2020	Expiry date:	4-Aug-2028
Material Details					
Production Order:	90161442	Material Code:	40040-AL-34	Cylinder No.:	D824432
Gas content:	6.60 M ³	Filling pressure:	137.0 bar	Valve:	CGA 590 BRASS
Cylinder Owner:	LINDE	Filling Material:	Aluminum	Cylinder Size:	50A

Laboratory Report

Analytical Result

Component	Normal Concentration	Analysis Result ¹	Uncertainty ²	Method of Analysis ³	Assay Date
Methane In Air	40.0 ppm	39.8 ppm	± 1% relative	(6) I-PB-352	4-Aug-2020

Reference Standard used in Assay

Reference Standard	Cylinder number	Concentration	Expiry date:
Methane In Nitrogen	2559995G	49.29 ± 0.39 ppm	4-Oct-2020

Analytical Instruments used in Assay

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
FTIR Spectrometers Nicolet iS50	FTIR-CH4	4-Aug-2020

Recommend usage condition

Minimum utilization: 5% of actual content or before expire date whichever comes first.

Storage condition: Keep in well ventilation and secure area

Comments

When reordering, please quote the material number

Note:

1. All results expressed in this report are on mole/mole basis, unless otherwise specified. The Assay of this Standard has been performed in accordance with the EPA Traceability Protocol EPA-600/8-12/531 for the Assay and Certification of Gaseous Calibration Standards using procedure G1
2. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The measurement of this material is traceable to the SI through the reference gas standard which is traceable to Swiss National Standard of Mass or other recognised national metrology institutes.
(1) Gas Chromatography, (2) Paramagnetic Oxygen Analyzer, (3) Electrochemical Oxygen Analyzer, (4) Electrochemical Moisture Analyzer,
(5) Total Hydrocarbon Analyzer, (6) Other - Specified

Sukanya Parinyasontorn
Signatory for and on behalf of Linde (Thailand) Co., Ltd.

PB-002/F006

Lindo (Thailand) Public Company Limited 81/1, 01 October 2019

Linde (Thailand) Public Company Limited/ 1, 6 October 2017

PLC Registration No. 0107337000705

15th Floor, Bangna Tower A, 2/3 Moo 14, Bangna Trad KM. 6.5 Road, Bangkaew

Bangphee, Samutprakarn 10540, Tel (66) 2338-6100 Fax (66) 2338-6333

Wellgrow Plant: 105 Moo 5, T.Bangsamak, A.Bangpakong, Chachoengsao 24180

Thailand, Tel (66) 38.570-479-93 Fax (66) 38.570-323

CERTIFICATE OF CALIBRATION

Certificate No. : CL-003-65

Page 1 of 2 Pages

MEASUREMENT ITEM : Top Load Orifice
MANUFACTURER : Tisch Environmental, Inc.
MODEL/TYPE : TE-5025A
SERIAL NUMBER : 3383
ID NUMBER : UAE.EFM.063/2560
CONDITION AS-RECEIVED : Used item
CUSTOMER : United Analyst and Engineering Consultant Co.,Ltd.
81 Soi Udomsuk 41, Sukhumvit Road, Bangchak, Phrakhanong,
Bangkok 10260

RECEIVED DATE : 15 Jul 2022
MEASUREMENT DATE : 25 Jul 2022
ISSUE DATE : 26 Jul 2022

ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature : 23.0 ± 3.0 °C
Relative Humidity : 55.0 ± 15.0 %RH
Atmospheric Pressure : 1010 ± 10 hPa

CALIBRATION CONDITION:

Preconditioning : 24 hours at ambient conditions.
Measurement Condition : The average values during measurement are 24.8 °C and 55.1 %RH.

TABULATION OF RESULTS:

The table on next page give the measured values.

Calibration procedure:
The Orifice gas flow device was calibrated against Standard Rotary Displacement Meter (Roots Meter) Model G65/IMC/W2-dp. The WI-CL-004 was used as a calibration guideline.

Traceability:
This certificate provides a traceability of The measurement to recognized the national standards, and to realization of the international system of units (SI) through the VSL (National Metrology Institute of Netherlands) via Certificate number: G2211901

Uncertainty of Measurement:
The reported uncertainty of measurement is based on the standard uncertainty multiplied by a coverage factor $k=2$. Which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty has been determined in accordance with the GUM 'Evaluation of measurement data - Guide to the expression of uncertainty in measurement'

Calibrated by:
☐ Mr. Sorawit Thachalad
☒ Miss Jitraporn Lertsomphol



Approved signatory:
Mr. Parinya Booncharoen
Calibration Department Manager

Continuation of Certificate of Calibration Number CL-003-65

Page 2 of 2 Pages

MEASUREMENT RESULTS:

The Orifice gas flow device was calibrated by direct comparison method with the Standard Rotary Displacement Meter (Roots Meter). The Humid air was used as a medium in the system. The standard conditions are 25°C (298.15 K) and 760 mmHg for standard temperature and standard pressure respectively.

Table 1: The results of Q Standard calibration data

Plate	Flow rate m ³ /min	Pressure [Pa] mmHg	Temperature [Ta] °C	Temperature [Tm] °C	Δp_meter mmHg	Δp_Orifice inH ₂ O	Y	Standard Flow [Q _s] m ³ /min
1	0.697	754.265	24.640	23.960	55.399	1.699	1.299	0.643
2	1.000	754.236	24.950	24.350	62.172	3.444	1.849	0.913
3	1.118	754.323	24.730	24.210	41.925	4.582	2.133	1.051
4	1.169	754.212	24.640	24.160	31.045	5.150	2.262	1.116
5	1.416	754.175	24.480	24.210	30.117	7.629	2.754	1.353

Slope (m): 2.04804
Intercept (b): -0.01939
Correlation coefficient (r): 0.99982
Uncertainty (k=2): 0.011 m³/min

Table 2: The results of Q actual calibration data

Plate	Flow rate m ³ /min	Pressure [Pa] mmHg	Temperature [Ta] °C	Temperature [Tm] °C	Δp_meter mmHg	Δp_Orifice inH ₂ O	Y	Standard Flow [Q _a] m ³ /min
1	0.697	754.265	24.640	23.960	55.399	1.699	0.819	0.647
2	1.000	754.236	24.950	24.350	62.172	3.444	1.167	0.919
3	1.118	754.323	24.730	24.210	41.925	4.582	1.345	1.058
4	1.169	754.212	24.640	24.160	31.045	5.150	1.426	1.123
5	1.416	754.175	24.480	24.210	30.117	7.629	1.735	1.361

Slope (m): 1.28277
Intercept (b): -0.01223
Correlation coefficient (r): 0.99982
Uncertainty (k=2): 0.012 m³/min

End of Certificate of Calibration





Certificate of Calibration

Certificate No. : 24P1250
Page : 1 of 2

Equipment : U Tube Manometer

Manufacturer: Dwyer

Model : 1221-36-W/M

Serial No.: -

ID No.: UAE.EFM.076/2566

Condition As-Received: Used Item

Received Date: 03 April 2024

Calibration Date: 10 April 2024

Reference: 2404-0118WSC

Submitted by: United Analyst and Engineering Consultant Co.,Ltd.

Ambient Temperature: (23 ± 2) °C

Relative Humidity: (50 ± 15) %

Atmospheric Pressure: 1007 mbar

81 Soi Udomsuk 41, Sukhumvit Road, Bangchak,
Phrakhanong, Bangkok 10260

Procedure used: The calibration was conducted by direct comparison method against Pressure Measuring Instruments Standard according to calibration procedure CP-P04, using " DKD-R 6-1 ; Calibration of Pressure Gauges " as a guidelines.

Condition of this result of calibration

1.Reference standards instruments :

<u>Instrument</u>	<u>Model</u>	<u>Serial No.</u>	<u>Certificate No.</u>	<u>Due Date</u>
1) Pressure Calibrator	PC106P	1189	MP-0176-23	12 Sep 2024

2.This result of calibration was made on requested at the point specified by customer.

3.Scale and conversion factor is 1 kPa = 4.0146293 inH₂O

4.This instrument was used clean air as pressure media.

5.This instrument was calibrated by applied pressure to high-port (+) side and low-port (-) side open to atmospheric pressure.

6.This instrument was installed in vertical orientation and top of the pressure port was used as the reference level.

7.The certificate is valid only to the item calibrated on date and place of calibration.

8.This Certification is traceable to the International System of Unit maintained through:-

-National Institute of Metrology (Thailand), NSC-ONSC Accredited No. Calibration 0144

Calibrated by : Suksan Khankaew
Issue Date : 17 April 2024

Approved Signatory : _____
[] Phalinee Prabpaipal
[] Sura Suwannasri
[✓] Attapol Panurach

เอกสารไม่ควบคุม



Cert.No.: 24P1250
Page: 2 of 2

Result of calibration:- Without adjustment

Function:- Pressure Measurement

Increasing Pressure

Range : 0 inH₂O to 36 inH₂O

Scale Interval : 0.1 inH₂O (The Second Estimate)

<u>Applied Pressure</u>	<u>High-port side</u>	<u>UUC Indication</u>		<u>ΔP</u>	<u>Error</u>
		<u>Low-port side</u>			
0.00	0.00	0.00		0.00	0.00
2.00	1.00	-1.00		2.00	0.00
4.00	2.00	-2.00		4.00	0.00
6.00	3.00	-3.00		6.00	0.00
8.00	4.00	-4.00		8.00	0.00
10.00	5.05	-4.95		10.00	0.00
12.00	6.05	-5.95		12.00	0.00
14.00	7.05	-6.95		14.00	0.00
16.00	8.10	-7.95		16.05	0.05
18.00	9.10	-8.95		18.05	0.05
20.00	10.10	-9.95		20.05	0.05
22.00	11.10	-10.95		22.05	0.05
24.00	12.10	-11.95		24.05	0.05
26.00	13.15	-12.95		26.10	0.10
28.00	14.15	-13.95		28.10	0.10
30.00	15.20	-14.95		30.15	0.15
32.00	16.20	-15.95		32.15	0.15
34.00	17.20	-16.95		34.15	0.15
35.50	18.00	-17.70		35.70	0.20

The uncertainty of measurement was ± 0.11 inH₂O

* ΔP = High-port side - Low-port side

* UUC = Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95 %.

-000-

เอกสารไม่ควบคุม



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG, BANGKOK 10250
TEL. 0-2717-3000-24 FAX. 0-2719-9484



Certificate of Calibration

Certificate No. : 24P1367
Page : 1 of 2

Equipment : Aneroid Barometer

Manufacturer: Barigo

Model : -

Serial No.: -

ID No.: UAE.ANV.152/2550

Condition As-Received: Used Item

Received Date: 05 April 2024

Calibration Date: 22 April 2024

Reference: 2404-0243WSC

Submitted by: United Analyst and Engineering Consultant Co.,Ltd.

Ambient Temperature: (23 ± 2) °C

Relative Humidity: (50 ± 15) %

Atmospheric Pressure: 1007 mbar

81 Soi Udomsuk 41, Sukhumvit Road, Bangchak,
Phrakhanong, Bangkok 10260

Procedure used: The calibration was conducted by direct comparison method against Pressure Measuring Instruments Standard according to calibration procedure CP-P10, using " DKD-R 6-1 ; Calibration of Pressure Gauges " as a guidelines.

Condition of this result of calibration

1.Reference standards instruments :

Instrument	Model	Serial No.	Certificate No.	Due Date
1) Standard Barometer	DPI142	1422505046	MP-0094-23	03 May 2024

2.This instrument was installed in vertical orientation and center of the dial was used as the reference level.

3.This result of calibration was made on requested at the point specified by customer.

4.This result of calibration instrument was in absolute pressure.

5.This instrument was used clean air as pressure media.

6.The certificate is valid only to the item calibrated on date and place of calibration.

7.This Certification is traceable to the International System of Unit maintained through:-

-National Institute of Metrology Thailand (NIMT)

Calibrated by : Suksan Khankaew
Issue Date : 23 April 2024

Approved Signatory : _____
[] Phalinee Prabpaipal
[] Sura Suwannasri
[✓] Attapol Panurach

เอกสารไม่ควบคุม



Cert.No.: 24P1367
Page: 2 of 2

Result of calibration:- Without adjustment

Range : 960 hPa to 1030 hPa

Function:- Absolute Pressure Measurement

Scale Interval : 1 hPa (The Fifth Estimate)

Increasing Pressure

Applied Pressure (hPa)	957.13	968.77	980.13	990.56	1001.26	1011.35	1022.10	1032.61
UUC* Indication (hPa)	960.0	970.0	980.0	990.0	1000.0	1010.0	1020.0	1030.0
Error (hPa)	2.87	1.23	-0.13	-0.56	-1.26	-1.35	-2.10	-2.61

Decreasing Pressure

Applied Pressure (hPa)	1032.61	1021.84	1010.88	1000.82	990.20	979.52	968.48	957.17
UUC* Indication (hPa)	1030.0	1020.0	1010.0	1000.0	990.0	980.0	970.0	960.0
Error (hPa)	-2.61	-1.84	-0.88	-0.82	-0.20	0.48	1.52	2.83

The uncertainty of measurement was ± 0.25 hPa

* UUC = Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95 %.

-000-

เอกสารไม่ควบคุม



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG, BANGKOK 10250
TEL. 0-2717-3000-24 FAX. 0-2719-9484



Certificate of Calibration

Certificate No. : 24H756
Page : 1 of 2

Equipment : Dial Thermo-Hygrometer

Manufacturer: Barigo

Model : -

Serial No.: -

ID No.: UAE.ANV.131/2550

Condition As-Received: Used Item

Received Date: 05 April 2024

Calibration Date: 10 April 2024
to 18 April 2024

Reference: 2404-0247WSC

Submitted by: United Analyst and Engineering Consultant Co.,Ltd.

Ambient Temperature: (25 ± 3) °C

Relative Humidity: (50 ± 20) %

81 Soi Udomsuk 41, Sukhumvit Road,
Bangchak, Phrakhanong, Bangkok 10260

Procedure used: Calibration were conducted using in-house calibration procedure CP-H02 according to comparison with standard chilled mirror sensor for humidity measurement function and comparison with standard temperature probe for temperature measurement function into humidity / temperature chamber.

Condition of this result of calibration

1.Reference standards instruments :

Instrument	Model	Serial No.	Certificate No.	Due Date
1) Chilled Mirror Hygrometer	Dew Master	44730	21656	02 Aug 2024
2) Handheld Thermometer With Sensor	1521	A5A339	2311238	16 Oct 2024

2.The certificate is valid only to the item calibrated on date and place of calibration.

3.This Certification is traceable to the International System of Unit maintained through:-

- Thunder Scientific Corporation, NVLAB Accreditation No. Calibration 200582-0
- Technology Promotion Association (Thailand-Japan), NSC-ONSC Accredited No. Calibration 0008

Calibrated by : Chakrit Waewwanjua
Issue Date : 18 April 2024

Approved Signatory :
[] Chakrit Waewwanjua
[✓] Viporn Tantiyawutti
[] Unnophol Harachai

เอกสารไม่ควบคุม



Cert. No.: 24H756
Page.: 2 of 2

Result of Calibration:- Without Adjustment
Function: Humidity Measurement.

Reference Temperature (°C)	Standard Humidity (%R.H.)	UUC* Reading (%R.H.)	Error (%R.H.)	Uncertainty of Measurement (±%R.H.)
25.0	40.1	44	3.9	1.6
25.0	60.0	61	1.0	1.7
25.0	80.0	76	-4.0	1.8

Result of Calibration:- Without Adjustment
Function: Temperature Measurement.

Standard Temperature (°C)	UUC* Reading (°C)	Error (°C)	Uncertainty of Measurement (±°C)
20.007	20.5	0.493	0.72
25.032	25.0	-0.032	0.72
29.997	30.0	0.003	0.72
35.010	34.5	-0.510	0.72
40.019	39.5	-0.519	0.72

UUC* : Unit Under Calibration

The reported uncertainty of measurement was base on standard uncertainty multiplied by coverage factor k = 2.00, providing confidence level approximately 95%.

-o0o-

เอกสารไม่ควบคุม

MULTI-POINT GAS TEST REPORT

Test Date : Nov 1, 2023

Equipment : Gas Analyzer (NO₂) Model : 42i
Manufacturer : Thermo Scientific Serial Number : CM08130002

Standard Gas Concentration


Sulphur Dioxide (SO₂) 44.68 PPM
Nitric Oxide (NO) 45.94 PPM
Methane (CH₄) - PPM
Carbon Monoxide (CO) 984.8 PPM
Cylinder No. : EB0143262
Expiration Date : Jun 21, 2024

Dilutor Detail

Manufacturer : Thermo Scientific
Model : 146i
Serial Number : 1180540071

Multi-point gas test data

Reference Value (ppb)			Analyzer Display (ppb)	Difference Error	Percent Error	[% Error]
Level 1	Zero	0.0	0.9	0.90	0.90	0.90
Level 2	20.00%	100.0	101.3	1.30	1.28	1.28
Level 3	40.00%	200.0	200.9	0.90	0.45	0.45
Level 4	60.00%	300.0	300.9	0.90	0.30	0.30
Level 5	80.00%	400.0	400.0	0.00	0.00	0.00
Remark : Measuring Range			500.0 ppb	Average Difference (%)		0.59



Calculate by

01 Nov 2023

Approve by

01 Nov 2023

MULTI-POINT GAS TEST REPORT

Test Date : Nov 13, 2023

Equipment : Gas Analyzer (NO₂) Model : 42i
Manufacturer : Thermo Scientific Serial Number : CM19050148

Standard Gas Concentration


Sulphur Dioxide (SO₂) 44.68 PPM
Nitric Oxide (NO) 45.94 PPM
Methane (CH₄) - PPM
Carbon Monoxide (CO) 984.8 PPM
Cylinder No. : EB0143262
Expiration Date : Jun 21, 2024

Dilutor Detail

Manufacturer : Thermo Scientific
Model : 146i
Serial Number : 1180540071

Multi-point gas test data

Reference Value (ppb)			Analyzer Display (ppb)	Difference Error	Percent Error	[% Error]
Level 1	Zero	0.0	0.8	0.80	0.80	0.80
Level 2	20.00%	100.0	100.7	0.70	0.70	0.70
Level 3	40.00%	200.0	200.9	0.90	0.45	0.45
Level 4	60.00%	300.0	300.9	0.90	0.30	0.30
Level 5	80.00%	400.0	400.0	0.00	0.00	0.00
Remark : Measuring Range		500.0 ppb	Average Difference (%)			0.45



Calculate by

13 Nov 2023

Approve by

13 Nov 2023

MULTI-POINT GAS TEST REPORT


Test Date : Nov 1, 2023

Equipment : Gas Analyzer (NO₂) Model : 42i
Manufacturer : Thermo Scientific Serial Number : CM19050149

Standard Gas Concentration		Dilutor Detail	
Sulphur Dioxide (SO ₂)	44.68	PPM	Manufacturer : Thermo Scientific
Nitric Oxide (NO)	45.94	PPM	Model : 146i
Methane (CH ₄)	-	PPM	Serial Number : 1180540071
Carbon Monoxide (CO)	984.8		
Cylinder No. :	EB0143262		
Expiration Date :	Jun 21, 2024		

Multi-point gas test data

Reference Value (ppb)			Analyzer Display (ppb)	Difference Error	Percent Error	[% Error]
Level 1	Zero	0.0	0.6	0.60	0.60	0.60
Level 2	20.00%	100.0	100.8	0.80	0.79	0.79
Level 3	40.00%	200.0	200.7	0.70	0.35	0.35
Level 4	60.00%	300.0	300.7	0.70	0.23	0.23
Level 5	80.00%	400.0	400.0	0.00	0.00	0.00
Remark : Measuring Range		500.0 ppb	Average Difference (%)			0.40



Calculate by

.....
01 / Nov / 2023

Approve by

.....
01 / Nov / 2023

CERTIFICATE OF ANALYSIS

Grade of Product: EPA Protocol

Part Number: E04NI99E15A01D3 Reference Number: 122-402135167-1
Cylinder Number: EB0143262 Cylinder Volume: 144.4 CF
Laboratory: 124 - Durham (SAP) - NC Cylinder Pressure: 2015 PSIG
PGVP Number: B22021 Valve Outlet: 660
Gas Code: CO,NO,NOX,SO₂,BALN Certification Date: Jun 21, 2021

Expiration Date: Jun 21, 2024

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 600/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a mole/mole basis unless otherwise noted.

Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

ANALYTICAL RESULTS					
Component	Requested Concentration	Actual Concentration	Protocol Method	Total Relative Uncertainty	Assay Dates
NOX	45.00 PPM	45.96 PPM	G1	+/- 1.4% NIST Traceable	06/14/2021, 06/21/2021
NITRIC OXIDE	45.00 PPM	45.94 PPM	G1	+/- 1.4% NIST Traceable	06/14/2021, 06/21/2021
SULFUR DIOXIDE	45.00 PPM	44.68 PPM	G1	+/- 1.0% NIST Traceable	06/14/2021, 06/21/2021
CARBON MONOXIDE	1000 PPM	984.8 PPM	G1	+/- 0.7% NIST Traceable	06/14/2021
NITROGEN	Balance				

CALIBRATION STANDARDS					
Type	Lot ID	Cylinder No	Concentration	Uncertainty	Expiration Date
NTRM	20061120	CC708068	49.82 PPM NITRIC OXIDE/NITROGEN	+/- 1.0%	Feb 02, 2025
PRM	12386	D685025	9.91 PPM NITROGEN DIOXIDE/AIR	+/- 2.0%	Feb 20, 2020
GMIS	401423838102	CC505581	4,348 PPM NITROGEN DIOXIDE/NITROGEN	+/- 2.1	Feb 18, 2023
NTRM	16011043	CC473277	49.02 PPM SULFUR DIOXIDE/NITROGEN	+/- 0.8%	Jun 17, 2022
NTRM	14080119	CC434277	990.9 PPM CARBON MONOXIDE/NITROGEN	+/-0.6%	Nov 15, 2025
The SRM, PRM or RGM noted above is only in reference to the GMIS used in the assay and not part of the analysis.					

ANALYTICAL EQUIPMENT		
Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
Nicolet 6700 AHR0801333 CO	FTIR	Jun 03, 2021
Nicolet 6700 AHR0801333 NO	FTIR	Jun 03, 2021
Nicolet 6700 AHR0801333 NO ₂	FTIR	Jun 03, 2021
Nicolet 6700 AHR0801333 SO ₂	FTIR	Jun 03, 2021

Triad Data Available Upon Request

NOTES: PO #5221002807

GROSS WT: 28.40kg

NET WT: 4.73kg



The analytical test results reported on this certificate relate only to the cylinder number specified above. This concludes the test report.

[Signature]

Approved for Release



CERT 3082.01

เอกสารไม่ควบคุม Page 3 of 12 2023-11-01

Certificate of Calibration

WL-21 Wireless Anemometer

Scarlet Tech Ltd. hereby certifies that the WL-21 wireless anemometer listed below was thoroughly calibrated, tested and inspected following the standard calibration procedure (st-wl-21) and is within manufacturer's specification at the time when the calibration is done.

Client: Envir Service Co., Ltd.

Serial No.: 2205DT0114

Calibration Date: 2022/9/14

Calibration Expiry Date: 2023/9/13

The Result of Calibration

Velocity				
Measured Value(m/s)	Actual Value (m/s)	Deviation	Tolerance	Result
1.0	1.0	0.0	0.9-1.1	Pass
1.9	2.0	0.1	1.8-2.2	Pass
5.1	5.0	0.1	4.7-5.3	Pass
7.0	7.0	0.0	6.0-8.0	Pass
10.1	10.0	0.1	9.5-10.5	Pass
19.6	20.0	0.4	19.0-21.0	Pass

Wind Direction				
Measured Value	Actual Value	Deviation	Tolerance	Result
45°	45°	0	42-48	Pass
136°	135°	1	132-138	Pass
227°	225°	2	222-228	Pass
316°	315°	1	312-318	Pass
358°	0°	2	357-3	Pass

Inspection Room Temp	Actual Value	Deviation	Tolerance	Result
22.5°C	22.5°C	0.0	21.5-23.5	Pass

Atmospheric Pressure Inspection	Actual Value	Deviation	Tolerance	Result
1005	1005	0	1001-1019	Pass

Environment conditions :

Air temperature: 22 °C
Relative humidity: 55 %
Static pressure: 102.2 kPa

Performed by: 

Certified by Head of Engineering department



This certificate may not be published or reproduced, except in full, unless obtaining permission in writing form from Scarlet Tech Ltd.
4F-3, No. 347, 2nd Sec., Heping E. Rd., Daan Dist. Taipei City 106, Taiwan

เอกสารไม่ควบคุม

Certificate of Calibration

WL-21 Wireless Anemometer

Scarlet Tech Ltd. hereby certifies that the WL-21 wireless anemometer listed below was thoroughly calibrated, tested and inspected following the standard calibration procedure (st-wl-21) and is within manufacturer's specification at the time when the calibration is done.

Client: Envir Service Co., Ltd.

Serial No.: 2205DT0116

Calibration Date: 2022/9/14

Calibration Expiry Date: 2023/9/13

The Result of Calibration

Velocity				
Measured Value(m/s)	Actual Value (m/s)	Deviation	Tolerance	Result
1.0	1.0	0.0	0.9-1.1	Pass
2.1	2.0	0.1	1.8-2.2	Pass
5.0	5.0	0.0	4.7-5.3	Pass
7.1	7.0	0.1	6.0-8.0	Pass
10.2	10.0	0.2	9.5-10.5	Pass
19.9	20.0	0.1	19.0-21.0	Pass

Wind Direction				
Measured Value	Actual Value	Deviation	Tolerance	Result
45°	45°	0	42-48	Pass
134°	135°	1	132-138	Pass
225°	225°	0	222-228	Pass
316°	315°	1	312-318	Pass
358°	0°	2	357-3	Pass

Inspection Room Temp	Actual Value	Deviation	Tolerance	Result
22.5°C	22.5°C	0.0	21.5-23.5	Pass

Atmospheric Pressure Inspection	Actual Value	Deviation	Tolerance	Result
1005	1005	0	1001-1019	Pass

Environment conditions :

Air temperature: 22 °C
Relative humidity: 55 %
Static pressure: 102.2 kPa

Performed by: 

Certified by Head of Engineering department



This certificate may not be published or reproduced, except in full, unless obtaining permission in writing form from Scarlet Tech Ltd.
4F-3, No. 347, 2nd Sec., Heping E. Rd., Daan Dist. Taipei City 106, Taiwan

เอกสารไม่ควบคุม



THAI METEOROLOGICAL DEPARTMENT

4353 Sukhumvit, Bangna, Bangkok 10260 Tel. 081-454-2804,0-2399-0469

Calibration Certificate

Issued by : Calibration & Test Section : Meteorological Instruments Bureau

Date of Issue 22 February, 2024

Certification No. 096/24

Page : 1 of 5

Object : Wind Speed & Wind Direction Data Logger

Manufacturer : SCARLET/TECH

Type : WL-21

Mfg Code : Wireless Receiver 2301DR0024

Wind Sensor 2301DT0024

Customer : United Analyst and Engineering Consultant Co.,Ltd.

81 Soi Udomsuk 41, Sukhumvit Road,

Bangchak, Prakanong, Bangkok 10260.

Calibration Condition : Temperature 25.1 °C Barometric Pressure 1010.1 hPa

NATIONAL STANDARD WIND TUNNEL : Wind Aloft Plotting Board

: Micromanometer Theodor Friedrichs FC014 Serial No. 9310119 : HOOK GAGE NO 1425

N.I.S.T. Test Reference Number 731/241460 : Standard Velocity at 20 - 30 m/sec

: Ultrasonic Anemometer Model DA-650-3TV (sensor TR-90AH)
Serial Number 110730029 (sensor 120629586)

JAPAN QUALITY ASSURANCE ORGANIZATION : Standard Velocity at 0 - 20 m/sec

STANDARD THERMOMETER : Theodor Friedrich : Dry No.8390/94 Wet No. 8389/94

: testo, testo 645 Serial No. 02848057 : Thermoschneider No.918802

STANDARD BAROMETER : Digital Barometer Vaisala Type PTB220 No. V1220015

: Digital Barometer Vaisala Type PTB330 No. V4320001

Calibrated by : Watchapol

Signed :

(Authorised Signatory)

Mr. Watcharapol Subwat

Mr. Pisood Promsut

for the Chief

Mechanical Engineer

Sub-Standard Instrument

เอกสารไม่ควบคุม



THAI METEOROLOGICAL DEPARTMENT

4353 Sukhumvit, Bangna, Bangkok 10260 Tel. 081-454-2804,0-2399-0469

The Result of Calibration

Certification No. 096/24

22 February, 2024

Page : 2 of 5

Standard Ultrasonic Anemometer m/sec	HOOK GAGE NO. 1425			TESTED ANEMOMETER	
	Pressure inches H2O	Vacuum inches H2O	Velocity m/sec	Velocity m/sec	Correction m/sec
1.00	-	-	-	1.0	0.00
3.02	-	-	-	3.0	0.02
5.00	-	-	-	5.0	0.00
7.04	-	-	-	7.0	0.04
9.02	-	-	-	9.0	0.02
11.02	-	-	-	10.9	0.12
13.01	-	-	-	12.9	0.11
15.01	-	-	-	14.9	0.11
17.02	-	-	-	17.0	0.02
20.02	-	-	-	19.9	0.12

Wind Aloft Plotting Board.	
US.DEPARTMENT OF COMMERCE WEATHER BUREAU	
WIND DIRECTION	TESTED WIND DIRECTION
0	0
90	90
180	180
270	270

Calibrated by :

Mr. Watcharapol Subwat

Mechanical Engineer

Calibration & Test Section

Meteorological Instruments Bureau

เอกสารไม่ควบคุม



The Result of Calibration

Certification No. 096/24

22 February, 2024

Page : 3 of 5

Standard Barometer	Tested Barometer	Correction
Pressure	Pressure	mbar
1010.84	1011	-0.16
1010.60	1011	-0.40
1011.71	1012	-0.29
1012.17	1012	0.17
1012.31	1012	0.31
1012.25	1012	0.25
1012.79	1013	-0.21
1012.95	1013	-0.05
1013.52	1014	-0.48
1014.16	1014	0.16
1015.79	1016	-0.21
1016.02	1016	0.02
1015.86	1016	-0.14
1015.69	1015	0.69
1011.51	1012	-0.49
1011.80	1012	-0.20
1012.06	1012	0.06
1012.81	1013	-0.19
1013.22	1013	0.22
1013.49	1013	0.49

Average


-0.02

Calibrated by :

Mr. Watcharapol Subwat

Mechanical Engineer

Calibration & Test Section
Meteorological Instruments Bureau



เอกสารไม่ควบคุม



The Result of Calibration

Certification No. 096/24

22 February, 2024

Page : 4 of 5

Standard Barometer	Tested Barometer	Correction
Pressure	Pressure	mmHg
758.19	758	0.19
758.01	758	0.01
758.84	759	-0.16
759.19	759	0.19
759.29	759	0.29
759.25	759	0.25
759.65	760	-0.35
759.77	760	-0.23
760.20	760	0.20
760.68	761	-0.32
761.90	762	-0.10
762.08	762	0.08
761.96	762	-0.04
761.83	762	-0.17
758.69	759	-0.31
758.91	759	-0.09
759.11	759	0.11
759.67	760	-0.33
759.98	760	-0.02
760.18	760	0.18

Average

-0.03

Calibrated by :

Mr. Watcharapol Subwat

Mechanical Engineer

Calibration & Test Section
Meteorological Instruments Bureau



เอกสารไม่ควบคุม



THAI METEOROLOGICAL DEPARTMENT

4353 Sukhumvit, Bangna, Bangkok 10260 Tel. 081-454-2804, 0-2399-0469

The Result of Calibration

Certification No. 096/24

22 February, 2024

Page : 5 of 5

Standard Temp. °C	Temperature Sensor Reading	
	Reading °C	Correction °C
45.2	45	0.2
30.3	30	0.3
15.8	16	-0.2

Calibrated by :

Watchapol

Mr. Watchapol Subwat
Mechanical Engineer



เอกสารไม่ควบคุม



ELECTRICAL AND ELECTRONICS INSTITUTE FOUNDATION FOR INDUSTRIAL DEVELOPMENT

975 Moo 4, Bangpoo Industrial Estate, Soi 8, Sukhumvit Road km 37,

Phraek Sa, Mueang Samut Prakan, Samut Prakan 10280

Tel: +66 2709 4860 Fax: +66 2324 0917



Certificate No.: CP20230294EA

Operation No.: CP2023070031

Certificate of Calibration

Equipment: Sound Level Meter

Manufacturer: 01dB (Meter), G.R.A.S. (Microphone), 01dB (Preamplifier)

Model/Type: CUBE (Meter), 40CD (Microphone), PRE22 (Preamplifier)

Serial No.: 11073 (Meter), 260921 (Microphone), 1605143 (Preamplifier)

ID No.: UAE.EFM.010/2560 (Meter), UAE.EFM.010/2560 (Extension cable)

Customer: United Analyst and Engineering Consultant Co.,Ltd.

Address: 81 Soi Udomsuk 41, Sukhumvit Road, Bangchak
Phrakhanong, Bangkok 10260

Received Date: 24 July 2023

Calibrated Date: 4 - 7 August 2023

Issued Date: 8 August 2023

Calibrated by: Ms. Juntaporn Kunhakom

Approved by: 

(Mr. Sittichai Swaksuriyawong)
Group Manager

This report was prepared electronically using applicable electronic signature. Printing or copy of file are considered as a copy of the document.

The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor (k) providing a level of confidence of approximately 95%. This certificate may not be reproduced other than in full except with the prior written approval of the Electrical and Electronics Institute, Foundation for Industrial Development.

เอกสารไม่ควบคุม

Certificate No.: CP20230294EA

Calibration Report

Equipment: Sound Level Meter
 Manufacturer: 01dB (Meter), G.R.A.S. (Microphone), 01dB (Preamplifier)
 Model/Type: CUBE (Meter), 40CD (Microphone), PRE22 (Preamplifier)
 Serial No.: 11073 (Meter), 260921 (Microphone), 1605143 (Preamplifier)
 ID No.: UAE.EFM.010/2560 (Meter), UAE.EFM.010/2560 (Extension cable)
 Ambient Temperature: (23 ± 2) °C
 Relative Humidity: (50 ± 15) %
 Pressure: (101.3 ± 1.5) kPa

Method of Calibration :-

IEC 61672-3:2013.

Condition of this result of calibration

1. Reference standards instrument :-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Standard microphone	4180	2787490	AA-1024-22	6 November 2023
2) Arbitrary Function Generator	AFG2021	C010063	CK20230040EA	26 June 2024
3) Programmable Attenuator	PA5	2755	EF-0034-22	30 October 2023
4) 6.5 Digit precision multimeter	8846A	9610014	CB20220223EA	14 November 2023
5) Pressure humidity and Temperature Transmitter	PTU301	F0640002	CL1-P230024 CD20230196EA	20 March 2024 23 July 2024
6) Pressure humidity and Temperature Transmitter	PTU301	F0640003	CL1-P230032 CD20230197EA	4 April 2024 23 July 2024
7) Performance Audio Analyzer	U8903B	MY56510003	CB20230038EA CK20220080EA	14 February 2024 8 September 2023

2. This result of calibration was found accurate as shown on date and place of calibration only.

3. This certification is traceable to the international system of unit maintained at :-

Reference standards instrument for Acoustic function

- National Institute of Metrology (Thailand)

Reference standards instrument for Electrical function

- National Institute of Metrology (Thailand)

- Electrical and Electronics Institute; NSC Accredited Calibration No.0119

Result of Calibration:-

Function : 1. Indication at the calibration check frequency

Reference Acoustic Signal (dB)	Correction for Microphone Model 40CD (dB)	Effective Calibration Level (dB)	Measured value (dB)	Deviation (dB)	Acceptance limits (dB)
94.1	-0.3	93.8	93.8	0.0	±0.7

Note : Absolute sensitivity was established by the use of the Sound Calibrator 01dB Type CAL31 S/N: 82795.

Certificate No.: CP20230294EA

Calibration Report

Function : 2. Self-generated Noise

2.1 Microphone Installed

Measured value (dB)
16.0

2.2 Microphone replaced by the electrical input signal device

Frequency Weighting	Measured value (dB)
A-weighting	8.8
C-weighting	9.9
Z-weighting	15.3

Function : 3. Acoustical signal tests of frequency weightings (With Windscreen)

Meter free-field acoustic response at a level of 84 dB.

Frequency (Hz)	Deviation from various Frequency Weighting Response Curve			
	C-Weighting (dB)	A-Weighting (dB)	Z-Weighting (dB)	Acceptance limits (dB)
125	0.1	0.0	0.1	±1.0
1000	0.0	0.0	0.0	±0.7
8000	-1.1	-1.0	-0.6	+1.5; -2.5

Function : 4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various Frequency Weighting Response Curve			
	C-Weighting (dB)	A-Weighting (dB)	Z-Weighting (dB)	Acceptance limits (dB)
63	0.0	-0.2	0.1	±1.0
125	0.1	-0.2	0.2	±1.0
250	0.1	0.1	0.2	±1.0
500	0.2	0.1	0.3	±1.0
1000	0.2	0.2	0.2	±0.7
2000	0.2	0.1	0.1	±1.0
4000	0.8	0.8	0.9	±1.0
8000	-0.1	-0.1	0.5	+1.5; -2.5
16000	-9.7	-9.7	-4.3	+2.5; -16.0

Certificate No.: CP20230294EA

Calibration Report

Function : 5. Frequency and time weighting at 1 kHz

5.1 Frequency weighting at 1 kHz

Frequency Weighting	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
C-weighting	94.0	0.0	±0.2
A-weighting	94.0	0.0	±0.2
Z-weighting	94.0	0.0	±0.2

5.2 Time weighting at 1 kHz

Time Weighting	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Fast	94.0	0.0	±0.1
Slow	94.0	0.0	±0.1
LAeq	94.0	0.0	±0.1

Function : 6. Long-Term Stability

Long-term stability over 30 minutes, with steady 1 kHz signal at reference level.

Time Period to Apply Signal (min)	Reference SPL (dB)	Record SPL at Conclusion of Time Period (dB)	Deviated value (dB)	Acceptance limits (dB)
30	94.0	94.0	0.0	±0.1

Function : 7. Level Linearity on the reference level range

7.1 Level Linearity on the reference level range, Upper

Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
94.0	94.0	0.0	±0.8
99.0	99.0	0.0	±0.8
104.0	104.0	0.0	±0.8
109.0	108.8	-0.2	±0.8
114.0	113.8	-0.2	±0.8
119.0	118.8	-0.2	±0.8
124.0	123.8	-0.2	±0.8
129.0	128.8	-0.2	±0.8

Certificate No.: CP20230294EA

Calibration Report

7.2 Level Linearity on the reference level range, Lower

Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
94.0	94.0	0.0	±0.8
89.0	89.0	0.0	±0.8
84.0	84.0	0.0	±0.8
79.0	79.0	0.0	±0.8
74.0	74.0	0.0	±0.8
69.0	69.0	0.0	±0.8
64.0	64.0	0.0	±0.8
59.0	59.0	0.0	±0.8
54.0	54.0	0.0	±0.8
49.0	49.0	0.0	±0.8
44.0	44.0	0.0	±0.8
39.0	39.0	0.0	±0.8
34.0	33.9	-0.1	±0.8
29.0	28.9	-0.1	±0.8
24.0	24.0	0.0	±0.8

Function : 8. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Fast	200	134.0	0.0	±0.5
	2	117.0	0.0	+1.0 ; -1.5
	0.25	107.8	-0.2	+1.0 ; -3.0
Slow	200	127.6	0.0	±0.5
	2	108.0	0.0	+1.0 ; -3.0
LAE	200	128.0	0.0	±0.5
	2	108.0	0.0	+1.0 ; -1.5
	0.25	98.9	-0.1	+1.0 ; -3.0

Function : 9. Peak C sound level

Number of cycles in test signal	Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Complete cycle	133.4	134.2	0.8	±2.0
Positive half cycle	132.4	131.5	-0.9	±1.0
Negative half cycle	132.4	131.6	-0.8	±1.0

Function : 10. Overload indication

Measured value (dB)		Deviated value (dB)	Acceptance limits (dB)
Positive one-half cycle	Negative one-half cycle		
139.3	139.9	0.6	±1.5

Certificate No.: CP20230294EA

Calibration Report

Function : 11. High-Level Stability

High-level stability over 5 minutes, with steady 1 kHz signal, 1 dB below upper boundary.

Time Period to Apply Signal (min)	Reference SPL (dB)	Record SPL at Conclusion of Time Period (dB)	Deviated value (dB)	Acceptance limits (dB)
5	137.0	137.0	0.0	±0.1

Uncertainty of measurement

Function	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1) Indication at the calibration check frequency	0.30	Not applicable
2) Self-generated Noise	0.10	Not applicable
3) Acoustical signal tests of frequency weightings - Free-field sound pressure response level	0.30	0.60 (10Hz to 4kHz) 0.70 (>4kHz to 10kHz)
4) Electrical signal tests of frequency weightings	0.20	0.20
5) Frequency and time weighting at 1 kHz	0.20	0.20
6) Long-Term Stability	0.10	0.10
7) Level Linearity on the reference level range	0.30	0.30
8) Tone burst response	0.20	0.30
9) Peak C sound level	0.20	0.35
10) Overload indication	0.20	0.25
11) High-Level Stability	0.10	0.10

Remarks: 1. The acceptance limit is for the deviated value.
2. Acceptance limits was IEC61672-3:2013 Class 1.
3. The coverage factor $k = 2.00$

-- End of Report --

Certificate No.: CP20230295EA

Operation No.: CP2023070032

Certificate of Calibration

Equipment: Sound Level Meter

Manufacturer: 01dB (Meter), G.R.A.S. (Microphone), 01dB (Preamplifier)

Model/Type: CUBE (Meter), 40CD (Microphone), PRE22 (Preamplifier)

Serial No.: 11113 (Meter), 260806 (Microphone), 1610398 (Preamplifier)

ID No.: UAE.EFM.023/2560 (Meter), UAE.EFM.023/2560 (Extension cable)

Customer: United Analyst and Engineering Consultant Co.,Ltd.

Address: 81 Soi Udomsuk 41, Sukhumvit Road, Bangchak
Phrakhanong, Bangkok 10260

Received Date: 24 July 2023

Calibrated Date: 4 - 7 August 2023

Issued Date: 8 August 2023

Calibrated by: Ms. Juntaporn Kunhakom

Approved by: _____

(Mr. Sittichai Swaksuriyawong)
Group Manager

This report was prepared electronically using applicable electronic signature. Printing or copy of file are considered as a copy of the document.

The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor (k) providing a level of confidence of approximately 95%. This certificate may not be reproduced other than in full except with the prior written approval of the Electrical and Electronics Institute, Foundation for Industrial Development.

Certificate No.: CP20230295EA

Calibration Report

Equipment: Sound Level Meter
Manufacturer: 01dB (Meter), G.R.A.S. (Microphone), 01dB (Preamplifier)
Model/Type: CUBE (Meter), 40CD (Microphone), PRE22 (Preamplifier)
Serial No.: 11113 (Meter), 260806 (Microphone), 1610398 (Preamplifier)
ID No.: UAE.EFM.023/2560 (Meter), UAE.EFM.023/2560 (Extension cable)
Ambient Temperature: (23 ± 2) °C
Relative Humidity: (50 ± 15) %
Pressure: (101.3 ± 1.5) kPa

Method of Calibration :-

IEC 61672-3:2013.

Condition of this result of calibration

1. Reference standards instrument :-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Standard microphone	4180	2787490	AA-1024-22	6 November 2023
2) Arbitrary Function Generator	AFG2021	C010063	CK20230040EA	26 June 2024
3) Programmable Attenuator	PA5	2755	EF-0034-22	30 October 2023
4) 6.5 Digit precision multimeter	8846A	9610014	CB20220223EA	14 November 2023
5) Pressure humidity and Temperature Transmitter	PTU301	F0640002	CL1-P230024	20 March 2024
			CD20230196EA	23 July 2024
6) Pressure humidity and Temperature Transmitter	PTU301	F0640003	CL1-P230032	4 April 2024
			CD20230197EA	23 July 2024
7) Performance Audio Analyzer	U8903B	MY56510003	CB20230038EA	14 February 2024
			CK20220080EA	8 September 2023

2. This result of calibration was found accurate as shown on date and place of calibration only.

3. This certification is traceable to the international system of unit maintained at :-

Reference standards instrument for Acoustic function

- National Institute of Metrology (Thailand)

Reference standards instrument for Electrical function

- National Institute of Metrology (Thailand)

- Electrical and Electronics Institute; NSC Accredited Calibration No.0119

Result of Calibration:-

Function : 1. Indication at the calibration check frequency

Reference Acoustic Signal (dB)	Correction for Microphone Model 40CD (dB)	Effective Calibration Level (dB)	Measured value (dB)	Deviation (dB)	Acceptance limits (dB)
94.1	-0.3	93.8	93.8	0.0	±0.7

Note : Absolute sensitivity was established by the use of the Sound Calibrator 01dB Type CAL31 S/N: 82795.

Certificate No.: CP20230295EA

Calibration Report

Function : 2. Self-generated Noise

2.1 Microphone Installed

Measured value (dB)
17.4

2.2 Microphone replaced by the electrical input signal device

Frequency Weighting	Measured value (dB)
A-weighting	12.2
C-weighting	12.2
Z-weighting	20.1

Function : 3. Acoustical signal tests of frequency weightings (With Windscreen)

Meter free-field acoustic response at a level of 84 dB.

Frequency (Hz)	Deviation from various Frequency Weighting Response Curve			
	C-Weighting (dB)	A-Weighting (dB)	Z-Weighting (dB)	Acceptance limits (dB)
125	0.1	0.0	0.1	±1.0
1000	0.0	0.0	0.0	±0.7
8000	-1.2	-1.2	-0.7	+1.5; -2.5

Function : 4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various Frequency Weighting Response Curve			
	C-Weighting (dB)	A-Weighting (dB)	Z-Weighting (dB)	Acceptance limits (dB)
63	0.0	-0.2	0.0	±1.0
125	0.1	-0.2	0.1	±1.0
250	0.1	0.1	0.2	±1.0
500	0.2	0.2	0.2	±1.0
1000	0.2	0.2	0.2	±0.7
2000	0.2	0.2	0.1	±1.0
4000	0.8	-0.1	0.9	±1.0
8000	-0.1	-0.1	0.4	+1.5; -2.5
16000	-9.7	-9.7	-4.3	+2.5; -16.0

Certificate No.: CP20230295EA

Calibration Report

Function : 5. Frequency and time weighting at 1 kHz

5.1 Frequency weighting at 1 kHz

Frequency Weighting	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
C-weighting	94.0	0.0	±0.2
A-weighting	94.0	0.0	±0.2
Z-weighting	94.0	0.0	±0.2

5.2 Time weighting at 1 kHz

Time Weighting	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Fast	94.0	0.0	±0.1
Slow	94.0	0.0	±0.1
LAeq	94.0	0.0	±0.1

Function : 6. Long-Term Stability

Long-term stability over 30 minutes, with steady 1 kHz signal at reference level.

Time Period to Apply Signal (min)	Reference SPL (dB)	Record SPL at Conclusion of Time Period (dB)	Deviated value (dB)	Acceptance limits (dB)
30	94.0	94.0	0.0	±0.1

Function : 7. Level Linearity on the reference level range

7.1 Level Linearity on the reference level range, Upper

Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
94.0	94.0	0.0	±0.8
99.0	99.0	0.0	±0.8
104.0	103.8	-0.2	±0.8
109.0	108.8	-0.2	±0.8
114.0	113.8	-0.2	±0.8
119.0	118.8	-0.2	±0.8
124.0	123.8	-0.2	±0.8
129.0	128.8	-0.2	±0.8

Certificate No.: CP20230295EA

Calibration Report

7.2 Level Linearity on the reference level range, Lower

Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
94.0	94.0	0.0	±0.8
89.0	89.0	0.0	±0.8
84.0	84.0	0.0	±0.8
79.0	79.0	0.0	±0.8
74.0	74.0	0.0	±0.8
69.0	69.0	0.0	±0.8
64.0	64.0	0.0	±0.8
59.0	59.0	0.0	±0.8
54.0	54.0	0.0	±0.8
49.0	49.0	0.0	±0.8
44.0	44.0	0.0	±0.8
39.0	39.0	0.0	±0.8
34.0	34.0	0.0	±0.8
29.0	29.0	0.0	±0.8
24.0	24.2	0.2	±0.8

Function : 8. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Fast	200	134.0	0.0	±0.5
	2	117.0	0.0	+1.0 ; -1.5
	0.25	107.8	-0.2	+1.0 ; -3.0
Slow	200	127.6	0.0	±0.5
	2	108.0	0.0	+1.0 ; -3.0
	0.25	128.0	0.0	±0.5
LAE	2	108.0	0.0	+1.0 ; -1.5
	0.25	98.9	-0.1	+1.0 ; -3.0

Function : 9. Peak C sound level

Number of cycles in test signal	Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Complete cycle	133.4	134.3	0.9	±2.0
Positive half cycle	132.4	131.5	-0.9	±1.0
Negative half cycle	132.4	131.7	-0.7	±1.0

Function : 10. Overload indication

Measured value (dB)		Deviated value (dB)	Acceptance limits (dB)
Positive one-half cycle	Negative one-half cycle		
139.8	140.3	0.5	±1.5

Certificate No.: CP20230295EA

Calibration Report

Function : 11. High-Level Stability

High-level stability over 5 minutes, with steady 1 kHz signal, 1 dB below upper boundary.

Time Period to Apply Signal (min)	Reference SPL (dB)	Record SPL at Conclusion of Time Period (dB)	Deviated value (dB)	Acceptance limits (dB)
5	137.0	137.0	0.0	±0.1

Uncertainty of measurement

Function	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1) Indication at the calibration check frequency	0.30	Not applicable
2) Self-generated Noise	0.10	Not applicable
3) Acoustical signal tests of frequency weightings - Free-field sound pressure response level	0.30	0.60 (10Hz to 4kHz) 0.70 (>4kHz to 10kHz)
4) Electrical signal tests of frequency weightings	0.20	0.20
5) Frequency and time weighting at 1 kHz	0.20	0.20
6) Long-Term Stability	0.10	0.10
7) Level Linearity on the reference level range	0.30	0.30
8) Tone burst response	0.20	0.30
9) Peak C sound level	0.20	0.35
10) Overload indication	0.20	0.25
11) High-Level Stability	0.10	0.10

Remarks: 1. The acceptance limit is for the deviated value.
2. Acceptance limits was IEC61672-3:2013 Class 1.
3. The coverage factor $k = 2.00$

-- End of Report --

Certificate No.: CP20230296EA

Operation No.: CP2023070033

Certificate of Calibration

Equipment: Sound Level Meter

Manufacturer: 01dB (Meter), G.R.A.S. (Microphone), 01dB (Preamplifier)

Model/Type: CUBE (Meter), 40CD (Microphone), PRE22 (Preamplifier)

Serial No.: 11135 (Meter), 161893 (Microphone), 1610285 (Preamplifier)

ID No.: UAE.EFM.024/2560 (Meter), UAE.EFM.024/2560 (Extension cable)

Customer: United Analyst and Engineering Consultant Co.,Ltd.

Address: 81 Soi Udomsuk 41, Sukhumvit Road, Bangchak
Phrakhanong, Bangkok 10260

Received Date: 24 July 2023

Calibrated Date: 5 - 7 August 2023

Issued Date: 8 August 2023

Calibrated by: Ms. Juntaporn Kunhakom

Approved by: 
(Mr. Sittichai Swaksuriyawong)
Group Manager

This report was prepared electronically using applicable electronic signature. Printing or copy of file are considered as a copy of the document.

The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor (k) providing a level of confidence of approximately 95%. This certificate may not be reproduced other than in full except with the prior written approval of the Electrical and Electronics Institute, Foundation for Industrial Development.

Certificate No.: CP20230296EA

Calibration Report

Equipment: Sound Level Meter
 Manufacturer: 01dB (Meter), G.R.A.S. (Microphone), 01dB (Preamplifier)
 Model/Type: CUBE (Meter), 40CD (Microphone), PRE22 (Preamplifier)
 Serial No.: 11135 (Meter), 161893 (Microphone), 1610285 (Preamplifier)
 ID No.: UAE.EFM.024/2560 (Meter), UAE.EFM.024/2560 (Extension cable)
 Ambient Temperature: (23 ± 2) °C
 Relative Humidity: (50 ± 15) %
 Pressure: (101.3 ± 1.5) kPa

Method of Calibration :-

IEC 61672-3:2013.

Condition of this result of calibration

1. Reference standards instrument :-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Standard microphone	4180	2787490	AA-1024-22	6 November 2023
2) Arbitrary Function Generator	AFG2021	C010063	CK20230040EA	26 June 2024
3) Programmable Attenuator	PA5	2755	EF-0034-22	30 October 2023
4) 6.5 Digit precision multimeter	8846A	9610014	CB20220223EA	14 November 2023
5) Pressure humidity and Temperature Transmitter	PTU301	F0640002	CL1-P230024 CD20230196EA	20 March 2024 23 July 2024
6) Pressure humidity and Temperature Transmitter	PTU301	F0640003	CL1-P230032 CD20230197EA	4 April 2024 23 July 2024
7) Performance Audio Analyzer	U8903B	MY56510003	CB20230038EA CK20220080EA	14 February 2024 8 September 2023

2. This result of calibration was found accurate as shown on date and place of calibration only.

3. This certification is traceable to the international system of unit maintained at :-

Reference standards instrument for Acoustic function

- National Institute of Metrology (Thailand)

Reference standards instrument for Electrical function

- National Institute of Metrology (Thailand)

- Electrical and Electronics Institute; NSC Accredited Calibration No.0119

Result of Calibration:-

Function : 1. Indication at the calibration check frequency

Reference Acoustic Signal (dB)	Correction for Microphone Model 40CD (dB)	Effective Calibration Level (dB)	Measured value (dB)	Deviation (dB)	Acceptance limits (dB)
94.1	-0.3	93.8	93.8	0.0	±0.7

Note : Absolute sensitivity was established by the use of the Sound Calibrator 01dB Type CAL31 S/N: 82795.

Certificate No.: CP20230296EA

Calibration Report

Function : 2. Self-generated Noise

2.1 Microphone Installed

Measured value (dB)
17.7

2.2 Microphone replaced by the electrical input signal device

Frequency Weighting	Measured value (dB)
A-weighting	14.1
C-weighting	13.3
Z-weighting	22.0

Function : 3. Acoustical signal tests of frequency weightings (With Windscreen)

Meter free-field acoustic response at a level of 84 dB.

Frequency (Hz)	Deviation from various Frequency Weighting Response Curve			
	C-Weighting (dB)	A-Weighting (dB)	Z-Weighting (dB)	Acceptance limits (dB)
125	0.2	0.1	0.1	±1.0
1000	0.1	0.1	0.1	±0.7
8000	-0.3	-0.3	0.2	+1.5; -2.5

Function : 4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various Frequency Weighting Response Curve			
	C-Weighting (dB)	A-Weighting (dB)	Z-Weighting (dB)	Acceptance limits (dB)
63	0.0	-0.1	0.1	±1.0
125	0.1	-0.1	0.1	±1.0
250	0.1	0.1	0.2	±1.0
500	0.2	0.2	0.3	±1.0
1000	0.2	0.2	0.2	±0.7
2000	0.2	0.2	0.2	±1.0
4000	0.9	0.8	0.9	±1.0
8000	0.0	0.0	0.5	+1.5; -2.5
16000	-9.7	-9.6	-4.3	+2.5; -16.0

Certificate No.: CP20230296EA

Calibration Report

Function : 5. Frequency and time weighting at 1 kHz

5.1 Frequency weighting at 1 kHz

Frequency Weighting	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
C-weighting	94.0	0.0	±0.2
A-weighting	94.0	0.0	±0.2
Z-weighting	94.0	0.0	±0.2

5.2 Time weighting at 1 kHz

Time Weighting	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Fast	94.0	0.0	±0.1
Slow	94.0	0.0	±0.1
LAeq	94.0	0.0	±0.1

Function : 6. Long-Term Stability

Long-term stability over 30 minutes, with steady 1 kHz signal at reference level.

Time Period to Apply Signal (min)	Reference SPL (dB)	Record SPL at Conclusion of Time Period (dB)	Deviated value (dB)	Acceptance limits (dB)
30	94.0	94.0	0.0	±0.1

Function : 7. Level Linearity on the reference level range

7.1 Level Linearity on the reference level range, Upper

Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
94.0	94.0	0.0	±0.8
99.0	99.0	0.0	±0.8
104.0	103.9	-0.1	±0.8
109.0	108.9	-0.1	±0.8
114.0	113.9	-0.1	±0.8
119.0	118.9	-0.1	±0.8
124.0	123.9	-0.1	±0.8
129.0	128.9	-0.1	±0.8
134.0	133.9	-0.1	±0.8

Certificate No.: CP20230296EA

Calibration Report

7.2 Level Linearity on the reference level range, Lower

Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
94.0	94.0	0.0	±0.8
89.0	89.0	0.0	±0.8
84.0	84.0	0.0	±0.8
79.0	79.0	0.0	±0.8
74.0	74.0	0.0	±0.8
69.0	69.0	0.0	±0.8
64.0	64.0	0.0	±0.8
59.0	59.0	0.0	±0.8
54.0	54.0	0.0	±0.8
49.0	49.0	0.0	±0.8
44.0	44.0	0.0	±0.8
39.0	39.0	0.0	±0.8
34.0	34.0	0.0	±0.8
29.0	29.0	0.0	±0.8
24.0	24.2	0.2	±0.8

Function : 8. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Fast	200	134.1	0.1	±0.5
	2	117.0	0.0	+1.0 ; -1.5
	0.25	107.8	-0.2	+1.0 ; -3.0
Slow	200	127.6	0.0	±0.5
	2	108.0	0.0	+1.0 ; -3.0
	0.25	98.9	-0.1	+1.0 ; -3.0

Function : 9. Peak C sound level

Number of cycles in test signal	Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Complete cycle	133.4	134.7	1.3	±2.0
Positive half cycle	132.4	131.6	-0.8	±1.0
Negative half cycle	132.4	131.6	-0.8	±1.0

Function : 10. Overload Indication

Measured value (dB)		Deviated value (dB)	Acceptance limits (dB)
Positive one-half cycle	Negative one-half cycle		
140.2	140.7	0.5	±1.5

Certificate No.: CP20230296EA

Calibration Report

Function : 11. High-Level Stability

High-level stability over 5 minutes, with steady 1 kHz signal, 1 dB below upper boundary.

Time Period to Apply Signal (min)	Reference SPL (dB)	Record SPL at Conclusion of Time Period (dB)	Deviated value (dB)	Acceptance limits (dB)
5	137.0	137.0	0.0	±0.1

Uncertainty of measurement

Function	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1) Indication at the calibration check frequency	0.30	Not applicable
2) Self-generated Noise	0.10	Not applicable
3) Acoustical signal tests of frequency weightings - Free-field sound pressure response level	0.30	0.60 (10Hz to 4kHz) 0.70 (>4kHz to 10kHz)
4) Electrical signal tests of frequency weightings	0.20	0.20
5) Frequency and time weighting at 1 kHz	0.20	0.20
6) Long-Term Stability	0.10	0.10
7) Level Linearity on the reference level range	0.30	0.30
8) Tone burst response	0.20	0.30
9) Peak C sound level	0.20	0.35
10) Overload indication	0.20	0.25
11) High-Level Stability	0.10	0.10

Remarks: 1. The acceptance limit is for the deviated value.
2. Acceptance limits was IEC61672-3:2013 Class 1.
3. The coverage factor $k = 2.00$

-- End of Report --

Certificate of Calibration

Customer

Name : UNITED ANALYST AND ENGINEERING

Certificate No : 23-ACT-116

CONSULTANT CO.,LTD.

Request No : Req-2023-1545

Address : 81 Soi Udomsuk 41, Sukhumvit Road, Bangchak,

Prakanong, Bangkok 10260

Unit Under Calibration Details

Measurement item : Acoustic Calibrator

Class : 1

Manufacturer : QUEST

Range : 94 , 114 dB / 250 , 1000 Hz

Model : QC-20

Instrument Status : Used

Serial Number : QOF110030

ID : UAE.EMA2.116/2555

Calibration Environment and Details

Temperature : (23 ±2 °C)

Humidity : (50 ± 20 %RH)

Barometric Pressure : (1013 ±10.0 hPa)

Received Date : 21 July 2023

Calibration Date : 4 August 2023

Location of Calibration : LAB 1 Acoustic


Calibration Procedure : In-house method CP-ACT-02 based on IEC 60942:2017 Electroacoustics - Sound calibrators


Reference Standard	Model	Serial Number	Traceable	Due Calibration
Sound Calibrator	SV 35A	58079	EI	31 May 2024
Sound Calibrator	AC-300	AC-300001087	EI	23 May 2024
THD Multimeter	2015	1047765	NIMT	31 January 2024

Traceability : This certificate provides traceability of measurement to recognized national standard, and to the realization of the international System of Units (SI).

Note

The reported uncertainty is based on standard uncertainty multiplied by the Coverage Factor $k=2$, providing a level of confidence approximately 95 %.

Calibrated By : 
Mr. Noppadon Luangart
Service Calibration Engineer

Approved By : 
Mr. Pacit Mathavorn
Calibration Engineer Supervisor

Issue Date : 4 August 2023

Sound pressure level

Calibration Results : Without Adjustment

Frequency of Sound pressure level

Total Harmonic Distortion plus Noise of Sound pressure level (THD+N %)

Note :

- Acceptance limit was IEC60942:2017 Class 1
- The calibration results exclude the calibrator pressure correction
- The calibration results exclude the microphone volume correction

The results related only to the item calibrated. The certificate shall not be reproduced except in full, without written approval of the Innovative Instrument Co., Ltd.

Customer

Name : UNITED ANALYST AND ENGINEERING

Certificate No : 23-ACT-111

Request No : Req-2023-1408

Address : 81 Soi Udomsuk 41, Sukhumvit Road, Bangchak.

Prakanong, Bangkok 10260

Unit Under Calibration Details

Measurement item : Acoustic Calibrator

Class : 1

Manufacturer : SVANTEK

Range : 94 , 114 dB / 1000 Hz

Model : SV 35A

Instrument Status : Used

Serial Number : 73249

ID : UAE.EFM.105/2561

Calibration Environment and Details

Temperature : (23 ±2 °C)

Humidity : (50 ± 20 %RH)

Barometric Pressure : (1013 \pm 10.0 hPa)

Received Date : 26 June 2023

Calibration Date : 27 June 2023

Location of Calibration : LAB 1 Acoustic

Calibration Procedure : In-house method CP-ACT-02 based on IEC 60942:2017 Electroacoustics - Sound calibrators

Traceability : This certificate provides traceability of measurement to recognized national standard, and to the realization of the international System of Units (SI).

Note

The reported uncertainty is based on standard uncertainty multiplied by the Coverage Factor $k=2$, providing a level of confidence approximately 95 %.

Calibrated By :

Mr. Noppadon Luangart

Service Calibration Engineer

Approved By :

Mr. Pacit Mathavorn

Calibration Engineer Supervisor

Issue Date : 27 June 2023

The results related only to the item calibrated. The certificate shall not be reproduced except in full, without written approval of the Institute of Metrology.

Certificate No : 23-ACT-111

Request No : Req-2023-1408

Sound pressure level

Calibration Results : Without Adjustment

Calibration Range (dB)	Without Adjustment (dB)		Adjustment (dB)		Uncertainty (± dB)	Acceptance limit Class 1 (± dB)
	Measured	Error	Measured	Error		
94 dB / 1000 Hz	93.84	-0.16	-	-	0.14	0.25
114 dB / 1000 Hz	113.79	-0.21	-	-	0.13	0.25

Frequency of Sound pressure level

Calibration Range (Hz)	Without Adjustment		Adjustment		Uncertainty (± %)	Acceptance limit Class 1 (± %)
	Measured (Hz)	Error (%)	Measured (Hz)	Error (%)		
94 dB / 1000 Hz	1000.00	0.00	-	-	0.01	0.70
114 dB / 1000 Hz	1000.00	0.00	-	-	0.01	0.70

Total Harmonic Distortion plus Noise of Sound pressure level (THD+N %)

Calibration Range (Hz)	Without Adjustment	Adjustment	Uncertainty (± %)	Acceptance limit Class 1 (± %)
	Measured (%)	Measured (%)		
94 dB / 1000 Hz	0.17	-	0.40	2.5
114 dB / 1000 Hz	0.04	-	0.40	2.5

Note :

- Acceptance limit was IEC60942:2017 Class 1
- The calibration results exclude the calibrator pressure correction
- The calibration results exclude the microphone volume correction

End of Calibration

Certificate of Calibration

Customer

Name UNITED ANALYST AND ENGINEERING CONSULTANT CO.,LTD.
Address 81 Soi Udomsuk 41, Sukhumvit Road, Bangchak, Prakanong, Bangkok 10260

Certificate No : 23-NDM-183
Request No : Req-2023-1488

Unit Under Calibration Details

Measurement item : Noise Dosimeter Microphone Class : 2
Manufacturer : SVANTEK Microphone Model : SV 27
Model : SV 104 Microphone S/N : 136867
Serial Number : 143229 Preamplifier Model : -
ID : - Preamplifier S/N : -
Resolution : 0.1 dB Instrument Status : New

Calibration Environment and Details


Temperature : 23 °C ± 2 °C
Humidity : 50 %RH ± 20 %RH
Barometric Pressure : 1013 hPa ± 10 hPa
Received Date : 12 July 2023
Calibrated Date : 7 August 2023
Calibration Procedure : In-house method CP-NDM-01 based on IEC 61252 : 2017
Location of Calibration : Lab Acoustic


Reference Standard

Instrument	Brand	Model	SN.	Due calibration	Traceability
Multifrequency Calibrator	Quest	Quest-cal	188272	25 July 2024	TSI
Standard Microphone	GRAS	40AN	188273	6 October 2023	GRAS
Sine Generator	Svantek	Svan401	131	12 October 2023	WK Electric
Timer	EXTECH	-	05-ACT	20 March 2024	TPA

Note

The reported uncertainty is based on standard uncertainty multiplied by the Coverage Factor $k = 2$, providing a level of confidence approximately 95 %.

Calibrated By : 
Mr. Noppadon Luangart
Calibration Officer

Approved By : 
Mr. Pacit Mathavorn
Calibration Engineer Supervisor
Issue Date : 7 August 2023

Certificate No : 23-NDM-183

Request No : Req-2023-1488

1. Absolute acoustical sensitivity

UUC Setting	Time		Exposure Measurement			UNCERTAINTY (%)	Tolerances Limit (%)
FAST / A / 55-140	Ref	UUC	Ref	UUC	Error		
Calibrator Setting	(s)	(s)	(Pa ² h)	(Pa ² h)	(%)		
1000 Hz 114 dB	120	120	3.18	3.13	-1.57	3.1	-21, +26

Note : Absolute sensitivity was established by the use of Sound Calibrator Brand SVANTEK, Model SV 35A, SN. 58079

2. Frequency weightings

UUC Setting	Deviation from various Frequency Weighting		UNCERTAINTY	Tolerances Limit
FAST / 55-140	A	C	(± dB)	(± dB)
STD Setting	(dB)	(dB)		
*63 Hz	0.1	0.2	0.40	2.0
125 Hz	-0.5	0.0	0.40	1.5
250 Hz	-0.6	-0.1	0.40	1.5
500 Hz	-0.3	0.1	0.40	1.5
1000 Hz	0.0	0.0	0.40	-
2000 Hz	0.1	0.5	0.40	2.0
4000 Hz	1.5	1.6	0.40	3.0
8000 Hz	-1.9	-2.0	0.40	5.0

Certificate No : 23-NDM-183

Request No : Req-2023-1488

3. Linearity of response to steady signals

a. Sound exposure meter, linearity of response for changes of input sinusoidal signal level

UUC Setting	FAST / A / High											
1000 Hz	Ref	(dB)	55.0	80.0	90.0	100.0	110.0	114.0	120.0	130.0	140.0	
	Level A	(dB)	54.9	80.1	90.1	100.0	110.0	114.0	120.0	130.0	140.0	
	Error	(dB)	-0.1	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.0	
8000 Hz	Ref	(dB)			88.9	98.9	108.9	112.9	118.9	128.9	138.9	
	Level A	(dB)			89.0	99.0	108.9	112.9	118.9	128.9	138.8	
	Error	(dB)			0.1	0.1	0.0	0.0	0.0	0.0	-0.1	
63 Hz	Ref	(dB)						87.8	93.8	103.8	113.8	
	Level A	(dB)						87.8	93.8	103.8	113.8	
	Error	(dB)						0.0	0.0	0.0	0.0	
Tolerances Limit		(±dB)	1.0									
UNCERTAINTY		(±dB)	0.3									

b. Sound exposure meter linearity of error

UUC Setting	Time		Exposure Measurement			UNCERTAINTY	Tolerance Limit
FAST / A / 55-140	Ref	UUC	Ref	UUC	Error		
Calibrator Setting	(s)	(s)	(Pa ² h)	(Pa ² h)	(%)	(%)	(%)
1000 Hz 110 dB	27	27	0.30	0.30	0.00	5.6	-21, +26
1000 Hz 110 dB	45	45	0.50	0.50	0.00		
1000 Hz 110 dB	90	90	1.00	0.99	-1.00		
1000 Hz 110 dB	180	180	2.00	1.99	-0.50		
1000 Hz 120 dB	36	36	4.00	4.03	+0.75		
1000 Hz 120 dB	72	72	8.00	8.05	+0.63	5.6	
1000 Hz 120 dB	90	90	10.00	9.90	-1.00		
1000 Hz 120 dB	180	180	20.00	20.22	+1.10		
1000 Hz 120 dB	360	360	40.00	40.34	+0.85		
1000 Hz 120 dB	720	720	80.00	80.49	+0.61		

Certificate No : 23-NDM-183
Request No : Req-2023-1488

4. Response to short duration
a. Response for sinusoidal signals - reference level

UUC Setting	Time		Exposure Measurement			UNCERTAINTY	Tolerances Limit
FAST / A / 55-140	Ref	UUC	Ref	UUC	Error		
Calibrator Setting	(s)	(s)	(Pa ² h)	(Pa ² h)	(Pa ² h)	(Pa ² h)	(Pa ² h)
4000 Hz 95 dB	2846	2846	1.00	0.98	-0.02	0.052	-0.29 ~ +0.41

b. Sound exposure meter response for series of toneburst impulses

UUC Setting	Time		Exposure Measurement			UNCERTAINTY	Tolerances Limit
FAST / A / 55-140	Ref	UUC	Ref	UUC	Error		
Calibrator Setting	(s)	(s)	(Pa ² h)	(Pa ² h)	(%)	(%)	(%)
Burst 1 ms, 95 dB	2846	2846	1.00	0.98	-2.00	5.6	-21 ~ +26
Burst 1 ms, 100 dB	900	900	1.00	0.98	-2.00		-29 ~ +41
Burst 1 ms, 108 dB	143	143	1.00	0.99	-1.00		-29 ~ +41

5. Response to unipolar pulse

UUC Setting	Time	Exposure Measurement		UNCERTAINTY	Tolerances Limit
FAST / A / 55-140	UUC	UUC	Different		
Calibrator Setting	(s)	(Pa ² h)	(%)	(%)	(%)
Continuous Rectangle +	29	10.13	0.00	3.7	-21 ~ +26
Continuous Rectangle -		10.13			

* Indicates non accredited

End of Certificate

Certificate of Calibration

Customer

Name : UNITED ANALYST AND ENGINEERING CONSULTANT CO.,LTD. Certificate No : 24-NDM-018
Address : 81 Soi Udumsuk 41, Sukhumvit Road, Bangehak, Prakanong, Bangkok 10260 Request No : Req-2023-2689

Unit Under Calibration Details

Measurement item : Noise Dosimeter Microphone Class : 2
Manufacturer : SVANTEK Microphone Model : SV 271S
Model : SV 1041S Microphone S/N : 106312
Serial Number : 106069 Preamplifier Model : -
ID : - Preamplifier S/N : -
Resolution : 0.1 dB Instrument Status : Used

Calibration Environment and Details


Temperature : 23 °C ± 2 °C
Humidity : 50 %RH ± 20 %RH
Barometric Pressure : 1013 hPa ± 10 hPa
Received Date : 21 December 2023
Calibrated Date : 25 January 2024
Calibration Procedure : In-house method CP-NDM-01 based on IEC 61252 : 2017
Location of Calibration : Lab Acoustic


Reference Standard

Instrument	Brand	Model	SN.	Due calibration	Traceability
Multifrequency Calibrator	Quest	Quest-cal	EFA000234	25 July 2024	TSI
Standard Microphone	GRAS	40AN	188273	21 August 2024	GRAS
Sine Generator	SvanteK	Svan401	131	9 October 2024	WK Electric
Timer	EXTECH	-	05-ACT	21 March 2024	TPA

Note

The reported uncertainty is based on standard uncertainty multiplied by the Coverage Factor $k = 2$, providing a level of confidence approximately 95 %.

Calibrated By : 
Mr. Noppadon Luangart
Service Calibration Engineer

Approved By : 
Mr. Pacit Mathavorn
Calibration Engineer Supervisor
Issue Date : 25 January 2024



Certificate No : 24-NDM-018
Request No : Req-2023-2689

1. Absolute acoustical sensitivity

UUC Setting	Time		Exposure Measurement			UNCERTAINTY	Tolerances Limit
FAST / A / 60-140	Ref	UUC	Ref	UUC	Error		
Calibrator Setting	(s)	(s)	(Pa ² h)	(Pa ² h)	(%)	(%)	(%)
1000 Hz 114 dB	120	120	3.18	3.13	-1.6	3.1	-21, +26

Note : Absolute sensitivity was established by the use of Sound Calibrator Brand SVANTEK, Model SV 35A, SN. 58079

2. Frequency weightings

UUC Setting	Deviation from various Frequency Weighting		UNCERTAINTY	Tolerances Limit
FAST / 60-140	A	C		
STD Setting	(dB)	(dB)	(± dB)	(± dB)
*63 Hz	0.1	0.1	0.40	2.0
125 Hz	0.3	0.4	0.40	1.5
250 Hz	0.0	0.1	0.40	1.5
500 Hz	0.0	0.0	0.40	1.5
1000 Hz	0.0	0.0	0.40	-
2000 Hz	0.5	0.5	0.40	2.0
4000 Hz	2.4	2.4	0.40	3.0
8000 Hz	-2.8	-2.8	0.40	5.0



Certificate No : 24-NDM-018
Request No : Req-2023-2689

3. Linearity of response to steady signals

a. Sound exposure meter, linearity of response for changes of input sinusoidal signal level

UUC Setting	FAST / A / High											
1000 Hz	Ref	(dB)	60.0	80.0	90.0	100.0	110.0	114.0	120.0	130.0	140.0	
	Level A	(dB)	60.0	80.1	90.1	100.0	110.0	114.0	120.0	129.9	139.9	
	Error	(dB)	0.0	0.1	0.1	0.0	0.0	0.0	0.0	-0.1	-0.1	
8000 Hz	Ref	(dB)			88.9	98.9	108.9	112.9	118.9	128.9	138.9	
	Level A	(dB)			89.0	98.9	108.9	112.9	118.9	128.9	138.8	
	Error	(dB)				0.1	0.0	0.0	0.0	0.0	-0.1	
63 Hz	Ref	(dB)						87.8	93.8	103.8	113.8	
	Level A	(dB)						87.8	93.8	103.8	113.8	
	Error	(dB)						0.0	0.0	0.0	0.0	
Tolerances Limit		(±dB)	1.0									
UNCERTAINTY		(±dB)	0.3									

b. Sound exposure meter linearity of error

UUC Setting	Time		Exposure Measurement			UNCERTAINTY	Tolerances
FAST / A / 60-140	Ref	UUC	Ref	UUC	Error		Limit
Calibrator Setting	(s)	(s)	(Pa ² h)	(Pa ² h)	(%)	(%)	(%)
1000 Hz 110 dB	27	27	0.30	0.30	0.00	5.6	-21, +26
1000 Hz 110 dB	45	45	0.50	0.50	0.00		
1000 Hz 110 dB	90	90	1.00	0.99	-1.00		
1000 Hz 110 dB	180	180	2.00	1.98	-1.00		
1000 Hz 120 dB	36	36	4.00	4.03	+0.75		
1000 Hz 120 dB	72	72	8.00	8.05	+0.63	5.6	
1000 Hz 120 dB	90	90	10.00	10.13	+1.30		
1000 Hz 120 dB	180	180	20.00	20.22	+1.10		
1000 Hz 120 dB	360	360	40.00	40.34	+0.85		
1000 Hz 120 dB	720	720	80.00	80.49	+0.61		

Certificate No : 24-NDM-018
Request No : Req-2023-2689

4. Response to short duration
a. Response for sinusoidal signals - reference level

UUC Setting	Time		Exposure Measurement			UNCERTAINTY	Tolerances
FAST / A / 60-140	Ref	UUC	Ref	UUC	Error	(Pa ² h)	Limit
Calibrator Setting	(s)	(s)	(Pa ² h)	(Pa ² h)	(Pa ² h)		(Pa ² h)
4000 Hz 95 dB	2846	2846	1.00	0.98	-0.02		-0.29 ~ +0.41

b. Sound exposure meter response for series of toneburst impulses

UUC Setting	Time		Exposure Measurement			UNCERTAINTY	Tolerances
FAST / A / 60-140	Ref	UUC	Ref	UUC	Error	(%)	Limit
Calibrator Setting	(s)	(s)	(Pa ² h)	(Pa ² h)	(%)		(%)
Burst 1 ms, 95 dB	2846	2846	1.00	0.98	-2.00		-21 ~ +26
Burst 1 ms, 100 dB	900	900	1.00	0.98	-2.00	5.6	-29 ~ +41
Burst 1 ms, 108 dB	143	143	1.00	0.99	-1.00		-29 ~ +41

5. Response to unipolar pulse

UUC Setting	Time	Exposure Measurement		UNCERTAINTY	Tolerances
FAST / A / 60-140	UUC	UUC	Different		Limit
Calibrator Setting	(s)	(Pa ² h)	(%)	(%)	(%)
Continuous Rectangle +	28	10.13	0.00	3.7	-21 ~ +26
Continuous Rectangle -		10.13			

* Indicates non accredited

End of Certificate

Certificate of Calibration

Customer
Name : UNITED ANALYST AND ENGINEERING CONSULTANT CO.,LTD. Certificate No : 24-NDM-073
Address : 81 Soi Udomsuk 41, Sukhumvit Road, Bangchak, Prakanong, Bangkok 10260 Request No : Req-2024-0526

Unit Under Calibration Details

Measurement item : Noise Dosimeter Microphone Class : 2
Manufacturer : SVANTEK Microphone Model : SV 27IS
Model : SV 104IS Microphone S/N : 133718
Serial Number : 128372 Preamplifier Model : -
ID : - Preamplifier S/N : -
Resolution : 0.1 dB Instrument Status : Used

Calibration Environment and Details


Temperature : 23 °C ± 2 °C
Humidity : 50 %RH ± 20 %RH
Barometric Pressure : 1013 hPa ± 10 hPa
Received Date : 5 March 2024
Calibrated Date : 21 March 2024
Calibration Procedure : In-house method CP-NDM-01 based on IEC 61252 : 2017
Location of Calibration : Lab Acoustic


Reference Standard

Instrument	Brand	Model	SN.	Due calibration	Traceability
Multifrequency Calibrator	Quest	Quest-cal	EFA000234	25 July 2024	TSI
Standard Microphone	GRAS	40AN	188273	21 August 2024	GRAS
Sine Generator	Svantek	Svan401	131	9 October 2024	WK Electric
Timer	EXTECH	-	05-ACT	21 March 2024	TPA

Note

The reported uncertainty is based on standard uncertainty multiplied by the Coverage Factor $k = 2$, providing a level of confidence approximately 95 %.

Calibrated By : 
Mr. Noppadon Luangart
Service Calibration Engineer

Approved By : 
Mr. Pacit Mathavorn
Calibration Engineer Supervisor
Issue Date : 21 March 2024

Certificate No : 24-NDM-073
Request No : Req-2024-0526

1. Absolute acoustical sensitivity

UUC Setting	Time		Exposure Measurement			UNCERTAINTY	Tolerances Limit
	Ref	UUC	Ref	UUC	Error		
Calibrator Setting	(s)	(s)	(Pa ² h)	(Pa ² h)	(%)	(%)	(%)
1000 Hz 114 dB	120	120	3.18	3.13	-1.6	3.1	-21, +26

Note : Absolute sensitivity was established by the use of Sound Calibrator Brand SVANTEK, Model SV 35A, SN. 58079

2. Frequency weightings

UUC Setting	Deviation from various Frequency Weighting		UNCERTAINTY	Tolerances Limit
	A	C		
FAST / 60-140	A	C	(± dB)	(± dB)
STD Setting	(dB)	(dB)		
*63 Hz	0.2	0.5	0.40	2.0
125 Hz	0.5	0.7	0.40	1.5
250 Hz	0.3	0.4	0.40	1.5
500 Hz	0.2	0.3	0.40	1.5
1000 Hz	0.0	0.0	0.40	-
2000 Hz	-0.8	-0.1	0.40	2.0
4000 Hz	1.1	1.0	0.40	3.0
8000 Hz	-1.7	-1.7	0.40	5.0

Certificate No : 24-NDM-073
Request No : Req-2024-0526

3. Linearity of response to steady signals

a. Sound exposure meter, linearity of response for changes of input sinusoidal signal level

UUC Setting	FAST / A / High										
	Ref	(dB)	60.0	80.0	90.0	100.0	110.0	114.0	120.0	130.0	140.0
1000 Hz	Level A	(dB)	59.8	80.1	90.1	100.1	110.0	114.0	120.0	130.0	140.0
	Error	(dB)	-0.2	0.1	0.1	0.1	0.0	0.0	0.0	0.0	0.0
	Ref	(dB)			88.9	98.9	108.9	112.9	118.9	128.9	138.9
8000 Hz	Level A	(dB)			89.0	98.9	108.9	112.9	118.9	128.9	138.8
	Error	(dB)			0.1	0.0	0.0	0.0	0.0	0.0	-0.1
	Ref	(dB)						87.8	93.8	103.8	113.8
63 Hz	Level A	(dB)						87.8	93.8	103.8	113.8
	Error	(dB)						0.0	0.0	0.0	0.0
Tolerances Limit		(±dB)	1.0								
UNCERTAINTY		(±dB)	0.3								

b. Sound exposure meter linearity of error

UUC Setting	Time		Exposure Measurement			UNCERTAINTY	Tolerances
FAST / A / 60-140	Ref	UUC	Ref	UUC	Error		Limit
Calibrator Setting	(s)	(s)	(Pa ² h)	(Pa ² h)	(%)	(%)	(%)
1000 Hz 110 dB	27	27	0.30	0.30	0.00	5.6	-21, +26
1000 Hz 110 dB	45	45	0.50	0.50	0.00		
1000 Hz 110 dB	90	90	1.00	1.01	+1.00		
1000 Hz 110 dB	180	180	2.00	2.02	+1.00		
1000 Hz 120 dB	36	36	4.00	4.03	+0.75		
1000 Hz 120 dB	72	72	8.00	8.05	+0.63	5.6	
1000 Hz 120 dB	90	90	10.00	9.90	-1.00		
1000 Hz 120 dB	180	180	20.00	19.76	-1.20		
1000 Hz 120 dB	360	360	40.00	39.42	-1.45		
1000 Hz 120 dB	720	720	80.00	80.49	+0.61		

Certificate No : 24-NDM-073
Request No : Req-2024-0526

4. Response to short duration

a. Response for sinusoidal signals - reference level

UUC Setting	Time		Exposure Measurement			UNCERTAINTY	Tolerances
FAST / A / 60-140	Ref	UUC	Ref	UUC	Error		Limit
Calibrator Setting	(s)	(s)	(Pa ² h)	(Pa ² h)	(Pa ² h)	(Pa ² h)	(Pa ² h)
4000 Hz 95 dB	2846	2846	1.00	1.00	0.00	0.052	-0.29 - +0.41

b. Sound exposure meter response for series of toneburst impulses

UUC Setting	Time		Exposure Measurement			UNCERTAINTY	Tolerances
FAST / A / 60-140	Ref	UUC	Ref	UUC	Error		Limit
Calibrator Setting	(s)	(s)	(Pa ² h)	(Pa ² h)	(%)	(%)	(%)
Burst 1 ms, 95 dB	2846	2846	1.00	1.00	0.00	5.6	-21 - +26
Burst 1 ms, 100 dB	900	900	1.00	1.00	0.00		-29 - +41
Burst 1 ms, 108 dB	143	143	1.00	1.01	+1.00		-29 - +41

5. Response to unipolar pulse

UUC Setting	Time	Exposure Measurement		UNCERTAINTY	Tolerances
FAST / A / 60-140	UUC	UUC	Different		Limit
Calibrator Setting	(s)	(Pa ² h)	(%)	(%)	(%)
Continuous Rectangle +	29	10.37	0.00	3.7	-21 - +26
Continuous Rectangle -		10.37			

* Indicates non accredited

End of Certificate



ELECTRICAL AND ELECTRONICS INSTITUTE
FOUNDATION FOR INDUSTRIAL DEVELOPMENT
975 Moo 4, Bangpoo Industrial Estate, Soi 8, Sukhumvit Road km 37,
Phraek Sa, Mueang Samut Prakan, Samut Prakan 10280
Tel: +66 2709 4860 Fax: +66 2324 0917



Certificate No.: CP20230291EA
Operation No.: CP2023070045

Certificate of Calibration

Equipment: Sound Level Meter
Manufacturer: RION
Model/Type: NL-62 (Meter), UC-59L (Microphone), NH-26 (Preamplifier)
Serial No.: 00130357 (Meter), 02373 (Microphone), 00391 (Preamplifier)
ID No.: UAE,EMA2.104/2556
Customer: United Analyst and Engineering Consultant Co.,Ltd.
Address: 81 Soi Udomsuk 41, Sukhumvit Road, Bangchak
Phrakhanong, Bangkok 10260
Received Date: 24 July 2023
Calibrated Date: 3 - 4 August 2023
Issued Date: 7 August 2023
Calibrated by: Ms. Juntaporn Kunhakom

Approved by: 
(Mr. Sittichai Swaksuriyawong)
Group Manager

This report was prepared electronically using applicable electronic signature. Printing or copy of file are considered as a copy of the document.

The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor (k) providing a level of confidence of approximately 95%. This certificate may not be reproduced other than in full except with the prior written approval of the Electrical and Electronics Institute, Foundation for Industrial Development.

Certificate No.: CP20230291EA

Calibration Report

Equipment: Sound Level Meter
Manufacturer: RION
Model/Type: NL-62 (Meter), UC-59L (Microphone), NH-26 (Preamplifier)
Serial No.: 00130357 (Meter), 02373 (Microphone), 00391 (Preamplifier)
ID No.: UAE.EMA2.104/2556
Ambient Temperature: (23 ± 2) °C
Relative Humidity: (50 ± 15) %
Pressure: (101.3 ± 1.5) kPa

Method of Calibration :-

IEC 61672-3:2013.

Condition of this result of calibration

1. Reference standards instrument :-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Standard microphone	4180	2787490	AA-1024-22	6 November 2023
2) Arbitrary Function Generator	AFG2021	C010063	CK20230040EA	26 June 2024
3) Programmable Attenuator	PA5	2755	EF-0034-22	30 October 2023
4) 6.5 Digit precision multimeter	8846A	9610014	CB20220223EA	14 November 2023
5) Pressure humidity and Temperature Transmitter	PTU301	F0640002	CL1-P230024 CD20230196EA	20 March 2024 23 July 2024
6) Pressure humidity and Temperature Transmitter	PTU301	F0640003	CL1-P230032 CD20230197EA	4 April 2024 23 July 2024
7) Performance Audio Analyzer	U8903B	MY56510003	CB20230038EA CK20220080EA	14 February 2024 8 September 2023

2. This result of calibration was found accurate as shown on date and place of calibration only.

3. This certification is traceable to the international system of unit maintained at :-

Reference standards instrument for Acoustic function

- National Institute of Metrology (Thailand)

Reference standards instrument for Electrical function

- National Institute of Metrology (Thailand)

- Electrical and Electronics Institute; NSC Accredited Calibration No.0119

Result of Calibration:-

Function : 1. Indication at the calibration check frequency

Reference Acoustic Signal (dB)	Measured value (dB)	Deviation (dB)	Acceptance limits (dB)
94.0	94.0	0.0	±0.7

Note : Absolute sensitivity was established by the use of the Sound Calibrator RION Type NC-74 S/N : 34615278.

Certificate No.: CP20230291EA

Calibration Report

Function : 2. Self-generated Noise

2.1 Microphone Installed

Measured value (dB)
16.3

2.2 Microphone replaced by the electrical input signal device

Frequency Weighting	Measured value (dB)
A-weighting	12.1
C-weighting	18.5
Z-weighting	27.5

Function : 3. Acoustical signal tests of frequency weightings (Without Windscreen)

Meter free-field acoustic response at a level of 84 dB.

Frequency (Hz)	Deviation from various Frequency Weighting Response Curve			
	C-Weighting (dB)	A-Weighting (dB)	Z-Weighting (dB)	Acceptance limits (dB)
125	0.1	0.0	0.2	±1.0
1000	0.0	0.0	0.0	±0.7
8000	0.0	0.0	0.0	+1.5; -2.5

Function : 4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various Frequency Weighting Response Curve			
	C-Weighting (dB)	A-Weighting (dB)	Z-Weighting (dB)	Acceptance limits (dB)
63	0.0	0.0	0.1	±1.0
125	0.1	-0.1	0.0	±1.0
250	0.0	0.0	0.1	±1.0
500	0.0	0.0	0.1	±1.0
1000	0.0	0.0	0.0	±0.7
2000	0.1	0.0	0.0	±1.0
4000	0.0	0.0	0.0	±1.0
8000	0.1	0.1	0.0	+1.5; -2.5
16000	-1.3	-1.3	0.1	+2.5; -16.0

Certificate No.: CP20230291EA

Calibration Report

Function : 5. Frequency and time weighting at 1 kHz

5.1 Frequency weighting at 1 kHz

Frequency Weighting	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
C-weighting	94.0	0.0	±0.2
A-weighting	94.0	0.0	±0.2
Z-weighting	94.0	0.0	±0.2

5.2 Time weighting at 1 kHz

Time Weighting	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Fast	94.0	0.0	±0.1
Slow	94.0	0.0	±0.1
LAeq	94.0	0.0	±0.1

Function : 6. Long-Term Stability

Long-term stability over 30 minutes, with steady 1 kHz signal at reference level.

Time Period to Apply Signal (min)	Reference SPL (dB)	Record SPL at Conclusion of Time Period (dB)	Deviated value (dB)	Acceptance limits (dB)
30	94.0	94.0	0.0	±0.1

Function : 7. Level Linearity on the reference level range

7.1 Level Linearity on the reference level range, Upper

Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
94.0	94.0	0.0	±0.8
99.0	99.0	0.0	±0.8
104.0	104.0	0.0	±0.8
109.0	109.0	0.0	±0.8
114.0	114.0	0.0	±0.8
119.0	119.0	0.0	±0.8
124.0	124.0	0.0	±0.8
129.0	129.0	0.0	±0.8
130.0	130.0	0.0	±0.8
131.0	131.0	0.0	±0.8
132.0	132.0	0.0	±0.8
133.0	133.0	0.0	±0.8
134.0	134.0	0.0	±0.8
135.0	135.0	0.0	±0.8
136.0	136.0	0.0	±0.8
137.0	137.0	0.0	±0.8

Certificate No.: CP20230291EA

Calibration Report

7.2 Level Linearity on the reference level range, Lower

Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
94.0	94.0	0.0	±0.8
89.0	89.0	0.0	±0.8
84.0	84.0	0.0	±0.8
79.0	79.0	0.0	±0.8
74.0	74.0	0.0	±0.8
69.0	69.0	0.0	±0.8
64.0	64.0	0.0	±0.8
59.0	59.0	0.0	±0.8
54.0	54.0	0.0	±0.8
49.0	49.0	0.0	±0.8
44.0	44.0	0.0	±0.8
39.0	39.0	0.0	±0.8
34.0	34.0	0.0	±0.8
29.0	28.9	-0.1	±0.8

Function : 8. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Fast	200	126.0	0.0	±0.5
	2	108.9	-0.1	+1.0 ; -1.5
	0.25	99.9	-0.1	+1.0 ; -3.0
Slow	200	119.6	0.0	±0.5
	2	100.0	0.0	+1.0 ; -3.0
	200	120.0	0.0	±0.5
LAE	2	100.0	0.0	+1.0 ; -1.5
	0.25	90.8	-0.2	+1.0 ; -3.0

Function : 9. Peak C sound level

Number of cycles in test signal	Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Complete cycle	125.4	125.2	-0.2	±2.0
Positive half cycle	124.4	124.1	-0.3	±1.0
Negative half cycle	124.4	124.1	-0.3	±1.0

Certificate No.: CP20230291EA

Calibration Report

Function : 10. Overload indication

Measured value (dB)		Deviated value (dB)	Acceptance limits (dB)
Positive one-half cycle	Negative one-half cycle		
139.5	139.5	0.0	±1.5

Function : 11. High-Level Stability

High-level stability over 5 minutes, with steady 1 kHz signal, 1 dB below upper boundary.

Time Period to Apply Signal (min)	Reference SPL (dB)	Record SPL at Conclusion of Time Period (dB)	Deviated value (dB)	Acceptance limits (dB)
5	129.0	129.0	0.0	±0.1

Uncertainty of measurement

Function	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1) Indication at the calibration check frequency	0.30	Not applicable
2) Self-generated Noise	0.10	Not applicable
3) Acoustical signal tests of frequency weightings - Free-field sound pressure response level	0.30	0.60 (10Hz to 4kHz) 0.70 (>4kHz to 10kHz)
4) Electrical signal tests of frequency weightings	0.20	0.20
5) Frequency and time weighting at 1 kHz	0.20	0.20
6) Long-Term Stability	0.10	0.10
7) Level Linearity on the reference level range	0.30	0.30
8) Tone burst response	0.20	0.30
9) Peak C sound level	0.20	0.35
10) Overload indication	0.20	0.25
11) High-Level Stability	0.10	0.10

Remarks: 1. The acceptance limit is for the deviated value.
2. Acceptance limits was IEC61672-3:2013 Class 1.
3. The coverage factor $k = 2.00$

-- End of Report --

Certificate of Calibration

Customer

Name : UNITED ANALYST AND ENGINEERING CONSULTANT CO.,LTD.
Address : 81 Soi Udomsuk 41, Sukhumvit Road, Bangchak, Prakanong, Bangkok 10260

Certificate No : 23-ASP-073

Request No : Req-2023-0422

Unit Under Calibration Details

Measurement Item : Air Sampling Pump
Manufacturer : SENSIDYNE
Model : GilAir 5
Serial Number : 20120301021
ID : *

Calibration Environment and Details

Temperature : 23 ± 3 °C
Humidity : 55 ± 15 %RH
Barometric : 1010 ± 10 hpa
Received Date : 16 February 2023
Calibration Date : 9 March 2023
Calibration By : Mr. Sittichok Jirapukdeesakun
Location of Calibration : LAB 4 Air Velocity
Calibration Procedure : In-house method CP-ASP-01 based on ISO 13137:2013

Reference Standard	Model	Serial Number	Traceble	Due Date
Air flow meter	Gilibrator 3 Standard flow	19031011003	Sensidyne	15 June 2023
Digital Vacuum Meter	Digi Mano	29508	PCAL	21 September 2023

Traceability : This certificate provides traceability of measurement to recognized national standard, and to the realization of the international System of Units (SI).

Note : The reported uncertainty is based on standard uncertainty multiplied by the Coverage Factor $k=2$, providing a level of confidence approximately 95 %.

Calibration By : 
Service Calibration Engineer

Approved By : 
Mr. Pacit Mathavorn

Calibration Engineer Supervisor

Issue Date : 9 March 2023

Certificate No : 23-ASP-073

Request No : Req-2023-0422

Constant Flow

Result of Calibration :

UUC Flow Setting	STD FLOW READING (LPM) at							Flow Acceptable Tolerance (LPM)	Uncertainty (LPM)	Evaluation (Pass / Fail)
(LPM)	BP 5±1 inH ₂ O	BP 10±1 inH ₂ O	BP 15±1 inH ₂ O	BP 20±1 inH ₂ O	BP 25±1 inH ₂ O	BP 30±1 inH ₂ O	BP 40±1 inH ₂ O			
HIGH										
1.0	1.034	1.045	1.042	1.038	1.040	1.044	-	0.950 - 1.049	0.0078	Pass
1.7	1.737	1.768	1.774	1.774	1.776	1.773	-	1.615 - 1.785	0.015	Pass
2.0	1.960	1.976	1.982	1.998	2.012	1.994	-	1.900 - 2.100	0.025	Pass

Note : - Reference Specifications ± 5% of set flow or ±3 cc/min whichever is higher

* Indicates non accredited

End of Certificate

Certificate of Calibration

Customer

Name : UNITED ANALYST AND ENGINEERING CONSULTANT CO.,LTD.

Address : 81 Soi Udomsuk 41, Sukhumvit Road, Bangchak, Prakanong, Bangkok 10260

Certificate No : 23-ASP-075

Request No : Req-2023-0422

Unit Under Calibration Details

Measurement Item : Air Sampling Pump

Manufacturer : SENSIDYNE

Model : GilAir 5

Serial Number : 20170701010

ID : -

Calibration Environment and Details

Temperature : 23 ± 3 °C

Humidity : 55 ± 15 %RH

Barometric : 1010 ± 10 hpa

Received Date : 16 February 2023

Calibration Date : 9 March 2023

Calibration By : Mr. Sittichok Jirapukdeesakun

Location of Calibration : LAB 4 Air Velocity

Calibration Procedure : In-house method CP-ASP-01 based on ISO 13137:2013

Reference Standard	Model	Serial Number	Traceble	Due Date
Air flow meter	Gilibrator 3 Standard flow	19031011003	Sensidyne	15 June 2023
Digital Vacuum Meter	Digi Mano	29508	PCAL	21 September 2023

Traceability : This certificate provides traceability of measurement to recognized national standard, and to the realization of the international System of Units (SI).

Note : The reported uncertainty is based on standard uncertainty multiplied by the Coverage Factor k=2, providing a level of confidence approximately 95 %.

Calibration By : 
Service Calibration Engineer

Approved By : 
Mr. Pacit Mathavorn

Calibration Engineer Supervisor

Issue Date : 9 March 2023


Request No : Req-2023-0422

Result of Calibration :

Note : - Reference Specifications $\pm 5\%$ of set flow or ± 3 cc/min whichever is higher

* Indicates non accredited

End of Certificate



This certificate may not be reproduced other than in full, except with the prior written Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.



Cert.No.: 23CH1488

Page.: 2 of 3

Condition of this result of calibration

1. Reference Standard Instrument :-

Instrument	Serial No.	ID No.	Certificate No.	Due date
1) Thermometer	1963878	130RC095	2311051	05 Sep 2024
2) Ref. Std. Thermometer	4982054	110RC044	231908	26 July 2024

- This Certification is traceable to SI Through Technology Promotion Association (Thailand - Japan)

2. Certified Reference Materials :-

- Conductivity calibration solution, CPA chem Ltd., The measurement results are traceable to SI through CPA chem Ltd., ANSI-ASQ National Accreditation Board, Accredited No. AR-1835

Conductivity Solution	Manufacturer	Lot No.	Exp. date
1413.0 $\mu\text{S/cm}$	CPA Chem	913596	14 July 2024
12.880 mS/cm	CPA Chem	913597	14 July 2024

- Control Conductivity calibration solution temperature by Water bath (25 \pm 0.1) $^{\circ}\text{C}$

3. This certificate is valid only to the item calibrated on date and place of calibration.

Calibration results

Function : Conductivity Measurement

(*) After Adjustment at 1413.0 $\mu\text{S/cm}$

Conductivity Electrode Serial No.: 18L100008

Standard Conductivity Solution	Before Adjustment UUC* Reading	After Adjustment UUC* Reading	Uncertainty of Measurement (\pm)	Coverage factor k
1413.0 $\mu\text{S/cm}$	1393 $\mu\text{S/cm}$	1413 $\mu\text{S/cm}$	9.2 $\mu\text{S/cm}$	2.00
12.880 mS/cm	12.36 mS/cm	12.42 mS/cm	0.086 mS/cm	2.00

Remark - UUC* = Unit Under Calibration*Saitip*เอกสารไม่ควบคุม
a 1191350

Cert.No.: 23CH1488

Page.: 3 of 3

Calibration Results

Function : Temperature Measurement

(*) Without adjustment

This equipment was connected with Temperature Probe;

- Model :	PRO 30 COND-T
- Serial No. :	18L100038

Dimension of probe;

- Length :	7 mm
- Diameter :	2.4 mm
- Immersion Depth :	95 mm

Calibration Point ($^{\circ}\text{C}$)	Standard Temperature ($^{\circ}\text{C}$)	UUC* Reading ($^{\circ}\text{C}$)	Error ($^{\circ}\text{C}$)	Uncertainty of Measurement (\pm $^{\circ}\text{C}$)	Coverage factor k
25.0	25.002	25.0	-0.002	0.13	2.00
30.0	30.002	30.0	-0.002	0.13	2.00
35.0	35.002	35.0	-0.002	0.13	2.00

Remark : - UUC* = Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

-o0o-

*Saitip*เอกสารไม่ควบคุม
a 1191349



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-29 FAX. 0-2719-9484



Cert.No.: 23CH1487
Page.: 1 of 3

Certificate of Calibration

Equipment : pH Meter
Manufacturer : EcoSense
Model : pH100A
Serial No. : JC03354
ID No. : UAE.EFM.063/2562(ENV.pH 03/62)
Condition As-Received: Used Item
Received Date : 21 November 2023
Calibration Date : 22 November 2023
Reference : 2311-0720WSC-1
Submitted by : United Analyst and Engineering Consultant Co.,Ltd.
3 Soi Udomsuk 41, Sukhumvi Road,
Bangchak, Phrakhanong, Bangkok 10260
Ambient Temperature : (25 ± 2.5) °C
Relative Humidity : (50 ± 15) %
Calibration Procedure : 1. - house method
- CP-CI-5 by direc. measurement with standard
voltage calibrator and direct measurement with
certified reference material (CRM)
- CP-CI-8 by comparison with standard thermometer

Calibrated by : Warakorn Lemgagtrakul

Approved by : 
Approved Signatory

(✓) Saitthip Meangmai
() Warakorn Lemgagtrakul
() Ponpan Paipim

Issue Date : 27 November 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

เอกสารไม่ควบคุม
A 0061266



Cert.No.: 23CH1487
Page.: 2 of 3

Condition of this calibration result

1. Reference Standard Instrument : -

Instrument	Serial No.	ID No.	Cert. No.	Due Date
1) Document Process Calibrator	54030049	130RC116	23E2802	27 Aug 2024
2) Ref. Standard Thermometer	4982054	110RC044	23I908	26 July 2024

This certification is traceable to the International System of Unit maintained through:-
- Technology Promotion Association (Thailand-Japan)

2. Certified Reference Materials : The measurement results are traceable to SI through CPA chem Ltd.,
ANSI-ASQ National Accreditation Board, Accredited No. AR-1835

Buffer Solution	Manufacturer	Lot No.	Exp. date
pH 4.039	CPA chem	913586	14 July 2025
pH 6.995	CPA chem	913589	14 July 2024
pH 9.997	CPA chem	943106	02 Nov 2024

3 This certificate is valid only to the item calibrated on date and place of calibration.

Calibration Results

Function : mV Measurement

Performing standard curve by Fluke at pH (4.7,7,10)

Unit Under Calibration	Nominal Value	Standard Voltage Input	Actual Reading		Uncertainty of Measurement (±mV)	Coverage factor k
			mV	pH		
pH Meter S/N.: JC03354	4.00	177.48	177	4.01	0.58	2.00
	7.00	0.00	0	7.00	0.58	2.00
	7.00	0.00	0	7.00	0.58	2.00
	10.00	-177.48	-178	10.01	0.58	2.00

เอกสารไม่ควบคุม
a 1191352



Cert.No.: 23CH1487

Page.: 3 of 3

Calibration Results**Function : pH Measurement****Performing three buffers standard curve by using buffer nominal pH (4,7)(7,10)**

Unit Under Calibration	Standard pH Buffer Solution	Actual pH Reading	Actual mV Reading (mV)	Uncertainty of pH measurement (\pm)	Coverage factor k
pH Electrode S/N.: 230906SIA605377	4.008	4.01	174	0.0085	2.05
	6.985	7.00	-2	0.0099	2.00
	6.985	7.00	-2	0.0093	2.00
	9.997	10.00	-177	0.0092	2.00

Function : Temperature Measurement**(*) Without adjustment**

This equipment was connected with Temperature Probe;

- Model : -
- Serial No. : 230906SIA605377

Dimension of probe;

- Length : 110 mm
- Diameter : 12 mm
- Immersion Depth : 100 mm

Calibration Point ($^{\circ}\text{C}$)	Standard Temperature ($^{\circ}\text{C}$)	UUC* Reading ($^{\circ}\text{C}$)	Error ($^{\circ}\text{C}$)	Uncertainty of measurement (\pm $^{\circ}\text{C}$)	Coverage factor k
25.0	25.002	25.1	0.098	0.13	2.00
30.0	30.001	30.1	0.099	0.13	2.00
35.0	35.003	35.0	-0.003	0.13	2.00

Remark : - UUC* = Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

-o0o-

Signature
เอกสารไม่ควบคุม
a 1191351

INNOVATIVE INSTRUMENT CALIBRATION LAB
INNOVATIVE INSTRUMENT CO., LTD. HEAD OFFICE
7/139 MOO 13, SOI SUNTINAKORN 11 TAMBON BANG KAE0,
AMPHOE BANG PHLI SAMUT PRAKAN PROVINCE 10540 THAILAND
TEL: (66)0-2116-5860-1 FAX: (66)0-2116-7140



Page 1 of 2.

Certificate of Calibration**Customer**

Name : UNITED ANALYST AND ENGINEERING

Certificate No : 23-ACT-118

CONSULTANT CO.,LTD.

Request No : Req-2023-1547

Address : 81 Soi Udomsuk 41, Sukhumvit Road, Bangchak,

Prakanong, Bangkok 10260

Unit Under Calibration Details

Measurement item : Acoustic Calibrator

Class : 2

Manufacturer : LARSON DAVIS

Range : 94 , 114 dB / 1000 Hz

Model : CAL150

Instrument Status : Used

Serial Number : 6171

ID : UAE.EFM.117/2562

Calibration Environment and DetailsTemperature : (23 ± 2 $^{\circ}\text{C}$)Humidity : (50 ± 20 %RH)Barometric Pressure : (1013 ± 10.0 hPa)

Received Date : 21 July 2023

Calibration Date : 4 August 2023

Location of Calibration : LAB 1 Acoustic

Calibration Procedure : In-house method CP-ACT-02 based on IEC 60942:2017 Electroacoustics - Sound calibrators

Reference Standard	Model	Serial Number	Traceable	Due Calibration
Sound Calibrator	SV 35A	58079	EEL	31 May 2024
THD Multimeter	2015	1047765	NIMT	31 January 2024

Traceability : This certificate provides traceability of measurement to recognized national standard, and to the realization of the international System of Units (SI).

Note

The reported uncertainty is based on standard uncertainty multiplied by the Coverage Factor $k=2$, providing a level of confidence approximately 95 %.

Calibrated By :

Signature
Mr. Noppadon Luangart
Service Calibration Engineer

Approved By :

Signature
Mr. Pacit Mathavorn
Calibration Engineer Supervisor

Issue Date : 4 August 2023

The results related only to the item calibrated. The certificate shall not be reproduced except in full, without written approval of the Innovative Instrument Co., Ltd.

FM-708-AC-1-02 Rev.001 Issue Date: 01/07/14

เอกสารไม่ควบคุม



Certificate No : 23-AC-T-118

Request No : Req-2023-1547

Sound pressure level

Calibration Results : Without Adjustment

Calibration Range (dB)	Without Adjustment (dB)		Adjustment (dB)		Uncertainty (\pm dB)	Acceptance limit Class 2 (\pm dB)
	Measured	Error	Measured	Error		
94 dB / 1000 Hz	94.14	0.14	-	-	0.13	0.40
114 dB / 1000 Hz	114.15	0.15	-	-	0.13	0.40

Frequency of Sound pressure level

Calibration Range (Hz)	Without Adjustment		Adjustment		Uncertainty (\pm %)	Acceptance limit Class 2 (\pm %)
	Measured (Hz)	Error (%)	Measured (Hz)	Error (%)		
94 dB / 1000 Hz	1000.00	0.00	-	-	0.01	1.7
114 dB / 1000 Hz	1000.00	0.00	-	-	0.01	1.7

Total Harmonic Distortion plus Noise of Sound pressure level (THD+N %)

Calibration Range (Hz)	Without Adjustment	Adjustment	Uncertainty (\pm %)	Acceptance limit Class 2 (\pm %)
	Measured (%)	Measured (%)		
94 dB / 1000 Hz	0.04	-	0.40	3.0
114 dB / 1000 Hz	0.20	-	0.40	3.0

Note :

- Acceptance limit was IEC60942:2017 Class 1
- The calibration results exclude the calibrator pressure correction
- The calibration results exclude the microphone volume correction

End of Calibration

SITHIPORN ASSOCIATES CO.,LTD. CALIBRATION LABORATORY



451-451/1 Sirinthorn Rd.,Bangbunru, Bangplud Bangkok 10700 THAILAND.
 Tel.0-2435-8800 Fax.0-2433-1679 e-mail:cal-center@sithiphorn.com http://www.sithiphorn.com

Cert. No. : ACL23199
 Pages : 1 of 8

Calibration Certificate

Equipment : SOUND LEVEL METER
 Manufacturer : RION
 Model : NL-62 / Microphone UC-59L / Preamplifier NH-26
 Serial No.: 00130355 / 02734 / 00389
 ID No.: UAE.EMA2.102/2556

Condition As Found : GOOD

Customer : UNITED ANALYST AND ENGINEERING CONSULTANT (UAE)
 81 SOI UDOMSUK 41, SUKHUMVIT ROAD,
 BANGCHAK SUB-DISTRICT,
 PHRAKHANONG DISTRICT, BANGKOK 10260
 THAILAND.

Location : -
 Ambient Temperature : (23.0 \pm 3) °C
 Pressure : (101.3 \pm 3) kPa
 Relative Humidity : (50.0 \pm 20) %

Received Date : 21 JUNE 2023
 Calibration Date : 26-27 JUNE 2023
 Date of Issue : 28 JUNE 2023

Calibrated by : Nathakorn Pisutpaisan

Approved by :

T. Petchurai
 (Thanakul Petchurai)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Continuation of Calibration Certificate

Cert. No. : ACL23199
Job No. : VC66AC0067
Pages : 2 of 8

Calibration Procedure : CP-AC-01

Calibration Method :

This equipment was calibrated by based on IEC-61672-3 (2013) Standard for sound level meter (SLM).
The SLM had tests to Acoustical and Electrical signal tests of frequency weighting with Anechoic chamber and Reference Standard Instruments.
For tests results of each items were made by observation of each Instruments display and also with SLM's display.

Condition of this result of calibration :

1. Reference Standard Instruments :

Instrument	Model	Serial No.	Cert. No.	Due Date
Waveform Generator	33210A	MY48017076	EF-0009-23	07-FEB-24
Waveform Generator	33511B	MY52302742	EF-0010-23	07-FEB-24
Digital Multimeter	33461A	MY53220104	EEL.BP 30/0266	13-FEB-24
Digital Multimeter	33461A	MY53220076	EEL.BP 29/0266	13-FEB-24
Digital Multimeter	34461A	MY60024273	EEL.BP 31/0266	14-FEB-24
Programmable Attenuator	MAT-1070	62100114	EF-0011-23	08-FEB-24
Condenser Microphone	4180	2977900	AA-1001-23	14-FEB-24
Measuring Amplifier	NA-42KAI	34560495	AA-3002-23	14-FEB-24

2. This result of calibration was found accurate as shown on date and place of calibration for this calibrated item only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 National Institute of Metrology (Thailand).

3.2 Thailand Institute of Scientific and Technological Research (TISTR).

Continuation of Calibration Certificate

Cert. No. : ACL23199
Job No. : VC66AC0067
Pages : 3 of 8

Summary of Measurement Result :

Parameter	Pass	Fail	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1. Absolute sensitivity	✓	-	0.2	N/A
2. Self-generated noise	✓	-	0.4	N/A
3. Acoustical signal tests of frequency weightings				
125 Hz	✓	-	0.3	0.6
1000 Hz	✓	-	0.3	0.6
8000 Hz	✓	-	0.3	0.7
4. Electrical signal tests of frequency weightings				
For 10 Hz to 4 kHz	✓	-	0.3	0.6
For > 4 kHz to 10 kHz	✓	-	0.3	0.7
For > 10 kHz to 20 kHz	✓	-	0.3	1.0
5. Frequency and time weightings at 1 kHz	✓	-	0.2	0.2
6. Long - term stability	✓	-	0.1	0.1
7. Level linearity on the reference level range	✓	-	0.2	0.3
8. Level linearity including the level range control	✓	-	0.2	0.3
9. Tone burst response	✓	-	0.2	0.3
10. Peak C sound level	✓	-	0.2	0.35
11. Overload indication	✓	-	0.2	0.25
12. High level stability	✓	-	0.1	0.1

Note : Pass/Fail evaluation for each parameter, will be considered together from the acceptance limit and the Maximum-permitted uncertainty of measurement.

Continuation of Calibration Certificate

Cert. No. : ACL23199
Job No. : VC66AC0067
Pages : 4 of 8

Result of calibration :

1. Absolute sensitivity

Reference Acoustic Signal (dB)	Measured Value (dB)	Deviation (dB)	Acceptance Limit (dB)
93,9 (93,98)	94,0	0,0	±0,3

2. Self-generated noise

2.1 Normal test

Measured Value (dB)
15,4

2.2 The microphone of the sound level meter was replaced by electrical signal input device.

Frequency Weighting	Measured value (dB)
A - weight	11,0
C - weight	15,7
Flat	24,8

3. Acoustical signal tests of frequency weightings

Meter free-field acoustic response at a level of 84 dB

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
125	0,1	0,1	0,1	± 1,0
1000	-0,1	-0,1	-0,1	± 0,7
8000	-0,7	-0,6	-0,6	+ 1,5, - 2,5

Continuation of Calibration Certificate

Cert. No. : ACL23199
Job No. : VC66AC0067
Pages : 5 of 8

4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various frequency weighting response curve (dB)			
	Flat	C-weight	A-weight	Acceptance Limits
63	0,1	0,0	0,0	±1,0
125	0,0	0,1	0,1	±1,0
250	0,0	0,0	0,0	±1,0
500	0,0	0,1	0,0	±1,0
1000	0,0	0,0	0,0	±1,0
2000	0,0	0,0	0,0	±1,0
4000	0,0	0,0	0,0	±1,0
8000	0,0	0,1	0,1	+ 1,5, - 2,5
16000	0,0	-1,2	-1,2	+ 2,5, -16,0

5. Frequency and time weightings at 1 kHz

5.1 Frequency weightings at 1 kHz

Frequency Weighting	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94,0	94,0	0,0	± 0,2
C - weight	94,0	94,0	0,0	± 0,2
Flat	94,0	94,0	0,0	± 0,2

5.2 Time weighting at 1 kHz

Frequency Weighting	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	94,0	94,0	0,0	± 0,1
Slow	94,0	94,0	0,0	± 0,1
Leq	94,0	94,0	0,0	± 0,1

6. Long - term stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	94,0	94,1	0,1	± 0,1

Continuation of Calibration Certificate

Cert. No. : ACL23199
Job No. : VC66AC0067
Pages : 6 of 8

7. Level linearity on the reference level range

Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
137.0	137.1	0.1	±0.8
136.0	136.1	0.1	±0.8
135.0	135.1	0.1	±0.8
134.0	134.1	0.1	±0.8
133.0	133.1	0.1	±0.8
132.0	132.1	0.1	±0.8
131.0	131.1	0.1	±0.8
129.0	129.1	0.1	±0.8
124.0	124.0	0.0	±0.8
119.0	119.0	0.0	±0.8
114.0	114.0	0.0	±0.8
109.0	109.0	0.0	±0.8
104.0	104.0	0.0	±0.8
99.0	99.0	0.0	±0.8
94.0	94.0	0.0	±0.8
89.0	89.0	0.0	±0.8
84.0	84.0	0.0	±0.8
79.0	79.0	0.0	±0.8
74.0	74.0	0.0	±0.8
69.0	69.0	0.0	±0.8
64.0	64.0	0.0	±0.8
59.0	59.0	0.0	±0.8
54.0	53.9	-0.1	±0.8
49.0	49.0	0.0	±0.8
44.0	44.0	0.0	±0.8
39.0	38.9	-0.1	±0.8
34.0	33.9	-0.1	±0.8
30.0	29.9	-0.1	±0.8
29.0	28.9	-0.1	±0.8
28.0	28.0	0.0	±0.8
27.0	26.9	-0.1	±0.8
26.0	25.9	-0.1	±0.8
25.0	25.0	0.0	±0.8

Continuation of Calibration Certificate

Cert. No. : ACL23199
Job No. : VC66AC0067
Pages : 7 of 8

8. Level linearity including the level range control

Range	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Auto	94.0	94.0	0.0	±0.8

9. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Cycle	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Fast	0.25	1	108.0	107.9	-0.1	1.0 ; -3.0
	2	8	117.0	117.0	0.0	1.0 ; -1.5
	200	800	134.0	134.0	0.0	±0.5
Slow	2	8	108.0	108.0	0.0	1.0 ; -3.0
	200	800	127.6	127.6	0.0	±0.5
SEL	0.25	1	99.0	98.8	-0.2	1.0 ; -3.0
	2	8	108.0	108.0	0.0	1.0 ; -1.5
	200	800	128.0	128.0	0.0	±0.5

10. Peak C sound level

Number of cycle in test signal	Anticipated Value (dB)	Measured Value, L _{peak} (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	±2.0
One	136.4	135.7	-0.7	±2.0

Number of cycle in test signal	Anticipated Value (dB)	Measured Value (dB)	Deviated Value (dB)	Acceptance Limits (dB)
Continuous	133.0	133.0	0.0	±1.0
Positive half cycle	135.4	135.2	-0.2	±1.0
Negative half cycle	135.4	135.1	-0.3	±1.0

Continuation of Calibration Certificate

Cert. No. : ACL23199
Job No. : VC66AC0067
Pages : 8 of 8

11. Overload indication

Measured value (dB)		Deviated Value (dB)	Acceptance Limits (dB)
Positive one-half cycle	Negative one-half cycle		
89.5	89.6	0.1	±1.5

12. High level stability

Frequency Weighting	SLM Display at initial (dB)	SLM Display at final (dB)	Deviated Value (dB)	Acceptance Limits (dB)
A - weight	137.0	137.0	0.0	±0.1

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor $k = 2$
or any value following calculation, providing a level of confidence of approximately 95 %

End of Calibration Certificate



ELECTRICAL AND ELECTRONICS INSTITUTE
FOUNDATION FOR INDUSTRIAL DEVELOPMENT

975 Moo 4, Bangpoo Industrial Estate, Soi 8, Sukhumvit Road km 37,

Phraek Sa, Mueang Samut Prakan, Samut Prakan 10280

Tel: +66 2709 4860 Fax: +66 2324 0917



Certificate No.: CP20230290EA
Operation No.: CP2023070044

Certificate of Calibration

Equipment: Sound Level Meter
Manufacturer: RION
Model/Type: NL-62 (Meter), UC-59L (Microphone), NH-26 (Preamplifier)
Serial No.: 00130356 (Meter), 01891 (Microphone), 00951 (Preamplifier)
ID No.: UAE.EMA2.103/2556
Customer: United Analyst and Engineering Consultant Co.,Ltd.
Address: 81 Soi Udomsuk 41, Sukhumvit Road, Bangchak
Phrakhanong, Bangkok 10260
Received Date: 24 July 2023
Calibrated Date: 3 - 4 August 2023
Issued Date: 7 August 2023
Calibrated by: Ms. Juntaporn Kunhakom

Approved by:
(Mr. Sittichai Swaksuriyawong)
Group Manager

This report was prepared electronically using applicable electronic signature. Printing or copy of file are considered as a copy of the document.

The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor (k)
providing a level of confidence of approximately 95%. This certificate may not be reproduced other than in full except
with the prior written approval of the Electrical and Electronics Institute, Foundation for Industrial Development.

Certificate No.: CP20230290EA

Calibration Report

Equipment: Sound Level Meter
Manufacturer: RION
Model/Type: NL-62 (Meter), UC-59L (Microphone), NH-26 (Preamplifier)
Serial No.: 00130356 (Meter), 01891 (Microphone), 00951 (Preamplifier)
ID No.: UAE.EMA2.103/2556
Ambient Temperature: (23 ± 2) °C
Relative Humidity: (50 ± 15) %
Pressure: (101.3 ± 1.5) kPa

Method of Calibration :-

IEC 61672-3:2013.

Condition of this result of calibration

1. Reference standards instrument :-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Standard microphone	4180	2787490	AA-1024-22	6 November 2023
2) Arbitrary Function Generator	AFG2021	C010063	CK20230040EA	26 June 2024
3) Programmable Attenuator	PA5	2755	EF-0034-22	30 October 2023
4) 6.5 Digit precision multimeter	8846A	9610014	CB20220223EA	14 November 2023
5) Pressure humidity and Temperature Transmitter	PTU301	F0640002	CL1-P230024 CD20230196EA	20 March 2024 23 July 2024
6) Pressure humidity and Temperature Transmitter	PTU301	F0640003	CL1-P230032 CD20230197EA	4 April 2024 23 July 2024
7) Performance Audio Analyzer	U8903B	MY56510003	CB20230038EA CK20220080EA	14 February 2024 8 September 2023

2. This result of calibration was found accurate as shown on date and place of calibration only.

3. This certification is traceable to the international system of unit maintained at :-

Reference standards instrument for Acoustic function

- National Institute of Metrology (Thailand)

Reference standards instrument for Electrical function

- National Institute of Metrology (Thailand)

- Electrical and Electronics Institute; NSC Accredited Calibration No.0119

Result of Calibration:-

Function : 1. Indication at the calibration check frequency

Reference Acoustic Signal (dB)	Measured value (dB)	Deviation (dB)	Acceptance limits (dB)
94.0	94.0	0.0	±0.7

Note : Absolute sensitivity was established by the use of the Sound Calibrator RION Type NC-74 S/N : 34615278.

Certificate No.: CP20230290EA

Calibration Report

Function : 2. Self-generated Noise

2.1 Microphone Installed

Measured value (dB)
16.3

2.2 Microphone replaced by the electrical input signal device

Frequency Weighting	Measured value (dB)
A-weighting	11.3
C-weighting	16.5
Z-weighting	25.1

Function : 3. Acoustical signal tests of frequency weightings (Without Windscreen)

Meter free-field acoustic response at a level of 84 dB.

Frequency (Hz)	Deviation from various Frequency Weighting Response Curve			
	C-Weighting (dB)	A-Weighting (dB)	Z-Weighting (dB)	Acceptance limits (dB)
125	0.0	-0.1	0.0	±1.0
1000	0.1	0.1	0.1	±0.7
8000	-0.2	-0.2	-0.3	+1.5; -2.5

Function : 4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various Frequency Weighting Response Curve			
	C-Weighting (dB)	A-Weighting (dB)	Z-Weighting (dB)	Acceptance limits (dB)
63	-0.2	-0.1	0.0	±1.0
125	0.0	-0.1	0.0	±1.0
250	0.0	-0.1	-0.1	±1.0
500	0.0	0.0	0.0	±1.0
1000	0.0	0.0	0.0	±0.7
2000	0.1	0.0	0.0	±1.0
4000	0.0	0.0	0.0	±1.0
8000	0.1	0.0	-0.1	+1.5; -2.5
16000	-1.4	-1.4	0.0	+2.5; -16.0

Certificate No.: CP20230290EA

Calibration Report

Function : 5. Frequency and time weighting at 1 kHz

5.1 Frequency weighting at 1 kHz

Frequency Weighting	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
C-weighting	94.0	0.0	±0.2
A-weighting	94.0	0.0	±0.2
Z-weighting	94.0	0.0	±0.2

5.2 Time weighting at 1 kHz

Time Weighting	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Fast	94.0	0.0	±0.1
Slow	94.0	0.0	±0.1
LAeq	94.0	0.0	±0.1

Function : 6. Long-Term Stability

Long-term stability over 30 minutes, with steady 1 kHz signal at reference level.

Time Period to Apply Signal (min)	Reference SPL (dB)	Record SPL at Conclusion of Time Period (dB)	Deviated value (dB)	Acceptance limits (dB)
30	94.0	94.0	0.0	±0.1

Function : 7. Level Linearity on the reference level range

7.1 Level Linearity on the reference level range, Upper

Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
94.0	94.0	0.0	±0.8
99.0	99.0	0.0	±0.8
104.0	104.0	0.0	±0.8
109.0	109.0	0.0	±0.8
114.0	114.0	0.0	±0.8
119.0	119.0	0.0	±0.8
124.0	124.0	0.0	±0.8
129.0	129.0	0.0	±0.8
130.0	130.0	0.0	±0.8
131.0	131.0	0.0	±0.8
132.0	132.0	0.0	±0.8
133.0	133.0	0.0	±0.8
134.0	134.0	0.0	±0.8
135.0	135.0	0.0	±0.8
136.0	136.0	0.0	±0.8
137.0	137.0	0.0	±0.8

Certificate No.: CP20230290EA

Calibration Report

7.2 Level Linearity on the reference level range, Lower

Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
94.0	94.0	0.0	±0.8
89.0	89.0	0.0	±0.8
84.0	84.0	0.0	±0.8
79.0	79.0	0.0	±0.8
74.0	74.0	0.0	±0.8
69.0	69.0	0.0	±0.8
64.0	64.0	0.0	±0.8
59.0	59.0	0.0	±0.8
54.0	54.0	0.0	±0.8
49.0	49.0	0.0	±0.8
44.0	44.0	0.0	±0.8
39.0	39.0	0.0	±0.8
34.0	33.9	-0.1	±0.8
29.0	28.9	-0.1	±0.8

Function : 8. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Fast	200	126.0	0.0	±0.5
	2	109.0	0.0	+1.0 ; -1.5
	0.25	99.9	-0.1	+1.0 ; -3.0
Slow	200	119.6	0.0	±0.5
	2	100.0	0.0	+1.0 ; -3.0
	0.25	120.0	0.0	±0.5
LAE	200	100.0	0.0	+1.0 ; -1.5
	2	100.0	0.0	+1.0 ; -1.5
	0.25	90.9	-0.1	+1.0 ; -3.0

Function : 9. Peak C sound level

Number of cycles in test signal	Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Complete cycle	125.4	125.2	-0.2	±2.0
Positive half cycle	124.4	124.0	-0.4	±1.0
Negative half cycle	124.4	124.0	-0.4	±1.0

Certificate No.: CP20230290EA

Calibration Report

Function : 10. Overload indication

Measured value (dB)		Deviated value (dB)	Acceptance limits (dB)
Positive one-half cycle	Negative one-half cycle		
139.4	139.5	0.1	±1.5

Function : 11. High-Level Stability

High-level stability over 5 minutes, with steady 1 kHz signal, 1 dB below upper boundary.

Time Period to Apply Signal (min)	Reference SPL (dB)	Record SPL at Conclusion of Time Period (dB)	Deviated value (dB)	Acceptance limits (dB)
5	129.0	129.0	0.0	±0.1

Uncertainty of measurement

Function	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1) Indication at the calibration check frequency	0.30	Not applicable
2) Self-generated Noise	0.10	Not applicable
3) Acoustical signal tests of frequency weightings - Free-field sound pressure response level	0.30	0.60 (10Hz to 4kHz) 0.70 (>4kHz to 10kHz)
4) Electrical signal tests of frequency weightings	0.20	0.20
5) Frequency and time weighting at 1 kHz	0.20	0.20
6) Long-Term Stability	0.10	0.10
7) Level Linearity on the reference level range	0.30	0.30
8) Tone burst response	0.20	0.30
9) Peak C sound level	0.20	0.35
10) Overload indication	0.20	0.25
11) High-Level Stability	0.10	0.10

- Remarks:
1. The acceptance limit is for the deviated value.
 2. Acceptance limits was IEC61672-3:2013 Class 1.
 3. The coverage factor $k = 2.00$

-- End of Report --

Certificate No.: CP20230291EA

Operation No.: CP2023070045

Certificate of Calibration

Equipment: Sound Level Meter

Manufacturer: RION

Model/Type: NL-62 (Meter), UC-59L (Microphone), NH-26 (Preamplifier)

Serial No.: 00130357 (Meter), 02373 (Microphone), 00391 (Preamplifier)

ID No.: UAE,EMA2.104/2556

Customer: United Analyst and Engineering Consultant Co.,Ltd.

Address: 81 Soi Udomsuk 41, Sukhumvit Road, Bangchak
Phrakhanong, Bangkok 10260

Received Date: 24 July 2023

Calibrated Date: 3 - 4 August 2023

Issued Date: 7 August 2023

Calibrated by: Ms. Juntaporn Kunhakom

Approved by:

(Mr. Sittichai Swaksuriyawong)
Group Manager

This report was prepared electronically using applicable electronic signature. Printing or copy of file are considered as a copy of the document.

The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor (k) providing a level of confidence of approximately 95%. This certificate may not be reproduced other than in full except with the prior written approval of the Electrical and Electronics Institute, Foundation for Industrial Development.

Certificate No.: CP20230291EA

Calibration Report

Equipment: Sound Level Meter
Manufacturer: RION
Model/Type: NL-62 (Meter), UC-59L (Microphone), NH-26 (Preamplifier)
Serial No.: 00130357 (Meter), 02373 (Microphone), 00391 (Preamplifier)
ID No.: UAE.EMA2.104/2556
Ambient Temperature: (23 ± 2) °C
Relative Humidity: (50 ± 15) %
Pressure: (101.3 ± 1.5) kPa

Method of Calibration :-

IEC 61672-3:2013.

Condition of this result of calibration

1. Reference standards instrument :-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Standard microphone	4180	2787490	AA-1024-22	6 November 2023
2) Arbitrary Function Generator	AFG2021	C010063	CK20230040EA	26 June 2024
3) Programmable Attenuator	PA5	2755	EF-0034-22	30 October 2023
4) 6.5 Digit precision multimeter	8846A	9610014	CB20220223EA	14 November 2023
5) Pressure humidity and Temperature Transmitter	PTU301	F0640002	CL1-P230024 CD20230196EA	20 March 2024 23 July 2024
6) Pressure humidity and Temperature Transmitter	PTU301	F0640003	CL1-P230032 CD20230197EA	4 April 2024 23 July 2024
7) Performance Audio Analyzer	U8903B	MY56510003	CB20230038EA CK20220080EA	14 February 2024 8 September 2023

2. This result of calibration was found accurate as shown on date and place of calibration only.

3. This certification is traceable to the international system of unit maintained at :-

Reference standards instrument for Acoustic function

- National Institute of Metrology (Thailand)

Reference standards instrument for Electrical function

- National Institute of Metrology (Thailand)

- Electrical and Electronics Institute; NSC Accredited Calibration No.0119

Result of Calibration:-

Function : 1. Indication at the calibration check frequency

Reference Acoustic Signal (dB)	Measured value (dB)	Deviation (dB)	Acceptance limits (dB)
94.0	94.0	0.0	±0.7

Note : Absolute sensitivity was established by the use of the Sound Calibrator RION Type NC-74 S/N : 34615278.

Certificate No.: CP20230291EA

Calibration Report

Function : 2. Self-generated Noise

2.1 Microphone Installed

Measured value (dB)
16.3

2.2 Microphone replaced by the electrical input signal device

Frequency Weighting	Measured value (dB)
A-weighting	12.1
C-weighting	18.5
Z-weighting	27.5

Function : 3. Acoustical signal tests of frequency weightings (Without Windscreen)

Meter free-field acoustic response at a level of 84 dB.

Frequency (Hz)	Deviation from various Frequency Weighting Response Curve			
	C-Weighting (dB)	A-Weighting (dB)	Z-Weighting (dB)	Acceptance limits (dB)
125	0.1	0.0	0.2	±1.0
1000	0.0	0.0	0.0	±0.7
8000	0.0	0.0	0.0	+1.5; -2.5

Function : 4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various Frequency Weighting Response Curve			
	C-Weighting (dB)	A-Weighting (dB)	Z-Weighting (dB)	Acceptance limits (dB)
63	0.0	0.0	0.1	±1.0
125	0.1	-0.1	0.0	±1.0
250	0.0	0.0	0.1	±1.0
500	0.0	0.0	0.1	±1.0
1000	0.0	0.0	0.0	±0.7
2000	0.1	0.0	0.0	±1.0
4000	0.0	0.0	0.0	±1.0
8000	0.1	0.1	0.0	+1.5; -2.5
16000	-1.3	-1.3	0.1	+2.5; -16.0

Certificate No.: CP20230291EA

Calibration Report

Function : 5. Frequency and time weighting at 1 kHz

5.1 Frequency weighting at 1 kHz

Frequency Weighting	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
C-weighting	94.0	0.0	±0.2
A-weighting	94.0	0.0	±0.2
Z-weighting	94.0	0.0	±0.2

5.2 Time weighting at 1 kHz

Time Weighting	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Fast	94.0	0.0	±0.1
Slow	94.0	0.0	±0.1
LAeq	94.0	0.0	±0.1

Function : 6. Long-Term Stability

Long-term stability over 30 minutes, with steady 1 kHz signal at reference level.

Time Period to Apply Signal (min)	Reference SPL (dB)	Record SPL at Conclusion of Time Period (dB)	Deviated value (dB)	Acceptance limits (dB)
30	94.0	94.0	0.0	±0.1

Function : 7. Level Linearity on the reference level range

7.1 Level Linearity on the reference level range, Upper

Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
94.0	94.0	0.0	±0.8
99.0	99.0	0.0	±0.8
104.0	104.0	0.0	±0.8
109.0	109.0	0.0	±0.8
114.0	114.0	0.0	±0.8
119.0	119.0	0.0	±0.8
124.0	124.0	0.0	±0.8
129.0	129.0	0.0	±0.8
130.0	130.0	0.0	±0.8
131.0	131.0	0.0	±0.8
132.0	132.0	0.0	±0.8
133.0	133.0	0.0	±0.8
134.0	134.0	0.0	±0.8
135.0	135.0	0.0	±0.8
136.0	136.0	0.0	±0.8
137.0	137.0	0.0	±0.8

Certificate No.: CP20230291EA

Calibration Report

7.2 Level Linearity on the reference level range, Lower

Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
94.0	94.0	0.0	±0.8
89.0	89.0	0.0	±0.8
84.0	84.0	0.0	±0.8
79.0	79.0	0.0	±0.8
74.0	74.0	0.0	±0.8
69.0	69.0	0.0	±0.8
64.0	64.0	0.0	±0.8
59.0	59.0	0.0	±0.8
54.0	54.0	0.0	±0.8
49.0	49.0	0.0	±0.8
44.0	44.0	0.0	±0.8
39.0	39.0	0.0	±0.8
34.0	34.0	0.0	±0.8
29.0	28.9	-0.1	±0.8

Function : 8. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Fast	200	126.0	0.0	±0.5
	2	108.9	-0.1	+1.0 ; -1.5
	0.25	99.9	-0.1	+1.0 ; -3.0
Slow	200	119.6	0.0	±0.5
	2	100.0	0.0	+1.0 ; -3.0
	200	120.0	0.0	±0.5
LAE	2	100.0	0.0	+1.0 ; -1.5
	0.25	90.8	-0.2	+1.0 ; -3.0

Function : 9. Peak C sound level

Number of cycles in test signal	Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Complete cycle	125.4	125.2	-0.2	±2.0
Positive half cycle	124.4	124.1	-0.3	±1.0
Negative half cycle	124.4	124.1	-0.3	±1.0

Certificate No.: CP20230291EA

Calibration Report

Function : 10. Overload indication

Measured value (dB)		Deviated value (dB)	Acceptance limits (dB)
Positive one-half cycle	Negative one-half cycle		
139.5	139.5	0.0	±1.5

Function : 11. High-Level Stability

High-level stability over 5 minutes, with steady 1 kHz signal, 1 dB below upper boundary.

Time Period to Apply Signal (min)	Reference SPL (dB)	Record SPL at Conclusion of Time Period (dB)	Deviated value (dB)	Acceptance limits (dB)
5	129.0	129.0	0.0	±0.1

Uncertainty of measurement

Function	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1) Indication at the calibration check frequency	0.30	Not applicable
2) Self-generated Noise	0.10	Not applicable
3) Acoustical signal tests of frequency weightings - Free-field sound pressure response level	0.30	0.60 (10Hz to 4kHz) 0.70 (>4kHz to 10kHz)
4) Electrical signal tests of frequency weightings	0.20	0.20
5) Frequency and time weighting at 1 kHz	0.20	0.20
6) Long-Term Stability	0.10	0.10
7) Level Linearity on the reference level range	0.30	0.30
8) Tone burst response	0.20	0.30
9) Peak C sound level	0.20	0.35
10) Overload indication	0.20	0.25
11) High-Level Stability	0.10	0.10

- Remarks:
1. The acceptance limit is for the deviated value.
 2. Acceptance limits was IEC61672-3:2013 Class 1.
 3. The coverage factor $k = 2.00$

-- End of Report --

Certificate No.: CP20230293EA

Operation No.: CP2023070030

Certificate of Calibration

Equipment: Sound Level Meter

Manufacturer: 01dB (Meter), G.R.A.S. (Microphone), 01dB (Preamplifier)

Model/Type: CUBE (Meter), 40CD (Microphone), PRE22 (Preamplifier)

Serial No.: 11070 (Meter), 260897 (Microphone), 1605132 (Preamplifier)

ID No.: UAE.EFM.009/2560 (Meter), UAE.EFM.009/2560 (Extension cable)

Customer: United Analyst and Engineering Consultant Co.,Ltd.

Address: 81 Soi Udomsuk 41, Sukhumvit Road, Bangchak
Phrakhanong, Bangkok 10260

Received Date: 24 July 2023

Calibrated Date: 4 - 7 August 2023

Issued Date: 8 August 2023

Calibrated by: Ms. Juntaporn Kunhakom

Approved by:

(Mr. Sittichai Swaksuriyawong)
Group Manager

This report was prepared electronically using applicable electronic signature. Printing or copy of file are considered as a copy of the document.

The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor (k) providing a level of confidence of approximately 95%. This certificate may not be reproduced other than in full except with the prior written approval of the Electrical and Electronics Institute, Foundation for Industrial Development.

Certificate No.: CP20230293EA

Calibration Report

Equipment: Sound Level Meter
 Manufacturer: 01dB (Meter), G.R.A.S. (Microphone), 01dB (Preamplifier)
 Model/Type: CUBE (Meter), 40CD (Microphone), PRE22 (Preamplifier)
 Serial No.: 11070 (Meter), 260897 (Microphone), 1605132 (Preamplifier)
 ID No.: UAE.EFM.009/2560 (Meter), UAE.EFM.009/2560 (Extension cable)
 Ambient Temperature: (23 ± 2) °C
 Relative Humidity: (50 ± 15) %
 Pressure: (101.3 ± 1.5) kPa

Method of Calibration :-

IEC 61672-3:2013.

Condition of this result of calibration

1. Reference standards instrument :-

Instrument	Model	Serial No.	Cert. No.	Due Date
1) Standard microphone	4180	2787490	AA-1024-22	6 November 2023
2) Arbitrary Function Generator	AFG2021	C010063	CK20230040EA	26 June 2024
3) Programmable Attenuator	PA5	2755	EF-0034-22	30 October 2023
4) 6.5 Digit precision multimeter	8846A	9610014	CB20220223EA	14 November 2023
5) Pressure humidity and Temperature Transmitter	PTU301	F0640002	CL1-P230024 CD20230196EA	20 March 2024 23 July 2024
6) Pressure humidity and Temperature Transmitter	PTU301	F0640003	CL1-P230032 CD20230197EA	4 April 2024 23 July 2024
7) Performance Audio Analyzer	U8903B	MY56510003	CB20230038EA CK20220080EA	14 February 2024 8 September 2023

2. This result of calibration was found accurate as shown on date and place of calibration only.

3. This certification is traceable to the international system of unit maintained at :-

Reference standards instrument for Acoustic function

- National Institute of Metrology (Thailand)

Reference standards instrument for Electrical function

- National Institute of Metrology (Thailand)

- Electrical and Electronics Institute; NSC Accredited Calibration No.0119

Result of Calibration:-

Function : 1. Indication at the calibration check frequency

Reference Acoustic Signal (dB)	Correction for Microphone Model 40CD (dB)	Effective Calibration Level (dB)	Measured value (dB)	Deviation (dB)	Acceptance limits (dB)
94.1	-0.3	93.8	93.8	0.0	±0.7

Note : Absolute sensitivity was established by the use of the Sound Calibrator 01dB Type CAL31 S/N: 82795.

Certificate No.: CP20230293EA

Calibration Report

Function : 2. Self-generated Noise

2.1 Microphone Installed

Measured value (dB)
16.5

2.2 Microphone replaced by the electrical input signal device

Frequency Weighting	Measured value (dB)
A-weighting	9.8
C-weighting	10.6
Z-weighting	17.4

Function : 3. Acoustical signal tests of frequency weightings (With Windscreen)

Meter free-field acoustic response at a level of 84 dB.

Frequency (Hz)	Deviation from various Frequency Weighting Response Curve			
	C-Weighting (dB)	A-Weighting (dB)	Z-Weighting (dB)	Acceptance limits (dB)
125	0.1	0.0	0.1	±1.0
1000	0.2	0.2	0.2	±0.7
8000	-0.3	-0.3	0.2	+1.5; -2.5

Function : 4. Electrical signal tests of frequency weightings

Weighting network response with relative to 1 kHz.

Frequency (Hz)	Deviation from various Frequency Weighting Response Curve			
	C-Weighting (dB)	A-Weighting (dB)	Z-Weighting (dB)	Acceptance limits (dB)
63	0.0	-0.2	0.1	±1.0
125	0.1	-0.1	0.1	±1.0
250	0.2	0.1	0.2	±1.0
500	0.2	0.2	0.2	±1.0
1000	0.2	0.2	0.2	±0.7
2000	0.1	0.1	0.1	±1.0
4000	0.8	0.8	0.9	±1.0
8000	-0.1	-0.1	0.5	+1.5; -2.5
16000	-9.7	-9.6	-4.3	+2.5; -16.0

Certificate No.: CP20230293EA

Calibration Report

Function : 5. Frequency and time weighting at 1 kHz

5.1 Frequency weighting at 1 kHz

Frequency Weighting	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
C-weighting	94.0	0.0	±0.2
A-weighting	94.0	0.0	±0.2
Z-weighting	94.0	0.0	±0.2

5.2 Time weighting at 1 kHz

Time Weighting	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Fast	94.0	0.0	±0.1
Slow	94.0	0.0	±0.1
LAeq	94.0	0.0	±0.1

Function : 6. Long-Term Stability

Long-term stability over 30 minutes, with steady 1 kHz signal at reference level.

Time Period to Apply Signal (min)	Reference SPL (dB)	Record SPL at Conclusion of Time Period (dB)	Deviated value (dB)	Acceptance limits (dB)
30	94.0	94.0	0.0	±0.1

Function : 7. Level Linearity on the reference level range

7.1 Level Linearity on the reference level range, Upper

Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
94.0	94.0	0.0	±0.8
99.0	99.0	0.0	±0.8
104.0	104.0	0.0	±0.8
109.0	108.8	-0.2	±0.8
114.0	113.8	-0.2	±0.8
119.0	118.8	-0.2	±0.8
124.0	123.8	-0.2	±0.8
129.0	128.8	-0.2	±0.8
134.0	133.8	-0.2	±0.8

Certificate No.: CP20230293EA

Calibration Report

7.2 Level Linearity on the reference level range, Lower

Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
94.0	94.0	0.0	±0.8
89.0	89.0	0.0	±0.8
84.0	84.0	0.0	±0.8
79.0	79.0	0.0	±0.8
74.0	74.0	0.0	±0.8
69.0	69.0	0.0	±0.8
64.0	64.0	0.0	±0.8
59.0	59.0	0.0	±0.8
54.0	54.0	0.0	±0.8
49.0	49.0	0.0	±0.8
44.0	44.0	0.0	±0.8
39.0	39.0	0.0	±0.8
34.0	33.9	-0.1	±0.8
29.0	29.0	0.0	±0.8
24.0	24.1	0.1	±0.8

Function : 8. Tone burst response

Time Weighting	Tone burst duration, Tb (ms)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Fast	200	134.0	0.0	±0.5
	2	117.0	0.0	+1.0 ; -1.5
	0.25	107.8	-0.2	+1.0 ; -3.0
Slow	200	127.6	0.0	±0.5
	2	108.0	0.0	+1.0 ; -3.0
	0.25	98.9	-0.1	+1.0 ; -3.0

Function : 9. Peak C sound level

Number of cycles in test signal	Anticipated Value (dB)	Measured value (dB)	Deviated value (dB)	Acceptance limits (dB)
Complete cycle	133.4	134.6	1.2	±2.0
Positive half cycle	132.4	131.6	-0.8	±1.0
Negative half cycle	132.4	131.7	-0.7	±1.0

Function : 10. Overload indication

Measured value (dB)		Deviated value (dB)	Acceptance limits (dB)
Positive one-half cycle	Negative one-half cycle		
140.5	141.1	0.6	±1.5

Certificate No.: CP20230293EA

Calibration Report

Function : 11. High-Level Stability

High-level stability over 5 minutes, with steady 1 kHz signal, 1 dB below upper boundary.

Time Period to Apply Signal (min)	Reference SPL (dB)	Record SPL at Conclusion of Time Period (dB)	Deviated value (dB)	Acceptance limits (dB)
5	137.0	137.0	0.0	±0.1

Uncertainty of measurement

Function	Uncertainty (dB)	Maximum-permitted uncertainty of measurement (dB)
1) Indication at the calibration check frequency	0.30	Not applicable
2) Self-generated Noise	0.10	Not applicable
3) Acoustical signal tests of frequency weightings - Free-field sound pressure response level	0.30	0.60 (10Hz to 4kHz) 0.70 (>4kHz to 10kHz)
4) Electrical signal tests of frequency weightings	0.20	0.20
5) Frequency and time weighting at 1 kHz	0.20	0.20
6) Long-Term Stability	0.10	0.10
7) Level Linearity on the reference level range	0.30	0.30
8) Tone burst response	0.20	0.30
9) Peak C sound level	0.20	0.35
10) Overload indication	0.20	0.25
11) High-Level Stability	0.10	0.10

Remarks: 1. The acceptance limit is for the deviated value.
2. Acceptance limits was IEC61672-3:2013 Class 1.
3. The coverage factor $k = 2.00$

-- End of Report --