

เอกสารเทียบเครื่องมือที่ใช้ในการวิเคราะห์



THAI HEART CALIBRATION CO., LTD.
112/1 Moo 5, Phraek Si, Muang, Samut Prakan 10780
Tel: 02-5942162, 02-5754292, 02-5754293 Fax: 02-5754807



CERTIFICATE OF CALIBRATION

Certificate No.: C0-2008004/21 Page 1 of total 4 pages

Customer WATER ANALYSIS CENTER CO., LTD.
30/5 Soi Vipavadee 50, Vipavadee Rangsit Road,
Kweng Taladbangklien, Khet Laksi, Bangkok 10210

Equipment pH Meter
Manufacturer METTLER TOLEDO
Model SevenCompact
Serial No. B327527211
ID No. WWL 0068
Description Range : 0 - 14 pH, Resolution : 0.01 pH

Environmental Conditions Ambient Temperature: (20 ± 2) °C
Relative Humidity: (50 ± 10) %

Calibration Location Jayhawks Laboratory (CL&GL)
Received Date 20 August 2021
Calibration Date 20 August 2021

Date of Issue 23 August 2021

Checked by
Act as Technical Manager

() (Krisyol K.) () (Sakda V.)
() (Putiphan K.) () (Onnape P.)
() (Pongsak H.) () (Nitiphong K.)
() (Kanung C.) () (Nonthuchai K.)
() (Pramong P.) () (Noppol P.)

Approved by
Representative of Managing Director

(Dr. Ekachai Puthitwong)



THAI HEART CALIBRATION CO., LTD.
112/1 Moo 5, Phraek Si, Muang, Samut Prakan 10780
Tel: 02-5942162, 02-5754292, 02-5754293 Fax: 02-5754807



Certificate No.: C0-2008004/21

Page 2 of total 4 pages

Reference Method:

- The calibration method used was CP-178 based on an in-house method.
- This certificate can be traceable to the national standards, which is realized the shown measurement units according to the International System of Units (SI Units).

Reference Standard:

Type	pH Value	Lot No.	Due Date	Traceability
pH Standard Solution	4.01	081020	Feb. 1, 2022	NIMT
	7.01	020221	Dec. 25, 2021	
	10.00	091020	Jan. 19, 2022	

Type	Model	Serial No.	Certificate No.	Due Date	Traceability
Documenting Process Calibrator	753	3101007	10-0804001/21	Apr. 7, 2022	THC
Digital Thermometer with Sensor	1523 / 5622	1709138 / 4605984-005	10-1006001/21	Jun. 10, 2022	

Remark: This certificate is traceable to the International System of Unit (SI Unit) through:
- NIMT, National Institute of Metrology (Thailand).
- THC, Thai Heart Calibration Co., Ltd.

Measurement Results:

1. Function Simulated pH Meter

Standard Applied (mV)	Nominal Value (pH)	UUC Reading		Uncertainty (± mV)
		pH	mV	
177.48	4.00	4.01	177.4	0.060
0.00	7.00	7.00	0.0	0.060
-177.48	10.00	10.01	-177.4	0.060

UUC : Unit Under Calibration
Note : Adjust Curve to simulate pH (4,7,10)

Certificate No.: CO-2008004/21

Page 3 of total 4 pages

Measurement Results (Cont.):

2. Calibration of pH Electrode

pH Standard Solution (pH)	Measured Value		Uncertainty (± pH)
	(pH)	(mV)	
4.01	4.00	187.0	0.013
7.01	7.00	11.1	0.013
10.00	10.02	-161.6	0.013

Note: Adjust Curve to Buffer Solution pH (4,7,10)

Temperature stability of micro bath: $25 \pm 0.2^\circ\text{C}$

The above reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor $k = 2.00$, providing a level of confidence approximately 95%.

ภาคผนวก ข-2

Certificate No.: CO-2008004/21

Page 4 of total 4 pages

Reference Method:

- The calibration method used was CP-096 based on an in-house method.
- The temperature scale used was an ITS-90.
- This certificate can be traceable to the national standards, which is realized the shown measurement units according to the International System of Units (SI Units).

Reference Standard Instruments:

Type	Model	Serial No.	Cert. No.	Due Date	Traceability
Thermometer Readout	1529-R	B7C853	20E3985	Nov. 9, 2021	TPA
Platinum Resistance Thermometer	5626	4853	COA30046	Oct. 28, 2023	FLUKE
Liquid Bath	XORTS-40A	XO111019	10-0306002/21	Jun. 3, 2023	THC

Remark: This certificate is traceable to the International System of Unit (SI Unit) through:

- TPA, Technology Promotion Association (Thailand-Japan).
- FLUKE, Fluke Corporation, U.S.A.
- THC, Thai Heart Calibration Co., Ltd.

Measurement Results:

(X) Without Adjustment

Dimension of probe : Diameter 4 mm. Sensor Type : RTD (PT100)

Immersion Depth (mm.)	Standard Reading ($^\circ\text{C}$)	UUC Reading ($^\circ\text{C}$)	Correction ($^\circ\text{C}$)	Uncertainty ($\pm ^\circ\text{C}$)
120	22.00	22.0	0.00	0.058
120	25.00	25.0	0.00	0.058
120	28.00	28.0	0.00	0.058

UUC : Unit Under Calibration

The above reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor $k = 2.00$, providing a level of confidence approximately 95%.

- End of Certificate -



THAI HEART CALIBRATION CO., LTD.

112/1 Moo 5, Puch SA, Bang, Samut Prakan 10280
Tel: 0-2394-3169, 0-2257-8343, 0-2257-8344, 0-2257-8307



CERTIFICATE OF CALIBRATION

Certificate No.: CO-2107005/21 Page 1 of total 2 pages

Customer: WATER ANALYSIS CENTER CO., LTD.
30/5 Soi Vipavadee 66, Vipavadee Rangsit Road,
Kwaeng Taladbangkhen, Khet Laksj, Bangkok 10210

Equipment: Conductivity Meter
Manufacturer: EUTECH
Serial No.: 2657889
Description: Model CON 2700 ID No. WWL 0136

Environmental Conditions: Ambient Temperature: (20 ± 2) °C
Relative Humidity: (50 ± 10) %
Atmospheric Pressure: -

Calibration Location: Jayhawks Laboratory (CL&GL)
Received Date: 21 July 2021
Calibration Date: 21 July 2021

Date of Issue: 22 July 2021

Checked by: [Signature] Approved by: [Signature]
Act as Technical Manager: Representative of Managing Director

() (Krisyos K.) () (Sakda Y.)
() (Patiphan K.) () (Onnapa P.)
() (Pongsak H.) () (Nitiphong K.)
() (Kanung C.) () (Nonthachai K.)
() (Pramong P.) () (Noppol P.)

This calibration certificate shall not be reproduced other than in full except with the prior written approval of the Thai Heart Calibration Co., Ltd.
FE-169

REV.02 02/24/21

Certificate No.: AD2012-017-0001

Environment: Ambient Temperature: (25 ± 2)°C

Relative Humidity: (50 ± 15)%RH

STD Reading (mg/l)	UUC Reading Before (mg/l)	UUC Reading After (mg/l)	Error (mg/l)	Uncertainty (± mg/l)
9.046	9.07	-	0.024	0.013

STD = Standard
UUC = Unit Under Calibration

Description of UUC: Range 0.00 to 60.00 mg/l
Resolution 0.01 mg/l

Measurement Standards Used & Traceability:

The International System of Units (SI) through
MIT Certificate No. L2002-756.L2002-757 for Data Logger (Lutron Temperature & Humid & Baro) Serial No. B014887, Due 28-Feb-21
MIT Certificate No. L2001-629 for Hi Accuracy Thermometer Serial No. 130508834, Due 07-Jan-21

End of Certificate

Page 2 of 2

929/029/1 Soi Pathanakarn 30, Pathanakarn Rd., Suanluang, Suanluang, Bangkok 10250
Head Office : Tel. 02-315-8894 ext.1 Fax.02-315-4051 E-mail : atco@automation.co.th
Rayong Branch : 115 Huaypong Rd., A. Muang, Rayong 21150 Tel. 038-632-152 Fax. 038-692-245
Lamphun Branch : 1225 M.4, T. Ban Nang, A. Muang, Lamphun 51000 Tel/Fax. 053-591-976
website : www.automation.co.th

SV 212001/2021

Cert No. WAC-065
Page 1 of 2

CERTIFICATE OF CALIBRATION

Instrument : DO Meter
Model : DO-31P
Serial No. : 780065
Manufacturer : TOA-DKK
Measuring Range : 0.00 ~ 20.00 mg/l

Machine : -
Location : -

Customer : Water Analysis Center Co.,Ltd.
1/94 Moo.5 T. Kaubam, A.U.-Thai
Ayuthaya 13210 Thailand

Date Of Received : 03 / 12 / 2021
Date Of Calibration : 03 / 12 / 2021

Ambient Condition : Temperature 24 °C
Humidity 47 % RH

Calibrated By :

(Ms. Phancee Yooyen)
Technician

Approved By :

(Mr. Nipon Phungsomsak)
Technical Manager

Date Of Issue : 03 / 12 / 2021

This Certificate may not be reproduced other than in full, except with the prior written approval of the head of the industrial instruments calibration center.

Instrument : DO Meter
Model : DO-31P
Serial No. : 780065

Cert. No. WAC-065
Page 2 of 2

Calibrate Procedure

- ☐ This instrument was calibrated by comparison with standard solution (PE/ORP)
- ☐ This instrument was calibrated by comparison with scattering plate value (Turbidity)
- ☐ This instrument was calibrated by comparison with conductivity (Conductivity)
- ☒ This instrument was calibrated by comparison with Sodium sulfite anhydrous (DO)

Condition of this result of calibration

1). Reference Standard Solution

Standard	Lot No	Batch	Cert. No.	Due Date
Sodium Sulfite Power	1.06657.0500	K52300357	-	31 Mar 2022

- 2). Traceability This certification is traceable to
- ☒ Merck KGaA 64271 Darmstadt
 - ☐ DKK Corporation

Result Of Calibration

Standard Solution (mg/l) at 26.0°C	Before Adjust		After Adjust	
	Indicator	Error	Indicator	Error
Zero	0.00	+ 0.10	0.00	-
Span	7.99	+ 0.22	7.99	-

DO Electrode No. OE270AA(S) S/N 111F0029

Calibrated By

(Ms. Phancee Yooyen)
Technician

Certificate of Calibration

TEMPERATURE CONTROLLER ENCLOSURES



Certificate No.: MC 2107214

Customer : Water Analysis Center Co., Ltd.
1/94 Moo 5, T.Kantham, A.U.-Thai, Ayuthaya 13210.

Reference Job No. : 21-1565 Received Date : 13 July 2021
Description : Refrigerator
Manufacturer : SANDENINTERCOOL Model : SEC-1500SBD
Serial No. : SEC1500201A-0708-30304 ID. No. : WWL0038
Marking : Additionally for the purpose of identification by this laboratory a label marked with this certificate number (MC 2107214) has been attached to the case.
Method : In-House calibration procedure MWI-T-033 (this method is reference to TLAS G-20 "Temperature Controlled Enclosures".

Location of Calibration : Water Analysis Center Co., Ltd. : Laboratory.

Environmental Conditions : Ambient Temperature : (26.3) °C

Relative Humidity : (56.4 to 59.3) %

Date of Calibration : 13 July 2021 Date of Issue : 14 July 2021

Checked by : Thangorn
Thangorn Limchichareon
(Calibration Supervisor)

Approved by : Aittipong
Aittipong Kanjanavasi
(Technical Manager)

The uncertainties are for a confidence probability of approximately 95%

This certificate is issued in accordance with the conditions of accreditation granted by the National Standardization Council of Thailand-Office of the National Standardization Council that has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of Master Calibration Co., Ltd.

Certificate No.: MC 2107214

Page 2 of 3

The Reference Standard :

Description : Certificate No. : Serial No. : Due date :
Data Acquisition/Switch Unit : MC 2009600 : MY44095818 : 8 August 2021
With Thermocouple Type " T " ID. No.6/1 to 6/9

This certificate is traceable to the international system of units maintained at:

- Master Calibration Co., Ltd. And Quality Reborn Co., Ltd.


1. Calibration Procedure:

This instrument was calibration according to TLAS G-20 by comparison with calibrated thermocouple type T under no load condition. The Thermocouples were placed on nine points and located one thermocouple in each of the eight corners of the chamber and was away from the each wall of 5 cm to 10 cm. And placed the ninth thermocouple within 2.5 cm of the geometric center of the chamber.

Temperature Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.

Temperature Stability - one-half of the greatest maximum difference of measured temperatures at any one sensor.

Overall Variation - The Difference of the maximum and minimum measured temperatures throughout observation.



Overall Ambient Temperature around the Chamber variation : 4.4 °C
Overall Line Voltage variation : 0.1 V
Chamber Size (W*H*D): 171 cm x 157 cm x 60 cm

Figure 1 : Sensor Installation Location

Checked by : Thangorn

Certificate No.: MC 2107214

2. Result of calibration :

Temperature Measurement Accuracy Test

Indicating Temperature (°C)	Measured Temperature (°C) at Spread Locations									Uncertainty (°C)
	#1	#2	#3	#4	#5	#6	#7	#8	#9	
2.6	4.0	4.0	4.1	4.0	3.9	3.8	3.7	3.8	3.4	1.2

Chamber Characterization Result

Controller Temperature (°C)	Indicating Temperature (°C)	Temperature Stability (°C)	Temperature Uniformity (°C)	Overall Variation (°C)
2.0	2.6	2.7	1.4	5.8

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95 %.

This report will certify of the calibrated equipment only.

End of Certificate

Checked by: Thanasorn



TEMPERATURE
CONTROLLER ENCLOSURES

Report No.: MC 2103787

Customer : Water Analysis Center Co., Ltd.
1/94 Moo 5, T.Kanham, A.U-Thai, Ayuthaya 13210.

Reference Job No. : 21-0710 Received Date : 25 March 2021
Description : Oven Model : UF260
Manufacturer : Menmert ID. No. : N/A
Serial No. : B620.0814
Marking : Additionally for the purpose of identification by this laboratory a label marked with this report number (MC 2103787) has been attached to the case.
Method : In-House calibration procedure MWI-T-033 this method is reference to "TLAS G-20 "Temperature Controlled Enclosures".
Location of Calibration : Water Analysis Center Co., Ltd. ; Laboratory.
Environmental Conditions : Ambient Temperature : (31.8 to 35.3) °C
Relative Humidity : (44.7 to 55.9) %
Date of Calibration : 25 March 2021 Date of Issue : 26 March 2021

Checked by: Thanasorn Approved by: Aitipong
Thanasorn Linchaicharn Aitipong Kadjanavasil
(Calibration Supervisor) (Technical Manager)

The uncertainties are for a confidence probability of approximately 95%

This certificate is issued in accordance with the condition of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.

Continuation of Report No. : MC 2103787

Page 2 of 3

The Reference Standard :

Description
Data Acquisition/Switch Unit
With Thermocouple Type "T" ID. No.17/1 to 17/9

Report No.
MC 2016027

Serial No.
MY41010916

Due date
10 January 2022

This certificate is traceable to the international system of units maintained at:

- Master Calibration Co., Ltd.

1. Calibration Procedure:

This instrument was calibration according to TLAS G-20 by comparison with calibrated thermocouple type T under no load condition. The Thermocouples were placed on nine points and located one thermocouple in each of the eight corners of the chamber and was away from the each wall of 5 cm to 10 cm. And placed the ninth thermocouple within 2.5 cm of the geometric center of the chamber.

Temperature Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.

Temperature Stability - one-half of the greatest maximum difference of measured temperatures at any one sensor.

Overall Variation - The Difference of the maximum and minimum measured temperatures throughout observation.




Figure 1: Sensor Installation Location

Checked by : Thanigorn

(MCF-Q-077 ; Rev.3 ; Date : 15/07/2014)

Continuation of Report No. : MC 2103787

Page 3 of 3

2. Result of calibration :

Temperature Measurement Accuracy Test

Indicating Temperature (°C)	Measured Temperature (°C) at Spread Locations									Uncertainty (±°C)
	#1	#2	#3	#4	#5	#6	#7	#8	Ref. #9	
104.0	103.7	103.7	103.7	103.2	104.2	104.3	104.3	104.3	104.0	0.67
180.0	179.1	179.1	179.0	179.2	180.4	180.5	180.6	180.6	180.2	0.99

Chamber Characterization Result

Controller Temperature (°C)	Indicating Temperature (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
104.0	104.0	0.27	0.44	1.0
180.0	180.0	0.29	1.31	1.9

3. Uncertainty of Measurement

The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95 %.

This report will certify of the calibrated equipment only.

End of Calibration Report

Checked by : Thanigorn

Calibration Report

TEMPERATURE
CONTROLLER ENCLOSURES



Report No. : MC 2103787

Customer

: Water Analysis Center Co., Ltd.
1/94 Moo 5, T.Kantnam, A.U-Thai, Ayuthaya 13210.

Reference Job No.

: 21-0710 Received Date : 25 March 2021

Description

: Oven

Manufacturer

: Model : UF260

Serial No.

: ID. No. : N/A

Marking

: Additionally for the purpose of identification by this laboratory a label marked with this report number (MC 2103787) has been attached to the case.

Method

: In-House calibration procedure MWI-T-033 this method is reference to TLAS G-20 "Temperature Controlled Enclosures".

Location of Calibration

: Water Analysis Center Co., Ltd. ; Laboratory.

Environmental Conditions

: Ambient Temperature : (31.8 to 35.1) °C

: Relative Humidity : (44.7 to 55.9) %

Date of Calibration

: 25 March 2021 Date of Issue : 26 March 2021

Checked by :

Tharagon
Tharagon Limchaisurien
(Calibration Supervisor)

Approved by :

Aittiong
Aittiong Karjanawisit
(Technical Manager)

The uncertainties are for a confidence probability of approximately 95%

This certificate is issued in accordance with the condition of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the issuing laboratory.

Continuation of Report No. : MC 2103787

The Reference Standard :

Description

Report No.

Serial No.

Due date

Data Acquisition/Switch Unit

MC 2016027

MY41010916

10 January 2022

With Thermocouple Type "T" ID. No. 17/1 to 17/9

This certificate is traceable to the international system of units maintained at:

- Master Calibration Co., Ltd.


1. Calibration Procedure:

This instrument was calibration according to TLAS G-20 by comparison with calibrated thermocouple type T under no load condition. The Thermocouples were placed on nine points and located one thermocouple in each of the eight corners of the chamber and was away from the each wall of 5 cm to 10 cm. And placed the ninth thermocouple within 2.5 cm of the geometric center of the chamber.

Temperature Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.

Temperature Stability - one-half of the greatest maximum difference of measured temperatures at any one sensor.

Overall Variation - The Difference of the maximum and minimum measured temperatures throughout observation.



Overall Ambient Temperature around the Chamber variation : 1.1 °C
Overall Line Voltage variation : 1.2 V
Chamber Size (W*H*D) : 65 cm x 80 cm x 50 cm

Figure 1 : Sensor Installation Location

Checked by : Tharagon