

ภาคผนวก ง

ใบรับรองการสอบเทียบเครื่องมือ

CERTIFICATE OF CALIBRATION

NO. 20240113128

Name of Product:	Sound Level Meter
Model:	ST-11D
Serial Number:	820888
Specification:	Class 1
Conclusion:	Pass
Date of calibration:	2024-01-31
Due Date:	2025-01-29

Calibrated by:



- This report certifies that all calibration equipment used in the test is traceable with the internal ISO9001 procedures and meets all specification given in the Manual(s) or respectively surpass then, and applies only to the unit identified above.
- This certificate is produced with advanced equipment & procedures which permit comprehensive quality assurance verification of all data supplied herein.
- This certificate of calibration shall not be reproduced except in full, without written permission of the Scarlet Tech Co Ltd Taiwan.

1. Preliminary inspection: OK

 2. Type & serial No. of Microphone: AWA14425-39024

3. Adjustments to indicated sound levels:

 Type of Calibrator B&K 4231 Sound

 Pressure Level 94.0 dB

 4. Measuring up limit: 140 dBA

5. Frequency weightings (Acoustic signal tests for Z weighting, other electric signal tests.)

 Equivalent Free-field Sound Level (reference environment conditions) 93.8 dB

Nominal frequency /Hz	Frequency weighting / dB			Nominal frequency /Hz	Frequency weighting / dB		
	A	C	Z		A	C	Z
10	-71.1	-14.2	-0.4	1000	0.0	0.1	-0.1
20	-50.1	-6.2	-0.2	2000	1.3	-0.2	-0.1
31.5	-39.2	-2.6	-0.1	4000	1.1	-0.5	-0.1
63	-26.1	-0.4	-0.1	8000	-1.0	-3.1	0.0
125	-16.2	-0.1	-0.1	12500	-11.5	-8.0	0.1
250	-8.6	0.1	-0.1	16000	-11.4	-13.8	0.1
500	-3.2	0.1	-0.2	20000	-23.8	-25.9	-0.1

6. Self-generated noise

Microphone replaced by electrical input signal device

9.1 dB(A)	9.4 dB(C)	17.1 dB(Z)
-----------	-----------	------------

7. F&S Weighting

Rate of the F weighting decrease (dB/s)	35.2
Rate of the S weighting decrease (dB/s)	4.4
Deviation of F&S	-0.1

8. Level Linearity (A-weighting at frequency 1 kHz)

Reference sound level 90.0 dB

Max error at 10dB steps upper reference sound level 0.1 dB

Max error at 1dB steps within 5dB of the upper limit linear operating range 0.0 dB

Max error at 10dB steps below reference sound level 0.1 dB

Max error at 1dB steps within 5dB upper the lower limit linear operating range 0.1 dB

9. Tone burst response (A Weighting) :

Single Toneburst duration /ms	Toneburst response /dB			
	LAFmax-LA	LASmax-LA	LAE-LA	LAeqT-LA
500	0.0	-4.0	-2.9	-7.0
200	-1.0	-7.4	-6.9	-7.0
2	-18.1	-26.9	-26.9	-7.0
0.25	-27.2	/	-36.0	-7.0

10. Peak C sound level (500Hz) :

Cycle	One cycle	nominal value	Positive half	nominal value	Negative half	nominal value
LCpeak-LC(dB)	3.5	3.5	2.4	2.4	2.3	2.4

11. Overload indication: Pass

12. Statistical analysis function

Sweep signal maximum indicated sound level: 112.8 dB

Sweep amplitude: 40 dB

Scan cycle time: 60 S; Measurement period: 180 S.

Items	Measured value/dB	Theoretical calculated value/dB	Error/dB
LAeq,T	103.2	103.2	0.0
L5	110.8	110.8	0.0
L10	108.8	108.8	0.0
L50	92.9	92.8	0.1
L90	76.9	76.8	0.1
L95	75.0	74.9	0.1

Uncertainty of measurement results: 0.4 dB (k=2)

Environment conditions:

Air temperature: 20 °C

Relative humidity: 50 %

Static pressure: 101.8 kPa

Reference equipment used in the calibration:

Description:	Model	Serial No.	Expiry Date	Traceable To
Microphone	B&K 4191	2929405	2024-12-15	NML
Multi function sound calibrator	B&K 4226	2288444	2024-10-15	CIGISMEC
Signal generator	DS 360	33873	2024-10-15	CEPREI

Test specifications:

1. All Scarlet's Sound level Meter has been calibrated in accordance with the requirements as specified in ISO 17025 and the lab calibration procedure SMTP004-CA-152.
2. The electrical tests were performed using an electrical signal substituted for the microphone which was removed and replaced by an equivalent capacitance within a tolerance of $\pm 20\%$.
3. The acoustic calibration was performed using an B&K 4226 sound calibrator and corrections was applied for the difference between the free-field and pressure responses of the Sound Level Meter.


References:

IEC 61672-3 Sound Level Meters Part 3: Periodic tests

CERTIFICATE OF CALIBRATION

NO. 20240113133

Name of Product:	Sound Level Meter
Model:	ST-11D
Serial Number:	820893
Specification:	Class 1
Conclusion:	Pass
Date of calibration:	2024-01-31
Due Date:	2025-01-29

Calibrated by: 

- I. This report certifies that all calibration equipment used in the test is traceable with the internal ISO9001 procedures and meets all specification given in the Manual(s) or respectively surpass then, and applies only to the unit identified above.
- II. This certificate is produced with advanced equipment & procedures which permit comprehensive quality assurance verification of all data supplied herein.
- III. This certificate of calibration shall not be reproduced except in full, without written permission of the Scarlet Tech Co Ltd Taiwan.

1. Preliminary inspection: OK

2. Type & serial No. of Microphone: AWA14425-58633

3. Adjustments to indicated sound levels:

Type of Calibrator: B&K 4231 SoundPressure Level 94.0 dB4. Measuring up limit: 140 dBA

5. Frequency weightings (Acoustic signal tests for Z weighting, other electric signal tests.)

Equivalent Free-field Sound Level (reference environment conditions) 93.8 dB

Nominal frequency /Hz	Frequency weighting / dB			Nominal frequency /Hz	Frequency weighting / dB		
	A	C	Z		A	C	Z
10	-71.2	-14.3	-0.2	1000	0.0	0.0	-0.1
20	-50.1	-6.2	-0.2	2000	1.3	-0.1	-0.1
31.5	-39.2	-2.5	-0.1	4000	1.1	-0.8	-0.1
63	-26.1	-0.2	-0.1	8000	-1.0	-3.2	0.0
125	-16.2	-0.1	0.1	12500	-11.5	-13.5	0.1
250	-8.7	0.1	-0.1	16000	-11.5	-13.3	0.1
500	-3.2	0.2	-0.1	20000	-23.9	-25.9	-0.1

6. Self-generated noise

Microphone replaced by electrical input signal device

7.0 dB(A)	8.6 dB(C)	13.8 dB(Z)
-----------	-----------	------------

7. F&S Weighting

Rate of the F weighting decrease (dB/s)	35.1
Rate of the S weighting decrease (dB/s)	4.3
Deviation of F&S	-0.1

8. Level Linearity (A-weighting at frequency 1 kHz)

Reference sound level 90.0 dB

Max error at 10dB steps upper reference sound level 0.1 dB

Max error at 1dB steps within 5dB of the upper limit linear operating range 0.0 dB

Max error at 10dB steps below reference sound level 0.1 dB

Max error at 1dB steps within 5dB upper the lower limit linear operating range 0.1 dB

9. Tone burst response (A Weighting) :

Single Toneburst duration /ms	Toneburst response /dB			
	LAFmax-LA	LASmax-LA	LAE-LA	LAeqT-LA
500	0.0	-4.0	-2.9	-7.0
200	-1.0	-7.4	-6.9	-7.0
2	-18.1	-26.9	-26.9	-7.0
0.25	-27.2	/	-36.0	-7.0

10. Peak C sound level (500Hz) :

Cycle	One cycle	nominal value	Positive half	nominal value	Negative half	nominal value
LCpeak-LC(dB)	3.5	3.5	2.4	2.4	2.3	2.4

11. Overload indication: Pass

12. Statistical analysis function

Sweep signal maximum indicated sound level: 112.8 dB

Sweep amplitude: 40 dB

Scan cycle time: 60 S; Measurement period: 180 S.

Items	Measured value/dB	Theoretical calculated value/dB	Error/dB
LAeq,T	103.2	103.2	0.0
L5	110.8	110.8	0.0
L10	108.8	108.8	0.0
L50	92.9	92.8	0.1
L90	76.9	76.8	0.1
L95	75.0	74.9	0.1

Uncertainty of measurement results: 0.4 dB (k=2)

Environment conditions:

Air temperature: 20 °C

Relative humidity: 50 %

Static pressure: 101.8 kPa

Reference equipment used in the calibration:

Description:	Model	Serial No.	Expiry Date	Traceable To
Microphone	B&K 4191	2929405	2024-12-15	NML
Multi function sound calibrator	B&K 4226	2288444	2024-10-15	CIGISMEC
Signal generator	DS 360	33873	2024-10-15	CEPREI

Test specifications:

1. All Scarlet's Sound level Meter has been calibrated in accordance with the requirements as specified in ISO 17025 and the lab calibration procedure SMTP004-CA-152.
2. The electrical tests were performed using an electrical signal substituted for the microphone which was removed and replaced by an equivalent capacitance within a tolerance of $\pm 20\%$.
3. The acoustic calibration was performed using an B&K 4226 sound calibrator and corrections was applied for the difference between the free-field and pressure responses of the Sound Level Meter.

References:

IEC 61672-3 Sound Level Meters Part 3: Periodic tests



CERTIFICATE OF CALIBRATION

NO. 20240113134

Name of Product:	Sound Level Meter
Model:	ST-11D
Serial Number:	820894
Specification:	Class 1
Conclusion:	Pass
Date of calibration:	2024-01-31
Due Date:	2025-01-29

Calibrated by:



- I. This report certifies that all calibration equipment used in the test is traceable with the internal ISO9001 procedures and meets all specification given in the Manual(s) or respectively surpass then, and applies only to the unit identified above.
- II. This certificate is produced with advanced equipment & procedures which permit comprehensive quality assurance verification of all data supplied herein.
- III. This certificate of calibration shall not be reproduced except in full, without written permission of the Scarlet Tech Co Ltd Taiwan.

1. Preliminary inspection: OK

2. Type & serial No. of Microphone: AWA14425-58728

3. Adjustments to indicated sound levels:

Type of Calibrator B&K 4231 Sound

Pressure Level 94.0 dB

4. Measuring up limit: 140 dBA

5. Frequency weightings (Acoustic signal tests for Z weighting, other electric signal tests.)

Equivalent Free-field Sound Level (reference environment conditions) 93.8 dB

Nominal frequency /Hz	Frequency weighting / dB			Nominal frequency /Hz	Frequency weighting / dB		
	A	C	Z		A	C	Z
10	-71.1	-14.2	-0.1	1000	0.0	0.0	-0.1
20	-50.1	-6.3	-0.2	2000	1.3	-0.1	-0.1
31.5	-39.2	-2.5	-0.1	4000	1.2	-0.8	-0.1
63	-26.1	-0.2	-0.1	8000	-1.0	-3.2	0.0
125	-16.2	-0.1	0.1	12500	-11.6	-13.5	0.1
250	-8.7	0.1	-0.1	16000	-11.5	-13.3	0.1
500	-3.3	0.2	-0.1	20000	-23.8	-25.9	-0.1

6. Self-generated noise

Microphone replaced by electrical input signal device

7.4 dB(A)	10.4 dB(C)	14.6 dB(Z)
-----------	------------	------------

7. F&S Weighting

Rate of the F weighting decrease (dB/s)	35.2
Rate of the S weighting decrease (dB/s)	4.4
Deviation of F&S	-0.1

8. Level Linearity (A-weighting at frequency 1 kHz)

Reference sound level 90.0 dB

Max error at 10dB steps upper reference sound level 0.1 dB

Max error at 1dB steps within 5dB of the upper limit linear operating range 0.0 dB

Max error at 10dB steps below reference sound level 0.1 dB

Max error at 1dB steps within 5dB upper the lower limit linear operating range 0.1 dB

9. Tone burst response (A Weighting) :

Single Toneburst duration /ms	Toneburst response /dB			
	LAFmax-LA	LASmax-LA	LAE-LA	LAeqT-LA
500	0.0	-4.0	-2.9	-7.0
200	-1.0	-7.4	-6.9	-7.0
2	-18.1	-26.9	-26.9	-7.0
0.25	-27.2	/	-36.0	-7.0

10. Peak C sound level (500Hz) :

Cycle	One cycle	nominal value	Positive half	nominal value	Negative half	nominal value
LCpeak-LC(dB)	3.5	3.5	2.4	2.4	2.3	2.4

11. Overload indication: Pass

12. Statistical analysis function

Sweep signal maximum indicated sound level: 112.8 dB

Sweep amplitude: 40 dB

Scan cycle time: 60 S; Measurement period: 180 S.

Items	Measured value/dB	Theoretical calculated value/dB	Error/dB
LAeq,T	103.2	103.2	0.0
L5	110.8	110.8	0.0
L10	108.8	108.8	0.0
L50	92.9	92.8	0.1
L90	76.9	76.8	0.1
L95	75.0	74.9	0.1

Uncertainty of measurement results: 0.4 dB (k=2)

Environment conditions:

Air temperature: 20 °C

Relative humidity: 50 %

Static pressure: 101.8 kPa

Reference equipment used in the calibration:

Description:	Model	Serial No.	Expiry Date	Traceable To
Microphone	B&K 4191	2929405	2024-12-15	NML
Multi function sound calibrator	B&K 4226	2288444	2024-10-15	CIGISMEC
Signal generator	DS 360	33873	2024-10-15	CEPREI

Test specifications:

1. All Scarlet's Sound level Meter has been calibrated in accordance with the requirements as specified in ISO 17025 and the lab calibration procedure SMTP004-CA-152.
2. The electrical tests were performed using an electrical signal substituted for the microphone which was removed and replaced by an equivalent capacitance within a tolerance of $\pm 20\%$.
3. The acoustic calibration was performed using an B&K 4226 sound calibrator and corrections was applied for the difference between the free-field and pressure responses of the Sound Level Meter.

References:

IEC 61672-3 Sound Level Meters Part 3: Periodic tests



SMILE
Laboratory Co., Ltd.

บริษัท สไมล์ แล็บอราทอรี จำกัด

Smile Laboratory Co., Ltd.

563/1 ถนนทองเอก แขวงบางหว้า เขตภาษีเจริญ กรุงเทพฯ 10160 โทรศัพท์ 02-227-0265 โทรสาร 02-454-0317
563/1 Thoei Thal Rd., Bangwa, Phasicharoen, Bangkok 10160 Tel. 02-227-0265 Fax. 02-454-0317

TSP HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sampler Location	โครงการ SML-0155-02-67	Date	12 February 2024
Project Site	ชุมชนหมู่ 4 บ้านสันป่าฝ้าย	Person	Mr. Warakorn Charoenkhun

Calibration Orifice

Transfer Standard Type	Orifice	Q_{std} Slope (m)	2.10372
Calibrator Model	TE-5025A	Q_{std} Intercept (b)	-0.03890
Calibrator Serial Number	3092		

Calibration Information


Sampler Number	TSP No.05	Motor Serial Number	1203-422	Recorder Serial Number	599
----------------	-----------	---------------------	----------	------------------------	-----

Test No.	Pressure Drop Across Orifice (ΔH_2O) (inH ₂ O)			(A)	(X)	(I)	(Y)	Temperature (°K = °C+273)	Barometric Pressure (mmHg)
	Positive	Negative	ΔH_2O	$[\Delta H_2O(P_a/P_{std})(T_{std}/T_a)]^{1/2}$	$Q_{std} = (1/m)[(A-b)]$ (m ³ /min)	Sample Flow Rate Indication (ft ³ /min)	$IC = I[(P_a/P_{std})(T_{std}/T_a)]^{1/2}$ (ft ³ /min)		
1	2.2	2.1	4.30	2.05646	0.99603	36.0	35.70	303.0	760.0
2	3.1	3.0	6.10	2.44936	1.18279	43.0	42.64	303.0	760.0
3	4.3	4.2	8.50	2.89132	1.39288	50.0	49.59	303.0	760.0
4	5.2	5.1	10.30	3.18277	1.53142	55.0	54.54	303.0	760.0
5	6.1	6.0	12.10	3.44969	1.65829	60.0	59.50	303.0	760.0
Average								303.0	760.0

Linear Regression : $y = mx + b$

Slope (m)	35.494588
Intercept (b)	0.397050
R-Square (R^2)	0.999329
Correlation Coefficient (r)	0.999664

Andersen Instruments, Inc.



Calibrated By

Mr. Warakorn Charoenkhun

SMILE
Laboratory Co., Ltd.



SMILE
Laboratory Co., Ltd.

บริษัท สไมล์ แล็บอราทอรี จำกัด

Smile Laboratory Co., Ltd.

563/1 ถนนเทอดไท แขวงบางหว้า เขตภาษีเจริญ กรุงเทพฯ 10160 โทรศัพท์ 02-227-0265 โทรสาร 02-454-0317

563/1 Thoei Thai Rd., Bangwa, Phasicharoen, Bangkok 10160 Tel. 02-227-0265 Fax. 02-454-0317

TSP HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sampler Location	โครงการ SML-0155-02-67	Date	12 February 2024
Project Site	วัดวังทอง	Person	Mr. Warakorn Charoenkhun

Calibration Orifice

Transfer Standard Type	Orifice	Q_{std} Slope (m)	2.10372
Calibrator Model	TE-5025A	Q_{std} Intercept (b)	-0.03890
Calibrator Serial Number	3092		

Calibration Information


Sampler Number	TSP No.03	Motor Serial Number	1203-426	Recorder Serial Number	600
----------------	-----------	---------------------	----------	------------------------	-----

Test No.	Pressure Drop Across Orifice (ΔH_2O) (inH ₂ O)			(A)	(X)	(I)	(Y)	Temperature (°K = °C+273)	Barometric Pressure (mmHg)
	Positive	Negative	ΔH_2O	$[\Delta H_2O(P_a/P_{std})(T_{std}/T_a)]^{1/2}$	$Q_{std} = (1/m)[(A-b)]$ (m ³ /min)	Sample Flow Rate Indication (ft ³ /min)	$IC = I[(P_a/P_{std})(T_{std}/T_a)]^{1/2}$ (ft ³ /min)		
1	2.0	1.9	3.90	1.95848	0.94945	36.0	35.70	303.0	760.0
2	3.3	3.1	6.40	2.50886	1.21107	43.0	42.64	303.0	760.0
3	4.4	4.3	8.70	2.92514	1.40895	50.0	49.59	303.0	760.0
4	5.5	5.4	10.90	3.27416	1.57486	55.0	54.54	303.0	760.0
5	6.3	6.1	12.40	3.49219	1.67850	60.0	59.50	303.0	760.0
Average								303.0	760.0

Linear Regression : $y = mx + b$

Slope (m)	32.249636
Intercept (b)	4.388908
R-Square (R^2)	0.992855
Correlation Coefficient (r)	0.996421

Andersen Instruments, Inc.



Calibrated By

Mr. Warakorn Charoenkhun

SMILE
Laboratory Co., Ltd.



SMILE
Laboratory Co., Ltd.

บริษัท สไมล์ แล็บอราทอรี จำกัด

Smile Laboratory Co., Ltd.

563/1 ถนนทองหล่อ แขวงบางหว้า เขตภาษีเจริญ กรุงเทพฯ 10160 โทรศัพท์ 02-227-0265 โทรสาร 02-454-0317
563/1 Thoei Thal Rd., Bangwa, Phasicharoen, Bangkok 10160 Tel. 02-227-0265 Fax. 02-454-0317

TSP HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sampler Location	โครงการ SML-0155-02-67	Date	12 February 2024
Project Site	โรงเรียนบ้านอ้อมกอม่วง	Person	Mr. Warakorn Charoenkhun

Calibration Orifice

Transfer Standard Type	Orifice	Q_{std} Slope (m)	2.10372
Calibrator Model	TE-5025A	Q_{std} Intercept (b)	-0.03890
Calibrator Serial Number	3092		

Calibration Information


Sampler Number	TSP No.02	Motor Serial Number	1203-421	Recorder Serial Number	598
----------------	-----------	---------------------	----------	------------------------	-----

Test No.	Pressure Drop Across Orifice (ΔH_2O) (inH ₂ O)			(A)	(X)	(I)	(Y)	Temperature (°K = °C+273)	Barometric Pressure (mmHg)
	Positive	Negative	ΔH_2O	$[\Delta H_2O(P_a/P_{std})(T_{std}/T_a)]^{1/2}$	$Q_{std} = (1/m)[(A-b)]$ (m ³ /min)	Sample Flow Rate Indication (ft ³ /min)	$IC = I[(P_a/P_{std})(T_{std}/T_a)]^{1/2}$ (ft ³ /min)		
1	2.2	2.1	4.30	2.05646	0.99603	36.0	35.70	303.0	760.0
2	3.1	3.0	6.10	2.44936	1.18279	43.0	42.64	303.0	760.0
3	4.3	4.2	8.50	2.89132	1.39288	50.0	49.59	303.0	760.0
4	5.2	5.1	10.30	3.18277	1.53142	55.0	54.54	303.0	760.0
5	6.1	6.0	12.10	3.44969	1.65829	60.0	59.50	303.0	760.0
Average								303.0	760.0

Linear Regression : $y = mx + b$

Slope (m)	35.494588
Intercept (b)	0.397050
R-Square (R^2)	0.999329
Correlation Coefficient (r)	0.999664

Andersen Instruments, Inc.



Calibrated By

Mr. Warakorn Charoenkhun

SMILE
Laboratory Co., Ltd.



SMILE
Laboratory Co., Ltd.

บริษัท สไมล์ แล็บอราทอรี จำกัด

Smile Laboratory Co., Ltd.

563/1 ถนนเทอดไท แขวงบางหว้า เขตภาษีเจริญ กรุงเทพฯ 10160 โทรศัพท์ 02-227-0265 โทรสาร 02-454-0317
563/1 Thoeet Thal Rd., Bangwa, Phasicharoen, Bangkok 10160 Tel. 02-227-0265 Fax. 02-454-0317

TSP HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sampler Location	โครงการ SML-0155-02-67	Date	12 February 2024
Project Site	สถานีตรวจวัดคุณภาพอากาศอัตโนมัติ (AAQMS)	Person	Mr. Warakorn Charoenkhun

Calibration Orifice

Transfer Standard Type	Orifice	Q_{std} Slope (m)	2.10372
Calibrator Model	TE-5025A	Q_{std} Intercept (b)	-0.03890
Calibrator Serial Number	3092		

Calibration Information


Sampler Number	TSP No.01	Motor Serial Number	1203-415	Recorder Serial Number	596
----------------	-----------	---------------------	----------	------------------------	-----

Test No.	Pressure Drop Across Orifice (ΔH_2O) (inH ₂ O)			(A)	(X)	(I)	(Y)	Temperature (°K = °C+273)	Barometric Pressure (mmHg)
	Positive	Negative	ΔH_2O	$[\Delta H_2O(P_a/P_{std})(T_{std}/T_a)]^{1/2}$	$Q_{std} = (1/m)[(A-b)]$ (m ³ /min)	Sample Flow Rate Indication (ft ³ /min)	$IC = I[(P_a/P_{std})(T_{std}/T_a)]^{1/2}$ (ft ³ /min)		
1	2.2	2.1	4.30	2.05646	0.99603	36.0	35.70	303.0	760.0
2	3.2	2.9	6.10	2.44936	1.18279	43.0	42.64	303.0	760.0
3	4.1	3.8	7.90	2.78741	1.34348	50.0	49.59	303.0	760.0
4	5.4	5.3	10.70	3.24398	1.56051	55.0	54.54	303.0	760.0
5	6.3	6.2	12.50	3.50624	1.68518	60.0	59.50	303.0	760.0
Average								303.0	760.0

Linear Regression : $y = mx + b$

Slope (m)	33.768326
Intercept (b)	2.686962
R-Square (R^2)	0.990356
Correlation Coefficient (r)	0.995166

Andersen Instruments, Inc.



Calibrated By

Mr. Warakorn Charoenkhun

SMILE
Laboratory Co., Ltd.



SMILE
Laboratory Co., Ltd.

บริษัท สไมล์ แล็บอราทอรี จำกัด

Smile Laboratory Co., Ltd.

563/1 ถนนเทอดไท แขวงบางหว้า เขตภาษีเจริญ กรุงเทพฯ 10160 โทรศัพท์ 02-227-0265 โทรสาร 02-454-0317
563/1 Thoei Thai Rd., Bangwa, Phasicharoen, Bangkok 10160 Tel. 02-227-0265 Fax. 02-454-0317

PM10 HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sampler Location	โครงการ SML-0155-02-67	Date	12 February 2024
Project Site	ชุมชนหมู่ 4 บ้านสันป่าฝ้าย	Person	Mr. Warakorn Charoenkhun

Calibration Orifice

Transfer Standard Type	Orifice	Q_{std} Slope (m)	2.10372
Calibrator Model	TE-5025A	Q_{std} Intercept (b)	-0.03890
Calibrator Serial Number	3092		

Calibration Information


Sampler Number	PM10 No.09	Motor Serial Number	1203-444	Recorder Serial Number	606
----------------	------------	---------------------	----------	------------------------	-----

Test No.	Pressure Drop Across Orifice (ΔH_2O) (inH ₂ O)			(A)	(X)	(I)	(Y)	Temperature (°K = °C+273)	Barometric Pressure (mmHg)
	Positive	Negative	ΔH_2O	$[\Delta H_2O(P_a/P_{std})(T_{std}/T_a)]^{1/2}$	$Q_{std} = (1/m)[(A-b)]$ (m ³ /min)	Sample Flow Rate Indication (ft ³ /min)	$IC = I[(P_a/P_{std})(T_{std}/T_a)]^{1/2}$ (ft ³ /min)		
1	1.8	1.7	3.50	1.84924	0.89752	32.0	31.63	305.0	760.0
2	2.9	2.7	5.60	2.33912	1.13039	42.0	41.52	305.0	760.0
3	4.2	4.1	8.30	2.84772	1.37215	52.0	51.40	305.0	760.0
4	5.8	5.6	11.40	3.33742	1.60493	58.0	57.33	305.0	760.0
5	6.5	6.4	12.90	3.55020	1.70607	62.0	61.28	305.0	760.0
Average								305.0	760.0

Linear Regression : $y = mx + b$

Slope (m)	36.049868
Intercept (b)	0.245557
R-Square (R^2)	0.991639
Correlation Coefficient (r)	0.995811

Andersen Instruments, Inc.



Calibrated By

Mr. Warakorn Charoenkhun

SMILE
Laboratory Co., Ltd.

PM10 HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sampler Location	โครงการ SML-0155-02-67	Date	12 February 2024
Project Site	วัดวังทอง	Person	Mr. Warakorn Charoenkhun

Calibration Orifice

Transfer Standard Type	Orifice	Q_{std} Slope (m)	2.10372
Calibrator Model	TE-5025A	Q_{std} Intercept (b)	-0.03890
Calibrator Serial Number	3092		

Calibration Information


Sampler Number	PM10 No.04	Motor Serial Number	1203-440	Recorder Serial Number	606
----------------	------------	---------------------	----------	------------------------	-----

Test No.	Pressure Drop Across Orifice (ΔH_2O) (inH ₂ O)			(A)	(X)	(I)	(Y)	Temperature (°K = °C+273)	Barometric Pressure (mmHg)
	Positive	Negative	ΔH_2O	$[\Delta H_2O(P_a/P_{std})(T_{std}/T_a)]^{1/2}$	$Q_{std} = (1/m)[(A-b)]$ (m ³ /min)	Sample Flow Rate Indication (ft ³ /min)	$IC = I[(P_a/P_{std})(T_{std}/T_a)]^{1/2}$ (ft ³ /min)		
1	1.9	1.8	3.70	1.90134	0.92229	32.0	31.63	305.0	760.0
2	3.1	3.0	6.10	2.44131	1.17896	42.0	41.52	305.0	760.0
3	4.2	4.0	8.20	2.83051	1.36397	52.0	51.40	305.0	760.0
4	5.7	5.5	11.20	3.30801	1.59095	58.0	57.33	305.0	760.0
5	6.1	6.0	12.10	3.43836	1.65291	62.0	61.28	305.0	760.0
Average								305.0	760.0

Linear Regression : $y = mx + b$

Slope (m)	39.949045
Intercept (b)	4.972153
R-Square (R^2)	0.990397
Correlation Coefficient (r)	0.995187

Andersen Instruments, Inc.



Calibrated By

Mr. Warakorn Charoenkhun



บริษัท สไมล์ แล็บอราทอรี จำกัด

Smile Laboratory Co., Ltd.

563/1 ถนนพหลโยธิน แขวงบางหว้า เขตภาษีเจริญ กรุงเทพฯ 10160 โทรศัพท์ 02-227-0265 โทรสาร 02-454-0317
563/1 Thoei Thai Rd., Bangwa, Phasicharoen, Bangkok 10160 Tel. 02-227-0265 Fax. 02-454-0317

PM10 HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sampler Location	โครงการ SML-0155-02-67	Date	12 February 2024
Project Site	โรงเรียนบ้านหนองกุ่มม่วง	Person	Mr. Warakorn Charoenkhun

Calibration Orifice

Transfer Standard Type	Orifice	Q_{std} Slope (m)	2.10372
Calibrator Model	TE-5025A	Q_{std} Intercept (b)	-0.03890
Calibrator Serial Number	3092		

Calibration Information


Sampler Number	PM10 No.03	Motor Serial Number	1203-449	Recorder Serial Number	608
----------------	------------	---------------------	----------	------------------------	-----

Test No.	Pressure Drop Across Orifice (ΔH_2O) (inH ₂ O)			(A)	(X)	(I)	(Y)	Temperature (°K = °C+273)	Barometric Pressure (mmHg)
	Positive	Negative	ΔH_2O	$[\Delta H_2O(P_a/P_{std})(T_{std}/T_a)]^{1/2}$	$Q_{std} = (1/m)[(A-b)]$ (m ³ /min)	Sample Flow Rate Indication (ft ³ /min)	$IC = I[(P_a/P_{std})(T_{std}/T_a)]^{1/2}$ (ft ³ /min)		
1	1.8	1.7	3.50	1.84924	0.89752	32.0	31.63	305.0	760.0
2	2.9	2.7	5.60	2.33912	1.13039	42.0	41.52	305.0	760.0
3	4.2	4.1	8.30	2.84772	1.37215	52.0	51.40	305.0	760.0
4	5.8	5.6	11.40	3.33742	1.60493	58.0	57.33	305.0	760.0
5	6.5	6.4	12.90	3.55020	1.70607	62.0	61.28	305.0	760.0
Average								305.0	760.0

Linear Regression : $y = mx + b$

Slope (m)	36.049868
Intercept (b)	0.245557
R-Square (R^2)	0.991639
Correlation Coefficient (r)	0.995811

Andersen Instruments, Inc.



Calibrated By

Mr. Warakorn Charoenkhun

SMILE
Laboratory Co., Ltd.



บริษัท สไมล์ แล็บอราทอรี จำกัด

Smile Laboratory Co., Ltd.

563/1 ถนนเทอดไท แขวงบางหว้า เขตภาษีเจริญ กรุงเทพฯ 10160 โทรศัพท์ 02-227-0265 โทรสาร 02-454-0317
563/1 Thoei Thai Rd., Bangwa, Phasicharoen, Bangkok 10160 Tel. 02-227-0265 Fax. 02-454-0317

PM10 HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sampler Location	โครงการ SML-0155-02-67	Date	12 February 2024
Project Site	สถานีตรวจวัดคุณภาพอากาศอัตโนมัติ (AAQMS)	Person	Mr. Warakorn Charoenkhun

Calibration Orifice

Transfer Standard Type	Orifice	Q_{std} Slope (m)	2.10372
Calibrator Model	TE-5025A	Q_{std} Intercept (b)	-0.03890
Calibrator Serial Number	3092		

Calibration Information


Sampler Number	PM10 No.03	Motor Serial Number	1203-449	Recorder Serial Number	608
----------------	------------	---------------------	----------	------------------------	-----

Test No.	Pressure Drop Across Orifice (ΔH_2O) (in H_2O)			(A)	(X)	(I)	(Y)	Temperature ($^{\circ}K = ^{\circ}C + 273$)	Barometric Pressure (mmHg)
	Positive	Negative	ΔH_2O	$[\Delta H_2O(P_a/P_{std})(T_{std}/T_a)]^{1/2}$	$Q_{std} = (1/m)[(A-b)]$ (m^3/min)	Sample Flow Rate Indication (ft^3/min)	$IC = I[(P_a/P_{std})(T_{std}/T_a)]^{1/2}$ (ft^3/min)		
1	1.8	1.7	3.50	1.84924	0.89752	32.0	31.63	305.0	760.0
2	2.9	2.7	5.60	2.33912	1.13039	42.0	41.52	305.0	760.0
3	4.2	4.1	8.30	2.84772	1.37215	52.0	51.40	305.0	760.0
4	5.8	5.6	11.40	3.33742	1.60493	58.0	57.33	305.0	760.0
5	6.5	6.4	12.90	3.55020	1.70607	62.0	61.28	305.0	760.0
Average								305.0	760.0

Linear Regression : $y = mx + b$

Slope (m)	36.049868
Intercept (b)	0.245557
R-Square (R^2)	0.991639
Correlation Coefficient (r)	0.995811

Andersen Instruments, Inc.



Calibrated By

Mr. Warakorn Charoenkhun

SMILE
Laboratory Co., Ltd.

Analyzer Performance Test

Calibrated Date: 12 February 2024

Instruments Information

Analyzer Type: SO ₂ Analyzer	Manufacturer Thermo Environmental
Model: 43C	S/N: 250818

Calibration System


Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008	NO Conc 46.05 ppm
S/N: 705	SO ₂ Conc 46.01 ppm
ZERO AIR Generator API MODEL 701	CO Conc 4,487 ppm
S/N: 1924	Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C

Humidity: 51 %RH

Calibration Report

Status	Zero			Span		
	Reference (ppb)	Reading (ppb)	Drift (ppb)	Reference (ppb)	Reading (ppb)	Drift%
Before	0.0	0.1	0.1	400.0	394.0	-1.5
After	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :

Mr.PASAGORN SAMOL

Analyzer Performance Test

Calibrated Date: 12 February 2024

Instruments Information

Analyzer Type: SO2 Analyzer Model: M100A	Manufacturer API Environmental S/N: 1810
---	---

Calibration System


Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008 S/N: 705 ZERO AIR Generator API MODEL 701 S/N: 1924	NO Conc 55.47 PPM SO2 Conc 55.11 PPM CO Conc 4,535 PPM Cylinder number EB0129027 Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C

Humidity: 51 %RH

Calibration Report

Status	Zero			Span		
	Reference (PPB)	Reading (PPB)	Drift (PPB)	Reference (PPB)	Reading (PPB)	Drift%
Before	0.0	0.1	0.1	400.0	395.1	-1.2
After	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :

Mr.PASAGORN SAMOL

Analyzer Performance Test

Calibrated Date: 12 February 2024

Instruments Information

Analyzer Type: SO2 Analyzer Model: 100A	Manufacturer API Environmental S/N: 1618
--	---

Calibration System


Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008 S/N: 705 ZERO AIR Generator API MODEL 701 S/N: 1924	NO Conc 55.47 PPM SO2 Conc 55.11 PPM CO Conc 4,535 PPM Cylinder number EB0129027 Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C

Humidity: 51 %RH

Calibration Report

Status	Zero			Span		
	Reference (PPB)	Reading (PPB)	Drift (PPB)	Reference (PPB)	Reading (PPB)	Drift%
Before	0.0	0.1	0.1	400.0	394.6	-1.3
After	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :

Mr.PASAGORN SAMOL

Analyzer Performance Test

Calibrated Date: 12 February 2024

Instruments Information

Analyzer Type: SO2 Analyzer Model: 100A	Manufacturer API Environmental S/N: 1192
--	---

Calibration System


Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008 S/N: 705 ZERO AIR Generator API MODEL 701 S/N: 1924	NO Conc 55.47 PPM SO2 Conc 55.11 PPM CO Conc 4,535 PPM Cylinder number EB0129027 Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C

Humidity: 51 %RH

Calibration Report

Status	Zero			Span		
	Reference (PPB)	Reading (PPB)	Drift (PPB)	Reference (PPB)	Reading (PPB)	Drift%
Before	0.0	0.1	0.1	400.0	395.5	-1.1
After	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :

Mr.PASAGORN SAMOL

Analyzer Performance Test

Calibrated Date: 12 February 2024

Instruments Information

Analyzer Type: NO/NO2/NOx Analyzer Model: APNA-360	Manufacturer: Horiba Environmental S/N: 8517870112
---	---

Calibration System

Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008 S/N: 705 ZERO AIR Generator API Model 701 S/N: 1924	NO Conc 55.47 PPM SO2 Conc 55.11 PPM CO Conc 4,535 PPM Cylinder number EB0129027 Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C


Humidity: 51 %RH

Calibration Check (Before adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.1	0.0	0.1	396.5	400.0	-0.9
NOx	0.1	0.0	0.1	400.0	400.0	0.0

Calibration Check (After adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.0	0.0	0.0	400.0	400.0	0.0
NOx	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By : Mr. Pasagorn Samol

Analyzer Performance Test

Calibrated Date: 12 February 2024

Instruments Information

Analyzer Type: NO/NO ₂ /NO _x Analyzer Model: 200A	Manufacturer TELEDYNE INSTRUMENTS S/N: 2609
--	--

Calibration System

Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008 S/N: 705 ZERO AIR Generator API Model 701 S/N: 1924	NO Conc 55.47 PPM SO ₂ Conc 55.11 PPM CO Conc 4.535 PPM Cylinder number EB0129027 Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C


Humidity: 51 %RH

Calibration Check (Before adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.1	0.0	0.1	395.2	400.0	-1.2
NO _x	0.1	0.0	0.1	400.0	400.0	0.0

Calibration Check (After adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.0	0.0	0.0	400.0	400.0	0.0
NO _x	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By : Mr. Pasagorn Samol

Analyzer Performance Test

Calibrated Date: 12 February 2024

Instruments Information

Analyzer Type: NO/NO ₂ /NO _x Analyzer Model: 200A	Manufacturer API Environmental S/N: 612
--	--

Calibration System

Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008 S/N: 705 ZERO AIR Generator API Model 701 S/N: 1924	NO Conc 55.47 PPM SO ₂ Conc 55.11 PPM CO Conc 4,535 PPM Cylinder number EB0129027 Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C


Humidity: 51 %RH

Calibration Check (Before adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.1	0.0	0.1	395.8	400.0	-1.1
NO _x	0.1	0.0	0.1	400.0	400.0	0.0

Calibration Check (After adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.0	0.0	0.0	400.0	400.0	0.0
NO _x	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By : Mr. Pasagorn Samol

Analyzer Performance Test

Calibrated Date: 12 February 2024

Instruments Information

Analyzer Type: NO/NO ₂ /NO _x Analyzer Model: 200A	Manufacturer API Environmental S/N: 605
--	--

Calibration System

Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008 S/N: 705 ZERO AIR Generator API Model 701 S/N: 1924	NO Conc 55.47 PPM SO ₂ Conc 55.11 PPM CO Conc 4,535 PPM Cylinder number EB0129027 Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C


Humidity: 51 %RH

Calibration Check (Before adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.1	0.0	0.1	394.7	400.0	-1.3
NO _x	0.1	0.0	0.1	400.0	400.0	0.0

Calibration Check (After adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.0	0.0	0.0	400.0	400.0	0.0
NO _x	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By : Mr. Pasagorn Samol



บริษัท เอ็นไวร์ เซอร์วิส จำกัด
42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201
ENVIR SERVICE CO., LTD. 42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bankok 10230 Tel : 02-9435814-5 Fax : 02-9438201

บริษัท เอ็นไวร์ เซอร์วิส จำกัด

Calibration Certificate

Issued by : Calibration & Test Section : Meteorological Instruments Bureau

Date of Issue 12 February 2024

Certificate No. 123/21

Page : 1 of 2

Manufacture Yong Instruments

Type four blade helicoid propeller

Model No. 05103

Mfg Code Logger 30908233

Transmitter -

Customer ENVIR SERVICE CO., LTD.

42 Raminthra 14 yeak 9, Tha Raeng,

Bangkhen, Bangkok 10230

Calibration Condition : Temperature 25.2 °C

Barometric Pressure 1012.8 hPa

NATIONAL STANDARD WIND TUNNEL : Thermal Anemometer 642 S/N 91563

: HOOK GAGE NO 1425 : Wind Aloft Plotting Board

N.I.S.T. Test Reference Number 731/241460

: Ultrasonic Anemometer Model DA-650-3TV (sensor TR-90AH)

Serial Number 110730029 (sensor 120629586)

JAPAN QUALITY ASSURANCE ORGANIZATION

STANDARD THERMOMETER

: Theodor Friedrich : Dry No. 8390/94 Wet No. 8389/94

: Thermoschneider No. 918802

STANDARD BAROMETER

: Digital Barometer Vaisaia Type RTB220 No. V1220015

Calibrated by :



Mr. Pasagorn Samol



บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201
บริษัท เอ็นไวร์ เซอร์วิส จำกัด
ENVIRO SERVICE CO., LTD. 42 Ramintra 14 yeak 9, Tha Rang, Bangkhen, Bankok 10230 Tel : 02-9435814-5 Fax : 02-9438201

The Result of Calibration

Date of Issue 12 February 2024


Certificate No. 123/21

Page : 2 of 2

Standard Ultrasonic Anemometer m/sec	HOOK GAGE NO 1425			TESTED ANEMOMETER			
	Pressure inches	Vacuum inches	Pressure hPa	Pressure hPa	Correction hPa	Velocity m/sec	Correction m/sec
1.00	-	-	-	-	-	0.7	0.30
3.02	-	-	-	-	-	2.7	0.32
5.04	-	-	-	-	-	4.7	0.34
7.03	-	-	-	-	-	6.7	0.33
9.01	-	-	-	-	-	8.5	0.51
11.03	-	-	-	-	-	10.7	0.33
13.01	-	-	-	-	-	12.4	0.61
15.03	-	-	-	-	-	14.1	0.93
17.05	-	-	-	-	-	16.4	0.65
20.02	-	-	-	-	-	19.1	0.92

Wind Aloft Plotting Board.	
US. DEPARTMENT OF COMMERCE WEATHER BUREAU	
WIND DIRECTION	TESTED WIND DIRECTION
0	0
90	90
180	180
270	270

Calibrated by :



Mr. Pasagorn Samol



บริษัท เอ็นไวร์ เซอร์วิส จำกัด
42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201
ENVIR SERVICE CO., LTD. 42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bangkok 10230 Tel : 02-9435814-5 Fax : 02-9438201

บริษัท เอ็นไวร์ เซอร์วิส จำกัด

Calibration Certificate

Issued by : Calibration & Test Section : Meteorological Instruments Bureau

Date of Issue 12 February 2024

Certificate No. 124/21

Page : 1 of 2

Manufacture Yong Instruments

Type four blade helicoid propeller

Model No. 05103

Mfg Code Logger 30908695

Transmitter -

Customer ENVIR SERVICE CO., LTD.

42 Raminthra 14 yeak 9, Tha Raeng,

Bangkhen, Bangkok 10230

Calibration Condition : Temperature 25.2 °C Barometric Pressure 1012.8 hPa

NATIONAL STANDARD WIND TUNNEL : Thermal Anemometer 642 S/N 91563

: HOOK GAGE NO 1425 : Wind Aloft Plotting Board

N.I.S.T. Test Reference Number 731/241460

: Ultrasonic Anemometer Model DA-650-3TV (sensor TR-90AH)

Serial Number 110730029 (sensor 120629586)

JAPAN QUALITY ASSURANCE ORGANIZATION

STANDARD THERMOMETER : Theodor Friedrich : Dry No. 8390/94 Wet No. 8389/94

: Thermoschneider No. 918802

STANDARD BAROMETER : Digital Barometer Vaisaia Type RTB220 No. V1220015

Calibrated by :



Mr. Pasagorn Samol



บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201
บริษัท เอ็นไวร์ เซอร์วิส จำกัด
ENVIRO SERVICE CO., LTD. 42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bankok 10230 Tel : 02-9435814-5 Fax : 02-9438201

The Result of Calibration

Date of Issue 12 February 2024

Certificate No. 124/21

Page : 2 of 2

Standard Ultrasonic Anemometer m/sec	HOOK GAGE NO 1425			TESTED ANEMOMETER			
	Pressure inches	Vacuum inches	Pressure hPa	Pressure hPa	Correction hPa	Velocity m/sec	Correction m/sec
1.00	-	-	-	-	-	0.7	0.30
3.02	-	-	-	-	-	2.7	0.32
5.04	-	-	-	-	-	4.7	0.34
7.03	-	-	-	-	-	6.7	0.33
9.01	-	-	-	-	-	8.5	0.51
11.03	-	-	-	-	-	10.7	0.33
13.01	-	-	-	-	-	12.4	0.61
15.03	-	-	-	-	-	14.1	0.93
17.05	-	-	-	-	-	16.4	0.65
20.02	-	-	-	-	-	19.1	0.92

Wind Aloft Plotting Board. US. DEPARTMENT OF COMMERCE WEATHER BUREAU	
WIND DIRECTION	TESTED WIND DIRECTION
0	0
90	90
180	180
270	270

Calibrated by :



Mr. Pasagorn Samol



บริษัท เอ็นไวร์ เซอร์วิส จำกัด
42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201
ENVIR SERVICE CO., LTD. 42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bankok 10230 Tel : 02-9435814-5 Fax : 02-9438201

บริษัท เอ็นไวร์ เซอร์วิส จำกัด

Calibration Certificate

Issued by : Calibration & Test Section : Meteorological Instruments Bureau

Date of Issue 12 February 2024

Certificate No. 122/21

Page : 1 of 2

Manufacture Yong Instruments

Type four blade helicoid propeller

Model No. 05103

Mfg Code Logger 30908794

Transmitter -

Customer ENVIR SERVICE CO., LTD.

42 Raminthra 14 yeak 9, Tha Raeng,

Bangkhen, Bangkok 10230

Calibration Condition : Temperature 25.2 °C

Barometric Pressure 1012.8 hPa

NATIONAL STANDARD WIND TUNNEL : Thermal Anemometer 642 S/N 91563

: HOOK GAGE NO 1425 : Wind Aloft Plotting Board

N.I.S.T. Test Reference Number 731/241460

: Ultrasonic Anemometer Model DA-650-3TV (sensor TR-90AH)

Serial Number 110730029 (sensor 120629586)

JAPAN QUALITY ASSURANCE ORGANIZATION

STANDARD THERMOMETER

: Theodor Friedrich : Dry No. 8390/94 Wet No. 8389/94

: Thermoschneider No. 918802

STANDARD BAROMETER

: Digital Barometer Vaisaia Type RTB220 No. V1220015

Calibrated by :



Mr. Pasagorn Samol



บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201
บริษัท เอ็นไวร์ เซอร์วิส จำกัด
ENVIR SERVICE CO., LTD. 42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bankok 10230 Tel : 02-9435814-5 Fax : 02-9438201

The Result of Calibration

Date of Issue 12 February 2024

Certificate No. 122/21

Page : 2 of 2

Standard Ultrasonic Anemometer m/sec	HOOK GAGE NO 1425			TESTED ANEMOMETER			
	Pressure inches	Vacuum inches	Pressure hPa	Pressure hPa	Correction hPa	Velocity m/sec	Correction m/sec
1.00	-	-	-	-	-	0.7	0.30
3.02	-	-	-	-	-	2.7	0.32
5.04	-	-	-	-	-	4.5	0.54
7.03	-	-	-	-	-	6.7	0.33
9.01	-	-	-	-	-	8.5	0.51
11.03	-	-	-	-	-	10.7	0.33
13.01	-	-	-	-	-	12.4	0.61
15.03	-	-	-	-	-	14.1	0.93
17.05	-	-	-	-	-	16.4	0.65
20.02	-	-	-	-	-	19.1	0.92

Wind Aloft Plotting Board.	
US. DEPARTMENT OF COMMERCE WEATHER BUREAU	
WIND DIRECTION	TESTED WIND DIRECTION
0	0
90	90
180	180
270	270

Calibrated by :



Mr. Pasagorn Samol



บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201
บริษัท เอ็นไวร์ เซอร์วิส จำกัด 42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bangkok 10230 Tel : 02-9435814-5 Fax : 02-9438201
ENVIRO SERVICE CO., LTD.

Calibration Certificate

Issued by : Calibration & Test Section : Meteorological Instruments Bureau

Date of Issue 12 February 2024

Certificate No. 114/21

Page : 1 of 2

Manufacture Yong Instruments

Type four blade helicoid propeller

Model No. 05103

Mfg Code Logger 309020206

Transmitter -

Customer ENVIR SERVICE CO., LTD.

42 Raminthra 14 yeak 9, Tha Raeng,

Bangkhen, Bangkok 10230

Calibration Condition : Temperature 25.2 °C

Barometric Pressure 1012.8 hPa

NATIONAL STANDARD WIND TUNNEL : Thermal Anemometer 642 S/N 91563

: HOOK GAGE NO 1425 : Wind Aloft Plotting Board

N.I.S.T. Test Reference Number 731/241460

: Ultrasonic Anemometer Model DA-650-3TV (sensor TR-90AH)

Serial Number 110730029 (sensor 120629586)

JAPAN QUALITY ASSURANCE ORGANIZATION

STANDARD THERMOMETER : Theodor Friedrich : Dry No. 8390/94 Wet No. 8389/94

: Thermoschneider No. 918802

STANDARD BAROMETER : Digital Barometer Vaisala Type RTB220 No. V1220015

Calibrated by :



Mr. Pasagorn Samol



บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201
บริษัท เอ็นไวร์ เซอร์วิส จำกัด
ENVIR SERVICE CO., LTD. 42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bangkok 10230 Tel : 02-9435814-5 Fax : 02-9438201

The Result of Calibration

Date of Issue 12 February 2024

Certificate No. 114/21

Page : 2 of 2

Standard	HOOK GAGE NO 1425			TESTED ANEMOMETER			
	Pressure	Vacuum	Pressure	Pressure	Correction	Velocity	Correction
Ultrasonic Anemometer m/sec	inches	inches	hPa	hPa	hPa	m/sec	m/sec
1.00	-	-	-	-	-	0.7	0.30
3.02	-	-	-	-	-	2.7	0.32
5.04	-	-	-	-	-	4.8	0.24
7.03	-	-	-	-	-	6.7	0.33
9.01	-	-	-	-	-	8.7	0.31
11.03	-	-	-	-	-	10.7	0.33
13.01	-	-	-	-	-	12.4	0.61
15.03	-	-	-	-	-	14.1	0.93
17.05	-	-	-	-	-	16.4	0.65
20.02	-	-	-	-	-	19.1	0.92

Wind Aloft Plotting Board.	
US. DEPARTMENT OF COMMERCE WEATHER BUREAU	
WIND DIRECTION	TESTED WIND DIRECTION
0	0
90	90
180	180
270	270

Calibrated by :



Mr. Pasagorn Samol



บริษัท สไมล์ แล็บอราทอรี จำกัด

Smile Laboratory Co., Ltd.

563/1 ถนนพหลโยธิน แขวงบางหว้า เขตภาษีเจริญ กรุงเทพฯ 10160 โทรศัพท์ 02-227-0265 โทรสาร 02-454-0317
563/1 Thoe Thai Rd., Bangwa, Phasicharoen, Bangkok 10160 Tel. 02-227-0265 Fax. 02-454-0317

TSP HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sampler Location	โครงการ SML-0981-06-67	Date	11 June 2024
Project Site	สถานีตรวจวัดคุณภาพอากาศอัตโนมัติ (AAQMS)	Person	Mr. Warakorn Charoenkhun

Calibration Orifice

Transfer Standard Type	Orifice	Q_{std} Slope (m)	2.10372
Calibrator Model	TE-5025A	Q_{std} Intercept (b)	-0.03890
Calibrator Serial Number	3092		

Calibration Information


Sampler Number	TSP No.01	Motor Serial Number	1203-415	Recorder Serial Number	604
----------------	-----------	---------------------	----------	------------------------	-----

Test No.	Pressure Drop Across Orifice (ΔH_2O) (inH ₂ O)			(A)	(X)	(I)	(Y)	Temperature (°K = °C+273)	Barometric Pressure (mmHg)
	Positive	Negative	ΔH_2O	$[\Delta H_2O(P_a/P_{std})(T_{std}/T_a)]^{1/2}$	$Q_{std} = (1/m)[(A-b)]$ (m ³ /min)	Sample Flow Rate Indication (ft ³ /min)	$IC = I[(P_a/P_{std})(T_{std}/T_a)]^{1/2}$ (ft ³ /min)		
1	2.2	2.1	4.30	2.05646	0.99603	36.0	35.70	303.0	760.0
2	3.2	2.9	6.10	2.44936	1.18279	43.0	42.64	303.0	760.0
3	4.1	3.8	7.90	2.78741	1.34348	50.0	49.59	303.0	760.0
4	5.4	5.3	10.70	3.24398	1.56051	55.0	54.54	303.0	760.0
5	6.3	6.2	12.50	3.50624	1.68518	60.0	59.50	303.0	760.0
Average								303.0	760.0

Linear Regression : $y = mX + b$

Slope (m)	33.768326
Intercept (b)	2.686962
R-Square (R^2)	0.990356
Correlation Coefficient (r)	0.995166

Andersen Instruments, Inc.



Calibrated By

Mr. Warakorn Charoenkhun

SMILE
Laboratory Co., Ltd.



บริษัท สไมล์ แล็บอราทอรี จำกัด
Smile Laboratory Co., Ltd.

563/1 ถนนทองเอก แขวงบางค้อ เขตภาษีเจริญ กรุงเทพฯ 10160 โทรศัพท์ 02-227-0265 โทรสาร 02-454-0317
563/1 Thoet Thai Rd., Bangwa, Phasicharoen, Bangkok 10160 Tel. 02-227-0265 Fax. 02-454-0317

TSP HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sampler Location	โครงการ SML-0981-06-67	Date	11 June 2024
Project Site	โรงเรียนบ้านช่องกอม่วง	Person	Mr. Warakorn Charoenkhun

Calibration Orifice

Transfer Standard Type	Orifice	Q_{std} Slope (m)	2.10372
Calibrator Model	TE-5025A	Q_{std} Intercept (b)	-0.03890
Calibrator Serial Number	3092		

Calibration Information


Sampler Number	TSP No.02	Motor Serial Number	1203-416	Recorder Serial Number	605
----------------	-----------	---------------------	----------	------------------------	-----

Test No.	Pressure Drop Across Orifice (ΔH_2O) (inH ₂ O)			(A)	(X)	(I)	(Y)	Temperature (°K = °C+273)	Barometric Pressure (mmHg)
	Positive	Negative	ΔH_2O	$[\Delta H_2O(P_a/P_{std})(T_{std}/T_a)]^{1/2}$	$Q_{std} = (1/m)[(A-b)]$ (m ³ /min)	Sample Flow Rate Indication (ft ³ /min)	$IC = I[(P_a/P_{std})(T_{std}/T_a)]^{1/2}$ (ft ³ /min)		
1	2.2	2.1	4.30	2.05646	0.99603	36.0	35.70	303.0	760.0
2	3.1	3.0	6.10	2.44936	1.18279	43.0	42.64	303.0	760.0
3	4.3	4.2	8.50	2.89132	1.39288	50.0	49.59	303.0	760.0
4	5.2	5.1	10.30	3.18277	1.53142	55.0	54.54	303.0	760.0
5	6.1	6.0	12.10	3.44969	1.65829	60.0	59.50	303.0	760.0
Average								303.0	760.0

Linear Regression : $y = mX + b$

Slope (m)	35.494588
Intercept (b)	0.397050
R-Square (R^2)	0.999329
Correlation Coefficient (r)	0.999664

Andersen Instruments, Inc.



Calibrated By

Mr. Warakorn Charoenkhun

SMILE
Laboratory Co., Ltd.



บริษัท สไมล์ แล็บอราทอรี จำกัด
Smile Laboratory Co., Ltd.

563/1 ถนนทองเอก แขวงบางค้อว เขตภาษีเจริญ กรุงเทพฯ 10160 โทรศัพท์ 02-227-0265 โทรสาร 02-454-0317
563/1 Thoei Thai Rd., Bangwa, Phasicharoen, Bangkok 10160 Tel. 02-227-0265 Fax. 02-454-0317

TSP HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sampler Location	โครงการ SML-0981-06-67	Date	11 June 2024
Project Site	วัดวังทอง	Person	Mr. Warakorn Charoenkhun

Calibration Orifice

Transfer Standard Type	Orifice	Q _{std} Slope (m)	2.10372
Calibrator Model	TE-5025A	Q _{std} Intercept (b)	-0.03890
Calibrator Serial Number	3092		

Calibration Information


Sampler Number	TSP No.03	Motor Serial Number	1203-426	Recorder Serial Number	606
----------------	-----------	---------------------	----------	------------------------	-----

Test No.	Pressure Drop Across Orifice (ΔH ₂ O) (inH ₂ O)			(A)	(X)	(I)	(Y)	Temperature (°K = °C+273)	Barometric Pressure (mmHg)
	Positive	Negative	ΔH ₂ O	$[\Delta H_2O(P_a/P_{std})(T_{std}/T_a)]^{1/2}$	Q _{std} = (1/m)[(A-b)] (m ³ /min)	Sample Flow Rate Indication (ft ³ /min)	IC = $I[(P_a/P_{std})(T_{std}/T_a)]^{1/2}$ (ft ³ /min)		
1	2.0	1.9	3.90	1.95848	0.94945	36.0	35.70	303.0	760.0
2	3.3	3.1	6.40	2.50886	1.21107	43.0	42.64	303.0	760.0
3	4.4	4.3	8.70	2.92514	1.40895	50.0	49.59	303.0	760.0
4	5.5	5.4	10.90	3.27416	1.57486	55.0	54.54	303.0	760.0
5	6.3	6.1	12.40	3.49219	1.67850	60.0	59.50	303.0	760.0
Average								303.0	760.0

Linear Regression : y= mX + b

Slope (m)	32.249636
Intercept (b)	4.388908
R-Square (R ²)	0.992855
Correlation Coefficient (r)	0.996421

Andersen Instruments, Inc.



Calibrated By

Mr. Warakorn Charoenkhun

SMILE
Laboratory Co., Ltd.



บริษัท สไมล์ แล็บอราทอรี จำกัด

Smile Laboratory Co., Ltd.

563/1 ถนนทองเอก แขวงบางหว้า เขตภาษีเจริญ กรุงเทพฯ 10160 โทรศัพท์ 02-227-0265 โทรสาร 02-454-0317

563/1 Thoei Thai Rd., Bangwa, Phasicharoen, Bangkok 10160 Tel. 02-227-0265 Fax. 02-454-0317

TSP HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sampler Location	โครงการ SML-0981-06-67	Date	11 June 2024
Project Site	ชุมชนหมู่ 4 บ้านสันป่าฝ้าย	Person	Mr. Warakorn Charoenkhun

Calibration Orifice

Transfer Standard Type	Orifice	Q_{std} Slope (m)	2.10372
Calibrator Model	TE-5025A	Q_{std} Intercept (b)	-0.03890
Calibrator Serial Number	3092		

Calibration Information


Sampler Number	TSP No.04	Motor Serial Number	1203-422	Recorder Serial Number	607
----------------	-----------	---------------------	----------	------------------------	-----

Test No.	Pressure Drop Across Orifice (ΔH_2O) (inH ₂ O)			(A)	(X)	(I)	(Y)	Temperature (°K = °C+273)	Barometric Pressure (mmHg)
	Positive	Negative	ΔH_2O	$[\Delta H_2O(P_s/P_{std})(T_{std}/T_a)]^{1/2}$	$Q_{std} = (1/m)[(A-b)]$ (m ³ /min)	Sample Flow Rate Indication (ft ³ /min)	$IC = I[(P_s/P_{std})(T_{std}/T_a)]^{1/2}$ (ft ³ /min)		
1	2.2	2.1	4.30	2.05646	0.99603	36.0	35.70	303.0	760.0
2	3.1	3.0	6.10	2.44936	1.18279	43.0	42.64	303.0	760.0
3	4.3	4.2	8.50	2.89132	1.39288	50.0	49.59	303.0	760.0
4	5.2	5.1	10.30	3.18277	1.53142	55.0	54.54	303.0	760.0
5	6.1	6.0	12.10	3.44969	1.65829	60.0	59.50	303.0	760.0
Average								303.0	760.0

Linear Regression : $y = mX + b$

Slope (m)	35.494588
Intercept (b)	0.397050
R-Square (R^2)	0.999329
Correlation Coefficient (r)	0.999664

Andersen Instruments, Inc.



Calibrated By

Mr. Warakorn Charoenkhun

SMILE
Laboratory Co., Ltd.



บริษัท สไมล์ แล็บอราทอรี จำกัด
Smile Laboratory Co., Ltd.

563/1 ถนนทองเอก แขวงบางท่าว เขตปทุมธานี กรุงเทพมหานคร 10160 โทรศัพท์ 02-227-0265 โทรสาร 02-454-0317
563/1 Thoeet Thai Rd., Bangwa, Phasicharoen, Bangkok 10160 Tel. 02-227-0265 Fax. 02-454-0317

PM10 HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sampler Location	โครงการ SML-0981-06-67	Date	11 June 2024
Project Site	สถานีตรวจวัดคุณภาพอากาศอัตโนมัติ (AAQMS)	Person	Mr. Warakorn Charoenkhun

Calibration Orifice

Transfer Standard Type	Orifice	Q_{std} Slope (m)	2.10372
Calibrator Model	TE-5025A	Q_{std} Intercept (b)	-0.03890
Calibrator Serial Number	3092		

Calibration Information


Sampler Number	PM10 No.01	Motor Serial Number	1203-449	Recorder Serial Number	658
----------------	------------	---------------------	----------	------------------------	-----

Test No.	Pressure Drop Across Orifice (ΔH_2O) (inH ₂ O)			(A)	(X)	(I)	(Y)	Temperature (°K = °C+273)	Barometric Pressure (mmHg)
	Positive	Negative	ΔH_2O	$[\Delta H_2O(P_a/P_{std})(T_{std}/T_a)]^{1/2}$	$Q_{std} = (1/m)[(A-b)]$ (m ³ /min)	Sample Flow Rate Indication (ft ³ /min)	$IC = I[(P_a/P_{std})(T_{std}/T_a)]^{1/2}$ (ft ³ /min)		
1	1.8	1.7	3.50	1.84924	0.89752	32.0	31.63	305.0	760.0
2	2.9	2.7	5.60	2.33912	1.13039	42.0	41.52	305.0	760.0
3	4.2	4.1	8.30	2.84772	1.37215	52.0	51.40	305.0	760.0
4	5.8	5.6	11.40	3.33742	1.60493	58.0	57.33	305.0	760.0
5	6.5	6.4	12.90	3.55020	1.70607	62.0	61.28	305.0	760.0
Average								305.0	760.0

Linear Regression : $y = mX + b$

Slope (m)	36.049868
Intercept (b)	0.245557
R-Square (R^2)	0.991639
Correlation Coefficient (r)	0.995811

Andersen Instruments, Inc.



Calibrated By

Mr. Warakorn Charoenkhun



บริษัท สไมล์ แล็บอราทอรี จำกัด

Smile Laboratory Co., Ltd.

563/1 ถนนพหลโยธิน แขวงบางกะปิ เขตปทุมธานี กรุงเทพมหานคร 10160 โทรศัพท์ 02-227-0265 โทรสาร 02-454-0317
563/1 Thot Thai Rd., Bangwa, Phasicharoen, Bangkok 10160 Tel. 02-227-0265 Fax. 02-454-0317

PM10 HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sampler Location	โครงการ SML-0981-06-67	Date	11 June 2024
Project Site	โรงเรียนบ้านหนองกอมม่วง	Person	Mr. Warakorn Charoenkhun

Calibration Orifice

Transfer Standard Type	Orifice	Q_{std} Slope (m)	2.10372
Calibrator Model	TE-5025A	Q_{std} Intercept (b)	-0.03890
Calibrator Serial Number	3092		

Calibration Information


Sampler Number	PM10 No.02	Motor Serial Number	1203-450	Recorder Serial Number	659
----------------	------------	---------------------	----------	------------------------	-----

Test No.	Pressure Drop Across Orifice (ΔH_2O) (inH ₂ O)			(A)	(X)	(I)	(Y)	Temperature (°K = °C+273)	Barometric Pressure (mmHg)
	Positive	Negative	ΔH_2O	$[\Delta H_2O(P_a/P_{std})(T_{std}/T_a)]^{1/2}$	$Q_{std} = (1/m)[(A-b)]$ (m ³ /min)	Sample Flow Rate Indication (ft ³ /min)	$IC = I[(P_a/P_{std})(T_{std}/T_a)]^{1/2}$ (ft ³ /min)		
1	1.8	1.7	3.50	1.84924	0.89752	32.0	31.63	305.0	760.0
2	2.9	2.7	5.60	2.33912	1.13039	42.0	41.52	305.0	760.0
3	4.2	4.1	8.30	2.84772	1.37215	52.0	51.40	305.0	760.0
4	5.8	5.6	11.40	3.33742	1.60493	58.0	57.33	305.0	760.0
5	6.5	6.4	12.90	3.55020	1.70607	62.0	61.28	305.0	760.0
Average								305.0	760.0

Linear Regression : $y = mX + b$

Slope (m)	36.049868
Intercept (b)	0.245557
R-Square (R^2)	0.991639
Correlation Coefficient (r)	0.995811

Andersen Instruments, Inc.



Calibrated By

Mr. Warakorn Charoenkhun

SMILE
Laboratory Co., Ltd.



บริษัท สไมล์ แล็บอราทอรี จำกัด

Smile Laboratory Co., Ltd.

563/1 ถนนพหลโยธิน แขวงบางหว้า เขตภาษีเจริญ กรุงเทพมหานคร 10160 โทรศัพท์ 02-227-0265 โทรสาร 02-454-0317
563/1 Thoe Thai Rd., Bangwa, Phasicharoen, Bangkok 10160 Tel. 02-227-0265 Fax. 02-454-0317

PM10 HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sampler Location	โครงการ SML-0981-06-67	Date	11 June 2024
Project Site	วัดวังทอง	Person	Mr. Warakorn Charoenkhun

Calibration Orifice

Transfer Standard Type	Orifice	Q_{std} Slope (m)	2.10372
Calibrator Model	TE-5025A	Q_{std} Intercept (b)	-0.03890
Calibrator Serial Number	3092		

Calibration Information


Sampler Number	PM10 No.03	Motor Serial Number	1203-440	Recorder Serial Number	660
----------------	------------	---------------------	----------	------------------------	-----

Test No.	Pressure Drop Across Orifice (ΔH_2O) (in H_2O)			(A)	(X)	(I)	(Y)	Temperature ($^{\circ}K = ^{\circ}C + 273$)	Barometric Pressure (mmHg)
	Positive	Negative	ΔH_2O	$[\Delta H_2O(P_a/P_{std})(T_{std}/T_a)]^{1/2}$	$Q_{std} = (1/m)[(A-b)]$ (m^3/min)	Sample Flow Rate Indication (ft^3/min)	$IC = I[(P_a/P_{std})(T_{std}/T_a)]^{1/2}$ (ft^3/min)		
1	1.9	1.8	3.70	1.90134	0.92229	32.0	31.63	305.0	760.0
2	3.1	3.0	6.10	2.44131	1.17896	42.0	41.52	305.0	760.0
3	4.2	4.0	8.20	2.83051	1.36397	52.0	51.40	305.0	760.0
4	5.7	5.5	11.20	3.30801	1.59095	58.0	57.33	305.0	760.0
5	6.1	6.0	12.10	3.43836	1.65291	62.0	61.28	305.0	760.0
Average								305.0	760.0

Linear Regression : $y = mX + b$

Slope (m)	39.949045
Intercept (b)	4.972153
R-Square (R^2)	0.990397
Correlation Coefficient (r)	0.995187

Andersen Instruments, Inc.



Calibrated By

Mr. Warakorn Charoenkhun

SMILE
Laboratory Co., Ltd.



บริษัท สไมล์ แล็บอราทอรี จำกัด

Smile Laboratory Co., Ltd.

563/1 ถนนทองเอก แขวงบางท่าว เขตป้อมปราบศัตรูพ่าย กรุงเทพมหานคร 10160 โทรศัพท์ 02-227-0265 โทรสาร 02-454-0317
563/1 Thoet Thai Rd., Bangwa, Phasicharoen, Bangkok 10160 Tel. 02-227-0265 Fax. 02-454-0317

PM10 HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Site Information

Sampler Location	โครงการ SML-0981-06-67	Date	11 June 2024
Project Site	ชุมชนหมู่บ้านสันป่าฝ้าย	Person	Mr. Warakorn Charoenkhun

Calibration Orifice

Transfer Standard Type	Orifice	Q_{std} Slope (m)	2.10372
Calibrator Model	TE-5025A	Q_{std} Intercept (b)	-0.03890
Calibrator Serial Number	3092		

Calibration Information


Sampler Number	PM10 No.04	Motor Serial Number	1203-444	Recorder Serial Number	661
----------------	------------	---------------------	----------	------------------------	-----

Test No.	Pressure Drop Across Orifice (ΔH_2O) (inH ₂ O)			(A)	(X)	(I)	(Y)	Temperature (°K = °C+273)	Barometric Pressure (mmHg)
	Positive	Negative	ΔH_2O	$[\Delta H_2O(P_a/P_{std})(T_{std}/T_a)]^{1/2}$	$Q_{std} = (1/m)[(A-b)]$ (m ³ /min)	Sample Flow Rate Indication (ft ³ /min)	$IC = [(P_a/P_{std})(T_{std}/T_a)]^{1/2}$ (ft ³ /min)		
1	1.8	1.7	3.50	1.84924	0.89752	32.0	31.63	305.0	760.0
2	2.9	2.7	5.60	2.33912	1.13039	42.0	41.52	305.0	760.0
3	4.2	4.1	8.30	2.84772	1.37215	52.0	51.40	305.0	760.0
4	5.8	5.6	11.40	3.33742	1.60493	58.0	57.33	305.0	760.0
5	6.5	6.4	12.90	3.55020	1.70607	62.0	61.28	305.0	760.0
Average								305.0	760.0

Linear Regression : $y = mX + b$

Slope (m)	36.049868
Intercept (b)	0.245557
R-Square (R^2)	0.991639
Correlation Coefficient (r)	0.995811

Andersen Instruments, Inc.



Calibrated By

Mr. Warakorn Charoenkhun

SMILE
Laboratory Co., Ltd.

Analyzer Performance Test

Calibrated Date: 11 June 2024

Instruments Information

Analyzer Type: SO ₂ Analyzer Model: 100A	Manufacturer API Environmental S/N: 1192
--	---

Calibration System


Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008 S/N: 705 ZERO AIR Generator API MODEL 701 S/N: 1924	NO Conc 46.05 ppm SO ₂ Conc 46.01 ppm CO Conc 4,487 ppm Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C

Humidity: 51 %RH

Calibration Report

Status	Zero			Span		
	Reference (ppb)	Reading (ppb)	Drift (ppb)	Reference (ppb)	Reading (ppb)	Drift%
Before	0.0	0.1	0.1	400.0	394.0	-1.5
After	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :

Mr.PASAGORN SAMOL

Analyzer Performance Test

Calibrated Date: 11 June 2024

Instruments Information

Analyzer Type: SO₂ Analyzer	Manufacturer API Environmental
Model: 100A	S/N: 1618

Calibration System


Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008	NO Conc 46.05 ppm
S/N: 705	SO ₂ Conc 46.01 ppm
ZERO AIR Generator API MODEL 701	CO Conc 4,487 ppm
S/N: 1924	Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C

Humidity: 51 %RH

Calibration Report

Status	Zero			Span		
	Reference (ppb)	Reading (ppb)	Drift (ppb)	Reference (ppb)	Reading (ppb)	Drift%
Before	0.0	0.1	0.1	400.0	394.0	-1.5
After	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :

Mr.PASAGORN SAMOL



บริษัท เอ็นไวร์ เซอร์วิส จำกัด
ENVIR SERVICE CO., LTD.

บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201

42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bankok 10230 Tel : 02-9435814-5 Fax : 02-9438201

Analyzer Performance Test

Calibrated Date: 11 June 2024

Instruments Information

Analyzer Type: SO ₂ Analyzer Model: 100A	Manufacturer API Environmental S/N: 1810
--	---

Calibration System


Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008 S/N: 705 ZERO AIR Generator API MODEL 701 S/N: 1924	NO Conc 46.05 ppm SO ₂ Conc 46.01 ppm CO Conc 4,487 ppm Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C

Humidity: 51 %RH

Calibration Report

Status	Zero			Span		
	Reference (ppb)	Reading (ppb)	Drift (ppb)	Reference (ppb)	Reading (ppb)	Drift%
Before	0.0	0.1	0.1	400.0	394.0	-1.5
After	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :



Mr.PASAGORN SAMOL

Analyzer Performance Test

Calibrated Date: 11 June 2024

Instruments Information

Analyzer Type: SO ₂ Analyzer Model: 43C	Manufacturer Thermo Environmental S/N: 250818
---	--

Calibration System


Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008 S/N: 705 ZERO AIR Generator API MODEL 701 S/N: 1924	NO Conc 46.05 ppm SO ₂ Conc 46.01 ppm CO Conc 4,487 ppm Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C

Humidity: 51 %RH

Calibration Report

Status	Zero			Span		
	Reference (ppb)	Reading (ppb)	Drift (ppb)	Reference (ppb)	Reading (ppb)	Drift%
Before	0.0	0.1	0.1	400.0	394.0	-1.5
After	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :

Mr.PASAGORN SAMOL



บริษัท เอ็นไวร์ เซอร์วิส จำกัด
ENVIR SERVICE CO., LTD.

บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201

42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bankok 10230 Tel : 02-9435814-5 Fax : 02-9438201

Analyzer Performance Test

Calibrated Date: 11 June 2024

Instruments Information

Analyzer Type: NO/NO ₂ /NO _x Analyzer	Manufacturer API Environmental
Model: 200A	S/N: 605

Calibration System

Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008	NO Conc 46.05 ppm
S/N: 705	SO ₂ Conc 46.01 ppm
ZERO AIR Generator API Model 701	CO Conc 4,487 ppm
S/N: 1924	Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C


Humidity: 51 %RH

Calibration Check (Before adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.1	0.0	0.1	393.4	400.0	-1.7
NO _x	0.1	0.0	0.1	400.0	400.0	0.0

Calibration Check (After adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.0	0.0	0.0	400.0	400.0	0.0
NO _x	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :

Mr. Pasagorn Samol

Analyzer Performance Test

Calibrated Date: 11 June 2024

Instruments Information

Analyzer Type: NO/NO ₂ /NO _x Analyzer	Manufacturer API Environmental
Model: 200A	S/N: 612

Calibration System

Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008	NO Conc 46.05 ppm
S/N: 705	SO ₂ Conc 46.01 ppm
ZERO AIR Generator API Model 701	CO Conc 4,487 ppm
S/N: 1924	Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C


Humidity: 51 %RH

Calibration Check (Before adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.1	0.0	0.1	393.4	400.0	-1.7
NO _x	0.1	0.0	0.1	400.0	400.0	0.0

Calibration Check (After adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.0	0.0	0.0	400.0	400.0	0.0
NO _x	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :

Mr. Pasagorn Samol

Analyzer Performance Test

Calibrated Date: 11 June 2024

Instruments Information

Analyzer Type: NO/NO ₂ /NO _x Analyzer	Manufacturer: TELEDYNE INSTRUMENTS
Model: 200A	S/N: 2609

Calibration System

Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008	NO Conc 46.05 ppm
S/N: 705	SO ₂ Conc 46.01 ppm
ZERO AIR Generator API Model 701	CO Conc 4,487 ppm
S/N: 1924	Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C


Humidity: 51 %RH

Calibration Check (Before adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.1	0.0	0.1	393.4	400.0	-1.7
NO _x	0.1	0.0	0.1	400.0	400.0	0.0

Calibration Check (After adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.0	0.0	0.0	400.0	400.0	0.0
NO _x	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :

Mr. Pasagorn Samol

Analyzer Performance Test

Calibrated Date: 11 June 2024

Instruments Information

Analyzer Type: NO/NO ₂ /NO _x Analyzer	Manufacturer Horiba Environmental
Model: APNA-360	S/N: 8517870112

Calibration System

Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008	NO Conc 46.05 ppm
S/N: 705	SO ₂ Conc 46.01 ppm
ZERO AIR Generator API Model 701	CO Conc 4,487 ppm
S/N: 1924	Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C


Humidity: 51 %RH

Calibration Check (Before adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.1	0.0	0.1	393.4	400.0	-1.7
NO _x	0.1	0.0	0.1	400.0	400.0	0.0

Calibration Check (After adjust)

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.0	0.0	0.0	400.0	400.0	0.0
NO _x	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :

Mr. Pasagorn Samol

Calibration Certificate

Issued by : Calibration & Test Section : Meteorological Instruments Bureau

Date of Issue 11 June 2024

Certificate No. 114/21

Page : 1 of 2

Manufacture Yong Instruments

Type four blade helicoid propeller

Model No. 05103

Mfg Code Logger 309020206

Transmitter -

Customer ENVIR SERVICE CO., LTD.

42 Raminthra 14 yeak 9, Tha Raeng,

Bangkhen, Bangkok 10230

Calibration Condition : Temperature 25.2 °C

Barometric Pressure 1012.8 hPa

NATIONAL STANDARD WIND TUNNEL : Thermal Anemometer 642 S/N 91563

: HOOK GAGE NO 1425 : Wind Aloft Plotting Board

N.I.S.T. Test Reference Number 731/241460

: Ultrasonic Anemometer Model DA-650-3TV (sensor TR-90AH)

Serial Number 110730029 (sensor 120629586)

JAPAN QUALITY ASSURANCE ORGANIZATION

STANDARD THERMOMETER : Theodor Friedrich : Dry No. 8390/94 Wet No. 8389/94

: Thermoschneider No. 918802

STANDARD BAROMETER : Digital Barometer Vaisala Type RTB220 No. V1220015

Calibrated by :



Mr. Pasagorn Samol

The Result of Calibration

Date of Issue 11 June 2024

Certificate No. 114/21

Page : 2 of 2

Standard	HOOK GAGE NO 1425			TESTED ANEMOMETER			
	Pressure	Vacuum	Pressure	Pressure	Correction	Velocity	Correction
Ultrasonic Anemometer m/sec	inches	inches	hPa	hPa	hPa	m/sec	m/sec
1.00	-	-	-	-	-	0.7	0.30
3.02	-	-	-	-	-	2.7	0.32
5.04	-	-	-	-	-	4.8	0.24
7.03	-	-	-	-	-	6.7	0.33
9.01	-	-	-	-	-	8.7	0.31
11.03	-	-	-	-	-	10.7	0.33
13.01	-	-	-	-	-	12.4	0.61
15.03	-	-	-	-	-	14.1	0.93
17.05	-	-	-	-	-	16.4	0.65
20.02	-	-	-	-	-	19.1	0.92

Wind Aloft Plotting Board.	
US. DEPARTMENT OF COMMERCE WEATHER BUREAU	
WIND DIRECTION	TESTED WIND DIRECTION
0	0
90	90
180	180
270	270

Calibrated by :



Mr. Pasagorn Samol



บริษัท เอ็นไวร์ เซอร์วิส จำกัด
ENVIR SERVICE CO., LTD.

บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201
42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bangkok 10230 Tel : 02-9435814-5 Fax : 02-9438201

Calibration Certificate

Issued by : Calibration & Test Section : Meteorological Instruments Bureau

Date of Issue 11 June 2024

Certificate No. 114/21

Page : 1 of 2

Manufacture Yong Instruments

Type four blade helicoid propeller

Model No. 05103

Mfg Code Logger 30908233

Transmitter -

Customer ENVIR SERVICE CO., LTD.

42 Raminthra 14 yeak 9, Tha Raeng,

Bangkhen, Bangkok 10230

Calibration Condition : Temperature 25.2 °C

Barometric Pressure 1012.8 hPa

NATIONAL STANDARD WIND TUNNEL : Thermal Anemometer 642 S/N 91563

: HOOK GAGE NO 1425 : Wind Aloft Plotting Board

N.I.S.T. Test Reference Number 731/241460

: Ultrasonic Anemometer Model DA-650-3TV (sensor TR-90AH)

Serial Number 110730029 (sensor 120629586)

JAPAN QUALITY ASSURANCE ORGANIZATION

STANDARD THERMOMETER : Theodor Friedrich : Dry No. 8390/94 Wet No. 8389/94

: Thermoschneider No. 918802

STANDARD BAROMETER : Digital Barometer Vaisala Type RTB220 No. V1220015

Calibrated by :



Mr. Pasagorn Samol

The Result of Calibration

Date of Issue 11 June 2024

Certificate No. 114/21

Page : 2 of 2

Standard	HOOK GAGE NO 1425			TESTED ANEMOMETER			
	Pressure	Vacuum	Pressure	Pressure	Correction	Velocity	Correction
Ultrasonic Anemometer m/sec	inches	inches	hPa	hPa	hPa	m/sec	m/sec
1.00	-	-	-	-	-	0.7	0.30
3.02	-	-	-	-	-	2.7	0.32
5.04	-	-	-	-	-	4.8	0.24
7.03	-	-	-	-	-	6.7	0.33
9.01	-	-	-	-	-	8.7	0.31
11.03	-	-	-	-	-	10.7	0.33
13.01	-	-	-	-	-	12.4	0.61
15.03	-	-	-	-	-	14.1	0.93
17.05	-	-	-	-	-	16.4	0.65
20.02	-	-	-	-	-	19.1	0.92

Wind Aloft Plotting Board.	
US. DEPARTMENT OF COMMERCE WEATHER BUREAU	
WIND DIRECTION	TESTED WIND DIRECTION
0	0
90	90
180	180
270	270

Calibrated by :



Mr. Pasagorn Samol



บริษัท เอ็นไวร์ เซอร์วิส จำกัด
ENVIr SERVICE CO., LTD.

บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201
42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bankok 10230 Tel : 02-9435814-5 Fax : 02-9438201

Calibration Certificate

Issued by : Calibration & Test Section : Meteorological Instruments Bureau

Date of Issue 11 June 2024

Certificate No. 114/21

Page : 1 of 2

Manufacture Yong Instruments

Type four blade helicoid propeller

Model No. 05103

Mfg Code Logger 30908695

Transmitter -

Customer ENVIR SERVICE CO., LTD.

42 Raminthra 14 yeak 9, Tha Raeng,

Bangkhen, Bangkok 10230

Calibration Condition : Temperature 25.2 °C

Barometric Pressure 1012.8 hPa

NATIONAL STANDARD WIND TUNNEL : Thermal Anemometer 642 S/N 91563

: HOOK GAGE NO 1425 : Wind Aloft Plotting Board

N.I.S.T. Test Reference Number 731/241460

: Ultrasonic Anemometer Model DA-650-3TV (sensor TR-90AH)

Serial Number 110730029 (sensor 120629586)

JAPAN QUALITY ASSURANCE ORGANIZATION

STANDARD THERMOMETER : Theodor Friedrich : Dry No. 8390/94 Wet No. 8389/94

: Thermoschneider No. 918802

STANDARD BAROMETER : Digital Barometer Vaisaia Type RTB220 No. V1220015

Calibrated by :



Mr. Pasagorn Samol

The Result of Calibration

Date of Issue 11 June 2024

Certificate No. 114/21

Page : 2 of 2

Standard	HOOK GAGE NO 1425			TESTED ANEMOMETER			
	Pressure	Vacuum	Pressure	Pressure	Correction	Velocity	Correction
Ultrasonic Anemometer m/sec	inches	inches	hPa	hPa	hPa	m/sec	m/sec
1.00	-	-	-	-	-	0.7	0.30
3.02	-	-	-	-	-	2.7	0.32
5.04	-	-	-	-	-	4.8	0.24
7.03	-	-	-	-	-	6.7	0.33
9.01	-	-	-	-	-	8.7	0.31
11.03	-	-	-	-	-	10.7	0.33
13.01	-	-	-	-	-	12.4	0.61
15.03	-	-	-	-	-	14.1	0.93
17.05	-	-	-	-	-	16.4	0.65
20.02	-	-	-	-	-	19.1	0.92

Wind Aloft Plotting Board.	
US. DEPARTMENT OF COMMERCE WEATHER BUREAU	
WIND DIRECTION	TESTED WIND DIRECTION
0	0
90	90
180	180
270	270

Calibrated by :



Mr. Pasagorn Samol



บริษัท เอ็นไวร์ เซอร์วิส จำกัด
ENVIRO SERVICE CO., LTD.

บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201

42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bankok 10230 Tel : 02-9435814-5 Fax : 02-9438201

Calibration Certificate

Issued by : Calibration & Test Section : Meteorological Instruments Bureau

Date of Issue 11 June 2024

Certificate No. 114/21

Page : 1 of 2

Manufacture Yong Instruments

Type four blade helicoid propeller

Model No. 05103

Mfg Code Logger 30908794

Transmitter -

Customer ENVIR SERVICE CO., LTD.

42 Raminthra 14 yeak 9, Tha Raeng,

Bangkhen, Bangkok 10230

Calibration Condition : Temperature 25.2 °C

Barometric Pressure 1012.8 hPa

NATIONAL STANDARD WIND TUNNEL : Thermal Anemometer 642 S/N 91563

: HOOK GAGE NO 1425 : Wind Aloft Plotting Board

N.I.S.T. Test Reference Number 731/241460

: Ultrasonic Anemometer Model DA-650-3TV (sensor TR-90AH)

Serial Number 110730029 (sensor 120629586)

JAPAN QUALITY ASSURANCE ORGANIZATION

STANDARD THERMOMETER : Theodor Friedrich : Dry No. 8390/94 Wet No. 8389/94

: Thermoschneider No. 918802

STANDARD BAROMETER : Digital Barometer Vaisala Type RTB220 No. V1220015

Calibrated by :

Mr. Pasagorn Samol

The Result of Calibration

Date of Issue 11 June 2024

Certificate No. 114/21

Page : 2 of 2

Standard	HOOK GAGE NO 1425			TESTED ANEMOMETER			
	Pressure	Vacuum	Pressure	Pressure	Correction	Velocity	Correction
Ultrasonic Anemometer m/sec	inches	inches	hPa	hPa	hPa	m/sec	m/sec
1.00	-	-	-	-	-	0.7	0.30
3.02	-	-	-	-	-	2.7	0.32
5.04	-	-	-	-	-	4.8	0.24
7.03	-	-	-	-	-	6.7	0.33
9.01	-	-	-	-	-	8.7	0.31
11.03	-	-	-	-	-	10.7	0.33
13.01	-	-	-	-	-	12.4	0.61
15.03	-	-	-	-	-	14.1	0.93
17.05	-	-	-	-	-	16.4	0.65
20.02	-	-	-	-	-	19.1	0.92

Wind Aloft Plotting Board.	
US. DEPARTMENT OF COMMERCE WEATHER BUREAU	
WIND DIRECTION	TESTED WIND DIRECTION
0	0
90	90
180	180
270	270

Calibrated by :



Mr. Pasagorn Samol