

ภาคผนวก ช

เอกสารสอบเทียบเครื่องมือที่ใช้ในการวิเคราะห์



CERTIFICATE OF CALIBRATION

Certificate No.: C0-1808005/23

Page 1 of total 4 pages

Customer WATER ANALYSIS CENTER CO., LTD.
1/94 Moo 5, T.Kanham,
A.Uthai, Ayutthaya 13210

Equipment pH Meter
Manufacturer METTLER TOLEDO
Serial No. B321527211
Description Range : 0 - 14 pH, Resolution : 0.01 pH

Environmental Conditions
Ambient Temperature: (20 ± 2) °C
Relative Humidity: (50 ± 10) %
Atmospheric Pressure: -

Calibration Location Jayhawks Laboratory (CL&GL)

Received Date 18 August 2023

Calibration Date 18 August 2023

Date of Issue 21 August 2023

Condition of Artifacts Used conditions but can be calibrated

Checked by 
Act as Technical Manager
Approved by 
Representative of Managing Director
(Dr. Ekachai Putitiwong)

- (Krisyos K.) (Sakda Y.)
- (Pariphan K.) (Onnapa P.)
- (Pongsak H.) (Nitiphong K.)
- (Kanung C.) (Nonthachai K.)
- (Prampong P.) (Noppol P.)

Measurement Results:

1. Function Simulated pH Meter

Standard Applied (mV)	Nominal Value (pH)	UUC Reading		Uncertainty (± mV)
		pH	mV	
177.48	4.00	4.01	177.4	0.060
0.00	7.00	7.00	0.0	0.060
-177.48	10.00	10.01	-177.4	0.060

UUC : Unit Under Calibration

Note : Adjust Curve to simulate pH (4,7,10)

This calibration certificate shall not be reproduced other than in full except with the prior written approval of the Thai Heart Calibration Co., Ltd.

FB-169

REV.02 02/24/21

FE-169

Page 2 of total 4 pages

Certificate No.: C0-1808005/23

Reference Method:
- The calibration method used was CP-178 based on an in-house method.

- This certificate can be traceable to the national standards, which is realized the shown measurement units according to the International System of Units (SI Units).

Reference Standard:	
Type	pH Value
pH Standard Solution	4.01
	7.01
	10.01

Reference Standard:	
Type	Model
Documenting Process Calibrator	754
Digital Thermometer with Sensor	1523 / 5622

Remark: This certificate is traceable to the International System of Unit (SI Unit) through:

- NIMT, National Institute of Metrology (Thailand),

- THC, Thai Heart Calibration Co., Ltd.

Calibrated by Kittipong REV.02 02/24/21



Certificate No.: C0-1808005/23

Page 3 of total 4 pages

Measurement Results (Cont.):

2. Calibration of pH Electrode (Serial No.: 3222623)

pH Standard Solution (pH)	Measured Value		Uncertainty (\pm pH)
	(pH)	(mV)	
4.01	4.01	180.0	0.013
7.01	7.00	4.0	0.013
10.01	10.01	-172.0	0.013

Note : Adjust Curve to Buffer Solution pH (4,7,10)
Temperature stability of micro bath : $25 \pm 0.2^\circ\text{C}$

The above reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor $k = 2.00$, providing a level of confidence approximately 95%.

Certificate No.: C0-1808005/23

Page 4 of total 4 pages

Reference Method:

- The calibration method used was CP-096 based on an in-house method.
- The temperature scale used was an ITS-90.
- This certificate can be traceable to the national standards, which is realized the shown measurement units according to the International System of Units (SI Units).

Reference Standard Instruments:

Type	Model	Serial No.	Cert. No.	Due Date	Traceability
Thermometer Readout	1529-R	B72853	10-0911001/22	Nov. 9, 2023	THC
Platinum Resistance Thermometer	5626	4854	C0A30047	Oct. 22, 2023	FLUKE
Liquid Bath	XORTS-40A	XO111019	10-2405001/23	May 25, 2025	THC

Remark: This certificate is traceable to the International System of Unit (SI Unit) through:

- THC, Thai Heart Calibration Co., Ltd.
- FLUKE, Fluke Corporation, U.S.A.

Measurement Results:

(X) Without Adjustment

Dimension of probe : Diameter 4 mm.	Sensor Type : RTD (PT100)
Immersion Depth (mm.)	Standard Reading ($^\circ\text{C}$)
120	22.00
120	25.00
120	28.00

UUC : Unit Under Calibration

The above reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor $k = 2.00$, providing a level of confidence approximately 95%.

- End of Certificate -

Calibrated by Kittipong REV.32 02/24/21

Calibrated by Pongsak REV.02 02/24/21

Certificate of Calibration

**TEMPERATURE
CONTROLLER ENCLOSURES**



Certificate No.: MC 2307702

Customer : Water Analysis Center Co., Ltd.
1/94 Moo 5, T.Kanthan, A.U-Thai, Ayutthaya 13210.

Reference Job No. : 23-157 Received Date : 11 July 2023
Description : Refrigerator
Manufacturer : SANDEN INTERCOOL Model : SEC-1500SBD
Serial No. : SEC1500201A-0708-00304 ID. No. : WWL0038
Marking : Additionally for the purpose of identification by this laboratory a label marked with this certificate number (MC 2307702) has been attached to the case.
Method : In-House calibration procedure MWI-T-033 this method is reference to TLAS G-20 "Temperature Controlled Enclosures".

Location of Calibration : Water Analysis Center Co., Ltd ; Laboratory.
Environmental Conditions : Ambient Temperature : (25.3 to 25.9) °C
Relative Humidity : (65.2 to 67.9) %

Date of Calibration : 11 July 2023 Date of Issue : 12 July 2023
Checked by : Thanagorn Limchaicharoen Approved by : Aitipong Kanjanawasit
(Calibration Supervisor) (Technical Manager)

The uncertainties are for a confidence probability of approximately 95%
This certificate is issued in accordance with the conditions of accreditation granted by the National Standardization Council of Thailand-Office of the National Standardization Council that has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of Master Calibration Co.,Ltd.

[MCF-Q-077 ; Rev6 ; Date : 22/04/2021]
[Page 1 of 3
Figure 1 : Sample Enclosed Enclosure
Overall Ambient Temperature variation : 3.2 °C
Overall Line Voltage variation : 0.1 V
Chamber Size (W*H*D) : 171 cm x 157 cm x 60 cm
Cheeked by : Thanagorn Limchaicharoen
[Page 2 of 3
[MCF-Q-077 ; Rev6 ; Date : 22/04/2021]

Certificate No.: MC 2307702

2. Result of calibration :

Temperature Measurement Accuracy Test

Measured Temperature (°C) at Spread Locations						
Indicating Temperature (°C)	#1	#2	#3	#4	#5	#6
2.5	4.4	4.2	4.2	4.2	4.0	3.9
					4.1	4.0

Chamber Characterization Result

Controller Temperature (°C)	Indicating Temperature (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
2.0	2.5	1.50	1.01	3.3

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95 %.

This certificate will certify of the calibrated equipment only.

End of Certificate

Page 3 of 3

CERTIFICATE OF CALIBRATION

Certificate No.: C0-190700723

Page 1 of total 2 pages

Customer: WATER ANALYSIS CENTER CO., LTD.
1/94 Moo 5, T.Kanham,
A.U-thai, Ayutthaya 13210

Equipment Manufacturer	CON 2700
Serial No.	EUTECH
Description	WWL.0136


Environmental Conditions	Ambient Temperature: (20 ± 2) °C
	Relative Humidity: (50 ± 10) %

Calibration Location	Jayhawks Laboratory (CL&GL)
Received Date	19 July 2023
Calibration Date	19 July 2023
Date of Issue	20 July 2023
Condition of Artifacts	Used conditions but can be calibrated

Checked by: 
Act as Technical Manager

Approved by: 
Representative of Managing Director

- () (Krisyos K.) () (Sakda Y.) (Dr. Ekachai Putititwong)
- () (Patiphan K.) () (Onnapa P.)
- () (Pongsak H.) () (Niphong K.)
- () (Kanung C.) () (Nonthachai K.)
- () (Pramong P.) () (Noppol P.)

Checked by: 

This calibration certificate shall not be reproduced other than in full except with the prior written approval of the Thai Heart Calibration Co., Ltd.

FE-169



THC
Thailand Calibration Co., Ltd.

Calibration Services

ANAB
ACCREDITED
CALIBRATION LABORATORY
AC-IRAB

Certificate No.: C0-190707/23

Page 2 of total 2 pages

Reference Method:

- The calibration method used was CP-177 based on an in-house method.

- This certificate can be traceable to the national standards, which is realized the shown measurement units according to the International System of Units (SI Units).

Reference Standard :

Material	Batch Value	Lot Number	Due Date	Traceability
Conductivity Standard Solution	1.47.8 µS/cm 1.425 mS/cm	S22/0611005 S22/0812006	Dec. 6, 2023 May 31, 2024	SCP Science
Note:	This certificate is traceable to the International System of Unit (SI Unit) through: - SCP Science.			

Measurement Results: (Probe Serial No.: 93X219065)

Conductivity Standard Solution	Measured Value	Correction	Uncertainty (±)
147.8 µS/cm	147.5 µS/cm	0.3 µS/cm	2.5 µS/cm
1.425 mS/cm	1.427 mS/cm	-0.002 mS/cm	0.0051 mS/cm

Note : Adjustment points: 147.8µS/cm, 1.425mS/cm

The above reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor $k = 2.00$, providing a level of confidence approximately 95%.

- End of Certificate -

Approved By :

(Mr. Nipon Phungsomsak)
Technician

Date Of Issue : 15/01/2024

This Certificate may not be reproduced other than in full, except with the prior written approval of the head of the industrial instruments calibration center.



AUTOMATION SERVICE CO., LTD.

CALIBRATION LABORATORY

Automation

Cert. No. WAC-065

Page 1 of 2

SV 201005/2024

CERTIFICATE OF CALIBRATION

Instrument : DO Meter
Model : DO-31P
Serial No. : 780065
Manufacturer : TOA-DKK
Measuring Range : 0.00 ~ 20.00 mg/l

Customer : Water Analysis Center Co.,Ltd.
1/94 Moo.5 T.Kanham, A.U.Thai
Ayutthaya 13210 Thailand

Date Of Received : 11/01/2024
Date Of Calibration : 11/01/2024

Ambient Condition : Temperature 26 °C
Humidity 58 % RH

Calibrated By : P. Yooey
(Ms. Phane Yooey)



AUTOMATION SERVICE CO., LTD.
CALIBRATION LABORATORY

Automation

Instrument : DO Meter
Model : DO-31P
Serial No. : 789065

Calibrate Procedure

- This instrument was calibrated by comparison with standard solution (PH/ORP)
- This instrument was calibrated by comparison with scattering plate value (Turbidity)
- This instrument was calibrated by comparison with conductivity (Conductivity)

Condition of this result of calibration
1). Reference Standard Solution

<u>Standard</u>	<u>Lot No</u>	<u>Batch</u>	<u>Cert. No.</u>	<u>Due Date</u>
Sodium Sulfite Power	408K1405	-	-	-

2). Traceability This certification is traceable to

Kanto Chemical Co.,INC.

DKK Corporation

Result Of Calibration

Standard Solution (mg/l) at 25.7°C	Before Adjust		After Adjust	
	Indicator	Error	Indicator	Error
Zero	0.00	+0.10	0.00	-
Span	8.02	6.45	-1.57	8.02

DO Electrode No. OE270AA(5) S/N 111F0029

Calibrated By P. Yooyen
(Ms. Phanee Yooyen)
Technician

Calibrated by : Mr.Yuttakorn Jamneansri

Issue date :

Apr 10, 2024

Automation Service Co. Ltd. 929/92/21 Soi Patakanakorn 30, Patakanakorn Rd, Suanluang, Bangkok 10250
Tel. : 02-319-5954 ext. 721-725 | E-mail : iso@automation.co.th services@automation.co.th | www.automation.co.th

Certificate of Calibration

Cert. No. WAC-065
Page 2 of 2

Incitech Metrological Center Co.,Ltd.
39/1 Soi 82, Sukhaoban 5 Rd, O NGOEN,
Salmaei, Bangkok 10220, Thailand
Tel. (662) 909-8820 (auto 10 lines) www.irminstrument.com

lac-MEA ACCREDITED
Calibration Cert. # X884.91
ISO/IEC 17025

Certificate No. : MT24-3208
Page : 1 of 2

Customer Address : Water Analysis Center Co.,Ltd.
: 194 M.5, Rojana Industrial Park, T.Karham, A.U-Thai, Ayutthaya 13210

Description	: Hot Air Oven
Manufacturer	: Memmert
Model	: UF 260
Serial No.	: BG20.0314
Identification No.	: WWL 0212
Calibration Place	: Customer Laboratory

Calibration Method : Calibration were conducted using In-house calibration procedure CR-NI-006 According to
comparison with LXI Data Acquisition Switch Unit with sensor. The calibration methods
based on Euramet Calibration Guide No.20 - guidelines on the Calibration of Temperature
and/or Humidity Controlled Enclosures.

Reference Standard Instruments :	<u>Instrument</u>	<u>Model</u>	<u>Serial No.</u>	<u>Certificate No.</u>	<u>Date</u>
LXI Data Acquisition Switch Unit with Sensor	34972A	MT49020096	MT23-7153	Nov 30, 2024	

The effect that the result relate only to the items calibrated. It was found accurate as shown on date and place
of calibration only.
Traceability : This measurement are traceable to the International System of Unit (SI), through
National Institute of Metrology Thailand (NIMT)



Approved by : Mr.Yuttakorn Jamneansri
Issue date : Apr 10, 2024

This calibration certificate shall not be reproduced other than in full except with the prior written
approval of Incitech Metrological Center Co.,Ltd

Rev.03 / Feb 2024

FIM-MT-013



Certificate No. : MT24-3208

Page : 2 of 2

Function : Temperature measurement
Calibration point : 104, 180 °C

Result : Without adjustment
Resolution : 0.1 °C

Temperature of UUC* at each position (°C)						
Calibration point (°C)	Uncertainty of measurement (+/- °C)					
	Ch.1	Ch.2	Ch.3	Ch.4	Ch.5	Ch.6
104	103.494	103.933	103.871	103.988	104.081	103.843
180	179.885	179.953	180.047	179.986	179.908	180.088

Measured stability (+/- °C) 0.34 0.34 0.34 0.34 0.34 0.34

Overall variation (°C) 1.3 1.3 1.3 1.3 1.3 1.3

Setting temperature (°C) 104.0 104.0 104.0 104.0 104.0 104.0

Indicating temperature (°C) 180.0 180.0 180.0 180.0 180.0 180.0

Measured uniformity (°C) 0.66 0.66 0.66 0.66 0.66 0.66

Overall variation (°C) 1.2 1.2 1.2 1.2 1.2 1.2

Setting temperature (°C)	Indicating temperature (°C)	Measured stability (+/- °C)	Measured uniformity (°C)	Overall variation (°C)
104.0	104.0	0.34	0.66	1.3
180.0	180.0	0.41	0.86	1.2


Customer: Water Analysis Center Co., Ltd.
1194 Moo 5, Rojana Industrial Park, Rojana Road,
Tambol Kanham, Amphur U-Thai, Ayutthaya 13210 Thailand

Environment Condition: Temperature 28 °C ± 0.2 °C
Humidity 50 %RH ± 2.6 %RH

Calibration Place: Water Analysis Center Co., Ltd. (วอเตอร์เอนะลิสเซนซ์)
1194 Moo 5, Rojana Industrial Park, Rojana Road,
Tambol Kanham, Amphur U-Thai, Ayutthaya 13210 Thailand

Calibration By: Mr. Polawat Ruamtip
Calibration Date: 05 June 2024
The Method used: In-house method, CAL-WL-47, based on UKAS Lab 14
Traceability: This certificate is traceable to the SI Units maintained by National Institute of Metrology (NIMT), Thailand through DKSH Technology Co., Ltd. Certificate No. C02240400

- #1 Lower Left Front
- #2 Lower Right Front
- #3 Lower Left Rear
- #4 Lower Right Rear
- #5 Upper Left Front
- #6 Upper Right Front
- #7 Upper Left Rear
- #8 Upper Right Rear
- #9 Geometric Center



Front view

UUC* = Unit under calibration
Uniformity = Maximum and Minimum difference of measured temperature at any probes and the measured temperature at the reference and same time.
Overall Variation = Difference of temperature value between the maximum and minimum any time.
Stability = One half of the maximum difference of measured temperatures at any one probe.

This certificate is issued to the units of measurement according to the International System of Units (SI). It provides traceability of measurement to International or national standard or other recognized national standard laboratories.
The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty specified by the coefficient factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).
These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. This report shall not be reproduced except in full without approval of DKSH Technology Limited.
Other laboratory certified by DKSH
DKSH Laboratory, 10th floor, 100/10 Rama 9 Road, Bangkok 10220
2222 Sathorn, Bangkok, Thailand. Phone: +66 2 553 2553, Email: info.calibration@dksh.com - Website: www.dksh.com/calibration/thailand
Delivering Growth - in Asia and Beyond.


CAL-FRA-C01-14; 12 Sep 2022

F.M-MT-013

Rev.03 / Feb 2024

Calibration Results:**Without Adjustment**

Eccentric Error: Weight to be 1/3 or 1/2 of Maximum capacity, taken from the center of the pan as a zero reference.



		Nominal Test Value	100	(g)			
		Reference Points (g)	A	B	C	D	E
	-	0.00000	0.0001	0.00000	-0.0002		

Repeatability: Determination of the standard deviation of weighing balance., Readability

Nominal test value (g)	Standard Deviation
20	0.00004
200	0.00006

Nominal Value	Conventional Mass	Displayed Value	Error of Indication	Uncertainty	k
(g)	(g)	(g)	(g)	(g)	
1	1.00001	1.0000	0.0000	0.00011	2.04
2	2.00002	2.0000	0.0000	0.00011	2.04
5	5.00002	5.0000	0.0000	0.00011	2.04
10	10.00001	10.0000	0.0000	0.00011	2.04
20	20.00001	20.0000	0.0000	0.00012	2.03
50	50.00003	50.0000	0.0000	0.00013	2.02
70	70.00004	70.0000	0.0000	0.00016	2.01
100	99.99996	100.0001	0.0001	0.00017	2.01
120	119.99997	120.0002	0.0002	0.00021	2.00
150	149.99999	150.0002	0.0002	0.00024	2.00
200	199.99996	200.0004	0.0004	0.00030	2.00

The End of Certificate

Within Master Calibration Co.,Ltd
DKSH Technology (Thailand) Co.,Ltd
2531 number 9/9, Kharunrat 3, Nonthaburi 10200
Phone: 66 239 7000 Email: ktc.calibration@dksh-thailand.com
Website: www.dksh.com/dksh-thailand
Delivering Growth - in Asia and Beyond.

Within Master Calibration Co.,Ltd
DKSH Technology (Thailand) Co.,Ltd
2531 number 9/9, Kharunrat 3, Nonthaburi 10200
Phone: 66 239 7000 Email: ktc.calibration@dksh-thailand.com
Website: www.dksh.com/dksh-thailand
Delivering Growth - in Asia and Beyond.


Checked by : Chalemkit Rakphada
(Calibration Engineer)

Approved by : Aittipong Kanchanaasit
(Technical Manager)

The uncertainties are for a confidence probability of approximately 95%

This certificate is issued in accordance with the conditions of accreditation granted by the National Standardization Council of Thailand-Office of the National Standardization Council that has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of Master Calibration Co.,Ltd.

Certificate of Calibration



Page 1 of 3

Liquid Bath

Certificate No.: MC 2314268

Customer : Water Analysis Center Co., Ltd.
1/94 Moo 5, T.Kantham, A.U-Thai, Ayutthaya 13210.

Received Date : 15 December 2023

Reference Job No. : 23-2833

Description : Water Bath

Manufacturer : ESSTELL

Model : EWB-122D

Serial No. : 20180508122

ID.No. : WWL 0214

Marking :

Additionally for the purpose of identification by this laboratory a label marked

with this certificate number (MC 2314268) has been attached to the case.

Method :

In-House calibration procedure MWL-T-029 this method is reference to

ASTM E715 "Liquid Bath".

Location of Calibration : Water Analysis Center Co., Ltd ; Laboratory.

Environmental Condition : Ambient Temperature : (29.4 to 29.8) °C

Relative Humidity : (49.0 to 52.0) %

Date of Calibration : 15 December 2023

Date of Issue : 19 December 2023

Certificate No.: MC 2314268

Page 2 of 3

Reference Standard Instrument:

Description	Certificate No.	Serial No.	Due date	Traceable thru
Data Acquisition/Switch Unit With Thermocouple Type "T" ID. No.271 to 275	MC 2301270	MY44020009	9 Mar 2024	MCAL

Traceability :

The measurement standard traceable to the international system of units (SI) through certificate as mentioned above

1. Calibration Procedure:


This instrument was calibration according to ASTM E715 - 2007 by comparison with calibrated sensor under no load condition. The sensor were placed on five points and located one sensor in each of the eight corners of the chamber and was away from the each wall of 5 cm to 10 cm. And placed the five sensor within 2.5 cm of the geometric center of the chamber.

Temperature Uniformity - the maximum difference of measured temperatures at any sensors and the

measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.

Temperature Stability - one-half of the greatest maximum difference of measured temperatures at any one sensor.

Overall Variation - The Difference of the maximum and minimum measured temperatures throughout observation.



Certificate No.: MC 2314268

Page 3 of 3

2. Result of calibration :

Temperature Measurement Accuracy Test

Indicating Temperature (°C)	Measured Temperature (°C) at Spread Locations				Uncertainty (±°C)
	#1	#2	#3	#4	
45.0	44.5	44.4	44.5	44.6	0.45

Chamber Characterization Result

Desired Temperature (°C)	Controller Temperature (°C)	Indicating Temperature (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
44.5	45.0	45.0	0.62	0.88	1.5

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2.0$, providing a level of confidence of approximately 95 %.

This certificate will certify of the calibrated equipment only.

End of Certificate

Checked by : *Chaleamka*

Checked by : *Chaleamka*



Master Calibration Co.,Ltd.

547 Soi Rachadanan, Keweng Samseenuk, Khet Huaykwang, Bangkok 10310

MASTER CALIBRATION CO.,LTD.

Tel : (02) 274 2978-9, (02) 2742987-8; Fax : (02) 274 2518, (02) 274 2989

Website : www.mastercalibration.com E-mail : calibrate@mastercalibration.com

Certificate of Calibration



**TEMPERATURE
CONTROLLER ENCLOSURES**

NSC:181HIS:17025
CALIBRATION 0183


Certificate No.: MC 2314270

Page 1 of 3

Customer : Water Analysis Center Co., Ltd.
1/94 Moo 5, T.Kanthan, A.U-Thai, Ayutthaya 13210.

Reference Job No.	: 23-2833	Received Date	: 15 December 2023
Description	: Incubator	Model	: IN260
Manufacturer	: Memmert	ID. No.	: WWL 0192
Serial No.	: D619.0170		
Marking	: Additionally for the purpose of identification by this laboratory a label marked with this certificate number (MC 2314270) has been attached to the case.		
Method	: In-House calibration procedure MWI-T-033 this method is reference to TLAS G-20 "Temperature Controlled Enclosures".		
Location of Calibration	: Water Analysis Center Co., Ltd. ; Laboratory.		
Environmental Conditions	: Ambient Temperature : (25.2 to 25.6) °C Relative Humidity : (65.4 to 66.2) %		
Date of Calibration	: 15 December 2023	Date of Issue	: 19 December 2023

Checked by : Chalemkit Rakphada
(Calibration Engineer)
Approved by : Aittipong Kajjanasusit
(Technical Manager)



The uncertainties are for a confidence probability of approximately 95%

This certificate is issued in accordance with the conditions of accreditation granted by the National Standardization Council of Thailand-Office of the National Standardization Council that has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of Master Calibration Co.,Ltd.



MASTER CALIBRATION CO.,LTD.

547 Soi Rachadanan, Keweng Samseenuk, Khet Huaykwang, Bangkok 10310
Tel. : (02) 274 2978-9, (02) 2742987-8; Fax : (02) 274 2518, (02) 274 2989
Website : www.mastercalibration.com E-mail : calibrate@mastercalibration.com

Certificate No.: MC 2314270

Page 2 of 3

Certificate No. : MC 2314270

Page 3 of 3

2. Result of calibration :

Temperature Measurement Accuracy Test

Measured Temperature (°C) at Spread Locations										
Indicating Temperature (°C)	#1	#2	#3	#4	#5	#6	#7	#8	Ref. #9	Uncertainty (±°C)
35.0	35.2	35.2	35.2	35.2	35.1	35.1	35.0	35.1	35.1	0.44

Chamber Characterization Result

Desired Temperature (°C)	Controller Temperature (°C)	Indicating Temperature (°C)	Temperature Stability (±°C)	Temperature (°C)	Uniformity (°C)	Temperature (°C)	Overall Variation (°C)
35.0	35.0	35.0	0.13	0.21	0.21	0.21	0.4

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95 %.

This certificate will certify of the calibrated equipment only.

End of Certificate

Customer :	Water Analysis Center Co., Ltd.	Received Date :	15 December 2023
Location of Calibration :	1/94 Moo 5, T.Kantham, A.U-Thai, Ayutthaya 13210.	Description :	Autoclave
Environmental Condition :	Ambient Temperature : (29.4 to 30.7) °C	Manufacturer :	TOMY
	Relative Humidity : (50.0 to 52.0) %	Serial No. :	51135128
Date of Calibration :	15 December 2023	Model :	Autoclave ES-315
		Marking :	W WL 0083
		Method :	In-House calibration procedure MWL-T-036 this method is reference to based on BS 2646 1993 Part 5 "Autoclave".

Date of Issue : 19 December 2023

Checked by : Chalermtipha
 Chalermtipha Rakphada
 (Calibration Engineer)

Approved by : Aittipong
 Aittipong Kamjantarakasit
 (Technical Manager)

The uncertainties are for a confidence probability of approximately 95%

This certificate is issued in accordance with the conditions of accreditation granted by the National Standardization Council of Thailand-Office of the National Standardization Council that has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of Master Calibration Co.,Ltd.

Checked by : Chalermtipha

Certificate No.: MC 2314269

Page 2 of 3

Certificate No.: MC 2314269

Page 3 of 3

Reference Standard Instrument :

Description	Certificate No.	Serial No.	Due date	Traceable thru
Temperature Recorder RTD 100 Ohm	MC 2300163	M79252	9 Jan 2024	MCAL
Temperature Recorder RTD 100 Ohm	MC 2300164	5978194	9 Jan 2024	MCAL
Temperature Recorder RTD 100 Ohm	MC 2300165	M79251	9 Jan 2024	MCAL

Traceability :

The measurement standard traceable to the international system of units (SI) through certificate as mentioned above

1. Calibration Procedure:

The equipment list above was calibrated an accuracy of temperature in a chamber of the sterilizer.

The calibration was performed by direct measurement of generated temperatures using the standard thermometer with three temperature sensors. The data was recorded in a period of fifteen minutes of the sterilizing status. The temperature scale used was based on ITS-90.

The calibration of sterilizer was carried out at the point indicated by following the In-house calibration method No. MWI-T-036 based on BS 2646 : 1993 : Part 5 in Tests for performance section.

The measurement standard traceable to the international system of units (SI) through certificate as mentioned above

2. Result of calibration :

Temperature Measurement Accuracy Test

Indicating Temperature (°C)	Measured Temperature (°C) at Spread Locations			Uncertainty (±°C)
	#1	#2	#3	
121	121.72	121.73	121.95	0.61

Characterization Result

Desired Temperature (°C)	Setting Temperature (°C)	Timer Setting (min)	Indicating Temperature (°C)	Indicating Pressure (kPa)	Measured Stability (±°C)	Measured Uniformity (°C)	Overall Variation (°C)
121	121	15.0	121	0.60	0.35	0.35	1.35

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95 %.

This certificate will certify of the calibrated equipment only.

End of Certificate

