


---

## เอกสารสอบเทียบเครื่องมือที่ใช้ในการวิเคราะห์



THAI HEART CALIBRATION CO., LTD.  
102/11 Moo 5, Phraek Sai, Bang Phai, Samut Prakan 10980  
Tel: 0289521021-0289521022-0289521023 Fax: 0289521027



ANAB  
ACCREDITED  
CALIBRATION LABORATORY  
AC-2095

## CERTIFICATE OF CALIBRATION

Certificate No.: CO-1808005/23

Page 1 of total 4 pages

**Customer**  
WATER ANALYSIS CENTER CO., LTD.  
1/94 Moo 5, T. Kanham,  
A.U-thai, Ayuthaya 13210

**Equipment**  
pH Meter  
**Manufacturer**  
METTLER TOLEDO  
**Model**  
SevenCompact S220  
**Serial No.**  
B327527211  
**ID No.**  
WWL 0068  
**Description**  
Range : 0 - 14 pH, Resolution : 0.01 pH

**Environmental Conditions**  
Ambient Temperature: (20 ± 2) °C  
Relative Humidity: (50 ± 10) %  
Atmospheric Pressure: -

**Calibration Location**  
Jayhawks Laboratory (CL&GL)  
**Received Date**  
18 August 2023  
**Calibration Date**  
18 August 2023  
**Date of Issue**  
21 August 2023

**Condition of Artifacts**  
Used conditions but can be calibrated

**Checked by**

**Approved by**

Act as Technical Manager

Representative of Managing Director

( ) (Krisyosil K.) ( ) (Sakda Y.)  
( ) (Patiphan K.) (✓) (Onnapa P.)  
( ) (Pongsak H.) ( ) (Nitiphong K.)  
( ) (Kanung C.) ( ) (Nonthachai K.)  
( ) (Pramong P.) ( ) (Noppol P.)

( Dr. Ekachai Puttitwong )

Certificate No.: CO-1808005/23

Page 2 of total 4 pages

Reference Method:

- The calibration method used was CP-178 based on an in-house method.
- This certificate can be traceable to the national standards, which is realized the shown measurement units according to the International System of Units (SI Units).

Reference Standard:

Type	pH Value	Lot No.	Due Date	Traceability
pH Standard Solution	4.01	030822	Feb. 9, 2024	NIMT
	7.01	300522	Feb. 9, 2024	
	10.01	230822	Feb. 7, 2024	

Type	Model	Serial No.	Certificate No.	Due Date	Traceability
Documenting Process Calibrator	754	2630521	10-2412001/22	Dec. 23, 2023	THC
Digital Thermometer with Sensor	1523 / 5622	1709138 / 4605984-005	10-0806001/23	Jun. 8, 2024	

Remark: This certificate is traceable to the International System of Unit (SI Unit) through:

- NIMT, National Institute of Metrology (Thailand).
- THC, Thai Heart Calibration Co., Ltd.

Measurement Results:

1. Function Simulated pH Meter

Standard Applied (mV)	Nominal Value (pH)	UUC Reading		Uncertainty (± mV)
		pH	mV	
177.48	4.00	4.01	177.4	0.060
0.00	7.00	7.00	0.0	0.060
-177.48	10.00	10.01	-177.4	0.060

UUC : Unit Under Calibration

Note : Adjust Curve to simulate pH (4,7,10)

This calibration certificate shall not be reproduced other than in full except with the prior written approval of the Thai Heart Calibration Co., Ltd.

FE-169

REV.02 02/24/21

FE-169

Calibrated by Kittipong  
REV.02 02/24/21

Certificate No.: C0-1808005/23

Page 3 of total 4 pages

Measurement Results (Cont.):

2. Calibration of pH Electrode (Serial No.: 3222623)

pH Standard Solution (pH)	Measured Value		Uncertainty (± pH)
	(pH)	(mV)	
4.01	4.01	180.0	0.013
7.01	7.00	4.0	0.013
10.01	10.01	-172.0	0.013

Note : Adjust Curve to Buffer Solution pH (4,7,10)

Temperature stability of micro bath :  $25 \pm 0.2^{\circ}\text{C}$

The above reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2.00$ , providing a level of confidence approximately 95%.

Certificate No.: C0-1808005/23

Page 4 of total 4 pages

Reference Method:

- The calibration method used was CP-096 based on an in-house method.
- The temperature scale used was an ITS-90.
- This certificate can be traceable to the national standards, which is realized the shown measurement units according to the International System of Units (SI Units).

Reference Standard Instruments:

Type	Model	Serial No.	Cert. No.	Due Date	Traceability
Thermometer Readout	1529-R	B7C853	10-0911001/22	Nov. 9, 2023	THC
Platinum Resistance Thermometer	5626	4854	C0A30047	Oct. 22, 2023	FLUKE
Liquid Bath	XORTS-40A	XO111019	10-2405001/23	May 25, 2025	THC

Remark: This certificate is traceable to the International System of Unit (SI Unit) through:

- THC, Thai Heart Calibration Co., Ltd.
- FLUKE, Fluke Corporation, U.S.A.

Measurement Results:

(X) Without Adjustment

Dimension of probe : Diameter 4 mm. Sensor Type : RTD (PT100)

Immersion Depth (mm.)	Standard Reading ( $^{\circ}\text{C}$ )	UUC Reading ( $^{\circ}\text{C}$ )	Correction ( $^{\circ}\text{C}$ )	Uncertainty ( $\pm ^{\circ}\text{C}$ )
120	22.00	22.2	-0.20	0.065
120	25.00	25.2	-0.20	0.065
120	28.00	28.2	-0.20	0.065

UUC : Unit Under Calibration

The above reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2.00$ , providing a level of confidence approximately 95%.

- End of Certificate -

Calibrated by Kittipong  
REV.02 02/24/21

Calibrated by Pongsak  
REV.02 02/24/21



Certificate No.: MC 2307702

**The Reference Standard Instrument :**

<b>Description</b>	<b>Certificate No.</b>	<b>Serial No.</b>	<b>Due date</b>	<b>Traceable thru</b>
Data Acquisition/Switch Unit	MC 2303173	MY41010916	9 Mar 2024	MCAL
With Thermocouple Type " T " ID. No.171 to 1719				

**Traceability :**

The measurement standard traceable to the international system of units (SI) through certificate as mentioned above

**1. Calibration Procedure:**

This Instrument was calibration according to TLAS G-20 by comparison with calibrated thermocouple type T under no load condition. The Thermocouples were placed on nine points and located one thermocouple in each of the eight corners of the chamber and was away from the each wall of 5 cm to 10 cm. And placed the ninth thermocouple within 2.5 cm of the geometric center of the chamber.

**Temperature Uniformity** - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.

**Temperature Stability** - one-half of the greatest maximum difference of measured temperatures at any one sensor.

**Overall Variation** - The Difference of the maximum and minimum measured temperatures throughout observation.




Figure 1 : Sensor Installation Location

Checked by : *Thanagim*

# Certificate of Calibration

## TEMPERATURE CONTROLLER ENCLOSURES



Certificate No.: MC 2307702

Customer : Water Analysis Center Co., Ltd.

1/94 Moo 5, T.Kantham, A.U.-Thai, Ayuthaya 13210.

Reference Job No. : 23-1577 Received Date : 11 July 2023

Description : Refrigerator

Manufacturer : SANDEN INTERCOOL Model : SEC-1500SBD

Serial No. : SEC1500201A-0708-00304 ID. No. : WWL0038

Marking : Additionally for the purpose of identification by this laboratory a label marked with this certificate number ( MC 2307702 ) has been attached to the case.

Method : In-House calibration procedure MWI-T-033 this method is reference to

TLAS G-20 "Temperature Controlled Enclosures".

Location of Calibration : Water Analysis Center Co., Ltd. ; Laboratory.

Environmental Conditions : Ambient Temperature : ( 25.3 to 25.9 ) °C

Relative Humidity : ( 65.2 to 67.9 ) %

Date of Calibration : 11 July 2023 Date of Issue : 12 July 2023

Checked by : *Thanagim*  
Thanagorn Limchaicharoen  
( Calibration Supervisor )

Approved by : *Aittipong*  
Aittipong Kanjanawasit  
( Technical Manager )

**The uncertainties are for a confidence probability of approximately 95%**

This certificate is issued in accordance with the conditions of accreditation granted by the National Standardization Council of Thailand-Office of the National Standardization Council that has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of Master Calibration Co.,Ltd.

Certificate No.: MC 2307702

Page 3 of 3

## 2. Result of calibration :

### Temperature Measurement Accuracy Test

Indicating Temperature (°C)	Measured Temperature (°C) at Spread Locations									Uncertainty (±°C)
	#1	#2	#3	#4	#5	#6	#7	#8	Ref. #9	
2.5	4.4	4.2	4.2	4.2	4.0	3.9	4.1	4.0	3.8	0.86

### Chamber Characterization Result

Controller Temperature (°C)	Indicating Temperature (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
2.0	2.5	1.50	1.01	3.3

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95 %.

This certificate will certify of the calibrated equipment only.

End of Certificate

Checked by: *Thanyan*

[MCF-Q-077 ; Rev.6 ; Date : 22/04/2021]



THAI HEART CALIBRATION CO., LTD.  
102/11 Moo 5, Phraek Sai, Bangna Suburb, Bangkok 10700  
Tel. 0-2394-9109, 0-2394-9108, 0-2394-9107, 0-2394-9106  
Fax 0-2394-9105, 0-2394-9104, 0-2394-9103, 0-2394-9102



# CERTIFICATE OF CALIBRATION

Certificate No.: C0-1907007/23 Page 1 of total 2 pages

**Customer**  
WATER ANALYSIS CENTER CO., LTD.  
1/94 Moo 5, T.Kanham,  
A.U-thai, Ayutthaya 13210

**Equipment** Conductivity Meter  
**Manufacturer** EUTECH  
**Serial No.** 2657889  
**Description** -  
**Model** CON 2700  
**ID No.** WWL 0136

**Environmental Conditions**  
Ambient Temperature: (20 ± 2) °C  
Relative Humidity: (50 ± 10) %  
Atmospheric Pressure: -

**Calibration Location** Jayhawks Laboratory (CL&GL)  
**Received Date** 19 July 2023  
**Calibration Date** 19 July 2023  
**Date of Issue** 20 July 2023

**Condition of Artifacts** Used conditions but can be calibrated

Checked by: *[Signature]* Approved by: *[Signature]*  
Representative of Managing Director

Act as Technical Manager

( ) (Krisyosl K.) ( ) (Sakda Y.)  
( ) (Patiphan K.) (✓) (Onnapa P.)  
( ) (Pongsak H.) ( ) (Nitiphong K.)  
( ) (Kanung C.) ( ) (Nonthachai K.)  
( ) (Pramong P.) ( ) (Noppol P.)

This calibration certificate shall not be reproduced other than in full except with the prior written approval of the Thai Heart Calibration Co., Ltd.

FE-169

REV.02 02/24/21



THEANIPART CALIBRATION CO., LTD.  
110/11 Moo 5, Phra Prachin, Muang Samut Prakan 10180  
Tel. 02-204-2162, 02-225-3165, 02-275-6743, 0643-02755-8507



Certificate No.: C0-1907007/23

Page 2 of total 2 pages

Reference Method:

- The calibration method used was CP-177 based on an in-house method.
- This certificate can be traceable to the national standards, which is realized the shown measurement units according to the International System of Units (SI Units).

Reference Standard :

Material	Batch Value	Lot Number	Due Date	Traceability
Conductivity Standard Solution	147.8 $\mu$ S/cm	S220611005	Dec. 6, 2023	SCP Science
	1.425 mS/cm	S220812006	May 31, 2024	

Remark: This certificate is traceable to the International System of Unit (SI Unit) through:

- SCP Science.

Measurement Results: (Probe Serial No. : 93X219065)

Conductivity Standard Solution	Measured Value	Correction	Uncertainty ( $\pm$ )
147.8 $\mu$ S/cm	147.5 $\mu$ S/cm	0.3 $\mu$ S/cm	2.5 $\mu$ S/cm
1.425 mS/cm	1.427 mS/cm	-0.002 mS/cm	0.0051 mS/cm

Note : Adjustment points: 147.8 $\mu$ S/cm 1.425mS/cm

The above reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2.00$ , providing a level of confidence approximately 95%.

- End of Certificate -

Calibrated by Onnapa  
REV.02.02/24/21



AUTOMATION SERVICE CO., LTD.  
CALIBRATION LABORATORY

SV 201005/2024

Cert. No. WAC-065  
Page 1 of 2

## CERTIFICATE OF CALIBRATION

Instrument : DO Meter  
Model : DO-31P  
Serial No. : 780065  
Manufacturer : TOA-DKK  
Measuring Range : 0.00 ~ 20.00 mg/l

Machine : -  
Location : -

Customer : Water Analysis Center Co.,Ltd.  
1/94 Moo.5 T.Kanham, A.U.-Thai  
Ayutthaya 13210 Thailand

Date Of Received : 11 / 01 / 2024  
Date Of Calibration : 11 / 01 / 2024

Ambient Condition : Temperature 26  $^{\circ}$ C  
Humidity 58 % RH

Calibrated By :

P. Yooyen  
(Ms. Phance Yooyen)  
Technician

Approved By :

N. Phung  
(Mr. Nipon Phungsomsak)  
Technical Manager

Date Of Issue : 15 / 01 / 2024

This Certificate may not be reproduced other than in full, except with the prior written approval of the head of the industrial instruments calibration center.





AUTOMATION SERVICE CO., LTD.  
CALIBRATION LABORATORY

Automation

Instrument : DO Meter  
Model : DO-31P  
Serial No. : 780065

Cert. No. WAC-065  
Page 2 of 2

Calibrate Procedure

- ☐ This instrument was calibrated by comparison with standard solution (PH/ORP)  
☐ This instrument was calibrated by comparison with scattering plate value (Turbidity)  
☐ This instrument was calibrated by comparison with conductivity (Conductivity)  
☒ This instrument was calibrated by comparison with Sodium sulfite anhydrous (DO)

Condition of this result of calibration

1). Reference Standard Solution

Standard	Lot No	Batch	Cert. No.	Due Date
----------	--------	-------	-----------	----------

Sodium Sulfite Power	408K1405	-	-	-
----------------------	----------	---	---	---

2). Traceability This certification is traceable to

- ☒ Kanto Chemical Co.,INC.  
☐ DKK Corporation

Result Of Calibration

Standard Solution (mg/l) at 25.7°C	Before Adjust		After Adjust	
	Indicator	Error	Indicator	Error
Zero	0.00	+ 0.10	0.00	-
Span	8.02	- 1.57	8.02	-

DO Electrode No. OE270AA(S) S/N 111F0029

Calibrated By

P. Yooyen

(Ms. Phanee Yooyen)  
Technician



Inctech Metrological Center Co.Ltd.  
39/1 Soi 82, Sukhaphan 5 Rd., O ngoen,  
Salmal, Bangkok 10220, Thailand  
Tel. (662) 909-8820 (Auto 10 lines) www.imcinstrument.com



## Certificate of Calibration

Certificate No. : MT24-3208  
Page : 1 of 2

Customer	:	Water Analysis Center Co.,Ltd.
Address	:	1/94 M.5, Rojana Industrial Park, T.Kanham, A.U-Thai, Ayuthaya 13210
Description	:	Hot Air Oven
Manufacturer	:	Memmert
Model	:	UF 260
Serial No.	:	B620.0814
Identification No.	:	WWL 0212
Calibration Place	:	Customer Laboratory
Order No.	:	1152/24
Received date	:	Mar 22, 2024
Calibration date	:	Mar 22, 2024
Environment Condition :		
Temperature	:	(25 $\pm$ 10) °C
Humidity	:	(50 $\pm$ 30) %RH

Calibration Method : Calibration were conducted using In-house calibration procedure CP-MT-006 According to comparison with LXI Data Acquisition Switch Unit with sensor. The calibration methods based on Euramet Calibration Guide No.20 - guidelines on the Calibration of Temperature and/or Humidity Controlled Enclosures.

Reference Standard Instruments :

Instrument	Model	Serial No.	Certificate No.	Due Date
LXI Data Acquisition Switch Unit with Sensor	34972A	MY49020096	MT23-7163	Nov 30, 2024

The effect that the result relate only to the items calibrated. It was found accurate as shown on date and place of calibration only.

Traceability : This measurement are traceable to the International System of Unit (SI), through National Institute of Metrology Thailand ( NIMT )

The reported expanded uncertainty of measurement was based on standard uncertainty multiplied by coverage factor 2, providing a level of confidence of not less than 95%



Calibrated by : Mr.Yuttakorn Jamneansri

Approved by :

(Mr.Panuwat Phuklan)

Issue date : Apr 10, 2024

This calibration certificate shall not be reproduced other than in full except with the prior written approval of Inctech Metrological Center Co.,Ltd

Rev.03 / Feb 2024

Automation Service Co.,Ltd. 929 829/1 Soi Pattanakarn30, Pattanakarn Rd., Suanluang, Suanluang, Bangkok 10250  
Tel. : 02-319-9994 ext. 721,725 | E-mail : iso@automation.co.th, service@automation.co.th | www.automation.co.th

FM-MT-013

Function

: Temperature measurement

Calibration point

: 104, 180 °C

Result

: Without adjustment


Certificate No. : MT24-3208

Page : 2 of 2

Resolution : 0.1 °C

Calibration point (°C)	Temperature of UUC* at each position (°C)									Uncertainty of measurement (+/- °C)
	Ch.1	Ch.2	Ch.3	Ch.4	Ch.5	Ch.6	Ch.7	Ch.8	Ch.9	
104	103.494	103.933	103.871	103.988	103.990	104.081	103.843	104.217	104.022	0.45
180	179.985	179.953	180.047	179.985	179.908	180.088	180.065	180.273	180.105	0.54

Setting temperature (°C)	Indicating Temperature (°C)	Measured stability (+/- °C)	Measured uniformity (°C)	Overall variation (°C)
104.0	104.0	0.34	0.66	1.3
180.0	180.0	0.41	0.86	1.2



UUC\* = Unit under calibration  
Uniformity = Maximum and Minimum difference of measured temperature at any probes and the measured temperature at the reference and same time.  
Overall Variation = Difference of temperature value between the maximum and minimum any time.  
Stability = One half of the maximum difference of measured temperatures at any one probe.



## Certificate of Calibration

Equipment: Balance  
Model: BL 210S  
Serial No. (or ID.): 15808131 (WWL 0022)  
Manufacturer: Sartorius  
Condition: In condition

Certificate No.: C01241754

Issued Date: 05 June 2024

Job No.: WO-00030302

Page: 1 of 2

Customer: Water Analysis Center Co., Ltd.  
1/94 Moo 5, Rojana Industrial Park, Rojana Road,  
Tambol Kanham, Amphur U-Thai, Ayutthaya 13210 Thailand

Environment Condition: Temperature 26 °C ± 0.2 °C  
Humidity 50 %RH ± 2.6 %RH

Calibration Place: Water Analysis Center Co., Ltd. ( ห้างเครื่องชั่ง )  
1/94 Moo 5, Rojana Industrial Park, Rojana Road,  
Tambol Kanham, Amphur U-Thai, Ayutthaya 13210 Thailand

Calibration By: Mr. Polawad Ruamrup

Calibration Date: 05 June 2024

The Method used: In-house method, CAL-WI-47, based on UKAS Lab 14

Traceability: This certificate is traceable to the SI Units maintained by National Institute of Metrology (NIMT), Thailand through DKSH Technology Co., Ltd. Certificate No. C02240400

(Mr. Polawad Ruamrup)

Person in charge

(Mr. Rungrod Jenkitrakulchai)

Authorized signatory

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national laboratories.  
The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).  
These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of DKSH Technology Limited.

บริษัท ดีเคเอส อีเซีย จำกัด  
DKSH Technology Limited  
2533 หมู่ที่ 5 ถนนสุขุมวิท แขวงคลองเตย เขตคลองเตย กรุงเทพมหานคร 10260  
Phone: +66 2039 7000 Email: info.calibration@dksh.com Website: www.dksh.com/scientific-asia/thailand

Delivering Growth - in Asia and Beyond.





Certificate No.: C01241754 Page: 2 of 2

#### Calibration Results:

##### Without Adjustment

Eccentric Error: Weight to be 1/3 or 1/2 of Maximum capacity, taken from the center of the pan as a zero reference.

Nominal Test Value	Reference Points (g)				
	A	B	C	D	E
-	0.0000	0.0001	0.0000	0.0000	-0.0002

Repeatability: Determination of the standard deviation of weighing balance., Readability 0.0001 (g)

Nominal test value (g)	Standard Deviation
20	0.00004
200	0.00006

Error of indication from nominal or conventional mass value., Readability 0.0001 (g)

Nominal Value (g)	Conventional Mass (g)	Displayed Value (g)	Error of indication (g)	Uncertainty (g)	k
1	1.00001	1.0000	0.0000	0.00011	2.04
2	2.00002	2.0000	0.0000	0.00011	2.04
5	5.00002	5.0000	0.0000	0.00011	2.04
10	10.00001	10.0000	0.0000	0.00011	2.04
20	20.00001	20.0000	0.0000	0.00012	2.03
50	50.00003	50.0000	0.0000	0.00013	2.02
70	70.00004	70.0000	0.0000	0.00016	2.01
100	99.99996	100.0001	0.0001	0.00017	2.01
120	119.99997	120.0002	0.0002	0.00021	2.00
150	149.99999	150.0002	0.0002	0.00024	2.00
200	199.99996	200.0004	0.0004	0.00030	2.00

The End of Certificate

DKSH Technology Limited  
2533 หมู่ 9 ตำบลบางนา อำเภอบางนา จังหวัดกรุงเทพมหานคร 10260  
Phone: +66 2639 7000 Email: info@dksh.com Website: www.dksh.com/vietnam-thailand

Delivering Growth - in Asia and Beyond.

CAL-FM-C01-14-12 Sep 2022



MASTER CALIBRATION CO.,LTD.

Master Calibration Co.,Ltd.


547 Soi Rachadaniwat, Kwang Sansenok, Khor Huaykwang, Bangkok 10310

Tel : (02) 274 2978-9, (02) 2742987-8 Fax : (02) 274 2518, (02) 274 2989

Website : www.mastercalibration.com E-mail : calibrate@mastercalibration.com

## Certificate of Calibration

### LIQUID BATH



Certificate No.: MC 2314268

Page 1 of 3



Customer : Water Analysis Center Co., Ltd.  
1/94 Moo 5, T.Kantham, A.U.-Thai, Ayutthaya 13210.

Reference Job No. : 23-2833 Received Date : 15 December 2023  
Description : Water Bath  
Manufacturer : ESSTELL Model : EWB-122D  
Serial No. : 20180508122 ID. No. : WWL 0214  
Marking : Additionally for the purpose of identification by this laboratory a label marked with this certificate number ( MC 2314268 ) has been attached to the case.  
Method : In-House calibration procedure MWI-T-029 this method is reference to ASTM E715 "Liquid Bath".  
Location of Calibration : Water Analysis Center Co., Ltd. ; Laboratory.  
Environmental Condition : Ambient Temperature : ( 29.4 to 29.8 ) °C  
Relative Humidity : ( 49.0 to 52.0 ) %  
Date of Calibration : 15 December 2023 Date of Issue : 19 December 2023

Checked by : Chalermkit Rakphada

( Calibration Engineer )

Approved by : Aittipong

Aittipong Karjanasit  
( Technical Manager )

The uncertainties are for a confidence probability of approximately 95%

This certificate is issued in accordance with the conditions of accreditation granted by the National Standardization Council of Thailand-Office of the National Standardization Council that has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of Master Calibration Co.,Ltd.

[MCF-Q-077 ; Rev.6 ; Date : 22/04/2021]

Certificate No.: MC 2314268

Page 2 of 3

#### Reference Standard Instrument :

Description	Certificate No.	Serial No.	Due date	Traceable thru
Data Acquisition/Switch Unit With Thermocouple Type " T " ID. No.27/1 to 27/5	MC 2301270	MY44020009	9 Mar 2024	MCAL

#### Traceability :

The measurement standard traceable to the international system of units (SI) through certificate as mentioned above


#### 1. Calibration Procedure:

This Instrument was calibration according to ASTM E715 - 2007 by comparison with calibrated sensor under no load condition. The sensor were placed on five points and located one sensor in each of the eight corners of the chamber and was away from the each wall of 5 cm to 10 cm. And placed the five sensor within 2.5 cm of the geometric center of the chamber.

*Temperature Uniformity* - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.

*Temperature Stability* - one-half of the greatest maximum difference of measured temperatures at any one sensor.

*Overall Variation* - The Difference of the maximum and minimum measured temperatures throughout observation.



Checked by : Chalermkit

[MCF-Q-077 ; Rev.6 ; Date : 22/04/2021]

Certificate No.: MC 2314268

Page 3 of 3

#### 2. Result of calibration :

##### Temperature Measurement Accuracy Test

Indicating Temperature (°C)	Measured Temperature (°C) at Spread Locations					Uncertainty (±°C)
	#1	#2	#3	#4	Ref. #5	
45.0	44.5	44.4	44.5	44.5	44.6	0.45

##### Chamber Characterization Result

Desired Temperature (°C)	Controller Temperature (°C)	Indicating Temperature (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
44.5	45.0	45.0	0.62	0.88	1.5

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor  $k = 2.0$ , providing a level of confidence of approximately 95 %.

**This certificate will certify of the calibrated equipment only.**


**End of Certificate**

Checked by : Chalermkit

[MCF-Q-077 ; Rev.6 ; Date : 22/04/2021]

## Certificate of Calibration

### TEMPERATURE CONTROLLER ENCLOSURES



Certificate No.: MC 2314270

Customer : Water Analysis Center Co., Ltd.  
1/94 Moo 5, T. Kantham, A.U-Thai, Ayuthaya 13210.

Reference Job No. : 23-2833  
Description : Incubator  
Manufacturer : Memmert  
Serial No. : D619/0170  
Model : IN260  
Marking : ID. No. : WWL 0192  
Method : Additionally for the purpose of identification by this laboratory a label marked with this certificate number ( MC 2314270 ) has been attached to the case.  
Location of Calibration : In-House calibration procedure MWI-T-033 this method is reference to TLAS G-20 "Temperature Controlled Enclosures".  
Environmental Conditions : Water Analysis Center Co., Ltd. ; Laboratory.  
Ambient Temperature : ( 25.2 to 25.6 ) °C  
Relative Humidity : ( 65.4 to 66.2 ) %  
Date of Calibration : 15 December 2023  
Date of Issue : 19 December 2023

Checked by : **Chalermkit**  
Chalermkit Rakphada  
( Calibration Engineer )

Approved by : **Aittipong**  
Aittipong Kahanjanwasit  
( Technical Manager )

**The uncertainties are for a confidence probability of approximately 95%**

This certificate is issued in accordance with the conditions of accreditation granted by the National Standardization Council of Thailand-Office of the National Standardization Council that has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of Master Calibration Co.,Ltd.

Certificate No.: MC 2314270

Page 2 of 3

### Reference Standard Instrument :

Description	Certificate No.	Serial No.	Due date	Traceable thru
Data Acquisition/Switch Unit	MC 2214032	MY41029992	26 Dec 2023	MCAL
With Thermocouple Type " T " ID. No.31/1 to 31/9				

### Traceability :

The measurement standard traceable to the international system of units (SI) through certificate as mentioned above


### 1. Calibration Procedure:

This Instrument was calibration according to TLAS G-20 by comparison with calibrated thermocouple type T under no load condition. The Thermocouples were placed on nine points and located one thermocouple in each of the eight corners of the chamber and was away from the each wall of 5 cm to 10 cm. And placed the ninth thermocouple within 2.5 cm of the geometric center of the chamber.

**Temperature Uniformity** - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.

**Temperature Stability** - one-half of the greatest maximum difference of measured temperatures at any one sensor.

**Overall Variation** - The Difference of the maximum and minimum measured temperatures throughout observation.



Checked by : **Chalermkit**



Certificate No.: MC 2314270

Page 3 of 3

## 2. Result of calibration :

### Temperature Measurement Accuracy Test

Indicating Temperature (°C)	Measured Temperature (°C) at Spread Locations									Uncertainty (±°C)
	#1	#2	#3	#4	#5	#6	#7	#8	Ref. #9	
35.0	35.2	35.2	35.2	35.2	35.1	35.1	35.0	35.1	35.1	0.44

### Chamber Characterization Result

Desired Temperature (°C)	Controller Temperature (°C)	Indicating Temperature (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
35.0	35.0	35.0	0.13	0.21	0.4

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95 %.

**This certificate will certify of the calibrated equipment only.**


**End of Certificate**

Checked by : **Chalermit**

[MCF-Q-077 ; Rev.6 ; Date : 22/04/2021]

# Certificate of Calibration

## AUTOCLAVE



Certificate No.: MC 2314269

Page 1 of 3



Customer : Water Analysis Center Co., Ltd.  
1/94 Moo 5, T.Kantham, A.U-Thai, Ayutthaya 13210.

Reference Job No. : 23-2833 Received Date : 15 December 2023  
Description : Autoclave  
Manufacturer : TOMY Model : Autoclave ES-315  
Serial No. : 51135128 ID. No. : WWL 0083  
Marking : Additionally for the purpose of identification by this laboratory a label marked with this certificate number (MC 2314269) has been attached to the case.  
Method : In-House calibration procedure MWI-TI-036 this method is reference to based on BS 2646 : 1993 Part 5 "Autoclave".  
Location of Calibration : Water Analysis Center Co., Ltd. ; Laboratory.  
Environmental Condition : Ambient Temperature : ( 29.4 to 30.7 ) °C  
Relative Humidity : ( 50.0 to 52.0 ) %  
Date of Calibration : 15 December 2023 Date of Issue : 19 December 2023

Checked by : **Chalermit**  
Chalermit Rakphada  
( Calibration Engineer )

Approved by : **Aittipong**  
Aittipong Kanjanakrasit  
( Technical Manager )

**The uncertainties are for a confidence probability of approximately 95%**

This certificate is issued in accordance with the conditions of accreditation granted by the National Standardization Council of Thailand-Office of the National Standardization Council that has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of Master Calibration Co.,Ltd.

[MCF-Q-077 ; Rev.6 ; Date : 22/04/2021]

Certificate No.: MC 2314269

Page 2 of 3

## Reference Standard Instrument :

Description	Certificate No.	Serial No.	Due date	Traceable thru
Temperature Recorder RTD 100 Ohm	MC 2300163	M79252	9 Jan 2024	MCAL
Temperature Recorder RTD 100 Ohm	MC 2300164	5978194	9 Jan 2024	MCAL
Temperature Recorder RTD 100 Ohm	MC 2300165	M79251	9 Jan 2024	MCAL

## Traceability :


The measurement standard traceable to the international system of units (SI) through certificate as mentioned above

## 1. Calibration Procedure:

The equipment list above was calibrated an accuracy of temperature in a chamber of the sterilizer.

The calibration was performed by direct measurement of generated temperatures using the standard thermometer with three temperature sensors. The data was recorded in a period of fifteen minutes of the sterilizing status. The temperature scale used was based on ITS-90.

The calibration of sterilizer was carried out at the point indicated by following the In-house calibration method No. MWI-T-036 based on BS 2646 : 1993 : Part 5 in Tests for performance section.



Checked by : Chalermthip

[MCF-Q-077 ; Rev.6 ; Date : 22/04/2021]

Certificate No.: MC 2314269

Page 3 of 3

## 2. Result of calibration :

### Temperature Measurement Accuracy Test

Indicating Temperature (°C)	Measured Temperature (°C) at Spread Locations			Uncertainty (±°C)
	#1	#2	#3	
121	121.72	121.73	121.95	0.61

### Characterization Result

Desired Temperature (°C)	Setting Temperature (°C)	Timer Setting (min)	Indicating Temperature (°C)	Indicating Pressure (kPa)	Measured Stability (±°C)	Measured Uniformity (°C)	Overall Variation (°C)
121	121	15.0	121	120	0.60	0.35	1.35

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95 %.

This certificate will certify of the calibrated equipment only.

End of Certificate

Checked by : Chalermthip


[MCF-Q-077 ; Rev.6 ; Date : 22/04/2021]

High Volume Air Sampler Calibration Worksheet

Page 1 of 1

Project Site : โนนเปลือย ร้อยตรี  
Location : ภายในพื้นที่โครงการ  
Date of measurement : 24/5/2024  
Worksheet No. : C-240524-WWL0094  
High Volume ID : WWL0103  
High Volume Model : TE-5170 (TSP)  
High Volume S/N : 2736  
Ambient Condition : 27/03/2024  
Temperature (°C) : 26  
Barometric Pressure (mmHg) : 756  
Quality Standard Intercept : -0.01922

Test No.	delta H <sub>2</sub> O (inch)	Q <sub>ad</sub> (m <sup>3</sup> /min)	I (Chart)	IC (Corrected)	Linear Regression
1	4.90	1.397	50.0	49.80	Slope : 33.22 Intercept : 3.155 Correlation Coefficient : 0.9995
2	4.40	1.324	47.0	46.81	
3	3.20	1.131	41.0	40.83	
4	2.40	0.981	36.0	35.85	
5	2.00	0.897	33.0	32.87	



Calibrated by :

Approved by :

Mr. JITTAWEE WONGMAKHEEB


Mr. RUNGSASIKORN KOSUM

High Volume Air Sampler Calibration Worksheet

Page 1 of 1

Project Site : โนนเปลือย ร้อยตรี  
Location : ภายในพื้นที่โครงการ  
Date of measurement : 24/5/2024  
Worksheet No. : C-240524-WWL0099  
High Volume ID : WWL0099  
High Volume Model : TE-6070 (PM10)  
High Volume S/N : 2732  
Ambient Condition : 27/03/2024  
Temperature (°C) : 26  
Barometric Pressure (mmHg) : 756  
Quality Standard Intercept : -0.01199

Test No.	delta H <sub>2</sub> O (inch)	Q <sub>ad</sub> (m <sup>3</sup> /min)	I (Chart)	IC (Corrected)	Linear Regression
1	5.00	1.422	50.0	31.44	Slope : 26.22 Intercept : -5.977 Correlation Coefficient : 0.9996
2	4.10	1.289	44.0	27.67	
3	3.50	1.192	40.0	25.15	
4	2.90	1.086	36.0	22.64	
5	2.20	0.947	30.0	18.87	



Calibrated by :

Approved by :

Mr. JITTAWEE WONGMAKHEEB


Mr. RUNGSASIKORN KOSUM



### Sulfur Dioxide Analyzer Calibration Worksheet

Project Site :	โนบิล รีโวลฟ์ รัชดา	Multi Gas Calibrator
Location :	ภายในพื้นที่โครงการ	Calibrator ID : WWL0124
Date of measurement :	24 May 2024	Calibrator Model : Series 6100
Worksheet No. :	C-240524-WWL 0111	Calibrator S/N : S/N 7462
Ambient SOx Analyzer ID :	WWL 0111	Calibrate Date : 08 March 2024
Manufacturer :	HORIBA	Cylinder Std. Gas
Ambient SOx Analyzer Model :	APSA-370	Std. Gas Concentration (PPM) : 50.90
Ambient SOx Analyzer S/N :	PGRKTBDX	Cylinder Pressure (psi) 2000
		Certified Date : 07 December 2021
		Expired Date : 07 December 2025
		Serial No. : CC241587

Point	CALIBRATION RESULTS			
	Ideal	Actual SOx	Error Sox	%Error Sox
ZERO	0.00	0.10	0.10	-
SPAN 200 ppb	200.00	200.20	0.20	0.10
SPAN 400 ppb	400.00	400.10	0.10	0.03
AVERAGE (%)				
0.06				




Calibrated by สุธิดา (Miss SUTHIDA SINGHAPHEN) Chemist  
Approved by Rungsasikorn Kosum (Mr. RUNGSASIKORN KOSUM) Technical Management

### Nitrogen Dioxide Analyzer Calibration Worksheet

Project Site :	โนบิล รีโวลฟ์ รัชดา	Multi Gas Calibrator
Location :	ภายในพื้นที่โครงการ	Calibrator ID : WWL0124
Date of measurement :	24 May 2024	Calibrator Model : Series 6100
Worksheet No. :	C-240524-WWL 0114	Calibrator S/N : S/N 7462
Ambient NOx Analyzer ID :	WWL 0114	Calibrate Date : 08 March 2024
Manufacturer :	HORIBA	Cylinder Std. Gas
Ambient NOx Analyzer Model :	APNA-370	Std. Gas Concentration (PPM) : 50.90
Ambient NOx Analyzer S/N :	PIEJ99E5	Cylinder Pressure (psi) 2000
		Certified Date : 07 December 2021
		Expired Date : 07 December 2025
		Serial No. : CC241587

Point	CALIBRATION RESULTS			
	Ideal	Actual NO	Error NO	%Error NO <sub>x</sub>
ZERO	0.00	0.10	0.10	0.00
SPAN 200 ppb	200.00	200.10	0.10	0.05
SPAN 400 ppb	400.00	400.10	0.10	0.20
AVERAGE (%)				
0.04				



Calibrated by สุธิดา (Miss SUTHIDA SINGHAPHEN) Chemist  
Approved by Rungsasikorn Kosum (Mr. RUNGSASIKORN KOSUM) Technical Management