

ภาคผนวก ค-3

เอกสารสอบเทียบเครื่องมือที่ใช้ในการวิเคราะห์



Certificate of Calibration

Certificate No. : 66-420087-1

Page : 1 of 2

Submitted by :

M Green Group Co.,Ltd.

188/46 Wisatsukhakhon 25, Pracha-Utd Rd., Thungkru Bangkok 10140 Thailand

Equipment :

pH Meter with electrode

pH meter

Manufacturer : Eutech

Model : pH 700

pH

Resolution : 0.01

ID No. : N/A

Serial No. : 2884323

Electrode

Model : N/A

Serial No. : 01X099320

Environment :

On site calibration was carried out at the Laboratory, M Green Group Co.,Ltd.

Ambient Temperature : (25.0 to 25.5)°C

Relative Humidity : (45 to 50) %

Date of Received : 20 September 2023

Date of Calibration : 20 September 2023

Date of Issue : 25 September 2023

Calibrated by : Permpon Chanpu

Calibration Method : In-house method CAL-M4201 direct measurement by using standard voltage calibrator and using certified reference material (CRM)

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Multiproduct Calibrator

ID No.	Cert.No.	Due.Date	Traceability
400005	SG-E-00307/66	23 Aug 2025	National Institute of Metrology Thailand (NIMT)
2. Standard Buffer Solution			
pH	Cert. No.	Lot No.	Exp.Date
4.008	61270213	915161	19 Jul 2025
6.985	61275614	898428	28 May 2024
9.997	61281073	915163	19 Jul 2024

Approved by

(Surachai Promthong)

Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 66-420087-1

Page : 2 of 2

Result of Calibration :

UUC Condition As-Received : Good

Function : Electrical measurement
pH meter

Performing standard curve by Multiproduct Calibrator at pH (4,7,10)

Adjustment Curve at nominal pH	Applied Voltage (mV)	Nominal Value (pH)	UUC Reading		Correction (mV)	Uncertainty (± mV)
			(pH)	(mV)		
4, 7, 10	177.4800	4	4.00	177.5	0.0	0.12
	0.0000	7	7.00	0.1	-0.1	0.086
	-177.4800	10	10.00	-177.4	-0.1	0.12

Function : pH meter with electrode

Performing a three - buffer standard curve using buffer nominal pH (4,7,10)

Adjustment Curve at nominal pH	Standard Buffer (pH)	UUC Reading		Correction (pH)	Uncertainty (± pH)
		(pH)	(pH)		
4, 7, 10	4.008	4.01	0.00	0.00	0.010
	6.985	7.00	-0.01	-0.01	0.011
	9.997	10.01	-0.01	-0.01	0.014

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

-o/o-



NSC-TIS-17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : **66-400519-1** Page : 1 of 2Submitted by : M Green Group Co.,Ltd.
188/46 Wisatsukhakhon 25, Pracha-Utd Rd., Thungkru Bangkok 10140 ThailandEquipment : Digital Thermometer with Thermistor probe
Temperature IndicatorManufacturer : Eutech Model : pH 700
Range : N/A °C Resolution : 0.1 °C
Serial No. : 2884323 ID No. : N/A

Thermistor probe

Model : N/A Sheath Material : Stainless
Diameter : 3.2 mm. Length : 100 mm.
Serial No. : PH5TEMB01P ID No. : N/A

Environment : On site calibration was carried out at the M Green Group Co.,Ltd.

Ambient Temperature : (25.0 to 26.0) °C
Relative Humidity : (56 to 60) %
Line Voltage : (224.0 to 225.2) VACDate of Received : 20 September 2023
Date of Calibration : 20 September 2023
Date of Issue : 25 September 2023
Calibrated by : Permpon ChanpuCalibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4003
by compared with PRT in the liquid bath at the constant controlled temperature.

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Platinum Resistance Thermometer (PRT)

ID No. Cert. No. Due Date Traceability
400002 TT-0074-22 20 Jun 2024 National Institute of Metrology Thailand (NIMT)

2. Standard Digital Thermometer

ID No. Cert. No. Due Date Traceability
400033 22E569 22 Feb 2024 National Institute of Metrology Thailand (NIMT)Approved by : 
(Surachai Promthong)
Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : **66-400519** Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

Immersion Depth (mm.)	Standard Reading (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
100	25.006	24.9	0.1	0.19

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$,
providing a level of confidence of approximately 95%

oDo





Certificate of Calibration

Certificate No. : 66-200300-1

Page : 1 of 2

Submitted by :

M Green Group Co., Ltd.

188/46 Wisetukulnakhon 25, Pracha-Utt Rd., Thungkru, Bangkok 10140 Thailand

Equipment :

Electronic Balance

Manufacturer : SHIMADZU

Model : AP225WD

Serial No. : D316300690

Capacity : 220 g

Resolution : 0.00001g/102g, 0.0001g/220g

Environment :

On site calibration was carried out at the Laboratory, M Green Group Co., Ltd.

Ambient Temperature : (25.6 to 26.7) °C

Relative Humidity : (54.4 to 56.6) %

Air Pressure : 1010.0 mbar

Date of Received :

20 September 2023

Date of Calibration :

20 September 2023

Date of Issue :

22 September 2023

Calibrated by :

Akaradath Thippichai

Calibration Method :

In-house method CAL-M2001 based on UKAS Publication ref : LAB 14

Edition 7 - November 2022

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Weights

ID No.

Cert. No.

Due Date

Traceability

E261-E264

C0222345

10 Nov 2023

National Institute of Metrology (Thailand), (NIMTT)

Approved by :

(Surachai Promthong)

Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.



Certificate of Calibration

Certificate No. : 66-200300-1

Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Departure of indication from nominal value

Nominal Value (g)	Correction (g)	Uncertainty ± (g)
0.001	0.00000	0.000012
0.01	0.00000	0.000013
0.1	0.00000	0.000015
1	0.00000	0.000026
10	0.00000	0.000053
20	-0.00003	0.000071
50	0.00004	0.00011
100	-0.00009	0.00020
150	0.0000	0.00038
200	-0.0001	0.00038

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2.15$ providing a level of confidence of approximately 95%

Eccentric error

Load test : 50 g



A B C D E

-0.00003 0.00000 0.00000 -0.00005 0.00000 g

Repeatability

Load test

200 g

Sidev.

0.000048 g

-0.00 -





Certificate of Calibration

Certificate No. : 66-400531-1

Page : 1 of 2

Submitted by :

M Green Group Co., Ltd.

188/46 Wisetuesuknakhon 25, Pracha-Uttd Rd., Thungkru, Bangkok 10140 Thailand

Equipment :

Liquid in Glass Thermometer

Manufacturer : N/A

Model : N/A

Range : 0 °C to 100 °C

Resolution : 1 °C

Serial No. : N/A

Immersion : Total

ID No. : 94-49747

Environment :

Ambient Temperature : (23 ± 2) °C

Relative Humidity : (50 ± 15) %

Line Voltage : (220 ± 22) VAC

Date of Received : 21 September 2023

Date of Calibration : 23 September to 26 September 2023

Date of Issue : 26 September 2022

Calibrated by : Chortip Sanchusri

Calibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4001 based on ASTM E77-07 by compared with PRT in the liquid bath at the constant controlled temperature.

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Platinum Resistance Thermometer (PRT)

ID No. Cert.No. Due Date

400001 TT-0016-22 07 Feb 2024

Traceability
National Institute of Metrology Thailand (NIMT)

2. Standard Digital Thermometer

ID No. Cert.No. Due Date

400003 23E1866 01 Jun 2025

Traceability
National Institute of Metrology Thailand (NIMT)

400004 23E1866 01 Jun 2025

Approved by :

(Surachai Promthong)

Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 66-400531-1

Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

Ice point check : UUC* reading 0 °C Standard reading 0.0352 °C

Standard Reading (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
39.7228	40	-0.3	0.31

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- o O o -





Certificate of Calibration

Certificate No. : 66-400520-1

Page : 1 of 2

Submitted by : M Green Group Co., Ltd.

188/46 Wisutesukhakhon 25, Prachia-Utd Rd., Thungkru Bangkok 10140 Thailand

Equipment : Air Chamber (Refrigerator)

Manufacturer : Biobase

Model : BXC-V250M (II)

Range : N/A °C

Resolution : 0.1 °C

Serial No. : YC025025190108

ID No. : N/A

Environment : On site calibration was carried out at the Laboratory, M Green Group Co., Ltd.

Ambient Temperature : (25.0 to 26.0) °C

Relative Humidity : (40 to 50) %

Line Voltage : (226.0 to 230.0) V

Date of Received : 20 September 2023

Date of Calibration : 20 September 2023

Date of Issue : 25 September 2023

Calibrated by : Permon Chanpu

Calibration Method : CAL-M4004, TLAS G-20

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Digital Thermometer with RTD Probe

ID No. : Cert.No. Due Date Traceability:

400046 & 400042 66-400453-1 31 Jan 2024 National Institute of Metrology Thailand (NIMT)

Approved by :

(Surachai Promthong)

Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.



Certificate of Calibration

Certificate No. : 66-400520-1


Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber
 W = 0.50 m
 D = 0.40 m
 H = 1.20 m
 Capacity = 0.24 m³

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (± °C)
			1	2	3	4	5	6	7	8	9	
4.0	2.0	2.0	4.05	4.04	4.27	4.89	4.10	4.05	4.92	4.37	4.43	0.46

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured		Overall Variation (°C)
			Uniformity (°C)	Stability (°C)	
4.0	2.0	2.0	0.60	0.21	1.2

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

-oOo-



NSG-TLSI-TS17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 66-400520-2 Page : 1 of 2

Submitted by : M Green Group Co., Ltd.
188/46 Wisetukhakhon 25, Pracha-Utt Rd., Thungkru Bangkok 10140 ThailandEquipment : Air Chamber (Oven)
Manufacturer : Memmert
Range : N/A °C
Serial No. : B419.1092
Model : UF110
Resolution : 0.1 °C
ID No. : N/A

Environment : On site calibration was carried out at the Laboratory, M Green Group Co., Ltd.

Ambient Temperature : (25.0 to 26.0) °C
Relative Humidity : (40 to 50) %
Line Voltage : (226.0 to 230.0) VDate of Received : 20 September 2023
Date of Calibration : 20 September 2023
Date of Issue : 25 September 2023
Calibrated by : Pernpon Chianpu

Calibration Method : CAL-M4004, TLAS G-20

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units
Standard Digital Thermometer with Thermocouple probeID No. 400046 & 400028
Cert. No. 66-400184-3
Due Date 04 Oct 2023
Traceability National Institute of Metrology Thailand (NIMT)

Approved by :

(Surachai Promthong)
Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd



Certificate of Calibration


Certificate No. : 66-400520-2 Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)

Inside of Chamber
W = 0.56 m
D = 0.40 m
H = 0.48 m
Capacity = 0.11 m³

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (± °C)
			1	2	3	4	5	6	7	8	9	
103.0	103.0	103.0	103.3	103.0	103.7	103.3	103.1	103.0	103.8	102.7	103.3	0.69
105.0	105.0	105.0	105.3	105.0	105.7	105.3	105.2	105.0	105.8	104.6	105.3	0.71
180.0	180.0	180.0	180.4	180.1	181.2	180.4	180.3	180.0	181.4	179.0	180.5	0.95

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Uniformity (°C)	Measured Stability (°C)	Overall Variation (°C)
103.0	103.0	103.0	0.8	0.1	1.3
105.0	105.0	105.0	0.9	0.1	1.4
180.0	180.0	180.0	1.7	0.2	2.7

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

-o0o-



Certificate of Calibration

Certificate No. : 66-400520-3
 Page : 1 of 2

Submitted by : M Green Group Co., Ltd.
 188/46 Wisatsukhakhon 25, Prachia-Uttd Rd., Thungkru Bangkok 10140 Thailand

Equipment : Water Bath
Manufacturer : Memmert
Model : WNB29
Range : N/A °C
Resolution : 0.1 °C
Serial No. : L619,0037
ID No. : N/A

Environment : On site calibration was carried out at the Laboratory, M Green Group Co., Ltd.
Ambient Temperature : (25.0 to 26.0) °C
Relative Humidity : (40 to 50) %
Line Voltage : (226.0 to 230.0) V

Date of Received : 20 September 2023
Date of Calibration : 20 September 2023
Date of Issue : 25 September 2023
Calibrated by : Permpoon Chanpu

Calibration Method : This instrument was calibrated by In-house method CAL-M4006 based on ASTM E715-80
 The temperature scale used was based on ITS-90


Reference Standard Instruments : This certification is traceable to the International System of Units
Standard Digital Thermometer with RTD probe
ID No. **Cert. No.** **Due Date** **Traceability**
 400046 & 400024 66-400184-2 06 Oct 2023 National Institute of Metrology Thailand (NIMT)

Approved by : 
 (Surachai Promthong)
 Laboratory Manager

Certificate of Calibration

Certificate No. : 66-400520-3
 Page : 2 of 2

Result of Calibration : Without Adjustment
UUC Condition As-Received : Good
Function : Temperature measurement



Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.					Uncertainty (± °C)	Measured Uniformity (°C)	Measured Stability (°C)
			1	2	3	4	5			
85.0	85.0	85.0	85.08	85.04	84.98	85.17	85.02	0.18	0.2	0.05

Remark The uncertainty is not combine uniformity of the water bath
 This result of calibration was found accurate as shown on date and place of calibration only.
 This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

-o0o-

AP



Certificate of Calibration

Certificate No. : 66-400520-4

Page : 1 of 2

Submitted by :

M Green Group Co., Ltd.

188/46 Wisatsukhakhon 25, Pracha-Utd Rd., Thungkru Bangkok 10140 Thailand

Equipment :

Air Chamber (Incubator)

Manufacturer : Biobase

Model : Biochemistry Incubator

Range : 0 °C to 65 °C

Resolution : 0.1 °C

Serial No. : KYP1502202003

ID No. : N/A

Environment :

On site calibration was carried out at the Laboratory, M Green Group Co., Ltd.

Ambient Temperature : (25.0 to 25.5) °C

Relative Humidity : (45 to 50) %

Line Voltage : (226.0 to 230.0) V

Date of Received : 20 September 2023

Date of Calibration : 20 September 2023

Date of Issue : 25 September 2023

Calibrated by : Permpon Chanpu

Calibration Method : CAL-M4004, TLAS G-20

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Digital Thermometer with RTD Probe

ID No.

Traceability

Cert.No.

National Institute of Metrology Thailand (NIMT)

400029 & 400043

Due Date

27 Oct 2023

Approved by :

(Surachai Promthong)

Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd



Certificate of Calibration

Certificate No. : 66-400520-4


Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (± °C)
			1	2	3	4	5	6	7	8	9	
20.0	20.0	20.0	20.14	20.04	19.91	19.97	20.03	19.96	19.91	19.96	19.92	0.70

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured		Overall Variation (°C)
			Uniformity (°C)	Stability (°C)	
20.0	20.0	20.0	0.28	0.37	0.8

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

-oOo-





Certificate of Calibration

Certificate No. : 66-300589-7

Page : 1 of 2

Submitted by

: M Green Group Co., Ltd.

188/46 Wisutesukrakhon 25, Pracha-Utd Rd., Thungkru, Bangkok 10140 Thailand

Equipment

: Burette

Manufacturer : GLASSCO

Class : A

Capacity : 10 ml Graduation : 0.05 ml

ID No. : 2212-0344-1

Environment

: Ambient Temperature : (20 ± 3) °C

Relative Humidity : (50 ± 10) %

Air Pressure : 1006.7 mbar.

Date of Received

: 20 September 2023

Date of Calibration

: 27 September 2023

Date of Issue

: 27 September 2023

Calibrated by

: Wipa Tovadee

Calibration Method : In-house method CAL-M3001 based on ASTM E 542-01

Reference Standard Instruments : This certification is traceable to the International System of Units

Electronic Balance

ID No.

Cert.No.

Due Date

Traceability

241003

66-200196-2

02 Dec 2023

National Institute of Metrology (Thailand) (NIMT)

Approved by :

(Wipa Tovadee)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.



Certificate of Calibration

Certificate No. : 66-300589-7

Page : 2 of 2

Result of Calibration : This result of true Volume is referred to standard temperature at 20 °C

UUC Condition As-Received : Good

Delivery Time : 21.33 sec.

Nominal Volume (ml)	Measuring Volume (ml)
10	9.9913

Uncertainty of measurement with in ± 0.0039 ml

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2.00 providing a level of confidence of approximately 95%

- o0o -





Certificate of Calibration

Certificate No. : 66-300589-8

Page : 1 of 2

Submitted by

: M Green Group Co., Ltd.

188/46 Wisatesuknakhon 25, Pracha-Utd Rd., Thungkru, Bangkok 10140 Thailand

Equipment

: Burette

Manufacturer : GLASSCO

Class : A

Capacity : 25 ml

Graduation : 0.1 ml

ID No. : 2212-0344-2

Environment

: Ambient Temperature : (20 ± 3) °CRelative Humidity : (50 ± 10) %

Air Pressure : 1006.7 mbar.

Date of Received

: 20 September 2023

Date of Calibration

: 27 September 2023

Date of Issue

: 27 September 2023

Calibrated by

: Wipa Tovadee

Calibration Method : In-house method CAL-M3001 based on ASTM E 542-01

Reference Standard Instruments : This certification is traceable to the International System of Units

Electronic Balance

ID No.

Cert.No.

Due Date

Traceability

241003

66-200196-2

02 Dec 2023

National Institute of Metrology (Thailand) (NIMT)

Approved by :

(Wipa Tovadee)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd



Certificate of Calibration

Certificate No. : 66-300589-8

Page : 2 of 2

Result of Calibration : This result of true Volume is referred to standard temperature at 20 °C

UUC Condition As-Received : Good

Delivery Time : 46.01 sec.

Nominal Volume (ml)	Measuring Volume (ml)
25	24.9741

Uncertainty of measurement with in $\pm$ 0.0066 ml

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2.00$, providing a level of confidence of approximately 95%

- o o o -



Certificate of Calibration

Certificate No. : 66-300590-1

Page : 1 of 2

Submitted by

: M Green Group Co., Ltd.

188/46 Wisatesuknakhon 25, Pracha-Uttd Rd., Thungkru, Bangkok 10140 Thailand

Equipment

: Imhoff Cone

Manufacturer : VITLAB

Capacity : 1000 ml Graduation : 50 ml

ID No. : CY1000/01/22

Environment

: Ambient Temperature : (20 ± 3) °C

Relative Humidity : (50 ± 10) %

Air Pressure : 1005.4 mbar.

Date of Received : 20 September 2023

Date of Calibration : 26 September 2023

Date of Issue : 26 September 2023

Calibrated by : Arcerat Sombun

Calibration Method : In-house method CAL-M3001 based on ASTM E 542-01

Reference Standard Instruments : This certification is traceable to the International System of Units

Electronic Balance

ID No. Cert.No. Due Date

241002 66-200196-1 02 Dec 2023

Traceability

National Institute of Metrology (Thailand) (NIMT)

Approved by :


(Wipa Tovadce)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.



Certificate of Calibration

Certificate No. : 66-300590-1

Page : 2 of 2

Result of Calibration : This result of true Volume is referred to standard temperature at 20 °C

UUC Condition As-Received : Good

Nominal Volume (ml)	Measuring Volume (ml)
500	501.19
1000	1010.67

Uncertainty of measurement with in ± 0.17 ml

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2.00$, providing a level of confidence of approximately 95%

-o0o-



การดูแลบำรุงรักษาเชิงป้องกัน Preventive Maintenance

เงื่อนไขการให้บริการ Preventive Maintenance

Type text here

บริษัทฯ จะส่งวิศวกรผู้ชำนาญ เพื่อให้บริการตามขอบข่ายของการบริการ เฉพาะ ในวันและเวลา ราชการ หากมีความประสงค์ที่จะรับบริการนอกเหนือจากวัน เวลา ราชการ (วันหยุดเสาร์ – อาทิตย์ หรือวันหยุด นักชดถกณ์) บริษัทฯ จะคิดค่าบริการเพิ่มเติมตามอัตราที่กฎหมายแรงงานกำหนดไว้

ขอบข่ายการบริการ

- ตรวจสอบสภาพการทำงานของต่าง ๆ ของเครื่องมือ
- ทดสอบประสิทธิภาพการทำงานของเครื่องมือ
- รายงานผลการตรวจสอบเครื่องมือ

หมายเหตุ

- ราคาไม่รวมถึงค่าบริการซ่อม หรือ เปลี่ยนอะไหล่ที่ชำรุดเสียหาย หรือหมดสภาพการใช้งาน
- ในกรณีที่ผู้ใช้บริการอยู่นอกเขตพื้นที่ให้บริการ บริษัทฯ จำเป็นต้องคิดค่าใช้จ่ายเพิ่มเติม ได้แก่ ค่าเดินทาง เป็นต้น
- บริษัทฯ ขอสงวนสิทธิ์ในการเปลี่ยนแปลงราคา โดยไม่แจ้งให้ทราบล่วงหน้า



บริษัท ดีเคเอสเอช เทคโนโลยี จำกัด

ฝ่ายบริการหลังการขาย

โทร 0 2 639 7000 E-mail: service.tec.th@dksh.com

ฝ่ายขายและการตลาด

โทร 0 2 639 7000 E-Mail : marketing.tec.th@dksh.com

Website : www.dksh.co.th/technology/scientific-thailand



ช่องทางการติดต่อ

DKSH Technology Limited (บริษัท ดีเคเอสเอช เทคโนโลยี จำกัด)
เลขที่ 2533 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง กรุงเทพฯ 10260
เลขประจำตัวผู้เสียภาษี 010-555-001-4547 (สำนักงานใหญ่)



Call center 0 2 639 7000



DKSH Scientific



www.dksh.com/scientific-thailand



marketing.th@dksh.com



@dkshscientific

Preventive Maintenance Contract

จำนวนใบการที่สัญญาบริการ ...ครั้งที่ ...
ครั้งที่ 1 วันที่ 19/04/2024

รายละเอียดผู้รับบริการ

หน่วยงาน	บริษัท เทสท์ เทคโนโลยี จำกัด
ที่อยู่	30-32 ซอยพระรามที่ 2 ซอย 63 ถนนพระรามที่ 2 แขวงแสมดำ เขตบางขุนเทียน กรุงเทพมหานคร 10150
โทรศัพท์	0-2893-4211-7 แฟกซ์ 0-2893-4218

ผู้ติดต่อ

ชื่อ - นามสกุล	คุณกรกมล ขุนพิทักษ์
ตำแหน่ง	หัวหน้างาน
โทรศัพท์	087 398 9274 เบอร์มือถือ - แฟกซ์ -
E-mail	lab_center@testtech.co.th

รายละเอียดผู้ให้บริการ

บริษัท ดีเคเอสเอช เทคโนโลยี จำกัด (ฝ่ายบริการหลังการขาย) (สำนักงานใหญ่)
เลขที่ 2533 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง กรุงเทพฯ 10260
โทรศัพท์ 0 2 693 7000 Email: marketing.th@dksh.com
เจ้าหน้าที่ประสานงาน : คุณจุฑารัตน์ สิริรัตน์ โทรศัพท์ 090 678 6925

เจ้าหน้าที่ผู้ให้บริการ	นายธิราชพล ผลิตอาด	
ตำแหน่ง	Specialist, Technical Service.	
โทรศัพท์	0938138736	แฟกซ์ -
E-mail	jirayut.pjs@dksh.com	


ลงนามผู้รับบริการ	กรกมล ขุนพิทักษ์	ลงนามผู้ให้บริการ	นายธิราชพล ผลิตอาด
ตัวจริง	(ลงนามลงชื่อ)	ตัวจริง	(นายธิราชพล ผลิตอาด)
ตำแหน่ง	หัวหน้างาน	ตำแหน่ง	Specialist Technical Service
วันที่ / ประทับตราบริษัท	19/04/2024	วันที่ / ประทับตราบริษัท	19/04/2024

JOB No: LSPR2402440.....MODEL: VAP300.....S/N: GER5300210095

Operational Qualification (OQ)

ตรวจสอบการทำงานของ


FRONT



No		PASS	FAIL	N/A
1	Quick clamping device with clamping block	☒	☐	☐
2	Distillation tube 250/300 ml	☒	☐	☐
3	PTFE steam inlet tubing	☒	☐	☐
4	Connection stopper, Vifion	☒	☐	☐
5	Screw cap GL18	☒	☐	☐
6	PTFE-inlet tubing NaOH	☒	☐	☐
7	Distribution head made of glass	☒	☐	☐
8	Screw cap GL32	☒	☐	☐
9	Distillation condenser made of glass	☒	☐	☐
10	Screw cap GL14	☒	☐	☐
11	Ventilation valve	☒	☐	☐
12	Control panel	☒	☐	☐
13	Operating Button	☒	☐	☐
14	USB interface (with protective cap)	☒	☐	☐
15	Silicone tubing 8/10 for distillate discharge **	☐	☐	☒
16	Veriprene tubing 4/8 , receiver suction **	☐	☐	☒
17	Cable duct for electrode cable + titration tube**	☐	☐	☒
18	Silicone tubing 4/7 , boric acid inlet**	☐	☐	☒
19	Sensor for level monitoring including connector**	☐	☐	☒
20	Agitator motor with propeller**	☐	☐	☒
21	Titration acid inlet tube **	☐	☐	☒
22	Receiver glass**	☐	☐	☒
23	Holder for pH electrode , removable**	☐	☐	☒
24	pH electrode (combined electrode)**	☐	☐	☒
25	Drip tray PP	☒	☐	☐


** only VAP 450

REAR




No		PASS	FAIL	N/A
1	Tube connection for sample H3BO3 supply	☐	☐	☒
2	Tube connection for sample H2O supply	☒	☐	☐
3	Tube connection for steam generator H2O supply	☒	☐	☐
4	Tube connection for NaOH supply	☒	☐	☐
5	Tube connection for receiver glass extraction	☒	☐	☐
6	Tube connection for sample waste extraction	☒	☐	☐
7	Tube connection , overpressure steam outlet	☒	☐	☐
8	Connection for cooling water supply (with cleaning sieve)	☒	☐	☐
9	Tube connection for cooling water outlet	☒	☐	☐
10	4 X USB interface	☒	☐	☐
11	1 X RS-232 Interface	☒	☐	☐
12	LAN interface	☒	☐	☐
13	Screw cap for Perspex cover	☒	☐	☐
14	Connection socket for sample waste tank level monitoring	☒	☐	☐
15	Connection (not used)	☐	☐	☒
16	Connection socket for H2O tank level monitoring	☒	☐	☐
17	Connection socket for H3BO3 tank level monitoring	☒	☐	☐
18	Connection socket for NaOH tank level monitoring	☒	☐	☐
19	Overcurrent circuit breaker	☒	☐	☐
20	Apparatus socket (mains cable connection)	☒	☐	☐
21	Rating plate with serial number	☒	☐	☐
22	Exhaust air fan	☒	☐	☐
23	Excess temperature switch	☒	☐	☐

Inside Steam generator




Module Pump holder VAP200 - 450 V3



No		PASS	FAIL	N/A
1	Steam generator	☒	☐	☐
2	Steam generator traverse	☒	☐	☐
3	Pinch valve	☒	☐	☐
4	Circuit board distributor	☒	☐	☐
5	Valve tubing connection	☒	☐	☐
6	Housing safety valve	☒	☐	☐
7	Safety valve SKT	☒	☐	☐
8	Excess temperature protection , steam generator	☒	☐	☐
9	Safety valve G 1/8 0.5 bar	☒	☐	☐
10	Ventilation glass pinch valve VAPODEST	☒	☐	☐
11	Hose clamp for ventilation clamp	☒	☐	☐
12	Distributor PP	☒	☐	☐
13	Angle connection PP	☒	☐	☐
14	Pressure transmitter	☒	☐	☐
15	Level switch	☒	☐	☐
16	Fixing bracket steam generator	☒	☐	☐
17	Relay HT+	☒	☐	☐
18	VA Hexagon nut 1/4"	☒	☐	☐
19	Angle connection 1/8"	☒	☐	☐
20	Bushing nipple 6-10-14	☒	☐	☐
21	VA Lens head screw M5 X 10	☒	☐	☐
22	Grounding connection , 2-pole	☒	☐	☐
23	VA Lens head screw M4 X 6	☒	☐	☐
24	Spacer bolt 5 mm	☒	☐	☐
25	VA Lens head screw M4 X 10	☒	☐	☐
26	Tubing connection	☒	☐	☐
27	Hose clamp 14.5 mm	☒	☐	☐
28	Module ball valve with nozzles	☒	☐	☐
29	Cross manifold with spout	☒	☐	☐
30	Seal corner G 1/8	☒	☐	☐
31	Locking screw 1/8"	☒	☐	☐
32	Pin strip	☒	☐	☐
33	Bundle clamp 12 H 4500	☒	☐	☐
34	Bundle clamp 12 H 4502	☒	☐	☐
35	Temperature switch 80°C	☒	☐	☐
36	VA Lens head screw M3 X 6	☒	☐	☐
37	VA Hexagon nut M4	☒	☐	☐
38	Lens head screw M4 X 8	☒	☐	☐
39	VA Spring washer	☒	☐	☐
40	Angle connection , reduced , 1/8" PP	☒	☐	☐

No		PASS	FAIL	N/A
1	Peristaltic pump	☒	☐	☐
2	Diaphragm pump NaOH with non return valve	☒	☐	☐
3	Circuit board	☒	☐	☐
4	Tubing connection module	☒	☐	☐
5	Flow controller	☒	☐	☐
6	Lens head screw M5 x 10	☒	☐	☐
7	Bushing nozzle	☒	☐	☐
8	Screw in socket	☒	☐	☐
9	Magnetic valve 2/2 way	☒	☐	☐
10	Circuit board distributor	☒	☐	☐
11	Bushing nozzle	☒	☐	☐
12	Screw 5 x 25	☒	☐	☐
13	Cylinder screw	☒	☐	☐
14	Screw 5 x 20	☒	☐	☐
15	Seal EPDM 1.5 x 4	☒	☐	☐
16	Tubing connection piece 51x10x6.5	☒	☐	☐
17	Tubing connection piece 51x10x10	☒	☐	☐
18	Screw M4x10	☒	☐	☐
19	Clamp	☒	☐	☐
20	Clamp	☒	☐	☐
21	Y-tube connector	☒	☐	☐
22	Spacer bolt 5 mm	☒	☐	☐
23	Bundle clamp	☒	☐	☐
24	Bundle clamp	☒	☐	☐
25	Retrofit earthing pump/v	☒	☐	☐
26	Snap ferrite	☒	☐	☐
27	Nut G 3/8"	☒	☐	☐
28	Pump holder plate	☒	☐	☐

Control panel



No		PASS	FAIL
1	Title bar	☒	☐
2	Status bar	☒	☐
3	Navigation button	☒	☐
4	Smart switch with multiple functions	☒	☐
5	USB interface	☒	☐

- รายละเอียดการตรวจสอบขั้นตอนการบริการ
- ตรวจสอบระบบไฟฟ้า (Electrical Test)
- ความดันทางไฟฟ้าของเครื่องกับกรวด
 - กระแสไฟฟ้าที่ใช้งาน
- ตรวจสอบสภาพเครื่อง (Optical Test)
- Main cable
 - Electric wiring
 - Pumps
 - Distribution Head
 - Condensor
 - Steam generator
 - Tubing
 - Viton cone
- ตรวจสอบ Function การทำงาน (The Function Test)
- ระบบสร้างและความดันคั้นของ Steam
 - ระบบการเติมน้ำเข้า Sample Tube
 - ระบบการเติมน้ำ Na OH
 - ระบบการเติม H3BO3

รายงานผลการให้บริการ

1. TECHNICAL DATA

Main Supply 220 volt + 10% 50 Hz with ground
Nominal current

Pass ☒ Fail ☐ N/A ☐
Remark
.....
...8a.....

1.1 COOLING WATER BATH

Temperature 15-20 °C
Cooling Water Outlet
Control Temperature

Pass ☒ Fail ☐ N/A ☐
Remark
.....
.....

1.2 OPTICAL TESTV/AF300

Screw cap GL14
Screw cap GL18
Screw cap GL32
Distillation Head
Condensor
Viton Cone
Ventilation Valve BV
Micro Switch Sample
Agitator motor for propeller

Pass ☒ Fail ☐ N/A ☐
Remark
.....
.....
.....
.....
.....
.....
.....
.....
.....
.....
.....

2. SYSTEM COOLING WATER INLET

Cooling Water Inlet
Cooling Water Outlet
Flow control valve

Pass ☒ Fail ☐ N/A ☐
Remark
.....
.....
.....

3.SYSTEM CONTROL

Display
Program
Adding NaOH
Adding H2O
Adding H3BO3
Suction Sample
Suction Receiver

Pass ☒ Fail ☐ N/A ☐
Remark
.....
.....
.....
.....
.....
.....
.....

4.SYSTEM DISTILLATION

Boiler
Level Sensor
Novopren
Solenoid Valve Shut-Off
Solenoid Valve Steam
Solenoid Valve soft steam
Ventilation Valve Premount
Excess Pressure Detector
Heating Element

Pass ☒ Fail ☐ N/A ☐
Remark
.....
.....
.....
.....
.....
.....
.....
.....
.....
.....

5. PUMP

Pump H₂O Steam
- Non-Return Valve
Pump H₂O Sample
- Non-Return Valve
Pump NaOH
- Non-Return Valve
Pump H3BO3
- Non-Return Valve
Pump suction
Pump suction receiver

Pass ☒ Fail ☐ N/A ☐
Remark
.....
.....
.....
.....
.....
.....
.....
.....
.....
.....

6. The Following Program Run :

Addition H2O 0-999 ml.
Addition NaOH 0-999 ml.
Addition H3BO3 0-999 ml.
Reaction Time 0-108 min
Distillation Time 0-108 min
Steam Capacity 10%-100%
Suction Sample
Suction Receiver

Pass ☒ Fail ☐ N/A ☐
Remark
.....
.....
.....
.....
.....
.....
.....
.....
.....

7. Measured pumps

Pump NaOH
Pump H2O
Pump H3BO3

Volume :20.40.....ml
Volume :10.00.....ml
Volume :ml

Remark :

.....
.....

ข้อมูลสนับสนุนด้านเทคนิค (General Technical Support)

การบำรุงรักษาทั่วไป (Basic maintenance)

Cleaning program

Glass parts and tubes must be rinsed only before starting analysis in order to prevent clogging by crystallizing chemicals.

The following entries are recommended for this:

Reagents	Notes
H ₂ O addition	120 ml
H ₂ SO ₄ addition	6 ml
Distillation time	7 min
Steam power	100 %
Reaction time	6 h
Sediment sample	30 g

→ Invert a digestion tube (without magnets) and start the program.

→ All liquid carrying parts are cleaned. In the case of drying, the sample is dried at 60 °C or sulphuric acid can also be added to the digestion tube.

General error message

[illegible]

Analytical errors

[illegible]



QUALITY CALIBRATION CO., LTD.
235 Pachkasem 63/2 Road, Laksong, Bangkok 10160
Tel (662) 421-3402, (662) 444-0152-3, Fax (662) 809-4584
www.qcalibration.com



CERTIFICATE No : 23T8796
REFERENCE No : 70515-4

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : WATER BATH
MANUFACTURER : MEMMERT
MODEL : WNE 45
SERIAL No : L720.0266
ID No : EQL-241
CONDITION AS RECEIVED : USED ITEM
SUBMITTED BY : TEST TECH CO., LTD.
30,32 RAMA II SOI 63, RAMA II RD., SAMAEADAM,
BANGKHUNTHIAN, BANGKOK 10150

CALIBRATED BY : CHAICHARN CH.

CALIBRATION DATE : 16-Aug-23

APPROVED BY :  PONGSAK J.

ISSUED DATE : 16-Aug-23

RECEIVED DATE : 16-Aug-23

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.

F-G010 REV : 03



QUALITY CALIBRATION CO., LTD.
235 Pachkasem 63/2 Road, Laksong, Bangkok 10160
Tel (662) 421-3402, (662) 444-0152-3, Fax (662) 809-4584
www.qcalibration.com

CERTIFICATE No : 23T8796

PAGE : 2 OF 2


Calibration Report

EQUIPMENT : WATER BATH
MANUFACTURER : MEMMERT
ID NUMBER : EQL-241
RECEIVED DATE : 16-Aug-23
AMBIENT TEMPERATURE : 25 °C ± 1 °C
RELATIVE HUMIDITY : 50 %RH ± 10 % RH
MODEL : WNE 45
SERIAL NUMBER : L720.0266
CALIBRATION DATE : 16-Aug-23

CONDITION OF THIS RESULTS OF CALIBRATION

1. THIS INSTRUMENT WAS CALIBRATED ACCORDING TO ASTM E715-80 (REAPPROVED 2001) BY COMPARISON WITH CALIBRATED RTD. THE PROBES WERE PLACED ON FIVE POINTS AND LOCATED ONE PROBE IN EACH OF THE FOUR CORNERS OF THE BATH AND PLACED THE FIFTH RTD WITHIN 2.5 cm. OF THE GEOMETRIC CENTER OF THE WATER VOLUME (REFERENCE LOCATION) UNDER NO LOAD CONDITION.
2. REFERENCE STANDARD INSTRUMENTS :-

1) DATA LOGGER WITH RTD
2) REFERENCE TEMPERATURE : 25 °C ± 1 °C
3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.
4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.
5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-
- NATIONAL INSTITUTE OF METROLOGY (THAILAND) THROUGH QUALITY CALIBRATION CO., LTD.
RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT



PROBE INSTALLATION
POSITION IN THE BATH

GENERAL INFORMATION

Overall Variation of Ambient Temperature around the Bath (°C) : 0.6
Overall Variation of Line Voltage (V) : 3
Instrument Condition : Normal
Bath Inner Size (W*L*H) : 59*35*20 cm

BATH PERFORMANCE

Calibrate Point (°C)	Average All Position Temp. (±°C)	Temperature Stability (±°C)	Radius Uniformity (°C)	Axial Uniformity (°C)	Overall Variation (°C)
83.0	83.09	0.05	0.07	0.05	0.16
92.0	92.13	0.11	0.06	0.06	0.28

TEMPERATURE MEASUREMENT ACCURACY TEST

Controller Temp (°C)	Indicating Temp (°C)	Measured Temperature (°C) at Spread Locations	Uncertainty (± °C)
83.0	83.08	#1 83.09 #2 83.06 #3 83.11 #4 83.12	Ref. 5 0.15
92.0	92.11	83.13 92.10 92.16	92.16 0.19

NOTE 1 : THE UNCERTAINTY OF MEASUREMENT EXCLUDED TEMPERATURE UNIFORMITY OF THE BATH.
NOTE 2 : THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA.
THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR k =2, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.

END OF CALIBRATION REPORT

F-G010 REV : 03



QUALITY CALIBRATION CO., LTD.
235 Petkasem 63/2 Road, Lakson, Bangkok 10160
Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4584
www.qcalibration.com



CERTIFICATE No : 24T1185
REFERENCE No : 72116-3

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : INCUBATOR
MANUFACTURER : MEMMERT
MODEL : IF 160
SERIAL No : D518.0082
ID No : EQL-205
CONDITION AS RECEIVED : USED ITEM
SUBMITTED BY : TEST TECH CO., LTD.
30,32 RAMA II SOI 63, RAMA II RD., SAMAEDAM,
BANGKHUNTHIAN, BANGKOK 10150

CALIBRATED BY : CHAICHARN CH.
CALIBRATION DATE : 09-Feb-24

APPROVED BY :
PONGSAK J.
ISSUED DATE : 12-Feb-24
RECEIVED DATE : 09-Feb-24

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.

F-G010 REV : 03



QUALITY CALIBRATION CO., LTD.
235 Petkasem 63/2 Road, Lakson, Bangkok 10160
Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4584

CERTIFICATE No : 24T1185

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : INCUBATOR
MANUFACTURER : MEMMERT
MODEL : IF 160
ID No : D518.0082
RECEIVED DATE : 09-Feb-24
AMBIENT TEMPERATURE : 25 °C ± 1 °C
RELATIVE HUMIDITY : 53 %RH ± 10 %RH
CALIBRATION DATE : 09-Feb-24
RELATIVE HUMIDITY : 53 %RH ± 10 %RH

CONDITION OF THIS RESULTS OF CALIBRATION

1. THIS INSTRUMENT WAS CALIBRATED ACCORDING TO TLAS G-20 BY COMPARISON WITH CALIBRATED RTD Pt100 UNDER NO LOAD CONDITION. THE TEMPERATURE PROBES WERE PLACED ON NINE POINTS AND LOCATED ONE THERMOMETER PROBE IN EACH OF THE EIGHT CORNERS OF THE CHAMBER AND WAS AWAY FROM THE EACH WALL OF 5 cm TO 10 cm. AND PLACED THE NINTH THERMOMETER PROBE WITHIN 2.5 cm. OF THE GEOMETRIC CENTER OF THE CHAMBER. THE UNIFORMITY WAS MEASURED BETWEEN REFERENCE PROBE AND OTHER PROBES AT THE SAME TIME.

2. REFERENCE STANDARD INSTRUMENTS :

1) DATA LOGGER WITH RTD
HYDRA 2635A
23T6636
10-Jul-24

3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.


4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.

5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-
- NATIONAL INSTITUTE OF METROLOGY (THAILAND) THROUGH QUALITY CALIBRATION CO.,LTD.

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT

GENERAL INFORMATION

Overall Ambient Temperature around the Chamber (°C) variation : 1
Overall Line Voltage (V) variation : 3
Instrument Condition : Normal
Chamber Size (W*L*H): 56*40*72 cm



CHAMBER PERFORMANCE

Calibrate Point (°C)	Average All Position Temp. (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
35.0	35.03	0.05	0.09	0.16
36.0	36.05	0.07	0.08	0.19
41.5	41.45	0.08	0.13	0.20

TEMPERATURE MEASUREMENT ACCURACY TEST

Controller Indicating Temp (°C)	Measured Temperature (°C) at Spread Locations									Uncertainty (± °C)
	#1	#2	#3	#4	#5	#6	#7	#8	#9	
35.0	35.0	34.98	35.01	35.00	35.02	35.08	35.07	35.04	35.10	0.25
36.0	36.0	36.00	36.03	36.02	36.04	36.09	36.10	36.04	36.12	0.25
41.5	41.5	41.45	41.39	41.46	41.46	41.47	41.43	41.44	41.49	0.36

NOTE 1 : THE UNCERTAINTY OF MEASUREMENT EXCLUDED TEMPERATURE UNIFORMITY OF THE CHAMBER.

NOTE 2 : LOCATION 5 WAS REFERENCE LOCATION.

NOTE 3 : THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA.
THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR k=2, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%
END OF CALIBRATION REPORT

F-G010 REV : 03



CERTIFICATE No : 24T1189
REFERENCE No : 72116-7

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : AUTOCLAVE

MANUFACTURER : HIRAYAMA

MODEL : HVE-50

SERIAL No : 30612085166

ID No : EQL-155

CONDITION AS RECEIVED : USED ITEM

SUBMITTED BY : TEST TECH CO., LTD.
30,32 RAMA II SOI 63, RAMA II RD.,
SAMAEDAM, BANGKHUNTHIAN, BANGKOK
10150

CALIBRATED BY : CHAICHARN CH.

CALIBRATION DATE : 09-Feb-24

APPROVED BY : 
PONGSAK J.

ISSUED DATE : 12-Feb-24

RECEIVED DATE : 09-Feb-24

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.



CERTIFICATE No : 24T1189

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : AUTOCLAVE
MANUFACTURER : HIRAYAMA
ID NUMBER : EQL-155
RECEIVED DATE : 09-Feb-24
AMBIENT TEMPERATURE : 30° C ± 1° C
MODEL : HVE-50
SERIAL NUMBER : 30612085166
CALIBRATION DATE : 09-Feb-24
RELATIVE HUMIDITY : 53 %RH ± 10 % RH

CONDITION OF THIS RESULTS OF CALIBRATION

1. THIS INSTRUMENT WAS CALIBRATED BASED ON ITS 2646-1:2021 BY COMPARISON WITH CALIBRATED RTD DATA LOGGERS UNDER NO LOAD CONDITION. THE SENSORS WERE PLACED ON FIVE LOCATIONS AS SHOWN IN THE PICTURE. THE SENSOR ON LOCATION 1 AND 2 WERE PLACED IN THE UPPER HALF AND LOWER HALF OF CHAMBER FREE SPACE RESPECTIVELY. THE THIRD SENSOR WAS PLACED WITHIN 2.5 cm. OF THE GEOMETRIC CENTER OF THE INSTRUMENT CHAMBER. SENSOR NUMBER 4 WAS ATTACHED TO THE LOAD TEMPERATURE PROBE, IF FITTED, WITHIN 15 mm OF ITS TIP. SENSOR NUMBER 5 WAS PLACED IN THE CHAMBER DRAIN OR VENT WITHIN 100 mm OF ITS CONNECTION TO THE CHAMBER.
2. REFERENCE STANDARD INSTRUMENTS :-

- 1) DATA LOGGER
INSTRUMENT : VALPROBE
MODEL : S350.S357.DV33.DN94
SERIAL No : 24T0890
CERTIFICATE No : 26-Jan-25
DUE DATE : 26-Jan-25
3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.
4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.

5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-
- NATIONAL INSTITUTE OF METROLOGY (THAILAND) THROUGH QUALITY CALIBRATION CO., LTD.

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT



GENERAL INFORMATION

Overall Ambient Temperature around the Chamber variation : 0.5 °C
Autoclave Condition : Normal
Chamber Size (Diameter*H): 30 * 71 cm

CHAMBER PERFORMANCE

Calibrate Point (°C)	Average all Position (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)	Pressure (MPa)	Holding time (min)	Operating Cycle time (min)
115	115.74	0.09	0.11	0.25	0.090	20	60
121	121.59	0.06	0.21	0.28	0.125	20	60

FRONT

TEMPERATURE MEASUREMENT ACCURACY TEST(°C)

Cont Temp	Ind Temp	Measured Temperature (°C) at Spread Locations					Uncertainty (± °C)
		#1	#2	#3	#4	#5	
115	115.72	115.74	115.79	115.79	115.71	115.71	0.59
121	121.59	121.62	121.56	121.56	121.58	121.59	0.59

NOTE 1 : THE UNCERTAINTY OF MEASUREMENT OF TEMPERATURE MEASUREMENT ACCURACY TEST EXCLUDED TEMPERATURE UNIFORMITY OF THE CHAMBER.

NOTE 2 : THE STABILITY TERM IN THE UNCERTAINTY BUDGET WAS REPLACED BY THE STANDARD REPEATABILITY.

NOTE 3 : LOCATION 3 WAS REFERENCE LOCATION.

NOTE 4 : THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA.
THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR k=2, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.

END OF CALIBRATION REPORT


F-G010 REV : 03



Bara Scientific Co., Ltd.
968 U Chu Liang Building Floor7 Rama4 Road
Silom Bangkok Bangkok Thailand 10500
Tel : 02-6324300 Fax : 02-6375496-7
www.barascientific.com



Certificate of Calibration

Number of Page(s) 1 of 3

Certificate No. BSCC-UV-166/24
Equipment UV/Vis Spectrophotometer
Model UV-1900i
Manufacturer Shimadzu
Serial No. A12535780311 ML
ID No. EQL-233
Date of receipt 26 April 2024
Date of calibration 26 April 2024
Date of issue 30 April 2024
Customer name Test Tech Co., Ltd.
Address 30.32 Rama II Soi 63, Rama II Road, Samae Dam, Bang Khun Thian, Bangkok 10150

Temperature (24.9 - 25.4) °C (On site)
Humidity (49.4 - 51.1) %RH (On site)

Equipment condition Good Operation

Calibration Location Water Room

Calibration Procedure In-house method WJ-UV-702-01 based on ASTM E275-01

Traceability
Wavelength Accuracy is traceable to certificate No. 106372 and 106371
Photometric Accuracy is traceable to certificate No. 106364 and 111398
Stray Light is traceable to certificate No. 106377
The above certificate are traceable to SI unit through Stama Scientific Ltd.
(UKAS accredited calibration laboratory NO. 0659)

Calibrated by Mr. Wanchana Janloey

Approved by

Mr. Sonthi Temboonsakdi
Service Manager

The above results are valid exclusively for the calibrated item(s) as mention in this report / certificate.
Advertising the report / Certificate and publicity of the results are prohibited and also shall not be reproduced
except in full, without written approval of the Bara Scientific Co., Ltd.



Bara Scientific Co., Ltd.
968 U Chu Liang Building Floor7 Rama4 Road
Silom Bangkok Bangkok Thailand 10500
Tel : 02-6324300 Fax : 02-6375496-7
www.barascientific.com



Certificate of Calibration

Number of Page(s) 2 of 3

Certificate No. BSCC-UV-166/24

Calibration Results:

1. Wavelength Accuracy

Certified Wavelength (nm)	UUC (nm)	Error (nm)	Uncertainty ($\pm$ nm)
279.44	279.18	-0.26	0.18
418.53	418.46	-0.07	0.18
536.52	536.54	0.02	0.18
684.50	684.63	0.13	0.18
879.41	879.43	0.02	0.18

2. Photometric Accuracy (UV)

Wavelength (nm)	Certified Absorbance (A)	UUC (A)	Error (A)	Uncertainty ($\pm$ A)
235	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
257	0.0000	0.0000	0.0000	0.0075
	0.8354	0.8333	-0.0021	0.0075
313	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
350	0.0000	-0.0001	-0.0001	0.0075
	0.6199	0.6190	-0.0009	0.0075

*CNR = Customer not request

The above results are valid exclusively for the calibrated item(s) as mention in this report / certificate.
Advertising the report / Certificate and publicity of the results are prohibited and also shall not be reproduced
except in full, without written approval of the Bara Scientific Co., Ltd.



Bara Scientific
988 U Chu Liang Building Floor 7 Rama4 Road
Siam Bangkok Bangkok Thailand 10500
Tel : 02-6324300 Fax : 02-6375496-7
www.barascientific.com



Certificate of Calibration

Certificate No. BSCC-UV-166/24 Number of Page(s) 3 of 3

Calibration Results:

3. Photometric Accuracy (Visible)

Wavelength (nm)	Certified Absorbance (A)	UUC (A)	Error (A)	Uncertainty (±A)
420.0	0.0000	0.0000	0.0000	0.0042
	0.5761	0.5791	0.0030	0.0042
	0.7119	0.7132	0.0013	0.0042
	1.0189	1.0221	0.0032	0.0042
440.0	0.0000	0.0000	0.0000	0.0042
	0.5610	0.5636	0.0026	0.0042
	0.7001	0.7012	0.0011	0.0042
	1.0026	1.0052	0.0026	0.0042
465.0	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
546.1	0.0000	0.0000	0.0000	0.0042
	0.5249	0.5260	0.0011	0.0042
	0.6975	0.6971	-0.0004	0.0042
	1.0009	1.0012	0.0003	0.0042
590.0	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
635.0	0.0000	0.0000	0.0000	0.0042
	0.5666	0.5673	0.0007	0.0042
	0.7620	0.7611	-0.0009	0.0042
	1.0982	1.0976	-0.0006	0.0042

*CNR = Customer not request

4. Stray Light*

Standard cut-off wavelength (nm)	Unit Under Calibration(UUC)	
	Wavelength (nm)	Absorbance (A)
200.85±0.1nm	200.76	0.9795
		2.0091

The Stray light transmission reference is less than 1.0%T and Stray light absorbance reference is greater than 2.00A

*Stray Light not NSC-ONSC Accredited.

The measurement uncertainty is base on a standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

End of Certificate

The above results are valid exclusively for the calibrated item(s) as mention in this report / certificate. Advertising the report / Certificate and publicity of the results are prohibited and also shall not be reproduced except in full, without written approval of the Bara Scientific Co., Ltd.



Certificate of Calibration

Equipment: SPECTROPHOTOMETER
Model: DR6000
Serial No. (or ID.): 1693421 (EQL-197)
Manufacturer: HACH
Condition: In Condition

Certificate No.: C06240153
Issued Date: 18 April 2024
Job No.: WO-00024683
Page: 1 of 3

Customer: TEST TECH CO., LTD.
30,32 Rama II Soi 63, Rama II Rd.,
Samaedam, Bangkhuntien Bangkok 10150 Thailand

Environment Condition: Temperature 29.8 °C ± 0.1 °C
Humidity 45.7 %RH ± 6.9 %RH

Calibration Place: TEST TECH CO., LTD. (แบบนัด)
30,32 Rama II Soi 63, Rama II Rd.,
Samaedam, Bangkhuntien Bangkok 10150 Thailand

Calibration By: Miss.Kaewkan Suradech
Calibration Date: 18 April 2024

The Method used: In house method, CAL-WI-24, base on ASTM E 275-08 and ASTM E 387-04
Traceability: This certificate is traceable to the CRM maintained by National Institute of Standards and Technology (NIST) through Starna Scientific Limited.

The standard for Wavelength Certificate No. 118106 and 118118
The standard for Photometric Certificate No. 118123 and 118113
The standard for Stray light Certificate No. 118110 and 118112
The standard for Spectral resolution Certificate No. 118104

Miss Kaewkan Suradech
(Miss Kaewkan Suradech)

Mr. Nitnun Srihawan
(Mr. Nitnun Srihawan)
Authorized signatory

Person in charge

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognised national standard laboratories.

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).

These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of DKSH Technology Limited.

NSC-IBITS 17925 Calibration
DKSH Technology Limited
2533 Sukhumvit Road, Bangkok, Thailand 10260
Phone: +66 2638 7000 Email: info@dksh.com Website: www.dksh.com/certificate-issuing

Delivering Growth - in Asia and Beyond.



Calibration Results:
Without Adjustment

Wavelength Accuracy (nm), The spectral bandwidth of Std at 2 nm and UUC at 2 nm				
Standard Wavelength	Unit Under Calibration	Correction	Uncertainty	
418.61	418.5	0.11	0.13	
536.66	536.7	-0.04	0.13	
637.98	637.9	0.08	0.13	
748.48	748.6	-0.12	0.13	
807.03	807.4	-0.37	0.13	
Photometric Accuracy (Absorbance)				
Wavelength	Standard absorbance	Unit Under Calibration	Correction	Uncertainty
420 nm	0.0000	0.000	0.0000	0.0045
	0.5772	0.576	0.0012	0.0045
	0.7198	0.719	0.0008	0.0045
	1.0394	1.039	0.0004	0.0045
440 nm	0.0000	0.000	0.0000	0.0045
	0.5608	0.560	0.0008	0.0045
	0.7062	0.705	0.0012	0.0045
	1.0189	1.018	0.0009	0.0045
465 nm	0.0000	0.000	0.0000	0.0045
	0.5214	0.521	0.0004	0.0045
	0.6662	0.664	0.0012	0.0045
	0.9577	0.957	0.0007	0.0045
546.1 nm	0.0000	0.000	0.0000	0.0045
	0.5192	0.518	0.0012	0.0045
	0.6907	0.689	0.0017	0.0045
	0.9949	0.993	0.0019	0.0045
590 nm	0.0000	0.000	0.0000	0.0045
	0.5530	0.551	0.0020	0.0045
	0.7555	0.753	0.0025	0.0045
	1.0761	1.073	0.0031	0.0045
635 nm	0.0000	0.000	0.0000	0.0045
	0.5604	0.559	0.0014	0.0045
	0.7418	0.739	0.0028	0.0045
	1.0467	1.044	0.0027	0.0045

DKSH Technology Limited
2533 Sukhumvit Road, Bangkok, Thailand 10260
Phone: +66 2639 7000 Email: info.calibration@dksh.com Website: www.dksh.com/scientific-thailand

Delivering Growth - in Asia and Beyond.



Calibration Results:
Without Adjustment

Photometric Accuracy (Absorbance)				
Wavelength	Standard absorbance	Unit Under Calibration	Correction	Uncertainty
235 nm	0.0000	0.000	0.0000	0.0080
	0.7533	0.748	0.0053	0.0080
257 nm	0.0000	0.000	0.0000	0.0080
	0.8745	0.869	0.0055	0.0080
313 nm	0.0000	0.000	0.0000	0.0080
	0.2926	0.293	-0.0004	0.0080
350 nm	0.0000	0.000	0.0000	0.0080
	0.6486	0.644	0.0046	0.0080
Stray light *				
Standard: cut-off		UUC: Wavelength (nm)	UUC: Transmission (%T)	Absorbance (A)
260.95 +/- 0.11 nm		261.0	0.9	2.046
392.04 +/- 0.11 nm		392.0	1.3	1.886
Spectral Resolution *				
Nominal Concentration 0.02 % v/v	Peak	Trough	Ratio	SBW
Standard Wavelength (nm)	268.74	266.81	1.29	2.00
UUC: Wavelength (nm)	268.6	266.6		
Std Absorbance (A)	0.5137	0.3473		
UUC: Absorbance (A)	0.463	0.359		

* Calibration Marked "Not TISI Accredited" in this Certificate have been included for completeness.

The End of Certificate

DKSH Technology Limited
2533 Sukhumvit Road, Bangkok, Thailand 10260
Phone: +66 2639 7000 Email: info.calibration@dksh.com Website: www.dksh.com/scientific-thailand

Delivering Growth - in Asia and Beyond.



บริษัท ไทยยูนิค จำกัด THAI UNIQUE CO., LTD.

80-82 ถนนประชาวิทย์ แขวงบางขุนพรหม เขตพระนคร กรุงเทพฯ 10200
80-82 Prachathipatai Rd., Bangkokhuphrom, Pranakorn, Bangkok 10200
Tel. 0-2629-0191-6, 0-2280-1787, Fax. 0-2280-1788, E-mail : thau@thaiunique.com, Website : www.thaiunique.com

ATOMIC ABSORPTION SPECTROMETER TEST CERTIFICATE

Certificate No : SV231021414
Instrument Type : Atomic Absorption Spectrometer
Model : AA240FS
Serial Number : EL08043418
Organization : Test Tech Co., Ltd.
Address : 30,32 Soi 66 Rama II Rd., Samaedam Bangkokhuphrom, Bangkok 10150
Date : 25 Oct 2023

Hollow cathode lamps used

Element	Lamp number	Comments
Arsenic	56-101003-00	
Copper	56-101014-00	
Potassium	56-101042-00	
Iron	56-101027-00	
Manganese	56-101337-00	

Test description	Specification	Result	Comments
Light throughput (%Gain) or (EHT)			
Cu at 324.8 nm	≤ 64 % or 380 V	32 %	Pass
As at 193.7 nm	≤ 80 % or 540 V	55 %	Pass
K at 766.5 nm*	≤ 84 % or 540 V	64 %	Pass
Fe at 248.3 nm	≤ 80 % or 540 V	59 %	Pass
Mn at 279.5 nm	≤ 64 % or 380 V	46 %	Pass
Photometric noise Cu BGC off			
STDV @ 0 Abs	≤ 0.0001	0.0001	Pass



บริษัท ไทยยูนิค จำกัด THAI UNIQUE CO., LTD.

80-82 ถนนประชาวิทย์ แขวงบางขุนพรหม เขตพระนคร กรุงเทพฯ 10200
80-82 Prachathipatai Rd., Bangkokhuphrom, Pranakorn, Bangkok 10200
Tel. 0-2629-0191-6, 0-2280-1787, Fax. 0-2280-1788, E-mail : thau@thaiunique.com, Website : www.thaiunique.com

Wavelength accuracy			
Cu at 324.8 nm	323.0 nm - 326.0 nm	324.8 nm	Pass
As 193.7 nm	192.0 nm - 195.0 nm	193.7 nm	Pass
K at 766.5 nm*	765.0 nm - 768.0 nm	766.6 nm	Pass
Fe at 248.3 nm	246.8 nm - 249.8 nm	248.3 nm	Pass
Mn at 279.5 nm	278.0 nm - 281.0 nm	279.5 nm	Pass
High solids nebulizer setting**			
Uptake rate	7.2 - 10.6 ml / min	9.8 ml/min	Pass
Max Abs	≥ 0.75 Abs	0.81 Abs	Pass
Precision(%RSD)	≤ 0.5 %	0.2 %	Pass
Zeeman Background Correction Accuracy (%)*			
BCA @ Au 242.8 nm	< 3.7 %	***	***
Zeeman Magnetic Sensitivity Ratio (%)***			
MSR @ Cu 324.7 nm	> 70 %	***	***
Characteristic mass and sensitivity ****			
Sensitivity	≥ 0.21 Abs	****	****
Precision (%RSD)	≤ 4.0 %	****	****

* for Wideband PMT (Wavelength 190nm - 900nm)

** for Flame system

*** for Zeeman system

**** for Graphite furnace system

CALIBRATED BY :

Signature:

Engineer : Suriya Nacharoen

Date : 25/Oct/2023

APPROVED BY :

Signature:

Engineer : Suchai Sangnankiatichai

Date : 25 / Oct / 2023



MAINTENANCE AND TEST CERTIFICATE MODEL OPTIMA 8000

Customer : บริษัท เทสท์ เทคโนโลยี จำกัด	Date Tested: May 18, 2023
Address : 30, 32 ซอยพระรามที่ 2 ซอย 63 ถนนพระรามที่ 2 แขวงสามเมีนบุรี เขตบางขุนเทียน กรุงเทพฯ 10150	Recommendation Recertification Period 12 Recertification Due: May 17, 2024 Date Last Certified: May 19, 2022 Visit Number: 1 of 1
User Name: คุณอัยยิฉ สรรค์จันทร์	PerkinElmer Phone: 02-719-6420 ext 206
Phone: 02-893-4211-17	PerkinElmer Fax: 02-318-5597
Fax: lab_center@testtech.co.th	

CONFIGURATION TESTED	ACCESSORIES/COMPONENT NOT INCLUDED
MODEL OPTIMA 8000	WinLab32 Version 5.5.0.0714
N0772045	PN:6150T21E4Q1E
EQL-180	
TESTED EQUIPMENT	EXPIRATION
IPV Methods	
TEST STANDARD USED	EXPIRATION DATE
Mixed standard 1/10	NOV 30, 2023
Mixed standard 1/100	NOV 30, 2023
CUSTOMER SUPPLIED	CUSTOMER INITIALS
2 % HNO3	
10 % HNO3	



MAINTENANCE AND TEST CERTIFICATE MODEL OPTIMA 8000

SERIAL NUMBER : 078S1411171C	DATE TESTED : May 18, 2023
1. MECHANICAL CHECKS	☐ OK
A. Inspect and clean all fans and filters.	☐ OK
B. Inspect and replace as necessary, all torch components including the RF coil.	☐ OK
C. Inspect all tubing for sign of clacking or leaking.	☐ OK
D. Adjust water and gas pressure regulator settings.	☐ OK
E. Inspect and leak check pneumatics drawers.	☐ OK
F. Clean the exterior of the instrument.	☐ OK
2. OPTICAL CHECKS	☐ OK
A. Inspect and clean all optical components.	☐ OK
B. As required, check and replace all purgefilters.	☐ OK
C. Recheck optical alignment.	☐ OK
3. COOLING SYSTEM CHECKS	☐ OK
A. Perform preventive maintenance on chiller.	☐ OK
B. Flush out the chiller every six months.	☐ OK
4. PERFORMANCE CHECKS	☐ OK
A. Torch View Alignment.	☐ OK
B. Wavelength Calibration.	☐ OK



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, BANGKOK 10250
TEL. 0-2717-3000-24 FAX. 0-2719-9484



Certificate of Calibration

Certificate No. : 23H2216
Page : 1 of 2

Equipment : Dial Thermo-Hygrometer

Manufacturer: Barigo

Model : -

Serial No.: -

ID No.: EQL-064

Condition As-Received: Used Item

Received Date: 12 October 2023

Calibration Date: 17 October 2023

Reference: 2310-047DN

Ambient Temperature: (25 ± 3) °C

Relative Humidity: (50 ± 20) %

This certificate may not be reproduced other than in full,
except with the prior written approval of the head of
Corporate Services 3: Equipment Calibration and Testing Services.

Submitted by: TEST TECH CO.,LTD. (HEAD Office)

30, 32 Rama II Soi 63, Rama II Rd.,
Samaedam, Bangkokthian, Bangkok 10150

Procedure used: Calibration were conducted using in-house calibration procedure CP-H02 according to comparison with standard chilled mirror sensor for humidity measurement function and comparison with standard temperature probe for temperature measurement function into humidity / temperature chamber.

Condition of this result of calibration

1. Reference standards instruments :

Instrument	Model	Serial No.	Certificate No.	Due Date
1) Handheld Thermometer With Sensor	1523	3240076	231305	15 Mar 2024
2) Dew Point Hygrometer	Optidew 401	164756	TH-0158-22	13 Dec 2023

2. The certificate is valid only to the item calibrated on date and place of calibration.

3. This Certification is traceable to the International System of Unit maintained through:-

-Technology Promotion Association (Thailand-Japan), NSC-ONSC Accredited No. Calibration 0008

-National Institute of Metrology Thailand (NIMT)

Calibrated by : Surasit Phansudnoi
Issue Date : 26 October 2023

Approved Signatory :

[] Chakrit Waewwanjua
[] Ponthippa Taneyakul
[x] Viporn Tantayawutti

Viporn

B 0327545



Cert. No.: 23H2216
Page.: 2 of 2

Result of Calibration:-
Function: Humidity Measurement Without Adjustment

Reference Temperature (°C)	Standard Humidity (%R.H.)	UUC* Reading (%R.H.)	Error (%R.H.)	Uncertainty of Measurement (±%R.H.)
25.0	30.1	30.0	-0.1	1.5
25.0	40.1	39.0	-1.1	1.5
25.0	50.1	49.0	-1.1	1.7
25.0	60.0	59.0	-1.0	1.7
25.0	75.2	75.5	0.3	1.8

Result of Calibration:-
Function: Temperature Measurement Without Adjustment

Standard Temperature (°C)	UUC* Reading (°C)	Error (°C)	Uncertainty of Measurement (±°C)
15.046	15.0	-0.046	0.72
19.975	20.0	0.025	0.72
25.022	25.0	-0.022	0.72
30.000	30.0	0.000	0.72

UUC* : Unit Under Calibration

The reported uncertainty of measurement was based on standard uncertainty multiplied by coverage factor k = 2.00, providing confidence level approximately 95%.

-000-

a 1185882



CERTIFICATE No : 23M6754
REFERENCE No : 69854-1

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : DIGITAL BALANCE
MANUFACTURER : SARTORIUS
MODEL : BP210S
SERIAL No : S0736477
ID No : EQL-008
CONDITION AS RECEIVED : USED ITEM

SUBMITTED BY : TEST TECH CO., LTD.
30,32 RAMA II SOI 63, RAMA II RD.,
SAMAEDAM, BANGKHUNTHIAN, BANGKOK
10150

CALIBRATED BY : PRASERT D.
CALIBRATION DATE : 13-Jul-23

APPROVED BY :
ISSUED DATE : 17-Jul-23
RECEIVED DATE : 13-Jul-23

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.

F-G010 REV 03



CERTIFICATE No : 23M6754

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : DIGITAL BALANCE
MANUFACTURER : SARTORIUS
ID No : EQL-008
AIR PRESSURE : 1011 mbar
AMBIENT TEMPERATURE : 23° C ± 1° C
MODEL : BP210S
SN : S0736477
RECEIVED DATE : 13-Jul-23
CALIBRATION DATE : 13-Jul-23
RELATIVE HUMIDITY : 50 %RH ± 10 % RH

CONDITION OF THIS RESULTS OF CALIBRATION

1. THIS INSTRUMENT WAS CALIBRATED BY ACCORDING TO UKAS LAB 14 EDITION 6:2019 BY USING KNOWN WEIGHT
STANDARD WEIGHT. THE BALANCE WAS ADJUSTED USING INTERNAL WEIGHT TO ADJUST. THE BALANCE HAS NO
ZERO TRACKING FUNCTION. REPEATABILITY WAS MEASURED BY USING 10 REPEATED MEASUREMENTS. LINEARITY
WAS MEASURED COVERING 10 POINTS, EVENLY SPREAD OVER THE RANGE. THE INSTRUMENT WAS SET ZERO BEFORE
PERFORMING THE LINEARITY TEST. OFF-CENTER LOADING WAS MEASURED BY USING STANDARD WEIGHTS PLACED
ON THE PAN AND MOVED TO VARIOUS POSITIONS ON THE PAN. THE INTERNAL WEIGHT WAS CHECKED BY USING
2. REFERENCE STANDARD INSTRUMENTS :-

INSTRUMENT MODEL SERIAL No CERTIFICATE No DUE DATE
1) STANDARD WEIGHT SET E2 QK-1-151 M230201HS 02-Feb-25
2) STANDARD WEIGHT E2 15843 M230201HS 02-Feb-25

3. THIS CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.

4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.

5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-

- NATIONAL INSTITUTE OF METROLOGY (THAILAND) THROUGH CENTRAL BUREAU OF WEIGHTS&MEASURES

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT

1. ZERO SETTING FUNCTION : NORMAL


2. TARE FUNCTION : NORMAL

3. REPEATABILITY OF READING AT 200 g WAS 0 g

4. DEPARTURE FROM NOMINAL VALUE/ LINEARITY

NOMINAL VALUE (g)	BALANCE READING (g)	CORRECTION (g)	UNCERTAINTY (±g)
0.0	0.0000	0.0000	0.000082
0.1	0.1000	0.0000	0.000083
0.2	0.2000	0.0000	0.000083
0.5	0.5000	0.0000	0.000083
1.0	1.0000	0.0000	0.000084
2.0	2.0000	0.0000	0.000084
5.0	5.0000	0.0000	0.000086
10.0	10.0000	0.0000	0.000089
20.0	20.0001	-0.0001	0.000094
50.0	49.9999	0.0001	0.00012
100.0	99.9999	0.0001	0.00019
200.0	199.9997	0.0003	0.00032

5. OFF CENTER LOADING ERROR



POINT	READING (g)
1	99.9998
2	99.9997
3	99.9998
4	99.9998
5	99.9998
OFF-CENTER LOADING	0.0001

6. INTERNAL WEIGHT ERROR : 0.000499999999988177 g

NOTE: THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA
THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A
COVERAGE FACTOR k=2, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.

END OF CALIBRATION REPORT

F-G010 REV 03



QUALITY CALIBRATION CO.,LTD.
235 Petchkasem 63/2 Road, Laksong, Bangkok, Bangkok 10160
Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4384
www.qcalibration.com



CERTIFICATE No : 23T8798
REFERENCE No : 70515-6

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : HOT AIR OVEN
MANUFACTURER : MEMMERT
MODEL : UFE 500
SERIAL No : G508.0791
ID No : EQL-128
CONDITION AS RECEIVED : USED ITEM
SUBMITTED BY : TEST TECH CO., LTD.
30,32 RAMA II SOI 63, RAMA II RD., SAMAEDAM,
BANGKHUNTHIAN, BANGKOK 10150

CALIBRATED BY : CHAICHARN CH.
CALIBRATION DATE : 11-Sep-23

APPROVED BY :
PONGK K. J.
ISSUED DATE : 15-Sep-23
RECEIVED DATE : 11-Sep-23

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.

F-G010 REV : 03



QUALITY CALIBRATION CO.,LTD.
235 Petchkasem 63/2 Road, Laksong, Bangkok, Bangkok 10160
Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4384

CERTIFICATE No : 23T8798

PAGE : 2 OF 2

Calibration Report


EQUIPMENT : HOT AIR OVEN
MANUFACTURER : MEMMERT
MODEL : UFE 500
ID No : EQL-128
RECEIVED DATE : 11-Sep-23
AMBIENT TEMPERATURE : 24 °C ± 1 °C
S/N : G508.0791
CALIBRATION DATE : 11-Sep-23
RELATIVE HUMIDITY : 51 %RH ± 10 %RH

CONDITION OF THIS RESULTS OF CALIBRATION

- THIS INSTRUMENT WAS CALIBRATED ACCORDING TO TLAS G-20 BY COMPARISON WITH CALIBRATED RTD Pt100 UNDER NO LOAD CONDITION. THE TEMPERATURE PROBES WERE PLACED ON NINE POINTS AND LOCATED ONE THERMOMETER PROBE IN EACH OF THE EIGHT CORNERS OF THE CHAMBER AND WAS AWAY FROM THE EACH WALL OF 5 cm TO 10 cm. AND PLACED THE NINTH THERMOMETER PROBE WITHIN 2.5 cm. OF THE GEOMETRIC CENTER OF THE CHAMBER. THE UNIFORMITY WAS MEASURED BETWEEN REFERENCE PROBE AND OTHER PROBES AT THE SAME TIME.
- REFERENCE STANDARD INSTRUMENTS :-
 - DATA LOGGER WITH RTD HYDRA 2635A 7301307 23T6636 10-Jul-24
 - THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.
 - THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.
 - THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-
- NATIONAL INSTITUTE OF METROLOGY (THAILAND) THROUGH QUALITY CALIBRATION CO.,LTD.
- RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT

GENERAL INFORMATION

Overall Ambient Temperature around the Chamber (°C) variation : 1
Overall Line Voltage (V) variation : 10
Instrument Condition : Normal
Chamber Size (W*L*H): 50*40*48 cm



CHAMBER PERFORMANCE

Calibrate	Average All	Temperature	Temperature	Overall
Point	Position Temp.	Stability	Uniformity	Variation
(°C)	(°C)	(±°C)	(°C)	(°C)
104.0	104.49	0.28	0.66	0.93
180.0	180.25	0.32	0.62	1.11

TEMPERATURE MEASUREMENT ACCURACY TEST

Controller Temp (°C)	Indicating Temp (°C)	Measured Temperature (°C) at Spread Locations									Uncertainty (± °C)
		#1	#2	#3	#4	Ref. 5	#6	#7	#8	#9	
104.0	104.0	104.46	104.13	104.45	104.28	104.57	104.67	104.60	104.58	104.67	0.38
180.0	180.0	180.27	179.85	180.41	179.93	180.19	180.54	180.41	180.51	180.13	1.1

NOTE 1 : THE UNCERTAINTY OF MEASUREMENT EXCLUDED TEMPERATURE UNIFORMITY OF THE CHAMBER.

NOTE 2 : LOCATION 5 WAS REFERENCE LOCATION.

NOTE 3 : THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA.
THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR k =2, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.
END OF CALIBRATION REPORT

F-G010 REV : 03



QUALITY CALIBRATION CO., LTD.
235 Petchkasem 63/2 Road, Laksong, Bangkok 10160
Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4384
www.qcalibration.com



CERTIFICATE No : 23T8799
REFERENCE No : 70515-7

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : HOT AIR OVEN
MANUFACTURER : MEMMERT
MODEL : UFE 500
SERIAL No : G512.2005
ID No : EQL-161
CONDITION AS RECEIVED : USED ITEM
SUBMITTED BY : TEST TECH CO., LTD.
30,32 RAMA II SOI 63, RAMA II RD., SAMAEADAM,
BANGKHUNTHIAN, BANGKOK 10150

CALIBRATED BY : CHAICHARN CH.
CALIBRATION DATE : 11-Sep-23

APPROVED BY :
ISSUED DATE : 15-Sep-23
RECEIVED DATE : 11-Sep-23

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.

F-G010 REV : 03



QUALITY CALIBRATION CO., LTD.
235 Petchkasem 63/2 Road, Laksong, Bangkok 10160
Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4384

CERTIFICATE No : 23T8799

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : HOT AIR OVEN
MANUFACTURER : MEMMERT
MODEL : UFE 500
ID No : EQL-161
RECEIVED DATE : 11-Sep-23
AMBIENT TEMPERATURE : 24 °C ± 1 °C
S/N : G512.2005
CALIBRATION DATE : 11-Sep-23
RELATIVE HUMIDITY : 51 %RH ± 10 %RH


CONDITION OF THIS RESULTS OF CALIBRATION

- THIS INSTRUMENT WAS CALIBRATED ACCORDING TO TLAS G-20 BY COMPARISON WITH CALIBRATED RTD Pt100 UNDER NO LOAD CONDITION. THE TEMPERATURE PROBES WERE PLACED ON NINE POINTS AND LOCATED ONE THERMOMETER PROBE IN EACH OF THE EIGHT CORNERS OF THE CHAMBER AND WAS AWAY FROM THE EACH WALL OF 5 cm TO 10 cm. AND PLACED THE NINTH THERMOMETER PROBE WITHIN 2.5 cm. OF THE GEOMETRIC CENTER OF THE CHAMBER. THE UNIFORMITY WAS MEASURED BETWEEN REFERENCE PROBE AND OTHER PROBES AT THE SAME TIME.
- REFERENCE STANDARD INSTRUMENTS :-
 - 1) DATA LOGGER WITH RTD HYDRA 2635A 7301307 23T6636 10-Jul-24
 3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.
 4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.
 5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-
- NATIONAL INSTITUTE OF METROLOGY (THAILAND) THROUGH QUALITY CALIBRATION CO., LTD.

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT

GENERAL INFORMATION

Overall Ambient Temperature around the Chamber (°C) variation : 1
Overall Line Voltage (V) variation : 10
Instrument Condition : Normal
Chamber Size (W*L*H): 56*40*48 cm



CHAMBER PERFORMANCE

Calibrate	Average All, Position Temp. (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
104.0	103.96	0.14	0.58	0.73
180.0	179.55	0.22	0.93	1.47

TEMPERATURE MEASUREMENT ACCURACY TEST

Controller Temp (°C)	Indicating Temp (°C)	Measured Temperature (°C) at Spread Locations									Uncertainty (± °C)
		#1	#2	#3	#4	Ref. 5	#6	#7	#8	#9	
104.0	104.0	104.16	104.13	104.20	103.98	103.76	103.76	104.06	103.71	103.93	0.38
180.0	180.0	179.73	179.89	180.04	179.54	179.30	178.98	179.75	178.97	179.77	1.1

NOTE 1 : THE UNCERTAINTY OF MEASUREMENT EXCLUDED TEMPERATURE UNIFORMITY OF THE CHAMBER.

NOTE 2 : LOCATION 5 WAS REFERENCE LOCATION.

NOTE 3 : THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA.

THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR k=2, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95% END OF CALIBRATION REPORT

F-G010 REV : 03



Certificate of Calibration

Equipment: TURBIDIMETER
Model: 2100N
Serial No. (or ID.): 97040003415 (EQL-024)
Manufacturer: HACH
Condition: In Condition
Certificate No.: C08230153
Issued Date: 15 September 2023
Job No.: WO-00005226
Page: 1 of 2
Customer: TEST TECH CO., LTD.
30,32 Rama II Soi 63, Rama II Rd.,
Samaedam, Bangkhuntien Bangkok 10150 Thailand

Environment Condition: Temperature 23 °C ± 2 °C
Humidity 50 %RH ± 15 %RH
Calibration Place: Environment Laboratory, DKSH Technology Limited.
2533 Sukhumvit Road, Bangkok,
Phrakhanong, Bangkok 10260 Thailand
Calibration By: Miss.Orawan Khlaiphloi
Calibration Date: 14 September 2023
The Method used: In house method, CAL-WI-23, base on Hach Manufacturer Method 8195
Traceability: This certificate is traceable to Primary standard Fomazin and StablCal accepted by United States Environmental Protection Agency (EPA) through Hach Company
Certificate No. A1075 , A1074 , A1091 , A1074 , A1074


(Miss Orawan Khlaiphloi)
Person in charge

(Mr. Nitiun Srihawan)
Authorized signatory

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.
The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).
These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of DKSH Technology Limited.

DKSH Technology Limited
2533 Sukhumvit Road, Bangkok, Phrakhanong, Bangkok 10260
Phone: +66 2639 7000 Email: info.calibration@dksh.com Website: www.dksh.com/calibration-thailand

Delivering Growth - In Asia and Beyond.

CAL-FM-C08-08: 20 Jul 2022

Calibration Results:

Before Adjustment

Std Turbidity (NTU)	UUC Reading	Correction	Deviation	Uncertainty
0.050	0.088	-0.038	0.0	0.070
20.40	19.1	1.30	0.0	1.0
205.0	195	10.0	0.5	10
1028.0	952	76.0	0.9	50
4068.0	3942	126.0	0.9	200

After Adjustment

Std Turbidity (NTU)	UUC Reading	Correction	Deviation	Uncertainty
0.050	0.084	-0.034	0.0	0.070
20.40	20.4	0.00	0.0	1.0
205.0	205	0.0	0.5	10
1028.0	1026	2.0	0.5	50
4068.0	4063	5.0	0.5	200

The End of Certificate

DKSH Technology Limited
2533 Sukhumvit Road, Bangkok, Phrakhanong, Bangkok 10260
Phone: +66 2639 7000 Email: info.calibration@dksh.com Website: www.dksh.com/calibration-thailand

Delivering Growth - In Asia and Beyond.

CAL-FM-C08-08: 20 Jul 2022



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
53/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-29 FAX. 0-2719-9484



Cert.No.: 24CH59
Page.: 1 of 3

Certificate of Calibration

Equipment : Conductivity Meter
Manufacturer : TOA
Model : CM-41X
Serial No. : 842572
ID No. : EQL-211
Condition As-Received: Used Item
Received Date : 11 January 2024
Calibration Date : 15 January 2024
Reference : 2401-0300DN-1
Submitted by : TEST TECH CO.,LTD. (HEAD Office)
30, 32 Rama II Soi 63, Rama II Rd.,
Samaedam, Bangkhuntian, Bangkok 10150

Ambient Temperature :
Relative Humidity :
Calibration Procedure:

(25 ± 2.5) °C
(50 ± 15) %
In-house method :
- CP-CH6 by direct measurement
with certified reference material (CRM)
- CP-CH8 by comparison with standard thermometer

Calibrated by :

Warakorn Lengagrakul

Approved by :

Saithip
Approved Signatory

(☒) Saithip Meangmai
() Warakorn Lengagrakul
() Ponpan Paipim

Issue Date :

17 January 2024

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

๒๒

ใบรับรองการสอบเทียบ “เครื่องวัดความนำไฟฟ้า”
(Calibration Certificate of Conductivity Meter)

A 0062587



Cert.No.: 24CH59

Page.: 2 of 3

Condition of this result of calibration

1. Reference Standard Instrument :-

Instrument	Serial No.	ID No.	Certificate No.	Due date
1) Thermometer	1963878	130RC095	2311051	05 Sep 2024
2) Ref. Std. Thermometer	4982054	110RC044	231908	26 Jul 2024

- This Certification is traceable to SI Through Technology Promotion Association (Thailand - Japan)

2. Certified Reference Materials :-

- Conductivity calibration solution, CPA chem Ltd., The measurement results are traceable to SI through CPA chem Ltd., ANSI-ASQ National Accreditation Board, Accredited No. AR-1835

Conductivity Solution

Conductivity Solution	Manufacturer	Lot No.	Exp. date
147.0 $\mu\text{S/cm}$	CPA Chem	913595	14 July 2024
1.413 mS/cm	CPA Chem	931955	30 Sep 2024
12.880 mS/cm	CPA Chem	913597	14 July 2024

- Control Conductivity calibration solution temperature by Water bath (25.0.1) $^{\circ}\text{C}$
3. This certificate is valid only to the item calibrated on date and place of calibration.

Calibration results

Function : Conductivity Measurement

(*) After Adjustment at 147.0, 1413.0, 12880 $\mu\text{S/cm}$

Conductivity Electrode Serial No.: 806F0005

Standard Conductivity Solution	After Adjustment UUC* Reading	Uncertainty of Measurement ($\pm$)	Coverage factor k
147.0 $\mu\text{S/cm}$	147.1 $\mu\text{S/cm}$	0.99 $\mu\text{S/cm}$	2.00
1.413 mS/cm	1.413 mS/cm	0.0092 mS/cm	2.00
12.880 mS/cm	12.88 mS/cm	0.086 mS/cm	2.00

Remark - UUC* = Unit Under Calibration

- Adjustment Cell constant = 147.0 $\mu\text{S/cm}$ 96.8 m^{-1} , 1.413 mS/cm = 98.0 m^{-1} , 12.880 mS/cm = 99.4 m^{-1}

Saitip

a 1197672



Cert.No.: 24CH59

Page.: 3 of 3

Calibration Results

Function : Temperature Measurement

(*) Without adjustment

This equipment was connected with Temperature Probe;

- Model : CT-58101B

- Serial No. : 806F0005

Dimension of probe;

- Length : 114 mm

- Diameter : 12 mm

- Immersion Depth : 100 mm

Calibration Point ($^{\circ}\text{C}$)	Standard Temperature ($^{\circ}\text{C}$)	UUC* Reading ($^{\circ}\text{C}$)	Error ($^{\circ}\text{C}$)	Uncertainty of Measurement ($\pm$ $^{\circ}\text{C}$)	Coverage factor k
25.0	25.003	25.0	-0.003	0.13	2.00

Remark : - UUC* = Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-000-

Saitip

a 1197671



Metrology

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhohi, Saraburi 18110, Thailand.

Saraburi Tel : +66 3627 3096 Fax : +66 3627 3100

Bangkok Tel : +668 9205 6851 , +669 8247 2360

Website : www.scieco.co.th E-Mail : calibrate@scg.com



Certificate No. T240070

Page 1 of 4

Certificate of Calibration

Equipment : Chamber (Cooling Room)

Manufacturer : -

Model : -

Serial No. : -

Customer Code : EQL-167

ID No. : T1447A1

Customer : Test Tech Co.,Ltd

30, 32 Rama II Soi 63, Rama II Rd., Samaedam,

Bangkhunthian Bangkok 10150

Customer Location : LABORATORY FLOOR 3

Date of Receipt : 12 January 2024

Calibrated By : Sujjar Naknakred (Site Calibration Manager)

Approved By :  / Boonchai Suriyawong (Site Calibration Manager)

Date of Issue : 24 JAN 2024

The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standard laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Metrology

๗9

ใบรับรองการสอบเทียบ "ห้องเย็น"
(Calibration Certificate of Cool Room)



Metrology

SCI ECO Services Company Limited

33/2 Moo 3, T. Banpa, A. Kaengkhroi, Saraburi 18110, Thailand.



Certificate No. T240070

Page 2 of 4

Calibration Report

Equipment : Chamber (Cooling Room)

Date of Calibration : 16 January 2024

Environment : Temperature : 19.4-24.1 °C

Line Voltage : 221.3-226.1 V

Relative Humidity : 55 - 65 %RH

Condition of this results of calibration :

1. This equipment was calibrated by insert nine standard thermocouples type T into its chamber , the other one standard thermocouples type T use for ambient temperature measurement . The calibration was done in according to WI-T20 (based on ASTM E145-94 (Reapproved 2001) and AS2853-1986).

All data show below were final values and the initial data from customer request . The temperature scale used was based on ITS - 90 .

2. Reference Standard Instrument :

Instrument	Model	Instrument No.	Certificate No.	Due Date
TC	TYPE T	TN161-TN170	T230773	10 April 2024
TC	TYPE T	TN161-TN170	T230773	10 April 2024
DATA LOGGER	34970A	T149	T230773	10 April 2024

3. This certificate is traceable to :

National Institute of Metrology (Thailand) through Metrological Center (NSC-TISI-TIS 17025 CALIBRATION 0244.)

4. Condition of calibrated item : good

Equipment Description :

Time Constant : Hour 37 Minute At 3 °C
Fresh Air Damper ☐ Open ☐ Min ☐ Medium ☐ Max
☐ Close
☒ Not Available

5. Adjustment :

() without adjustment (X) after adjustment

Approved By:

Approved By:



Metrology

SCI ECO Services Company Limited


33/2 Moo 3, T. Banpa, A. Kaengkhroi, Saraburi 18110, Thailand.



Certificate No. T240070

Page 3 of 4

Calibration Report



C = Centre, F = Centre of Face, A = Corner, E = Centre of Edge

1C	=	TN161
2A	=	TN162
3A	=	TN163
4F	=	TN164
5A	=	TN165
6A	=	TN166
7F	=	TN167
8F	=	TN168
9A	=	TN169
10A	=	TN170

11F	=	TN161
12F	=	TN162
13A	=	TN163
14A	=	TN164
15F	=	TN165



Metrology

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhoi, Saraburi 18110, Thailand.



Certificate No. T240070

Page 4 of 4

Calibration Report

Measurement Results:

Calibration Point	Average Standard Reading at each position (°C)									
	TN161	TN162	TN163	TN164	TN165	TN166	TN167	TN168	TN169	TN170
	3.17	3.11	3.11	3.33	2.94	3.06	2.95	3.17	2.86	2.59
3	TN161	TN162	TN163	TN164	TN165					
	2.74	2.95	2.75	2.95	2.85					

Chamber (Cooling Room)		Temperature Distribution				
Setting (°C)	Reading (°C)		Average (°C)		Stability (± °C)	Coverage
	Min , Max	Average	Uniformity (°C)		Uncertainty (± °C)	Factor k
3.0	2.9 , 3.1	3.0	2.97	0.29	0.64	0.80
						2.00

* The quoted uncertainty exclude " uniformity "

The calibration result apply only the above calibrated item.

The result of test was found accurate as shown on date and place of test only.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k which for a t-distribution, providing

a level of confidence of approximately 95 % .

Approved By. 

PM-L15 118/18-08-66



Metrology

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhoi, Saraburi 18110, Thailand.

Saraburi Tel : +66 3627 3096 Fax : +66 3627 3100

Bangkok Tel : +668 9205 6851 , +669 8247 2360

Website : www.scieco.co.th E-Mail : calibrate@scg.com



Certificate No. T240161

Page 1 of 4

Certificate of Calibration

Equipment : Chamber (Cooling Room)

Manufacturer : -

Model : -

Serial No. : -

Customer Code : EQL-181

ID No. : T0399A5

Customer : Test Tech Co.,Ltd

30, 32 Rama II Soi 63, Rama II Rd., Samadnam,

Bangkhunthian Bangkok 10150

Customer Location : LABORATORY FLOOR 4

Date of Receipt : 24 January 2024

Calibrated By : Preecha Phisassuthikul (Temperature Calibration Manager)

Approved By :  / Boonchai Suriyawong (Site Calibration Manager)

Date of Issue : 31 JAN 2024

The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standard laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Metrology.

PM-L14 119/18-08-66

Calibration Report

Equipment : Chamber (Cooling Room)

Date of Calibration : 29 January 2024

Environment

Environment : Temperature : 25.4-27.9 °C

Line Voltage : 223.4-227.1 V

Relative Humidity : 45 - 49 %RH

Condition of this results of calibration :

1. This equipment was calibrated by insert 15 standard thermocouples type T into its chamber, the other one standard thermocouples type T use for ambient temperature measurement. The calibration was done in according to WI-720 (based on ASTM E145-94 (Reapproved 2001) and AS2853-1986).

All data show below were final values and the initial data from customer request. The temperature scale used

was based on ITS - 90.

2. Reference Standard Instrument :

Instrument	Model	Instrument No.	Certificate No.	Due Date
TC	TYPE I	TN161-TN170	T230773	10 April 2024
TC	TYPE I	TN171-TN180	T230773	10 April 2024
DATA LOGGER	34970A	T149	T230773	10 April 2024

3. This certificate is traceable to :
National Institute of Metrology (Thailand) through Metrological Center (NSC-TISI-TIS 1702.5 CALIBRATION 0244.)

4. Condition of calibrated item: good

Condition of captured nemt. : good

Equipment Description :

Time Constant, 1 Hour


Time Constant	1 Hour	30 Minute At	3 °C
---------------	--------	--------------	------

Fresh Air Damper ☐ Open ☐ Min ☐ Medium ☐ Max ☐

☐ Open ☐ Close

☐ Close ☒ Not Available

5. Adjustment : (X) without adjustment () after adjustment



C = Centre, F = Centre of Face, A = Corner, E = Centre of Edge

1C	=	TN161
2A	=	TN162
3A	=	TN163
4F	=	TN164
5A	=	TN165
6A	=	TN166
7F	=	TN167
8F	=	TN168
9A	=	TN169
10A	=	TN170
11F	=	TN171

12F	=	TN172
13A	=	TN173
14A	=	TN174
15F	=	TN175

Approved By: Forlin

Approved By: Don Liu



Metrology

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhroi, Saraburi 18110, Thailand.



Certificate No. T240161

Page 4 of 4

Calibration Report

Measurement Results

Calibration Point	Average Standard Reading at each position (°C)									
	TN161	TN162	TN163	TN164	TN165	TN166	TN167	TN168	TN169	TN170
	2.81	3.01	2.99	2.87	2.92	3.08	3.04	2.93	3.31	3.10
	TN171	TN172	TN173	TN174	TN175					
	3.08	3.10	3.40	3.00	3.24					

Chamber (Cooling Room)		Temperature Distribution			
		Setting (°C)	Reading (°C)	Average (°C)	Coverage Factor k
3.0	Min	2.8	3.1	3.0	1.07
	Max	3.0	3.06	0.40	0.92
		Stability (± °C)		Uniformity (°C)	Uncertainty (± °C)
				0.92	1.07
				0.40	2.00

* The Annotated uncertainty exclude "uniformity"
The calibration result apply only the above calibrated item.
The result of test was found accurate as shown on date and place of test only.
The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k which for a t -distribution, providing a level of confidence of approximately 95 % .

Approved By _____

[Signature]

FM-LJ15 118/18-08-66



PerkinElmer
For the Refer

ICPWO-02834138

MAINTENANCE AND IPV TEST CERTIFICATE MODEL OPTIMA 8000

Customer : บริษัท เทสท์ เทคโนโลยี จำกัด	Date Tested: May 14, 2024
Address : 30.32 ซอยพระรามที่ 2 ซอย 63	Recommendation Recertification Period 12 Months
เบอร์โทรศัพท์ 2 แล่นมรณศา	Recertification Due: May 14, 2025
เบอร์โทรแจ้งเหตุฉุกเฉิน กรุณาโทรหาเราตรง 10150	Date Last Certified: May 18, 2023
User Name: คุณสุวิรัตน์ อธิราชธนกิจ	Visit Number: 1 of 1
Phone: 02-893-4211-17	PerkinElmer Phone: 02-719-6420 ext 206
Fax: lab_center@testtech.co.th	PerkinElmer Fax: 02-318-5597

CONFIGURATION TESTED	MODEL	SERIAL NUMBER	SOFTWARE
	OPTIMA 8000	078S141171C	ICP WinLab32 version 5
	(EQL-180)		
TESTED EQUIPMENT	CALIBRATION NUMBER	EXPIRATION	
IPV Methods			
TEST STANDARD USED	PART NUMBER	EXPIRATION DATE	
Multielement Standard	N069-1579	Dec 30, 2024	
Instrument Cal. STD4	N930-0221	Nov 30, 2024	
CUSTOMER SUPPLIED	COMMENTS	CUSTOMER INITIALS	
2 % HNO3			
10 % HNO3			

Page 1 of 4

PerkinElmer Scientific (Thailand) Co., Ltd. 290 Soi Soonvijai 4, Bangkepl, Huay Kwang, Bangkok 10310 Head Office



MAINTENANCE AND IPV TEST CERTIFICATE MODEL
OPTIMA 8000

SERIAL NUMBER: 078S1411171C	DATE TESTED: May 14, 2024
1. MECHANICAL CHECKS	
A. Inspect and clean all fans and filters.	☐
B. Inspect and replace as necessary, all torch components including the RF coil.	☐
C. Inspect all tubing for sign of clacking or leaking.	☐
D. Adjust water and gas pressure regulator settings.	☐
E. Inspect and leak check pneumatics drawers.	☐
F. Clean the exterior of the instrument.	☐
2. OPTICAL CHECKS	
A. Inspect and clean all optical components.	☐
B. As required, check and replace all purgefilters.	☐
C. Recheck optical alignment.	☐
3. COOLING SYSTEM CHECKS	
A. Perform preventive maintenance on chiller.	☐
B. Flush out the chiller every six months.	☐
4. PERFORMANCE CHECKS	
A. Torch View Alignment.	☐
B. Wavelength Calibration.	☐



MAINTENANCE AND IPV TEST CERTIFICATE MODEL
OPTIMA 8000

SERIAL NUMBER: 078S1411171C	DATE TESTED: May 14, 2024	
PARAMETER	SPECIFICATION	FINAL VALUE
Spectral Resolution : UV		
As 193.696 nm	≤ 0.009 nm	0.00735 nm
Ni 231.804 nm	≤ 0.011 nm	0.00913 nm
Ni 341.476 nm	≤ 0.015 nm	0.01386 nm
Spectral Resolution : VIS		
Ba 455.403 nm	≤ 0.020 nm	0.01721 nm
Precision		
Zn 206.200 nm	% RSD ≤ 1.0 %	0.35 %
Mg 280.271 nm	% RSD ≤ 1.0 %	0.27 %
Mg 285.213 nm	% RSD ≤ 1.0 %	0.46 %
Ba 455.403 nm	% RSD ≤ 1.0 %	0.48 %
Detection Limits : Axial		
Tl 190.801 nm	3(sd) ≤ 10.0 ppb	1.00 ppb
As 193.696 nm	3(sd) ≤ 10.0 ppb	3.32 ppb
Se 196.026 nm	3(sd) ≤ 5.0 ppb	3.88 ppb
Pb 220.353 nm	3(sd) ≤ 3.0 ppb	1.45 ppb
Detection Limits : Radial		
As 193.696 nm	3(sd) ≤ 60.0 ppb	3.41 ppb
Zn 213.837 nm	3(sd) ≤ 2.0 ppb	0.30 ppb
Mn 257.610 nm	3(sd) ≤ 1.0 ppb	0.03 ppb
La 379.478 nm	3(sd) ≤ 3.0 ppb	0.27 ppb
Ba 455.403 nm	3(sd) ≤ 0.3 ppb	0.05 ppb
Ba 493.408 nm	3(sd) ≤ 0.6 ppb	0.06 ppb
BEC : Axial (IB X 1000)/(IS-IB)		
Mn 257.610 nm	≤ 30 ppb	10.70 ppb
BEC : Radial (IB X 1000)/(IS-IB)		
Mn 257.610 nm	≤ 30 ppb	21.54 ppb

MAINTENANCE AND IPV TEST CERTIFICATE MODEL
OPTIMA 8000

SERIAL NUMBER: 078S1411171C DATE TESTED: May 14, 2024

Remarks :

Commissioning follow as commissioning performance sheets.

This is to certify that the above tests have been performed and the configuration tested

☒	☐
-------------------------------------	--------------------------

meets
does not meet

the PerkinElmer Specifications listed on this certificate.

This certificate does not modify PerkinElmer's standard terms and condition of sale, including warranty terms.

Service Department PerkinElmer Scientific (Thailand)Co.,Ltd.

Customer Service Engineer:

