

Certificate No.: T/O 660198

Date of issue : 11-Oct-2023

Equipment Description : Refrigerator
Equipment Model : P1010
Equipment Serial No. : P1010-1020-0005
I.D. No. or Control No. : TNP.LAB.01
Manufacturer : Entech Industrial Solution Co.,Ltd.
Customer Name : TNP ENVIRONMENT CO.,LTD.
Customer Address : 332/173 Moo. 3 Tambon Bang Rak Phatthana, Amphoe Bang Bua Thong,
Nonthaburi 11110
Total pages of certificate : 2 pages
Instrument Receiving Date : 9-Oct-2023
Receiving No. : O-230230
Environmental Conditions : All of the measurement were carried out in the working area
Temperature : (25 ± 15) °C
Humidity : (55 ± 30) % RH
Voltage : (220 ± 22) VAC
Calibration Place : 332/173 Moo. 3 Tambon Bang Rak Phatthana, Amphoe Bang Bua Thong,
Nonthaburi 11110
Calibration Procedure No. : This instrument was calibrated by comparison of reference radiation source standard
according to calibration work instration no WI-CL-18-C

The calibration certificate expended uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k , which for a normal distribution corresponds to a coverage probability of approximately 95%

*The standard uncertainty of measurement has been determined in accordance with M 3003
The expression uncertainty and confidence in measurement.*

This certificate is applied only to item under test environmental condition.

*This calibration certificate may not be reproduced other than in full except with the permission of the issuing laboratory.
Calibration certificates without signature and seal are not valid.*

This calibration certificate documents are traceability to national standards, which realize the unit of measurement according to the International system of units (SI).

Date of Calibration : 9-Oct-2023



Mr. Kittipong Kaewsai
Calibration Engineer



Ms. Nongluck Wongsettee
Technical Manager

Certificate No. : T/O 660198


The Reference Standard Instrument :-

Instrument	Model	Serial No.	Cert No.	
1) Data logger with RTD Probe	Agilent 34972A	MY41187730	PSL-T 0651-1/66	21-Apr-2024
		MY60008352	PSL-T 0651-3/66	21-Apr-2024

Measured room conditions

Temperature :	Minimum: 30.8 °C	Maximum: 31.9 °C
Humidity :	Minimum: 50.7 %RH	Maximum: 57.2 %RH
Voltage :	Minimum: 219.8 VAC	Maximum: 223.4 VAC
Fresh Air Setting:	off	

Sensor Position :



Working Space of chamber :

(Inside Dimensions) W x D x H : 1560 mm x 500 mm x 1380 mm

Sensor Installation Details :

- Sensor Number 1 to 12 installed approximately 50 mm From each wall.
- Sensor Number 13 installed approximately geometric of the chamber.

Results : The measurement results of the calibration were reported in the table below.

(*) Without adjustment

() After adjustment

UUC* Setting	UUC* Reading	Temperature Reading of Standard Sensor								
(°C)	(°C)	Sensor Position								
		1	2	3	4	5	6	7	8	9
		4.02	4.35	4.01	4.20	4.37	4.22	4.17	4.39	4.05
		Sensor Position								
		10	11	12	13					
		4.29	4.30	4.28	4.19					

UUC* Setting	UUC* Reading	Temperature Uniformity	Temperature Stability	Overall Variation	Uncertainty of Measurement	Coverage Factor K
(°C)	(°C)	(°C)	(± °C)	(°C)	(± °C)	
4.0	4.1	1.19	1.08	2.47	1.5	2

UUC* = Unit Under Calibration

Remark :-

- Temperature reading of Standard Sensors shown in the table were taken from the average of Standard reading at each position.
- Temperature Uniformity was calculated from the difference between the maximum and minimum of actual temperature reading from all reference sensors at the same time.
- Temperature Stability was calculated from the maximum stability of nine positions, and formula of Stability is $[(\text{Maximum Temperature Value} - \text{Minimum Temperature Value}) / 2]$
- Overall Variation was calculated from the difference between the maximum and minimum measured temperature throughout observation time.

End of Report

CERTIFICATE OF CALIBRATION

FOR

NOMENCLATURE : pH METER
MANUFACTURER : HORIBA
MODEL / TYPE : LAQUA-PH1100/9615S
SERIAL NO. : B80A0042/9X0B0575
CLID. NO. : 272001452
JOB CONTROL NO. : 230911100397

CUSTOMER : TNP ENVIRONMENT CO., LTD.
332/173 MOO 3 TAMBON BANG RAK PHATTANA,
AMPHOE BANG BUA THONG, NONTABURI 11110

DATE OF RECEIVED : 11 September 2023

DATE OF ISSUED : 14 September 2023

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Sukgasem Seehanart
Calibration Engineer



Approved By : Mongkol Yotsoontorn
Authorized Signatory
14 September 2023



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to
the International System of Units (SI)

Certificate No. Q23100397

F3-011-04/01-12

page 1 of 3



@clccalibration

REPORT OF CALIBRATION

FOR

NOMENCLATURE : pH METER
MANUFACTURER : HORIBA
MODEL / TYPE : LAQUA-PH1100/9615S
SERIAL NO. : B80A0042/9X0B0575
DATE OF CALIBRATION : 12 September 2023

ENVIRONMENT CONDITIONS :

Temperature : $(25 \pm 2.5) ^\circ\text{C}$

Relative Humidity : $(50 \pm 15) \% \text{ RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **CLC-CPCH-01**. The calibration was performed by direct measurement with Certified Reference Material (CRM).

REFERENCE STANDARD USED :

1. pH Standard Solution, NIMT TRM CODE TRM-S-2003, TRM CODE TRM-S-2007.
2. pH Standard Solution, Control Company Catalog Number 06664263,11784256, Lot Number CC752722.

TRACEABILITY :

1. The measurements are traceable to International System of Units (SI) , through National Institute of Metrology (Thailand).
Lot Number. 040822 , 230822. Due Date 26 April 2024.
2. The measurements are traceable to International System of Units (SI) , through Control Company.
Certificate No. 4288-13355261 , Due Date 06 May 2024.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.

It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2022)"

Certificate No. Q23100397

F3-011-04/01-12

page 2 of 3





CLC
Accredited
ISO/IEC 17025

CALIBRATION LABORATORY Co., LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment () adjustment

The table in the following gives the calibration results and associated measurement uncertainties of pH meter.

CALIBRATION DATA

pH METER RESULT @ 25 °C

Standard pH Buffer Solution (pH)	pH Meter Reading (pH)	pH Meter Reading (mV)	Correction (pH)	Uncertainty of pH Measurement (\pm pH)	k Factor
4.003	4.01	150.2	-0.007	0.010	2,00
7.000	7.00	-26.1	0.000	0.015	2,06
10.003	10.01	-187.1	-0.007	0.016	2,05

Technical Note. Setting function CAL 3 point (4,7,10).

The Scope of Accredited ANAB Certificate No. ACDM-2814 Version 008 Page 2,3 of 54

This report is valid for the above stated instrument/s only.

CLC

End of Certificate

Certificate No. Q23100397

F3-011-04/01-12

page 3 of 3



@clccalibration

CERTIFICATE OF CALIBRATION

FOR

NOMENCLATURE : DIGITAL THERMO-HYGROMETER
MANUFACTURER : EXTECH INSTRUMENTS
MODEL / TYPE : 445814
SERIAL NO. : PONPE5816745
CLID. NO. : 232303263
JOB CONTROL NO. : 230911100396

CUSTOMER : TNP ENVIRONMENT CO., LTD.
332/173 MOO 3 TAMBON BANG RAK PHATTANA,
AMPHOE BANG BUA THONG, NONTABURI 11110

DATE OF RECEIVED : 11 September 2023

DATE OF ISSUED : 15 September 2023

Report of calibration screening must not be taken in part. Except complete. Without the approval of the Calibration Laboratory Co., Ltd.

Calibrated By :

Oranut Kamchatphai

Calibration Engineer



Approved By :

Mongkol Yotsoontorn

Authorized Signatory

15 September 2023



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the
International System of Units (SI)

Certificate No. Q23100396

F3-011-04/01-12

page 1 of 3



@clccalibration

REPORT OF CALIBRATION

FOR

NOMENCLATURE : **DIGITAL THERMO-HYGROMETER**
MANUFACTURER : **EXTECH INSTRUMENTS**
MODEL / TYPE : **445814**
SERIAL NO. : **PONPE5816745**
DATE OF CALIBRATION : **13 September 2023**

ENVIRONMENT CONDITIONS :

Temperature : $(23 \pm 2) ^\circ\text{C}$

Relative Humidity : $(55 \pm 10) \% \text{RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **CLC-CPTH-11**. The calibration was performed by using Chilled Mirror Hygrometer and Temperature & Humidity Chamber which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

Chilled Mirror Hygrometer, Edgetech Model Dew Master S/N. 36151.

Temperature & Humidity Chamber, PGC Model 9141-5114 S/N.0802282.

TRACEABILITY :

The measurements are traceable to International System of Units (SI) , through Thunder Scientific Corporation.

Certificate No. 21028, Due Date 09 December 2023.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k = 2,00$ which for a normal distribution corresponds to a coverage probability of approximately 95 %.

It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2022)"

Certificate No. Q23100396

F3-011-04/01-12

page 2 of 3



CONDITION OF CALIBRATION ITEM : GOOD

MEASUREMENT RESULTS : (X) without adjustment () adjustment

The table in the following gives the calibration results and associated measurement uncertainties of the measuring digital thermo-hygrometer.

CALIBRATION DATA

1. CORRECTION OF TEMPERATURE

Test point (° C)	Actual Temperature (° C)	DUC Reading (° C)	Correction (° C)	Uncertainty ± (° C)
20.0	20.01	19.9	+0.11	0.27
25.0	25.01	25.2	-0.19	

2. CORRECTION OF HUMIDITY

STD Temperature (° C)	STD Reading (%RH)	DUC Reading (%RH)	Correction (%RH)	Uncertainty ± (%RH)
25	50.0	47	+3.0	0.8

Note. The Scope of Accredited ANAB Certificate No. ACDM-2814 Version 008 Page 49 of 54

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q23100396

F3-011-04/01-12

page 3 of 3



@clccalibration

Certificate of Calibration

Certificate No. : 67-400049-1

Page : 1 of 2

Submitted by : TNP Environment Co., Ltd.
332/173 Moo 3 Bang Rak Phatthana, Bang Bua Thong, Nonthaburi 11110

Equipment : Liquid in Glass Thermometer

Manufacturer : SK Model : N/A
Range : 0 °C to 200 °C Resolution : 1 °C
Serial No. : N/A Immersion : Total
ID No. : TNP.LAB.12

Environment : Ambient Temperature : (23 ± 2) °C
Relative Humidity : (50 ± 15) %
Line Voltage : (220 ± 22) VAC

Date of Received : 26 January 2024

Date of Calibration : 01 February to 02 February 2024

Date of Issue : 02 February 2024

Calibrated by : Chortip Samchusri

Calibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4001 based on ASTM E77-07 by compared with PRT in the liquid bath at the constant controlled temperature.

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Platinum Resistance Thermometer (PRT)

ID No.	Cert. No.	Due Date	Traceability
400001	TT-0016-22	07 Feb 2024	National Institute of Metrology Thailand (NIMT)

2. Standard Digital Thermometer

ID No.	Cert. No.	Due Date	Traceability
400003	23E1866	01 Jun 2025	National Institute of Metrology Thailand (NIMT)
400004	23E1866	01 Jun 2025	National Institute of Metrology Thailand (NIMT)

Approved by :

(Surachai Promthong)

Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 67-400049-1

Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

Ice point check : UUC* reading 0 ° C Standard reading 0.8789 ° C

Standard Reading (° C)	UUC Reading (° C)	Correction (° C)	Uncertainty (± ° C)
21.2064	20	1.2	0.31
31.3084	30	1.3	0.31

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- o0o -



Certificate no: H/T 670338

Date of issue : 21-Mar-24

Instrument description : Thermo-Hygrometer
Instrument model : Extech 445815
Instrument serial no. : PONPE5899554
ID no. or control no. : TNP.LAB.21
Manufacturer : Extech Instruments
Probe description : -
Probe model : -
Probe serial : -
Customer name : TNP ENVIRONMENT CO.,LTD.
Customer address : 332/173 Moo 3 Tambon Bang Rak Phatthana, Amphoe Bang Bua Thong, Nonthaburi 11110

Total pages of certificate : 2 Pages
Receiving no. : L-241004-1
Receiving date. : 08-Mar-24
Parameter of calibration : Temperature Calibration
Condition of UUC. : Used
Ambient condition : All of the Measurement were carried out the stabilized laboratory
Temperature : 23 ± 5 °C
Humidity : 55 ± 15 %RH
Calibration place : 17/121 Soi Ngamwongwan 47 Yaek 48, Toongsonghong, Laksi, Bangkok 10210
Calibration procedure no. : This instrument was calibrated by comparison of indication with the Standard Thermo- hygrometer according to calibration Work Instruction no .WI-CL-11-C

The calibration certificate expanded uncertainty of measurement is stated as the standard uncertainty of measurent Multiplied by coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.

This certificate is applied only to item under test Environmental condition.

This Calibration Certificate may not be reporduced other than in full except with the permission of the issuing laboratory. Calibration certificates without signature and seal not valid.

This calibration certificate documents are tracebility to national standards, which realize measurement according to the International System of Units (SI).

Date of calibration : 20-Mar-24



Mr. Peerawat Thongbai

Calibration Technician



Mrs. Nongluck Wongsettee

Technical Manager

Certificate no: H/T 670338

Standard references

Standard	Reference No.	Vendor	Due Date
ARALAB 300ECP,Fitoclima	S2023070040-001	MIT	07-Jul-24
Thermo HygroPalm HP 23-A	SG-H-00579/66	Success Gateway	16-Aug-24

Measured room conditions

Temperature : 22.1 °C

Humidity : 55.9 %RH

Pressure : 1019.3 mbar

Calibration results (Without Adjustment)

Reference temperature : - °C

Parameter of standard	Standard values	Mean of UUC.	Error	Uncertainty (±)
Temperature (°C)	19.97	20.1	0.13	0.50
Temperature (°C)	25.02	25.2	0.18	0.50
Temperature (°C)	29.99	30.2	0.21	0.50

Remark : -

End of Report

Certificate no: H 670285

Date of issue : 21-Mar-24

Instrument description : Thermo-Hygrometer
Instrument model : Extech 445815
Instrument serial no. : PONPE5899554
ID no. or control no. : TNP.LAB.21
Manufacturer : Extech Instruments
Probe description : -
Probe model : -
Probe serial : -
Customer name : TNP ENVIRONMENT CO.,LTD.
Customer address : 332/173 Moo 3 Tambon Bang Rak Phatthana, Amphoe Bang Bua Thong, Nonthaburi 11110

Total pages of certificate : 2 Pages
Receiving no. : L-241004
Receiving date. : 08-Mar-24
Parameter of calibration : Humidity Calibration
Condition of UUC. : Used
Ambient condition : All of the Measurement were carried out the stabilized laboratory
Temperature : 23 ± 5 °C
Humidity : 55 ± 15 %RH
Calibration place : 17/121 Soi Ngamwongwan 47 Yaek 48, Toongsonghong, Laksi, Bangkok 10210

Calibration procedure no. : This instrument was calibrated by comparison of indication with the Standard Thermo- hygrometer according to calibration Work Instruction no .WI-CL-11-C

The calibration certificate expanded uncertainty of measurement is stated as the standard uncertainty of measurent Multiplied by coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.

This certificate is applied only to item under test Environmental condition.

This Calibration Certificate may not be reporduced other than in full except with the permission of the issuing laboratory. Calibration certificates without signature and seal not valid.

This calibration certificate documents are tracebility to national standards, which realize measurement according to the International System of Units (SI).

Date of calibration : 19-Mar-24

Mr. Peerawat Thongbai

Calibration Technician

Mrs. Nongluck Wongsettee

Technical Manager

Certificate no: H 670285

Standard references

Standard	Reference No.	Vendor	Due Date
Thermo HygroPalm HP 23-A	SG-H-00579/66	Success Gateway	16-Aug-24
Hydrogen 2-XL	Performance	Entech	24-Mar-24

Measured room conditions

Temperature : 22.7 °C **Humidity :** 56.7 %RH **Pressure :** 1013.3 mbar.

Calibration results (Without Adjustment)

Reference temperature : 25.0 °C

Parameter of standard	Standard values	Mean of UUC.	Error	Uncertainty (±)
Humidity (%RH)	35.09	31	-4.09	1.3
Humidity (%RH)	50.03	48	-2.03	1.5
Humidity (%RH)	64.97	68	3.03	1.5

Remark : -

End of Report



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG, BANGKOK 10250
TEL. 0-2717-3000-24 FAX. 0-2719-9484



Certificate of Calibration

Certificate No. : 23M455

Page : 1 of 2

Equipment : Standard Weight Set

Manufacturer: -

Model : Class:F1

Serial No.: 15022021-01

ID No.: TNP.LAB.25

Condition As-Received: Used Item

Received Date: 02 March 2023

Calibration Date: 04 March 2023

Reference: 2303-0104WN

Submitted by: TNP ENVIRONMENT CO.,LTD.

Ambient Temperature: (23 ± 2) °C

Relative Humidity: (50 ± 15) %

Atmospheric Pressure: 1015.25 mbar

This certificate may not be reproduced other than in full,
except with the prior written approval of the head of
Corporate Services 3: Equipment Calibration and Testing Services.

332/173 Moo 3, Bang Rak Phatthana, Bang Bua Thong,
Nonthaburi 11110

Procedure used: Calibration were conducted using in-house calibration procedure CP-M01 according to comparison method against standard weights on the basis of weighings at an average air density of 1.2 kg/m³ and a temperature of 23.4 °C material density of weight is 8000 kg/m³.

Condition of this result of calibration

1.Reference standards instruments :

<u>Instrument</u>	<u>Model</u>	<u>Serial No.</u>	<u>Certificate No.</u>	<u>Due Date</u>
1) Standard Weight Set (E2)	73336	20026	MM-0018-22	28 Feb 2024
2) Standard Weight Set (E2)	73338	20028	MM-0019-22	28 Feb 2024

2.This certificate is not certified for any commercial transaction.

3.The certificate is valid only to the item calibrated on date and place of calibration.

4.This Certification is traceable to the International System of Unit maintained at:-

-National Institute of Metrology Thailand (NIMT)

Calibrated by : Thapakorn Thammachai

Issue Date : 07 March 2023

Approved Signatory : _____

☐ Phalinee Prabpaipal

☒ Sura Suwannasri

☐ Chaowalit Rittirak

B 0309848



Cert No.: 23M455

Page: 2 of 2

Result of calibration

Nominal Value	Conventional mass		Uncertainty of Measurement (\pm)	Maximum Permissible error (\pm)
	Before Adjustment	After Adjustment		
200 g	199.99986 g	-	0.30 mg	1.0 mg
100 g	100.00015 g	-	0.16 mg	0.50 mg
50 g	50.00015 g	-	0.10 mg	0.30 mg
20 g	20.000116 g	-	0.080 mg	0.25 mg
10 g	10.000041 g	-	0.060 mg	0.20 mg
5 g	5.000010 g	-	0.050 mg	0.16 mg
2 g	1.999936 g	-	0.040 mg	0.12 mg
1 g	0.999973 g	-	0.030 mg	0.10 mg
200 mg	200.059 mg	200.007 mg	0.020 mg	0.060 mg
100 mg	100.037 mg	99.981 mg	0.016 mg	0.050 mg

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95 %.

-o0o-

a 1151188



THAI CALIBRATION SERVICES CO., LTD.

19/8 Moo 9 Soi Raiking 30 Puttamonthon 5 Rd., Sampran, Nakornpatom 73210

Tel. 0-3439-7682-5 Fax: 0-3439-7687

www.thaical.com E-mail : sale@thaicalibration.com, lab@thaicalibration.com



CALIBRATION CERTIFICATE

Certificate No.S2306518S

page 1 of 2

Customer : TNP ENVIRONMENT CO., LTD.
332/173 Moo 3 Tambon Bang Rak Phatthana,
Amphoe Bang Bua Thong, Nonthaburi 11110

Equipment : Non-automatic weighing instrument (Electronic instrument)

Manufacturer : Shimadzu **Order No. :** 66S2523-1

Model : AP225WD **Ambient temperature :** $(26.9 \pm 5.0) ^\circ\text{C}$

Accuracy class : - **Relative humidity :** $(52.0 \pm 10.0) \%$

Capacity : 10 g / 220 g **Received date :** 21-Jun-2023

Resolution : 0.00001 g / 0.0001 g **Date of calibration :** 21-Jun-2023

Serial No. : D316301848 **Date of issue :** 24-Jun-2023

ID No. : TNP.LAB.30 **Condition of the balance :** Good working conditions

Place of calibration : ห้อง LAB

Calibration method

This instrument was calibrated according to the EURAMET Calibration Guide No. 18.


Condition of reference standard weight

<u>Instrument</u>	<u>Nominal value</u>	<u>Serial No.</u>	<u>Certificate No.</u>	<u>Due-date</u>	<u>Density (kg/m³)</u>
1 Standard weight set	1 mg to 2 kg	15885+15849	M2210001S	8-Oct-2023	7950

Traceability of the reference standard weight

This certificate is traceable to SI unit through Mass Calibration Laboratory Thai Calibration Services Co., Ltd., NSC-ONSC accredited no. Calibration 0189.

Calibrated By : Aekhasak Silarut
Technician

Approved Signatory : 
Chonlatee Pongwatvisanon

This calibration certificate may not be reproduced other than in full,
except with the prior written approval of the head of TCS calibration laboratory.



THAI CALIBRATION SERVICES CO., LTD.

19/8 Moo 9 Soi Raiking 30 Puttamonthon 5 Rd., Sampran, Nakornpatom 73210

Tel. 0-3439-7682-5 Fax: 0-3439-7687

www.thaical.com E-mail : sale@thaicalibration.com, lab@thaicalibration.com



CALIBRATION CERTIFICATE

Certificate No.S2306518S


page 2 of 2

The repeatability of indication

Nominal Value (g)	Standard Deviation of reading (g)	Maximum difference between suscensive reading (g)	n
100	0.000005	0.00001	5
200	0.00005	0.0001	5

The effect of eccentric application of a load on the indication (test load : 100 g)

Position	Balance Reading (g)
Point 1	100.0000
Point 2	100.0002
Point 3	100.0001
Point 4	100.0000
Point 5	100.0001
Eccentric Value	0.0002



The error of indication

Nominal Value (g)	Value of Reference Standard Weight (g)	Balance Reading (g)	Correction (g)	Uncertainty (±) (g)	k
Unload	0.00000	0.00000	0.00000	0.000016	2.32
0.1	0.10000	0.10003	-0.00003	0.000019	2.10
0.5	0.50000	0.50001	-0.00001	0.000023	2.04
1	1.00001	1.00000	+0.00001	0.000026	2.00
5	5.00000	5.00001	-0.00001	0.000038	2.00
10	9.99999	10.00001	-0.00002	0.000046	2.00
20	20.0000	20.0000	0.0000	0.000085	2.00
50	50.0000	50.0001	-0.0001	0.00011	2.00
100	100.0000	100.0000	0.0000	0.00018	2.00
200	200.0000	200.0004	-0.0004	0.00034	2.00

Remark : Adjustment, External weight nominal value 100 g, Standard weight of Lab

Uncertainty of measurement

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor (k), which for a normal distribution corresponds to a coverage probability of approximately 95% (confidence level).

This report will certify of the calibrated equipment only.

--End--



THAI CALIBRATION SERVICES CO., LTD.

19/8 Moo 9 Soi Raiking 30 Puttamonthon 5 Rd., Sampran, Nakornpatom 73210

Tel. 0-3439-7682-5 Fax: 0-3439-7687

www.thaical.com E-mail : sale@thaicalibration.com, lab@thaicalibration.com



CALIBRATION CERTIFICATE

Certificate No.S2306519S

page 1 of 2

Customer : TNP ENVIRONMENT CO., LTD.

332/173 Moo 3 Tambon Bang Rak Phatthana,

Amphoe Bang Bua Thong, Nonthaburi 11110

Equipment : Non-automatic weighing instrument (Electronic instrument)

Manufacturer : Sartorius

Order No. : 66S2523-2

Model : SECURA224-1S

Ambient temperature : $(26.8 \pm 5.0) ^\circ\text{C}$

Accuracy class : -

Relative humidity : $(52.0 \pm 10.0) \%$

Capacity : 220 g

Received date : 21-Jun-2023

Resolution : 0.0001 g

Date of calibration : 21-Jun-2023

Serial No. : 0041305301

Date of issue : 24-Jun-2023

ID No. : TNP.LAB.31

Condition of the balance : Good working conditions

Place of calibration : ห้อง LAB

Calibration method

This instrument was calibrated according to the EURAMET Calibration Guide No. 18.

Condition of reference standard weight

<u>Instrument</u>	<u>Nominal value</u>	<u>Serial No.</u>	<u>Certificate No.</u>	<u>Due-date</u>	<u>Density (kg/m³)</u>
1 Standard weight set	1 mg to 2 kg	15885+15849	M2210001S	8-Oct-2023	7950

Traceability of the reference standard weight

This certificate is traceable to SI unit through Mass Calibration Laboratory Thai Calibration Services Co., Ltd., NSC-ONSC accredited no. Calibration 0189.

Calibrated By : Aekhasak Silarut
Technician

Approved Signatory :

Chonlatee Pongwatvisanon

This calibration certificate may not be reproduced other than in full,
except with the prior written approval of the head of TCS calibration laboratory.



THAI CALIBRATION SERVICES CO., LTD.

19/8 Moo 9 Soi Raiking 30 Puttamonthon 5 Rd., Sampran, Nakornpatom 73210

Tel. 0-3439-7682-5 Fax: 0-3439-7687

www.thaical.com E-mail : sale@thaicalibration.com, lab@thaicalibration.com



CALIBRATION CERTIFICATE

Certificate No.S2306519S


page 2 of 2

The repeatability of indication

Nominal Value (g)	Standard Deviation of reading (g)	Maximum difference between successive reading (g)	n
200	0.00000	0.0000	5

The effect of eccentric application of a load on the indication (test load : 100 g)

Position	Balance Reading (g)
Point 1	100.0000
Point 2	100.0000
Point 3	100.0000
Point 4	99.9998
Point 5	99.9998
Eccentric Value	0.0002



The error of indication

Nominal Value (g)	Value of Reference Standard Weight (g)	Balance Reading (g)	Correction (g)	Uncertainty (±) (g)	k
Unload	0.0000	0.0000	0.0000	0.000082	2.00
0.1	0.1000	0.1000	0.0000	0.000083	2.00
0.5	0.5000	0.5000	0.0000	0.000084	2.00
1	1.0000	0.9999	+0.0001	0.000085	2.00
5	5.0000	5.0000	0.0000	0.000090	2.00
10	10.0000	10.0000	0.0000	0.000094	2.00
20	20.0000	20.0001	-0.0001	0.00011	2.00
50	50.0000	50.0001	-0.0001	0.00013	2.00
100	100.0000	100.0000	0.0000	0.00019	2.00
200	200.0000	199.9998	+0.0002	0.00033	2.00

Remark : Without adjustment

Uncertainty of measurement

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor (k), which for a normal distribution corresponds to a coverage probability of approximately 95% (confidence level).

This report will certify of the calibrated equipment only.

--End--

CERT.No.: HS-U039F

Certificate of Calibration

Calibration Date : 20 Jun 23
Submitted by : TNP ENVIRONMENT COMPANY LIMITED.
332/173 Moo. 3, Tambon Bang Rak Phatthana,
Amphoe Bang Bua Thong, Nonthaburi 11110

Avg Room Temp : 20 °C
Avg Water Temp : 20 °C
Air Pressure : 757.00 mmHg
Salinity : 0 ppt

Model : YSI 4010-2W
S/N : 22051520
Probe : YSI 4100
S/N : 22C102711
ID NO. : -
Air Temp ref : S/N. E00522
Barometric ref : S/N. E00522
Water Temp ref : S/N. 11431
Technician : Kittipong M.

Calibration Details

Calibration Point	100% air sat. (@20 °C, DO = 9.09 mg/l)	(status)	(status)
Measurement 1 (mg/l)	9.05	(PASS)	-
Measurement 2 (mg/l)	9.05	(PASS)	-
Measurement 3 (mg/l)	9.04	(PASS)	-
Measurement 4 (mg/l)	9.03	(PASS)	-
Measurement 5 (mg/l)	9.04	(PASS)	-
Measurement 6 (mg/l)	9.04	(PASS)	-
Measurement 7 (mg/l)	9.04	(PASS)	-
Measurement 8 (mg/l)	9.03	(PASS)	-
Measurement 9 (mg/l)	9.03	(PASS)	-
Measurement 10 (mg/l)	9.03	(PASS)	-
Mean Measurement	9.04	mg/l	-
Inaccuracy	0.05	mg/l	-

Overall Status (PASS)

Manufacturer Specification

Accuracy = +/- 0.2 mg/l

- 1) This certificate is issued based on the result that are found as shown on date and place of test only.
- 2) The calibration procedure followed in accordance with Harikul Science Co., Ltd.
- 3) This result shall not be used for advertising purpose.



Technician Signature

(Kittipong Maekwong)



Laboratory Manager

(Natenapha Pisatkunchon)

Certificate of Calibration

Certificate No. : 66-400687-1

Page : 1 of 2

Submitted by : TNP Environment Co., Ltd.

332/173 Moo 3 Bang Rak Phatthana, Bang Bua Thong, Nonthaburi 11110

Equipment : Air Chamber (Oven)

Manufacturer : Memmert

Model : UF75

Range : N/A °C

Resolution : 0.1 °C

Serial No. : B320.0251

ID No. : N/A

Environment : On site calibration was carried out at the Laboratory, TNP Environment Co., Ltd.

Ambient Temperature : (27.0 to 28.0) °C

Relative Humidity : (45 to 50) %

Line Voltage : (228.0 to 230.0) V

Date of Received : 11 December 2023

Date of Calibration : 11 December 2023

Date of Issue : 14 December 2023

Calibrated by : Permpon Chanpu

Calibration Method : CAL-M4004, TLAS G-20

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Digital Thermometer with Thermocouple probe

ID No.

Cert. No.

Due Date

Traceability

400046 & 400023 66-400547-1

04 Apr 2024

National Institute of Metrology Thailand (NIMT)

Approved by :



(Surachai Promthong)

Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 66-400687-1


Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber

W = 0.40 m

D = 0.33 m

H = 0.56 m

Capacity = 0.07 m³

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (± °C)
			1	2	3	4	5	6	7	8	9	
85.0	85.0	85.0	85.3	85.1	85.2	85.1	85.0	84.8	84.7	84.7	85.0	0.64
104.0	104.0	104.0	104.3	104.2	104.3	104.1	104.0	103.8	103.7	103.7	104.0	0.69
180.0	180.0	180.0	179.5	179.9	180.0	179.8	179.9	179.4	178.9	179.4	180.0	0.95

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Uniformity (°C)	Measured Stability (°C)	Overall Variation (°C)
85.0	85.0	85.0	0.4	0.1	0.8
104.0	104.0	104.0	0.4	0.1	0.8
180.0	180.0	180.0	1.4	0.3	1.6

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- o0o -

Handwritten signature





ID LINE : IEC17025



Certificate of Calibration

Certificate Number : SPR23110535-2

Page : 1 of 3

Customer : TNP ENVIRONMENT CO.,LTD.

332/173 Moo.3, Bang Rak Phatthana, Bang Bua Thong, Nonthaburi
11110

Equipment Name : Water Quality Meter

Manufacturer : Digicon

Model : WA-48SD

Serial Number : T.075714

ID. Number : TNP-LAB-46

Environmental Conditions

Ambient Temperature : $23^{\circ}\text{C} \pm 2^{\circ}\text{C}$

Relative Humidity : $50\% \pm 15\%$

Location of Calibration : In-Lab

Calibration Procedure : SP-CPC-04-01,
SP-CPC-04-02,

Method of Calibration : SP-CPC-04-11

Received Date : 30 Nov 2023

Calibration Date : 19 Jan 2024

Recommend Due Date : 19 Jan 2025

Date of Issue : 20 Jan 2024

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs.

The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Sarawut Khitmai

Calibration Officer

Approved by :

(Mr.Prayoon Topart)

Authorized Signatory



ID LINE : IEC17025



Calibration Report

Certificate Number : SPR23110535-2

Page : 2 of 3

Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Zero Oxygen Solution	HI7040L	Lot S0027-23 _	21C31	21 Mar 2028
Standard pH Solution	PH016.L5	Lot No.882984	61267077	20 Mar 2024
Standard pH Solution	PH107.L5	Lot No.882985	61275614	13 Apr 2024
Standard pH Solution	PH020.L5	Lot No.882986	61268050	20 Mar 2024
Conductivity Standard 84 uS/cm	CS84M0S.L5	Lot No.882987	61247444	20 Mar 2024
Conductivity Standard 1413 uS/cm	CS1413M0S.L5	Lot No.882988	61267992	20 Mar 2024
Conductivity Standard 12880 uS/cm	CS1288P1S.L5	Lot No.882989	61247253	20 Mar 2024
Sodium Chloride Standard Solution	RM003461L25	Lot No.841770	97756699	23 Sep 2024

Traceability

This certification is traceable to the International System of Unit maintained at :
HANNA - Hanna Instruments (Thailand) Ltd.

C.P.A. Chem - ANAB#AT-1836 (ISO/IEC 17025:2017) and ANAB#AR-1835 (ISO/IEC
17034:2016)



ID LINE : IEC17025



Result of Calibration

Certificate No. : SPR23110535-2

Page : 3 of 3

pH Measurement @ 25 °C

Unit : pH

Standard Solution	UUC Reading	Error	Uncertainty (±)
4.008	3.98	-0.028	0.014
6.984	7.01	0.026	0.012
10.011	10.02	0.009	0.018

Conductivity Measurement @ 25 °C

Standard Solution	UUC Reading	Error	Uncertainty (±)
84 µS/cm	84.6 µS/cm	0.6 µS/cm	0.60 µS/cm
1.413 mS/cm	1.415 mS/cm	0.002 mS/cm	0.0082 mS/cm
12.88 mS/cm	12.89 mS/cm	0.01 mS/cm	0.075 mS/cm

* Dissolved Oxygen Permanance Test

Unit : mg/L

Actual Standard	UUC Reading	Error	Uncertainty (±)
0.0	0.0	0.0	0.13
5.0	4.9	-0.1	0.13
8.3	8.2	-0.1	0.13

Salinity Measurement

Unit : % Salinity

Actual Standard	UUC Reading	Error	Uncertainty (±)
5.0269	5.04	0.0131	0.015

Note:

The result of calibration was found accurate as show on date and place of calibration only.
This Certificate is not certified for any commercial transaction.
Calibration Marked (*) "Not ANAB Accredited " in this Certificate have been included for completeness.

Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor $k = 2$, providing a level of confidence approximately 95%.

- End of Certificate -



CALIBRATION CERTIFICATE

Date of Issue Jun 28, 2024

Cert No. 24/2419

Site Calibration

Order No. 24060337

Customer Bureau Veritas AQ Lab (Thailand) Limited
111 Thailand Science Park
Moo 9 Paholyothin Rd., Klong 1, Klong Luang, Pathumthani 12120 Thailand

Place of Calibration Prepare Culture Media

Description Autoclave

Model VE-65

Serial No. 7202

ID.No. CHM000206

Date of Receipt Jun 26, 2024

Date of Calibration Jun 26, 2024

Environment

Temperature	(Min)	22.1	°C	(Max)	25.5	°C
Relative Humidity	(Min)	42.7	%RH	(Max)	56.3	%RH

Calibration Method

WI-22 : The reference thermometer was placing into the chamber and measurement based on BS-2646.
The temperature scale in use at this laboratory is the International Temperature Scale of 1990.



CALIBRATION CERTIFICATE


Date of Issue Jun 28, 2024

Cert No. 24/2419

Site Calibration

Order No. 24060337

Results (without adjustment)



Position of reference thermometers were placed

Note.

- 1). Duration of the holding time was 45 minutes.
- 2). Stability - greatest one half of difference between max peak and min peak of each reference probe measured temperature obtained during the calibration interval.
- 3). Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.



CALIBRATION CERTIFICATE

Date of Issue Jun 28, 2024

Cert No. 24/2419

Site Calibration

Order No. 24060337

Results (without adjustment)

UUC Setting (°C)	UUC Reading		Reference Thermometer (°C)		Stability ±(°C)	Uniformity (°C)	Uncertainty ±(°C)
	(°C)	(KPa)					
115.0	115.3	179.6	Position 1	115.58	0.25	0.15	0.70
			Position 2	115.58			
			Position 3	115.48			

UUC Setting (°C)	UUC Reading		Reference Thermometer (°C)		Stability ±(°C)	Uniformity (°C)	Uncertainty ±(°C)
	(°C)	(KPa)					
121.0	121.4	211.0	Position 1	121.59	0.20	0.13	0.70
			Position 2	121.60			
			Position 3	121.52			

Standard

1) Standard Datalogger with RTD Model. HiTemp. 140-PT S/N. R32397, Certificate No. QR23-1962, Calibrated by Quality Reborn Co., Ltd., ONAC Calibration No. 0292. Due Date Aug 07, 2024.

2) Standard Datalogger with RTD Model. HiTemp. 140-PT S/N. R32398, Certificate No. QR23-1963, Calibrated by Quality Reborn Co., Ltd., ONAC Calibration No. 0292. Due Date Aug 07, 2024.

3) Standard Datalogger with RTD Model. HiTemp. 140-PT S/N. R32399, Certificate No. QR23-1964, Calibrated by Quality Reborn Co., Ltd., ONAC Calibration No. 0292. Due Date Aug 07, 2024.

This certificate is traceable to SI unit.



CALIBRATION CERTIFICATE

Date of Issue Jun 28, 2024

Site Calibration

Cert No. 24/2419


Order No. 24060337

The stability and uniformity was taken into account in the measurement uncertainty stated.

The above results are valid exclusively for calibration samples as mentioned in the report.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with ONAC requirements.

APPROVED SIGNATORY :


-  
- ☐ MR. PRAJUCKPETCH THONGSOOKCHOTE
☐ MR. DAMRONG MULSING
☒ MR. JATURAPAT THONGSOOKCHOTE

Thermology

Autoclave

Model. VE-65 S/N. 7202 ID.No. CHM000206

Cert.No. 24/2419



28.7



CALIBRATION CERTIFICATE

Date of Issue Jun 28, 2024

Cert No. 24/2418

Site Calibration

Order No. 24060337

Customer Bureau Veritas AQ Lab (Thailand) Limited
111 Thailand Science Park
Moo 9 Paholyothin Rd., Klong 1, Klong Luang, Pathumthani 12120 Thailand

Place of Calibration Incubation Room

Description Water Bath

Model SC100

Serial No. 0152187501160414

ID.No. CHM000205

Date of Receipt Jun 24, 2024

Date of Calibration Jun 24, 2024

Environment

Temperature	(Min)	22.8	°C	(Max)	25.2	°C
Relative Humidity	(Min)	44.1	%RH	(Max)	58.5	%RH
Line Voltage	(Min)	227.2	VAC	(Max)	229.6	VAC

Calibration Method

WI-18 : The reference thermometer was placed into the chamber and measurement was performed based on AS-2853.

The temperature scale in use at this laboratory is the International Temperature Scale of 1990.

Standard

1) Data Acquisition with Sensor Model 34972A S/N. MY49007789, Certificate No. QR24-0186, Calibrated by Quality Reborn Co., Ltd., ONAC Calibration No. 0292. Due Date Jan 23, 2025.

This certificate is traceable to SI unit.




CALIBRATION CERTIFICATE

Date of Issue Jun 28, 2024
Site Calibration

Cert No. 24/2418
Order No. 24060337

Results (without adjustment)



Position of reference thermometers were placed

Note.

- 1). Dimension (W x L x H) is 30 x 34 x 20 cm.
- 2). Stability - greatest one half of difference between max peak and min peak of each reference probe measured temperature obtained during the calibration interval.
- 3). Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.



CALIBRATION CERTIFICATE

Date of Issue Jun 28, 2024

Cert No. 24/2418

Site Calibration

Order No. 24060337

Results (without adjustment)


Cal Point (°C)	UUC Setting (°C)	UUC Reading (°C)	Reference Thermometer (°C)		Stability \pm (°C)	Uniformity (°C)	Uncertainty \pm (°C)
44.5	44.5	44.5	Position 1	44.490	0.020	0.043	0.13
			Position 2	44.494			
			Position 3	44.491			
			Position 4	44.499			
			Position 5	44.503			

The stability and uniformity was taken into account in the measurement uncertainty stated.

The above results are valid exclusively for calibration samples as mentioned in the report.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with ONAC requirements.


APPROVED SIGNATORY :


☐ MR. PRAJUCKPETCH THONGSOOKCHOTE
☐ MR. DAMRONG MULSING
☒ MR. JATURAPAT THONGSOOKCHOTE

Water Bath

Cert.No. 24/2418

Model. SC100 S/N. 0152187501160414 ID.No. CHM000205



Handwritten signature and initials.



CALIBRATION CERTIFICATE

Date of Issue Jun 28, 2024

Cert No. 24/2415

Site Calibration

Order No. 24060337

Customer Bureau Veritas AQ Lab (Thailand) Limited
111 Thailand Science Park
Moo 9 Paholyothin Rd., Klong 1, Klong Luang, Pathumthani 12120 Thailand

Place of Calibration Incubation Room

Description Incubator

Model IN110

Serial No. D415.0797

ID.No. CHM000181

Date of Receipt Jun 24, 2024

Date of Calibration Jun 24, 2024

Environment

Temperature (Min) 22.8 °C (Max) 25.2 °C

Relative Humidity (Min) 44.1 %RH (Max) 58.5 %RH

Calibration Method

WI-17 : The reference thermometer was placed into the chamber and measurement was performed based on AS-2853.

The temperature scale in use at this laboratory is the International Temperature Scale of 1990.

Standard

1) Data Acquisition with Sensor Model 34972A S/N. MY49010059, Certificate No. QR24-0874, Calibrated by Quality Reborn Co., Ltd., ONAC Calibration No. 0292. Due Date Apr 24, 2025.

This certificate is traceable to SI unit.



CALIBRATION CERTIFICATE


Date of Issue Jun 28, 2024

Site Calibration

Cert No. 24/2415

Order No. 24060337

Results (without adjustment)



Position of reference thermometers were placed

Note.

- 1). Dimension (W x L x H) is 56 x 40 x 48 cm.
- 2). Stability - greatest one half of difference between max peak and min peak of each reference probe measured temperature obtained during the calibration interval.
- 3). Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.



CALIBRATION CERTIFICATE

Date of Issue Jun 28, 2024

Cert No. 24/2415

Site Calibration

Order No. 24060337

Results (without adjustment)

Cal Point (°C)	UUC Setting (°C)	UUC Reading (°C)	Reference Thermometer (°C)		Stability \pm (°C)	Uniformity (°C)	Uncertainty \pm (°C)
35.0	35.0	35.0	Position 1	35.138	0.067	0.253	0.30
			Position 2	35.099			
			Position 3	35.075			
			Position 4	35.187			
			Position 5	35.173			
			Position 6	34.988			
			Position 7	34.878			
			Position 8	34.965			
			Position 9	34.970			

The stability and uniformity was taken into account in the measurement uncertainty stated.

The above results are valid exclusively for calibration samples as mentioned in the report.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with ONAC requirements.

APPROVED SIGNATORY :


☐ MR. PRAJUCKPETCH THONGSOOKCHOTE

☐ MR. DAMRONG MULSING

☒ MR. JATURAPAT THONGSOOKCHOTE

Incubator
Model: IN110 S/N: D415.0797 ID.No: CHM000181

Cert.No. 24/2415



Handwritten signature



CRYSTAL CALIBRATION SALES AND SERVICE CO., LTD.

45/48 Soi Salathammasop31, Salathammasop Rd.,
Salathammasop, Thawewatthana, Bangkok 10170 Thailand
Tel : 0-2408-8474-5 Fax : 0-2408-8477 Email : info@crystalcal.com www.crystalcal.com



CERTIFICATE OF CALIBRATION

Certificate No. : 23-0879-019

Issue Date : 30 June 2023

Work Order No. : 23/0879

Customer Name : BUREAU VERITAS AQ LAB (THAILAND) COMPANY LIMITED
111 Thailand Science Park, Moo 9 Paholyotin Road,
Klong Nueng, Klong Luang, Pathumthani 12120, Thailand

Date of Received : 28 June 2023

Date of Calibration : 28 June 2023

Instrument Details : Description : Water Bath
Manufacturer : ThremoFisher
Model : N/A
Serial No. : 0152187501160414
ID No. : CHM000205
Resolution : 0.1 °C
Location : Laboratory

Calibration Method : This instrument was calibrated by insert standard thermometer into the liquid bath according to calibration procedure CWI-T-11 in-house methods based on ASTM E715-80 (Reapproved 2006)


Environmental Conditions :

Temperature : Area Monitoring between 15°C to 40°C
Humidity : Area Monitoring between 30%RH to 85%RH
Line Voltage : Area Monitoring 220 VAC \pm 10%

Traceability of Measurement :

This certificate of calibration documents the traceability to national standard, which realize the unit of measurement according to the International system of Units (SI) and The temperature scale in use at this laboratory is The International Temperature scale of 1990.

Calibrated by : Mr. Thichakorn Srisupob
Calibration Engineer

Approved by : 
(Mr. Anuwat Yaklermjit)
Laboratory Manager

This certificate may not be reproduced other than in full except with the prior written approval of Crystal Calibration Sales and Service co., Ltd.

Crystal Calibration Sales and Service Co., Ltd.

PAGE 1/3

45/48 Salathammasop 31, Salathammasop Rd., Salathammasop, Thawewatthana, Bangkok 10170

Phone : 0-2408-8474 Fax : 0-2408-8477 <http://www.crystalcal.com> Email : info@crystalcal.com





CRYSTAL CALIBRATION SALES AND SERVICE CO., LTD.

45/48 Soi Salathammasop31, Salathammasop Rd.,
Salathammasop, Thawewatthana, Bangkok 10170 Thailand
Tel : 0-2408-8474-5 Fax : 0-2408-8477 Email : info@crystalcal.com www.crystalcal.com



CERTIFICATE OF CALIBRATION

Certificate No. : 23-0879-019

Issue Date : 30 June 2023

Work Order No. : 23/0879

Details of calibration

1. Reference Standards Instrument

Instrument	Model	Serial No. / ID No.	Certificate No.	Due Date
Data Acquisition unit	34972A	MY59002085	22-1146-021	22 November 2023
Sensor type	RTD	Channel 101 to 106	22-1146-021	22 November 2023

2. Certificate traceble

: This certificate traceable to The International System of Unit refer to
Crystal Calibration Sales and Service Co., Ltd. , NAC Calibration No. 0260

3. Condition of item

: Used

4. Calibration site

: On-site

5. Result of Calibration

: Without Adjustment

6. Evaluate Condition


: Time Constant : - Hour 33 Minute At Cal. point 44.5 °C
Type of Control : PID Control

Testing liquid bath use media is Water

7. Calibration note

: The results reported in this certificate refer to the condition of instrument on
the process into the standby state of Liquid Bath

8. Sensors Installation Diagram



Position Diagrams

Av

**CRYSTAL CALIBRATION SALES AND SERVICE CO., LTD.**

45/48 Soi Salathammasop31, Salathammasop Rd.,
Salathammasop, Thawewatthana, Bangkok 10170 Thailand
Tel : 0-2408-8474-5 Fax : 0-2408-8477 Email : info@crystalcal.com www.crystalcal.com



CERTIFICATE OF CALIBRATION

Issue Date : 30 June 2023

Certificate No. : 23-0879-019

Work Order No. : 23/0879

Result of Temperature Distribution and Performance Check

Table 1 : Reporting of Temperature

Calibration point (°C)	Average Measured Temperature (°C) @ Sensor No. (Sensor No.5 is REF)					Uncertainty ± (°C)
	#1	#2	#3	#4	#5	
44.5	44.50	44.50	44.50	44.50	44.51	0.13

Table 2 : Reporting of Characterization Result

Indicator Set point (°C)	Indicator Reading (°C)			Stability ± (°C)	Uniformity (°C)	Overall variation (°C)
	MAX	MIN	Average			
44.4	44.4	44.4	44.4	0.04	0.07	0.07

Note :

Calibrate items in good condition and this report customer request and accepted in certificate

The reference sensor is preferably located at the center of bath

The measured temperature data readout by software "Benchlink Datalogger 3"

The quoted uncertainty include " Stability " and exclude " Loading effect (20% of Temp Uniformity) "

Stability - one-half of the greatest maximum difference of measured temperatures at any one sensor.

Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the bath under steady state conditions.

Overall Variation - The difference of the maximum and minimum measured temperatures throughout observation time.

Indicating Temperature - the average reading of indicating device that forms the integral part of the enclosure.

Av

This result of calibration was found accurate as shown on date and place of calibration only.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k=2$ providing a level of confidence of approximately 95%.



CRYSTAL CALIBRATION SALES AND SERVICE CO., LTD.

45/48 Soi Salathammasop31, Salathammasop Rd.,
Salathammasop, Thawewatthana, Bangkok 10170 Thailand
Tel : 0-2408-8474-5 Fax : 0-2408-8477 Email : info@crystalcal.com www.crystalcal.com



CERTIFICATE OF CALIBRATION

Certificate No. : 23-0879-008

Issue Date : 30 June 2023

Work Order No. : 23/0879

Customer Name : BUREAU VERITAS AQ LAB (THAILAND) COMPANY LIMITED
111 Thailand Science Park, Moo 9 Paholyotin Road,
Klong Nueng, Klong Luang, Pathumthani 12120, Thailand

Date of Received : 26 June 2023

Date of Calibration : 26 June 2023

Instrument Details : **Description** : Temperature Controlled Enclosures [Incubator]
Manufacturer : memmert
Model : INE 500
Serial No. : E512.0738
ID No. : CHM000151
Resolution : 0.1 °C
Location : Laboratory

Calibration Method : This instrument was calibrated by insert standard thermometer into the chamber according to calibration procedure no. CWI-T-10 follow up to TLAS G-20-1/02-08 (E) : Guidelines for Calibration and Checks of Temperature Controlled Enclosures.


Environmental Conditions :

Temperature : Area Monitoring between 15°C to 40°C
Humidity : Area Monitoring between 30%RH to 85%RH
Line Voltage : Area Monitoring 220 VAC \pm 10%

Traceability of Measurement :

This certificate of calibration documents the traceability to national standard, which realize the unit of measurement according to the International system of Units (SI) and The temperature scale in use at this laboratory is The International Temperature scale of 1990.

Calibrated by : Mr. Wuttinun Yindeepot
Calibration Engineer

Approved by : 
(Mr. Anuwat Yaklarmjit)
Laboratory Manager

This certificate may not be reproduced other than in full except with the prior written approval of Crystal Calibration Sales and Service co., Ltd.





CRYSTAL CALIBRATION SALES AND SERVICE CO., LTD.

45/48 Soi Salathammasop31, Salathammasop Rd.,
Salathammasop, Thawewatthana, Bangkok 10170 Thailand
Tel : 0-2408-8474-5 Fax : 0-2408-8477 Email : info@crystalcal.com www.crystalcal.com



CERTIFICATE OF CALIBRATION

Issue Date : 30 June 2023

Certificate No. : 23-0879-008

Work Order No. : 23/0879

Details of Calibration

1. Reference Standards Instrument

Instrument	Model	Serial No./Ins No.	Certificate No.	Due Date
Data Acquisition unit	34972A	MY59002085	22-1146-021	22 November 2023
Sensor type	RTD	RTD# 101-109	22-1146-021	22 November 2023

2. Certificate traceable : This certificate traceable to The International System of Unit refer to
Crystal Calibration Sales and Service Co., Ltd. , NAC Calibration No. 0260
3. Condition of item : Used
4. Calibration site : On - Site
5. Result of Calibration : Without adjustment
6. Evaluate Condition : Time Constant : - Hour 33 Minute At cal. point 41.5 °C
Air vent : Off
Fan speed status : None Fan Speed
7. Calibration note : The results reported in this certificate refer to the condition of instrument on
the process into the steady state of chamber
8. Sensors Installation Diagram : When ; Sensor installation location in Chamber @ Working Space
A = Distance between sensor and wall of chamber is 5 cm
9. Dimensions of chamber : W = 0.56 m ; D = 0.4 m ; H = 0.48 m




Diagram of Chamber

Signature

**CRYSTAL CALIBRATION SALES AND SERVICE CO., LTD.**

45/48 Soi Salathammasop31, Salathammasop Rd.,
 Salathammasop, Thawewatthana, Bangkok 10170 Thailand
 Tel : 0-2408-8474-5 Fax : 0-2408-8477 Email : info@crystalcal.com www.crystalcal.com

**CERTIFICATE OF CALIBRATION**

Certificate No. : 23-0879-008

Issue Date : 30 June 2023

Work Order No. : 23/0879

Result of Temperature Distribution and Performance Check

Table1 : Reporting of Temperature Distribution

Calibration point (°C)	Average Measured Temperature (°C) @ Sensor No. (Sensor No.9 is REF)									Uncertainty ± (°C)
	#1	#2	#3	#4	#5	#6	#7	#8	#9	
41.5	41.91	41.99	41.77	41.86	41.62	42.18	41.66	41.76	41.80	0.29
42.0	42.46	42.52	42.29	42.39	42.16	42.66	42.19	42.27	42.33	0.26

Table 2 : Reporting of Performance check

Indicator Set Point (°C)	Indicator Reading (°C)			Stability ± (°C)	Uniformity (°C)	Overall variation (°C)
	MAX	MIN	Average			
41.5	41.5	41.5	41.5	0.14	0.51	0.74
42.0	42.0	42.0	42.0	0.10	0.44	0.64

Note

Calibrate items it good condition and this report customer request and accepted in certificate

The reference sensor is preferably located of the geometric center of chamber

The measured temperature data readout by software "Benchlink Datalogger 3"

The quoted uncertainty include " Stability " and " Loading effect (20% of Temp Uniformity) "

Stability - one-half of the greatest maximum difference of measured temperatures at any one sensor.

Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions.

Overall Variation - The difference of the maximum and minimum measured temperatures througout observation time.

Indicating Temperature - the average reading of indicating device that forms the integral part of the enclosure.

This result of calibration was found accurate as shown on date and place of calibration only.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k=2$ providing a level of confidence of approximately 95%.



CRYSTAL CALIBRATION SALES AND SERVICE CO., LTD.

45/48 Soi Salathammasop31, Salathammasop Rd.,
Salathammasop, Thawewatthana, Bangkok 10170 Thailand
Tel : 0-2408-8474-5 Fax : 0-2408-8477 Email : info@crystalcal.com www.crystalcal.com



CERTIFICATE OF CALIBRATION

Certificate No. : 23-0879-010

Issue Date : 30 June 2023

Work Order No. : 23/0879

Customer Name : BUREAU VERITAS AQ LAB (THAILAND) COMPANY LIMITED
111 Thailand Science Park, Moo 9 Paholyotin Road,
Klong Nueng, Klong Luang, Pathumthani 12120, Thailand

Date of Received : 26 June 2023

Date of Calibration : 26 June 2023

Instrument Details : **Description** : Temperature Controlled Enclosures [Incubator]
Manufacturer : memmert
Model : IN110
Serial No. : D415.0797
ID No. : CHM000181
Resolution : 0.1 °C
Location : Laboratory

Calibration Method : This instrument was calibrated by insert standard thermometer into the chamber according to calibration procedure no. CWI-T-10 follow up to TLAS G-20-1/02-08 (E) : Guidelines for Calibration and Checks of Temperature Controlled Enclosures.


Environmental Conditions :

Temperature : Area Monitoring between 15°C to 40°C
Humidity : Area Monitoring between 30%RH to 85%RH
Line Voltage : Area Monitoring 220 VAC \pm 10%

Traceability of Measurement :

This certificate of calibration documents the traceability to national standard, which realize the unit of measurement according to the International system of Units (SI) and The temperature scale in use at this laboratory is The International Temperature scale of 1990.

Calibrated by : Mr. Wuttinun Yindeepot
Calibration Engineer

Approved by : 
(Mr. Anuwat Yaklermjit)
Laboratory Manager

This certificate may not be reproduced other than in full except with the prior written approval of Crystal Calibration Sales and Service co., Ltd.





CRYSTAL CALIBRATION SALES AND SERVICE CO., LTD.

45/48 Soi Salathammasop31, Salathammasop Rd.,
Salathammasop, Thawewatthana, Bangkok 10170 Thailand
Tel : 0-2408-8474-5 Fax : 0-2408-8477 Email : info@crystalcal.com www.crystalcal.com



CERTIFICATE OF CALIBRATION

Certificate No. : 23-0879-010

Issue Date : 30 June 2023

Work Order No. : 23/0879

Details of Calibration

1. Reference Standards Instrument

Instrument	Model	Serial No./Ins No.	Certificate No.	Due Date
Data Acquisition unit	34972A	MY49024826	22-1485-003	20 November 2023
Sensor type	RTD	RTD# 301-308, 310	22-1485-003	20 November 2023

2. Certificate traceable

: This certificate traceable to The International System of Unit refer to
Crystal Calibration Sales and Service Co., Ltd. , NAC Calibration No. 0260

3. Condition of item

: Used

4. Calibration site

: On - Site

5. Result of Calibration

: Without adjustment

6. Evaluate Condition

: Time Constant : - Hour 33 Minute At cal. point 35 °C
Air vent : Off
Fan speed status : None Fan Speed

7. Calibration note

: The results reported in this certificate refer to the condition of instrument on
the process into the steady state of chamber

8. Sensors Installation Diagram

: When ; Sensor installation location in Chamber @ Working Space
A = Distance between sensor and wall of chamber is 5 cm

9. Dimensions of chamber

: W = 0.56 m ; D = 0.4 m ; H = 0.48 m




Diagram of Chamber

**CRYSTAL CALIBRATION SALES AND SERVICE CO., LTD.**

45/48 Soi Salathammasop31, Salathammasop Rd.,
 Salathammasop, Thawewatthana, Bangkok 10170 Thailand
 Tel : 0-2408-8474-5 Fax : 0-2408-8477 Email : info@crystalcal.com www.crystalcal.com

**CERTIFICATE OF CALIBRATION**

Issue Date : 30 June 2023

Certificate No. : 23-0879-010

Work Order No. : 23/0879

Result of Temperature Distribution and Performance Check

Table1 : Reporting of Temperature Distribution

Calibration point (°C)	Average Measured Temperature (°C) @ Sensor No. (Sensor No.9 is REF)									Uncertainty ± (°C)
	#1	#2	#3	#4	#5	#6	#7	#8	#9	
35.0	35.05	35.03	35.04	35.07	34.80	34.87	34.78	34.86	35.07	0.26

Table 2 : Reporting of Performance check

Indicator Set Point (°C)	Indicator Reading (°C)			Stability ± (°C)	Uniformity (°C)	Overall variation (°C)
	MAX	MIN	Average			
34.8	34.8	34.8	34.8	0.11	0.33	0.48

Note

Calibrate items it good condition and this report customer request and accepted in certificate

The reference sensor is preferably located of the geometric center of chamber

The measured temperature data readout by software "Benchlink Datalogger 3"

The quoted uncertainty include " Stability " and " Loading effect (20% of Temp Uniformity) "

Stability - one-half of the greatest maximum difference of measured temperatures at any one sensor.

Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions.

Overall Variation - The difference of the maximum and minimum measured temperatures througout observation time.

Indicating Temperature - the average reading of indicating device that forms the integral part of the enclosure.

Zv

This result of calibration was found accurate as shown on date and place of calibration only.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k=2$ providing a level of confidence of approximately 95%.



สำนักมาตรฐานห้องปฏิบัติการ
กระทรวงสาธารณสุข

หนังสือฉบับนี้ให้ไว้เพื่อแสดงว่า

ห้องปฏิบัติการ

บริษัท บุโร เวริทัส เอคิวิ แล็บ (ประเทศไทย) จำกัด

เลขที่ 111 หมู่ที่ 9 ถนนพหลโยธิน ตำบลคลองหนึ่ง

อำเภอคลองหลวง จังหวัดปทุมธานี 12120

ได้รับการขึ้นทะเบียนเป็นห้องปฏิบัติการที่ผ่านการรับรองความสามารถ
ตามมาตรฐาน ISO/IEC 17025 : 2017 และข้อกำหนดและเงื่อนไขการรับรองความสามารถ
ห้องปฏิบัติการทดสอบด้านการแพทย์และสาธารณสุขของสำนักมาตรฐานห้องปฏิบัติการ
ตามรายการและวิธีทดสอบที่กำหนดในเอกสารแนบท้ายในด้าน

การทดสอบอาหารและเครื่องมือแพทย์

(ดร.ภัทรวีร์ สร้อยสังวาลย์)

ผู้อำนวยการสำนักมาตรฐานห้องปฏิบัติการ

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

หมายเลขทะเบียน 1018/46

ห้องปฏิบัติการ บริษัท บุโร เวริทัส เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
1	อาหาร *	1. Total Plate Count / Aerobic Plate Count (CFU)	- FDA BAM <i>Online</i> , 2001 (Chapter 3) - AOAC (2019) 966.23
		2. Coliform Bacteria (MPN, CFU, Detected or not detected)	FDA BAM <i>Online</i> , 2020 (Chapter 4)
		3. <i>Escherichia coli</i> (MPN, CFU, Detected or not detected)	FDA BAM <i>Online</i> , 2020 (Chapter 4)
		4. Fecal Coliforms (MPN)	FDA BAM <i>Online</i> , 2020 (Chapter 4)
		5. <i>Staphylococcus aureus</i> (MPN, CFU, Detected or not detected)	- FDA BAM <i>Online</i> , 2016 (Chapter12) - AOAC (2019) 987.09 - AOAC (2019) 975.55
		6. <i>Vibrio cholerae</i> (Detected or not detected)	FDA BAM <i>Online</i> , 2004 (Chapter 9)
		7. <i>Vibrio parahaemolyticus</i> (MPN, Detected or not detected)	FDA BAM <i>Online</i> , 2004 (Chapter 9)
		8. <i>Vibrio</i> spp. (Detected or not detected)	FDA BAM <i>Online</i> , 2004 (Chapter 9)

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 1 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ ห่มั่นพล)

ห้องปฏิบัติการ บริษัท บุโร เวอริทัส เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
1	อาหาร *	9. Yeasts and Molds (CFU)	FDA BAM <i>Online</i> 2001, (Chapter18)
		10. <i>Listeria monocytogenes</i> (Detected or not detected)	FDA BAM <i>Online</i> , 2022 (Chapter 10)
		11. <i>Listeria</i> spp. (Detected or not detected)	FDA BAM <i>Online</i> , 2022 (Chapter 10)
		12. <i>Bacillus cereus</i> (CFU)	ISO 7932 : 2004/Amd.1:2020
		13. <i>Bacillus cereus</i> (MPN, CFU)	FDA BAM <i>Online</i> , 2020 (Chapter 14)
		14. <i>Salmonella</i> spp. (Detected or not detected)	FDA BAM <i>Online</i> , 2022 (Chapter 5)
		15. <i>Salmonella</i> spp. (Detected or not detected)	ISO 6579-1:2017/Amd.1:2020
		16. <i>Shigella</i> (Detected or not detected)	Compendium of Methods for the Microbiological Examination of Foods, (APHA), 5 th Edition 2015 , Chapter 37
		17. <i>Clostridium perfringens</i> (CFU, Detected or not detected)	FDA BAM <i>Online</i> , 2001 (Chapter 16)

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 2 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ หมีนพล)

ห้องปฏิบัติการ บริษัท บุโร เวอริทัส เอคิว แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
1	อาหาร *	18. Enterococci/ <i>Enterococcus</i> spp. (CFU)	Compendium of Methods for the Microbiological Examination of Foods, (APHA), 5 th Edition 2015 , Chapter 10
		19. Lactic acid bacteria (CFU)	ISO 15214 : 1998
		20. <i>Campylobacter</i> spp./ <i>Campylobacter jejuni</i> / <i>Campylobacter coli</i> (Detected or not detected)	FDA BAM Online, 2001 (Chapter 7)
		21. Enterobacteriaceae (CFU)	ISO 21528-2:2017
2	-อาหารทะเลและผลิตภัณฑ์ (สด แช่เย็น แช่แข็ง)	22. Aerobic Plate Count (CFU)	AOAC (2019) 2015.13
	-เนื้อสัตว์และผลิตภัณฑ์ (สด แช่เย็น แช่แข็ง) -อาหารพร้อมรับประทาน (ผ่านกรรมวิธี แช่เย็น แช่แข็ง) -เครื่องมือ	23. Coliform (CFU)	AOAC (2019) 991.14

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 3 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00


วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ

 (นายสุรศักดิ์ หมั่นพล)

ห้องปฏิบัติการ บริษัท บุโร เวอริทัส เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
3	- เนื้อสัตว์ (สด ผ่านกรรมวิธี) - กุ้ง (สด แช่แข็ง)	24. <i>Salmonella</i> spp. (Detected or not detected)	AOAC (2019) 2014.01
4	- อาหารพร้อมรับประทาน (ผ่านกรรมวิธี แช่เย็น แช่แข็ง) - เครื่องดื่ม	25. <i>Escherichia coli</i> (CFU)	AOAC (2019) 991.14
5	- อาหารทะเลและผลิตภัณฑ์ (สด แช่เย็น แช่แข็ง) - เนื้อสัตว์และผลิตภัณฑ์ (สด แช่เย็น แช่แข็ง)	26. <i>Escherichia coli</i> (CFU)	AOAC (2019) 998.08
6	อาหารในภาชนะบรรจุ ที่ปิดสนิทรวมอาหารกระป๋อง Commercial sterile อาหารที่มีความเป็นกรดต่ำ (pH > 4.6)	- Incubation test (Normal or Abnormal) 27. Flat sour mesophile bacteria (Detected or not detected) 28. Flat sour thermophile bacteria (Detected or not detected)	- FDA BAM Online, 2001 (Chapter 21A) - Compendium of Methods for the Microbiological Examination of Foods, (APHA), 5 th Edition 2015, Chapter 61

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 4 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ หมั่นพล)

ห้องปฏิบัติการ บริษัท บุโร เวอริทัส เอคิว แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
6	อาหารในภาชนะบรรจุ ที่ปิดสนิทรวมอาหารกระป๋อง Commercial sterile อาหารที่มีความเป็นกรดต่ำ (pH > 4.6)	29. Mesophile anaerobe (Detected or not detected) 30. Thermophile anaerobe (Detected or not detected) 31. Mesophile aerobe (Detected or not detected) 32. Thermophile aerobe (Detected or not detected)	- FDA BAM <i>Online</i> , 2001 (Chapter 21A) - Compendium of Methods for the Microbiological Examination of Foods, (APHA), 5 th Edition 2015, Chapter 61
7.	อาหารในภาชนะบรรจุ ที่ปิดสนิทรวมอาหาร กระป๋อง Commercial sterile อาหารที่มีความเป็นกรดสูง (pH ≤ 4.6)	- Incubation test (Normal or Abnormal) 33. Flat sour mesophile bacteria (Detected or not detected) 34. Flat sour thermophile bacteria (Detected or not detected)	- FDA BAM <i>Online</i> , 2001 (Chapter 21A) - Compendium of Methods for the Microbiological Examination of Foods, (APHA), 5 th Edition 2015, Chapter 61

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 5 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ ห่มั่นพล)

**ห้องปฏิบัติการ บริษัท บูโร เวิร์ทส เอคิว แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้**

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
7.	อาหารในภาชนะบรรจุ ที่ปิดสนิทรวมอาหาร กระป๋อง Commercial sterile อาหารที่มีความเป็นกรดสูง (pH ≤ 4.6)	35. Aciduric spoilage bacteria (Detected or not detected) 36. Mesophile aerobe (Detected or not detected) 37. Thermophile aerobe (Detected or not detected)	- FDA BAM <i>Online</i> , 2001 (Chapter 21A) - Compendium of Methods for the Microbiological Examination of Foods, (APHA), 5 th Edition 2015, Chapter 61
8	<ul style="list-style-type: none"> ● น้ำบริโภคน้ำดื่ม - น้ำใช้ในกระบวนการผลิต - น้ำประปา - น้ำบริโภคในภาชนะบรรจุ ปิดสนิท ● น้ำอุปโภค - น้ำจากแหล่งน้ำ ธรรมชาติ - น้ำผิวดิน - น้ำสระว่ายน้ำ - น้ำใช้ในโรงงานที่ ไม่สัมผัสอาหาร ● น้ำแข็ง 	38. Total Plate Count/ Aerobic Plate Count/ Heterotrophic Plate Count (CFU) 39. Coliform Bacteria (MPN, Detected or not detected) 40. <i>Escherichia coli</i> (MPN, Detected or not detected)	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, Part 9215 B Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, Part 9221 B Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, Part 9221 F

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 6 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ หมั่นพล)

**ห้องปฏิบัติการ บริษัท บุโร เวริทัส เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้**

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
8	<ul style="list-style-type: none"> ● น้ำบริโภค <ul style="list-style-type: none"> - น้ำดื่ม - น้ำใช้ในกระบวนการผลิต - น้ำประปา - น้ำบริโภคในภาชนะบรรจุปิดสนิท ● น้ำอุปโภค <ul style="list-style-type: none"> - น้ำจากแหล่งน้ำธรรมชาติ - น้ำผิวดิน - น้ำสระว่ายน้ำ - น้ำใช้ในโรงงานที่ไม่สัมผัสอาหาร ● น้ำแข็ง 	41. Fecal Coliform Bacteria (MPN, Detected or not detected)	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, Part 9221 E
		42. <i>Staphylococcus aureus</i> (CFU, Detected or not detected)	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, Part 9213 B
		43. <i>Salmonella</i> spp. (Detected or not detected)	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, Part 9260 B
		44. <i>Clostridium perfringens</i> (CFU, Detected or not detected)	Environment agency Methods for the Examination of waters and Associated Materials, 2020, Part 6, A-B ,UK
9	<ul style="list-style-type: none"> - Swab - น้ำล้างภาชนะบรรจุอาหาร 	45. Total Plate Count (CFU)	In-housed method TPT-FS-150TM based on FDA BAM <i>Online</i> , 2001 (Chapter 3)
		46. Coliform (CFU, Detected or not detected)	In-housed method TPT-FS-151TM based on FDA BAM <i>Online</i> , 2020 (Chapter 4)
		47. <i>Escherichia coli</i> (CFU, Detected or not detected)	In-housed method TPT-FS-151TM based on FDA BAM <i>Online</i> , 2020 (Chapter 4)

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 7 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ.....(นายสุรศักดิ์ ห่มั่นพล)

ห้องปฏิบัติการ บริษัท บูโร เวิร์ทส เอคิว แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
9	-Swab -น้ำล้างภาชนะบรรจุอาหาร	48. Fecal Coliforms (CFU, Detected or not detected)	In-housed method TPT-FS-151TM based on FDA BAM <i>Online</i> , 2020 (Chapter 4)
		49. <i>Staphylococcus aureus</i> (Detected or not detected)	In-housed method TPT-FS-152TM based on FDA BAM <i>Online</i> , 2016 (Chapter12)
		50. <i>Listeria monocytogenes</i> (Detected or not detected)	FDA BAM <i>Online</i> , 2022 (Chapter 10)
		51. <i>Listeria</i> spp. (Detected or not detected)	FDA BAM <i>Online</i> , 2022 (Chapter 10)
		52. <i>Salmonella</i> spp. (Detected or not detected)	ISO 6579-1:2017/Amd.1:2020
		53. <i>Bacillus cereus</i> (Detected or not detected)	ISO 7932: 2004/Amd.1:2020
		54. <i>Clostridium perfringens</i> (Detected or not detected)	In-housed method TPT-FS-154TM based on FDA BAM <i>Online</i> , 2001 (Chapter 16)
		55. Enterobacteriaceae (CFU, Detected or not detected)	ISO 21528-2:2017

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 8 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ (นายสุรศักดิ์ ห่มั่นพล)

ห้องปฏิบัติการ บริษัท บูโร เวกีทส์ เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
10	<ul style="list-style-type: none"> ● น้ำบริโภค <ul style="list-style-type: none"> - น้ำดื่ม - น้ำใช้ในกระบวนการผลิต ● น้ำบริโภคในภาชนะบรรจุปิดสนิท ● น้ำใช้ในโรงงานที่ไม่สัมผัสอาหาร 	56. Total Solid	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, Part 2540B.
		57. Total Hardness (as CaCO ₃) / Total Hardness	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, Part 2340C.
		58. Turbidity	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, Part 2130B.
		59. pH	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, Part 4500-H+ B.
		60. Odor	TIS 257 Part2-2521:1978
		61. Phenols	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, Part 5530 C.
		62. Chloride	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, Part 4110B.
		63. Fluoride	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, Part 4110B.

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 9 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ หมั่นพล)

ห้องปฏิบัติการ บริษัท บูโร เวกีทส์ เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
10	<ul style="list-style-type: none"> ● น้ำบริโภค <ul style="list-style-type: none"> - น้ำดื่ม - น้ำใช้ในกระบวนการผลิต - น้ำบริโภคในภาชนะบรรจุปิดสนิท ● น้ำใช้ในโรงงานที่ไม่สัมผัสอาหาร 	64. - Anionic Surfactants as Methylene Blue Active Substances (MBAS): MBAS, calculated as Linear Alkylbenzene Sulfonate (LAS) MW = 238.38 - Alkylbenzene Sulfonate	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, Part 5540C.
		65. Nitrate	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, Part 4110B.
		66. Sulfate	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, Part 4110B.
		67. Cyanide	Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, Part 4500-CN E

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 10 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ หมั่นพล)

**ห้องปฏิบัติการ บริษัท บูโร เวกีทส์ เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้**

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
10	<ul style="list-style-type: none"> ● น้ำบริโภค - น้ำดื่ม - น้ำใช้ในกระบวนการผลิต - น้ำบริโภคในภาชนะบรรจุปิดสนิท ● น้ำใช้ในโรงงานที่ไม่สัมผัสอาหาร 	68. Silver (Ag) 69. Aluminium (Al) 70. Barium (Ba) 71. Cadmium (Cd) 72. Chromium (Cr) 73. Copper (Cu) 74. Iron (Fe) 75. Manganese (Mn) 76. Nickel (Ni) 77. Lead (Pb) 78. Zinc (Zn) 79. Arsenic (As) 80. Selenium (Se) 81. Mercury (Hg)	In-house method TPT-FS-233 TM based on Standard Methods for the Examination of Water and Wastewater APHA, AWWA & WEF, 23 rd Edition, 2017, Part 3030E
11	<ul style="list-style-type: none"> ● น้ำบริโภค - น้ำดื่ม - น้ำประปา - น้ำใช้ในกระบวนการผลิต - น้ำบาดาล ● น้ำแข็ง ● น้ำเสีย 	82. Silver (Ag) 83. Aluminium (Al) 84. Barium (Ba) 85. Cadmium (Cd) 86. Chromium (Cr) 87. Copper (Cu) 88. Iron (Fe)	In-house method TPT-FS-233 TM based on Standard Methods for the Examination of Water and Wastewater APHA, AWWA & WEF, 23 rd Edition, 2017, part 3030E

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 11 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ หมั่นพล)

ห้องปฏิบัติการ บริษัท บูโร เวกิธส์ เอคิฟ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
11	<ul style="list-style-type: none"> ● น้ำบริโภค - น้ำดื่ม - น้ำประปา - น้ำใช้ในกระบวนการผลิต - น้ำบาดาล ● น้ำแข็ง ● น้ำเสีย 	89. Manganese (Mn) 90. Nickel (Ni) 91. Lead (Pb) 92. Zinc (Zn) 93. Arsenic (As) 94. Selenium (Se) 95. Mercury (Hg)	In-house method TPT-FS-233 TM based on Standard Methods for the Examination of Water and Wastewater APHA, AWWA & WEF, 23 rd Edition, 2017, part 3030E
12.	<ul style="list-style-type: none"> ● น้ำบริโภค - น้ำดื่ม - น้ำบริโภคในภาชนะบรรจุปิดสนิท - น้ำใช้ในกระบวนการผลิต - น้ำประปา ● น้ำแข็ง ● น้ำเสีย ● น้ำใช้ในโรงงานที่ไม่สัมผัสอาหาร 	96. Silver (Ag) 97. Aluminium (Al) 98. Barium (Ba) 99. Cadmium (Cd) 100. Chromium (Cr) 101. Copper (Cu) 102. Iron (Fe) 103. Manganese (Mn) 104. Nickel (Ni) 105. Lead (Pb) 106. Zinc (Zn) 107. Arsenic (As) 108. Selenium (Se) 109. Mercury (Hg)	In-house method TPT-FS-281 TM based on Standard Methods for the Examination of Water and Wastewater APHA, AWWA&WEF, 23 rd Edition, 2017, part 3125 (A)

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 12 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ  (นายสุรศักดิ์ หมั่นพล)

ห้องปฏิบัติการ บริษัท บุโร เวริทัส เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
13	น้ำแข็ง	110. pH	Standard Methods for the Examination of water and wastewater, APHA, AWWA, WEF, 23 rd Edition ,2017, part 4500-H+B
		111. Total Hardness (as CaCO ₃) / Total Hardness	Standard Methods for the Examination of water and wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, part 2340C
		112. Total Solid	Standard Methods for the Examination of water and wastewater, APHA, AWWA, WEF, 23 rd Edition, 2017, part 2540B
14	เครื่องดื่ม (น้ำและผง)	113. Saccharin	In-house method TPT-FS-265TM based on Bull. Dept. Med. Sci. 1992; 34 (1): p. 31-6
15	-แป้งและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี) -สัตว์น้ำและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี)	114. Vitamin A	In-house method TPT-FS-262TM based on ISO 12080-2: 2000 (E)
		115. Vitamin B1	In-house method TPT-FS-271TM based on AOAC (2019), 942.23 ,953.17 and Food Chemistry (1996), Vol.56, No.1, pp.81-86.
		116. Vitamin B2	

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 13 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ หมั่นพล)

ห้องปฏิบัติการ บริษัท บุโร เวอริทัส เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
16.	-แป้งและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี) -สัตว์น้ำและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี) -นม (น้ำและผง)	117. Vitamin D	In-house method TPT-FS-277TM based on AOAC (2019) 2002.05 and ISO 12080-2 : 2000 (E)
17	- แป้งและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี) - เครื่องดื่ม (น้ำและผง)	118. Vitamin C	In-house method TPT-FS-275TM based on Bull. Dep. med. Sci. 1998; 40 (3):347-57
18	-แป้งและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี) -สัตว์น้ำและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี)	119. Total dietary fiber	AOAC (2019) 985.29
19	-แป้งและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี) -สัตว์น้ำและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี)	120. Total sugar 121. Sugar profiles - Fructose - Glucose - Sucrose - Maltose - Lactose	In-house method TPT-FS-259TM based on AOAC (2019) 982.14 and J. AOAC (1992) Vol 75 No.3

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 14 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ หมั่นพล)

ห้องปฏิบัติการ บริษัท บุโร เวริทัส เอคิว แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
20	-แป้งและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี) -สัตว์น้ำและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี)	122. Saturated Fatty Acid, 123. Trans Fat, 124. Mono unsaturated fat 125. Poly unsaturated fat 126. Fatty acid profiles	In-house method TPT-FS-261TM based on AOAC (2019) 991.39 and ISO 12966- 2:2017
		127. Cholesterol	In-house method TPT-FS-260TM based on AOAC (2019) 994.10 and 976.26
21	-แป้งและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี)	128. Fat	- AOAC (2019) 922.06
	-สัตว์น้ำและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี)		- AOAC (2019) 948.15
	-เนื้อสัตว์และผลิตภัณฑ์		- AOAC (2019) 960.39
	-นมและผลิตภัณฑ์		- AOAC (2019) 989.05
	-อาหารสัตว์และวัตถุดิบ		- AOAC (2019) 954.04
	-อาหารสัตว์เลี้ยง		- AOAC (2019) 948.15

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 15 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ ห่มั่นพล)

ห้องปฏิบัติการ บริษัท บูโร เวิร์ทส เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
22	-แป้งและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี)	129. Protein	- AOAC (2019) 920.87
	-สัตว์น้ำและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี)		- In-house method TPT-FS-257TM based on AOAC (2019) 981.10
	-เนื้อสัตว์และผลิตภัณฑ์		- AOAC (2019) 981.10
	-นมและผลิตภัณฑ์		- AOAC (2019) 991.20
	-อาหารสัตว์และวัตถุดิบ		- AOAC (2019) 984.13
	-อาหารสัตว์เลี้ยง		- AOAC (2019) 984.13
23	-แป้งและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี)	130. Ash	- AOAC (2019) 923.03
	-สัตว์น้ำและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี)		- AOAC (2019) 938.08
	-เนื้อสัตว์และผลิตภัณฑ์		- AOAC (2019) 920.153
	-นมและผลิตภัณฑ์		- AOAC (2019) 945.46
	-อาหารสัตว์และวัตถุดิบ		- AOAC (2019) 942.05
	-อาหารสัตว์เลี้ยง		- AOAC (2019) 942.05

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 16 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ  (นายสุรศักดิ์ หมั่นพล)

**ห้องปฏิบัติการ บริษัท บูโร เวกีส์ เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้**

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
24	-แป้งและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี)	131. Total solid and Moisture	- AOAC (2019) 925.10
	-สัตว์น้ำและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี)		- AOAC (2019) 945.39 (A)
	-เนื้อสัตว์และผลิตภัณฑ์		- AOAC (2019) 952.08
	-นมและผลิตภัณฑ์		- AOAC (2019) 950.46 (B)
	-อาหารสัตว์และวัตถุดิบ		- AOAC (2019) 990.20
	-อาหารสัตว์เลี้ยง		- AOAC (2019) 930.15
25	-แป้งและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี)	132. Total carbohydrate	Ralph Shapiro, 1995 Nutrition Labeling Handbook. Marcel Dekker Inc. New York
	-สัตว์น้ำและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี)	133. Available Total carbohydrate	
	-เนื้อสัตว์และผลิตภัณฑ์	134. Available Energy/Calories	
	-นมและผลิตภัณฑ์นม	135. Energy/Calories	
	-อาหารสัตว์และวัตถุดิบ	136. Calories from fat	
	-อาหารสัตว์เลี้ยง	137. Energy from fat	
26	สัตว์น้ำและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี)	138. Cadmium (Cd)	In- house method TPT-FS-234 TM based on AOAC (2019) 999.10
		139. Lead (Pb)	
		140. Mercury (Hg)	

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 17 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ หมั่นพล)

**ห้องปฏิบัติการ บริษัท บุโร เวริทัส เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้**

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
27	-ผลิตภัณฑ์ส้วม -ผลิตภัณฑ์ผัก -ผลิตภัณฑ์ผลไม้ (สด แช่แข็ง ผ่านกรรมวิธี)	141. Cadmium (Cd) 142. Lead (Pb) 143. Copper (Cu) 144. Iron (Fe) 145. Zinc (Zn) 146. Tin (Sn) 147. Mercury (Hg) 148. Total Arsenic (As)	In- house method TPT-FS-240TM based on AOAC (2019) 999.10
28	-แป้งและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี) -ข้าว (แช่เย็น แช่แข็ง ผ่านกรรมวิธี)	149. Cadmium (Cd) 150. Lead (Pb) 151. Copper (Cu) 152. Iron (Fe) 153. Zinc (Zn) 154. Tin (Sn) 155. Mercury (Hg) 156. Arsenic (As)	In- house method TPT-FS-240TM based on AOAC (2019) 999.10
29	เครื่องดื่ม (น้ำและผง)	157. Lead (Pb) 158. Copper (Cu) 159. Iron (Fe) 160. Zinc (Zn) 161. Tin (Sn) 162. Arsenic (As)	In- house method TPT-FS-272TM based on AOAC (2019) 999.10

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 18 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ ห่มั่นพล)

ห้องปฏิบัติการ บริษัท บุโร เวิร์ทส เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
30	-แป้งและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี) -สัตว์น้ำและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี)	163. Sodium 164. Calcium 165. Iron 166. Potassium	In- house method TPT-FS-252TM based on AOAC (2019) 984.27
31	อาหาร*	167. Cadmium (Cd) 168. Lead (Pb) 169. Copper (Cu) 170. Iron (Fe) 171. Zinc (Zn) 172. Tin (Sn) 173. Mercury (Hg) 174. Arsenic (As)	In-house method TPT-FS-282TM based on AOAC (2019) 2015.01
32	-เครื่องสำอางค์ (น้ำและผง) -ลูกอม	- Synthetic Color 175. Tartrazine 176. Amaranth 177. Indigotine or Indigo Carmine 178. Ponceau 4RC 179. Brilliant Black BN	In-house method TPT-FS-273TM based on AOAC (2019) 930.38 and TIS 696:1987

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 19 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ หมั่นพล)

**ห้องปฏิบัติการ บริษัท บุโร เวริทัส เอคิว แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้**

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
32	-เครื่องดื่ม (น้ำ และ พง) -ลูกอม	180. Sunset Yellow 181. Allura Rad AC 182. Fast Green FCF 183. Brilliant Blue FCF 184. Azorubin or Carmoisine 185. Quinoline Yellow 186. Erythrosin	In-house method TPT-FS-273TM based on AOAC (2019) 930.38 and TIS 696 : 1987
33	-น้ำกะทิ -ผลไม้กระป๋อง	187. Total Acidity	AOAC (2019) 942.15
34	-น้ำกะทิ -ผลไม้กระป๋อง -สัตว์น้ำและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี)	188. Salt (Chlorine as Sodium chloride)	AOAC (2019) 937.09
35	-เครื่องดื่ม (น้ำและพง) -แป้งและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี) -กุ้ง	189. Sulfur dioxide 190. Potassium sulfite 191. Sodium bisulfate 192. Potassium bisulfate 193. Sodium metabisulfite 194. Potassium metabisulfite 195. Sulfite	AOAC (2019) 990.28

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 20 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ หมั่นพล)

ห้องปฏิบัติการ บริษัท บุโร เวอริทัส เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
36	-แป้ง และผลิตภัณฑ์ -เนื้อสัตว์และผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี)	196. Total Phosphorus (Total Phosphorus as P ₂ O ₅ and PO ₄)	In-house method TPT-FS-270TM based on AOAC (2019) 969.31
37	ธัญพืชและผลิตภัณฑ์	197. Aflatoxin (B1, B2, G1, G2)	In-house method TPT-FS-247TM based on AOAC (2019) 991.31 and 994.08
38	-แป้งและผลิตภัณฑ์ (สด แช่แข็ง ผ่านกรรมวิธี) -เครื่องดื่ม	198. Benzoic acid 199. Sodium Benzoate 200. Sorbic Acid 201. Potassium Sorbate	In-house method TPT-FS-264TM based on Bull. Dept. Med. Sci. 1992; 34 (1): p. 31-6
39	อาหารทะเล	202. Chloramphenicol	In-house method TPT-FS-204TM based on USFDA Laboratory Information Bulletin no.4303 Vol.19, No.4 April 2003
40	กุ้งและปลา	203. Malachite green, 204. Leuco Malachite Green 205. Crystal Violet 206. Leuco crystal Violet	In-house method TPT-FS-230TM based on Journal of Chromatography B,788 (2003) 351-359

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 21 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ ห่มนพล)

ห้องปฏิบัติการ บริษัท บุโร เวอริทัส เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
40	กึ่งและปลา	Nitrofurans (Metabolites) 207. 3-amino-2-oxazolididone (AOZ) 208. 3-amino-5-morpholinomethyl-2-oxazolidinone (AMTZ) 209. 1-aminohydantoin (AHD) 210. Semicarbazide (SEM)	In-house method TPT-FS-203TM based on Journal of Chromatography B, 691 (1997) 87-94
		Fluoroquinolones 211. Norfloxacin 212. Ciprofloxacin 213. Danofloxacin 214. Enrofloxacin 215. Sarafloxacin 216. Difloxacin	In-house method TPT-FS-266TM based on Journal of Chromatography A, 974 (2002)
		217. Oxolinic acid 218. Nalidixic acid 219. Flumequine 220. Sparfloxacin 221. Lomefloxacin	In-house method TPT-FS-266TM based on Journal of chromatography A, 974 (2002)

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 22 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ (นายสุรศักดิ์ หมีนพล)

ห้องปฏิบัติการ บริษัท บุโร เวอริทัส เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
41	-ผลไม้และผลิตภัณฑ์ -ผักและผลิตภัณฑ์ -ข้าว -ธัญพืชและผลิตภัณฑ์	Pesticides residue group : -Organochlorine Group 222. aldrin 223. alpha-BHC 224. beta-BHC 225. delta-BHC 226. gamma-BHC 227. cis-Chlordane 228. trans-Chlordane 229. o,p'-DDT 230. p,p'-DDT 231. dicofol 232. dieldrin 233. endosulfan I 234. endosulfan II 235. endosulfan sulfate 236. endrin	In-house method TPT-FS-229TM based on AOAC (2019) 2007.01

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 23 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ หมั่นพล)

ห้องปฏิบัติการ บริษัท บุโร เวริทัส เอคิว แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
41	-ผลไม้และผลิตภัณฑ์ -ผักและผลิตภัณฑ์ -ข้าว -ธัญพืชและผลิตภัณฑ์	Pesticides residue group : -Organochlorine Group 237. heptachlor 238. heptachlor-epoxide 239. methoxychlor 240. o,p'-DDE 241. p,p'-DDE 242. o,p'-DDD 243. p,p'-DDD 244. mirex 245. endrin ketone -Organophosphate Group 246. acephate 247. azinphos-ethyl 248. azinphos-methyl 249. chlorpyrifos 250. chlorpyrifos-methyl 251. dichlorvos 252. diazinon 253. disulfoton 254. dicrotophos 255. dimethoate 256. EPN	In-house method TPT-FS-229TM based on AOAC (2019) 2007.01

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 24 ของทั้งหมด 28 หน้า


แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ  (นายสุรศักดิ์ หมั่นพล)

ห้องปฏิบัติการ บริษัท บุโร เวอริทัส เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
41	-ผลไม้และผลิตภัณฑ์ -ผักและผลิตภัณฑ์ -ข้าว -ธัญพืชและผลิตภัณฑ์	257. ethion 258. fenitrothion 259. malathion 260. methamidophos 261. methidathion 262. mevinphos 263. monocrotophos 264. omethoate 265. parathion-ethyl 266. parathion-methyl 267. phosalone 268. pirimiphos-ethyl 269. pirimiphos - methyl 270. profenofos 271. prothiophos 272. triazophos 273. phosphamidon - Pyrethroids Pesticides 274. bifenthrin 275. cyfluthrin 276. cypermethrin 277. deltamethrin 278. fenvalerate	In-house method TPT-FS-229TM based on AOAC (2019) 2007.01

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 25 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ หมั่นพล)

ห้องปฏิบัติการ บริษัท บุโร เวอริทัส เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
41	-ผลไม้และผลิตภัณฑ์ -ผักและผลิตภัณฑ์ -ข้าว -ธัญพืชและผลิตภัณฑ์	279. lambda-cyhalothrin 280. permethrin 281. fenpropathrin	In-house method TPT-FS-229TM based on AOAC (2019) 2007.01
		- Carbamate Group 282. aldicarb 283. aldicarb sulfone 284. aldicarb sulfoxide 285. carbofuran 286. carbofuran-3-hydroxy 287. carbaryl 288. fenobucarb 289. isoprocarb 290. methiocarb 291. methomyl 292. metolcarb 293. oxamyl 294. propoxur 295. promecarb	In-house method TPT-FS-241TM based on AOAC (2019) 2007.01
		- Other 296. atrazine 297. azoxystrobin 298. iprodione 299. vinclozolin	In-house method TPT-FS-229TM based on AOAC (2019) 2007.01

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 26 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ หมั่นพล)

ห้องปฏิบัติการ บริษัท บุโร เวอริทัส เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

อาหาร* ดังรายละเอียดต่อไปนี้

1. นมและผลิตภัณฑ์
2. ผักและผลิตภัณฑ์ (สด, แห้ง, แช่เย็น, แช่แข็ง)
3. ผลไม้และผลิตภัณฑ์ (สด, แห้ง, แช่เย็น, แช่แข็ง)
4. อาหารทะเลและผลิตภัณฑ์ (สด, แห้ง, แช่เย็น, แช่แข็ง)
5. เนื้อสัตว์และผลิตภัณฑ์ (สด, แห้ง, แช่เย็น, แช่แข็ง)
6. แป้งและผลิตภัณฑ์
7. ถั่วเขียว, บะหมี่
8. บะหมี่กึ่งสำเร็จรูป
9. แยม เนย มาการีน
10. อาหารพร้อมรับประทาน (ผ่านกรรมวิธี, แช่เย็น, แช่แข็ง)
11. สัตว์ปีกและผลิตภัณฑ์ (สด, แห้ง, แช่เย็น, แช่แข็ง)
12. ไข่และผลิตภัณฑ์
13. ไอศกรีม
14. เครื่องปรุงรส
15. ธัญพืช และผลิตภัณฑ์
16. อาหารกระป๋อง และ อาหารในภาชนะที่บรรจุสุญญากาศ
17. เครื่องดื่ม

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 27 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ

(นายสุรศักดิ์ หมั่นพล)

ห้องปฏิบัติการ บริษัท บูโร เวิร์ทส เอคิวิ แล็บ (ประเทศไทย) จำกัด ได้รับการรับรอง
ความสามารถในการทดสอบอาหารและเครื่องมือแพทย์ ดังรายการต่อไปนี้

การทดสอบเครื่องมือแพทย์

ลำดับ	ชนิดผลิตภัณฑ์/ตัวอย่าง	รายการทดสอบ	วิธีทดสอบ
1	- ถุงมือ - รากฟันเทียม - ผ้าปิดปาก - กาวน้/ชุดการแพทย์ - ถุงยางอนามัย - หมวกคลุมผม - ผ้าคลุมทางการแพทย์	300. Bioburden (CFU)	ISO 11737-1:2018/Amd.1:2021

สำนักมาตรฐานห้องปฏิบัติการ

หน้า 28 ของทั้งหมด 28 หน้า

แก้ไขครั้งที่ 00

วันที่แก้ไข 24 มิถุนายน 2565

หมายเลขทะเบียน 1018/46

ให้ไว้ ณ วันที่ 24 มิถุนายน 2565

ถึงวันที่ 23 มิถุนายน 2569

ตรวจสอบความถูกต้องโดย หัวหน้ากลุ่มรับรองห้องปฏิบัติการ..... (นายสุรศักดิ์ หมั่นพล)