

ภาคผนวก 9

สำเนาเอกสารขึ้นทะเบียน

ห้องปฏิบัติการวิเคราะห์เอกชนและ

เอกสารสอบเทียบเครื่องมือตรวจวัด



Ref No. : 0303/16367

CERTIFICATE OF TESTING LABORATORY ACCREDITATION

This is to certify that

HVE Co., Ltd.

**603 Soi Jarunsanitwong 46, Jarunsanitwong Road,
Bangyeekhan, Bang Phlat, Bangkok 10700**

has successfully undergone assessment according to ISO/IEC 17025 : 2017
and under the Bureau of Laboratory Accreditation, Department of Science Service
for the requirements, regulations and criteria for the competence of testing laboratories

LABORATORY ACCREDITATION
Accreditation Number TESTING - 0090
BLA-DSS

The scope of accreditation is as annexed hereto

Issue date : 22nd November 2021

Expired date : 21st November 2025

Signature :

(Mrs. Pochaman Tagheen)

Director of Bureau of Laboratory Accreditation

Bureau of Laboratory Accreditation, Department of Science Service,
Ministry of Higher Education, Science, Research and Innovation



ที่ อก ๐๓๑๐(๑)/ ๑ ๒ ๓ ๗ ๒

กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท
เขตราชเทวี กรุงเทพฯ ๑๐๕๐๐

๒ ๘ สิงหาคม ๒๕๖๖

เรื่อง ขันทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

เรียน กรรมการผู้จัดการ บริษัท เอชวีอี จำกัด

อ้างถึง คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และชนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน
ลงวันที่ ๑๘ พฤษภาคม ๒๕๖๖

สิ่งที่ส่งมาด้วย เอกสารแนบท้ายหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
บริษัท เอชวีอี จำกัด จำนวน ๒ แผ่น

ตามหนังสือที่อ้างถึง บริษัท เอชวีอี จำกัด ขอขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน พร้อม
รายชื่อผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์และรายการสารมลพิษ
ที่จะทำการวิเคราะห์ ต่อกรมโรงงานอุตสาหกรรม นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว ให้บริษัท เอชวีอี จำกัด ขันทะเบียนห้องปฏิบัติการ
วิเคราะห์เอกชน มีเลขทะเบียน ว-๓๕๘ สถานที่ตั้งเลขที่ ๖๐๓ ซอยจรัญสนิทวงศ์ ๔๖ แขวงบางยี่ขัน
เขตบางพลัด กรุงเทพมหานคร โดยมีองค์ประกอบดังนี้

ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์

- | | |
|-----------------------------|----------------------------|
| ๑) นายเอกลักษณ์ สีสาบิหาร | ทะเบียนเลขที่ ว-๓๕๘-ค-๐๐๐๑ |
| ๒) นางสาวทิพวรรณ วงศ์บุญตัน | ทะเบียนเลขที่ ว-๓๕๘-ค-๐๐๐๒ |
| ๓) นายพล ม่วงใหญ่ | ทะเบียนเลขที่ ว-๓๕๘-ค-๐๐๐๓ |

ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์

- | | |
|------------------------------|----------------------------|
| ๑) นางสาวสุปรียา หล้าอิน | ทะเบียนเลขที่ ว-๓๕๘-จ-๐๐๐๑ |
| ๒) นางสาวจิราภรณ์ ผงผานอก | ทะเบียนเลขที่ ว-๓๕๘-จ-๐๐๐๒ |
| ๓) นายธรรมรัตน์ จรรย์วัฒนสุข | ทะเบียนเลขที่ ว-๓๕๘-จ-๐๐๐๓ |
| ๔) นางสาวธนภรณ์ คำทา | ทะเบียนเลขที่ ว-๓๕๘-จ-๐๐๐๔ |
| ๕) นางสาวณัฐธิดา คงบัน | ทะเบียนเลขที่ ว-๓๕๘-จ-๐๐๐๕ |
| ๖) นายฐานันท์ นิภารัตน์ | ทะเบียนเลขที่ ว-๓๕๘-จ-๐๐๐๖ |
| ๗) นายมนโรด สุดจันทร์ | ทะเบียนเลขที่ ว-๓๕๘-จ-๐๐๐๗ |

ค. ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนให้วิเคราะห์ในน้ำเสียและอากาศเสีย ตามสิ่งที่


ส่งมาด้วย

หนังสือฉบับ...

หนังสือฉบับนี้มีอายุครั้งละ ๓ ปี นับจากวันที่กรมโรงงานอุตสาหกรรมออกหนังสือหากประสงค์จะต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ให้ยื่นคำขอต่ออายุพร้อมเอกสารประกอบคำขอต่อกรมโรงงานอุตสาหกรรม ภายใน ๓๐ วัน ก่อนวันสิ้นอายุของหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชนทั้งนี้สามารถยื่นคำขอผ่านระบบอิเล็กทรอนิกส์ได้ที่หน้าเว็บไซต์กรมโรงงานอุตสาหกรรม

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ



(นายประสม ดำรงพงษ์)

ผู้อำนวยการกองวิจัยและเตือนภัยมลพิษโรงงาน
ปฏิบัติราชการแทนอธิบดีกรมโรงงานอุตสาหกรรม

กองวิจัยและเตือนภัยมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๐๓-๕

โทรสาร ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๑๙

ไปรษณีย์อิเล็กทรอนิกส์ saraban@diw.mail.go.th



"อุตสาหกรรมก้าวไกล ประเทศไทยก้าวหน้า ร่วมกันพัฒนา อุตสาหกรรมสีเขียว"



เอกสารแนบท้ายหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท เอชวีอี จำกัด

เลขทะเบียน ว-๓๕๘

ที่ อก ๐๓๑๐(๑)/ ๑ ๒ ๓ ๗ ๒

ลงวันที่ ๒๘ สิงหาคม ๒๕๖๖

ขอขายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๓๑ รายการ

น้ำเสีย จำนวน 30 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Aldicarb	High-Performance Liquid Chromatographic Method ⁽¹⁾
2	Aldicarb sulfone	High-Performance Liquid Chromatographic Method ⁽¹⁾
3	Aldicarb sulfoxide	High-Performance Liquid Chromatographic Method ⁽¹⁾
4	Aldrin	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽¹⁾
5	α-BHC	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽¹⁾
6	Biochemical Oxygen Demand	5-Day BOD Test, Azide Modification Method ⁽¹⁾
7	Carbaryl	High-Performance Liquid Chromatographic Method ⁽¹⁾
8	Carbofuran	High-Performance Liquid Chromatographic Method ⁽¹⁾
9	Chemical Oxygen Demand	Closed Reflux, Titrimetric Method ⁽¹⁾
10	4,4'-DDD	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽¹⁾
11	4,4'-DDE	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽¹⁾
12	Dieldrin	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽¹⁾
13	Endosulfan I	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽¹⁾
14	Endosulfan II	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽¹⁾
15	Endrin	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽¹⁾
16	Heptachlor	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽¹⁾
17	Heptachlo Epoxide	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽¹⁾
18	3-Hydroxycarbofuran	High-Performance Liquid Chromatographic Method ⁽¹⁾
19	Methiocarb	High-Performance Liquid Chromatographic Method ⁽¹⁾
20	Methomyl	High-Performance Liquid Chromatographic Method ⁽¹⁾
21	1-Naphthol	High-Performance Liquid Chromatographic Method ⁽¹⁾
22	Oil & Grease	Liquid-Liquid, Partition-Gravimetric Method ⁽¹⁾
23	Oxamyl	High-Performance Liquid Chromatographic Method ⁽¹⁾
24	pH	Electrometric Method ⁽¹⁾
25	Propoxur	High-Performance Liquid Chromatographic Method ⁽¹⁾
26	Sulfide	Iodometric Method ⁽¹⁾
27	Temperature	Laboratory and Field Methods ⁽¹⁾
28	Total Dissolved Solids	Dried at 180 °C ⁽¹⁾
29	Total Kjeldahl Nitrogen	Macro Kjeldahl Method ⁽¹⁾
30	Total Suspended Solids	Dried at 103-105 °C ⁽¹⁾

อากาศเสีย ...

อากาศเสีย (ปล่องระบาย) จำนวน 1 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Opacity	Ringelmann's Method ^[2] วิกิ

เอกสารอ้างอิง

1. APHA, AWWA, WEF. Standard Methods for the Examination of Water and Wastewater. 23rd ed. Washington, DC: APHA, 2017.
2. กระทรวงอุตสาหกรรม. ประกาศกระทรวงอุตสาหกรรม เรื่อง กำหนดค่าปริมาณเขม่าควันที่เจือปนในอากาศที่ระบายออกจากปล่องของหม้อน้ำของโรงงาน พ.ศ. 2549. ราชกิจจานุเบกษา. 4 ธันวาคม 2549. เล่มที่ 123 ตอนพิเศษ 125ง.

ที่ E10091220091219



สำนักงานทะเบียนหุ้นส่วนบริษัทกลาง
กรมพัฒนาธุรกิจการค้า กระทรวงพาณิชย์

หนังสือรับรอง

ขอรับรองว่าบริษัทนี้ ได้จดทะเบียนเป็นนิติบุคคล ตามประมวลกฎหมายแพ่งและพาณิชย์

เมื่อวันที่ 21 มีนาคม 2549 ทะเบียนนิติบุคคลเลขที่ 0105549037865

ปรากฏข้อความในรายการตามเอกสารทะเบียนนิติบุคคล ณ วันออกหนังสือนี้ ดังนี้

1. ชื่อบริษัท บริษัท เอชวีอี จำกัด
2. กรรมการของบริษัทมี 1 คน ตามรายชื่อต่อไปนี้
 1. นางสาวจินตนา อภิวัฒนสังการ/
3. จำนวนหรือชื่อกรรมการซึ่งลงชื่อผูกพันบริษัทได้คือ นางสาวจินตนา อภิวัฒนสังการ ลงลายมือชื่อและประทับตราสำคัญของบริษัท/
- 4.ทุนจดทะเบียน 2,000,000.00 บาท / สองล้านบาทถ้วน/
5. สำนักงานใหญ่ ตั้งอยู่เลขที่ 603 ซอยเจริญสนิทวงศ์ 46 ถนนเจริญสนิทวงศ์ แขวงบางยี่ขัน เขตบางพลัด กรุงเทพมหานคร/
6. วัตถุประสงค์ของบริษัทมี 26 ข้อ ดังปรากฏในสำเนาเอกสารแนบท้ายหนังสือรับรองนี้ จำนวน 3 แผ่น โดยมีลายมือชื่อนาย

ทะเบียนซึ่งรับรองเอกสารเป็นสำคัญ

ออกให้ ณ วันที่ 31 เดือน มกราคม พ.ศ. 2567

(นางวาสนา พูลสวัสดิ์)

นายทะเบียน

คำเตือน : หนังสือรับรองฉบับนี้พิมพ์ออกจากรุ่นฉบับที่เป็นไฟล์อิเล็กทรอนิกส์ การสั่งพิมพ์ถือเป็นสำเนาเอกสาร



กรมพัฒนาธุรกิจการค้า กระทรวงพาณิชย์
Department of Business Development
Ministry of Commerce

หนังสือรับรองฉบับนี้สร้างจากระบบไฟล์อิเล็กทรอนิกส์ ผู้ใช้ควรตรวจสอบข้อความที่ปรากฏบนหนังสือรับรองฉบับนี้ทุกครั้ง
สามารถตรวจสอบภายในระบบผ่านทาง QR Code และเว็บไซต์กรม (www.dbd.go.th) ได้ไม่เกิน 90 วัน
นับจากวันที่ออกหนังสือรับรอง

ก้าวสู่อนาคต
สู่ยุคดิจิทัล

Leading Business
Towards Digital
Transformation



Ref:E6710091220091219

ออกให้ ณ วันที่ : 2024-01-31 T11:07:20+0700

1/5

ที่ E10091220091219



สำนักงานทะเบียนหุ้นส่วนบริษัทกลาง
กรมพัฒนาธุรกิจการค้า กระทรวงพาณิชย์

หนังสือรับรอง

ข้อควรทราบ ประกอบหนังสือรับรอง ฉบับที่ E10091220091219

- นิติบุคคลนี้ได้ส่งงบการเงินปี 2565
- หนังสือรับรองเฉพาะข้อความที่ห้าง/บริษัทได้นำมาจดทะเบียนไว้เพื่อผลทางกฎหมายเท่านั้น ข้อเท็จจริงเป็นสิ่งที่ควรหาไว้พิจารณา
- นายทะเบียนอาจเพิกถอนการจดทะเบียน ถ้าปรากฏว่าข้อความอันเป็นสาระสำคัญที่จดทะเบียนไม่ถูกต้อง หรือเป็นเท็จ



กรมพัฒนาธุรกิจการค้า กระทรวงพาณิชย์
Department of Business Development
Ministry of Commerce

หนังสือรับรองฉบับนี้สร้างจากระบบแบบไฟล์อิเล็กทรอนิกส์ ผู้ใช้ควรตรวจสอบข้อควรทราบท้ายหนังสือรับรองฉบับนี้ทุกครั้ง
สามารถตรวจสอบภายในระบบ/ผ่านทาง QR Code และเว็บไซต์กรม (www.dbd.go.th) ได้ไม่เกิน 90 วัน
นับจากวันที่ออกหนังสือรับรอง

ก้าวสู่ธุรกิจ
สู่ยุคดิจิทัล

Leading Business
Towards Digital
Transformation



Ref:E6710091220091219

ออกให้ ณ วันที่ : 2024-01-31 T11:07:20+0700

2/5

วบ (พิเศษ)

รายละเอียดวัตถุประสงค์

วัตถุประสงค์ทั่วไป

- (1) ซื้อ จัดหา รับ เช่า เช่าซื้อ ถือกรรมสิทธิ์ ครอบครอง ปรับปรุง ใช้ และการจัดการโดยประการอื่น ซึ่งทรัพย์สินใดๆ ตลอดจนดอกผลของทรัพย์สินนั้น
- (2) ขาย โอน จำนอง จำนำ แลกเปลี่ยน และจำหน่ายทรัพย์สินโดยประการอื่น
- (3) เป็นนายหน้า ตัวแทน ตัวแทนค้าต่างในกิจการและธุรกิจทุกประเภท เว้นแต่ในธุรกิจประกันภัย การหาสมาชิกให้สมาคม และการค้าหลักทรัพย์
- (4) กู้ยืมเงิน เบิกเงินเกินบัญชีจากธนาคาร นิติบุคคล หรือสถาบันการเงินอื่น และให้กู้ยืมเงินหรือให้เครดิต ด้วยวิธีการอื่น โดยจะมีหลักประกันหรือไม่ก็ตาม รวมทั้งการรับ ออก โอน และสละหลังตัวเงิน หรือตราสารที่เปลี่ยนมือได้อย่างอื่น เว้นแต่ในธุรกิจธนาคาร ธุรกิจเงินทุน และธุรกิจเครดิตฟองซิเอร์
- (5) ทำการจัดตั้งสำนักงานสาขาหรือแต่งตั้งตัวแทน ทั้งภายในและภายนอกประเทศ
- (6) เข้าเป็นหุ้นส่วนจำกัดความรับผิดชอบในห้างหุ้นส่วนจำกัด เป็นผู้ถือหุ้นในบริษัทจำกัด และบริษัทมหาชนจำกัด

วัตถุประสงค์ประกอบธุรกิจบริการ

- (7) ประกอบกิจการรับเหมาก่อสร้างอาคาร อาคารพาณิชย์ อาคารที่พักอาศัย สถานที่ทำการ ถนน สะพาน เขื่อน อุโมงค์ และงานก่อสร้างอย่างอื่นทุกชนิด รวมทั้งรับทำงานโยธาทุกประเภท
- (8) ประกอบกิจการโรงแรม กภัตตาคาร บาร์ ไนท์คลับ
- (9) ประกอบกิจการขนส่งและขนถ่ายสินค้า และคนโดยสารทั้งทางบก ทางน้ำ ทางอากาศ ทั้งภายในประเทศ และระหว่างประเทศ รวมทั้งรับบริการนำของออกจากท่าเรือตามพิธีศุลกากรและการจัดระวางการขนส่งทุกชนิด
- (10) ประกอบกิจการนำเที่ยว รวมทั้งธุรกิจที่เกี่ยวข้องกับการนำเที่ยวทุกชนิด
- (11) ประกอบกิจการบริการจัดเก็บ รวบรวม จัดทำ จัดพิมพ์และเผยแพร่สถิติ ข้อมูลในทางเกษตรกรรม อุตสาหกรรม พาณิชยกรรม การเงิน การตลาด รวมทั้งวิเคราะห์และประเมินผลในการดำเนินธุรกิจ
- (12) ประกอบกิจการบริการทางด้านกฎหมาย ทางบัญชี ทางวิศวกรรม ทางสถาปัตยกรรม รวมทั้งกิจการโฆษณา
- (13) ประกอบธุรกิจบริการรับคำปรึกษาหารือ รับผิด และการปฏิบัติตามสัญญาของบุคคลอื่น รวมทั้งรับบริการคำประกันบุคคล ซึ่งเดินทางเข้ามาในประเทศหรือเดินทางออกไปต่างประเทศตามกฎหมายว่าด้วยคนเข้าเมือง กฎหมายว่าด้วยภาษีอากร และกฎหมายอื่น
- (14) ประกอบธุรกิจบริการรับเป็นที่ปรึกษาและให้คำแนะนำปัญหาเกี่ยวกับด้านบริหารงานพาณิชยกรรม อุตสาหกรรม รวมทั้งปัญหาการผลิต การตลาดและจัดจำหน่าย
- (15) ประกอบธุรกิจบริการรับเป็นผู้จัดการและดูแลผลประโยชน์ เก็บผลประโยชน์และจัดการทรัพย์สินให้บุคคลอื่น
- (16) ประกอบกิจการโรงพยาบาลเอกชน สถานพยาบาล รักษาคนไข้และผู้ป่วยเจ็บ รับทำการฝึกสอน และอบรมทางด้านวิชาการเกี่ยวกับการแพทย์ การอนามัย
- (17) ประกอบกิจการจัดสร้างและจัดจำหน่ายภาพยนตร์ โรงภาพยนตร์ และโรงมหรสพอื่น สถานที่ตากอากาศ สนามกีฬา สระว่ายน้ำ โบว์ลิง
- (18) ประกอบกิจการให้บริการซ่อมแซม บำรุงรักษา ตรวจสอบ อัดฉีด พ่นน้ำยาแก๊สสำหรับยานพาหนะทุกประเภท รวมทั้งบริการติดตั้ง ตรวจสอบ และแก้ไขอุปกรณ์ ป้องกันวินาศภัยทุกประเภท
- (19) ประกอบกิจการซักรีดเสื้อผ้า ตัดผม แต่งผม เสริมสวย
- (20) ประกอบกิจการรับจ้างถ่ายรูป ล้างอัดขยายรูป รวมทั้งเอกสาร
- (21) ประกอบกิจการสถานบริการอาบอบนวด



กรมพัฒนาธุรกิจการค้า กระทรวงพาณิชย์
Department of Business Development
Ministry of Commerce

หนังสือรับรองฉบับนี้สร้างจากระบบไฟล์อิเล็กทรอนิกส์ ผู้ใช้ควรตรวจสอบข้อความที่ปรากฏบนหนังสือรับรองฉบับนี้ทุกครั้ง
สามารถตรวจสอบภายในระบบ/ผ่านทาง QR Code และเว็บไซต์กรม (www.dbd.go.th) ได้ไม่เกิน 90 วัน
นับจากวันที่ออกหนังสือรับรอง

ก้าวสู่ธุรกิจ
สู่ยุคดิจิทัล

Leading Business
Towards Digital
Transformation




Ref:E6710091220091219

ออกให้ ณ วันที่ : 2024-01-31 T11:07:20+0700

3/5

(22) ประกอบกิจการประมูลเพื่อรับจ้างทำของ ตามวัตถุประสงค์ทั้งหมด ให้แก่บุคคล คณะบุคคล นิติบุคคล ส่วนราชการ และองค์การของรัฐ



กรมพัฒนาธุรกิจการค้า กระทรวงพาณิชย์
Department of Business Development
Ministry of Commerce

หนังสือรับรองฉบับนี้สร้างในรูปแบบไฟล์อิเล็กทรอนิกส์ ผู้ใช้ควรตรวจสอบข้อความที่ปรากฏบนหนังสือรับรองฉบับนี้ทุกครั้ง
สามารถตรวจสอบภายในระบบผ่านทาง QR Code และเว็บไซต์กรม (www.dbd.go.th) ได้ไม่เกิน 90 วัน
นับจากวันที่ออกหนังสือรับรอง

ก้าวสู่ธุรกิจ
สู่ยุคดิจิทัล

Leading Business
Towards Digital
Transformation



Ref:E6710091220091219

ออกให้ ณ วันที่ : 2024-01-31 T11:07:20+0700

4/5

.....
.....
.....

ของ
ห้างหุ้นส่วนบริษัท เอชวีอี จำกัด
ทะเบียนเลขที่ 01055549037865

วัตถุประสงค์ของ ห้างหุ้นส่วนบริษัท นี้ มี 26 ข้อ ดังนี้

(23) ประกอบกิจการรับเหมาก่อสร้างและงานระบบ เช่น ระบบสุขาภิบาล ไฟฟ้า แอร์ ฯลฯ

(24) ประกอบกิจการก่อสร้างและปรับปรุงงานทางด้านสิ่งแวดล้อม เช่น ระบบบำบัดมลพิษทางน้ำ ระบบบำบัดมลพิษทาง

อากาศ ระบบการจัดการมูลพิษจากอุตสาหกรรม ระบบผลิตน้ำประปา ระบบการจัดการมลพิษทางเสียง ระบบดับเพลิง
และงานเกี่ยวกับระบบทอภายในอาคาร รับออกแบบและให้คำปรึกษาเกี่ยวกับกิจการข้างต้น

(25) ประกอบกิจการควบคุมดูแลและซ่อมบำรุงเครื่องจักรอุปกรณ์ทางด้านสิ่งแวดล้อม เช่น ระบบบำบัดน้ำเสีย ระบบบำบัด
มลพิษทางอากาศ ระบบเครื่องทำอากาศ ระบบผลิตน้ำประปา ระบบกำจัดกากไขมันกากตะกอนและกากอุตสาหกรรม
ระบบการจัดการมลพิษทางเสียงและระบบดับเพลิง

(26) รับจ้างวิเคราะห์ทดสอบมลพิษสิ่งแวดล้อมทุกชนิด ทั้งด้านมลพิษทางน้ำ มลพิษทางอากาศ มลพิษจากตะกอน มลพิษ
ด้านโลหะหนัก และมลพิษในดิน



กรมพัฒนาธุรกิจการค้า กระทรวงพาณิชย์
Department of Business Development
Ministry of Commerce

หนังสือรับรองฉบับนี้สร้างในรูปแบบไฟล์อิเล็กทรอนิกส์ ผู้ใช้ควรตรวจสอบข้อความที่ปรากฏบนหนังสือรับรองฉบับนี้ทุกครั้ง
สามารถตรวจสอบภายในระบบ/ช่องทาง QR Code และเว็บไซต์กรม (www.dbd.go.th) ได้ไม่เกิน 90 วัน
นับจากวันที่ออกหนังสือรับรอง

ก้าวสู่ธุรกิจ
สู่ยุคดิจิทัล






Leading Business
Towards Digital
Transformation



Ref:E6710091220091219

ออกให้ ณ วันที่ : 2024-01-31 T11:07:20+0700

5/5

<div style="text-align: center;">  <p>Calibratech Co.,Ltd. 7/106-7 Moo 2, Sukhprachasan 3 Rd., Bangpoo, Pakkred, Nonthaburi 11120 Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com</p> </div>	  <p>NSC-TISI-TIS 17025 CALIBRATION 0030</p>								
<h2 style="color: #0070C0;">Certificate of Calibration</h2>									
Certificate No. :	67-200134-1								
Page : 1 of 2									
Submitted by :	HVE Co.,Ltd. 603 Soi Jarunsanitwong 46, Jarunsanitwong Road, Bangyeekun, Bangplad, Bangkok 10700								
Equipment :	Electronic Balance <div style="display: flex; justify-content: space-between;"> <div> Manufacturer : SHIMADZU Serial No. : D432620040 Capacity : 200 g </div> <div> Model : AX200 ID No. : 114 Resolution : 0.0001 g </div> </div>								
Environment :	On site calibration was carried out at the Laboratory, HVE Co., Ltd. Ambient Temperature : (30.8 to 31.6) °C Relative Humidity : (50.4 to 53.4) % Air Pressure : 1008.0 mbar								
Date of Received :	17 April 2024								
Date of Calibration :	17 April 2024								
Date of Issue :	24 April 2024								
Calibrated by :	Akaradath Thippichai								
Calibration Method :	In-house method CAL-M2001 based on UKAS Publication ref : LAB 14 Edition 7 - November 2022								
Reference Standard Instruments : This certification is traceable to the International System of Units									
Standard Weights									
<table border="0" style="width: 100%;"> <thead> <tr> <th style="text-align: left;">ID No.</th> <th style="text-align: left;">Cert. No.</th> <th style="text-align: left;">Due Date</th> <th style="text-align: left;">Traceability</th> </tr> </thead> <tbody> <tr> <td>E261-E2624</td> <td>C02232088</td> <td>08 Nov 2024</td> <td>National Institute of Metrology (Thailand), (NIMT)</td> </tr> </tbody> </table>	ID No.	Cert. No.	Due Date	Traceability	E261-E2624	C02232088	08 Nov 2024	National Institute of Metrology (Thailand), (NIMT)	
ID No.	Cert. No.	Due Date	Traceability						
E261-E2624	C02232088	08 Nov 2024	National Institute of Metrology (Thailand), (NIMT)						
<div style="display: flex; align-items: center;"> <div style="margin-right: 10px;">Approved by :</div>  </div> <div style="text-align: center; margin-top: 5px;"> (Surachai Promthong) Laboratory Manager </div>									
<div style="display: flex; justify-content: space-between;"> <div> The Uncertainties are for a confidence probability of approximately 95% This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd. </div> <div style="text-align: right;">  </div> </div>									
CAL-F0031-03									

Certificate of Calibration

Certificate No. : 67-200134-1

Page : 2 of 2

Result of Calibration : After Adjustment

UUC Condition As-Received : Good

Departure of indication from nominal value

Nominal Value (g)	Correction (g)	Uncertainty \pm (g)	Error before Adjustment (g)
0.01	0.0000	0.00012	0.0000
0.1	0.0000	0.00012	0.0000
0.5	0.0000	0.00013	0.0000
1	-0.0001	0.00013	0.0000
10	0.0000	0.00013	-0.0002
20	0.0000	0.00014	-0.0003
50	0.0001	0.00015	-0.0004
100	0.0001	0.00020	-0.0007
150	0.0002	0.00038	-0.0014
200	0.0005	0.00038	-0.0019

This result of calibration was found accurate as shown on date and place of calibration only.


This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2.06$, providing a level of confidence of approximately 95%

Eccentric error

Load test : 50 g

A B C D E

0.0003 -0.0006 -0.0003 0.0006 0.0000 g



Repeatability

Load test : 200 g

Stdev. : 0.00005 g

- o0o -

AB





SCIMET Co., Ltd.
1194 Soi Wachirathamsathit 57, Bangchak,
Phrakhanong, Bangkok 10260 Thailand
Email: scimet2022@gmail.com, Tel: 02 460 9239
<https://www.scimet.co.th>



Certificate No. C17240065

Calibration Certificate

Equipment:	Oven	Job No.:	KSMT2400663
Model:	UNB 500	Received Date:	01 April 2024
Serial No.(or ID):	C507.1007 (012)	Issued Date:	03 April 2024
Manufacturer:	Memmert	Page:	1 of 4
Condition:	In Condition		
Ventilation Valve:	Closed	Shelves(pc):	2

Customer

HVE Co., Ltd.
603 Soi Charansanitwong 46, Charansanitwong Road Bang Yi Khan, Bang Phlat, Bangkok 10700

Calibration Place

HVE Co., Ltd. (Laboratory)
603 Soi Charansanitwong 46, Charansanitwong Road Bang Yi Khan, Bang Phlat, Bangkok 10700

Calibration Date

01 April 2024

Environment Condition

Temperature: 30.1 °C ± 1.3 °C
Humidity: 60.9 %RH ± 3.3 %RH

The Method used

In-house method, W17, based on TLAS-G20

Traceability

This certificate is traceable to the SI Units maintained by
National Institute of Metrology (NIMT), Thailand through
Quality Reborn Co.,Ltd.Certificate No. QR23-1906

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor ($k=2$) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).

These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of SCIMET Co., Ltd.

(Mr. Hattapong Purnnil)


Person in charge



(Mr. Thalemgkeat Pongngam)

Authorized signatory

FC17-02: 30 MAY 2023



Standard Installation Locations

Volume (Calibration Zone)= 6 (Liters)

Inside chamber: W = 56 (cm) D = 40 (cm) H = 48 (cm)

Standard Locations (#1, #2, #3, #4): w = 20 (cm) d = 10 (cm) h = 15 (cm)

Standard Locations (#5, #6, #7, #8): w = 20 (cm) d = 10 (cm) h = 15 (cm)

#9: Geometric center of the chamber

Position of Std	#1	#2	#3	#4	#5	#6	#7	#8	#9
Channel of Logger	101	102	103	104	105	106	107	108	109

Definitions

Indicating Temperature: The average reading of indicating device which forms the integral part of the enclosure.

Measured Temperature: The average reading of standards at any positions or location.

Measured Uniformity: The maximum difference of measured temperatures between of any probes and the measured temperature at the reference location which are observed at same time or at close observation time as possible to determine the temperature pattern or homogeneity with the chamber at steady-state. The reference probe is preferably located in the geometric center of the chamber.

Measured Stability: The one-half of greatest maximum difference of measured temperatures at any one probe.

Overall Variation: The difference of maximum and minimum measured temperatures throughout observation time.

บริษัท ชายนัมเมท จำกัด (SCIMET CO., LTD.)

1194 Soi Wachirathamsathit 57, Bangchak, Phrakhanong, Bangkok 10260 Thailand
Email: scimet2022@gmail.com, Tel: 02 460 9239

FC17-02: 30 MAY 2023

Calibration Results:

Pre-Calibration

Desired	Setting	Indicating	#1	#2	#3	#4	#5	#6	#7	#8	#9
104.0	104.0	104.0	103.23	103.17	103.10	103.10	101.81	101.68	101.89	101.61	102.51

Without adjustment

Measurement Temperature at Spread Locations, Indicating of Unit Under Calibration: 105.5 °C

Locations	Measured Temperature (°C)	Correction (°C)	Uncertainty (± °C)
#1	104.48	0.48	0.39
#2	104.51	0.51	0.39
#3	104.43	0.43	0.39
#4	104.45	0.45	0.39
#5	103.20	-0.80	0.39
#6	103.11	-0.89	0.39
#7	103.27	-0.73	0.39
#8	103.07	-0.93	0.39
#9	103.87	-0.13	0.39

Temperature Distribution

Desired (°C)	Setting (°C)	Indicating (°C)	Measured Temperature at Spread Locations (°C)									Uncertainty (± °C)*
			#1	#2	#3	#4	#5	#6	#7	#8	#9	
104.0	105.5	105.5	104.48	104.51	104.43	104.45	103.20	103.11	103.27	103.07	103.87	0.39

Chamber Characterization

Indicating (°C)	Measured Uniformity (°C)	Measured Stability (± °C)	Overall Variation (°C)
105.5	0.89	0.12	1.64

Note: * Maximum uncertainty of the each position

บริษัท ชายนีเมท จำกัด (SCIMET CO., LTD.)

1194 Soi Wachirathamsthit 57, Bangchak, Phrakhanong, Bangkok 10260 Thailand
Email: scimet2022@gmail.com, Tel: 02 460 9239

FC17-02: 30 MAY 2023

Without adjustment (Cont.)

Measurement Temperature at Spread Locations, Indicating of Unit Under Calibration: 182.0 °C

Locations	Measured Temperature (°C)	Correction (°C)	Uncertainty (± °C)
#1	181.05	1.05	0.49
#2	181.24	1.24	0.49
#3	180.99	0.99	0.49
#4	181.18	1.18	0.49
#5	179.64	-0.36	0.50
#6	179.63	-0.37	0.51
#7	179.84	-0.16	0.50
#8	180.00	0.00	0.50
#9	180.18	0.18	0.50

Temperature Distribution

Desired (°C)	Setting (°C)	Indicating (°C)	Measured Temperature at Spread Locations (°C)									Uncertainty (± °C)*
			#1	#2	#3	#4	#5	#6	#7	#8	#9	
180.0	182.0	182.0	181.05	181.24	180.99	181.18	179.64	179.63	179.84	180.00	180.18	0.51

Chamber Characterization

Indicating (°C)	Measured Uniformity (°C)	Measured Stability (± °C)	Overall Variation (°C)
182.0	1.17	0.16	1.83

Note: * Maximum uncertainty of the each position

The End of Certificate

บริษัท ชายนันท์ จำกัด (SCIMET CO., LTD.)

1194 Soi Wachirathamsathit 57, Bangchak, Phrakhanong, Bangkok 10260 Thailand
Email: scimet2022@gmail.com, Tel: 02 460 9239

FC17-02: 30 MAY 2023



Refer to Certificate No.: C17240065

Page: 1 of 2

Statements of conformity:

This conformity certificate documents the validity of the following statements of conformity based on the measurement results of corresponding calibration certificate:

The correction of indication determined during calibration are under given measurement and environmental conditions and considering the expanded measurement uncertainty (coverage probability 95%) within the specification. The given measurement uncertainty already includes other all effects by according to the standard method, TLAS-G20. Therefore, those parameters have not

Tolerance and Decision rules:

Assessment of the conformity of the measurement device are done based on direct comparison of the relevant measurement results with the tolerances and decision rule are prescribed by the customer.

Decision rule : ☐ Choice A Binary Statement for Simple Acceptance Rule ($w = 0$), Specific Risk < 50% PFA.

☒ Choice B Non-binary statement with guard band ($w = 1$ U), Pass or Fail Specific Risk < 2.5% PFA and Condition Pass or Condition Fail Specific Risk < 50% PFA.

☐ Choice C Customer defined, Customers may define arbitrary multiple of r to have applied as guard band ($w = r$ U).

; PFA: Probability of False Accept



(Mr. Thalemgkeat Pongngam)

Authorized signatory

Without adjustment**Desired Temperature : 104.0°C****Tolerances : 1.0 °C****Measurement Temperature at Spread Locations, Indicating of Unit Under Calibration: 105.5 °C**

Locations	Measured (°C)	Correction of UUC. (°C)	Guard band (W) (± °C)	Tolerance (± °C)	Conformity
#1	104.48	0.48	0.39	1.0	Pass
#2	104.51	0.51	0.39	1.0	Pass
#3	104.43	0.43	0.39	1.0	Pass
#4	104.45	0.45	0.39	1.0	Pass
#5	103.20	-0.80	0.39	1.0	Condition Pass
#6	103.11	-0.89	0.39	1.0	Condition Pass
#7	103.27	-0.73	0.39	1.0	Condition Pass
#8	103.07	-0.93	0.39	1.0	Condition Pass
#9	103.87	-0.13	0.39	1.0	Pass

Correction of UUC.* = Measured Temperature - Desired Temperature

The validity of the statements of conformity cannot be guaranteed for different places of use, environmental conditions or improper use.

บริษัท ชัยนิเมท จำกัด (SCIMET CO., LTD.)

1194 Soi Wachirathamsathit 57, Bangchak, Phrakhanong, Bangkok 10260 Thailand
Email: scimet2022@gmail.com, Tel: 02 460 9239

FC17-02: 30 MAY 2023



Refer to Certificate No.: C17240065

Page: 2 of 2

Statements of conformity:(Cont.)

Without adjustment (Cont.)

Desired Temperature : 180.0°C

Tolerances : 2 °C

Measurement Temperature at Spread Locations, Indicating of Unit Under Calibration: 182.0 °C

Locations	Measured (°C)	Correction of UUC. (°C)	Guard band (W) (± °C)	Tolerance (± °C)	Conformity
#1	181.05	1.05	0.49	2	Pass
#2	181.24	1.24	0.49	2	Pass
#3	180.99	0.99	0.49	2	Pass
#4	181.18	1.18	0.49	2	Pass
#5	179.64	-0.36	0.50	2	Pass
#6	179.63	-0.37	0.51	2	Pass
#7	179.84	-0.16	0.50	2	Pass
#8	180.00	0.00	0.50	2	Pass
#9	180.18	0.18	0.50	2	Pass

Correction of UUC.* = Measured Temperature - Desired Temperature



The validity of the statements of conformity cannot be guaranteed for different places of use, environmental conditions or improper use.


The End of Statements of Conformity

บริษัท ชายนันเมก จำกัด (SCIMET CO., LTD.)

1194 Soi Wachirathamsathit 57, Bangchak, Phrakhanong, Bangkok 10260 Thailand
Email: scimet2022@gmail.com, Tel: 02 460 9239

FC17-02: 30 MAY 2023







ใบตรวจสอบสภาพเครื่องควบคุมอุณหภูมิ

เลขที่ใบงาน: KSMT2400663

ชนิดเครื่องมือ: Oven

รุ่น: UNB 500

หมายเลขเครื่อง: C507.1007 (012)

ตรวจสอบ (รับ)		รายการตรวจเช็ค	ตรวจสอบ (ส่ง)		หมายเหตุ
01 Apr 2024			01 Apr 2024		
ปกติ	ไม่ปกติ		ปกติ	ไม่ปกติ	
		General			
<input checked="" type="checkbox"/>	<input type="checkbox"/>	1. สายไฟ	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	2. การทำงาน Main Switch	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	3. การทำงาน Selector Key	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	4. การแสดงผล Display	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	5. การทำงาน พัดลม	<input type="checkbox"/>	<input type="checkbox"/>	ไม่มี
<input checked="" type="checkbox"/>	<input type="checkbox"/>	6. สภาพ Lever of Ventilation valve	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	7. สภาพ Lever door open / close	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	8. สภาพ Door seal	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	9. การทำงานของระบบ Safety	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input type="checkbox"/>	<input type="checkbox"/>	10. การทำงานของระบบทำความเย็น	<input type="checkbox"/>	<input type="checkbox"/>	ไม่มี
<input type="checkbox"/>	<input type="checkbox"/>	11. การทำงานของระบบทำความชื้น	<input type="checkbox"/>	<input type="checkbox"/>	ไม่มี
<input checked="" type="checkbox"/>	<input type="checkbox"/>	12. สภาพตัวเครื่อง	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<input checked="" type="checkbox"/>	<input type="checkbox"/>	13. สภาพแวดล้อม ณ สถานที่ตั้งเครื่อง	<input checked="" type="checkbox"/>	<input type="checkbox"/>	

ข้อแนะนำ :

Mr. Hattapong Pumnil
Service Engineer

บริษัท ชายนีเมท จำกัด (SCIMET CO., LTD.)

1194 Soi Wachirathamsethit 57, Bangchak, Phrakhanong, Bangkok 10260 Thailand
Email: scimet2022@gmail.com, Tel: 02 460 9239

F117-00: 08 MAR 2023

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhaphrachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com



NSC-TISI-TIS 17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 67-400216-6

Page : 1 of 2

Submitted by : HVE Co., Ltd.

603 Soi Jarunsanitwong 46, Jarunsanitwong Road, Bangyeekun, Bangplad, Bangkok 10700

Equipment : Temperature controlled enclosure (Incubator)

Manufacturer : Lovibond

Model : ET636-6

Range : N/A °C

Resolution : 0.1 °C

Serial No. : 9982523-03

ID No. : 011

Environment : On site calibration was carried out at the Laboratory, HVE Co., Ltd.

Ambient Temperature : (27.0 to 28.0) °C

Relative Humidity : (50 to 55) %

Line Voltage : (229.0 to 230.0) V

Date of Received : 17 April 2024

Date of Calibration : 18 April 2024

Date of Issue : 19 April 2024

Calibrated by : Kittisak Kokaao

Calibration Method : CAL-M4004, TLAS G-20

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units
Standard Digital Thermometer with RTD Probe

ID No.

Cert. No.

Due Date

Traceability

400046 & 400047

67-400047-2

26 Jul 2024

National Institute of Metrology Thailand (NIMT)

Approved by :

(Surachai Promthong)

Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



CAL-F0031-03

Certificate of Calibration

Certificate No. : 67-400216-6


Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber

W = 0.53 m

D = 0.43 m

H = 1.40 m

Capacity = 0.32 m³

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (± °C)
			1	2	3	4	5	6	7	8	9	
20.0	21.0	21.0	20.24	20.19	20.28	20.16	20.15	20.08	19.95	19.87	19.98	0.36

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Uniformity (°C)	Measured Stability (°C)	Overall Variation (°C)
20.0	21.0	21.0	0.3	0.1	0.6

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- ๐0๐ -

AA



Certificate of Calibration

Certificate No. : 67-420044-2

Page : 1 of 2

Submitted by : HVE Co., Ltd.

603 Soi Jarunsanitwong 46 Jarunsanitwong Road, Bangyeekun, Bangplad, Bangkok 10700

Equipment : pH Meter with electrode
pH meter

Manufacturer : Hanna Model : HI 2211

Range : N/A pH Resolution : 0.01 pH

Serial No. : 08376721 ID No. : N/A

Electrode

Model : HI 1131 Serial No. : 084809EN

Environment : On site calibration was carried out at the Laboratory, HVE Co., Ltd.

Ambient Temperature : (25.0 to 26.0) °C

Relative Humidity : (40 to 45) %

Date of Received : 18 April 2024

Date of Calibration : 18 April 2024

Date of Issue : 19 April 2024

Calibrated by : Permpoon Chanpu

Calibration Method : In-house method CAL-M4201 direct measurement by using standard voltage calibrator and using certified reference material (CRM)

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Multiproduct Calibrator

ID No.	Cert. No.	Due Date	Traceability
400005	SG-E-00307/66	23 Aug 2025	National Institute of Metrology Thailand (NIMT)

2. Certified Reference Material (CRM)

pH	Cert. No.	Lot No.	Exp. Date	Traceability
4.008	61293328	944535	27 Nov 2025	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025
6.986	61281486	944537	17 Nov 2024	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025
9.997	61281073	944536	17 Nov 2024	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025

Approved by :

(Surachai Promthong)

Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 67-420044-2

Page : 2 of 2

Result of Calibration :

UUC Condition As-Received : Good

Function : Electrical measurement

pH meter

Performing standard curve by Multiproduct Calibrator at pH (4,7) and (7,10)

Adjustment Curve at nominal pH	Applied Voltage (mV)	Nominal Value (pH)	UUC Reading		Correction (mV)	Uncertainty (± mV)
			(pH)	(mV)		
4, 7	177.4800	4	4.00	177.3	0.2	0.12
	0.0000	7	7.00	0.0	0.0	0.086
7,10	0.0000	7	7.00	0.0	0.0	0.086
	-177.4800	10	10.00	-177.4	-0.1	0.12

Function : pH meter with electrode

Performing a three - buffer standard curve using buffer nominal pH (4,7) and (7,10)

Adjustment Curve at nominal pH	Standard Buffer (pH)	UUC Reading (pH)	Correction (pH)	Uncertainty (± pH)
4, 7	4.008	4.01	0.00	0.010
	6.986	7.01	-0.02	0.011
7, 10	6.986	7.01	-0.02	0.011
	9.997	10.01	-0.01	0.014

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- ๐0๐ -

AB



CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhprachusan 3 Rd., Bangpood, Pekkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com



Certificate of Calibration

Certificate No. : 67-400222-2

Page : 1 of 2

Submitted by : HVE Co.,Ltd.

603 Soi Jarunsanitwong 46 Jarunsanitwong Road, Bangyeekun Bangplad Bangkok 10700

Equipment : Digital Thermometer with Thermistor Probe (Temp pH)

Temperature Indicator

Manufacturer : Hanna Model : HI 2211

Range : N/A °C Resolution : 0.1 °C

Serial No. : 08376721 ID No. : N/A

Thermistor Probe

Model : N/A Sheath Material : Stainless

Diameter : 3.5 mm. Length : 100 mm.

Serial No. : N/A ID No. : 08376721

Environment : On site calibration was carried out at the Laboratory, HVE Co., Ltd.

Ambient Temperature : (25.0 to 26.0) °C

Relative Humidity : (40 to 45) %

Line Voltage : (229.0 to 230.0) VAC

Date of Received : 18 April 2024

Date of Calibration : 18 April 2024

Date of Issue : 19 April 2024

Calibrated by : Permpon Chanpu

Calibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4003 by compared with PRT in the liquid bath at the constant controlled temperature.

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Platinum Resistance Thermometer (PRT)

ID No.	Cert. No.	Due Date	Traceability
400002	TT-0074-22	20 Jun 2024	National Institute of Metrology Thailand (NIMT)

2. Standard Digital Thermometer

ID No.	Cert. No.	Due Date	Traceability
400033	24E633	21 Feb 2026	National Institute of Metrology Thailand (NIMT)

Approved by :

(Surachai Promthong)

Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



CAL-F0031-03

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhaphrasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

Certificate of Calibration

Certificate No. : 67-400222-2

Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

Immersion Depth (mm.)	Standard Reading (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
100	25.002	24.8	0.2	0.19

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- o0o -

APD



CAL-F0031-03



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3 : EQUIPMENT CALIBRATION AND TESTING SERVICES

534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250

TEL. 0-2717-3000 FAX. 0-2719-9484

Cert.No.: 24TW74

Page.: 1 of 2

Certificate of Testing

Equipment : DO Meter
Manufacturer : Digicon
Model : DO-552SD
Serial No. : AG.35318
ID No. : -
Received Date : 05 April 2024
Test Date : 09 April 2024
Reference : 2404-0175DN-1
Submitted by : HVE Co.,Ltd
603 Soi Jarransanitwong 46, Jarransanitwong Road,
Bang Yi Khan, Bang Phlat Bangkok 10700
Laboratory Condition : Temperature (25 ± 5) °C
Humidity (50 ± 20) %
Test Procedure : In - house method : CP-CH9
by Comparison Technique with Azide Modification Method
Tested by : Walalak Sirithean
Approved by : 
Approved Signatory
() Unnophol Harachai
(✓) Ponpan Paipim
() Saithip Meangmai
Issue Date : 10 April 2024

B 0338488



Cert.No.: 24TW74

Page.: 2 of 2

Condition of this result of calibration

1. Reference Standard Instruments :

This certification is traceable to the International System of Unit through the reference standards laboratory of Industrial Calibration Center, Technology Promotion Association (Thailand-Japan).

<u>Instruments</u>	<u>Serial No.</u>	<u>ID No.</u>	<u>Certificate No.</u>	<u>Due Date</u>
1. Burette	-	130BU10	23CG1172	22 Mar 2025
2. Balance	14233821	110RC001	23MM405	16 July 2024

2. Standard Material :-

<u>Material</u>	<u>Manufacturer</u>	<u>Lot.No.</u>	<u>Assay</u>
Sodium Thiosulfate pentahydrate	Merck	AM1763316	100.2%

Result : Dissolved Oxygen Meter Adjustment With Air 100 %

Dissolved Oxygen Probe No.: 07-07

Titration Method (Azide Modification Method) (mg/L)	DO Meter Reading (mg/L)	Standard Deviation (mg/L)
8.20	8.2	0.045

This report was certified only for the instrument we tested. It is allowable to use for study. Intend to use for advertising and referral purpose is prohibited. This report may not be reproduced other in full, without written approval of the laboratory.

-o0o-

a 1209346



BECTHAI BANGKOK EQUIPMENT & CHEMICAL CO., LTD.
CALIBRATION LABORATORY

300 Phaholyothin Road, Phayathai, Bangkok 10400, Thailand Tel: +66 2615-2929 Fax: +66 2615-2350-1
E-mail: bkk@becthai.com Website: www.becthai.com



Certificate No. : CAL-23-754

Page : 1 of 4

CERTIFICATE OF CALIBRATION

Equipment : Spectrophotometer
Manufacturer : Thermo Scientific
Model : Genesys 10S UV-VIS
Serial No. : 2L9Q310003
ID No. : 071
Customer : HVE CO., Ltd.
: 603 Soi Jarunsanitwong 46, Jarunsanitwong Road,
: Bangyeekun, Bangplad, Bangkok 10700
Location : แผนกน้ำโรค
Date of Receipt : 21 November 2023
Date of Calibration : 21 November 2023
Date of Issue : 21 November 2023
Ambient Temperature : (25±10) °C
Relative Humidity : (60±20) %
Condition As-Received : Used Item

Calibrated by

(Mr.Somphop Duangnguan)

Calibration Engineer

Approved by

(Ms. Jintana Sangthajaroenlap)

Calibration Manager

The reported expanded uncertainty of measurement was based on a combined standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%.

This certificate may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Indicated values are valid for the state of the Spectrophotometer at the time of calibration only.

ISSUE: 5 REV:4

FM-CAL-33/2

15/05/61



BECTHAI BANGKOK EQUIPMENT & CHEMICAL CO., LTD.
CALIBRATION LABORATORY

300 Phaholyothin Road, Phayathai, Bangkok 10400, Thailand Tel: +66 2615-2929 Fax: +66 2615-2350-1
E-mail: bkk@becthai.com Website: www.becthai.com



Certificate No. : CAL-23-754

Page : 2 of 4

CALIBRATION REPORT

Conditions of this result of calibration

1. Reference Standard Material :

<u>Material</u>	<u>Model</u>	<u>Serial No.</u>	<u>Cert.No.</u>	<u>Due date</u>
Holmium Glass Filter	RM-HG	24563	109211	13 Feb 25
Didymium Glass Filter	RM-DG	24562	109212	13 Feb 25
Neutral Density Filter	RM-1N2N3N	24568	109249	14 Feb 25
Potassium Dichromate Solution	RM-06	24567	109222	13 Feb 25

2. Traceability : This certification is traceable to the International System of Unit maintained at;

The Sarna Scientific Ltd. Accredited Calibration Laboratory No. 0659.

3. Method of calibration :

The calibration procedure was carried out according to ASTM E275-08 (2022) and ASTM E925-09 (2014).

4. Result of calibration :

(☒) without adjustment

(☐) after adjustment

5. Equipment Specifications:

Spectral Bandwidth :	1.8	nm
Data Interval :	0.1	nm
Scan Speed :	Slow	nm/min

ISSUE: 5 REV:4

FM-CAL-33/2

15/05/61



BECTHAI BANGKOK EQUIPMENT & CHEMICAL CO., LTD.
CALIBRATION LABORATORY

300 Phaholyothin Road, Phayathai, Bangkok 10400, Thailand Tel: +66 2615-2929 Fax: +66 2615-2350-1
E-mail: bkk@becthai.com Website: www.becthai.com



Certificate No. : CAL-23-754

Page : 3 of 4

CALIBRATION REPORT

Wavelength Calibration

Certified Values of Reference Material (nm)	Nominal Value (nm)	UUC*Reading (nm)	Error (nm)	Uncertainty of Measurement (\pm nm)
361.00	361.00	360.7	-0.30	0.13
536.66	536.66	536.7	0.04	0.13
879.27	879.27	879.8	0.53	0.13

Photometric Calibration for Visible

Wavelength (nm)	Certified Values of Reference Material (A)	UUC* Reading (A)	Error (A)	Uncertainty of Measurement (\pm A)
420.0	Zero	0.000	0.0000	0.0028
	0.5835	0.585	0.0015	0.0044
	0.725	0.725	0.0000	0.0040
	1.0367	1.037	0.0003	0.0039
440.0	Zero	0.000	0.0000	0.0028
	0.5662	0.567	0.0008	0.0042
	0.7106	0.710	-0.0006	0.0037
	1.0159	1.016	0.0001	0.0037
465.0	Zero	0.000	0.0000	0.0028
	0.5257	0.527	0.0013	0.0044
	0.6682	0.668	-0.0002	0.0039
	0.9547	0.954	-0.0007	0.0034
546.1	Zero	0.000	0.0000	0.0028
	0.5226	0.523	0.0004	0.0036
	0.6939	0.693	-0.0009	0.0039
	0.9919	0.991	-0.0009	0.0032
590.0	Zero	0.000	0.0000	0.0028
	0.5567	0.556	-0.0007	0.0035
	0.7502	0.748	-0.0022	0.0037
	1.0732	1.071	-0.0022	0.0033
635.0	Zero	0.000	0.0000	0.0028
	0.5643	0.563	-0.0013	0.0035
	0.7299	0.728	-0.0019	0.0038
	1.0437	1.042	-0.0017	0.0034

Remark : Each individual filter is measured against the empty filter holder (blank) used to zero the Spectrophotometer.

Note:

UUC* : Unit Under Calibration

ISSUE: 5 REV:4

FM-CAL-33/2

 15/05/61



BECTHAI BANGKOK EQUIPMENT & CHEMICAL CO., LTD.
CALIBRATION LABORATORY

300 Phaholyothin Road, Phayathai, Bangkok 10400, Thailand Tel: +66 2615-2929 Fax: +66 2615-2350-1
E-mail: bkk@becthai.com Website: www.becthai.com



Certificate No. : CAL-23-754

Page : 4 of 4

CALIBRATION REPORT

Photometric Calibration for UV

Wavelength (nm)	Certified Values of Reference Material (A)	UUC* Reading (A)	Error (A)	Uncertainty of Measurement (\pm A)
235.0	Zero	0.000	0.0000	0.0050
	0.7385	0.735	-0.0035	0.0076
257.0	Zero	0.000	0.0000	0.0050
	0.8556	0.851	-0.0046	0.0077
313.0	Zero	0.000	0.0000	0.0050
	0.2882	0.288	-0.0002	0.0059
350.0	Zero	0.000	0.0000	0.0050
	0.6346	0.631	-0.0036	0.0069

Remark : The Potassium Dichromate Filled cells are measured against a Perchloric acid blank.

Note:

UUC* : Unit Under Calibration

- End of Report -

ISSUE: 5 REV:4

FM-CAL-33/2

15/05/61