

ภาคผนวก จ

เอกสารสอบเทียบเครื่องมือ





Certificate of Calibration

Cert.No.: 23CHO644
Page: 1 of 2

Equipment : pH Meter
Manufacturer : Horiba
Model : LAQUA-PH1300
Serial No. : B06D0012
ID No. : Ins-LAB-028
Condition As-Received: Used Item
Received Date : 31 October 2023
Calibration Date : 01 November 2023
Reference : 2310-0843OC-7
Submitted by : Thai Environmental Technic Limited
1/8 Soi Ramkhamhaeng 145,
Khwaeng/Khet Saphan Sung,
Bangkok 10240
Calibration Place : Laboratory (Thai Environment Technic Limited)
Ambient Temperature : (25.4 - 24.2) °C
Relative Humidity : (69.3 - 66.7) %
Calibration Procedure : In-house method :
- CP-0CH2 by direct measurement with standard
voltage calibrator and direct measurement
with certified reference material (CRM)

Calibrated by : KNT Bullenatrasapachai

Approved by : 
Approved Signatory

(✓) Sathip Meangmai
() Warakorn Lemgagatrakul
() Ponpan Paipim

Issue Date : 10 November 2023

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

Condition of this calibration result

1. Reference Standard Instrument :-

Instrument	Serial No.	ID No.	Cert. No.	Due Date
1) Document Process Calibrator	43160095	130RC082	23E1284	10 Apr 2024
2) Digital Thermometer	-	130RC018	23T1595	13 Sep 2024

This certification is traceable to the International System of Unit maintained through:-
- Technology Promotion Association (Thailand - Japan)

2. Certified Reference Materials : The measurement results are traceable to SI through CPA chem Ltd.,
ANSI-ASQ National Accreditation Board, Accredited No. AR-1835

Buffer Solution	Manufacturer	Lot No.	Exp. date
pH 1.679	CPA chem	823319	20 Jun 2024
pH 4.008	CPA chem	931958	01 Oct 2025
pH 6.865	CPA chem	788996	01 Jan 2024
pH 9.181	CPA chem	931960	01 Oct 2024
*pH 12.45	Hach Lange GmbH	C02902	19 Nov 2023

3. This certificate is valid only to the item calibrated on date and place of calibration.

Calibration Results

Function : mV Measurement

Performing standard curve by Fluke at pH (1.7,4,7,10)

Unit Under Calibration	Nominal Value	Standard Voltage Input	Actual Reading		Uncertainty of Measurement (±mV)	Coverage factor k
	pH	mV	mV	pH		
pH Meter S/N: B06D0012	1.680	314.73	314.7	1.680	0.058	2.00
	4.000	177.48	177.4	4.000	0.058	2.00
	6.860	8.28	8.3	6.860	0.058	2.00
	7.000	0.00	0.0	7.000	0.058	2.00
	9.180	-128.97	-129.0	9.180	0.058	2.00
	10.000	-177.48	-177.4	10.000	0.058	2.00

Function : pH Measurement

Performing four buffers standard curve by using buffer nominal pH (1.7,4,7,8)

Unit Under Calibration	Standard pH Buffer Solution	Actual pH Reading	Actual mV Reading (mV)	Uncertainty of pH measurement (±)	Coverage factor k
pH Electrode S/N: SX3D0537	1.679	1.686	296.3	0.0071	2.13
	4.008	3.992	159.1	0.0099	2.25
	6.865	6.845	-10.1	0.015	2.20
	9.181	9.138	-143.9	0.014	2.00
	*12.45	12.427	-335.9	0.056	2.00

Remark : * : Not NSC-ONSC Accredited

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-000-

A 0060438

a 1188741



Certificate of Calibration

Cert.No.: 24CHO222
Page: 1 of 3

Equipment : Spectrophotometer
Manufacturer : Labtech
Model : Blue Star A
Serial No. : 1606UV1507
ID No. : Ins-LAB-004
Condition As-Received: Used Item
Received Date : 09 April 2024
Calibration Date : 09 April 2024
Reference : 2404-0113OC-2
Submitted by : Thai Environmental Technic Limited
1/8 Soi Ramkhamhaeng 145,
Khwaeng/Khet Saphan Sung,
Bangkok 10240
Calibration Place : Laboratory (Thai Environment Technic Limited)
Ambient Temperature : (29.2 - 31.4) °C (On-Site)
Relative Humidity : (45.2 - 40.3) % (On-Site)
Calibration Procedure : In-house method :
CP-0CH4 based on ASTM E 275-01

Calibrated by : Sathip Meangmai

Approved by : 
Approved Signatory

() Umpophol Harschai
(✓) Ponpan Paipim
() Sathip Meangmai

Issue Date : 17 April 2024

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.

Condition of calibration result

1. Reference Standard Material :

Material	Serial No.	Certificate No.	Due date
1. Absorbance Standard set	42527	116228	08 Nov 2025
2. Wavelength Standard set	29829	114509	11 Sep 2025
3. Wavelength Standard set	29829	114510	11 Sep 2025
4. Stray Light Standard set	14004	108964	01 Feb 2025

2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certificate is traceable to the International System of Unit maintained through :

- Starna Scientific Ltd.
Spectral Bandwidth : 2 nm
Scan Speed : Slow

Calibration Results : without adjustment

Wavelength Accuracy

Certified Values of Reference Material (nm)	UUC Reading (nm)	Uncertainty of Measurement (± nm)	Coverage Factor k
361.00	360.6	0.16	2.00
472.47	471.6	0.16	2.00
536.66	536.2	0.16	2.00
748.48	748.4	0.16	2.00
879.27	879.0	0.16	2.00



Cert. No. : 24CHO222
Page : 3 of 3

Calibration Results : without adjustment

Photometric Accuracy

Wavelength (nm)	Certified Values of Reference Material (Abs)	UUC Reading (Abs)	Uncertainty of Measurement (\pm Abs)	Coverage Factor k
420.0	Zero	0.0002	0.0028	2.00
	0.5739	0.5722	0.0028	2.00
	0.7085	0.7074	0.0030	2.00
	1.0109	1.0146	0.0028	2.00
546.1	Zero	-0.0001	0.0028	2.00
	0.5214	0.5211	0.0028	2.00
	0.6935	0.6926	0.0030	2.00
	0.9978	0.9960	0.0028	2.00
635.0	Zero	0.0000	0.0028	2.00
	0.5626	0.5623	0.0028	2.00
	0.7577	0.7570	0.0030	2.00
	1.0946	1.0927	0.0028	2.00

Stray Light

* Straylight at 260.49 nm \pm 0.11 nm	Reading at 260.49 nm \pm 0.11 nm
Abs	2.2284
%T	0.57

Remark

- Each individual filter is measured against the empty filter holder (blank) used to zero the spectrophotometer.
- Cut-off wavelength of stray light reference material (Potassium Iodide) at Wavelength
- Result = Pass, If Absorbance > 2.00 Abs and Transmission < 1.0 %T at Wavelength
- * : Not NSC-ONS Accredited

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

-o0o-



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL.0-2717-3000-29 FAX.0-2719-9484



Certificate of Calibration

Cert.No.: 24MM272
Page: 1 of 3

Equipment : Electronic Balance
Manufacturer : Mettler Toledo
Model : AB204
Serial No. : 1116392227
ID No. : Ins-LAB-033
Submitted by : Thai Environmental Technic Limited
1/6 Soi Ramkhamhaeng 145,
Khwaeng/Khet Saphan Sung,
Bangkok 10240
Location : Balance Room
Received order : 09 April 2024
Calibration Date : 10 April 2024
Ambient Temperature : 15 °C to 40 °C
Relative Humidity : 30 % to 90 %
Calibrated by : Khiti Ruttanaprapachai
Approved by :
() Ponpan Paipim
() Suwit Imjai
(✓) Kunchit Promprat

Issue Date : 12 April 2024

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.



Equipment : Electronic Balance
Condition As-Received : Used Item
Reference : 2404-0113OC-14

Cert.No.: 24MM272
Page: 2 of 3

Procedure used :-

Calibration were conducted using in-house calibration procedure CP-OB01 based on UKAS LAB 14 according to direct measurement method against standard weight.

Condition of this result of calibration

1. Reference standard instruments:-

Instruments	Model	Serial No.	ID No.	Test report No.	Due date
1) Standard Weight Set (E2)	15884	-	70RC138	MM-0020-23	30 Jan 2025

- This certificate is valid only to the item calibrated on date and place of calibration.
- This result of calibration was made on requested at the point specified by customer.
- This certificate is not certified for any commercial transaction.
- This certification is traceable to the International System of Unit.

Result of calibration () Without Adjustment (*) After Adjustment by External Calibration

Range capacity : 0 g to 210 g Resolution 0.0001 g

Before Adjustment :

Applied Weight (g)	Balance Reading (g)	Correction (g)	Measurement Uncertainty (\pm mg)	Coverage Factor (k)
100	100.0000	0.0000	0.19	2
200	200.0001	-0.0001	0.30	2

After Adjustment :

1. Determination of the standard deviation of weighing machine (n = 10)

Applied Weight (g)	Standard Deviation of Reading (g)
100	0.00007
200	0.00008



Equipment : Electronic Balance
Condition As-Received : Used Item
Reference : 2404-0113OC-14

Cert.No.: 24MM272
Page: 3 of 3

Result of calibration

2. Effect of off center loading

A mass of 100 g was placed to various position on the pan.
The weighing machine reading error obtained is given in the table

Position 1 (g)	Position 2 (g)	Position 3 (g)	Position 4 (g)	Position 5 (g)	Maximum difference between off-center and central loading (g)
0.0000	+0.0001	0.0000	+0.0001	+0.0003	0.0003

3. Departure from nominal value

Applied Weight (g)	Balance Reading (g)	Correction (g)	Measurement Uncertainty (\pm mg)	Coverage Factor (k)
Unload	0.0000	0.0000	0.14	2.11
0.01	0.0101	-0.0001	0.14	2.11
0.1	0.1001	-0.0001	0.14	2.11
0.5	0.5002	-0.0002	0.14	2.11
1	1.0002	-0.0002	0.14	2.11
5	5.0000	0.0000	0.14	2.11
10	10.0001	-0.0001	0.14	2.11
25	25.0000	0.0000	0.15	2.07
50	49.9999	+0.0001	0.15	2.06
100	100.0002	-0.0002	0.19	2
200	200.0002	-0.0002	0.30	2

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

-o0o-



Certificate of Calibration

Cert. No.: 24TM702
Page : 1 of 3

Equipment : BOD Incubator
Manufacturer : Accuplus
Model : i250
Serial No. : 0408-0115-0008
ID No. : Ins-LAB-046
Submitted by : Thai Environmental Technic Limited
1/8 Soi Ramkhamhaeng 145,
Khwang/Khet Saphan Sung,
Bangkok 10240
Location : Laboratory (Thai Environmental Technic Limited)
Received Order : 08 April 2024
Calibration Date : 08 April 2024
Ambient Temperature : (26 ± 10) °C
Relative Humidity : (50 ± 30) %
Calibrated by : Khit Rutanapropachai
Approved by :
() Ponpan Palpin
() Sukit Imjai
(✓) Kunchit Promprat
Issue Date : 26 April 2024

The Uncertainties are for a confidence probability of approximately 95%.

This certificate may not be reproduced other than in full, except with the prior written approval of the head of Corporate Services 3: Equipment Calibration and Testing Services.



Equipment : BOD Incubator
Condition As-Received : Used Item
Reference : 2404-0113OC-11
Procedure Used :-

Cert. No.: 24TM702
Page : 2 of 3

Calibration was conducted using calibration procedure CP-OT02 based on TLAS G-20 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector (RTD).
The temperature scale used was based on ITS-90.

Condition of this result of calibration


- Reference standard instrument:-

Instrument	Serial No.	Cert. No.	Traceable	Due Date
1) Data Acquisition	MY49001451	24UM44	TPA	17 Mar 2025
- This certificate is valid only to the item calibrated on date and place of calibration.
- This certification is traceable to the International System of Unit.

Remark : TPA : Technology Promotion Association (Thailand - Japan)

Result of Calibration : (*) Without Adjustment
Function of UUC* : Temperature Source
Fresh air setting : Not Available

Environment during calibration		
	Beginning	Finished
Temp. (°C)	24	25
REL.Humid. (%)	50	52
AC Supply (Volt)	221	220



Probe Installation Details :
a = 10 cm
b = 10 cm
c = 10 cm
Dimension of Chamber :
D = 0.48 m
W = 0.50 m
H = 1.1 m
Capacity = 0.26 m³

Position :	Ref. Std. ID No.:
1	19RTD-2/1
2	19RTD-2/2
3	19RTD-2/3
4	19RTD-2/4
5	19RTD-2/5
6	24-19RTD-2/6
7	19RTD-2/7
8	19RTD-2/8
9 (ref.)	19RTD-2/9



Equipment : BOD Incubator
Condition As-Received : Used Item
Reference : 2404-0113OC-11
Result of Calibration : (*) Without Adjustment
Function of UUC* : Temperature Source
Fresh air setting : Not Available

Cert. No.: 24TM702
Page : 3 of 3

Calibration Point (°C)	UUC* Setting (°C)	UUC* Reading (°C)	Temperature stability (± °C)	Temperature uniformity (°C)	Overall Variation (°C)	Coverage Factor k
20.0	20.0	20.0	0.30	0.27	0.77	2

Calibration Point (°C)	Measured Temperature (°C)									Uncertainty (± °C)
	1	2	3	4	5	6	7	8	9 (ref.)	
20.0	20.232	20.184	20.128	20.214	20.126	20.102	19.987	20.053	20.128	0.49

Average* : The average of 30 values in each position.

Temperature stability : One-half of the greatest maximum difference of measured temperature at any one sensor.
Temperature uniformity : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

Overall Variation : The Difference of the maximum and minimum measured temperatures throughout observation.

UUC* : Unit Under Calibration

Note : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95 %.

-o0o-



Thai Environmental Technic Limited
บริษัท เทคโนโลยีสิ่งแวดล้อมไทย จำกัด

High Volume TSP&PM-10 Calibration Report

Location : Thai Environmental Technic Site ID : Bangkok Date : 3-Jul-24
ITEM : TSP Serial No : (No. 29) Calibrate By : Pipet

Site Conditions

Barometric Pressure (mm Hg) : 760.00 Corrected Pressure (mm Hg) : 760.0
Temperature (°C) : 25.0 Temperature (deg K) : 298.0
Average Press. (mm Hg) : 754.4 Corrected Average (mm Hg) :
Average Temp (°C) : 31.5 Average Temp (Deg K) : :

Calibration Orifice

Make : Tisch Qstd Slope : 1.99045
Model : TS-5025A Qstd Intercept : -0.00789
Serial# : 0068 Calibration Due Date : 15-Aug-24

Calibration Information

Plate or Test #	ORIFICE (in H ₂ O)	Qstd (m³/min)	Indicate (CFM)	IC (corrected)	Linear Regression Slope : 29.7752 Intercept : 5.4130 Corr. Coeff : 0.9918 # of Observations : 5
1	12.30	1.766	60.0	57.00	
2	9.80	1.577	54.0	52.00	
3	7.40	1.371	50.0	48.00	
4	5.00	1.127	40.0	40.00	
5	3.00	0.874	30.0	30.00	

Calculations

$$Qstd = 1/m[\sqrt{P(H_2O/Pstd)(Tstd/Ta)} - b]$$
$$IC = m[\sqrt{P(H_2O/Pstd)(Tstd/Ta)}]$$

Qstd = standard flow rate
IC = corrected chart response
I = actual chart response

m = calibrator Qstd slope
b = calibrator Qstd intercept
Ta = actual temperature during calibration (deg K)
Pa = actual pressure during calibration (mm Hg)
Tstd = 298 deg K
Pstd = 760 mm Hg

For subsequent calculation of sampler flow:
 $1/m[\sqrt{P(298/Ta)(Pstd/Pa)} - b]$

NOTE: Ensure calibration orifice has been certified within 12 months of use

m = sampler slope
b = sampler intercept
I = chart response
Ta = daily average temperature
Pa = daily average pressure

Calibrate By :

Approve By :



High Volume TSP&PM-10 Calibration Report

Location: Thai Environmental Technic Site ID: Bangkok Date: 1-Jul-24
ITEM: TSP Serial No.: (No. 30) Calibrate By: Pipat
Site Conditions
Barometric Pressure (mm Hg): 760.00 Corrected Pressure (mm Hg): 760.0
Temperature (°C): 25.0 Temperature (deg K): 298.0
Average Press. (mm Hg): 754.4 Corrected Average (mm Hg):
Average Temp (°C): 29.8 Average Temp (deg K):

Calibration Office

Make: Tisch Qstd Slope: 1.99045
Model: TE-5025A Qstd Intercept: -0.00789
Serial#: 0068 Calibration Due Date: 15-Aug-24

Calibration Information

Plate or Test #	ORIFICE (in H ₂ O)	Qstd (m ³ /min)	Indicate (CFM)	IC (corrected)	Linear Regression Slope: 29.6691 Intercept: 5.6700 Corr. Coeff: 0.9993 # of Observations: 5
1	12.40	1.787	60.0	57.00	
2	9.30	1.536	54.0	52.00	
3	7.40	1.371	50.0	48.00	
4	5.00	1.127	40.0	40.00	
5	3.00	0.876	30.0	30.00	

Calculations

Qstd = $1/m[\sqrt{P(P+P_{std})}/(T_{std}/T_a)] - b$
IC = $1/m[\sqrt{P(P+P_{std})}/(T_{std}/T_a)]$

Qstd = standard flow rate
IC = corrected chart response
I = actual chart response

m = calibrator Qstd slope
b = calibrator Qstd intercept
Ta = actual temperature during calibration (deg K)
Pa = actual pressure during calibration (mm Hg)
Tstd = 298 deg K
Pstd = 760 mm Hg

For subsequent calculation of sampler flow:
 $1/m[I][\sqrt{P(P+P_{std})}/(T_{std}/T_a)] - b$

NOTE: Ensure calibration office has been certified within 12 months of use

m = sampler slope
b = sampler intercept
I = chart response
Tav = daily average temperature
Pav = daily average pressure

Calibrate By: [REDACTED]

Approve By: [REDACTED]



Analyzer Calibration Report


Calibrate Date: 23-Apr-24 Temperature (°C): 25°C
Analyzer Type: SO₂ Barometer (mmHg): 760.0
Brand: Teledyne Humidity (50±15 %): 50.0 %RH
Model: TML-50 Dilutor: API M700 S/N 625
Serial Number: S02870 (No. 19) Zero Air: API M701 S/N 1926
Range: 500 ppb Standard gas: D636157

Calibration of Span

Supply Gas	Ref Value(ppb)	Before of Span(ppb)	After of Span(ppb)	Abs% diff of Span
Zero	0.0	2.1	0.0	0.0
Span	400.0	398.0	400.0	0.0

Multi Point Calibration

Ref Value(ppb)	Analyzer Disp.(ppb)	Output Difference		
		Diff (ppb)	Percent Diff	Abs Percent Diff
0.0	0.2	0.2	0.00	0.05
100.0	101.1	1.1	0.01	1.10
200.0	199.8	-0.2	0.00	0.10
400.0	400.5	0.5	0.00	0.13
Average Diff (%)				0.34



Calibrate by: [REDACTED]

Approved by: [REDACTED]

Certificate of Analysis Special Gases Mixture

Customer Details
Name: Thai Environmental Technic Limited Address: 1/6 Soi Ramkhamhaeng 45, Saphan Sung, Khet Saphan Sung, Bangkok 10240 Customer Tag No.:
Certificate Details
Number: 2500/23 Date of issue: 18-Sep-2023 Expiry date: 18-Sep-2027
Material Details
Production Order: 90179846 Material Code: 608400-SK-44 Cylinder No.: D636157
Gas content: 5.520 M² Filling pressure: 145 bar Valve: CGA 660 SS
Cylinder Owner: LINDE Cylinder Material: Spectra seal Cylinder Size: 40 L

Component	Nominal Concentration	Analytical Result	Uncertainty ²	Method of Analysis ¹	Assay Date
Sulphur Dioxide in Nitrogen	40.0 ppm	41.1 ppm	± 1 % relative	(6) I-PB-352	8-Sep-18-Sep-23

Reference Standard used in Assay
Cylinder number: BOC15062956 Concentration: 25.35 ± 0.25 ppm Expiry date: 9-Jun-2024
Sulphur Dioxide in Nitrogen

Analytical Instruments used in Assay
Instrument/Make/Model: FTIR Spectrometers Nicolet 650 Analytical Principle: FTIR-SO2 Last Multipoint Calibration: 6-Sep-2023

Recommend usage condition

Minimum utilization: 5% of actual content or before expiry date whichever comes first.
Storage condition: Keep in well ventilation and secure area.

Comments

When reordering, please quote the material number

Notes:

- All results expressed in this report are on a 'best estimate' basis, unless otherwise specified. The assay of this Standard has been performed in accordance with the EPA Traceability Protocol EPA-821-R-12-011 for the Assay and Certification of Gaseous Calibration Standards using procedure 6.1.
- The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The measurement of this material is traceable to the SI through the reference gas standard which is traceable to Swiss National Standard of Mass or other recognized national metrology institutes.
- (1) Gas Chromatography, (2) Paramagnetic Oxygen Analyser, (3) Electrochemical Oxygen Analyser, (4) Electrochemical Moisture Analyser, (5) Total Hydrocarbon Analyser, (6) Other - Specified



Analyzer Calibration Report


Calibrate Date: 23-Apr-24 Temperature (°C): 25°C
Analyzer Type: SO₂ Barometer (mmHg): 760.0
Brand: Thermo Humidity (50±15 %): 50.0 %RH
Model: 43C Dilutor: API M700 S/N 625
Serial Number: 43C57277312 (No. 14) Zero Air: API M701 S/N 1926
Range: 500 ppb Standard gas: D636157

Calibration of Span

Supply Gas	Ref Value(ppb)	Before of Span(ppb)	After of Span(ppb)	Abs% diff of Span
Zero	0.0	2.1	0.0	0.0
Span	400.0	400.0	400.0	0.0

Multi Point Calibration

Ref Value(ppb)	Analyzer Disp.(ppb)	Output Difference		
		Diff (ppb)	Percent Diff	Abs Percent Diff
0.0	0.4	0.4	0.00	0.10
100.0	99.1	-0.9	-0.01	0.90
200.0	199.2	-0.8	0.00	0.40
400.0	399.3	-0.7	0.00	0.17
Average Diff (%)				0.39



Calibrate by: [REDACTED]

Approved by: [REDACTED]

Certificate of Analysis
Special Gases Mixture

Customer Details
Name: Thai Environmental Technic Limited
Address: 1/6 Soi Ramkhamhaeng 45, Saphan Song, Khet Saphan Song, Bangkok 10240
Customer Tag No.:

Certificate Details
Number: 1734/23
Date of Issue: 5-Jul-2023
Expiry date: 5-Jul-2026
Material Details
Production Order: 90178560
Material Code: 640300-SK-44
Cylinder No.: A00917SK
Gas content: 5.520 M³
Filling pressure: 145.0 bar
Valve: CGA 660 SS
Cylinder Owner: LINDE
Cylinder Material: Spectra seal
Cylinder Size: 40 L

Laboratory Report
Component: Nitric Oxide
Normal Concentration: 40.0 ppm
Analysis Result¹: 40.5 ppm
Uncertainty²: ± 1% relative
Method of Analysis³: (6) I-PB-352
Assay Date: 28-Jun & 5-Jul-2023
Other NO_x impurity in Nitrogen: Less than 2.0 ppm

Reference Standard
Nitric Oxide in Nitrogen
Reference Standard used in Assay
Cylinder number: 2580135G
Concentration: 25.32 ± 0.25 ppm
Expiry date: 13-Dec-2024

Analytical Instruments used in Assay
Instrument/Make/Model: FTIR Spectrometers Nicolet i550
Analytical Principle: FTIR-NO
Last Multipoint Calibration: 28-Jun-2023

Recommend usage condition

Minimum utilization: 5% of actual content or before expiry date whichever comes first.
Storage condition: Keep in well ventilation and secure area.

Comments
When reordering, please quote the material number.

Note:

1. All results expressed in this report are on mole/mole basis, unless otherwise specified.
2. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%.
3. (1) Gas Chromatography, (2) Paramagnetic Oxygen Analyser, (3) Electrochemical Oxygen Analyser, (4) Electrochemical Nitrogen Analyser, (5) Total Hydrocarbon Analyser, (6) Other - Specified.

Page 1 of 1

This report shall not be reproduced except in full.

ฉบับนี้ (ฉบับนี้) (ฉบับนี้) (ฉบับนี้)

No. 15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

NO_x Analyzer Calibration Report

Calibrate Date: 22-Apr-24
Analyzer Type: NO_x
Brand: API
Model: 200 S
Serial Number: 737 (No. 27)
Range: 500 ppb


Temperature (°C): 25°C
Barometer (mmHg): 760.0
Humidity (50±15 %): 50.0%RH
Dilutor: API M700 S/N 625
Zero Air: API M701 S/N 1926
Standard gas: A00917 SK

Calibration of Span

Supply Gas	Ref Value(ppb)	Before of Span(ppb)			After of Span(ppb)			% diff of Span
		NO _x	NO	NO ₂	NO _x	NO	NO ₂	
Zero	0.0	0.9	0.8	0.1	0.0	0.0	0.0	0.0
Span	400.0	388.0	389.0	-1.0	400.0	400.0	0.0	0.0

Multi Point Calibration

Ref Value(ppb)	Analyzer Disp(ppb)			Output Difference		
	NO _x	NO	NO ₂	Diff(ppb)	% Diff	Abs (%) Diff
0.0	0.3	0.2	0.2	0.20	0.001	0.05
100.0	99.5	99.1	0.4	-0.90	-0.009	0.90
200.0	199.7	199.3	0.4	-0.70	-0.003	0.35
400.0	401.0	400.5	0.5	0.50	0.001	0.13
Average Diff (%)						0.46



Calibrate by:

Approved by:

แก้ไขครั้งที่: 00

วันที่อนุมัติ: 02/09/15

เลขที่แบบฟอร์ม: QP-015-05

Thai Environmental Technic Limited 1/6 Soi Ramkhamhaeng 45, Saphan Song, Bangkok 10240 Thailand
Tel: +66(0)2373-7799(Auto) Fax: +66(0)2373-7979 E-mail: info@tet1995.com www.tet1995.com

NO_x Analyzer Calibration Report

Calibrate Date: 19-Apr-24
Analyzer Type: NO_x
Brand: Teledyne
Model: T200
Serial Number: 5154 (No. 30)
Range: 500 ppb


Temperature (°C): 25°C
Barometer (mmHg): 760.0
Humidity (50±15 %): 50.0%RH
Dilutor: API M700 S/N 625
Zero Air: API M701 S/N 1926
Standard gas: A00917 SK

Calibration of Span

Supply Gas	Ref Value(ppb)	Before of Span(ppb)			After of Span(ppb)			% diff of Span
		NO _x	NO	NO ₂	NO _x	NO	NO ₂	
Zero	0.0	0.5	0.8	-0.3	0.0	0.0	0.0	0.0
Span	400.0	394.0	392.0	2.0	400.0	400.0	0.0	0.0

Multi Point Calibration

Ref Value(ppb)	Analyzer Disp(ppb)			Output Difference		
	NO _x	NO	NO ₂	Diff(ppb)	% Diff	Abs (%) Diff
0.0	0.3	0.1	0.2	0.000	0.000	0.03
100.0	99.7	100.1	-0.4	0.10	0.001	0.10
200.0	199.7	199.1	0.6	-0.90	-0.005	0.45
400.0	401.2	400.5	0.7	0.50	0.001	0.13
Average Diff (%)						0.22



Calibrate by:

Approved by:

แก้ไขครั้งที่: 00

วันที่อนุมัติ: 02/09/15

เลขที่แบบฟอร์ม: QP-015-06

Thai Environmental Technic Limited 1/6 Soi Ramkhamhaeng 45, Saphan Song, Bangkok 10240 Thailand
Tel: +66(0)2373-7799(Auto) Fax: +66(0)2373-7979 E-mail: info@tet1995.com www.tet1995.com

ฉบับนี้ (ฉบับนี้) (ฉบับนี้) (ฉบับนี้)

No. 15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

Linde (Thailand) Public Company Limited

No. 15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

ฉบับนี้ (ฉบับนี้) 02/15 (ฉบับนี้) 02/15 (ฉบับนี้) 02/15

CERTIFICATE OF ANALYSIS

Analytical Result

Component	Request Concentration	Certified Concentration	Certified Uncertainty	Method	Assay Date
Carbon Monoxide	800 ppm	793 ppm	± 1 % relative	(6) I-PB-352	04-Oct & 11-Oct-2016
Sulphur Dioxide	800 ppm	804 ppm	± 1 % relative	(6) I-PB-352	04-Oct & 11-Oct-2016
In Nitrogen					

Reference Standard used in Assay

Reference Standard	Cylinder No.	Concentration	Expired Date
Sulphur Dioxide	118499SG	504.5 ± 2.5 ppm	02-Jul-2018
Carbon Monoxide	113882SG	504.3 ± 1.0 ppm	28-Apr-2019
In Nitrogen			

Analytical Instruments used in Assay

Instrument/Make/Model	Analytical Principle	Last Multistep Calibration
Digi LAB Excalibur IIE Series	FIR-002	16-Sep-2016
Digi LAB Excalibur IIE Series	FIR-410	31-Oct-2016

Method of Analysis:
 1. Gas Chromatography
 2. Paramagnetic Oxygen Analyzer
 3. Electrochemical Oxygen Analyzer
 4. Electrochemical Moisture Analyzer
 5. Total Hydrocarbon Analyzer
 6. Other specified

Cylinder Number: D271305
 Production Order Number: 90137638

Certification Date: 11-Oct-2016
 Expiration Date: 11-Oct-2024

Page 2 of 2

บริษัท ลินด์ (ประเทศไทย) จำกัด (มหาชน)
 15 อาคารทรีทรี 2/3 หมู่ 14 ถนนพหลโยธิน กม. 6.5 แขวง
 บางพลีใหญ่ เขตภาษีเจริญ กรุงเทพฯ 10540 โทรศัพท์ (66) 2338-6100 โทรสาร (66) 2338-6333
 โทรสารแฟกซ์: 105 หมู่ 5 แขวงบางนา เขตภาษีเจริญ กรุงเทพฯ 10150
 โทรศัพท์ (66) 38.579-479-93 โทรสาร (66) 38.579-323

Linde (Thailand) Public Company Limited
 15 อาคารทรีทรี 2/3 หมู่ 14 ถนนพหลโยธิน กม. 6.5 แขวง
 บางพลีใหญ่ เขตภาษีเจริญ กรุงเทพฯ 10540 โทรศัพท์ (66) 2338-6100 โทรสาร (66) 2338-6333
 โทรสารแฟกซ์: 105 หมู่ 5 แขวงบางนา เขตภาษีเจริญ กรุงเทพฯ 10150
 โทรศัพท์ (66) 38.579-479-93 โทรสาร (66) 38.579-323

CERTIFICATE OF ANALYSIS

Analytical Result

Component	Request Concentration	Certified Concentration	Certified Uncertainty	Method	Assay Date
Carbon Monoxide	400 ppm	404 ppm	± 1 % relative	(6) I-PB-352	03-Oct & 10-Oct-2016
Sulphur Dioxide	400 ppm	406 ppm	± 1 % relative	(6) I-PB-352	03-Oct & 10-Oct-2016
In Nitrogen					

Reference Standard used in Assay

Reference Standard	Cylinder No.	Concentration	Expired Date
Sulphur Dioxide	D832461	201 ± 1 ppm	12-Apr-2018
Carbon Monoxide	D832461	208 ± 1 ppm	12-Apr-2018
In Nitrogen			

Analytical Instruments used in Assay

Instrument/Make/Model	Analytical Principle	Last Multistep Calibration
Digi LAB Excalibur IIE Series	FIR-002	17-Sep & 18-Oct-2016
Digi LAB Excalibur IIE Series	FIR-CO	15-Oct-2016

Method of Analysis:
 1. Gas Chromatography
 2. Paramagnetic Oxygen Analyzer
 3. Electrochemical Oxygen Analyzer
 4. Electrochemical Moisture Analyzer
 5. Total Hydrocarbon Analyzer
 6. Other specified

Cylinder Number: D824500
 Production Order Number: 90137639

Certification Date: 11-Oct-2016
 Expiration Date: 11-Oct-2024

Page 2 of 2

บริษัท ลินด์ (ประเทศไทย) จำกัด (มหาชน)
 15 อาคารทรีทรี 2/3 หมู่ 14 ถนนพหลโยธิน กม. 6.5 แขวง
 บางพลีใหญ่ เขตภาษีเจริญ กรุงเทพฯ 10540 โทรศัพท์ (66) 2338-6100 โทรสาร (66) 2338-6333
 โทรสารแฟกซ์: 105 หมู่ 5 แขวงบางนา เขตภาษีเจริญ กรุงเทพฯ 10150
 โทรศัพท์ (66) 38.579-479-93 โทรสาร (66) 38.579-323

Linde (Thailand) Public Company Limited
 15 อาคารทรีทรี 2/3 หมู่ 14 ถนนพหลโยธิน กม. 6.5 แขวง
 บางพลีใหญ่ เขตภาษีเจริญ กรุงเทพฯ 10540 โทรศัพท์ (66) 2338-6100 โทรสาร (66) 2338-6333
 โทรสารแฟกซ์: 105 หมู่ 5 แขวงบางนา เขตภาษีเจริญ กรุงเทพฯ 10150
 โทรศัพท์ (66) 38.579-479-93 โทรสาร (66) 38.579-323

CERTIFICATE OF ANALYSIS

Customer Detail:
 Thai Environmental Technic Ltd.

Production Order Number: 90137639
 Material Number: 498800-AL-44
 Certification Date: 11-Oct-2016
 Expiry Date: 11-Oct-2024

Cylinder Description:
 ALU 50 L

The measurement of this reference material is traceable to SI through the reference standard which is traceable to the National Standard of Mass. The Assay of this Standard has been performed in accordance with the EPA Traceability Protocol EPA-800-B-12-031 for the Assay and Certification of Gaseous Calibration Standards using procedure G1. The results are expressed on a mole/mole basis, unless otherwise specified. The reported uncertainty is based on a standard uncertainty multiplied by coverage factor k=2, providing a level of confidence of approximately 95%.

Certificate Number:
 3111/16

Analyst:

Cylinder Number:
 D824500

Nominal Cylinder Content:
 6.900 M³

Approver:

Nominal Pressure:
 145.0 Bar

Valve Outlet:
 CGA 600 SS

To Re-Order Please Quote:
 498800-AL-44

Comment:

- It is recommended that this product be not used below 5% of actual content or should not be used when its gas pressure is below 150psig.
- Other impurities that detect by analytical condition of this mixture shall be report if it is more than 10% of minimum minor component.
- Keep and use in well-ventilated and secure area.

Page 1 of 2

บริษัท ลินด์ (ประเทศไทย) จำกัด (มหาชน)
 15 อาคารทรีทรี 2/3 หมู่ 14 ถนนพหลโยธิน กม. 6.5 แขวง
 บางพลีใหญ่ เขตภาษีเจริญ กรุงเทพฯ 10540 โทรศัพท์ (66) 2338-6100 โทรสาร (66) 2338-6333
 โทรสารแฟกซ์: 105 หมู่ 5 แขวงบางนา เขตภาษีเจริญ กรุงเทพฯ 10150
 โทรศัพท์ (66) 38.579-479-93 โทรสาร (66) 38.579-323

Linde (Thailand) Public Company Limited
 15 อาคารทรีทรี 2/3 หมู่ 14 ถนนพหลโยธิน กม. 6.5 แขวง
 บางพลีใหญ่ เขตภาษีเจริญ กรุงเทพฯ 10540 โทรศัพท์ (66) 2338-6100 โทรสาร (66) 2338-6333
 โทรสารแฟกซ์: 105 หมู่ 5 แขวงบางนา เขตภาษีเจริญ กรุงเทพฯ 10150
 โทรศัพท์ (66) 38.579-479-93 โทรสาร (66) 38.579-323



Thai Environmental Technic Limited
 บริษัท เทคนิคสิ่งแวดล้อมไทย จำกัด

Analyzer Calibration Report


Calibrate Date	19-Apr-24	Temperature (°C)	26°C
Analyzer Type	CO	Barometer (mmHg)	760
Brand	Thermo	Humidity (50±15 %)	50.0
Model	42C	Dilutor	API M700 S/N625
Serial Number	48062-846337 (No. 3)	Zero Air	API M701 S/N1926
Range	1000 ppm	Standard gas	D824500, D271305

Calibration of Span

Supply Gas	Ref Value(ppm)	Before of Span (ppm)	After of Span(ppm)	Abs% diff of Span
Zero	0.0	0.6	0.0	0.00
Span	793.0	792.2	793.0	0.00

Multi Point Calibration

Ref Value(ppm)	Analyzer Disp (ppm)	Diff (ppm)	Percent Diff	Abs Percent Diff
0.0	0.4	0.4	0.00	0.05
404.0	404.8	0.8	0.00	0.20
793.0	793.2	0.2	0.00	0.03
Average Diff (%)				0.09



Calibrate by:

Approved by:

วันที่จัดทำ: 00

วันที่อนุมัติ: 02/09/15

หมายเลขใบฟอร์ม: QF-QP16-06

Gas Detector Certificate

Issued By: Sithiporn Associates Company Limited
Owner Name: Thai Environmental Technic Limited
Reference Number: SVC027504
Product Brand: BW Technologies
Type Systematic: Portable
Battery Type: 3 V Lithium Energizer (CR2) 1 pcs.

Calibration Date: 29-Jan-24
Calibration Due: 28-Jan-25
Calibration Temp.: 25°C
Calibration Humidity: 65%RH

Product Model: Gas Alert Extreme CO
Sensor Model: SR-M04-SC
Detection Gases: Measuring Range

Serial Number: J017-M02B499
Visual Gas Alarm: Red Light
Detection Principles: Electrochemical cell

Carbon monoxide (CO): 0-1000 ppm in 1 ppm increments

Calibration Standard equipment: Std Gas Mixtures Cylinder Number 955-643209 Expired Date 04-Oct-25

Components Concentration
Methane (CH₄): 50.0 %LEL (2.5 %vol.)
Hydrogen Sulfide (H₂S): 25.0 ppm
Carbon Monoxide (CO): 100.0 ppm
Oxygen (O₂): 18.0 %Vol.

Calibration Result

Item Calibration	Fresh Air			Standard Gas			Standard Drift	Unit	%90/T (second)
	Before	Calibration	After	Before	Calibration	After			
Carbon monoxide (CO)	0	0	0	96	100	100	0	ppm	18

*90 %T is response time reading to 90% of standard gas

Gas Alarm Preset: Low Alarm High Alarm TWA STEL Unit

Carbon monoxide (CO): 35 200 35 200 ppm

Operation Test

Function	Battery Indicate	Sampling Module	Self-Test	Gas Display	Alarm Report	PC/Data Collection	Alarm Functional		
Judgement	Pass	N/A	Pass	Pass	Pass	N/A	Audible	Visual	Vibration

Remarks: บริษัท สิทธีพรเอเซีย จำกัด
SITHIPORN ASSOCIATES COMPANY LIMITED

Signature

Approved

451-451/1 มานูฟาส ออโตมัติคัล แอวาทิส ออแกไนซ์ 10700
451-451/1 Sithiporn Road, Bangbunru, Bangkok 10700, Thailand

Email: center@sithiporn.com
Tel: +66 2433 8331 www.sithiporn.com

BANGKOK INDUSTRIAL GAS CO., LTD.
11th Floor, Rajanikam Building
3 South Sathorn Rd, Yanawa, Sathorn
Bangkok 10120, Thailand
Tel: (662) 685-6789 Fax: (662) 685-6780-1

CERTIFICATE OF CONFORMITY
(For Package Gases)

Customer Name: Sithiporn Associates Co., Ltd.
Product Name: Nitrogen Certificate No.: OC1555-4298
Date of Issue: 12 DEC '21 Gas Content: 17 M³
Lot No.: 171215/201/DO3300016058 Shelf Life: 36 months
Page no.: 1/1 Cylinder Valve Type: CGA 580

Components	Specification
Oxygen	< 2 ppm
Moisture	< 3 ppm
Carbon Dioxide	< 1 ppm
Carbon Monoxide	< 1 ppm
Total Hydrocarbon as CH ₄	< 1 ppm
Nitrogen	> 99.999 %

Cylinder Number

M5281014 13D007140

บริษัท สิทธีพรเอเซีย จำกัด
SITHIPORN ASSOCIATES COMPANY LIMITED

Certificate Of Composition WO420382 - 5

Part Code: C006552
10ALQUAD-003-LUX

Customer: Calgaz International LLC
Customer Order Number: PO23789/8033613 SEI001

Cylinder No: 955-643239
Cylinder Valve: C10
Gross Weight: 1.5
Net Weight: 0.12

Component	Requested Values	Certified Values	Accuracy
Carbon Monoxide	100 ppm	100 ppm	± 2%
Oxygen	18 %	18 %	± 2%
Methane	2.5 %	2.5 %	± 2%
Hydrogen Sulphide	25 ppm	25 ppm	± 5%
Nitrogen	Balance	Balance	

Pressure: 1000 psi Volume: 1.6 ltr Size: 10AL

Please note all units are in mol% and methods used in analysis per VME
include: Chromatography, Paramagnetic, NDIR, UV-VIS, GC (FID/CO),
Electrochemical cells, Zirconia, PID and Gravimetric. Product composition is
verified by direct composition to calibration standards traceable to NPL, NIST
or equivalent National Standard weights or gas mixture reference materials.
Traceable reference: 55603595 / 27603229 / 27605263

Manufactured Date: 04/10/2023

Valid Until: 04/10/2025

UN 1955 Compressed gas, n.o.s.

(Oxygen, Nitrogen Mixture)

Certified By: Craig Nichols

All Gas Mixtures that are quoted with a balance of Air contain 20.9% (19.2%) Oxygen relative to the overall Gas Mixture.
The gases contained in this cylinder are not breathing gases.
DO NOT INHALE

บริษัท สิทธีพรเอเซีย จำกัด
SITHIPORN ASSOCIATES COMPANY LIMITED



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL: 0-2717-3000-29 FAX: 0-2719-9484



Certificate of Calibration

Cert.No.: 24MM273
Page.: 1 of 3

Equipment: Electronic Balance

Manufacturer: Mettler Toledo

Model: XP205DR

Serial No.: 1129273885

ID No.: Ins-LAB-035

Submitted by: Thai Environmental Technic Limited
1/6 Soi Ramkhamhaeng 145,
Khwaeng/Khet Saphan Sung,
Bangkok 10240

Location: Balance Room

Received order: 09 April 2024

Calibration Date: 10 April 2024

Ambient Temperature: 15 °C to 40 °C

Relative Humidity: 30 % to 90 %

Calibrated by: Khit Ruttanaprapachai

Approved by:

() Ponpan Palpin

() Suwit Imjai

(✓) Kunchit Prompratt

Issue Date: 12 April 2024

The Uncertainties are for a confidence probability of approximately 95%
This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3 - Equipment Calibration and Testing Services.



Equipment : Electronic Balance
 Condition As-Received : Used Item
 Reference : 2404-0113OC-15
 Procedure used :-

Cert.No.: 24MM273
 Page: 2 of 3

Calibration were conducted using in-house calibration procedure CP-0B01 based on UKAS LAB 14 according to direct measurement method against standard weight.

Condition of this result of calibration

1. Reference standard instruments:-

Instruments	Model	Serial No.	ID No.	Test report No.	Due date
1) Standard Weight Set (E2)	15884	-	70RC138	MM-0020-23	30 Jan 2025

- This certificate is valid only to the item calibrated on date and place of calibration.
- This result of calibration was made on requested at the point specified by customer.
- This certificate is not certified for any commercial transaction.
- This certificate is traceable to the International System of Unit.

Result of calibration () Without Adjustment (*) After Adjustment by Internal Calibration

Range capacity :	0 g to 81 g	Resolution	0.00001 g
	81 g to 220 g	Resolution	0.0001 g

Before Adjustment :

Applied Weight	Balance Reading	Correction	Measurement Uncertainty	Coverage Factor
(g)	(g)	(g)	(± mg)	(k)
80	79.99997	+0.00003	0.15	2
200	199.9998	+0.0002	0.29	2

After Adjustment :


1. Determination of the standard deviation of weighing machine (n = 10)

Applied Weight	Standard Deviation of Reading (g)
(g)	
80	0.000016
200	0.00005



Equipment : Electronic Balance
 Condition As-Received : Used Item
 Reference : 2404-0113OC-15

Cert.No.: 24MM273
 Page: 3 of 3



2. Effect of off center loading

A mass of 100 g was placed various position on the pan.
 The weighing machine reading error obtained is given in the table

Position 1	Position 2	Position 3	Position 4	Position 5	Maximum difference between off-center and central loading
(g)	(g)	(g)	(g)	(g)	(g)
+0.0001	+0.0001	0.0000	0.0000	+0.0002	0.0001

3. Departure from nominal value

Applied Weight	Balance Reading	Correction	Measurement Uncertainty	Coverage Factor
(g)	(g)	(g)	(± mg)	(k)
Unload	0.00000	0.00000	0.028	2.28
0.01	0.00999	+0.00001	0.029	2.28
0.05	0.04999	+0.00001	0.029	2.23
1	0.99999	+0.00001	0.030	2.17
2	1.99999	+0.00001	0.030	2.15
5	4.99999	+0.00001	0.034	2.09
10	10.00000	0.00000	0.036	2.06
20	19.99999	+0.00001	0.045	2
50	49.99999	+0.00001	0.080	2
80	79.99999	+0.00001	0.15	2
200	199.9998	+0.0002	0.29	2

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

-o0o-

ภาคผนวก ช

หนังสืออนุญาตขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน



ที่อก ๐๓๑๐(๑)/ ๕ ๘ ๗ ๖



กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท
เขตราชเทวี กรุงเทพฯ ๑๐๕๐๐

๒๒ มิถุนายน ๒๕๖๖

เรื่อง ต่ออายุหนังสือขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน


เรียน กรรมการผู้จัดการ บริษัท เทคโนโลยีสิ่งแวดล้อมไทย จำกัด

อ้างถึง คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และชนิดสารเคมีของห้องปฏิบัติการวิเคราะห์เอกชน
ลงวันที่ ๑ มีนาคม ๒๕๖๖

สิ่งที่ส่งมาด้วย เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
บริษัท เทคโนโลยีสิ่งแวดล้อมไทย จำกัด จำนวน ๒๘ แผ่น

ตามหนังสือที่อ้างถึง บริษัท เทคโนโลยีสิ่งแวดล้อมไทย จำกัด ขอต่ออายุหนังสือขึ้นทะเบียน
ห้องปฏิบัติการวิเคราะห์เอกชน เลขทะเบียน ๖-๒๓๖ สถานที่ตั้งเลขที่ ๑/๖ ซอยรามคำแหง ๑๕๕ แขวงสะพานสูง
เขตสะพานสูง กรุงเทพมหานคร ต่อกรมโรงงานอุตสาหกรรม นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว ใต้บริษัท เทคโนโลยีสิ่งแวดล้อมไทย จำกัด ต่ออายุหนังสือขึ้น
ทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน โดยมีองค์ประกอบดังนี้



๑๓) นายจิรวัฒน์...

ค. ขอบข่ายสารเคมีที่ได้รับขึ้นทะเบียนให้วิเคราะห์ในน้ำเสีย น้ำใต้ดิน อากาศเสีย สิ่งปฏิกูลหรือ
วัสดุที่ไม่ใช้แล้ว และดิน ตามสิ่งที่ส่งมาด้วย

หนังสือฉบับนี้จะหมดอายุในวันที่ ๓๑ มีนาคม ๒๕๖๘ หากประสงค์จะต่ออายุหนังสือ
รับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ให้ยื่นคำขอต่ออายุพร้อมเอกสารประกอบคำขอต่อ
กรมโรงงานอุตสาหกรรม ภายใน ๓๐ วัน ก่อนวันสิ้นสุดอายุของหนังสือขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
ทั้งนี้ สามารถยื่นคำขอผ่านระบบอิเล็กทรอนิกส์ได้ทั้งหน้าเว็บไซต์กรมโรงงานอุตสาหกรรม

จึงเรียนมาเพื่อทราบ

ผู้อำนวยการกองจัดและควบคุมมลพิษ
กรมโรงงานอุตสาหกรรม

กองวิจัยและเตือนภัยมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๕๓๐ ๖๓๒๖ ต่อ ๒๑๐๓๕

โทรสาร ๐ ๒๕๓๐ ๖๓๒๖ ต่อ ๒๑๐๔๕

ไปรษณีย์อิเล็กทรอนิกส์ sarabang@dlw.mail.go.th



"อุตสาหกรรมก้าวไกล ประเทศไทยก้าวหน้า ร่วมกันพัฒนา อุตสาหกรรมสีเขียว"



เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท เทคโนโลยีสิ่งแวดล้อมไทย จำกัด

ที่อก ๐๓๑๐(๑)/ ๕ ๘ ๗ ๖

เลขทะเบียน ๖-๒๓๖

ลงวันที่ ๒๒ มิถุนายน ๒๕๖๖

ขอขยายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๓๓๗ รายการ

น้ำเสีย จำนวน 40 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Aldrin	Liquid-Liquid Extraction, Gas Chromatographic Method ^(๑)
2	Arsenic	Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^(๑)
3	Barium	1) Digestion, Direct Nitrous Oxide-Acetylene Flame Method ^(๑) 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^(๑) 3) Digestion, Inductively Coupled Plasma Method ^(๑)
4	α-BHC	Liquid-Liquid Extraction, Gas Chromatographic Method ^(๑)
5	γ-BHC	Liquid-Liquid Extraction, Gas Chromatographic Method ^(๑)
6	Biochemical Oxygen Demand	5-Day BOD Test, Azide Modification Method ^(๑)
7	Cadmium	1) Digestion, Direct Air-Acetylene Flame Method ^(๑) 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^(๑) 3) Digestion, Inductively Coupled Plasma Method ^(๑)
8	Chemical Oxygen Demand	Closed Reflux, Titrimetric Method ^(๑)
9	Chlordane	Liquid-Liquid Extraction, Gas Chromatographic Method ^(๑)
10	Chromium	1) Digestion, Direct Air-Acetylene Flame Method ^(๑) 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^(๑) 3) Digestion, Inductively Coupled Plasma Method ^(๑)
11	Color	ADMI Weighted-Ordinate Spectrophotometric Method ^(๑)
12	Copper	1) Digestion, Direct Air-Acetylene Flame Method ^(๑) 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^(๑) 3) Digestion, Inductively Coupled Plasma Method ^(๑)
13	Cyanide	Distillation, Colorimetric Method ^(๑)
14	4,4'-DDE	Liquid-Liquid Extraction, Gas Chromatographic Method ^(๑)
15	4,4'-DDT	Liquid-Liquid Extraction, Gas Chromatographic Method ^(๑)
16	Dieldrin	Liquid-Liquid Extraction, Gas Chromatographic Method ^(๑)

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
17	Endosulfan I	Liquid-Liquid Extraction, Gas Chromatographic Method ^(๑)
18	Endosulfan II	Liquid-Liquid Extraction, Gas Chromatographic Method ^(๑)
19	Endosulfan Sulfate	Liquid-Liquid Extraction, Gas Chromatographic Method ^(๑)
20	Endrin	Liquid-Liquid Extraction, Gas Chromatographic Method ^(๑)
21	Formaldehyde	Distillation, Colorimetric Method ^(๑)
22	Free Chlorine	DPD Ferrous Titrimetric Method ^(๑)
23	Heptachlor	Liquid-Liquid Extraction, Gas Chromatographic Method ^(๑)
24	Heptachlor Epoxide	Liquid-Liquid Extraction, Gas Chromatographic Method ^(๑)
25	Hexavalent Chromium	Colorimetric Method ^(๑)
26	Lead	1) Digestion, Direct Air-Acetylene Flame Method ^(๑) 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^(๑) 3) Digestion, Inductively Coupled Plasma Method ^(๑)
27	Manganese	1) Digestion, Direct Air-Acetylene Flame Method ^(๑) 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^(๑) 3) Digestion, Inductively Coupled Plasma Method ^(๑)
28	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^(๑)
29	Nickel	1) Digestion, Direct Air-Acetylene Flame Method ^(๑) 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ^(๑) 3) Digestion, Inductively Coupled Plasma Method ^(๑)
30	Oil & Grease	1) Liquid-Liquid, Partition-Gravimetric Method ^(๑) 2) Soxhlet Extraction Method ^(๑)
31	pH	Electrometric Method ^(๑)
32	Phenols	Distillation, Direct Photometric Method ^(๑)
33	Selenium	Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^(๑)
34	Sulfide	1) Iodometric Method ^(๑) 2) Methylene Blue Method ^(๑)
35	Temperature	Laboratory and Field Methods ^(๑)
36	Total Dissolved Solids	Dried at 180 °C ^(๑)
37	Total Kjeldahl Nitrogen	Macro-Kjeldahl Method ^(๑)
38	Total Suspended Solids	Dried at 103-105 °C ^(๑)

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
39	Trivalent Chromium	Digestion, Inductively Coupled Plasma Method; Colorimetric Method; Calculation ⁽⁴⁾
40	Zinc	1) Digestion, Direct Air-Acetylene Flame Method ⁽⁴⁾ 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ⁽⁴⁾ 3) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾

นำได้ต้น จำนวน 122 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Acenaphthene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
2	Acetone	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
3	Aldrin	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
4	Anthracene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
5	Antimony	1) Digestion, Direct Air-Acetylene Flame Method ⁽⁴⁾ 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ⁽⁴⁾ 3) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾
6	Arsenic	Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ⁽⁴⁾
7	Atrazine	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
8	Barium	1) Digestion, Direct Nitrous Oxide-Acetylene Flame Method ⁽⁴⁾ 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ⁽⁴⁾ 3) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾
9	Benz(a)anthracene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
10	Benzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
11	Benzo(b)fluoranthene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
12	Benzo(k)fluoranthene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾

13 Benzoic acid...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
13	Benzoic acid	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
14	Benzo(a)pyrene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
15	Benzo(g,h,i)perylene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
16	Beryllium	1) Digestion, Electrothermal Atomic Absorption Spectrometric Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾
17	Bis(2-chloroethyl)ether	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
18	Bis(2-ethylhexyl)phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
19	Bromodichloromethane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
20	Bromoform	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
21	Butanol	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
22	Butyl benzyl phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
23	Cadmium	1) Digestion, Electrothermal Atomic Absorption Spectrometric Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾
24	Carbazole	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
25	Carbon disulfide	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
26	Carbon tetrachloride	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
27	Chlordane	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
28	p-Chloroaniline	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
29	Chlorobenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
30	Chlorodibromomethane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
31	Chloroform	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾

32 Chromium...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
32	Chromium	1) Digestion, Direct Air-Acetylene Flame Method ⁽⁴⁾ 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ⁽⁴⁾ 3) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾
33	Chromium (III)	1) Digestion, Direct Air-Acetylene Flame Method; Colorimetric Method; Calculation ⁽⁴⁾ 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method; Colorimetric Method; Calculation ⁽⁴⁾ 3) Digestion, Inductively Coupled Plasma Method; Colorimetric Method; Calculation ⁽⁴⁾
34	Chromium (VI)	Colorimetric Method ⁽⁴⁾
35	Chrysene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
36	Cyanide	Distillation, Colorimetric Method ⁽⁴⁾
37	2,4-D	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
38	DDD	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
39	DDE	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
40	DDT	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
41	Dibenz(a,h)anthracene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
42	Di-n-butyl phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
43	1,2-Dichlorobenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
44	1,3-Dichlorobenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
45	1,4-Dichlorobenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
46	1,1-Dichloroethane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
47	1,2-Dichloroethane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
48	1,1-Dichloroethylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
49	cis-1,2-Dichloroethylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾

50 trans-1,2-Dichloroethylene...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
50	trans-1,2-Dichloroethylene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
51	1,2-Dichloropropane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
52	1,3-Dichloropropane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
53	1,3-Dichloropropene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
54	Dieldrin	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
55	Diethyl phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
56	2,4-Dimethylphenol	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
57	2,4-Dinitrophenol	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
58	2,4-Dinitrotoluene	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
59	2,6-Dinitrotoluene	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
60	Di-n-Octyl phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
61	Endosulfan	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
62	Endrin	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
63	Ethylbenzene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
64	Fluoranthene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
65	Fluorene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
66	Heptachlor	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
67	Heptachlor epoxide	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
68	Hexachloro-1,3-butadiene	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
69	n-Hexane	Purge and Trap Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾
70	α-HCH	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
71	β-HCH	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
72	γ-HCH	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
73	Hexachlorocyclopentadiene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ⁽⁴⁾

74 Hexachloroethane...

ลำดับที่	สารเคมี	วิธีวิเคราะห์
74	Hexachloroethane	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
75	Indeno(1,2,3-cd)pyrene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
76	Isophorone	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
77	Lead	1) Digestion, Electrothermal Atomic Absorption Spectrometric Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾
78	Manganese	1) Digestion, Direct Air-Acetylene Flame Method ⁽⁴⁾ 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ⁽⁴⁾ 3) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾
79	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ⁽⁴⁾
80	Methanol	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
81	Methoxychlor	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
82	Methyl bromide	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
83	Methylene chloride	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
84	2-Methylphenol	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
85	2-Methylnaphthalene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
86	Methyl tert-butyl ether	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
87	Naphthalene	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
88	Nickel	1) Digestion, Electrothermal Atomic Absorption Spectrometric Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾
89	Nitrobenzene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
90	N-Nitrosodiphenylamine	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾

91 N-Nitrosodi-n-propylamine...

ลำดับที่	สารเคมี	วิธีวิเคราะห์
91	N-Nitrosodi-n-propylamine	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
92	Polychlorinated Biphenyls PCB-1016 PCB-1221 PCB-1232 PCB-1242 PCB-1248 PCB-1254 PCB-1260	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
93	Pentachlorophenol	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
94	pH	Electrometric Method ⁽⁴⁾
95	Phenanthrene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
96	Phenol	1) Distillation, Direct Photometric Method ⁽⁴⁾ 2) Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
97	Pyrene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
98	Selenium	Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ⁽⁴⁾
99	Silver	1) Digestion, Direct Air-Acetylene Flame Method ⁽⁴⁾ 2) Digestion, Electrothermal Atomic Absorption Spectrometric Method ⁽⁴⁾ 3) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾
100	Styrene	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
101	1,1,2,2-Tetrachloroethane	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
102	Tetrachloroethylene	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
103	Toluene	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
104	Toxaphene	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
105	TPH (C ₅ -C ₉)	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(12,22)

106 TPH (C₅-C₁₀)...

ลำดับที่	สารเคมี	วิธีวิเคราะห์
106	TPH (C ₅ -C ₁₀)	Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^(9,22)
107	TPH (C ₅ -C ₁₀)	Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^(9,22)
108	1,2,4-Trichlorobenzene	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
109	1,1,1-Trichloroethane	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
110	1,1,2-Trichloroethane	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
111	Trichloroethylene	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
112	2,4,5-Trichlorophenol	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
113	2,4,6-Trichlorophenol	Liquid-Liquid Extraction, Gas Chromatographic Method ⁽⁴⁾
114	1,3,5-Trimethylbenzene	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
115	Vanadium	1) Digestion, Electrothermal Atomic Absorption Spectrometric Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾
116	Vinyl acetate	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
117	Vinyl chloride	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
118	m-Xylene	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
119	o-Xylene	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
120	p-Xylene	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
121	Xylene (Total)	Purge and Trap Gas Chromatographic/ Mass Spectrometric Method ⁽⁴⁾
122	Zinc	1) Digestion, Direct Air-Acetylene Flame Method ⁽⁴⁾ 2) Digestion, Inductively Coupled Plasma Method ⁽⁴⁾

อากาศเสีย...

อากาศเสีย (ปล่อยระบาย) จำนวน 18 รายการ

ลำดับที่	สารเคมี	วิธีวิเคราะห์
1	Antimony	1) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ⁽⁵⁾ 2) Isokinetic Sampling, Digestion, Direct Air-Acetylene Flame Method ⁽⁵⁾ 3) Isokinetic Sampling, Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ⁽⁵⁾
2	Arsenic	Isokinetic Sampling, Digestion, Hydride Generation/ Atomic Absorption Spectrometric Method ⁽⁵⁾
3	Carbon monoxide	Instrumental Analyzer Method ⁽⁵⁾
4	Chlorine	Absorption Sampling, Ion Chromatographic Method ⁽⁵⁾
5	Copper	1) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ⁽⁵⁾ 2) Isokinetic Sampling, Digestion, Direct Air-Acetylene Flame Method ⁽⁵⁾
6	Cresol	Adsorption Sampling, Gas Chromatographic Method ⁽⁵⁾
7	Dioxins/Furans	Isokinetic Sampling, Analysis by ISO/IEC 17025 Accredited Laboratory or Analysis by Department of Industrial Works Registered Laboratory (Dioxins/Furans Analysis Approved) ⁽⁵⁾
8	Hydrogen Chloride	Absorption Sampling, Ion Chromatographic Method ⁽⁵⁾
9	Hydrogen Fluoride	Absorption Sampling, Ion Chromatographic Method ⁽⁵⁾
10	Hydrogen Sulfide	Absorption Sampling, Iodometric Method ⁽⁵⁾
11	Lead	1) Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method ⁽⁵⁾ 2) Isokinetic Sampling, Digestion, Direct Air-Acetylene Flame Method ⁽⁵⁾ 3) Isokinetic Sampling, Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ⁽⁵⁾
12	Mercury	Isokinetic Sampling, Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ⁽⁵⁾
13	Opacity	Ringelmann's Method ⁽²⁾
14	Oxides of Nitrogen	1) Absorption Sampling, Phenoldisulfonic acid Method ⁽⁵⁾ 2) Instrumental Analyzer Method ⁽⁵⁾

15 Sulfur dioxide...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
15	Sulfur dioxide	1) Absorption Sampling, Barium-Thorin Titrimetric Method ⁽⁵⁾ 2) Instrumental Analyzer Method ⁽⁵⁾
16	Sulfuric acid	Isokinetic Sampling, Barium-Thorin Titrimetric Method ⁽⁵⁾
17	Total Suspended Particulate	Isokinetic Sampling, Gravimetric Method ⁽⁵⁾
18	Xylene	Absorption Sampling, Gas Chromatographic Method ⁽⁵⁾

สิ่งบ่งชี้หรือตัววัดที่ใช้แล้ว จำนวน 36 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Aldrin	1) Waste Extraction, Solid-Phase Extraction, Gas Chromatographic Method ^(1,10,24) 2) Solid-Phase Extraction, Gas Chromatographic Method ^(10,24) 3) Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
2	Antimony	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^(1,4,13) 2) Waste Extraction, Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(1,4,14) 3) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,4,14) 4) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 5) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,14) 6) Digestion, Inductively Coupled Plasma Method ^(7,14)
3	Arsenic	1) Waste Extraction, Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^(1,4,17) 2) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^(7,17)
4	Barium	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^(1,4,15) 2) Waste Extraction, Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(1,4,16) 3) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,4,14)

4) Digestion...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
5	Beryllium	4) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 5) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,14) 6) Digestion, Inductively Coupled Plasma Method ^(7,14) 1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^(1,4,15) 2) Waste Extraction, Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(1,4,16) 3) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,4,14) 4) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 5) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,14) 6) Digestion, Inductively Coupled Plasma Method ^(7,14)
6	Cadmium	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^(1,4,15) 2) Waste Extraction, Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(1,4,16) 3) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,4,14) 4) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 5) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,14) 6) Digestion, Inductively Coupled Plasma Method ^(7,14)
7	Chlordane	1) Waste Extraction, Solid-Phase Extraction, Gas Chromatographic Method ^(1,10,24) 2) Solid-Phase Extraction, Gas Chromatographic Method ^(10,24) 3) Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
8	Chromium	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^(1,4,15) 2) Waste Extraction, Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(1,4,16)

3) Waste Extraction...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
9	Chromium (III)	3) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,4,14) 4) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 5) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,14) 6) Digestion, Inductively Coupled Plasma Method ^(7,14) 1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method; Waste Extraction, Colorimetric Method; Calculation ^(1,4,13,18) 2) Waste Extraction, Digestion, Graphite Furnace Atomic Absorption Spectrometric Method; Waste Extraction, Colorimetric Method; Calculation ^(1,4,14,18) 3) Waste Extraction, Digestion, Inductively Coupled Plasma Method; Waste Extraction, Colorimetric Method; Calculation ^(1,4,14,18) 4) Digestion, Flame Atomic Absorption Spectrometric Method; Alkaline Digestion, Colorimetric Method; Calculation ^(7,15,18) 5) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method; Alkaline Digestion, Colorimetric Method; Calculation ^(7,14,18) 6) Digestion, Inductively Coupled Plasma Method; Alkaline Digestion, Colorimetric Method; Calculation ^(7,14,18)
10	Chromium (VI)	1) Waste Extraction, Colorimetric Method ^(8,18) 2) Alkaline Digestion, Colorimetric Method ^(8,18)
11	Cobalt	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^(1,4,15) 2) Waste Extraction, Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(1,4,16) 3) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,4,14) 4) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 5) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,14) 6) Digestion, Inductively Coupled Plasma Method ^(7,14)

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
12	Copper	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^(1,4,15) 2) Waste Extraction, Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(1,4,16) 3) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,4,14) 4) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 5) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,14) 6) Digestion, Inductively Coupled Plasma Method ^(7,14)
13	2,4-D	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^(1,10,24) 2) Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
14	DDD	1) Waste Extraction, Solid-Phase Extraction, Gas Chromatographic Method ^(1,10,24) 2) Solid-Phase Extraction, Gas Chromatographic Method ^(10,24) 3) Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
15	DOE	1) Waste Extraction, Solid-Phase Extraction, Gas Chromatographic Method ^(1,10,24) 2) Solid-Phase Extraction, Gas Chromatographic Method ^(10,24) 3) Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
16	DOT	1) Waste Extraction, Solid-Phase Extraction, Gas Chromatographic Method ^(1,10,24) 2) Solid-Phase Extraction, Gas Chromatographic Method ^(10,24) 3) Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
17	Dieldrin	1) Waste Extraction, Solid-Phase Extraction, Gas Chromatographic Method ^(1,10,24) 2) Solid-Phase Extraction, Gas Chromatographic Method ^(10,24) 3) Soxhlet Extraction, Gas Chromatographic Method ^(11,24)

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
18	Endrin	1) Waste Extraction, Solid-Phase Extraction, Gas Chromatographic Method ^(1,9,24) 2) Solid-Phase Extraction, Gas Chromatographic Method ^(10,24) 3) Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
19	Heptachlor	1) Waste Extraction, Solid-Phase Extraction, Gas Chromatographic Method ^(1,10,24) 2) Solid-Phase Extraction, Gas Chromatographic Method ^(10,24) 3) Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
20	Lead	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^(1,6,15) 2) Waste Extraction, Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(1,6,16) 3) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,6,14) 4) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 5) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,16) 6) Digestion, Inductively Coupled Plasma Method ^(7,14)
21	Lindane	1) Waste Extraction, Solid-Phase Extraction, Gas Chromatographic Method ^(1,10,24) 2) Solid-Phase Extraction, Gas Chromatographic Method ^(10,24) 3) Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
22	Mercury	1) Waste Extraction, Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^(1,6,19) 2) Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ⁽²⁴⁾
23	Methoxychlor	1) Waste Extraction, Solid-Phase Extraction, Gas Chromatographic Method ^(1,10,24) 2) Solid-Phase Extraction, Gas Chromatographic Method ^(10,24)

3) Soxhlet...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
24	Mirex	3) Soxhlet Extraction, Gas Chromatographic Method ^(11,24) 1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^(1,9,24) 2) Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
25	Molybdenum	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^(1,6,15) 2) Waste Extraction, Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(1,6,16) 3) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,6,14) 4) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 5) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,16) 6) Digestion, Inductively Coupled Plasma Method ^(7,14)
26	Nickel	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^(1,6,15) 2) Waste Extraction, Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(1,6,16) 3) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,6,14) 4) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 5) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,16) 6) Digestion, Inductively Coupled Plasma Method ^(7,14)
27	Polychlorinated Biphenyls Aroclor 1016 Aroclor 1221 Aroclor 1232 Aroclor 1242 Aroclor 1248 Aroclor 1254 Aroclor 1260 2,4,4'-Trichlorobiphenyl 2,2',5,5'-Tetrachlorobiphenyl	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^(1,9,25) 2) Waste Extraction, Solid-Phase Extraction, Gas Chromatographic Method ^(1,10,25) 3) Soxhlet Extraction, Gas Chromatographic Method ^(11,25)

2,2',4,5,5'...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
28	2,2',4,5,5'-Pentachlorobiphenyl 2,2',3,4,4',5'-Hexachlorobiphenyl 2,2',4,4',5,5'-Hexachlorobiphenyl 2,2',3,4,4',5,5'-Heptachlorobiphenyl Pentachlorophenol	1) Waste Extraction, Separatory Funnel Liquid-Liquid Extraction, Gas Chromatographic Method ^(1,9,24) 2) Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
29	Selenium	1) Waste Extraction, Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^(1,6,21) 2) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^(7,21)
30	Silver	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^(1,6,15) 2) Waste Extraction, Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(1,6,16) 3) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,6,14) 4) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 5) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,16) 6) Digestion, Inductively Coupled Plasma Method ^(7,14)
31	Thallium	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^(1,6,15) 2) Waste Extraction, Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(1,6,16) 3) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,6,14) 4) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 5) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,16) 6) Digestion, Inductively Coupled Plasma Method ^(7,14)

32 Toxaphene...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
32	Toxaphene	1) Waste Extraction, Solid-Phase Extraction, Gas Chromatographic Method ^(1,10,24) 2) Solid-Phase Extraction, Gas Chromatographic Method ^(10,24) 3) Soxhlet Extraction, Gas Chromatographic/Mass Spectrometric Method ^(11,27)
33	Trichloroethylene	1) Waste Extraction, Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^(1,12,28) 2) Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^(11,28)
34	Vanadium	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^(1,6,15) 2) Waste Extraction, Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(1,6,16) 3) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,6,14) 4) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 5) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,16) 6) Digestion, Inductively Coupled Plasma Method ^(7,14)
35	Vinyl chloride	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^(11,28)
36	Zinc	1) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^(1,6,15) 2) Waste Extraction, Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(1,6,16) 3) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^(1,6,14) 4) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 5) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,16) 6) Digestion, Inductively Coupled Plasma Method ^(7,14)

33...

เดิม จำนวน 121 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Acenaphthene	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
2	Acetone	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
3	Aldrin	Soxhlet Extraction, Gas Chromatographic Method ^(11,26)
4	Anthracene	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
5	Antimony	1) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 2) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,16) 3) Digestion, Inductively Coupled Plasma Method ^(7,14)
6	Arsenic	Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^(7,17)
7	Atrazine	Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
8	Barium	1) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 2) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,16) 3) Digestion, Inductively Coupled Plasma Method ^(7,14)
9	Benz(a)anthracene	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
10	Benzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
11	Benzo(b)fluoranthene	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
12	Benzo(k)fluoranthene	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
13	Benzoic acid	Soxhlet Extraction, Gas Chromatographic Method ^(11,23)
14	Benzo(a)pyrene	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
15	Benzo(g,h,i)perylene	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
16	Beryllium	1) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15)

2) Digestion...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
		2) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,16) 3) Digestion, Inductively Coupled Plasma Method ^(7,14)
17	Bis(2-chloroethyl)ether	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
18	Bis(2-ethylhexyl)phthalate	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
19	Bromodichloromethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
20	Bromoform	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
21	Butanol	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
22	Butyl benzyl phthalate	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
23	Cadmium	1) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 2) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,16) 3) Digestion, Inductively Coupled Plasma Method ^(7,14)
24	Carbazole	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
25	Carbon disulfide	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
26	Carbon tetrachloride	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
27	Chlordane	Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
28	p-Chloroaniline	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
29	Chlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
30	Chlorodibromomethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
31	Chloroform	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
32	Chromium	1) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15)

2) Digestion...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
33	Chromium (III)	2) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,16) 3) Digestion, Inductively Coupled Plasma Method ^(7,14) 1) Digestion, Flame Atomic Absorption Spectrometric Method; Alkaline Digestion, Colorimetric Method; Calculation ^(7,13,18) 2) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method; Alkaline Digestion, Colorimetric Method; Calculation ^(7,13,18) 3) Digestion, Inductively Coupled Plasma Method; Alkaline Digestion, Colorimetric Method; Calculation ^(7,13,18)
34	Chromium (VI)	Alkaline Digestion, Colorimetric Method ^(8,18)
35	Chrysene	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
36	Cyanide	1) Extraction, Distillation, Titrimetric Method ^(28,29,30) 2) Extraction, Distillation, Colorimetric Method ^(28,29,30)
37	2,4-D	Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
38	DDD	Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
39	DDE	Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
40	DDT	Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
41	Dibenz(a,h)anthracene	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
42	Di-n-butyl phthalate	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
43	1,2-Dichlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
44	1,3-Dichlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
45	1,4-Dichlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
46	1,1-Dichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
47	1,2-Dichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
48	1,1-Dichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)

49 cis-1,2-Dichloroethylene...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
49	cis-1,2-Dichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
50	trans-1,2-Dichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
51	1,2-Dichloropropane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
52	1,3-Dichloropropane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
53	1,3-Dichloropropene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
54	Dieldrin	Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
55	Diethyl phthalate	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
56	2,4-Dimethylphenol	Soxhlet Extraction, Gas Chromatographic Method ^(11,23)
57	2,4-Dinitrophenol	Soxhlet Extraction, Gas Chromatographic Method ^(11,23)
58	2,4-Dinitrotoluene	Soxhlet Extraction, Gas Chromatographic Method ^(11,23)
59	2,6-Dinitrotoluene	Soxhlet Extraction, Gas Chromatographic Method ^(11,23)
60	Di-n-Octyl phthalate	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
61	Endosulfan	Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
62	Endrin	Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
63	Ethylbenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
64	Fluoranthene	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
65	Fluorene	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
66	Heptachlor	Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
67	Heptachlor epoxide	Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
68	Hexachloro-1,3-butadiene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
69	n-Hexane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
70	α-HCH	Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
71	β-HCH	Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
72	γ-HCH	Soxhlet Extraction, Gas Chromatographic Method ^(11,24)

73 Hexachlorocyclopentadiene...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
73	Hexachlorocyclopentadiene	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
74	Hexachloroethane	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
75	Indeno(1,2,3-cd)pyrene	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
76	Isophorone	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
77	Lead	1) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 2) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,16) 3) Digestion, Inductively Coupled Plasma Method ^(7,14)
78	Manganese	1) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 2) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,16) 3) Digestion, Inductively Coupled Plasma Method ^(7,14)
79	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ⁽²⁴⁾
80	Methanol	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
81	Methoxychlor	Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
82	Methyl bromide	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
83	Methylene chloride	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
84	2-Methylphenol	Soxhlet Extraction, Gas Chromatographic Method ^(11,23)
85	2-Methylnaphthalene	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
86	Methyl tert-butyl ether	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
87	Naphthalene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
88	Nickel	1) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 2) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,16) 3) Digestion, Inductively Coupled Plasma Method ^(7,14)

89 Nitrobenzene...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
89	Nitrobenzene	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
90	N-Nitrosodiphenylamine	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
91	N-Nitrosodi-n-propylamine	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
92	Polychlorinated Biphenyls	Soxhlet Extraction, Gas Chromatographic Method ^(11,25)
	Aroclor 1016	
	Aroclor 1221	
	Aroclor 1232	
	Aroclor 1242	
	Aroclor 1248	
	Aroclor 1254	
	Aroclor 1260	
	2,2',5,5'-Tetrachlorobiphenyl	
	2,2',4,5'-Pentachlorobiphenyl	
	2,2',3,4,4',5'-Hexachlorobiphenyl	
	2,2',4,4',5,5'-Hexachlorobiphenyl	
	2,2',3,4,4',5,5'-Heptachlorobiphenyl	
93	Pentachlorophenol	Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
94	Phenanthrene	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
95	Phenol	Soxhlet Extraction, Gas Chromatographic Method ^(11,23)
96	Pyrene	Soxhlet Extraction, Gas Chromatographic/ Mass Spectrometric Method ^(11,27)
97	Selenium	Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^(7,21)
98	Silver	1) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 2) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,16) 3) Digestion, Inductively Coupled Plasma Method ^(7,14)
99	Styrene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)

100 1,1,2,2-Tetrachloroethane...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
100	1,1,2,2-Tetrachloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
101	Tetrachloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
102	Toluene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
103	Toxaphene	Soxhlet Extraction, Gas Chromatographic Method ^(11,24)
104	TPH (C ₈ -C ₁₄)	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
105	TPH (C ₈ -C ₁₄)	Soxhlet Extraction, Gas Chromatographic Method ^(11,23)
106	TPH (C ₁₆ -C ₃₅)	Soxhlet Extraction, Gas Chromatographic Method ^(11,23)
107	1,2,4-Trichlorobenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
108	1,1,1-Trichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
109	1,1,2-Trichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
110	Trichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
111	2,4,5-Trichlorophenol	Soxhlet Extraction, Gas Chromatographic Method ^(11,23)
112	2,4,6-Trichlorophenol	Soxhlet Extraction, Gas Chromatographic Method ^(11,23)
113	1,3,5-Trimethylbenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
114	Vanadium	1) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 2) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,16) 3) Digestion, Inductively Coupled Plasma Method ^(7,14)
115	Vinyl acetate	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
116	Vinyl chloride	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
117	m-Xylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
118	o-Xylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
119	p-Xylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)

120 Xylene (Total)

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
120	Xylene (Total)	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^(13,26)
121	Zinc	1) Digestion, Flame Atomic Absorption Spectrometric Method ^(7,15) 2) Digestion, Graphite Furnace Atomic Absorption Spectrometric Method ^(7,16) 3) Digestion, Inductively Coupled Plasma Method ^(7,14)

เอกสารอ้างอิง

- กระทรวงอุตสาหกรรม. ประกาศกระทรวงอุตสาหกรรม, พ.ศ. 2548. เรื่อง การกำจัดสิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว.ราชกิจจานุเบกษา. 25 มกราคม 2549. เล่มที่ 123 ตอนพิเศษ 11ง.
- กระทรวงอุตสาหกรรม. ประกาศกระทรวงอุตสาหกรรม, พ.ศ. 2549. เรื่อง กำหนดค่าปริมาณแอมโมเนียในอากาศที่ระบายออกจากปล่องของหม้อน้ำโรงสีข้าวที่ใช้กลบเป็นเชื้อเพลิง.ราชกิจจานุเบกษา. 4 ธันวาคม 2549. เล่มที่ 123 ตอนพิเศษ 125ง.
- สมาคมวิศวกรรมสิ่งแวดล้อมแห่งประเทศไทย. คู่มือวิเคราะห์น้ำเสีย. พิมพ์ครั้งที่ 4. กรุงเทพฯ: เอ็นเนกการพิมพ์, 2547.
- APHA, AWWA, WEF. Standard Methods for the Examination of Water and Wastewater. 23rd ed. Washington, DC: APHA, 2017.
- United States Environmental Protection Agency. Standards of Performance for New Stationary Sources. 40 CFR 60. Appendix A, 2022.
- United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. SW-846, 1997.
- United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Acid Digestion of Sludges and Sediments and Soils. SW-846 Method 3050B, 1996.
- United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Alkaline Digestion for Hexavalent Chromium. SW-846 Method 3060A, 1996.
- United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Separatory Funnel Liquid-Liquid Extraction. SW-846 Method 3510C, 1996.
- United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Solid Phase Extraction. SW-846 Method 3535A, 2007.
- United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Soxhlet Extraction. SW-846 Method 3540C, 1996.

12. United States...

12. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Purge-and-Trap for Aqueous Samples. SW-846 Method 5030C, 2003.
13. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Closed-System Purge-and-Trap and Extraction for Volatile Organics in Soil and Waste Samples. SW-846 Method 5035A, 2007.
14. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Inductively Coupled Plasma-Optical Emission Spectrometry. SW-846 Method 6010D, 2018.
15. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Flame Atomic Absorption Spectrophotometry. SW-846 Method 7000B, 2007.
16. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Graphite Furnace Atomic Absorption Spectrophotometry. SW-846 Method 7010, 2007.
17. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Arsenic (Atomic Absorption, Gaseous Hydride). SW-846 Method 7061A, 1992.
18. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Chromium, Hexavalent (Colorimetric). SW-846 Method 7196A, 1992.
19. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Mercury in Liquid Waste (Manual Cold-Vapor Technique). SW-846 Method 7470A, 1994.
20. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Mercury in Solid or Semisolid Waste (Manual Cold-Vapor Technique). SW-846 Method 7471A, 1994.
21. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Selenium (Atomic Absorption, Borohydride Reduction). SW-846 Method 7742, 1994.
22. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Nonhalogenated Organics Using GC/FID. SW-846 Method 8015D, 2003.
23. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Phenols by Gas Chromatography. SW-846 Method 8041, 1996. *Small*

24. United States...

24. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Organochlorine Pesticides by Gas Chromatography. SW-846 Method 8081B, 2007.
25. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Polychlorinated Biphenyls (PCBs) by Gas Chromatography. SW-846 Method 8082A, 2007.
26. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Volatile Organic Compounds by Gas Chromatography/Mass Spectrometry (GC/MS). SW-846 Method 8260C, 2006.
27. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Semivolatile Organic Compounds by Gas Chromatography/Mass Spectrometry. SW-846 Method 8270D, 2014.
28. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Total and Amenable Cyanide: Distillation. SW-846 Method 9010C, 2004.
29. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Cyanide Extraction Procedure for Solids and Oil. SW-846 Method 9013A, 1996.
30. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Cyanide in Waters and Extracts Using Titrimetric and Manual Spectrophotometric Procedures. SW-846 Method 9014, 2014. *Small*