


ภาคผนวก ข-32

เอกสารการสอบเทียบ

THE LINDE GROUP



Work order no.

83754772

DATE

13-Jun-24

PTWC No.

: 45

Standard Reference

: BOC Limited

Certificate No.

: LC-0018/23, 21/017565

PERMISSIBLE ERROR

: ± 0.20 % VOL.

CHECKED

Range : 0-100 % VOL.

Tag No.

: AT-16011

Model

: URAS 26

Maker

: ABB

Serial No.

: 3.372791.7

Measured

: PRODUCT CABON MONOXIDE (%)

Reference	Reading	Error	Remark
Zero Gas UHP NITROGEN PURITY> 99.9999% N ₂ CYLINDER No. 39542 Expired Date 13 May 2026	0.0	0.00	11.30-12.00
Span Gas 99.997 % UHP CO CYLINDER No. 3991552 Expired Date 30 Jun 2026	99.90	-0.10	12.00-13.30


CALIBRATION

Reference	Reading	Error	Remark
Zero Gas UHP NITROGEN PURITY> 99.9999% N ₂ CYLINDER No. 39542 Expired Date 13 May 2026	0.00	0.00	
Span Gas 99.997 % UHP CO CYLINDER No. 3991552 Expired Date 30 Jun 2026	100.00	0.00	


Result

☒ Accepted Note
☐ Not Accepted
☐ Accepted as note

Tested By



Approved By



Date

13 / Jun / 2024

Date

13 / Jun / 2024

PM-CD/F-005

1-Nov-95