

# เอกสารแนบ 6

เอกสารสอบเทียบเครื่องมือ  
ที่ใช้ในการวิเคราะห์



# CALIBRATION LABORATORY Co., LTD.

2/10-11, 14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



## REPORT OF CALIBRATION

### FOR

NOMENCLATURE : BALANCE  
MANUFACTURER : METTLER TOLEDO  
MODEL / TYPE : MS204TS/00  
SERIAL NO. : B935191252[LA-002]  
LOCATION SITE : LABORATORY - BALANCE ROOM  
DATE OF CALIBRATION : 04 April 2024

#### ENVIRONMENT CONDITIONS :

Temperature : 24 °C to 25 °C  
Relative Humidity : 50 % to 51 %

#### PROCEDURE USED :

This instrument was calibrated under procedure No. WI-305-46 based on EURAMET cg-18 Version 4.0 (11/2015).  
The calibration was performed by Comparison with Weight Set which maintained by the Calibration Laboratory Co., Ltd.

#### REFERENCE STANDARD USED :

1. Weight Set, Phoenix Class E2 S/N. WBS-SET-E2-01.
2. Weight, Sartorius Class E2 S/N. 44329129, 43529037, 44329167, 43529293.

#### TRACEABILITY :

1. The measurements are traceable to International System of Units (SI), through National Institute of Metrology (Thailand).  
Certificate No. MM-0123-22, Due Date 22 August 2024.
2. The measurements are traceable to International System of Units (SI), through Sartorius Lab Instruments GmbH & Co. KG.  
Certificate No. M141607, M141608, M141609, M141611. Due Date 15 September 2025.

#### UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95%. It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2022)"

Certificate No. Q24024790

F3-011-05/12-23

page 2 of 3



@clcalibration



# CALIBRATION LABORATORY Co., LTD.

2/10-11, 14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



## CERTIFICATE OF CALIBRATION

### FOR

NOMENCLATURE : BALANCE  
MANUFACTURER : METTLER TOLEDO  
MODEL / TYPE : MS204TS/00  
SERIAL NO. : B935191252[LA-002]  
CLID. NO. : 362200356  
JOB CONTROL NO. : 240307024790  
CALIBRATION SERVICE : ☐ IN-LABORATORY ☒ ON-SITE

CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.

5/45 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BON SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 07 March 2024

DATE OF ISSUED : 08 April 2024

The report of calibration shall not be reproduced except in full without approval of the Calibration Laboratory Co., Ltd.

Calibrated By :



Approved By :

08 April 2024

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q24024790

F3-011-05/12-23

page 1 of 3



@clcalibration

## CERTIFICATE OF CALIBRATION

### FOR

NOMENCLATURE : WATER BATH  
MANUFACTURER : MEMMERT  
MODEL / TYPE : WNB14  
SERIAL NO. : L418.0758[LA-004]  
CLID. NO. : 332100157  
JOB CONTROL NO. : 240307024794  
CALIBRATION SERVICE : ☐ IN-LABORATORY ☒ ON-SITE

CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.


5/45 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BON SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 07 March 2024

DATE OF ISSUED : 06 April 2024

The report of calibration shall not be reproduced except in full without approval of the Calibration Laboratory Co., Ltd.

Calibrated By :



Approved By :

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q24024794  
F3-011-05/12-23

page 1 of 4



@clcalibration

CONDITION OF CALIBRATION ITEM : RECEIVED IN GOOD OPERATIONAL CONDITION  
MEASUREMENT RESULTS : (X) without adjustment ( ) adjustment

### CALIBRATION DATA


#### 1. Error of indications

Nominal Test Value (g)	Conventional mass (g)	Display Value (g)	Error of Balance (g)	Uncertainty $\pm$ (mg)	Coverage factor k
Unload	0.0000	0.0000	0.0000	0.07	2,00
0.1000	0.1000	0.0999	-0.0001	0.11	2,00
0.5000	0.5000	0.5000	0.0000	0.11	2,00
1.0000	1.0000	1.0000	0.0000	0.11	2,00
2.0000	2.0000	2.0000	0.0000	0.11	2,00
5.0000	5.0000	4.9999	-0.0001	0.11	2,00
10.0000	10.0000	9.9999	-0.0001	0.11	2,00

#### 2. Repeatability of indications

Nominal Test Value (g)	Standard Deviation of Reading (g)
200.0000	0.00005

#### 3. Effect of eccentric application of a load on the indication

<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Nominal Test Value (g)	Display Value (g)	Maximum Difference of Center Value (g)
100.0000	Position 1 100.0000	Position 2 100.0001
	Position 3 100.0001	Position 4 99.9999
	Position 5 100.0001	100.0001
		0.0001

Note. The Scope of Accredited TISI Certificate No. 23-LB0092 Issue 02 Page 116 of 138

This report is valid for the above stated instrument/s only.

### End of Certificate ###

Certificate No. Q24024790  
F3-011-05/12-23

page 3 of 3



@clcalibration



# CALIBRATION LABORATORY CO., LTD.

2/10-11, 14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2872 www.cal-laboratory.com E-mail: sale@cal-laboratory.com



NSC-TISI-TIS 17025  
CALIBRATION 0659  
CLC

## CONDITION OF CALIBRATION ITEM : RECEIVED IN GOOD OPERATIONAL CONDITION

### MEASUREMENT RESULTS : ( X ) without adjustment ( ) adjustment

The table in the following gives the calibration results and associated measurement uncertainties of the measuring water bath.

#### CALIBRATION DATA

##### 1. WATER BATH PERFORMANCE

Test Point ( °C )	DUC Reading ( °C )	Uniformity ( °C )	Stability ( °C )
95.0	95.0	0.50	0.21

Certificate No. Q24024794

F3-011-05/12-23

page 3 of 4



@clcalibration



# CALIBRATION LABORATORY CO., LTD.

2/10-11, 14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2872 www.cal-laboratory.com E-mail: sale@cal-laboratory.com



NSC-TISI-TIS 17025  
CALIBRATION 0659  
CLC

## REPORT OF CALIBRATION

### FOR

NOMENCLATURE : WATER BATH  
MANUFACTURER : MEMMERT  
MODEL / TYPE : WNB14  
SERIAL NO. : L418.0758[LA-004]  
LOCATION SITE : LABORATORY-HOT ZONE  
DATE OF CALIBRATION : 04 April 2024

#### ENVIRONMENT CONDITIONS :

Temperature : 24 °C to 25 °C

Relative Humidity : 50% to 52%

#### PROCEDURE USED :

This instrument was calibrated under procedure No. WI-305-135 based on ASTM E 715-80:2016 as calibration guidelines.

The calibration was performed by using Hydra Series II which maintained by the Calibration Laboratory Co., Ltd.

#### REFERENCE STANDARD USED :

Hydra Series II, Fluke Model 2635A S/N. 7107303.

#### TRACEABILITY :

The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd.

Certificate No. Q24026699, Due Date 13 September 2024.

#### UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k = 2.00$  which for a normal distribution corresponds to a coverage probability of approximately 95 %. It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2022)"

Certificate No. Q24024794

F3-011-05/12-23

page 2 of 4



@clcalibration



# CALIBRATION LABORATORY Co., LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2872 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



NSC-TISI-TIS 17025  
CALIBRATION 0659  
CLC

## CERTIFICATE OF CALIBRATION

### FOR


NOMENCLATURE : WATER BATH  
MANUFACTURER : M-LAB  
MODEL / TYPE : WBN 15  
SERIAL NO. : 0335[LA-007]  
CLID. NO. : 332300657  
JOB CONTROL NO. : 240307024795  
CALIBRATION SERVICE : ☐ IN-LABORATORY ☒ ON-SITE  
CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.  
5/45 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BON SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 07 March 2024

DATE OF ISSUED : 06 April 2024

The report of calibration shall not be reproduced except in full without approval of the Calibration Laboratory Co., Ltd.

Calibrated By :



Approved By :

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q24024795  
F3-011-05/12-23

page 1 of 4



# CALIBRATION LABORATORY Co., LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2872 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



NSC-TISI-TIS 17025  
CALIBRATION 0659  
CLC


## CALIBRATION DATA

### 2. TEMPERATURE DISTRIBUTION

Test Point ( ° C )	DUC Reading ( ° C )	STD Reading ( ° C )					Uncertainty ± ( ° C )
		Probe No. 1	Probe No. 2	Probe No. 3	Probe No. 4	Probe No. 5	
95.0	95.0	96.62	96.74	96.93	96.68	96.66	0.65

Technical Note : W = 35 cm, D = 29 cm, H = 14 cm.

The Scope of Accredited TISI Certificate No. 23-LB0092 Issue 02 Page 128 of 138



This report is valid for the above stated instrument's only.

### End of Certificate ###

Certificate No. Q24024794  
F3-011-05/12-23

page 4 of 4



@clcalibration



@clcalibration



# CONDITION OF CALIBRATION ITEM : RECEIVED IN GOOD OPERATIONAL CONDITION

## MEASUREMENT RESULTS : ( X ) without adjustment ( ) adjustment

The table in the following gives the calibration results and associated measurement uncertainties of the measuring water bath.

### CALIBRATION DATA

#### 1. WATER BATH PERFORMANCE

Test Point ( °C )	DUC Reading ( °C )	Uniformity ( °C )	Stability ( °C )
85.0	85.0	0.48	0.17

Certificate No. Q24024795

F3-011-05/12-23

page 3 of 4



@clcalibration

# REPORT OF CALIBRATION

## FOR

NOMENCLATURE : WATER BATH  
MANUFACTURER : M-LAB  
MODEL / TYPE : WBN 15  
SERIAL NO. : 0335[LA-007]  
LOCATION SITE : LABORATORY-HOT ZONE  
DATE OF CALIBRATION : 04 April 2024

### ENVIRONMENT CONDITIONS :

Temperature : 24 °C to 25 °C

Relative Humidity : 50% to 52%

### PROCEDURE USED :

This instrument was calibrated under procedure No. WI-305-135 based on ASTM E 715-80:2016 as calibration guidelines.

The calibration was performed by using Hydra Series II which maintained by the Calibration Laboratory Co., Ltd.

### REFERENCE STANDARD USED :

Hydra Series II, Fluke Model 2635A S/N. 7107303.

### TRACEABILITY :

The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd.

Certificate No. Q24026699, Due Date 13 September 2024.

### UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k = 2.00$  which for a normal distribution corresponds to a coverage probability of approximately 95 %. It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2022)"

Certificate No. Q24024795

F3-011-05/12-23

page 2 of 4



@clcalibration



## CERTIFICATE OF CALIBRATION

### FOR

NOMENCLATURE : DIGITAL THERMOHYGRO METER  
MANUFACTURER : DIGICON  
MODEL / TYPE : TH-02A  
SERIAL NO. : 1919E0284980[DTH-02]  
CLID. NO. : 232100201  
JOB CONTROL NO. : 240408038370  
CALIBRATION SERVICE : ☒ IN-LABORATORY ☐ ON-SITE

CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.  
5/45 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BON SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 08 April 2024

DATE OF ISSUED : 17 April 2024

The report of calibration shall not be reproduced except in full without approval of the Calibration Laboratory Co., Ltd.

Calibrated By :

Approved By :

17 April 2024

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q24038370

F3-011-05/12-23

page 1 of 3



@ckcalibration




### CALIBRATION DATA

#### 2. TEMPERATURE DISTRIBUTION

Test Point ( ° C )	DUC Reading ( ° C )	STD Reading ( ° C )					Uncertainty ± ( ° C )
		Probe No. 1	Probe No. 2	Probe No. 3	Probe No. 4	Probe No. 5	
85.0	85.0	84.37	84.87	84.67	84.86	84.71	0.60

Technical Note : W = 35 cm, D = 30 cm, H = 15 cm.

The Scope of Accredited TISI Certificate No. 23-LB0092 Issue 02 Page 128 of 138



This report is valid for the above stated instrument/s only.

### End of Certificate ###

Certificate No. Q24024795

F3-011-05/12-23

page 4 of 4



@ckcalibration

## CONDITION OF CALIBRATION ITEM : RECEIVED IN GOOD OPERATIONAL CONDITION MEASUREMENT RESULTS : ( X ) without adjustment ( ) adjustment

The table in the following gives the calibration results and associated measurement uncertainties of the measuring digital thermohygro meter.

### CALIBRATION DATA

#### 1. CORRECTION OF TEMPERATURE

Test point ( ° C )	Actual Temperature ( ° C )	DUC Reading ( ° C )	Correction ( ° C )	Uncertainty ± ( ° C )
20.0	20.00	19.7	+0.30	0.27
25.0	25.00	24.7	+0.30	
30.0	30.00	29.6	+0.40	

#### 2. CORRECTION OF HUMIDITY

STD Temperature ( ° C )	STD Reading ( %RH )	DUC Reading ( %RH )	Correction ( %RH )	Uncertainty ± ( %RH )
25	40.0	33	+7.0	0.8
25	60.0	53	+7.0	0.9

Note. The Scope of Accredited ANAB Certificate No. ACDM-2814 Version 012 Page 59 of 67

This report is valid for the above stated instrument/s only.

### End of Certificate ###



@clcalibration

## REPORT OF CALIBRATION

### FOR

NOMENCLATURE : DIGITAL THERMOHYGRO METER  
MANUFACTURER : DIGICON  
MODEL / TYPE : TH-02A  
SERIAL NO. : 1919E0284980[DTH-02]  
DATE OF CALIBRATION : 10 April 2024

#### ENVIRONMENT CONDITIONS :

Temperature : (23 ± 2) °C Relative Humidity : (55 ± 10) %RH

#### PROCEDURE USED :

This instrument was calibrated under procedure No. CLC-CPTH-11. The calibration was performed by using Chilled Mirror Hygrometer which maintained by the Calibration Laboratory Co., Ltd.

#### REFERENCE STANDARD USED :

Chilled Mirror Hygrometer, Edgetech Model Dew Master S/N. 44602.  
Temperature & Humidity Chamber, PGC Model 9141-5116 S/N. 1304261.

#### TRACEABILITY :

The measurements are traceable to International System of Units (SI), through Thunder Scientific Corporation.  
Certificate No. 21594, Due Date 06 July 2024.

#### UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k = 2.00$  which for a normal distribution corresponds to a coverage probability of approximately 95 %.  
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2022)"



@clcalibration





# CALIBRATION LABORATORY Co., LTD.

2/10-11, 14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax. 02-578-2672 www.cal-laboratory.com E-mail: sale@cal-laboratory.com



ACCM-2814

## REPORT OF CALIBRATION

### FOR

NOMENCLATURE : DIGITAL THERMOHYGRO METER  
MANUFACTURER : DIGICON  
MODEL / TYPE : TH-02A  
SERIAL NO. : 1919E0284991[DTH-01]  
DATE OF CALIBRATION : 10 April 2024

#### ENVIRONMENT CONDITIONS :

Temperature :  $(23 \pm 2) ^\circ\text{C}$  Relative Humidity :  $(55 \pm 10) \% \text{RH}$

#### PROCEDURE USED :

This instrument was calibrated under procedure No. CLC-CPTH-11. The calibration was performed by using Chilled Mirror Hygrometer which maintained by the Calibration Laboratory Co., Ltd.

#### REFERENCE STANDARD USED :

Chilled Mirror Hygrometer, Edgetech Model Dew Master S/N. 44602.  
Temperature & Humidity Chamber, PGC Model 9141-5116 S/N. 1304261.

#### TRACEABILITY :

The measurements are traceable to International System of Units (SI), through Thunder Scientific Corporation.  
Certificate No. 21594, Due Date 06 July 2024.

#### UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k = 2.00$  which for a normal distribution corresponds to a coverage probability of approximately 95 %.  
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2022)"

Certificate No. Q24038369  
F3-011-05/12-23



# CALIBRATION LABORATORY Co., LTD.

2/10-11, 14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax. 02-578-2672 www.cal-laboratory.com E-mail: sale@cal-laboratory.com



ACCM-2814

## CERTIFICATE OF CALIBRATION

### FOR

NOMENCLATURE : DIGITAL THERMOHYGRO METER  
MANUFACTURER : DIGICON  
MODEL / TYPE : TH-02A  
SERIAL NO. : 1919E0284991[DTH-01]  
CLID. NO. : 232100200  
JOB CONTROL NO. : 240408038369  
CALIBRATION SERVICE : ☒ IN-LABORATORY ☐ ON-SITE

CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.


5/45 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BON SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 08 April 2024


DATE OF ISSUED : 17 April 2024

The report of calibration shall not be reproduced except in full without approval of the Calibration Laboratory Co., Ltd.

Calibrated By :



Approved By :



17 April 2024

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q24038369  
F3-011-05/12-23



## CERTIFICATE OF CALIBRATION

### FOR

NOMENCLATURE : DIGITAL THERMOMETER WITH PROBE  
MANUFACTURER : LUTRON  
MODEL / TYPE : MTM-380SD  
SERIAL NO. : 1570147/N/A[LA-0013/LA-0013/A]  
CLID. NO. : 232204019  
JOB CONTROL NO. : 240408038373  
CALIBRATION SERVICE : ☒ IN-LABORATORY ☐ ON-SITE

CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.  
5/45 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BON SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 08 April 2024 DATE OF ISSUED : 10 April 2024  
The report of calibration shall not be reproduced except in full without approval of the Calibration Laboratory Co., Ltd.

Calibrated By :

Approved By :

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q24038373  
F3-011-05/12-23

page 1 of 3



@clcalibration

## CONDITION OF CALIBRATION ITEM : RECEIVED IN GOOD OPERATIONAL CONDITION MEASUREMENT RESULTS : ( X ) without adjustment ( ) adjustment

The table in the following gives the calibration results and associated measurement uncertainties of the measuring digital thermohygro meter.

### CALIBRATION DATA

#### 1. CORRECTION OF TEMPERATURE

Test point ( ° C )	Actual Temperature ( ° C )	DUC Reading ( ° C )	Correction ( ° C )	Uncertainty ± ( ° C )
20.0	20.00	19.6	+0.40	0.27
25.0	25.00	24.6	+0.40	
30.0	30.00	29.5	+0.50	

#### 2. CORRECTION OF HUMIDITY

STD Temperature ( ° C )	STD Reading ( %RH )	DUC Reading ( %RH )	Correction ( %RH )	Uncertainty ± ( %RH )
25	40.0	32	+8.0	0.8
25	60.0	52	+8.0	0.9

Note. The Scope of Accredited ANAB Certificate No. ACDM-2814 Version 012 Page 59 of 67

This report is valid for the above stated instrument/s only.

### End of Certificate ###

Certificate No. Q24038369  
F3-011-05/12-23

page 3 of 3



@clcalibration

CONDITION OF CALIBRATION ITEM : RECEIVED IN GOOD OPERATIONAL CONDITION

MEASUREMENT RESULTS : ( X ) without adjustment ( ) adjustment

The DUC Reading were recorded and the means value were reported of five times measurement in the table below.

CALIBRATION DATA

CORRECTION OF TEMPERATURE : T1 [ THERMOCOUPLE TYPE K ]

Immersion depth (mm)	Actual Temperature ( °C )	DUC Reading ( °C )	Correction ( °C )	Uncertainty ± ( °C )
200	4.00	4.2	-0.20	0.52
	20.01	20.3	-0.29	
	95.04	95.6	-0.56	
	104.02	104.9	-0.88	
	180.03	181.4	-1.37	

Note. The Scope of Accredited ANAB Certificate No. ACDM-2814 Version 012 Page 56 of 67

This report is valid for the above stated instrument/s only.

### End of Certificate ###

Certificate No. Q24038373

F3-011-05/12-23

page 3 of 3



@cccalibration

REPORT OF CALIBRATION

FOR

NOMENCLATURE : DIGITAL THERMOMETER WITH PROBE  
MANUFACTURER : LUTRON  
MODEL / TYPE : MTM-380SD  
SERIAL NO. : I.570147N/A[LA-0013/LA-0013/A]  
DATE OF CALIBRATION : 09 April 2024

ENVIRONMENT CONDITIONS :

Temperature : (23 ± 2) °C Relative Humidity : (55 ± 10) % RH

PROCEDURE USED :

This instrument was calibrated under procedure No. CLC-CPTH-06 based on ASTM E 220-86 as calibration guidelines.

The calibration was performed by using Calibration Bath, Precision Thermometer and IPKT which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

1. Calibration Bath, Kambic Model OB-22/2 ULT, OB-22/2 S/N. 17115653, 17115654.
2. Precision Thermometer, ASL Model F200-A-8 S/N. 014433/03.
3. IPKT, ASL Model T100-250-1D, T100-450-1D S/N. L0193A-1-1, L1123A-1-5.

TRACEABILITY :

1. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd. Certificate No. Q23136342, Q23126517. Due Date 20 December 2024, 20 November 2024.
2. The measurements are traceable to International System of Units (SI), through Thailand Institute of Scientific and Technological Research (TISTR). Certificate No. PSL-T 0203/67, Due Date 07 December 2024.
3. The measurements are traceable to International System of Units (SI), through National Institute of Metrology (Thailand). Certificate No. TT-0136-23, TT-0132-23. Due Date 12 December 2024, 22 November 2024.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k = 2,00$  which for a normal distribution corresponds to a coverage probability of approximately 95 %. It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2022)"

Certificate No. Q24038373

F3-011-05/12-23

page 2 of 3



@cccalibration



# CALIBRATION LABORATORY Co., LTD.

2/10-11, 14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



## REPORT OF CALIBRATION

### FOR

NOMENCLATURE : REFRIGERATOR  
MANUFACTURER : MEDICOOL  
MODEL / TYPE : BB-117  
SERIAL NO. : BB117-190725001[LA-003]  
LOCATION SITE : LABORATORY  
DATE OF CALIBRATION : 04 April 2024

### ENVIRONMENT CONDITIONS :

Temperature : 24 °C to 25 °C

Relative Humidity : 50 % to 52 %

### PROCEDURE USED :

This instrument was calibrated under procedure No. WI-305-165 based on TLAS G-20-1/02-08 as calibration guidelines.  
The calibration was performed by using Hydra Data Logger which maintained by the Calibration Laboratory Co., Ltd.

### REFERENCE STANDARD USED :

Hydra Data Logger, Fluke Model 2620 S/N. 5592550.

### TRACEABILITY :

The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd.  
Certificate No. Q23065868, Due Date 22 June 2024.

### UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.  
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2022)"

Certificate No. Q24024793

F3-011-05/12-23



# CALIBRATION LABORATORY Co., LTD.

2/10-11, 14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



## CERTIFICATE OF CALIBRATION

### FOR

NOMENCLATURE : REFRIGERATOR  
MANUFACTURER : MEDICOOL  
MODEL / TYPE : BB-117  
SERIAL NO. : BB117-190725001[LA-003]  
CLID. NO. : 332100156  
JOB CONTROL NO. : 240307024793  
CALIBRATION SERVICE : ☐ IN-LABORATORY ☒ ON-SITE

### CUSTOMER :



ENVIRONMENTAL MEASUREMENTS CO., LTD.  
5/45 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BON SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 07 March 2024

DATE OF ISSUED : 06 April 2024

The report of calibration shall not be reproduced except in full without approval of the Calibration Laboratory Co., Ltd.

Calibrated By :



Approved By :

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q24024793

F3-011-05/12-23






### CALIBRATION DATA

#### 2. TEMPERATURE DISTRIBUTION

DUC		Measured Temperature ( °C ) @Probe No.9 is Ref.								Uncertainty ± ( °C )	Coverage factor <i>k</i>
Setting ( °C )	Indicating ( °C )	1	2	3	4	5	6	7	8	9	
4.0	4.0	3.22	3.93	3.21	3.93	3.52	3.29	3.02	3.18	3.31	2.00

Technical Note : W = 50 cm, D = 50 cm, H = 120 cm.

The Scope of Accredited TISI Certificate No. 23-LB0092 Issue 02 Page 129 of 138



This report is valid for the above stated instrument/s only.

### End of Certificate ###



### CONDITION OF CALIBRATION ITEM : RECEIVED IN GOOD OPERATIONAL CONDITION

#### MEASUREMENT RESULTS : ( X ) without adjustment ( ) adjustment

The table in the following gives the calibration results and associated measurement uncertainties of the measuring refrigerator.

### CALIBRATION DATA

#### 1. REFRIGERATOR PERFORMANCE

DUC		Measured Uniformity ( °C )	Measured Stability ( °C )	Measured Overall Variation ( °C )
Setting ( °C )	Indicating ( °C )			
4.0	4.0	0.82	1.14	2.49







# CALIBRATION LABORATORY Co., LTD.

210-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



NSC-TIS-TIS 17025  
CALIBRATION 0059  
CLC

## REPORT OF CALIBRATION

### FOR

NOMENCLATURE : pH METER  
MANUFACTURER : APERA  
MODEL / TYPE : PH700/201T-F  
SERIAL NO. : PH700X1019061009/N/A [PH-02]  
DATE OF CALIBRATION : 22 May 2024

#### ENVIRONMENT CONDITIONS :

Temperature :  $(25 \pm 2.5) ^\circ\text{C}$  Relative Humidity :  $(50 \pm 15) \% \text{ RH}$

#### PROCEDURE USED :

This instrument was calibrated under procedure No. WI-305-128, 238. The calibration was performed by direct measurement with Certified Reference Material (CRM) and comparison with Calibration Bath, Precision Thermometer and IPRT which maintained by the Calibration Laboratory Co., Ltd.

#### REFERENCE STANDARD USED :

1. pH Standard Solution, NIMT TRM CODE TRM-S-2003, TRM CODE TRM-S-2007.
2. pH Standard Solution, Control Company Catalog Number 06664260,11754256, Lot Number CC787362.
3. Calibration Bath, Kambic Model OB-22/2 ULT S/N. 171115653.
4. Precision Thermometer, ASL Model F200-A-8 S/N. 014433/03.
5. IPRT, ASL Model T100-250-1D S/N. L0193A-1-1.

Certificate No. Q24051980

F3-011-05/12-23

page 2 of 4



@clcalibration



NSC-TIS-TIS 17025  
CALIBRATION 0059  
CLC

# CALIBRATION LABORATORY Co., LTD.

210-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



NSC-TIS-TIS 17025  
CALIBRATION 0059  
CLC

## CERTIFICATE OF CALIBRATION

### FOR

NOMENCLATURE : pH METER  
MANUFACTURER : APERA  
MODEL / TYPE : PH700/201T-F  
SERIAL NO. : PH700X1019061009/N/A [PH-02]  
CLID. NO. : 272401000  
JOB CONTROL NO. : 240521051980  
CALIBRATION SERVICE : ☒ IN-LABORATORY ☐ ON-SITE

CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.


5/45 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BON SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 21 May 2024


DATE OF ISSUED : 23 May 2024

The report of calibration shall not be reproduced except in full without approval of the Calibration Laboratory Co., Ltd.

Calibrated By :



Approved By :



23 May 2024

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q24051980

F3-011-05/12-23

page 1 of 4



@clcalibration



CLC  
Accredited  
ISO/IEC 17025



ILAC-MRA  
NSC-TISI-TIS 17025  
CALIBRATION 0059  
CLC



CALIBRATION LABORATORY Co., LTD.  
210-11,14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



CALIBRATION LABORATORY Co., LTD.  
210-11,14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com

## CONDITION OF CALIBRATION ITEM : RECEIVED IN GOOD OPERATIONAL CONDITION

### MEASUREMENT RESULTS : ( X ) without adjustment ( ) adjustment

The table in the following gives the calibration results and associated measurement uncertainties of pH meter.

#### CALIBRATION DATA

##### 1. pH METER RESULT @ 25 °C

Standard pH Buffer Solution (pH)	pH Meter Reading (pH)	pH Meter Reading (mV)	Correction (pH)	Uncertainty of Measurement (± pH)	k Factor
4.003	4.00	180	+0.003	0.014	2.00
7.005	7.00	1	+0.005	0.014	2.00
10.015	10.01	-168	+0.005	0.100	2.00

Technical Note. Setting function CAL.3 point (4,7,10).

Note. The Scope of Accredited TISI Certificate No. 23-LB0092 Issue 02 Page 91 of 138

##### \*2. TEMPERATURE RESULT [ pH PROBE ]

Immersion depth (mm)	Actual Temperature (°C)	DUC Reading (°C)	Correction (°C)	Uncertainty ± (°C)
100	25.00	25.0	0.00	0.07

Note. Probe Ø 12 mm

\* means Calibrations marked " Not TISI Accredited " in this Certificate have been included for completeness.

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor of  $k = 2.00$ .

This report is valid for the above stated instrument/s only.

### End of Certificate ###

Certificate No. Q24051980

F3-011-05/12-23

page 4 of 4



@cccalibration

#### TRACEABILITY :

1. The measurements are traceable to International System of Units (SI), through National Institute of Metrology (Thailand).  
Lot Number. 040822, 120124, Due Date 04 March 2025.

2. The measurements are traceable to International System of Units (SI), through Control Company.

Certificate No. 4281-14495731, Due Date 27 September 2025.

3. The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd.

Certificate No. Q23136342, Due Date 20 December 2024.

4. The measurements are traceable to International System of Units (SI), through Thailand Institute of Scientific and Technological Research (TISTR). Certificate No. PSL-T 0203/67, Due Date 07 December 2024.

5. The measurements are traceable to International System of Units (SI), through National Institute of Metrology (Thailand).  
Certificate No. TT-0136-23, Due Date 12 December 2024.

#### UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.

It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2022)"

Certificate No. Q24051980

F3-011-05/12-23

page 3 of 4



@cccalibration



# CALIBRATION LABORATORY Co., LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



ISO/IEC 17025  
CALIBRATION 0659  
CLC

## REPORT OF CALIBRATION

### FOR

NOMENCLATURE : INCUBATOR  
MANUFACTURER : ACCUPLUS  
MODEL / TYPE : SMART i250  
SERIAL NO. : 2059-0718-0010[LA-002]  
LOCATION SITE : LABORATORY  
DATE OF CALIBRATION : 04 April 2024

### ENVIRONMENT CONDITIONS :

Temperature : 24 °C to 25 °C  
Relative Humidity : 50 % to 52 %

### PROCEDURE USED :

This instrument was calibrated under procedure No. WI-305-165 based on TLAS G-20-1/02-08 as calibration guidelines.  
The calibration was performed by using Hydra Data Logger which maintained by the Calibration Laboratory Co., Ltd.

### REFERENCE STANDARD USED :

Hydra Data Logger, Fluke Model 2620 S/N. 5592550.

### TRACEABILITY :

The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd.  
Certificate No. Q23065868, Due Date 22 June 2024.

### UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.  
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2022)"

Certificate No. Q24024792

F3-011-05/12-23

page 2 of 4



@cdcalibration



# CALIBRATION LABORATORY Co., LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



ISO/IEC 17025  
CALIBRATION 0659  
CLC

## CERTIFICATE OF CALIBRATION

### FOR


NOMENCLATURE : INCUBATOR  
MANUFACTURER : ACCUPLUS  
MODEL / TYPE : SMART i250  
SERIAL NO. : 2059-0718-0010[LA-002]  
CLID. NO. : 332100155  
JOB CONTROL NO. : 240307024792  
CALIBRATION SERVICE : ☐ IN-LABORATORY ☒ ON-SITE  
CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.  
5/45 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BON SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 07 March 2024

DATE OF ISSUED : 06 April 2024

The report of calibration shall not be reproduced except in full without approval of the Calibration Laboratory Co., Ltd.

Calibrated By :



Approved By :

06 April 2024

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q24024792

F3-011-05/12-23

page 1 of 4



@cdcalibration


### CALIBRATION DATA

#### 2. TEMPERATURE DISTRIBUTION

DUC	Measured Temperature ( °C ) @ Probe No. 9 is Ref.									Uncertainty ± ( °C )	Coverage factor k
	Setting ( °C )	Indicating ( °C )	1	2	3	4	5	6	7	8	9
20.0	20.0	20.58	20.70	20.70	20.70	20.46	20.54	20.53	20.53	20.36	20.39

Technical Note : W = 50 cm, D = 48 cm, H = 110 cm.

The Scope of Accredited TISI Certificate No. 23-LB0092 Issue 02 Page 129 of 138



This report is valid for the above stated instrument/s only.

### End of Certificate ###



### CONDITION OF CALIBRATION ITEM : RECEIVED IN GOOD OPERATIONAL CONDITION

#### MEASUREMENT RESULTS : ( X ) without adjustment ( ) adjustment

The table in the following gives the calibration results and associated measurement uncertainties of the measuring incubator.

#### CALIBRATION DATA

#### 1. INCUBATOR PERFORMANCE

	DUC		Measured Uniformity ( °C )	Measured Stability ( °C )	Measured Overall Variation ( °C )
	Setting ( °C )	Indicating ( °C )			
	20.0	20.0	0.52	0.46	1.06





# CALIBRATION LABORATORY Co., LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2872 www.cal-laboratory.com E-mail:sale@cal-laboratory.com

CLC  
Accredited  
ISO/IEC 17025



ANAB National Accreditation Board  
ACCREDITED  
CALIBRATION AND  
DIMENSIONAL MEASUREMENT  
ACDM-2014

## REPORT OF CALIBRATION

### FOR

NOMENCLATURE : HOT AIR OVEN  
MANUFACTURER : MEMMERT  
MODEL / TYPE : UF110  
SERIAL NO. : B422.0026[L A-0012]  
LOCATION SITE : LABORATORY-HOT ZONE  
DATE OF CALIBRATION : 04 April 2024

#### ENVIRONMENT CONDITIONS :

Temperature : 24 °C to 25 °C

Relative Humidity : 50% to 52 %

#### PROCEDURE USED :

This instrument was calibrated under procedure No. CLC-CPTH-07 based on TLAS G-20 as calibration guidelines.  
The calibration was performed by using Hydra Data Logger which maintained by the Calibration Laboratory Co., Ltd.

#### REFERENCE STANDARD USED :

Hydra Data Logger, Fluke Model 2620 S/N: 5592550.

#### TRACEABILITY :

The measurements are traceable to International System of Units (SI), through Calibration Laboratory Co., Ltd.  
Certificate No. Q23065868, Due Date 22 June 2024.

#### UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.  
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2022)"

Certificate No. Q24024791

F3-011-05/12-23

page 2 of 4



@clcalibration



# CALIBRATION LABORATORY Co., LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2872 www.cal-laboratory.com E-mail:sale@cal-laboratory.com

CLC  
Accredited  
ISO/IEC 17025



ANAB National Accreditation Board  
ACCREDITED  
CALIBRATION AND  
DIMENSIONAL MEASUREMENT  
ACDM-2814

## CERTIFICATE OF CALIBRATION

### FOR

NOMENCLATURE : HOT AIR OVEN  
MANUFACTURER : MEMMERT  
MODEL / TYPE : UF110  
SERIAL NO. : B422.0026[L A-0012]  
CLID. NO. : 332202464  
JOB CONTROL NO. : 240307024791  
CALIBRATION SERVICE : ☐ IN-LABORATORY ☒ ON-SITE

CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.


5/45 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BON SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 07 March 2024

DATE OF ISSUED : 06 April 2024

The report of calibration shall not be reproduced except in full without approval of the Calibration Laboratory Co., Ltd.

Calibrated By :



Approved By :

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q24024791

F3-011-05/12-23

page 1 of 4



@clcalibration





# CALIBRATION LABORATORY Co., LTD.

2/10-11, 14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel: 02-578-0353-4 Fax: 02-578-2672 www.ccl-laboratory.com E-mail: sale@cal-laboratory.com



Accredited  
ISO/IEC 17025



ANAB  
ACCREDITED  
CALIBRATION AND  
DIMENSIONAL MEASUREMENT  
ACDM-2814



Accredited  
ISO/IEC 17025

# CALIBRATION LABORATORY Co., LTD.

2/10-11, 14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel: 02-578-0353-4 Fax: 02-578-2672 www.ccl-laboratory.com E-mail: sale@cal-laboratory.com



ANAB  
ACCREDITED  
CALIBRATION AND  
DIMENSIONAL MEASUREMENT  
ACDM-2814


## CALIBRATION DATA

### 2. TEMPERATURE DISTRIBUTION

DUC		Measured Temperature ( °C )@Probe No.9 is Ref.									Uncertainty ± ( °C )	Coverage factor <i>k</i>
Setting ( °C )	Indicating ( °C )	1	2	3	4	5	6	7	8	9		
104.0	104.0	103.11	104.12	103.56	103.79	103.72	103.88	103.57	104.18	103.88	0.45	2.00
180.0	180.0	178.33	180.32	178.77	179.54	179.26	179.74	179.19	180.15	179.58	0.55	2.00

Technical Note : W = 56 cm, D = 40 cm, H = 48 cm.

Note. The Scope of Accredited ANAB Certificate No. ACDM-2814 Version 012 Page 58 of 67



This report is valid for the above stated instrument/s only.

### End of Certificate ###

Certificate No. Q24024791

F3-011-05/12-23

Certificate No. Q24024791

F3-011-05/12-23

page 4 of 4



@cclcalibration

page 3 of 4



@cclcalibration



Accredited  
ISO/IEC 17025

## CALIBRATION LABORATORY CO., LTD.

2/10-11, 14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail: sale@cal-laboratory.com



ANAB Accredited  
A C C R E D I T E D  
CALIBRATION AND  
DIMENSIONAL MEASUREMENT  
ACDM-2814

### REPORT OF CALIBRATION

#### FOR

NOMENCLATURE : DO METER  
MANUFACTURER : YSI  
MODEL / TYPE : 5000-230V/5010  
SERIAL NO. : 16D101626/19D100367[DOM-M-01]  
DATE OF CALIBRATION : 09 April 2024

#### ENVIRONMENT CONDITIONS :

Temperature :  $(25 \pm 2.5) ^\circ\text{C}$  Relative Humidity :  $(50 \pm 15) \% \text{RH}$

#### PROCEDURE USED :

This instrument was calibrated under procedure No. CLC-CPCH-06. The calibration was performed by direct measurement with Certified Reference Material (CRM).

#### REFERENCE STANDARD USED :

Dissolved Oxygen, Sigma-Aldrich Product ID QC3077-500ML.

#### TRACEABILITY :

The measurements are traceable to International System of Units (SI), through Merck Co., Ltd.  
Lot LRAD3840, Due Date October 2024.

#### UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k = 2$ , which for a normal distribution corresponds to a coverage probability of approximately 95 %.  
It has been evaluated according to the "Evaluation of the Uncertainty of Measurement (EA-4/02 M:2022)"

Certificate No. Q24038371

F3-011-05/12-23

page 2 of 3



@clccalibration



Accredited  
ISO/IEC 17025

## CALIBRATION LABORATORY CO., LTD.

2/10-11, 14, 55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail: sale@cal-laboratory.com



ANAB Accredited  
A C C R E D I T E D  
CALIBRATION AND  
DIMENSIONAL MEASUREMENT  
ACDM-2814

### CERTIFICATE OF CALIBRATION

#### FOR

NOMENCLATURE : DO METER  
MANUFACTURER : YSI  
MODEL / TYPE : 5000-230V/5010  
SERIAL NO. : 16D101626/19D100367[DOM-M-01]  
CLID. NO. : 272100329  
JOB CONTROL NO. : 240408038371  
CALIBRATION SERVICE : ☒ IN-LABORATORY ☐ ON-SITE

CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.

5/45 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BON SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 08 April 2024

DATE OF ISSUED : 10 April 2024

The report of calibration shall not be reproduced except in full without approval of the Calibration Laboratory Co., Ltd.

Calibrated By :



Approved By :

10 April 2024

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q24038371

F3-011-05/12-23

page 1 of 3



@clccalibration



**CALIBRATION LABORATORY Co., LTD.**  
2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



## CERTIFICATE OF CALIBRATION

### FOR

NOMENCLATURE : BALANCE  
MANUFACTURER : SHIMADZU  
MODEL / TYPE : AP225WD  
SERIAL NO. : D316300692[LA-001]  
CLID. NO. : 362100172  
JOB CONTROL NO. : 240307024789  
CALIBRATION SERVICE : ☐ IN-LABORATORY ☒ ON-SITE

CUSTOMER : ENVIRONMENTAL MEASUREMENTS CO., LTD.  
5/45 BAAN KLANG KRUNG BIZ TOWN, SOI SRINAGARINDRA 46/1 (PRAMOTE),  
NONG BON SUB-DISTRICT, PRAWET DISTRICT, BANGKOK 10250

DATE OF RECEIVED : 07 March 2024 DATE OF ISSUED : 05 April 2024

The report of calibration shall not be reproduced except in full without approval of the Calibration Laboratory Co., Ltd.

Calibrated By :



Approved By :

This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q24024789  
F3-011-05/12-23

page 1 of 3



@cccalibration



**CALIBRATION LABORATORY Co., LTD.**  
2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230  
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



CONDITION OF CALIBRATION ITEM : RECEIVED IN GOOD OPERATIONAL CONDITION  
MEASUREMENT RESULTS : (X) without adjustment ( ) adjustment

The table in the following gives the calibration results and associated measurement uncertainties of Do Meter.

### CALIBRATION DATA

#### CORRECTION OF DO METER @ 20°C

Nominal Value (mg/L)	DUC Reading (mg/L)	Correction (mg/L)	Uncertainty (mg/L)
7.78	7.80	-0.02	± 0.38

Note: The Scope of Accredited ANAB Certificate No. ACDM-2814 Version 012 Page 5 of 67

This report is valid for the above stated instrument/s only.

### End of Certificate ###

Certificate No. Q24038371  
F3-011-05/12-23

page 3 of 3

u.6/20



**CONDITION OF CALIBRATION ITEM : RECEIVED IN GOOD OPERATIONAL CONDITION**  
**MEASUREMENT RESULTS : ( X ) without adjustment ( ) adjustment**  
**CALIBRATION DATA**

**1. Error of indications**

Nominal Test Value ( g )	Conventional mass ( g )	Display Value ( g )	Error of Balance ( g )	Uncertainty $\pm$ ( mg )	Coverage factor k
Unload	0.0000	0.0000	0.0000	0.07	2.00
5.0000	5.0000	5.0000	0.0000	0.11	2.00
10.0000	10.0000	10.0001	+0.0001	0.11	2.00
20.0000	20.0000	20.0001	+0.0001	0.11	2.00
40.0000	40.0000	40.0001	+0.0001	0.13	2.00
60.0000	60.0000	60.0001	+0.0001	0.17	2.00
80.0000	80.0000	80.0002	+0.0002	0.17	2.00
100.0000	100.0000	100.0001	+0.0001	0.17	2.00
120.0000	120.0000	120.0001	+0.0001	0.29	2.00
140.0000	140.0000	140.0002	+0.0002	0.29	2.00
160.0000	160.0000	160.0002	+0.0002	0.29	2.00
180.0000	180.0000	180.0001	+0.0001	0.29	2.00
200.0000	200.0001	200.0001	0.0000	0.29	2.00

**2. Repeatability of indications**

Nominal Test Value ( g )	Standard Deviation of Reading ( g )
200.0000	0.00005


**3. Effect of eccentric application of a load on the indication**

<input type="checkbox"/>					
	<input checked="" type="checkbox"/>				
Nominal Test Value ( g )	Display Value ( g )				
100.0000	Position 1	Position 2	Position 3	Position 4	Position 5
	100.0001	100.0001	100.0002	100.0002	100.0002
Maximum Difference of Center Value ( g )					
0.0001					

Note. The Scope of Accredited TISI Certificate No. 23-LB0092 Issue 02 Page 116 of 138

This report is valid for the above stated instrument/s only.

### End of Certificate ###



**REPORT OF CALIBRATION**

**FOR**

NOMENCLATURE	:	BALANCE
MANUFACTURER	:	SHIMADZU
MODEL / TYPE	:	AP225WD
SERIAL NO.	:	D316300692[LA-001]
LOCATION SITE	:	LABORATORY-BALANCE ROOM
DATE OF CALIBRATION	:	04 April 2024

**ENVIRONMENT CONDITIONS :**

Temperature : 23 °C to 24 °C

Relative Humidity : 49 % to 50 %

**PROCEDURE USED :**

This instrument was calibrated under procedure No. **WI-305-46** based on **EURAMET/eg-18/Version 4.0 (11/2015)**.

The calibration was performed by Comparison with Weight Set which maintained by the Calibration Laboratory Co., Ltd.

**REFERENCE STANDARD USED :**

1. Weight Set, Phoenix Class E2 S/N. WBS-SET-E2-01.
2. Weight, Sartorius Class E2 S/N. 44329129, 43529037, 44329167, 43529293.

**TRACEABILITY :**

1. The measurements are traceable to International System of Units (SI), through National Institute of Metrology (Thailand). Certificate No. MM-01/23-22, Due Date 22 August 2024.
2. The measurements are traceable to International System of Units (SI), through Sartorius Lab Instruments GmbH & Co. KG. Certificate No. M141607, M141608, M141609, M141611. Due Date 15 September 2025.

**UNCERTAINTY :**

The reported expanded uncertainty of measurement is stated as the standard uncertainty multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95%. It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2022)"

Certificate No. Q24024789

F3-011-05/12-23

