

ภาคผนวกที่ 5

เอกสารสอบเทียบความถูกต้องของเครื่องมือ

- | | | |
|-----------|-----|---|
| เอกสารแนบ | 5-1 | เอกสารสอบเทียบเครื่องมือการตรวจวัดคุณภาพอากาศในบรรยากาศ |
| เอกสารแนบ | 5-2 | เอกสารสอบเทียบเครื่องมือการตรวจวิเคราะห์คุณภาพน้ำ |

ตารางสรุปรายการเอกสารการสอบเทียบความถูกต้องของเครื่องมือเก็บตัวอย่างและตรวจวิเคราะห์คุณภาพสิ่งแวดล้อม

รายการตรวจวัด	เครื่องมือเก็บตัวอย่าง	เครื่องมือตรวจวิเคราะห์
	ชื่อเครื่องมือ	ชื่อเครื่องมือ
คุณภาพอากาศในบรรยากาศ TSP	High Volume Air Sampler No. B21, B35, B37, B38, B40, B42, B43	Digital Balance
CO	CO Analyzer No. B01, B03, B06, B10, B11, B12, B15	CO Analyzer No. B01, B03, B06, B10, B11, B12, B15
THC	Personal Pump SKC No. B10, B21, B40, B44, B48, B53, B70 Rotameter No. H-B01	THC Analyzer No. B01
SO ₂	SO ₂ Analyzer No. B01, B02, B04, B09, B10, B11, B13	SO ₂ Analyzer No. B01, B02, B04, B09, B10, B11, B13
NO ₂	NO Analyzer No. B11, B12, B14, B15, B18, B19, R07	NO Analyzer No. B11, B12, B14, B15, B18, B19, R07
Lead	High Volume Air Sampler No. B21, B35, B37, B38, B40, B42, B43	ICP
คุณภาพน้ำใต้ดิน Lead	-	AAS

เอกสารแนบ 5-1

เอกสารสอบเทียบเครื่องมือการตรวจวัดคุณภาพอากาศในบรรยากาศ



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
Tel : (662) 939-4370-72 Fax : (662) 513-4221 E-mail : sales@spsc.com www.spsc.com

High Volume Air Sampler Calibration Report

Calibration Method : Multipoint Orifice Flow Transfer Standard Model : TE 5025A S/N : 3611

Calibration Data

High Volume Air Sampler Data		Calibration Data		
Recorder No.	Blower No.	Date	Actual Flowrate (ft ³ /min)	R ²
B01	B01	05/02/2025	y = 1.190x-4.759	0.999
B02	B02	05/02/2025	y = 1.167x-1.802	0.999
B03	B03	03/02/2025	y = 1.142x-3.352	0.997
B04	B04	06/02/2025	y = 1.160x-3.139	0.998
B05	B05	06/02/2025	y = 1.155x-5.601	0.996
B06	B06	06/02/2025	y = 1.150x-1.476	0.999
B07	B07	03/02/2025	y = 1.143x-3.035	0.998
B08	B08	03/02/2025	y = 1.161x-4.459	0.999
B09	B09	05/02/2025	y = 1.177x-3.970	0.996
B10	B10	03/02/2025	y = 1.144x-2.471	0.998
B11	B11	03/02/2025	y = 1.195x-5.384	0.996
B12	B12	04/02/2025	y = 1.168x-4.228	0.998
B13	B13	04/02/2025	y = 1.165x-3.801	0.999
B14	B14	04/02/2025	y = 1.148x-3.248	0.996
B15	B15	04/02/2025	y = 1.173x-4.773	0.997
B16	B16	04/02/2025	y = 1.156x-4.042	0.998
B17	B17	06/02/2025	y = 1.140x-2.730	0.999
B18	B18	06/02/2025	y = 1.171x-4.178	0.999
B19	B19	06/02/2025	y = 1.151x-3.979	0.999
B20	B20	04/02/2025	y = 1.129x-1.255	0.999
B21	B21	04/02/2025	y = 1.132x-3.156	0.999
B22	B22	04/02/2025	y = 1.147x-2.649	0.997
B23	B23	03/02/2025	y = 1.158x-3.223	0.999
B24	B24	05/02/2025	y = 1.144x-3.476	0.997
B25	B25	03/02/2025	y = 1.071x-1.478	0.997
B26	B26	04/02/2025	y = 1.142x-4.263	0.999
B27	B27	04/02/2025	y = 1.175x-5.192	0.996
B28	B28	04/02/2025	y = 1.173x-5.127	0.999
B29	B29	04/02/2025	y = 1.145x-1.952	0.996
B30	B30	06/02/2025	y = 1.162x-3.062	0.999
B31	B31	03/02/2025	y = 1.182x-5.652	0.998
B32	B32	03/02/2025	y = 1.167x-3.993	0.999
B33	B33	05/02/2025	y = 1.168x-4.451	0.998
B34	B34	05/02/2025	y = 1.127x-3.203	0.999

Calibrated by : Adul Dangklom
(Mr. Adul Dangklom)

Approved by : Peera Detudom
(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
Tel : (662) 939-4370-72 Fax : (662) 513-4221 E-mail : sales@spsc.com www.spsc.com

High Volume Air Sampler Calibration Report

Calibration Method : Multipoint Orifice Flow Transfer Standard Model : TE 5025A S/N : 3611

Calibration Data

High Volume Air Sampler Data		Calibration Data		
Recorder No.	Blower No.	Date	Actual Flowrate (ft ³ /min)	R ²
B35	B35	05/02/2025	y = 1.163x-3.579	0.997
B36	B36	05/02/2025	y = 1.130x-2.116	0.999
B37	B37	04/02/2025	y = 1.146x-2.265	0.996
B38	B38	04/02/2025	y = 1.156x-6.034	0.998
B39	B39	03/02/2025	y = 1.151x-3.366	0.998
B40	B40	03/02/2025	y = 1.174x-4.582	0.999
B41	B41	06/02/2025	y = 1.123x-1.633	0.997
B42	B42	03/02/2025	y = 1.149x-3.382	0.997
B43	B43	03/02/2025	y = 1.137x-2.074	0.997
B44	B44	03/02/2025	y = 1.155x-1.460	0.999
R01	R01	04/02/2025	y = 1.121x-3.007	0.999
R02	R02	03/02/2025	y = 1.159x-5.099	0.999
R03	R03	05/02/2025	y = 1.138x-2.774	0.998
R04	R04	05/02/2025	y = 1.118x-2.575	0.999
R05	R05	03/02/2025	y = 1.136x-1.720	0.998
R06	R06	05/02/2025	y = 1.154x-2.706	0.997
R07	R07	03/02/2025	y = 1.037x+1.361	0.999
R08	R08	03/02/2025	y = 1.146x-3.762	0.996
R09	R09	05/02/2025	y = 1.121x-2.360	0.997
R10	R10	05/02/2025	y = 1.180x-4.626	0.999
R11	R11	05/02/2025	y = 1.147x-3.861	0.996
R12	R12	03/02/2025	y = 1.128x-4.676	0.998
R13	R13	04/02/2025	y = 1.135x-4.055	0.999
R14	R14	04/02/2025	y = 1.153x-3.122	0.997
R15	R15	03/02/2025	y = 1.161x-5.223	0.998
R16	R16	03/02/2025	y = 1.187x-6.674	0.999
R17	R17	03/02/2025	y = 1.120x-1.730	0.999
R18	R18	03/02/2025	y = 1.146x-2.347	0.998
R19	R19	06/02/2025	y = 1.161x-5.195	0.999
R20	R20	06/02/2025	y = 1.134x-3.449	0.998

Calibrated by : Adul Dangklom
(Mr. Adul Dangklom)

Approved by : Peera Detudom
(Mr. Peera Detudom)



CERTIFICATE No : 25M2254
REFERENCE No : 76365-1

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : DIGITAL BALANCE
MANUFACTURER : METTLER TOLEDO
MODEL : XS105DU
SERIAL No : 1126422905
ID No : BA05/50
CONDITION AS RECEIVED : USED ITEM
SUBMITTED BY : S.P.S. CONSULTING SERVICE CO., LTD.
7 SOI PHAHOLYOTHIN 24, PHAHOLYOTHIN RD.,
JOMPOL, CHATUCHAK, BANGKOK 10900

CALIBRATED BY : ATSAWIN Y.
CALIBRATION DATE : 07-Mar-25

APPROVED BY : PONGSAK J.
ISSUED DATE : 13-Mar-25
RECEIVED DATE : 07-Mar-25

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.

F-G010 REV 03



CERTIFICATE No : 25M2254

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : DIGITAL BALANCE MODEL : XS105DU
MANUFACTURER : METTLER TOLEDO S/N : 1126422905
ID No : BA05/50 RECEIVED DATE : 07-Mar-25
AIR PRESSURE : 1009mbar \pm 1mbar CALIBRATION DATE : 07-Mar-25
AMBIENT TEMPERATURE : 24°C \pm 1°C RELATIVE HUMIDITY : 54%RH \pm 10%RH

CONDITION OF THIS RESULTS OF CALIBRATION

1. THIS INSTRUMENT WAS CALIBRATED BY ACCORDING TO UKAS LAB 14 EDITION 6:2019 BY USING KNOWN WEIGHT STANDARD WEIGHT. THE BALANCE WAS NOT ADJUSTED BEFORE CALIBRATION. THE BALANCE HAS NO ZERO TRACKING FUNCTION. REPEATABILITY WAS MEASURED BY USING 10 REPEATED MEASUREMENTS. LINEARITY WAS MEASURED COVERING 10 POINTS, EVENLY SPREAD OVER THE RANGE. THE INSTRUMENT WAS SET ZERO BEFORE PERFORMING THE LINEARITY TEST. OFF-CENTER LOADING WAS MEASURED BY USING STANDARD WEIGHTS PLACED ON THE PAN AND MOVED TO VARIOUS POSITIONS ON THE PAN.

2. REFERENCE STANDARD INSTRUMENTS :-

INSTRUMENT	MODEL	SERIAL No	CERTIFICATE No	DUE DATE
1) STANDARD WEIGHT SET	E2	QK-I-151	C02250116	28-Jan-27
2) STANDARD WEIGHT	E2	15843	C02250117	29-Jan-27

3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.

4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.

5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-
- NATIONAL INSTITUTE OF METROLOGY (THAILAND)

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT

- ZERO SETTING FUNCTION : NORMAL
- TARE FUNCTION : NORMAL
- REPEATABILITY OF READING AT 120 g WAS 0.000055 g
- DEPARTURE FROM NOMINAL VALUE/ LINEARITY

NOMINAL VALUE (g)	BALANCE READING (g)	CORRECTION (g)	UNCERTAINTY (\pm g)
0.00	0.00000	0.00000	0.000065
0.02	0.01999	0.00001	0.000065
0.10	0.10001	-0.00001	0.000066
0.20	0.20001	-0.00001	0.000066
0.50	0.50002	-0.00002	0.000065
1.00	1.00003	-0.00003	0.000066
2.00	2.00001	-0.00001	0.000067
5.00	5.00002	-0.00002	0.000068
10.00	10.00000	0.00000	0.000070
20.00	20.00004	-0.00004	0.000078
50.00	50.00000	0.00000	0.00013
100.00	100.0001	-0.0001	0.00019
120.00	120.0002	-0.0002	0.00022

5. OFF CENTER LOADING ERROR



POINT	READING (g)
1	50.0000
2	50.0000
3	50.0000
4	50.0000
5	50.0000
OFF-CENTER LOADING	0.0000

NOTE: THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA
THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR $k=2$, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.

END OF CALIBRATION REPORT



F-G010 REV 03

Calibration Report				
Non-Dispersive Infrared CO Analyzer				
Date :	17 April 2025	Brand :	API	Model :
No.	CO-B01	Serial No.	782	
Calibrator (Dilution System)				
Brand : Teledyne		Model : 700		
Last Cal. Date : 29 October 2024		Serial No. : 421		
Reference Standard Gas				
Standard Gas : Carbon Monoxide (CO)		Cylinder No. : D711839		
Certified Date : 14 March 2024	Expired Date : 14 March 2032	Cylinder Conc. : 4,580 ppm		
Calibrating Condition				
Pressure	1011 mmbar	Temp.	24.5 °C	% RH
Calibration Setting				
Span	Initial Reading (Before Adj.), PPM		Final Reading (After Adj.), PPM	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response
Zero	0	0.11	-	0
CO Span	40.00	40.04	0.100	40.00
API Model 300E CO Analyzer Check List				
Parameter	Observed Value	Units	Nominal Range	
Range	50	PPM	0-1000 ppm	
Stability	0.10	PPM	< 1 ppm With Zero Air	
CO Measure	4014.2	mV	2500-4800 mV	
CO Reference	3948.5	mV	2500-4800 mV	
Measure/Reference Ratio	1.179	-	1.1-1.3 W/Zero Air	
Sample Pressure	28.5	In-Hg-A	~2"± Ambient Absolute Pressure	
Sample Flow	804	CC/Min	800 ± 10%	
Sample Temperature	48.3	°C	48 ± 4	
Bench Temperature	48.1	°C	48 ± 2	
Wheel Temperature	68.2	°C	68 ± 2	
Box Temperature	30.9	°C	Ambient Temp + 7 ± 10	
Photo-Drive	3032.4	mV	250 mV to 4750 mV	
Slope	1.017	-	1.0 ± 0.3	
Offset	0.2	-	0 ± 0.3	

Calibrated by : Adul Dangklorn
(Mr. Adul Dangklorn)

Approved by : Peera Detudom
(Mr. Peera Detudom)

Calibration Report				
Non-Dispersive Infrared CO Analyzer				
Date :	17 April 2025	Brand :	API	Model :
No.	CO-B03	Serial No.	3019	
Calibrator (Dilution System)				
Brand : Teledyne		Model : 700		
Last Cal. Date : 29 October 2024		Serial No. : 421		
Reference Standard Gas				
Standard Gas : Carbon Monoxide (CO)		Cylinder No. : D711839		
Certified Date : 14 March 2024	Expired Date : 14 March 2032	Cylinder Conc. : 4,580 ppm		
Calibrating Condition				
Pressure	1011 mmbar	Temp.	24.5 °C	% RH
Calibration Setting				
Span	Initial Reading (Before Adj.), PPM		Final Reading (After Adj.), PPM	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response
Zero	0	0.10	-	0
CO Span	40.00	39.96	-0.100	40.00
API Model 300E CO Analyzer Check List				
Parameter	Observed Value	Units	Nominal Range	
Range	50	PPM	0-1000 ppm	
Stability	0.10	PPM	< 1 ppm With Zero Air	
CO Measure	4017.1	mV	2500-4800 mV	
CO Reference	3948.9	mV	2500-4800 mV	
Measure/Reference Ratio	1.180	-	1.1-1.3 W/Zero Air	
Sample Pressure	28.6	In-Hg-A	~2"± Ambient Absolute Pressure	
Sample Flow	809	CC/Min	800 ± 10%	
Sample Temperature	48.2	°C	48 ± 4	
Bench Temperature	48.0	°C	48 ± 2	
Wheel Temperature	68.4	°C	68 ± 2	
Box Temperature	30.7	°C	Ambient Temp + 7 ± 10	
Photo-Drive	3043.2	mV	250 mV to 4750 mV	
Slope	1.017	-	1.0 ± 0.3	
Offset	0.2	-	0 ± 0.3	

Calibrated by : Adul Dangklorn
(Mr. Adul Dangklorn)

Approved by : Peera Detudom
(Mr. Peera Detudom)

Calibration Report			
Non-Dispersive Infrared CO Analyzer			
Date :	17 April 2025	Brand :	API
No.	CO-B06	Model :	300E
		Serial No.	3117
Calibrator (Dilution System)			
Brand : Teledyne		Model : 700	
Last Cal. Date : 29 October 2024		Serial No. : 421	
Reference Standard Gas			
Standard Gas : Carbon Monoxide (CO)		Cylinder No. : D711839	
Certified Date : 14 March 2024	Expired Date : 14 March 2032	Cylinder Conc. : 4,580 ppm	
Calibrating Condition			
Pressure	1011 mmbar	Temp.	24.5 °C
		% RH	49
Calibration Setting			
Span	Initial Reading (Before Adj.), PPM		Final Reading (After Adj.), PPM
Set Point	Expected Concentration	Analyzer Response	%Diff
Zero	0	0.10	-
CO Span	40.00	40.07	0.175
API Model 300E CO Analyzer Check List			
Parameter	Observed Value	Units	Nominal Range
Range	50	PPM	0-1000 ppm
Stability	0.10	PPM	< 1 ppm With Zero Air
CO Measure	4013.4	mV	2500-4800 mV
CO Reference	3946.9	mV	2500-4800 mV
Measure/Reference Ratio	1.180	-	1.1-1.3 W/Zero Air
Sample Pressure	28.4	In-Hg-A	~2" x Ambient Absolute Pressure
Sample Flow	810	CC/Min	800 ± 10%
Sample Temperature	48.5	°C	48 ± 4
Bench Temperature	48.2	°C	48 ± 2
Wheel Temperature	68.3	°C	68 ± 2
Box Temperature	30.6	°C	Ambient Temp + 7 ± 10
Photo-Drive	3025.9	mV	250 mV to 4750 mV
Slope	1.017	-	1.0 ± 0.3
Offset	0.2	-	0 ± 0.3

Calibrated by : Adul Dangklom
(Mr. Adul Dangklom)

Approved by : (Signature)
(Mr. Peera Detudom)

Calibration Report			
Non-Dispersive Infrared CO Analyzer			
Date :	17 April 2025	Brand :	API
No.	CO-B10	Model :	300E
		Serial No.	199-S
Calibrator (Dilution System)			
Brand : Teledyne		Model : 700	
Last Cal. Date : 29 October 2024		Serial No. : 421	
Reference Standard Gas			
Standard Gas : Carbon Monoxide (CO)		Cylinder No. : D711839	
Certified Date : 14 March 2024	Expired Date : 14 March 2032	Cylinder Conc. : 4,580 ppm	
Calibrating Condition			
Pressure	1011 mmbar	Temp.	24.5 °C
		% RH	49
Calibration Setting			
Span	Initial Reading (Before Adj.), PPM		Final Reading (After Adj.), PPM
Set Point	Expected Concentration	Analyzer Response	%Diff
Zero	0	-0.10	-
CO Span	40.00	39.95	-0.125
API Model 300E CO Analyzer Check List			
Parameter	Observed Value	Units	Nominal Range
Range	50	PPM	0-1000 ppm
Stability	0.10	PPM	< 1 ppm With Zero Air
CO Measure	4016.9	mV	2500-4800 mV
CO Reference	3949.3	mV	2500-4800 mV
Measure/Reference Ratio	1.179	-	1.1-1.3 W/Zero Air
Sample Pressure	28.7	In-Hg-A	~2" x Ambient Absolute Pressure
Sample Flow	808	CC/Min	800 ± 10%
Sample Temperature	48.3	°C	48 ± 4
Bench Temperature	48.1	°C	48 ± 2
Wheel Temperature	68.5	°C	68 ± 2
Box Temperature	30.8	°C	Ambient Temp + 7 ± 10
Photo-Drive	3023.7	mV	250 mV to 4750 mV
Slope	1.017	-	1.0 ± 0.3
Offset	0.2	-	0 ± 0.3

Calibrated by : Adul Dangklom
(Mr. Adul Dangklom)


Approved by : (Signature)
(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Calibration Report					
Non-Dispersive Infrared CO Analyzer					
Date :	17 April 2025	Brand :	Thermo	Model :	48C
No.	CO-811	Serial No.	401304262		
Calibrator (Dilution System)					
Brand :	Teledyne	Model :	700		
Last Cal. Date :	29 October 2024	Serial No. :	421		
Reference Standard Gas					
Standard Gas :	Carbon Monoxide (CO)	Cylinder No. :	D711839		
Certified Date :	14 March 2024	Expired Date :	14 March 2032	Cylinder Conc. :	4,580 ppm
Calibrating Condition					
Pressure	1011	mmbar	Temp.	24.5	°C
% RH					49
Calibration Setting					
Span	Initial Reading (Before Adj.),PPM			Final Reading (After Adj.),PPM	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response	
Zero	0	-0.10	-	0	
CO Span	40.00	39.93	-0.175	40.00	
Instrument Status					
Chamber Temp	47.2	°C	Flow	1.5 LPM	
Pressure	730.7	mm Hg	Motor Speed	100.00%	

Calibrated by : Adul Dangklom
(Mr.Adul Dangklom)


Approved by : 
(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Calibration Report					
Non-Dispersive Infrared CO Analyzer					
Date :	17 April 2025	Brand :	Thermo	Model :	48C
No.	CO-812	Serial No.	TL-65343-348		
Calibrator (Dilution System)					
Brand :	Teledyne	Model :	700		
Last Cal. Date :	29 October 2024	Serial No. :	421		
Reference Standard Gas					
Standard Gas :	Carbon Monoxide (CO)	Cylinder No. :	D711839		
Certified Date :	14 March 2024	Expired Date :	14 March 2032	Cylinder Conc. :	4,580 ppm
Calibrating Condition					
Pressure	1011	mmbar	Temp.	24.5	°C
% RH					49
Calibration Setting					
Span	Initial Reading (Before Adj.),PPM			Final Reading (After Adj.),PPM	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response	
Zero	0	0.10	-	0	
CO Span	40.00	40.05	0.125	40.00	
Instrument Status					
Chamber Temp	47.4	°C	Flow	1.5 LPM	
Pressure	730.6	mm Hg	Motor Speed	100.00%	

Calibrated by : Adul Dangklom
(Mr.Adul Dangklom)

Approved by : 
(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจตุจักร เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Calibration Report					
Non-Dispersive Infrared CO Analyzer					
Date :	17 April 2025	Brand :	API		
No.	CO-B15	Model :	300E		
		Serial No.	226		
Calibrator (Dilution System)					
Brand	: Teledyne		Model	: 700	
Last Cal. Date	: 29 October 2024		Serial No.	: 421	
Reference Standard Gas					
Standard Gas	: Carbon Monoxide (CO)		Cylinder No.	: D711839	
Certified Date	: 14 March 2024	Expired Date	: 14 March 2032	Cylinder Conc.	: 4,580 ppm
Calibrating Condition					
Pressure	1011	mmbar	Temp.	24.5	°C
			% RH	49	
Calibration Setting					
Span	Initial Reading (Before Adj.), PPM			Final Reading (After Adj.), PPM	
Set Point	Expected Concentration	Analyzer Response	%Diff	Analyzer Response	
Zero	0	-0.10	-	0	
CO Span	40.00	39.97	-0.075	40.00	
API Model 300E CO Analyzer Check List					
Parameter	Observed Value	Units	Nominal Range		
Range	50	PPM	0-1000 ppm		
Stability	0.10	PPM	< 1 ppm With Zero Air		
CO Measure	4015.6	mV	2500-4800 mV		
CO Reference	3947.2	mV	2500-4800 mV		
Measure/Reference Ratio	1.180	-	1.1-1.3 W/Zero Air		
Sample Pressure	28.6	In-Hg-A	~2"± Ambient Absolute Pressure		
Sample Flow	807	CC/Min	800 ± 10%		
Sample Temperature	48.5	°C	48 ± 4		
Bench Temperature	48.2	°C	48 ± 2		
Wheel Temperature	68.3	°C	68 ± 2		
Box Temperature	30.9	°C	Ambient Temp + 7 ± 10		
Photo-Drive	3030.5	mV	250 mV to 4750 mV		
Slope	1.017	-	1.0 ± 0.3		
Offset	0.2	-	0 ± 0.3		

Calibrated by : Adul Dangklom
(Mr. Adul Dangklom)

Approved by : 
(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจตุจักร เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jomjol, Chatchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Personal Pump Calibration Report

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

Environmental Conditions

Temperature : 25 ± 3 °C
Pressure : 1010 ± 15 mmbar

Personal Pump Data				Calibration Data									
No.	Brand	Model	Serial No.	Date	Flow Rate (mL/min)						Value From Calibration Curve		
					Setting			Actual (Q std.)					
					1	2	3	1	2	3		R ²	
B01	SKC	224-PCXR4	262101	01/04/2025	1,000	1,500	2,000	998	1,494	2,002	1.001x - 3.594	1.000	
B02	SKC	224-PCXR4	626166	01/04/2025	1,000	1,500	2,000	995	1,508	2,003	1.009x - 12.605	1.000	
B03	SKC	224-PCXR4	612968	03/04/2025	1,000	1,500	2,000	1,003	1,502	2,006	1.006x - 7.796	1.000	
B04	SKC	224-PCXR4	602804	03/04/2025	1,000	1,500	2,000	1,001	1,499	2,013	1.004x - 7.040	0.999	
B05	SKC	224-PCXR4	612693	03/04/2025	1,000	1,500	2,000	1,004	1,498	2,007	1.003x - 2.455	1.000	
B06	SKC	224-PCXR4	262188	03/04/2025	1,000	1,500	2,000	1,003	1,512	2,004	1.000x + 6.696	1.000	
B07	SKC	224-PCXR4	626262	03/04/2025	1,000	1,500	2,000	1,012	1,504	1,996	0.994x + 10.330	0.999	
B08	SKC	224-PCXR4	626100	02/04/2025	1,000	1,500	2,000	996	1,511	2,007	1.010x - 14.048	1.000	
B09	SKC	224-PCXR4	626479	03/04/2025	1,000	1,500	2,000	999	1,510	2,003	1.003x - 4.677	1.000	
B10	SKC	224-PCXR4	091950	03/04/2025	1,000	1,500	2,000	1,002	1,498	2,004	1.004x - 6.548	1.000	
B11	SKC	224-PCXR8	568315	04/04/2025	1,000	1,500	2,000	1,013	1,505	2,010	1.002x + 2.171	1.000	
B12	SKC	224-PCXR4	034656	04/04/2025	1,000	1,500	2,000	1,004	1,506	2,009	1.008x - 9.391	1.000	
B13	SKC	224-PCXR4	602073	03/04/2025	1,000	1,500	2,000	1,001	1,497	2,012	1.009x - 9.643	1.000	
B14	SKC	224-PCXR4	626313	03/04/2025	1,000	1,500	2,000	1,004	1,515	1,997	1.002x - 1.275	0.999	
B15	SKC	224-PCXR4	626474	03/04/2025	1,000	1,500	2,000	999	1,497	1,996	1.000x - 2.511	1.000	
B16	SKC	224-PCXR4	626477	03/04/2025	1,000	1,500	2,000	1,012	1,504	2,007	0.997x + 8.160	1.000	
B17	SKC	224-PCXR4	626860	01/04/2025	1,000	1,500	2,000	997	1,506	1,999	1.001x - 0.455	1.000	
B18	SKC	224-PCXR4	691484	02/04/2025	1,000	1,500	2,000	1,007	1,493	2,005	0.998x + 4.350	1.000	
B19	SKC	224-PCXR4	691599	03/04/2025	1,000	1,500	2,000	1,004	1,513	2,001	1.003x - 2.043	1.000	
B20	SKC	224-PCXR4	691587	03/04/2025	1,000	1,500	2,000	999	1,504	1,998	0.999x + 0.556	1.000	
B21	SKC	224-PCXR4	691531	03/04/2025	1,000	1,500	2,000	1,004	1,499	1,997	1.003x - 7.572	0.999	
B22	SKC	224-PCXR4	691654	04/04/2025	1,000	1,500	2,000	1,008	1,504	2,006	1.005x - 4.941	1.000	
B23	SKC	224-PCXR4	798393	04/04/2025	1,000	1,500	2,000	995	1,499	1,998	1.002x - 4.953	1.000	
B24	SKC	224-PCXR4	626363	04/04/2025	1,000	1,500	2,000	1,002	1,501	1,996	0.999x - 1.539	1.000	
B25	SKC	224-PCXR4	798489	04/04/2025	1,000	1,500	2,000	1,010	1,515	2,001	0.990x + 16.203	0.999	
B26	SKC	224-PCXR8	798479	03/04/2025	1,000	1,500	2,000	999	1,492	1,999	0.990x - 0.596	1.000	
B27	SKC	224-PCXR4	691673	03/04/2025	1,000	1,500	2,000	996	1,498	2,002	1.004x - 6.496	1.000	
B28	SKC	224-PCXR4	691570	03/04/2025	1,000	1,500	2,000	1,004	1,499	1,994	0.993x + 8.068	1.000	
B29	SKC	224-PCXR4	626472	03/04/2025	1,000	1,500	2,000	1,001	1,501	1,996	0.994x + 9.367	1.000	
B30	SKC	224-PCXR4	691489	01/04/2025	1,000	1,500	2,000	996	1,507	2,003	1.006x - 12.489	1.000	
B31	SKC	224-PCXR4	691509	02/04/2025	1,000	1,500	2,000	1,010	1,509	2,008	1.003x - 3.858	0.999	
B32	SKC	224-PCXR4	091567	02/04/2025	1,000	1,500	2,000	996	1,497	1,997	0.998x - 0.764	1.000	
B33	SKC	224-PCXR4	091756	03/04/2025	1,000	1,500	2,000	998	1,508	2,006	1.007x - 13.441	0.999	
B34	SKC	224-PCXR4	612962	03/04/2025	1,000	1,500	2,000	1,004	1,494	1,995	0.993x + 8.471	1.000	
B35	SKC	224-PCXR4	602682	03/04/2025	1,000	1,500	2,000	999	1,491	2,001	1.000x - 2.275	1.000	
B36	SKC	224-PCXR4	626104	03/04/2025	1,000	1,500	2,000	1,004	1,499	1,997	0.995x + 5.109	1.000	
B37	SKC	224-PCXR4	626256	03/04/2025	1,000	1,500	2,000	998	1,503	1,996	0.996x + 5.729	1.000	
B38	SKC	224-PCXR4	626167	03/04/2025	1,000	1,500	2,000	996	1,509	2,004	1.008x - 15.248	0.999	
B39	SKC	224-PCXR4	034637	02/04/2025	1,000	1,500	2,000	1,006	1,505	2,010	1.011x - 15.064	0.999	
B40	SKC	224-PCXR4	798349	03/04/2025	1,000	1,500	2,000	997	1,510	2,008	1.012x - 19.381	1.000	

Calibrated by :

Abul Dangkum
(Mr. Abul Dangkum)

Approved by :

Peers Detudom
(Mr. Peers Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจตุจักร เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jomjol, Chatchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Personal Pump Calibration Report

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

Environmental Conditions

Temperature : 25 ± 3 °C
Pressure : 1010 ± 15 mmbar

Personal Pump Data				Calibration Data									
No.	Brand	Model	Serial No.	Date	Flow Rate (mL/min)						Value From Calibration Curve		
					Setting			Actual (Q std.)					
					1	2	3	1	2	3	y	R ²	
B61	SKC	224-PCXR4	612669	03/04/2025	1,000	1,500	2,000	1,005	1,502	2,004	1.005x - 8.923	1.000	
B62	SKC	224-PCXR4	626041	03/04/2025	1,000	1,500	2,000	1,004	1,501	2,008	1.009x - 13.856	1.000	
B43	SKC	224-PCXR4	034636	01/04/2025	1,000	1,500	2,000	1,012	1,497	1,996	0.990x + 15.132	1.000	
B44	SKC	224-PCXR8	529341	01/04/2025	1,000	1,500	2,000	1,011	1,511	2,008	1.002x - 0.860	0.999	
B45	SKC	224-PCXR8	529594	04/04/2025	1,000	1,500	2,000	993	1,512	2,003	1.009x - 14.476	1.000	
B46	SKC	224-PCXR8	566743	04/04/2025	1,000	1,500	2,000	1,008	1,508	2,008	1.000x - 0.100	0.999	
B47	SKC	224-PCXR8	566747	04/04/2025	1,000	1,500	2,000	999	1,510	2,010	1.010x - 14.444	1.000	
B48	SKC	224-PCXR8	566753	01/04/2025	1,000	1,500	2,000	1,010	1,506	2,006	0.999x + 2.782	1.000	
B49	SKC	224-PCXR8	566780	04/04/2025	1,000	1,500	2,000	1,003	1,504	2,004	1.003x - 2.183	1.000	
B50	SKC	224-PCXR8	500400	04/04/2025	1,000	1,500	2,000	1,002	1,493	1,995	0.994x + 5.841	1.000	
B51	SKC	224-PCXR8	500363	04/04/2025	1,000	1,500	2,000	998	1,511	2,011	1.013x - 19.465	0.999	
B52	SKC	224-PCXR8	093186	02/04/2025	1,000	1,500	2,000	997	1,505	2,006	1.008x - 12.641	1.000	
B53	SKC	224-PCXR8	707670	02/04/2025	1,000	1,500	2,000	1,004	1,503	2,007	1.007x - 7.992	1.000	
B54	SKC	224-PCXR3	509821	02/04/2025	1,000	1,500	2,000	1,005	1,504	2,008	1.010x - 15.060	0.999	
B55	SKC	224-PCXR3	510710	02/04/2025	1,000	1,500	2,000	1,001	1,495	1,997	0.996x + 5.073	1.000	
B56	SKC	224-PCXR3	511450	02/04/2025	1,000	1,500	2,000	1,005	1,494	1,996	0.991x - 13.585	1.000	
B57	SKC	224-PCXR3	510798	03/04/2025	1,000	1,500	2,000	997	1,511	2,009	1.014x - 21.340	0.999	
B58	SKC	224-PCXR3	509852	03/04/2025	1,000	1,500	2,000	1,006	1,493	2,002	1.001x - 4.094	1.000	
B59	SKC	224-PCXR3	509862	03/04/2025	1,000	1,500	2,000	995	1,502	2,003	1.012x - 21.564	1.000	
B60	SKC	224-PCXR3	512655	03/04/2025	1,000	1,500	2,000	998	1,507	2,004	1.010x - 18.510	0.999	
B61	SKC	224-PCXR3	503915	03/04/2025	1,000	1,500	2,000	997	1,499	2,001	1.002x - 4.374	1.000	
B62	SKC	224-PCXR3	505975	01/04/2025	1,000	1,500	2,000	1,002	1,503	2,005	1.008x - 11.138	1.000	
B63	SKC	224-PCXR3	511432	04/04/2025	1,000	1,500	2,000	998	1,502	1,996	0.996x + 3.970	1.000	
B64	SKC	224-PCXR3	508302	04/04/2025	1,000	1,500	2,000	1,005	1,509	2,008	1.009x - 10.402	1.000	
B65	SKC	224-PCXR3	508310	04/04/2025	1,000	1,500	2,000	1,004	1,503	2,007	1.010x - 14.088	1.000	
B66	SKC	224-PCXR3	509861	04/04/2025	1,000	1,500	2,000	1,003	1,504	2,010	1.008x - 12.369	1.000	
B67	SKC	224-PCXR3	506295	04/04/2025	1,000	1,500	2,000	1,002	1,498	2,004	0.998x + 4.290	1.000	
B68	SKC	224-PCXR3	505872	04/04/2025	1,000	1,500	2,000	999	1,504	1,998	1.000x + 0.436	1.000	
B69	SKC	224-PCXR3	508375	02/04/2025	1,000	1,500	2,000	1,004	1,498	2,002	0.996x + 5.501	1.000	
B70	SKC	224-PCXR3	510623	02/04/2025	1,000	1,500	2,000	994	1,497	2,005	1.005x - 8.735	1.000	
B71	SKC	224-PCXR3	508367	02/04/2025	1,000	1,500	2,000	1,013	1,505	2,009	1.000x + 3.394	0.999	
B72	SKC	224-PCXR3	505977	02/04/2025	1,000	1,500	2,000	997	1,494	2,003	1.006x - 11.330	1.000	
B73	SKC	224-PCXR3	512606	01/04/2025	1,000	1,500	2,000	1,010	1,507	2,004	0.998x + 5.129	1.000	
B74	SKC	224-PCXR3	505993	01/04/2025	1,000	1,500	2,000	998	1,499	2,010	1.009x - 11.942	1.000	
B75	SKC	224-PCXR3	509820	01/04/2025	1,000	1,500	2,000	995	1,511	2,004	1.011x - 18.966	0.999	
B76	SKC	224-PCXR3	509811	01/04/2025	1,000	1,500	2,000	999	1,504	2,010	1.012x - 20.993	0.999	
B77	SKC	224-PCXR3	508301	03/04/2025	1,000	1,500	2,000	1,007	1,509	2,008	1.001x + 3.750	1.000	
B78	SKC	224-PCXR3	510677	04/04/2025	1,000	1,500	2,000	998	1,508	2,001	1.003x - 3.278	1.000	
B79	SKC	224-PCXR3	510920	04/04/2025	1,000	1,500	2,000	1,001	1,501	1,994	0.999x - 1.819	1.000	



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chaiyachak, Bangkok 10900
Tel : (662) 939-4370-72 Fax : (662) 513-4321, E-mail : sales@spscn.com, www.spscn.com

Personal Pump Calibration Report

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

Environmental Conditions

Temperature : 25 ± 3 °C
Pressure : 1010 ± 15 mmbar

Personal Pump Data

Calibration Data

No.	Brand	Model	Serial No.	Date	Flow Rate (ml/min)						Value From Calibration Curve	
					Setting			Actual (Q std.)				
					1	2	3	1	2	3	y	R ²
880	SKC	224-PCXR3	504569	04/04/2025	1,000	1,500	2,000	1,010	1,515	1,999	0.989x + 16.683	0.999
881	SKC	224-PCXR3	503480	02/04/2025	1,000	1,500	2,000	1,007	1,499	1,998	0.997x + 2.890	1.000
882	SKC	224-PCXR3	505673	02/04/2025	1,000	1,500	2,000	999	1,511	2,004	1.007x - 11.710	1.000
883	SKC	224-PCXR3	510785	02/04/2025	1,000	1,500	2,000	1,005	1,504	2,008	1.005x - 3.353	1.000
884	SKC	224-PCXR3	508333	03/04/2025	1,000	1,500	2,000	998	1,508	2,002	1.003x - 4.482	1.000
885	SKC	224-PCXR3	505757	03/04/2025	1,000	1,500	2,000	1,010	1,499	2,006	0.999x + 0.820	0.999
886	SKC	224-PCXR3	512625	04/04/2025	1,000	1,500	2,000	1,003	1,494	1,998	0.993x + 6.616	1.000
887	SKC	224-PCXR3	504324	04/04/2025	1,000	1,500	2,000	1,004	1,506	2,000	1.000x - 1.787	1.000
888	SKC	224-PCXR3	508307	04/04/2025	1,000	1,500	2,000	1,002	1,511	2,009	1.009x - 12.753	0.999
889	SKC	224-PCXR3	509860	04/04/2025	1,000	1,500	2,000	999	1,504	1,997	0.998x + 1.835	1.000
890	SKC	224-PCXR3	508366	04/04/2025	1,000	1,500	2,000	1,004	1,498	2,004	0.997x + 4.382	1.000
891	SKC	224-PCXR3	510919	02/04/2025	1,000	1,500	2,000	997	1,495	2,002	1.005x - 9.911	1.000
892	SKC	224-PCXR3	510987	02/04/2025	1,000	1,500	2,000	1,012	1,507	2,004	0.997x + 7.928	1.000
893	SKC	224-PCXR3	509845	02/04/2025	1,000	1,500	2,000	998	1,499	2,010	1.009x - 11.942	1.000

Calibrated by :

Adul Dangklom
(Mr. Adul Dangklom)

Approved by :

Mr. Peters Detudom
(Mr. Peters Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin Rd., Jompol, Chulachak, Bangkok 10900
Tel : (662) 939-4370-72 Fax : (662) 513-4221 E-mail : sale@spscn.com, www.spscn.com

Rotameter Calibration Report (For Personal Pump Low Flow Adjust)

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

Rotameter Data			Calibration Data								
No.	Brand	Model	Date	Flow Rate (mL/min)						Value From Calibration Curve	
				Flow Rate (Reading)			Actual (Q std.)				
				1	2	3	1	2	3	y	R ²
L-801	Dwyer	VFA-21	01/04/2025	50	100	200	49.7	99.0	200.9	0.993x + 0.389	1.000
L-802	Dwyer	VFA-21	03/04/2025	50	100	200	50.1	100.2	199.3	0.995x + 1.127	0.999
L-803	Dwyer	VFA-21	04/04/2025	50	100	200	50.4	100.7	201.1	0.992x + 1.381	1.000
L-804	Dwyer	VFA-21	01/04/2025	50	100	200	50.8	99.0	198.2	0.998x + 0.410	0.999
L-805	Dwyer	VFA-21	02/04/2025	50	100	200	50.6	99.5	198.9	0.993x + 0.837	1.000
L-806	Dwyer	VFA-21	03/04/2025	50	100	200	50.1	100.2	201.7	1.002x - 0.009	1.000
L-807	Dwyer	VFA-21	03/04/2025	50	100	200	50.3	101.7	200.5	0.996x + 1.584	1.000
L-808	Dwyer	VFA-21	01/04/2025	50	100	200	50.8	99.2	199.4	0.991x + 0.972	1.000
L-809	Dwyer	VFA-21	04/04/2025	50	100	200	50.5	99.6	199.7	0.996x + 1.101	0.999
L-810	Dwyer	VFA-21	04/04/2025	50	100	200	50.4	101.5	202.0	0.998x + 1.243	1.000

Calibrated by :

Adul Dangklom
(Mr. Adul Dangklom)

Approved by :

(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin Rd., Phaholyothin Rd., Jompol, Chulachak, Bangkok 10900
Tel : (662) 939-4370-72 Fax : (662) 513-4221 E-mail : sale@spscn.com, www.spscn.com

Rotameter Calibration Report (For Personal Pump High Flow Adjust)

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

Calibration Data											
Rotameter Data			Calibration Data								
No.	Brand	Model	Date	Flow Rate (mL/min)						Value From Calibration Curve	
				Flow Rate (Reading)			Actual (Q std.)				
				1	2	3	1	2	3	y	R ²
H-801	Dwyer	VFB-65	01/04/2025	500	1,000	2,000	499.1	997.5	1996.2	0.992x + 10.557	1.000
H-802	Dwyer	VFB-65	03/04/2025	500	1,000	2,000	501.5	996.9	2004.4	1.002x - 0.966	1.000
H-803	Dwyer	VFB-65	03/04/2025	500	1,000	2,000	498.9	997.4	1996.5	0.997x - 0.674	1.000
H-804	Dwyer	VFB-65	01/04/2025	500	1,000	2,000	498.0	996.5	2007.8	1.001x - 8.142	0.999
H-805	Dwyer	VFB-65	02/04/2025	500	1,000	2,000	501.2	998.6	1993.7	0.994x + 6.199	1.000
H-806	Dwyer	VFB-65	03/04/2025	500	1,000	2,000	499.7	995.3	1989.1	0.995x + 1.374	0.999
H-807	Dwyer	VFB-65	03/04/2025	500	1,000	2,000	500.1	999.7	2006.4	0.998x - 1.014	1.000
H-808	Dwyer	VFB-65	01/04/2025	500	1,000	2,000	499.8	997.4	1994.8	0.993x + 6.689	1.000
H-809	Dwyer	VFB-65	04/04/2025	500	1,000	2,000	498.2	997.1	2005.6	0.999x + 0.065	0.999
H-810	Dwyer	VFB-65	04/04/2025	500	1,000	2,000	501.2	998.4	2009.2	0.998x + 3.713	1.000

Calibrated by :

Adul Dangklom
(Mr. Adul Dangklom)

Approved by :

(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ซอยพหลโยธิน แขวงจตุจักร เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompet, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sales@spscs.com, www.spscs.com

Calibration Report Total Hydrocarbon Analyzer			
Date :	04 March 2025	Brand :	HORIBA
No. :	B01	Model :	APHA-360CE
		Serial No. :	4211954001
Calibrator (Dilution System)			
Brand :	Teledyne	Model :	700
Last Cal. Date :	29 October 2024	Serial No. :	421
Reference Standard Gas			
Standard Gas :	Methane (CH ₄)	Cylinder No. :	D612165
Certified Date :	25 February 2023	Expired Date :	25 February 2031
		Cylinder Conc. :	453 ppm
Calibrating Condition			
Pressure :	1011 mmbar	Temp. :	24.5 °C
		% RH :	50
		Start Time :	9:00 AM
Pre-Calibration Checks			
Change Particulate Filter :	Yes	Station Temp. :	25.0 °C
Leak Test :	Yes		
Calibration Setting			
Span Set Point	Initial Reading (Before Adj)		Final Reading (After Adj)
	Expected Concentration (PPM)	Analyzer Response (PPM)	Analyzer Response (PPM)
Zero	0	0.11	0
Span	10	10.05	10
Calibration Setting (Final)			
Span Instrument Gain:	0.995	Finish Time:	10:00 AM
APHA-360 Total Hydrocarbon Analyzer			
Test Values	Observed Value	Units	Nominal Range
Signal (CH ₄)	910.4	mV	800-1,350
Signal (THC)	916.2	mV	800-1,350
Detector	78.0	kPa	((Pressure Air/1013)x100)-20 ± 4 kPa
Purifier	19.2	kPa	8 - 25
NMC	259.1	°C	260 ± 10
Bypass	0.9	L / min	0.9 ± 0.3
Over Flow	0.8	L / Min	0.8

Calibrated by : Adul Dangklom
(Mr. Adul Dangklom)

Approved by : Peera Detudom
(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ซอยพหลโยธิน แขวงจตุจักร เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompet, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sales@spscs.com, www.spscs.com

Calibration Report Total Hydrocarbon Analyzer			
Date :	02 April 2025	Brand :	HORIBA
No. :	B01	Model :	APHA-360CE
		Serial No. :	4211954001
Calibrator (Dilution System)			
Brand :	Teledyne	Model :	700
Last Cal. Date :	29 October 2024	Serial No. :	421
Reference Standard Gas			
Standard Gas :	Methane (CH ₄)	Cylinder No. :	D612165
Certified Date :	25 February 2023	Expired Date :	25 February 2031
		Cylinder Conc. :	453 ppm
Calibrating Condition			
Pressure :	1011 mmbar	Temp. :	24.6 °C
		% RH :	50
		Start Time :	1:00 PM
Pre-Calibration Checks			
Change Particulate Filter :	Yes	Station Temp. :	25.0 °C
Leak Test :	Yes		
Calibration Setting			
Span Set Point	Initial Reading (Before Adj)		Final Reading (After Adj)
	Expected Concentration (PPM)	Analyzer Response (PPM)	Analyzer Response (PPM)
Zero	0	-0.10	0
Span	10	10.03	10
Calibration Setting (Final)			
Span Instrument Gain:	0.998	Finish Time:	2:00 PM
APHA-360 Total Hydrocarbon Analyzer			
Test Values	Observed Value	Units	Nominal Range
Signal (CH ₄)	911.2	mV	800-1,350
Signal (THC)	915.8	mV	800-1,350
Detector	78.1	kPa	((Pressure Air/1013)x100)-20 ± 4 kPa
Purifier	19.0	kPa	8 - 25
NMC	258.5	°C	260 ± 10
Bypass	0.9	L / min	0.9 ± 0.3
Over Flow	0.8	L / Min	0.8

Calibrated by : Adul Dangklom
(Mr. Adul Dangklom)

Approved by : Peera Detudom
(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

CALIBRATION REPORT					
SO ₂ FLUORESCENT ANALYZER					
DATE :	17 April 2025	BRAND :	API	MODEL :	100A
NO.	SO ₂ -B01	SERIAL NO.	1749		
Calibrator (Dilution System)					
Brand : Teledyne			Model : 700		
Last Cal. Date : 29 October 2024			Serial No. : 421		
Reference Standard Gas					
Standard Gas : Sulphur Dioxide (SO ₂)			Cylinder No. : A008145K		
Certified Date : 21 June 2021			Expired Date : 21 June 2029		
			Cylinder Conc. : 49.8 ppm		
CALIBRATING CONDITION					
Pressure	1011	mmbar	Temp.	24.5	°C
		% RH	49		
CALIBRATION SETTING					
Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response	Slope
Zero	0	-0.10	-	0	-
SO ₂ Span	400.0	400.2	0.050	400.0	1.013
API Model 100A SO ₂ Analyzer Check list					
Test Values	Observed Value	Units	Nominal Range		
RANGE	500	PPB	0-500		
SAMPLE PRESS	28.4	in-Hg	25-35		
SAMPLE FLOW	658	cc/min	650 ± 10%		
PMT	103.2	mV	-20-150 with Zero Air		
UV LAMP	3027.5	mV	1000-4900		
STR. LGT	61.8	PPB	<100		
DRK PMT	63.3	mV	-50 - 200		
DRK LMP	58.0	mV	-50 - 200		
HVPS	674	V	550-900 constant		
DCPS	2528	mV	2500 ± 200		
RCCELL TEMP	50.1	°C	50 ± 1		
BOX TEMP	29.3	°C	5-40		
PMT TEMP	7.0	°C	7 ± 2.0		
SO ₂ Span Conc	400	PPB	20-20,000		
SO ₂ Slope	1.013	-	1.0 ± 0.3		
SO ₂ Offset	21.9	mV	<250		
Stability at Zero	0.1	PPB	<0.2		
Stability at Span	0.2	PPB	0.5% of reading (above 50 ppb)		

Calibrated by : Adul Dangklom
(Mr.Adul Dangklom)

Approved by : (Signature)
(Mr.Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

CALIBRATION REPORT					
SO ₂ FLUORESCENT ANALYZER					
DATE :	17 April 2025	BRAND :	API	MODEL :	100A
NO.	SO ₂ -B02	SERIAL NO.	1847		
Calibrator (Dilution System)					
Brand : Teledyne			Model : 700		
Last Cal. Date : 29 October 2024			Serial No. : 421		
Reference Standard Gas					
Standard Gas : Sulphur Dioxide (SO ₂)			Cylinder No. : A008145K		
Certified Date : 21 June 2021			Expired Date : 21 June 2029		
			Cylinder Conc. : 49.8 ppm		
CALIBRATING CONDITION					
Pressure	1011	mmbar	Temp.	24.5	°C
		% RH	49		
CALIBRATION SETTING					
Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response	Slope
Zero	0	0.11	-	0	-
SO ₂ Span	400.0	400.1	0.025	400.0	1.011
API Model 100A SO ₂ Analyzer Check list					
Test Values	Observed Value	Units	Nominal Range		
RANGE	500	PPB	0-500		
SAMPLE PRESS	28.5	in-Hg	25-35		
SAMPLE FLOW	654	cc/min	650 ± 10%		
PMT	103.1	mV	-20-150 with Zero Air		
UV LAMP	3020.4	mV	1000-4900		
STR. LGT	61.6	PPB	<100		
DRK PMT	63.0	mV	-50 - 200		
DRK LMP	57.7	mV	-50 - 200		
HVPS	670	V	550-900 constant		
DCPS	2519	mV	2500 ± 200		
RCCELL TEMP	50.3	°C	50 ± 1		
BOX TEMP	29.0	°C	5-40		
PMT TEMP	7.2	°C	7 ± 2.0		
SO ₂ Span Conc	400	PPB	20-20,000		
SO ₂ Slope	1.011	-	1.0 ± 0.3		
SO ₂ Offset	22.2	mV	<250		
Stability at Zero	0.1	PPB	<0.2		
Stability at Span	0.2	PPB	0.5% of reading (above 50 ppb)		

Calibrated by : Adul Dangklom
(Mr.Adul Dangklom)

Approved by : (Signature)
(Mr.Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจตุจักร เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chaluchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

CALIBRATION REPORT					
SO ₂ FLUORESCENT ANALYZER					
DATE :	17 April 2025	BRAND :	API	MODEL :	100E
NO.	SO ₂ -804	SERIAL NO.	3159		
Calibrator (Dilution System)					
Brand : Teledyne			Model : 700		
Last Cal. Date : 29 October 2024			Serial No. : 421		
Reference Standard Gas					
Standard Gas : Sulphur Dioxide (SO ₂)			Cylinder No. : A00814SK		
Certified Date : 21 June 2021			Expired Date : 21 June 2029		
			Cylinder Conc. : 49.8 ppm		
CALIBRATING CONDITION					
Pressure	1011	mmbar	Temp.	24.5	°C
			% RH	49	
CALIBRATION SETTING					
Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response	Slope
Zero	0	-0.10	-	0	-
SO ₂ Span	400.0	399.6	-0.100	400.0	1.005
Test Values		Observed Value	Units	Nominal Range	
RANGE		500	PPB	0-500	
SAMPLE PRESS		28.7	in-Hg	25-35	
SAMPLE FLOW		656	cc/min	650 ± 10%	
PMT		103.4	mV	-20-150 with Zero Air	
UV LAMP		3038.1	mV	1000-4900	
STR. LGT		61.9	PPB	<100	
DRK PMT		63.4	mV	-50 - 200	
DRK LMP		58.1	mV	-50 - 200	
HVPS		673	V	550-900 constant	
DCPS		2526	mV	2500 ± 200	
RCCELL TEMP		50.5	°C	50 ± 1	
BOX TEMP		29.4	°C	5-40	
PMT TEMP		7.1	°C	7 ± 2.0	
SO ₂ Span Conc		400	PPB	20-20,000	
SO ₂ Slope		1.005	-	1.0 ± 0.3	
SO ₂ Offset		21.8	mV	<250	
Stability at Zero		0.1	PPB	<0.2	
Stability at Span		0.2	PPB	0.5% of reading (above 50 ppb)	

Calibrated by : Adul Dangklom
(Mr.Adul Dangklom)

Approved by : Peera Detudom
(Mr.Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจตุจักร เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chaluchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

CALIBRATION REPORT						
SO ₂ FLUORESCENT ANALYZER						
DATE :	17 April 2025	BRAND :	Thermo	MODEL :	43C	
NO.	SO ₂ -809	SERIAL NO.	43C-59325-322			
Calibrator (Dilution System)						
Brand : Teledyne			Model : 700			
Last Cal. Date : 29 October 2024			Serial No. : 421			
Reference Standard Gas						
Standard Gas : Sulphur Dioxide (SO ₂)			Cylinder No. : A00814SK			
Certified Date : 21 June 2021			Expired Date : 21 June 2029			
			Cylinder Conc. : 49.8 ppm			
CALIBRATING CONDITION						
Pressure	1011	mmbar	24.5	°C	% RH	49
CALIBRATION SETTING						
Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB		
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response		
Zero	0	0.11	-	0		
SO ₂ Span	400.0	400.3	0.075	400.0		
INSTRUMENT STATUS						
CHAMBER TEMP	44.5	°C	FLOW	1.0	LPM	
PRESSURE	728.6	mm Hg				

Calibrated by : Adul Dangklom
(Mr.Adul Dangklom)

Approved by : Peera Detudom
(Mr.Peera Detudom)

CALIBRATION REPORT				
SO ₂ FLUORESCENT ANALYZER				
DATE :	17 April 2025	BRAND :	Thermo	MODEL :
NO.	SO2-B10	SERIAL NO.	43C-69604-364	
Calibrator (Dilution System)				
Brand :	Teledyne	Model :	700	
Last Cal. Date :	29 October 2024	Serial No. :	421	
Reference Standard Gas				
Standard Gas :	Sulphur Dioxide (SO ₂)	Cylinder No. :	A00814SK	
Certified Date :	21 June 2021	Expired Date :	21 June 2029	
Cylinder Conc. :	49.8 ppm			
CALIBRATING CONDITION				
Pressure	1011 mmbar	24.5 °C	% RH	49
CALIBRATION SETTING				
Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response
Zero	0	0.10	-	0
SO ₂ Span	400.0	400.2	0.050	400.0
INSTRUMENT STATUS				
CHAMBER TEMP	44.3 °C	FLOW	1.0 LPM	
PRESSURE	728.7 mm Hg			

Calibrated by : Adul Dangklom
(Mr.Adul Dangklom)

Approved by : Mr.Peera Detudom
(Mr.Peera Detudom)

CALIBRATION REPORT					
SO ₂ FLUORESCENT ANALYZER					
DATE :	17 April 2025	BRAND :	TELEDYNE	MODEL :	TML-50
NO.	SO ₂ -B11	SERIAL NO.	SO2187		
Calibrator (Dilution System)					
Brand :	Teledyne	Model :	700		
Last Cal. Date :	29 October 2024	Serial No. :	421		
Reference Standard Gas					
Standard Gas :	Sulphur Dioxide (SO ₂)	Cylinder No. :	A00814SK		
Certified Date :	21 June 2021	Expired Date :	21 June 2029		
Cylinder Conc. :	49.8 ppm				
CALIBRATING CONDITION					
Pressure	1011 mmbar	Temp.	24.5 °C	% RH	49
CALIBRATION SETTING					
Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response	Slope
Zero	0	0.10	-	0	-
SO ₂ Span	400.0	399.7	-0.075	400.0	1.007
API Model TML-50 SO ₂ Analyzer Check List					
Test Values	Observed Value	Units	Nominal Range		
RANGE	500	PPB	0-500		
SAMPLE PRESS	28.6	in-Hg	25-35		
SAMPLE FLOW	659	cc/min	650 ± 10%		
PMT	103.3	mV	-20-150 with Zero Air		
UV LAMP	3031.7	mV	1000-4900		
STR. LGT	61.7	PPB	<100		
DRK PMT	63.1	mV	-50 - 200		
DRK LMP	57.8	mV	-50 - 200		
HVPS	672	V	550-900 constant		
DCPS	2517	mV	2500 ± 200		
RCELL TEMP	50.0	°C	50 ± 1		
BOX TEMP	29.2	°C	5-40		
PMT TEMP	7.4	°C	7 ± 2.0		
SO ₂ Span Conc	400	PPB	20-20,000		
SO ₂ Slope	1.007	-	1.0 ± 0.3		
SO ₂ Offset	22.1	mV	<250		
Stability at Zero	0.1	PPB	<0.2		
Stability at Span	0.2	PPB	0.5% of reading (above 50 ppb)		

Calibrated by : Adul Dangklom
(Mr.Adul Dangklom)

Approved by : Mr.Peera Detudom
(Mr.Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

CALIBRATION REPORT					
SO ₂ FLUORESCENT ANALYZER					
DATE :	17 April 2025	BRAND :	TELEDYNE	MODEL :	TML-50
NO.	SO ₂ -B13	SERIAL NO.	1891		
Calibrator (Dilution System)					
Brand	: Teledyne		Model	: 700	
Last Cal. Date	: 29 October 2024		Serial No.	: 421	
Reference Standard Gas					
Standard Gas	: Sulphur Dioxide (SO ₂)		Cylinder No.	: A008145K	
Certified Date	: 21 June 2021		Expired Date	: 21 June 2029	
			Cylinder Conc.	: 49.8 ppm	
CALIBRATING CONDITION					
Pressure	1011	mmbar	Temp.	24.5	°C
			% RH	49	
CALIBRATION SETTING					
Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response	Slope
Zero	0	-0.10	-	0	-
SO ₂ Span	400.0	399.8	-0.050	400.0	1.009
API Model TML-50 SO ₂ Analyzer Check list					
Test Values	Observed Value	Units	Nominal Range		
RANGE	500	PPB	0-500		
SAMPLE PRESS	28.5	in-Hg	25-35		
SAMPLE FLOW	660	cc/min	650 ± 10%		
PMT	103.0	mV	-20-150 with Zero Air		
UV LAMP	3016.4	mV	1000-4900		
STR. LGT	61.5	PPB	<100		
DRK PMT	62.9	mV	-50 - 200		
DRK LMP	57.6	mV	-50 - 200		
HVPS	671	V	550-900 constant		
DCPS	2523	mV	2500 ± 200		
RCCELL TEMP	50.4	°C	50 ± 1		
BOX TEMP	29.1	°C	5-40		
PMT TEMP	7.3	°C	7 ± 2.0		
SO ₂ Span Conc	400	PPB	20-20,000		
SO ₂ Slope	1.009	-	1.0 ± 0.3		
SO ₂ Offset	22.0	mV	<250		
Stability at Zero	0.1	PPB	<0.2		
Stability at Span	0.2	PPB	0.5% of reading (above 50 ppb)		

Calibrated by : Adul Dangklom
(Mr.Adul Dangklom)

Approved by : Peera Detudom
(Mr.Peera Detudom)

CALIBRATION REPORT					
CHEMILUMINESCENT NO / NO ₂ / NO _x ANALYZER					
DATE :	17 April 2025	BRAND :	API	MODEL :	200E
NO.	NOX-B11	SERIAL NO.	4467		
Calibrator (Dilution System)					
Brand	: Teledyne		Model	: 700	
Last Cal. Date	: 29 October 2024		Serial No.	: 421	
Reference Standard Gas					
Standard Gas	: Nitric Oxide (NO)		Cylinder No.	: A00726SV	
Certified Date	: 05 January 2023	Expired Date	: 05 January 2026	Cylinder Conc.	: 48.8 ppm
CALIBRATING CONDITION					
Pressure	1011	mmbar	Temp.	24.5	°C
% RH	49				
CALIBRATION SETTING					
Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response	Slope
Zero	0	-0.10	-	0	-
NO Span	400	399.6	-0.100	400.0	1.004
NO _x Span	400	399.8	-0.050	400.0	1.007
API Model 200E NO _x Analyzer Check List					
Test Values	Observed Value	Units	Nominal Range		
RANGE	500	PPB	500 standard		
STABILITY (Zero Gas)	0.1	PPB	< 2 with zero air		
SAMPLE FLOW	505	cc/min	500 ± 50		
OZONE FLOW	78	cc/min	80 ± 15		
PMT	103.2	mV	-20 - 150		
AZERO	93.9	mV	-20 - 150		
HVPS	673	V	420 - 900 constant		
RCCELL TEMP	50.1	°C	50 ± 1		
BOX TEMP	28.9	°C	8 - 48		
PMT TEMP	7.2	°C	7 ± 2		
MOLY TEMP	315.4	°C	315 ± 5		
RCCELL PRESS	8.4	IN-Hg-A	2 - 10 constant		
SAMPLE PRESS	28.6	IN-Hg-A	25 - 30 constant		
NO Span Conc	400	PPB	20 - 20,000		
NO _x Span Conc	400	PPB	20 - 20,000		
NO Slope	1.004	-	1.0 ± 0.3		
NO _x Slope	1.007	-	1.0 ± 0.3		
NO Offset	1.0	mV	-20 to +150		
NO _x Offset	0.6	mV	-20 to 150		
Stability at Zero	0.1	PPB	< 0.2		
Stability at Span	0.2	PPB	< 2 ppb @ 400 ppb span gas		

Calibrated by : Adul Dangklom
(Mr.Adul Dangklom)

Approved by : Peera Detudom
(Mr.Peera Detudom)

CALIBRATION REPORT					
CHEMILUMINESCENT NO / NO ₂ / NO _x ANALYZER					
DATE :	17 April 2025	BRAND :	API	MODEL :	200A
NO.	NOX-B12	SERIAL NO.	2675		
Calibrator (Dilution System)					
Brand	: Teledyne		Model	: 700	
Last Cal. Date	: 29 October 2024		Serial No.	: 421	
Reference Standard Gas					
Standard Gas	: Nitric Oxide (NO)		Cylinder No.	: A00726SV	
Certified Date	: 05 January 2023	Expired Date	: 05 January 2026	Cylinder Conc.	: 48.8 ppm
CALIBRATING CONDITION					
Pressure	1011	mmbar	Temp.	24.5	°C
% RH	49				
CALIBRATION SETTING					
Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response	Slope
Zero	0	0.11	-	0	-
NO Span	400	400.1	0.025	400.0	1.009
NO _x Span	400	400.2	0.050	400.0	1.012
API Model 200A NO _x Analyzer Check List					
Test Values	Observed Value	Units	Nominal Range		
RANGE	500	PPB	500 standard		
STABILITY (Zero Gas)	0.1	PPB	< 2 with zero air		
SAMPLE FLOW	511	cc/min	500 ± 50		
OZONE FLOW	79	cc/min	80 ± 15		
PMT	103.0	mV	-20 - 150		
AZERO	93.7	mV	-20 - 150		
HVPS	670	V	420 - 900 constant		
RCCELL TEMP	50.4	°C	50 ± 1		
BOX TEMP	29.3	°C	8 - 48		
PMT TEMP	7.0	°C	7 ± 2		
MOLY TEMP	315.2	°C	315 ± 5		
RCCELL PRESS	8.5	IN-Hg-A	2 - 10 constant		
SAMPLE PRESS	28.7	IN-Hg-A	25 - 30 constant		
NO Span Conc	400	PPB	20 - 20,000		
NO _x Span Conc	400	PPB	20 - 20,000		
NO Slope	1.009	-	1.0 ± 0.3		
NO _x Slope	1.012	-	1.0 ± 0.3		
NO Offset	1.7	mV	-20 to +150		
NO _x Offset	1.0	mV	-20 to 150		
Stability at Zero	0.1	PPB	< 0.2		
Stability at Span	0.2	PPB	< 2 ppb @ 400 ppb span gas		

Calibrated by : Adul Dangklom
(Mr.Adul Dangklom)

Approved by : Peera Detudom
(Mr.Peera Detudom)

CALIBRATION REPORT					
CHEMILUMINESCENT NO / NO ₂ / NO _x ANALYZER					
DATE :	17 April 2025	BRAND :	API	MODEL :	200A
NO.	NOX-B14	SERIAL NO.	212		
Calibrator (Dilution System)					
Brand :	Teledyne	Model :	700		
Last Cal. Date :	29 October 2024	Serial No. :	421		
Reference Standard Gas					
Standard Gas :	Nitric Oxide (NO)	Cylinder No. :	A00726SV		
Certified Date :	05 January 2023	Expired Date :	05 January 2026	Cylinder Conc. :	48.8 ppm
CALIBRATING CONDITION					
Pressure	1011	mmbar	Temp.	24.5	°C
			% RH	49	
CALIBRATION SETTING					
Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response	Slope
Zero	0	-0.10	-	0	-
NO Span	400	399.7	-0.075	400.0	1.006
NO _x Span	400	400.1	0.025	400.0	1.010
API Model 200A NO _x Analyzer Check List					
Test Values	Observed Value	Units	Nominal Range		
RANGE	500	PPB	500 standard		
STABILITY (Zero Gas)	0.1	PPB	< 2 with zero air		
SAMPLE FLOW	507	cc/min	500 ± 50		
OZONE FLOW	78	cc/min	80 ± 15		
PMT	102.9	mV	-20 - 150		
AZERO	93.6	mV	-20 - 150		
HVPS	674	V	420 - 900 constant		
RCELL TEMP	50.5	°C	50 ± 1		
BOX TEMP	29.4	°C	8 - 48		
PMT TEMP	7.3	°C	7 ± 2		
MOLY TEMP	315.1	°C	315 ± 5		
RCELL PRESS	8.2	IN-Hg-A	2 - 10 constant		
SAMPLE PRESS	28.5	IN-Hg-A	25 - 30 constant		
NO Span Conc	400	PPB	20 - 20,000		
NO _x Span Conc	400	PPB	20 - 20,000		
NO Slope	1.006	-	1.0 ± 0.3		
NO _x Slope	1.010	-	1.0 ± 0.3		
NO Offset	1.2	mV	-20 to +150		
NO _x Offset	0.8	mV	-20 to 150		
Stability at Zero	0.1	PPB	< 0.2		
Stability at Span	0.2	PPB	< 2 ppb @ 400 ppb span gas		

Calibrated by : Adul Dangklom
(Mr.Adul Dangklom)

Approved by : Peera Detudom
(Mr.Peera Detudom)

CALIBRATION REPORT					
CHEMILUMINESCENT NO / NO ₂ / NO _x ANALYZER					
DATE :	17 April 2025	BRAND :	API	MODEL :	200A
NO.	NOX-B15	SERIAL NO.	213		
Calibrator (Dilution System)					
Brand :	Teledyne	Model :	700		
Last Cal. Date :	29 October 2024	Serial No. :	421		
Reference Standard Gas					
Standard Gas :	Nitric Oxide (NO)	Cylinder No. :	A00726SV		
Certified Date :	05 January 2023	Expired Date :	05 January 2026	Cylinder Conc. :	48.8 ppm
CALIBRATING CONDITION					
Pressure	1011	mmbar	Temp.	24.5	°C
			% RH	49	
CALIBRATION SETTING					
Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response	Slope
Zero	0	0.10	-	0	-
NO Span	400	399.8	-0.050	400.0	1.007
NO _x Span	400	400.1	0.025	400.0	1.011
API Model 200A NO _x Analyzer Check List					
Test Values	Observed Value	Units	Nominal Range		
RANGE	500	PPB	500 standard		
STABILITY (Zero Gas)	0.1	PPB	< 2 with zero air		
SAMPLE FLOW	509	cc/min	500 ± 50		
OZONE FLOW	78	cc/min	80 ± 15		
PMT	103.3	mV	-20 - 150		
AZERO	94.0	mV	-20 - 150		
HVPS	669	V	420 - 900 constant		
RCELL TEMP	50.0	°C	50 ± 1		
BOX TEMP	28.8	°C	8 - 48		
PMT TEMP	7.1	°C	7 ± 2		
MOLY TEMP	314.7	°C	315 ± 5		
RCELL PRESS	8.4	IN-Hg-A	2 - 10 constant		
SAMPLE PRESS	28.7	IN-Hg-A	25 - 30 constant		
NO Span Conc	400	PPB	20 - 20,000		
NO _x Span Conc	400	PPB	20 - 20,000		
NO Slope	1.007	-	1.0 ± 0.3		
NO _x Slope	1.011	-	1.0 ± 0.3		
NO Offset	1.3	mV	-20 to +150		
NO _x Offset	0.9	mV	-20 to 150		
Stability at Zero	0.1	PPB	< 0.2		
Stability at Span	0.2	PPB	< 2 ppb @ 400 ppb span gas		

Calibrated by : Adul Dangklom
(Mr.Adul Dangklom)

Approved by : Peera Detudom
(Mr.Peera Detudom)

CALIBRATION REPORT					
CHEMILUMINESCENT NO / NO ₂ / NO _x ANALYZER					
DATE :	17 April 2025	BRAND :	API	MODEL :	TML-41M
NO.	NOX-B18	SERIAL NO.	N07543		
Calibrator (Dilution System)					
Brand	: Teledyne		Model	: 700	
Last Cal. Date	: 29 October 2024		Serial No.	: 421	
Reference Standard Gas					
Standard Gas	: Nitric Oxide (NO)		Cylinder No.	: A007265V	
Certified Date	: 05 January 2023	Expired Date	: 05 January 2026	Cylinder Conc.	: 48.8 ppm
CALIBRATING CONDITION					
Pressure	1011	mmbar	Temp.	24.5	°C
% RH	49				
CALIBRATION SETTING					
Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response	Slope
Zero	0	-0.10	-	0	-
NO Span	400	399.7	-0.075	400.0	1.005
NO _x Span	400	399.9	-0.025	400.0	1.008
API Model TML-41M NO _x Analyzer Check List					
Test Values	Observed Value	Units	Nominal Range		
RANGE	500	PPB	500 standard		
STABILITY (Zero Gas)	0.1	PPB	< 2 with zero air		
SAMPLE FLOW	510	cc/min	500 ± 50		
OZONE FLOW	79	cc/min	80 ± 15		
PMT	103.4	mV	-20 - 150		
AZERO	94.1	mV	-20 - 150		
HVPS	675	V	420 - 900 constant		
RCELL TEMP	50.1	°C	50 ± 1		
BOX TEMP	29.3	°C	8 - 48		
PMT TEMP	7.5	°C	7 ± 2		
MOLY TEMP	314.9	°C	315 ± 5		
RCELL PRESS	8.2	IN-Hg-A	2 - 10 constant		
SAMPLE PRESS	28.4	IN-Hg-A	25 - 30 constant		
NO Span Conc	400	PPB	20 - 20,000		
NO _x Span Conc	400	PPB	20 - 20,000		
NO Slope	1.005	-	1.0 ± 0.3		
NO _x Slope	1.008	-	1.0 ± 0.3		
NO Offset	1.1	mV	-20 to +150		
NO _x Offset	0.7	mV	-20 to 150		
Stability at Zero	0.1	PPB	< 0.2		
Stability at Span	0.2	PPB	< 2 ppb @ 400 ppb span gas		

Calibrated by : Adul Dangklom
(Mr.Adul Dangklom)

Approved by : Mr.Peera Detudom
(Mr.Peera Detudom)

CALIBRATION REPORT					
CHEMILUMINESCENT NO / NO ₂ / NO _x ANALYZER					
DATE :	17 April 2025	BRAND :	API	MODEL :	200E
NO.	NOX-B19	SERIAL NO.	353		
Calibrator (Dilution System)					
Brand	: Teledyne		Model	: 700	
Last Cal. Date	: 29 October 2024		Serial No.	: 421	
Reference Standard Gas					
Standard Gas	: Nitric Oxide (NO)		Cylinder No.	: A007265V	
Certified Date	: 05 January 2023	Expired Date	: 05 January 2026	Cylinder Conc.	: 48.8 ppm
CALIBRATING CONDITION					
Pressure	1011	mmbar	Temp.	24.5	°C
% RH	49				
CALIBRATION SETTING					
Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response	Slope
Zero	0	0.11	-	0	-
NO Span	400	400.1	0.025	400.0	1.010
NO _x Span	400	400.3	0.075	400.0	1.013
API Model 200E NO _x Analyzer Check List					
Test Values	Observed Value	Units	Nominal Range		
RANGE	500	PPB	500 standard		
STABILITY (Zero Gas)	0.1	PPB	< 2 with zero air		
SAMPLE FLOW	512	cc/min	500 ± 50		
OZONE FLOW	79	cc/min	80 ± 15		
PMT	103.1	mV	-20 - 150		
AZERO	93.8	mV	-20 - 150		
HVPS	672	V	420 - 900 constant		
RCELL TEMP	50.3	°C	50 ± 1		
BOX TEMP	29.5	°C	8 - 48		
PMT TEMP	7.2	°C	7 ± 2		
MOLY TEMP	314.8	°C	315 ± 5		
RCELL PRESS	8.3	IN-Hg-A	2 - 10 constant		
SAMPLE PRESS	28.6	IN-Hg-A	25 - 30 constant		
NO Span Conc	400	PPB	20 - 20,000		
NO _x Span Conc	400	PPB	20 - 20,000		
NO Slope	1.010	-	1.0 ± 0.3		
NO _x Slope	1.013	-	1.0 ± 0.3		
NO Offset	1.8	mV	-20 to +150		
NO _x Offset	1.1	mV	-20 to 150		
Stability at Zero	0.1	PPB	< 0.2		
Stability at Span	0.2	PPB	< 2 ppb @ 400 ppb span gas		

Calibrated by : Adul Dangklom
(Mr.Adul Dangklom)

Approved by : Mr.Peera Detudom
(Mr.Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

CALIBRATION REPORT					
CHEMILUMINESCENT NO / NO ₂ / NO _x ANALYZER					
DATE :	17 April 2025	BRAND :	API	MODEL :	200E
NO.	NOX-R07	SERIAL NO.	4468		
Calibrator (Dilution System)					
Brand	: Teledyne		Model	: 700	
Last Cal. Date	: 29 October 2024		Serial No.	: 421	
Reference Standard Gas					
Standard Gas	: Nitric Oxide (NO)		Cylinder No.	: A00726SV	
Certified Date	: 05 January 2023		Expired Date	: 05 January 2026	
Cylinder Conc.	: 48.8 ppm				
CALIBRATING CONDITION					
Pressure	1011	mmbar	Temp.	24.5	°C
% RH	49				
CALIBRATION SETTING					
Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response	Slope
Zero	0	0.10	-	0	-
NO Span	400	399.9	-0.025	400.0	1.008
NO _x Span	400	400.2	0.050	400.0	1.012
API Model 200E NO _x Analyzer Check List					
Test Values	Observed Value	Units	Nominal Range		
RANGE	500	PPB	500 standard		
STABILITY (Zero Gas)	0.1	PPB	< 2 with zero air		
SAMPLE FLOW	506	cc/min	500 ± 50		
OZONE FLOW	78	cc/min	80 ± 15		
PMT	103.5	mV	-20 - 150		
AZERO	94.2	mV	-20 - 150		
HVPS	671	V	420 - 900 constant		
RCCELL TEMP	50.2	°C	50 ± 1		
BOX TEMP	29.0	°C	8 - 48		
PMT TEMP	7.3	°C	7 ± 2		
MOLY TEMP	315.3	°C	315 ± 5		
RCCELL PRESS	8.5	IN-Hg-A	2 - 10 constant		
SAMPLE PRESS	28.7	IN-Hg-A	25 - 30 constant		
NO Span Conc	400	PPB	20 - 20,000		
NO _x Span Conc	400	PPB	20 - 20,000		
NO Slope	1.008	-	1.0 ± 0.3		
NO _x Slope	1.012	-	1.0 ± 0.3		
NO Offset	1.5	mV	-20 to +150		
NO _x Offset	0.9	mV	-20 to 150		
Stability at Zero	0.1	PPB	< 0.2		
Stability at Span	0.2	PPB	< 2 ppb @ 400 ppb span gas		

Calibrated by : Adul Dangklorn
(Mr.Adul Dangklorn)

Approved by : (Signature)
(Mr.Peera Detudom)



MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

Customer :	S.P.S.Consulting Service Co.,Ltd	Date Tested:	January 6, 2025
		Recommendation Recertification	
Address :	7 Soi Phaholyothin 24	Period	6 Months
	Paholyothin Road	Recertification Due:	July 6, 2025
	Jompoi Chatuchak, Bangkok 1090	Date Last Certified:	July 4, 2024
User Name:	K.Phenpha Viphashtawat	Visit Number:	2 of 2
Phone:	083-9269252	PerkinElmer Phone:	02-719-6420 ext 206
Fax:	02-513-4221	PerkinElmer Fax:	02-318-5597

CONFIGURATION TESTED	ACCESSORIES/COMPONENT NOT INCLUDED	
MODEL	SERIAL NUMBER	
OPTIMA 5300DV	077C7042401	
TESTED EQUIPMENT	CALIBRATION NUMBER	EXPIRATION
IPV Methods		
TEST STANDARD USED	PART NUMBER	EXPIRATION DATE
Multielement Standard	N069-1579	December 30, 2025
Wavecal Solution	N058-2152	April 30, 2025
VIS Wavecal solution	N930-2946	December 30, 2025
Instrument Cal. STD4	N930-0221	August 30, 2025
CUSTOMER SUPPLIED	COMMENTS	CUSTOMER INITIALS
2 % HNO3		
10 % HNO3		



MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

SERIAL NUMBER	077C7042401	DATE TESTED	January 6, 2025
1. MECHANICAL CHECKS			
A. Inspect and clean all fans and filters.			<input type="checkbox"/> OK
B. Inspect and replace as necessary, all torch components including the RF coil.			<input type="checkbox"/> OK
C. Inspect all tubing for sign of clacking or leaking.			<input type="checkbox"/> OK
D. Adjust water and gas pressure regulator settings.			<input type="checkbox"/> OK
E. Inspect and leak check pneumatics drawers.			<input type="checkbox"/> OK
F. Clean the exterior of the instrument.			<input type="checkbox"/> OK
2. OPTICAL CHECKS			
A. Inspect and clean all optical components.			<input type="checkbox"/> OK
B. As required, check and replace all purgefilters.			<input type="checkbox"/> OK
C. Recheck optical alignment.			<input type="checkbox"/> OK
3. COOLING SYSTEM CHECKS			
A. Perform preventive maintenance on chiller.			<input type="checkbox"/> OK
B. Flush out the chiller every year.			<input type="checkbox"/> N/A
4. PERFORMANCE CHECKS			
A. Torch View Alignment.			<input type="checkbox"/> OK
B. Wavelength Calibration.			<input type="checkbox"/> OK



MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

SERIAL NUMBER : <u>077C7042401</u>		DATE TESTED : <u>January 6, 2025</u>	
PARAMETER	SPECIFICATION		FINAL VALUE
Spectral Resolution : UV	As 193.696 nm	≤ 0.007	0.00519
	Ni 231.604 nm	≤ 0.008	0.00667
	Ni 341.476 nm	≤ 0.012	0.00757
Spectral Resolution : VIS	La 408.672 nm	≤ 0.020	0.01621
	Ba 455.403 nm	≤ 0.025	0.02183
Precision			
	As 193.656 nm	% RSD < 1.0	0.51 %
	Zn 213.856 nm	% RSD < 1.0	0.48 %
	Mn 257.610 nm	% RSD < 1.0	0.03 %
	La 379.478 nm	% RSD < 1.0	0.05 %
	Ba 455.403 nm	% RSD < 1.0	0.07 %
	Ba 493.408 nm	% RSD < 1.0	0.04 %
Detection Limits : Axial	Tl 190.080 nm	3(sd)	10.65 ppb
	As 193.696 nm	3(sd)	2.48 ppb
	Pb 220.353 nm	3(sd)	3.09 ppb
Detection Limits : Radial	As 193.696 nm	3(sd)	12.41 ppb
	Zn 213.856 nm	3(sd)	0.91 ppb
	Mn 257.610 nm	3(sd)	0.13 ppb
	La 379.478 nm	3(sd)	4.74 ppb
	Ba 455.403 nm	3(sd)	0.10 ppb
	Ba 493.408 nm	3(sd)	0.18 ppb
BEC : Axial (IB X 500)/(IS-IB)	Cd 226.502 nm	≤ 150 ppb	14.22
BEC : Radial (IB X 1000)/(IS-IB)	Mn 257.610 nm	≤ 45 ppb	6.14



MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

SERIAL NUMBER	<u>077C7042401</u>	DATE TESTED	<u>January 6, 2025</u>
---------------	--------------------	-------------	------------------------

Remarks :

Commissioning follow as commissioning performance sheets.

This is to certify that the above tests have been performed and the configuration tested

☒ meets
☐ does not meet

the PerkinElmer Specifications listed on this certificate.

This certificate does not modify PerkinElmer's standard terms and condition of sale, including warranty terms.

Service Department PerkinElmer Ltd.

Authorized Representative: *Wiphan Promlumda*

(Wiphan Promlumda)
Service Engineer

เอกสารแนบ 5-2

เอกสารสอบเทียบเครื่องมือการตรวจวิเคราะห์คุณภาพน้ำ

MAINTENANCE REPORT AND CALIBRATION CERTIFICATE

ATOMIC ABSORPTION SPECTROPHOTOMETER MODEL

PinAAcle 900T

Customer :	S.P.S.Consulting Service Co.,Ltd	Date Tested:	January 6, 2025
		Recommendation	Recertification
Address :	7 Soi Phaholyothin 24	Period	6 Months
	Paholyothin Road	Recertification Due:	July 6, 2025
	Jompol Chatuchak, Bangkok 10900	Date Last Certified:	July 4, 2024
User Name:	K.Phenpha Vipasthawatt	Visit Number:	1 OF 2
Phone:	083-9269252	PerkinElmer Phone:	02-719-6420 ext 204
Email:		PerkinElmer Fax:	02-318-5597

CONFIGURATION TESTED		
MODEL	SERIAL NUMBER	SOFTWARE
PinAAcle 900T	PTCS14111103	Wiblab V5.1
AS 900		
TEST STANDARD USED	PART NUMBER	EXPIRATION DATE
Copper	N9300183	APR 30 2025
GFAAS Mixed standard	N9300244	FEB 28 2025
MG0-042	N101-3000	
MG2-045	N101-3002	

MAINTENANCE REPORT AND CALIBRATION CERTIFICATE

ATOMIC ABSORPTION SPECTROPHOTOMETER MODEL

PinAAcle 900T

SERIAL NUMBER	PTCS14111103	DATE TESTED	January 6, 2025
1. INSTRUMENT CHECKS			
A. The Mirror and Lenses Condition			<input type="checkbox"/> OK
B. Grating Condition			<input type="checkbox"/> OK
C. Replace or Clean Dust Filter			<input type="checkbox"/> OK
D. Cleaning the Contact Cylinders			<input type="checkbox"/> OK
E. Cleaning the Furnace Windows			<input type="checkbox"/> OK
F. Cleaning the Burner Head			<input type="checkbox"/> OK
G. Cleaning the Nebulizer			<input type="checkbox"/> OK
H. Cleaning the Drain System			<input type="checkbox"/> OK
2. AUTOSAMPLE CHECK			
A. Sampling and Arm			<input type="checkbox"/> OK
B. Sampling & Rinse Pump			<input type="checkbox"/> OK
C. Sample Position & Clean			<input type="checkbox"/> OK
3. COOLING SYSTEM CHECKS			
A. Clean and Change Distill water			<input type="checkbox"/> OK
B. Themensor			<input type="checkbox"/> OK
4. FIAS CHECKS			
A. Pump and 5 Port Valve			<input type="checkbox"/> N/A
B. Chemifold and Tubing			<input type="checkbox"/> N/A
C. Power Supply			<input type="checkbox"/> N/A
D. Flow meter and Gas system			<input type="checkbox"/> N/A

MAINTENANCE REPORT AND CALIBRATION CERTIFICATE

ATOMIC ABSORPTION SPECTROPHOTOMETER MODEL

PinAAcle 900T

SERIAL NUMBER	PTCS14111103	DATE TESTED	January 6, 2025
PARAMETER	SPECIFICATION	ACTUAL VAULE	
A. Flame Mode Tests			
1. Detector-Linearity with Barium (553.55 nm)			
Neutral Density Filter 0.2 :	0.2042 Abs. + 5%	0.2029 Abs.	
Neutral Density Filter 1.0 :	0.9798 Abs. + 5%	1.0137 Abs.	
2. Baseline Noise at 1 Abs with Barium (553.55 nm) (at an integration time of 0.5 seconds and 99 replicates)	SD ≤ 0.010 Abs.	0.0016 Abs.	
3. AA Baseline with Copper (Cu 324.75 nm) (at an integration time of 0.5 seconds and 99 replicates)	SD ≤ 0.001 Abs.	0.0002 Abs.	
4. D ₂ Background Compensation (Copper 324.75 nm) with Neutral Density Filter 1.0	Absorbance ≤ 0.010 Abs	0.0020 Abs.	
5. AA-BG Baseline Noise with Copper (324.75 nm) (at an integration time of 2.0 seconds and 99 replicates)	SD ≤ 0.005 Abs.	0.0002 Abs.	
6. AA-BG Baseline Noise with Arsenic (193.70 nm) (at an integration time of 2.0 seconds and 99 replicates)	SD ≤ 0.005 Abs.	0.0007 Abs.	

MAINTENANCE REPORT AND CALIBRATION CERTIFICATE

ATOMIC ABSORPTION SPECTROPHOTOMETER MODEL

PinAAcle 900T

SERIAL NUMBER	PTCS14111103	DATE TESTED	January 6, 2025
PARAMETER	SPECIFICATION	ACTUAL VAULE	
7. Flame Interlock Shutdown	Shutdown correct?	<input checked="" type="checkbox"/> OK	
8. Flame Sensitivity with Copper (324.75 nm) (5 mg/L Cu Standard a read time of 10 seconds 10 replicates, standard burner and Stainless stell nebulizer)	Sensitivity ≥ 0.250 Abs.	0.3115 Abs.	
(2 mg/L Cu Standard a read time of 10 seconds 10 replicates, standard burner and High sensitivity nebulizer)	Sensitivity ≥ 0.250 Abs.	N/A Abs.	

MAINTENANCE REPORT AND CALIBRATION CERTIFICATE

ATOMIC ABSORPTION SPECTROPHOTOMETER MODEL


PinAAcle 900T

SERIAL NUMBER	PTCS14111103		DATE TESTED	January 6, 2025
PARAMETER	SPECIFICATION	ACTUAL VAULE		
B. THGA Tests				
1. Furnace Gas Flows				
Internal Flow	250 ± 25 mL/min	250	mL/min	
External Flow	100 ± 10 mL/min	100	mL/min	
2. Chromium Baseline Noise (357.87 nm)				
(mesure 5 furnace dry firings without any sample)				
	Baseline ≤ 0.005 Int.Abs	0.0012		
	SD ≤ 0.005 Int.Abs	0.0002	Int.Abs.	
3. Chromium Characteristic Mass(m ₀) and Precition (357.87 nm)				
(measure 5 furnace firing using 20 ul sample injections of 10 ug/L Cr standard)				
	m0 Results ≤ 7.0 pg/0.0044A-s	5.4	pg/0.0044A-s	
	Precision ≤ 2.0%	1.15	%	
4. Copper Characteristic Mass(m ₀) and Zeeman Ratio (324.75 nm)				
(measure 5 furnace firing using 20 ul sample injections of 25 ug/L Cu standard)				
	m0 Results ≤ 16.5 pg/0.0044A-s	14.4	pg/0.0044A-s	
	Zeeman Ratio 0.52 ± 0.04	0.542		

MAINTENANCE REPORT AND CALIBRATION CERTIFICATE

ATOMIC ABSORPTION SPECTROPHOTOMETER MODEL

PinAAcle 900T

SERIAL NUMBER	PTCS14111103		DATE TESTED	January 6, 2025
Remarks :				
- Neutral Density Filter refer to data sheet				
- Zeeman Ratio = $\frac{\text{Atomic Signal(peak area)}}{\text{Atomic Signal(peak area)+Background Signal(peak area)}}$				
= 0.1635/0.1635+0.1378				
0.542				
This is to certify that the above tests have been performed and the configuration tested				
<input checked="" type="checkbox"/> meets <input type="checkbox"/> does not meet				
the PerkinElmer Specifications listed on this certificate.				
This certificate does not modify PerkinElmer's standrd terms and condition of sale, including warranty terms.				
Service Department PerkinElmer Ltd.  (Wiphan Promlumda) Service Engineer				
Customer Service Engineer:				