

ภาคผนวก จ

เอกสารแสดงการตรวจเทียบเครื่องมือการตรวจวัดคุณภาพสิ่งแวดล้อม



CONTROL UNIT CALIBRATION

(Metric units, mm)

Date 6 Jan 25

Barometric press, Pb

Initial	Final	Average
758	758	758

mmHg

Dry Gas Meter Data

Console No. M50-07

Metering System ID

DGM Number 90331

DGM Model MST-C2-1

Calibrated by Montri P.

Reference Dry Gas Meter Data

Serial No. 358794

Model S110

Correction factor (Yr) 1.0077

Last Calibration Date 25 Oct 24

Orifice manometer setting, ΔH mm H2O	Ref.	DGM Volume V _m Liters	Temperature (°C)			Time Θ min	DGM Correction factor (Y)	ΔH@ mm	
	DGM		Dry Gas Meter						
	Volume V _r Liters		Ref DGM T _r	Inlet T _i	Outlet T _o				Avg T _m
12.5	100.1	101.0	25	25	24	24.5	8.67	0.9958	42.5842
25.0	99.9	100.8	25	25	24	24.5	6.23	0.9946	44.2513
50.0	100.0	100.9	25	25	24	24.5	4.62	0.9920	48.4414
76.0	100.1	99.3	25	25	24	24.5	3.63	1.0074	45.4868
100.0	100.2	100.7	25	25	24	24.5	3.63	0.9921	47.7831
150.0	99.9	99.4	25	25	24	24.5	2.62	0.9970	46.7598

Average

0.9965	45.8844
--------	---------

Approved by :



PITOT TUBE CALIBRATION REPORT

Calibration Location: SECOT

Calibration Date : 04-01-2025

Calibration Duct No.: CD-0123

Calibration Standard Pitot tube data

Pitot No. : Std-02

Coefficient (Cp) : 0.99

Type S Pitot No. : LL10-01

Calibrated by : Mr. Montri P.

A Side Calibration

Run No.	ΔPstd (mm H ₂ O)	ΔPs (mm H ₂ O)	Cp(s)	Deviation, δ Cp(s) - Cp(A)
1	15.0	21.0	0.8367	-0.0068
2	15.0	20.5	0.8468	0.0034
3	15.0	20.5	0.8468	0.0034

C_{P(A),avg} 0.8435

B Side Calibration

Run No.	ΔPstd (mm H ₂ O)	ΔPs (mm H ₂ O)	Cp(s)	Deviation, δ Cp(s) - Cp(B)
1	15.0	20.5	0.8468	0.0034
2	15.0	21.0	0.8367	-0.0068
3	15.0	20.5	0.8468	0.0034

C_{P(B),avg} 0.8435

|CP(A)-CP(B)| = 0.0000

C_{P(Avg)} = 0.8435

Approved by :

*** δ must be ≤ 0.01 for the test to be acceptable ***
*** | Cp(A)-Cp(B) | must also be < 0.01 if average of Cp(A) and Cp(B) is to be used ***



CONTROL UNIT CALIBRATION

(Metric units, mm)

Date 9 Jan 25

Barometric press, Pb

Initial	Final	Average
758	758	758

mmHg

Dry Gas Meter Data

Console No. M50-09

Metering System ID

DGM Number 333249

DGM Model ES-110

Calibrated by : Montri P.

Reference Dry Gas Meter Data

Serial No. 358794

Model S110

Correction factor (Yr) 1.0077

Last Calibration Date 25 Oct 24

Orifice manometer setting, ΔH mm H2O	Ref.	DGM Volume V _m Liters	Temperature (°C)				Time ⊙ min	DGM Correction factor (Y)	ΔH@ mm
	DGM		Ref DGM T _r	Dry Gas Meter					
	Volume V _r Liters			Inlet T _i	Outlet T _o	Avg T _m			
12.5	100.1	101.3	25	25	24	24.5	8.57	0.9926	41.6238
25.0	100.2	100.4	25	25	24	24.5	6.23	1.0012	44.0131
50.0	100.1	100.5	25	25	24	24.5	4.42	0.9965	44.2732
76.0	100.2	99.7	25	25	24	24.5	3.58	1.0037	44.1905
100.0	100.3	99.6	25	25	24	24.5	3.58	1.0034	45.3098
150.0	100.3	99.2	25	25	24	24.5	2.60	1.0029	45.7895

Average

1.0000	44.2000
--------	---------

Approved by :



PITOT TUBE CALIBRATION REPORT

Calibration Location: SECOT

Calibration Duct No.: CD-0123

Calibration Standard Pitot tube data

Pitot No. : Std-02

Type S Pitot No. : LL10-02

Calibration Date : 04-01-2025

Coefficient (Cp) : 0.99

Calibrated by : Mr. Montri P.

A Side Calibration

Run No.	ΔPstd (mm H ₂ O)	ΔPs (mm H ₂ O)	Cp(s)	Deviation, δ Cp(s) - Cp(A)
1	15.0	20.5	0.8468	-0.0035
2	15.0	20.5	0.8468	-0.0035
3	15.0	20.0	0.8574	0.0070

C_{P(A),avg} 0.8504

B Side Calibration

Run No.	ΔPstd (mm H ₂ O)	ΔPs (mm H ₂ O)	Cp(s)	Deviation, δ Cp(s) - Cp(B)
1	15.0	20.5	0.8468	0.0034
2	15.0	21.0	0.8367	-0.0068
3	15.0	20.5	0.8468	0.0034

C_{P(B),avg} 0.8435

|CP(A)-CP(B)| = 0.0069

C_{P(Avg)} = 0.8469

Approved by :

*** δ must be ≤ 0.01 for the test to be acceptable ***
 *** |CP(A)-CP(B)| must also be < 0.01 if average of Cp(A) and Cp(B) is to be used ***

THE LINDE GROUP

Linde

Certificate Of Analysis

Special Gases Mixture

Customer Details

Name: Secot Co., Ltd.
Address: 239 Rimklongprapa Rd. Bangsue Khet Bangsue Bangkok 10800
Customer Tag No.:

Certificate Details

Number: 0483/23 Date of Issue: 22-Feb-2023 Expiry date: 21-Feb-2027
Material Details
Production Order: 90176403 Material Code: 478100-J-62 Cylinder No.: 51108
Gas content: 6.520 M³ (nominal) Filling pressure: 145 bar (g) Valve: CGA 590 BRASS
Cylinder Owner: LINDE Cylinder Material: STEEL Cylinder Size: 47 L

Laboratory Report

Component	Normal Concentration	Analysis Result ¹	Uncertainty ²	Method of Analysis ³
Oxygen	8.00%	7.93%	± 2% relative	(1) SG-O-01
In Nitrogen				

Recommend usage condition

Minimum utilization: 5% of actual content or before expiry date whichever comes first.

Storage condition: Keep in well ventilation and secure area.

Comments

Note:

- All results expressed in this report are on mole/mole basis, unless otherwise specified.
- The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The measurement of this material is traceable to the SI through the reference gas standard which is traceable to Swiss National Standard of Mass or other recognised national metrology institutes.
- (1) Gas Chromatography, (2) Paramagnetic Oxygen Analyzer, (3) Electrochemical Oxygen Analyzer, (4) Electrochemical Moisture Analyzer, (5) Total Hydrocarbon Analyzer, (6) Other - Specified

Sukanya Parinyasoonorn

Signatory for and on behalf of Linde (Thailand) Co., Ltd.

Page 1 of 1

This report shall not be reproduced except in full

PB-002/F004

Iss:K/2, 15 Oct 2021

บริษัท ลินด์ (ประเทศไทย) จำกัด (มหาชน)

P.L.C. Registration No. 0107537000785

ชั้น 15 อาคารทาวเวอร์ 10 2/3 หมู่ 14 ถนนบางนา-ตราด กม. 6.5 แขวงบางนา

เขตคลองเตย กรุงเทพมหานคร 10540 โทรศัพท์ (66) 2338-6100 โทรสาร (66) 2338-6333

โรงงานเวลโกรว์: 105 หมู่ 5 ต.บางนาใหม่ อ.บางนาเขต คลองเตย กรุงเทพฯ 10540

โทรศัพท์ (66) 38.570-479-93

โทรสาร (66) 38.570-323

Linde (Thailand) Public Company Limited

P.L.C. Registration No. 0107537000785

15th Floor, Bangna Tower A, 2/3 Moo 14, Bangna Trad KM. 6.5 Road, Bangkaew

Bangplee, Samutprakarn 10540, Tel (66) 2338-6100 Fax (66) 2338-6333

Wellgrow Plant: 105 Moo 5, T.Bangnamak, A.Bangpakong, Chachoengsao 24180

Thailand, Tel (66) 38.570-479-93

Fax (66) 38.570-323

THE LINDE GROUP

Linde

Certificate Of Analysis

Special Gases Mixture

Customer Details

Name: Secot Co., Ltd.
Address: 239, Rimklongprapa Rd., Bangsue, Bangkok 10800
Customer Tag No.:

Certificate Details

Number: 0527/23 Date of Issue: 8-Mar-2023 Expiry date: 8-Mar-2026
Material Details
Production Order: 90176406 Material Code: 511600-SK-34 Cylinder No.: A00878SK
Gas content: 5.20 M³ Filling pressure: 137.0 bar Valve: CGA 660 SS
Cylinder Owner: LINDE Cylinder Material: Spectra seal Cylinder Size: 40 L

Laboratory Report

Analytical Result

Component	Normal Concentration	Analysis Result ¹	Uncertainty ²	Method of Analysis ³	Assay Date
Nitric Oxide	40.0 ppm	39.8 ppm	± 1% relative	(6) I-PB-352	1-Mar & 8-Mar-23
Other NOx impurity		Less than 1.9 ppm			
Carbon Monoxide	40.0 ppm	42.0 ppm	± 1% relative	(6) I-PB-352	1-Mar-2023
In Nitrogen					

Reference Standard used in Assay

Reference Standard	Cylinder number	Concentration	Expiry date:
Nitric Oxide	1332615G	25.61 ± 0.13 ppm	6-May-2023
Carbon Monoxide	ND52320	25.03 ± 0.13 ppm	7-Oct-2023
In Nitrogen			

Analytical Instruments used in Assay

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
FTIR Spectrometers Nicolet iS50	FTIR-NO	28-Feb-2023
FTIR Spectrometers Nicolet iS50	FTIR-CO	25-Feb-2023

Recommend usage condition

Minimum utilization: 5% of actual content or before expiry date whichever comes first.

Storage condition: Keep in well ventilation and secure area.

Comments

When reordering, please quote the material number

Note:

- All results expressed in this report are on mole/mole basis, unless otherwise specified. The Assay of this Standard has been performed in accordance with the EPA Traceability Protocol EPA-600/R-12/531 for the Assay and Certification of Gaseous Calibration Standards using procedure G1
- The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The measurement of this material is traceable to the SI through the reference gas standard which is traceable to Swiss National Standard of Mass or other recognised national metrology institutes.
- (1) Gas Chromatography, (2) Paramagnetic Oxygen Analyzer, (3) Electrochemical Oxygen Analyzer, (4) Electrochemical Moisture Analyzer, (5) Total Hydrocarbon Analyzer, (6) Other - Specified

Page 1 of 1

This report shall not be reproduced except in full

บริษัท ลินด์ (ประเทศไทย) จำกัด (มหาชน)

P.L.C. Registration No. 0107537000785

ชั้น 15 อาคารทาวเวอร์ 10 2/3 หมู่ 14 ถนนบางนา-ตราด กม. 6.5 แขวงบางนา

เขตคลองเตย กรุงเทพมหานคร 10540 โทรศัพท์ (66) 2338-6100 โทรสาร (66) 2338-6333

โรงงานเวลโกรว์: 105 หมู่ 5 ต.บางนาใหม่ อ.บางนาเขต คลองเตย กรุงเทพฯ 10540

โทรศัพท์ (66) 38.570-479-93

โทรสาร (66) 38.570-323

Sukanya Parinyasoonorn

Signatory for and on behalf of Linde (Thailand) Co., Ltd.

PB-002/F006

Linde (Thailand) Public Company Limited

P.L.C. Registration No. 0107537000785

15th Floor, Bangna Tower A, 2/3 Moo 14, Bangna Trad KM. 6.5 Road, Bangkaew

Bangplee, Samutprakarn 10540, Tel (66) 2338-6100 Fax (66) 2338-6333

Wellgrow Plant: 105 Moo 5, T.Bangnamak, A.Bangpakong, Chachoengsao 24180

Thailand, Tel (66) 38.570-479-93

Fax (66) 38.570-323

THE LINDE GROUP

Linde

Certificate Of Analysis
Special Gases Mixture

Customer Details

Name: Secot Co., Ltd. Address: 239, Rimklongprapa Rd., Bangsue, Barigkok 10800 Customer Tag No.:

Certificate Details

Number: 0742/23 Date of Issue: 29-Mar-2023 Expiry date: 29-Mar-2027
Material Details
Production Order: 90176408 Material Code: 608400-SK-44 Cylinder No.: A00940SK
Gas content: 5.52 M³ Filling pressure: 145.0 bar Valve: CGA 660 SS
Cylinder Owner: LINDE Cylinder Material: Spectra seal Cylinder Size: 40 L

Laboratory Report

Analytical Result

Component	Nominal Concentration	Analysis Result ¹	Uncertainty ²	Method of Analysis ³	Assay Date
Sulphur Dioxide In Nitrogen	40.0 ppm	40.2 ppm	± 1% relative	(6) I-PB-352	22-Mar & 29-Mar-23

Reference Standard used in Assay

Reference Standard	Cylinder number	Concentration	Expiry date:
Sulphur Dioxide In Nitrogen	256240	52.73 ± 0.42 ppm	6-May-2023

Analytical Instruments used in Assay

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
FTIR Spectrometers Nicolet i550	FTIR-SO2	17-Mar-2023

Recommend usage condition

Minimum utilization: 5% of actual content or before expire date whichever comes first.
Storage condition: Keep in well ventilation and secure area.

Comments

When reordering, please quote the material number

Note:

1. All results expressed in this report are on mole/mole basis, unless otherwise specified. The Assay of this Standard has been performed in accordance with the EPA Traceability Protocol EPA-600/R-12/531 for the Assay and Certification of Gaseous Calibration Standards using procedure G1
2. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The measurement of this material is traceable to the SI through the reference gas standard which is traceable to Swiss National Standard of Mass or other recognised national metrology institutes.
3. (1) Gas Chromatography, (2) Paramagnetic Oxygen Analyzer, (3) Electrochemical Oxygen Analyzer, (4) Electrochemical Moisture Analyzer, (5) Total Hydrocarbon Analyzer, (6) Other - Specified

Page 1 of 1

This report shall not be reproduced except in full

บริษัท ลินด์ (ประเทศไทย) จำกัด (มหาชน)

เลขที่ใบแจ้งหนี้: 010753/000785

ชั้น 15 อาคารทาวเวอร์ เอ 2/3 หมู่ 14 ถนนลาดพร้าว-คลองหลวง กม. 6.5 แขวงคลองหลวง

บางพลี, สมุทรปราการ 10540 โทรศัพท์ (66) 2338-6100 โทรสาร (66) 2338-6333

โรงงานผลิต: 105 หมู่ 5 ตำบลบางพลี อำเภอบางพลี จังหวัดสมุทรปราการ 24180

โทรศัพท์ (66) 38.570-479-93

โทรสาร (66) 38.570-323

Sukanya Parinyasoonorn

Signatory for and on behalf of Linde (Thailand) Co., Ltd.

Linde (Thailand) Public Company Limited

PB-002/F006

P.L.C. Registration no. 010753/000785

15th Floor, Bangna Tower A, 2/3 Moo 14, Bangna Trad KM. 6.5 Road, Bangnae

Bangplee, Samutprakarn 10540, Tel (66) 2338-6100 Fax (66) 2338-6333

Wellgrow Plant : 105 Moo 5, T.Bangsamak, A.Bangpakong, Chachoengsao 24180

Thailand, Tel (66) 38.570-479-93

Fax (66) 38.570-323

Making our world more productive

Linde

Certificate Of Analysis
Special Gases Mixture

Customer Details

Name: Secot Co., Ltd. Address: 239, Rimklongprapa Rd., Bangsue, Bangkok 10800 Customer Tag No.:

Certificate Details

Number: 1393/24 Date of Issue: 24-May-2024 Expiry date: 24-May-2027
Material Details
Production Order: 90183672 Material Code: 436700-SK-34 Cylinder No.: A00987SK
Gas content: 6.900 M³ Filling pressure: 145 bar Valve: CGA 660 SS
Cylinder Owner: LINDE Cylinder Material: Spectra seal Cylinder Size: 40 L

Laboratory Report

Analytical Result

Component	Nominal Concentration	Analysis Result ¹	Uncertainty ²	Method of Analysis ³	Assay Date
Nitric Oxide	80.0 ppm	79.0 ppm	± 1% relative	(6) I-PB-352	15 & 23-May-2024
Other NOx impurity		Less Than 3.9 ppm			
Carbon Monoxide In Nitrogen	80.0 ppm	81.7 ppm	± 1% relative	(6) I-PB-352	15 & 23-May-2024

Reference Standard used in Assay

Reference Standard	Cylinder number	Concentration	Expiry date:
Carbon Monoxide	D619725	70.6 ± 0.2 ppm	20-Sep-2026
Nitric Oxide In Nitrogen	D619725	70.6 ± 0.2 ppm	20-Sep-2026

Analytical Instruments used in Assay

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
FTIR Spectrometers Nicolet i550	FTIR-CO	7-May-2024
FTIR Spectrometers Nicolet i550	FTIR-NO	7-May-2024

Recommend usage condition

Minimum utilization: 5% of actual content or before expire date whichever comes first.
Storage condition: Keep in well ventilation and secure area.

Comments

When reordering, please quote the material number

Note:

1. All results expressed in this report are on mole/mole basis, unless otherwise specified. The Assay of this Standard has been performed in accordance with the EPA Traceability Protocol EPA-600/R-12/531 for the Assay and Certification of Gaseous Calibration Standards using procedure G1
2. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The measurement of this material is traceable to the SI through the reference gas standard which is traceable to Swiss National Standard of Mass or other recognised national metrology institutes.
3. (1) Gas Chromatography, (2) Paramagnetic Oxygen Analyzer, (3) Electrochemical Oxygen Analyzer, (4) Electrochemical Moisture Analyzer, (5) Total Hydrocarbon Analyzer, (6) Other - Specified

Page 1 of 1

This report shall not be reproduced except in full

Sukanya Parinyasoonorn

Signatory for and on behalf of Linde (Thailand) Co., Ltd.

PB-002/F006
Iss-M/1, 01 December 2023

บริษัท ลินด์ (ประเทศไทย) จำกัด (มหาชน)

เลขที่ใบแจ้งหนี้: 010753/000785

ชั้น 15 อาคารทาวเวอร์ เอ 2/3 หมู่ 14 ถนนลาดพร้าว-คลองหลวง กม. 6.5 แขวงคลองหลวง

บางพลี, สมุทรปราการ 10540 โทรศัพท์ (66) 2338-6100 โทรสาร (66) 2338-6333

โรงงานผลิต: 105 หมู่ 5 ตำบลบางพลี อำเภอบางพลี จังหวัดสมุทรปราการ 24180

โทรศัพท์ (66) 38.570-479-93

โทรสาร (66) 38.570-323

Linde (Thailand) Public Company Limited

P.L.C. Registration no. 010753/000785

15th Floor, Bangna Tower A, 2/3 Moo 14, Bangna Trad KM. 6.5 Road, Bangnae

Bangplee, Samutprakarn 10540, Tel (66) 2338-6100 Fax (66) 2338-6333

Wellgrow Plant : 105 Moo 5, T.Bangsamak, A.Bangpakong, Chachoengsao 24180

Thailand, Tel (66) 38.570-479-93

Fax (66) 38.570-323

THE LINDE GROUP

Linde

Certificate Of Analysis
Special Gases Mixture

Customer Details

Name: Secot Co., Ltd. Address: 239, Rimklongprapa Rd., Bangsue, Bangkok 10800 Customer Tag No.:

Certificate Details

Number: 0273/22 Date of Issue: 4-Feb-2022 Expiry date: 4-Feb-2030
Material Details
Production Order: 90169723 Material Code: 445100-SK-44 Cylinder No.: D636047
Gas content: 5.52 M³ Filling pressure: 145.0 bar Valve: CGA 660 SS
Cylinder Owner: LINDE Cylinder Material: Spectra seal Cylinder Size: 40 L

Laboratory Report

Analytical Result

Component	Normal Concentration	Analysis Result ¹	Uncertainty ²	Method of Analysis ³	Assay Date
Sulphur Dioxide In Nitrogen	80.0 ppm	81.0 ppm	± 1% relative	(6) I-PB-352	28-Jan & 4-Feb-22

Reference Standard used in Assay

Reference Standard	Cylinder number	Concentration	Expiry date
Sulphur Dioxide In Nitrogen	256240	52.73 ± 0.42 ppm	6-May-2023

Analytical Instruments used in Assay

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
FTIR Spectrometers Nicolet iS50	FTIR-SO2	10-Jan-2022

Recommend usage condition

Minimum utilization: 5% of actual content or before expiry date whichever comes first.
Storage condition: Keep in well ventilation and secure area.

Comments

When reordering, please quote the material number

Note:

1. All results expressed in this report are on mole/mole basis, unless otherwise specified. The Assay of this Standard has been performed in accordance with the EPA Traceability Protocol EPA-600/R-12/531 for the Assay and Certification of Gaseous Calibration Standards using procedure G1.
2. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The measurement of this material is traceable to the SI through the reference gas standard which is traceable to Swiss National Standard of Mass or other recognised national metrology institutes.
3. (1) Gas Chromatography, (2) Paramagnetic Oxygen Analyzer, (3) Electrochemical Oxygen Analyzer, (4) Electrochemical Moisture Analyzer, (5) Total Hydrocarbon Analyzer, (6) Other - Specified

Page 1 of 1

This report shall not be reproduced except in full

บริษัท ลินด์ (ประเทศไทย) จำกัด (มหาชน)

เลขที่ใบขึ้นทะเบียนมาตรฐาน 0107537000785

ชั้น 15 อาคารทาวเวอร์ เอ 2/3 หมู่ 14 ถนนบางนา-ตราด กม. 6.5 แขวงคลอง

บางพลี เขตอุตสาหกรรม 10540 กรุงเทพมหานคร โทร (66) 2338-6100 โทรสาร (66) 2338-6333

โรงงานผลิตก๊าซ: 105 หมู่ 5 ตำบลบางพลีใหญ่ อำเภอบางพลี จังหวัดสมุทรปราการ 24180

โทรศัพท์ (66) 38.570-479-93

โทรสาร (66) 38.570-323

Linde (Thailand) Public Company Limited

PLC Registration no. 0107537000785

15th Floor, Bangna Tower A, 2/3 Moo 14, Bangna Trad KM. 6.5 Road, Bangkaew

Bangplee, Samutprakarn 10540, Tel (66) 2338-6100 Fax (66) 2338-6333

Wellgrow Plant: 105 Moo 5, T.Bangsamak, A.Bangpakong, Chachoengsao 24180

Thailand, Tel (66) 38.570-479-93

Fax (66) 38.570-323

PB-002/F006

Iss:K/2, 15 Oct 2021

Sukanya Parinyasoonorn

Signatory for and on behalf of Linde (Thailand) Co., Ltd.

THE LINDE GROUP

Linde

Certificate Of Analysis
Special Gases Mixture

Customer Details

Name: Secot Co., Ltd. Address: 239 Rimklongprapa Rd. Bangsue Khet Bangsue Bangkok 10800 Customer Tag No.:

Certificate Details

Number: 0484/23 Date of Issue: 22-Feb-2023 Expiry date: 21-Feb-2027
Material Details
Production Order: 90176403 Material Code: 478100-J-62 Cylinder No.: 12360
Gas content: 6.520 M³ (nominal) Filling pressure: 145 bar (g) Valve: CGA 590 BRASS
Cylinder Owner: LINDE Cylinder Material: STEEL Cylinder Size: 47 L

Laboratory Report

Component	Normal Concentration	Analysis Result ¹	Uncertainty ²	Method of Analysis ³
Oxygen In Nitrogen	8.00%	7.94%	± 2% relative	(1) SG-O-01

Recommend usage condition

Minimum utilization: 5% of actual content or before expiry date whichever comes first.
Storage condition: Keep in well ventilation and secure area.

Comments

Note:

1. All results expressed in this report are on mole/mole basis, unless otherwise specified.
2. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The measurement of this material is traceable to the SI through the reference gas standard which is traceable to Swiss National Standard of Mass or other recognised national metrology institutes.
3. (1) Gas Chromatography, (2) Paramagnetic Oxygen Analyzer, (3) Electrochemical Oxygen Analyzer, (4) Electrochemical Moisture Analyzer, (5) Total Hydrocarbon Analyzer, (6) Other - Specified

Sukanya Parinyasoonorn

Signatory for and on behalf of Linde (Thailand) Co., Ltd.

Page 1 of 1

This report shall not be reproduced except in full

PB-002/F004

Iss:K/2, 15 Oct 2021

บริษัท ลินด์ (ประเทศไทย) จำกัด (มหาชน)

เลขที่ใบขึ้นทะเบียนมาตรฐาน 0107537000785

ชั้น 15 อาคารทาวเวอร์ เอ 2/3 หมู่ 14 ถนนบางนา-ตราด กม. 6.5 แขวงคลอง

บางพลี เขตอุตสาหกรรม 10540 กรุงเทพมหานคร โทร (66) 2338-6100 โทรสาร (66) 2338-6333

โรงงานผลิตก๊าซ: 105 หมู่ 5 ตำบลบางพลีใหญ่ อำเภอบางพลี จังหวัดสมุทรปราการ 24180

โทรศัพท์ (66) 38.570-479-93

โทรสาร (66) 38.570-323

Linde (Thailand) Public Company Limited

PLC Registration no. 0107537000785

15th Floor, Bangna Tower A, 2/3 Moo 14, Bangna Trad KM. 6.5 Road, Bangkaew

Bangplee, Samutprakarn 10540, Tel (66) 2338-6100 Fax (66) 2338-6333

Wellgrow Plant: 105 Moo 5, T.Bangsamak, A.Bangpakong, Chachoengsao 24180

Thailand, Tel (66) 38.570-479-93

Fax (66) 38.570-323

Certificate Of Analysis
Special Gases Mixture

Customer Details

Name: Address: Customer Tag No.:
Secot Co., Ltd. 239, Rimklongprapa Rd., Bangsue, Bangkok 10800

Certificate Details

Number: 0528/23 Date of Issue: 8-Mar-2023 Expiry date: 8-Mar-2026
Material Details
Production Order: 90176406 Material Code: 511600-SK-34 Cylinder No.: A007225K
Gas content: 5.20 M³ Filling pressure: 137.0 bar Valve: CGA 660 SS
Cylinder Owner: LINDE Cylinder Material: Spectra seal Cylinder Size: 40 L

Laboratory Report

Analytical Result

Component	Normal Concentration	Analysis Result ¹	Uncertainty ²	Method of Analysis ³	Assay Date
Nitric Oxide	40.0 ppm	39.6 ppm	± 1% relative	(6) I-PB-352	1-Mar & 8-Mar-23
Other NOx impurity		Less than 1.9 ppm			
Carbon Monoxide in Nitrogen	40.0 ppm	41.9 ppm	± 1% relative	(6) I-PB-352	1-Mar-2023

Reference Standard used in Assay

Reference Standard	Cylinder number	Concentration	Expiry date:
Nitric Oxide	133261SG	25.61 ± 0.13 ppm	6-May-2023
Carbon Monoxide in Nitrogen	ND52320	25.03 ± 0.13 ppm	7-Oct-2023

Analytical Instruments used in Assay

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
FTIR Spectrometers Nicolet iS50	FTIR-NO	28-Feb-2023
FTIR Spectrometers Nicolet iS50	FTIR-CO	25-Feb-2023

Recommend usage condition

Minimum utilization: 5% of actual content or before expiry date whichever comes first.
Storage condition: Keep in well ventilation and secure area.

Comments

When reordering, please quote the material number

Note:

- All results expressed in this report are on mole/mole basis, unless otherwise specified. The Assay of this Standard has been performed in accordance with the EPA Traceability Protocol EPA-600/R-12/531 for the Assay and Certification of Gaseous Calibration Standards using procedure G1
- The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The measurement of this material is traceable to the SI through the reference gas standard which is traceable to Swiss National Standard of Mass or other recognised national metrology institutes.
- (1) Gas Chromatography, (2) Paramagnetic Oxygen Analyzer, (3) Electrochemical Oxygen Analyzer, (4) Electrochemical Moisture Analyzer, (5) Total Hydrocarbon Analyzer, (6) Other - Specified

Sukanya Parinyasoontorn
Signatory for and on behalf of Linde (Thailand) Co., Ltd.

Certificate Of Analysis
Special Gases Mixture

Customer Details

Name: Address: Customer Tag No.:
Secot Co., Ltd. 239, Rimklongprapa Rd., Bangsue, Bangkok 10800

Certificate Details

Number: 0275/22 Date of Issue: 4-Feb-2022 Expiry date: 4-Feb-2026
Material Details
Production Order: 90169722 Material Code: 631500-SK-44 Cylinder No.: D636195
Gas content: 5.52 M³ Filling pressure: 145.0 bar Valve: CGA 660 SS
Cylinder Owner: LINDE Cylinder Material: Spectra seal Cylinder Size: 40 L

Laboratory Report

Analytical Result

Component	Normal Concentration	Analysis Result ¹	Uncertainty ²	Method of Analysis ³	Assay Date
Sulphur Dioxide in Nitrogen	20.0 ppm	20.4 ppm	± 1% relative	(6) I-PB-352	28-Jan & 4-Feb-22

Reference Standard used in Assay

Reference Standard	Cylinder number	Concentration	Expiry date:
Sulphur Dioxide in Nitrogen	145754SG	25.03 ± 0.25 ppm	18-Aug-2022

Analytical Instruments used in Assay

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
FTIR Spectrometers Nicolet iS50	FTIR-SO2	27-Jan-2022

Recommend usage condition

Minimum utilization: 5% of actual content or before expiry date whichever comes first.
Storage condition: Keep in well ventilation and secure area.

Comments

When reordering, please quote the material number

Note:

- All results expressed in this report are on mole/mole basis, unless otherwise specified. The Assay of this Standard has been performed in accordance with the EPA Traceability Protocol EPA-600/R-12/531 for the Assay and Certification of Gaseous Calibration Standards using procedure G1
- The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The measurement of this material is traceable to the SI through the reference gas standard which is traceable to Swiss National Standard of Mass or other recognised national metrology institutes.
- (1) Gas Chromatography, (2) Paramagnetic Oxygen Analyzer, (3) Electrochemical Oxygen Analyzer, (4) Electrochemical Moisture Analyzer, (5) Total Hydrocarbon Analyzer, (6) Other - Specified

Sukanya Parinyasoontorn
Signatory for and on behalf of Linde (Thailand) Co., Ltd.

Making our world more productive



Certificate Of Analysis

Special Gases Mixture

Customer Details

Name: Secot Co., Ltd. Address: 239, Rimklongprapa Rd., Bangsue, Bangkok 10800 Customer Tag No.:

Certificate Details

Number: 1394/24 Date of Issue: 24-May-2024 Expiry date: 24-May-2027
Material Details
Production Order: 90183672 Material Code: 436700-SK-34 Cylinder No.: A00896SK
Gas content: 5.230 M³ Filling pressure: 137 bar Valve: CGA 660 SS
Cylinder Owner: LINDE Cylinder Material: Spectra seal Cylinder Size: 40 L

Laboratory Report

Analytical Result

Component	Nominal Concentration	Analysis Result ¹	Uncertainty ²	Method of Analysis ³	Assay Date
Nitric Oxide	80.0 ppm	79.6 ppm	± 1% relative	(6) I-PB-352	15823-May-2024
Other NO _x impurity		Less Than 3.9 ppm			
Carbon Monoxide In Nitrogen	80.0 ppm	81.7 ppm	± 1% relative	(6) I-PB-352	15823-May-2024

Reference Standard used in Assay

Reference Standard	Cylinder number	Concentration	Expiry date
Carbon Monoxide	D619725	70.6 ± 0.2 ppm	20-Sep-2026
Nitric Oxide In Nitrogen	D619725	70.6 ± 0.2 ppm	20-Sep-2026

Analytical Instruments used in Assay

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
FTIR Spectrometers Nicolet iS50	FTIR-CO	7-May-2024
FTIR Spectrometers Nicolet iS50	FTIR-NO	7-May-2024

Recommend usage condition

Minimum utilization: 5% of actual content or before expire date whichever comes first.
Storage condition: Keep in well ventilation and secure area.

Comments

When reordering, please quote the material number

Note:

- All results expressed in this report are on mole/mole basis, unless otherwise specified. The Assay of this Standard has been performed in accordance with the EPA Traceability Protocol EPA-600/R-12/531 for the Assay and Certification of Gaseous Calibration Standards using procedure G1
- The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The measurement of this material is traceable to the SI through the reference gas standard which is traceable to Swiss National Standard of Mass or other recognised national metrology institutes.
- (1) Gas Chromatography, (2) Paramagnetic Oxygen Analyzer, (3) Electrochemical Oxygen Analyzer, (4) Electrochemical Moisture Analyzer, (5) Total Hydrocarbon Analyzer, (6) Other - Specified

Page 1 of 1

This report shall not be reproduced except in full

Sukanya Parinyasoontorn
Signatory for and on behalf of Linde (Thailand) Co., Ltd.

PB-002/F006
Iss-M/1, 01 December 2023

บริษัท ลินด์ (ประเทศไทย) จำกัด (มหาชน)

เลขที่จดทะเบียนการค้า: 010731000783

ชั้น 15 อาคารทาวเวอร์ เอ 2/3 หมู่ 14 ถนนบางนา-ตราด กม. 6.5 แขวงบางนา
อ.บางนาเขตสมุทรปราการ 10540 โทรศัพท์ (66) 2338-6100 โทรสาร (66) 2338-6333
โทรสารและโทรสาร: 105 หมู่ 5 แขวงบางนา เขตบางนา 24180
โทรศัพท์ (66) 38.570-479-93 โทรสาร (66) 38.570-323

Linde (Thailand) Public Company Limited

P.L.C. Registration No. 010731000783

15th Floor, Bangna Tower A, 2/3 Moo 14, Bangna Trad Km. 6.5 Road, Bangkaew
Bangplee, Samutprakarn 10540, Tel (66) 2338-6100 Fax (66) 2338-6333
Wellgrow Plant: 105 Moo 5, T.Bangsamak, A.Bangpakong, Chachoengsao 24180
Thailand, Tel (66) 38.570-479-93 Fax (66) 38.570-323

THE LINDE GROUP

Certificate Of Analysis

Special Gases Mixture

Customer Details

Name: Secot Co., Ltd. Address: 239 Rimklongprapa Rd., Bangsue, Bangkok 10800 Customer Tag No.:

Certificate Details

Number: 0610/23 Date of Issue: 21-Mar-2023 Expiry date: 21-Mar-2031
Material Details
Production Order: 90176409 Material Code: 445100-SK-44 Cylinder No.: D869384
Gas content: 5.52 M³ Filling pressure: 145.0 bar Valve: CGA 660 SS
Cylinder Owner: LINDE Cylinder Material: Spectra seal Cylinder Size: 40 L

Laboratory Report

Analytical Result

Component	Nominal Concentration	Analysis Result ¹	Uncertainty ²	Method of Analysis ³	Assay Date
Sulphur Dioxide In Nitrogen	80.0 ppm	83.5 ppm	± 1% relative	(6) I-PB-352	14-Mar & 21-Mar-23

Reference Standard used in Assay

Reference Standard	Cylinder number	Concentration	Expiry date
Sulphur Dioxide In Nitrogen	256240	52.73 ± 0.42 ppm	6-May-2023

Analytical Instruments used in Assay

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
FTIR Spectrometers Nicolet iS50	FTIR-SO2	16-Feb & 17-Mar-23

Recommend usage condition

Minimum utilization: 5% of actual content or before expire date whichever comes first.
Storage condition: Keep in well ventilation and secure area.

Comments

When reordering, please quote the material number

Note:

- All results expressed in this report are on mole/mole basis, unless otherwise specified. The Assay of this Standard has been performed in accordance with the EPA Traceability Protocol EPA-600/R-12/531 for the Assay and Certification of Gaseous Calibration Standards using procedure G1
- The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The measurement of this material is traceable to the SI through the reference gas standard which is traceable to Swiss National Standard of Mass or other recognised national metrology institutes.
- (1) Gas Chromatography, (2) Paramagnetic Oxygen Analyzer, (3) Electrochemical Oxygen Analyzer, (4) Electrochemical Moisture Analyzer, (5) Total Hydrocarbon Analyzer, (6) Other - Specified

Page 1 of 1

This report shall not be reproduced except in full

บริษัท ลินด์ (ประเทศไทย) จำกัด (มหาชน)

เลขที่จดทะเบียนการค้า: 010731000783

ชั้น 15 อาคารทาวเวอร์ เอ 2/3 หมู่ 14 ถนนบางนา-ตราด กม. 6.5 แขวงบางนา
อ.บางนาเขตสมุทรปราการ 10540 โทรศัพท์ (66) 2338-6100 โทรสาร (66) 2338-6333
โทรสารและโทรสาร: 105 หมู่ 5 แขวงบางนา เขตบางนา 24180
โทรศัพท์ (66) 38.570-479-93 โทรสาร (66) 38.570-323

Sukanya Parinyasoontorn
Signatory for and on behalf of Linde (Thailand) Co., Ltd.

Linde (Thailand) Public Company Limited

P.L.C. Registration No. 010731000783

15th Floor, Bangna Tower A, 2/3 Moo 14, Bangna Trad Km. 6.5 Road, Bangkaew
Bangplee, Samutprakarn 10540, Tel (66) 2338-6100 Fax (66) 2338-6333
Wellgrow Plant: 105 Moo 5, T.Bangsamak, A.Bangpakong, Chachoengsao 24180
Thailand, Tel (66) 38.570-479-93 Fax (66) 38.570-323

CERTIFICATE OF ANALYSIS

Grade of Product: EPA Protocol

Part Number: E04NI99E15AC084 Reference Number: 82-401409170-1
Cylinder Number: EB0102326 Cylinder Volume: 144.4 CF
Laboratory: 124 - Riverton (SAP) - NJ Cylinder Pressure: 2015 PSIG
PGVP Number: B52019 Valve Outlet: 660
Gas Code: CO,NO,NOX,SO2,BALN Certification Date: Feb 05, 2019

Expiration Date: Feb 05, 2027

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 800/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.

Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

ANALYTICAL RESULTS					
Component	Requested Concentration	Actual Concentration	Protocol Method	Total Relative Uncertainty	Assay Dates
NOX	50.00 PPM	51.01 PPM	G1	+/- 0.9% NIST Traceable	01/28/2019, 02/05/2019
NITRIC OXIDE	50.00 PPM	50.86 PPM	G1	+/- 0.9% NIST Traceable	01/28/2019, 02/05/2019
SULFUR DIOXIDE	50.00 PPM	50.87 PPM	G1	+/- 1.0% NIST Traceable	01/28/2019, 02/05/2019
CARBON MONOXIDE	0.5000 %	0.5050 %	G1	+/- 0.7% NIST Traceable	01/31/2019
NITROGEN	Balance				

CALIBRATION STANDARDS					
Type	Lot ID	Cylinder No	Concentration	Uncertainty	Expiration Date
NTRM	13080206	CC401947	4950 PPM CARBON MONOXIDE/NITROGEN	+/- 0.4%	Feb 15, 2019
PRM	12367	APEX1099237	9.82 PPM NITROGEN DIOXIDE/AIR	+/- 2.0%	Jun 02, 2017
NTRM	12010724	KAL004497	50.03 PPM NITRIC OXIDE/NITROGEN	+/- 0.8%	Mar 12, 2024
GMIS	1114201601	CC506710	4.971 PPM NITROGEN DIOXIDE/NITROGEN	+/- 2.0%	Nov 14, 2019
NTRM	14010327	KAL004376	49.08 PPM SULFUR DIOXIDE/NITROGEN	+/- 1.0%	Apr 17, 2024

The SRM, PRM or RGM noted above is only in reference to the GMIS used in the assay and not part of the analysis.

ANALYTICAL EQUIPMENT		
Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
Siemens Ultramat 6 J3-599 COHIGH	NDIR	Jan 18, 2019
Nicolet 6700 APW1100391 NO	FTIR	Jan 10, 2019
Nicolet 6700 APW1100391 NO2	FTIR	Jan 10, 2019
Nicolet 6700 APW1100391 SO2	FTIR	Jan 10, 2019

Triad Data Available Upon Request

PERMANENT NOTES: PRODUCED IN ACCORDANCE WITH ISO17025 REQUIREMENTS

NOTES:

Gross Weight: 27806.3 grams

Net Weight: 4733.2 grams

This calibration std. has been certified in accordance with the May 2012 EPA Traceability Protocol Document EPA-600/R-12/531. All testing processes and measurements conform to the requirements of ISO/IEC 17025 and to Airgas ISO 9001:2008 and relate only to items identified on this certificate. All measurements are certified to be NIST Traceable with total uncertainty as detailed under Analytical Uncertainty. This document shall not be reproduced in full without written approval of the issuer.



TESTING CERT No. 3082.05

[Signature]
Approved for Release

Page 1 of 82-401409170-1



SOUND LEVEL METER CALIBRATION

Calibration Location: SECOT

Calibration Date: Feb 6, 25

ACOUSTIC CALIBRATOR

Brand	Model	Serial No.	Frequency (Hz)	Ref. Calibrated (dB)	Eff. Calibrated (dB)
Cirrus	CR:515	97097	1000.00	94.0	93.7

No.	Brand	Model	Serial No.	Reading (dB)	dB Adjust
14	Cirrus	CR162B	G300709	94.6	-0.9
24	Cirrus	CR162C	G300832	93.7	0.0
26	Cirrus	CR162C	G300841	93.1	0.6
27	Cirrus	CR162C	G301029	93.7	0.0
28	Cirrus	CR162C	G301065	94.0	-0.3
40	Cirrus	CR162B	G302740	94.0	-0.3
43	Cirrus	CR162B	G302741	93.6	0.1
50	Cirrus	CR162B	G302333	93.7	0.0

Calibrated by :

[Signature]

Approved by :

[Signature]

CERTIFICATE OF CALIBRATION

ISSUED BY Noisemeters

DATE OF ISSUE 29 April 2024

CERTIFICATE NUMBER 213338

NoiseMeters

NoiseMeters
Acoustic House
Bridlington Road
Hunmanby
YO14 0PH
United Kingdom
www.noisemeters.com

Page 1 of 2

Approved signatory
N.Smith
Electronically signed:



doseBadge Reader : IEC 60942:2003

Instrument information

Manufacturer: Pulsar Instruments

Notes:

Model: Model 22R

Serial number: 79781

Class: 2

Test summary

Date of calibration: 29 April 2024

The doseBadge reader detailed above has been calibrated to the published data as described in the operating manual and in the half-inch configuration. The procedures and techniques used are as described in IEC60942_2003 Annex B – Periodic Tests and three determinations of the sound pressure level, frequency and total distortion were made.

The sound pressure level was measured using a WS2F condenser microphone type MK:224 manufactured by Cirrus Research plc.

The results have been corrected to the reference pressure of 101.33 kPa using the manufacturer's data.

The doseBadge Reader has been shown to conform to the Class 2 requirements for periodic testing, described in Annex B of IEC 60942:2003 for the sound pressure level(s) and frequency(ies) stated, for the environmental conditions under which the tests were performed.

However, as public evidence was not available, from a testing organisation responsible for pattern approval, to demonstrate that the model of doseBadge Reader conformed to the requirements for pattern evaluation described in Annex A of IEC 60942:2003, no general statement or conclusion can be made about conformance of the doseBadge Reader to the requirements of IEC 60942:2003.

Notes:

This certificate provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory. The results within this certificate relate only to the items calibrated. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%.

CERTIFICATE OF CALIBRATION

Certificate Number:
213338

Page 2 of 2

Environmental conditions

The following conditions were recorded at the time of the test:

Before Pressure: 100.34 kPa Temperature: 22.4 °C Humidity: 38.5 %

After Pressure: 100.34 kPa Temperature: 22.7 °C Humidity: 36.3 %

Test equipment

Equipment	Manufacturer	Model	Serial number
Distortion Meter	Keithley	2015	0839263
Acoustic Calibrator	Bruel and Kjaer	4231	2610257
Environmental Monitor	Comet	T7510	21962628

Initial Acoustic Results

	Expected	Sample 1	Sample 2	Sample 3	Average	Deviation	Tolerance	Uncertainty
Level (dB)	114.00	113.94	113.93	113.94	113.94	-0.06	±0.75	0.11 dB
Distortion (%)	< 4.00	0.49	0.49	0.49	0.49	0.49	+4.00	0.13 %
Frequency (Hz)	1000.0	998.9	998.9	998.9	998.9	-1.1	±20.0	0.1 Hz

The measured quantities or deviations (as applicable), extended by the expanded combined uncertainty of measurement, must not exceed the corresponding tolerance.

Adjusted Acoustic Results

	Expected	Sample 1	Sample 2	Sample 3	Average	Deviation	Tolerance	Uncertainty
Level (dB)	114.00	114.01	113.98	113.98	113.99	-0.01	±0.75	0.11 dB
Distortion (%)	< 4.00	0.49	0.49	0.49	0.49	0.49	+4.00	0.13 %
Frequency (Hz)	1000.0	998.9	999.0	998.9	998.9	-1.1	±20.0	0.1 Hz

Functionality Results

Function	Result
Keypad	Pass
Battery Power	Pass
Display	Pass
Communication	Pass
2 way IR link	Pass
Clock	Pass

End of results

CERTIFICATE OF CALIBRATION

ISSUED BY Noisemeters

DATE OF ISSUE 26 February 2025 CERTIFICATE NUMBER 234084

NoiseMeters

NoiseMeters
Acoustic House
Bridlington Road
Hunmanby
YO14 0PH
United Kingdom
www.noisemeters.com

Page 1 of 2

Approved signatory
N.Smith
Electronically signed:



doseBadge Reader : IEC 60942:2003

Instrument information

Manufacturer: Cirrus Research plc

Notes:

Model: RC:110A

Serial number: 95167

Class: 2

Test summary

Date of calibration: 21 February 2025

The doseBadge reader detailed above has been calibrated to the published data as described in the operating manual and in the half-inch configuration. The procedures and techniques used are as described in IEC60942_2003 Annex B – Periodic Tests and three determinations of the sound pressure level, frequency and total distortion were made.

The sound pressure level was measured using a WS2F condenser microphone type MK:224 manufactured by Cirrus Research plc.

The results have been corrected to the reference pressure of 101.33 kPa using the manufacturer's data.

The doseBadge Reader has been shown to conform to the Class 2 requirements for periodic testing, described in Annex B of IEC 60942:2003 for the sound pressure level(s) and frequency(ies) stated, for the environmental conditions under which the tests were performed.

However, as public evidence was not available, from a testing organisation responsible for pattern approval, to demonstrate that the model of doseBadge Reader conformed to the requirements for pattern evaluation described in Annex A of IEC 60942:2003, no general statement or conclusion can be made about conformance of the doseBadge Reader to the requirements of IEC 60942:2003.

Notes:

This certificate provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory. The results within this certificate relate only to the items calibrated. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a coverage probability of approximately 95%.

CERTIFICATE OF CALIBRATION

Certificate Number:

234084

Page 2 of 2

Environmental conditions

The following conditions were recorded at the time of the test:

Before Pressure: 99.38 kPa Temperature: 25.0 °C Humidity: 40.4 %
After Pressure: 99.39 kPa Temperature: 25.1 °C Humidity: 37.9 %

Test equipment

Equipment	Manufacturer	Model	Serial number
Distortion Meter	Keithley	2015	1063074
Environmental Monitor	Comet	T7510	21962628

Initial Acoustic Results

	Expected	Sample 1	Sample 2	Sample 3	Average	Deviation	Tolerance	Uncertainty
Level (dB)	114.00	113.79	113.78	113.79	113.79	-0.21	±0.75	0.11 dB
Distortion (%)	< 4.00	1.52	0.51	0.51	0.84	0.84	+4.00	0.13 %
Frequency (Hz)	1000.0	990.4	990.4	990.3	990.4	-9.6	±20.0	0.1 Hz

The measured quantities or deviations (as applicable), extended by the expanded combined uncertainty of measurement, must not exceed the corresponding tolerance.

Adjusted Acoustic Results

	Expected	Sample 1	Sample 2	Sample 3	Average	Deviation	Tolerance	Uncertainty
Level (dB)	114.00	113.97	113.97	113.97	113.97	-0.03	±0.75	0.11 dB
Distortion (%)	< 4.00	0.51	0.50	0.51	0.51	0.51	+4.00	0.13 %
Frequency (Hz)	1000.0	990.3	990.3	990.3	990.3	-9.7	±20.0	0.1 Hz

Functionality Results

Function	Result
Keypad	Pass
Battery Power	Pass
Display	Pass
Communication	Pass
2 way IR link	Pass
Clock	Pass

End of results



SOUND LEVEL METER CALIBRATION

Calibration Location: SECOT

Calibration Date: 06-03-2025

ACOUSTIC CALIBRATOR

Brand	Model	Serial No.	Frequency (Hz)	Ref.Calibrated (dB)	Eff.Calibrated (dB)
Cirrus	CR:515	97097	1000.00	94	93.8

No.	Brand	Model	Serial No.	Reading (dB)	dB Adjust
1	SCARLET TECH	ST-21D	820722	93.8	0.0
2	SCARLET TECH	ST-21D	820725	93.8	0.0
3	SCARLET TECH	ST-21D	820723	93.8	0.0

Calibrated by :

Approved by : *Sukh Suthanon*

THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Mechanical Engineering Standards Laboratory Soi 1, Bangpoo Industrial Estate, Muang, Samutprakan 10280, Thailand.

Request No.23-68/0279

MTC.No.23-68/0279-01

Number of page(s) 2

CALIBRATION CERTIFICATE

Nomenclature : DRYCAL

Manufacturer : Mesa Labs

Serial No.: 160100

Model : Defender 520-L

Scale range : 5 ml/min to 500 ml/min

Subdivision : (0.001, 0.01) ml/min

Submitted by : SECOT CO.,LTD.

239, Rimklongprapa Road, Bangsue,

Bangkok 10800, Thailand.

Received date : 13 February 2025 Condition of measured item : Normal

Calibration date : 24 February 2025

Standard :

Standard	Certificate No.	Date due	Traceability
RTD Thermometer	PSL-T 0811/67	3-Jul-26	TISTR
Molbox/PressureTransducer/UpStream	MP-0076-23	2-Apr-25	NIMT
Primary Flow Calibrator S/N 117982	MW-0034-23	11-Jun-25	NIMT

Calibrated by : *Terasak Panna*

(Mr.Terasak Panna)

Approved by :

(Ms.Kirana Luanghirun)

Director

Mechanical Engineering Standards Laboratory

Ref. 2013268021300656001

Issued Date 28 February 2025

The results relate only to the items tested/calibrated or value assigned.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

Head Office

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9036
Fax. (66) 0 2577 9009

Office/Laboratory

668 Mu 2 Tambon Bangpoomai, Amphoe Muang Samutprakan,
Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
(66) 08 3219 9440
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office

196 Phahonyothin Road, Ladyao, Chatuchak,
Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
(66) 08 1889 6827



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Mechanical Engineering Standards Laboratory Soi 1, Bangpoo Industrial Estate, Muang, Samutprakan 10280, Thailand.

Request No.23-68/0279

2/2

MTC.No.23-68/0279-01

Calibration point : (20, 50, 100, 200, 400) ml/min

Ambient condition : Temperature (23 ± 3) °C , Relative humidity (55 ± 15) %

Atmospheric pressure (1010 ± 13) hPa

Calibration method : The flowmeter (UUC) was calibrated by comparison method with standard flowmeter according to CP-370.01.

The reported value is the value that converted to value at reference condition within pressure and temperature of the actual gas entering the UUC

Measurement data :

UUC Value	Standard Value	Temperature	Pressure	Deviation	Uncertainty
(ml/min)	(ml/min)	(°C)	(hPa)	(%)	(%)
20.473*	20.340	24.275	1011.42	+0.65	0.94
49.952	50.732	24.057	1011.52	-1.54	0.95
99.449	99.622	24.102	1011.62	-0.17	0.93
200.34	199.94	24.133	1011.77	+0.20	0.93
401.89	397.98	24.140	1012.07	+0.98	0.93

The reported expanded uncertainties are based on standard uncertainties multiplied by a coverage factor $k=2$, which provides a level of confidence of approximately 95%.

* : The calibration point is not the scope of accreditation.

The end of calibration certificate.

Ts.

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5

Head Office

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9036
Fax. (66) 0 2577 9009

Office/Laboratory

668 Mu 2 Tambon Bangpoomai, Amphoe Muang Samutprakan,
Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
(66) 08 3219 9440
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office

196 Phahonyothin Road, Ladyao, Chatuchak,
Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
(66) 08 1889 6827