





ที่ อก ๐๓๒๐/ ๖๗๕๐๗

กรมโรงงานอุตสาหกรรม  
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท  
เขตราชเทวี กรุงเทพฯ ๑๐๕๐๐

๒๗ ธันวาคม ๒๕๖๖

เรื่อง ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

เรียน กรรมการผู้จัดการ บริษัท อีสเทิร์น ไทย คอนซัลติ้ง ๑๙๙๒ จำกัด

อ้างถึง คำขอต่ออายุของห้องปฏิบัติการวิเคราะห์เอกชน ลงวันที่ ๑๒ ธันวาคม ๒๕๖๖

สิ่งที่ส่งมาด้วย เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน  
บริษัท อีสเทิร์น ไทย คอนซัลติ้ง ๑๙๙๒ จำกัด จำนวน ๑ แผ่น

ตามหนังสือที่อ้างถึง บริษัท อีสเทิร์น ไทย คอนซัลติ้ง ๑๙๙๒ จำกัด ขอต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน เลขทะเบียน ว-๐๐๓/๒ สถานที่ตั้งเลขที่ ๑๒๙ หมู่ที่ ๑ ตำบลนนทรี อำเภอบินทร์บุรี จังหวัดปทุมธานี ต่อกรมโรงงานอุตสาหกรรม นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้วให้ บริษัท อีสเทิร์น ไทย คอนซัลติ้ง ๑๙๙๒ จำกัด ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน โดยมีองค์ประกอบดังนี้

ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์

- |                               |                              |
|-------------------------------|------------------------------|
| ๑) นางสาวพัชรพรรณ จันทร์สำโรง | ทะเบียนเลขที่ ว-๐๐๓/๒-ค-๐๐๐๑ |
| ๒) นางสาวสุภาวดี สาธุภาค      | ทะเบียนเลขที่ ว-๐๐๓/๒-ค-๐๐๐๒ |
| ๓) นางสาวยุภา กะชินรัมย์      | ทะเบียนเลขที่ ว-๐๐๓/๒-ค-๐๐๐๓ |
| ๔) นายสิทธิพงษ์ หัตถรักไชย    | ทะเบียนเลขที่ ว-๐๐๓/๒-ค-๐๐๐๔ |

ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์

- |                               |                              |
|-------------------------------|------------------------------|
| ๑) นางสาวสุภาณัน กุลศรี       | ทะเบียนเลขที่ ว-๐๐๓/๒-จ-๐๐๐๑ |
| ๒) นางสาวเพชรรัตน์ ชิตสวัสดิ์ | ทะเบียนเลขที่ ว-๐๐๓/๒-จ-๐๐๐๒ |

ค. ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนให้วิเคราะห์ในน้ำเสีย จำนวน ๑๑ รายการ น้ำใต้ดิน จำนวน ๒ รายการ รวมทั้งสิ้นจำนวน ๑๓ รายการ ตามสิ่งที่ส่งมาด้วย

หนังสือฉบับนี้จะหมดอายุในวันที่ ๘ มกราคม ๒๕๗๐ หากประสงค์จะต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ให้ยื่นคำขอต่ออายุพร้อมเอกสารประกอบการคำขอต่อกรมโรงงานอุตสาหกรรม ภายใน ๓๐ วัน ก่อนวันสิ้นอายุของหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ทั้งนี้ สามารถยื่นคำขอผ่านระบบอิเล็กทรอนิกส์ได้ที่หน้าเว็บไซต์กรมโรงงานอุตสาหกรรม

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ

(นายทวี อำพาพันธ์)

ผู้อำนวยการศูนย์วิจัยและเตือนภัยมลพิษโรงงานภาคตะวันออก

ปฏิบัติราชการแทนอธิบดีกรมโรงงานอุตสาหกรรม

ศูนย์วิจัยและเตือนภัยมลพิษโรงงานภาคตะวันออก

โทร. ๐ ๓๓๑๓ ๖๐๕๕ ต่อ ๕๐๐๑-๒

ไปรษณีย์อิเล็กทรอนิกส์ eirw@diw.mail.go.th



เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท อีสเทิร์น ไทย คอนซัลติ้ง ๑๙๙๒ จำกัด

เลขทะเบียน ว-๐๐๓/๒

ที่ อก ๐๓๒๐/๑๗/๕๐๗

ลงวันที่ ๒๗ มิ.ย. ๒๕๖๖

ขอขยายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๑๓ รายการ  
น้ำเสีย จำนวน 11 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Biochemical Oxygen Demand	5-Day BOD Test, Membrane Electrode Method
2	Chemical Oxygen Demand	Closed Reflux, Titrimetric Method
3	Color	ADMI Weight-Ordinate Spectrophotometric Method
4	Hexavalent Chromium	Filtration, Colorimetric Method
5	Oil and Grease	Liquid-Liquid Partition, Gravimetric Method
6	pH	Electrometric Method
7	Sulfide	ZnS Precipitation, Iodometric Method
8	Temperature	Field Method
9	Total Dissolved Solids	Dried at 180 °C
10	Total Kjeldahl Nitrogen	Macro-Kjeldahl Method
11	Total Suspended Solids	Dried at 103-105 °C

น้ำใต้ดิน จำนวน 2 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Hexavalent Chromium	Filtration, Colorimetric Method
2	pH	Electrometric Method

เอกสารอ้างอิง

APHA, AWWA, WEF. **Standard Methods for the Examination of Water and Wastewater**. 24<sup>th</sup> ed. Washington, DC: APHA, 2023





ที่ อก ๐๓๒๐/๑๑๓๔๒

กรมโรงงานอุตสาหกรรม  
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท  
เขตราชเทวี กรุงเทพฯ ๑๐๔๐๐

๒๗ กรกฎาคม ๒๕๖๖

เรื่อง ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

เรียน กรรมการผู้จัดการ บริษัท อีสเทิร์น ไทย คอนซัลติ้ง ๑๙๙๒ จำกัด

อ้างถึง คำขอต่ออายุของห้องปฏิบัติการวิเคราะห์เอกชน ลงวันที่ ๗ มิถุนายน ๒๕๖๖

สิ่งที่ส่งมาด้วย เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

๑. รายชื่อผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ จำนวน ๔๐ ราย
๒. รายชื่อเจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๒๕ ราย
๓. ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๒๙๒ รายการ  
จำนวน ๑๙ แผ่น

ตามหนังสือที่อ้างถึง บริษัท อีสเทิร์น ไทย คอนซัลติ้ง ๑๙๙๒ จำกัด ขอต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน เลขทะเบียน ว-๐๐๓ สถานที่ตั้งเลขที่ ๖๘๓ หมู่ที่ ๑๑ ถนนสุขาภิบาล ๘ ตำบลหนองขาม อำเภอศรีราชา จังหวัดชลบุรี ต่อกรมโรงงานอุตสาหกรรม นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว ให้บริษัท อีสเทิร์น ไทย คอนซัลติ้ง ๑๙๙๒ จำกัด ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน โดยมีองค์ประกอบดังนี้

- ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ จำนวน ๔๐ ราย ตามสิ่งที่ส่งมาด้วย ๑
- ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๒๕ ราย ตามสิ่งที่ส่งมาด้วย ๒
- ค. ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนให้วิเคราะห์ในน้ำเสีย จำนวน ๔๗ รายการ  
อากาศเสีย (ปล่องระบาย) จำนวน ๒๑ รายการ น้ำใต้ดิน จำนวน ๑๑๑ รายการ สิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว  
จำนวน ๑๘ รายการ และดิน จำนวน ๙๕ รายการ รวมทั้งสิ้นจำนวน ๒๙๒ รายการ ตามสิ่งที่ส่งมาด้วย ๓

หนังสือฉบับนี้จะหมดอายุในวันที่ ๕ กรกฎาคม ๒๕๖๙ หากประสงค์จะต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ให้ยื่นคำขอต่ออายุพร้อมเอกสารประกอบคำขอต่อกรมโรงงานอุตสาหกรรม ภายใน ๓๐ วัน ก่อนวันสิ้นอายุของหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ทั้งนี้ สามารถยื่นคำขอผ่านระบบอิเล็กทรอนิกส์ได้ที่หน้าเว็บไซต์กรมโรงงานอุตสาหกรรม

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ


(นายทวี อำพาพันธ์)

ผู้อำนวยการศูนย์วิจัยและเตือนภัยมลพิษโรงงานภาคตะวันออก  
ปฏิบัติราชการแทนอธิบดีกรมโรงงานอุตสาหกรรม

ศูนย์วิจัยและเตือนภัยมลพิษโรงงานภาคตะวันออก

โทร. ๐ ๓๓๑๓ ๖๐๕๕ ต่อ ๕๐๐๑-๒

ไปรษณีย์อิเล็กทรอนิกส์ [eirw@diw.mail.go.th](mailto:eirw@diw.mail.go.th)



“อุตสาหกรรมก้าวไกล ประเทศไทยก้าวหน้า ร่วมกันพัฒนา อุตสาหกรรมสีเขียว”



เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท อีสเทิร์น ไทย คอนซัลติง ๑๙๙๒ จำกัด เลขทะเบียน ว-๐๐๓

ที่ อก ๐๓๒๐/๑๑๓๔๒

ลงวันที่ ๒๗ กรกฎาคม ๒๕๖๖

ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ จำนวน ๔๐ ราย

๑) นางสาวมาลีเกษ เลขะวัจกุล	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๐๑
๒) นายวัฒนา โคตรหล้า	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๐๒
๓) นางวรรณเพ็ญ เหลาจินดาวรรณ	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๐๓
๔) นายกะวีร์ สุธาทรัพย์	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๐๔
๕) นางสาวนันท์ณภัส แบนขุนทด	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๐๕
๖) นางสาวพรนภา หลงคำหงษ์	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๐๖
๗) นางสาวอภิตี ชื่นอารมย์	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๐๗
๘) นางสาวอัจฉรี จิตตะยโสธร	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๐๘
๙) นางสาวจิรพร ปานคง	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๐๙
๑๐) นายสุทธา สองธินัย	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๑๐
๑๑) นางสาวนันประภา อูยสูงเนิน	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๑๑
๑๒) นายธงไชย บุญศักดิ์	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๑๒
๑๓) นางสาวธนิชพร กลิ่นโสภณ	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๑๓
๑๔) นายธีระพงษ์ นวลอินทร์	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๑๔
๑๕) นางสาวแพรว พลเสน	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๑๕
๑๖) นายทรงพล ผิวอ้วน	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๑๖
๑๗) นายภาคภูมิ บัวสวัสดิ์	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๑๗
๑๘) นางสาวจันทนี สายพันธ์	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๑๘
๑๙) นายภาณุพงศ์ บำรุงรส	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๑๙
๒๐) นางสาวภาณิน จันดีสอน	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๒๐
๒๑) นายวรกร ไวทยะเสวี	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๒๑
๒๒) นางสาววรรณภา ไชยศิริ	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๒๒
๒๓) นางสาวพรพิมล ภูมิคอนสาร	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๒๓
๒๔) นางสาวธมลวรรณ ผลอ้อ	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๒๔
๒๕) นางสาวบุญเรือง บุญถม	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๒๕
๒๖) นางสาวภัสนันท์ ป้อมน้อย	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๒๖
๒๗) นายชานวัฒน์ โชตะวงศ์	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๒๗
๒๘) นางสาวพจณีย์ งามวิสัย	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๒๘
๒๙) นายวิญญ์วัช สิงห์โต	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๒๙
๓๐) นางสาวนุกูล อามศรี	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๓๐
๓๑) นายศุภฤกษ์ พาดกลาง	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๓๑
๓๒) นายณิขพล ทองหล่อ	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๓๒
๓๓) นายธรรมรัตน์ โพธิ์ตันคำ	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๓๓
๓๔) นายโอชา ขวัญศิริมงคล	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๓๔
๓๕) นายเมธี สุขประเสริฐ	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๓๕

๓๖) นางสาวพรพินันท์...

๓๖) นางสาวพรพินันท์ วิริยกุลกุล	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๓๖
๓๗) นางสาวอาภาภรณ์ เสริมสนธิ	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๓๗
๓๘) นางสาวนภัทร์ธมณต์ ประดิษฐ์นุช	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๓๘
๓๙) นางสาวสุนิษา เอ็งเส้ง	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๓๙
๔๐) นางสาวระพีณ อินัน	ทะเบียนเลขที่	ว-๐๐๓-ค-๐๐๔๐

ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๒๕ ราย

๑) นางสาวดวงกมล เนื้อทอง	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๐๑
๒) นางสาววัชรภรณ์ อินทสุข	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๐๒
๓) นางสาวกัญจน์ธวิภา จันทร์ขอดแก้ว	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๐๓
๔) นางสาวฉัตรสุดา มงคลโกชนัน	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๐๔
๕) นางสาวณัฐวดี อำมาตย์ทัศน์	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๐๕
๖) นางสาวนอรอุมา ปาระ	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๐๖
๗) นางสาวธัญลักษณ์ ชันโต	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๐๗
๘) นางสาวสุทธิดา สร้างแก้ว	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๐๘
๙) นายอุดมทรัพย์ เจนจบจริง	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๐๙
๑๐) นายนราธิป สงวนศิลป์	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๑๐
๑๑) นายวีระชัย พอใจ	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๑๑
๑๒) นายอัญชลี ทะพงษ์	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๑๒
๑๓) นางสาวสุมลิตรา มีแก่น	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๑๓
๑๔) นางสาวสวรรยา เพชรประไพ	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๑๔
๑๕) นางสาวจุฑามาศ เจริญพรหม	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๑๕
๑๖) นางสาวนิภาพร คำชมภู	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๑๖
๑๗) นางสาวอรชา พันธุ์เมือง	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๑๗
๑๘) นายกิตติ ไพโรจน์	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๑๘
๑๙) นายชาญณรงค์ ตั้งธรรมรักษ์	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๑๙
๒๐) นางสาวปวีศา เอสันเทียะ	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๒๐
๒๑) นางสาวจุฑาทิพย์ กิจดี	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๒๑
๒๒) นางสาวสุภาวดี ศรีละออง	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๒๒
๒๓) นางสาวณัฐชยา บรรพบุตร	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๒๓
๒๔) นางสาวณัฐนิช นนตานอก	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๒๔
๒๕) นางสาวดวงสุดา แสนวันดี	ทะเบียนเลขที่	ว-๐๐๓-จ-๐๐๒๕

เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท อีสเทิร์น ไทย คอนซัลติ้ง ๑๙๙๒ จำกัด เลขทะเบียน ว-๐๐๓

ที่ อก ๐๓๒๐/๑๑๓๔๒

ลงวันที่ ๒๗ กรกฎาคม ๒๕๖๖

ขอขยายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๒๙๒ รายการ

น้ำเสีย จำนวน 47 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Aldrin	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
2	Arsenic	1) Continuous Hydride Generation/Atomic Absorption Spectrometric Method <sup>[4]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
3	Barium	Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
4	$\alpha$ -BHC	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
5	$\beta$ -BHC	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
6	$\delta$ -BHC	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
7	$\gamma$ -BHC	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
8	Biochemical Oxygen Demand	1) 5-Day BOD Test, Membrane Electrode Method <sup>[4]</sup> 2) 5-Day BOD Test, Azide Modification Method <sup>[4]</sup>
9	Cadmium	Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
10	Chemical Oxygen Demand	Closed Reflux, Titrimetric Method <sup>[4]</sup>
11	cis-Chlordane	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
12	trans-Chlordane	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
13	Chromium	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
14	Color	ADMI Weighted-Ordinate Spectrophotometric Method <sup>[4]</sup>
15	Copper	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
16	Cyanide	Distillation, Colorimetric Method <sup>[4]</sup>
17	4,4'-DDD	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
18	4,4'-DDE	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
19	DDT	Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
20	Dieldrin	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
21	Endosulfan I	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
22	Endosulfan II	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
23	Endosulfan sulfate	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
24	Endrin	Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
25	Endrin aldehyde	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
26	Endrin ketone	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
27	Formaldehyde	Distillation, Colorimetric Method <sup>[3]</sup>
28	Free Chlorine	1) Iodometric Method <sup>[4]</sup> 2) Colorimetric Method <sup>[4]</sup>



ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
29	Heptachlor	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
30	Heptachlor Epoxide	1) Liquid-Liquid Extraction, Gas Chromatographic Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
31	Hexavalent Chromium	Filtration, Colorimetric Method <sup>[4]</sup>
32	Lead	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
33	Manganese	Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
34	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method <sup>[4]</sup>
35	Methoxychlor	Liquid-Liquid Extraction, Gas Chromatographic / Mass Spectrometric Method <sup>[4]</sup>
36	Nickel	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
37	Oil and Grease	Liquid-Liquid, Partition-Gravimetric Method <sup>[4]</sup>
38	pH	Electrometric Method <sup>[4]</sup>
39	Phenols	Distillation, Direct Photometric Method <sup>[4]</sup>
40	Selenium	Digestion, Hydride Generation/Atomic Absorption Spectrometric Method <sup>[4]</sup>
41	Sulfide	ZnS Precipitation, Iodometric Method <sup>[4]</sup>
42	Temperature	Field Method <sup>[4]</sup>
43	Trivalent Chromium	1) Digestion, Direct Air-Acetylene Flame Method; Filtration, Colorimetric Method; Calculation <sup>[4]</sup> 2) Digestion, Inductively Coupled Plasma Method; Colorimetric Method; Calculation <sup>[4]</sup>
44	Total Dissolved Solids	Dried at 180 °C <sup>[4]</sup>
45	Total Kjeldahl Nitrogen	Macro Kjeldahl Method <sup>[4]</sup>
46	Total Suspended Solids	Dried at 103-105 °C <sup>[4]</sup>
47	Zinc	Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>

อากาศเสีย (ปล่อยระบาย) จำนวน 21 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Antimony	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
2	Arsenic	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
3	Cadmium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
4	Carbon Monoxide	1) Bag, Non-Dispersive Infrared Method <sup>[5]</sup> 2) Instrumental Analyzer Method <sup>[5]</sup>
5	Chromium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
6	Cobalt	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
7	Copper	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
8	Hydrogen Sulfide	Absorption Sampling, Iodometric Method <sup>[5]</sup>
9	Lead	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
10	Manganese	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
11	Mercury	Isokinetic Sampling, Digestion, Cold-Vapor Atomic Absorption Spectrometric Method <sup>[5]</sup>
12	Nickel	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
13	Opacity	Ringelmann's Method <sup>[1,5]</sup>
14	Oxides of Nitrogen	1) Absorption Sampling, Phenoldisulfonic Acid Method <sup>[8]</sup> 2) Instrumental Analyzer Method <sup>[7]</sup>
15	Selenium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
16	Sulfur Dioxide	1) Absorption Sampling , Barium-Thorin Titrimetric Method <sup>[5]</sup> 2) Instrumental Analyzer Method <sup>[5]</sup>
17	Sulfuric Acid	Isokinetic Sampling, Barium-Thorin Titrimetric Method <sup>[6]</sup>
18	Tin	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
19	Total Suspended Particulate	Isokinetic Sampling, Gravimetric Method <sup>[6]</sup>
20	Vanadium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[5]</sup>
21	Xylene	Adsorption Sampling, Gas Chromatographic Method <sup>[6]</sup>

**น้ำใต้ดิน จำนวน 111 รายการ**

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Acenaphthene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
2	Acetone	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
3	Aldrin	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
4	Anthracene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
5	Antimony	Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
6	Arsenic	1) Continuous Hydride Generation/Atomic Absorption Spectrometric Method <sup>[4]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
7	Barium	Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
8	Benz(a)anthracene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
9	Benzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
10	Benzo(b)fluoranthene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
11	Benzo(k)fluoranthene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
12	Benzo(a)pyrene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
13	Benzo[g,h,i]perylene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
14	Beryllium	Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
15	Bis(2-chloroethyl)ether	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
16	Bis(2-ethylhexyl)phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
17	Bromodichloromethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
18	Bromoform	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
19	Butanol	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
20	Butyl benzyl phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
21	Cadmium	Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
22	Carbazole	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
23	Carbon disulfide	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
24	Carbon tetrachloride	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
25	Chlordane	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
26	p-Chloroaniline	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
27	Chlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
28	Chlorodibromomethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
29	Chloroform	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
30	2-Chlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
31	Chromium	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
32	Chromium (III)	1) Digestion, Direct Air-Acetylene Flame Method; Filtration, Colorimetric Method; Calculation <sup>[4]</sup> 2) Digestion, Inductively Coupled Plasma Method; Filtration, Colorimetric Method; Calculation <sup>[4]</sup>

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
33	Chromium (VI)	Filtration, Colorimetric Method <sup>[4]</sup>
34	Chrysene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
35	Cyanide	Distillation, Colorimetric Method <sup>[4]</sup>
36	DDD	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
37	DDE	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
38	DDT	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
39	Dibenz(a,h)anthracene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
40	Di-n-butyl phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
41	1,2-Dichlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
42	1,3-Dichlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
43	1,4-Dichlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
44	1,1-Dichloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
45	1,2-Dichloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
46	1,1-Dichloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
47	cis-1,2-Dichloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
48	trans-1,2-Dichloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
49	2,4-Dichlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
50	1,2-Dichloropropane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
51	1,3-Dichloropropane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
52	Dieldrin	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
53	Diethyl phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
54	2,4-Dimethylphenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
55	2,4-Dinitrotoluene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
56	2,6-Dinitrotoluene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
57	Di-n-octyl phthalate	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
58	Endosulfan	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
59	Endrin	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
60	Ethylbenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
61	Fluoranthene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
62	Fluorene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
63	Heptachlor	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
64	Heptachlor epoxide	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
65	Hexachlorobenzene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
66	Hexachloro-1,3-butadiene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
67	n-Hexane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
68	$\alpha$ -HCH	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
69	$\beta$ -HCH	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>



ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
70	$\gamma$ -HCH	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
71	Hexachlorocyclopentadiene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
72	Hexachloroethane	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
73	Indeno(1,2,3-cd)pyrene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
74	Isophorone	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
75	Lead	1) Digestion, Direct Air-Acetylene Flame Method <sup>[4]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
76	Manganese	Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
77	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method <sup>[4]</sup>
78	Methoxychlor	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
79	Methylene chloride	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
80	2-Methylphenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
81	2-Methylnaphthalene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
82	Methyl tert-butyl ether	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
83	Naphthalene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
84	Nickel	Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
85	Nitrobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
86	N-Nitrosodi-n-propylamine	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
87	pH	Electrometric Method <sup>[4]</sup>
88	Phenanthrene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
89	Phenol	1) Distillation, Direct Photometric Method <sup>[4]</sup> 2) Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
90	Pyrene	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
91	Selenium	Digestion, Hydride Generation/Atomic Absorption Spectrometric Method <sup>[4]</sup>
92	Silver	Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
93	Styrene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
94	1,1,2,2-Tetrachloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
95	Trichloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
96	1,2,4-Trichlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
97	1,1,1-Trichloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
98	1,1,2-Trichloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
99	Tetrachloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
100	Toluene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
101	2,4,5-Trichlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
102	2,4,6-Trichlorophenol	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
103	1,3,5-Trimethylbenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
104	Vanadium	Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>
105	Vinyl acetate	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
106	Vinyl chloride	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
107	m-Xylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
108	o-Xylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
109	p-Xylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
110	Xylene (Total)	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[4]</sup>
111	Zinc	Digestion, Inductively Coupled Plasma Method <sup>[4]</sup>

**สิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว จำนวน 18 รายการ**

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Antimony	Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
2	Arsenic	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,9,10]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
3	Barium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,9,10]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
4	Beryllium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,9,10]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
5	Cadmium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,9,10]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
6	Chromium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,9,10]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
7	Chromium (VI)	1) Waste Extraction, Digestion, Colorimetric Method <sup>[2,13]</sup> 2) Alkaline Digestion, Colorimetric Method <sup>[9,13]</sup>
8	Cobalt	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,9,10]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
9	Copper	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,9,10]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
10	Lead	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,9,10]</sup>
11	Mercury	2) Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup> 1) Waste Extraction, Digestion, Cold Vapor Atomic Absorption Spectrometric Method <sup>[2,11]</sup> 2) Digestion, Cold vapor Atomic Absorption Spectrometric Method <sup>[9,11]</sup>
12	Nickel	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,9,10]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
13	Molybdenum	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,9,10]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
14	Selenium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,9,10]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
15	Silver	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,9,10]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
16	Thallium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,9,10]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
17	Vanadium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,9,10]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
18	Zinc	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method <sup>[2,9,10]</sup> 2) Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>

ดิน จำนวน 95 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Acenaphthene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
2	Acetone	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
3	Anthracene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
4	Antimony	Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
5	Arsenic	Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
6	Barium	Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
7	Benz(a)anthracene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
8	Benzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
9	Benzo(b)fluoranthene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
10	Benzo(k)fluoranthene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
11	Benzo(a)pyrene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
12	Benzo[g,h,i]perylene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
13	Beryllium	Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
14	Bis(2-chloroethyl)ether	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
15	Bis(2-ethylhexyl)phthalate	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
16	Bromodichloromethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
17	Bromoform	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
18	Butanol	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
19	Butyl benzyl phthalate	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
20	Cadmium	Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
21	Carbazole	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
22	Carbon disulfide	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
23	Carbon tetrachloride	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
24	p-Chloroaniline	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
25	Chlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
26	Chlorodibromomethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
27	Chloroform	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
28	2-Chlorophenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
29	Chromium	Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
30	Chromium (III)	Digestion, Inductively Coupled Plasma Method; Filtration, Colorimetric Method; Calculation <sup>[9,10]</sup>
31	Chromium (VI)	Alkaline Digestion, Colorimetric Method <sup>[12,13]</sup>
32	Chrysene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
33	Dibenz(a,h)anthracene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
34	Di-n-butyl phthalate	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
35	1,2-Dichlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
36	1,3-Dichlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
37	1,4-Dichlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>



ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
38	1,1-Dichloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
39	1,2-Dichloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
40	1,1-Dichloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
41	cis-1,2-Dichloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
42	trans-1,2-Dichloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
43	2,4-Dichlorophenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
44	1,2-Dichloropropane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
45	1,3-Dichloropropane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
46	Diethyl phthalate	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
47	2,4-Dimethylphenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
48	2,4-Dinitrotoluene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
49	2,6-Dinitrotoluene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
50	Di-n-octyl phthalate	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
51	Ethylbenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
52	Fluoranthene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
53	Fluorene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
54	Hexachlorobenzene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
55	Hexachloro-1,3-butadiene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
56	n-Hexane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
57	Hexachlorocyclopentadiene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
58	Hexachloroethane	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
59	Indeno(1,2,3-cd)pyrene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
60	Isophorone	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
61	Lead	Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
62	Manganese	Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
63	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method <sup>[9,11]</sup>
64	Methylene chloride	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
65	2-Methylphenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
66	2-Methylnaphthalene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
67	Methyl tert-butyl ether	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
68	Naphthalene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
69	Nickel	Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
70	Nitrobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
71	N-Nitrosodi-n-propylamine	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
72	Phenanthrene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
73	Phenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
74	Pyrene	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
75	Selenium	Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
76	Silver	Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
77	Styrene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
78	1,1,2,2-Tetrachloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
79	Tetrachloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
80	Toluene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
81	1,2,4-Trichlorobenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
82	1,1,1-Trichloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
83	1,1,2-Trichloroethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
84	Trichloroethylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
85	2,4,5-Trichlorophenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
86	2,4,6-Trichlorophenol	Ultrasonic Extraction, Gas Chromatographic/Mass Spectrometric Method <sup>[15,17]</sup>
87	1,3,5-Trimethylbenzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
88	Vanadium	Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>
89	Vinyl acetate	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
90	Vinyl chloride	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
91	m-Xylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
92	o-Xylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
93	p-Xylene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>



ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
94	Xylene (Total)	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method <sup>[14,16]</sup>
95	Zinc	Digestion, Inductively Coupled Plasma Method <sup>[9,10]</sup>

### เอกสารอ้างอิง

1. กระทรวงอุตสาหกรรม. **ประกาศกระทรวงอุตสาหกรรม พ.ศ.2549** เรื่องกำหนดค่าปริมาณเขม่าควันที่เจือปนในอากาศที่ระบายออกจากปล่องของหม้อน้ำโรงสีข้าวที่ใช้กลบเป็นเชื้อเพลิง. ราชกิจจานุเบกษา. 4 ธันวาคม 2549. เล่มที่ 123 ตอนพิเศษ 125 ง.
2. กระทรวงอุตสาหกรรม. **ประกาศกระทรวงอุตสาหกรรม พ.ศ.2548** เรื่อง การกำจัดสิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว. ราชกิจจานุเบกษา. 25 มกราคม 2549. เล่มที่ 123 ตอนพิเศษ 11ง.
3. สมาคมวิศวกรรมสิ่งแวดล้อมแห่งประเทศไทย. คู่มือวิเคราะห์น้ำเสีย. พิมพ์ครั้งที่ 4. กรุงเทพฯ: เรือนแก้วการพิมพ์, 2547.
4. APHA, AWWA, WEF. **Standard Methods for the Examination of Water and Wastewater**. 23<sup>rd</sup> ed. Washington, DC : APHA, 2017
5. United States Environmental Protection Agency. **Standard of Performance for New Stationary Sources**. 40 CFR Part 60. Appendix A, 2017.
6. United States Environmental Protection Agency. **Standard of Performance for New Stationary Sources**. 40 CFR Part 60. Appendix A, 2019.
7. United States Environmental Protection Agency. **Standard of Performance for New Stationary Sources**. 40 CFR Part 60. Appendix A, 2020.
8. United States Environmental Protection Agency. **Standard of Performance for New Stationary Sources**. 40 CFR Part 60. Appendix A, 2023.
9. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Acid Digestion of Sediments Sludge and Soils. SW-846 Method 3050B**, 1996.
10. United States Environment Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Inductively Coupled Plasma-Atomic Emission spectrometry. SW-846 Method 6010C**, 2007.
11. United States Environment Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Mercury in Solid or Semisolid Waste (Manual Cold-Vapor Technique). SW-846 Method 7471B**, 2007.
12. United States Environment Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Alkaline digestion for Hexavalent Chromium. SW-846 Method 3060A**, 1996.

13. United States Environment Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Chromium. Hexavalent (Colorimetric). SW-846 Method 7196A**, 1992

14. United States Environment Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Closed-System Purge-and-Trap and Extraction for Volatile Organics in Soil and Waste Samples. SW-846 Method 5035A**, 2002

15. United States Environment Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Ultrasonic Extraction. SW-846 Method 3550C**, 2007

16. United States Environment Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Volatile Organic Compounds by Gas Chromatography/Mass Spectrometry (GC/MS). SW-846 Method 8260D**, 2018

17. United States Environment Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Semivolatile Organic Compounds by Gas Chromatography/Mass Spectrometry. SW-846 Method 8270E**, 2018





ที่ อก ๐๓๑๐(๑)/ ๖๕๕ ๗



กรมโรงงานอุตสาหกรรม  
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท  
เขตราชเทวี กรุงเทพฯ ๑๐๔๐๐

๐๕ กรกฎาคม ๒๕๖๗

เรื่อง ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

เรียน กรรมการผู้จัดการ บริษัท เอ็ม อี ที จำกัด

อ้างถึง คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และชนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน  
ลงวันที่ ๒๙ เมษายน ๒๕๖๗

สิ่งที่ส่งมาด้วย เอกสารแนบท้ายหนังสือต่ออายุรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน  
บริษัท เอ็ม อี ที จำกัด จำนวน ๔ แผ่น

ตามคำขอที่อ้างถึง บริษัท เอ็ม อี ที จำกัด ขอต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน เลขทะเบียน ว-๑๐๐ สถานที่ตั้งเลขที่ ๓๖/๖๕๕ หมู่ที่ ๖ ตำบลบางรักพัฒนา อำเภอบางบัวทอง จังหวัดนนทบุรี ต่อกรมโรงงานอุตสาหกรรม นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว ให้บริษัท เอ็ม อี ที จำกัด ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน โดยมีองค์ประกอบดังนี้


ก. ผู้ควบคุมห้องปฏิบัติการวิเคราะห์เอกชน

- |                                  |                            |
|----------------------------------|----------------------------|
| ๑) นางสาวศศิธร สุวรรณวิโก        | ทะเบียนเลขที่ ว-๑๐๐-ค-๐๐๐๑ |
| ๒) นางสาวประภาพร เงะผล           | ทะเบียนเลขที่ ว-๑๐๐-ค-๐๐๐๒ |
| ๓) นางสาวภคธฤตา สุนทรอำไพ        | ทะเบียนเลขที่ ว-๑๐๐-ค-๐๐๐๓ |
| ๔) นางสาวลัดดาวัลย์ วงศ์คำจันทร์ | ทะเบียนเลขที่ ว-๑๐๐-ค-๐๐๐๔ |
| ๕) นายชิษณุพล ตู่ทอง             | ทะเบียนเลขที่ ว-๑๐๐-ค-๐๐๐๕ |
| ๖) นางสาวกาญจนา ไตรวงศ์          | ทะเบียนเลขที่ ว-๑๐๐-ค-๐๐๐๖ |
| ๗) นายอนุพงศ์ นามศรีฐาน          | ทะเบียนเลขที่ ว-๑๐๐-ค-๐๐๐๗ |
| ๘) นางสาวณิศรา พนานิกิตร์        | ทะเบียนเลขที่ ว-๑๐๐-ค-๐๐๐๘ |
| ๙) นางสาวปิยนุช ผุดผ่อง          | ทะเบียนเลขที่ ว-๑๐๐-ค-๐๐๐๙ |
| ๑๐) นางสาวศิริวรรณ บุญเพ็ง       | ทะเบียนเลขที่ ว-๑๐๐-ค-๐๐๑๐ |

ข. เจ้าหน้าที่ห้องปฏิบัติการวิเคราะห์เอกชน

- |                                |                            |
|--------------------------------|----------------------------|
| ๑) นายจิรายุทธ์ สรรพช่วง       | ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๐๑ |
| ๒) นายศรัทธา ชัยกิจตระกูล      | ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๐๒ |
| ๓) นายเมธา บุญหล้า             | ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๐๓ |
| ๔) นายจักริน ชินชะง้อ          | ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๐๔ |
| ๕) นายประภากร เกิดเกรียงไกร    | ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๐๕ |
| ๖) นายอภิสิทธิ์ แสงจินดา       | ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๐๖ |
| ๗) นางสาวอนงนาถ มหาเมฆรัตน์ฤดี | ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๐๗ |
| ๘) นายธนพล สิทธิพล             | ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๐๘ |
| ๙) นายพนาสีทธิ์ ใจเย็น         | ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๐๙ |
| ๑๐) นางสาวมณีรัตน์ พรหมอักษร   | ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๑๐ |

สำเนาถูกต้อง



๑๑) นางสาวชาลิณี...



๑๑) นางสาวชาลิณี ชูยน้ำเที่ยง	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๑๑
๑๒) นางสาวสุณาดา วงศ์อนุ	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๑๒
๑๓) นางสาวชนิกานต์ หล้าแสน	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๑๓
๑๔) นางสาวสุภัครินทร์ ก้อนมณี	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๑๔
๑๕) นางสาวอมรรัตน์ เพียรชนะ	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๑๕
๑๖) นางสาวกาญจนา คงคุณ	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๑๖
๑๗) นางสาวทิพย์สุตา เพชรเกิด	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๑๗
๑๘) นายพนัษกร เสียงชื่น	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๑๘
๑๙) นายปิยะชัย สิริอินทร์	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๑๙
๒๐) นางสาวภัทราวดี ศรีละออ	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๒๐
๒๑) นายสิทธิชัย เขียวไกร	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๒๑
๒๒) นางสาวแรมจันทร์ พิมพาเรียน	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๒๒
๒๓) นายदनัยฤทธิ์ ทองอ่อน	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๒๓
๒๔) นายเกษม อ่อนคำมา	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๒๔
๒๕) นายสันติภาพ ขาวนวล	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๒๕
๒๖) นายอานนท์ นนทเกียรติกุล	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๒๖
๒๗) นายพงษ์ธรณ์ เพียสา	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๒๗
๒๘) นายหัตถชัย บุญสว่าง	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๒๘
๒๙) นางสาวพานทิพย์ สีดาบุตร	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๒๙
๓๐) นางสาววิภารัตน์ ประณต	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๓๐
๓๑) นางสาวจิตติมา แก้วโสภาค	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๓๑
๓๒) นางสาวพานแก้ว สีดาบุตร	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๓๒
๓๓) นางสาวอริยา วิงวอน	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๓๓
๓๔) นายจิรายุทธ สีหาบุตร	ทะเบียนเลขที่ ว-๑๐๐-จ-๐๐๓๔

ค. ขอบข่ายชนิดสารมลพิษที่ได้รับขึ้นทะเบียนให้วิเคราะห์ในน้ำ/น้ำเสีย น้ำใต้ดิน อากาศเสีย สิ่งปฏิกูลหรือวัสดุที่ไม่ใช่แล้ว และดิน ตามสิ่งที่ส่งมาด้วย

หนังสือฉบับนี้จะหมดอายุในวันที่ ๒ มิถุนายน ๒๕๗๑ หากประสงค์จะต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ให้ยื่นคำขอต่ออายุพร้อมเอกสารประกอบคำขอต่อกรมโรงงานอุตสาหกรรม ภายใน ๖๐ วัน ก่อนวันสิ้นอายุของหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ

น

(นายพรยศ กลั่นกรอง)

รองอธิบดี ปฏิบัติราชการแทน  
อธิบดีกรมโรงงานอุตสาหกรรม

ว.จ.น.



สำเนาถูกต้อง

กองวิจัยและเตือนภัยมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๐๓-๕

โทรสาร ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๙๙

ไปรษณีย์อิเล็กทรอนิกส์ saraban@diw.mail.go.th



“อุตสาหกรรมก้าวไกล ประเทศไทยก้าวหน้า ร่วมกันพัฒนา อุตสาหกรรมสีเขียว”



เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท เอ็ม อี ที จำกัด

เลขทะเบียน ว-๑๐๐

ที่ อก ๐๓๑๐(๑)/ ๖๕๙๗

ลงวันที่ ๐๕ กรกฎาคม ๒๕๖๗

ขอขยายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรมจำนวน ๘๘ รายการ

น้ำ/น้ำเสีย จำนวน 21 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Arsenic	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
2	Barium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
3	Biochemical Oxygen Demand	1) 5-Day BOD Test, Azide Modification Method <sup>[2]</sup> 2) 5-Day BOD Test, Membrane Electrode Method <sup>[2]</sup>
4	Cadmium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
5	Chemical Oxygen Demand	Closed Reflux, Titrimetric Method <sup>[2]</sup>
6	Chromium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
7	Copper	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
8	Hexavalent Chromium	Colorimetric Method <sup>[2]</sup>
9	Lead	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
10	Manganese	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
11	Nickel	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
12	Oil & Grease	Liquid-Liquid, Partition Gravimetric Method <sup>[2]</sup>
13	pH	Electrometric Method <sup>[2]</sup>
14	Selenium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
15	Sulfide	Iodometric Method <sup>[2]</sup>
16	Temperature	Laboratory and Field Methods <sup>[2]</sup>
17	Total Dissolved Solids	Dried at 180 °C <sup>[2]</sup>
18	Total Kjeldahl Nitrogen	Semi-Micro-Kjeldahl Method <sup>[2]</sup>
19	Total Suspended Solids	Dried at 103-105 °C <sup>[2]</sup>
20	Trivalent Chromium	Digestion, Digestion, Inductively Coupled Plasma Method, Colorimetric Method ; Calculation <sup>[2]</sup>
21	Zinc	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>

น้ำใต้ดิน จำนวน 14 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Antimony	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
2	Arsenic	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
3	Barium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
4	Beryllium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
5	Cadmium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
6	Chromium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
7	Lead	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>

8 Manganese...

สำเนาถูกต้อง






ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
8	Manganese	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
9	Nickel	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
10	pH	Laboratory and Field Methods <sup>[2]</sup>
11	Selenium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
12	Silver	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
13	Vanadium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
14	Zinc	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>

อากาศเสีย (ปล่อยระบาย) จำนวน 23 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Antimony	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
2	Arsenic	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
3	Cadmium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
4	Chromium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
5	Cobalt	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
6	Copper	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
7	Chlorine	Absorption Sampling, Ion Chromatographic Method <sup>[3]</sup>
8	Cresol	Adsorption Sampling, Gas Chromatographic Method <sup>[3]</sup>
9	Hydrogen Chloride	Absorption Sampling, Ion Chromatographic Method <sup>[3]</sup>
10	Hydrogen Fluoride	Absorption Sampling, Ion Chromatographic Method <sup>[3]</sup>
11	Hydrogen Sulfide	Absorption Sampling, Ion Chromatographic Method <sup>[3]</sup>
12	Lead	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
13	Manganese	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>

สำเนาถูกต้อง



ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
14	Nickel	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
15	Opacity	Ringelmann's Method <sup>[1]</sup>
16	Selenium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
17	Sulfur Dioxide	Absorption Sampling, Barium-Thorin Titrimetric Method <sup>[3]</sup> Isokinetic Sampling, Barium-Thorin Titrimetric Method <sup>[3]</sup>
18	Sulfuric Acid	Isokinetic Sampling, Barium-Thorin Titrimetric Method <sup>[3]</sup>
19	Tellurium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
20	Tin	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
21	Total Suspended Particulate	Isokinetic Sampling, Gravimetric Method <sup>[3]</sup>
22	Vanadium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
23	Xylene	Adsorption Sampling, Gas Chromatographic Method <sup>[3]</sup>

สิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว จำนวน 17 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Antimony	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
2	Arsenic	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
3	Barium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
4	Beryllium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
5	Cadmium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
6	Chromium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
7	Cobalt	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
8	Copper	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
9	Lead	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
10	Molybdenum	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
11	Nickel	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
12	pH	Electrometric Method <sup>[6,7]</sup>
13	Selenium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
14	Silver	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
15	Thallium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
16	Vanadium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
17	Zinc	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>

สำนักงานกคค.



ดิน...



ดิน จำนวน 13 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Antimony	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
2	Arsenic	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
3	Barium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
4	Beryllium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
5	Cadmium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
6	Chromium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
7	Lead	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
8	Manganese	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
9	Nickel	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
10	Selenium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
11	Silver	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
12	Vanadium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
13	Zinc	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>

เอกสารอ้างอิง

1. กระทรวงอุตสาหกรรม. ประกาศกระทรวงอุตสาหกรรม เรื่อง กำหนดค่าปริมาณเขม่าควันที่เจือปนในอากาศที่ระบายออกจากปล่องของหม้อน้ำของโรงงาน พ.ศ. 2549. ราชกิจจานุเบกษา. 4 ธันวาคม 2549. เล่มที่ 123 ตอนพิเศษ 125 ง.

2. APHA, AWWA, WEF. Standard Methods for the Examination of Water and Wastewater. 23<sup>rd</sup> ed. Washington, DC: APHA, 2017.

3. United States Environmental Protection Agency. Standard of Performance for New Stationary Source. 40 CFR 60. Appendix A, 2023.

4. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Acid Digestion of Sediments, Sludges, and Soils. SW-846 Method 3050B, 1996.

5. United States Environmental Protection Agency. Test Method for Evaluation Solid Waste Physical/Chemical Methods. Inductively Coupled Plasma-Optical Emission Spectrometry. SW-846 Method 6010D, 2018.

6. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. pH Electrometric Measurement. SW-846 Method 9040C, 2004.

7. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Solid and Waste pH. SW-846 Method 9045D, 2004.

Q. J. M.  
**สำเนาถูกต้อง**



ที่ อก ๐๓๑๐(๑)/ ๘๘๗ ๑



กรมโรงงานอุตสาหกรรม  
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท  
เขตราชเทวี กรุงเทพฯ ๑๐๔๐๐

๐๒ กันยายน ๒๕๖๗

เรื่อง แก้ไขวิธีวิเคราะห์ในข้อบ่งชี้อากาศเสีย (ปล่อยระบาย)

เรียน กรรมการผู้จัดการ บริษัท เอ็ม อี ที จำกัด

อ้างถึง ๑. หนังสือกรมโรงงานอุตสาหกรรม ที่ อก ๐๓๑๐(๑)/๖๕๕๗ ลงวันที่ ๕ กรกฎาคม ๒๕๖๗

๒. คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และชนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน  
ลงวันที่ ๕ สิงหาคม ๒๕๖๗

ตามหนังสือที่อ้างถึง ๑ กรมโรงงานอุตสาหกรรมได้ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชนของ บริษัท เอ็ม อี ที จำกัด เลขทะเบียน ว-๑๐๐ สถานที่ตั้งเลขที่ ๓๖/๖๕๕ หมู่ที่ ๖ ตำบลบางรักพัฒนา อำเภอบางบัวทอง จังหวัดนนทบุรี และตามหนังสือที่อ้างถึง ๒ บริษัทฯ ได้ขอแก้ไขวิธีวิเคราะห์ในหนังสือดังกล่าว ความละเอียดแจ้งแล้ว นั้น

กรมโรงงานอุตสาหกรรม ได้แก้ไขวิธีวิเคราะห์ในข้อบ่งชี้อากาศเสีย (ปล่อยระบาย) ดังนี้

๑. ลำดับที่ ๔ รายการ Chromium และ ลำดับที่ ๕ รายการ Cobalt ให้ใช้วิธีวิเคราะห์  
“Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method”

๒. ลำดับที่ ๑๑ รายการ Hydrogen Sulfide แก้ไขวิธีวิเคราะห์ จาก “Absorption Sampling, Ion Chromatographic Method และ Isokinetic Sampling, Ion Chromatographic Method” เป็น  
“Absorption Sampling, Iodometric Method”

อนึ่ง หนังสือฉบับนี้จะหมดอายุพร้อมหนังสือต่ออายุรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน  
ในวันที่ ๒ มิถุนายน ๒๕๗๑

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ

นค

(นายพรยศ กลั่นกรอง)

รองอธิบดี ปฏิบัติราชการแทน  
อธิบดีกรมโรงงานอุตสาหกรรม

กองวิจัยและเตือนภัยมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๐๔-๕

โทรสาร ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๙๙

ไปรษณีย์อิเล็กทรอนิกส์ saraban@diw.mail.go.th



สำเนาถูกต้อง

๐  
๖. ๖๓.







๐๙ กันยายน ๒๕๖๗

เรื่อง ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

เรียน กรรมการผู้จัดการ บริษัท ท็อปส์-แลบ คอนซัลแตนท์ จำกัด

อ้างถึง คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และชนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน  
ลงวันที่ ๑๐ มิถุนายน ๒๕๖๗

สิ่งที่ส่งมาด้วย เอกสารแนบท้ายหนังสือต่ออายุรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน  
บริษัท ท็อปส์-แลบ คอนซัลแตนท์ จำกัด จำนวน ๔ แผ่น

ตามคำขอที่อ้างถึง บริษัท ท็อปส์-แลบ คอนซัลแตนท์ จำกัด ขอต่ออายุหนังสือรับขึ้นทะเบียน  
ห้องปฏิบัติการวิเคราะห์เอกชน เลขทะเบียน ว-๓๒๖ สถานที่ตั้งเลขที่ ๑๘๙ หมู่ที่ ๓ ตำบลบางรักพัฒนา  
อำเภอบางบัวทอง จังหวัดนนทบุรี ต่อกรมโรงงานอุตสาหกรรม นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว ให้บริษัท ท็อปส์-แลบ คอนซัลแตนท์ จำกัด ต่ออายุ  
หนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน โดยมีองค์ประกอบดังนี้


ก. ผู้ควบคุมห้องปฏิบัติการวิเคราะห์เอกชน

- |                        |                            |
|------------------------|----------------------------|
| ๑) นางภักชนิตา พัสระ   | ทะเบียนเลขที่ ว-๓๒๖-ค-๐๐๐๒ |
| ๒) นางสาวเมธาวี คุ่มขำ | ทะเบียนเลขที่ ว-๓๒๖-ค-๐๐๐๓ |

ข. เจ้าหน้าที่ห้องปฏิบัติการวิเคราะห์เอกชน

- |                             |                            |
|-----------------------------|----------------------------|
| ๑) นายภาณุวัฒน์ ขาวชายโขง   | ทะเบียนเลขที่ ว-๓๒๖-จ-๐๐๐๑ |
| ๒) นายภควัต เทียมระกิจ      | ทะเบียนเลขที่ ว-๓๒๖-จ-๐๐๐๒ |
| ๓) นายธีรพงษ์ ชลวิริยะกุล   | ทะเบียนเลขที่ ว-๓๒๖-จ-๐๐๐๓ |
| ๔) นางสาวสุกษา จันทาโท      | ทะเบียนเลขที่ ว-๓๒๖-จ-๐๐๐๔ |
| ๕) นางสาวปณิตา จันทะสม      | ทะเบียนเลขที่ ว-๓๒๖-จ-๐๐๐๕ |
| ๖) นางสาวสุภาพร น้อยลา      | ทะเบียนเลขที่ ว-๓๒๖-จ-๐๐๐๖ |
| ๗) นางสาวนลพรรณ บัวหุ่น     | ทะเบียนเลขที่ ว-๓๒๖-จ-๐๐๐๗ |
| ๘) นายสุวิวัฒน์ อินทร์ช่วย  | ทะเบียนเลขที่ ว-๓๒๖-จ-๐๐๐๘ |
| ๙) นายพีระศักดิ์ ชูแก้ว     | ทะเบียนเลขที่ ว-๓๒๖-จ-๐๐๐๙ |
| ๑๐) นายทรงภพ ศรีทราบุญ      | ทะเบียนเลขที่ ว-๓๒๖-จ-๐๐๑๑ |
| ๑๑) นายศุภณัฐ ไชยลาภ        | ทะเบียนเลขที่ ว-๓๒๖-จ-๐๐๑๒ |
| ๑๒) นายฉัตรชัย ยาทะเล       | ทะเบียนเลขที่ ว-๓๒๖-จ-๐๐๑๓ |
| ๑๓) นางสาวมณิกา บุตรศรี     | ทะเบียนเลขที่ ว-๓๒๖-จ-๐๐๑๔ |
| ๑๔) นางสาวกุลภัสสร์ เขยโชติ | ทะเบียนเลขที่ ว-๓๒๖-จ-๐๐๑๕ |
| ๑๕) นายพุดพิงษ์ ภาคภูมิ     | ทะเบียนเลขที่ ว-๓๒๖-จ-๐๐๑๖ |
| ๑๖) นายทัศนัย มอญจัตรัส     | ทะเบียนเลขที่ ว-๓๒๖-จ-๐๐๑๗ |
| ๑๗) นางสาวจิรัชญา รอยรัตน   | ทะเบียนเลขที่ ว-๓๒๖-จ-๐๐๑๘ |



From   
สำเนาถูกต้อง

ค. ขอบข่ายชนิดสารมลพิษที่ได้รับขึ้นทะเบียนให้วิเคราะห์ในน้ำ/น้ำเสีย น้ำใต้ดิน อากาศเสีย  
สิ่งปฏิกูลหรือวัสดุที่ไม่ใช่แล้ว และดิน ตามสิ่งที่ส่งมาด้วย

หนังสือฉบับนี้จะหมดอายุในวันที่ ๒๕ สิงหาคม ๒๕๗๑ หากประสงค์จะต่ออายุหนังสือรับขึ้น  
ทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ให้ยื่นคำขอต่ออายุพร้อมเอกสารประกอบคำขอต่อกรมโรงงานอุตสาหกรรม  
ภายใน ๖๐ วัน ก่อนวันสิ้นอายุของหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ

(นายพรยศ กลั่นกรอง)

รองอธิบดี ปฏิบัติราชการแทน  
อธิบดีกรมโรงงานอุตสาหกรรม

กองวิจัยและเตือนภัยมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๐๓-๕

โทรสาร ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๙๙

ไปรษณีย์อิเล็กทรอนิกส์ saraban@diw.mail.go.th



TPC com firm

สำเนาถูกต้อง



เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท ท็อปส์-แลบ คอนซัลแตนท์ จำกัด

เลขทะเบียน ว-๓๒๖

ที่ ออก ๐๓๑๐(๑)/ ๙๑๒ ๔

ลงวันที่ ๐๙ กันยายน ๒๕๖๗

ขอขยายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรมจำนวน ๗๕ รายการ

น้ำ/น้ำเสีย จำนวน 22 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Arsenic	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
2	Barium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
3	Biochemical Oxygen Demand	1) 5-Day BOD Test, Azide Modification Method <sup>[2]</sup> 2) 5-Day BOD Test, Membrane Electrode Method <sup>[2]</sup>
4	Cadmium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
5	Chemical Oxygen Demand	Closed Reflux, Titrimetric Method <sup>[2]</sup>
6	Chromium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
7	Color	ADMI Weighted-Ordinate Spectrophotometric Method <sup>[2]</sup>
8	Copper	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
9	Hexavalent Chromium	Colorimetric Method <sup>[2]</sup>
10	Lead	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
11	Manganese	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
12	Nickel	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
13	Oil & Grease	Liquid-Liquid, Partition Gravimetric Method <sup>[2]</sup>
14	pH	Electrometric Method <sup>[2]</sup>
15	Phenols	Distillation, Direct Photometric Method <sup>[2]</sup>
16	Selenium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
17	Sulfide	Iodometric Method <sup>[2]</sup>
18	Temperature	Laboratory and Field Methods <sup>[2]</sup>
19	Total Dissolved Solids	Dried at 180 °C <sup>[2]</sup>
20	Total Kjeldahl Nitrogen	Macro-Kjeldahl Method <sup>[2]</sup>
21	Total Suspended Solids	Dried from 103 to 105 °C <sup>[2]</sup>
22	Zinc	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>

น้ำใต้ดิน จำนวน 17 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Antimony	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
2	Arsenic	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
3	Barium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
4	Beryllium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
5	Cadmium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>



ชื่อนาม  
สำนักงาน

อนุมัติ

6 Chromium...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
6	Chromium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
7	Chromium (III)	Digestion, Inductively Coupled Plasma Method; Colorimetric Method; Calculation <sup>[2]</sup>
8	Chromium (VI)	Colorimetric Method <sup>[2]</sup>
9	Lead	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
10	Manganese	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
11	Nickel	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
12	pH	Electrometric Method <sup>[2]</sup>
13	Phenol	Distillation, Direct Photometric Method <sup>[2]</sup>
14	Selenium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
15	Silver	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
16	Vanadium	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>
17	Zinc	Digestion, Inductively Coupled Plasma Method <sup>[2]</sup>

**อากาศเสีย (ปล่อยระบาย) จำนวน 21 รายการ**

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Antimony	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
2	Arsenic	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
3	Beryllium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
4	Cadmium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
5	Carbon Monoxide	Instrument Analyzer Method <sup>[3]</sup>
6	Chromium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
7	Cobalt	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
8	Copper	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
9	Lead	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
10	Manganese	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>



อนุมัติ

Ilcom firm  
ดำเนินการต้อง

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
11	Nickel	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
13	Opacity	Ringelmann's Method <sup>[1]</sup>
14	Oxides of Nitrogen	Instrument Analyzer Method <sup>[3]</sup>
15	Selenium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
16	Sulfur Dioxide	1) Absorption Sampling, Barium-Thorin Titrimetric Method <sup>[3]</sup> 2) Instrument Analyzer Method <sup>[3]</sup>
17	Sulfuric Acid	Isokinetic Sampling, Barium-Thorin Titrimetric Method <sup>[3]</sup>
18	Tellurium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
19	Tin	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
20	Total Suspended Particulate	Isokinetic Sampling, Gravimetric Method <sup>[3]</sup>
21	Vanadium	Isokinetic Sampling, Digestion, Inductively Coupled Plasma Method <sup>[3]</sup>
22	Xylene	Adsorption Sampling, Gas Chromatographic Method <sup>[3]</sup>

สิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว จำนวน 1 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	pH	Electrometric Method <sup>[6]</sup>

ดิน จำนวน 14 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Antimony	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
2	Arsenic	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
3	Barium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
4	Beryllium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
5	Cadmium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
6	Chromium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
7	Lead	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
8	Manganese	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>



จาก...  
สำนักงานท้องถิ่น

๑๗

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
9	Mercury	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
10	Nickel	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
11	Selenium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
12	Silver	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
13	Vanadium	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>
14	Zinc	Digestion, Inductively Coupled Plasma Method <sup>[4,5]</sup>

#### เอกสารอ้างอิง

1. กระทรวงอุตสาหกรรม. ประกาศกระทรวงอุตสาหกรรม เรื่อง กำหนดค่าปริมาณเขม่าควันที่เจือปนในอากาศที่ระบายออกจากปล่องของหม้อน้ำของโรงงาน พ.ศ. 2549. ราชกิจจานุเบกษา. 4 ธันวาคม 2549. เล่มที่ 123 ตอนพิเศษ 125 ง.
2. APHA, AWWA, WEF. **Standard Methods for the Examination of Water and Wastewater**. 24<sup>th</sup> ed. Washington, DC: APHA, 2023.
3. United States Environmental Protection Agency. **Standard of Performance for New Stationary Source**. 40 CFR 60. Appendix A, 2019.
4. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Acid Digestion of Sediments, Sludges, and Soils. SW-846 Method 3050B**, 1996.
5. United States Environmental Protection Agency. Test Method for Evaluation Solid Waste Physical/Chemical Methods. **Inductively Coupled Plasma-Optical Emission Spectrometry. SW-846 Method 6010D**, 2018.
6. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. **Solid and Waste pH. SW-846 Method 9045D**, 2004.

๑๗



Tom Kim  
สำเนาถูกต้อง





แบบ กก.บญ  
มี.ศ.๒๕๖๕

กรมสวัสดิการและคุ้มครองแรงงาน

ใบอนุญาต

เป็นนิติบุคคลผู้ให้บริการตรวจวัดระดับความเข้มข้นของสารเคมีอันตราย  
ในบรรยากาศของสถานที่ทำงานและสถานที่เก็บรักษาสารเคมีอันตราย

ใบอนุญาตเลขที่ ๐๒๐๑-๐๓-๒๕๖๕-๐๐๕๕

อนุญาตให้ บริษัท เวอร์ดีคอน เป็นหน่วยงานแทน พล.เชอร์วิส จำกัด

เลขทะเบียนนิติบุคคล ๐๑๐๕๕๕๕๐๙๙๙๙๙๖

ตั้งอยู่ เลขที่ ๑๙๙/๑๑๙ หมู่ที่ ๑ ถนนสุขุมวิท-คลองเตย ตำบลคลองเตย อำเภอคลองเตย จังหวัดนนทบุรี  
เป็นนิติบุคคลผู้ให้บริการด้านความปลอดภัย อาชีวอนามัย และสภาพแวดล้อมในการทำงาน ตามกฎกระทรวง  
กำหนดมาตรฐานในการบริหาร จัดการ และดำเนินการด้านความปลอดภัย อาชีวอนามัย และสภาพแวดล้อม  
ในการทำงานเกี่ยวกับสารเคมีอันตราย พ.ศ. ๒๕๕๖ ในการเป็นผู้ให้บริการตรวจวัดระดับความเข้มข้น  
ของสารเคมีอันตรายในบรรยากาศของสถานที่ทำงานและสถานที่เก็บรักษาสารเคมีอันตราย ประกอบกับ  
กฎกระทรวงการขึ้นทะเบียนและกำกับดูแลผู้ประกอบการเพื่อส่งเสริมความปลอดภัย อาชีวอนามัย และสภาพแวดล้อม  
ในการทำงาน พ.ศ. ๒๕๖๔ แห่งพระราชบัญญัติความปลอดภัย อาชีวอนามัย และสภาพแวดล้อมในการทำงาน  
พ.ศ. ๒๕๕๔ โดยมีเลขาธิการ จำนวน ๒ ราย ดำเนินการมอบหมายใบอนุญาตนี้

ทั้งนี้ ตั้งแต่วันที่ ๘ กรกฎาคม พ.ศ. ๒๕๖๕ ถึงวันที่ ๗ กรกฎาคม พ.ศ. ๒๕๖๙

ให้ไว้ ณ วันที่ ๘ กรกฎาคม พ.ศ. ๒๕๖๕

(นายสมพงษ์ กวางแก้ว)  
รองอธิบดี ปฏิบัติราชการแทน  
อธิบดีกรมสวัสดิการและคุ้มครองแรงงาน

รายชื่อบุคลากรแทนท้ายใบอนุญาต  
เป็นนิติบุคคลผู้ให้บริการตรวจวัดระดับความเข้มข้นของสารเคมีอันตรายในบรรยากาศของสถานที่ทำงาน  
และสถานที่เก็บรักษาสารเคมีอันตราย

ขอบริษัท เวอร์ดีคอน เป็นหน่วยงานแทน พล.เชอร์วิส จำกัด

ใบอนุญาตเลขที่ ๐๒๐๑-๐๓-๒๕๖๕-๐๐๕๕

๑. นายประกาย บุญเกิด
๒. นายมานะ สิกหัด

ทั้งนี้ ตั้งแต่วันที่ ๘ กรกฎาคม พ.ศ. ๒๕๖๕ ถึงวันที่ ๗ กรกฎาคม พ.ศ. ๒๕๖๙

ให้ไว้ ณ วันที่ ๘ กรกฎาคม พ.ศ. ๒๕๖๕

(นายสมพงษ์ กวางแก้ว)  
รองอธิบดี ปฏิบัติราชการแทน  
อธิบดีกรมสวัสดิการและคุ้มครองแรงงาน



รายการเครื่องมือวิเคราะห์แบบหัตถ์โยนยูนิท  
 เป็นมีบุคคลผู้ให้บริการวิเคราะห์ความเข้มข้นของสารเคมีอันตรายในบรรยากาศของสถานที่ทำงาน  
 และสถานที่เก็บรักษาสารเคมีอันตราย  
 ของบริษัท เอ็มเม็กซ์ แอสโซซิเอชั่น จำกัด  
 ใบอนุญาตเลขที่ ๐๒๐๒-๐๓-๒๕๖๕-๐๐๒๒

ลำดับที่	รายการเครื่องมือ	รายละเอียด	จำนวน (เครื่อง)
๑	Spectrophotometer (UV/VIS)	ยี่ห้อ HACH รุ่น DR6000 Serial No. 1744836	๑
๒	Atomic Absorption Spectrometer (AAS)	ยี่ห้อ Thermo รุ่น GENESYS 10S VS Serial No. 2E8R020104	๑
๓	Inductively Coupled Plasma Optical Emission Spectroscopy (ICP-OES)	ยี่ห้อ PerkinElmer รุ่น Analyst 200 Serial No. 20059030303	๑
๔	Ion Chromatography (IC)	ยี่ห้อ PerkinElmer รุ่น Optima8300 Serial No. 07851604262	๑
		ยี่ห้อ METROHM รุ่น Compact IC 881/IC Detector 850 Serial No. 1881000010137	๑
๕	Gas Chromatography Flame Ionization Detector (GC-FID)	ยี่ห้อ Thermo รุ่น Dionex Aquion Serial No. 180344663	๑
๖	Gas Chromatography Mass Spectrometer (GC-MS)	ยี่ห้อ Agilent Technologies รุ่น 7890A Serial No. CN12211142	๑
		ยี่ห้อ Agilent Technologies รุ่น 7890 Serial No. CN15263109	๑

ลำดับที่	รายการเครื่องมือ	รายละเอียด	จำนวน (เครื่อง)
๗	เครื่องชั่ง (Electronic Balance)	ยี่ห้อ METTLER TOLEDO รุ่น XP105DR Serial No. B138280195	๑
		ยี่ห้อ METTLER TOLEDO รุ่น XSR105DU Serial No. B902059485	๑
๘	Ion Specific Electrode (ISE)	ยี่ห้อ METTLER TOLEDO รุ่น SevenCompact S220 Serial No. B224014992	๑

นี้ ตั้งแต่วันที่ ๒๓ มีนาคม พ.ศ. ๒๕๖๘ ถึงวันที่ ๒๒ มีนาคม พ.ศ. ๒๕๗๑

ให้ไว้ ณ วันที่ ๒๓ มีนาคม พ.ศ. ๒๕๖๘



(นายศักดิ์ศิลป์ ตูลาธร)  
 ผู้ตรวจราชการกรม ปฏิบัติราชการแทน  
 อธิบดีกรมสวัสดิการและคุ้มครองแรงงาน



แบบ กบ.บญ  
นิติบุคคล

กรมสวัสดิการและคุ้มครองแรงงาน

ใบอนุญาต

เป็นนิติบุคคลผู้ให้บริการตรวจวัดและวิเคราะห์สภาพการทำงานเกี่ยวกับระดับความร้อน

ใบอนุญาตเลขที่ ๐๔๐๑-๐๓-๒๕๖๕-๐๐๕๒

อนุญาตให้ บริษัท เวิร์ดิเยน เอ็นไวรอนเม้นทอล เซอร์วิส จำกัด

เลขทะเบียนนิติบุคคล ๐๓๐๕๕๕๕๕๕๕๕๕๖

ตั้งอยู่เลขที่ ๑๔๔/๑๔๔ หมู่ที่ ๒ ถนนกรุงธนบุรี แขวงบางมดใต้ อำเภอบางมด กรุงเทพมหานคร จังหวัดนนทบุรี  
เป็นนิติบุคคลผู้ให้บริการด้านความปลอดภัย อาชีวอนามัย และสภาพแวดล้อมในการทำงาน ตามกฎกระทรวง  
กำหนดมาตรฐานในการบริหาร จัดการ และดำเนินการด้านความปลอดภัย อาชีวอนามัย และสภาพแวดล้อม  
ในการทำงานเกี่ยวกับความร้อน แสงสว่าง และเสียง พ.ศ. ๒๕๕๙ ในการตรวจวัดและวิเคราะห์สภาพการทำงาน  
เกี่ยวกับระดับความร้อน ประกอบกับกฎกระทรวงการขึ้นทะเบียนและการอนุญาตให้บริการเพื่อส่งเสริม  
ความปลอดภัย อาชีวอนามัย และสภาพแวดล้อมในการทำงาน พ.ศ. ๒๕๖๔ แห่งพระราชบัญญัติความปลอดภัย  
อาชีวอนามัย และสภาพแวดล้อมในการทำงาน พ.ศ. ๒๕๕๔ โดยมีบุคลากร จำนวน ๑ ราย และรายการเครื่องมือ  
ตรวจวัด จำนวน ๑๐ เครื่อง ดังรายละเอียดแนบท้ายใบอนุญาตนี้

ทั้งนี้ ตั้งแต่วันที่ ๒๔ เมษายน พ.ศ. ๒๕๖๕ ถึงวันที่ ๒๔ เมษายน พ.ศ. ๒๕๖๕

ให้ไว้ ณ วันที่ ๒๐ มีนาคม พ.ศ. ๒๕๖๕

(นายศักดิ์ศิลป์ ตุลาธร)

ผู้ตรวจราชการกรม ปฏิบัติราชการแทน  
อธิบดีกรมสวัสดิการและคุ้มครองแรงงาน

กรมสวัสดิการและคุ้มครองแรงงาน  
กระทรวงแรงงาน

รายชื่อบุคลากรแนบท้ายใบอนุญาต  
เป็นนิติบุคคลผู้ให้บริการตรวจวัดและวิเคราะห์สภาพการทำงานเกี่ยวกับระดับความร้อน  
ของบริษัท เวิร์ดิเยน เอ็นไวรอนเม้นทอล เซอร์วิส จำกัด

ใบอนุญาตเลขที่ ๐๔๐๑-๐๓-๒๕๖๕-๐๐๕๒

๑. นายประกาย บุญเกิด

ทั้งนี้ ตั้งแต่วันที่ ๒๔ เมษายน พ.ศ. ๒๕๖๕ ถึงวันที่ ๒๔ เมษายน พ.ศ. ๒๕๖๕

ให้ไว้ ณ วันที่ ๒๐ มีนาคม พ.ศ. ๒๕๖๕

(นายศักดิ์ศิลป์ ตุลาธร)  
ผู้ตรวจราชการกรม ปฏิบัติราชการแทน  
อธิบดีกรมสวัสดิการและคุ้มครองแรงงาน



รายการเครื่องมือตรวจวัดแบบห้วยโป่ง  
เป็นต้นเหตุผู้ให้บริการตรวจวัดและวิเคราะห์ผลการทำงานเกี่ยวกับความ  
ของ บริษัท เอนเนอร์ยี เอ็นไวรอนเม้นทอล เซอร์วิส จำกัด  
ใบอนุญาตเลขที่ ๐๔๐๑-๐๓-๒๕๖๕-๐๐๔๒

ลำดับที่	รายการเครื่องมือ	ยี่ห้อ	รายละเอียด	จำนวน (เครื่อง)
๑	อุปกรณ์ตรวจวัดระดับความชื้น ชนิดอิเล็กทรอนิกส์ที่สามารถอ่าน และคำนวณค่าอุณหภูมิแวดล้อมกลับ (WBGT)	รุ่น	JANTYTECH JT2011-F2A	๑๐
		Serial No.	3522210206	
			3522210207	
			3522211245	
			3522211246	
			3522406010	
			3522406011	
			3522406012	
		3522406013		
		3522406014		
		3522406015		
		มาตรฐาน	ISO 7243	

ทั้งนี้ ตั้งแต่วันที่ ๒๙ เมษายน พ.ศ. ๒๕๖๕ ถึงวันที่ ๒๘ เมษายน พ.ศ. ๒๕๖๕

ให้ไว้ ณ วันที่ ๒๐ มีนาคม พ.ศ. ๒๕๖๕

*[Signature]*

(นายศักดิ์ศิลป์ ตูสาร)  
ผู้ตรวจราชการกรม ปฏิบัติราชการแทน  
อธิบดีกรมสวัสดิการและคุ้มครองแรงงาน

*[Signature]*



แบบ กบญ  
ชนิดบุคคล

### กรมสวัสดิการและคุ้มครองแรงงาน

ใบอนุญาต

เป็นนิติบุคคลผู้ให้บริการตรวจวัดและวิเคราะห์สภาพการทำงานเกี่ยวกับระดับแสงสว่าง

ใบอนุญาตเลขที่ ๐๔๐๒-๐๓-๒๕๖๕-๐๐๕๓

อนุญาตให้... บริษัท เวิร์คเดย์ เน้นไวรอนแมนทอล เซอร์วิส จำกัด

เลขทะเบียนนิติบุคคล ๐๑๐๕๕๕๕๕๕๕๕๖

ตั้งอยู่เลขที่ ๑๔๔/๑๑๔ หมู่ที่ ๒ ถนนสุขุมวิท-ถนนพหลโยธิน ตำบลหนองปรือ อำเภอนนทบุรี จังหวัดนนทบุรี  
เป็นนิติบุคคลผู้ให้บริการด้านความปลอดภัย อาชีวอนามัย และสภาพแวดล้อมในการทำงาน ตามกฎกระทรวง  
กำหนดมาตรฐานในการบริหาร จัดการ และดำเนินการด้านความปลอดภัย อาชีวอนามัย และสภาพแวดล้อม  
ในการทำงานเกี่ยวกับความร้อน แสงสว่าง และเสียง พ.ศ. ๒๕๕๙ ในการตรวจวัดและวิเคราะห์สภาพการทำงาน  
เกี่ยวกับระดับแสงสว่าง ประกอบกับกฎกระทรวงการขึ้นทะเบียนและการอนุญาตให้บริการเพื่อส่งเสริม  
ความปลอดภัย อาชีวอนามัย และสภาพแวดล้อมในการทำงาน พ.ศ. ๒๕๖๔ แห่งพระราชบัญญัติความปลอดภัย  
อาชีวอนามัย และสภาพแวดล้อมในการทำงาน พ.ศ. ๒๕๕๔ โดยมีบุคลากร จำนวน ๑ ราย และรายการเครื่องมือ  
ตรวจวัด จำนวน ๒ เครื่อง ดังรายละเอียดแนบท้ายใบอนุญาตนี้

ทั้งนี้ ตั้งแต่วันที่ ๒๙ เมษายน พ.ศ. ๒๕๖๕ ถึงวันที่ ๒๘ เมษายน พ.ศ. ๒๕๖๕

ให้ไว้ ณ วันที่ ๒๐ มีนาคม พ.ศ. ๒๕๖๕

(นายศักดิ์ศิลป์ ตูลาสกร)

ผู้ตรวจราชการกรม ปฏิบัติราชการแทน

อธิบดีกรมสวัสดิการและคุ้มครองแรงงาน



รายชื่อบุคลากรแนบท้ายใบอนุญาต  
เป็นนิติบุคคลผู้ให้บริการตรวจวัดและวิเคราะห์สภาพการทำงานเกี่ยวกับระดับแสงสว่าง  
ของบริษัท เวิร์คเดย์ เน้นไวรอนแมนทอล เซอร์วิส จำกัด

ใบอนุญาตเลขที่ ๐๔๐๒-๐๓-๒๕๖๕-๐๐๕๓

๑. นายประกาย บุญเกิด

ทั้งนี้ ตั้งแต่วันที่ ๒๙ เมษายน พ.ศ. ๒๕๖๕ ถึงวันที่ ๒๘ เมษายน พ.ศ. ๒๕๖๕

ให้ไว้ ณ วันที่ ๒๐ มีนาคม พ.ศ. ๒๕๖๕

(นายศักดิ์ศิลป์ ตูลาสกร)

ผู้ตรวจราชการกรม ปฏิบัติราชการแทน

อธิบดีกรมสวัสดิการและคุ้มครองแรงงาน





รายการเครื่องมือตรวจวัดแบบพกพาใบอนุญาต  
เป็นนิติบุคคลผู้ให้บริการตรวจวัดและวิเคราะห์สภาวะการทำงานเกี่ยวกับระดับแสงสว่าง  
ของบริษัท เวิร์ธเวิลด์ เอ็นไวรอนเม้นทอล เซอร์วิส จำกัด  
ใบอนุญาตเลขที่ ๐๔๐๒-๐๓-๒๕๖๕-๐๐๔๓

ลำดับที่	รายการเครื่องมือ	รายละเอียด	จำนวน (เครื่อง)
๑	เครื่องวัดแสง	ยี่ห้อ EXTECH	๒
		รุ่น 407026	
		Serial No. A.043002	
		A.052153	
		มาตรฐาน CIE	

ทั้งนี้ ตั้งแต่วันที่ ๒๔ เมษายน พ.ศ. ๒๕๖๕ ถึงวันที่ ๒๘ เมษายน พ.ศ. ๒๕๖๕

ให้ไว้ ณ วันที่ ๒๐ มีนาคม พ.ศ. ๒๕๖๕



(นายศักดิ์ศิลป์ ตูลาธร)

ผู้ตรวจราชการกรม ปฏิบัติราชการแทน  
อธิบดีกรมสวัสดิการและคุ้มครองแรงงาน







แบบ กภ.บญ  
นิติบุคคล

กรมสวัสดิการและคุ้มครองแรงงาน

ใบอนุญาต

เป็นนิติบุคคลผู้ให้บริการตรวจวัดและวิเคราะห์สภาพการทำงานเกี่ยวกับระดับเสียง

ใบอนุญาตเลขที่ ๐๔๐๓-๐๓-๒๕๖๕-๐๐๔๒

อนุญาตให้...บริษัท เวิร์คเดย์ เอ็นไวรอนเม้นท์โฮล เซอร์วิส จำกัด

เลขทะเบียนนิติบุคคล ๐๑๐๕๕๕๘๘๘๕๕๖

ตั้งอยู่เลขที่ ๑๔๔/๑๔๔ หมู่ที่ ๒ ถนนพหลโยธิน ตำบลหนองแขม กรุงเทพมหานคร จังหวัดนนทบุรี  
เป็นนิติบุคคลผู้ให้บริการด้านความปลอดภัย อาชีวอนามัย และสภาพแวดล้อมในการทำงาน ตามกฎหมายที่กำหนดมาตรฐานในการบริหาร จัดการ และดำเนินการด้านความปลอดภัย อาชีวอนามัย และสภาพแวดล้อมในการทำงานเกี่ยวกับความรบกวน แสงสว่าง และเสียง พ.ศ. ๒๕๕๔ ในการตรวจวัดและวิเคราะห์สภาพการทำงานเกี่ยวกับการทำงานเกี่ยวกับระดับเสียง ประกอบกับกฎกระทรวงกำหนดการอนุญาตให้บริการเพื่อส่งเสริมความปลอดภัย อาชีวอนามัย และสภาพแวดล้อมในการทำงาน พ.ศ. ๒๕๖๔ แห่งพระราชบัญญัติความปลอดภัย อาชีวอนามัย และสภาพแวดล้อมในการทำงาน พ.ศ. ๒๕๕๔ โดยมีบุคลากร จำนวน ๑ ราย และรายการเครื่องมือตรวจวัด จำนวน ๒๒ เครื่อง ดังรายละเอียดแนบท้ายใบอนุญาตนี้

ทั้งนี้ ตั้งแต่วันที่ ๒๔ เมษายน พ.ศ. ๒๕๖๕ ถึงวันที่ ๒๔ เมษายน พ.ศ. ๒๕๖๕

ให้ไว้ ณ วันที่ ๒๐ มีนาคม พ.ศ. ๒๕๖๕

(นายศักดิ์ศิลป์ ตูลาธร)

ผู้ตรวจราชการกรม ปฏิบัติราชการแทน  
อธิบดีกรมสวัสดิการและคุ้มครองแรงงาน

กรมสวัสดิการและคุ้มครองแรงงาน  
กระทรวงแรงงาน  
100

รายชื่อบุคลากรแบบท้ายใบอนุญาต  
เป็นนิติบุคคลผู้ให้บริการตรวจวัดและวิเคราะห์สภาพการทำงานเกี่ยวกับระดับเสียง  
ของบริษัท เวิร์คเดย์ เอ็นไวรอนเม้นท์โฮล เซอร์วิส จำกัด  
ใบอนุญาตเลขที่ ๐๔๐๓-๐๓-๒๕๖๕-๐๐๔๒

๑. นายประกาย บุญเกิด

ทั้งนี้ ตั้งแต่วันที่ ๒๔ เมษายน พ.ศ. ๒๕๖๕ ถึงวันที่ ๒๔ เมษายน พ.ศ. ๒๕๖๕

ให้ไว้ ณ วันที่ ๒๐ มีนาคม พ.ศ. ๒๕๖๕

(นายศักดิ์ศิลป์ ตูลาธร)  
ผู้ตรวจราชการกรม ปฏิบัติราชการแทน  
อธิบดีกรมสวัสดิการและคุ้มครองแรงงาน

กรมสวัสดิการและคุ้มครองแรงงาน  
กระทรวงแรงงาน  
100

รายการเครื่องมือตรวจวัดแบบพกพาใบอนุญาต  
เป็นต้นบุคคลผู้ให้บริการตรวจวัดและวิเคราะห์การทำงานเกี่ยวกับระดับเสียง  
ของ บริษัท เวอร์เนี่ยน เอ็นไวรอนเม้นทอล เซอร์วิส จำกัด  
ใบอนุญาตเลขที่ ๐๔๐๓-๐๓-๒๕๖๕-๐๐๔๒

ลำดับที่	รายการเครื่องมือ	รายละเอียด	จำนวน (เครื่อง)
๑	เครื่องวัดเสียง และ เครื่องวัดเสียงกะทบหรือเสียงสะท้อน	ยี่ห้อ SCARLET TECH	๘
		รุ่น ST-21D	
		Serial No.	
		820402	
		820471	
		820472	
		820473	
๒	เครื่องวัดปริมาณเสียงสะสม	820474	๒
		820475	
		820476	
		821061	
		มาตรฐาน IEC 61672	
		ยี่ห้อ SCARLET TECH	
		รุ่น ST-11D	
		Serial No.	
		820861	
		820862	
		มาตรฐาน IEC 61672	
		ยี่ห้อ CIRRUS	
		รุ่น CR-110A	
		Serial No.	๑๐
		CB0451	
		CB0452	
		CB0453	
		CB0454	
		CB0455	
		CB0631	
		CB0632	
		CB0633	
		CB0634	
		มาตรฐาน IEC 61252	
		CB0635	

ลำดับที่	รายการเครื่องมือ	รายละเอียด	จำนวน (เครื่อง)
๓	อุปกรณ์ตรวจสอบความถูกต้อง	ยี่ห้อ QUEST ELECTRONICS	๑
		รุ่น QC-10	
		Serial No.	
		QE3020266	
		มาตรฐาน IEC 60942	
		ยี่ห้อ CIRRUS	
		รุ่น RC-110A	๑
		Serial No.	
		48669	
		มาตรฐาน IEC 60942	

ทั้งนี้ ตั้งแต่วันที่ ๒๙ เมษายน พ.ศ. ๒๕๖๘ ถึงวันที่ ๒๘ เมษายน พ.ศ. ๒๕๖๙

ให้ไว้ ณ วันที่ ๒๐ มีนาคม พ.ศ. ๒๕๖๘

(นายศักดิ์ศิลป์ ตูลาธร)  
ผู้ตรวจราชการกรม ปฏิบัติราชการแทน  
อธิบดีกรมสวัสดิการและคุ้มครองแรงงาน

หนังสือสอบเทียบอุปกรณ์ในการตรวจวัด (Calibration Sheet)

ใบรับรองการสอบเทียบเครื่องมือ

น้ำหนัก/น้ำหนักเสีย

Analytical Balances

Model : XS205DU

Serial No : B814654693




Mettler-Toledo (Thailand) Ltd.  
846/4 - 846/5 Lasalle Rd., Bangna Tai Sub-District  
Bangna District, Bangkok 10260  
+662 723 0382  
MT-TH.ServiceSupport@mt.com



## Accuracy Calibration Certificate

### Customer

Company: Eastern Thai Consulting 1992 Co., Ltd.  
Address: 129 Moo 1, Nonsi  
City: Kabin Buri Contact: Tassawan Chansamrong  
Zip / Postal: 25110  
State / Province: Prachinburi  
Order Number:   
0 3 3 3 4 3 3 1 4

### Weighing Device

Manufacturer: Mettler Toledo Instrument Type: Weighing Instrument  
Model: XS205DU Asset Number: KB-LAB-61/002  
Serial No.: B814654693 Terminal Model: SAT  
Building: Office Laboratory Terminal Serial No.: B814654693  
Floor: 1 Terminal Asset No.: N/A  
Room: Laboratory

Range	Max. Capacity	Readability (d)
1	81 g	0.00001 g
2	220 g	0.0001 g

### Procedure

Calibration Guideline: EURAMET cg-18 v. 4.0 (11/2015)  
METTLER TOLEDO Work Instruction: CP/W002/20

This calibration certificate contains measurements for As Found calibration. No As Left calibration was performed because the device was not modified after As Found calibration. Therefore, results for As Left correspond to As Found.

The sensitivity/span of the weighing instrument was adjusted before calibration with a built-in weight.

In accordance with EURAMET cg-18 (11/2015), the test loads were selected to reflect the specific use of the weighing device or to accommodate specific calibration conditions.

As Found	Temperature		Humidity	
	Start: 24.7 °C	End: 23.4 °C	Start: 55.4 %	End: 48.9 %

As Found Calibration Date: 21-Mar-2025  
As Left Calibration Date: N/A  
Issue Date: 24-Mar-2025

Calibrator:

Pairada Meephet

Approved Signatory:

Technical Manager / Head of Calibration Center

COPY

## Measurement Results


### Repeatability

Test Load: 70 g

	As Found	As Left
1	69.99999 g	N/A
2	70.00000 g	N/A
3	70.00001 g	N/A
4	69.99999 g	N/A
5	70.00000 g	N/A
6	69.99998 g	N/A
7	69.99999 g	N/A
8	70.00000 g	N/A
9	69.99999 g	N/A
10	70.00001 g	N/A

Standard Deviation	0.000010 g	N/A
--------------------	------------	-----

○ As Found  
◆ As Left



The "d" in the graph represents the readability of the range/interval in which the test was performed.


The results of this graph are based upon the absolute values of the differences from the mean value.

### Eccentricity

Test Load: 100 g

Position	As Found	As Left
1	100.0001 g	N/A
2	100.0001 g	N/A
3	100.0000 g	N/A
4	100.0000 g	N/A
5	100.0001 g	N/A

Maximum Deviation	0.0001 g	N/A
-------------------	----------	-----



As Found

The "d" in the graph represents the readability of the range/interval in which the test was performed.


**COPY**

## Error of Indication

### As Found

	Reference Value	Indication	Error of Indication	Expanded Uncertainty	k
1	0.00000 g	0.00000 g	0.00000 g	0.020 mg	2
2	0.05001 g	0.05001 g	0.00000 g	0.024 mg	2
3	0.10001 g	0.10001 g	0.00000 g	0.025 mg	2
4	0.50001 g	0.50002 g	0.00001 g	0.031 mg	2
5	1.00001 g	1.00001 g	0.00000 g	0.034 mg	2
6	5.00001 g	5.00000 g	-0.00001 g	0.049 mg	2
7	10.00001 g	9.99999 g	-0.00002 g	0.062 mg	2
8 <sup>1</sup>	50.00000 g	49.99995 g	-0.00005 g	0.12 mg	2
9	100.0000 g	100.0000 g	0.0000 g	0.21 mg	2
10	150.0000 g	150.0001 g	0.0001 g	0.32 mg	2
11	200.0001 g	200.0002 g	0.0001 g	0.37 mg	2

<sup>1</sup>The calculated uncertainty was replaced by the CMC (Calibration and Measurement Capabilities) value because the calculated uncertainty was smaller than the CMC value.



○ As Found

◆ As Left

For improved legibility of the graphics only increasing measurement points are shown and measurement points close to zero are not displayed.

The expanded measurement uncertainty is reported as the standard measurement uncertainty multiplied by the coverage factor  $k$  such that the coverage probability corresponds to approximately 95 %.

The user is responsible for maintaining environmental conditions and the settings of the weighing instrument when it was calibrated. The results of this calibration certificate relate only to the calibrated item.

**COPY**

## Test Equipment

All weights used for metrological testing are traceable to national or international standards. The weights were calibrated and certified by an accredited calibration laboratory.

### Weight Set 1: OIML E2

Weight Set No.:	<u>WS79</u>	Date of Issue:	<u>19-Feb-2025</u>
Certificate Number:	<u>25-022267</u>	Calibration Due Date:	<u>16-Aug-2026</u>

### Thermo Hygrometer

Equipment No.:	<u>IN322</u>	Date of Issue:	<u>23-Jan-2025</u>
Certificate Number:	<u>SG-H-00170/68</u>	Calibration Due Date:	<u>20-Jan-2026</u>

## Remarks

FACT adjustment functionality activated  
Equipment condition: Good  
Next calibration according to customer's procedure  
Calibration data not decide by calibration laboratory  
XS205DU is Analytical Balances

### End of Accredited Section

The information below and any attachments to this calibration certificate are not part of the accredited calibration.

**COPY**

## Measurement Uncertainty of the Weighing Instrument in Use

Stated is the expanded uncertainty with  $k=2$  in use. The formula shall be used for the estimation of the uncertainty under consideration of the errors of indication. The value  $R$  represents the net load indication in the unit of measure of the device.

Temperature coefficient for the evaluation of the measurement uncertainty in use:  $1.5 \cdot 10^{-6} / K$

Temperature range on site for the evaluation of the measurement uncertainty in use:  $3 K$


### Linearization of Uncertainty Equation

	Range		As Found	As Left
	d	Max		
1	0.00001 g	81 g	$U_1 = 0.021 \text{ mg} + 0.00465 \text{ mg/g} \cdot R$	N/A
2	0.0001 g	220 g	$U_2 = 0.06 \text{ mg} + 0.00464 \text{ mg/g} \cdot R$	N/A


To optimize the stability of the linearization, besides of the zero load only increasing measurement points with a test load of 5% of the measurement range or larger are taken for the calculation of the linear equation.

### Absolute and Relative Measurement Uncertainty in Use for Various Net Indications (Examples)

Net Indication	As Found		As Left	
0.00220 g	0.021 mg	0.96%	N/A	N/A
0.02200 g	0.021 mg	0.096%	N/A	N/A
0.22000 g	0.022 mg	0.010%	N/A	N/A
2.20000 g	0.031 mg	0.0014%	N/A	N/A
220.0000 g	1.1 mg	0.00049%	N/A	N/A



As Found



As Left

The weighing range shown in the absolute uncertainty graph refers to the first interval/range of the device.

**COPY**

**Thermo Reactor**

**Model : RD125**

**Serial No. : 0220/003514**





# Metrological Center

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhoi, Saraburi 18110

Telephone : +66 2 586 5792-4 Fax : +66 2 586 5109

Website : www.scieco.co.th E-Mail : calibrate@scg.co.th

Certificate No. T241379I02

"Substitute for Calibration Certificate Number T241379I01"

Page 1 of 4

## Certificate of Calibration

**Equipment** : Thermoreactor

**Manufacturer** : Lovibond

**Model** : RD125

**Serial No.** : 0220/003514

**Customer Code** : KB-LAB-63/004


**ID No.** : T2603A5

**Customer** : Eastern Thai Consulting 1992 Co.,Ltd.  
129 Moo.1 Nonsi,  
Kabinburi, Prachinburi 25110

**Customer Location** : Laboratory Room

**Date of Receipt** : 17 July 2024

**Calibrated By** : Atiphong Rongrat ( Technician )

**Approved By** :  / Boonchai Suriyawong (Site Calibration Manager)

**Date of Issue** : 27 AUG 2024

The uncertainties are for a confidence probability of approximately 95%.



This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standard laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Metrological Center.



# Metrological Center

## SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhoi, Saraburi 18110

Telephone : +66 2 586 5792-4 Fax : +66 2 586 5109

Website : www.scieco.co.th

E-Mail : calibrate@scg.co.th

Certificate No. T241379I02

Page 2 of 4

## Calibration Report

**Equipment** : Thermoreactor  
**Date of Calibration** : 25 July 2024  
**Environment** : Temperature : 27.4-28.9 °C  
Line Voltage : 222.7-227.8 V  
Relative Humidity : 55 - 65 %RH

### Condition of this results of calibration :

1. This equipment was calibrated by insert 20 standard thermocouples type T into its chamber , the other one standard thermocouples type T use for ambient temperature measurement . The calibration was done in according to WI-T20 ( based on ASTM E145-94 ( Reapproved 2001) and AS2853-1986 ).

All data show below were final values and the initial data from customer request . The temperature scale used was based on ITS - 90 .

### 2. Reference Standard Instrument :

Instrument	Model	Instrument No.	Certificate No.	Due Date
TC	TYPE T	TN241-TN250	T240401	16 March 2025
TC	TYPE T	TN251-TN260	T240401	16 March 2025
DATA LOGGER	34970A	T193	T240401	16 March 2025

### 3. This certificate is traceable to :

National Institute of Metrology ( Thailand ) through Metrological Center ( NSC-TISI-TIS 17025 CALIBRATION 0244.)

### 4. Condition of calibrated item : good

#### Equipment Description :

Time Constant 1 Hour 30 Minute At 150 °C  
Fresh Air Damper ☐ Open ☐ Min ☐ Medium ☐ Max  
☐ Close  
☒ Not Available

### 5. Adjustment :

( X ) without adjustment

( ) after adjustment

Approved By. \_\_\_\_\_

COPY

Certificate No T241379I02


Page 3 of 4

## Calibration Report

Row 6	H 21		H 22		H 23		H 24
Row 5	H 17		H 18		H 19		H 20
Row 4	HOLD		H 14		H 15		H 16
Row 3	H 9		H 10		H 11		H 12
Row 2	H 5		H 6		H 7		H 8
Row 1	H 1		H 2		H 3		H 4

H: STANDARD THERMOCOUPLE TYPE T

H1	=	TN241	H9	=	TN249	H17	=	TN257
H2	=	TN242	H10	=	TN250	H18	=	TN258
H3	=	TN243	H11	=	TN251	H19	=	TN259
H4	=	TN244	H12	=	TN252	H20	=	TN260
H5	=	TN245	H13	=	Hold	H21	=	TN241
H6	=	TN246	H14	=	TN254	H22	=	TN242
H7	=	TN247	H15	=	TN255	H23	=	TN243
H8	=	TN248	H16	=	TN256	H24	=	TN244



Approved By. \_\_\_\_\_



COPY

Certificate No.    **T241379I02**

Page   4   of   4

## Calibration Report

### Measurement Results

Calibration Point			Average Standard Reading at each position ( ° C )									
			TN241	TN242	TN243	TN244	TN245	TN246	TN247	TN248	TN249	TN250
			H1	H2	H3	H4	H5	H6	H7	H8	H9	H10
Point	Setting	Max	150.23	149.91	150.53	149.99	150.89	150.09	150.14	150.64	149.95	150.31
150	150.0	Min	149.93	149.70	150.37	149.75	150.73	150.03	150.01	150.52	150.15	150.16
		Average	150.08	149.82	150.45	149.88	150.80	150.06	150.07	150.58	149.93	150.24
			TN251	TN252	TN253	TN254	TN255	TN256	TN257	TN258	TN259	TN260
			H11	H12	H13	H14	H15	H16	H17	H18	H19	H20
		Max	150.23	150.77	HOLD	150.82	150.31	150.45	150.84	150.84	150.33	150.22
		Min	150.12	150.62		150.76	150.20	150.39	150.77	150.75	150.22	150.13
		Average	150.17	150.69		150.79	150.25	150.42	150.80	150.80	150.27	150.17
			TN241	TN242	TN243	TN244						
			H21	H22	H23	H24						
		Max	149.98	149.98	150.54	150.01						
		Min	149.91	149.92	150.48	149.95						
		Average	149.94	149.95	150.51	149.98						

Thermoreactor			Temperature Distribution	
Setting ( ° C )	Reading ( ° C )		Stability ( ± C )	Uncertainty ( ± ° C )
	Min , Max	Average		
150		150	0.20	1.20

The calibration result apply only the above calibrated item.

The result of test was found accurate as shown on date and place of test only.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor *k* which for a t-distribution, providing a level of confidence of approximately 95 % .

By.   
 Approved \_\_\_\_\_

COPY

**Chamber (Incubator)**

**Model : TC445s**

**Serial No. : 0922/006981**



# Metrology

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhoi, Saraburi 18110, Thailand.

Saraburi Tel : +66 3627 3096 Fax : +66 3627 3100

Bangkok Tel : +668 9205 6851 , +669 8247 2360

Website : www.scieco.co.th E-Mail : calibrate@scg.com



Certificate No. T250467

Page 1 of 3

## Certificate of Calibration

**Equipment** : Chamber ( Incubator )

**Manufacturer** : Lovibond

**Model** : TC445s

**Serial No.** : 0922/006981

**Customer Code** : KB-LAB-66/003

**ID No.** : T4983A5

**Customer** : Eastern Thai Consulting 1992 Co.,Ltd.


129 Moo.1 Nonsi,

Kabinburi, Prachinburi 25110

**Customer Location** : Laboratory

**Date of Receipt** : 12 March 2025

**Calibrated By** : Sujjar Naknakred ( Site Calibration Manager )

**Approved By** :  Boonchai Suriyawong (Site Calibration Manager)

**Date of Issue** : 24 MAR 2025

The uncertainties are for a confidence probability of approximately 95%.

**COPY**

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standard laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Metrology.



Certificate No. T250467

Page 2 of 3

## Calibration Report

**Equipment** : Chamber ( Incubator )  
**Date of Calibration** : 20 March 2025  
**Environment** : Temperature : 27.8-28.3 °C  
Line Voltage : 223.9-231.3 V  
Relative Humidity : 55 - 65 %RH

### Condition of this results of calibration :

1. This equipment was calibrated by insert nine resistance thermometer detectors into its chamber , the other one resistance thermometer detector use for ambient temperature measurement . The calibration was done in according to WI-T20 ( based on ASTM E145-94 ( Reapproved 2019) and AS2853-1986 ).  
All data show below were final values and the initial data from customer request . The temperature scale used was based on ITS - 90 .

2. Reference Standard Instrument :

Instrument	Model	Instrument No.	Certificate No.	Due Date
RTD	100 ohm	28-(CH1-10)	T240709	19 April 2025
DATA LOGGER	34970A	T149	T240709	19 April 2025

3. This certificate is traceable to :

National Institute of Metrology ( Thailand ) through Metrological Center ( NSC-TISI-TIS 17025 CALIBRATION 0244.)

4. Condition of calibrated item : good

Equipment Description :

Time Constant 1 Hour 57 Minute At 20 °C  
Fresh Air Damper ☐ Open ☐ Min ☐ Medium ☐ Max  
☐ Close  
☒ Not Available

5. Adjustment :

( X ) without adjustment

( ) after adjustment


Approved By \_\_\_\_\_

**COPY**

Certificate No. T250467

Page 3 of 3

## Calibration Report



**Remark :** Internal Dimensions of Chamber : W (Width) = 66 cm. , H (Height) = 145 cm. and D (Depth) = 56 cm.  
Size of Installed Standard sensor number 28-CH1 to number 28-CH8 : a = 5 cm. ,b = 5 cm. and c = 5 cm.  
Size of Installed Standard sensor number 28-CH9 : W/2 = 66 cm./2 , H/2 = 145 cm./2 and D/2 = 56cm./2

### Measurement Results

Average Standard Reading at each position (°C)									
Calibration Point	28-CH1	28-CH2	28-CH3	28-CH4	28-CH5	28-CH6	28-CH7	28-CH8	28-CH9
20	20.18	20.07	19.87	19.83	19.89	19.95	19.82	20.01	19.95

Chamber ( Incubator )			Temperature Distribution				
Setting °C	**Reading (°C)		Average ( °C )	Stability (± °C )	Uniformity ( °C )	Uncertainty (± °C )	Coverage Factor k
	Min , Max	Average					
20.8	19.9 , 20.0	20.0	19.95	0.31	0.94	0.52	2.00

\*\* Reading Temperature in Chamber with " Digital thermometer No.KB-LAB-63/022 "

\* The quoted uncertainty exclude "uniformity"

The calibration result apply only the above calibrated item.

The result of test was found accurate as shown on date and place of test only.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k which for a t-distribution, providing a level of confidence of approximately 95 % .

End of Certificate.

Approved By. \_\_\_\_\_



**DO METER**

**Model : YSI5000**

**Serial No. : 16H102702**

CERT.No.: HS-V071J

**Certificate of Calibration**

Calibration Date **4 Oct 24**  
Submitted by : **Eastern Thai Consulting 1992 Co., Ltd.**  
129 Moo 1, Nonsi Subdistrict, Kabinburi  
District Prachinburi Province 25110

Model **YSI 5000**  
S/N **16H102702**  
Probe **YSI 5010**  
S/N **22A100334**  
ID NO. **-**  
Air Temp ref **\$/N. F8065C26**  
Barometric ref **\$/N. F8065C26**  
Water Temp ref **-**  
ID NO. **HS001**  
Technician **Kittipong M**

Avg Room Temp : **20 °C**  
Avg Water Temp : **20 °C**  
Air Pressure : **760.00 mmHg**  
Salinity : **0 ppt**

**Calibration Details**

Calibration Point	100% air sat. (@20 °C, DO = 9.09 mg/l)	(status)	(status)
Measurement 1 (mg/l)	9.09	(PASS)	-
Measurement 2 (mg/l)	9.09	(PASS)	-
Measurement 3 (mg/l)	9.09	(PASS)	-
Measurement 4 (mg/l)	9.09	(PASS)	-
Measurement 5 (mg/l)	9.09	(PASS)	-
Measurement 6 (mg/l)	9.08	(PASS)	-
Measurement 7 (mg/l)	9.08	(PASS)	-
Measurement 8 (mg/l)	9.08	(PASS)	-
Measurement 9 (mg/l)	9.08	(PASS)	-
Measurement 10 (mg/l)	9.09	(PASS)	-

Mean Measurement	9.09	mg/ l	-	-
Inaccuracy	0.00	mg/ l	-	-

Overall Status **(PASS)**

**Manufacturer Specification**

Accuracy = +/- 0.02 mg/l

- 1) This certificate is issued based on the result that are found as shown on date and place of test only.
- 2) The calibration procedure followed in accordance with Harikul Science Co., Ltd
- 3) This result shall not be used for advertising purpose.



Technician Signature  
(Kittipong Maekwong)




Laboratory Manager  
(Supreecha Sumaritam)

**WATER BATH**

**Model : WNB45**

**Serial No. : L719.0236**



## CERTIFICATE OF CALIBRATION

Page 1 of 3

Certificate No. : 24-128424

Sample Code : 24-51593-001

**Customer** : Eastern Thai Consulting 1992 Co., Ltd.  
129 Moo 1, Suwannasorn Rd., Nonsri,  
Kabinburi, Prachinburi 25110

**Location of Calibration** : Eastern Thai Consulting 1992 Co., Ltd.  
(Laboratory)

**Equipment** : Liquid bath (Water bath)

**Manufacturer** : Memmert

**Model** : WNB 45

**Serial No.** : L719.0236

**ID No.** : KB-LAB-63/017

**Date of Receipt** : 03 October 2024

**Date of Calibration** : 03 October 2024

**Condition of Calibration**

1. Environment	1.1 Ambient temperature	: Maximum	31.9 °C	; Minimum	30.3 °C
	1.2 Relative humidity	: Maximum	74.4 %	; Minimum	63.0 %
	1.3 Line voltage supplied	: Maximum	223.8 VAC	; Minimum	218.8 VAC

**2. Calibration method**

In-house method WI-CL-023 based on ASTM E 715-80 (Reapproved 2022).

**3. Reference standard instrument**

Instrument	ID No.	Certificate No.	Due Date
Data acquisition with sensor (RTD-Pt100)	LB-DA-10 (RTD-220 to RTD-224)	24-025104	28 February 2025

**4. This certificate is traceable to the international system of unit (SI Unit).**

The measurement is traceable to Asia Medical and Agricultural Laboratory and Research Center Public Company Limited.

**5. This result of calibration was found accurate as shown on date and place of calibration only.****6. Condition of calibration item** : Normal

**Calibrated by** Mr. Sarut Sa-nguansin  
Scientist

**Approved by** (Mr. Somchai Neampunt)  
Signed for Director

**Issue date** 04 October 2024

The uncertainties are for a confidence probability of approximately 95%.

The calibration result is applied only to the above calibrated item and was found accurate as shown on date and place of calibration only.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the unit of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Asia Medical and Agricultural Laboratory and Research Center Public Company Limited (AMARC).

## REPORT OF CALIBRATION

Page 2 of 3

Certificate No. : 24-128424

Sample Code : 24-51593-001

## Results of Calibration

Resolution : 0.1 °C

## 1. Reporting of Temperature

Calibration point (°C)	UUC* setting (°C)	UUC* reading (°C)	Measured temperature at each positions (°C)					Uncertainty ± (°C)	Coverage factor k
			# 1	# 2	# 3	# 4	# 5 <sup>Ref.</sup>		
85	85.9	85.9	85.067	85.120	85.046	85.068	85.094	0.17	2.00
95	95.0	95.0	93.941	93.919	93.908	93.881	93.919	0.17	2.00

## 2. Characterization results

Calibration point (°C)	Stability ± (°C)	Uniformity (°C)	Overall variation (°C)
85	0.087	0.156	0.202
95	0.078	0.144	0.237

## Notes

- UUC\* = Unit Under Calibration



## REPORT OF CALIBRATION

Page 3 of 3


Certificate No. : 24-128424

Sample Code : 24-51593-001

### Results of Calibration

#### Notes

1. Sensor installation locations
  - 1.1 Place five calibrated temperature sensors in the unloaded water bath with diffuser plate in place and at lowest position and water level approximately 38 mm from the top.
  - 1.2 Locate one sensor in each of the four corners of the bath approximately 50 mm from each wall and midway between the diffuser plate and the water surface.
  - 1.3 Locate the fifth sensor within 25 mm of the geometric center of the bath.
2. The quoted uncertainty includes "Stability of bath and loading effect in bath at 20% of uniformity".
3. Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time.
4. Stability - one-half of the greatest maximum difference of measured temperatures at any one sensor.
5. Overall variation - the difference of the maximum and the minimum measured temperatures throughout observation time.
6. UUC\* reading - the average reading of indicating device that forms the integral part of the bath.
7. Controlled circulation or stirrer motor setting : N/A
8. Cooling system : N/A
9. Calibration results without adjustment.



**Figure: Example of sensor  
installation Positions**

The result expanded uncertainty of measurement  $U$  is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k$ , which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with UKAS M3003.

- End of Report -

**COPY**

## **Chamber ( Refrigerator )**

**Model : SCR-1320SAD**

**Serial No. : 0508-00065**



# Metrology

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhoi, Saraburi 18110, Thailand.

Saraburi Tel : +66 3627 3096 Fax : +66 3627 3100

Bangkok Tel : +668 9205 6851 , +669 8247 2360

Website : www.scieco.co.th E-Mail : calibrate@scg.com



Certificate No. T250468

Page 1 of 3

## Certificate of Calibration

**Equipment** : Chamber ( Refrigerator )

**Manufacturer** : Sanden Intercool

**Model** : SRC-1320SAD

**Serial No.** : 0508-00065

**Customer Code** : KB-LAB-48/034

**ID No.** : T8421A2

**Customer** : Eastern Thai Consulting 1992 Co.,Ltd.  
129 Moo.1 Nonsi,  
Kabinburi, Prachinburi 25110

**Customer Location** : Laboratory

**Date of Receipt** : 12 March 2025

**Calibrated By** : Sujjar Naknakred ( Site Calibration Manager )

**Approved By** :  Boonchai Suriyawong (Site Calibration Manager)

**Date of Issue** : 24 MAR 2025

The uncertainties are for a confidence probability of approximately 95%.

**COPY**

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standard laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Metrology.



Certificate No. T250468

Page 2 of 3

## Calibration Report

**Equipment** : Chamber ( Refrigerator )  
**Date of Calibration** : 20-21 March 2025  
**Environment** : Temperature : 27.8-28.3 °C  
Line Voltage : 223.9-231.3 V  
Relative Humidity : 55 - 65 %RH

### Condition of this results of calibration :

1. This equipment was calibrated by insert nine resistance thermometer detectors into its chamber , the other one resistance thermometer detector use for ambient temperature measurement . The calibration was done in according to WI-T20 ( based on ASTM E145-94 ( Reapproved 2019) and AS2853-1986 ).

All data show below were final values and the initial data from customer request . The temperature scale used was based on ITS - 90 .

2. Reference Standard Instrument :

Instrument	Model	Instrument No.	Certificate No.	Due Date
RTD	100 ohm	28-(CH1-10)	T240709	19 April 2025
DATA LOGGER	34970A	T149	T240709	19 April 2025

3. This certificate is traceable to :

National Institute of Metrology ( Thailand ) through Metrological Center ( NSC-TISI-TIS 17025 CALIBRATION 0244.)

4. Condition of calibrated item : good

Equipment Description :

Time Constant 4 Hour 37 Minute At 3 °C

Fresh Air Damper ☐ Open ☐ Min ☐ Medium ☐ Max  
☐ Close  
☒ Not Available

5. Adjustment :

( ) without adjustment

( X ) after adjustment


Approved By. 

**COPY**

Certificate No. T250468

Page 3 of 3

## Calibration Report



**Remark :** Internal Dimensions of Chamber : W (Width) = 100 cm. , H (Height) = 135 cm. and D (Depth) = 45 cm.  
 Size of Installed Standard sensor number 28-CH1 to number 28-CH8 : a = 5 cm. ,b = 5 cm. and c = 5 cm.  
 Size of Installed Standard sensor number 28-CH9 : W/2 = 100 cm./2 , H/2 = 135 cm./2 and D/2 = 45cm./2

### Measurement Results

Average Standard Reading at each position (°C)									
Calibration Point	28-CH1	28-CH2	28-CH3	28-CH4	28-CH5	28-CH6	28-CH7	28-CH8	28-CH9
3	3.18	2.90	2.99	3.03	3.01	2.82	2.79	3.07	3.22

Chamber ( Refrigerator )			Temperature Distribution				
Setting °C	Reading (°C )		Average ( °C )	Stability (± °C )	Uniformity ( °C )	Uncertainty (± °C )	Coverage Factor k
	Min , Max	Average					
3.0	2.9 , 3.8	3.1	3.00	0.76	1.20	1.21	2.00

\* The quoted uncertainty exclude "uniformity"

The calibration result apply only the above calibrated item.

The result of test was found accurate as shown on date and place of test only.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k which for a t-distribution, providing a level of confidence of approximately 95 % .

End of Certificate.

Approved By. \_\_\_\_\_



**pH METER**

**Model : SevenCompact**

**Serial No. : B824972289**



## CERTIFICATE OF CALIBRATION

Page 1 of 3

Certificate No. : 24-108219

Sample Code : 24-43250-001

**Customer** : Eastern Thai Consulting 1992 Co., Ltd.  
129 Moo 1, Suwannasorn Rd., Nonsri, Kabinburi,  
Prachinburi 25110

**Location of Calibration** : Eastern Thai Consulting 1992 Co., Ltd.  
(Calibration laboratory)

**Equipment** : pH Meter

**Manufacturer** : METTLER TOLEDO

**Model** : SevenCompact S220

**Serial No.** : B824972289

**ID No.** : KB-LAB-61/003

**Date of Receipt** : 26 August 2024

**Date of Calibration** : 26 August 2024

**Condition of Calibration**

1. **Environment** Ambient temperature : 15.0 to 35.0 °C Relative humidity : 35.0 to 80.0 %RH  
1.1 Start time : 24.5 °C ; End time : 25.8 °C 1.2 Start time : 76.8 %RH ; End time : 79.5 %RH

**2. Calibration method**

In house method WI-CL-019 : Direct measurement with standard voltage calibrator and direct measurement with certified reference material (CRM).

**3. Reference standard / Certified reference material**

Instrument	ID No.	Certificate No.	Due Date
3.1 Voltage Calibrator	LB-DPC-01	LF24-0280	25 June 2025
3.2 Digital Thermometer	LB-DPC-01	24-096498	11 August 2025
Certified Reference Material	Lot. No.	Ref No.	Expire Date
3.3 Buffer Solution pH 4.008	991689	PH216.L5	03 May 2026
3.4 Buffer Solution pH 6.999	C03142	S11M004	12 February 2026
3.5 Buffer Solution pH 9.997	980686	PH220.L5	25 April 2025

**4. This certificate is traceable to the international system of unit (SI Unit).**

- 4.1 Instrument No. 3.1 through Measuretronix Limited.  
4.2 Instrument No. 3.2 through Asia Medical and Agricultural Laboratory and Research Center Public Company Limited.  
4.3 Buffer Solution No. 3.3 and No. 3.5 traceable to CPA chem (through primary measurement method-Harned cell using calibrated thermometer, barometer, and nanovoltmeter Accredited laboratory ISO/IEC 17025 and ISO 17034).  
4.4 Buffer Solution No. 3.4 traceable to HACH (Certified Reference Material from DFM and PTB Certificate Nr. CRM-P1118 and Certificate Nr. PTB-PHOB-555/30620/22. According to DIN EN ISO 17034:2017).

**5. This result of calibration was found accurate as shown on date and place of calibration only.****6. Condition of calibration item : Normal****Calibrated by**Mr. Nuttaput Timula  
Scientist**Approved by**(Mr. Somchai Neampunt)  
Signed for Director**Issue date**

27 August 2024

The uncertainties are for a confidence probability of approximately 95%.

The calibration result is applied only to the above calibrated item and was found accurate as shown on date and place of calibration only.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the unit of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Asia Medical and Agricultural Laboratory and Research Center Public Company Limited (AMARC).

NSC-TISI-TIS17025  
CALIBRATION 0152

## REPORT OF CALIBRATION

Page 2 of 3

Certificate No. : 24-108219

Sample Code : 24-43250-001

Equipment : pH Meter Resolution : 0.001 pH ; 0.1 mV ; 0.1°C  
Manufacturer : METTLER TOLEDO Model : SevenCompact S220  
Serial No. : B824972289 ID No. : KB-LAB-61/003  
Range : -2.000 pH to 20.000 pH ;  $\pm 2000.0$  mV ; -5.0°C to 130.0°C

## Results of Calibration

## Part 1. DC Voltage measurement

pH Meter Serial No. : B824972289

Nominal Value pH	Applied DC Voltage mV	Average indicator reading		Uncertainty mV	Coverage factor k
		mV	pH		
0	414.113	414.0	0.00	$\pm 0.083$	2.00
4	177.477	177.5	4.00	$\pm 0.083$	2.00
7	0.000	0.0	7.00	$\pm 0.083$	2.00
10	-177.477	-177.5	10.00	$\pm 0.083$	2.00
14	-414.113	-414.0	14.00	$\pm 0.083$	2.00

## Part 2. Performance of Electrode system

Electrode Manufacturer : METTLER TOLEDO Model : InLab Expert Pro-ISM

Electrode Serial No. : 4293264

Three-Point Calibration at pH4 and pH7 Percent Slope : 99.4 , at pH7 and pH10 Percent Slope : 98.7

Standard Buffer Solution pH (@ 25 °C)	Average indicator reading		Error Value pH	Uncertainty pH	Coverage factor k
	pH	mV			
4.008	4.00	181.2	-0.008	$\pm 0.0083$	2.00
6.999	6.99	4.8	-0.009	$\pm 0.0083$	2.00
9.997	9.99	-170.4	-0.007	$\pm 0.0083$	2.00

The result expanded uncertainty of measurement  $U$  is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k$ , which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with UKAS M3003.

COPY





## REPORT OF CALIBRATION

Certificate No. : 24-108219

Sample Code : 24-43250-001

Equipment : pH Meter (Digital Thermometer with sensor)

## Thermometer readout

Manufacturer : METTLER TOLEDO Model : SevenCompact S220  
 Serial No. : B824972289 ID No. : KB-LAB-61/003  
 Resolution : 0.1 °C Range : -30.0 °C to 130.0 °C

## Thermometer sensor

Manufacturer : METTLER TOLEDO Model : InLab Expert Pro-ISM  
 Serial No. : 4293264 ID No. : N/A

## Condition of Calibration

1. Environment  
 1.1 Ambient temperature : 25.2 °C ± 0.1 °C  
 1.2 Relative humidity : 66.3 % ± 7.2 %

## 2. Calibration method

- 2.1 The calibration use in house method WI-CL-021 : by comparison with standard thermometer  
 2.2 The calibration by comparison unit under calibration (UUC) to the standard thermometer in a calibration bath at the controlled temperature.  
 2.3 The temperature scale in use of this laboratory is the international temperature scale of 1990 (ITS-90).

## 3. Reference standard instrument

Instrument	Model	ID. No.	Certificate No.	Due date
3.1 Resistance Thermometer	PT-100	RTD-366	24-096498	11 August 2025
3.2 Thermometer Readout	753	LB-DPC-01	24-096498	11 August 2025

## 4. This certificate is traceable to the international system of unit (SI Unit).

Asia Medical and Agricultural Laboratory and Research Center Public Company Limited (Accreditation Under TLAS Laboratory Calibration No.0152)

## 5. This result of calibration was found accurate as shown on date and place of calibration only.

## 6. Condition of Calibration item : Normal

## Results of Calibration

Calibration point °C	Average of standard reading °C	Unit under calibration			Expanded uncertainty °C	Coverage factor k
		Immersion depth mm	Average reading °C	Correction value °C		
23	22.98	100	22.8	+ 0.18	± 0.14	2.00
25	24.98	100	24.8	+ 0.18	± 0.14	2.00
27	27.00	100	26.8	+ 0.20	± 0.14	2.00

## Notes

- Calibration results without adjustment

COPY

The result expanded uncertainty of measurement  $U$  is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k$ , which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with UKAS M3003.

- End of report -

**HOT AIR OVEN**

**Model : UF55**

**Serial No. : B224.1762**

## CERTIFICATE OF CALIBRATION

Page 1 of 3

Certificate No. : 25-024434

Sample Code : 25-10468-003

**Customer** : Eastern Thai Consulting 1992 Co., Ltd.  
129 Moo 1, Suwannasorn Rd., Nonsri,  
Kabinburi, Prachinburi 25110

**Location of Calibration** : Eastern Thai Consulting 1992 Co., Ltd.  
(Hot Room)

**Equipment** : Temperature controlled enclosures (Hot air oven)

**Manufacturer** : Memmert

**Model** : UF55

**Serial No.** : B224.1762

**ID No.** : KB-LAB-68/002

**Date of Receipt** : 18 February 2025

**Date of Calibration** : 18-19 February 2025

**Condition of Calibration**

- 1. Environment**
- |                           |   |
|---------------------------|---|
| 1.1 Ambient temperature   | : Maximum 31.8 °C ; Minimum 30.5 °C     |
| 1.2 Relative humidity     | : Maximum 59.9 % ; Minimum 57.6 %       |
| 1.3 Line voltage supplied | : Maximum 224.0 VAC ; Minimum 220.5 VAC |

**2. Calibration method**

TLAS-G-20: Guidelines for calibration and checks of temperature controlled enclosures.

**3. Reference standard instrument**

Instrument	ID No.	Certificate No.	Due Date
Data Acquisition With Sensor (RTD-Pt100)	LB-DA-08 (RTD-239 to RTD-247)	24-099333	31 July 2025

**4. This certificate is traceable to the international system of unit (SI Unit).**

The measurement is traceable to Asia Medical and Agricultural Laboratory and Research Center Public Company Limited.

**5. This result of calibration was found accurate as shown on date and place of calibration only.**

**6. Condition of calibration item** : Normal

**Calibrated by**

Mr. Natthanan Phosri  
Scientist

**Approved by**

(Mr. Somchai Neampunt)  
Signed for Director

**Issue date**

19 February 2025

The uncertainties are for a confidence probability of approximately 95%.

The calibration result is applied only to the above calibrated item and was found accurate as shown on date and place of calibration only.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the unit of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Asia Medical and Agricultural Laboratory and Research Center Public Company Limited (AMARC).



## REPORT OF CALIBRATION

Certificate No. : 25-024434

Sample Code : 25-10468-003

## Results of Calibration

Resolution : 0.1 °C

## 1. Reporting of Temperature

Calibration point (°C)	UUC* setting (°C)	UUC* reading (°C)	Measured temperature at each positions (°C)									Uncertainty ± (°C)	Coverage factor k
			# 1	# 2	# 3	# 4	# 5	# 6	# 7	# 8	# 9 <sup>Ref</sup>		
85	85.0	85.0	85.30	85.14	85.25	85.29	85.17	85.17	85.07	85.19	85.27	0.25	2.00
104	104.0	104.0	104.29	104.08	104.23	104.29	104.13	104.15	104.02	104.21	104.30	0.47	2.00
150	150.0	150.0	150.27	149.88	150.24	150.28	150.07	150.04	149.82	150.27	150.31	0.47	2.00
180	180.0	180.0	180.32	179.86	180.34	180.40	180.12	180.07	179.79	180.47	180.39	0.48	2.00

## 2. Characterization results

Calibration point (°C)	Stability ± (°C)	Uniformity (°C)	Overall variation (°C)
85	0.07	0.24	0.34
104	0.27	0.37	0.80
150	0.20	0.54	0.85
180	0.17	0.67	0.99

## Notes

- UUC\* = Unit Under Calibration



COPY

## REPORT OF CALIBRATION

Page 3 of 3


Certificate No. : 25-024434

Sample Code : 25-10468-003

### Results of Calibration

#### Notes


1. Sensor installation locations
  - 1.1 All sensors at any corners or walls should be positioned 5 cm (a x b x c) from the wall.
  - 1.2 The reference sensor is preferably located of the geometric center of the chamber.
2. Interior dimensions approx of chamber :  
W = 40 cm ; D = 33 cm ; H = 40 cm
3. Air valve or fresh air level : Off
4. Fan level : Open
5. The quoted uncertainty includes " Stability of chamber and loading effect in chamber at 20% of uniformity ".
6. Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time.
7. Stability - one-half of the greatest maximum difference of measured temperatures at any one sensor.
8. Overall variation - the difference of the maximum and the minimum measured temperatures throughout observation time.
9. UUC\* reading - the average reading of indicating device that forms the integral part of the enclosure.
10. Calibration results without adjustment.



**Figure: Example of sensor  
installation Positions**

The result expanded uncertainty of measurement  $U$  is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k$ , which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty of measurement has been determined in accordance with UKAS M3003.

- End of Report -



**COPY**



**SPECTROPHOTOMETER**

**Model : Merck Prove 300**

**Serial No : 2013314847**



Bangkok High Lab Co.,Ltd.  
4/176 Soi Ladplakao 66, Ladplakao Rd., Anusawari, Bangkok, Bangkok 10220  
Tel: (662) 971-5800 Fax: (662) 971-5300  
Website: www.bangkokhighlab.com E-mail: info@bangkokhighlab.com



# CERTIFICATE OF CALIBRATION

Certificate No : S2024/151

Page : 1/5

Order No : 210/2024

Customer : Eastern Thai Consulting 1992 Co.,Ltd.  
Address : 129 Moo 1, Nonsi, Kabinburi, Prachinburi 25110  
Instrument : UV/VIS spectrophotometer  
Manufacture : Merck  
Model : Prove 300  
Serial Number : 2013314847  
Environment : Temperature (25.9 - 25.8) °C  
Humidity (51 - 57) %RH  
Received Date : August 19, 2024  
Calibration Date : August 19, 2024  
Issued Date : August 26, 2024  
Calibrate Status : No Adjustment  
Calibration Area : Customer area  
Roomname : Laboratory Room of Eastern Thai Consulting 1992 Co.,Ltd.

Calibrated By : Pacharapol  
( Mr. Pacharapol Kwanbang )  
Calibration Engineer

Approved By : Wanchai  
( Mr.Wanchai Meesiri )  
Manager

**COPY**

This calibration certificate shall not be reproduced other than in full except with the prior written approval of the Bangkok High Lab Co.,Ltd.



Bangkok High Lab Co.,Ltd.

4/176 Soi Ladplakao 66, Ladplakao Rd., Anusawari, Bangkhen, Bangkok 10220

Tel: (662) 971-5800

Fax: (662) 971-5300

Website: www.bangkokhighlab.com

E-mail: info@bangkokhighlab.com



Certificate No : S2024/151

Page : 2/5

## 1. Photometric Accuracy

CRMs: Neutral Density Glass Filters

CRMs Serial Number: 10563

Traceability: Traceable to NIST, U.S.A. through Neutral density filters NIST SRM 930e & 1930, Double Aperture method through Starna certificate report no.113594

Spectral slit width : 4.00 nm

### 1.1 Reading scale at 420.0 nm

Filter STDs (Abs) Certificate	Average Measured Value (A)	Correction (A)	Uncertainty ± (A)
0.0000	0.000	0.0000	0.0028
0.5604	0.560	0.0004	0.0044
1.0723	1.073	-0.0007	0.0038
2.1753	2.176	-0.0007	0.0064

### 1.2 Reading scale at 440.0 nm

Filter STDs (Abs) Certificate	Average Measured Value (A)	Correction (A)	Uncertainty ± (A)
0.0000	0.000	0.0000	0.0028
0.5503	0.550	0.0003	0.0040
1.0467	1.046	0.0007	0.0040
2.1117	2.110	0.0017	0.0064

### 1.3 Reading scale at 465.0 nm

Filter STDs (Abs) Certificate	Average Measured Value (A)	Correction (A)	Uncertainty ± (A)
0.0000	0.000	0.0000	0.0028
0.4996	0.499	0.0006	0.0034
0.9649	0.964	0.0009	0.0040
1.9646	1.964	0.0006	0.0060

### 1.4 Reading scale at 546.1 nm

Filter STDs (Abs) Certificate	Average Measured Value (A)	Correction (A)	Uncertainty ± (A)
0.0000	0.000	0.0000	0.0028
0.5136	0.513	0.0006	0.0028
0.9765	0.976	0.0005	0.0028
1.9848	1.982	0.0028	0.0064

COPY



Bangkok High Lab Co.,Ltd.

4/176 Soi Ladplakao 66, Ladplakao Rd., Anusawari, Bangkok, Bangkok 10220

Tel: (662) 971-5800

Fax: (662) 971-5300

Website: www.bangkokhighlab.com

E-mail: info@bangkokhighlab.com



NSC-TISI-TIS 17025  
CALIBRATION 0366

Certificate No : S2024/151

Page : 3/5

1.5 Reading scale at 590.0 nm

Filter STDs (Abs) Certificate	Average Measured Value (A)	Correction (A)	Uncertainty ± (A)
0.0000	0.000	0.0000	0.0028
0.5424	0.542	0.0004	0.0029
1.0130	1.012	0.0010	0.0029
2.0238	2.021	0.0028	0.0061

1.6 Reading scale at 635.0 nm

Filter STDs (Abs) Certificate	Average Measured Value (A)	Correction (A)	Uncertainty ± (A)
0.0000	0.000	0.0000	0.0028
0.5265	0.526	0.0005	0.0030
0.9667	0.965	0.0017	0.0031
1.9145	1.913	0.0015	0.0062

## 2. Photometric Accuracy

CRMs: Potassium Dichromate in Perchloric acid

CRMs Serial Number: 132023

Blank Serial Number: 128038

Traceability: Traceable to NIST, U.S.A. through crystalline potassium dichromate NIST SRM 935a through Starna certificate report no.120920

Spectral slit width : 4.00 nm

Wavelength (nm)	Certificate (Abs)	Average Measured Value (A)	Correction (A)	Uncertainty ± (A)
235	0.0000	0.000	0.0000	0.0050
	0.7351	0.737	-0.0019	0.0056
257	0.0000	0.000	0.0000	0.0050
	0.8564	0.853	0.0034	0.0055
313	0.0000	0.000	0.0000	0.0050
	0.2855	0.286	-0.0005	0.0054
350	0.0000	0.000	0.0000	0.0050
	0.6363	0.634	0.0023	0.0056

COPY



**Bangkok High Lab Co.,Ltd.**

4/176 Soi Ladplakao 66, Ladplakao Rd., Anusawari, Bangkhen, Bangkok 10220

Tel: (662) 971-5800

Fax: (662) 971-5300

Website: [www.bangkokhighlab.com](http://www.bangkokhighlab.com)

E-mail: [info@bangkokhighlab.com](mailto:info@bangkokhighlab.com)



**Certificate No : S2024/151**

**Page : 4/5**

### 3. Wavelength Accuracy

**Spectral slit width : 4.00 nm**

#### 3.1 CRMs: Holmium Glass Filter

**CRMs Serial Number: 10763**

**Traceability** Traceable to NIST Holmium oxide filter NIST SRM 2034, through Starna certificate report no. 113607

Filter STDs (nm) Certificate	Average Measured Value (nm)	Correction (nm)	Uncertainty ± (nm)
241.54	241.4	0.14	0.12
279.40	279.1	0.30	0.12
288.70	288.3	0.40	0.12
334.22	333.8	0.42	0.12
361.26	360.9	0.36	0.12
418.48	418.8	-0.32	0.12
453.20	453.2	0.00	0.12
460.06	460.4	-0.34	0.12
536.90	536.9	0.00	0.12
637.94	638.1	-0.16	0.12

#### 3.2 CRMs: Didymium Glass Filter

**CRMs Serial Number: 10764**

**Traceability** Traceable to NIST Didymium filter NIST SRM 2034, through Starna certificate report no. 113608

Filter STDs (nm) Certificate	Average Measured Value (nm)	Correction (nm)	Uncertainty ± (nm)
585.48	585.5	-0.02	0.12
684.63	684.8	-0.17	0.12
740.27	740.5	-0.23	0.12
748.28	748.7	-0.42	0.12
807.16	807.4	-0.24	0.12
879.70	879.6	0.10	0.12

**COPY**





**Bangkok High Lab Co.,Ltd.**

4/176 Soi Ladplakao 66, Ladplakao Rd., Anusawari, Bangkhen, Bangkok 10220

Tel: (662) 971-5800

Fax: (662) 971-5300

Website: [www.bangkokhighlab.com](http://www.bangkokhighlab.com)

E-mail: [info@bangkokhighlab.com](mailto:info@bangkokhighlab.com)



**Certificate No : S2024/151**

**Page : 5/5**

#### 4. \*Stray Light

**CRMs:** Potassium Chloride aqueous solution

**CRMs Serial Number: 14912**

**Blank Serial Number: 14958**

**Traceability** Traceable to NIST, U.S.A. potassium chloride NIST SRM2032, through Starna certificate report no.113597

**Spectral slit width : 4.00 nm**

Wavelength (nm)	Certificate	Average Measured
201.13	>2A	#N/A
201.13	<1%T	#N/A

#### 5. \*Spectral Resolution

**CRMs:** Toluene in Hexane

**CRMs Serial Number: 14812**

**Blank Serial Number: 14803**

**Traceability** Traceable to toluene in hexane NIST SRM2034,through Starna certificate report no. 113598

Spectral slit width (nm)	Abs Ratio
0.5	#N/A
1.0	#N/A
1.5	#N/A
2.0	#N/A
3.0	#N/A

**Note :** \* "Not TISI Accredited" in this certificate have been included for completeness

**Remark:** 1 Calibrate Method

1.1 Photometric and Wavelength accuracy: In-house method W-SER-001 based on ASTM E925-02 and ASTM E275-01

1.2 Stray light: Measuring the CRMs in both absorbance and transmittance unit at wavelength 201.23 nm. Base on European Pharmacopoeia V.6.19.3 1984

1.3 Spectral resolution: Measuring the CRMs. The maximum absorbance values were read at closest to 268.7nm and the minimum absorbance values were read at closest 267.0 nm. Refer to European Pharmacopoeia V.6.19.3 1984

2. N/A = not available.

3. Uncertainty of Measurement: The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95%.

4. This result of calibration was found accurate as shown on date and place of calibration only.

5. This report will certify of calibrated equipment only.

**- End of Report -**

**COPY**

**SPECTROPHOTOMETER**

**Model : UV-1900i**

**Serial No : A12535880464**



**Bangkok High Lab Co.,Ltd.**

**4/176 Soi Ladplakao 66, Ladplakao Rd., Anusawari, Bangkok, Bangkok 10220**

**Tel: (662) 971-5800**

**Fax: (662) 971-5300**

**Website: [www.bangkokhighlab.com](http://www.bangkokhighlab.com)**

**E-mail: [info@bangkokhighlab.com](mailto:info@bangkokhighlab.com)**



# CERTIFICATE OF CALIBRATION

**Certificate No : S2024/152**

**Page : 1/5**

**Order No : 210/2024**

**Customer : Eastern Thai Consulting 1992 Co., Ltd.**

**Address : 129 Moo. 1, Nonsi, Kabinburi, Prachinburi 25110**

**Instrument : UV/VIS spectrophotometer**

**Manufacture : SHIMADZU**

**Model : UV-1900i**

**Serial Number : A12535880464(ID.NoKB-LAB-63/007)**

**Environment : Temperature (25.4 - 25.3) °C**

**: Humidity (57 - 52) %RH**

**Received Date : August 19, 2024**

**Calibration Date : August 19, 2024**

**Issued Date : August 26, 2024**

**Calibrate Status : No Adjustment**

**Calibration Area : Customer area**

**Roomname : Laboratory Room of Eastern Thai Consulting 1992 Co., Ltd.**

**Calibrated By : Pannawat Pungsaard**  
( Mr. Pannawat Pungsaard )  
**Calibration Engineer**

**Approved By : [Signature]**  
( Mr.Wanchai Meesiri )  
**Manager**

**COPY**

**This calibration certificate shall not be reproduced other than in full except with the prior written approval of the Bangkok High Lab Co.,Ltd.**



Certificate No : S2024/152

Page : 2/5

## 1. Photometric Accuracy

CRMs: Neutral Density Glass Filters

CRMs Serial Number: A404

**Traceability:** Traceable to NIST, U.S.A. through Neutral density filters NIST SRM 930e & 1930, Double Aperture method through Starna certificate report no. 9113946

**Spectral slit width : 1.00 nm**

### 1.1 Reading scale at 420.0 nm

Filter STDs (Abs) Certificate	Average Measured Value (A)	Correction (A)	Uncertainty ± (A)
0.0000	0.0000	0.0000	0.0028
0.4967	0.4957	0.0010	0.0044
0.9639	0.9638	0.0001	0.0038
2.0368	2.0392	-0.0024	0.0064

### 1.2 Reading scale at 440.0 nm

Filter STDs (Abs) Certificate	Average Measured Value (A)	Correction (A)	Uncertainty ± (A)
0.0000	0.0000	0.0000	0.0028
0.4868	0.4856	0.0012	0.0040
0.9436	0.9435	0.0001	0.0040
1.9667	1.9679	-0.0012	0.0064

### 1.3 Reading scale at 465.0 nm

Filter STDs (Abs) Certificate	Average Measured Value (A)	Correction (A)	Uncertainty ± (A)
0.0000	0.0000	0.0000	0.0028
0.4532	0.4525	0.0007	0.0034
0.8778	0.8777	0.0001	0.0040
1.8421	1.8440	-0.0019	0.0060

### 1.4 Reading scale at 546.1 nm

Filter STDs (Abs) Certificate	Average Measured Value (A)	Correction (A)	Uncertainty ± (A)
0.0000	0.0000	0.0000	0.0028
0.4709	0.4707	0.0002	0.0028
0.9093	0.9088	0.0005	0.0028
1.8759	1.8762	-0.0003	0.0064

**COPY**



**Bangkok High Lab Co.,Ltd.**

4/176 Soi Ladplakao 66, Ladplakao Rd., Anusawari, Bangkok, Bangkok 10220

Tel: (662) 971-5800

Fax: (662) 971-5300

Website: [www.bangkokhighlab.com](http://www.bangkokhighlab.com)

E-mail: [info@bangkokhighlab.com](mailto:info@bangkokhighlab.com)



**Certificate No : S2024/152**

**Page : 3/5**

**1.5 Reading scale at 590.0 nm**

Filter STDs (Abs) Certificate	Average Measured Value (A)	Correction (A)	Uncertainty ± (A)
0.0000	-0.0001	0.0001	0.0028
0.4895	0.4896	-0.0001	0.0029
0.9468	0.9462	0.0006	0.0029
1.9022	1.9005	0.0017	0.0061

**1.6 Reading scale at 635.0 nm**

Filter STDs (Abs) Certificate	Average Measured Value (A)	Correction (A)	Uncertainty ± (A)
0.0000	0.0000	0.0000	0.0028
0.4639	0.4644	-0.0005	0.0030
0.8995	0.8999	-0.0004	0.0031
1.7824	1.7830	-0.0006	0.0062

**2. Photometric Accuracy**

**CRMs:** Potassium Dichromate in Perchloric acid

**CRMs Serial Number: 15086**

**Blank Serial Number: 15178**

**Traceability:** Traceable to NIST, U.S.A. through crystalline potassium dichromate NIST SRM 935a through Starna certificate report no. 107758

**Spectral slit width : 1.00 nm**

Wavelength (nm)	Certificate (Abs)	Average Measured Value (A)	Correction (A)	Uncertainty ± (A)
235	0.0000	0.0000	0.0000	0.0050
	0.7348	0.7350	-0.0002	0.0056
257	0.0000	0.0000	0.0000	0.0050
	0.8537	0.8536	0.0001	0.0055
313	0.0000	0.0002	-0.0002	0.0050
	0.2886	0.2871	0.0015	0.0054
350	0.0000	0.0003	-0.0003	0.0050
	0.6347	0.6322	0.0025	0.0056

**COPY**





**Certificate No : S2024/152**  
**Page : 4/5**

### 3. Wavelength Accuracy

**Spectral slit width : 1.00 nm**

#### 3.1 CRMs: Holmium Glass Filter

**CRMs Serial Number: W184/H**

**Traceability** Traceable to NIST Holmium oxide filter NIST SRM 2034, through Starna certificate report no. 9113944

Filter STDs (nm) Certificate	Average Measured Value (nm)	Correction (nm)	Uncertainty ± (nm)
241.70	241.4	0.30	0.12
279.44	279.4	0.04	0.12
287.71	287.6	0.11	0.12
334.02	334.0	0.02	0.12
360.89	361.0	-0.11	0.12
418.53	418.6	-0.07	0.12
453.67	453.6	0.07	0.12
459.99	460.2	-0.21	0.12
536.52	536.4	0.12	0.12
638.00	637.6	0.40	0.12

#### 3.2 CRMs: Didymium Glass Filter

**CRMs Serial Number: W184/D**

**Traceability** Traceable to NIST Holmium oxide filter NIST SRM 2034, through Starna certificate report no. 9113945

Filter STDs (nm) Certificate	Average Measured Value (nm)	Correction (nm)	Uncertainty ± (nm)
585.19	585.4	-0.21	0.12
684.50	684.6	-0.10	0.12
741.02	740.8	0.22	0.12
748.56	748.6	-0.04	0.12
807.02	807.2	-0.18	0.12
879.41	879.4	0.01	0.12

**COPY**



**Bangkok High Lab Co.,Ltd.**

4/176 Soi Ladplakao 66, Ladplakao Rd., Anusawari, Bangkok, Bangkok 10220

Tel: (662) 971-5800

Fax: (662) 971-5300

Website: [www.bangkokhighlab.com](http://www.bangkokhighlab.com)

E-mail: [info@bangkokhighlab.com](mailto:info@bangkokhighlab.com)



**Certificate No : S2024/152**

**Page : 5/5**

#### 4. \*Stray Light

**CRMs:** Potassium Chloride aqueous solution

**CRMs Serial Number: 5469**

**Blank Serial Number: 8745**

**Traceability** Traceable to NIST, U.S.A. potassium chloride NIST SRM2032, through Starna certificate report no. 107759

**Spectral slit width : 1.00 nm**

Wavelength (nm)	Certificate	Average Measured
201.05	>2A	2.0331
201.05	<1%T	0.95

#### 5. \*Spectral Resolution

**CRMs:** Toluene in Hexane

**CRMs Serial Number: 8697**

**Blank Serial Number: 8716**

**Traceability** Traceable to toluene in hexane NIST SRM2034, through Starna certificate report no. 107760

Spectral slit width (nm)	Abs Ratio
0.5	#N/A
1.0	1.8691
1.5	#N/A
2.0	#N/A
3.0	#N/A

**Note :** \* "Not TISI Accredited" in this certificate have been included for completeness

**Remark:**

Calibrate Method

1.1 Photometric and Wavelength accuracy: In-house method W-SER-001 based on ASTM E925-02 and ASTM E275-01

1.2 Stray light: Measuring the CRMs in both absorbance and transmittance unit at wavelength 201.23 nm. Base on European Pharmacopoeia V.6.19.3 1984

1.3 Spectral resolution: Measuring the CRMs. The maximum absorbance values were read at closest to 268.7nm and the minimum absorbance values were read at closest 267.0 nm. Refer to European Pharmacopoeia V.6.19.3 1984

2. N/A = not available.

3. Uncertainty of Measurement: The reported uncertainty of measurement was based on standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95%.

4. This result of calibration was found accurate as shown on date and place of calibration only.

5. This report will certify of calibrated equipment only.

**- End of Report -**

**COPY**

**Dissolved Oxygen and BOD**

**Model : HI98193**

**Serial No : 03500057101**

Certificate No. : HIT-2524-0940

Page : 1 of 2

**CERTIFICATE OF CALIBRATION**

**Equipment :** Dissolved Oxygen and BOD Meter

**Meter Model :** HI98193      **Serial No. :** 03500057101

**Probe Model :** HI764073      **Serial No. :** KC1N42MCK

**Manufacturer :** Hanna Instruments      **Made in :** Romania

**Condition As-Received :** Used Product      **Reference :** RE250987

**Ambient Temperature :**  $(25 \pm 2)^{\circ}\text{C}$       **Relative Humidity :**  $(50 \pm 15)\% \text{ RH}$

**Customer name :** Eastern Thai Consulting 1992 Co., Ltd.  
129 Moo. 1, Nonsi Kabinburi, Kabinburi,  
Prachinburi 25110

**Received date :** 29 May 2025

**Calibrate date :** 11 June 2025

**Issue date :** 11 June 2025

**Calibrated Location :** Hanna Instruments (Thailand) Ltd.

**Calibration Procedure :** This calibrator was conducted by using in-house: calibration procedure  
CP-11 by using certified reference material (CRM).

**Calibrated by :** ☒ Mr. Pichit Petthong**Approved by :** 

Mr. Anan Suwanchaisakul

Authorized Signatory



This certificate was certified only for the instrument we calibrated.

This result of calibration was found accurate on date and place of calibration only.

\*\* This certificate may not be reproduced other than in full, except with the prior written \*\*

approval of the head of Hanna Instrument (Thailand)



**Condition of this calibration result**

1. Reference Standard Instruments : This certification is traceable to the international unit of thru Technology Promotion Association (Thailand-Japan).

Instruments	Model	Serial No.	Certificate No.
Thermometer with sensor	HI98509	39643D	24T1281
Digital Thermo-Hygrometer	HT-771SD	AI.07155	25H171

2. Reference Standard Materials : DO calibration standard traceable to Hanna Instrument Ltd.

Buffer Solution	Manufacture	Certified Value	Lot Number	Exp. date
Zero Oxygen Solution	Hanna	$0.0 \pm 0.1$ @25°C	S0028/23	March 2028

**Calibration Result**

Inspection the accuracy of the Dissolved Oxygen (DO) Meter by using the following certificate reference material value.

Unit Under Calibration	CRM Standard DO	Actual value Reading	Error value Reading	Uncertainty of Measurement ( ± )
DO Electrode S/N KC1N42MCK	0.0 mg/L	0.00 mg/L	0.00 mg/L	N/A
	8.3 mg/L	8.26 mg/L	-0.04 mg/L	0.33 mg/L

The report uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95%

**\*\* End of certificate \*\***

**COPY**



**Inductively Coupled Plasma Spectrometer**

**Model : AVIO 550**

**Serial No : M81S2210101**

# Avio550 Preventive Maintenance Report

Company Name: Eastern Thai Consulting 1992 Ci.,Ltd.

Instrument Location: 683 Moo 11 Sukapibal 8 Rd. Nong Kham,  
Si Racha, Chonburi 20230


Instrument Serial No.: M81S2210101

Date: 25-Oct-2024

**COPY**

## ICP-OES/Avio550 Preventive Maintenance (PM)

<b>Company Name:</b>	Eastern Thai Consulting 1992 Ci.,Ltd.		
<b>Address (Instrument Location):</b>	683 Moo 11 Sukapibal 8 Rd. Nong Kham,Si Racha, Chonburi 20230		
<b>Serial Number:</b>	M81S2210101	<b>PM Number:</b>	2 of 2
<b>Customer Name (if applicable):</b>		<b>Telephone Number:</b>	
<b>Service Engineer Name:</b>	Khwanchai	<b>Service Order Number:</b>	WO-02963150
<b>Date PM Performed: (DD-MMM-YYYY)</b>	25-Oct-2024	<b>Next PM Due Date: (DD-MMM-YYYY)</b>	25-Apr-2025
<b>Standard Labor Hours to Complete PM :</b>		<b>4 hours</b>	

Part Number	Release	Publication Date	
TH09370188 Rev.2	B	July 2020	

### Scope

The purpose of this PM is to ensure the continued functionality of the PerkinElmer / Avio550 by inspecting and replacing any worn or damaged parts. This service should only be performed by a trained representative of PerkinElmer. The customer should save their method before the PM begins.

### General Instructions:

The customer must provide the engineer operational data to demonstrate recent instrument performance prior to starting the PM.

Always check with the customer before making any changes that may affect the customer's analysis or calibration, including a current back-up of system software and/or data files. The completed document should be signed by an authorized PerkinElmer and customer representative and left with the customer. Update the PM sticker and instrument logbook as required.

### Copyright Information

This document contains proprietary information that is protected by copyright. All rights are reserved. No part of this publication may be reproduced in any form whatsoever or translated into any language without the prior, written permission of PerkinElmer, Inc. **Copyright © 2013 PerkinElmer, Inc.**

### Trademarks

Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are protected by law. PerkinElmer is a registered trademark of PerkinElmer, Inc. All other trademarks and registered trademarks not owned by PerkinElmer, Inc. or its subsidiaries that are depicted herein are the property of their respective owners.

**Except as specifically set forth in its terms and conditions of sale, PerkinElmer makes no Warranty of any kind with regard to this document, including, but not limited to, the implied warranties of merchantability and fitness for a particular purpose.**

PerkinElmer shall not be liable for incidental or consequential damages in connection with the furnishing or use of this document.

**COPY**

## Component List

Component / Specific Model	Serial #	Configuration Notes
NA	NA	NA

## Parts Lists

Parts Included with the PM		
Part Number (if applicable)	Description	Quantity
09995098	Air Fiter-Spectrometer	N/A
N077520	Air Fiter-RF Generetor	N/A
09992731	Axial Window	N/A
B0810377	Radial Window	N/A
N0770438	O-ring kit, injector support adapter	N/A
N0780437	O-ring kit, torch	N/A

Additional Reagents and Standards Required for PM				
Part Number (if applicable)	Description	Quality	Batch/Lot #	Expired Date (MM/YY)
N0691579	Muti-Element Standard	AR	61-176CRX1	06/2025
N9300221	DL Standard diluted 100 X	AR	59-091CRY1	11/2024
N0582152	Wave Cal Solution	AR	61-023CRX1	02/2025
N9302946	VIS Wavecal Solution	AR	58-145CRT1	04/2025

**COPY**

# Procedure Checklist

Use (✓) to check off those steps in the checklist that have been completed.

## 1. General:

- ☒ Ask customer about unit's performance since last visit.
- ☒ Check incoming AC line voltage under load for proper levels and grounding.
- ☒ Is the instrument operational? If not, please comment.

## 2. Mechanical:

- ☒ Inspect and clean all fans and filters.
- ☒ Inspect and replace torch components and necessary.

Torch Components Replaced: ☐Yes ☒No

- ☐ Inspect all tubing for signs of cracking or leaking and replace as necessary.

Tubing Replaced: ☐Yes ☒No

- ☒ Inspect the peristaltic pump for proper operation.
- ☒ Check and adjust if necessary, the external nitrogen, argon shear gas and water supply pressures.
- ☒ Check and adjust if necessary, the internal nitrogen, main argon, torch argon and shear gas pressures.

Regulator	Measured Pressure	Set Pressure
Nitrogen	NA	NA (calibrated in Factory)
Main Argon	76	76 psig
Torch Argon	67	67 psig
Shear Gas	65	65 psig
Water	35	35 psig

- ☒ Check shear gas nozzle for blockages and proper, uniform flow.
- ☒ Inspect nitrogen Hi/Low purge and shear gas solenoids for proper function.
  - ☒ Inspect the function of all spectrometer motors. Drive the motors from the Spectrometer DCM. (slits, XY motor)
- ☒ Inspect the function of the pneumatic shutter for proper operation.
  - ☒ Perform preventative maintenance on the chiller as required. Make the customer aware of the importance of maintaining the chiller fluid level and filter replacement.
- ☒ Drain air compressor surge tank.
- ☒ Clean exterior of instrument.
- ☒ Visually inspect all PC boards for cleanliness and signs of corrosion.

**COPY**



### 3. Electronical

- ☒ Check all RF generator and spectrometer power supply voltages.
- ☒ Run instrument diagnostic checks from the appropriate Device Control Module.

#### **RF Generator:**

- ☒ Check the RF generator status screens.
- ☒ Check the function of all interlocks.

#### **Spectrometer:**

- ☒ Check the spectrometer status screens. Ensure Ready mode with no fatal errors.
- ☒ Check the spectrometer optical tub temperatures (top, bottom, fin, optical base).
- ☒ Check detector temperatures.
- ☒ Check TEC voltages (6.5VDC)

### 4. Optical:

- ☒ Clean or replace the axial and radial view windows as necessary.
  - Axial Window Replaced: ☐ Yes ☒ No
  - Radial Window Replaced: ☐ Yes ☒ No

### 5. PM Performance Tests:

- ☒ Perform View Align.

#### **Test Spectral Resolution:**

- ☒ Measure the spectrometers ability to separate two adjacent wavelengths.

Parameter	Specification	Test Result	Pass/Fail
As 193.696 - Resolution	$\leq 0.007$	0.00528	Passed
Ni 231.604 - Resolution	$\leq 0.008$	0.00724	Passed
Ni 341.476 - Resolution	$\leq 0.012$	0.00911	Passed
La 408.672 - Resolution	$\leq 0.020$	0.01596	Passed
Ba 455.403 - Resolution	$\leq 0.025$	0.02165	Passed

#### **Test Precision:**

- ☒ Test for reproducibility of a set of measurement.

Parameter	Specification	Test Result	Pass/Fail
As 193.696	%RSD $\leq 1\%$	0.26	Passed
Zn 213.856	%RSD $\leq 1\%$	0.21	Passed
Mn 257.610	%RSD $\leq 1\%$	0.20	Passed
La 379.478	%RSD $\leq 1\%$	0.21	Passed
Ba 455.403	%RSD $\leq 1\%$	0.21	Passed
Ba 493.408	%RSD $\leq 1\%$	0.19	Passed

**COPY**

- ☒ Run an Axial & Radial BEC according to the A&T spec.

**Test Axial BEC Cd:**

Method "BEC-XL" For Samples "IB (2%HNO3)" and "IS (N930-0221/100)", record intensities.

Calculated BEC:  $BEC = (IB * \text{Conc of Std}) / (IS - IB)$ . Where Conc of Std = 500 PPB

Element	Conc.	IB	IS	
<b>Cd 226</b>	500	523.1	223029.5	
IB*Conc	IS-IB	BEC	Spec	Pass/Fail
261550	222506.4	1.18	<150 PPB	Passed

**Test Radial BEC Mn:**

Method "BEC-RL" For Samples "IB (2%HNO3)" and "IS (N069-1579)", record intensities.

Calculated BEC:  $BEC = (IB * \text{Conc of Std}) / (IS - IB)$ . Where Conc of Std = 1,000 PPB

Element	Conc.	IB	IS	
<b>Mn 257</b>	1,000	586.9	253416.6	
IB*Conc	IS-IB	BEC	Spec	Pass/Fail
586900	252829.7	2.32	<45 PPB	Passed

**6. Review:**

- ☒ Review with the customer PM work performed.
- ☒ Discuss recommended customer supplied materials to have on hand.
- ☒ Attach PM sticker.

**COPY**

## Additional Comments

### Additional Comments Regarding the PM

None

## Review

*The preventive maintenance checks and if applicable performance tests for ICP-OES/Avio550 have been completed.*

*This ICP-OES/Avio550   Passes ☒   Fails ☐ the preventive maintenance.*

### Review of Preventive Maintenance:

Authorized PerkinElmer Representative:

KL S.

Date:

25-Oct-2024

(DD-MMM-YYYY)

Authorized Customer Representative:

Date:

25-Oct-2024

(DD-MMM-YYYY)

**COPY**

ใบรับรองการสอบเทียบเครื่องมือ

ปล่อยระบายอากาศ

Envi Equipment Service Co., Ltd.  
110/254 Moo 3, Tumbon Bang Rak Phatthana, Amphur Bang Bua Thong, Nonthaburi 11110  
Tel. 098 362 9152, 089 478 7885  
E-mail: sales@envi-ees.com

Certificate No.: E25-010015  
Page.: 1 of 6

## CERTIFICATE OF CALIBRATION

**Customer** : M E T COMPANY LIMITED  
**Address** : 36/659 Moo 6, Bangrakpatthana, Bangbuathong, Nonthaburi 11110  
**Description of Equipment** : Console meter  
**Manufacturer** : Apex Instrument  
**Model Number** : TMC-572-V  
**Serial Number** : A2202103  
**ID./Control No.** : -  
**Environment Conditions** : Temperature (25 ± 2) °C  
: Humidity (50 ± 15) % RH  
**Cal. Date** : 21/01/2025  
**Issue Date** : 21/01/2025

### Calibration Method or Calibration Procedure Used

US EPA Method (United State Environmental Protection Agency)

This certificate is traceable to national standard, which realize the units of measurement according to the International System of Units (IS).

### Result of Calibration

This certificate may not be reproduced other than in full except with prior Written approval of the Technical Manager, Envi Equipment Service Company Limited.

These reported uncertainties of measurement are expanded by a coverage factor of k=2, providing a 95% confidence level

Calibrated by : Mr. Sanya Sangnil

Approved by :



(Mr. Mana Fucksun)  
Technical Manger

Certificate No. : E25-010015  
Page : 2 of 6

## METHOD 5 CONSOLE CALIBRATION USING REFERENCE WET GAS METER W-NK-2.5-B-Z No.547425 5-POINT METRIC UNIT

Meter Console Information		Calibration Conditions				Factors/Conversions		
Console Model Number	TMC-572-V	Date	Time	21/01/2025	09:50 AM	Std Temp	293	K
Console Serial Number	A2202103	Calibration Reference No.	SER25-010006			Std Press	760	mm Hg
DGM Model Number	SK25EX	Barometric Pressure	759.74		mmHg	K <sub>1</sub>	0.386	
DGM Serial Number	00006345	Calibration Meter Gamma	1.001			Console Leak Check	PASS	

Calibration Data									
Run Time	Metering Console					Calibration Meter			
Elapsed	DGM Orifice DH	Volume Initial	Volume Final	Outlet Temp Initial	Outlet Temp Final	Volume Initial	Volume Final	Outlet Temp Initial	Outlet Temp Final
(Q)	(P <sub>m</sub> )	(V <sub>mi</sub> )	(V <sub>mf</sub> )	(t <sub>mi</sub> )	(t <sub>mf</sub> )	(V <sub>wi</sub> )	(V <sub>wf</sub> )	(t <sub>wi</sub> )	(t <sub>wf</sub> )
min	mm H <sub>2</sub> O	m <sup>3</sup>	m <sup>3</sup>	°C	°C	m <sup>3</sup>	m <sup>3</sup>	°C	°C
11.87	13.0	530.4710	530.6110	24	24	273.36358	273.50440	25	25
11.82	13.0	530.6110	530.7510	24	24	273.50440	273.64412	25	25
8.45	26.0	530.7590	530.8990	25	25	273.65214	273.79212	25	25
8.45	26.0	530.8990	531.0390	25	25	273.79212	273.93118	25	25
13.67	40.0	531.0480	531.3280	25	25	273.94070	274.21964	25	25
13.60	40.0	531.3280	531.6080	25	25	274.21964	274.49774	24	24
10.35	70.0	531.6220	531.9020	26	26	274.51106	274.78932	24	24
10.33	70.0	531.9020	532.1820	26	26	274.78932	275.06664	24	24
9.15	90.0	532.1950	532.4750	26	26	275.07944	275.35590	24	24
9.13	90.0	532.4750	532.7550	26	26	275.35590	275.63104	24	24




**METHOD 5 CONSOLE CALIBRATION  
USING REFERENCE WET GAS METER W-NK-2.5-B-Z No.547425  
5-POINT METRIC UNIT**

Meter Console Information		Calibration Conditions				Factors/Conversions		
Console Model Number	TMC-572-V	Date	Time	21/01/2025	09:50 AM	Std Temp	293	K
Console Serial Number	A2202103	Calibration Reference No.	SER25-010006			Std Press	760	mm Hg
DGM Model Number	SK25EX	Barometric Pressure	759.74			K <sub>1</sub>	0.386	
DGM Serial Number	00006345	Calibration Meter Gamma	1.001			Console Leak Check	PASS	

Calibration Data Results									
Standardized Data				Dry Gas Meter					
Dry Gas Meter		Calibration Meter		Calibration Factor		Flowrate		Variation	
(V <sub>m(std)</sub> )	(Q <sub>m(std)</sub> )	(V <sub>w(std)</sub> )	(Q <sub>w(std)</sub> )	Value (Y)	Variation (ΔY)	Std & Corr (Q <sub>m(std)(corr)</sub> )	.0212 m <sup>3</sup> <sub>std</sub> /min (ΔH <sub>@</sub> )	Variation (ΔH <sub>@</sub> )	
m <sup>3</sup>	m <sup>3</sup> /min	m <sup>3</sup>	m <sup>3</sup> /min			m <sup>3</sup> /min	mm H <sub>2</sub> O		
0.138	0.012	0.139	0.012	1.006	0.015	0.012	42.373	-1.896	
0.138	0.012	0.137	0.012	0.998	0.007	0.012	42.680	-1.588	
0.138	0.016	0.138	0.016	0.998	0.008	0.016	43.597	-0.671	
0.138	0.016	0.137	0.016	0.992	0.001	0.016	44.176	-0.092	
0.276	0.020	0.274	0.020	0.993	0.003	0.020	44.304	0.036	
0.277	0.020	0.275	0.020	0.990	0.000	0.020	43.990	-0.278	
0.278	0.027	0.275	0.027	0.988	-0.002	0.027	44.792	0.524	
0.278	0.027	0.274	0.026	0.985	-0.006	0.026	44.951	0.683	
0.279	0.030	0.273	0.030	0.980	-0.011	0.030	45.773	1.505	
0.279	0.030	0.272	0.030	0.975	-0.015	0.030	46.046	1.777	
				0.990	Y Average		44.268	ΔH@ Average	


**Note:** For Calibration Factor Y, the ratio of the reading of the calibration meter to the dry gas meter, acceptable tolerance of individual values from the average is ±0.02.

For ΔH<sub>@</sub>, orifice pressure differential that equates to 0.75 cfm (0.0212 m<sup>3</sup>/min) at standard temperature and pressure, acceptable tolerance of individual values from the average is ±0.2 inches (5.1mm) H<sub>2</sub>O.



Meter Console Information		Calibration Conditions				Factors/Conversions		
Console Model Number	TMC-572-V	Date	Time	21/01/2025	09:50 AM	Std Temp	293	K
Console Serial Number	A2202103	Calibration Reference No.	SER25-010006			Std Press	760	mm Hg
DGM Model Number	SK25EX	Barometric Pressure	759.74			K <sub>1</sub>	0.386	
DGM Serial Number	00006345	Calibration Meter Gamma	1.001			Console Leak Check	PASS	

Calibration Date: 21-1-2025 Calibration Reference No: SER25-010006



Console Serial: A2202103

Console Model: TMC-572-V





Meter Console Information	
Console Model Number	TMC-572-V
Console Serial Number	A2202103
DGM Model Number	SK25EX
DGM Serial Number	00006345


Calibration Conditions			
Date	Time	21/01/2025	09:50 AM
Calibration Reference No.	SER25-010006		
Barometric Pressure	759.74	mmHg	
Calibration Meter Gamma	1.001		

Factors/Conversions		
Std Temp	293	K
Std Press	760	mm Hg
K <sub>1</sub>	0.386	
Console Leak Check	PASS	

Calibration Date: 21-1-2025

Calibration Reference No: SER25-010006

Meter Pressure vs Flowrate



Console Serial:

Console Model: TMC-572-V



## THERMOCOUPLES SYSTEM CALIBRATION

Sampling System Equipment Information	
Console Model Number	TMC-572-V
Console Serial Number	A2202103
DGM Model Number	SK25EX
DGM Serial Number	00006345
Meter Box Model Number	JENCO 765 KF
Meter Box Serial Number	JC 18920

Calibration Conditions			
Date	Time	21/01/2025	00:00 PM
Calibration Reference No.	SER25-010006		
Reference Thermometer	DIGICON		
Serial Number	183169105		

Results											
Console Thermocouple Simulator											
Channel and test point	Meter Box Channel Temperature Reading (°C)										
	-18.0	25.0	38.0	93.0	149.0	260.0	371.0	482.0	593.0	816.0	1038.0
Stack	-17.0	24.0	37.0	92.0	148.0	258.0	371.0	482.0	593.0	816.0	1039.0
Aux	-17.0	24.0	37.0	92.0	148.0						
Probe	-17.0	24.0	37.0	92.0	148.0						
Filter	-17.0	24.0	37.0	92.0	148.0						
Exit	-17.0	24.0	37.0								

Stack ± 1.50% Absolute  
 Probe ± 3.0 °C  
 Filter ± 3.0 °C

### Tolerance Range

Meter ± 3.0 °C  
 Exit ± 2.0 °C



Envi Equipment Service Co., Ltd.

110/254 Moo 3, Tumbon Bang Rak Phatthana, Amphur Bang Bua Thong, Nonthaburi 11110

Tel. 098 362 9152, 089 478 7885

E-mail: sales@envi-ees.com

Certificate No. : E25-010016

Page. : 1 of 3

## CERTIFICATE OF CALIBRATION

**Customer** : M E T COMPANY LIMITED

**Address** : 36/659 Moo 6, Bangrakpatthana, Bangbuathong, Nonthaburi 11110

**Description of Equipment** : Standard Probe Method 5

**Manufacturer** : Apex Instrument

**Model Number** : PS-3HV

**Serial Number** : -

**ID./Control No.** : -

**Environment Conditions** : Temperature (25 ± 2) °C  
Humidity (50 ± 15) % RH

**Cal. Date** : 21/01/2025

**Issue Date** : 21/01/2025

### Calibration Method or Calibration Procedure Used

US EPA Method (United State Environmental Protection Agency)

This certificate is traceable to national standard, which realize the units of measurement according to the International System of Units (IS).

### Result of Calibration

This certificate may not be reproduced other than in full except with prior Written approval of the Technical Manager, Envi Equipment Service Company Limited.

These reported uncertainties of measurement are expanded by a coverage factor of k=2, providing a 95% confidence level

Calibrated by : Mr. Sanya Sangnil

Approved by :

(Mr. Mana Fuekha)

Technical Manger



Certificate No. : E25-010016

Page.: 2 of 3

## CALIBRATION RESULTS

### S-Type Geometric Pitot Tube Calibration

#### Sampling System Equipment Information

**Probe Model** : PS-3HV

**Probe Number** : -

**Pitot Number** : A10664

**Pitot Tube Type** : S-type

#### Calibration Condition

**Date** : 21 January 2025

**Barometric Pressure** : 759.74 mm Hg


**Digital Caliper** : CD-6" ASX

**Serial number** : A18008059

Pitot tube/Probe: # PS-3HV			
Parameter	Value	Allowable Range	Check
Assembly level?	Yes	Yes	Pass
Ports Damage?	No	No	Pass
α1	0	-10° < α1 < +10°	Pass
α2	1	-10° < α2 < +10°	Pass
β1	0	-5° < β1 < +5°	Pass
β2	0	-5° < β2 < +5°	Pass
γ	0	N/A	-
θ	0	N/A	-
Dt	0.375	.188" to .375"	Pass
A	0.8585	2.1Dt ≤ A ≤ 3Dt	Pass
A/2Dt	1.144	1.05 ≤ A/Dt ≤ 1.5	Pass
Z = A tan γ	0.045	Z ≤ .125"	Pass
W = A tan θ	0.018	W ≤ .031"	Pass

#### Remark:

I certified that probe model: **PS-3HV** and **Pitot tube no. A10664** meets or exceeds all specifications, criteria and/or applicable design and is hereby assigned a pitot tube certification factor of 0.84. See 40 CFR Pt. 60, Appendix A-1, EPA Method 2.



## THERMOCOUPLES SYSTEM CALIBRATION

Sampling System Equipment Information	
Probe Model Number	PS-3HV
Probe Serial Number	-
Meter Box Model Number	JENCO 765 KF
Meter Box Serial Number	JC 18920

Calibration Conditions			
Date	Time	21/01/2025	01:00 PM
Calibration Reference No.	E25-010006		
Reference Thermometer	DIGICON		
Serial Number	183169105		

Thermocouple of Standard Probe method 5 = length 3 foot			
Set Point	Reference Thermocouple	Probe Thermocouple	Difference
100	100.0	99.0	0.27
250	250.0	249.0	0.19
300	300.0	298.0	0.35
350	350.0	348.0	0.32



## CERTIFICATE OF CALIBRATION

Customer : M E T COMPANY LIMITED  
Address : 36/659 Moo 6, Bangrakpatthana, Bangbua Thong, Nonthaburi 11110  
Description of Equipment : Stainless Steel Nozzle  
Manufacturer : Apex Instrument  
Model Number : NS-SET  
Serial Number : -  
ID./Control No. : -  
Environment Conditions : Temperature (25 ± 2) °C  
Humidity (50 ± 15) % RH  
Cal. Date : 21/01/2025  
Issue Date : 21/01/2025

### Calibration Method or Calibration Procedure Used

US EPA Method (United State Environmental Protection Agency)


This certificate is traceable to national standard, which realize the units of measurement according to the International System of Units (IS).

### Result of Calibration

This certificate may not be reproduced other than in full except with prior Written approval of the Technical Manager, Envi Equipment Service Company Limited.

These reported uncertainties of measurement are expanded by a coverage factor of k=2, providing a 95% confidence level

Calibrated by : Mr. Sanya Sangnil

Approved by :   
(Mr. Mana Fuchai)  
Technical Manger



## CALIBRATION RESULTS

### Sampling System Equipment Information

Nozzle Model : NS-SET  
Nozzle Number : -  
Nozzle Type : Stainless Steel

### Calibration Condition

Date : 21 January 2025  
Barometric Pressure : 759.74 mm Hg  
Calibration Device : Vernier, 0-150 mm  
Method Reference : US, EPA Method


Nozzle ID	Nozzle Diameter				Different	(D1 + D2 + D3) / 3
Size	mm	D1	D2	D3	$\Delta D$	Davg
		mm	mm	mm	mm	mm
NS-4	3.18	2.91	2.91	2.91	0.000	2.910
NS-5	3.97	4.01	4.03	4.03	0.012	4.023
NS-7	5.56	5.42	5.43	5.42	0.006	5.423
NS-8	6.35	6.06	6.06	6.08	0.012	6.067
NS-9	7.14	7.21	7.21	7.21	0.000	7.210
NS-12	9.53	9.15	9.16	9.17	0.010	9.160
NS-16	12.70	12.15	12.13	12.14	0.010	12.140

### Remark:

D1, D2, D3 = There difference nozzle diameters, mm; diameter must be within 0.025 mm

$\Delta D$  = Maximum difference between any two diameters, must be  $\leq 0.100$  mm

Davg =  $(D_1 + D_2 + D_3) / 3$





## Certificate of Calibration

Certificate No. : 67-400505-2

Page : 1 of 2

Submitted by : M E T Company Limited

36/659 Moo 6, T.Bangrakpattana, A.Bangbuatong, Nonthaburi 11110

Equipment : Temperature controlled enclosure (Oven)

Manufacturer : Binder

Model : ED53

Range : N/A °C

Resolution : 1 °C

Serial No. : 13-07419

ID No. : MET-OV02/57

Environment : On site calibration was carried out at the Laboratory, M E T Company Limited

Ambient Temperature : (30.0 to 31.0) °C

Relative Humidity : (50 to 55) %

Line Voltage : (210.0 to 210.8) V

Date of Received : 20 August 2024

Date of Calibration : 20 August 2024

Date of Issue : 21 August 2024

Calibrated by : Permpon Chanpu

Calibration Method : CAL-M4004, TLAS G-20

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Digital Thermometer with Thermocouple probe

ID No.

Cert. No.

Due Date

Traceability

400029 &amp; 400030 67-400246-1 25 Oct 2024

National Institute of Metrology Thailand (NIMT)

Approved by :

( Permpon Chanpu )

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



## Certificate of Calibration

Certificate No. : 67-400505-2


Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber

W = 0.40 m

D = 0.33 m

H = 0.40 m

Capacity = 0.05 m<sup>3</sup>

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (± °C)
			1	2	3	4	5	6	7	8	9	
104	110	110	103.4	105.0	104.8	105.0	104.1	103.8	104.2	104.4	104.2	0.96
180	184	184	179.3	182.0	180.1	180.6	180.1	180.4	180.0	180.7	179.9	1.3

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Uniformity (°C)	Measured Stability (°C)	Overall Variation (°C)
104	110	110	1.1	0.3	2.0
180	184	184	2.5	0.4	3.3

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2 , providing a level of confidence of approximately 95%

- o0o -





**Certificate No:** G 670680

**Date of issue :** 27-Sep-24

**Instrument description :** Flue Gas Analyzer  
**Instrument model :** Testo 350 New  
**Instrument serial no. :** 02685557/311  
**Control unit serial no. :** 02669071/311  
**ID no. or control no. :** -  
**Manufacturer :** Testo SE & Co. KGaA  
**Probe description :** -  
**Probe model :** -  
**Probe serial no. :** -  
**Customer name :** TOPS-LAB CONSULTANTS CO.,LTD.  
**Customer address :** 189 Moo. 3 Bangrakphatthana Bangbuathong Nonthaburi 11110

**Total pages of certificate :** 2 Pages  
**Receiving no. :** L-243691  
**Receiving date. :** 24-Sep-24  
**Parameter of calibration :** Gas Calibration (Carbon dioxide 40.1 %Vol)

**Condition of UUC. :** Used  
**Ambient condition :** All of the Measurment ware caried out the stabilized labotary  
 Temperature : 23 ±5 °C  
 Humidity : 55 ± 15 %RH

**Calibration place :** 17/121 Soi Ngamwongwan 47 Yaek 48, Toongsonghong, Laksi, Bangkok 10210

**Calibration procedure no. :** This instrument was calibrated by comparison with Standard gas mixture according to calibration Work Instruction no. WI-CL-19-C

*The calibration certificate expanded uncertainty of measurement is stated as the standard uncertainty of measurent Multiplied by coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.*

*This certificate is applied only to item under test Environmental condition.*

*This Calibration Certificate may not be reporduced other than in full except with the permission of the issuing laboratory. Calibration certificates without signature and seal not valid and The results relate only to the items tested/calibrated.*

*This calibration certificate documents are tracebility to national standards, which realize measurement according to the International System of Units (SI).*

**Date of calibration :** 26-Sep-24



Mr. Kwanchai Khamdoun  
Calibration Technician



Mrs. Nongluck Wongsettee  
Technical Manager



Certificate No.: G 670680

Standard References (Table 1)

Standard	Certificate No.	Vendor	Due date
Carbon dioxide ( CO2 ) 40.1 %Vol	0204/23	Linde	29-Jan-27

Measured room conditions

Temperature : 22.7 °C Humidity : 65.8 %RH Pressure : 1010.5 mbar

Calibration conditions

Gas Temperature : 23 °C Flow rate : 1,200 ml/min Gas pressure : 1013.8 mbar

Calibration Results (Without adjustment) (Table 2)

Parameter of Standard	Standard Values	Mean of UUC	Error	Uncertainty ( ± )
CO2 (%Vol)	40.10	40.24	0.14	0.41

Remark : 1 cmol/mol = 1 %vol.

End of Report



Testing Report

General Information

Measuring Site : TOPS-LAB CONSULTANTS CO.,LTD.

Address : 189 Moo. 3 Bangrakphatthana Bangbuathong Nonthaburi 11110

Parameter : Gas

Report No. : R 670004

Date of Issue : 07-Jun-24

Date of testing : 06-Jun-24

Customer Devices Information

Instrument Description : Flue Gas Analyzer

Instrument Model : Testo 350 New

Instrument S/N : 02685557/311

ID no. or Control no. : -

Manufacturer : Testo SE & Co. KGaA

Receiving no : L-241705

Standard Reference

Standard	Reference No.	Traceability	Due Date
Carbon Monoxide ( CO ) 80.18 ppm	CG-0002-24	Nimt	11-Jan-29
Carbon Monoxide ( CO ) 302 ppm	1915/23	Linde	16-Jun-25
Carbon Monoxide ( CO ) 1001 ppm	CG-0085-24	Nimt	22-May-29

Results : The testing results were reported in the table below

No.	Gas	Standard Gas	UUT Mean	Error
1	CO	80.18	79	-1.18
2	CO	302	299	-3
3	CO	1001	998	-3



Tested by  
Mr.Kwanchai Khamdoun



Approved by  
Mrs. Nongluck Wongsettee





## Calibration & Test Certificate

### FOR

Equipment Name : VOC Meter

Received Date : Mar 20, 2025

Manufacturer : E Instruments

Calibration Date : Mar 24, 2025

Model : AQ VOC-V2

Recommended Due Date : N/A

Serial Number : 14656

Location of Calibration : In Lab

Customer Name : M E T CO., LTD.

36/659 Moo 6 Bang Rak Phatthana, Bang Bua Thong,  
Nonthaburi 11110

Parameter of Calibration : Gas Calibration (Nitrogen (Zero) ppm),  
Iso Butylene 103.13 ppm

### CONDITION AS RECIVED : Normal

Environmental Conditions

Ambient Temperature : (  $25 \pm 2$  ) °C

Relative Humidity : (  $50 \pm 15$  ) %RH

Result : Adjustment (See data attached in page 2 to the end of certificate)

**ENVIR SERVICE CO., LTD.**

1. The report applies to the item and shall not be reproduced except in full, without written approval by Calibration Laboratory, Envir Service Co., Ltd.
2. The working standard is indicated in page 2 of this certificate.
3. This result of this report only to the item calibrated.

Date of Calibration : Mar 24, 2025

Valid to : Mar 23, 2026

Calibrated By : 

Kittisak Jansangwattana  
(Technician)



บริษัท เอ็นไวร์ เซอร์วิส จำกัด  
ENVIR SERVICE CO., LTD.

Approved By : 

Pasagorn Samol  
(Technician Manager)

Certificate No. EN-250323-01

We hereby certify that the instrument under mentioned has been certainly calibrated according to our calibration standard and the testing result in the calibration procedure has been good enough within the tolerance regulated in our specification.

#### Reference Standard

Standard Gas Concentrations	Reference No.	Due Date
Nitrogen (Zero) ppm	8241	20-Apr-27
Iso Butylene 103.13 ppm	Lot#GAP-248-100-2	10-Apr-25

#### Result of Calibration

##### Calibration Results (Before Adjustment)

Parameter of Standard	ZERO			SPAN			Drift%
	STD	UUC	Drift	STD	UUC	Drift	
Nitrogen (Zero)	0.0	0.2	-0.2	0.0	0.6	-0.60	-
Iso Butylen(VOC) 103.13 ppm	0.0	0.3	-0.3	103.13	98.8	-4.33	4.20

##### Calibration Results (After Adjustment)

Parameter of Standard	ZERO			SPAN			Drift%
	STD	UUC	Drift	STD	UUC	Drift	
Nitrogen (Zero)	0.0	0.2	-0.2	0.0	0.0	0.00	0.00
Iso Butylen(VOC) 103.13 ppm	0.0	0.3	-0.3	103.13	103.1	-0.03	0.03

STD = Standard

UUC = Unit Under Calibration



## Certificate of Calibration

Certificate No. : 68-200092-2

Page : 1 of 2

Submitted by : M E T Company Limited

36/659 Moo 6, T.Bangrakpattana, A.Bangbuatong, Nonthaburi 11110

Equipment : Electronic Balance

Manufacturer : AND

Model : FX-2000i

Serial No. : 15639789

ID No. : MET-EB03/61

Capacity : 2200 g

Resolution : 0.01 g

Environment : On site calibration was carried out at the Laboratory, M E T Company Limited

Ambient Temperature : (26.0 to 26.2) °C

Relative Humidity : (48.4 to 51.4) %

Air Pressure : 1012.0 mbar

Date of Received : 10 March 2025

Date of Calibration : 10 March 2025

Date of Issue : 11 March 2025

Calibrated by : Akaradath Thippichai

Calibration Method : In-house method CAL-M2001 based on UKAS Publication ref : LAB 14

Edition 7 - November 2022

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Weights

ID No.	Cert. No.	Due Date	Traceability
F181-F1821	67-210457-1	12 Jun 2025	National Institute of Metrology (Thailand), (NIMT)

Approved by :

  
( Satja Sangkhum )

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



## Certificate of Calibration

Certificate No. : 68-200092-2

Page : 2 of 2

Result of Calibration : After Adjustment

UUC Condition As-Received : Good

Departure of indication from nominal value


Nominal Value (g)	Correction (g)	Uncertainty ± (g)	Error before Adjustment (g)
200	0.00	0.011	0.00
500	0.00	0.011	0.00
600	0.00	0.011	0.00
700	0.00	0.011	0.00
800	0.00	0.011	0.00
1000	0.01	0.011	0.00
1200	0.01	0.012	0.01
1500	0.01	0.012	0.02
2000	0.00	0.014	0.03
2200	0.00	0.023	0.06

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor  $k = 2.00$ , providing a level of confidence of approximately 95%

Eccentric error Load test : 500 g

A	B	C	D	E	g
0.00	-0.01	-0.01	0.00	0.00	



Repeatability Load test : 2000 g

Stdev. : 0.000 g

- o o o -







ใบรับรองการสอบเทียบเครื่องมือ

คุณภาพอากาศในบรรยากาศ

## Certificate of Calibration

Certificate No. : 68-200092-2

Page : 1 of 2

Submitted by : M E T Company Limited

36/659 Moo 6, T.Bangrakpattana, A.Bangbuatong, Nonthaburi 11110

Equipment : Electronic Balance

Manufacturer : AND

Model : FX-2000i

Serial No. : 15639789

ID No. : MET-EB03/61

Capacity : 2200 g

Resolution : 0.01 g

Environment : On site calibration was carried out at the Laboratory, M E T Company Limited

Ambient Temperature : (26.0 to 26.2) °C

Relative Humidity : (48.4 to 51.4) %

Air Pressure : 1012.0 mbar

Date of Received : 10 March 2025

Date of Calibration : 10 March 2025

Date of Issue : 11 March 2025

Calibrated by : Akaradath Thippichai

Calibration Method : In-house method CAL-M2001 based on UKAS Publication ref : LAB 14

Edition 7 - November 2022

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Weights

ID No.	Cert. No.	Due Date	Traceability
F181-F1821	67-210457-1	12 Jun 2025	National Institute of Metrology (Thailand), (NIMT)

Approved by :

( Satja Sangkhum )

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



## Certificate of Calibration

Certificate No. : 68-200092-2

Page : 2 of 2

Result of Calibration : After Adjustment

UUC Condition As-Received : Good

Departure of indication from nominal value


Nominal Value (g)	Correction (g)	Uncertainty ± (g)	Error before Adjustment (g)
200	0.00	0.011	0.00
500	0.00	0.011	0.00
600	0.00	0.011	0.00
700	0.00	0.011	0.00
800	0.00	0.011	0.00
1000	0.01	0.011	0.00
1200	0.01	0.012	0.01
1500	0.01	0.012	0.02
2000	0.00	0.014	0.03
2200	0.00	0.023	0.06

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor  $k = 2.00$ , providing a level of confidence of approximately 95%

Eccentric error Load test : 500 g

A	B	C	D	E	g
0.00	-0.01	-0.01	0.00	0.00	



Repeatability Load test : 2000 g

Stdev. : 0.000 g

- o o o -





บริษัท เอ็นไวร์ เซอร์วิส จำกัด  
ENVIR SERVICE CO., LTD.

## บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201

42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bankok 10230 Tel : 02-9435814-5 Fax : 02-9438201

### Analyzer Performance Test

Calibrated Date: 23 May 2025

#### Instruments Information

Analyzer Type: Total Hydrocarbon Analyzer Model: 51C	Manufacturer Thermo Enviromental S/N: 51HT-50961-286
---	---

#### Calibration System

Calibrator Unit	Standard Gas
Dilutor Model Dasibi S/N: 705 ZERO AIR Generator API Model 701 S/N: 1924	Mrteane 180 PPM Propane 181 PPM Cylinder AAL5888 Expire Date: 24 May, 2027

Environment: Temperature 25 °C


Humidity: 51 %RH

#### Calibration Check ( Before adjust )

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NH4	0.1	0.0	0.1	17.8	18.0	-0.9
NMHC	0.1	0.0	0.1	17.8	18.0	-0.9

#### Calibration Check ( After adjust )

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NH4	0.0	0.0	0.0	18.0	18.0	0.0
NMHC	0.0	0.0	0.0	18.0	18.0	0.0



*Mr. Pasagorn Samol*

Calibrate By : Mr. Pasagorn Samol



บริษัท เอ็นไวร์ เซอร์วิส จำกัด  
ENVIR SERVICE CO., LTD.

## บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201

42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bankok 10230 Tel : 02-9435814-5 Fax : 02-9438201

### Analyzer Performance Test

Calibrated Date: 23 May 2025

#### Instruments Information

Analyzer Type: NO/NO <sub>2</sub> /NO <sub>x</sub> Analyzer Model: 42C	Manufacturer Thermo Environmental S/N: 42C-33500-371
---	---

#### Calibration System

Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008 S/N: 705 ZERO AIR Generator API Model 701 S/N: 1924	NO Conc 55.47 PPM SO <sub>2</sub> Conc 55.11 PPM CO Conc 4,535 PPM Cylinder number EB0129027 Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C


Humidity: 51 %RH

#### Calibration Check ( Before adjust )

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.1	0.0	0.1	393.4	400.0	-1.7
NO <sub>x</sub>	0.1	0.0	0.1	396.7	400.0	-0.8

#### Calibration Check ( After adjust )

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.0	0.0	0.0	400.0	400.0	0.0
NO <sub>x</sub>	0.0	0.0	0.0	400.0	400.0	0.0



*Signature*

Calibrate By : Mr. Pasagorn Samol





บริษัท เอ็ม อี ที จำกัด MET Company Limited

36/659 หมู่ 6 ต.บางรักพัฒนา อ.บางบัวทอง จ. นนทบุรี 11110

36/659 Moo 6 Tambol Bangrakpattana Amphur Bangbuatong Nontaburi 11110

Tel : 0 2920 1458-9 Fax : 0 2920 1460 E-mail : met\_jj@yahoo.com

## PM10 HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Sampler Location				Date	May 26, 2025
ที่ทำการผู้ใหญ่บ้าน หมู่ 3				Start Time	3:20 PM
Sampler Number	PM-10 No.7	Transfer Standard Type	Orifice	Stop Time	3:25 PM
Motor Serial Number	HVL-07	Calibrator Model	TE-5025A	Person	Mr.Jirayut Seehabut
Recorder Serial Number	-	Calibrator Serial Number	1		


Plate No.	(Delta H)			(A)	(X)	(I)	(Y)	Temperature	Barometric Pressure	Start Meter	Stop Meter
	Pressure Drop Across Orifice (inH <sub>2</sub> O)			$[\Delta H_2O(Pa/P_{std})(T_{std}/Ta)]^{1/2}$		Qstd = (1/m)[(A-b)]		IC = I[(Pa/P <sub>std</sub> )(T <sub>std</sub> /Ta)] <sup>1/2</sup>			
	Positive	Negative	$\Delta H_2O$			(m <sup>3</sup> /min)		(inch)			
5	1.4	1.4	2.8	1.65074		0.84821		0.9		0.89	
7	1.9	2.0	3.9	1.94819		1.00366		1.3		1.28	
10	2.4	2.5	4.9	2.18372		1.12676		1.9		1.87	
13	3.7	3.7	7.4	2.68358		1.38799		2.4		2.37	
18	4.5	4.5	9.0	2.95952		1.53220		3.2		3.16	

Linear Regression Y ON X : Y= mX + b

1	Slope ( m )	1.91345	Linear Equation		Average	305.0	757.0		
2	Intercept ( b )	0.02773	Set Point Flow Rate ( X ) (m <sup>3</sup> /min)	1.133	r <sup>2</sup>	0.993306	Pstd(mmHg)	760.0	
3	Correlation Coefficient ( r )	0.99995	Final Set Flow Rate = ( I )	0	r	0.9966474	T <sub>NTP</sub>	298.0	
Result							C=(Pa/Pstd)*(Tstd/Ta)		0.973192407
							C=(Pa/Pstd)*(Tstd/Ta)^0.5		0.986505148

COMMENT

Andersen Instruments, Inc.



Calibrated By .....

(Mr.Jirayut Seehabut)  
Field Environmental

Approved By .....

(Mr.Jarung Jamnongbut)  
Division Manager



## บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201  
บริษัท เอ็นไวร์ เซอร์วิส จำกัด 42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bankok 10230 Tel : 02-9435814-5 Fax : 02-9438201  
ENVIR SERVICE CO., LTD.

### Analyzer Performance Test

Calibrated Date: 23 May 2025

#### Instruments Information

Analyzer Type: SO2 Analyzer Model: 43C	Manufacturer Thermo Environmental S/N: 43C-33500-719
---	---

#### Calibration System


Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008 S/N: 705 ZERO AIR Generator API MODEL 701 S/N: 1924	NO Conc 55.47 PPM SO2 Conc 55.11 PPM CO Conc 4,535 PPM Cylinder number EB0129027 Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C

Humidity: 51 %RH

#### Calibration Report

Status	Zero			Span		
	Reference (PPB)	Reading (PPB)	Drift (PPB)	Reference (PPB)	Reading (PPB)	Drift%
Before	0.0	0.1	0.1	400.0	393.3	-1.7
After	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :

Mr.PASAGORN SAMOL



บริษัท เอ็ม อี ที จำกัด MET Company Limited

36/659 หมู่ 6 ต.บางรักพัฒนา อ.บางบัวทอง จ. นนทบุรี 11110

36/659 Moo 6 Tambol Bangrakpattana Amphur Bangbuatong Nontaburi 11110

Tel : 0 2920 1458-9 Fax : 0 2920 1460 E-mail : met\_jj@yahoo.com

## TSP HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Sampler Location				Date	May 26, 2025
ที่ทำการผู้ใหญ่บ้าน หมู่ 3				Start Time	3:20 PM
Sampler Number	TSP No.7	Transfer Standard Type	Orifice	Stop Time	3:25 PM
Motor Serial Number	BL-07	Calibrator Model	TE-5025A	Person	Mr.Jirayut Seehabut
Recorder Serial Number	-	Calibrator Serial Number	1		


Plate No.	(Delta H)			(A)	(X)	(I)	(Y)	Temperature	Barometric Pressure	Start Meter	Stop Meter
	Pressure Drop Across Orifice (inH <sub>2</sub> O)			$[\Delta H_2O(Pa/P_{std})(T_{std}/Ta)]^{1/2}$	Qstd = (1/m)[(A-b)] (m <sup>3</sup> /min)	Sample Flow Rate Indication (inch)	IC = I[(Pa/P <sub>std</sub> )(T <sub>std</sub> /Ta)] <sup>1/2</sup>	°K = °C+273)	(mmHg)		
	Positive	Negative	ΔH <sub>2</sub> O								
5	1.5	1.5	3.0	1.70868	0.87849	1.3	1.28	305.0	757.0		
7	2.6	2.6	5.2	2.24958	1.16117	2.3	2.27	305.0	757.0		
10	3.8	3.8	7.6	2.71961	1.40682	3.1	3.06	305.0	757.0		
13	5.0	5.0	10.0	3.11960	1.61586	4.2	4.14	305.0	757.0		
18	6.4	6.4	12.8	3.52943	1.83004	5.1	5.03	305.0	757.0		

Linear Regression Y ON X : Y= mX + b

1	Slope ( m )	1.91345	Linear Equation			$r^2$	0.963579	Pstd(mmHg)	760.0
2	Intercept ( b )	0.02773	Set Point Flow Rate ( X ) (m <sup>3</sup> /min)	1.133	r	0.9816206	T <sub>NTP</sub>	298.0	
3	Correlation Coefficient ( r )	0.99995	Final Set Flow Rate = ( I )	0	(Pa/Pstd)*(Tstd/Ta)		0.973192407		
Result						C=(Pa/Pstd)*(Tstd/Ta)^0.5	0.986505148		

COMMENT

Andersen Instruments, Inc.



Calibrated By .....

(Mr.Jirayut Seehabut)  
Field Environmental

Approved By .....

(Mr.Jarung Jamnongbut)  
Division Manager



บริษัท เอ็นไวร์ เซอร์วิส จำกัด  
ENVIR SERVICE CO., LTD.

## บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201

42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bankok 10230 Tel : 02-9435814-5 Fax : 02-9438201

### Analyzer Performance Test

Calibrated Date: 23 May 2025

#### Instruments Information

Analyzer Type: CO Analyzer Model: 300	Manufacturer API S/N: 200-S
--	--------------------------------

#### Calibration System


Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008 S/N: 705 ZERO AIR Generator API MODEL 701 S/N: 1924	NO Conc 55.47 PPM SO2 Conc 55.11 PPM CO Conc 4,535 PPM Cylinder number EB0129027 Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C

Humidity: 51 %RH

#### Calibration Report

Status	Zero			Span		
	Reference (PPM)	Reading (PPM)	Drift (PPM)	Reference (PPM)	Reading (PPM)	Drift%
Before	0.0	0.2	0.2	44.3	45.0	1.7
After	0.0	0.0	0.0	45.0	45.0	0.0



Calibrate By :

Mr. PASAGORN SAMOL





บริษัท เอ็นไวร์ เซอร์วิส จำกัด  
ENVIR SERVICE CO., LTD.

## บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201

42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bankok 10230 Tel : 02-9435814-5 Fax : 02-9438201

### Analyzer Performance Test

Calibrated Date: 23 May 2025

#### Instruments Information

Analyzer Type: NO/NO <sub>2</sub> /NO <sub>x</sub> Analyzer Model: 42C	Manufacturer Thermo Environmental S/N: 42C-601114773
---	---

#### Calibration System

Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008 S/N: 705 ZERO AIR Generator API Model 701 S/N: 1924	NO Conc 55.47 PPM SO <sub>2</sub> Conc 55.11 PPM CO Conc 4,535 PPM Cylinder number EB0129027 Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C


Humidity: 51 %RH

#### Calibration Check ( Before adjust )

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.1	0.0	0.1	393.3	400.0	-1.7
NO <sub>x</sub>	0.1	0.0	0.1	396.4	400.0	-0.9

#### Calibration Check ( After adjust )

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.0	0.0	0.0	400.0	400.0	0.0
NO <sub>x</sub>	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By : Mr. Pasagorn Samol



บริษัท เอ็ม อี ที จำกัด MET Company Limited

36/659 หมู่ 6 ต.บางรักพัฒนา อ.บางบัวทอง จ. นนทบุรี 11110

36/659 Moo 6 Tambol Bangrakpattana Amphur Bangbuatong Nontaburi 11110

Tel : 0 2920 1458-9 Fax : 0 2920 1460 E-mail : met\_jj@yahoo.com

## PM10 HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Sampler Location				Date	May 26, 2025
โรงเรียนวัดดอนขี้เหล็ก				Start Time	2:20 PM
Sampler Number	PM-10 No.1	Transfer Standard Type	Orifice	Stop Time	2:25 PM
Motor Serial Number	HVL-01	Calibrator Model	TE-5025A	Person	Mr.Jirayut Seehabut
Recorder Serial Number	-	Calibrator Serial Number	1		


Plate No.	(Delta H)			(A)	(X)	(I)	(Y)	Temperature	Barometric Pressure	Start Meter	Stop Meter
	Pressure Drop Across Orifice (inH <sub>2</sub> O)			$[\Delta H_2O(Pa/P_{std})(T_{std}/T_a)]^{1/2}$	Qstd = (1/m)[(A-b)] (m <sup>3</sup> /min)	Sample Flow Rate Indication (inch)	IC = I[(Pa/P <sub>std</sub> )(T <sub>std</sub> /T <sub>a</sub> )] <sup>1/2</sup>	(°K = °C+273)	(mmHg)		
	Positive	Negative	ΔH <sub>2</sub> O								
5	1.4	1.4	2.8	1.65074	0.84821	1.2	1.18	305.0	757.0		
7	2.5	2.5	5.0	2.20589	1.13834	2.2	2.17	305.0	757.0		
10	3.4	3.5	6.9	2.59134	1.33978	3.1	3.06	305.0	757.0		
13	4.6	4.6	9.2	2.99222	1.54929	4.2	4.14	305.0	757.0		
18	5.8	5.8	11.6	3.35992	1.74145	5.2	5.13	305.0	757.0		

Linear Regression Y ON X : Y= mX + b

Linear Regression: Pstd*(Tstd/Ta)^0.5				Average		305.0	757.0	
1	Slope ( m )	1.91345	Linear Equation		$r^2$	0.987743	Pstd(mmHg)	760.0
2	Intercept ( b )	0.02773	Set Point Flow Rate ( X ) (m <sup>3</sup> /min)	1.133	r	0.9938526	T <sub>NTP</sub>	298.0
3	Correlation Coefficient ( r )	0.99995	Final Set Flow Rate = ( I )	0	(Pa/Pstd)*(Tstd/Ta)	0.973192407		
Result						C=(Pa/Pstd)*(Tstd/Ta)^0.5	0.986505148	

COMMENT

Andersen Instruments, Inc.



Calibrated By .....  
(Mr.Jirayut Seehabut)  
Field Environmental

Approved By .....  
(Mr.Jarung Jamnongbut)  
Division Manager



## บริษัท เอ็นไวร์ เซอร์วิส จำกัด

บริษัท เอ็นไวร์ เซอร์วิส จำกัด  
ENVIR SERVICE CO., LTD.

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201

42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bankok 10230 Tel : 02-9435814-5 Fax : 02-9438201

### Analyzer Performance Test

Calibrated Date: 23 May 2025

#### Instruments Information

Analyzer Type: SO2 Analyzer Model: 43C	Manufacturer Thermo Environmental S/N: 43C-71354-368
---	---

#### Calibration System


Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008 S/N: 705 ZERO AIR Generator API MODEL 701 S/N: 1924	NO Conc 55.47 PPM SO2 Conc 55.11 PPM CO Conc 4,535 PPM Cylinder number EB0129027 Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C

Humidity: 51 %RH

#### Calibration Report

Status	Zero			Span		
	Reference (PPB)	Reading (PPB)	Drift (PPB)	Reference (PPB)	Reading (PPB)	Drift%
Before	0.0	0.1	0.1	400.0	393.4	-1.7
After	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :

Mr.PASAGORN SAMOL



บริษัท เอ็ม อี ที จำกัด MET Company Limited

36/659 หมู่ 6 ต.บางรักพัฒนา อ.บางบัวทอง จ. นนทบุรี 11110

36/659 Moo 6 Tambol Bangrakpattana Amphur Bangbuatong Nontaburi 11110

Tel : 0 2920 1458-9 Fax : 0 2920 1460 E-mail : met\_jj@yahoo.com

## TSP HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Sampler Location				Date	May 26, 2025
โรงเรียนวัดดอนขี้เหล็ก				Start Time	1:20 PM
Sampler Number	TSP No.13	Transfer Standard Type	Orifice	Stop Time	1:25 PM
Motor Serial Number	BL-13	Calibrator Model	TE-5025A	Person	Mr.Jirayut Seehabut
Recorder Serial Number	-	Calibrator Serial Number	1		


Plate No.	(Delta H)			(A)	(X)	(I)	(Y)	Temperature	Barometric Pressure	Start Meter	Stop Meter
	Positive	Negative	$\Delta H_2O$	$[\Delta H_2O(Pa/P_{std})(T_{std}/T_a)]^{1/2}$	$Q_{std} = (1/m)[(A-b)]$ ( $m^3/min$ )	sample Flow Rate Indication (inch)	$IC = I[(Pa/P_{std})(T_{std}/T_a)]^{1/2}$	( $^{\circ}K = ^{\circ}C + 273$ )	(mmHg)		
5	1.5	1.5	3.0	1.70868	0.87849	1.2	1.18	305.0	757.0		
7	2.6	2.7	5.3	2.27111	1.17242	1.7	1.68	305.0	757.0		
10	4.2	4.2	8.4	2.85916	1.47975	2.7	2.66	305.0	757.0		
13	5.3	5.4	10.7	3.22694	1.67196	3.5	3.45	305.0	757.0		
18	6.6	6.7	13.3	3.59770	1.86573	4.4	4.34	305.0	757.0		

Linear Regression Y ON X :  $Y = mX + b$

Linear Regression: $P_{std} = f(X)$				Average		$r^2$	Pstd	
1	Slope ( m )	1.91345	Linear Equation				Pstd(mmHg)	760.0
2	Intercept ( b )	0.02773	Set Point Flow Rate ( X ) (m <sup>3</sup> /min)	1.133	r	0.930028	T <sub>NTP</sub>	298.0
3	Correlation Coefficient ( r )	0.99995	Final Set Flow Rate = ( I )	0	(Pa/Pstd)*(Tstd/Ta)		0.973192407	
Result						C=(Pa/Pstd)*(Tstd/Ta)^0.5		0.986505148

COMMENT

Andersen Instruments, Inc.



Calibrated By .....

(Mr.Jirayut Seehabut)  
Field Environmental

Approved By .....

(Mr.Jarung Jamnongbut)  
Division Manager





บริษัท เอ็นไวร์ เซอร์วิส จำกัด  
ENVIR SERVICE CO., LTD.

## บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201

42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bangkok 10230 Tel : 02-9435814-5 Fax : 02-9438201

### Analyzer Performance Test

Calibrated Date: 23 May 2025

#### Instruments Information

Analyzer Type: CO Analyzer Model: 300	Manufacturer API S/N: 203-S
--	--------------------------------

#### Calibration System


Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008 S/N: 705 ZERO AIR Generator API MODEL 701 S/N: 1924	NO Conc 55.47 PPM SO2 Conc 55.11 PPM CO Conc 4,535 PPM Cylinder number EB0129027 Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C

Humidity: 51 %RH

#### Calibration Report

Status	Zero			Span		
	Reference (PPM)	Reading (PPM)	Drift (PPM)	Reference (PPM)	Reading (PPM)	Drift%
Before	0.0	0.2	0.2	44.2	45.0	1.7
After	0.0	0.0	0.0	45.0	45.0	0.0



Calibrate By :

Mr. PASAGORN SAMOL



บริษัท เอ็นไวร์ เซอร์วิส จำกัด  
ENVIR SERVICE CO., LTD.

## บริษัท เอ็นไวร์ เซอร์วิส จำกัด

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201

42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bankok 10230 Tel : 02-9435814-5 Fax : 02-9438201

### Analyzer Performance Test

Calibrated Date: 23 May 2025

#### Instruments Information

Analyzer Type: NO/NO <sub>2</sub> /NO <sub>x</sub> Analyzer Model: 42C	Manufacturer Thermo Environmental S/N: 42C-601114783
---	---

#### Calibration System

Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008 S/N: 705 ZERO AIR Generator API Model 701 S/N: 1924	NO Conc 55.47 PPM SO <sub>2</sub> Conc 55.11 PPM CO Conc 4,535 PPM Cylinder number EB0129027 Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C


Humidity: 51 %RH

#### Calibration Check ( Before adjust )

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.1	0.0	0.1	393.5	400.0	-1.6
NO <sub>x</sub>	0.1	0.0	0.1	396.2	400.0	-1.0

#### Calibration Check ( After adjust )

GAS	Zero			Span		
	Reading Value (ppb)	Expected Value (ppb)	Drift (ppb)	Reading Value (ppb)	Expected Value (ppb)	Drift%
NO	0.0	0.0	0.0	400.0	400.0	0.0
NO <sub>x</sub>	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By : Mr. Pasagorn Samol



บริษัท เอ็ม อี ที จำกัด MET Company Limited

36/659 หมู่ 6 ต.บางรักพัฒนา อ.บางบัวทอง จ. นนทบุรี 11110

36/659 Moo 6 Tambol Bangrakpattana Amphur Bangbuatong Nontaburi 11110

Tel : 0 2920 1458-9 Fax : 0 2920 1460 E-mail : met\_jj@yahoo.com

## PM10 HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Sampler Location				Date	May 26, 2025
วัดบวรยงสุวรรณาาราม (ม่วงโพรง)				Start Time	2:20 PM
Sampler Number	PM-10 No.2	Transfer Standard Type	Orifice	Stop Time	2:25 PM
Motor Serial Number	HVL-02	Calibrator Model	TE-5025A	Person	Mr.Jirayut Seehabut
Recorder Serial Number	-	Calibrator Serial Number	1		


Plate No.	(Delta H)			(A)	(X)	(I)	(Y)	Temperature	Barometric Pressure	Start Meter	Stop Meter
	Pressure Drop Across Orifice (inH <sub>2</sub> O)			$[\Delta H_2O(Pa/P_{std})(T_{std}/Ta)]^{1/2}$	$Q_{std} = (1/m)[(A-b)]$ (m <sup>3</sup> /min)	sample Flow Rate Indication (inch)	$IC = I[(Pa/P_{std})(T_{std}/Ta)]^{1/2}$	(°K = °C+273)	(mmHg)		
	Positive	Negative	$\Delta H_2O$								
5	1.8	1.8	3.6	1.88165	0.96889	1.1	1.09	303.0	760.0		
7	2.9	2.9	5.8	2.38837	1.23371	1.9	1.88	303.0	760.0		
10	4.5	4.5	9.0	2.97514	1.54037	2.9	2.88	303.0	760.0		
13	5.7	5.7	11.4	3.34841	1.73544	3.8	3.77	303.0	760.0		
18	7.2	7.2	14.4	3.76329	1.95227	4.6	4.56	303.0	760.0		

Linear Regression Y ON X : Y= mX + b

1	Slope ( m )	1.91345	Linear Equation		Average	303.0	760.0		
2	Intercept ( b )	0.02773	Set Point Flow Rate ( X ) (m <sup>3</sup> /min)	1.133	r <sup>2</sup>	0.950727	Pstd(mmHg)	760.0	
3	Correlation Coefficient ( r )	0.99995	Final Set Flow Rate = ( I )	0	r	0.9750523	T <sub>NTP</sub>	298.0	
Result							C=(Pa/Pstd)*(Tstd/Ta)	0.98349835	
							C=(Pa/Pstd)*(Tstd/Ta)^0.5	0.991714853	

COMMENT

Andersen Instruments, Inc.



Calibrated By .....  
(Mr.Jirayut Seehabut)  
Field Environmental

Approved By .....  
(Mr.Jarung Jamnongbut)  
Division Manager



## บริษัท เอ็นไวร์ เซอร์วิส จำกัด

บริษัท เอ็นไวร์ เซอร์วิส จำกัด  
ENVIR SERVICE CO., LTD.

42 รามอินทรา 14 แยก 9 แขวงท่าแร้ง เขตบางเขน กรุงเทพฯ 10230 โทรศัพท์ 02-9435814-5 โทรสาร 02-9438201

42 Raminthra 14 yeak 9, Tha Rang, Bangkhen, Bankok 10230 Tel : 02-9435814-5 Fax : 02-9438201

### Analyzer Performance Test

Calibrated Date: 23 May 2025

#### Instruments Information

Analyzer Type: SO2 Analyzer Model: 100A	Manufacturer API S/N: 193
--	------------------------------

#### Calibration System


Calibrator Unit	Standard Gas
Dilutor Model Dasibi Model 5008 S/N: 705 ZERO AIR Generator API MODEL 701 S/N: 1924	NO Conc 55.47 PPM SO2 Conc 55.11 PPM CO Conc 4,535 PPM Cylinder number EB0129027 Expire Date: 29 Oct. 2027

Environment: Temperature 25.5 °C

Humidity: 51 %RH

#### Calibration Report

Status	Zero			Span		
	Reference (PPB)	Reading (PPB)	Drift (PPB)	Reference (PPB)	Reading (PPB)	Drift%
Before	0.0	0.1	0.1	400.0	393.4	-1.7
After	0.0	0.0	0.0	400.0	400.0	0.0



Calibrate By :

Mr.PASAGORN SAMOL





บริษัท เอ็ม อี ที จำกัด MET Company Limited

36/659 หมู่ 6 ต.บางรักพัฒนา อ.บางบัวทอง จ. นนทบุรี 11110

36/659 Moo 6 Tambol Bangrakpattana Amphur Bangbuatong Nontaburi 11110

Tel : 0 2920 1458-9 Fax : 0 2920 1460 E-mail : met\_jj@yahoo.com

## TSP HIGH VOLUME AIR SAMPLER CALIBRATION REPORT

Sampler Location				Date	May 26, 2025
วัดบวรสังสุวรณาราม (ม่วงโพรง)				Start Time	2:20 PM
Sampler Number	TSP No.6	Transfer Standard Type	Orifice	Stop Time	2:25 PM
Motor Serial Number	BL-06	Calibrator Model	TE-5025A	Person	Mr.Jirayut Seehabut
Recorder Serial Number	-	Calibrator Serial Number	1		


Plate No.	(Delta H)			(A)	(X)	(I)	(Y)	Temperature	Barometric Pressure	Start Meter	Stop Meter
	Positive	Negative	$\Delta H_2O$	$[\Delta H_2O(Pa/P_{std})(T_{std}/Ta)]^{1/2}$	$Q_{std} = (1/m)[(A-b)]$ ( $m^3/min$ )	sample Flow Rate Indication (inch)	$IC = I[(Pa/P_{std})(T_{std}/Ta)]^{1/2}$	( $^{\circ}K = ^{\circ}C + 273$ )	(mmHg)		
5	1.3	1.3	2.6	1.59069	0.81683	0.5	0.49	305.0	757.0		
7	2.5	2.5	5.0	2.20589	1.13834	1.5	1.48	305.0	757.0		
10	4.5	4.5	9.0	2.95952	1.53220	2.3	2.27	305.0	757.0		
13	6.2	6.2	12.4	3.47384	1.80099	3.1	3.06	305.0	757.0		
18	8.0	8.0	16.0	3.94602	2.04776	4.0	3.95	305.0	757.0		

Linear Regression Y ON X :  $Y = mX + b$

Linear Regression: $P_{std} = f(X)$				Average		$r^2$	Pstd (mmHg)	
1	Slope ( m )	1.91345	Linear Equation			0.992656	Pstd(mmHg)	760.0
2	Intercept ( b )	0.02773	Set Point Flow Rate ( X ) (m <sup>3</sup> /min)	1.133	r	0.9963212	T <sub>NTP</sub>	298.0
3	Correlation Coefficient ( r )	0.99995	Final Set Flow Rate = ( I )	0	(Pa/Pstd)*(Tstd/Ta)		0.973192407	
Result					C=(Pa/Pstd)*(Tstd/Ta)^0.5		0.986505148	

COMMENT

Andersen Instruments, Inc.
----------------------------



Calibrated By .....  
(Mr.Jirayut Seehabut)  
Field Environmental

Approved By .....  
(Mr.Jarung Jamnongbut)  
Division Manager

ใบรับรองการสอบเทียบเครื่องมือ  
ตรวจวัดสภาพแวดล้อมในพื้นที่ทำงาน



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL.0-2717-3000-29 FAX.0-2719-9484

## Certificate of Calibration

Cert. No.: 25MD556

Page : 1 of 3

Equipment : Gas Flow Meter  
Model : DCL-M  
Serial No. : 103343  
ID No. : -  
Manufacturer : Bios International Corp.  
Submitted by : Viridian Environmental Service Co., Ltd.  
149/119 Moo 2, Krung Non-Chong Thanom Rd.,  
Mahasawat, Bang Kruay, Nonthaburi 11130  
Place of calibration : TPA Medical Equipment Calibration Lab.  
Ambient temperature :  $(23 \pm 2) ^\circ\text{C}$   
Relative humidity :  $(50 \pm 15) \%$   
Atmospheric pressure :  $(1010 \pm 13) \text{ mbar}$

Calibrated by : Kanda Traisin

Approved by :

Approved signatory

- ( ) Pornthippa Tameyakul  
(✓) Surin Yenprasert  
( ) Nattachai Sawangkunnopchai

Issue date : 2 April 2025

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Calibration and Testing Equipment Services.



Received order : 13 March 2025  
Condition as-received : Used item  
Calibration date : 27 March 2025  
Reference : 2503-0475WN-1

Cert. No.: 25MD556

Page : 2 of 3

### Procedure used :-

Calibration was conducted using in-house calibration procedure : CP-MD11, according to  
comparison method, using dry air as gas media.

### Conditions of this result of calibration

#### 1. Reference standard instrument :-

Instrument	Model	Serial No.	Cert. No.	Due date
1.1 Drum-Type Gas Meter	TG3/9	0.541.F57	MW-0018-25	24 Feb 2028
1.2 Digital Pressure Gauge	681	211H16340004	25P815	27 Feb 2026
1.3 Thermometer	HH376	140808579	24I607	10 Jun 2025
1.4 Thermometer	HH 376.	230409515	24I1183	30 Oct 2025
1.5 Digital Test Gauge	30PSIXP2i	354478	24P3804	4 Nov 2025
1.6 Digital Test Gauge	100PSIXP2i.	470307	25P251	20 Jan 2026

2. The certificate is valid only to the item calibrated on date and place of calibration.  
3. This result of calibration was made on requested at the point specified by customer.  
4. This certification is traceable to the International System of Units, through :-  
- National Institute of Metrology (Thailand)  
- National Institute of Metrology (Thailand), through Technology Promotion Association (Thailand-Japan)



**Result of calibration**

Cert. No.: 25MD556

Condition of UUC\* :-

Reading : Volumetric flow

Gas type : Dry air

Warm-up : Power on at least 30 minutes before calibration

Page : 3 of 3

Function : Flow measurement

UUC* Gas Temperature ( °C )	Applied Gas flow ( L/min )	UUC* Reading Without adjust ( L/min )	UUC* Error Without adjust ( L/min )	Uncertainty ( ± L/min )	Uncertainty ± %reading
22.59	5.012	5.012	0.000	0.056	1.12
22.84	4.060	4.044	-0.016	0.042	1.04
22.77	3.008	3.019	0.011	0.032	1.06
22.87	2.074	2.060	-0.014	0.025	1.21
22.95	1.527	1.523	-0.004	0.018	1.18
22.81	1.011	1.010	-0.001	0.011	1.09
22.60	0.400	0.4007	0.0007	0.0051	1.27

UUC\* : Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor (  $k = 2$  ), providing a level of confidence of approximately 95 %.

-000-





Ref. No.H-PP 05-2025

PERSONAL AIR SAMPLER CALIBRATION REPORT : HI-FLOW

Calibration Date :  
Expire Date : May 01, 2025  
Condition of Calibration :  
Temperature : 28 °C  
Ambient Pressure : 758 mmHg  
Humidity : 59 % RH  
Reference Equipment : Electronic Primary Gas Flow Meter (Graphite)  
Calibrator Description : DryCal Model DCL-M  
Serial No. : 103343  
Cert. No. : 25M0556  
Calibration Date : March 27, 2025  
Expire Date : March 26, 2026  
Equipment Detail :  
Sensidyne Gillian Model BDX II  
Gillian Model GilAir 3  
Gillian Model GilAir 5  
YINHEKE/J Model QCD-1500

Personal Pump ID No. H-01 to H-09  
Personal Pump ID No. H-10, H-17 to H-24  
Personal Pump ID No. H-11 to H-16  
Personal Pump ID No. PH-27 to PH-31

ID. No.	Description	Serial No.	Flow Rate Reading (LPM)									
H-01	20150205030	0.5	1.0	1.5	2.0	2.5	2.5	1.7	1.7	1.7	1.7	1.7
H-02	20150205033	0.8	1.2	1.5	2.0	2.5	2.5	1.8	1.8	1.8	1.8	1.8
H-03	20150205033	0.8	1.2	1.6	2.0	2.5	2.5	1.8	1.8	1.8	1.8	1.8
H-04	20150602005	0.8	1.2	1.6	2.0	2.5	2.5	1.8	1.8	1.8	1.8	1.8
H-05	20200504091	0.7	1.1	1.6	1.9	2.5	2.5	1.8	1.8	1.8	1.8	1.8
H-06	20200504095	0.6	1.1	1.4	1.9	2.5	2.5	1.7	1.7	1.7	1.7	1.7
H-07	20200504096	0.7	1.2	1.5	2.0	2.5	2.5	1.8	1.8	1.8	1.8	1.8
H-08	20200504097	0.8	1.1	1.5	2.0	2.5	2.5	1.8	1.8	1.8	1.8	1.8
H-09	20200504100	0.7	1.1	1.5	1.9	2.5	2.5	1.7	1.7	1.7	1.7	1.7
H-10	19980306759	0.5	1.2	1.6	2.0	2.5	2.5	1.7	1.7	1.7	1.7	1.7
H-11	19920601089	0.7	1.1	1.6	2.1	2.6	2.6	1.8	1.8	1.8	1.8	1.8
H-12	19920601083	0.5	1.0	1.4	1.9	2.5	2.5	1.7	1.7	1.7	1.7	1.7
H-13	19920701189	0.7	1.1	1.5	2.1	2.6	2.6	1.8	1.8	1.8	1.8	1.8
H-14	19920701190	0.7	1.2	1.6	2.1	2.6	2.6	1.8	1.8	1.8	1.8	1.8
H-15	19920901283	0.4	1.0	1.4	1.8	2.5	2.5	1.7	1.7	1.7	1.7	1.7
H-16	19940103770	0.5	1.0	1.5	2.1	2.5	2.5	1.7	1.7	1.7	1.7	1.7
H-17	20080321655	0.6	1.1	1.5	2.0	2.4	2.4	1.7	1.7	1.7	1.7	1.7
H-18	20120512608	0.6	1.2	1.6	2.0	2.4	2.4	1.7	1.7	1.7	1.7	1.7
H-19	20121113114	0.6	1.0	1.6	2.1	2.5	2.5	1.8	1.8	1.8	1.8	1.8
H-20	20130415277	0.6	1.2	1.6	2.2	2.6	2.6	1.8	1.8	1.8	1.8	1.8
H-21	20130415279	0.7	1.2	1.5	2.0	2.5	2.5	1.8	1.8	1.8	1.8	1.8
H-22	20130415371	0.6	1.1	1.4	1.9	2.5	2.5	1.7	1.7	1.7	1.7	1.7
H-23	20130415394	0.6	1.1	1.5	1.9	2.5	2.5	1.7	1.7	1.7	1.7	1.7
H-24	20130415640	0.6	1.2	1.5	2.1	2.5	2.5	1.7	1.7	1.7	1.7	1.7
PH-27	233234	0.6	1.2	1.5	2.1	2.5	2.5	1.7	1.7	1.7	1.7	1.7
PH-28	233237	0.5	1.2	1.6	2.2	2.6	2.6	1.8	1.8	1.8	1.8	1.8
PH-29	233242	0.7	1.2	1.5	2.0	2.5	2.5	1.7	1.7	1.7	1.7	1.7
PH-30	233295	0.4	1.1	1.4	1.9	2.5	2.5	1.7	1.7	1.7	1.7	1.7
PH-31	233299	0.6	1.2	1.5	2.1	2.5	2.5	1.7	1.7	1.7	1.7	1.7

\* Flow 1.7 L/min for Nylon Cyclone on Cyclone Calibrating Jar

Calibrated by

Prakati Boonkerd



Ref. No.L-PP 05 2025

PERSONAL AIR SAMPLER CALIBRATION REPORT : LOW-FLOW

Calibration Date :  
Expire Date : May 01, 2025  
Condition of Calibration :  
Temperature : 28 °C  
Ambient Pressure : 758 mmHg  
Humidity : 59 % RH  
Reference Equipment : Electronic Primary Gas Flow Meter (Graphite)  
Calibrator Description : DryCal Model DCL-M  
Serial No. : 103343  
Cert. No. : 25M0556  
Calibration Date : March 27, 2025  
Expire Date : March 26, 2026  
Equipment Detail :  
Gillian Model LFS-1130C  
YINHEKE/J Model QCD-1500

Personal Pump ID No. L-01 to L-11  
Personal Pump ID No. PH-27 to PH-31

ID. No.	Description	Serial No.	Type of Sorbent tubes									
L-01	11593	226-01	226-81A	226-17A	226-10-03	226-10-03	226-10-03	226-10-03	226-10-03	226-10-03	226-10-03	226-10-03
L-02	14759	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19
L-03	14760	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19
L-04	14762	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19
L-05	14808	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19
L-06	14809	0.20	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19
L-07	14812	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19
L-08	14814	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19
L-09	14852	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19
L-10	15878	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19
L-11	15880	0.18	0.18	0.18	0.18	0.18	0.18	0.18	0.18	0.18	0.18	0.18
PH-27	233234	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19
PH-28	233237	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19
PH-29	233242	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19
PH-30	233295	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19	0.19
PH-31	233299	0.18	0.18	0.18	0.18	0.18	0.18	0.18	0.18	0.18	0.18	0.18

Ref. Flow Rate on 0.2 LPM for all Sorbent tubes

Calibrated by

Prakati Boonkerd



# CRYSTAL CALIBRATION SALES AND SERVICE CO., LTD.

45/48 Soi Salathammassop31, Salathammassop Rd.,  
Salathammassop, Thawewattana, Bangkok 10170 Thailand  
Tel : 0-2408-8474-5 Fax : 0-2408-8477 Email : info@crystalcal.com www.crystalcal.com



NSC-TS1-ITS 17025  
CALIBRATION 0260

## CERTIFICATE OF CALIBRATION

Issue Date

: 31 January 2025

Certificate No. : 25-0133-001

Work Order No. : 25/0133

Customer Name

: Viridian Environmental Service Co., Ltd.  
149/199 Moo 2, Krung Non-Chong Thanom Rd.,  
Mahachulalongkornrajavidyalaya, Nonthaburi 11130

Date of Received

: 31 January 2025

Date of Calibration

: 31 January 2025

Instrument Details

: Description : Electronic Balance

: Manufacturer : arzet

: Model : CY 224

: Serial No. : 17308352

: ID No. : N/A

: Resolution : 0.0001 g

: Capacity : 220 g

: Location : Preparation Room

Calibration Method

: This calibration was conducted by using in-house method according to calibration  
procedure no. CWI-B-01 based on UKAS LAB14 edition 6, October 2019

Environmental Condition

Temperature : Maximum 27.9°C / Minimum 27.4°C

Humidity : Maximum 53%RH / Minimum 50%RH

Air Pressure : Maximum 1001.2hPa / Minimum 1001.2hPa

Traceability of Measurement

: This certificate of calibration documents the traceability to national standard,  
which realize the unit of measurement according to the International system of  
Units (SI)

Calibrated by :

Mr. Jaturong Dabkaew

Calibration Engineer

Approved by :

( Mr. Anuwat Yaklemjit )

Laboratory Manager

This certificate may not be reproduced other than in full except with the prior written approval of Crystal Calibration Sales and Service Co., Ltd.

Crystal Calibration Sales and Service Co., Ltd.

45/48 Salathammassop 31, Salathammassop Rd., Salathammassop, Thawewattana, Bangkok 10170

Phone : 0-2408-8474 Fax : 0-2408-8477 http://www.crystalcal.com Email : info@crystalcal.com

PAGE 1/3



# CRYSTAL CALIBRATION SALES AND SERVICE CO., LTD.

45/48 Soi Salathammassop31, Salathammassop Rd.,  
Salathammassop, Thawewattana, Bangkok 10170 Thailand  
Tel : 0-2408-8474-5 Fax : 0-2408-8477 Email : info@crystalcal.com www.crystalcal.com



NSC-TS1-ITS 17025  
CALIBRATION 0260

## CERTIFICATE OF CALIBRATION

Issue Date

: 31 January 2025

Certificate No. : 25-0133-001

Work Order No. : 25/0133

Details of Calibration

1. Reference Standards Instrument

Instrument	Capacity of Weight	Serial No. / ID No.	Certificate No.	Due date
Weight Set E2	1mg to 200g	43529071	M300098	13 May 2026

2. Certificate traceable

: This certificate traceable to The International System of Unit refer to

Sartorius Lab Instrument GmbH & Co.KG., DAWS Calibration No. DK-19398-01

3. Condition of item

: Used

4. Calibration site

: On-site

Result of Calibration

1. Calibration result : Check performance before calibration

Applied Weight g	Balance Reading g	Correction Value g	Uncertainty (±) g	Coverage Factor (k)
100.0000	99.9986	0.0014	0.00019	2.00
200.0001	199.9965	0.0036	0.00032	2.00

2. The result of check performance in first step has to Reset span

3. Calibration result : After set span by External Reset span by weight 200 g ID No. WE2-10

3.1 Repeatability number of repeatability is 10 times

Normal Value ( g )	Standard Deviation of Reading ( g )
100	0.0000463
200	0.0000422

This result of calibration was found accurate as shown on date and place of calibration only.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2 providing  
a level of confidence of approximately 95%.

PAGE 2/3



## CERTIFICATE OF CALIBRATION

Issue Date


: 31 January 2025

Certificate No. : 25-0133-001

Work Order No. : 25/0133

## 3. Calibration result : After set span by External Reset span by weight 200 g ID No. WEZ-10 (continued)

3.2 Eccentric or Off-center Error A mass of 100 g was placed and moved to various position on pan.



Result of Eccentric Error	
Position 1	100.0005 g
Position 2	100.0005 g
Position 3	100.0014 g
Position 4	100.0003 g
Position 5	100.0006 g
(Maximum Difference)	0.0009 g

## 3.3 Departure of indication from nominal value

Applied Weight	Balance Reading	Correction Value	Uncertainty	Coverage Factor (k)
Unit load	0.0000	0.0000	0.00010	2.05
0.0010	0.0010	0.0000	0.00010	2.05
0.0100	0.0100	0.0000	0.00010	2.05
0.0200	0.0200	0.0000	0.00010	2.05
0.0500	0.0500	0.0000	0.00010	2.05
0.1000	0.1000	0.0000	0.00010	2.05
0.2000	0.2000	0.0000	0.00010	2.05
0.5000	0.5000	0.0000	0.00010	2.05
1.0000	1.0000	0.0000	0.00010	2.05
10.0000	10.0007	-0.0007	0.00011	2.00
50.0000	50.0004	-0.0004	0.00013	2.00
100.0000	100.0005	-0.0005	0.00019	2.00
200.0001	200.0001	0.0000	0.00032	2.00

## Note

Calibrate items in good condition and this report customer request and accepted in certificate.

This result of calibration was found accurate as shown on date and place of calibration only.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

--END--

PAGE 3/3

## Statements of conformity report

Refer to Certificate No. : 25-0133-001

Work Order No. : 25/0133

Issue Date : 31 January 2025

## Detail of Equipment

Description : Electronic Balance

Manufacturer : aczet

ID No. : N/A

Model : CY 224

Resolution : 0.0001 g

Serial no. : 17308352

Capacity : 220 g

## Result of Calibration

This result of calibration : Adjustment

Applied Weight	Balance Reading	Correction Value	Uncertainty	Uncertainty + Correction	Limit of performance	Evaluation
0.0010	0.0010	0.0000	0.00010	0.00010	0.00091	Pass
0.0100	0.0100	0.0000	0.00010	0.00010	0.00091	Pass
0.0200	0.0200	0.0000	0.00010	0.00010	0.00091	Pass
0.0500	0.0500	0.0000	0.00010	0.00010	0.00091	Pass
0.1000	0.1000	0.0000	0.00010	0.00010	0.00091	Pass
0.2000	0.2000	0.0000	0.00010	0.00010	0.00091	Pass
0.5000	0.5000	0.0000	0.00010	0.00010	0.00091	Pass
1.0000	1.0000	0.0000	0.00010	0.00010	0.00091	Pass
10.0000	10.0007	-0.0007	0.00011	0.00081	0.00092	Pass
50.0000	50.0004	-0.0004	0.00013	0.00053	0.00094	Pass
100.0000	100.0005	-0.0005	0.00019	0.00069	0.00100	Pass
200.0001	200.0001	0.0000	0.00032	0.00032	0.00113	Pass

Unit : g

The conformity certificate documents validity following ISO/IEC Guide 98-4 : Role of measurement uncertainty in conformity assessment based on statement with guard band refer to NATA User checks and maintenance of laboratory balances consider expanded measurement uncertainty (k=95%).

## The tolerance and decision rules :

Limit of performance (LoP) =  $2.26 \text{ sw} + |C_{\text{cor}}| + U(C_{\text{cal}})$ 

Evaluation of decision = Measurement uncertainty + Absolute correction ; s LoP is pass , &gt; LoP is Fail

Statements of conformity decision by :

(Mr. Anuwat Yakkernijit)

Laboratory Manager

-- END --

Certificate No. : CAL-24-317

Page : 1 of 3

## CERTIFICATE OF CALIBRATION

Equipment : Spectrophotometer  
Manufacturer : Thermo Scientific  
Model : Genesys 10S VIS  
Serial No. : 2E8R020104  
ID No. : 0402-0101-13  
Customer : Emex Association Co., Ltd.  
27, 29 Soi Rama 2, Soi 30,  
Rama Road, Bangmod,  
Jomthong, Bangkok 10150  
Location : ห้องปฏิบัติการ 4  
Date of Receipt : 19 July 2024  
Date of Calibration : 19 July 2024  
Date of Issue : 23 July 2024  
Ambient Temperature : (25±10) °C  
Relative Humidity : (60±20) %  
Condition As-Received : Used Item

Calibrated by

Mr. Somphop Duangguan

Calibration Engineer

Approved by

( Ms. Jintana Sangthajarnlap )

Calibration Manager

The reported expanded uncertainty of measurement was based on a combined standard uncertainty multiplied by a coverage factor  $k$  providing a level of confidence of approximately 95%.

This certificate may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory. Indicated values are valid for the state of the Spectrophotometer at the time of calibration only.

Certificate No. : CAL-24-317

Page : 2 of 3

## CALIBRATION REPORT

Conditions of this result of calibration

1. Reference Standard Material :

Material	Model	Serial No.	Cert No.	Due date
Holmium Glass Filter	RM-HG	24563	109211	13 February 2025
Didymium Glass Filter	RM-DG	24562	109212	13 February 2025
Neutral Density Filter	RM-1N2N3N	24568	109249	14 February 2025

2. Traceability : This certification is traceable to the International System of Unit maintained at

The Sterna Scientific Ltd. Accredited Calibration Laboratory No. 0659.

3. Method of calibration :

The calibration procedure was carried out according to ASTM E275-08 (2022) and ASTM E925-09 (2014).

4. Result of calibration :

( ✓ ) without adjustment

( ) after adjustment

5. Equipment Specifications:

Spectral Bandwidth : 5 nm  
Data Interval : 1 nm  
Scan Speed : Slow nm/min





**BECTHAI**  
**BECTHAI BANGKOK EQUIPMENT & CHEMICAL CO., LTD.**  
**CALIBRATION LABORATORY**

99/9 Moo 2, Min Buri, Suwattana, Pathum Thani 12110, Thailand Tel: 06-3424 5099 Fax: 06-3424 5050  
E-mail: bde@becthai.com Website: www.becthai.com

Certificate No. : CAL-24-317

Page : 3 of 3

**CALIBRATION REPORT**

**Wavelength Calibration**

Certified Values of Reference Material	Nominal Value (nm)	UUC*Reading (nm)	Error (nm)	Uncertainty of Measurement ( $\pm$ nm)	k Factor
361.40	362	362	0.60	0.59	2.00
528.59	529	529	0.41	0.59	2.00
879.68	879	879	-0.68	0.59	2.00

**Photometric Calibration for Visible**

Wavelength (nm)	Certified Values of Reference Material (A)	UUC* Reading (A)	Error (A)	Uncertainty of Measurement ( $\pm$ A)	k Factor
420.0	Zero	0.000	0.0000	0.0028	2.00
	0.5835	0.587	0.0035	0.0045	2.00
	0.725	0.726	0.0010	0.0045	2.00
	1.0367	1.036	-0.0007	0.0045	2.00
440.0	Zero	0.000	0.0000	0.0028	2.00
	0.5662	0.569	0.0028	0.0045	2.00
	0.7106	0.710	-0.0006	0.0045	2.00
	1.0159	1.015	-0.0009	0.0045	2.00
465.0	Zero	0.000	0.0000	0.0028	2.00
	0.5257	0.528	0.0023	0.0045	2.00
	0.6682	0.669	0.0008	0.0045	2.00
	0.9547	0.955	0.0003	0.0045	2.00
546.1 (546.0)	Zero	0.000	0.0000	0.0028	2.00
	0.5226	0.524	0.0014	0.0045	2.00
	0.6939	0.694	0.0001	0.0045	2.00
	0.9919	0.991	-0.0009	0.0045	2.00
590.0	Zero	0.000	0.0000	0.0028	2.00
	0.5567	0.558	0.0013	0.0045	2.00
	0.7502	0.750	-0.0002	0.0045	2.00
	1.0732	1.072	-0.0012	0.0045	2.00
635.0	Zero	0.000	0.0000	0.0028	2.00
	0.5643	0.566	0.0017	0.0045	2.00
	0.7299	0.730	0.0001	0.0045	2.00
	1.0437	1.044	0.0003	0.0045	2.00

Remark : Each individual filter is measured against the empty filter holder (blank) used to zero the Spectrophotometer.

Note:

UUC\* : Unit Under Calibration

ISSUE: 6 REV:5

FM-CAL-33/2

- End of Report  
เลขาฯ เป็น 7-244  
20/02/24





#### Instrument information

Name **WET BULB GLOBE TEMPERATURE (WBGT) METER**  
Series No **3522406014**  
Type **JT2011-E2A**

#### Integrity check of instrument

Appearance ✓  
Parts integrity ✓  
Screen display or touch ✓  
Instrument button ✓  
Power supply ✓  
battery ✓  
Data storage and export ✓  
Deviation degree of comparison test with standard instrument ✓

#### Calibration Results

UUC Sensor	Standard Temperature (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (±°C)
WET	25.0	25.1	-0.1	0.2
	30.0	29.9	0.1	0.2
	35.0	34.9	0.1	0.2
	40.0	40.2	-0.2	0.2
DRY	45.0	44.9	0.1	0.2
	25.0	24.8	0.2	0.2
	30.0	29.8	0.2	0.2
	35.0	35.2	-0.2	0.2
GLOBE	40.0	40.1	-0.1	0.2
	45.0	44.8	0.2	0.2
	25.0	24.9	0.1	0.2
	30.0	30.1	-0.1	0.2
	35.0	34.8	0.2	0.2
	40.0	39.8	0.2	0.2
	45.0	44.9	0.1	0.2

Environmental conditions: temperature: 26 °C±2°C, relative humidity: 30% RH±10RH%

Reference Standard : Standard Mercury Thermometers, Manufacturer: BGRI, Model: STA, SN: 2-56,

Calibrated Date: 30 March 2021, Calibration Certificate No.: RA21H-AB1000009

This Certificate is traceable to NCMT North China, Certificate No.: RA20J-AK000073

Calibration Engineer:

Date:



## Factory Calibration Certificate

BEIJING J.T TECHNOLOGY CO.,LTD  
www.bjtitec.com  
www.janitech.com



#### Instrument information

Name **WET BULB GLOBE TEMPERATURE (WBGT) METER**  
Series No **3522406015**  
Type **JT2011-E2A**

#### Integrity check of instrument

Appearance ✓  
Parts integrity ✓  
Screen display or touch ✓  
Instrument button ✓  
Power supply ✓  
battery ✓  
Data storage and export ✓  
Deviation degree of comparison test with standard instrument ✓

#### Calibration Results

UUC Sensor	Standard Temperature (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (±°C)
WET	25.0	25.1	-0.1	0.2
	30.0	30.1	-0.1	0.2
	35.0	34.8	0.2	0.2
	40.0	40.1	-0.1	0.2
DRY	45.0	45.1	-0.1	0.2
	25.0	24.8	0.2	0.2
	30.0	29.9	0.1	0.2
	35.0	35.1	-0.1	0.2
GLOBE	40.0	40.2	-0.2	0.2
	45.0	44.9	0.1	0.2
	25.0	25.2	-0.2	0.2
	30.0	29.8	0.2	0.2
	35.0	34.8	0.2	0.2
	40.0	40.1	-0.1	0.2
	45.0	44.8	0.2	0.2

Environmental conditions: temperature: 26 °C±2°C, relative humidity: 30% RH±10RH%

Reference Standard : Standard Mercury Thermometers, Manufacturer: BGRI, Model: STA, SN: 2-56,

Calibrated Date: 30 March 2021, Calibration Certificate No.: RA21H-AB1000009

This Certificate is traceable to NCMT North China, Certificate No.: RA20J-AK000073

Calibration Engineer:

Date:





39/0867  
 39/0867

## 112

[illegible]

2019-08-20  
 2019-08-20  
 2019-08-20  
 2019-08-20

PA4.5L.MTC.001 Rev.5

1994年11月17日 星期四

252256 : 07-28-1991 08:26:33  
 12130  
 252257 : 08-06-1991 08:26:33  
 252258 : 08-06-1991 08:26:33

Figure 1: A schematic diagram of a vertical column of cells. The column is divided into three main sections. The top section is labeled 'APIC' and contains a cell with a nucleus and a flagellum. The middle section is labeled 'STEM' and contains a cell with a nucleus. The bottom section is labeled 'BASAL' and contains a cell with a nucleus. The column is flanked by two vertical lines, one on the left and one on the right. The left line is labeled 'L' and the right line is labeled 'R'. The column is also labeled with 'APIC', 'STEM', and 'BASAL' at the top, middle, and bottom respectively.

2014年12月10日

2019-05-21 10:00:00





**ADVANTAGE CENTER CO., LTD.**  
58/494 M.6, Frakham Road, Tukhrot, Lumiookkar, Pathumthani 12130 Thailand  
Tel. (66-2) 9873248-50 Fax: (66-2) 9873252 E-mail: info.act12662@gmail.com  
pornsak2008@yahoo.co.th

Certificate No.: RA-2408011-1

## Reference Standards

Equipment Name	Serial No.	Certificate No.	Due Date	Traceability to
Sound Calibrator	170603302	EEL BP 108/1066	Oct 30, 2024	TISTR

### Traceability

This calibration is traceable to the International System of Unit via :  
 - TISTR : Thailand Institute of Scientific and Technological Research

[illegible]

**ADVANTAGE CENTER CO., LTD.**  
58/564 56/6, Frachtan Road, I.Kushe, Lumpini, Pathumthani 12130 Thailand.  
Tel. (66-2) 9873248-50 Fax (66-2) 9873252 E-mail: advco.1234@gmail.com  
pmsk2008@yahoo.co.th

# CALIBRATION LABORATORY

Certificate No. RA-2408011-1

Job No. RA-2408011

# Certificate of Calibration

FOR

Equipment Name : Sound Level Meter  
Manufacturer : Sencel Tech  
Model : ST-21D  
Serial Number : SN-20402  
Customer Code : NO.21  
Location of Calibration : In Lab  
Calibration Procedure : CPE-04-01  
Received Date : Jun 14, 2024  
Calibration Date : Jun 15, 2024  
Recommended Due Date : N/A  
Customer Name : Virsham Environmental Service Co., Ltd.

CONDITION AS RECEIVED: Normal

### Environmental Conditions

Ambient Temperature : (25 ± 2)°C

[illegible]

Relative Humidity: 50 to 100% RH

No. 10. Investment of \$500,000 at 4% interest for 10 years. The interest is to be paid quarterly.

**NEW! A NEW APPROACH TO WORKING CAPITAL MANAGEMENT**

SI-Units. The calibration procedure documented is intended to implement the requirements of ISO/IEC 17025:2017

3. The working standard is indicated in page 2 of this certificate.

4. This report applies to the item calibrated and shall not be reproduced except in full, without written approval by

Calibration Laboratory, Advantage Center Co., Ltd.



California and Nevada. U. S. Census.

1870

(Pernak Sukawati)

Laboratory Management

Date of Issue: Aug 13, 2024

Page 1 of 3

[illegible]





# ADVANTAGE CENTER CO., LTD.

59/494 M.6, Prakhom Road, T.Kuchot, Lumphini, Pathumthani 12130 Thailand.  
Tel: (66-2) 9873248-50 Fax: (66-2) 9873252 E-mail: info.accl2662@gmail.com  
pomsak2008@yahoo.co.th

Certificate No.: RA-2408011-1

## Result of Calibration

Calibration Range : 94 dB, 114 dB

Function : Measurement @ 1 kHz

Select A Fast response

STD	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement ( $\pm$ dB)
Setting			
94.56 dB	94.5	0.06	0.88
114.52 dB	114.5	0.02	0.88

Select A Slow response

STD	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement ( $\pm$ dB)
Setting			
94.56 dB	94.3	0.16	0.88
114.52 dB	114.5	0.02	0.88

Select C Fast response

STD	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement ( $\pm$ dB)
Setting			
94.56 dB	94.5	0.06	0.88
114.52 dB	114.5	0.02	0.88

Select C Slow response

STD	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement ( $\pm$ dB)
Setting			
94.56 dB	94.5	0.06	0.88
114.52 dB	114.5	0.02	0.88

STD = Standard

UUC = Unit Under Calibration

- End of Certificate -







**ADVANTAGE CENTER CO., LTD.**  
59/494 M6, Frakham Road, Tukhoh, Lumchokkar, Pathumthani 12130 Thailand  
Tel: (66-2) 9873248-50 Fax: (66-2) 9873252 E-mail: info.accl2652@gmail.com  
pomsak2008@yahoo.co.th


Certificate No.: RA-2408011-2

### Reference Standards

Equipment Name	Serial No.	Certificate No.	Due Date	Traceability to
Sound Calibrator	170603302	EEL BP 1081066	Oct 30, 2024	TISTR

### Traceability

This calibration is traceable to the International System of Unit via :  
- TISTR : Thailand Institute of Scientific and Technological Research



**ADVANTAGE CENTER CO., LTD.**  
59/494 M6, Frakham Road, Tukhoh, Lumchokkar, Pathumthani 12130 Thailand  
Tel: (66-2) 9873248-50 Fax: (66-2) 9873252 E-mail: info.accl2652@gmail.com  
pomsak2008@yahoo.co.th

## CALIBRATION LABORATORY

Certificate No. RA-2408011-2  
Job No. RA-2408011

# Certificate of Calibration

FOR

Equipment Name : Sound Level Meter  
Manufacturer : Sanyo Tech  
Model : ST-21D  
Serial Number : SN 820471  
Customer Code : NO.22  
Location of Calibration : In Lab  
Customer Name : Viradish Environmental Service Co., Ltd.  
149/119 Moo 2, Krung Non-Chong Pathom Rd., Mueangsat, Bang Krue, Nonthaburi 11159

Calibration Procedure : CPE-04-01  
Received Date : Jun 14, 2024  
Calibration Date : Jun 15, 2024  
Recommended Due Date : N/A

### CONDITION AS RECEIVED : Normal

### Environmental Conditions

Ambient Temperature : (25 ± 2) °C

### Result

Relative Humidity : (50 ± 15) %RH

1. No Adjustment (See data attached in page 3 to the end of certificate)

1. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95%.

2. The Unit Under Calibration (UUC) has been calibrated by using the working standard which is traceable to SI-Units. The calibration procedure documented is intended to implement the requirements of ISO IEC 17025 : 2017

3. The working standard is indicated in page 2 of this certificate.

4. This report applies to the item calibrated and shall not be reproduced except in full, without written approval by Calibration Laboratory, Advantage Center Co., Ltd.



Calibrated by : V. Surasak

Approved by :

( Pomsak Sukasawang )

Laboratory Management

Date of Issue : Aug 13, 2024





# ADVANTAGE CENTER CO., LTD.

59/494 M.5, Frakham Road, T.Kukhok, Lumlokkor, Pathumthani 12130 Thailand.  
Tel. (66-2) 9873248-50 Fax: (66-2) 9873252 E-mail: info@accl2662@gmail.com  
pornsak2008@yahoo.co.th

Certificate No. : RA-2408011-2

## Result of Calibration

Calibration Range : 94 dB, 114 dB

Function : Measurement @ 1 kHz

Select A Fast response

STD	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement ( ± dB)
Setting			
94.56 dB	94.6	-0.04	0.88
114.52 dB	114.5	0.02	0.88

Select A Slow response

STD	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement ( ± dB)
Setting			
94.56 dB	94.5	0.06	0.88
114.52 dB	114.5	0.02	0.88

Select C Fast response

STD	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement ( ± dB)
Setting			
94.56 dB	94.5	0.06	0.88
114.52 dB	114.5	0.02	0.88

Select C Slow response

STD	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement ( ± dB)
Setting			
94.56 dB	94.6	-0.04	0.88
114.52 dB	114.5	0.02	0.88

STD = Standard

UUC = Unit Under Calibration

End of Certificate







**ADVANTAGE CENTER CO., LTD.**  
59/494 M.6, Fraikham Road, Tuksoth, Lumlokkan, Pathumthani 12130 Thailand  
Tel: (66-2) 9873248-50 Fax: (66-2) 9873252 E-mail: info.accl12662@gmail.com  
pornsak2008@yahoo.co.th


Certificate No: RA-2408011-3

### Reference Standards

Equipment Name	Serial No.	Certificate No.	Due Date	Traceability to
Sound Calibrator	170603302	EHL BP.108/1066	Oct 30, 2024	TISTR

### Traceability

This calibration is traceable to the International System of Unit via:  
TISTR : Thailand Institute of Scientific and Technological Research



**ADVANTAGE CENTER CO., LTD.**  
59/494 M.6, Fraikham Road, Tuksoth, Lumlokkan, Pathumthani 12130 Thailand  
Tel: (66-2) 9873248-50 Fax: (66-2) 9873252 E-mail: info.accl12662@gmail.com  
pornsak2008@yahoo.co.th

## CALIBRATION LABORATORY

Certificate No. RA-2408011-3  
Job No. RA-2408011

# Certificate of Calibration

FOR

Equipment Name : Sound Level Meter  
Manufacturer : Scaier Tech  
Model : SF-21D  
Serial Number : SN R20472  
Customer Code : NO-23  
Location of Calibration : In Lab  
Customer Name : Yonson Environmental Service Co., Ltd.  
149/119 Moo 2, Krung Non-Chong Thangon Rd., Mahasarakham, Bangkok 17150

**CONDITION AS RECEIVED : Normal**  
**Environmental Conditions**  
Ambient Temperature : (25 ± 2) °C  
Relative Humidity : (50 ± 15) %RH  
**Result**  
1. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95%.  
2. The Unit Under Calibration (UUC) has been calibrated by using the working standard which is traceable to SI-Units. The calibration procedure documented is intended to implement the requirements of ISO/IEC 17025:2017.  
3. The working standard is indicated in page 2 of this certificate.  
4. This report applies to the item calibrated and shall not be reproduced except in full, without written approval by Calibration Laboratory, Advantage Center Co., Ltd.



Calibrated by : V. Surasak  
Approved by : (Pornsak Suk-awong) Laboratory Management

Date of Issue : Aug 13, 2024  
Page 1 of 3





# ADVANTAGE CENTER CO., LTD.

59/494 M.6, Prakhom Road, T.Kukhot, Lumphini, Pathumthani 12130 Thailand.  
Tel. (66-2) 9873248-50 Fax: (66-2) 9873252 E-mail: info.accl2682@gmail.com  
pornsak2008@yahoo.co.th

Certificate No. : RA-2408011-3

## Result of Calibration

Calibration Range : 94 dB, 114 dB Function : Measurement @ 1 kHz

Select A Fast response

STD Setting	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement (± dB)
94.56 dB	94.5	0.06	0.88
114.52 dB	114.6	-0.08	0.88

Select A Slow response

STD Setting	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement (± dB)
94.56 dB	94.5	0.06	0.88
114.52 dB	114.6	-0.08	0.88

Select C Fast response

STD Setting	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement (± dB)
94.56 dB	94.5	0.06	0.88
114.52 dB	114.5	0.02	0.88

Select C Slow response

STD Setting	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement (± dB)
94.56 dB	94.5	0.06	0.88
114.52 dB	114.5	0.02	0.88

STD = Standard

UUC = Unit Under Calibration

End of Certificate







## ADVANTAGE CENTER CO., LTD.

59/494 M6, Frakham Road, T.Kukhot, Lumbookkar, Pathumthani 12130 Thailand.  
Tel: (66-2) 9873248-50 Fax: (66-2) 9873252 E-mail: info.accl@advantage.co.th  
pornsak2008@yahoo.co.th


Certificate No.: RA-2408011-4

### Reference Standards

Equipment Name	Serial No.	Certificate No.	Due Date	Traceability to
Sound Calibrator	170603302	EEL BP 108 1966	Oct 30, 2024	TISTR

### Traceability

This calibration is traceable to the International System of Unit via:  
- TISTR : Thailand Institute of Scientific and Technological Research



## ADVANTAGE CENTER CO., LTD.

59/494 M6, Frakham Road, T.Kukhot, Lumbookkar, Pathumthani 12130 Thailand.  
Tel: (66-2) 9873248-50 Fax: (66-2) 9873252 E-mail: info.accl@advantage.co.th  
pornsak2008@yahoo.co.th

## CALIBRATION LABORATORY

Certificate No. RA-2408011-4

Job No. RA-2408011

# Certificate of Calibration

FOR

Equipment Name : Sound Level Meter

Manufacturer : Scalet Tech

Model : ST-21D

Serial Number : SN820473

Customer Code : NO.24

Location of Calibration : In Lab

Customer Name : Viridian Environmental Service Co., Ltd.

149/119 Moo 2, Krung Non-Chong Thaporn Rd., Mahasarakham, Bang Kruei, Nonthaburi 11130

Calibration Procedure : CIP-04-01

Received Date : Jun 14, 2024

Calibration Date : Jun 15, 2024

Recommended Due Date : N/A

### CONDITION AS RECEIVED : Normal

#### Environmental Condition

Ambient Temperature :  $25 \pm 2^\circ\text{C}$

Relative Humidity :  $(50 \pm 5)\%$  RH

Result : No Adjustment (See data attached in page 3 to the end of certificate)

1. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95%.

2. The Unit Under Calibration (UUC) has been calibrated by using the working standard which is traceable to SI-Unit. The calibration procedure documented is intended to implement the requirements of ISO/IEC 17025 : 2017

3. The working standard is indicated in page 2 of this certificate.

4. This report applies to the item calibrated and shall not be reproduced except in full, without written approval by Calibration Laboratory, Advantage Center Co., Ltd.

Calibration Laboratory, Advantage Center Co., Ltd.



Calibrated by : V. Surasak

Approved by :

(Pornsak Sukswang)

Date of Issue : Aug 13, 2024

Laboratory Management





# ADVANTAGE CENTER CO., LTD.

59/494 M.6, Frakham Road, Tukhok, Lumlokkar, Pathumthani 12130 Thailand  
Tel: (66-2) 9873248-50 Fax: (66-2) 9873252 E-mail: info.accl2662@gmail.com  
pornsak2008@yahoo.co.th

Certificate No. : RA-2408011-4

## Result of Calibration

Calibration Range : 94 dB, 114 dB Function : Measurement @ 1 kHz

Select A Fast response

STD Setting	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement (± dB)
94.56 dB	94.5	0.06	0.88
114.52 dB	114.6	-0.08	0.88

Select A Slow response

STD Setting	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement (± dB)
94.56 dB	94.5	0.06	0.88
114.52 dB	114.5	0.02	0.88

Select C Fast response

STD Setting	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement (± dB)
94.56 dB	94.5	0.06	0.88
114.52 dB	114.6	-0.08	0.88

Select C Slow response

STD Setting	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement (± dB)
94.56 dB	94.5	0.06	0.88
114.52 dB	114.5	0.02	0.88

STD = Standard

UUC = Unit Under Calibration

- End of Certificate -







## ADVANTAGE CENTER CO., LTD.

59/494 M6, Frakham Road, Tukhrot, Lumlookkar, Pathumthani 12130 Thailand.  
Tel: (66-2) 9873248-50 Fax: (66-2) 9873252 E-mail: info.accl2662@gmail.com  
pomsal2008@yahoo.co.th

Certificate No.: RA-2408011-6

### Reference Standards

Equipment Name	Serial No.	Certificate No.	Due Date	Traceability to
Sound Calibrator	170603302	EEL BP-1081066	Oct 30, 2024	TISTR

### Traceability

This calibration is traceable to the International System of Unit via:  
- TISTR : Thailand Institute of Scientific and Technological Research



## ADVANTAGE CENTER CO., LTD.

59/494 M6, Frakham Road, Tukhrot, Lumlookkar, Pathumthani 12130 Thailand.  
Tel: (66-2) 9873248-50 Fax: (66-2) 9873252 E-mail: info.accl2662@gmail.com  
pomsal2008@yahoo.co.th

## CALIBRATION LABORATORY

Certificate No. RA-2408011-6

Job No. RA-2408011

# Certificate of Calibration

FOR

Equipment Name : Sound Level Meter  
Manufacturer : Scalet Tech  
Model : ST-21D  
Serial Number : SN-K0475  
Customer Code : NO.26  
Location of Calibration : In Lab  
Customer Name : Viridian Environmental Service Co., Ltd.  
489/119 Moo 2, Kung Vor-Chong Thanon Rd, Mahasarak, Bang Krury, Nonthaburi 11190

Calibration Procedure : CPE-04-01

Received Date : Jun 14, 2024

Calibration Date : Jun 15, 2024

Recommended Due Date : N/A

### CONDITION AS RECEIVED : Normal

### Environmental Conditions

Ambient Temperature : (25 ± 2) °C

Relative Humidity : (50 ± 15) %RH

### Result

- The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95%.
- The Unit Under Calibration (UUC) has been calibrated by using the working standard which is traceable to SI-Units. The calibration procedure documented is intended to implement the requirements of ISO/IEC 17025 : 2017.
- The working standard is indicated in page 2 of this certificate.
- This report applies to the item calibrated and shall not be reproduced except in full, without written approval by Calibration Laboratory, Advantage Center Co., Ltd.



Calibrated by : V. Surasak

Approved by :

(Pomsak Sudaewasing)

Laboratory Management

Date of Issue : Aug 13, 2024





# ADVANTAGE CENTER CO., LTD.

59/494 M6, Frakham Road, Tukhet, Lumlokkar, Pathumthani 12130 Thailand.  
Tel. (66-2) 9873248-50 Fax (66-2) 9873252 E-mail: info.accl2662@gmail.com  
pomsak2008@yahoo.co.th

Certificate No. : RA-2408011-6

## Result of Calibration

Calibration Range : 94 dB, 114 dB Function : Measurement @ 1 kHz

Select A Fast response

STD	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement (± dB)
Setting			
94.56 dB	94.5	0.06	0.88
114.52 dB	114.5	0.02	0.88

Select A Slow response

STD	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement (± dB)
Setting			
94.56 dB	94.5	0.06	0.88
114.52 dB	114.5	0.02	0.88

Select C Fast response

STD	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement (± dB)
Setting			
94.56 dB	94.5	0.06	0.88
114.52 dB	114.5	0.02	0.88

Select C Slow response

STD	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement (± dB)
Setting			
94.56 dB	94.5	0.06	0.88
114.52 dB	114.5	0.02	0.88

STD = Standard

UUC = Unit Under Calibration

- End of Certificate -






Certificate No.: RA-2408011-7

### Reference Standards

Equipment Name	Serial No.	Certificate No.	Due Date	Traceability to
Sound Calibrator	170603302	EEL BP 1081066	Oct 30, 2024	TISTR

### Traceability

This calibration is traceable to the International System of Unit via :  
TISTR : Thailand Institute of Scientific and Technological Research



# CALIBRATION LABORATORY

Certificate No. RA-2408011-7

Job No. RA-2408011

# Certificate of Calibration

**FOR**

Equipment Name : Sound Level Meter

Manufacturer : Scario Tech

Model: ST-21D

Serial Number : SN 820476

Customer Code : NO 77

Location of Calibration: In Lab

**CANON**

149,119 Map 2, Kuang Nan-Cheng, Taiwan R.O.C. Map 1874, Hsing Kwei, Nankai-shan 11170

CONDITION 48 RECEIVED - Normal

### Environmental Conditions

Ambient Temperature:  $(25 \pm 2)^{\circ}\text{C}$ 

of confidence

1. The reported expanded uncertainty is based on a constant measurement wavelength.

providing a level of confidence of approximately 0.65/

The First Under Construction At 11000 Ave. L, Suite 100, Houston, TX 77036

[illegible]

Figure 1. The effect of the concentration of the *Agrobacterium* strain on the transformation efficiency of *Agrobacterium* strain 104. The concentration of the *Agrobacterium* strain 104 was varied from 10<sup>6</sup> to 10<sup>9</sup> cells/ml. The transformation efficiency was determined by the number of transformants per 10<sup>6</sup> cells of the *Agrobacterium* strain 104. The data are the mean  $\pm$  SD of three independent experiments.

Calibration Laboratory, Advantage Center Co., Ltd.

ACCL

Calibrated by: V. Swasek

Approved by:

(Pomník Subwacze)

Laboratory Management

Date of Issue : Aug 13, 2024

Page 1 of 3

[illegible]





# ADVANTAGE CENTER CO., LTD.

59/494 M.6, Prakhom Road, T.Kukhet Lumlokkor, Pathumthani 12130 Thailand  
Tel: (66-2) 9873248-50 Fax: (66-2) 9873252 E-mail: info.aaccl2662@gmail.com  
pomsak2008@yahoo.co.th

Certificate No. : RA-2408011-7

## Result of Calibration

Calibration Range : 94 dB, 114 dB Function : Measurement @ 1 kHz

Select A Fast response

STD	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement (± dB)
Setting			
94.56 dB	94.5	0.06	0.88
114.52 dB	114.5	0.02	0.88

Select A Slow response

STD	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement (± dB)
Setting			
94.56 dB	94.6	-0.04	0.88
114.52 dB	114.6	-0.08	0.88

Select C Fast response

STD	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement (± dB)
Setting			
94.56 dB	94.5	0.06	0.88
114.52 dB	114.5	0.02	0.88

Select C Slow response

STD	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement (± dB)
Setting			
94.56 dB	94.6	-0.04	0.88
114.52 dB	114.6	-0.08	0.88

STD = Standard

UUC = Unit Under Calibration

- End of Certificate -









## Result of Calibration

Select A Fast response

STD Setting	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement ( $\pm$ dB)
94.42 dB	94.4	0.02	0.88
114.32 dB	114.3	0.02	0.88

Select A Slow response

STD Setting	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement ( $\pm$ dB)
94.42 dB	94.3	0.12	0.88
114.32 dB	114.4	-0.08	0.88

Select C Fast response

STD Setting	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement ( $\pm$ dB)
94.42 dB	94.3	0.12	0.88
114.32 dB	114.4	-0.08	0.88

Select C Slow response

STD Setting	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement ( $\pm$ dB)
94.42 dB	94.3	0.12	0.88
114.32 dB	114.3	0.02	0.88

STD = Standard

UUC = Unit Under Calibration

- End of Certificate -





# ADVANTAGE CENTER CO., LTD.

59/494 M.6, Frakham Road, T.Kukhot, Lumlookkar, Pathumthani 12130 Thailand.  
Tel: (66-2) 9873248-50 Fax: (66-2) 9873252 E-mail: info.accl2862@gmail.com  
pornsak2008@yahoo.co.th

Certificate No.: RA-2412024-2

## Reference Standards

Equipment Name	Serial No.	Certificate No.	Due Date	Traceability to
Sound Calibrator	170603302	EEL BP. 32/1167	Nov 14, 2025	TISTR

### Traceability

This calibration is traceable to the International System of Unit via :

- TISTR : Thailand Institute of Scientific and Technological Research



The equipment is traceable to the International System of Unit via :  
- TISTR : Thailand Institute of Scientific and Technological Research



# ADVANTAGE CENTER CO., LTD.

59/494 M.6, Frakham Road, T.Kukhot, Lumlookkar, Pathumthani 12130 Thailand.  
Tel: (66-2) 9873248-50 Fax: (66-2) 9873252 E-mail: info.accl2862@gmail.com  
pornsak2008@yahoo.co.th

## CALIBRATION LABORATORY

Certificate No. RA-2412024-2

# Certificate of Calibration

## FOR

Equipment Name : Sound Level Meter

Manufacturer : Scalet Tech

Model : ST11D

Serial Number : 820862

Customer Code : SLM-29

Location of Calibration : In Lab

Customer Name : Vinidian Environmental Service Co., Ltd.

149/119 Moo 2, Krung Non-Chong Thanom Rd., Melhasawat, Bang Kruiay, Nonthaburi 11130

Calibration Procedure : CPE-04-01

Received Date : Dec 13, 2024

Calibration Date : Dec 17, 2024

Recommended Due Date : N/A

### CONDITION AS RECEIVED : Normal

#### Environmental Conditions

Ambient Temperature : (25 ± 2) °C

Result : No Adjustment (See data attached in page 3 to the end of certificate)

Relative Humidity : (50 ± 15) %RH

1. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k = 2.

providing a level of confidence of approximately 95%.

2. The Unit Under Calibration (UUC) has been calibrated by using the working standard which is traceable to

SI-Units. The calibration procedure documented is intended to implement the requirements of ISO/IEC 17025:2017

3. The working standard is indicated in page 2 of this certificate.

4. This report applies to the item calibrated and shall not be reproduced except in full, without written approval by

Calibration Laboratory, Advantage Center Co., Ltd.



Calibrated by : P. Noppadon

Approved by :

(Pornsak Suksaewang)

Laboratory Management

Date of Issue : Dec 18, 2024

The equipment is traceable to the International System of Unit via :  
- TISTR : Thailand Institute of Scientific and Technological Research





## ADVANTAGE CENTER CO., LTD.

59/494 M.6, Frakham Road, T.Kukhot, Lumlookkar, Pathumthani 12130 Thailand.  
Tel. (66-2) 9873248-50 Fax: (66-2) 9873252 E-mail: info.accl2662@gmail.com  
pornsak2008@yahoo.co.th

Certificate No. : RA-2412024-2

### Result of Calibration

Calibration Range : 94 dB, 114 dB

Function : Measurement @ 1 kHz

Select A Fast response

STD Setting	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement ( ± dB)
94.42 dB	94.4	0.02	0.88
114.32 dB	114.3	0.02	0.88

Select A Slow response

STD Setting	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement ( ± dB)
94.42 dB	94.4	0.02	0.88
114.32 dB	114.3	0.02	0.88

Select C Fast response

STD Setting	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement ( ± dB)
94.42 dB	94.3	0.12	0.88
114.32 dB	114.3	0.02	0.88

Select C Slow response

STD Setting	UUC Reading (dB)	Correction (dB)	Uncertainty of Measurement ( ± dB)
94.42 dB	94.3	0.12	0.88
114.32 dB	114.3	0.02	0.88

STD = Standard

UUC = Unit Under Calibration

- End of Certificate -






SCARLET TECH



## CERTIFICATE OF CALIBRATION

NO. 20250110167

Name of Product:	Sound Level Meter
Model:	ST-21D
Serial Number:	821061
Specification:	Class 2
Conclusion:	Pass
Date of calibration:	2024-12-26
Due Date:	2025-12-25

Calibrated by:   
SCARLET TECH CO., LTD.  
SCARLET TECH  
SCARLET TECH CO., LTD.

- I. This report certifies that all calibration equipment used in the test is traceable with the internal ISO9001 procedures and meets all specification given in the Manual(s) or respectively surpass them, and applies only to the unit identified above.  
II. This certificate is produced with advanced equipment & procedures which permit comprehensive quality assurance verification of all data supplied herein.  
III. This certificate of calibration shall not be reproduced except in full, without written permission of the Scarlet Tech Co Ltd Taiwan.

1. Preliminary inspection: OK  
2. Type & serial No. of Microphone: AWA14421A-000247  
3. Adjustments to indicated sound levels:  
Type of Calibrator B&K 4231  
Sound Pressure Level 94.0 dB  
Equivalent Free-field Sound Level (reference environment conditions) 93.8 dB  
4. Measuring up limit: 139 dBA  
5. Frequency weightings (Acoustic signal tests for Z weighting, other electric signal tests.)

Nominal frequency /Hz	Frequency weighting / dB			Nominal frequency /Hz	Frequency weighting / dB		
	A	C	Z		A	C	Z
20	-50.4	-6.1	-0.1	1000	0.1	0.0	0.0
31.5	-39.4	-3.0	-0.1	2000	1.3	-0.1	0.0
63	-26.2	-0.8	0.0	4000	1.3	-0.6	0.0
125	-16.2	-0.2	0.0	8000	-1.2	-3.1	0.1
250	-8.6	0.0	0.0	12500	-11.0	-13.0	0.0
500	-3.2	0.0	0.0	/	/	/	/

6. Self-generated noise  
Microphone replaced by electrical input signal device

25 dB(A)	27.5 dB(C)	36.1 dB(Z)
----------	------------	------------

### 7. F&S Weighting

Rate of the F weighting decrease (dB/s)	34.4
Rate of the S weighting decrease (dB/s)	4.2
Deviation of F&S	-0.1

### 8. Level Linearity (A-weighting at frequency 1 kHz)

Reference sound level 90.0 dB  
Max error at 10dB steps upper reference sound level 0.1 dB  
Max error at 1dB steps within 5dB of the upper limit linear operating range 0.0 dB  
Max error at 10dB steps below reference sound level 0.1 dB  
Max error at 1dB steps within 5dB upper the lower limit linear operating range 0.1 dB

### 9. Tone burst response (A Weighting) :

Single Toneburst duration /ms	Toneburst response /dB			
	LAF-mdB	LASmax-LA	LAE-LA	LAEqLA
500	0.0	-4.0	-2.9	-7.0
200	-1.0	-7.4	-6.9	-7.0
2	-18.1	-26.9	-26.9	-7.0
0.25	-27.1	/	-36.0	-7.0

### 10. Peak C sound level (500Hz) :

Cycle	One cycle	nominal value	Positive half	nominal value	Negative half	nominal value
LCpeak-LC(dB)	3.4	3.5	2.3	2.4	2.3	2.4

### 11. Overload indication: Pass

### 12. Statistical analysis function

Sweep signal maximum indicated sound level: 123.0 dB  
Sweep amplitude: 40 dB  
Scan cycle time: 60 S; Measurement period: 180 S.

Items	Measured value/dB	Theoretical calculated value/dB	Error/dB
LAeq,T	113.3	113.3	-0.1
L5	121.0	121.0	0.0
L10	119.0	119.0	0.0
L50	103.0	103.0	0.0
L90	87.1	87.0	0.1
L95	85.1	85.0	0.1

Uncertainty of measurement results: 0.4 dB (k=2)

Environment conditions :

Air temperature : 25 °C  
Relative humidity : 60 %  
Static pressure : 101.8 kPa

Test specifications:

- 1. All Scalet's Sound Level Meter has been calibrated in accordance with the requirements as specified in ISO 17025 and the lab calibration procedure SMP004-CA-152.
- 2. The electrical tests were performed using an electrical signal substituted for the microphone which was removed and replaced by an equivalent capacitance within a tolerance of +20%.
- 3. The acoustic calibration was performed using an B&K 4226 sound calibrator and corrections was applied for the difference between the free-field and pressure responses of the Sound Level Meter.

References :

IEC 61672-3 Sound Level Meters Part 3: Periodic tests



CERTIFICATE OF CALIBRATION

ISSUED BY Cirrus Research

DATE OF ISSUE 23 December 2024 CERTIFICATE NUMBER 229416

Cirrus Research  
Acoustic House  
Bridlington Road  
Hummanby  
North Yorkshire  
YO14 0PH  
United Kingdom

Approved Signatory  
R. Thomas  
Electronically signed:

Page 1 of 2

doseBadge Reader : IEC 60942:2003

Instrument information

Manufacturer: Cirrus Research plc  
Model: RC-110A  
Serial number: 48669  
Class: 2  
Notes: Viridian Environment Service Co., Ltd.  
149/119 Moo 2  
Krung Non-Chong Thanom Rd.  
Mahasawat  
Bang Kruay

Test summary

Date of calibration: 19 December 2024

The doseBadge reader detailed above has been calibrated to the published data as described in the operating manual and in the half-inch configuration. The procedures and techniques used are as described in IEC60942:2003 Annex B – Periodic Tests and three determinations of the sound pressure level, frequency and total distortion were made.

The sound pressure level was measured using a WS2F condenser microphone type MK-224 manufactured by Cirrus Research plc.

The results have been corrected to the reference pressure of 101.33 kPa using the manufacturer's data.

The doseBadge Reader has been shown to conform to the Class 2 requirements for periodic testing, described in Annex B of IEC 60942:2003 for the sound pressure level(s) and frequency(ies) stated, for the environmental conditions under which the tests were performed.

However, as public evidence was not available, from a testing organisation responsible for pattern approval, to demonstrate that the model of doseBadge Reader conformed to the requirements for pattern evaluation described in Annex A of IEC 60942:2003, no general statement or conclusion can be made about conformance of the doseBadge Reader to the requirements of IEC 60942:2003.

Notes:

This certificate provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory. The results within this certificate relate only to the items calibrated. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a coverage probability of approximately 95%.

CERTIFICATE OF CALIBRATION

Certificate Number:  
229416  
Page 2 of 2

Environmental conditions

The following conditions were recorded at the time of the test:

Before Pressure: 99.70 kPa Temperature: 23.4 °C Humidity: 32.1 %  
After Pressure: 99.71 kPa Temperature: 23.5 °C Humidity: 31.3 %

Test equipment

Equipment	Manufacturer	Model	Serial number
Distortion Meter	Keithley	2015	1053426
Environmental Monitor	Comet	T7510	21962628

Initial Acoustic Results

	Expected	Sample 1	Sample 2	Sample 3	Average	Deviation	Tolerance	Uncertainty
Level (dB)	114.00	114.10	114.12	114.11	114.11	0.11	±0.75	0.11 dB
Distortion (%)	< 4.00	0.33	0.34	0.35	0.34	0.34	+4.00	0.13 %
Frequency (Hz)	1000.0	995.4	995.4	995.4	995.4	-4.6	±20.0	0.1 Hz

The measured quantities or deviations (as applicable), extended by the expanded combined uncertainty of measurement, must not exceed the corresponding tolerance.

Adjusted Acoustic Results

	Expected	Sample 1	Sample 2	Sample 3	Average	Deviation	Tolerance	Uncertainty
Level (dB)	114.00	114.01	114.00	114.01	114.01	0.01	±0.75	0.11 dB
Distortion (%)	< 4.00	0.37	0.34	0.36	0.36	0.36	+4.00	0.13 %
Frequency (Hz)	1000.0	995.4	995.4	995.4	995.4	-4.6	±20.0	0.1 Hz

Functionality Results

Function	Result
Keypad	Pass
Battery Power	Pass
Display	Pass
Communication	Pass
2 way IR link	Pass
Clock	Pass

End of results

Test	Result
Absolute Acoustic Sensitivity	Complies
Linearity	Complies
Short Duration	Complies
Overload Latching	Complies
Frequency weighting	Complies



CERTIFICATE OF CALIBRATION

ISSUED BY      Cirrus Research

DATE OF ISSUE    23 December 2024    CERTIFICATE NUMBER 229412



Cirrus Research  
Acoustic House  
Bridlington Road  
Hunmanby  
North Yorkshire  
YO14 0PH  
United Kingdom

Page 1 of 2  
Approved signatory  
R. Thomas  
Electronically signed:

*R. Thomas*

Dosimeter : IEC 61252-1993+A1:2000

Instrument information

Manufacturer:      Cirrus Research plc  
Model:                CR:110A  
Serial number:      CB0452  
Firmware version:   5.4

Notes:      Viridian Environment Service Co. Ltd  
              149/119 Moo 2  
              Krung Non-Chong Thanom Rd.  
              Mahasawat  
              Bang Kruay

Test summary

Date of calibration:    20 December 2024

The calibration was performed respecting the requirements of ISO/IEC 17025:2017.

The dosimeter submitted for testing successfully completed the periodic tests of IEC 61252-1993+A1:2000.

The dosimeter submitted for testing conforms to the specifications in IEC 61252-1993+A1:2000.

Test equipment

Equipment	Manufacturer	Model	Serial number
Signal Generator	KEYSIGHT	33511B	MY58001613
Attenuator	Cirrus Research	ZE:952	64370
Environmental Monitor	Comet	T7310	24792823
doseBadge Reader	Cirrus Research plc	RC:110A	100498

Notes

This certificate provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory. The results within this certificate relate only to the items calibrated. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a coverage probability of approximately 95%.

CERTIFICATE OF CALIBRATION

Certificate Number:  
229412

Page 2 of 2

Environmental conditions

The following conditions were recorded at the time of the test:

Before    Pressure: 100.46 kPa    Temperature: 20.5 °C    Humidity: 31.1 %  
After     Pressure: 100.42 kPa    Temperature: 20.7 °C    Humidity: 30.4 %

Test results summary

Test	Result
Absolute Acoustic Sensitivity	Complies
Linearity	Complies
Short Duration	Complies
Overload Latching	Complies
Frequency weighting	Complies






CERTIFICATE OF CALIBRATION

ISSUED BY      Cirrus Research

DATE OF ISSUE    23 December 2024    CERTIFICATE NUMBER 229415



Cirrus Research  
Acoustic House  
Bridlington Road  
Hummanby  
North Yorkshire  
YO14 0PH  
United Kingdom

Page 1 of 2

Approved signatory  
R. Thomas  
Electronically signed:

Dosimeter : IEC 61252-1993+A1:2000

Instrument information

Manufacturer:      Cirrus Research plc  
Model:              CR 110A  
Serial number:      CB0454  
Firmware version:    5.4

Notes:              Viridian Environment Service Co. Ltd  
                         149/119 Moo 2  
                         Krung Non-Chong Thanom Rd.  
                         Mahasawat  
                         Bang Kruay

Test summary

Date of calibration:    23 December 2024

The calibration was performed respecting the requirements of ISO/IEC 17025:2017.

The dosimeter submitted for testing successfully completed the periodic tests of IEC 61252-1993+A1:2000.

The dosimeter submitted for testing conforms to the specifications in IEC 61252-1993+A1:2000.

Test equipment

Equipment	Manufacturer	Model	Serial number
Signal Generator	KEYSIGHT	33511B	MY68001613
Attenuator	Cirrus Research	ZE-952	64370
Environmental Monitor	Comet	T7310	24792823
doseBadge Reader	Cirrus Research plc	RC-110A	100498

Notes

This certificate provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory. The results within this certificate relate only to the items calibrated. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a coverage probability of approximately 95%.

CERTIFICATE OF CALIBRATION

Certificate Number:  
229415

Page 2 of 2

Environmental conditions

The following conditions were recorded at the time of the test:

Before    Pressure: 101.33 kPa    Temperature: 20.0 °C    Humidity: 27.9 %  
After    Pressure: 101.34 kPa    Temperature: 20.0 °C    Humidity: 25.8 %

Test results summary

Test	Result
Absolute Acoustic Sensitivity	Complies
Linearity	Complies
Short Duration	Complies
Overload Latching	Complies
Frequency weighting	Complies



INTERNATIONAL TESTING SERVICE CO., LTD

1213/388 Ladprao 94 Ladprao Rd. Wangtonglang Bangkok 10310  
Tel 0-2559-2095 Fax 0-2559-2096

E-mail : [sale@itest-lab.com](mailto:sale@itest-lab.com) web site : [www.itest-lab.com](http://www.itest-lab.com)



MSC-7818-7819 17025  
CALIBRATION 128

CALIBRATION CERTIFICATE

Issued date: 4 October 2024

Client Name : VIRIDIAN ENVIRONMENTAL SERVICE CO., LTD.

Address : 149/119 Moo.2, Krungnon-Chong Thanom Road, T.Maha Sawat, A.Bang Kruai,  
Nonthaburi 11130.

Request No: C-2410 - 346

Laboratory No.: CAL - 346

Date of Request: 4 October 2024.

Date of Calibration: 4 October 2024.

1. Unit Under Calibration (UUC) :

Nomenclature : Digital Light Meter

Maker : EXTECH

Serial No. : A 052153

Model : 407026

2. Place of Calibration: Photometry Standard Laboratory, INTERNATIONAL TESTING SERVICE CO., LTD

3. Range of Calibration: 1 Range

4. Condition of Laboratory: Ambient temperature ( $25 \pm 2$ ) °C and relative humidity ( $60 \pm 20$ ) %.

5. Reference Standard: Standard Tungsten Halogen Lamp, Serial No.: 504011, which was calibrated on  
22 April 2024, can be traceable to International System of Unit (SI) through National Institute of  
Metrology (Thailand), Certificate No. TP-1014-24.

6. Support Equipment:

1. Photometric bench, 6.3 meter long.
2. DC, power supply, Serial No.: EJ 19A 009, Model: GPR-25H 300, Maker: GW INSTEK.
3. Digital Multimeter, Model: 34401A, S/N: MY44011212 and MY44011215.
4. Foot Candle / Lux Meter, Model: 407026, S/N: Q 558437, Maker: EXTECH

7. Calibration Procedure:

The measurement was done in accordance with WI-CP-01. The reported uncertainty is based on a standard  
uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95 %.

The Results shown in this certification report refer only to the equipment(s) calibrated. The results are not valid for any other equipment(s) not calibrated. This Calibration Certificate cannot be reproduced, except in full, without permission of company.

Page 2 of 2

INTERNATIONAL TESTING SERVICE CO., LTD



INTERNATIONAL TESTING SERVICE CO., LTD  
1213/388 Ladprao 94 Ladprao Rd. Wangtonglang Bangkok 10310  
Tel 0-2559-2095 Fax 0-2559-2096

E-mail : [sale@itest-lab.com](mailto:sale@itest-lab.com) web site : [www.itest-lab.com](http://www.itest-lab.com)



MSC-7818-7819 17025  
CALIBRATION 128

Request No: C-2410 - 346

Laboratory No.: CAL - 346

Serial No.: A052153

Results :

UUC Range	Standard (lx)	UUC Reading (lx)		Correction (lx)	Uncertainty of Measurement ( $\pm$ lx)
		Before adjust	After adjust		
2000	0	0	0	0	2.1 % of Reading
	100	93	100	0	
	499	464	498	+1	
	998	929	994	+4	
	1496	1391	1490	+6	
	1994	1819	1954	+40	

Note : 1. The results relate only to the items calibrated.  
2. Zero adjust before used.

Calibration result approved by

(Mr. Uttana Tholueg)

Approved on behalf of  
International Testing Service Co., Ltd

(Mr. Pichit Vivat-Anant)  
Managing Director



# INTERNATIONAL TESTING SERVICE CO., LTD

1213/388 Ladprao 94 Ladprao Rd. Wangtonglang Bangkok 10310  
Tel 0-2559-2095 Fax 0-2559-2096  
E-mail : sale@itest-lab.com web site : www.itest-lab.com



## CALIBRATION CERTIFICATE

Issued date: 4 October 2024

Client Name : VIRIDIAN ENVIRONMENTAL SERVICE CO., LTD.

Address : 149/119 Moo.2, Krungnon-Chong Thanom Road, T.Maha Sawat, A.Bang Kruai,

Northaburi 11130.

Request No : C-2410 - 345

Laboratory No : CAL - 345

Date of Request: 4 October 2024.

Date of Calibration: 4 October 2024.

1. Unit Under Calibration (UUC) :

Nomenclature : Digital Light Meter

Maker : EXTECH

Serial No. : A043002

Model : 407026

2. Place of Calibration: Photometry Standard Laboratory, INTERNATIONAL TESTING SERVICE CO., LTD.

3. Range of Calibration: 1 Range

4. Condition of Laboratory: Ambient temperature:  $(25 \pm 2) ^\circ\text{C}$  and relative humidity  $(60 \pm 20) \%$ .

5. Reference Standard: Standard Tungsten Halogen Lamp, Serial No.: 504011, which was calibrated on 22 April 2024, can be traceable to International System of Unit (SI) through National Institute of Metrology (Thailand), Certificate No.: TP-1014-24.

6. Support Equipment:

1. Photometric bench, 6.3 meter long.
2. DC, power supply, Serial No.: EJ 19A 009, Model: GPR-25H 300, Maker: GW INSTEK.
3. Digital Multimeter, Model: 34401A, S/N: MY44011212 and MY44011215
4. Foot Candle / Lux Meter, Model: 407026, S/N: Q 558437, Maker: EXTECH.

7. Calibration Procedure

The measurement was done in accordance with WI-CP-01. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95 %.

The Results shown in this certification report refer only to the equipment(s) calibrated by the company. This Calibration Certificate cannot be reproduced, except in full, without permission of the company.

Viridian Environmental Service Co., Ltd.



# INTERNATIONAL TESTING SERVICE CO., LTD

1213/388 Ladprao 94 Ladprao Rd. Wangtonglang Bangkok 10310  
Tel 0-2559-2095 Fax 0-2559-2096  
E-mail : sale@itest-lab.com web site : www.itest-lab.com

ISO 17025:2017 CALIBRATION 175

Request No: C-2410 - 345

Laboratory No: CAL - 345

Serial No.: A043002

Results :

UUC Range	Standard (ix)	UUC Reading (ix)		Correction (ix)	Uncertainty of Measurement ( $\pm$ ix)
		Before adjust	After adjust		
2000	0	0	0	0	2.1 % of Reading
	100	99	102	-2	
	500	494	498	+2	
	999	986	996	+3	
	1498	1474	1494	+4	
	1996	1959	1992	+4	

Note : 1. The results relate only to the items calibrated.  
2. Zero adjust before used.

Calibration result approved by

Approved on behalf of  
International Testing Service Co., Ltd

(Mr. Yuktana Tholueung)  
(Mr. Pichit Vivat-Anant)  
Managing Director