


ภาคผนวก จ  
เอกสารสอบเทียบเครื่องมือ  
ที่ใช้ในการตรวจวิเคราะห์  
(Calibration)



ตารางการสอบเทียบเครื่องมือที่ใช้ในการตรวจวัดและวิเคราะห์

Item	Description	Parameter	List of Equipment	Equipment No.	Calibration	Next Calibration
1.	Ambient Air	TSP	ORIFICE TRANSFER STANDARD/Tisch	S/N 0068	17/08/2023	August 2024
			High Volume Air Sampler/TET	S/N TSP-14	03/07/2024	July 2025
			High Volume Air Sampler/TET	S/N TSP-18	02/07/2024	July 2025
			High Volume Air Sampler/TET	S/N TSP-13	02/07/2024	July 2025
			High Volume Air Sampler/TET	S/N TSP-23	02/07/2024	July 2025
		PM-10	Electronic Balance/METTLER TOLEDO	S/N 1116392227	10/04/2024	April 2025
			ORIFICE TRANSFER STANDARD/Tisch	S/N 0068	17/08/2023	August 2024
			High Volume Air Sampler/TET	S/N PM10-13	01/07/2024	July 2025
			High Volume Air Sampler/TET	S/N PM10-25	04/07/2024	July 2025
			High Volume Air Sampler/TET	S/N PM10-14	01/07/2024	July 2025
		NO <sub>2</sub>	High Volume Air Sampler/TET	S/N PM10-16	02/07/2024	July 2025
			Electronic Balance/METTLER TOLEDO	S/N 1116392227	10/04/2024	April 2025
			Certificate of Analysis/Linde	S/N A00917SK	05/07/2023	July 2026
			NO <sub>x</sub> Analyzer/API TML-41-H-02	S/N 495	03/10/2024	April 2025
			NO <sub>x</sub> Analyzer/API 200A	S/N 777	08/10/2024	April 2025
		SO <sub>2</sub>	NO <sub>x</sub> Analyzer/API 200A	S/N 1978	25/09/2024	March 2025
			NO <sub>x</sub> Analyzer/API 200A	S/N 542	08/10/2024	April 2025
			Certificate of Analysis/Linde	S/N D636157	18/09/2023	September 2027
			SO <sub>x</sub> Analyzer/API 100A	S/N 856	26/10/2024	April 2025
			SO <sub>x</sub> Analyzer/API 100A	S/N 1563	28/09/2024	March 2025
		CO	SO <sub>x</sub> Analyzer/Thermo 43C	S/N 43C57277312	27/09/2024	March 2025
			CERTIFICATE OF ANALYSIS : Linde	S/N D271305	11/10/2016	11/10/2024
			CERTIFICATE OF ANALYSIS : Linde	S/N D824500	11/10/2016	11/10/2024

ตารางการสอบเทียบเครื่องมือที่ใช้ในการตรวจวัดและวิเคราะห์ (ต่อ)

Item	Description	Parameter	List of Equipment	Equipment No.	Calibration	Next Calibration
1.	Ambient Air (Cont.)	CO	CO Analyzer/Teledyne 300E	S/N 1066	02/10/2024	April 2025
			CO Analyzer/Horiba APMA 360CE	S/N 42088-7001	02/10/2024	April 2025
		HC as Methane	Personal Air Sampler/Gilian	S/N 20140505076	08/01/2025	February 2025
			Personal Air Sampler/Gilian	S/N 20140505076	14/02/2025	March 2025
			Methane NIMHC Analyzer/Model 55C	S/N 55C-72555-371	09/01/2025	January 2026
2.	Wastewater	pH	pH Meter/Horiba	S/N B06D0012	30/10/2024	October 2025
		SS	Electronic Balance/METTLER TOLEDO	S/N 1116392227	10/04/2024	April 2025
		TDS	Electronic Balance/METTLER TOLEDO	S/N 1116392227	10/04/2024	April 2025
		BOD	BOD Incubator/Model i250	S/N 0408-0115-0008	09/04/2024	April 2025
		Oil & Grease	Electronic Balance/METTLER TOLEDO	S/N 1116392227	10/04/2024	April 2025
		Sulfide	Spectrophotometer/Blue Star A	S/N 1606UV1507	09/04/2024	April 2025
		Total Coliform Bacteria	Incubator Model INE 500	S/N E.505.0595	09-10/04/2024	April 2025
3.	Sound Level	Leq 24 hr	Sound Level Calibrator/TENMARS TM-100	S/N ST120C0263E	12/10/2024	October 2025
			Integrated Sound Level/ACO TYPE 6226	S/N 110102	02/01/2025	31/01/2025
			Integrated Sound Level/ACO TYPE 6226	S/N 100101	02/01/2025	31/01/2025
			Integrated Sound Level/ACO TYPE 6226	S/N 160099	31/01/2025	28/02/2025
			Integrated Sound Level/ACO TYPE 6226	S/N 160212	31/01/2025	28/02/2025
4.	Vibration	Vibration	Vibration Meter/Instantel	S/N UM16046	12/02/2024	February 2025
			Vibration Meter/Instantel	S/N UM10831	18/10/2024	October 2025




JIRANATEE ASSOCIATES CO., LTD.

Jiranatee Associates Co., Ltd  
63/14-15, 67/35-36  
Petchkasem 7, 7/1, Rd. Watthapra, Bangkokyai,  
Bangkok 10600 (Thailand)  
Tel: +6608680812  
Mobile: +66863999453  
E-mail: jnac-calibration@jiranatee.com  
Web site: www.jiranatee.com

Accredited calibration laboratory  
ISO/IEC 17025:2017  
NSC-TISI-TIS 17025  
CALIBRATION 0367

Flow measurement laboratory  
Calibration services department.



NSC - TISI - TIS 17025  
CALIBRATION 0367

## CERTIFICATE OF CALIBRATION

Certificate No. : COF-033-67

Page 1 of 2 Pages

MEASUREMENT ITEM	: Top Load Orifice
MANUFACTURER	: TISCH
MODEL/TYPE	: TE-5025A
SERIAL NUMBER	: 0068
ID NUMBER	: -
CONDITION AS-RECEIVED	: Used item
CUSTOMER	: Thai Environmental Technic Limited. 1/6 Soi Ramkhamhaeng 145, Khwaeng/Khet Saphan Sung, Bangkok 10240
RECEIVED DATE	: 18 Jul 2024
MEASUREMENT DATE	: 30 Jul 2024
ISSUE DATE	: 30 Jul 2024

### ENVIRONMENTAL CONDITIONS:

Ambient condition in the laboratory are as follow:

Temperature	: $23.0 \pm 3.0$	°C
Relative Humidity	: $55.0 \pm 15.0$	%RH
Atmospheric Pressure	: $1010 \pm 10$	hPa

### CALIBRATION CONDITION:

Preconditioning	: 24 hours at ambient conditions.
Measurement Condition	: The average values during measurement are $23.2^{\circ}\text{C}$ and $52.0\% \text{RH}$ .

**NOTED:** The certificate is valid only to the item calibrated on date and place of calibration.

### TABULATION OF RESULTS:

The table on next page give the measured values.

### Calibration procedure:

The Orifice gas flow device was calibrated against Standard Rotary Displacement Meter (Roots Meter) Model G65/IMC/W2-dp. The WI-CL-004 was used as a calibration guideline.

### Traceability:

This certificate provides a traceability of the measurement to recognized the national standards, and to realization of the international system of units (SI) through the NIMT (National Metrology Institute of Thailand) via Certificate number: MW-0063-23.

### Uncertainty of Measurement:

The reported uncertainty of measurement is based on the standard uncertainty multiplied by a coverage factor  $k=2$ , Which for a normal distribution corresponds to a coverage probability of approximately 95%. The standard uncertainty has been determined in accordance with the GUM 'Evaluation of measurement data - Guide to the expression of uncertainty in measurement'

Calibrated by:

- ☐ Mr. Sorawit Thachalad  
☒ Miss Jitraporn Lertsomphol



Approved signatory: \_\_\_\_\_

Mr. Parinya Booncharoen  
Calibration Department Manager

**MEASUREMENT RESULTS:**

The Orifice gas flow device was calibrated by direct comparison method with the Standard Rotary Displacement Meter (Roots Meter). The Humid air was used as a medium in the system. The standard conditions are 25°C (298.15 K) and 760 mmHg for standard temperature and standard pressure respectively.

Table 1: The results of  $Q$  Standard calibration data

Plate	Flow rate $m^3/min$	Pressure [Pa] mmHg	Temperature [Ta] °C	Temperature [Tm] °C	$\Delta p_{meter}$ mmHg	$\Delta p_{Orifice}$ inH <sub>2</sub> O	$\gamma$	Standard Flow [ $Q_s$ ] $m^3/min$
1	0.699	755.089	23.18	22.32	50.227	1.674	1.293	0.654
2	1.001	755.019	23.13	22.49	55.196	3.410	1.846	0.930
3	1.120	754.920	23.15	22.58	38.094	4.483	2.117	1.065
4	1.168	754.911	23.24	22.61	27.632	5.018	2.239	1.127
5	1.412	754.933	23.37	22.78	27.237	7.365	2.712	1.362


Slope ( $m$ ): 2.00395  
 Intercept ( $b$ ): -0.01781  
 Correlation coefficient ( $r$ ): 0.99988  
 Uncertainty ( $k=2$ ): 0.015  $m^3/min$

Table 2: The results of  $Q$  actual calibration data

Plate	Flow rate $m^3/min$	Pressure [Pa] mmHg	Temperature [Ta] °C	Temperature [Tm] °C	$\Delta p_{meter}$ mmHg	$\Delta p_{Orifice}$ inH <sub>2</sub> O	$\gamma$	Standard Flow [ $Q_s$ ] $m^3/min$
1	0.699	755.089	23.18	22.32	50.227	1.674	0.810	0.655
2	1.001	755.019	23.13	22.49	55.196	3.410	1.157	0.930
3	1.120	754.920	23.15	22.58	38.094	4.483	1.326	1.066
4	1.168	754.911	23.24	22.61	27.632	5.018	1.404	1.128
5	1.412	754.933	23.37	22.78	27.237	7.365	1.701	1.364

Slope ( $m$ ): 1.25514  
 Intercept ( $b$ ): -0.01115  
 Correlation coefficient ( $r$ ): 0.99988  
 Uncertainty ( $k=2$ ): 0.015  $m^3/min$

\*\*\*End of Certificate of Calibration\*\*\*





Thai Environmental Technic Limited  
บริษัท เทคนิคสิ่งแวดล้อมไทย จำกัด

## High Volume TSP&PM-10 Calibration Report

Location : Thai Environmental Technic

Site ID : Bangkok

Date : 2-Jul-24

ITEM : TSP

Serial No : (No.18 )

Calibrate By : Pipat

### Site Conditions

Barometric Pressure (mm Hg) : 760.00

Temperature (°C) : 25.0

Average Press. (mm Hg) : 754.5

Average Temp (°C) : 30.6

Corrected Pressure (mm Hg) : 760.0

Temperature (deg K) : 298.0

Corrected Average (mm Hg) : -

Average Temp: (Deg K) : -

### Calibration Orifice

Make : Tisch

Model : TE-5025A

Serial# : 0068

Qstd Slope : 1.99045

Qstd Intercept : -0.00789

Calibration Due Date : 16-Aug-24

### Calibration Information

Plate or Test #	ORIFICE (in H <sub>2</sub> O)	Qstd (m3/min)	Indicate (CFM)	IC (corrected)	Linear Regression
1	12.40	1.773	60.0	57.00	Slope : 29.7516
2	9.60	1.561	54.0	52.00	Intercept : 5.6088
3	7.20	1.352	50.0	48.00	Corr. Coeff : 0.9890
4	5.00	1.127	40.0	40.00	
5	3.00	0.874	30.0	30.00	# of Observations: 5

### Calculations

$$Qstd = 1/m[\text{Sqrt}(H_2O(Pa/Pstd)(Tstd/Ta))-b]$$

$$IC = I[\text{Sqrt}(Pa/Pstd)(Tstd/Ta)]$$

Qstd = standard flow rate

IC = corrected chart response

I = actual chart response

m = calibrator Qstd slope

b = calibrator Qstd intercept

Ta = actual temperature during calibration (deg K)

Pa = actual pressure during calibration (mm Hg)

Tstd = 298 deg K

Pstd = 760 mm Hg

For subsequent calculation of sampler flow:

$$1/m(I)[\text{Sqrt}(298/Tav)(Pav/760))-b]$$

NOTE: Ensure calibration orifice has been certified within 12 months of use


m = sampler slope


b = sampler intercept

I = chart response

Tav = daily average temperature

Pav = daily average pressure

Calibrate By : 

Approve By : 



Thai Environmental Technic Limited  
บริษัท เทคนิคสิ่งแวดล้อมไทย จำกัด

## High Volume TSP&PM-10 Calibration Report

Location : Thai Environmental Technic

Site ID : Bangkok

Date : 3-Jul-24

ITEM : TSP

Serial No : (No. 14 )

Calibrate By : Pipat

### Site Conditions

Barometric Pressure (mm Hg) : 760.00

Temperature (°C) : 25.0

Average Press. (mm Hg) : 754.5

Average Temp (°C) : 29.8

Corrected Pressure (mm Hg) : 760.0

Temperature (deg K) : 298.0

Corrected Average (mm Hg) : -

Average Temp: (Deg K) : -

### Calibration Orifice

Make : Tisch

Model : TE-5025A

Serial# : 0068

Qstd Slope : 1.99045

Qstd Intercept : -0.00789

Calibration Due Date : 16-Aug-24

### Calibration Information

Plate or Test #	ORIFICE (in H <sub>2</sub> O)	Qstd (m3/min)	Indicate (CFM)	IC (corrected)	Linear Regression Slope : 29.7516 Intercept : 5.6088 Corr. Coeff : 0.9890 # of Observations: 5
1	12.40	1.773	60.0	57.00	
2	9.60	1.561	54.0	52.00	
3	7.20	1.352	50.0	48.00	
4	5.00	1.127	40.0	40.00	
5	3.00	0.874	30.0	30.00	

### Calculations

$$Qstd = 1/m[\text{Sqrt}(H_2O(Pa/Pstd)(Tstd/Ta))-b]$$

$$IC = [(\text{Sqrt}(Pa/Pstd)(Tstd/Ta))]$$

Qstd = standard flow rate

IC = corrected chart response

I = actual chart response

m = calibrator Qstd slope

b = calibrator Qstd intercept

Ta = actual temperature during calibration (deg K)

Pa = actual pressure during calibration (mm Hg)

Tstd = 298 deg K

Pstd = 760 mm Hg

For subsequent calculation of sampler flow:

$$1/m((I)[\text{Sqrt}(298/Tav)(Pav/760)]-b)$$

NOTE: Ensure calibration orifice has been certified within 12 months of use

m = sampler slope


b = sampler intercept

I = chart response

Tav = daily average temperature

Pav = daily average pressure

Calibrate By : 

Approve By : 



Thai Environmental Technic Limited  
บริษัท เทคนิคสิ่งแวดล้อมไทย จำกัด

## High Volume TSP&PM-10 Calibration Report

Location: Thai Environmental Technic

Site ID: Bangkok

Date: 2-Jul-24

ITEM: TSP

Serial No: (No. 13)

Calibrate By: Pipat

### Site Conditions

Barometric Pressure (mm Hg): 760.00

Temperature (°C): 25.0

Average Press. (mm Hg): 754.4

Average Temp (°C): 31.2

Corrected Pressure (mm Hg): 760.0

Temperature (deg K): 298.0

Corrected Average (mm Hg): -

Average Temp: (Deg K): -

### Calibration Orifice

Make: Tisch

Model: TE-5025A

Serial#: 0068

Qstd Slope: 1.99045

Qstd Intercept: -0.00789

Calibration Due Date: 16-Aug-24

### Calibration Information

Plate or Test #	ORIFICE (in H <sub>2</sub> O)	Qstd (m <sup>3</sup> /min)	Indicate (CFM)	IC (corrected)	Linear Regression Slope: 28.6103 Intercept: 6.6833 Corr. Coeff: 0.9889 # of Observations: 5
1	12.80	1.801	60.0	57.00	
2	10.00	1.593	54.0	52.00	
3	7.40	1.371	50.0	48.00	
4	5.00	1.127	40.0	40.00	
5	3.00	0.874	30.0	30.00	

### Calculations

$$Qstd = 1/m[\text{Sqrt}(H_2O(Pa/Pstd)(Tstd/Ta)) - b]$$

$$IC = I[\text{Sqrt}(Pa/Pstd)(Tstd/Ta)]$$

Qstd = standard flow rate

IC = corrected chart response

I = actual chart response

m = calibrator Qstd slope

b = calibrator Qstd intercept

Ta = actual temperature during calibration (deg K)

Pa = actual pressure during calibration (mm Hg)

Tstd = 298 deg K

Pstd = 760 mm Hg

For subsequent calculation of sampler flow:

$$1/m((I) [\text{Sqrt}(298/Tav)(Pav/760)] - b)$$

NOTE: Ensure calibration orifice has been certified within 12 months of use

m = sampler slope

b = sampler intercept

I = chart response

Tav = daily average temperature

Pav = daily average pressure

Calibrate By : 

Approve By : 





Thai Environmental Technic Limited  
บริษัท เทคนิคสิ่งแวดล้อมไทย จำกัด

## High Volume TSP&PM-10 Calibration Report

Location: Thai Environmental Technic

Site ID: Bangkok

Date: 2-Jul-24

ITEM: TSP

Serial No: (No. 23 )

Calibrate By: Pipat

### Site Conditions

Barometric Pressure (mm Hg) : 760.00

Temperature (°C) : 25.0

Average Press. (mm Hg) : 754.4

Average Temp (°C) : 29.6

Corrected Pressure (mm Hg) : 760.0

Temperature (deg K) : 298.0

Corrected Average (mm Hg) : -

Average Temp: (Deg K) : -

### Calibration Orifice

Make: Tisch

Model: TE-5025A

Serial#: 0068

Qstd Slope : 1.99045

Qstd Intercept : -0.00789

Calibration Due Date : 16-Aug-24

### Calibration Information

Plate or Test #	ORIFICE (in H <sub>2</sub> O)	Qstd (m3/min)	Indicate (CFM)	IC (corrected)	Linear Regression
1	12.40	1.773	60.0	57.00	Slope : 29.7516
2	9.60	1.561	54.0	52.00	Intercept : 5.6088
3	7.20	1.352	50.0	48.00	Corr. Coeff : 0.9890
4	5.00	1.127	40.0	40.00	
5	3.00	0.874	30.0	30.00	# of Observations: 5

### Calculations

$$Qstd = 1/m[\sqrt{H_2O(Pa/Pstd)}(Tstd/Ta)] - b$$

$$IC = I[\sqrt{Pa/Pstd}(Tstd/Ta)]$$

Qstd = standard flow rate

IC = corrected chart response

I = actual chart response

m = calibrator Qstd slope

b = calibrator Qstd intercept

Ta = actual temperature during calibration (deg K)

Pa = actual pressure during calibration (mm Hg)

Tstd = 298 deg K

Pstd = 760 mm Hg

For subsequent calculation of sampler flow:

$$1/m((I)[\sqrt{298/Tav}(Pav/760)] - b)$$

NOTE: Ensure calibration orifice has been certified within 12 months of use

m = sampler slope


b = sampler intercept

I = chart response

Tav = daily average temperature

Pav = daily average pressure

Calibrate By : 

Approve By : 



Thai Environmental Technic Limited

บริษัท เทคนิคสิ่งแวดล้อมไทย จำกัด

## High Volume TSP&PM-10 Calibration Report

Location: Thai Environmental Technic

Site ID: Bangkok

Date: 1-Jul-24

ITEM: PM10

Serial No: (No. 13 )

Calibrate By: Pipat

### Site Conditions

Barometric Pressure (mm Hg) : 760.00

Temperature (°C) : 25.0

Average Press. (mm Hg) : 754.4

Average Temp (°C) : 30.5

Corrected Pressure (mm Hg) : 760.0

Temperature (deg K) : 298.0

Corrected Average (mm Hg) : -

Average Temp: (Deg K) : -

### Calibration Orifice

Make: Tisch

Model: TE-5025A

Serial#: 0068

Qstd Slope : 1.99045

Qstd Intercept : -0.00789

Calibration Due Date : 16-Aug-24

### Calibration Information

Plate or Test #	ORIFICE (in H <sub>2</sub> O)	Qstd (m3/min)	Indicate (CFM)	IC (corrected)	Linear Regression Slope : 34.4221 Intercept : 1.3310 Corr. Coeff : 0.9878 # of Observations: 5
1	12.20	1.759	60.0	60.00	
2	9.00	1.511	54.0	54.00	
3	7.00	1.333	50.0	50.00	
4	5.00	1.127	40.0	40.00	
5	3.00	0.874	30.0	30.00	

### Calculations

$$Qstd = 1/m[\text{Sqrt}(H_2O(Pa/Pstd)(Tstd/Ta))-b]$$

$$IC = I[\text{Sqrt}(Pa/Pstd)(Tstd/Ta)]$$

Qstd = standard flow rate

IC = corrected chart response

I = actual chart response

m = calibrator Qstd slope

b = calibrator Qstd intercept

Ta = actual temperature during calibration (deg K)

Pa = actual pressure during calibration (mm Hg)

Tstd = 298 deg K

Pstd = 760 mm Hg

For subsequent calculation of sampler flow:

$$1/m([I][\text{Sqrt}(298/Tav)(Pav/760))-b]$$

NOTE: Ensure calibration orifice has been certified within 12 months of use

m = sampler slope

b = sampler intercept

I = chart response

Tav = daily average temperature

Pav = daily average pressure

Calibrate By : 

Approve By : 



Thai Environmental Technic Limited  
บริษัท เทคนิคสิ่งแวดล้อมไทย จำกัด

## High Volume TSP&PM-10 Calibration Report

Location : Thai Environmental Technic

Site ID : Bangkok

Date : 2-Jul-24

ITEM : PM10

Serial No : (No. 16 )

Calibrate By : Pipat

### Site Conditions

Barometric Pressure (mm Hg) : 760.00

Temperature (°C) : 25.0

Average Press. (mm Hg) : 754.4

Average Temp (°C) : 30.2

Corrected Pressure (mm Hg) : 760.0

Temperature (deg K) : 298.0

Corrected Average (mm Hg) : -

Average Temp (Deg K) : -

### Calibration Orifice

Make : Tisch

Model : TE-5025A

Serial# : 0068

Qstd Slope : 1.99045

Qstd Intercept : -0.00789

Calibration Due Date : 16-Aug-24

### Calibration Information

Plate or Test #	ORIFICE (in H <sub>2</sub> O)	Qstd (m3/min)	Indicate (CFM)	IC (corrected)	Linear Regression
1	12.20	1.759	60.0	60.00	Slope : 34.4886
2	9.20	1.528	54.0	54.00	Intercept : 0.9749
3	7.00	1.333	50.0	50.00	Corr. Coeff : 0.9878
4	5.20	1.150	40.0	40.00	
5	3.00	0.874	30.0	30.00	# of Observations : 5

### Calculations

$$Qstd = 1/m[\sqrt{H_2O(Pa/Pstd)}(Tstd/Ta)] - b$$

$$IC = I[\sqrt{Pa/Pstd}(Tstd/Ta)]$$

Qstd = standard flow rate

IC = corrected chart response

I = actual chart response

m = calibrator Qstd slope

b = calibrator Qstd intercept

Ta = actual temperature during calibration (deg K)

Pa = actual pressure during calibration (mm Hg)

Tstd = 298 deg K

Pstd = 760 mm Hg

For subsequent calculation of sampler flow:

$$1/m((I)[\sqrt{298/Tav}(Pav/760)] - b)$$

NOTE: Ensure calibration orifice has been certified within 12 months of use


m = sampler slope

b = sampler intercept

I = chart response

Tav = daily average temperature

Pav = daily average pressure

Calibrate By : 

Approve By : 



Thai Environmental Technic Limited  
บริษัท เทคโนโลยีสิ่งแวดล้อมไทย จำกัด

## High Volume TSP&PM-10 Calibration Report

Location : Thai Environmental Technic

Site ID : Bangkok

Date : 1-Jul-24

ITEM : PM10

Serial No : (No. 14 )

Calibrate By : Pipat

### Site Conditions

Barometric Pressure (mm Hg) : 760.00

Temperature (°C) : 25.0

Average Press. (mm Hg) : 754.4

Average Temp (°C) : 29.5

Corrected Pressure (mm Hg) : 760.0

Temperature (deg K) : 298.0

Corrected Average (mm Hg) : -

Average Temp (Deg K) : -

### Calibration Orifice

Make : Tisch

Model : TE-5025A

Serial# : 0068

Qstd Slope : 1.99045

Qstd Intercept : -0.00789

Calibration Due Date : 16-Aug-24

### Calibration Information

Plate or Test #	ORIFICE (in H <sub>2</sub> O)	Qstd (m3/min)	Indicate (CFM)	IC (corrected)	Linear Regression Slope : 34.7047 Intercept : 0.9424 Corr. Coeff : 0.9897 # of Observations: 5
1	12.00	1.744	60.0	60.00	
2	9.20	1.528	54.0	54.00	
3	7.00	1.333	50.0	50.00	
4	5.00	1.127	40.0	40.00	
5	3.00	0.874	30.0	30.00	

### Calculations

$$Qstd = 1/m[\text{Sqrt}(H_2O(P_a/P_{std}))(T_{std}/T_a)] - b$$

$$IC = I[\text{Sqrt}(P_a/P_{std})(T_{std}/T_a)]$$

Qstd = standard flow rate

IC = corrected chart response

I = actual chart response

m = calibrator Qstd slope

b = calibrator Qstd intercept

Ta = actual temperature during calibration (deg K)

Pa = actual pressure during calibration (mm Hg)

Tstd = 298 deg K

Pstd = 760 mm Hg

For subsequent calculation of sampler flow:

$$1/m((I)[\text{Sqrt}(298/T_a)(P_a/760)] - b)$$

NOTE: Ensure calibration orifice has been certified within 12 months of use

m = sampler slope

b = sampler intercept

I = chart response

Tav = daily average temperature

Pav = daily average pressure

Calibrate By : 

Approve By : 



Thai Environmental Technic Limited  
บริษัท เทคนิคสิ่งแวดล้อมไทย จำกัด

## High Volume TSP&PM-10 Calibration Report

Location : Thai Environmental Technic

Site ID : Bangkok

Date : 4-Jul-24

ITEM : PM10

Serial No : (No. 25 )

Calibrate By : Pipat

### Site Conditions

Barometric Pressure (mm Hg) : 760.00

Temperature (°C) : 25.0

Average Press. (mm Hg) : 754.4

Average Temp (°C) : 31.4

Corrected Pressure (mm Hg) : 760.0

Temperature (deg K) : 298.0

Corrected Average (mm Hg) : -

Average Temp: (Deg K) : -

### Calibration Orifice

Make : Tisch

Model : TE-5025A

Serial# : 0068

Qstd Slope : 1.99045

Qstd Intercept : -0.00789

Calibration Due Date : 16-Aug-24

### Calibration Information

Plate or Test #	ORIFICE (in H <sub>2</sub> O)	Qstd (m <sup>3</sup> /min)	Indicate (CFM)	IC (corrected)	Linear Regression Slope : 34.1977 Intercept : 1.5135 Corr. Coeff : 0.9883 # of Observations: 5
1	12.20	1.759	60.0	60.00	
2	9.20	1.528	54.0	54.00	
3	7.00	1.333	50.0	50.00	
4	5.00	1.127	40.0	40.00	
5	3.00	0.874	30.0	30.00	

### Calculations

$$Qstd = 1/m[\text{Sqrt}(H_2O(P_a/P_{std}))(T_{std}/T_a)] - b$$

$$IC = I[\text{Sqrt}(P_a/P_{std})(T_{std}/T_a)]$$

Qstd = standard flow rate

IC = corrected chart response

I = actual chart response

m = calibrator Qstd slope

b = calibrator Qstd intercept

Ta = actual temperature during calibration (deg K)

Pa = actual pressure during calibration (mm Hg)

Tstd = 298 deg K

Pstd = 760 mm Hg

For subsequent calculation of sampler flow:

$$1/m(I)[\text{Sqrt}(298/T_a)(P_{av}/760)] - b$$

NOTE: Ensure calibration orifice has been certified within 12 months of use

m = sampler slope

b = sampler intercept


I = chart response

Tav = daily average temperature

Pav = daily average pressure

Calibrate By : 

Approve By : 



Thai Environmental Technic Limited  
บริษัท เทคนิคสิ่งแวดล้อมไทย จำกัด

## NOx Analyzer Calibration Report

Calibrate Date : 3-Oct-24  
Analyzer Type : NOx  
Brand : API  
Model : TML-41-H-02  
Serial Number : 495 (No. 23)  
Range : 500 ppb


Temperature (°C) : 25°C  
Barometer (mmHg) : 759.9  
Humidity (50±15 %) : 50.0%RH  
Dilutor : API M700 S/N 625  
Zero Air : API M701 S/N 1926  
Standard gas : A00917 SK

### Calibration of Span

Supply Gas	Ref Value(ppb)	Before of Span.(ppb)			After of Span.(ppb)			% diff of Span
		NOx	NO	NO <sub>2</sub>	NOx	NO	NO <sub>2</sub>	
Zero	0.0	3.6	3.1	0.5	0.0	0.0	0.0	0.0
Span	400.0	408.0	401.0	7.0	400.0	400.0	0.0	0.0

### Multi Point Calibration

Ref Value(ppb)	Analyzer Disp.(ppb)			Output Difference		
	NOx	NO	NO <sub>2</sub>	Diff(ppb)	% Diff	Abs (%) Diff
0.0	0.4	0.3	0.1	0.30	0.001	0.08
100.0	101.3	100.5	0.8	0.50	0.005	0.50
200.0	200.7	199.5	1.2	-0.50	-0.003	0.25
400.0	401.2	399.8	1.4	-0.20	0.000	0.05
Average Diff (%)						0.22




Calibrate by:

*[Signature]*

Approved by:

*[Signature]*



Thai Environmental Technic Limited  
บริษัท เทคนิคสิ่งแวดล้อมไทย จำกัด

## NOx Analyzer Calibration Report

Calibrate Date : 25-Sep-24  
Analyzer Type : NOx  
Brand : API  
Model : 200A  
Serial Number : 1978 (No. 15)  
Range : 500 ppb


Temperature (°C) : 25°C  
Barometer (mmHg) : 759.9  
Humidity (50±15 %) : 50.0%RH  
Dilutor : API M700 S/N 625  
Zero Air : API M701 S/N 1926  
Standard gas : A00917 SK

### Calibration of Span

Supply Gas	Ref Value(ppb)	Before of Span.(ppb)			After of Span.(ppb)			% diff of Span
		NOx	NO	NO <sub>2</sub>	NOx	NO	NO <sub>2</sub>	
Zero	0.0	0.4	0.2	0.2	0.0	0.0	0.0	0.0
Span	400.0	382.0	380.0	2.0	400.0	400.0	0.0	0.0

### Multi Point Calibration

Ref Value(ppb)	Analyzer Disp.(ppb)			Output Difference		
	NOx	NO	NO <sub>2</sub>	Diff(ppb)	% Diff	Abs (%) Diff
0.0	0.2	0.3	-0.1	0.31	0.001	0.08
100.0	99.8	99.4	0.4	-0.60	-0.006	0.60
200.0	200.5	200.2	0.3	0.20	0.001	0.10
400.0	401.0	400.5	0.5	0.50	0.001	0.13
Average Diff (%)						0.23




Calibrate by:

*[Signature]*

Approved by:

*[Signature]*



Thai Environmental Technic Limited  
บริษัท เทคนิคสิ่งแวดล้อมไทย จำกัด

## NOx Analyzer Calibration Report

Calibrate Date : 8-Oct-24  
Analyzer Type : NOx  
Brand : API  
Model : 200 A  
Serial Number : 542 (No. 29)  
Range : 500 ppb


Temperature (°C) : 25°C  
Barometer (mmHg) : 759.9  
Humidity (50±15 %) : 50.0%RH  
Dilutor : API M700 S/N 625  
Zero Air : API M701 S/N 1926  
Standard gas : A00917 SK

### Calibration of Span

Supply Gas	Ref Value(ppb)	Before of Span.(ppb)			After of Span.(ppb)			% diff of Span
		NOx	NO	NO <sub>2</sub>	NOx	NO	NO <sub>2</sub>	
Zero	0.0	1.5	1.3	2.0	0.0	0.0	0.0	0.0
Span	400.0	381.0	388.0	-7.0	400.0	400.0	0.0	0.0

### Multi Point Calibration

Ref Value(ppb)	Analyzer Disp.(ppb)			Output Difference		
	NOx	NO	NO <sub>2</sub>	Diff(ppb)	% Diff	Abs (%) Diff
0.0	0.3	0.2	0.1	0.20	0.001	0.05
100.0	100.3	99.8	0.5	-0.20	-0.002	0.20
200.0	199.8	199.7	0.1	-0.30	-0.002	0.15
400.0	400.4	400.2	0.2	0.20	0.000	0.05
Average Diff (%)						0.13




Calibrate by:

*[Signature]*

Approved by:

*[Signature]*





Thai Environmental Technic Limited  
บริษัท เทคนิคสิ่งแวดล้อมไทย จำกัด

## NOx Analyzer Calibration Report

Calibrate Date : 8-Oct-24  
Analyzer Type : NOx  
Brand : API  
Model : 200 A  
Serial Number : 777 (No. 25)  
Range : 500 ppb


Temperature (°C) : 25°C  
Barometer (mmHg) : 759.9  
Humidity (50±15 %) : 50.0%RH  
Dilutor : API M700 S/N 625  
Zero Air : API M701 S/N 1926  
Standard gas : A00917 SK

### Calibration of Span

Supply Gas	Ref Value(ppb)	Before of Span.(ppb)			After of Span.(ppb)			% diff of Span
		NOx	NO	NO <sub>2</sub>	NOx	NO	NO <sub>2</sub>	
Zero	0.0	1.2	0.5	0.7	0.0	0.0	0.0	0.0
Span	400.0	355.0	350.0	5.0	400.0	400.0	0.0	0.0

### Multi Point Calibration

Ref Value(ppb)	Analyzer Disp.(ppb)			Output Difference		
	NOx	NO	NO <sub>2</sub>	Diff(ppb)	% Diff	Abs (%) Diff
0.0	0.5	0.4	0.0	0.42	0.001	0.11
100.0	101.5	100.3	1.2	0.30	0.003	0.30
200.0	201.3	200.1	1.2	0.10	0.000	0.05
400.0	399.8	399.5	0.3	-0.50	-0.001	0.13
Average Diff (%)						0.14




Calibrate by:

*James James*

Approved by:

*Ramual M*

**TET**

Thai Environmental Technic Limited  
บริษัท เทคนิคสิ่งแวดล้อมไทย จำกัด

## Analyzer Calibration Report

Calibrate Date : 26-Oct-24  
Analyzer Type : SO<sub>2</sub>  
Brand : API  
Model : 100 A  
Serial Number : 856 (No. 5)  
Range : 500 ppb


Temperature (°C) : 25°C  
Barometer (mmHg) : 760.0  
Humidity (50±15 %) : 50.0 %RH  
Dilutor : API M700 S/N 625  
Zero Air : API M701 S/N 1926  
Standard gas : D636157

### Calibration of Span

Supply Gas	Ref Value(ppb)	Before of Span.(ppb)	After of Span.(ppb)	Abs% diff of Span
Zero	0.0	-2.9	0.0	0.0
Span	400.0	364.0	400.0	0.0

### Multi Point Calibration

Ref Value(ppb)	Analyzer Disp.(ppb)	Output Difference		
		Diff (ppb)	Percent Diff	Abs Percent Diff
0.0	0.4	0.4	0.00	0.10
100.0	99.3	-0.7	-0.01	0.70
200.0	199.1	-0.9	0.00	0.45
400.0	398.6	-1.4	0.00	0.35
Average Diff (%)				0.40




Calibrate by: \_\_\_\_\_

Approved by: \_\_\_\_\_

แก้ไขครั้งที่ : 00

วันที่อนุมัติ 02/09/15

เลขที่แบบฟอร์ม : QF-QP16-06



Thai Environmental Technic Limited  
บริษัท เทคนิคสิ่งแวดล้อมไทย จำกัด

## Analyzer Calibration Report


Calibrate Date	28-Sep-24	Temperature (°C)	: 25°C
Analyzer Type	SO <sub>2</sub>	Barometer (mmHg)	: 755.0
Brand	API	Humidity (50±15 %)	: 50.0 %RH
Model	100A	Dilutor	: API M700 S/N 625
Serial Number	1563 (No. 15)	Zero Air	: API M701 S/N 1926
Range	500 ppb	Standard gas	: D636157

### Calibration of Span

Supply Gas	Ref Value(ppb)	Before of Span.(ppb)	After of Span.(ppb)	Abs% diff of Span
Zero	0.0	10.2	0.0	0.0
Span	400.0	403.0	400.0	0.0

### Multi Point Calibration

Ref Value(ppb)	Analyzer Disp.(ppb)	Output Difference		
		Diff (ppb)	Percent Diff	Abs Percent Diff
0.0	0.7	0.7	0.00	0.18
100.0	100.5	0.5	0.01	0.50
200.0	199.8	-0.2	0.00	0.10
400.0	399.3	-0.7	0.00	0.17
Average Diff (%)				0.24




Calibrate by:

Approved by:

แก้ไขครั้งที่ : 00

วันที่อนุมัติ 02/09/15

เลขที่แบบฟอร์ม : QF-QP16-06



Thai Environmental Technic Limited

บริษัท เทคนิคสิ่งแวดล้อมไทย จำกัด

## Analyzer Calibration Report


Calibrate Date	27-Sep-24	Temperature (°C)	25°C
Analyzer Type	SO <sub>2</sub>	Barometer (mmHg)	760.0
Brand	Thermo	Humidity (50±15 %)	50.0 %RH
Model	43C	Dilutor	API M700 S/N 625
Serial Number	43C57277312 (No. 14)	Zero Air	API M701 S/N 1926
Range	500 ppb	Standard gas	D636157

### Calibration of Span

Supply Gas	Ref Value(ppb)	Before of Span.(ppb)	After of Span.(ppb)	Abs% diff of Span
Zero	0.0	2.6	0.0	0.0
Span	400.0	362.0	400.0	0.0

### Multi Point Calibration

Ref Value(ppb)	Analyzer Disp.(ppb)	Output Difference		
		Diff (ppb)	Percent Diff	Abs Percent Diff
0.0	0.4	0.4	0.00	0.10
100.0	99.1	-0.9	-0.01	0.90
200.0	199.2	-0.8	0.00	0.40
400.0	399.3	-0.7	0.00	0.17
Average Diff (%)				0.39



Calibrate by:

*[Signature]*

Approved by:

*[Signature]*

แก้ไขครั้งที่ : 00

วันที่อนุมัติ 02/09/15

เลขที่แบบฟอร์ม : QF-QP16-06



Thai Environmental Technic Limited  
บริษัท เทคนิคสิ่งแวดล้อมไทย จำกัด

## Analyzer Calibration Report

Calibrate Date : 2-Oct-24  
Analyzer Type : CO  
Brand : Teledyne  
Model : 300E  
Serial Number : 1066 (No.2)  
Range : 1000 ppm


Temperature (°C) : 25°C  
Barometer (mmHg) : 760  
Humidity (50±15 %) : 50.0 %RH  
Dilutor : API M700 S/N625  
Zero Air : API M701 S/N1926  
Standard gas : D824500, D271305

### Calibration of Span

Supply Gas	Ref Value(ppm)	Before of Span.(ppm)	After of Span.(ppm)	Abs% diff of Span
Zero	0.0	1.2	0.0	0.00
Span	793.0	791	793.0	0.00

### Multi Point Calibration

Ref Value(ppm)	Analyzer Disp.(ppm)	Output Difference		
		Diff (ppm)	Percent Diff	Abs Percent Diff
0.0	0.4	0.4	0.00	0.05
404.0	405.1	1.1	0.00	0.27
793.0	793.2	0.2	0.00	0.03
Average Diff (%)				0.12



Calibrate by :

*[Signature]*


Approved by :

*[Signature]*

แก้ไขครั้งที่ : 00

วันที่อนุมัติ 02/09/15

เลขที่แบบฟอร์ม : QF-QP16-06



Thai Environmental Technic Limited  
บริษัท เทคนิคสิ่งแวดล้อมไทย จำกัด

## Analyzer Calibration Report


Calibrate Date	: 2-Oct-24	Temperature (°C)	: 26°C
Analyzer Type	: CO	Barometer (mmHg)	: 760
Brand	: Horiba	Humidity (50±15 %)	: 50.0
Model	: APMA 360CE	Dilutor	: API M700 S/N625
Serial Number	: 42088-7001 (No.1)	Zero Air	: API M701 S/N1926
Range	: 1000 ppm	Standard gas	: D824500, D271305

### Calibration of Span

Supply Gas	Ref Value(ppm)	Before of Span.(ppm)	After of Span.(ppm)	Abs% diff of Span
Zero	0.0	0.70	0.00	0.00
Span	793.0	799.00	793.00	0.00

### Multi Point Calibration

Ref Value(ppm)	Analyzer Disp.(ppm)	Output Difference		
		Diff (ppm)	Percent Diff	Abs Percent Diff
0.0	0.42	0.4	0.00	0.05
404.0	403.80	-0.2	0.00	0.05
793.0	793.50	0.5	0.00	0.06
Average Diff (%)				0.06



Calibrate by:

*Janet Janet*

Approved by

*Pramual M*

แก้ไขครั้งที่ : 00

วันที่อนุมัติ 02/09/15

เลขที่แบบฟอร์ม : QF-QP16-06

# Personal Pump Calibration Report

*Equipment Type* : Personal Pump/Parameter

*Equipment Range* : 0.1-7.0 V/min

Calibration Range : 0.1-4.0 V/min

Calibration Type : Drycal

Calibration S/N : 4491

[illegible]

Calibration Date 08 / 01 / 68

Calibration By √5.250

Remark : Uncertainty Type A =  $\sigma =$  SD

$\sqrt{n}$

: SD = Standard deviation

$$\bar{X} = \text{Mean}$$

## Personal Pump Calibration Report

*Equipment Type* : Personal Pump/Parameter  
*Equipment Range* : 0.1-7.0 l/min  
*Calibration Range* : 0.1-4.0 l/min  
*Calibration Type* : Drycal  
*Calibration S/N* : 4491

Item	Personal Pump S/N	Hi Flow/Low Flow	ครั้งที่ 1	ครั้งที่ 2	ครั้งที่ 3	Average	Uncertainty
1.	20140505076	2.0	1.9970	1.9980	1.9990	1.9980	±0.0010

Calibration Date 14 / 02 / 68

Calibration By ป/รณนัฏ

Remark : Uncertainty Type A =  $\frac{\sigma}{\sqrt{n}}$  = SD

: SD = Standard deviation

:  $\bar{X}$  = Mean





## Preventive Maintenance and Performance Report

### Methane-NMHC Analyzer

#### CONFIGURATION TESTED

<u>MODEL</u>	<u>SERIAL NUMBER</u>	<u>DATE TEST</u>	<u>DUE DATE</u>
55C	55C-72555-371	09/01/2025	08/01/2026

#### Preventive Maintenance List:

##### 1. Clean and inspect Analyzer

- ✓ Unplug power cord from the power source.
- ✓ Wipe/remove any dust.
- ✓ Inspect internal connectors for proper contact and placement.
- ✓ Verify operation of all replaceable parts.

##### 2. Restore Analyzer

- ✓ Restore the normal operating conditions.
- ✓ Check and record the post PM detector signal output values. Results should be similar or lower than the detector output recorded prior to PM.

#### Calibration System :

Standard gas					
GASES	Conc.	Uncertainty	Cer. No	Cyl. No	Exp. Date
Methane/Propane in Air	2.0 ppm	± 0.2 ppm	1672/23	682621	26/6/25
Methane/Propane in Air	20.0 ppm	± 1.0 ppm	2182/24	14M343067	15/7/26
Methane/Propane in Air	200.0 ppm	± 4.0 ppm	2969/23	53214	18/10/27



Environmental: Temperature 24.0 °C

Humidity 51 % RM

**Test Results Table :**

The calibration was performed following the triple point by Standard gas mixed Methane-Propane in Air at concentration 2 , 20 and 200 ppm and verified by Standard gas mixed Methane-Propane in Air as following :

Calibration Check (Before adjust)						
Std. gas	Zero			Span		
	Reading (ppm)	Expected (ppm)	Drift (ppm)	Reading (ppm)	Expected (ppm)	Drift (%)
Methane	0.0	0.0	0.0	1.91	2.0	4.61
	0.0	0.0	0.0	13.50	20.0	32.50
	0.0	0.0	0.0	182.68	200.0	8.66
NMHC	0.0	0.0	0.0	2.03	2.0	1.55
	0.0	0.0	0.0	36.06	20.0	80.32
	0.0	0.0	0.0	194.99	200.0	2.51


Calibration Check (After adjust)							
Std. gas	Zero			Span			Evaluated (≤ 2 %)
	Reading (ppm)	Expected (ppm)	Drift (ppm)	Reading (ppm)	Expected (ppm)	Drift (%)	
Methane	0.0	0.0	0.0	1.99	2.0	0.50	pass
	0.0	0.0	0.0	19.84	20.0	0.79	pass
	0.0	0.0	0.0	198.40	200.0	0.80	pass
NMHC	0.0	0.0	0.0	1.99	2.0	0.42	pass
	0.0	0.0	0.0	19.97	20.0	0.16	pass
	0.0	0.0	0.0	198.62	200.0	0.69	pass




Linearity Check :

Conc. (ppm)	Methane		Propane	
	Reading	Height	Reading	Height
2.0	1.99	41.36	1.99	17.85
20.0	19.84	283.11	19.97	309.90
200.0	198.40	2629.40	198.62	3059.80

Methane Response



Non-Methane Response



PM Operations by 100 118457  
( Mr. Jer Saewa )  
Scientist

PM Date 9/1/25

Approve by [Signature]  
( Mrs. Phornthip Phetshee )  
Laboratory Manager

Approve Date 9/1/25

THE LINDE GROUP

Linde

# Certificate Of Analysis Special Gases Mixture

## Customer Details

Name:

Thai Environmental Technic Ltd.

Address:

1/6 Soi Ramkhamhaeng 145, Saphansoong,  
Saphansoong, Bangkok 10240 Thailand

Customer Tag No.:

## Certificate Details

Number:

1672/23

Date of Issue:

27-Jun-2023

Expiry date:

26-Jun-2025

Material Details

Production Order:

90178337

Material Code:

470200-J-32

Cylinder No.:

682621

Gas content:

6.24 M<sup>3</sup> (nominal)

Filling pressure:

137.0 bar (g)

Valve:

CGA 590 BRASS

Cylinder Owner:

LINDE

Cylinder Material:

STEEL

Cylinder Size:

47 L

## Laboratory Report

Component	Normal Concentration	Analysis Result <sup>1</sup>	Uncertainty <sup>2</sup>	Method of Analysis <sup>3</sup>
Methane	2.0 ppm	1.6 ppm	± 10% relative	(1) ACC-FID-01
Propane	2.0 ppm	1.9 ppm	± 10% relative	(1) ACC-FID-01
In Air				

## Recommend usage condition

Minimum utilization: 5% of actual content or before expire date whichever comes first.

Storage condition: Keep in well ventilation and secure area.

## Comments

## Note:

- All results expressed in this report are on mole/mole basis, unless otherwise specified.
- The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The measurement of this material is traceable to the SI through the reference gas standard which is traceable to Swiss National Standard of Mass or other recognised national metrology institutes.
- (1) Gas Chromatography, (2) Paramagnetic Oxygen Analyzer, (3) Electrochemical Oxygen Analyzer, (4) Electrochemical Moisture Analyzer, (5) Total Hydrocarbon Analyzer, (6) Other - Specified

Sukanya Parinyasontorn

Signatory for and on behalf of Linde (Thailand) Co., Ltd.

Page 1 of 1

This report shall not be reproduced except in full

PB-002/F004

Iss:K/2, 15 Oct 2021

บริษัท ลินด์ (ประเทศไทย) จำกัด (มหาชน)

ทะเบียนบริษัทเลขที่ 0107537000785

ชั้น 15 อาคารทาวเวอร์ เอ 2/3 หมู่ 14 ถนนบางนา-ตราด กม. 6.5 ต.บางนาใต้

อ.บางพลี จ.สมุทรปราการ 10540 โทรศัพท์ (66) 2338-6100 โทรสาร (66) 2338-6333

โรงงานเวลโกรว์: 105 หมู่ 5 ต.บางพลี อ.บางพลี จ.สมุทรปราการ 24180


Linde (Thailand) Public Company Limited

PLC Registration no. 0107537000785

15<sup>th</sup> Floor, Bangna Tower A, 2/3 Moo 14, Bangna Trad KM. 6.5 Road, Bangkaew

Bangplee, Samutprakarn 10540, Tel (66) 2338-6100 Fax (66) 2338-6333

Wellgrow Plant: 105 Moo 5, T.Bangsamak, A.Bangpakong, Chachoengsao 24180



## Certificate Of Analysis Special Gases Mixture

**Customer Details**

Name:	Address:	Customer Tag No.:
Thai Environmental Technic Limited.	House number 1/6 Soi Ramkhamhaeng 45, Sapansoong, Khet Saphan Sung, Bangkok 10240	-

**Certificate Details**

Number:	2182/24	Date of Issue:	16-Jul-2024	Expiry date:	15-Jul-2026
<b>Material Details</b>					
Production Order:	90184949	Material Code:	470300-V-32	Cylinder No.:	14M343067
Gas content:	1.330 M <sup>3</sup> (nominal)	Filling pressure:	137 bar (g)	Valve:	CGA 590 BRASS
Cylinder Owner:	LINDE	Cylinder Material:	STEEL	Cylinder Size:	10 L

**Laboratory Report**

Component	Nominal Concentration	Analysis Result <sup>1</sup>	Uncertainty <sup>2</sup>	Method of Analysis <sup>3</sup>
Methane	20.0 ppm	20.1 ppm	± 5% relative	(1) ACC-FID-01
Propane	20.0 ppm	20.4 ppm	± 5% relative	(1) ACC-FID-01
In Air				

**Recommend usage condition**

Minimum utilization: 5% of actual content or before expire date whichever comes first.

Storage condition: Keep in well ventilation and secure area.

**Comments**
**Note:**

1. All results expressed in this report are on mole/mole basis, unless otherwise specified.
2. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The measurement of this material is traceable to the SI through the reference gas standard which is traceable to Swiss National Standard of Mass or other recognised national metrology institutes.
3. (1) Gas Chromatography, (2) Paramagnetic Oxygen Analyzer, (3) Electrochemical Oxygen Analyzer,
- (4) Electrochemical Moisture Analyzer, (5) Total Hydrocarbon Analyzer, (6) Other - Specified

Sukanya Parinyasontorn

Signatory for and on behalf of Linde (Thailand) Co., Ltd.

**บริษัท ลินด์ (ประเทศไทย) จำกัด (มหาชน)**

หมายเลขใบอนุญาตประกอบธุรกิจ 0107537000785

ชั้น 15 ถนนพหลโยธิน แขวงจตุจักร เขตจตุจักร กรุงเทพฯ 10110

อ.บางพลี จ.สมุทรปราการ 10540 โทรศัพท์ (66) 2338-6100 โทรสาร (66) 2338-6333

โรงงานเวลโกรว์: 105 หมู่ 5 ต.บางพลี จ.สมุทรปราการ 24180

โทรศัพท์ (66) 38.570-479-93

โทรสาร (66) 38.570-323

**Linde (Thailand) Public Company Limited**

PLC Registration no. 0107537000785

 15<sup>th</sup> Floor, Bangna Tower A, 2/3 Moo 14, Bangna Trad KM. 6.5 Road, Bangkaew

Bangplee, Samutprakarn 10540, Tel (66) 2338-6100 Fax (66) 2338-6333

Wellgrow Plant : 105 Moo 5, T.Bangsamak, A.Bangpakong, Chachoengsao 24180

Thailand, Tel (66) 38.570-479-93

Fax (66) 38.570-323

Certificate Of Analysis  
Special Gases Mixture

## Customer Details

Name:

Thai Environmental Technic Limited

Address:

1/6 Soi Ramkhamhaeng 45 Sapansoong  
District., Khet Saphan Sung Bangkok 10240  
Thailand

Customer Tag No.:

## Certificate Details

Number:

2969/23

Date of Issue:

19-Oct-2023

Expiry date:

18-Oct-2027

Material Details

Production Order:

90180549

Material Code:

470400-J-32

Cylinder No.:

53214

Gas content:

6.24 M<sup>3</sup> (nominal)

Filling pressure:

137.0 bar (g)

Valve:

CGA 590 BRASS

Cylinder Owner:

LINDE

Cylinder Material:

STEEL

Cylinder Size:

47 L

## Laboratory Report

Component	Nominal Concentration	Analysis Result <sup>1</sup>	Uncertainty <sup>2</sup>	Method of Analysis <sup>3</sup>
Methane	200 ppm	195 ppm	± 2% relative	(1) ACC-RGA-01
Propane	200 ppm	200 ppm	± 2% relative	(1) ACC-RGA-01
In Air				

## Recommend usage condition

Minimum utilization:

5% of actual content or before expire date whichever comes first.

Storage condition:

Keep in well ventilation and secure area.

Comments

## Note:

1. All results expressed in this report are on mole/mole basis, unless otherwise specified.
2. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The measurement of this material is traceable to the SI through the reference gas standard which is traceable to Swiss National Standard of Mass or other recognised national metrology institutes.
3. (1) Gas Chromatography, (2) Paramagnetic Oxygen Analyzer, (3) Electrochemical Oxygen Analyzer,
- (4) Electrochemical Moisture Analyzer, (5) Total Hydrocarbon Analyzer, (6) Other - Specified

Sukanya Parinyasontorn

Signatory for and on behalf of Linde (Thailand) Co., Ltd.

Page 1 of 1

This report shall not be reproduced except in full

PB-002/F004

Iss:L/2, 01 August 2023

บริษัท ลินด์ (ประเทศไทย) จำกัด (มหาชน)

เลขที่ใบอนุญาตประกอบกิจการ 0107537000785

ชั้น 15 แขวงบางนาแถม เขต 2/3 หมู่ 14 ถนนบางนา-ตราด กม. 6.5 ตำบลบางนา

อ.บางพลี จ.สมุทรปราการ 10540 โทรศัพท์ (66) 2338-6100 โทรสาร (66) 2338-6333

โรงงานเวลโกรว์: 105 หมู่ 5 ต.บางสนธิ์ อ.บางปะกง จ.ฉะเชิงเทรา 24180

โทรศัพท์ (66) 38.570-479-93

โทรสาร (66) 38.570-323

Linde (Thailand) Public Company Limited

PLC Registration no. 0107537000785


15<sup>th</sup> Floor, Bangna Tower A, 2/3 Moo 14, Bangna Trad KM. 6.5 Road, Bangkaew

Bangplee, Samutprakarn 10540, Tel (66) 2338-6100 Fax (66) 2338-6333

Wellgrow Plant: 105 Moo 5, T.Bangsamak, A.Bangpakong, Chachoengsao 24180

Thailand, Tel (66) 38.570-479-93

Fax (66) 38.570-323




TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL.0-2717-3000-29 FAX.0-2719-9484



## Certificate of Calibration

Cert.No.: 24CHO574

Page.: 1 of 2

Equipment : pH Meter  
Manufacturer : Horiba  
Model : LAQUA-PH1300  
Serial No. : B06D0012  
ID No. : Ins-LAB-026  
Condition As-Received: Used Item  
Received Date : 30 October 2024  
Calibration Date : 30 October 2024  
Reference : 2410-0784OC-6  
Submitted by : Thai Environmental Technic Limited  
1/6 Soi Ramkhamhaeng 145,  
Khwaeng/Khet Saphan Sung,  
Bangkok 10240  
Calibration Place : Laboratory (Thai Environmental Technic Limited)  
Ambient Temperature : ( 25.3 to 24.8 ) °C (On-Site)  
Relative Humidity : ( 71.7 to 77.5 ) % (On-Site)  
Calibration Procedure : In - house method :  
- CP-OCH2 by direct measurement with DC voltage  
standard and direct measurement with  
certified reference material (CRM)  
Calibrated by : Saithip Meangmai  
Approved by :   
Approved Signatory  
( ) Unnopphol Harachai  
(✓) Ponpan Paipim  
( ) Saithip Meangmai  
Issue Date : 2 November 2024

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.



Cert.No.: 24CHO574

Page.: 2 of 2

**Condition of this calibration result****1. Reference Standard Instrument**

<b>Instrument</b>	<b>Serial No.</b>	<b>ID No.</b>	<b>Cert. No.</b>	<b>Due Date</b>
1) Document Process Calibrator	46530031	130RC098	24E3004	12 Sep 2025
2) Digital Thermometer	307901	70RC137	24I973	01 Sep 2025

- This Certification is traceable to SI Through Technology Promotion Association (Thailand - Japan)

2. Certified Reference Materials : The measurement results are traceable to SI through CPA chem Ltd.,  
ANSI-ASQ National Accreditation Board, Accredited No. AR-1835  
: The measurement results are traceable to SI through Hach Lenge GmbH Ltd.,  
The measurement results are traceable to SI through Hach Lenge GmbH Ltd.,

<b>Buffer Solution</b>	<b>Manufacturer</b>	<b>Lot No.</b>	<b>Exp. date</b>
pH 1.685	CPA chem	1005300	15 June 2026
pH 4.008	CPA chem	1034203	27 Sep 2026
pH 6.876	CPA chem	1005301	15 June 2026
pH 9.174	CPA chem	1005302	15 June 2025
*pH 12.42	Hach Lenge GmbH	C03178	07 June 2026

3. This certificate is valid only to the item calibrated on date and place of calibration.

**Calibration Results****Function : mV Measurement**

Performing standard curve by Document Process Calibrator at pH (1.7,4,7,10,12)

<b>Unit Under Calibration</b>	<b>Nominal Value</b>	<b>Standard Voltage Input</b>	<b>Actual Reading</b>		<b>Uncertainty of Measurement (<math>\pm</math>mV)</b>	<b>Coverage factor k</b>
	<b>pH</b>	<b>mV</b>	<b>mV</b>	<b>pH</b>		
pH Meter S/N.: V3B1F8H3	1.680	314.73	314.7	1.680	0.058	2.00
	4.000	177.48	177.4	4.000	0.058	2.00
	6.860	8.28	8.3	6.860	0.058	2.00
	7.000	0.00	0.0	7.000	0.058	2.00
	9.180	-128.97	-128.9	9.180	0.058	2.00
	10.000	-177.48	-177.4	10.000	0.058	2.00
	12.000	-295.80	-295.8	12.000	0.058	2.00

**Function : pH Measurement**

Performing five buffers standard curve by using buffer nominal pH (1.68,4.01,6.86,9.18,12.42)

<b>Unit Under Calibration</b>	<b>Standard pH Buffer Solution</b>	<b>Actual pH Reading</b>	<b>Actual mV Reading (mV)</b>	<b>Uncertainty of pH Measurement (<math>\pm</math>)</b>	<b>Coverage factor k</b>
pH Electrode S/N.: 9X3D0537	1.685	1.679	293.0	0.0074	2.18
	4.008	3.981	154.7	0.0057	2.09
	6.876	6.842	-13.2	0.0075	2.05
	9.174	9.151	-148.8	0.013	2.07
	*12.42	12.423	-337.1	0.059	2.05

**Remark** - \* = Not NSC-ONSC Accredited.

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor, providing a level of confidence of approximately 95 %.





TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL.0-2717-3000-29 FAX.0-2719-9484



## Certificate of Calibration

Cert.No.: 24MM272

Page.: 1 of 3

Equipment : Electronic Balance

Manufacturer : Mettler Toledo

Model : AB204

Serial No. : 1116392227

ID No. : Ins-LAB-033

Submitted by : Thai Environmental Technic Limited  
1/6 Soi Ramkhamhaeng 145,  
Khwaeng/Khet Saphan Sung,  
Bangkok 10240

Location : Balance Room

Received order : 09 April 2024

Calibration Date : 10 April 2024

Ambient Temperature : 15 °C to 40 °C

Relative Humidity : 30 % to 90 %

Calibrated by : Khit Ruttanaprapachai

Approved by : Kunchit  
Approved Signatory

( ) Ponpan Paipim  
( ) Suwit Imjai  
(✓) Kunchit Promprat

Issue Date :

12 April 2024

**The Uncertainties are for a confidence probability of approximately 95%**

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.



**Equipment :** Electronic Balance  
**Condition As-Received :** Used Item  
**Reference :** 2404-01130C-14  
**Procedure used :-**

**Cert.No.:** 24MM272

**Page:** 2 of 3

Calibration were conducted using in-house calibration procedure CP-OB01 based on UKAS LAB 14 according to direct measurement method against standard weight.

**Condition of this result of calibration**

**1. Reference standard instruments:-**

<u>Instruments</u>	<u>Model</u>	<u>Serial No.</u>	<u>ID No.</u>	<u>Test report No.</u>	<u>Due date</u>
1) Standard Weight Set (E2)	15884		70RC138	MM-0020-23	30 Jan 2025

2. This certificate is valid only to the item calibrated on date and place of calibration.
3. This result of calibration was made on requested at the point specified by customer.
4. This certificate is not certified for any commercial transaction.
5. This certification is traceable to the International System of Unit.

**Result of calibration** ( ) Without Adjustment ( \* ) After Adjustment by External Calibration

**Range capacity :** 0 g to 210 g **Resolution** 0.0001 g

**Before Adjustment :**

<u>Applied Weight</u>	<u>Balance Reading</u>	<u>Correction</u>	<u>Measurement Uncertainty</u>	<u>Coverage Factor</u>
( g )	( g )	( g )	( $\pm$ mg )	( k )
100	100.0000	0.0000	0.19	2
200	200.0001	-0.0001	0.30	2

**After Adjustment :**

**1. Determination of the standard deviation of weighing machine** ( n = 10 )

<u>Applied Weight</u>	<u>Standard Deviation of Reading ( g )</u>
( g )	
100	0.00007
200	0.00008



Equipment : Electronic Balance  
 Condition As-Received : Used Item  
 Reference : 2404-0113OC-14


Cert.No.: 24MM272

Page: 3 of 3

### Result of calibration

#### 2. Effect of off center loading

A mass of 100 g was placed to various position on the pan.  
 The weighing machine reading error obtained is given in the table



Maximum difference between  
 off-center and central loading

Position 1 (g)	Position 2 (g)	Position 3 (g)	Position 4 (g)	Position 5 (g)	(g)
0.0000	+0.0001	0.0000	+0.0001	+0.0003	0.0003

#### 3. Departure from nominal value

Applied Weight (g)	Balance Reading (g)	Correction (g)	Measurement Uncertainty ( $\pm$ mg)	Coverage Factor (k)
Unload	0.0000	0.0000	0.14	2.11
0.01	0.0101	-0.0001	0.14	2.11
0.1	0.1001	-0.0001	0.14	2.11
0.5	0.5002	-0.0002	0.14	2.11
1	1.0002	-0.0002	0.14	2.11
5	5.0000	0.0000	0.14	2.11
10	10.0001	-0.0001	0.14	2.11
25	25.0000	0.0000	0.15	2.07
50	49.9999	+0.0001	0.15	2.06
100	100.0002	-0.0002	0.19	2
200	200.0002	-0.0002	0.30	2

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor  $k$ , providing a level of confidence of approximately 95 %.

-o0o-



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL.0-2717-3000-29 FAX.0-2719-9484



## Certificate of Calibration

Cert. No.: 24TM702

Page : 1 of 3

Equipment : BOD Incubator

Manufacturer : Accuplus

Model : i250

Serial No. : 0408-0115-0008


ID No. : Ins-LAB-046

Submitted by : Thai Environmental Technic Limited  
1/6 Soi Ramkhamhaeng 145,  
Khwaeng/Khet Saphan Sung,  
Bangkok 10240

Location : Laboratory (Thai Environmental Technic Limited)

Received Order : 09 April 2024  
Calibration Date : 09 April 2024  
Ambient Temperature : (  $26 \pm 10$  ) °C  
Relative Humidity : (  $50 \pm 30$  ) %

Calibrated by : Khit Ruttanaprapachai

Approved by :   
Approved Signatory

( ) Ponpan Paipim  
( ) Suwit Imjai  
(✓) Kunchit Promprat

Issue Date : 26 April 2024

**The Uncertainties are for a confidence probability of approximately 95%**

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.



**Equipment :** BOD Incubator  
**Condition As-Received :** Used Item  
**Reference :** 2404-0113OC-11  
**Procedure Used :-**

**Cert. No.:** 24TM702  
**Page :** 2 of 3

Calibration were conducted using calibration procedure CP-OT02 based on TLAS G-20 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector ( RTD ).

The temperature scale used was based on ITS-90.

**Condition of this result of calibration**

1. Reference standard instrument:-

<u>Instrument</u>	<u>Serial No.</u>	<u>Cert. No.</u>	<u>Traceable</u>	<u>Due Date</u>
1 ) Data Acquisition	MY49001451	24LM44	TPA	17 Mar 2025

2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.


**Remark :** TPA : Technology Promotion Association ( Thailand - Japan )

**Result of Calibration :-** ( \* ) Without Adjustment

**Function of UUC\* :** Temperature Source

**Fresh air setting :** Not Available

<b>Environment during calibration</b>		
	<b>Beginning</b>	<b>Finished</b>
Temp. ( °C )	24	25
REL.Humid. ( % )	50	52
AC Supply ( Volt )	221	220



<b>Position :</b>	<b>Ref. Std. ID No.:</b>
1	19RTD-2/1
2	19RTD-2/2
3	19RTD-2/3
4	19RTD-2/4
5	19RTD-2/5
6	24-19RTD-2/6
7	19RTD-2/7
8	19RTD-2/8
9 (ref.)	19RTD-2/9

**Probe Installation Details :**

a = 10 cm  
 b = 10 cm  
 c = 10 cm

**Dimension of Chamber :**

D = 0.48 m  
 W = 0.50 m  
 H = 1.1 m  
 Capacity = 0.26 m<sup>3</sup>



**Equipment :** BOD Incubator  
**Condition As-Received :** Used Item  
**Reference :** 2404-0113OC-11  
**Result of Calibration :-** ( \* ) Without Adjustment  
**Function of UUC\* :** Temperature Source  
**Fresh air setting :** Not Available

**Cert. No.:** 24TM702  
**Page :** 3 of 3

Calibration Point ( °C )	UUC* Setting ( °C )	UUC* Reading ( °C )	Temperature stability ( ± °C )	Temperature uniformity ( °C )	Overall Variation ( °C )	Coverage Factor <i>k</i>
20.0	20.0	20.0	0.30	0.27	0.77	2

Calibration Point ( °C )	Measured Temperature ( °C )									Uncertainty  ( ±°C )
	Position									
	1	2	3	4	5	6	7	8	9 (ref.)	
20.0	20.232	20.184	20.129	20.214	20.126	20.102	19.987	20.053	20.128	0.49

**Average\* :** The average of 30 values in each position.

**Temperature stability :** One-half of the greatest maximum difference of measured temperature at any one sensor.

**Temperature uniformity :** The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

**Overall Variation :** The Difference of the maximum and minimum measured temperatures throughout observation.

**UUC\* :** Unit Under Calibration

**Note :** The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k* , providing a level of confidence of approximately 95 %.

-o0o-




TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL.0-2717-3000-29 FAX.0-2719-9484



Cert.No.: 24CHO222

Page.: 1 of 3

## Certificate of Calibration

Equipment :	Spectrophotometer
Manufacturer :	Labtech
Model :	Blue Star A
Serial No. :	1606UV1507
ID No. :	Ins-LAB-004
Condition As-Received:	Used Item
Received Date :	09 April 2024
Calibration Date :	09 April 2024
Reference :	2404-0113OC-2
Submitted by :	Thai Environmental Technic Limited 1/6 Soi Ramkhamhaeng 145, Khwaeng/Khet Saphan Sung, Bangkok 10240
Calibration Place :	Laboratory ( Thai Environment Technic Limited)
Ambient Temperature :	( 29.2 - 31.4 ) °C (On-Site)
Relative Humidity :	( 45.2 - 40.3 ) % (On-Site)
Calibration Procedure :	In - house method : CP-OCH4 based on ASTM E 275-01
Calibrated by :	Saithip Meangmai
Approved by :	 Approved Signatory
( ) Unriopphol Harachai	
(✓) Ponpan Paipim	
( ) Saithip Meangmai	
Issue Date :	17 April 2024

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.



Cert. No. : 24CHO222

Page : 2 of 3

**Condition of calibration result**

**1. Reference Standard Material :**

<u>Material</u>	<u>Serial No.</u>	<u>Certificate No.</u>	<u>Due date</u>
1. Absorbance Standard set	42527	116226	08 Nov 2025
2. Wavelength Standard set	29829	114509	11-Sep 2025
3. Wavelength Standard set	29829	114510	11 Sep 2025
4. Stray Light Standard set	14004	108964	01 Feb 2025

2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certificate is traceable to the International System of Unit maintained through :

→ Starna Scientific Ltd.

4. Spectral BandWidth : 2 nm

Scan Speed : Slow

**Calibration Results : without adjustment**

**Wavelength Accuracy**

<b>Certified Values of Reference Material ( nm )</b>	<b>UUC Reading ( nm )</b>	<b>Uncertainty of Measurement ( ± nm )</b>	<b>Coverage Factor k</b>
361.00	360.6	0.16	2.00
472.47	471.6	0.16	2.00
536.66	536.2	0.16	2.00
748.48	748.4	0.16	2.00
879.27	879.0	0.16	2.00





Cert. No. : 24CHO222

Page : 3 of 3

**Calibration Results : without adjustment**

**Photometric Accuracy**

Wavelength (nm)	Certified Values of Reference Material ( Abs )	UUC Reading ( Abs )	Uncertainty of Measurement ( $\pm$ Abs )	Coverage Factor <i>k</i>
420.0	Zero	0.0002	0.0028	2.00
	0.5739	0.5722	0.0028	2.00
	0.7085	0.7074	0.0030	2.00
	1.0169	1.0146	0.0028	2.00
546.1	Zero	-0.0001	0.0028	2.00
	0.5214	0.5211	0.0028	2.00
	0.6935	0.6926	0.0030	2.00
	0.9978	0.9960	0.0028	2.00
635.0	Zero	0.0000	0.0028	2.00
	0.5626	0.5623	0.0028	2.00
	0.7577	0.7570	0.0030	2.00
	1.0946	1.0927	0.0028	2.00

**Stray Light**

* Straylight at 260.49 nm $\pm$ 0.11 nm	Reading at 260.49 nm $\pm$ 0.11 nm
Abs	2.2284
%T	0.57

**Remark**

- Each individual filter is measured against the empty filter holder (blank) used to zero the spectrophotometer
- Cut-off wavelength of stray light reference material (Potassium Iodide) at Wavelength
- Result = Pass, if Absorbance > 2.00 Abs and Transmission < 1.0 %T at Wavelength
- \* : Not NSC-ONSC Accredited

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)  
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES  
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250  
TEL.0-2717-3000-29 FAX.0-2719-9484



## Certificate of Calibration

Cert. No.: 24TM619

Page : 1 of 3

Equipment : Incubator

Manufacturer : Memmert

Model : INE 500

Serial No. : E505.0595

ID No. : Ins-LAB-041

Submitted by : Thai Environmental Technic Limited  
1/6 Soi Ramkhamhaeng 145,  
Khwaeng/Khet Saphan Sung,  
Bangkok 10240

Location : Bacteria Room


Received Order : 09 April 2024

Calibration Date : 09 - 10 April 2024

Ambient Temperature : ( 26 ± 10 ) °C

Relative Humidity : ( 50 ± 30 ) %

Calibrated by : Preecha Hlahib

Approved by :   
Approved Signatory

( ) Ponpan Paipim

( ) Suwit Imjai

(✓) Kunchit Promprat

Issue Date : 12 April 2024

**The Uncertainties are for a confidence probability of approximately 95%**

This certificate may not be reproduced other than in full, except with the prior written  
Approval of the head of Corporate Services 3 : Equipment Calibration and Testing Services.



Equipment : Incubator  
Condition As-Received : Used Item  
Reference : 2404-0113OC-3

Cert. No.: 24TM619  
Page : 2 of 3

**Procedure Used :-**

Calibration were conducted using calibration procedure CP-OT02 based on TLAS G-20 according to direct measurement method with Data Acquisition which connected with Resistance Temperature Detector ( RTD ).

The temperature scale used was based on ITS-90.

**Condition of this result of calibration**

**1. Reference standard instrument:-**

<u>Instrument</u>	<u>Serial No.</u>	<u>Cert. No.</u>	<u>Traceable</u>	<u>Due Date</u>
1 ) Data Acquisition	MY49023932	23LM122	TPA	26 Jul 2024

2. This certificate is valid only to the item calibrated on date and place of calibration.

3. This certification is traceable to the International System of Unit.


**Remark :** TPA : Technology Promotion Association ( Thailand - Japan )

**Result of Calibration :-** ( \* ) Without Adjustment

**Function of UUC\* :** Temperature Source

**Fresh air setting :** Close

Environment during calibration		
	Beginning	Finished
Temp. ( °C )	26	26
REL.Humid. ( % )	43	46
AC Supply ( Volt )	220	222



Position :	Ref. Std. ID No.:
1	19-16RTD-01
2	19-16RTD-02
3	19-16RTD-03
4	19-16RTD-04
5	19-16RTD-05
6	19-16RTD-06
7	21-16RTD-07
8	19-16RTD-08
9 (ref.)	19-16RTD-09

**Probe Installation Details :**

a = 5.0 cm  
b = 5.0 cm  
c = 5.0 cm

**Dimension of Chamber :**

D = 0.40 m  
W = 0.56 m  
H = 0.48 m  
Capacity = 0.11 m<sup>3</sup>



Equipment : Incubator  
Condition As-Received : Used Item  
Reference : 2404-0113OC-3  
**Result of Calibration :-** ( \* ) Without Adjustment  
Function of UUC\* : Temperature Source  
Fresh air setting : Close

Cert. No.: 24TM619

Page : 3 of 3

Calibration Point ( °C )	UUC* Setting ( °C )	UUC* Reading ( °C )	Temperature stability ( ± °C )	Temperature uniformity ( °C )	Overall Variation ( °C )	Coverage Factor <i>k</i>
35.0	35.0	35.0	0.022	0.27	0.50	2
41.5	41.5	41.5	0.062	0.29	0.53	2
44.5	44.5	44.5	0.033	0.60	1.2	2

Calibration Point ( °C )	Measured Temperature ( °C )									Uncertainty  ( ± °C )
	Position									
	1	2	3	4	5	6	7	8	9 (ref.)	
35.0	35.037	35.081	35.018	35.039	34.634	34.962	34.620	34.990	34.854	0.30
41.5	41.873	41.868	41.845	41.803	41.479	41.667	41.437	41.684	41.610	0.30
44.5	44.899	44.986	44.845	44.827	43.898	44.270	43.883	44.311	44.410	0.30

**Average\*** : The average of 30 values in each position.

**Temperature stability** : One-half of the greatest maximum difference of measured temperature at any one sensor.

**Temperature uniformity** : The maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady-state conditions.

**Overall Variation** : The Difference of the maximum and minimum measured temperatures throughout observation.

**UUC\*** : Unit Under Calibration

**Note** : The reported uncertainty of measurement was included stability and excluded uniformity .

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor *k*, providing a level of confidence of approximately 95 %.

-o0o-



## Certificate of Calibration

Certificate Number : SPR24100208-9 Page : 1 of 3

Customer : Thai Environmental Technic Limited.  
1/6 Soi Ramkhamhaeng 145, Khwaeng Saphan Sung, Khet Saphan  
Sung, Bangkok 10240, Thailand.

Equipment Name : Vibration Meter  
Manufacturer : InstanTel  
Model : 721A2501/721A2901  
Serial Number : UM10831  
ID. Number : No.06

### Environmental Conditions


Ambient Temperature	: 23 °C $\pm$ 3 °C	Received Date	: 11 Oct 2024
Relative Humidity	: 50 % $\pm$ 15 %	Calibration Date	: 18 Oct 2024
Location of Calibration	: In-Lab	Recommend Due Date	: 18 Oct 2025
Calibration Procedure	: In-House Method	Date of Issue	: 19 Oct 2024

### Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs.

The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr. Munin Khumpum  
Calibration Officer

Approved by :   
( Mr. Pootthipong A. )

Authorized Signatory



## Calibration Report

Certificate Number : SPR24100208-9

Page : 2 of 3

### Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Vibration Calibrator	VC-02	2007014	AV-0032-24	20 Jul 2025

### Traceability

This certification is traceable to the International System of Unit maintained at :  
NIMT - The National Institute of Metrology, Thailand.



## Result of Calibration

Certificate No. : SPR24100208-9

Page : 3 of 3

Frequency Response Performance Test @ 1 mm/s

Unit : mm/s

Frequency (Hz)	STD Reading	UUC. Reading	Error	Uncertainty (±)
20.0	1.008	1.030	0.022	0.042
40.0	1.005	1.027	0.022	0.042
50.0	1.004	1.025	0.021	0.042
80.0	1.003	1.022	0.019	0.042
100.0	1.004	1.023	0.019	0.042
160.0	1.006	1.026	0.020	0.042
200.0	1.009	1.032	0.023	0.042

Linearity Performance Test

Unit : mm/s

Frequency (Hz)	STD Reading	UUC. Reading	Error	Uncertainty (±)
100	0.501	0.519	0.018	0.041
	1.000	1.024	0.024	0.042
	1.501	1.529	0.028	0.044
	2.001	2.032	0.031	0.047
	3.002	3.036	0.034	0.053
	5.001	5.039	0.038	0.070

### Note:

The result of calibration was found accurate as show on date and place of calibration only.

This Certificate is not certified for any commercial transaction.

### Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2$ , providing a level of confidence approximately 95%

- End of Certificate -



## Certificate of Calibration

Certificate Number : SPR24020097-2

Page : 1 of 3

Customer : Thai Environmental Technic Limited.

1/6 Soi Ramkhamhaeng 145, Khwaeng Saphan Sung, Khet Saphan  
Sung, Bangkok 10240, Thailand.

Equipment Name : Vibration

Manufacturer : Instatel

Model : Micromate

Serial Number : UM16046

ID. Number : No.13

### Environmental Conditions

Ambient Temperature :  $23^{\circ}\text{C} \pm 3^{\circ}\text{C}$

Received Date : 07 Feb 2024

Relative Humidity :  $50\% \pm 15\%$

Calibration Date : 12 Feb 2024

Location of Calibration : In-Lab

Recommend Due Date : 12 Feb 2025

Calibration Procedure : In-House Method

Date of Issue : 13 Feb 2024

### Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs.

The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Munin Khumpum

Approved by :

Calibration Officer

( Mr.Yodyaim Chansang )

Authorized Signatory





# Calibration Report

Certificate Number : SPR24020097-2

Page : 2 of 3

## Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Vibration Calibrator	VC-02	2007014	AV-0048-23	13 Aug 2024

## Traceability

This certification is traceable to the International System of Unit maintained at :  
NIMT - The National Institute of Metrology, Thailand.



## Result of Calibration

Certificate No. : SPR24020097-2

Page : 3 of 3

Geophone P/N 721A3301 S/N UM15363

Functional Performance Test@160Hz

Function	STD Reading	UUC. Reading	Error	Uncertainty ( ± )
Velocity (mm/s)	5.003	4.989	-0.014	0.059

Frequency Response Performance Test @ 5 mm/s

Unit : mm/s

Frequency (Hz)	STD Reading	UUC. Reading	Error	Uncertainty ( ± )
10.0	5.014	4.989	-0.025	0.058
20.0	5.012	4.987	-0.025	0.058
50.0	5.010	4.988	-0.022	0.058
80.0	5.008	4.990	-0.018	0.058
100.0	5.005	4.990	-0.015	0.058
160.0	5.008	4.983	-0.025	0.058
200.0	5.012	4.980	-0.032	0.058

Linearity Performance Test

Unit : mm/s<sup>2</sup>

Frequency (Hz)	STD Reading	UUC. Reading	Error	Uncertainty ( ± )
160.0	0.502	0.511	0.009	0.0060
160.0	1.001	1.010	0.009	0.012
160.0	1.503	1.493	-0.010	0.017
160.0	2.002	1.989	-0.013	0.023
160.0	3.005	2.985	-0.020	0.035
160.0	5.003	4.987	-0.016	0.058

### Note:

The result of calibration was found accurate as show on date and place of calibration only.  
This Certificate is not certified for any commercial transaction.

### Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2$ , providing a level of confidence approximately 95%.

- End of Certificate -



Thai Environmental Technic Limited  
บริษัท เทคนิคสิ่งแวดล้อมไทย จำกัด

## Sound Level Meter Calibration Report

Equipment Type : Sound Level Meter  
Calibrator : TENMARS Sound Calibrator TM-100  
Standard : IEC 60942  
Accuracy : 94.0 ±0.3 dB and 114.0±0.5 dB  
Frequency : at 1,000 Hz ±1%  
Calibrator Serial NO. : ST120C0263E

Calibration Date : 31-Jan-2025  
Barometric pressure (mmHg) : 759.0 mmHg  
Temperature (23±3)°C : 25.00 °C  
Relative Humidity(50±15 %) : 50.0 % RH  
Dued Date of Calibrate : 28-Feb-2025

Item	Instrument Calibrated			Reference Acoustic dB	Before Adjust				After Adjust ± dB	Deviation ± dB	Result Calibrate
	Brand	Model	Serial NO.		ครั้งที่ 1	ครั้งที่ 2	ครั้งที่ 3	เฉลี่ย			
57	ACO	6226	160099	94.0	94.0	94.0	94.0	94.0	94.0	0.0	PASS
				114.0	114.0	114.0	114.0	114.0			
58	ACO	6226	160143	94.0	94.0	94.0	94.0	94.0	94.0	0.0	PASS
				114.0	114.0	114.0	114.0	114.0			
59	ACO	6226	160203	94.0	94.1	94.1	94.1	94.1	94.0	0.1	PASS
				114.0	114.1	114.1	114.1	114.1			
60	ACO	6226	160204	94.0	93.8	93.8	93.8	93.8	94.0	0.2	PASS
				114.0	113.7	113.7	113.7	113.7			
61	ACO	6226	160205	94.0	94.1	94.1	94.1	94.1	94.0	0.1	PASS
				114.0	114.2	114.2	114.2	114.2			
62	ACO	6226	160211	94.0	94.1	94.1	94.1	94.1	94.0	0.1	PASS
				114.0	1140.0	1140.0	1140.0	1140.0			
63	ACO	6226	160212	94.0	94.1	94.1	94.1	94.1	94.0	0.1	PASS
				114.0	114.0	114.0	114.0	114.0			
64	ACO	6226	160213	94.0	93.8	93.8	93.8	93.8	94.0	0.2	PASS
				114.0	113.8	113.8	113.8	113.8			
66	ACO	6226	160215	94.0	93.9	93.9	93.9	93.9	94.0	0.1	PASS
				114.0	113.9	113.9	113.9	113.9			
67	ACO	6226	160216	94.0	94.3	94.3	94.3	94.3	94.0	0.3	PASS
				114.0	114.2	114.2	114.2	114.2			

Calibration By :

Approve by :

*[Signature]*  
Kramud M.



## Sound Level Meter Calibration Report

Equipment Type	: Sound Level Meter	Calibration Date	: 2-Jan-2025
Calibrator	: TENMARS Sound Calibrator TM-100	Barometric pressure (mmHg)	: 759.0 mmHg
Standard	: IEC 60942	Temperature (23±3)°C	: 25.00 °C
Accuracy	: 94.0 ±0.3 dB and 114.0±0.5 dB	Relative Humidity(50±15 %)	: 50.0 % RH
Frequency	: at 1,000 Hz ±1%	Dued Date of Calibrate	: 31-Jan-2025
Calibrator Serial NO.	: ST120C0263E		

Item	Instrument Calibrated			Reference Acoustic dB	Before Adjust				After Adjust ± dB	Deviation ± dB	Result Calibrate
	Brand	Model	Serial NO.		ครั้งที่ 1	ครั้งที่ 2	ครั้งที่ 3	เฉลี่ย			
21	ACO	6226	070049	94.0	94.0	94.0	94.0	94.0	94.0	0.0	PASS
				114.0	114.0	114.0	114.0	114.0			
23	RION	NL-21	00487676	94.0	94.1	94.1	94.1	94.1	94.0	0.1	PASS
				114.0	114.1	114.1	114.1	114.1			
25	ACO	6226	100098	94.0	94.2	94.2	94.2	94.2	94.0	0.2	PASS
				114.0	114.2	114.2	114.2	114.2			
26	ACO	6226	100099	94.0	93.9	93.9	93.9	93.9	94.0	0.1	PASS
				114.0	113.9	113.9	113.9	113.9			
28	ACO	6226	100101	94.0	94.1	94.1	94.1	94.1	94.0	0.1	PASS
				114.0	114.1	114.1	114.1	114.1			
29	ACO	6226	100102	94.0	93.9	93.9	93.9	93.9	94.0	0.1	PASS
				114.0	114.0	114.0	114.0	114.0			
30	ACO	6226	100106	94.0	93.9	93.9	93.9	93.9	94.0	0.1	PASS
				114.0	113.9	113.9	113.9	113.9			
31	ACO	6226	110098	94.0	94.2	94.2	94.2	94.2	94.0	0.2	PASS
				114.0	114.2	114.2	114.2	114.2			
32	ACO	6226	110105	94.0	94.1	94.1	94.1	94.1	94.0	0.1	PASS
				114.0	114.1	114.1	114.1	114.1			
34	ACO	6226	110099	94.0	94.2	94.2	94.2	94.2	94.0	0.2	PASS
				114.0	114.2	114.2	114.2	114.2			

Calibration By :

Approve by :

*[Signature]*  
Prasarn M.





Thai Environmental Technic Limited  
บริษัท เทคโนโลยีสิ่งแวดล้อมไทย จำกัด

## Sound Level Meter Calibration Report

Equipment Type	: Sound Level Meter	Calibration Date	: 2-Jan-2025
Calibrator	: TENMARS Sound Calibrator TM-100	Barometric pressure (mmHg)	: 759.0 mmHg
Standard	: IEC 60942	Temperature (23±3)°C	: 25.00 °C
Accuracy	: 94.0±0.3 dB and 114.0±0.5 dB	Relative Humidity(50±15 %)	: 50.0 % RH
Frequency	: at 1,000 Hz ±1%	Dued Date of Calibrate	: 31-Jan-2025
Calibrator Serial NO.	: ST120C0263E		

Item	Instrument Calibrated			Reference Acoustic dB	Before Adjust				After Adjust ± dB	Deviation ± dB	Result Calibrate
	Brand	Model	Serial NO.		ครั้งที่ 1	ครั้งที่ 2	ครั้งที่ 3	เฉลี่ย			
57	ACO	6226	160099	94.0	94.1	94.1	94.1	94.1	94.0	0.1	PASS
				114.0	114.1	114.1	114.1	114.1			
58	ACO	6226	160143	94.0	94.0	94.0	94.0	94.0	94.0	0.0	PASS
				114.0	114.0	114.0	114.0	114.0			
59	ACO	6226	160203	94.0	93.8	93.8	93.8	93.8	94.0	0.2	PASS
				114.0	113.8	113.8	113.8	113.8			
60	ACO	6226	160204	94.0	94.1	94.1	94.1	94.1	94.0	0.1	PASS
				114.0	114.1	114.1	114.1	114.1			
61	ACO	6226	160205	94.0	93.9	93.9	93.9	93.9	94.0	0.1	PASS
				114.0	113.9	113.9	113.9	113.9			
62	ACO	6226	160211	94.0	93.8	93.8	93.8	93.8	94.0	0.2	PASS
				114.0	113.8	113.8	113.8	113.8			
63	ACO	6226	160212	94.0	94.0	94.0	94.0	94.0	94.0	0.0	PASS
				114.0	114.0	114.0	114.0	114.0			
64	ACO	6226	160213	94.0	94.1	94.1	94.1	94.1	94.0	0.1	PASS
				114.0	114.1	114.1	114.1	114.1			
66	ACO	6226	160215	94.0	94.1	94.1	94.1	94.1	94.0	0.1	PASS
				114.0	114.1	114.1	114.1	114.1			
67	ACO	6226	160216	94.0	93.9	93.9	93.9	93.9	94.0	0.1	PASS
				114.0	113.9	113.9	113.9	113.9			

Calibration By : 

Approve by : 



## Certificate of Calibration

Certificate Number : SPR24100208-5

Page : 1 of 3

Customer : Thai Environmental Technic Limited.

1/6 Soi Ramkhamhaeng 145, Khwaeng Saphan Sung, Khet Saphan  
Sung, Bangkok 10240, Thailand.

Equipment Name : Sound Calibrator

Manufacturer : Scarlet Tech

Model : ST-120

Serial Number : ST120C0263E

ID. Number : No.8

### Environmental Conditions

Ambient Temperature :  $23\text{ }^{\circ}\text{C} \pm 3\text{ }^{\circ}\text{C}$

Received Date : 11 Oct 2024

Relative Humidity :  $50\% \pm 15\%$

Calibration Date : 12 Oct 2024

Location of Calibration : In-Lab

Recommend Due Date : 12 Oct 2025

Calibration Procedure : In-House Method

Date of Issue : 13 Oct 2024

### Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs.

The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Nanthawat Wanasit

Calibration Officer

Approved by :

( Mr.Pootthipong A. )

Authorized Signatory



# Calibration Report

Certificate Number : SPR24100208-5

Page : 2 of 3

## Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Measuring Receiver	8902A	2950A02471	E3U2401129	05 Sep 2025
AUDIO Analyzer	8903B	3011A09975	EL02442/24	23 Jan 2025

## Traceability

This certification is traceable to the International System of Unit maintained at :

NA - NA Caltechnologies Co., Ltd.

PCAL - Professional Calibration &amp; Services Co.,Ltd



## Result of Calibration

Certificate No. : SPR24100208-5

Page : 3 of 3

Function : Sound Level Calibrator

UUC Setting ( $\pm$ dB)	Standard Reading (dB)	Error (dB)	Uncertainty ( $\pm$ dB)
94	93.9	0.1	1.5
114	113.9	0.1	1.5

Note:

The result of calibration was found accurate as show on date and place of calibration only.  
This Certificate is not certified for any commercial transaction.


### Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor  $k = 2.00$ , providing a level of confidence approximately 95%.

- End of Certificate -



# CERTIFICATE OF ANALYSIS

<b>Customer Detail:</b> <b>Thai Environmental Technic Ltd.</b>		<b>Production Order Number:</b> 90137639 <b>Material Number:</b> 498800-AL-44 <b>Certification Date:</b> 11-Oct-2016 <b>Expiry Date:</b> 11-Oct-2024	
<b>Cylinder Description:</b> <b>ALU 50 L</b>		<p>The measurement of this reference material is traceable to SI through the reference standard which is traceable to Swiss National Standard of Mass. The Assay of this Standard has been performed in accordance with the EPA Traceability Protocol EPA-600/R-12/531 for the Assay and Certification of Gaseous Calibration Standards using procedure G1. The results are expressed on a mole/mole basis, unless otherwise specified. The reported uncertainty is based on a standard uncertainty multiplied by coverage factor k=2, providing a level of confidence of approximately 95%.</p>	
<b>Certificate Number:</b> <b>3111/16</b>		<b>Analyst:</b>  <b>THITIRAT LOYRAT</b>	
<b>Cylinder Number:</b> <b>D824500</b>			
<b>Nominal Cylinder Content:</b> <b>6.900 M<sup>3</sup></b>		<b>Approve:</b>  <b>SUKANYA KAMUTHARAT</b>	
<b>Nominal Pressure:</b> <b>145.0 Bar</b>			
<b>Valve Outlet:</b> <b>CGA 660 SS</b>		<b>To Re-Order Please Quote:</b> <b>498800-AL-44</b>	
<b>Comment:</b>	<ul style="list-style-type: none"> <li>● It is recommended that this product be not used below 5% of actual contents or should not be used when its gas pressure is below 150psig.</li> <li>● Other impurities that detect by analytical condition of this mixture shall be report if it is more than 10% of minimum minor component.</li> <li>● Keep and use in well-ventilated and secure area.</li> </ul>		

## CERTIFICATE OF ANALYSIS

### Analytical Result

<u>Component</u>	<u>Request Concentration</u>	<u>Certified Concentration</u>	<u>Certified Uncertainty</u>	<u>Method</u>	<u>Assay Date</u>
Carbon Monoxide	400 ppm	404 ppm	± 1 % relative	(6) I-PB-352	03-Oct & 10-Oct-2016
Sulphur Dioxide	400 ppm	406 ppm	± 1 % relative	(6) I-PB-352	03-Oct & 10-Oct-2016
In Nitrogen					

### Reference Standard used in Assay

<u>Reference Standard</u>	<u>Cylinder No.</u>	<u>Concentration</u>	<u>Expired Date</u>
Sulphur Dioxide	D832461	201 ± 1 ppm	12-Apr-2018
Carbon Monoxide	D832461	208 ± 1 ppm	12-Apr-2018
In Nitrogen			

### Analytical Instruments used in Assay

<u>Instrument/Make/Model</u>	<u>Analytical Principle</u>	<u>Last Multipoint Calibration</u>
Digi LAB Excalibur IIE Series	FTIR-SO2	07-Sep & 10-Oct-2016
Digi LAB Excalibur IIE Series	FTIR-CO	03-Oct-2016

#### Method of Analysis

1. Gas Chromatograph
2. Paramagnetic Oxygen Analyser
3. Electrochemical Oxygen Analyser
4. Electrochemical Moisture Analyser
5. Total Hydrocarbon Analyser
6. Other specified

Cylinder Number **D824500**  
 Production Order Number **90137639**

Certification Date: **11-Oct-2016**  
 Expiration Date: **11-Oct-2024**

# CERTIFICATE OF ANALYSIS

<b>Customer Detail:</b> <b>Thai Environmental Technic Ltd.</b>		<b>Production Order Number:</b> 90137638 <b>Material Number:</b> 498700-AL-44 <b>Certification Date:</b> 11-Oct-2016 <b>Expiry Date:</b> 11-Oct-2024	
<b>Cylinder Description:</b> <b>ALU 50 L</b>		The measurement of this reference material is traceable to SI through the reference standard which is traceable to Swiss National Standard of Mass. The Assay of this Standard has been performed in accordance with the EPA Traceability Protocol EPA-600/R-12/531 for the Assay and Certification of Gaseous Calibration Standards using procedure G1. The results are expressed on a mole/mole basis, unless otherwise specified. The reported uncertainty is based on a standard uncertainty multiplied by coverage factor k=2, providing a level of confidence of approximately 95%.	
<b>Certificate Number:</b> <b>3112/16</b>		<b>Analyst:</b>  <b>THITIRAT LOYRAT</b>	
<b>Cylinder Number:</b> <b>D271305</b>			
<b>Nominal Cylinder Content:</b> <b>6.900 M<sup>3</sup></b>		<b>Approve:</b>  <b>SUKANYA KAMUTHARAT</b>	
<b>Nominal Pressure:</b> <b>145.0 Bar</b>			
<b>Valve Outlet:</b> <b>CGA 660 SS</b>		<b>To Re-Order Please Quote:</b> <b>498700-AL-44</b>	
<b>Comment:</b>		<ul style="list-style-type: none"> <li>● It is recommended that this product be not used below 5% of actual contents or should not be used when its gas pressure is below 150psig.</li> <li>● Other impurities that detect by analytical condition of this mixture shall be report if it is more than 10% of minimum minor component.</li> <li>● Keep and use in well-ventilated and secure area.</li> </ul>	

# CERTIFICATE OF ANALYSIS

## Analytical Result

Component	Request Concentration	Certified Concentration	Certified Uncertainty	Method	Assay Date
Carbon Monoxide	800 ppm	793 ppm	± 1 % relative	(6) I-PB-352	04-Oct & 11-Oct-2016
Sulphur Dioxide	800 ppm	804 ppm	± 1 % relative	(6) I-PB-352	04-Oct & 11-Oct-2016
In Nitrogen					

## Reference Standard used in Assay

Reference Standard	Cylinder No.	Concentration	Expired Date
Sulphur Dioxide	118499SG	504.5 ± 2.5 ppm	02-Jul-2018
Carbon Monoxide	113882SG	504.3 ± 1.0 ppm	28-Apr-2019
In Nitrogen			

## Analytical Instruments used in Assay

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
Digi LAB Excalibur IIE Series	FTIR-SO <sub>2</sub>	16-Sep-2016
Digi LAB Excalibur IIE Series	FTIR-CO	04-Oct-2016

### Method of Analysis

1. Gas Chromatograph
2. Paramagnetic Oxygen Analyser
3. Electrochemical Oxygen Analyser
4. Electrochemical Moisture Analyser
5. Total Hydrocarbon Analyser
6. Other specified

Cylinder Number D271305  
Production Order Number 90137638

Certification Date: 11-Oct-2016  
Expiration Date: 11-Oct-2024

Certificate Of Analysis  
Special Gases Mixture

## Customer Details

Name:	Address:	Customer Tag No.:
Thai Environmental Technic Limited.	1/6 Soi Ramkhamhaeng 45, Sapansoong, Khet Saphan Sung, Bangkok 10240	-

## Certificate Details

Number:	2500/23	Date of Issue:	18-Sep-2023	Expiry date:	18-Sep-2027
Material Details					
Production Order:	90179846	Material Code:	608400-SK-44	Cylinder No.:	D636157
Gas content:	5.520 M <sup>3</sup>	Filling pressure:	145 bar	Valve:	CGA 660 SS
Cylinder Owner:	LINDE	Cylinder Material:	Spectra seal	Cylinder Size:	40 L

## Laboratory Report

## Analytical Result

Component	Nominal Concentration	Analysis Result <sup>1</sup>	Uncertainty <sup>2</sup>	Method of Analysis <sup>3</sup>	Assay Date
Sulphur Dioxide In Nitrogen	40.0 ppm	41.1 ppm	± 1% relative	(6) I-PB-352	8-Sep & 18-Sep-23

## Reference Standard used in Assay

Reference Standard	Cylinder number	Concentration	Expiry date:
Sulphur Dioxide In Nitrogen	BOC1506295G	25.35 ± 0.25 ppm	9-Jun-2024

## Analytical Instruments used in Assay

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
FTIR Spectrometers Nicolet iS50	FTIR-SO2	6-Sep-2023

## Recommend usage condition

Minimum utilization: 5% of actual content or before expire date whichever comes first.  
Storage condition: Keep in well ventilation and secure area.

## Comments

When reordering, please quote the material number

## Note:

- All results expressed in this report are on mole/mole basis, unless otherwise specified. The Assay of this Standard has been performed in accordance with the EPA Traceability Protocol EPA-600/R-12/531 for the Assay and Certification of Gaseous Calibration Standards using procedure G1
- The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The measurement of this material is traceable to the SI through the reference gas standard which is traceable to Swiss National Standard of Mass or other recognised national metrology institutes.
- (1) Gas Chromatography, (2) Paramagnetic Oxygen Analyzer, (3) Electrochemical Oxygen Analyzer, (4) Electrochemical Moisture Analyzer, (5) Total Hydrocarbon Analyzer, (6) Other - Specified

Sukanya Parinyaasontorn  
Signatory for and on behalf of Linde (Thailand) Co., Ltd.

Page 1 of 1

This report shall not be reproduced except in full

PB-002/F006

Iss: L/2, 01 August 2023

บริษัท ลินด์ (ประเทศไทย) จำกัด (มหาชน)

หมายเลขประจำตัวผลิตภัณฑ์ 0107537000765

ชั้น 15 อาคารทาวเวอร์ เอ 2/3 หมู่ 14 ถนนบางนา-ตราด กม. 6.5 เขตบางนา

อ.บางพลี จ.สมุทรปราการ 10540 โทรศัพท์ (66) 2338-6100 โทรสาร (66) 2338-6333

โรงงานเวลโกรว์: 105 หมู่ 5 ต.บางพลี อ.บางพลี จ.สมุทรปราการ 24180

โทรศัพท์ (66) 38.570-479-93

โทรสาร (66) 38.570-323

Linde (Thailand) Public Company Limited

PLC Registration no. 0107537000765

15<sup>th</sup> Floor, Bangna Tower A, 2/3 Moo 14, Bangna Trad KM. 6.5 Road, Bangkaew

Bangplee, Samutprakarn 10540, Tel (66) 2338-6100 Fax (66) 2338-6333

Wellgrow Plant : 105 Moo 5, T.Bangsamak, A.Bangpakong, Chachoengsao 24180

Thailand, Tel (66) 38.570-479-93

Fax (66) 38.570-323

Certificate Of Analysis  
Special Gases Mixture

## Customer Details

Name:

Thai Environmental Technic Limited

Address:

1/6 Soi Ramkhamhaeng 45, Sapansoong,  
Khet Saphan Sung, Bangkok 10240

Customer Tag No.:

## Certificate Details

Number:	1734/23	Date of Issue:	5-Jul-2023	Expiry date:	5-Jul-2026
Material Details					
Production Order:	90178560	Material Code:	640300-SK-44	Cylinder No.:	A00917SK
Gas content:	5 520 M <sup>3</sup>	Filling pressure:	145.0 bar	Valve:	CGA 660 SS
Cylinder Owner:	LINDE	Cylinder Material:	Spectra seal	Cylinder Size:	40 L

## Laboratory Report

## Analytical Result

Component	Normal Concentration	Analysis Result <sup>1</sup>	Uncertainty <sup>2</sup>	Method of Analysis <sup>3</sup>	Assay Date
Nitric Oxide	40.0 ppm	40.5 ppm	± 1% relative	(6) I-PB-352	28-Jun & 5-Jul-2023
Other NOx impurity In Nitrogen		Less than 2.0 ppm			

## Reference Standard used in Assay

Reference Standard	Cylinder number	Concentration	Expiry date
Nitric Oxide In Nitrogen	258013SG	25.32 ± 0.25 ppm	13-Dec-2024

## Analytical Instruments used in Assay

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
FTIR Spectrometers Nicolet iS50	FTIR-NO	28-Jun-2023

## Recommend usage condition

Minimum utilization: 5% of actual content or before expire date whichever comes first.

Storage condition: Keep in well ventilation and secure area.

## Comments

When reordering, please quote the material number

## Note:

1. All results expressed in this report are on mole/mole basis, unless otherwise specified. Assay of this standard has been performed in accordance with the EPA Traceability Protocol EPA-600/R-12/531 for the Assay and Certification of Gaseous Calibration Standards using procedure G1.
2. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The measurement of this material is traceable to the SI through the reference gas standard, which is traceable to Swiss National Standard of Mass or other recognised national metrology institutes.
3. (1) Gas Chromatography, (2) Paramagnetic Oxygen Analyzer, (3) Electrochemical Oxygen Analyzer, (4) Electrochemical Moisture Analyzer, (5) Total Hydrocarbon Analyzer, (6) Other - Specified

Page 1 of 1

This report shall not be reproduced except in full

บริษัท ลินด์ (ประเทศไทย) จำกัด (มหาชน)

เลขที่เอกสาร: 0105370000

Box 15, Bangna Tower A, 2/3 Moo 14, Bangna Trad KM 6.5 Road, Bangnaew

Bangkok 10540 โทรสาร (66) 2338-6100 โทร (66) 2338-6333

Wellgrow Plant: 105 Moo 5, T.Bangsamak, A.Bangpakong, Chachoengsao 24180

โทรสาร (66) 38 570 479 93

โทร (66) 38 570 323

Sukanya Parinyasontorn

Signatory for and on behalf of Linde (Thailand) Co., Ltd.

Linde (Thailand) Public Company Limited

เลขที่เอกสาร: 0105370000

15<sup>th</sup> Floor, Bangna Tower A, 2/3 Moo 14, Bangna Trad KM 6.5 Road, Bangnaew

Bangkok, Samutprakarn 10540, Tel (66) 2338 6100 Fax (66) 2338 6333

Wellgrow Plant: 105 Moo 5, T.Bangsamak, A.Bangpakong, Chachoengsao 24180

Thailand, Tel (66) 38 570 479 93

Fax (66) 38 570 323