ภาคผนวกที่ 4

ใบรับรองการสอบเทียบเครื่องมือ

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

NSC-TISI-TIS17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : **67-400505-6** **Page : 1 of 2**

Submitted by : M E T Company Limited
36/659 Moo 6, T.Bangrakpattana, A.Bangbuatong, Nonthaburi 11110

Equipment : Temperature controlled enclosure (Refrigerator)
Manufacturer : Sanden Intercool Model : SRR3-0687 AR
Range : N/A °C Resolution : 1 °C
Serial No. : SRR3675A-201200111 R ID No. : MET-RE05/64

Environment : On site calibration was carried out at the Laboratory, M E T Company Limited
Ambient Temperature : (29.0 to 30.0) °C
Relative Humidity : (55 to 60) %
Line Voltage : (210.0 to 210.8) V

Date of Received : 20 August 2024
Date of Calibration : 20 August 2024
Date of Issue : 21 August 2024
Calibrated by : Permpon Chanpu
Calibration Method : CAL-M4004, TLAS G-20
The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units
Standard Digital Thermometer with RTD Probe

ID No.	Cert. No.	Due Date	Traceability
400046 & 400042	67-400442-1	26 Jan 2025	National Institute of Metrology Thailand (NIMT)

The Uncertainties are for a confidence probability of approximately 95%

This cerificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

Certificate of Calibration


Certificate No. : 67-400505-6

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber

W = 0.58 m

D = 0.60 m

H = 1.45 m

Capacity = 0.50 m^3

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (±°C)
			1	2	3	4	5	6	7	8	9	
3	3	3	4.68	3.63	4.12	3.05	3.11	2.96	2.97	2.89	3.02	0.95

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Uniformity (°C)	Measured Stability (°C)	Overall Variation (°C)
3	3	3	1.77	0.33	2.23

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2 , providing a level of confidence of approximately 95%

- o0o -

CAL

Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

NSC-TISI-TIS17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : **68-420008-1** **Page : 1 of 2**

Submitted by : M E T Company Limited
36/659 Moo 6, T.Bangrakpattana, A.Bangbuatong, Nonthaburi 11110

Equipment : pH Meter with electrode
pH meter
Manufacturer : Eutech Model : pH 150
Range : -2.00 to 16.00 pH Resolution : 0.01 pH
Serial No. : 2657036 ID No. : MET-PH04/60
Electrode
Model : ECFC7252101B Serial No. : 2603242122
ID No. : MET-PH04/60

Environment : Ambient Temperature : (25 ± 2) °C
Relative Humidity : (50 ± 15) %

Date of Received : 16 January 2025

Date of Calibration : 23 January 2025

Date of Issue : 23 January 2025

Calibrated by : Permpon Chanpu

Calibration Method : In-house method CAL-M4201 direct measurement by using standard voltage calibrator and using certified reference material (CRM)

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Multiproduct Calibrator

ID No.	Cert. No.	Due Date	Traceability
440001	23E1240	24 Mar 2025	National Institute of Metrology Thailand (NIMT)

2. Standard Buffer Solution

pH	Cert. No.	Lot No.	Exp. Date	Traceability
4.008	61293328	1027612	15 Sep 2026	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025
6.987	61297593	1027614	15 Sep 2025	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025
10.010	61306165	1027613	15 Sep 2025	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025

The Uncertainties are for a confidence probability of approximately 95%

This cerificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

Certificate of Calibration

Certificate No. : 68-420008-1

Result of Calibration :

UUC Condition As-Received : Good

Function : Electrical measurement

pH meter

Performing standard curve by Multiproduct Calibrator at pH (4,7,10)

Adjustment Curve at nominal pH	Applied Voltage (mV)	Nominal Value (pH)	UUC Reading		Correction (mV)	Uncertainty (± mV)
			(pH)	(mV)		
4, 7, 10	177.4800	4	4.00	177.1	0.4	0.060
	0.0000	7	6.98	-0.1	0.1	0.060
	-177.4800	10	10.00	-177.3	-0.2	0.060

Function : pH meter with electrode

Performing a three - buffer standard curve using buffer nominal pH (4,7,10)

Adjustment Curve at nominal pH	Standard Buffer (pH)	UUC Reading (pH)	Correction (pH)	Uncertainty (± pH)
4, 7, 10	4.008	4.01	0.00	0.0097
	6.987	7.00	-0.01	0.011
	10.010	10.01	0.00	0.014

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurment was based on a standard uncertainty multiplied by a coverage factor k = 2 , providing a level of confidence of approximately 95%

- o0o -

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

NSC-TISI-TIS17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : **67-400708-1** **Page : 1 of 2**

Submitted by : M E T Company Limited
36/659 Moo 6, T.Bangrakpattana, A.Bangbuatong, Nonthaburi 11110

Equipment : Temperature controlled enclosure (Incubator)
Manufacturer : M-LAB Model : BIC-140
Range : N/A °C Resolution : 0.1 °C
Serial No. : 1022 ID No. : MET-BI02/64

Environment : On site calibration was carried out at the Laboratory, M E T Company Limited
Ambient Temperature : (24.0 to 24.8) °C
Relative Humidity : (50 to 55) %
Line Voltage : (220.0 to 221.8) V

Date of Received : 20 December 2024
Date of Calibration : 20 December 2024
Date of Issue : 20 December 2024
Calibrated by : Permpon Chanpu
Calibration Method : CAL-M4004, TLAS G-20
The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Digital Thermometer with RTD Probe

ID No.	Cert. No.	Due Date	Traceability
400046 & 400047	67-400442-2	27 Jan 2025	National Institute of Metrology Thailand (NIMT)

The Uncertainties are for a confidence probability of approximately 95%

This cerificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

Certificate of Calibration


Certificate No. : 67-400708-1

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber

W = 0.37 m

D = 0.33 m

H = 1.14 m

Capacity = 0.14 m^3

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (±°C)
			1	2	3	4	5	6	7	8	9	
20.0	20.0	20.0	20.13	19.70	19.68	19.68	20.35	20.19	20.15	20.10	20.38	0.35

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Uniformity (°C)	Measured Stability (°C)	Overall Variation (°C)
20.0	20.0	20.0	0.8	0.1	0.8

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2 , providing a level of confidence of approximately 95%

- o0o -

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

NSC-TISI-TIS17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 67-200293-1 **Page : 1 of 2**

Submitted by : M E T Company Limited
36/659 Moo 6, T. Bangrakpattana, A. Bangbuatong, Nonthaburi 11110

Equipment : Electronic Balance

Manufacturer :	Sartorius	Model :	BSA224S-CW
Serial No. :	35090472	ID No. :	MET-EB 02/60
Capacity :	220 g	Resolution :	0.0001 g

Environment : On site calibration was carried out at the Laboratory, M E T Company Limited

Ambient Temperature :	(25.9 to 26.2)	°C
Relative Humidity :	(32.6 to 45.5)	%
Air Pressure :	1007.0	mbar

Date of Received : 20 August 2024
Date of Calibration : 20 August 2024
Date of Issue : 21 August 2024
Calibrated by : Akaradath Thippichai

Calibration Method : In-house method CAL-M2001 based on UKAS Publication ref : LAB 14 Edition 7 - November 2022

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Weights

ID No.	Cert. No.	Due Date	Traceability
E261-E2624	C02232088	08 Nov 2024	National Institute of Metrology (Thailand), (NIMT)

The Uncertainties are for a confidence probability of approximately 95%

This cerificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

Certificate of Calibration

Certificate No. : 67-200293-1

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Departure of indication from nominal value

Nominal Value (g)	Correction (g)	Uncertainty ± (g)
0.05	0.0001	0.00012
0.1	0.0001	0.00012
0.5	0.0001	0.00013
1	0.0000	0.00013
5	0.0000	0.00013
10	0.0000	0.00013
50	0.0001	0.00015
100	0.0000	0.00020
150	0.0000	0.00038
200	0.0001	0.00038


This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty mutiplied by a coverage factor k = 2.06 , providing a level of confidence of approximately 95%

Eccentric error

Load test : 50 g

A	B	C	D	E	
0.0001	-0.0001	-0.0001	0.0001	0.0000	g



Repeatability

Load test : 200 g

Stdev. : 0.00005 g

- o0o -

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

NSC-TISI-TIS17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 68-200092-2 **Page : 1 of 2**

Submitted by : M E T Company Limited
36/659 Moo 6, T.Bangrakpattana, A.Bangbuatong, Nonthaburi 11110

Equipment : Electronic Balance

Manufacturer :	AND	Model :	FX-2000i
Serial No. :	15639789	ID No. :	MET-EB03/61
Capacity :	2200 g	Resolution :	0.01 g

Environment : On site calibration was carried out at the Laboratory, M E T Company Limited

Ambient Temperature :	(26.0 to 26.2)	°C
Relative Humidity :	(48.4 to 51.4)	%
Air Pressure :	1012.0	mbar

Date of Received : 10 March 2025

Date of Calibration : 10 March 2025

Date of Issue : 11 March 2025

Calibrated by : Akaradath Thippichai

Calibration Method : In-house method CAL-M2001 based on UKAS Publication ref : LAB 14 Edition 7 - November 2022

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Weights

ID No.	Cert. No.	Due Date	Traceability
F181-F1821	67-210457-1	12 Jun 2025	National Institute of Metrology (Thailand), (NIMT)

The Uncertainties are for a confidence probability of approximately 95%

This cerificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

Certificate of Calibration

Certificate No. : 68-200092-2

Result of Calibration : After Adjustment

UUC Condition As-Received : Good

Departure of indication from nominal value

Nominal Value (g)	Correction (g)	Uncertainty ± (g)	Error before Adjustment (g)
200	0.00	0.011	0.00
500	0.00	0.011	0.00
600	0.00	0.011	0.00
700	0.00	0.011	0.00
800	0.00	0.011	0.00
1000	0.01	0.011	0.00
1200	0.01	0.012	0.01
1500	0.01	0.012	0.02
2000	0.00	0.014	0.03
2200	0.00	0.023	0.06


CAL

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty mutiplied by a coverage factor k = 2.00 , providing a level of confidence of approximately 95%

Eccentric error Load test : 500 g

A	B	C	D	E	
0.00	-0.01	-0.01	0.00	0.00	g



Repeatability Load test : 2000 g

Stdev. : 0.000 g

- o0o -

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

NSC-TISI-TIS17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : **67-400505-2** **Page : 1 of 2**

Submitted by : M E T Company Limited
36/659 Moo 6, T.Bangrakpattana, A.Bangbuatong, Nonthaburi 11110

Equipment : Temperature controlled enclosure (Oven)
Manufacturer : Binder Model : ED53
Range : N/A °C Resolution : 1 °C
Serial No. : 13-07419 ID No. : MET-OV02/57

Environment : On site calibration was carried out at the Laboratory, M E T Company Limited
Ambient Temperature : (30.0 to 31.0) °C
Relative Humidity : (50 to 55) %
Line Voltage : (210.0 to 210.8) V

Date of Received : 20 August 2024
Date of Calibration : 20 August 2024
Date of Issue : 21 August 2024
Calibrated by : Permpon Chanpu
Calibration Method : CAL-M4004, TLAS G-20
The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units
Standard Digital Thermometer with Thermocouple probe

ID No.	Cert. No.	Due Date	Traceability
400029 & 400030	67-400246-1	25 Oct 2024	National Institute of Metrology Thailand (NIMT)

The Uncertainties are for a confidence probability of approximately 95%

This cerificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

Certificate of Calibration


Certificate No. : 67-400505-2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber
W = 0.40 m
D = 0.33 m
H = 0.40 m
Capacity = 0.05 m^3

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (±°C)
			1	2	3	4	5	6	7	8	9	
104	110	110	103.4	105.0	104.8	105.0	104.1	103.8	104.2	104.4	104.2	0.96
180	184	184	179.3	182.0	180.1	180.6	180.1	180.4	180.0	180.7	179.9	1.3

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Uniformity (°C)	Measured Stability (°C)	Overall Variation (°C)
104	110	110	1.1	0.3	2.0
180	184	184	2.5	0.4	3.3

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2 , providing a level of confidence of approximately 95%

- o0o -

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

NSC-TISI-TIS17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : **67-400505-1** **Page : 1 of 2**

Submitted by : M E T Company Limited
36/659 Moo 6, T.Bangrakpattana, A.Bangbuatong, Nonthaburi 11110

Equipment : Temperature controlled enclosure (Oven)
Manufacturer : Memmert Model : UM 100
Range : N/A °C Resolution : 0.1 °C
Serial No. : b197.0985 ID No. : MET-OV01/46

Environment : On site calibration was carried out at the Laboratory, M E T Company Limited
Ambient Temperature : (30.0 to 31.0) °C
Relative Humidity : (50 to 55) %
Line Voltage : (210.0 to 210.8) V

Date of Received : 20 August 2024
Date of Calibration : 20 August 2024
Date of Issue : 21 August 2024
Calibrated by : Permpon Chanpu
Calibration Method : CAL-M4004, TLAS G-20
The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units
Standard Digital Thermometer with Thermocouple probe

ID No.	Cert. No.	Due Date	Traceability
400029 & 400032	67-400247-1	26 Oct 2024	National Institute of Metrology Thailand (NIMT)

The Uncertainties are for a confidence probability of approximately 95%

This cerificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

Certificate of Calibration


Certificate No. : 67-400505-1

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber

W = 0.32 m

D = 0.18 m

H = 0.24 m

Capacity = 0.01 m^3

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (±°C)
			1	2	3	4	5	6	7	8	9	
180.0	180.0	180.0	181.6	181.2	181.9	180.7	180.7	181.9	179.2	179.1	180.8	0.95

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Uniformity (°C)	Measured Stability (°C)	Overall Variation (°C)
180.0	180.0	180.0	1.9	0.2	3.1

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2 , providing a level of confidence of approximately 95%

- o0o -

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

NSC-TISI-TIS17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : **68-300172-2** **Page : 1 of 2**

Submitted by : M E T Company Limited
36/659 Moo 6, T. Bangrakpattana, A. Bangbuatong, Nonthaburi 11110

Equipment : Burette
Manufacturer : Witeg Class : B
Capacity : 50 ml Graduation : 0.1 ml
ID No. : MET-BU50:01/64

Environment : Ambient Temperature : (20 ± 3) °C
Relative Humidity : (50 ± 10) %
Air Pressure : 1016.8 mbar.

Date of Received : 20 February 2025
Date of Calibration : 24 February 2025
Date of Issue : 24 February 2025
Calibrated by : Wipa Tovadee

Calibration Method : In-house method CAL-M3001 based on ASTM E 542-22

Reference Standard Instruments : This certification is traceable to the International System of Units

Electronic Balance

ID No.	Cert. No.	Due Date	Traceability
241003	67-200410-2	02 Jun 2025	National Institute of Metrology (Thailand) (NIMT)

The Uncertainties are for a confidence probability of approximately 95%

This cerificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

Certificate of Calibration

Certificate No. : 68-300172-2

Result of Calibration : This result of true Volume is referred to standard temperature at 20 °C

UUC Condition As-Received : Good

Delivery Time : 32.90 sec.

Nominal Volume (ml)	Measuring Volume (ml)
10	10.0231
25	24.9988
50	49.9766

Uncertainty of measurement with in ± 0.011 ml

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2.00 , providing a level of confidence of approximately 95%

- o0o -

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

ilac-MRA
THAILAND
NSC-TISI-TIS17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 67-410047-1 **Page : 1 of 2**

Submitted by : M E T Company Limited
36/659 Moo 6, T. Bangrakpattana, A. Bangbuatong, Nonthaburi 11110

Equipment : Digital Thermo-Hygrometer

Manufacturer :	Digicon	Model :	TH-03A
Range Temperature :	-10 °C to 50 °C	Resolution :	0.1 °C
Range Humidity :	20 %R.H. to 99 %R.H.	Resolution :	1 %R.H.
Serial No. :	425048554	ID No. :	MET-TH07/66

Environment : Ambient Temperature : (23 ± 2) °C
Relative Humidity : (50 ± 15) %

Date of Received : 25 April 2024
Date of Calibration : 27 April to 30 April 2024
Date of Issue : 30 April 2024
Calibrated by : Chortip Samchusri

Calibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4013 by compared with standard probe sensor humidity/temperature into humidity/temperature chamber.

Reference Standard Instruments : This certification is traceable to the International System of Units

Digital Indicator with Standard Probe Temp&Hum

ID No.	Cert. No.	Due Date	Traceability
400034 & 400036	SG-H-00020/67	05 Jul 2024	Success Gateway Co., Ltd., Accredited by TISI Calibration No.0268

The Uncertainties are for a confidence probability of approximately 95%

This cerificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

Certificate of Calibration

Certificate No. : 67-410047-1

UUC Condition As-Received : Good

Result of Calibration : Without Adjustment

Function : Temperature measurement (Mode : In)

Reference Humidity @ 50 %R.H.

Standard Temperature (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
19.96	20.4	-0.4	0.46
25.02	25.3	-0.3	0.46
30.01	29.8	0.2	0.46

Result of Calibration : Without Adjustment

Function : Humidity measurement

Reference Temperature @ 25 °C

Standard Humidity (%R.H.)	UUC Reading (%R.H.)	Correction (%R.H.)	Uncertainty (± %R.H)
40.01	47	-7	2.2
49.99	56	-6	2.2
59.97	65	-5	2.3

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2 , providing a level of confidence of approximately 95%

- oOo -

BUCHI (Thailand) Ltd.
ASEAN Competence Center

77/121 Sin Sathorn Tower
28th FL, Krungthonburi Rd.
Klongtonsai, Klongsan
Bangkok 10600
P +662 862 0851
F +662 862 0854
bacc@buchi.com
www.buchi.com

Preventive Maintenance Scrubber

Service No. PM25-S08-106

1. Customer Information

Customer Name	Instrument	Serial Number	Service Date
บริษัท เอ็ม อี ที จำกัด 36/659 หมู่ที่ 6 ถนนตลิ่งชัน-สุพรรณฯ ตำบลบางรักพัฒนา อำเภอบางบัวทอง นนทบุรี 11110 ติดต่อคุณ ศศิธร 087 4985239	K-415	1000281005	30 Apr 2025 PM 2/2

2. Instrument

2.1 Cooling water (If it connected)	OK	NOT OK	Remark
- Temperature 10 – 20 °C	/		
- Cooling water inlet	/		
- Cooling water outlet	/		

2.2 Cleaning	DONE	NOT DONE	Remark
- Housing	/		
- Condenser	/		
- Swirl disc	/		

BUCHI (Thailand) Limited

BUCHI (Thailand) Ltd.
ASEAN Competence Center

77/121 Sin Sathorn Tower
28th FL, Krungthonburi Rd.
Klongtonsai, Klongsan
Bangkok 10600
P +662 862 0851
F +662 862 0854
bacc@buchi.com
www.buchi.com

Preventive Maintenance Scrubber

2.3 Visual Check	OK	NOT OK	Remark
- Hose connection to suction	/		
- Glassware	/		
- Lip gasket	/		
- GL-14 connector	/		
- Activated charcoal	/		

2.4 Flush Pump

Make sure, the bypass valve is closed completely (for maximum suction power).

- Disconnect the silencer, move it down (or take it away from the instrument), and flush out the pump with at least 500 mL of distilled water through the pump inlet, until the collected washing water is clean.
- Switch on the instrument and collect the waste water from the pump output in a suitable vessel.

Flush pump ☑ **OK** ☐ **NOT OK**

2.5 Washing Solution

- Sodium hydroxide 8-10 %, max. 20 %
- Sodium carbonate
  - o dissolve 600 g Na_2CO_3 in 3 L distilled warm water, or
  - o dissolve 1.7 kg Na_2CO_3 in 10 H_2O in 3 L distilled warm water

Washing solution ☑ **OK** ☐ **NOT OK**

BUCHI (Thailand) Limited

BUCHI (Thailand) Ltd.
ASEAN Competence Center

77/121 Sin Sathorn Tower
28th FL, Krungthonburi Rd.
Klongtonsai, Klongsan
Bangkok 10600
P +662 862 0851
F +662 862 0854
bacc@buchi.com
www.buchi.com

Preventive Maintenance Scrubber

3. Summary

All specifications OK	Specification not OK
OK	

Comments

- เครื่องทำงานปกติ

- PM 2/2 [illegible] 4Hg

M 1/2 [illegible] 4Hg

/2 [illegible] 4Hg

[illegible] 4Hg

4Hg

g

Signature BUCHI

30 Apr 2025

02 May 2025

BUCHI (Thailand) Limited

เลขที่ 36/659 ม.6 ต.บางรักพัฒนา อ.บางบัวทอง จ.นนทบุรี 11110 Tel. : 0 29201458-9 Fax. : 0 2920 1460 Email : met_jj@yahoo.com