

ภาคผนวก จ

เอกสารรับรองเครื่องมือการตรวจวัด

NSC-764-765-7025
CALIBRATION 0000

Certificate of Calibration

Certificate No. : 67-420106-1

Page : 1 of 2

Submitted by :

M Green Group Co., Ltd.

188/46 Wisetukhakhon 25, Pracha-Uttid Rd., Thungkru Bangkok 10140 Thailand

Equipment :

pH Meter with electrode

pH meter

Manufacturer : Eutech

Model : pH 700

Range : N/A pH

Resolution : 0.01 pH

Serial No. : 2884323

ID No. : N/A

Electrode

Model : ECFC7252101B

Serial No. : 01X099320

Environment : On site calibration was carried out at the Laboratory, M Green Group Co., Ltd.

Ambient Temperature : (26.3 to 26.7) °C

Relative Humidity : (40 to 45) %

Date of Received : 20 September 2024

Date of Calibration : 20 September 2024

Date of Issue : 21 September 2024

Calibrated by : Permpon Chanpu

Calibration Method : In-house method CAL-M4201 direct measurement by using standard voltage calibrator and using certified reference material (CRM)

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Multiproduct Calibrator

ID No.	Cert. No.	Due Date	Traceability
400005	SG-E-00307/66	23 Aug 2025	National Institute of Metrology Thailand (NIMT)

2. Standard Buffer Solution

pH	Cert. No.	Lot No.	Exp. Date	Traceability
4.008	61293328	1027612	15 Sep 2026	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025
6.987	61297593	1027614	15 Sep 2025	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025
10.010	61306165	1027613	15 Sep 2025	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025

Approved by :

(Permpon Chanpu)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%.

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.



Certificate of Calibration

Certificate No. : 67-420106-1

Page : 2 of 2

Result of Calibration :

UUC Condition As-Received : Good

Function : Electrical measurement

pH meter

Performing standard curve by Multiproduct Calibrator at pH (4,7,10)

Adjustment Curve at nominal pH	Applied Voltage (mV)	Nominal Value (pH)	UUC Reading		Correction (mV)	Uncertainty (± mV)
			(pH)	(mV)		
4, 7, 10	177.4800	4	4.00	177.5	0.0	0.12
	0.0000	7	7.00	0.0	0.0	0.086
	-177.4800	10	10.00	-177.4	-0.1	0.12

Function : pH meter with electrode

Performing a three - buffer standard curve using buffer nominal pH (4,7,10)

Adjustment Curve at nominal pH	Standard Buffer (pH)	UUC Reading (pH)	Correction (pH)	Uncertainty (± pH)
4, 7, 10	4.008	4.01	0.00	0.0097
	6.987	7.00	-0.01	0.011
	10.010	10.01	0.00	0.014

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2 ,
providing a level of confidence of approximately 95%

- o f o -

D



Certificate of Calibration

Certificate No. : 67-400547-1 Page : 1 of 2

Submitted by :

M Green Group Co., Ltd.

188/46 Wisetuekhaekhon 25, Pracha-Uitd Rd., Thungkru Bangkok 10140 Thailand

Equipment :

Temperature Indicator with Thermistor Probe

Temperature Indicator

Manufacturer : Eutech Model : pH 700

Range : N/A °C Resolution : 0.1 °C

Serial No. : 2884323 ID No. : N/A

Thermistor probe

Model : N/A Sheath Material : Stainless

Diameter : 3 mm. Length : 115 mm.

Serial No. : PH5TEM/B01P 279 ID No. : N/A

Environment : On site calibration was carried out at the Laboratory, M Green Group Co., Ltd.

Ambient Temperature : (26.3 to 26.7) °C

Relative Humidity : (40 to 45) %

Line Voltage : (220.0 to 223.0) VAC

Date of Received : 20 September 2024

Date of Calibration : 20 September 2024

Date of Issue : 21 September 2024

Calibrated by : Permpun Champu

Calibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4003 by compared with PRT in the liquid bath at the constant controlled temperature.

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Platinum Resistance Thermometer (PRT)

ID No. Cert. No. Due Date Traceability

400002 TT-0095-24 01 Jul 2026 National Institute of Metrology Thailand (NIMT)

2. Standard Digital Thermometer

ID No. Cert. No. Due Date Traceability

400033 24E633 21 Feb 2026 National Institute of Metrology Thailand (NIMT)

Approved by :

(Permpun Champu)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.



Certificate of Calibration

Certificate No. : 67-400547-1 Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

Immersion Depth (mm.)	Standard Reading (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
130	25.004	24.9	0.1	0.19

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- o/lo -





Certificate of Calibration

Certificate No. : 67-200334-1

Page : 1 of 2

Submitted by :

M Green Group Co., Ltd.

188/46 Wisetusuknakdion 25, Pracha-Ud Rd., Thungkru, Bangkok 10140 Thailand

Equipment :

Electronic Balance

Manufacturer : SHIMADZU Model : AP225WD

Serial No. : D316300690

Capacity : 220 g Resolution : 0.00001g/102g, 0.0001g/220g

Environment :

On site calibration was carried out at the Laboratory, M Green Group Co., Ltd.

Ambient Temperature : (24.9 to 26.3) °C

Relative Humidity : (40.8 to 55.4) %

Air Pressure : 1003.0 mbar

Date of Received :

20 September 2024

Date of Calibration :

20 September 2024

Date of Issue :

21 September 2024

Calibrated by :

Akaradath Thippichai

Calibration Method :

In-house method CAL-M2001 based on UKAS Publication ref : LAB 14

Edition 7 - November 2022

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Weights

ID No.

Cart. No.

Due Date

Traceability

E261-E2624

C02232088

08 Nov 2024

National Institute of Metrology (Thailand), (NIMT)

Approved by :

(Saia Sangthum)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%.

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.



Certificate of Calibration

Certificate No. : 67-200334-1

Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Departure of indication from nominal value

Nominal Value (g)	Correction (g)	Uncertainty $\pm$ (g)
0.001	0.00000	0.000015
0.01	0.00000	0.000016
0.1	0.00000	0.000018
1	0.00000	0.000026
10	0.00000	0.000053
20	-0.00003	0.000071
50	0.00006	0.00011
100	0.00004	0.00020
150	0.0001	0.00038
200	0.0000	0.00038

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2.13$, providing a level of confidence of approximately 95%

Eccentric error

Load test : 50 g

A B C D E

-0.00005 0.00001 0.00002 0.00000 0.00000 g

C
B E D
A

Repeatability

Load test : 200 g

Sidev. : 0.000000 g

-o0o-

Saia



CAL

Calibratech Co., Ltd.

7/106-7 Moo 2, Suddhaphrasaen 3 Rd., Banggood, Pakkred, Nonthaburi 11120
Tel: (02) 964-6211 Fax: (02) 964-5155, e-mail: calibratech.co@yahoocom, calibratech.co@btmail.com

Certificate of Calibration

Certificate No. : 67-410112-1

Page : 1 of 2

Submitted by : M Green Group Co., Ltd.

188/46 Watsueknaikhon 25, Pracha-Uthit Rd., Thunglu, Bangkok 10140 Thailand

Equipment : Digital Thermo-Hygrometer

Manufacturer : Digicon

Model : TH-02A

Range Temperature : 0 °C to 50 °C Resolution : 0.1 °C

Range Humidity : 20 %R.H. to 99 %R.H. Resolution : 1 %R.H.

Serial No. : 1819A071796 ID No. : N/A

Environment : Ambient Temperature : (23 ± 2) °C

Relative Humidity : (50 ± 15) %

Date of Received : 20 September 2024

Date of Calibration : 23 September to 24 September 2024

Date of Issue : 24 September 2024

Calibrated by : Chortip Sanchusri

Calibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4013 by compared with standard probe sensor humidity/temperature into humidity/temperature chamber.

Reference Standard Instruments : This certification is traceable to the International System of Units

Digital Indicator with Standard Probe Temp&Hum

ID No. Cert. No. Due Date Traceability

400034 & 400015 SG-H-00611/67 04 Jan 2025 Success Gateway Co., Ltd., Accredited by TISI Calibration No.0268

Approved by :

(Permpoon Champo)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.

CAL-F0031-03



CAL

Calibratech Co., Ltd.

7/106-7 Moo 2, Suddhaphrasaen 3 Rd., Banggood, Pakkred, Nonthaburi 11120
Tel: (02) 964-6211 Fax: (02) 964-5155, e-mail: calibratech.co@yahoocom, calibratech.co@btmail.com

Certificate of Calibration

Certificate No. : 67-410112-1

Page : 2 of 2

UUC Condition As-Received : Good

Result of Calibration : Without Adjustment

Function : Temperature measurement

Reference Humidity @ 50 %R.H.

Standard Temperature (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
19.99	19.7	0.3	0.46
25.00	24.7	0.3	0.46
30.01	29.7	0.3	0.46

Result of Calibration : Without Adjustment

Function : Humidity measurement

Reference Temperature @ 25 °C

Standard Humidity (%R.H.)	UUC Reading (%R.H.)	Correction (%R.H.)	Uncertainty (± %R.H.)
40.01	38	2	2.2
60.00	56	4	2.3

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

- ๐๕๐ -

P



CAL-F0031-03



Certificate of Calibration

Certificate No. : 67-400548-1

Page : 1 of 2

Submitted by : M Green Group Co., Ltd.

18846 Wisatesukkhon 25, Pracha-Utd Rd., Thungkru Bangkok 10140 Thailand

Equipment : Temperature controlled enclosure (Refrigerator)

Manufacturer : Biohase

Model : BXC-V230M (II)

Range : N/A °C

Resolution : 0.1 °C

Serial No. : YC025025190108

ID No. : N/A

Environment : On site calibration was carried out at the Laboratory, M Green Group Co., Ltd.

Ambient Temperature : (26.3 to 27.4) °C

Relative Humidity : (39 to 47) %

Line Voltage : (226.0 to 230.0) V

Date of Received : 20 September 2024

Date of Calibration : 20 September 2024

Date of Issue : 21 September 2024

Calibrated by : Permpon Chamru

Calibration Method : CAL-M4004, TLAS G-20

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Digital Thermometer with RTD Probe

ID No. Cert. No.

Traceability

400046 & 400042 67-400442-1

Due Date

National Institute of Metrology Thailand (NIMT)

26 Jan 2025

Approved by :

(Permpon Chamru)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.

Certificate of Calibration

Certificate No. : 67-400548-1


Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber

W = 0.50 m

D = 0.40 m

H = 1.20 m

Capacity = 0.24 m³

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (± °C)
			1	2	3	4	5	6	7	8	9	
4.0	2.0	2.0	4.70	4.42	5.04	4.99	4.52	4.22	4.74	4.67	3.89	0.45

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured		Overall Variation (°C)
			Uniformity (°C)	Stability (°C)	
4.0	2.0	2.00	1.23	0.13	1.40

Remark: The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

-o0o-





Certificate of Calibration

Page : 1 of 2

Certificate No. : 67-400548-2

Submitted by :

M Green Group Co., Ltd.

188/46 Wisatsukhakhon 25, Pracha-Utd Rd., Thungkru Bangkok 10140 Thailand

Equipment :

Temperature controlled enclosure (Oven)

Manufacturer : Memmert

Model : UF110

Range : N/A °C

Resolution : 0.1 °C

ID No. : N/A

Serial No. : B419.1092

Environment :

On site calibration was carried out at the Laboratory, M Green Group Co., Ltd.

Ambient Temperature : (26.3 to 27.4) °C

Relative Humidity : (39 to 47) %

Line Voltage : (226.0 to 230.0) V

Date of Received :

20 September 2024

Date of Calibration :

20 September 2024

Date of Issue :

21 September 2024

Calibrated by :

Permpoon Chanpu

Calibration Method :

CAL-M4004, TLAS G-20

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Digital Thermometer with Thermocouple probe

ID No.

Cert. No.

Due Date

Traceability

400046 & 400023

67-400198-1

01 Oct 2024 National Institute of Metrology Thailand (NIMT)

Approved by :

(Permpoon Chanpu)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd



Certificate of Calibration


Certificate No. : 67-400548-2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber
 W = 0.56 m
 D = 0.40 m
 H = 0.48 m
 Capacity = 0.11 m³

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (± °C)
			1	2	3	4	5	6	7	8	9	
103.0	103.0	103.0	102.3	102.9	103.7	103.7	102.3	102.3	103.8	102.2	102.7	0.82
104.0	104.0	104.0	103.3	103.9	104.8	104.6	103.2	103.3	104.8	103.2	103.7	0.84
105.0	105.0	105.0	104.3	104.9	105.8	105.7	104.3	104.3	105.8	104.3	104.7	0.83
150.0	150.0	150.0	149.3	149.9	151.3	151.2	149.1	149.2	151.2	149.3	149.5	1.1
180.0	180.0	180.0	179.0	179.8	181.4	181.1	178.8	179.0	181.1	178.8	179.3	1.1

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured		Overall Variation (°C)
			Uniformity (°C)	Stability (°C)	
103.0	103.0	103.0	1.3	0.3	2.0
104.0	104.0	104.0	1.5	0.3	2.1
105.0	105.0	105.0	1.4	0.3	2.0
150.0	150.0	150.0	2.3	0.5	2.7
180.0	180.0	180.0	2.5	0.5	3.3

Remark: The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

-o0o-



CAL

Calibratech Co., Ltd.

7/106-7 Moo 2, Sukhprachasan 3 Rd., Banggood, Pakkred, Nonthaburi 11120
Tel: (02) 964-6211 Fax: (02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com



NSC-TSI-7817025
CALIBRATION 0030

Certificate of Calibration

Page : 1 of 2

Certificate No. : 67-400548-3

Submitted by :

M Green Group Co., Ltd.

188/46 Wistatukhakhon 25, Pracha-Utd Rd., Thungru Bangkok 10140 Thailand

Equipment :

Water Bath

Manufacturer : Memmert

Model : WNB29

Range : N/A °C Resolution : 0.1 °C

Serial No. : L619.0037

ID No. : N/A

Environment : On site calibration was carried out at the Laboratory, M Green Group Co., Ltd.

Ambient Temperature : (26.3 to 27.4) °C

Relative Humidity : (39 to 47) %

Line Voltage : (226.0 to 230.0) V

Date of Received : 20 September 2024

Date of Calibration : 20 September 2024

Date of Issue : 21 September 2024

Calibrated by : Permpon Charpu

Calibration Method : This instrument was calibrated by In-house method CAL-M4006 based on ASTM E715-80
The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units
Standard Digital Thermometer with RTD probe

ID No. Cert. No. Due Date Traceability

400046 & 400024 67-400198-2 30 Sep 2024

National Institute of Metrology Thailand (NIMT)

Approved by :

(Permpon Charpu)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.

CAL-00031-03



CAL

Calibratech Co., Ltd.

7/106-7 Moo 2, Sukhprachasan 3 Rd., Banggood, Pakkred, Nonthaburi 11120
Tel: (02) 964-6211 Fax: (02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

Certificate of Calibration


Page : 2 of 2

Certificate No. : 67-400548-3

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement



Front

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.					Uncertainty (± °C)	Measured Uniformity (°C)	Measured Stability (°C)
			1	2	3	4	5			
85.0	85.0	85.0	84.55	84.44	84.49	84.46	84.62	0.20	0.25	0.06

Remark The uncertainty is not combine uniformity of the water bath

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

-o0o-



CAL-00031-03



Certificate of Calibration

Page : 1 of 2

Certificate No. : 67-400548-4

Submitted by : M Green Group Co., Ltd.

188/46 Wisetukhakhon 25, Pracha-Uttid Rd., Thungkru Bangkok 10140 Thailand

Equipment : Temperature controlled enclosure (Incubator)

Manufacturer : Biobase

Model : Biochemistry Incubator

Range : 0 °C to 65 °C

Resolution : 0.1 °C

Serial No. : KYP1502202003

ID No. : N/A

Environment : On site calibration was carried out at the Laboratory, M Green Group Co., Ltd.

Ambient Temperature : (31.0 to 32.0) °C

Relative Humidity : (50 to 55) %

Line Voltage : (226.0 to 230.0) V

Date of Received : 20 September 2024

Date of Calibration : 20 September 2024

Date of Issue : 21 September 2024

Calibrated by : Permpon Chappu

Calibration Method : CAL-M4004, TLAS G-20

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Digital Thermometer with RTD Probe

ID No.

Cert. No.

Due Date

Traceability

400029 & 400048

67-400444-1

02 Feb 2025

National Institute of Metrology Thailand (NIMT)

Approved by :

(Permpon Chappu)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.



Certificate of Calibration


Certificate No. : 67-400548-4

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)

Inside of Chamber
W = 0.45 m
D = 0.41 m
H = 0.85 m
Capacity = 0.16 m³

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (± °C)
			1	2	3	4	5	6	7	8	9	
20.0	20.4	20.4	20.23	20.18	20.08	20.15	19.90	19.92	19.96	19.90	19.96	0.82
Test Point (°C)			Indicating Temperature (°C)			Measured Uniformity (°C)			Measured Stability (°C)			Overall Variation (°C)
20.0			20.4			0.38			0.43			0.96

Remark: The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

-o00-

P





Certificate of Calibration

Certificate No. : 67-400560-1

Page : 1 of 2

Submitted by :

M Green Group Co., Ltd.

188/46 Wisatesuknakhon 25, Pracha-Uthit Rd., Thungkru, Bangkok 10140 Thailand

Equipment :

Liquid in Glass Thermometer

Manufacturer :

N/A

Range :

0 °C to 100 °C

Serial No. :

N/A

ID No. :

94-49747

Environment :

Ambient Temperature : (23 ± 2) °C

Relative Humidity : (50 ± 15) %

Line Voltage : (220 ± 22) VAC

Date of Received :

20 September 2024

Date of Calibration :

23 September 2024

Date of Issue :

23 September 2024

Calibrated by :

Chorrip Sanchusri

Calibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4001 based on ASTM E77-07 by compared with PRT in the liquid bath at the constant controlled temperature.

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Platinum Resistance Thermometer (PRT)

ID No. Cert.No. Due Date

400001 TT-0023-24 16 Feb 2026

Traceability

National Institute of Metrology Thailand (NIMT)

2. Standard Digital Thermometer

ID No. Cert.No. Due Date

400003 23E1866 01 Jun 2025

Traceability

National Institute of Metrology Thailand (NIMT)

400004 23E1866 01 Jun 2025

National Institute of Metrology Thailand (NIMT)

Approved by :

(Permpoon Chantpu)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 67-400560-1

Page : 2 of 2

Result of Calibration :

Without Adjustment

UUC Condition As-Received : Good

Function :

Temperature measurement

Ice point check : UUC® reading 0 °C Standard reading -0.0321 °C

Standard Reading (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
39.8223	40	-0.2	0.31

Remark


UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

- o/c -





Certificate of Calibration

Certificate No. : 67-300569-1 Page : 1 of 2

Submitted by : M Green Group Co., Ltd.

188/46 Wisetasukrakhon 25, Pracha-Utd Rd., Thungkru, Bangkok 10140 Thailand

Equipment : Burette

Manufacturer : GLASSCO Class : A

Capacity : 10 ml Graduation : 0.05 ml

ID No. : 2212-0344-1

Environment : Ambient Temperature : (20 ± 3) °C

Relative Humidity : (50 ± 10) %

Air Pressure : 1008.0 mbar.

Date of Received : 20 September 2024

Date of Calibration : 23 September 2024

Date of Issue : 23 September 2024

Calibrated by : Areearat Sombun

Calibration Method : In-house method CAL-M3001 based on ASTM E 542-22

Reference Standard Instruments : This certification is traceable to the International System of Units

Electronic Balance

ID No.	Cert. No.	Due Date	Traceability
241005	67-200210-4	02 Dec 2024	National Institute of Metrology (Thailand) (NIMT)

Approved by : 
(Wipa Tovadee)
Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.



Certificate of Calibration

Certificate No. : 67-300569-1 Page : 2 of 2

Result of Calibration : This result of true Volume is referred to standard temperature at 20 °C

UUC Condition As-Received : Good

Delivery Time : 19.92 sec.

Nominal Volume (ml)	Measuring Volume (ml)
2	2.0044
5	5.0024
10	9.9915

Uncertainty of measurement with in $\pm$ 0.0039 ml

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2.00$,
providing a level of confidence of approximately 95%

-o0o-



CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Subhacharasri 3 Rd., Banggood, Pakkred, Nonthaburi 11120
Tel:(02) 964-6211 Fax:(02) 964-5155, e-mail : calibratech.co@yahoo.com, calibratech.co@gmail.com



REC-TB-1781702's
CALIBRATION 0026

Certificate of Calibration

Certificate No. : 67-300569-2

Submitted by : M Green Group Co., Ltd.

188/46 Wisatsukmakhon 25, Pracha-Utd Rd., Thungkru, Bangkok 10140 Thailand

Page : 1 of 2

Equipment

: Burette

Manufacturer : GLASSCO

Class : A

Capacity : 25 ml

Graduation : 0.1 ml

ID No. : 2212-0344-2

Environment

: Ambient Temperature : (20 ± 3) °C

: Relative Humidity : (50 ± 10) %

: Air Pressure : 1008.0 mbar.

Date of Received : 20 September 2024

Date of Calibration : 23 September 2024

Date of Issue : 23 September 2024

Calibrated by : Areearat Sonbun

Calibration Method : In-house method CAL-M3001 based on ASTM E 542-22

Reference Standard Instruments : This certification is traceable to the International System of Units

Electronic Balance

ID No. Cert. No.

241005 67-200210-4

Due Date

02 Dec 2024

Traceability

National Institute of Metrology (Thailand) (NIMT)

Approved by :

(Wipha Tovadee)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd

CAL-P0051-03



CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Subhacharasri 3 Rd., Banggood, Pakkred, Nonthaburi 11120
Tel:(02) 964-6211 Fax:(02) 964-5155, e-mail : calibratech.co@yahoo.com, calibratech.co@gmail.com

Certificate of Calibration

Certificate No. : 67-300569-2

Result of Calibration : This result of true Volume is referred to standard temperature at 20 °C

UUC Condition As-Received : Good

Delivery Time : 46.13 sec.

Nominal Volume (ml)	Measuring Volume (ml)
5	5.0039
15	14.9824
25	24.9757

Uncertainty of measurement with in ± 0.0066 ml

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2.00 , providing a level of confidence of approximately 95%

-o0o-

Wipha Tovadee



CAL-P0051-03



Certificate of Calibration

Certificate No. : 67-300569-3

Page : 1 of 2

Submitted by : M Green Group Co., Ltd.

188/46 Wisatsukunkhon 25, Prachin-Utd Rd., Thungkru, Bangkok 10140 Thailand

Equipment : Measuring Pipette

Manufacturer : GLASSCO

Class : A

Capacity : 10 ml

Graduation : 0.1 ml

ID No. : MP1001/19

Environment : Ambient Temperature : (20 ± 3) °C

Relative Humidity : (50 ± 10) %

Air Pressure : 1000.8 mbar.

Date of Received : 20 September 2024

Date of Calibration : 23 September 2024

Date of Issue : 23 September 2024

Calibrated by : Areearat Sombun

Calibration Method : In-house method CAL-M3001 based on ASTM E 542-22

Reference Standard Instruments : This certification is traceable to the International System of Units

Electronic Balance

ID.No. : 241005

Cert.No. : 67-200210-4

Date/Date : 02 Dec 2024

Traceability : National Institute of Metrology (Thailand) (NIMT)

Approved by :

(Wipa Tovades)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.



Certificate of Calibration

Certificate No. : 67-300569-3

Page : 2 of 2

Result of Calibration : This result of true Volume is referred to standard temperature at 20 °C

UIUC Condition As-Received : Good

Delivery Time : 12.20 sec.

Nominal Volume (ml)	Measuring Volume (ml)
2	1.9998
5	4.9871
10	9.9742

Uncertainty of measurement with in ± 0.0039 ml

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2.00 , providing a level of confidence of approximately 95%

-o0o-



CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhprachasan 3 Rd., Banggood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax (02) 964-5155, e-mail : calibratech.co@yahoo.com, calibratech.co@gmail.com



NSG-TSI-TIS17025
CALIBRATION 0050

Certificate of Calibration

Certificate No. : 67-300569-4

Page : 1 of 2

Submitted by : M Green Group Co., Ltd.

188/46 Wisatesuknakhon 25, Pracha-Utd Rd., Thungkru, Bangkok 10140 Thailand

Equipment : Measuring Pipette
Manufacturer : GLASSCO Class : A
Capacity : 25 ml Graduation : 0.1 ml
ID No. : MP25/01/19

Environment : Ambient Temperature : $(20 \pm 3) ^\circ\text{C}$
Relative Humidity : $(50 \pm 10) \%$
Air Pressure : 1000.8 mbar.

Date of Received : 20 September 2024
Date of Calibration : 23 September 2024
Date of Issue : 23 September 2024
Calibrated by : Arcarat Sombun

Calibration Method : In-house method CAL-M3001 based on ASTM E 542-22

Reference Standard Instruments : This certification is traceable to the International System of Units

Electronic Balance

ID No. : 241005 Cert.No. : 67-210210-4 Due Date : 02 Dec 2024
Traceability : National Institute of Metrology (Thailand) (NIMT)

Approved by :

(Signature)
(Wipa Tavee)
Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



CAL 67031-03

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhprachasan 3 Rd., Banggood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax (02) 964-5155, e-mail : calibratech.co@yahoo.com, calibratech.co@gmail.com

Certificate of Calibration

Certificate No. : 67-300569-4

Page : 2 of 2

Result of Calibration : This result of true Volume is referred to standard temperature at $20 ^\circ\text{C}$

UUC Condition As-Received : Good

Delivery Time : 15.20 sec.

Nominal Volume (ml)	Measuring Volume (ml)
5	5.0302
15	15.0832
25	25.0398

Uncertainty of measurement with in ± 0.0067 ml

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2.00$, providing a level of confidence of approximately 95%

-o0o-

(Signature)



CAL 67031-03



Certificate of Calibration

Certificate No. : 67-300571-1

Page : 1 of 2

Submitted by

: M Green Group Co., Ltd.

188/46 Wisetutkarnkhon 25, Pracha-Utd Rd., Thungkru, Bangkok 10140 Thailand

Equipment

: Cylinder

Manufacturer : GLASSCO

Class : A

Capacity : 100 ml

Graduation : 1 ml

ID No. : CY100/01/24

Environment

: Ambient Temperature : (20 ± 3) °C

: Relative Humidity : (50 ± 10) %

: Air Pressure : 1003.9 mbar.

Date of Received : 20 September 2024

Date of Calibration : 24 September 2024

Date of Issue : 24 September 2024

Calibrated by : Accurat Sombun

Calibration Method : In-house method CAL-M3001 based on ASTM E 542-22

Reference Standard Instruments : This certification is traceable to the International System of Units

Electronic Balance

ID No.

Cert. No.

Due Date

Traceability

241002

67-200210-1

02 Dec 2024

National Institute of Metrology (Thailand) (NIMT)

Approved by :

(Signature)
(Wipa Towadee)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.



Certificate of Calibration

Certificate No. : 67-300571-1

Page : 2 of 2

Result of Calibration :

This result of true Volume is referred to standard temperature at 20 °C

UUC Condition As-Received : Good

Nominal Volume (ml)	Measuring Volume (ml)
50	50.45
100	100.52

Uncertainty of measurement with in ± 0.063 ml

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2.00 , providing a level of confidence of approximately 95%

-o0o-

(Signature)



CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Subhaphrachasan 3 Rd., Bangood, Pathum, Nonthaburi 11120
Tel:(02) 964-6211 Fax (02) 964-5155, e-mail : calibratech.co@yahoo.com, calibratech.co@hotmail.com



NSG-TIS-TIS17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 67-300571-2

Submitted by

: M Green Group Co., Ltd.

188/46 Wisatsuknakhon 25, Pracha-Utd Rd., Thungkru, Bangkok 10140 Thailand

Page : 1 of 2

Equipment

: Cylinder

Manufacturer : GLASSCO

Class : A

Capacity : 250 ml

Graduation : 2 ml

ID No. : CY25001/19

Environment

: Ambient Temperature :

(20 ± 3) °C

Relative Humidity :

(50 ± 10) %

Air Pressure :

1004.1 mbar.

Date of Received : 20 September 2024

Date of Calibration : 24 September 2024

Date of Issue : 24 September 2024

Calibrated by : Arccrat Sombun

Calibration Method : In-house method CAL-M3001 based on ASTM E 542-22

Reference Standard Instruments : This certification is traceable to the International System of Units

Electronic Balance

ID No.

Cert. No.

Due Date

Traceability

241002

67-300210-1

02 Dec 2024

National Institute of Metrology (Thailand) (NIMT)

Approved by :

(Wipa Towadee)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.

Cal-003-03



CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Subhaphrachasan 3 Rd., Bangood, Pathum, Nonthaburi 11120
Tel:(02) 964-6211 Fax (02) 964-5155, e-mail : calibratech.co@yahoo.com, calibratech.co@hotmail.com

Certificate of Calibration

Certificate No. : 67-300571-2

Result of Calibration : This result of true Volume is referred to standard temperature at 20 °C

UUC Condition As-Received : Good

Page : 2 of 2

Nominal Volume (ml)	Measuring Volume (ml)
150	151.02
250	251.16

Uncertainty of measurement with in ± 0.087 ml

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2.00 , providing a level of confidence of approximately 95%

-o-o-

Wipa Towadee



Cal-003-03



Certificate of Calibration

Certificate No. : 67-3005711-3

Page : 1 of 2

Submitted by : M Green Group Co., Ltd.

188/46 Wisatsukrakchon 25, Prachin-Ud Rd., Thungkru, Bangkok 10140 Thailand

Equipment

: Cylinder

Manufacturer : GLASSCO

Class : A

Capacity : 1000 ml

Graduation : 10 ml

ID No. : CY1000/01/24

Environment

: Ambient Temperature : (20 ± 3) °C

: Relative Humidity : (50 ± 10) %

: Air Pressure : 1001.7 mbar.

Date of Received : 20 September 2024

Date of Calibration : 24 September 2024

Date of Issue : 24 September 2024

Calibrated by : Areerat Sombun

Calibration Method : In-house method CAL-M3001 based on ASTM E 542-22

Reference Standard Instruments : This certification is traceable to the International System of Units

Electronic Balance

ID No. : 241002
Cert.No. : 67-200210-1
Due Date : 02 Dec 2024

Traceability

National Institute of Metrology (Thailand) (NIMT)

Approved by :

(Wipa Tovadee)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%.

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.



Certificate of Calibration

Certificate No. : 67-3005711-3

Page : 2 of 2

Result of Calibration : This result of true Volume is referred to standard temperature at 20 °C

UUC Condition As-Received : Good

Nominal Volume (ml)	Measuring Volume (ml)
500	500.74
1000	1003.10

Uncertainty of measurement with in ± 0.17 ml

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2.00 , providing a level of confidence of approximately 95%

-o0o-

mij



Certificate of Calibration

Certificate No. : 67-300572-1 Page : 1 of 2

Submitted by : M Green Group Co., Ltd.
188/46 Wisatsubhakkhon 25, Pracha-Utd Rd., Thungkru, Bangkok 10140 Thailand

Equipment : Imhoff Cone
Manufacturer : VITLAB
Capacity : 1000 ml Graduation : 0.1, 50 ml
ID No. : CY1000/01/22

Environment : Ambient Temperature : $(20 \pm 3) ^\circ\text{C}$
Relative Humidity : $(50 \pm 10) \%$
Air Pressure : 1001.5 mbar.

Date of Received : 20 September 2024
Date of Calibration : 23 September 2024
Date of Issue : 23 September 2024
Calibrated by : Areerat Sombun

Calibration Method : In-house method CAL-M3001 based on ASTM E 542-01

Reference Standard Instruments : This certification is traceable to the International System of Units

Electronic Balance
ID No. 2411002 Cert.No. 67-200210-1 Due Date 02 Dec 2024 Traceability
National Institute of Metrology (Thailand) (NIMT)

Approved by : 
(Wipha T. Phadree)
Supervisor

The Uncertainties are for a confidence probability of approximately 95%
This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.



Certificate of Calibration

Certificate No. : 67-300572-1 Page : 2 of 2

Result of Calibration : This result of true Volume is referred to standard temperature at $20 ^\circ\text{C}$

UUC Condition As-Received : Good

Nominal Volume (ml)	Measuring Volume (ml)
500	500.74
1000	1009.16

Uncertainty of measurement with in ± 0.17 ml

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2.00$, providing a level of confidence of approximately 95%

- o O o -





CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhprachassin 3 Rd., Banggood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com



NSC-TSI-TS17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 67-300570-1

Page : 1 of 2

Submitted by : M Green Group Co., Ltd.

188/46 Wisetrunakulchon 25, Pracha-Uth Rd., Thungkru, Bangkok 10140 Thailand

Equipment : Volumetric Flask

Manufacturer : GLASSCO Class : A

Capacity : 100 ml

ID No. : VF100/01/19

Environment : Ambient Temperature : (20 ± 3) °C

Relative Humidity : (50 ± 10) %

Air Pressure : 1003.9 mbar.

Date of Received : 20 September 2024

Date of Calibration : 24 September 2024

Date of Issue : 24 September 2024

Calibrated by : Areearat Sombun

Calibration Method : In-house method CAL-M3001 based on ASTM E 542-22

Reference Standard Instruments : This certification is traceable to the International System of Units

Electronic Balance

ID No. 241005


Cert. No. 67-200210-4

Due Date 02 Dec 2024

Traceability

National Institute of Metrology (Thailand) (NIMT)

Approved by :


(Wipa Tovadee)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



CAL-00001-03

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhprachassin 3 Rd., Banggood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com

Certificate of Calibration

Certificate No. : 67-300570-1

Page : 2 of 2

Result of Calibration : This result of true Volume is referred to standard temperature at 20 °C

UNC Condition As-Received : Good

Nominal Volume (ml)	Measuring Volume (ml)
100	100.063

Uncertainty of measurement with in ± 0.018 ml

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2.00$, providing a level of confidence of approximately 95%

-o-o-





CAL-00001-03

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Subhacharasri 3 Rd., Bangpood, Pakred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.co@yahoo.com, calibratech.co@hotmail.com



NSC-TIS-71517075
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 67-300570-2

Page : 1 of 2

Submitted by : M Green Group Co., Ltd.

188/46 Wisatsukhakhon 25, Pracha-Ud Rd., Thungkrn, Bangkok 10140 Thailand

Equipment : Volumetric Flask

Manufacturer : GLASSCO

Class : A

Capacity : 250 ml

ID No. : VF25001/19

Environment

Ambient Temperature : (20 ± 3) °C

Relative Humidity : (50 ± 10) %

Air Pressure : 1003.8 mbar.

Date of Received : 20 September 2024

Date of Calibration : 24 September 2024

Date of Issue : 24 September 2024

Calibrated by : Areearat Sombun

Calibration Method : In-house method CAL-M3001 based on ASTM E 542-22

Reference Standard Instruments : This certification is traceable to the International System of Units

Electronic Balance

ID No. : 241002

Cart. No. : 67-200210-1

Due Date : 02 Dec 2024

Traceability : National Institute of Metrology (Thailand) (NIMT)

Approved by :

(Wipa Tovaatec)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.



CAL-190011-01

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Subhacharasri 3 Rd., Bangpood, Pakred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.co@yahoo.com, calibratech.co@hotmail.com

Certificate of Calibration

Certificate No. : 67-300570-2

Page : 2 of 2

Result of Calibration : This result of true Volume is referred to standard temperature at 20 °C

UUC Condition As-Received : Good

Nominal Volume (ml)	Measuring Volume (ml)
250	250.04

Uncertainty of measurement with in ± 0.051 ml

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2.00 , providing a level of confidence of approximately 95%

-oOo-

(Signature)



CAL-190011-01



Certificate of Calibration

Certificate No. : 67-300570-3

Page : 1 of 2

Submitted by

: M Green Group Co., Ltd.

188/46 Wisetnukrachon 25, Pracha-Uthit Rd., Thungkru, Bangkok 10140 Thailand

Equipment

: Volumetric Flask

Manufacturer : GLASSCO

Class : A

Capacity : 1000 ml

ID No. : VF1000/01/19

Environment

: Ambient Temperature : (20 ± 3) °C

Relative Humidity : (50 ± 10) %

Air Pressure : 1003.3 mbar.

Date of Received : 20 September 2024

Date of Calibration : 24 September 2024

Date of Issue : 24 September 2024

Calibrated by : Arcarat Sornbun

Calibration Method : In-house method CAL-M3001 based on ASTM E 542-22

Reference Standard Instruments : This certification is traceable to the International System of Units

Electronic Balance

ID No.

Cert. No.

Due Date

Traceability

241002

67-210210-1

02 Dec 2024

National Institute of Metrology (Thailand) (NIMT)

Approved by :

(Wipa Towadee)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.



Certificate of Calibration

Certificate No. : 67-300570-3

Page : 2 of 2

Result of Calibration :

This result of true Volume is referred to standard temperature at 20 °C

UUC Condition As-Received : Good

Nominal Volume (ml)	Measuring Volume (ml)
1000	1000.25

Uncertainty of measurement with in ± 0.14 ml

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2.00 , providing a level of confidence of approximately 95%

-o0o-





Certificate of Calibration

Certificate No. : 67-210362-1 Page : 1 of 2

Submitted by :

M Green Group Co., Ltd.

188/46 Wisatsukrakhon 25, Pracha-Utd Rd., Thungkru, Bangkok 10140 Thailand

Equipment :

Weight

Manufacturer : N/A

Material : Stainless Steel

Weight size : 1 g

ID No. : 63-210391-1

Assumed density of weight : 7950 kg / m³Assured Air density : 1.2 kg / m³

Environment :

Ambient Temperature :

(20 ± 2) °C

Relative Humidity :

(50 ± 10) %

Air Pressure :

1005.7

mbar

Date of Received :

20 September 2024

Date of Calibration :

23 September 2024

Date of Issue :

23 September 2024

Calibrated by :

Wuttichai Swarphong

Calibration Method : In-house method CAL-M2101 based on OIML R 111-1 : 2004(E)

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Weights

Cert. No.

Due Date

Traceability

MM4-0042-22

21 Mar 2025

National Institute of Metrology (Thailand), (NIMT)

Approved by :

(Sufja Sangkhum)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.



Certificate of Calibration

Certificate No. : 67-210362-1 Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

No.	Nominal Value	Id.Mark	Conventional mass Value	Measuring Uncertainty
1	1 g	none	1 g +0.003 mg	± 0.023 mg

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

-o0o-





Certificate of Calibration

Certificate No. : 67-210362-2 Page : 1 of 2

Submitted by : M Green Group Co., Ltd.
188/46 Wisatsuknakhon 25, Pracha-Utd Rd., Thungkru, Bangkok 10140 Thailand

Equipment : Weight
Manufacturer : N/A Material : Stainless Steel
Weight size : 100 g

ID No. : 63-210391-2

Assumed density of weight : 7950 kg / m³Assumed Air density : 1.2 kg / m³

Environment : Ambient Temperature : (20 ± 2) °C

Relative Humidity : (50 ± 10) %

Air Pressure : 1005.7 mbar

Date of Received : 20 September 2024

Date of Calibration : 23 September 2024

Date of Issue : 23 September 2024

Calibrated by : Wuttichai Swatphong

Calibration Method : In-house method CAL-M2101 based on OIML R 111-1 : 2004(E)

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Weights

ID No.	Cert. No.	Due Date	Traceability
E221-E2210	MM-0042-22	21 Mar 2025	National Institute of Metrology (Thailand), (NIMT)

Approved by :
(Saja Sangkhunt)
Supervisor

The Uncertainties are for a confidence probability of approximately 95%.

This certificate may not be reproduced after than in full except with the prior written approval of the Calibratech Co., Ltd.



Certificate of Calibration

Certificate No. : 67-210362-2 Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

No.	Nominal Value	Id. Mark	Conventional mass Value	Measuring Uncertainty
1	100 g	none	100 g : +0.06 mg	± 0.11 mg

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

-o0o-





MSC-TSI-TS17025
CALIBRATION 0050

Certificate of Calibration

Certificate No. : 67-210362-3 Page : 1 of 2

Submitted by : M Green Group Co., Ltd.

188/46 Wisatesukrakhoon 25, Pracha-Uth Rd., Thungkru, Bangkok 10140 Thailand

Equipment :

Weight
Manufacturer : N/A Material : Stainless Steel
Weight size : 200 g

ID No. : 63-210391-3

Assumed density of weight : 7950 kg / m³

Assumed Air density : 1.2 kg / m³

Ambient Temperature : (20 ± 2) ° C

Relative Humidity : (50 ± 10) %

Air Pressure : 1005.7 mbar

Date of Received : 20 September 2024

Date of Calibration : 23 September 2024

Date of Issue : 23 September 2024

Calibrated by : Wutichai Swatphong

Calibration Method : In-house method CAL-M2101 based on OIML R 111-1 : 2004(E)

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Weights

ID No.	Cert. No.	Due Date	Traceability
E221-E2210	MM-0042-22	21 Mar 2025	National Institute of Metrology (Thailand), (NIMT)

Approved by : 
(Sattia Sangkhum)
Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co., Ltd.



Certificate of Calibration

Certificate No. : 67-210362-3 Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

No.	Nominal Value	Id. Mark	Conventional mass Value	Measuring Uncertainty
1	200 g	none	200 g +0.06 mg	± 0.17 mg

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

-oOb-







บริษัท ไทยยูนิค จำกัด THAI UNIQUE CO., LTD.

80-82 ถนนประชาวิทย์ แขวงบางขุนพรหม เขตพระนคร กรุงเทพฯ 10200
80-82 Prachathipatani Rd., Bangkokkhunprom, Pranakorn, Bangkok 10200
Tel. 0-2629-0191-6, 0-2280-1787, Fax. 0-2280-1788, E-mail : thawan@thaiunique.com, Website : www.thaiunique.com

ATOMIC ABSORPTION SPECTROMETER TEST CERTIFICATE

Certificate No : SV2410/22223
Instrument Type : Atomic Absorption Spectrophotometers
Model : AA240FS
Serial Number : EL08043418
Organization : Test Tech Co., Ltd.
Address : 30.32, Soi 63 Rama II Rd. Samedam Bangkokkhuntem, Bangkok 10150
Date : 17 Oct. 2024

Hollow cathode lamps used

Element	Lamp number	Comments
Arsenic	56-101003-00	
Copper	56-101014-00	
Potassium	56-101042-00	
Iron	56-101027-00	
Manganese	56-101337-00	

Test description	Specification	Result	Comments
Light throughput (%Gain) or (EHT)			
Cu at 324.8 nm	≤ 64 % or 380 V	32 %	Pass
As at 193.7 nm	≤ 80 % or 540 V	54 %	Pass
K at 766.5 nm*	≤ 84 % or 540 V	64 %	Pass
Fe at 248.3 nm	≤ 80 % or 540 V	63 %	Pass
Mn at 279.5 nm	≤ 64 % or 380 V	53 %	Pass
Photometric noise Cu BGC off			
STDV @ 0 Abs	≤ 0.0005	0.0001	Pass



บริษัท ไทยยูนิค จำกัด THAI UNIQUE CO., LTD.

80-82 ถนนประชาวิทย์ แขวงบางขุนพรหม เขตพระนคร กรุงเทพฯ 10200
80-82 Prachathipatani Rd., Bangkokkhunprom, Pranakorn, Bangkok 10200
Tel. 0-2629-0191-6, 0-2280-1787, Fax. 0-2280-1788, E-mail : thawan@thaiunique.com, Website : www.thaiunique.com

Wavelength accuracy

Cu at 324.8 nm	323.0 nm - 326.0 nm	324.7 nm	Pass
As 193.7 nm	192.0 nm - 195.0 nm	193.7 nm	Pass
K at 766.5 nm*	765.0 nm - 768.0 nm	766.4 nm	Pass
Fe at 248.3 nm	246.8 nm - 249.8 nm	248.2 nm	Pass
Mn at 279.5 nm	278.0 nm - 281.0 nm	279.5 nm	Pass

High solids nebulizer setting**

Uptake rate	7.2 - 10.6 ml / min	7.8 ml/min	Pass
Max Abs	≥ 0.75 Abs	0.79 Abs	Pass
Precision(%RSD)	≤ 0.5 %	0.4 %	Pass

Zeeman Background Correction Accuracy (%)**

BCA @ Au 242.8 nm	< 3.7 %	***	***
-------------------	---------	-----	-----

Zeeman Magnetic Sensitivity Ratio (%)***

MSR @ Cu 324.7 nm	> 70 %	***	***
-------------------	--------	-----	-----

Characteristic mass and sensitivity ****

Sensitivity	≥ 0.21 Abs	****	****
Precision (%RSD)	≤ 4.0 %	****	****

* for Wideband PMT (Wavelength 190nm - 900nm)

** for Flame system

*** for Zeeman system

**** for Graphite furnace system



CALIBRATED BY : 

Signature : 

Engineer : Suriya Nacharoen

Date : 17 / Oct / 2024

APPROVED BY :

Signature : 

Service Manager : Suchai Sangrakiatichai

Date : 17 / Oct / 2024



PinAAcle 900F Preventive Maintenance Report

Company Name: TEST TECH CO., LTD.
Instrument Location: 30,32 RAMA2 SOI 63, RAMA 2 RD.,
SAMMAEDAM, BANGKHUNTEN, BANGKOK 10150
Instrument Serial No.: PFBS21091601
Date: 22-Feb-2024

PinAAcle 900F Preventive Maintenance (PM)

Company Name:		TEST TECH CO., LTD.	
Address		30,32 RAMA2 SOI 63, RAMA 2 RD., BANGKHUNTEN BANGKOK 10150	
(Instrument Location):			
Serial Number:		PFBS21091601	PM Number: 1 of 1
Customer Name		Juralrat Jongprakobkitt	Telephone Number: 087-5199005
(If applicable):		Chainarong	Service Order Number: WO-02710284
Customer Support Engineer Name:		22-Feb-2024	Next PM Due Date: 22-Feb-2025
Date PM Performed: (DD-MMM-YY)		Standard Labor Hours to Complete PM: 5 hours	

Port Number	Release	Publication Date	
09370145 Rev.9	A	January 2018	

Scope
The purpose of this PM is to ensure the continued functionality of the PinAAcle 900F by inspecting and replacing any worn or damaged parts. This service should only be performed by a trained representative of PerkinElmer.
The customer should save their method before the PM begins.

General Instructions:
The customer must provide the engineer operational data to demonstrate recent instrument performance prior to starting the PM.
Always check with the customer before making any changes that may affect the customer's analysis or calibration, including a current back-up of system software and/or data files.
The completed document should be signed by an authorized PerkinElmer and customer representative and left with the customer.
Update the PM sticker and instrument logbook as required.

Copyright Information
This document contains proprietary information that is protected by copyright. All rights are reserved. No part of this publication may be reproduced in any form whatsoever or translated into any language without the prior, written permission of PerkinElmer, Inc.
Copyright © 2013 PerkinElmer, Inc.

Trademarks
Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are protected by law. PerkinElmer is a registered trademark of PerkinElmer, Inc. All other trademarks and registered trademarks not owned by PerkinElmer, Inc. or its subsidiaries that are depicted herein are the property of their respective owners. Except as specifically set forth in its terms and conditions of sale, PerkinElmer makes no Warranty of any kind with regard to this document, including, but not limited to, the implied warranties of merchantability and fitness for a particular purpose.
PerkinElmer shall not be liable for incidental or consequential damages in connection with the furnishing or use of this document.

Component List

Component / Specific Model	Serial #	Configuration Notes
FIAS-400	400521100101	Syngistix version 4.0

Parts Lists

Parts Included with the PM		
Part Number (If applicable)	Description	Quantity
80501696	Fan Filters	2
N3160156	O-Ring Kits for Sampling Introduction (Stainless Steels Nebulizer)	N/A
N3160157	O-Ring Kits for Sampling Introduction (Plastic Nebulizer)	1
N9301714	Replacement Acetylene Filter Cartridge	1
TH001022	Replacement Air Filter Cartridge	2

Additional Reagents and Standards Required for PM			
Part Number (If applicable)	Description	Quantity	Expiry Date (MM/YY)
N9300183	1000 mg/L Copper Standard	AR	APR-2025

Additional Reagents and Standards Required for PM (Customer Support Solution)			
Part Number (If applicable)	Description	Quantity	Expiration Date (MM/YY)
N/A	DI Water	250 mL	AR
N/A	0.5% HNO ₃	250 mL	AR

Additional Tools Required for PM			
Part Number (If applicable)	Description	Quantity	Serial #
N1013000	0.2A Neutral density filter	1	MG0-141
N1013002	1.0A Neutral density filter	1	MG2-045
03030997	System 2 EDL Driver	1	839936
N3050605	As System 2 EDL	1	06261
N3050121	Cu Lumina HCL	1	101615-01080
N3050109	Ba Lumina HCL	1	858AD8
N3050139	K Lumina HCL	1	011604-41713
N3050152	Ni Lumina HCL	1	050914-01060

Procedure Checklist

Use (✓) to check off those steps in the checklist that have been completed.

1. General:

- ☒ Review the instrument performance with the customer and document any recent problems.
- ☒ Inspect the customer log/book and make any appropriate PM entries.
- ☒ Perform general inspection of system for cleanliness.

2. PC Instrument Software:

- ☒ Instrument Software user files/databases archived, packed, and/or deleted as needed.

3. Mechanical:

- ☒ Inspect and clean all fans and filters. Replace filters if necessary.
- ☒ Inspect all gas lines for leaks and/or wear. Replace if needed.
- ☒ Clean exterior of the instrument.
- ☒ Inspect the burner head, burner chamber, and nebulizer. Clean if needed as stated in the Hardware Guide.
- ☒ Check burner head dimensions with the feeler gauge as stated in the Hardware Guide in the Maintenance chapter section on cleaning the burner head and checking sloth width. Replace if out of specification.
- ☒ Check the condition of the end cap, burner head, and nebulizer O-rings. Replace if necessary.
- ☒ Check the drain system for signs of wear. Replace worn or damaged parts.
- ☒ Visually check for proper flame conditions when lighting the Air-C2H2 and N2O-C2H2 flames (if applicable).

4. Electrical:

- ☒ Inspect PC boards. Clean if necessary.
- ☒ Carefully check all internal and external cable connections.
- ☒ Check instrument firmware revisions upgrade to current levels (if necessary)
- ☒ Run Diagnostics Test within the Advanced function of the Spectrometer page. Check the results in the service log folder in the Spectrometer BM Log Viewer.

5. Optics:

- ☒ Inspect and clean the sample compartment windows, if needed.
- ☒ Inspect optics. Clean or replace if necessary.

6. Gasses:

- ☒ Verify that the Gasses supplied to the instrument are within the pressure and purity specifications found in the PinAAcle 900F Series Pre-Installation Checklist SDB.
- ☒ Verify that the acetylene filter and air filter element is dry. Replace if necessary.

7. Flame Interlock Check:

Description: Check to ensure that all safety interlocks are closed.

Parameter	Specification	Test Results	Pass/Fail
Flame Sensor	Air/C ₂ H ₂ Flame correctly shuts down	Active	Passed
Drain Sensor	Air/C ₂ H ₂ Flame correctly shuts down	Active	Passed
Nebulizer Sensor	Air/C ₂ H ₂ Flame correctly shuts down	Active	Passed
C ₂ H ₂ Pressure Sensor	Air/C ₂ H ₂ Flame correctly shuts down	Active	Passed
Air Pressure Sensor	Air/C ₂ H ₂ Flame correctly shuts down	Active	Passed
Burner Head Sensor	Choosing Nitrous Oxide as the oxidant should trigger an interlock shuts down	Active	Passed

8. After PM Performance tests:

8.1 Detector Linearity with Barium

Description: Ensures that the detector is linear in the Visible Range.

Parameter	Specification	Certificate Value at 553.8 nm (Abs.)	Test Results	Pass/Fail
1.0 A ND Filter	± 5% from Cert.	1.0547	1.0432	Passed
0.2 A ND Filter	± 5% from Cert.	0.1903	0.1832	Passed

8.2 Baseline Noise at 1.0 Absorbance with Barium

Description: Ensures that a high absorbance will not produce excessive noise.

Parameter	Specification	Results	Pass/Fail
Standard Deviation	≤ 0.010	0.0038	Passed

8.3 AA Baseline Noise with Copper

Description: Check baseline noise.

Parameter	Specification	Results	Pass/Fail
Standard Deviation	≤ 0.001	0.0001	Passed

8.4 D₂ Background Compensation with Copper

Description: Verifies the instruments ability to compensate for Background absorption.

Parameter	Specification	Results	Pass/Fail
Standard Deviation	≤ 0.020	0.0090	Passed

8.5 AA-BG Baseline Noise with Copper

Description: Ensures that background correction does not produce excessive noise.

Parameter	Specification	Results	Pass/Fail
Standard Deviation	≤ 0.005	0.0001	Passed

8.6 AA-BG Baseline Noise with Arsenic

Description: Ensures that background correction does not produce excessive noise at a low wavelength.

Parameter	Specification	Results	Pass/Fail
Standard Deviation	≤ 0.005	0.0002	Passed

8.7 Flame Sensitivity

Description: Instrument Sensitivity checked against Copper standard.

Standard Copper Sensitivity	Specification	Results (Abs.)	Pass/Fail
5 mg/L Sensitivity SS Neb (if applicable)	> 0.250 Abs.	N/A	Not Applicable
2 mg/L Sensitivity HS Neb (if applicable)	> 0.250 Abs.	0.3366	Passed

10. Review:

- ☒ Review with the customer PM work performed.
- ☒ Review with the customer routine maintenance procedures.
- ☒ Discuss recommended customer supplied materials to have on hand.
- ☒ Attach PM sticker.

Additional Comments

Additional Comments Regarding the PM	
PERFORMANCE TEST FOR FIAS	
1. Characteristics mass for Mercury. (500ul of 10ug/L Hg and 5 replicates) Characteristics mass %RSD	307.3 pg/0.0044 A 2.09 %
2. Characteristics mass for Arsenic. (500ul of 10ug/L As and 5 replicates) Characteristics mass %RSD	134.7 pg/0.0044 A 1.63 %
2. Characteristics mass for Selenium. (500ul of 10ug/L Se and 5 replicates) Characteristics mass %RSD	80.4 pg/0.0044 A 1.02 %
Wavelength Accuracy Check	
As 193.700 nm +/- 0.3nm	193.700 nm
Cu 324.750 nm +/- 0.3nm	324.700 nm
Ba 553.550 nm +/- 0.3nm	553.650 nm
K 766.490 nm +/- 0.3nm	766.490 nm

Review

The preventive maintenance checks and if applicable performance tests for PinAAcle 900F have been completed.

This PinAAcle 900F Passes ☒ Fails ☐ the preventive maintenance.

Review of Preventive Maintenance:

Authorized PerkinElmer Representative:	Date: 22-Feb-2024 (DD-MMM-YYYY)
Authorized Customer Representative:	Date: 22-Feb-2024 (DD-MMM-YYYY)



MAINTENANCE AND IPV TEST CERTIFICATE MODEL
OPTIMA 8000

Customer : บริษัท เอสที เทคโนโลยี	Date Tested: May 14, 2024
Address : 30,32 ซอยพหลโยธิน 2 แขวง 63 กรุงเทพมหานคร 2 แขวง	Recommendation Recertification Period 12 Months
เบอร์โทรศัพท์ 2 แขวง	Recertification Due: May 14, 2025
เบอร์โทรมือถือ กรุงเทพมหานคร 70150	Date Last Certified: May 18, 2023
User Name: คุณวิรัตน์ ศรีประเสริฐ	Visit Number: 1 of 1
Phone: 02-993-4211-17	PerkinElmer Phone: 02-719-6420 ext 206
Fax: lab_center@testtech.co.th	PerkinElmer Fax: 02-318-5597

CONFIGURATION TESTED		
MODEL	SERIAL NUMBER	SOFTWARE
OPTIMA 8000 (EQL-180)	078S1411171C	ICP WinLab32 version 5
TESTED EQUIPMENT	CALIBRATION NUMBER	EXPIRATION
IPV Methods		
TEST STANDARD USED	PART NUMBER	EXPIRATION DATE
Multielement Standard Instrument Cal. STD4	N069-1579 N930-0221	Dec 30, 2024 Nov 30, 2024
CUSTOMER SUPPLIED	COMMENTS	CUSTOMER INITIALS
2 % HNO3		
10 % HNO3		



MAINTENANCE AND IPV TEST CERTIFICATE MODEL
OPTIMA 8000

SERIAL NUMBER: 078S1411171C	DATE TESTED: May 14, 2024
1. MECHANICAL CHECKS	
A. Inspect and clean all fans and filters.	OK
B. Inspect and replace as necessary, all torch components including the RF coil.	OK
C. Inspect all tubing for sign of clacking or leaking.	OK
D. Adjust water and gas pressure regulator settings.	OK
E. Inspect and leak check pneumatics drawers.	OK
F. Clean the exterior of the instrument.	OK
2. OPTICAL CHECKS	
A. Inspect and clean all optical components.	OK
B. As required, check and replace all purge filters.	OK
C. Recheck optical alignment.	OK
3. COOLING SYSTEM CHECKS	
A. Perform preventive maintenance on chiller.	OK
B. Flush out the chiller every six months.	OK
4. PERFORMANCE CHECKS	
A. Torch View Alignment	OK
B. Wavelength Calibration.	OK



MAINTENANCE AND IPV TEST CERTIFICATE MODEL
OPTIMA 8000

SERIAL NUMBER: 078S1411171C		DATE TESTED: May 14, 2024	
PARAMETER		SPECIFICATION	
Spectral Resolution : UV			
As 193.696 nm	≤ 0.009 nm	nm	0.00735 nm
Ni 231.604 nm	≤ 0.011 nm	nm	0.00913 nm
Ni 341.476 nm	≤ 0.015 nm	nm	0.01386 nm
Spectral Resolution : VIS			
Ba 455.403 nm	≤ 0.020 nm	nm	0.01721 nm
Precision			
Zn 206.200 nm	% RSD ≤ 1.0 %		0.35 %
Mg 280.271 nm	% RSD ≤ 1.0 %		0.27 %
Mg 285.213 nm	% RSD ≤ 1.0 %		0.46 %
Ba 455.403 nm	% RSD ≤ 1.0 %		0.48 %
Detection Limits : Axial			
Tl 190.801 nm	3(sd) ≤ 10.0 ppb		1.00 ppb
As 193.696 nm	3(sd) ≤ 10.0 ppb		3.32 ppb
Se 196.026 nm	3(sd) ≤ 5.0 ppb		3.88 ppb
Pb 220.353 nm	3(sd) ≤ 3.0 ppb		1.45 ppb
Detection Limits : Radial			
As 193.696 nm	3(sd) ≤ 60.0 ppb		3.41 ppb
Zn 213.857 nm	3(sd) ≤ 2.0 ppb		0.30 ppb
Mn 257.610 nm	3(sd) ≤ 1.0 ppb		0.03 ppb
La 379.478 nm	3(sd) ≤ 3.0 ppb		0.27 ppb
Ba 455.403 nm	3(sd) ≤ 0.3 ppb		0.05 ppb
Ba 493.408 nm	3(sd) ≤ 0.6 ppb		0.06 ppb
BEC : Axial (IB X 1000)/(IS-IB)			
Mn 257.610 nm	≤ 30 ppb		10.70 ppb
BEC : Radial (IB X 1000)/(IS-IB)			
Mn 257.610 nm	≤ 30 ppb		21.54 ppb



MAINTENANCE AND IPV TEST CERTIFICATE MODEL
OPTIMA 8000

SERIAL NUMBER: 078S1411171C DATE TESTED: May 14, 2024

Remarks :
Commissioning follow as commissioning performance sheets.

☒ meets

☐ does not meet

This is to certify that the above tests have been performed and the configuration tested the PerkinElmer Specifications listed on this certificate.
This certificate does not modify PerkinElmer's standard terms and condition of sale, including warranty terms.

Service Department PerkinElmer Scientific (Thailand)Co.,Ltd.

Customer Service Engineer: (Narong Wajrakit)
Service Engineer



QUALITY CALIBRATION CO., LTD.
235 Petchkasem 63/2 Road, Lakson, Bangkok, Bangkok 10160
Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4584
www.qcalibration.com



CERTIFICATE No : 25T0970
REFERENCE No : 76012-2

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : INCUBATOR
MANUFACTURER : MEMMERT
MODEL : IF 160
SERIAL No : D518.0082
ID No : EQL-205
CONDITION AS RECEIVED : USED ITEM
SUBMITTED BY : TEST TECH CO., LTD.
30,32 RAMA II RD., SAMAEDAM,
BANGKHUNTHIAN, BANGKOK 10150

CALIBRATED BY : CHAICHARN CH.
CALIBRATION DATE : 04-Feb-25

APPROVED BY :
PONGSAK J.
ISSUED DATE : 12-Feb-25
RECEIVED DATE : 04-Feb-25

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.



F-G010 REV : 03



QUALITY CALIBRATION CO., LTD.
235 Petchkasem 63/2 Road, Lakson, Bangkok, Bangkok 10160
Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4584

CERTIFICATE No : 25T0970

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : INCUBATOR
MANUFACTURER : MEMMERT
MODEL : IF 160
ID No : EQL-205
RECEIVED DATE : 04-Feb-25
AMBIENT TEMPERATURE : 25 °C ± 1 °C
SN : D518.0082
CALIBRATION DATE : 04-Feb-25
RELATIVE HUMIDITY : 50 %RH ± 10 %RH

CONDITION OF THIS RESULTS OF CALIBRATION

1. THIS INSTRUMENT WAS CALIBRATED ACCORDING TO TLAS G-20 BY COMPARISON WITH CALIBRATED RTD Pt100 UNDER NO LOAD CONDITION. THE TEMPERATURE PROBES WERE PLACED ON NINE POINTS AND LOCATED ONE THERMOMETER PROBE IN EACH OF THE EIGHT CORNERS OF THE CHAMBER AND WAS AWAY FROM THE EACH WALL OF 5 cm TO 10 cm. AND PLACED THE NINTH THERMOMETER PROBE WITHIN 2.5 cm. OF THE GEOMETRIC CENTER OF THE CHAMBER. THE UNIFORMITY WAS MEASURED BETWEEN REFERENCE PROBE AND OTHER PROBES AT THE SAME TIME.


2. REFERENCE STANDARD INSTRUMENTS :-

- INSTRUMENT MODEL SERIAL No CERTIFICATE No DUE DATE
1) DATA LOGGER WITH RTD HYDRA 2635A 7301307 24T6467 26-Jun-25
3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.
4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.
5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-
- NATIONAL INSTITUTE OF METROLOGY (THAILAND) THROUGH QUALITY CALIBRATION CO., LTD.

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT

GENERAL INFORMATION

Overall Ambient Temperature around the Chamber (°C) variation : 2
Overall Line Voltage (V) variation : 11
Instrument Condition : Normal
Chamber Size (W*L*H): 56*40*72 cm



CHAMBER PERFORMANCE

Calibration Point (°C)	Controller Temperature (°C)	Average All Position (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation
35.0	35.0	34.98	0.03	0.08	0.12
36.0	36.0	35.99	0.02	0.16	0.16
41.5	41.6	41.51	0.03	0.19	0.23

TEMPERATURE MEASUREMENT ACCURACY TEST

Controller Temp (°C)	Indicating Temp (°C)	Measured Temperature (°C) at Spread Locations									Uncertainty (± °C)
		#1	#2	#3	#4	#5	#6	#7	#8	#9	
35.0	35.0	34.96	35.01	34.94	35.00	35.01	34.99	34.98	34.97	35.01	0.25
36.0	36.0	35.98	36.03	35.95	36.01	36.04	36.01	35.91	35.99	36.04	0.25
41.6	41.6	41.49	41.53	41.43	41.52	41.59	41.53	41.42	41.53	41.58	0.36

NOTE 1 : THE UNCERTAINTY OF MEASUREMENT EXCLUDED TEMPERATURE UNIFORMITY OF THE CHAMBER.

NOTE 2 : LOCATION 5 WAS REFERENCE LOCATION.

NOTE 3 : THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA.

THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR k=2, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.
END OF CALIBRATION REPORT

F-G010 REV : 03



QUALITY CALIBRATION CO., LTD.
235 Petcharasem 63/2 Road, Laksoeng, Bangkok 10160
Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4584



CERTIFICATE No : 24E11174
REFERENCE No : 75187-2

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : pH METER
MANUFACTURER : TOA DKK
MODEL : HM-41X
SERIAL No : 858942
ID No : EQL-225
CONDITION AS RECEIVED : USED ITEM
SUBMITTED BY : TEST TECH CO., LTD.
30,32 RAMA II SOI 63, RAMA II RD., SAMAEADAM,
BANGKHUNTHIAN, BANGKOK 10150

CALIBRATED BY : PRASERT D.
CALIBRATION DATE : 04-Nov-24

APPROVED BY :
PONGSAK J.
ISSUED DATE : 05-Nov-24
RECEIVED DATE : 04-Nov-24

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.

F-G010 REV 03



QUALITY CALIBRATION CO., LTD.
235 Petcharasem 63/2 Road, Laksoeng, Bangkok 10160
Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4584

CERTIFICATE No : 24E11174

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : pH METER
MANUFACTURER : TOA DKK
ID No : HM-41X
RECEIVED DATE : EQL-225
SERIAL NUMBER : 858942
AMBIENT TEMPERATURE : 04-Nov-24
RELATIVE HUMIDITY : 49 %RH ± 10 % RH

CONDITION OF THIS RESULTS OF CALIBRATION

1. THIS INSTRUMENT WAS CALIBRATED BY DIRECT MEASUREMENT METHOD BASED ON WI-TQ-062 AND WI-TQ-063. THE
DISPLAY UNIT WAS TESTED BY GENERATING STANDARD VOLTAGE TO THE UNIT AND READ THE VALUE COMPARED WITH
CALCULATED VALUE. THE DISPLAY AND ELECTRODE WAS CALIBRATED BY USING STANDARD pH BUFFER SOLUTION.

2. REFERENCE STANDARD INSTRUMENTS :-

INSTRUMENT	MODEL	SERIAL No	CERTIFICATE No	DUE DATE
1) pH STANDARD SOLUTION	00651-06	CC801395	4880-15044130	18-Apr-26
2) pH STANDARD SOLUTION	00651-08	CC801899	4881-15078560	01-May-26
3) pH STANDARD SOLUTION	00651-10	CC803826	4882-15161223	31-May-26
4) PROCESS CALIBRATOR	CA150	91S6079	24E1251	09-Apr-25
5) BATH	260014	1247 48074	24E1251	12-Sep-25
6) THERMOMETER WITH PROBE	421504	5300379	24E1251	12-Sep-25

3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.

4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.

5. THIS CERTIFICATE IS TRACEABLE TO SI UNIT MAINTAINED AT :-

- NATIONAL INSTITUTE OF STANDARD AND TECHNOLOGY, USA.

- NATIONAL INSTITUTE OF METROLOGY (THAILAND)

RESULT OF CALIBRATION : WITHOUT ADJUSTMENT

1. DISPLAY UNIT ONLY

SLOPE FACTOR $k = 2.303 \text{ RT/F} = 59 \text{ mV/pH}$

mV APPLIED	UUC READING (mV)	CORRECTION (mV)	UNCERTAINTY OF MEASUREMENT (±mV)	COVERAGE FACTOR k
177.48	177	0.48	0.59	2.0
0.00	0	0.00	0.59	2.0
-177.48	-178	0.52	0.59	2.0

2. DISPLAY UNIT WITH pH ELECTRODE SN: 903F0080MK

STANDARD pH BUFFER SOLUTION (pH)	UUC READING (pH)	CORRECTION (pH)	UNCERTAINTY OF MEASUREMENT (±pH)	COVERAGE FACTOR k
4.0060	4.01	-0.004	0.013	2.0
7.0000	7.00	0.000	0.013	2.0
10.0070	10.01	-0.003	0.014	2.0

3. DISPLAY UNIT MEASUREMENT TEMPERATURE WITH PROBE

STANDARD READING (°C)	UUC* READING (°C)	IMMERSION DEPTH (mm)	CORRECTION (°C)	UNCERTAINTY OF MEASUREMENT (±°C)
25.004	25.1	80	-0.096	0.21

UUC : UNIT UNDER CALIBRATION

THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA.

THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE
FACTOR k, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.

END OF CALIBRATION REPORT

F-G010 REV 03



BECTHAI BANGKOK EQUIPMENT & CHEMICAL CO., LTD.
CALIBRATION LABORATORY

99/9 Moo 2, Muea Sawet, Phuthamonthon, Nakhon Pathom 73070, Thailand. Tel: +66 3424 5299 Fax: +66 3424 5250
E-mail: bld@becthai.com Website: www.becthai.com



NSC-T18-T18 17825
CALIBRATION 8131

Certificate No. : CAL-24-196

Page : 1 of 3

CERTIFICATE OF CALIBRATION

Equipment : Spectrophotometer
Manufacturer : Thermo Scientific
Model : Genesys 30
Serial No. : 9A1Z089145
ID No. : EQL-251
Customer : TEST TECH CO., LTD.
: 30,32 Rama II Soi 63, Rama II Rd.,
: Samaedam, Bangkhunthian,
: Bangkok 10150
Location : Wastewater Room 3
Date of Receipt : 10 June 2024
Date of Calibration : 10 June 2024
Date of Issue : 13 June 2024
Ambient Temperature : (25±10) °C
Relative Humidity : (60±20) %
Condition As-Received : Used Item

Calibrated by

Mr. Palawat Lunchak

Calibration Engineer

Approved by

(Ms. Jintana Sangthajaroentap)

Calibration Manager

The reported expanded uncertainty of measurement was based on a combined standard uncertainty multiplied by a coverage factor k providing a level of confidence of approximately 95%.
This certificate may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.
Indicated values are valid for the state of the Spectrophotometer at the time of calibration only.

ISSUE: 6 REV:5

FM-CAL-33/2

20/02/24



BECTHAI BANGKOK EQUIPMENT & CHEMICAL CO., LTD.
CALIBRATION LABORATORY

99/9 Moo 2, Muea Sawet, Phuthamonthon, Nakhon Pathom 73070, Thailand. Tel: +66 3424 5299 Fax: +66 3424 5250
E-mail: bld@becthai.com Website: www.becthai.com



NSC-T18-T18 17825
CALIBRATION 8131

Certificate No. : CAL-24-196

Page : 2 of 3

CALIBRATION REPORT

Conditions of this result of calibration

1. Reference Standard Material :

Material	Model	Serial No.	Cert.No.	Due date
Holmium Glass Filter	RM-HG	12705	117342	13 December 2025
Dicynium Glass Filter	RM-DG	13498	117323	13 December 2025
Neutral Density Filter	RM-1N2N3N	8323	117341	13 December 2025

2. Traceability : This certification is traceable to the International System of Unit maintained at:

The Starna Scientific Ltd. Accredited Calibration Laboratory No. 0659.

3. Method of calibration :

The calibration procedure was carried out according to ASTM E275-08 (2022) and ASTM E925-09 (2014).

4. Result of calibration :

(✓) without adjustment : () after adjustment

5. Equipment Specifications:

Spectral Bandwidth : 5 nm
Data Interval : 1 nm
Scan Speed : 1200 nm/min

ISSUE: 6 REV:5

FM-CAL-33/2

20/02/24



BECTHAI BANGKOK EQUIPMENT & CHEMICAL CO., LTD.
CALIBRATION LABORATORY

2992 Moo 2, Mueang Samet, Phraethon, Nakhon Phanom, 29000 Thailand Tel: +66 9424 5259 Fax: +66 9424 5250
E-mail: bkg@becthai.com Website: www.becthai.com



Certificate No. : CAL-24-196

Page : 3 of 3

CALIBRATION REPORT

Wavelength Calibration

Certified Values of Reference Material	Nominal Value (nm)	UUC* Reading (nm)	Error (nm)	Uncertainty of Measurement ($\pm$ nm)	k Factor
361.40	361.40	361	-0.40	0.59	2.00
537.00	537.00	537	0.00	0.59	2.00
879.68	879.68	879	-0.68	0.59	2.00

Photometric Calibration for Visible

Wavelength (nm)	Certified Values of Reference Material (A)	UUC* Reading (A)	Error (A)	Uncertainty of Measurement ($\pm$ A)	k Factor
420.0	Zero	0.000	0.0000	0.0028	2.00
	0.5703	0.569	-0.0013	0.0045	2.00
	0.7336	0.733	-0.0006	0.0045	2.00
440.0	1.0709	1.070	-0.0009	0.0045	2.00
	Zero	0.000	0.0000	0.0028	2.00
	0.5592	0.559	-0.0002	0.0045	2.00
465.0	0.716	0.718	0.0020	0.0045	2.00
	1.0454	1.045	-0.0004	0.0045	2.00
	Zero	0.000	0.0000	0.0028	2.00
546.1 (546.0)	0.5094	0.511	0.0016	0.0045	2.00
	0.6601	0.663	0.0029	0.0045	2.00
	0.963	0.966	0.0030	0.0045	2.00
590.0	Zero	0.000	0.0000	0.0028	2.00
	0.5206	0.522	0.0014	0.0045	2.00
	0.6677	0.670	0.0023	0.0045	2.00
635.0	0.9763	0.977	0.0007	0.0045	2.00
	Zero	0.000	0.0000	0.0028	2.00
	0.5522	0.553	0.0008	0.0045	2.00
	0.6966	0.698	0.0014	0.0045	2.00
	1.0201	1.021	0.0009	0.0045	2.00
	Zero	0.000	0.0000	0.0028	2.00
	0.5377	0.540	0.0023	0.0045	2.00
	0.6649	0.667	0.0021	0.0045	2.00
	0.9736	0.974	0.0004	0.0045	2.00

Remark : Each individual filter is measured against the empty filter holder (blank) used to zero the Spectrophotometer.

Note:

UUC* : Unit Under Calibration

- End of Report -
FM-CAL-33/2

ISSUE: 6 REV:5

20/02/24



Bara Scientific Co., Ltd.
888 U Chu Liang Building Floor7 Rama4 Road
Siam Bangkok Bangkok Thailand 10500
Tel : 02-6324300 Fax : 02-6375486-7
www.barascientific.com



Certificate of Calibration

Number of Page(s) 1 of 3

Certificate No. BSCC-UV-16624
Equipment UV/Vis Spectrophotometer
Model UV-1900i
Manufacturer Shimadzu
Serial No. A12535780311 ML
ID No. EQL-233
Date of receipt 28 April 2024
Date of calibration 28 April 2024
Date of issue 30 April 2024

Customer name Test Tech Co., Ltd.
Address 30,32 Rama II Sol 63, Rama II Road, Sarnae Dam, Bang Khun Thian, Bangkok 10150

Temperature (24.9 - 25.4) °C (On site)
Humidity (49.4 - 51.1) %RH (On site)

Equipment condition Good Operation

Calibration Location Water Room

Calibration Procedure In-house method WI-UV-702-01 based on ASTM E275-01

Traceability
Wavelength Accuracy is traceable to certificate No. 106372 and 106371
Photometric Accuracy is traceable to certificate No. 106364 and 111398
Stray Light is traceable to certificate No. 106377
The above certificate are traceable to SI unit through Stama Scientific Ltd.
(UKAS accredited calibration laboratory NO. 0659)

Calibrated by Mr. Wanchana Janloey

Approved by

Mr. Sonthi Temboonsakdi
Service Manager

The above results are valid exclusively for the calibrated item(s) as mention in this report / certificate.
Advertising the report / certificate and publicity of the results are prohibited and also shall not be reproduced
except in full, without written approval of the Bara Scientific Co., Ltd.



Bara Scientific Co., Ltd.
968 U Chu Liang Building Floor7 Rama4 Road
Siloam Bangkok Thailand 10500
Tel : 02-6324300 Fax : 02-6375498-7
www.barascientific.com

Certificate of Calibration

Certificate No. BSCC-UV-166/24 Number of Page(s) 3 of 3

Calibration Results:

3. Photometric Accuracy (Visible)

Wavelength (nm)	Certified Absorbance (A)	UUC (A)	Error (A)	Uncertainty (±A)
420.0	0.0000	0.0000	0.0000	0.0042
	0.5761	0.5781	0.0030	0.0042
	0.7119	0.7132	0.0013	0.0042
	1.0189	1.0221	0.0032	0.0042
440.0	0.0000	0.0000	0.0000	0.0042
	0.5610	0.5636	0.0026	0.0042
	0.7001	0.7012	0.0011	0.0042
	1.0028	1.0082	0.0026	0.0042
465.0	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
546.1	0.0000	0.0000	0.0000	0.0042
	0.5249	0.5260	0.0011	0.0042
	0.6975	0.6971	-0.0004	0.0042
	1.0009	1.0012	0.0003	0.0042
590.0	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
635.0	0.0000	0.0000	0.0000	0.0042
	0.5668	0.5673	0.0007	0.0042
	0.7620	0.7611	-0.0009	0.0042
	1.0882	1.0876	-0.0006	0.0042

*CNR = Customer not request

4. Stray Light*

Standard cut-off wavelength (nm)	Unit Under Calibration(UUC)	
	Wavelength (nm)	Absorbance (A)
200.85±0.11nm	200.76	0.9795
		2.0091

The Stray light transmission reference is less than 1.0%T and Stray light absorbance reference is greater than 2.00A
*Stray Light not NSC-ONSC Accredited.

The measurement uncertainty is based on a standard uncertainty multiplied by a coverage factor (k=2) providing a level of confidence of approximately 95%.

End of Certificate

The above results are valid exclusively for the calibrated item(s) as mention in this report / certificate. Advertising the report / Certificate and publicity of the results are prohibited and also shall not be reproduced except in full, without written approval of the Bara Scientific Co., Ltd.



Bara Scientific Co., Ltd.
968 U Chu Liang Building Floor7 Rama4 Road
Siloam Bangkok Thailand 10500
Tel : 02-6324300 Fax : 02-6375498-7
www.barascientific.com

Certificate of Calibration

Certificate No. BSCC-UV-166/24 Number of Page(s) 2 of 3

Calibration Results:

1. Wavelength Accuracy

Certified Wavelength (nm)	UUC (nm)	Error (nm)	Uncertainty (±nm)
279.44	279.18	-0.26	0.18
418.53	418.46	-0.07	0.18
536.52	536.54	0.02	0.18
684.50	684.63	0.13	0.18
879.41	879.43	0.02	0.18

2. Photometric Accuracy (UV)

Wavelength (nm)	Certified Absorbance (A)	UUC (A)	Error (A)	Uncertainty (±A)
235	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
257	0.0000	0.0000	0.0000	0.0075
	0.8354	0.8333	-0.0021	0.0075
313	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
350	0.0000	-0.0001	-0.0001	0.0075
	0.6199	0.6190	-0.0009	0.0075

*CNR = Customer not request

The above results are valid exclusively for the calibrated item(s) as mention in this report / certificate. Advertising the report / Certificate and publicity of the results are prohibited and also shall not be reproduced except in full, without written approval of the Bara Scientific Co., Ltd.



บริษัท ซีจี ไซแอนติฟิค จำกัด
CG SCIENTIFIC CO., LTD.

การดูแลบำรุงรักษาเชิงป้องกัน Preventive Maintenance

Customer Name: บริษัท เพรสท์ เทคโนโลยี จำกัด
Product : Distillation Unit
Brand : GERHARDT
Model : Vapodest 30
Serial number : GER003718



บริษัท ซีจี ไซแอนติฟิค จำกัด
CG SCIENTIFIC CO., LTD.

- Part 1: สัญญาการให้บริการ (Service Contact)
Part 2: ข้อมูลพื้นฐานของเครื่องมือ (Details of Instrument)
Part 3: ตรวจสอบประสิทธิภาพเครื่อง
Part 4: รายละเอียดตารางงานผลการให้บริการ Preventive Maintenance
4.1 ขั้นตอนการบริการ
4.2 รายงานผลการให้บริการ
Part 5: ข้อมูลสนับสนุนด้านเทคนิค (General Technical Support)
5.1 Care and Maintenance
5.1.1 การบำรุงรักษาทั่วไป (Basic maintenance)
5.1.2 General error message



บริษัท ซีอี ไซแอนติฟิค จำกัด
CG SCIENTIFIC CO., LTD.

1. ข้อมูลการให้บริการ (Service Contact)

หน่วยงานลูกค้า :	บริษัท เทสท์ เทค จำกัด
ที่อยู่ :	30/32 ซอยพรสวรรค์ 2 ซอย 63 แขวงสามยุค เขตบางขุนเทียน กรุงเทพมหานคร 10150
โทรศัพท์ :	
อีเมล :	
บุคคลติดต่อ :	
ชื่อ-นามสกุล :	คุณ กรรณกาน ขุนพิทักษ์
ตำแหน่ง :	
โทรศัพท์ :	02-893-4211-17
อีเมล :	lab_center@testtech.co.th

สัญญาการบริการจำนวน 1 ครั้ง ต่อ ปี

ครั้งที่ 1 วันที่ : 19 มีนาคม 2568

ครั้งที่ 2 วันที่ :

ครั้งที่ 3 วันที่ :

2. ข้อมูลพื้นฐานของเครื่องมือ (Details of Instrument)

2.1 รายละเอียดเครื่องมือ (Instrument Description)

ประเภทเครื่องมือ :	Distillation Unit
ผลิตภัณฑ์ :	GERHARDT
รุ่น :	Vapodest 30
หมายเลขเครื่อง :	GER003718
หมายเลขซีรียัล :	EQL-062

2.2 ผู้ดำเนินการ

ดำเนินการทำ PM โดย

ชื่อ-นามสกุล :	คุณกรรณกาน ขุนพิทักษ์
ตำแหน่ง :	วิศวกร
แผนก :	บริการหลังการขาย
ฝ่าย :	บริการหลังการขาย




บริษัท ซีอี ไซแอนติฟิค จำกัด
CG SCIENTIFIC CO., LTD.

Part 3: Operational Qualification (OQ)

3.1 ตรวจสอบสภาพเครื่อง


FRONT



No	Description	PASS	FAIL	N/A
1.	Quick clamping device with wedge	☒	☐	☐
2.	Kjeldatherm digestion tube	☒	☐	☐
3.	Holder for steam inlet tubing	☒	☐	☐
4.	PTFP-Inlet tubing, steam	☒	☐	☐
5.	Viton-cone	☒	☐	☐
6.	Clamping for glassware	☒	☐	☐
7.	Screw cap GL18 with silicone seal	☒	☐	☐
8.	PTFP-Inlet tubing, NaOH	☒	☐	☐
9.	PP-Distributor with PP-threaded joint	☒	☐	☐
10.	Distribution head, PP	☒	☐	☐
11.	Screw cap GL32 with silicone seal	☒	☐	☐
12.	Distillation condenser	☒	☐	☐
13.	Screw cap GL14 with plastic screw connection	☒	☐	☐
14.	Display	☒	☐	☐
15.	Keyboard, chemical-resistant	☒	☐	☐
16.	Main switch, green	☒	☐	☐
17.	Ventilation valve	☒	☐	☐
18.	Distillate outlet tubing	☒	☐	☐
19.	Erlenmeyer flask	☒	☐	☐
20.	Platform	☒	☐	☐
21.	Drip tray	☒	☐	☐



REAR



No	Description	PASS	FAIL	N/A
1.	Diaphragm pump NaOH	☒	☐	☐
2.	Diaphragm pump H ₂ O ₂	☒	☐	☐
3.	Diaphragm pump H ₂ O for steam generator	☒	☐	☐
4.	Diaphragm pump H ₂ O for sample	☒	☐	☐
5.	Peristaltic pump for suction sample	☒	☐	☐
6.	Peristaltic pump for suction receiver	☒	☐	☐
7.	Pinch-solenoid valve, steam	☒	☐	☐
8.	Magnetic valve with pressure control	☒	☐	☐
9.	Pinch-solenoid valve, shut-off	☒	☐	☐
10.	Verprene-tubing 4x8 mm.	☒	☐	☐
11.	Verprene-tubing 4x8 mm.	☒	☐	☐
12.	Non-return valve for diaphragm pumps	☒	☐	☐
13.	Tubing reduction PP 51x10x3 mm.	☒	☐	☐
14.	Silicone tubing 4x7 mm.	☒	☐	☐
15.	Silicone tubing 4x7 mm.	☒	☐	☐
16.	Silicone-tubing 4x7 mm.	☒	☐	☐
17.	Verprene-tubing 8x12 mm.	☒	☐	☐
18.	Verprene tubing 4x7 mm.	☒	☐	☐
19.	Silicone tubing 4x7 mm.	☒	☐	☐
20.	Ventilation glass	☒	☐	☐
21.	Novoprene-tubing 4.8x8 mm.	☒	☐	☐
22.	Tubing reduction	☒	☐	☐
23.	Silicone tubing 6x10 mm.	☒	☐	☐
24.	PP-distributor with PP-thread	☒	☐	☐
25.	SKT-valve (built in with brass fitting)	☒	☐	☐
26.	Silicone tubing 8x16x80 mm.	☒	☐	☐
27.	Steam generator	☒	☐	☐
28.	PTFE-inlet tubing NaOH	☒	☐	☐
29.	Silicone tubing 8x16 for cooling water inlet	☒	☐	☐
30.	Silicone tubing 8x16 for cooling water outlet	☒	☐	☐
31.	Viton-tubing 6x12*50 mm.	☒	☐	☐
32.	Silicone tubing 4x7 mm.	☒	☐	☐



Part 4:

รายละเอียดการตรวจสอบ

4.1 ขั้นตอนการบริการ

ตรวจสอบระบบไฟฟ้า (Electrical Test)

- ความต้านทานทางไฟฟ้าของเครื่องกับกราวด์
- กระแสไฟฟ้าที่ใช้งาน

PASS	FAIL	Remark
☒	☐	
☒	☐	

ตรวจสอบสภาพเครื่อง (Optical Test)

- Main cable
- Electric wiring
- Pumps
- Distribution Head
- Condensor
- Steam generator
- Tubing
- Viton cone

PASS	FAIL	Remark
☒	☐	
☒	☐	
☒	☐	
☒	☐	
☒	☐	
☒	☐	
☒	☐	
☒	☐	

ตรวจสอบ Function การทำงาน (The Function Test)

- ระบบเครื่องและควบคุมความดันของ Steam
- ระบบการต้มน้ำเข้า Sample Tube
- ระบบการต้ม NaOH
- ระบบการ Suction ตั้ง Sample Tube และ Receiver

PASS	FAIL	Remark
☒	☐	
☒	☐	
☒	☐	
☒	☐	

4.2 รายงานผลการให้บริการ

1. TECHNICAL DATA

Main Supply 220 volt + 10% 50 Hz
Nominal current

PASS	FAIL	Remark
☒	☐	
☒	☐	

1.1 COOLING WATER BATH

Temperature 15-20 °C
Cooling Water Outlet
Control Temperature

PASS	FAIL	Remark
☒	☐	
☒	☐	
☒	☐	

1.2 OPTICAL TEST VAP 30

Screw cap GL14
Screw cap GL18
Screw cap GL32
Distillation Head
Condensor
Viton Cone
Ventilation Valve
Micro Switch Sample

PASS	FAIL	Remark
☒	☐	
☒	☐	
☒	☐	
☒	☐	
☒	☐	
☒	☐	
☒	☐	
☒	☐	

2. SYSTEM COOLING WATER INLET

Cooling Water Inlet
Cooling Water Outlet
Flow control valve

PASS	FAIL	Remark
☒	☐	
☒	☐	
☒	☐	

3. SYSTEM CONTROL

	PASS	FAIL	Remark
Key Board	☒	☐	
Display	☒	☐	
Program	☒	☐	
Adding H ₂ O	☒	☐	*Vap30/40
Adding NaOH	☒	☐	
Adding H ₃ BO ₃	☒	☐	*Vap40
Suction Sample	☒	☐	*Vap30/40

4. SYSTEM DISTILLATION

	PASS	FAIL	Remark
Boiler	☒	☐	
Level Sensor	☒	☐	
Neoprene-Tubing	☒	☐	
Solenoid Valve Shut-Off	☒	☐	
Solenoid Valve Steam	☒	☐	
Excess Pressure Detector	☒	☐	
Ventilation Valve	☒	☐	
Heater	☒	☐	

5. PUMP

	PASS	FAIL	Remark
Pump H ₂ O Steam	☒	☐	
- Non-Return Valve	☒	☐	
Pump H ₂ O Sample	☒	☐	
- Non-Return Valve	☒	☐	
Pump NaOH	☒	☐	
- Non-Return Valve	☒	☐	
Pump H ₃ BO ₃	☒	☐	
- Non-Return Valve	☒	☐	N/A
Pump suction	☒	☐	

6. THE FOLLOWING PROGRAM RUN

	PASS	FAIL	Remark
Addition H ₂ O	☒	☐	0-99 sec.
Addition NaOH	☒	☐	0-99 sec.
Addition H ₃ BO ₃	☒	☐	0-99 sec.
Reaction Time	☒	☐	0-99 min
Distillation Time	☒	☐	0-99 min
Steam Capacity	☒	☐	30%-100%
Suction Time	☒	☐	0-99 sec.

The instrument is in perfect technical shape

Remark :



4.2 Error Code

The micro-processor continually surveys all the functions of the distillation system. As soon as an error arises it is shown on the display and accompanied by an acoustic signal.

Error message	Measures
No tap water	Check cooling water inlet for blockages. Ensure the tap is turned on. → Error
No sample tube	Insert tube. → Error
Check chemical	Check set of tanks. → Error
Low water Press Error	Check the water inlet distilled H ₂ O. → Error
Filling steam generator	This message disappears as soon as steam generator is filled.

After the above mentioned errors are corrected, the following message is displayed.

Error message	Measures
Stop Prog.No.x continue=Enter	Enter = continue of interrupted program Reset = Standby-mode

Other error messages

Error message	Measures
Wait for steam	Message disappears as soon as stand-by is reached
Add sol. > 1min Continue=Enter	Check programming Enter=continue of interrupted program Reset=Standby-mode
Program undefined	Check programming → Reset
Excess steam pressure	Switch the system off and call service
Sensor error	Switch the system off and call service



SCIMET Co., Ltd.
1194 Soi Wachitrathansathit 57, Bangkok,
Phrakhanong, Bangkok 10260 Thailand
Email: scimet2022@gmail.com, Tel: 02 460 9239
https://www.scimet.co.th

Certificate No. C06240053

Calibration Certificate

Equipment: TURBIDIMETER
Model: 2100N
Serial No.(or ID): 970400003415 (EQL-024)
Manufacturer: HACH
Condition: In Condition
Customer: TEST TECH CO., LTD.
30,32 Rama II Soi 63, Rama II Rd., Samaedam, Bangkokthunien Bangkok 10150 Thailand

Calibration Place
TEST TECH CO., LTD.(503 ซอยทดสอบใหม่)
30,32 Rama II Soi 63, Rama II Rd., Samaedam, Bangkokthunien Bangkok 10150 Thailand


Calibration Date
12 September 2024

Environment Condition
Temperature: 24.4 °C ± 0.2 °C
Humidity: 58.4 %RH ± 1.1 %RH

The Method used
In-house method, W106, based on Hach Manufacturer Method 8195


Traceability
This certificate is traceable to Primary standard Fromazin and Stabical accepted by United States Environmental Protection Agency (EPA) through Hach Company Certificate No. A3304, A3312, A3305, A3304, A3305

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.
The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor ($k=2$) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).
These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of SCIMET Co., Ltd.



(Mr. Dumrong Boonsopon)
Person in charge

(Mr. Thalengkeat Pongngam)
Authorized signatory



Calibration Results:

Before Adjustment

Std Turbidity (NTU)	UUC Reading	Correction	Deviation	Uncertainty
0.040	0.121	-0.081	0.002	0.070
20.40	20.0	0.40	0.0	1.0
205.0	198	7.0	0.3	10
1026.0	995	31.0	0.0	50
4114.0	4017	97.0	5.0	200

After Adjustment

Std Turbidity (NTU)	UUC Reading	Correction	Deviation	Uncertainty
0.040	0.122	-0.082	0.003	0.070
20.40	20.5	-0.10	0.0	1.0
205.0	205	0.0	0.3	10
1026.0	1027	-1.0	0.3	50
4114.0	4113	1.0	1.1	200

The End of Certificate