ภาคผนวก ก.-9

Ref.เพิ่มเติม

หนังสือรับรองจดทะเบียนเป็นนิติบุคคล_EN MECH

HVE สำเนาใบรับรองห้องปฏิบัติการทดสอบ ISO-IEC 17025 : 2017

HVE สำเนาใบอนุญาตการขึ้นทะเบียนห้องปฏิบัติการ

HVE เอกสารเทียบเครื่องมือวัด



ที่ E10091220869421

สำนักงานทะเบียนหุ้นส่วนบริษัทกลาง
กรมพัฒนาธุรกิจการค้า กระทรวงพาณิชย์

หนังสือรับรอง

ขอรับรองว่าบริษัทนี้ ได้จดทะเบียนเป็นนิติบุคคล ตามประมวลกฎหมายแพ่งและพาณิชย์ เมื่อวันที่ 15 สิงหาคม 2551 ทะเบียนนิติบุคคลเลขที่ 0105551092070 ปรากฏข้อความในรายการตามเอกสารทะเบียนนิติบุคคล ณ วันออกหนังสือนี้ ดังนี้

1. ชื่อบริษัท บริษัท เอ็น เมค จำกัด
2. กรรมการของบริษัทมี 1 คน ตามรายชื่อดังต่อไปนี้
    1. นางสาวจินตนา อภิวัฒนลังการ/
3. จำนวนหรือชื่อกรรมการซึ่งลงชื่อผูกพันบริษัทได้คือ กรรมการหนึ่งคนลงลายมือชื่อและประทับตราสำคัญของบริษัท/
4. ทุนจดทะเบียน 2,000,000.00 บาท / สองล้านบาทถ้วน/
5. สำนักงานแห่งใหญ่ ตั้งอยู่เลขที่ 601 ซอยจรัญสนิทวงศ์46 ถนนจรัญสนิทวงศ์ แขวงบางยี่ขัน เขตบางพลัด กรุงเทพมหานคร/
6. วัตถุที่ประสงค์ของบริษัทมี 36 ข้อ ดังปรากฏในสำเนาเอกสารแนบท้ายหนังสือรับรองนี้ จำนวน 4 แผ่น โดยมีลายมือชื่อนายทะเบียนซึ่งรับรองเอกสารเป็นสำคัญ

ออกให้ ณ วันที่ 30 เดือน กันยายน พ.ศ. 2568

(นายวีรวิทย์ พันธุรัตน์)
นายทะเบียน

คำเตือน : หนังสือรับรองฉบับนี้พิมพ์ออกจากต้นฉบับที่เป็นไฟล์อิเล็กทรอนิกส์ การสั่งพิมพ์ถือเป็นสำเนาเอกสาร

DBD กรมพัฒนาธุรกิจการค้า กระทรวงพาณิชย์
Department of Business Development
Ministry of Commerce

ก้าวล้ำนำธุรกิจ สู่ยุคดิจิทัล
Leading Business Towards Digital Transformation

หนังสือรับรองฉบับนี้ถูกจัดทำด้วยวิธีการทางอิเล็กทรอนิกส์ โดยเป็นข้อมูล ณ วันที่ออกเอกสาร ทั้งนี้ ในการใช้งาน ผู้ใช้ควรตรวจสอบข้อควรทราบท้ายหนังสือรับรองฉบับนี้ทุกครั้ง และสามารถตรวจสอบเอกสารฉบับนี้ผ่านทาง QR Code และเว็บไซต์กรม (www.dbd.go.th) ได้ภายใน 1 ปี นับจากวันที่ออกหนังสือรับรอง

Ref:E6810091220869421
ออกให้ ณ วันที่ : 2025-09-30 T09:42:02+0700



ที่ E10091220869421

สำนักงานทะเบียนหุ้นส่วนบริษัทกลาง
กรมพัฒนาธุรกิจการค้า กระทรวงพาณิชย์

หนังสือรับรอง

ข้อควรทราบ ประกอบหนังสือรับรอง ฉบับที่ E10091220869421

1. นิติบุคคลนี้ได้ส่งงบการเงินปี 2567

2. หนังสือนี้รับรองเฉพาะข้อความที่ห้าง/บริษัทได้นำมาจดทะเบียนไว้เพื่อผลทางกฎหมายเท่านั้น ข้อเท็จจริงเป็นสิ่งที่ควรหาไว้พิจารณาฐานะ

3. นายทะเบียนอาจเพิกถอนการจดทะเบียน ถ้าปรากฏว่าข้อความอันเป็นสาระสำคัญที่จดทะเบียนไม่ถูกต้อง หรือเป็นเท็จ

ว.บ (พิเศษ)

รายละเอียดวัตถุที่ประสงค์

วัตถุที่ประสงค์ทั่วไป

(1) ซื้อ จัดหา รับ เช่า เช่าซื้อ ถือกรรมสิทธิ์ ครอบครอง ปรับปรุง ใช้ และการจัดการโดยประการอื่น ซึ่งทรัพย์สินใดๆ ตลอดจนดอกผลของทรัพย์สินนั้น

(2) ขาย โอน จำนอง จำนำ แลกเปลี่ยน และจำหน่ายทรัพย์สินโดยประการอื่น

(3) เป็นนายหน้า ตัวแทน ตัวแทนค้าต่างในกิจการและธุรกิจทุกประเภท เว้นแต่ในธุรกิจประกันภัย การหาสมาชิกให้สมาคม และการค้าหลักทรัพย์

(4) กู้ยืมเงิน เบิกเงินเกินบัญชีจากธนาคาร นิติบุคคล หรือสถาบันการเงินอื่น และให้กู้ยืมเงินหรือให้เครดิต ด้วยวิธีการอื่น โดยจะมีหลักประกันหรือไม่ก็ตาม รวมทั้งการรับ ออก โอน และสลักหลังตั๋วเงิน หรือตราสารที่เปลี่ยนมือได้อย่างอื่น เว้นแต่ในธุรกิจธนาคาร ธุรกิจเงินทุน และธุรกิจเครดิตฟองซิเอร์

(5) ทำการจัดตั้งสำนักงานสาขาหรือแต่งตั้งตัวแทน ทั้งภายในและภายนอกประเทศ

(6) เข้าเป็นหุ้นส่วนจำกัดความรับผิดในห้างหุ้นส่วนจำกัด เป็นผู้ถือหุ้นในบริษัทจำกัด และบริษัทมหาชนจำกัด

วัตถุที่ประสงค์ประกอบธุรกิจบริการ

(7) ประกอบกิจการรับเหมาก่อสร้างอาคาร อาคารพาณิชย์ อาคารที่พักอาศัย สถานที่ทำการ ถนน สะพาน เขื่อน อุโมงค์ และงานก่อสร้างอย่างอื่นทุกชนิด รวมทั้งรับทำงานโยธาทุกประเภท

(8) ประกอบกิจการโรงแรม ภัตตาคาร บาร์ ไนท์คลับ

(9) ประกอบกิจการขนส่งและขนถ่ายสินค้า และคนโดยสารทั้งทางบก ทางน้ำ ทางอากาศ ทั้งภายในประเทศ และระหว่างประเทศ รวมทั้งรับบริการนำของออกจากท่าเรือตามพิธีศุลกากรและการจัดระวางการขนส่งทุกชนิด

(10) ประกอบกิจการนำเที่ยว รวมทั้งธุรกิจที่เกี่ยวข้องกับการนำเที่ยวทุกชนิด

(11) ประกอบกิจการบริการจัดเก็บ รวบรวม จัดทำ จัดพิมพ์และเผยแพร่สถิติ ข้อมูลในทางเกษตรกรรม อุตสาหกรรม พาณิชยกรรม การเงิน การตลาด รวมทั้งวิเคราะห์และประเมินผลในการดำเนินธุรกิจ

(12) ประกอบกิจการบริการทางด้านกฎหมาย ทางบัญชี ทางวิศวกรรม ทางสถาปัตยกรรม รวมทั้งกิจการโฆษณา

(13) ประกอบธุรกิจบริการรับค้ำประกันหนี้สิน ความรับผิด และการปฏิบัติตามสัญญาของบุคคลอื่น รวมทั้งรับบริการค้ำประกันบุคคล ซึ่งเดินทางเข้ามาในประเทศหรือเดินทางออกไปต่างประเทศตามกฎหมายว่าด้วยคนเข้าเมือง กฎหมายว่าด้วยภาษีอากร และกฎหมายอื่น

(14) ประกอบธุรกิจบริการรับเป็นที่ปรึกษาและให้คำแนะนำปัญหาเกี่ยวกับด้านบริหารงานพาณิชยกรรม อุตสาหกรรม รวมทั้งปัญหาการผลิต การตลาดและจัดจำหน่าย

(15) ประกอบธุรกิจบริการรับเป็นผู้จัดการและดูแลผลประโยชน์ เก็บผลประโยชน์และจัดการทรัพย์สินให้บุคคลอื่น

(16) ประกอบกิจการโรงพยาบาลเอกชน สถานพยาบาล รับรักษาคนไข้และผู้ป่วยเจ็บ รับทำการฝึกสอน และอบรมทางด้านวิชาการเกี่ยวกับการแพทย์ การอนามัย

(17) ประกอบกิจการจัดสร้างและจัดจำหน่ายภาพยนตร์ โรงภาพยนตร์ และโรงมหรสพอื่น สถานพักตากอากาศ สนามกีฬา สระว่ายน้ำ โบว์ลิ่ง

(18) ประกอบกิจการให้บริการซ่อมแซม บำรุงรักษา ตรวจสอบ อัดฉีด พ่นน้ำยากันสนิมสำหรับยานพาหนะทุกประเภท รวมทั้งบริการติดตั้ง ตรวจสอบ และแก้ไขอุปกรณ์ ป้องกันวินาศภัยทุกประเภท

(19) ประกอบกิจการซักรีดเสื้อผ้า ตัดผม แต่งผม เสริมสวย

(20) ประกอบกิจการรับจ้างถ่ายรูป ล้าง อัด ขยายรูป รวมทั้งเอกสาร

(21) ประกอบกิจการสถานบริการอาบอบนวด

(22) ประกอบกิจการประมูลเพื่อรับจ้างทำของ ตามวัตถุที่ประสงค์ทั้งหมด ให้แก่บุคคล คณะบุคคล นิติบุคคล ส่วนราชการและองค์การของรัฐ

ห้างหุ้นส่วน/บริษัท เอ็น เมค จำกัด
ทะเบียนเลขที่ 0105551092070

วัตถุที่ประสงค์ของ ห้างหุ้นส่วน/บริษัท นี้ มี 36 ข้อ ดังนี้

(23) ประกอบกิจการค้า ผ้า ผ้าทอจากใยสังเคราะห์ ด้าย ด้ายยางยืด เส้นใยไนล่อน ใยสังเคราะห์ เส้นด้ายยืด เครื่องนุ่งห่ม เสื้อผ้าสำเร็จรูป เครื่องแต่งกาย เครื่องประดับกาย ถุงเท้า ถุงน่อง เครื่องหนัง รองเท้า กระเป๋า เครื่องอุปโภคอื่น สิ่งทอ อุปกรณ์การเล่นกีฬา

(24) ประกอบกิจการค้า เครื่องเคหะภัณฑ์ เครื่องเรือน เฟอร์นิเจอร์ เครื่องแก้ว เครื่องครัว ตู้เย็น เครื่องปรับอากาศ เครื่องฟอกอากาศ พัดลม เครื่องดูดอากาศ หม้อหุงข้าวไฟฟ้า เตารีดไฟฟ้า เครื่องทำความร้อน เครื่องทำความเย็น เตาอบไมโครเวฟ เครื่องใช้ไฟฟ้า อุปกรณ์ไฟฟ้า รวมทั้งอะไหล่และอุปกรณ์ของสินค้าดังกล่าว

(25) ประกอบกิจการค้า วัสดุก่อสร้าง อุปกรณ์และเครื่องมือใช้ในการก่อสร้าง เครื่องมือช่างทุกประเภท สี เครื่องมือทาสี เครื่องตกแต่งอาคาร เครื่องเหล็ก เครื่องทองแดง เครื่องทองเหลือง เครื่องเคลือบ เครื่องสุขภัณฑ์ อุปกรณ์ประปา รวมทั้งอะไหล่และอุปกรณ์ของสินค้าดังกล่าว

(26) ประกอบกิจการค้า เครื่องจักร เครื่องยนต์ เครื่องมือกล เครื่องทุ่นแรง ยานพาหนะ เครื่องกำเนิดไฟฟ้า เครื่องสูบน้ำ เครื่องบำบัดน้ำเสีย และเครื่องกำจัดขยะ

(27) ประกอบกิจการค้า ยา ยารักษาโรค เภสัชภัณฑ์ เคมีภัณฑ์ เครื่องมือแพทย์ เครื่องมือเครื่องใช้ทางวิทยาศาสตร์ ปุ๋ย ยาปราบศัตรูพืช ยาบำรุงพืชและสัตว์ทุกชนิด

(28) ประกอบกิจการค้า กระดาษ เครื่องเขียน แบบเรียน แบบพิมพ์ หนังสือ อุปกรณ์การเรียนการสอน อุปกรณ์การถ่ายภาพและภาพยนตร์ เครื่องคำนวณ เครื่องพิมพ์ อุปกรณ์การพิมพ์ สิ่งพิมพ์ หนังสือพิมพ์ ตู้เก็บเอกสาร เครื่องใช้สำนักงาน เครื่องมือสื่อสาร คอมพิวเตอร์ รวมทั้งอุปกรณ์และอะไหล่ของสินค้าดังกล่าว

(29) ออกแบบ ผลิต จำหน่าย นำเข้า ส่งออก เครื่องจักรกล อะไหล่เครื่องจักรกล ทุกประเภท

(30) ประกอบกิจการรับเหมางานก่อสร้าง งานระบบต่างๆเช่นระบบไฟฟ้า สุขาภิบาล และปรับอากาศ

(31) ประกอบกิจการ การบริหารจัดการดูแลงานระบบอาคาร เช่น ระบบปรับอากาศ, ไฟฟ้า เครื่องกล ไฟฟ้าสำรองและฉุกเฉิน ระบบป้องกันอัคคีภัย และอื่นๆ ที่เกี่ยวข้องกับอาคาร

หนังสือรับรองฉบับนี้ถูกจัดทำด้วยวิธีการทางอิเล็กทรอนิกส์ โดยเป็นข้อมูล ณ วันที่ออกเอกสาร ทั้งนี้ ในการใช้งาน ผู้ใช้ควรตรวจสอบข้อควรทราบท้ายหนังสือรับรองฉบับนี้ทุกครั้ง และสามารถตรวจสอบเอกสารฉบับนี้ผ่านทาง QR Code และเว็บไซต์กรม (www.dbd.go.th) ได้ภายใน 1 ปี นับจากวันที่ออกหนังสือรับรอง

ห้างหุ้นส่วน/บริษัท เอ็น เมค จำกัด
ทะเบียนเลขที่ 0105561092070

วัตถุที่ประสงค์ของ ห้างหุ้นส่วน/บริษัท นี้ มี 36 ข้อ ดังนี้

(32) ประกอบกิจการ ควบคุมดูแลและซ่อมแซมเครื่องจักรอุปกรณ์ทางด้านสิ่งแวดล้อม เช่น ระบบบำบัดน้ำเสีย ระบบบำบัดมลพิษทางอากาศ ระบบเครื่องทำอากาศ ระบบผลิตน้ำประปา ระบบกำจัดกากไขมัน กากตะกอนและกากอุตสาหกรรม ระบบการจัดการมลพิษทางเสียงและระบบดับเพลิง

(33) ประกอบกิจการรับเหมาก่อสร้าง ปรับปรุง ซ่อมแซม ดูแล งานทางด้านสิ่งแวดล้อม เช่น ระบบบำบัดมลพิษทางน้ำ ระบบบำบัดมลพิษทางอากาศ ระบบการจัดการมลพิษกากอุตสาหกรรม ระบบผลิตน้ำประปา ระบบการจัดการมลพิษทางเสียง ระบบดับเพลิง และงานเกี่ยวกับระบบท่อภายในอาคาร รับออกแบบและให้คำปรึกษาเกี่ยวกับกิจการข้างต้น

(34) สั่งเข้ามาจำหน่ายในประเทศและส่งออกไปจำหน่ายยังต่างประเทศซึ่งสินค้าตามที่กำหนดไว้ในวัตถุที่ประสงค์

(35) ทำการประมูลเพื่อขายสินค้าตามวัตถุที่ประสงค์ให้แก่บุคคล คณะบุคคล นิติบุคคล ส่วนราชการ และองค์การของรัฐทั้งภายในและภายนอกประเทศ

(36) บริษัทมีสิทธิออกหุ้นในราคา สูงกว่าค่าหุ้นที่กำหนดไว้

หนังสือรับรองฉบับนี้ถูกจัดทำด้วยวิธีการทางอิเล็กทรอนิกส์ โดยเป็นข้อมูล ณ วันที่ออกเอกสาร ทั้งนี้ ในการใช้งาน ผู้ใช้ควรตรวจสอบข้อควรทราบท้ายหนังสือรับรองฉบับนี้ทุกครั้ง และสามารถตรวจสอบเอกสารฉบับนี้ผ่านทาง QR Code และเว็บไซต์กรม (www.dbd.go.th) ได้ภายใน 1 ปี นับจากวันที่ออกหนังสือรับรอง



ที่ อก ๐๓๑๐(๑)/ ๒๖๘๒

กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท
เขตราชเทวี กรุงเทพฯ ๑๐๔๐๐

๒๕ มีนาคม ๒๕๖๘

เรื่อง ยกเลิกบุคลากรของห้องปฏิบัติการวิเคราะห์

เรียน กรรมการผู้จัดการ บริษัท เอชวีอี จำกัด

อ้างถึง คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และขนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน ลงวันที่ ๑๗ มีนาคม ๒๕๖๘

ตามคำขอที่อ้างถึง บริษัท เอชวีอี จำกัด ห้องปฏิบัติการวิเคราะห์เอกชน เลขทะเบียน ว-๓๕๘ สถานที่ตั้งเลขที่ ๖๐๓ ซอยจรัญสนิทวงศ์ ๔๖ ถนนจรัญสนิทวงศ์ แขวงบางยี่ขัน เขตบางพลัด กรุงเทพมหานคร ขอยกเลิกบุคลากร ความละเอียดแจ้งแล้ว นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว มีความเห็นดังนี้

๑. ให้ยกเลิกผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ จำนวน ๑ ราย
นายเอกลักษณ์ ลีลาบริหาร ทะเบียนเลขที่ ว-๓๕๘-ค-๐๐๐๑

๒. ให้ยกเลิกเจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๒ ราย
๑) นางสาวจิราภรณ์ ผงผานอก ทะเบียนเลขที่ ว-๓๕๘-จ-๐๐๐๒
๒) นางสาวณัฐรดา คงบัน ทะเบียนเลขที่ ว-๓๕๘-จ-๐๐๐๕

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ

(นายธีรทัศน์ อิศรางกูร ณ อยุธยา)
รองอธิบดี ปฏิบัติราชการแทน
อธิบดีกรมโรงงานอุตสาหกรรม

กองวิจัยและเตือนภัยมลพิษโรงงาน
กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ
โทร. ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๐๓-๕
โทรสาร ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๙๙
ไปรษณีย์อิเล็กทรอนิกส์ saraban@diw.mail.go.th



ที่ อก ๐๓๑๐(๑)/ ๑ ๒ ๓ ๗ ๒

กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท
เขตราชเทวี กรุงเทพฯ ๑๐๔๐๐

๒ ๘ สิงหาคม ๒๕๖๖

เรื่อง ขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

เรียน กรรมการผู้จัดการ บริษัท เอชวีอี จำกัด

อ้างถึง คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และขนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน ลงวันที่ ๑๘ พฤษภาคม ๒๕๖๖

สิ่งที่ส่งมาด้วย เอกสารแนบท้ายหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน บริษัท เอชวีอี จำกัด จำนวน ๒ แผ่น

ตามหนังสือที่อ้างถึง บริษัท เอชวีอี จำกัด ขอขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน พร้อมรายชื่อผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์และรายการสารมลพิษที่จะทำการวิเคราะห์ ต่อกรมโรงงานอุตสาหกรรม นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว ให้บริษัท เอชวีอี จำกัด ขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน มีเลขทะเบียน ว-๓๕๘ สถานที่ตั้งเลขที่ ๖๐๓ ซอยจรัญสนิทวงศ์ ๔๖ แขวงบางยี่ขัน เขตบางพลัด กรุงเทพมหานคร โดยมีองค์ประกอบดังนี้

ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์

๑) นายเอกลักษณ์ ลีลาบริหาร	ทะเบียนเลขที่ ว-๓๕๘-ค-๐๐๐๑
๒) นางสาวทิพวรรณ วงค์บุญตัน	ทะเบียนเลขที่ ว-๓๕๘-ค-๐๐๐๒
๓) นายพล ม่วงใหญ่	ทะเบียนเลขที่ ว-๓๕๘-ค-๐๐๐๓

ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์

๑) นางสาวสุปรียา หล้าอิน	ทะเบียนเลขที่ ว-๓๕๘-จ-๐๐๐๑
๒) นางสาวจิราภรณ์ ผงผานอก	ทะเบียนเลขที่ ว-๓๕๘-จ-๐๐๐๒
๓) นายธรรมรัตน์ จริยวัฒนสุข	ทะเบียนเลขที่ ว-๓๕๘-จ-๐๐๐๓
๔) นางสาวธนภรณ์ ก่ำทา	ทะเบียนเลขที่ ว-๓๕๘-จ-๐๐๐๔
๕) นางสาวณัฐรดา คงบัน	ทะเบียนเลขที่ ว-๓๕๘-จ-๐๐๐๕
๖) นายฐานันท์ นิภารัตน์	ทะเบียนเลขที่ ว-๓๕๘-จ-๐๐๐๖
๗) นายมโนรถ สุดจันทร์	ทะเบียนเลขที่ ว-๓๕๘-จ-๐๐๐๗

ค. ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนให้วิเคราะห์ในน้ำเสียและอากาศเสีย ตามสิ่งที่ส่งมาด้วย

หนังสือฉบับ...

หนังสือฉบับนี้มีอายุครั้งละ ๓ ปี นับจากวันที่กรมโรงงานอุตสาหกรรมออกหนังสือหากประสงค์จะต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ให้ยื่นคำขอต่ออายุพร้อมเอกสารประกอบคำขอต่อกรมโรงงานอุตสาหกรรม ภายใน ๓๐ วัน ก่อนวันสิ้นอายุของหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชนทั้งนี้ สามารถยื่นคำขอผ่านระบบอิเล็กทรอนิกส์ได้ที่หน้าเว็บไซต์กรมโรงงานอุตสาหกรรม

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ

(นายประสม ตำรงพงษ์)
ผู้อำนวยการกองวิจัยและเตือนภัยมลพิษโรงงาน
ปฏิบัติราชการแทนอธิบดีกรมโรงงานอุตสาหกรรม

กองวิจัยและเตือนภัยมลพิษโรงงาน
กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ
โทร. ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๐๓-๕
โทรสาร ๐ ๒๔๓๐ ๖๓๑๒ ต่อ๒๑๙๙
ไปรษณีย์อิเล็กทรอนิกส์ saraban@diw.mail.go.th

เอกสารแนบท้ายหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
บริษัท เอชวีอี จำกัด เลขทะเบียน ว-๓๕๘
ที่ อก ๐๓๑๐(๑)/ ๑ ๒ ๓ ๗ ๒ ลงวันที่ ๒ ๘ สิงหาคม ๒๕๖๖

ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๓๑ รายการ

น้ำเสีย จำนวน 30 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Aldicarb	High-Performance Liquid Chromatographic Method [1]
2	Aldicarb sulfone	High-Performance Liquid Chromatographic Method [1]
3	Aldicarb sulfoxide	High-Performance Liquid Chromatographic Method [1]
4	Aldrin	Liquid-Liquid Extraction, Gas Chromatographic Method [1]
5	α-BHC	Liquid-Liquid Extraction, Gas Chromatographic Method [1]
6	Biochemical Oxygen Demand	5-Day BOD Test, Azide Modification Method [1]
7	Carbaryl	High-Performance Liquid Chromatographic Method [1]
8	Carbofuran	High-Performance Liquid Chromatographic Method [1]
9	Chemical Oxygen Demand	Closed Reflux, Titrimetric Method [1]
10	4,4′-DDD	Liquid-Liquid Extraction, Gas Chromatographic Method [1]
11	4,4′-DDE	Liquid-Liquid Extraction, Gas Chromatographic Method [1]
12	Dieldrin	Liquid-Liquid Extraction, Gas Chromatographic Method [1]
13	Endosulfan I	Liquid-Liquid Extraction, Gas Chromatographic Method [1]
14	Endosulfan II	Liquid-Liquid Extraction, Gas Chromatographic Method [1]
15	Endrin	Liquid-Liquid Extraction, Gas Chromatographic Method [1]
16	Heptachlor	Liquid-Liquid Extraction, Gas Chromatographic Method [1]
17	Heptachlo Epoxide	Liquid-Liquid Extraction, Gas Chromatographic Method [1]
18	3-Hydroxycarbofuran	High-Performance Liquid Chromatographic Method [1]
19	Methiocarb	High-Performance Liquid Chromatographic Method [1]
20	Methomyl	High-Performance Liquid Chromatographic Method [1]
21	1-Naphthol	High-Performance Liquid Chromatographic Method [1]
22	Oil & Grease	Liquid-Liquid, Partition-Gravimetric Method [1]
23	Oxamyl	High-Performance Liquid Chromatographic Method [1]
24	pH	Electrometric Method [1]
25	Propoxur	High-Performance Liquid Chromatographic Method [1]
26	Sulfide	Iodometric Method [1]
27	Temperature	Laboratory and Field Methods [1]
28	Total Dissolved Solids	Dried at 180 °C [1]
29	Total Kjeldahl Nitrogen	Macro Kjeldahl Method [1]
30	Total Suspended Solids	Dried at 103-105 °C [1]

อากาศเสีย ...

อากาศเสีย (ปล่องระบาย) จำนวน 1 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Opacity	Ringelmann's Method [2]

เอกสารอ้างอิง

1. APHA, AWWA, WEF.Standard Methods for the Examination of Water and Wastewater, 23rd ed. Washington, DC: APHA, 2017.

2. กระทรวงอุตสาหกรรม. ประกาศกระทรวงอุตสาหกรรม เรื่อง กำหนดค่าปริมาณเขม่าควันที่เจือปนในอากาศที่ระบายออกจากปล่องของหม้อน้ำของโรงงาน พ.ศ. 2549. ราชกิจจานุเบกษา. 4 ธันวาคม 2549. เล่มที่ 123 ตอนพิเศษ 125ง.

DIGITAL CALIBRATION CO.,LTD.
103, Soi Ekachai 132, Ekachai Rd.,
Bang Bon Nuea, Bang Bon, Bangkok 10150
Phone : 02-862-5226-7 Email : Calibration@digitalscale.co.th

NSC – TISI – TIS 17025
CALIBRATION 0262

CALIBRATION CERTIFICATE

Certificate No. : 24EB 542
Job No. : CI2412-043

Issue Date	:	26 December 2024
Location	:	Lab Calibration
Customer Name	:	HVE CO., LTD. 603 Soi Jarunsanitwong 46 Jarunsanitwong Road Bangyeekhan Bangplad Bangkok 10700
Equipment Name	:	Electronic Balance
Manufacturer	:	SHIMADZU
Model	:	ATX224R
Serial No.	:	D326005191
ID No.	:	-
Weighing Capacity	:	220 g
Resolution	:	0.0001 g
Received Date	:	25 December 2024
Condition of calibrated Item	:	Good
Calibration Date	:	26 December 2024
Ambient Temperature	:	(22 - 28) °C
Relative Humidity	:	(30 - 70) %
Atmospheric Pressure	:	(990 - 1030) mbar
Procedure Used	:	This calibration was conducted by using in-house calibration procedure number CP-B01-01 based on UKAS LAB14 Edition 7 November 2022

Reference Standard	:	Instrument	ID No.	Certificate No.	Due date
		Weight Set E2 (1 mg to 1 kg)	DCC 0001-13	M2306115S	21-Jun-2026

DIGITAL CALIBRATION

This certification is traceable to the International System of Unit

Calibrated by : Krissana Thongklang

Approved by : _____

( ) Mr.Chaiyapatr (Laboratory Manager)
( ) Mr.Boonchuay Muenchaisit (Technical Manager)
(✓) Mr.Pisut Nantipawan (Quality Manager)

The Uncertainties are for a Confidence Probability of Approximately 95%.
This certificate may not be reproduced other than in full except with the prior written approval of the head of Calibration Laboratory Department.

DIGITAL CALIBRATION CO.,LTD.
103, Soi Ekachai 132, Ekachai Rd.,
Bang Bon Nuea, Bang Bon, Bangkok 10150
Phone : 02-862-5225-7 Email : Calibration@digitalscale.co.th

NSC – TISI – TIS 17025
CALIBRATION 0262

CALIBRATION CERTIFICATE

Certificate No. : 24EB 542
Job No. : CI2412-043

Result of calibration : Before Adjustment

Nominal Value (g)	True Value (g)	Average Balance Reading (g)	Correction Value (g)	Uncertainty (± g)	Coverage Factor (k)
0	0.00000	0.00000	0.00000	0.00025	2.00
20	20.00000	20.00000	0.00000	0.00025	2.00
100	100.00001	100.00000	0.00001	0.00025	2.00
200	200.00010	200.00000	0.00010	0.00035	2.00

Adjustment By : Without Calibration

Result of calibration ; After Adjustment

1. Repeatability (n = 10 number of measurement)

Nominal Value (g)	Standard deviation of reading (g)
200	0.000042

2. Departure of indication

Nominal Value (g)	True Value (g)	Average Balance Reading (g)	Correction Value (g)	Uncertainty (± g)	Coverage Factor (k)
0	0.00000	0.00000	0.00000	0.00025	2.00
20	20.00000	20.00000	0.00000	0.00025	2.00
40	39.99999	40.00000	-0.00001	0.00025	2.00
60	59.99997	60.00000	-0.00003	0.00025	2.00
80	79.99996	80.00000	-0.00004	0.00025	2.00
100	100.00001	100.00000	0.00001	0.00025	2.00
120	120.00001	120.00000	0.00001	0.00030	2.00
140	140.00001	140.00000	0.00001	0.00030	2.00
160	159.99998	160.00000	-0.00002	0.00030	2.00
180	179.99998	180.00006	-0.00007	0.00035	2.00
200	200.00010	200.00000	0.00010	0.00035	2.00

3. Eccentric or off-centre loading (Figure. 1)




Figure. 1

Nominal Value	100 g
Position 1	Off-Centre
Position 2 (g)	0.00000
Position 3 (g)	-0.00010
Position 4 (g)	0.00000
Position 5 (g)	0.00000
Maximum difference between off-centre loading	0.00010 g

Approved by : _____

This result of calibration was found accurate as shown on date and place of calibration only.
The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k, providing a level of confidence of approximately 95%.

-End of report-

SCIMET Co., Ltd.
818/124 Udomsuk Rd., Bangna-Nuea,
Bangna, Bangkok 10260 Thailand
Email:scimet2022@gmail.com, Tel: 02 460 9239
https://www.scimet.co.th

NSC-TISI-TIS 17025
CALIBRATION 0454

Certificate No. C17250174

Calibration Certificate

Equipment:	Oven		
Model:	UNB 500	Job No.:	KSMT2501286
Serial No.(or ID):	C507.1007 (012)	Received Date:	28 March 2025
Manufacturer:	Memmert	Issued Date:	28 March 2025
Ventilation Valve:	Closed	Page:	1 of 3
Shelves(pc.):	1		

Customer

HVE Co., Ltd.

603 Soi Charansanitwong 46, Charansanitwong Road Bang Yi Khan, Bang Phlat, Bangkok 10700

Calibration Place

HVE Co., Ltd. (Laboratory)

603 Soi Charansanitwong 46, Charansanitwong Road Bang Yi Khan, Bang Phlat, Bangkok 10700

Calibration Date

28 March 2025

Environment Condition

Temperature:	22.0 °C	±	0.4 °C
Humidity:	55.1 %RH	±	3.5 %RH

The Method used

In-house method, WI17, based on G-20-1/02-08 (E)

Traceability

This certificate is traceable to the SI Units maintained by National Institute of Metrology (NIMT), Thailand through SCIMET Co.,Ltd. Certificate No. C23250016

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).

These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of SCIMET Co., Ltd.

(Mr. Siwapan Srijan)
Person in charge


SCIMET CO.,LTD.

(Mr. Thalerngkeat Poungngam)
Authorized signatory

Condition of reference standards Instruments:

Instruments	Model	S/N or ID.	Certificate No.	Due Date
Datalogger 1	34972A	MY59003249	C23250016	30-Jan-2026

Condition of Calibration Item : In Condition



Standard Installation Locations

Volume (Calibration Zone)= 6 (Liters)

Inside chamber:	W = 56 (cm)	D = 40 (cm)	H = 48 (cm)
Standard Locations (#1, #2, #3, #4):	w = 20 (cm)	d = 10 (cm)	h = 15 (cm)
Standard Locations (#5, #6, #7, #8):	w = 20 (cm)	d = 10 (cm)	h = 15 (cm)

#9: Geometric center of the chamber

Position of Std	#1	#2	#3	#4	#5	#6	#7	#8	#9
Channel of Logger	101	102	103	104	105	106	107	108	109

Definitions

Indicating Temperature: The average reading of indicating device which forms the integral part of the enclosure.

Measured Temperature: The average reading of standards at any positions or location.

Measured Uniformity: The maximum difference of measured temperatures between of any probes and the measured temperature at the reference location which are observed at same time or at close observation time as possible to determine the temperature pattern or homogeneity with the chamber at steady-state. The reference probe is preferably located in the geometric center of the chamber.

Measured Stability: The one-half of greatest maximum difference of measured temperatures at any one probe.

Overall Variation: The difference of maximum and minimum measured temperatures throughout observation time.

Calibration Results:

Without adjustment

Measurement Temperature at Spread Locations, Indicating of Unit Under Calibration: 105.5 °C

Locations	Measured Temperature (°C)	Correction (°C)	Uncertainty (± °C)
#1	104.47	0.47	0.39
#2	104.43	0.43	0.39
#3	104.38	0.38	0.39
#4	104.34	0.34	0.39
#5	103.17	-0.83	0.39
#6	103.31	-0.69	0.39
#7	103.20	-0.80	0.39
#8	103.25	-0.75	0.39
#9	103.93	-0.07	0.39

Temperature Distribution

Desired (°C)	Setting (°C)	Indicating (°C)	Measured Temperature at Spread Locations (°C)									Uncertainty (± °C)*
			#1	#2	#3	#4	#5	#6	#7	#8	#9	
104.0	105.5	105.5	104.47	104.43	104.38	104.34	103.17	103.31	103.20	103.25	103.93	0.39

Chamber Characterization

Indicating (°C)	Measured Uniformity (°C)	Measured Stability (± °C)	Overall Variation (°C)
105.5	0.83	0.13	1.48

Note: * Maximum uncertainty of the each position

The End of Certificate

บริษัท ซายน์เมท จำกัด (SCIMET CO., LTD.)
818/124 Udomsuk rd., Bangna-Nuea, Bangna, Bangkok 10260 Thailand
Email: scimet2022@gmail.com, Tel: 02 460 9239

SCIMET

Statements of conformity:

This conformity certificate documents the validity of the following statements of conformity based on the measurement results of corresponding calibration certificate:

The correction of indication determined during calibration are under given measurement and environmental conditions and considering the expanded measurement uncertainty (coverage probability 95%) within the specification. The given measurement uncertainty already includes other all effects by according to the standard method, G20-1/02-08(E). Therefore, those parameters have not

Tolerance and Decision rules:

Assessment of the conformity of the measurement device are done based on direct comparison of the relevant measurement results with the tolerances and decision rule are prescribed by the customer.

Decision rule : ☐ Choice A Binary Statement for Simple Acceptance Rule (w = 0), Specific Risk < 50% PFA.

☑ Choice B Non-binary statement with guard band (w = 1 U), Pass or Fail Specific Risk < 2.5% PFA and Condition Pass or Condition Fail Specific Risk < 50% PFA.

☐ Choice C Customer defined, Customers may define arbitrary multiple of r to have applied as guard band (w = r U) .

; PFA: Probability of False Accept

SCIMET CO.,LTD บริษัท ซายน์เมท จำกัด

(Mr. Thalerngkeat Poungngam)
Authorized signatory

Without adjustment

Desired Temperature : 104.0°C **Tolerances : 1.0 °C**

Measurement Temperature at Spread Locations, Indicating of Unit Under Calibration: 105.5 °C


Locations	Measured (°C)	Correction of UUC, (°C)	Guard band (W) (± °C)	Tolerance (± °C)	Conformity
#1	104.47	0.47	0.39	1.0	Pass
#2	104.43	0.43	0.39	1.0	Pass
#3	104.38	0.38	0.39	1.0	Pass
#4	104.34	0.34	0.39	1.0	Pass
#5	103.17	-0.83	0.39	1.0	Condition Pass
#6	103.31	-0.69	0.39	1.0	Condition Pass
#7	103.20	-0.80	0.39	1.0	Condition Pass
#8	103.25	-0.75	0.39	1.0	Condition Pass
#9	103.93	-0.07	0.39	1.0	Pass


Correction of UUC,* = Measured Temperature - Desired Temperature

The validity of the statements of conformity cannot be guaranteed for different places of use, environmental conditions or improper use.

The End of Statements of Conformity

บริษัท ซายน์เมท จำกัด (SCIMET CO., LTD.)
818/124 Udomsuk rd., Bangna-Nuea, Bangna, Bangkok 10260 Thailand
Email: scimet2022@gmail.com, Tel: 02 460 9239


Corr_Distribution & Max_Measurement Uncertainty
Job_No. KSMT2501286
Without adjustment
Correction (°C)
7.0
6.0
5.0
4.0
3.0
2.0
1.0
0.0
-1.0
-2.0
-3.0
-4.0
-5.0
-6.0
-7.0
0.0
20.0
40.0
60.0
80.0
100.0
120.0
UUC (°C)
#1
#2
#3
#4
#5
#6
#7
#8
#9
Uncer (+)
Uncer (-)
Lower Acceptance limit
Upper Acceptance limit
Lower Specification
Upper Specification


Temperature Distribution @ 104.0°C
Job_No. KSMT2501286
Without adjustment
Std(°C)
111.0
110.0
109.0
108.0
107.0
106.0
105.0
104.0
103.0
102.0
101.0
100.0
99.0
98.0
97.0
0
20
40
60
80
100
120
140
Time (Interval= 15 sec)
#1
#2
#3
#4
#5
#6
#7
#8
#9
Lower Acceptance limit
Upper Acceptance limit
Lower Specification
Upper Specification

SCIMET Co., Ltd.
818/124 Udomsuk Rd., Bangna-Nuea,
Bangna, Bangkok 10260 Thailand
Email:scimet2022@gmail.com, Tel: 02 460 9239
https://www.scimet.co.th

NSC-TISI-TIS 17025
CALIBRATION 0454

Certificate No. C17250320

Calibration Certificate

Equipment:	Oven
Model:	UNB500
Serial No.(or ID):	C507.1007 (012)
Manufacturer:	Memmert
Ventilation Valve:	Closed
Shelves(pc.):	1

Job No.:	KSMT2502204
Received Date:	09 June 2025
Issued Date:	10 June 2025
Page:	1 of 3

Customer

HVE Co., Ltd.
603 Soi Charansanitwong 46, Charansanitwong Road Bang Yi Khan, Bang Phlat, Bangkok 10700

Calibration Place

HVE Co., Ltd. (Laboratory)
603 Soi Charansanitwong 46, Charansanitwong Road Bang Yi Khan, Bang Phlat, Bangkok 10700

Calibration Date

09 June 2025

Environment Condition

Temperature:	24.3 °C	±	0.9 °C
Humidity:	60.6 %RH	±	4.5 %RH

The Method used

In-house method, WI17, based on G-20-1/02-08 (E)

Traceability

This certificate is traceable to the SI Units maintained by National Institute of Metrology (NIMT), Thailand through SCIMET Co.,Ltd. Certificate No. C23250016

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).

These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of SCIMET Co., Ltd.

(Mr. Siwapan Srijan)
Person in charge


SCIMET CO.,LTD.

(Mr. Thaleerngkeat Poungngam)
Authorized signatory

Condition of reference standards instruments:

Instruments	Model	S/N or ID.	Certificate No.	Due Date
Datalogger 1	34972A	MY59003249	C23250016	30-Jan-2026

Condition of Calibration Item : In Condition



Standard Installation Locations

Volume (Calibration Zone)= 13 (Liters)

Inside chamber:	W = 56 (cm)	D = 40 (cm)	H = 48 (cm)
Standard Locations (#1, #2, #3, #4):	w = 10 (cm)	d = 10 (cm)	h = 15 (cm)
Standard Locations (#5, #6, #7, #8):	w = 10 (cm)	d = 10 (cm)	h = 15 (cm)

#9: Geometric center of the chamber

Position of Std	#1	#2	#3	#4	#5	#6	#7	#8	#9
Channel of Logger	101	102	103	104	105	106	107	108	109

Definitions

Indicating Temperature: The average reading of indicating device which forms the integral part of the enclosure.

Measured Temperature: The average reading of standards at any positions or location.

Measured Uniformity: The maximum difference of measured temperatures between of any probes and the measured temperature at the reference location which are observed at same time or at close observation time as possible to determine the temperature pattern or homogeneity with the chamber at steady-state. The reference probe is preferably located in the geometric center of the chamber.

Measured Stability: The one-half of greatest maximum difference of measured temperatures at any one probe.

Overall Variation: The difference of maximum and minimum measured temperatures throughout observation time.

Calibration Results:

Without adjustment

Measurement Temperature at Spread Locations, Indicating of Unit Under Calibration: 182.5 °C

Locations	Measured Temperature (°C)	Correction (°C)	Uncertainty (± °C)
#1	181.64	1.64	0.87
#2	181.67	1.67	0.87
#3	180.83	0.83	0.87
#4	180.69	0.69	0.87
#5	175.35	-4.65	0.88
#6	175.72	-4.28	0.97
#7	176.01	-3.99	0.89
#8	175.84	-4.16	0.90
#9	178.23	-1.77	0.87

Temperature Distribution

Desired (°C)	Setting (°C)	Indicating (°C)	Measured Temperature at Spread Locations (°C)									Uncertainty (± °C)*
			#1	#2	#3	#4	#5	#6	#7	#8	#9	
180.0	182.5	182.5	181.64	181.67	180.83	180.69	175.35	175.72	176.01	175.84	178.23	0.97

Chamber Characterization

Indicating (°C)	Measured Uniformity (°C)	Measured Stability (± °C)	Overall Variation (°C)
182.5	3.54	0.38	6.54

Note: * Maximum uncertainty of the each position

The End of Certificate

SCIMET

Statements of conformity:

This conformity certificate documents the validity of the following statements of conformity based on the measurement results of corresponding calibration certificate:

The correction of indication determined during calibration are under given measurement and environmental conditions and considering the expanded measurement uncertainty (coverage probability 95%) within the specification. The given measurement uncertainty already includes other all effects by according to the standard method, G20-1/02-08(E). Therefore, those parameters have not

Tolerance and Decision rules:

Assessment of the conformity of the measurement device are done based on direct comparison of the relevant measurement results with the tolerances and decision rule are prescribed by the customer.

Decision rule : ☑ Choice A Binary Statement for Simple Acceptance Rule (w = 0), Specific Risk < 50% PFA.

☐ Choice B Non-binary statement with guard band (w = 1 U), Pass or Fail Specific Risk < 2.5% PFA and Condition Pass or Condition Fail Specific Risk < 50% PFA.

☐ Choice C Customer defined, Customers may define arbitrary multiple of r to have applied as guard band (w = r U) .

; PFA: Probability of False Accept

SCIMET CO.,LTD. บริษัท ซายน์เมท จำกัด

(Mr. Thaleringkeat Poungngam)
Authorized signatory

Without adjustment

Desired Temperature : 180.0°C **Tolerances : 2.0 °C**

Measurement Temperature at Spread Locations, Indicating of Unit Under Calibration: 182.5 °C


Locations	Measured (°C)	Correction of UUC. (°C)	Guard band (W) (± °C)	Tolerance (± °C)	Conformity
#1	181.64	1.64	0	2.0	Pass
#2	181.67	1.67	0	2.0	Pass
#3	180.83	0.83	0	2.0	Pass
#4	180.69	0.69	0	2.0	Pass
#5	175.35	-4.65	0	2.0	Fail
#6	175.72	-4.28	0	2.0	Fail
#7	176.01	-3.99	0	2.0	Fail
#8	175.84	-4.16	0	2.0	Fail
#9	178.23	-1.77	0	2.0	Pass


Correction of UUC.* = Measured Temperature - Desired Temperature

The validity of the statements of conformity cannot be guaranteed for different places of use, environmental conditions or improper use.

The End of Statements of Conformity

บริษัท ซายน์เมท จำกัด (SCIMET CO., LTD.)
818/124 Udomsuk rd., Bangna-Nuea, Bangna, Bangkok 10260 Thailand
Email: scimet2022@gmail.com, Tel: 02 460 9239


Corr_Distribution & Max_Measurement Uncertainty
Job_No. KSMT2502204
Without adjustment
Correction (°C)
12.0
10.0
8.0
6.0
4.0
2.0
0.0
-2.0
-4.0
-6.0
-8.0
-10.0
-12.0
0.0
20.0
40.0
60.0
80.0
100.0
120.0
140.0
160.0
180.0
200.0
UUC (°C)
#1
#2
#3
#4
#5
#6
#7
#8
#9
Uncer (+)
Uncer (-)
Lower Acceptance limit
Upper Acceptance limit
Lower Specification
Upper Specification


Temperature Distribution @ 180.0°C
Job_No. KSMT2502204
Without adjustment
Std(°C)
192.0
190.0
188.0
186.0
184.0
182.0
180.0
178.0
176.0
174.0
172.0
170.0
168.0
Time (Interval= 15 sec)
0
20
40
60
80
100
120
140
#1
#2
#3
#4
#5
#6
#7
#8
#9
Lower Acceptance limit
Upper Acceptance limit
Lower Specification
Upper Specification

CAL

Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

ilac-MRA

NSC-TISI-TIS17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 68-400210-7 **Page : 1 of 2**

Submitted by : HVE Co., Ltd.
603 Soi Jarunsanitwong 46, Jarunsanitwong Road, Bangyeekun, Bangplad, Bangkok 10700

Equipment : Temperature controlled enclosure (Incubator)
Manufacturer : Lovibond Model : TC 445 S
Range : N/A °C Resolution : 0.1 °C
Serial No. : 0515/001080 ID No. : 112

Environment : On site calibration was carried out at the Laboratory, HVE Co., Ltd.
Ambient Temperature : (22.5 to 23.0) °C
Relative Humidity : (60 to 65) %
Line Voltage : (229.0 to 230.0) V

Date of Received : 17 April 2025
Date of Calibration : 17 April 2025
Date of Issue : 25 April 2025
Calibrated by : Pernpon Chanpu
Calibration Method : CAL-M4004, TLAS G-20
The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units
Standard Digital Thermometer with RTD Probe

ID No.	Cert. No.	Due Date	Traceability
400029 & 400043	67-400585-1	26 Apr 2025	National Institute of Metrology Thailand (NIMT)

Approved by :

(Pernpon Chanpu)
Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.

CAL

Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

Certificate of Calibration


Certificate No. : 68-400210-7

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber
W = 0.65 m
D = 0.60 m
H = 1.50 m
Capacity = 0.59 m^3

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (±°C)
			1	2	3	4	5	6	7	8	9	
20.0	20.7	20.7	20.15	20.03	20.13	20.08	20.02	20.09	20.04	20.10	20.01	0.42

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Uniformity (°C)	Measured Stability (°C)	Overall Variation (°C)
20.0	20.7	20.7	0.26	0.18	0.41

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

-oOo-

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

ilac-MRA
NSC-TISI-TI817025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 68-400210-6 **Page : 1 of 2**

Submitted by : HVE Co., Ltd.
603 Soi Jarunsanitwong 46, Jarunsanitwong Road, Bangyeekun, Bangplad, Bangkok 10700

Equipment : Temperature controlled enclosure (Incubator)
Manufacturer : Lovibond Model : ET636-6
Range : N/A °C Resolution : 0.1 °C
Serial No. : 6686 0707 ID No. : 011

Environment : On site calibration was carried out at the Laboratory, HVE Co., Ltd.
Ambient Temperature : (22.5 to 23.0) °C
Relative Humidity : (60 to 65) %
Line Voltage : (229.0 to 230.0) V

Date of Received : 17 April 2025
Date of Calibration : 17 April 2025
Date of Issue : 25 April 2025
Calibrated by : Permpon Chanpu
Calibration Method : CAL-M4004, TLAS G-20
The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units
Standard Digital Thermometer with RTD Probe

ID No.	Cert. No.	Due Date	Traceability
400029 & 400048	68-400063-1	01 Aug 2025	National Institute of Metrology Thailand (NIMT)

Approved by :
(Permpon Chanpu)
Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

Certificate of Calibration


Certificate No. : 68-400210-6

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber
W = 0.53 m
D = 0.43 m
H = 1.40 m
Capacity = 0.32 m³

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (±°C)
			1	2	3	4	5	6	7	8	9	
20.0	21.0	21.0	20.23	20.28	20.26	20.19	20.12	20.08	19.89	19.81	19.96	0.41

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Uniformity (°C)	Measured Stability (°C)	Overall Variation (°C)
20.0	21.0	21.0	0.34	0.16	0.77

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

-oOo-

CAL

Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail ; calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

ilac-MRA

NSC-TISI-TIS17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 68-420036-2 **Page : 1 of 2**

Submitted by : HVE Co., Ltd.
603 Soi Jarunsanitwong 46 Jarunsanitwong Road, Bangyeekun, Bangplad, Bangkok 10700

Equipment : pH Meter with electrode
pH meter
Manufacturer : Hanna Model : HI 2211
Range : N/A pH Resolution : 0.01 pH
Serial No. : 08376721 ID No. : N/A
Electrode
Model : HI 1131 Serial No. : 084809EN

Environment : On site calibration was carried out at the Laboratory, HVE Co., Ltd.
Ambient Temperature : (24.5 to 25.0)° C
Relative Humidity : (45 to 50) %

Date of Received : 18 April 2025
Date of Calibration : 18 April 2025
Date of Issue : 24 April 2025
Calibrated by : Permpon Champu

Calibration Method : In-house method CAL-M4201 direct measurement by using standard voltage calibrator and using certified reference material (CRM)

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Multiproduct Calibrator

ID No.	Cert. No.	Due Date	Traceability
400005	SG-E-00307/66	23 Aug 2025	National Institute of Metrology Thailand (NIMT)

2. Certified Reference Material (CRM)

pH	Cert. No.	Lot No.	Exp. Date	Traceability
4.008	61314276	1081108	28 Feb 2027	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025
6.965	61318175	1081110	28 Feb 2026	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025
10.010	61325043	1081109	28 Feb 2026	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025

Approved by :
(Permpon Champu)
Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.

CAL
Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibmtech_cal@yahoo.com, calibratech_cal@hotmail.com

Certificate of Calibration

Certificate No. : 68-420036-2

Result of Calibration :

UUC Condition As-Received : Good

Function : Electrical measurement

pH meter

Performing standard curve by Multiproduct Calibrator at pH (4,7) and (7,10)

Adjustment Curve at nominal pH	Applied Voltage (mV)	Nominal Value (pH)	UUC Reading (pH)	UUC Reading (mV)	Correction (mV)	Uncertainty (± mV)
4, 7	177.4800	4	4.00	177.3	0.2	0.12
	0.0000	7	7.00	-0.1	0.1	0.086
7,10	0.0000	7	7.00	-0.1	0.1	0.086
	-177.4800	10	10.00	-177.4	-0.1	0.12

Function : pH meter with electrode

Performing a three - buffer standard curve using buffer nominal pH (4,7) and (7,10)

Adjustment Curve at nominal pH	Standard Buffer (pH)	UUC Reading (pH)	Correction (pH)	Uncertainty (± pH)
4, 7	4.008	4.01	0.00	0.0097
	6.965	7.01	-0.04	0.011
7, 10	6.965	7.01	-0.04	0.011
	10.010	10.01	0.00	0.014

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurment was based on a standard uncertainty multiplied by a coverage factor k = 2 , providing a level of confidence of approximately 95%

- o0o -

CAL

Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail ; calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

NSC-TISI-TIS17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 68-400231-2 **Page : 1 of 2**

Submitted by : HVE Co., Ltd.
603 Soi Jarunsanitwong 46 Jarunsanitwong Road, Bangyeekun, Bangplad, Bangkok 10700

Equipment : Temperature Indicator with Thermistor Probe

Temperature Indicator

Manufacturer :	Hanna	Model :	HI 2211
Range :	N/A °C	Resolution :	0.1 °C
Serial No. :	08376721	ID No. :	N/A

Thermistor probe

Model :	N/A	Sheath Material :	Stainless
Diameter :	3 mm.	Length :	100 mm.
Serial No. :	TH036368	ID No. :	N/A

Environment : On site calibration was carried out at the Laboratory, HVE Co., Ltd.

Ambient Temperature :	(24.5 to 25.0)	°C
Relative Humidity :	(45 to 50)	%
Line Voltage :	(229.0 to 230.0)	VAC

Date of Received : 18 April 2025
Date of Calibration : 18 April 2025
Date of Issue : 24 April 2025
Calibrated by : Pernpon Chanpu

Calibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4003 by compared with PRT in the liquid bath at the constant controlled temperature.
The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Platinum Resistance Thermometer (PRT)

ID No.	Cert. No.	Due Date	Traceability
400002	TT-0095-24	01 Jul 2026	National Institute of Metrology Thailand (NIMT)

2. Standard Digital Thermometer

ID No.	Cert. No.	Due Date	Traceability
400033	24E633	21 Feb 2026	National Institute of Metrology Thailand (NIMT)

Approved by :

(Pernpon Chanpu)
Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.

CAL

Calibratech Co.,Ltd.
7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120
Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech_cal@yahoo.com, calibratech_cal@hotmail.com

Certificate of Calibration

Certificate No. : 68-400231-2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

Immersion Depth (mm.)	Standard Reading (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (±°C)
130	25.002	24.8	0.2	0.19

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2 , providing a level of confidence of approximately 95%

-oOo-

TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3 : EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000 FAX. 0-2719-9484

Certificate of Testing

Cert.No.: 25TW66

Equipment :	DO Meter
Manufacturer :	Digicon
Model :	DO-552SD
Serial No. :	AG.35318
ID No. :	-
Received Date :	01 April 2025
Test Date :	02 April 2025
Reference :	2504-0013DN-1
Submitted by :	HVE Co.,Ltd 603 Soi Jarransanitwong 46, Jarransanitwong Road, Bang Yi Khan, Bang Phlat Bangkok 10700
Laboratory Condition :	Temperature (25 ± 5) °C Humidity (50 ± 20) %
Test Procedure :	In - house method : CP-CH9 by Comparison Technique with Azide Modification Method
Tested by :	Walalak Sirithean
Approved by :	Saithip Approved Signatory

( ) Chakrit Waewwanjua
( ) Ponpan Palpim
(✓) Saithip Meangmal

Issue Date : 3 April 2025

Condition of this result of calibration

1. Reference Standard Instruments :
   This certification is traceable to the International System of Unit through the reference standards laboratory of Industrial Calibration Center, Technology Promotion Association (Thailand-Japan).

Instruments	Serial No.	ID No.	Certificate No.	Due Date
1. Burette	-	130BU10	25CG1126	18 Mar 2027
2. Balance	14233821	110RC001	24MM131	04 July 2025

2. Standard Material :-

Material	Manufacturer	Lot.No.	Assay
Sodium Thiosulfate 5-Hydrate AR	KEMAUS	2203162447	99.6%

Result : **Dissolved Oxygen Meter Adjustment With Air 100 %**
Dissolved Oxygen Probe No.: 07-07

Titration Method (Azide Modification Method) (mg/L)	DO Meter Reading (mg/L)	Standard Deviation (mg/L)
8.20	8.1	0.045

This report was certified only for the instrument we tested.It is allowable to use for study, intend to use for advertising and referral purpose is prohibited.This report may not be reproduced other in full, without written approval of the laboratory

-o0o-

BECTHAI BANGKOK EQUIPMENT & CHEMICAL CO., LTD.
CALIBRATION LABORATORY
99/9 Moo 2, Maha Sawat, Phutthamonthon, Nakhon Pathom 73170, Thailand Tel +66 3424 5299 Fax: +66 3424 5250
E-mail: bkk@becthai.com Website: www.becthai.com

NSC-TISI-TIS 17025
CALIBRATION 0131

Certificate No. : CAL-24-567

CERTIFICATE OF CALIBRATION

Equipment	:	Spectrophotometer
Manufacturer	:	Thermo Scientific
Model	:	Genesys 10S UV-VIS
Serial No.	:	2L9Q310003
ID No.	:	071
Customer	:	HVE CO., Ltd.
	:	603 Soi Jarunsanitwong 46, Jarunsanitwong Road,
	:	Bangyeekun, Bangplad, Bangkok 10700
Location	:	แผนกน้ำบริโภค
Date of Receipt	:	18 November 2024
Date of Calibration	:	18 November 2024
Date of Issue	:	19 November 2024
Ambient Temperature	:	(25±10) °C
Relative Humidity	:	(60±20) %
Condition As-Received	:	Used Item

Calibrated by

Mr.Somphop Duangnguan

Calibration Engineer

Approved by

.....

(Ms. Jintana Sangthaljaroenlap)

Calibration Manager

The reported expended uncertainty of measurement was based on a combined standard uncertainty multiplied by a coverage factor *k* providing a level of confidence of appoximately 95%.

This certificate may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Indicated values are valid for the state of the Spectrophotometer at the time of calibration only.

BECTHAI BANGKOK EQUIPMENT & CHEMICAL CO., LTD.
CALIBRATION LABORATORY
99/9 Moo 2, Maha Sawat, Phuthamonthon, Nakhon Pathom 73170, Thailand, Tel +66 3424 5299 Fax: +66 3424 5250
E-mail: bkk@becthai.com Website: www.becthai.com

NSC-TISI-TIS 17025
CALIBRATION 0131

CALIBRATION REPORT

Conditions of this result of calibration

1. Reference Standard Material :

Material	Model	Serial No.	Cert.No.	Due date
Holmium Glass Filter	RM-HG	24563	109211	13 February 2025
Didymium Glass Filter	RM-DG	24562	109212	13 February 2025
Neutral Density Filter	RM-1N2N3N	24568	109249	14 February 2025
Potassium Dichromate Solution	RM-06	24567	109222	13 February 2025

2. Traceability : This certification is traceable to the International System of Unit maintained at;

The Starna Scientific Ltd. Accredited Calibration Laboratory No. 0659.

3. Method of calibration :

The calibration procedure was carried out according to ASTM E275-08 (2022) and ASTM E925-09 (2014).

4. Result of calibration ;

(√) without adjustment ( ) after adjustment

5. Equipment Specifications;

Spectral Bandwidth :	1.8	nm
Data Interval :	0.1	nm
Scan Speed :	Slow	nm/min

20/02/24

BECTHAI BANGKOK EQUIPMENT & CHEMICAL CO., LTD.
CALIBRATION LABORATORY
99/9 Moo 2, Maha Sawat, Phutthamonthon, Nakhon Pathom 73170, Thailand. Tel: +66 3424 5299 Fax: +66 3424 5250
E-mail: bkk@becthai.com Website: www.becthai.com

NSC-TISI-TIS 17025
CALIBRATION 0131

Certificate No. : CAL-24-567

CALIBRATION REPORT

Wavelength Calibration

Certified Values of Reference Material	Nominal Value (nm)	UUC*Reading (nm)	Error (nm)	Uncertainty of Measurement (± nm)	k Factor
361.00	361.00	360.7	-0.27	0.13	2.00
536.66	536.66	536.6	-0.09	0.13	2.00
879.27	879.27	879.8	0.51	0.13	2.00

Photometric Calibration for Visible

Wavelength (nm)	Certified Values of Reference Material (A)	UUC* Reading (A)	Error (A)	Uncertainty of Measurement (± A)	k Factor
420.0	Zero	0.000	0.0000	0.0028	2.00
	0.5835	0.588	0.0046	0.0045	2.00
	0.725	0.726	0.0010	0.0045	2.00
	1.0367	1.038	0.0013	0.0045	2.00
440.0	Zero	0.000	0.0000	0.0028	2.00
	0.5662	0.567	0.0008	0.0045	2.00
	0.7106	0.709	-0.0016	0.0045	2.00
	1.0159	1.014	-0.0019	0.0045	2.00
465.0	Zero	0.000	0.0000	0.0028	2.00
	0.5257	0.529	0.0033	0.0045	2.00
	0.6682	0.669	0.0008	0.0045	2.00
	0.9547	0.955	0.0003	0.0045	2.00
546.1	Zero	0.000	0.0000	0.0028	2.00
	0.5226	0.524	0.0014	0.0045	2.00
	0.6939	0.693	-0.0009	0.0045	2.00
	0.9919	0.991	-0.0009	0.0045	2.00
590.0	Zero	0.000	0.0000	0.0028	2.00
	0.5567	0.558	0.0013	0.0045	2.00
	0.7502	0.749	-0.0012	0.0045	2.00
	1.0732	1.071	-0.0022	0.0045	2.00
635.0	Zero	0.000	0.0000	0.0028	2.00
	0.5643	0.565	0.0007	0.0045	2.00
	0.7299	0.729	-0.0009	0.0045	2.00
	1.0437	1.043	-0.0007	0.0045	2.00

Remark ; Each individual filter is measured against the empty filter holder (blank) used to zero the Spectrophotometer.

Note:

UUC* : Unit Under Calibration

20/02/24

BECTHAI BANGKOK EQUIPMENT & CHEMICAL CO., LTD.
CALIBRATION LABORATORY
99/9 Moo 2, Maha Sawat, Phutthamonthon, Nakhon Pathom 73170 Thailand. Tel +66 3424 5299 Fax: +66 3424 5250
E-mail: bkk@becthai.com Website: www.becthai.com

NSC-TISI-TIS 17025
CALIBRATION 0131

Certificate No. : CAL-24-567

CALIBRATION REPORT

Photometric Calibration for UV

Wavelength (nm)	Certified Values of Reference Material (A)	UUC* Reading (A)	Error (A)	Uncertainty of Measurement (± A)	*k* Factor
235.0	Zero	0.000	0.0000	0.0050	2.00
	0.7385	0.738	-0.0005	0.0081	2.00
257.0	Zero	0.000	0.0000	0.0050	2.00
	0.8556	0.851	-0.0046	0.0081	2.00
313.0	Zero	0.000	0.0000	0.0050	2.00
	0.2882	0.286	-0.0022	0.0081	2.00
350.0	Zero	0.000	0.0000	0.0050	2.00
	0.6346	0.632	-0.0026	0.0081	2.00

Remark : The Potassium Dichromate Filled cells are measured against a Perchloric acid blank.

Note:

UUC* : Unit Under Calibration

- End of Report -



Ref No. : 0303/16972

CERTIFICATE OF TESTING LABORATORY ACCREDITATION

This is to certify that

Laboratory of HVE Co., Ltd.
603 Soi Jarunsanitwong 46, Jarunsanitwong Road,
Bangyeekhan, Bang Phlat, Bangkok 10700

has successfully undergone assessment according to ISO/IEC 17025 : 2017
and under the Bureau of Laboratory Accreditation, Department of Science Service
for the requirements, regulations and criteria for the competence of testing laboratories

LABORATORY ACCREDITATION
Accreditation Number TESTING - 0090
BLA-D88

The scope of accreditation is as annexed hereto

Issue date : *26th November 2024*
Expired date : *25th November 2028*
Signature : Chantarat Vorasapavit

(Mrs. Chantarat Vorasapavit)
Director of Bureau of Laboratory Accreditation

Bureau of Laboratory Accreditation, Department of Science Service,
Ministry of Higher Education, Science, Research and Innovation

Certificate of Registration

This certificate has been awarded to

HVE CO., LTD.

603 Soi Jarunsanitwong 46, Jarunsanitwong Rd.,
Bangyeekan, Bangplud, Bangkok, 10700, Thailand

In recognition of the organization's Quality Management System which complies with

ISO 9001:2015

The scope of activities covered by this certificate is defined below

Water Monitoring Laboratory Service

Certificate Number 32670/A/0001/UK/En

Date of Issue of Certification Cycle	Issue Number	Certificate Expiry Date	Certification Cycle
22 October 2023	8	21 October 2026	6
Revision Date	**Revision Number**	**Original Certificate Issue Date**	**Scheme Number**
09 October 2023	0	22 October 2008	n/a

For detailed explanation for the data fields above, refer to http://www.urs-holdings.com/logos-and-regulations

Issued by

Mukesh Singhal - On behalf of the Schemes Manager

Certificate of Registration

This certificate has been awarded to

HVE CO., LTD.

603 Soi Jarunsanitwong 46, Jarunsanitwong Rd.,
Bangyeekan, Bangplud, Bangkok, 10700, Thailand

In recognition of the organization's Environmental Management System which complies with

ISO 14001:2015

The scope of activities covered by this certificate is defined below

Water Monitoring Laboratory Service

Certificate Number 32670/B/0001/UK/En

Date of Issue of Certification Cycle	Issue Number	Certificate Expiry Date	Certification Cycle
23 October 2023	**6**	**22 October 2026**	**6**
Revision Date	**Revision Number**	**Original Certificate Issue Date**	**Scheme Number**
10 October 2023	**0**	**28 October 2008**	**n/a**

For detailed explanation for the data fields above, refer to http://www.urs-holdings.com/logos-and-regulations

Issued by

Mukesh Singhal - On behalf of the Schemes Manager

If there is any doubt as to the authenticity of this certificate, please do not hesitate to contact the Head Office of the Group on info@urs-certification.com