

ภาคผนวกที่ 4

เอกสารสอบเทียบความถูกต้องของเครื่องมือตรวจวัดคุณภาพสิ่งแวดล้อม

**ตารางสรุปรายการเอกสารการสอบเทียบความถูกต้องของเครื่องมือเก็บตัวอย่าง
และเครื่องมือตรวจวิเคราะห์คุณภาพสิ่งแวดล้อม**

รายการตรวจวัด	เครื่องมือเก็บตัวอย่าง	เครื่องมือตรวจวิเคราะห์
	ชื่อเครื่องมือ	ชื่อเครื่องมือ
1. คุณภาพอากาศจากปล่อง Total Suspended Particulate	Console No. B01 Pitot Tube No. B35	Digital Balance
Oxides of Nitrogen (NO _x)	Vacuum Gauge	Spectrophotometer
Carbon Monoxide	Personal Pump SKC No. B05, B74 Rotameter No. H-B09	CO Analyzer No. B01
Aluminium	Console No. B01 Pitot Tube No. B35	ICP
VOCs	Personal Pump SKC No. B05 Rotameter No. L-B09	GC/MS
2. คุณภาพอากาศในบรรยากาศ Total Suspended Particulate	High Volume Air Sampler Rec No. B44, Blow No. B44	Digital Balance
PM ₁₀	High Volume PM ₁₀ Air Sampler Rec No. B06, Blow No. B06	Digital Balance
Nitrogen Dioxide	NO ₂ Analyzer No. B20	NO ₂ Analyzer No. B21
Carbon Monoxide	CO Analyzer No. B04	CO Analyzer No. B05
Aluminium	High Volume Air Sampler Rec No. B26, Blow No. B26	ICP
3. ระดับเสียงบริเวณแนวรั้วโรงงาน L _{eq} 24 hr, L _{max} และ L ₉₀	Acoustic Calibrator Sound Level Meter No. ACO-B22, B30, B33, R04, R11, R19, R50, R52	-
4. คุณภาพน้ำ pH	-	pH Meter
Total Dissolved Solids	-	Digital Balance
Total Suspended Solids	-	Digital Balance
BOD ₅	-	BOD Analyzer
Grease & Oil	-	Digital Balance
Lead	-	ICP
Aluminum	-	ICP
5. คุณภาพอากาศในสถานประกอบการ Total Dust	Personal Pump SKC No. B64, B76, B72 Rotameter No. H-B07	Digital Balance
Respirable Dust	Personal Pump SKC No. B69, B83, B93, B45 Rotameter No. H-B07	Digital Balance
Aluminium	Personal Pump SKC No. B03, B73, B79 Rotameter No. H-B07	ICP
Hydrogen Fluoride	Personal Pump SKC No. B62, B32, B89 Rotameter No. L-B07	Ion Chromatography
Hydrogen Chloride	Personal Pump SKC No. B71, B96, B91 Rotameter No. L-B07	Ion Chromatography

ตารางสรุปรายการเอกสารการสอบเทียบความถูกต้องของเครื่องมือเก็บตัวอย่าง
และเครื่องมือตรวจวิเคราะห์คุณภาพสิ่งแวดล้อม (ต่อ)

รายการตรวจวัด	เครื่องมือเก็บตัวอย่าง	เครื่องมือตรวจวิเคราะห์
	ชื่อเครื่องมือ	ชื่อเครื่องมือ
6. ระดับเสียงในสถานประกอบการ L _{eq} 8 hr และ L _{max}	Acoustic Calibrator Sound Level Meter ACO No. B41, B43, R51, R52	-
7. ปริมาณเสียงสะสมแบบติดตัวบุคคล TWA	Acoustic Calibrator Sound Level Meter NMD No. B13, B15	-
8. ระดับความร้อนในสถานประกอบการ WBGT	Heat Stress WBGT Meter No. B05, B07, B11, B12	-
9. กลิ่นในสถานประกอบการ Ammonia	Personal Pump SKC No. B91 Rotameteer No. H-B07	Spectrophotometer

คุณภาพอากาศจากปล่อง



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด

S.P.S. CONSULTING SERVICE CO., LTD.

7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900

7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900

Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com., www.spscon.com

Console Calibration Report

Calibration Method

Critical Orifices

Calibration Data

Console Data		Calibration Data		
No.	Serial No.	Date	y	ΔH_{\oplus} (mmH ₂ O)
B01	1563	05/09/2025	1.004	49.67
B02	8002514	01/09/2025	1.002	49.85
B03	1503016	01/09/2025	1.005	49.77
B04	00006659	04/09/2025	0.997	49.93
B05	00007428	02/09/2025	1.003	49.51
R01	1561	01/09/2025	0.999	49.82
R02	8002513	01/09/2025	0.996	49.94
R03	1570	04/09/2025	0.998	50.02
R04	8002519	04/09/2025	1.002	49.89
R05	1503015	02/09/2025	0.996	50.10

Remark : Accept Value of y (test) is $0.97 < y < 1.03$

Accept Value of ΔH_{\oplus} (test) is 46.7 ± 6.4 (mmH₂O)

Calibrated by :

(Mr. Adul Dangklom)

Approved by :

(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด

S.P.S. CONSULTING SERVICE CO., LTD.

7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900

7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900

Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Pitot Tube Calibration Report

Calibration Method

Standard Pitot Tube

Calibration Data

Pitot Tube Data			Calibration Data		
No.	Type of Pitot	Coefficient of Standard Pitot	Date	Avg. of Cp (test)	
				Side A	Side B
B03	S	0.99	04/11/2025	0.84	0.83
B04	S	0.99	03/11/2025	0.85	0.84
B05	S	0.99	03/11/2025	0.84	0.83
B07	S	0.99	04/11/2025	0.84	0.84
B08	S	0.99	05/11/2025	0.84	0.85
B09	S	0.99	04/11/2025	0.84	0.84
B11	S	0.99	03/11/2025	0.84	0.85
B16	S	0.99	03/11/2025	0.84	0.84
B18	S	0.99	03/11/2025	0.85	0.84
B19	S	0.99	03/11/2025	0.84	0.84
B21	S	0.99	04/11/2025	0.85	0.84
B24	S	0.99	04/11/2025	0.84	0.85
B27	S	0.99	04/11/2025	0.84	0.84
B30	S	0.99	05/11/2025	0.84	0.85
B31	S	0.99	05/11/2025	0.84	0.84
B33	S	0.99	05/11/2025	0.84	0.83
B35	S	0.99	03/11/2025	0.83	0.84

Remark : Accept value of Cp (test) is 0.84 ± 0.01

Calibrated by :

(Mr. Adul Dangklom)


Approved by :

(Mr. Peera Detudom)



CALIBRATION LABORATORY Co.,LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230
Tel. 02-578-G353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



CERTIFICATE OF CALIBRATION FOR

NOMENCLATURE	VACUUM GAUGE
MANUFACTURER	HI-LIGHT
MODEL/TYPE	N/A
SERIAL NO.	N/A[64-220066-2]
CLID.NO.	212301420
JOB CONTROL NO.	240720076549

CUSTOMER S.P.S. CONSULTING SERVICE CO., LTD.
7 SOI PHAHOLYOTHIN 24 ROAD., JOMPOL,
CHATUCHAK, BANGKOK 10900

DATE OF RECEIVED : 19 July 2025

DATE OF ISSUED: 24 July 2025

The report or calibration shall not be reproduced except in full without approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Sittipong Pimdee
Calibration Engineer



Approved By : Mongkol Yotsoontorn
Authorized Signatory
24 July 2025



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q24076549

F3-011-05/12-23

page 1 of 3



CALIBRATION LABORATORY Co., LTD.

211Q-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230
Tel. 02-578-Q353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



CALIBRATION AND
DIMENSIONAL MEASUREMENT
ACDM-2814

REPORT OF CALIBRATION

FOR

NOMENCLATURE	VACUUM GAUGE
MANUFACTURER	HI-LIGHT
MODEL/TYPE	N/A
SERIAL NO.	N/A [64-220066-2]
DATE OF CALIBRATION	23 July 2025
DUE DATE OF CALIBRATION	23 July 2026

ENVIRONMENT CONDITIONS

Temperature : $(23 \pm 2) ^\circ\text{C}$

Relative Humidity $(55 \pm 10) \% \text{RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. CLC-CPPP-05 according to DKD-R 6-1 as calibration guidelines.

The calibration was performed by direct measurement with Document Process Calibrator and Pressure Module which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

Document Process Calibrator, Fluke Model 74 1B *S/N.* 8295020 with Pressure Module Model 700PD5 *S/N.* 89404505.

TRACEABILITY :

The measurement s are traceable to International System of Units (SI), through National Institute of Metrology (Thailand).
Certificate No. MP-0040-24.

UNCERTAINTY :

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor of $k=2$. It has been evaluated according to the "Calibration of Pressure Gauges (DKD-R 6-1)" which provides a level of confidence approximately 95%.

Certificate No. Q24076549

F3-011-05/12-23

page 2 of 3

CONDITION OF CALIBRATION ITEM :RECEIVED IN GOOD OPERATIONAL CONDITION

MEASUREMENT RESULTS: (X) without adjustment () adjustment

The DUC was exercised by applying a known pressure from its zero to full scale 1 times. Then 2 series of known gauge pressure were applied. The STD reading were recorded and the means value were reported in the table below.

CALIBRATION DATA

CORRECTION OF PRESSURE

DUC Test point (inHg)	STD Reading (kPa)		Conversion to inHg		Correction (inHg)	
	Up	Down	Up	Down	Up	Down
0	0.00	0.00	0.0	0.0	0.0	0.0
-5	-17.61	-17.95	-5.2	-5.3	+0.2	+0.3
-10	-34.54	-34.54	-10.2	-10.2	+0.2	+0.2
-15	-51.13	-51.47	-15.1	-15.2	+0.1	+0.2
-20	-67.72	-68.06	-20.0	-20.1	+0.0	+0.1
-25	-84.31	-84.31	-24.9	-24.9	+0.1	+0.1
-30	-101.24	-101.24	-29.9	-29.9	+0.1	+0.1

Uncertainty of measurement ± 0.2 inHg

Transmitting fluid : Air.

Technical Note. Conversion factor 1 kPa ; 0.2953003 inHg

Note. The Scope of Accredited ANAB Certificate No. ACDM-2814 Version 012 Page 43 of 67

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q24076549

F3-011-05/ 12-23



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com., www.spscon.com

Personal Pump Calibration Report

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

Environmental Conditions

Temperature : 25 \pm 3 $^{\circ}$ C
Pressure : 1010 \pm 15 mmbar

Personal Pump Data					Calibration Data							
No.	Brand	Model	Serial No.	Date	Flow Rate (mL/min)						Value From Calibration Curve	
					Setting			Actual (Q std.)			y	R ²
					1	2	3	1	2	3		
B01	SKC	224-PCXR4	262101	03/10/2025	1,000	1,500	2,000	998	1,490	1,997	1.000x - 7.191	1.000
B02	SKC	224-PCXR4	626166	03/10/2025	1,000	1,500	2,000	1,007	1,500	2,008	0.999x + 2.537	1.000
B03	SKC	224-PCXR4	612968	03/10/2025	1,000	1,500	2,000	1,003	1,503	2,001	0.997x + 0.810	0.999
B04	SKC	224-PCXR4	602804	02/10/2025	1,000	1,500	2,000	998	1,494	1,993	1.001x - 6.035	1.000
B05	SKC	224-PCXR4	612693	02/10/2025	1,000	1,500	2,000	999	1,495	2,001	0.999x - 2.481	1.000
B06	SKC	224-PCXR4	262188	02/10/2025	1,000	1,500	2,000	997	1,510	2,000	0.998x + 0.064	0.999
B07	SKC	224-PCXR4	626262	01/10/2025	1,000	1,500	2,000	1,004	1,492	2,007	1.002x - 4.778	1.000
B08	SKC	224-PCXR4	626100	02/10/2025	1,000	1,500	2,000	1,005	1,500	2,005	1.004x - 7.223	1.000
B09	SKC	224-PCXR4	626479	01/10/2025	1,000	1,500	2,000	1,001	1,501	1,986	0.996x + 3.462	0.999
B10	SKC	224-PCXR4	091950	01/10/2025	1,000	1,500	2,000	997	1,504	2,000	1.003x - 8.822	1.000
B11	SKC	224-PCXR8	564315	03/10/2025	1,000	1,500	2,000	1,001	1,503	1,995	0.995x + 2.449	1.000
B12	SKC	224-PCXR4	034656	03/10/2025	1,000	1,500	2,000	997	1,506	2,003	1.003x - 9.062	0.999
B13	SKC	224-PCXR4	602073	03/10/2025	1,000	1,500	2,000	1,003	1,497	2,006	1.002x - 5.013	1.000
B14	SKC	224-PCXR4	626313	03/10/2025	1,000	1,500	2,000	998	1,501	1,992	1.005x - 11.702	0.999
B15	SKC	224-PCXR4	626474	03/10/2025	1,000	1,500	2,000	1,001	1,502	2,004	1.006x - 11.694	1.000
B16	SKC	224-PCXR4	626477	03/10/2025	1,000	1,500	2,000	996	1,498	1,992	1.007x - 16.329	0.999
B17	SKC	224-PCXR4	626860	02/10/2025	1,000	1,500	2,000	1,001	1,503	1,998	1.001x - 4.838	1.000
B18	SKC	224-PCXR4	691484	01/10/2025	1,000	1,500	2,000	997	1,514	1,996	0.996x + 5.360	1.000
B19	SKC	224-PCXR4	691599	01/10/2025	1,000	1,500	2,000	998	1,499	2,003	0.998x + 0.399	1.000
B20	SKC	224-PCXR4	691587	01/10/2025	1,000	1,500	2,000	1,001	1,501	1,999	0.995x + 1.520	0.999
B21	SKC	224-PCXR4	691531	03/10/2025	1,000	1,500	2,000	996	1,502	2,001	1.003x - 7.151	1.000
B22	SKC	224-PCXR4	691654	03/10/2025	1,000	1,500	2,000	1,001	1,500	1,998	0.997x - 0.666	1.000
B23	SKC	224-PCXR4	798393	03/10/2025	1,000	1,500	2,000	993	1,507	1,999	1.007x - 17.505	0.999
B24	SKC	224-PCXR4	626363	03/10/2025	1,000	1,500	2,000	994	1,498	1,995	1.000x - 3.941	1.000
B25	SKC	224-PCXR4	798489	01/10/2025	1,000	1,500	2,000	1,003	1,490	2,001	0.997x + 1.703	1.000
B26	SKC	224-PCXR4	798479	01/10/2025	1,000	1,500	2,000	1,001	1,509	1,995	1.002x - 8.057	0.999
B27	SKC	224-PCXR4	691673	01/10/2025	1,000	1,500	2,000	998	1,510	2,002	1.005x - 9.656	1.000
B28	SKC	224-PCXR4	691570	01/10/2025	1,000	1,500	2,000	1,011	1,508	2,009	0.999x + 3.729	0.999
B29	SKC	224-PCXR4	626472	01/10/2025	1,000	1,500	2,000	1,002	1,503	1,998	1.002x - 6.066	1.000
B30	SKC	224-PCXR4	691489	01/10/2025	1,000	1,500	2,000	997	1,506	2,001	1.004x - 8.049	1.000
B31	SKC	224-PCXR4	691509	02/10/2025	1,000	1,500	2,000	995	1,497	1,992	0.998x - 2.293	1.000
B32	SKC	224-PCXR4	091567	01/10/2025	1,000	1,500	2,000	1,002	1,500	2,003	1.008x - 15.778	0.999
B33	SKC	224-PCXR4	091756	02/10/2025	1,000	1,500	2,000	1,003	1,501	1,997	1.003x - 6.509	1.000
B34	SKC	224-PCXR4	612962	01/10/2025	1,000	1,500	2,000	996	1,512	1,996	1.001x - 5.867	0.999
B35	SKC	224-PCXR4	602682	01/10/2025	1,000	1,500	2,000	1,008	1,494	1,999	0.993x + 6.992	1.000
B36	SKC	224-PCXR4	626164	01/10/2025	1,000	1,500	2,000	997	1,502	1,992	0.999x - 3.235	1.000
B37	SKC	224-PCXR4	626256	01/10/2025	1,000	1,500	2,000	1,003	1,490	1,997	0.994x + 5.093	1.000
B38	SKC	224-PCXR4	626167	02/10/2025	1,000	1,500	2,000	998	1,513	1,995	1.000x - 5.277	0.999
B39	SKC	224-PCXR4	034637	03/10/2025	1,000	1,500	2,000	1,007	1,504	2,004	0.996x + 8.240	1.000
B40	SKC	224-PCXR4	798349	03/10/2025	1,000	1,500	2,000	998	1,510	2,002	0.998x + 3.905	1.000

Calibrated by :

(Mr. Abdul Dangklom)

Approved by :

(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด

S.P.S. CONSULTING SERVICE CO., LTD.

7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900

7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900

Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Personal Pump Calibration Report

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

Environmental Conditions

Temperature : 25 ± 3 °C
Pressure : 1010 ± 15 mmbar

Personal Pump Data				Calibration Data								
No.	Brand	Model	Serial No.	Date	Flow Rate (ml/min)						Value From Calibration Curve	
					Setting			Actual (Q std.)				
					1	2	3	1	2	3	y	R ²
B41	SKC	224-PCXR4	612669	01/10/2025	1,000	1,500	2,000	1,010	1,508	2,009	1.000x + 2.612	0.999
B42	SKC	224-PCXR4	626041	02/10/2025	1,000	1,500	2,000	1,004	1,494	1,994	0.997x + 1.344	1.000
B43	SKC	224-PCXR4	034636	01/10/2025	1,000	1,500	2,000	998	1,505	2,002	1.001x - 5.177	1.000
B44	SKC	224-PCXR8	529341	01/10/2025	1,000	1,500	2,000	999	1,496	1,998	0.996x + 0.909	1.000
B45	SKC	224-PCXR8	529594	01/10/2025	1,000	1,500	2,000	996	1,510	1,992	1.005x - 11.543	1.000
B46	SKC	224-PCXR8	566743	01/10/2025	1,000	1,500	2,000	1,003	1,488	1,997	0.994x + 3.717	1.000
B47	SKC	224-PCXR8	566747	01/10/2025	1,000	1,500	2,000	1,004	1,500	1,993	0.996x + 2.230	1.000
B48	SKC	224-PCXR8	566753	01/10/2025	1,000	1,500	2,000	1,002	1,501	1,991	1.000x - 4.116	0.999
B49	SKC	224-PCXR8	566780	01/10/2025	1,000	1,500	2,000	995	1,502	1,990	0.997x - 1.978	1.000
B50	SKC	224-PCXR8	500400	02/10/2025	1,000	1,500	2,000	997	1,503	2,001	1.004x - 10.178	1.000
B51	SKC	224-PCXR8	500363	01/10/2025	1,000	1,500	2,000	1,001	1,502	1,993	0.995x + 2.848	1.000
B52	SKC	224-PCXR8	093186	03/10/2025	1,000	1,500	2,000	996	1,510	1,999	1.005x - 12.252	0.999
B53	SKC	224-PCXR8	707670	03/10/2025	1,000	1,500	2,000	1,002	1,496	2,004	1.003x - 8.791	1.000
B54	SKC	224-PCXR3	509821	03/10/2025	1,000	1,500	2,000	999	1,501	1,995	0.999x - 2.090	1.000
B55	SKC	224-PCXR3	510710	02/10/2025	1,000	1,500	2,000	1,002	1,503	2,006	1.007x - 13.250	0.999
B56	SKC	224-PCXR3	511450	01/10/2025	1,000	1,500	2,000	995	1,505	1,997	1.002x - 7.594	1.000
B57	SKC	224-PCXR3	510798	02/10/2025	1,000	1,500	2,000	998	1,500	1,994	0.999x - 7.163	0.999
B58	SKC	224-PCXR3	509852	03/10/2025	1,000	1,500	2,000	1,002	1,494	1,996	0.993x + 6.485	1.000
B59	SKC	224-PCXR3	509862	01/10/2025	1,000	1,500	2,000	1,006	1,505	1,998	0.996x + 5.117	1.000
B60	SKC	224-PCXR3	512655	02/10/2025	1,000	1,500	2,000	1,004	1,501	2,003	1.010x - 14.223	0.999
B61	SKC	224-PCXR3	503915	03/10/2025	1,000	1,500	2,000	993	1,495	1,994	0.999x - 4.942	1.000
B62	SKC	224-PCXR3	505975	03/10/2025	1,000	1,500	2,000	995	1,500	2,005	1.009x - 16.396	1.000
B63	SKC	224-PCXR3	511432	03/10/2025	1,000	1,500	2,000	996	1,497	1,991	0.998x - 3.171	1.000
B64	SKC	224-PCXR3	508302	03/10/2025	1,000	1,500	2,000	1,008	1,506	1,998	0.992x + 8.667	0.999
B65	SKC	224-PCXR3	508310	03/10/2025	1,000	1,500	2,000	1,006	1,492	2,003	1.000x - 4.355	1.000
B66	SKC	224-PCXR3	509861	03/10/2025	1,000	1,500	2,000	994	1,496	1,994	0.997x - 0.275	1.000
B67	SKC	224-PCXR3	506295	01/10/2025	1,000	1,500	2,000	997	1,505	2,001	1.004x - 10.258	1.000
B68	SKC	224-PCXR3	505872	03/10/2025	1,000	1,500	2,000	998	1,512	1,992	0.999x - 3.554	0.999
B69	SKC	224-PCXR3	508375	01/10/2025	1,000	1,500	2,000	997	1,489	1,996	0.997x - 2.309	1.000
B70	SKC	224-PCXR3	510623	03/10/2025	1,000	1,500	2,000	1,001	1,496	1,991	0.992x + 7.131	1.000
B71	SKC	224-PCXR3	508367	03/10/2025	1,000	1,500	2,000	999	1,498	1,995	0.994x + 6.433	1.000
B72	SKC	224-PCXR3	505977	03/10/2025	1,000	1,500	2,000	996	1,507	1,999	1.003x - 7.490	1.000
B73	SKC	224-PCXR3	512606	03/10/2025	1,000	1,500	2,000	1,004	1,503	2,003	1.001x - 5.285	0.999
B74	SKC	224-PCXR3	505993	01/10/2025	1,000	1,500	2,000	1,006	1,501	1,997	0.993x + 8.232	1.000
B75	SKC	224-PCXR3	509820	02/10/2025	1,000	1,500	2,000	1,005	1,494	1,995	0.997x - 2.979	0.999
B76	SKC	224-PCXR3	509811	02/10/2025	1,000	1,500	2,000	996	1,503	1,999	1.005x - 10.613	1.000
B77	SKC	224-PCXR3	508301	01/10/2025	1,000	1,500	2,000	1,003	1,490	1,998	0.994x + 4.694	1.000
B78	SKC	224-PCXR3	510677	01/10/2025	1,000	1,500	2,000	1,004	1,492	1,997	0.995x + 4.036	1.000
B79	SKC	224-PCXR3	510920	01/10/2025	1,000	1,500	2,000	1,008	1,504	2,006	1.006x - 9.588	0.999

Calibrated by :

(Mr. Adul Dangklom)

Approved by :

(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com., www.spscon.com

Rotameter Calibration Report (For Personal Pump High Flow Adjust)

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

Calibration Data

Calibration Data											
Rotameter Data			Calibration Data								
No.	Brand	Model	Date	Flow Rate (ml/min)						Value From Calibration Curve	
				Flow Rate (Reading)			Actual (Q std.)				
				1	2	3	1	2	3	y	R ²
H-B01	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	502.3	998.1	1996.9	0.999x + 2.995	1.000
H-B02	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	501.2	998.9	1998.1	0.998x + 4.159	1.000
H-B03	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	500.4	999.5	2000.3	1.000x - 1.574	0.999
H-B04	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	499.5	1000.6	1999.5	0.999x + 0.880	1.000
H-B05	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	499.1	1001.3	2001.6	1.001x - 7.095	0.999
H-B06	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	499.4	999.8	1994.1	1.000x + 2.760	1.000
H-B07	Dwyer	VFB-65	02/10/2025	500	1,000	2,000	501.0	999.4	2000.7	0.997x + 1.623	1.000
H-B08	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	499.9	1001.7	2002.8	0.998x - 1.887	1.000
H-B09	Dwyer	VFB-65	02/10/2025	500	1,000	2,000	499.3	999.2	1996.6	0.999x + 1.428	0.999
H-B10	Dwyer	VFB-65	03/10/2025	500	1,000	2,000	498.7	1001.1	1997.8	0.996x + 6.669	1.000

Calibrated by :

(Mr. Abdul Dangklom)

Approved by :

(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com., www.spscon.com

Rotameter Calibration Report (For Personal Pump Low Flow Adjust)

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

Calibration Data

Rotameter Data			Calibration Data								
No.	Brand	Model	Date	Flow Rate (mL/min)						Value From Calibration Curve	
				Flow Rate (Reading)			Actual (Q std.)				
				1	2	3	1	2	3	y	R ²
L-B01	Dwyer	VFA-21	01/10/2025	50	100	200	50.7	99.8	199.6	1.000x + 0.067	1.000
L-B02	Dwyer	VFA-21	01/10/2025	50	100	200	50.2	99.4	198.7	1.001x - 1.181	0.999
L-B03	Dwyer	VFA-21	01/10/2025	50	100	200	50.8	99.2	201.5	1.002x + 0.053	1.000
L-B04	Dwyer	VFA-21	01/10/2025	50	100	200	49.7	101.6	200.8	1.001x + 0.344	1.000
L-B05	Dwyer	VFA-21	01/10/2025	50	100	200	50.5	100.4	201.6	0.998x + 0.225	0.999
L-B06	Dwyer	VFA-21	01/10/2025	50	100	200	50.1	100.5	201.8	1.003x - 0.103	1.000
L-B07	Dwyer	VFA-21	02/10/2025	50	100	200	50.6	100.8	201.3	0.998x + 0.877	1.000
L-B08	Dwyer	VFA-21	01/10/2025	50	100	200	49.8	101.3	198.9	1.000x - 0.165	1.000
L-B09	Dwyer	VFA-21	02/10/2025	50	100	200	50.1	99.6	200.7	1.002x - 0.766	0.999
L-B10	Dwyer	VFA-21	03/10/2025	50	100	200	50.9	100.8	201.2	1.003x + 0.694	1.000

Calibrated by :

(Mr. Adul Dangklom)

Approved by :

(Mr. Peera Detudom)



CERTIFICATE No : 25M2254
REFERENCE No : 76365-1

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : DIGITAL BALANCE
MANUFACTURER : METTLER TOLEDO
MODEL : XS105DU
SERIAL No : 1126422905
ID No : BA05/50
CONDITION AS RECEIVED : USED ITEM
SUBMITTED BY : S.P.S. CONSULTING SERVICE CO., LTD.
7 SOI PHAHOLYOTHIN 24, PHAHOLYOTHIN RD.,
JOMPOL, CHATUCHAK, BANGKOK 10900

CALIBRATED BY : ATSAWIN Y.

CALIBRATION DATE : 07-Mar-25

APPROVED BY : 
PONGSAK J.

ISSUED DATE : 13-Mar-25

RECEIVED DATE : 07-Mar-25

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.





CERTIFICATE No : 25M2254

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : DIGITAL BALANCE MODEL : XS105DU
MANUFACTURER : METTLER TOLEDO S/N : 1126422905
ID No : BA05/50 RECEIVED DATE : 07-Mar-25
AIR PRESSURE : 1009mbar \pm 1mbar CALIBRATION DATE : 07-Mar-25
AMBIENT TEMPERATURE : 24°C \pm 1°C RELATIVE HUMIDITY : 54 %RH \pm 10 % RH

CONDITION OF THIS RESULTS OF CALIBRATION

1. THIS INSTRUMENT WAS CALIBRATED BY ACCORDING TO UKAS LAB 14 EDITION 6:2019 BY USING KNOWN WEIGHT STANDARD WEIGHT. THE BALANCE WAS NOT ADJUSTED BEFORE CALIBRATION. THE BALANCE HAS NO ZERO TRACKING FUNCTION. REPEATABILITY WAS MEASURED BY USING 10 REPEATED MEASUREMENTS. LINEARITY WAS MEASURED COVERING 10 POINTS, EVENLY SPREAD OVER THE RANGE. THE INSTRUMENT WAS SET ZERO BEFORE PERFORMING THE LINEARITY TEST. OFF-CENTER LOADING WAS MEASURED BY USING STANDARD WEIGHTS PLACED ON THE PAN AND MOVED TO VARIOUS POSITIONS ON THE PAN.

2. REFERENCE STANDARD INSTRUMENTS :-

INSTRUMENT	MODEL	SERIAL No	CERTIFICATE No	DUE DATE
1) STANDARD WEIGHT SET	E2	QK-I-151	C02250116	28-Jan-27
2) STANDARD WEIGHT	E2	15843	C02250117	29-Jan-27

3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.

4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.

5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-
- NATIONAL INSTITUTE OF METROLOGY (THAILAND)

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT

1. ZERO SETTING FUNCTION : NORMAL


2. TARE FUNCTION : NORMAL

3. REPEATABILITY OF READING AT 120 g WAS 0.000055 g

4. DEPARTURE FROM NOMINAL VALUE/ LINEARITY

NOMINAL VALUE (g)	BALANCE READING (g)	CORRECTION (g)	UNCERTAINTY (\pm g)
0.00	0.00000	0.00000	0.000065
0.02	0.01999	0.00001	0.000065
0.10	0.10001	-0.00001	0.000066
0.20	0.20001	-0.00001	0.000066
0.50	0.50002	-0.00002	0.000065
1.00	1.00003	-0.00003	0.000066
2.00	2.00001	-0.00001	0.000067
5.00	5.00002	-0.00002	0.000068
10.00	10.00000	0.00000	0.000070
20.00	20.00004	-0.00004	0.000078
50.00	50.00000	0.00000	0.00013
100.00	100.0001	-0.0001	0.00019
120.00	120.0002	-0.0002	0.00022

5. OFF CENTER LOADING ERROR



POINT	READING (g)
1	50.0000
2	50.0000
3	50.0000
4	50.0000
5	50.0000
OFF-CENTER LOADING	0.0000

NOTE: THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA

THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR $k=2$, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.

END OF CALIBRATION REPORT



Cert. No. : SP25026

Pages : 1 of 4


Calibration Certificate

Equipment :	UV-VIS SPECTROPHOTOMETER
Manufacturer :	PERKINELMER
Model :	LAMBDA 25
Serial No.:	501S14123010
ID No.:	SP03/58
Calibration Mode :	WAVELENGTH ACCURACY PHOTOMETRIC ACCURACY STRAY LIGHT
Condition As Found :	GOOD
Customer :	S.P.S CONSULTING SERVICE CO., LTD. 7 SOI PHAHOLYOTHIN 24, PHAHOLYOTHIN ROAD, CHOMPHON SUB-DISTRICT, CHATUCHAK DISTRICT, BANGKOK PROVINCE 10900 THAILAND.
Location :	ORGANIC LABORATORY IV
Ambient Temperature :	(22.9 \pm 5) °C
Relative Humidity :	(53.7 \pm 25) %
Received Date :	22 AUGUST 2025
Calibration Date :	22 AUGUST 2025
Date of Issue :	25 AUGUST 2025

Calibrated by :

Nitinun Srihawan

Approved by :

( Wichok Ekpongpradit)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Cert. No. : SP25026

Job No. : VC68SP0019

Pages : 2 of 4

Calibration Method :

This instrument was calibrated by using on-site calibration procedure In-house method : CP-SP-01

The calibration procedure to direct measurement wavelength accuracy by using wavelength standard solution, Photometric accuracy by using absorbance standard filter and absorbance standard solution

The calibration procedure used was based on ASTM E275-01, ASTM E925-02

Condition of this result of calibration :

1. Certified reference materials

<u>Material</u>	<u>Ref. type</u>	<u>Cell serial No.</u>	<u>Cert. No.</u>	<u>Due Date</u>
Holmium liquid	RM-HL	29706	126461	24/10/2026
Didymium liquid	RM-DL	28912	126462	24/10/2026
Neutral density filter	RM-1N2N3N	13877	126457	24/10/2026
Potassium dichromate solutions	RM-0204060810	14204	126497	25/10/2026
Potassium Iodide solution	-	KI-0701-001	CI-0185-24	14/05/2026

2. This result of calibration was found accurate as shown on date and place of calibration only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 The UK National Physical Laboratory (NPL)

Result of calibration : Wavelength Accuracy

(Without adjustment)

<u>Material</u>	<u>Certified Values of Reference Material (nm)</u>	<u>UUC* Reading (nm)</u>	<u>Error (nm)</u>	<u>Uncertainty ± (nm)</u>	<u>k Factor</u>
RM-HL	278.13	278.21	0.08	0.16	2.00
	361.25	361.39	0.14	0.16	2.00
	467.82	467.71	-0.11	0.16	2.00
	536.56	536.50	-0.06	0.16	2.00
	640.50	640.36	-0.14	0.16	2.00
RM-DL	740.09	739.85	-0.24	0.16	2.00
	864.94	865.12	0.18	0.16	2.00

UUC* = Unit Under Calibration

Cert. No. : SP25026

Job No. : VC68SP0019

Pages : 3 of 4

Result of calibration : Photometric Accuracy

Material	Wavelength (nm)	Filter S/N	Nominal Absorbance (A)	Certified Absorbance (A)	UUC* Reading Absorbance (A)	Error (A)	Uncertainty ± (A)	k Factor
Neutral Density glass filter	440.0	29381	0.5	0.5443	0.5413	-0.0030	0.0043	2.00
		29914	0.7	0.7484	0.7455	-0.0029	0.0054	2.00
		29360	1.0	1.0527	1.0535	0.0008	0.0032	2.00
	465.0	29381	0.5	0.4948	0.4922	-0.0026	0.0041	2.00
		29914	0.7	0.6906	0.6877	-0.0029	0.0050	2.00
		29360	1.0	0.9695	0.9709	0.0014	0.0031	2.00
	546.1	29381	0.5	0.5090	0.5068	-0.0022	0.0036	2.00
		29914	0.7	0.6985	0.6960	-0.0025	0.0041	2.00
		29360	1.0	0.9814	0.9825	0.0011	0.0031	2.00
	590.0	29381	0.5	0.5375	0.5353	-0.0022	0.0034	2.00
		29914	0.7	0.7256	0.7231	-0.0025	0.0037	2.00
		29360	1.0	1.0213	1.0219	0.0006	0.0032	2.00
	635.0	29381	0.5	0.5223	0.5202	-0.0021	0.0033	2.00
		29914	0.7	0.6927	0.6901	-0.0026	0.0036	2.00
		29360	1.0	0.9744	0.9750	0.0006	0.0032	2.00

UUC* = Unit Under Calibration

Cert. No. : SP25026

Job No. : VC68SP0019

Pages : 4 of 4

Result of calibration : Photometric Accuracy

(Without adjustment)

Material	Wavelength (nm)	Solution (mg/l)	Certified Absorbance (A)	UUC* Reading Absorbance (A)	Error (A)	Uncertainty ± (A)	k Factor
Potassium dichromate solutions	235.0	20	0.2415	0.2443	0.0028	0.0101	2.00
		40	0.4866	0.4871	0.0005	0.0115	2.00
		60	0.7415	0.7295	-0.0120	0.0067	2.00
		80	0.9854	0.9844	-0.0010	0.0071	2.00
		100	1.2444	1.2425	-0.0019	0.0073	2.00

UUC* = Unit Under Calibration

Condition of this result of calibration : Spectrophotometer PERKINELMER Model LAMBDA 25 S/N 501S14123010

Resolution of Wavelength Mode 0.1 nm

Resolution of Photometric Mode 0.001 A

Parameter Setting

Measurement Mode Wavelength, Absorbance

Wavelength Scan 190 nm - 1100 nm

Scanning Speed 7.5 nm/min

Band width(Wavelength) 1.0

Band width(Vis) 1.0

Band width(Uv) 1.0

Stray Light** UUC* Reading at 220.0 nm	
Transimission T(%)	Absorbance(A)
0.020	3.7032

**Specific Acceptance :

Transmission ≤ 1.0 T(%), Absorbance ≥ 2.0 A

**Stray light not TISI Accredited

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95%

End of Calibration Certificate



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72 Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Calibration Report					
Non-Dispersive Infrared CO Analyzer					
Date :	01 October 2025	Brand :	API	Model :	300E
No.	CO-B01			Serial No.	782
Calibrator (Dilution System)					
Brand : Teledyne			Model : 700E		
Last Cal. Date : 28 October 2024			Serial No. : 201-S		
Reference Standard Gas					
Standard Gas : Carbon Monoxide (CO)			Cylinder No. : D711839		
Certified Date : 14 March 2024		Expired Date : 14 March 2032		Cylinder Conc. : 4,580 ppm	
Calibrating Condition					
Pressure	1011	mmbar	Temp.	24.6	°C
			% RH	50	
Calibration Setting					
Span	Initial Reading (Before Adj.), PPM			Final Reading (After Adj.), PPM	
Set Point	Expected Concentration	Analyzer Response		%Dif	
Zero	0	0.11		-	
CO Span	40.00	40.04		0.100	
API Model 300E CO Analyzer Check List					
Parameter	Observed Value	Units	Nominal Range		
Range	50	PPM	0-1000 ppm		
Stability	0.10	PPM	< 1 ppm With Zero Air		
CO Measure	4014.1	mV	2500-4800 mV		
CO Reference	3948.4	mV	2500-4800 mV		
Measure/Reference Ratio	1.179	-	1.1-1.3 W/Zero Air		
Sample Pressure	28.7	In-Hg-A	~2" < Ambient Absolute Pressure		
Sample Flow	809	CC/Min	800 ± 10%		
Sample Temperature	48.3	°C	48 ± 4		
Bench Temperature	48.0	°C	48 ± 2		
Wheel Temperature	68.5	°C	68 ± 2		
Box Temperature	30.6	°C	Ambient Temp + 7 ± 10		
Photo-Drive	3036.3	mV	250 mV to 4750 mV		
Slope	1.017	-	1.0 ± 0.3		
Offset	0.2	-	0 ± 0.3		

Calibrated by :

[Redacted Signature]

(Mr. Adul Dangklom)

Approved by :

[Redacted Signature]

(Mr. Peera Detudom)



MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

Customer : <u>S.P.S.Consulting Service Co.,Ltd</u>	Date Tested: <u>December 18, 2025</u>	
	Recommendation Recertification	
Address : <u>7 Soi Phaholyothin 24</u>	Period <u>6</u> Months	
<u>Paholyothin Road</u>	Recertification Due: <u>June 28, 2026</u>	
<u>Jompol Chatuchak, Bangkok 10900</u>	Date Last Certified: <u>July 1, 2025</u>	
User Name: <u>K.Phenpha Viphashawat</u>	Visit Number: <u>2 of 2</u>	
Phone: <u>083-9269252</u>	PerkinElmer Phone: <u>02-719-6420 ext 206</u>	
Fax: <u>02-513-4221</u>	PerkinElmer Fax: <u>02-318-5597</u>	

CONFIGURATION TESTED		ACCESSORIES/COMPONENT NOT INCLUDED
MODEL	SERIAL NUMBER	
<u>OPTIMA 5300DV</u>	<u>077C7042401</u>	
TESTED EQUIPMENT	CALIBRATION NUMBER	EXPIRATION
<u>IPV Methods</u>		
TEST STANDARD USED	PART NUMBER	EXPIRATION DATE
<u>Multielement Standard</u>	<u>N069-1579</u>	<u>November 30, 2026</u>
<u>Wavecal Solution</u>	<u>N058-2152</u>	<u>July 30, 2026</u>
<u>VIS Wavecal solution</u>	<u>N930-2946</u>	<u>August 30, 2026</u>
<u>Instrument Cal. STD4</u>	<u>N930-0221</u>	<u>November 30, 2026</u>
CUSTOMER SUPPLIED	COMMENTS	CUSTOMER INITIALS
<u>2 % HNO3</u>		
<u>10 % HNO3</u>		



MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

SERIAL NUMBER 077C7042401
DATE TESTED December 18, 2025
1. MECHANICAL CHECKS

A. Inspect and clean all fans and filters.

☐ OK

B. Inspect and replace as necessary, all torch components including the RF coil.

☐ OK

C. Inspect all tubing for sign of clacking or leaking.

☐ OK

D. Adjust water and gas pressure regulator settings.

☐ OK

E. Inspect and leak check pneumatics drawers.

☐ OK

F. Clean the exterior of the instrument.

☐ OK

2. OPTICAL CHECKS

A. Inspect and clean all optical components.

☐ OK

B. As required, check and replace all purgefilters.

☐ OK

C. Recheck optical alignment.

☐ OK

3. COOLING SYSTEM CHECKS

A. Perform preventive maintenance on chiller.

☐ OK

B. Flush out the chiller every year.

☐ N/A

4. PERFORMANCE CHECKS

A. Torch View Alignment.

☐ OK

B. Wavelength Calibration.

☐ OK



MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

SERIAL NUMBER : 077C7042401			DATE TESTED : December 18, 2025		
PARAMETER	SPECIFICATION			FINAL VALUE	
Spectral Resolution : UV	As	193.696 nm	≤ 0.007	0.00530	
	Ni	231.604 nm	≤ 0.008	0.00708	
	Ni	341.476 nm	≤ 0.012	0.00776	
Spectral Resolution : VIS	La	408.672 nm	≤ 0.020	0.01614	
	Ba	455.403 nm	≤ 0.025	0.02377	
Precision					
	As	193.656 nm	% RSD < 1.0	0.67	%
	Zn	213.856 nm	% RSD < 1.0	0.62	%
	Mn	257.610 nm	% RSD < 1.0	0.88	%
	La	379.478 nm	% RSD < 1.0	0.63	%
	Ba	455.403 nm	% RSD < 1.0	0.65	%
	Ba	493.408 nm	% RSD < 1.0	0.45	%
Detection Limits : Axial	Tl	190.080 nm	3(sd)	3.21	ppb
	As	193.696 nm	3(sd)	6.06	ppb
	Pb	220.353 nm	3(sd)	0.92	ppb
Detection Limits : Radial	As	193.696 nm	3(sd)	17.35	ppb
	Zn	213.856 nm	3(sd)	1.79	ppb
	Mn	257.610 nm	3(sd)	0.18	ppb
	La	379.478 nm	3(sd)	0.76	ppb
	Ba	455.403 nm	3(sd)	0.11	ppb
	Ba	493.408 nm	3(sd)	0.56	ppb
BEC : Axial (IB X 500)/(IS-IB)	Cd	226.502 nm	≤ 150 ppb	40.52	
BEC : Radial (IB X 1000)/(IS-IB)	Mn	257.610 nm	≤ 45 ppb	42.33	



MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

SERIAL NUMBER 077C7042401**DATE TESTED** December 18, 2025**Remarks :**

Commissioning follow as commissioning performance sheets.

This is to certify that the above tests have been performed and the configuration tested



meets



does not meet

the PerkinElmer Specifications listed on this certificate.

This certificate does not modify PerkinElmer's standard terms and condition of sale,
including warranty terms.

Service Department PerkinElmer Ltd


Authorized Representative:

(Wiphan Promlumda)

Service Engineer

GC Clarus 600/680 Preventive Maintenance (PM)

Company Name:	S.P.S. Consulting Service Co.,Ltd		
Address (Instrument Location):	7 Soi Phaholyothin24 Phaholyothin Road, Jompol, Chatuchak, Bangkok, 10900.		
Serial Number:	680S14042502	Service Tag:	N68APSSFxMP
Customer Name (if applicable):	Ms.Naruecha	PM number:	2 of 2
Service Engineer Name:	Monchai Kitcharoenkeat	Service Order Number:	WO-06815714
Date PM Performed: (DD-MMM-YYYY)	13-Aug-2025	Next PM Due Date: (DD-MMM-YYYY)	13-Feb-2026

Part Number	Release	Publication Date	
TH09370070	C	August 2016	

Scope

The purpose of this PM is to ensure the continued functionality of the Clarus 600 and Clarus 680 GC by inspecting and replacing any worn or damaged parts. This service should only be performed by a trained representative of PerkinElmer. The customer should save their method before the PM begins.

General Instructions:

The customer must provide the engineer operational data to demonstrate recent instrument performance prior to starting the PM. Always check with the customer before making any changes that may affect the customer's analysis or calibration, including a current back-up of system software and/or data files. The completed document should be signed by an authorized PerkinElmer and customer representative and left with the customer. Update the PM sticker and instrument logbook as required.

Copyright Information

This document contains proprietary information that is protected by copyright. All rights are reserved. No part of this publication may be reproduced in any form whatsoever or translated into any language without the prior, written permission of Perkin Elmer, Inc. **Copyright © 2013 PerkinElmer, Inc.**

Trademarks

Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are protected by law. PerkinElmer is a registered trademark of PerkinElmer, Inc. All other trademarks and registered trademarks not owned by PerkinElmer, Inc. or its subsidiaries that are depicted herein are the property of their respective owners. **Except as specifically set forth in its terms and conditions of sale, PerkinElmer makes no Warranty of any kind with regard to this document, including, but not limited to, the implied warranties of merchantability and fitness for a particular purpose.** PerkinElmer shall not be liable for incidental or consequential damages in connection with the furnishing or use of this document.

Component List

Component / Specific Model	Serial #	Software Version	Configuration Notes
Clarus680	680S14042502	Totalchrom6.3.2	PSS, PSS, FID,
Clarus SQ8T	648N4050804	Turbomass 6.4	
AtomX	US14113002	Tekma AtomX	

Parts Lists

Additional Tools Required for PM				
Part Number (if applicable)	Description	Quantity	Serial #	Calibration Due Date (MM/YY)
N/A				
Additional Reagents and Standards Required for PM				
Part Number (if applicable)	Description	Quantity	Batch/Lot #	Expiration Date (MM/YY)
N/A				

Procedure Checklist

Use (✓) to check off those steps in the checklist that have been completed.

1. General:

- ☒ Review the instrument performance with the customer and document any recent problems.

- ☒ Check incoming AC line voltage for proper levels and grounding.

L-N 220 Volt

L-G 220 Volt

N-G 0.32 Volt

**Neutral to ground not more than 0.5 volts peak to peak*

- ☒ Inspect all gas line filters and traps; Replace if necessary with customer supplied spares.

Carrier gas ☒ Helium ☐ Nitrogen ☐ Hydrogen

Moisture level ☒ Good ☐ Need to replace ☐ Other _____

Detector gas ☒ Air Zero ☒ Hydrogen ☐ Nitrogen ☐ Helium

Moisture level ☒ Good ☐ Need to replace ☐ Other _____

- ☒ Inspect the customer log book and make any appropriate PM entries.

- ☒ Leak check all fittings from the gas source to instrument.

Gas leakage ☒ Pass ☐ Fail Comment _____

- ☒ Perform general inspection of system for cleanliness.

- ☒ Inspect for functional and clean electronic cooling and oven vent fans

Electronic cooling fan ☒ Yes ☐ No

Oven cooling fan ☒ Yes ☐ No

2. Electronic :

- ☒ Check oven temperature. Calibrate if necessary.

Oven temperature set point 150 °C ☒ Pass ☐ Fail

- ☐ Check sub-ambient option. (If installed).

Oven temperature set point 5 °C ☐ Pass ☐ Fail

- ☒ Perform routine maintenance on detector/injector. Replace parts as necessary with customer supplied spares.

- ☒ Check flows, including split flows if applicable. Calibrate if necessary.

Carrier flow	Pass
Split flow	Pass
- ☒ Check detector gas flows and adjust if necessary.

Detector flow	Pass
---------------	------
- ☒ Autosampler installed ☒ Yes ☐ No

Check autosampler sensor for wear and replace if necessary.	
Vial sensor	Pass
Door sensor	Pass
Tower sensor	Pass
Plunger sensor	Pass
Elevator sensor	Pass
- ☒ Remove syringe, manually flush. Replace with customer supplied spare if necessary.
- ☒ Check firmware version. Upgrade to current levels if necessary.

Firmware version	<u>6.5</u>
------------------	------------
- ☒ Measure all accessible power supply voltages.

5 Volt	Pass
+15 Volt	Pass
-15 Volt	Pass
24 Volt	Pass
- ☒ Record all detector voltage signal.

Detector Channel A	<u>0.98</u>	mV.
Detector Channel B	<u>NA</u>	mV.

3. Diagnostics Tests:

- ☒ Run instrument diagnostics.

<input checked="" type="checkbox"/> BRAM	Pass
<input checked="" type="checkbox"/> EPROM	Pass
- ☒ Run Autosampler diagnostics.

<input checked="" type="checkbox"/> BRAM	Pass
<input checked="" type="checkbox"/> EPROM	Pass

4. Review:

- ☒ Review with the customer PM work performed.
- ☒ Review with the customer routine maintenance procedures.
- ☒ Discuss recommended customer-supplied materials to have on hand
- ☒ Attach PM sticker.
- ☒ Update Logbook.

Additional Comments

Additional Comments Regarding the PM

Review

<p><i>The preventive maintenance checks and if applicable performance tests for Clarus600/680 GC have been completed.</i></p>	
<p><i>This Clarus600/680 GC Pass the preventive maintenance.</i></p>	
<p>Review of Preventive Maintenance:</p>	
<p>Authorized PerkinElmer Representative:</p> <p>Monchai Kitcharoenkeat [REDACTED]</p>	<p>Date:</p> <p>13-Aug-2025 (DD-MMM-YYYY)</p>
<p>Authorized Customer Representative:</p> <p>Ms.Naruecha [REDACTED]</p>	<p>Date:</p> <p>13-Aug-2025 (DD-MMM-YYYY)</p>

คุณภาพอากาศในบรรยากาศ



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72 Fax : (662) 513-4221 E-mail : sale@spscon.com, www.spscon.com

High Volume Air Sampler Calibration Report

Calibration Method : Multipoint Orifice Flow Transfer Standard

Model : TE 5025A

S/N : 3440

Calibration Data

High Volume Air Sampler Data		Calibration Data		
Recorder No.	Blower No.	Date	Actual Flowrate (ft ³ /min)	R ²
B35	B35	01/08/2025	$y = 1.126x - 2.314$	0.997
B36	B36	01/08/2025	$y = 1.158x - 3.625$	0.999
B37	B37	01/08/2025	$y = 1.071x - 0.714$	0.998
B38	B38	07/08/2025	$y = 1.138x - 6.470$	0.999
B39	B39	07/08/2025	$y = 1.074x - 2.233$	0.999
B40	B40	01/08/2025	$y = 1.137x - 4.281$	0.998
B41	B41	01/08/2025	$y = 1.124x - 3.061$	0.999
B42	B42	01/08/2025	$y = 1.130x - 3.831$	0.998
B43	B43	04/08/2025	$y = 1.098x - 1.647$	0.999
B44	B44	07/08/2025	$y = 1.107x - 2.029$	0.997
R01	R01	01/08/2025	$y = 1.027x + 1.685$	0.998
R02	R02	01/08/2025	$y = 1.154x - 5.444$	0.998
R03	R03	01/08/2025	$y = 1.174x - 5.934$	0.999
R04	R04	04/08/2025	$y = 1.125x - 3.465$	0.997
R05	R05	01/08/2025	$y = 1.097x + 0.437$	0.999
R06	R06	04/08/2025	$y = 1.138x - 2.560$	0.997
R07	R07	01/08/2025	$y = 1.046x - 0.699$	0.999
R08	R08	01/08/2025	$y = 1.109x - 3.582$	0.997
R09	R09	01/08/2025	$y = 1.088x - 1.852$	0.999
R10	R10	01/08/2025	$y = 1.134x - 4.535$	0.996
R11	R11	01/08/2025	$y = 1.170x - 6.929$	0.998
R12	R12	01/08/2025	$y = 1.151x - 4.183$	0.999
R13	R13	01/08/2025	$y = 1.117x - 4.198$	0.999
R14	R14	01/08/2025	$y = 1.109x - 2.662$	0.998
R15	R15	01/08/2025	$y = 1.126x - 5.806$	0.996
R16	R16	01/08/2025	$y = 1.149x - 7.086$	0.996
R17	R17	01/08/2025	$y = 1.120x - 5.050$	0.997
R18	R18	04/08/2025	$y = 1.155x - 5.737$	0.997
R19	R19	04/08/2025	$y = 1.131x - 5.715$	0.997
R20	R20	01/08/2025	$y = 1.152x - 5.912$	0.996

Calibrated by :

(Mr. Adul Dangklom)

Approved by :

(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด

S.P.S. CONSULTING SERVICE CO., LTD.

7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900

7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900

Tel : (662) 939-4370-72 Fax : (662) 513-4221 E-mail : sale@spscon.com. www.spscon.com

High Volume PM-10 Air Sampler Calibration Report

Calibration Method : Multipoint Orifice Flow Transfer Standard

Model : TE 5025A

S/N : 3440

Calibration Data

High Volume PM-10 Data		Calibration Data		
Recorder No.	Blower No.	Date	Actual Flowrate (ft ³ /min)	R ²
B01	B01	01/08/2025	$y = 1.114x - 2.914$	0.997
B02	B02	07/08/2025	$y = 1.013x + 1.223$	0.998
B03	B03	01/08/2025	$y = 1.161x - 6.637$	0.997
B04	B04	01/08/2025	$y = 1.104x - 4.741$	0.999
B05	B05	01/08/2025	$y = 1.139x - 4.983$	0.999
B06	B06	07/08/2025	$y = 1.115x - 4.334$	0.997
B07	B07	01/08/2025	$y = 1.134x - 5.274$	0.999
B08	B08	07/08/2025	$y = 1.118x - 2.369$	0.999
B09	B09	01/08/2025	$y = 1.043x - 0.834$	0.999
B10	B10	01/08/2025	$y = 1.096x - 2.892$	0.998
B11	B11	01/08/2025	$y = 1.114x - 3.605$	0.997
B12	B12	06/08/2025	$y = 1.096x - 2.892$	0.998
B13	B13	04/08/2025	$y = 1.112x - 4.752$	0.996
B14	B14	01/08/2025	$y = 1.104x - 3.418$	0.997
B15	B15	01/08/2025	$y = 1.119x - 2.509$	0.996
B16	B16	01/08/2025	$y = 1.012x + 1.776$	0.996
B17	B17	04/08/2025	$y = 1.094x - 0.874$	0.999
B18	B18	07/08/2025	$y = 1.140x - 5.779$	0.997
B19	B19	04/08/2025	$y = 1.087x - 0.543$	0.999
B20	B20	01/08/2025	$y = 1.108x - 3.582$	0.997
B21	B21	01/08/2025	$y = 1.138x - 4.442$	0.996
B22	B22	01/08/2025	$y = 1.097x - 3.833$	0.999
B23	B23	01/08/2025	$y = 1.127x - 4.713$	0.999
B24	B24	01/08/2025	$y = 1.117x - 4.019$	0.999
B25	B25	01/08/2025	$y = 1.137x - 5.745$	0.996
B26	B26	01/08/2025	$y = 1.029x - 0.023$	0.998
B27	B27	01/08/2025	$y = 1.136x - 6.732$	0.996
B28	B28	01/08/2025	$y = 1.114x - 4.531$	0.999
B29	B29	01/08/2025	$y = 1.126x - 5.420$	0.999
B30	B30	01/08/2025	$y = 1.119x - 4.736$	0.998
B31	B31	01/08/2025	$y = 1.011x + 2.394$	0.998
B32	B32	01/08/2025	$y = 1.047x - 0.534$	0.999
B33	B33	01/08/2025	$y = 1.052x - 0.474$	0.998
B34	B34	07/08/2025	$y = 1.028x + 2.008$	0.997

Calibrated by :

(Mr. Adul Dangklom)

Approved by :

(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด

S.P.S. CONSULTING SERVICE CO., LTD.

7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900

7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chulachak, Bangkok 10900

Tel : (662) 939-4370-72 Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

CALIBRATION REPORT					
CHEMILUMINESCENT NO / NO ₂ / NO _x ANALYZER					
DATE :	02 November 2025	BRAND :	API	MODEL :	TML-41M
NO.	NOX-B20	SERIAL NO.	N02782		
Calibrator (Dilution System)					
Brand	: Teledyne			Model	: 700
Last Cal. Date	: 24 October 2025			Serial No.	: 421
Reference Standard Gas					
Standard Gas	: Nitric Oxide (NO)			Cylinder No.	: A00726SV
Certified Date	: 05 January 2023	Expired Date	: 05 January 2026	Cylinder Conc.	: 48.8 ppm
CALIBRATING CONDITION					
Pressure	1011	mmbar	Temp.	24.6	°C
			% RH	50	
CALIBRATION SETTING					
Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response	Slope
Zero	0	0.10	-	0	-
NO Span	400	399.7	-0.075	400.0	1.006
NO _x Span	400	400.1	0.025	400.0	1.009
API Model TML-41M NO _x Analyzer Check List					
Test Values	Observed Value	Units	Nominal Range		
RANGE	500	PPB	500 standard		
STABILITY (Zero Gas)	0.1	PPB	< 2 with zero air		
SAMPLE FLOW	511	cc/min	500 ± 50		
OZONE FLOW	79	cc/min	80 ± 15		
PMT	103.1	mV	-20 - 150		
AZERO	93.7	mV	-20 - 150		
HVPS	673	V	420 - 900 constant		
RCELL TEMP	50.1	°C	50 ± 1		
BOX TEMP	29.2	°C	8 - 48		
PMT TEMP	7.3	°C	7 ± 2		
MOLY TEMP	315.1	°C	315 ± 5		
RCELL PRESS	8.4	IN-Hg-A	2 - 10 constant		
SAMPLE PRESS	28.6	IN-Hg-A	25 - 30 constant		
NO Span Conc	400	PPB	20 - 20,000		
NO _x Span Conc	400	PPB	20 - 20,000		
NO Slope	1.006	-	1.0 ± 0.3		
NO _x Slope	1.009	-	1.0 ± 0.3		
NO Offset	1.3	mV	-20 to +150		
NO _x Offset	0.9	mV	-20 to 150		
Stability at Zero	0.1	PPB	< 0.2		
Stability at Span	0.2	PPB	< 2 ppb @ 400 ppb span gas		

Calibrated by :

(Mr.Adul Dangklom)


Approved by :


(Mr.Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72. Fax : (662) 513-4221. E-mail : sale@spscon.com, www.spscon.com

Calibration Report					
Non-Dispersive Infrared CO Analyzer					
Date :	02 November 2025	Brand :	API	Model :	300E
No.	CO-804			Serial No.	3089
Calibrator (Dilution System)					
Brand : Teledyne			Model : 700		
Last Cal. Date : 24 October 2025			Serial No. : 421		
Reference Standard Gas					
Standard Gas : Carbon Monoxide (CO)			Cylinder No. : D711839		
Certified Date : 14 March 2024		Expired Date : 14 March 2032		Cylinder Conc. : 4,580 ppm	
Calibrating Condition					
Pressure : 1011 mmbar		Temp. : 24.6 °C		% RH : 50	
Calibration Setting					
Span	Initial Reading (Before Adj.), PPM			Final Reading (After Adj.), PPM	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response	
Zero	0	0.10	-	0	
CO Span	40.00	39.96	-0.100	40.00	
API Model 300E CO Analyzer Check List					
Parameter	Observed Value	Units	Nominal Range		
Range	50	PPM	0-1000 ppm		
Stability	0.10	PPM	< 1 ppm With Zero Air		
CO Measure	4015.2	mV	2500-4800 mV		
CO Reference	3947.7	mV	2500-4800 mV		
Measure/Reference Ratio	1.180	-	1.1-1.3 W/Zero Air		
Sample Pressure	28.7	In-Hg-A	~2" < Ambient Absolute Pressure		
Sample Flow	808	CC/Min	800 ± 10%		
Sample Temperature	48.4	°C	48 ± 4		
Bench Temperature	48.2	°C	48 ± 2		
Wheel Temperature	68.3	°C	68 ± 2		
Box Temperature	30.7	°C	Ambient Temp + 7 ± 10		
Photo-Drive	3030.8	mV	250 mV to 4750 mV		
Slope	1.017	-	1.0 ± 0.3		
Offset	0.2	-	0 ± 0.3		

Calibrated by : 
(Mr. Abdul Dangklom)

Approved by : 
(Mr. Peera Detudom)



CERTIFICATE No : 25M2254
REFERENCE No : 76365-1

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : DIGITAL BALANCE
MANUFACTURER : METTLER TOLEDO
MODEL : XS105DU
SERIAL No : 1126422905
ID No : BA05/50
CONDITION AS RECEIVED : USED ITEM
SUBMITTED BY : S.P.S. CONSULTING SERVICE CO., LTD.
7 SOI PHAHOLYOTHIN 24, PHAHOLYOTHIN RD.,
JOMPOL, CHATUCHAK, BANGKOK 10900

CALIBRATED BY : ATSAWIN Y.

CALIBRATION DATE : 07-Mar-25

APPROVED BY : 
PONGSAK J.

ISSUED DATE : 13-Mar-25

RECEIVED DATE : 07-Mar-25

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.





CERTIFICATE No : 25M2254

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : DIGITAL BALANCE MODEL : XS105DU
MANUFACTURER : METTLER TOLEDO S/N : 1126422905
ID No : BA05/50 RECEIVED DATE : 07-Mar-25
AIR PRESSURE : 1009mbar \pm 1mbar CALIBRATION DATE : 07-Mar-25
AMBIENT TEMPERATURE : 24°C \pm 1°C RELATIVE HUMIDITY : 54 %RH \pm 10 % RH

CONDITION OF THIS RESULTS OF CALIBRATION

1. THIS INSTRUMENT WAS CALIBRATED BY ACCORDING TO UKAS LAB 14 EDITION 6:2019 BY USING KNOWN WEIGHT STANDARD WEIGHT. THE BALANCE WAS NOT ADJUSTED BEFORE CALIBRATION. THE BALANCE HAS NO ZERO TRACKING FUNCTION. REPEATABILITY WAS MEASURED BY USING 10 REPEATED MEASUREMENTS. LINEARITY WAS MEASURED COVERING 10 POINTS, EVENLY SPREAD OVER THE RANGE. THE INSTRUMENT WAS SET ZERO BEFORE PERFORMING THE LINEARITY TEST. OFF-CENTER LOADING WAS MEASURED BY USING STANDARD WEIGHTS PLACED ON THE PAN AND MOVED TO VARIOUS POSITIONS ON THE PAN.

2. REFERENCE STANDARD INSTRUMENTS :-

INSTRUMENT	MODEL	SERIAL No	CERTIFICATE No	DUE DATE
1) STANDARD WEIGHT SET	E2	QK-I-151	C02250116	28-Jan-27
2) STANDARD WEIGHT	E2	15843	C02250117	29-Jan-27

3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.

4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.

5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-
- NATIONAL INSTITUTE OF METROLOGY (THAILAND)

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT

1. ZERO SETTING FUNCTION : NORMAL


2. TARE FUNCTION : NORMAL

3. REPEATABILITY OF READING AT 120 g WAS 0.000055 g

4. DEPARTURE FROM NOMINAL VALUE/ LINEARITY

NOMINAL VALUE (g)	BALANCE READING (g)	CORRECTION (g)	UNCERTAINTY (\pm g)
0.00	0.00000	0.00000	0.000065
0.02	0.01999	0.00001	0.000065
0.10	0.10001	-0.00001	0.000066
0.20	0.20001	-0.00001	0.000066
0.50	0.50002	-0.00002	0.000065
1.00	1.00003	-0.00003	0.000066
2.00	2.00001	-0.00001	0.000067
5.00	5.00002	-0.00002	0.000068
10.00	10.00000	0.00000	0.000070
20.00	20.00004	-0.00004	0.000078
50.00	50.00000	0.00000	0.00013
100.00	100.00001	-0.00001	0.00019
120.00	120.00002	-0.00002	0.00022

5. OFF CENTER LOADING ERROR



POINT	READING (g)
1	50.0000
2	50.0000
3	50.0000
4	50.0000
5	50.0000
OFF-CENTER LOADING	0.0000

NOTE: THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA

THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR $k=2$, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.

END OF CALIBRATION REPORT





MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

Customer : <u>S.P.S.Consulting Service Co.,Ltd</u>	Date Tested: <u>December 18, 2025</u>	
	Recommendation Recertification	
Address : <u>7 Soi Phaholyothin 24</u>	Period <u>6</u> Months	
<u>Paholyothin Road</u>	Recertification Due: <u>June 28, 2026</u>	
<u>Jompol Chatuchak, Bangkok 1090</u>	Date Last Certified: <u>July 1, 2025</u>	
User Name: <u>K.Phenpha Viphasathawat</u>	Visit Number: <u>2 of 2</u>	
Phone: <u>083-9269252</u>	PerkinElmer Phone: <u>02-719-6420 ext 206</u>	
Fax: <u>02-513-4221</u>	PerkinElmer Fax: <u>02-318-5597</u>	

CONFIGURATION TESTED		ACCESSORIES/COMPONENT NOT INCLUDED
MODEL	SERIAL NUMBER	
<u>OPTIMA 5300DV</u>	<u>077C7042401</u>	
TESTED EQUIPMENT	CALIBRATION NUMBER	EXPIRATION
<u>IPV Methods</u>		
TEST STANDARD USED	PART NUMBER	EXPIRATION DATE
<u>Multielement Standard</u>	<u>N069-1579</u>	<u>November 30, 2026</u>
<u>Wavecal Solution</u>	<u>N058-2152</u>	<u>July 30, 2026</u>
<u>VIS Wavecal solution</u>	<u>N930-2946</u>	<u>August 30, 2026</u>
<u>Instrument Cal. STD4</u>	<u>N930-0221</u>	<u>November 30, 2026</u>
CUSTOMER SUPPLIED	COMMENTS	CUSTOMER INITIALS
<u>2 % HNO3</u>		
<u>10 % HNO3</u>		



MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

SERIAL NUMBER 077C7042401**DATE TESTED** December 18, 2025**1. MECHANICAL CHECKS**

A. Inspect and clean all fans and filters.

☐ OK

B. Inspect and replace as necessary, all torch components including the RF coil.

☐ OK

C. Inspect all tubing for sign of clacking or leaking.

☐ OK

D. Adjust water and gas pressure regulator settings.

☐ OK

E. Inspect and leak check pneumatics drawers.

☐ OK

F. Clean the exterior of the instrument.

☐ OK**2. OPTICAL CHECKS**

A. Inspect and clean all optical components.

☐ OK

B. As required, check and replace all purgefilters.

☐ OK

C. Recheck optical alignment.

☐ OK**3. COOLING SYSTEM CHECKS**

A. Perform preventive maintenance on chiller.

☐ OK

B. Flush out the chiller every year.

☐ N/A**4. PERFORMANCE CHECKS**

A. Torch View Alignment.

☐ OK

B. Wavelength Calibration.

☐ OK



MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

SERIAL NUMBER : 077C7042401			DATE TESTED : December 18, 2025		
PARAMETER	SPECIFICATION			FINAL VALUE	
Spectral Resolution : UV	As	193.696 nm	≤ 0.007	0.00530	
	Ni	231.604 nm	≤ 0.008	0.00708	
	Ni	341.476 nm	≤ 0.012	0.00776	
Spectral Resolution : VIS	La	408.672 nm	≤ 0.020	0.01614	
	Ba	455.403 nm	≤ 0.025	0.02377	
Precision					
	As	193.656 nm	% RSD < 1.0	0.67	%
	Zn	213.856 nm	% RSD < 1.0	0.62	%
	Mn	257.610 nm	% RSD < 1.0	0.88	%
	La	379.478 nm	% RSD < 1.0	0.63	%
	Ba	455.403 nm	% RSD < 1.0	0.65	%
	Ba	493.408 nm	% RSD < 1.0	0.45	%
Detection Limits : Axial	Tl	190.080 nm	3(sd)	3.21	ppb
	As	193.696 nm	3(sd)	6.06	ppb
	Pb	220.353 nm	3(sd)	0.92	ppb
Detection Limits : Radial	As	193.696 nm	3(sd)	17.35	ppb
	Zn	213.856 nm	3(sd)	1.79	ppb
	Mn	257.610 nm	3(sd)	0.18	ppb
	La	379.478 nm	3(sd)	0.76	ppb
	Ba	455.403 nm	3(sd)	0.11	ppb
	Ba	493.408 nm	3(sd)	0.56	ppb
BEC : Axial (IB X 500)/(IS-IB)	Cd	226.502 nm	≤ 150 ppb	40.52	
BEC : Radial (IB X 1000)/(IS-IB)	Mn	257.610 nm	≤ 45 ppb	42.33	



MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

SERIAL NUMBER 077C7042401**DATE TESTED** December 18, 2025**Remarks :**

Commissioning follow as commissioning performance sheets.

This is to certify that the above tests have been performed and the configuration tested



meets



does not meet

the PerkinElmer Specifications listed on this certificate.

This certificate does not modify PerkinElmer's standard terms and condition of sale,
including warranty terms.

Service Department PerkinElmer Ltd.

Authorized Representative:

(Wiphan Promlumda)

Service Engineer

ระดับเสียงบริเวณริมรั้วโรงงาน



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-68/0220

MTC No. EEL. BP. 44/0268

CALIBRATION CERTIFICATE

Submitted by : S.P.S.Consulting Service Co.,Ltd.

Address : 7 Soi Phaholyothin 24, Phaholyothin Road, Jompol, Chatuchak, Bangkok 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

Instrument Calibrated :

Description : Sound Calibrator

Manufacturer : ACO

Model : 2127

Serial No. : 130006

Ambient Environment

Temperature : $(23 + 3) ^\circ\text{C}$

Relative Humidity : $(50 \pm 15) \%$

Ambient Pressure : $(101.325 \pm 1.500) \text{ kPa}$

Standards used :

1. Digital Function Synthesizer NF Electronic DF-193A S/N 122037.
2. Measuring Amplifier Bruel&Kjaer 2636 S/N 1537484.
3. Programmable Attenuator Tamagawa TPA-303A S/N OF 2214.
4. Digital Multimeter Agilent 34401A S/N MY44005560.
5. Pressure Transmitter Vaisala PTB202AD S/N T0650001.
6. Audio Analyzer Panasonic VP-7722A S/N 041477D122.
7. Condenser Microphone B&K 4180 S/N 2889871.

Calibration Procedure: CP-102-04 based on IEC 60942-2003; The sound pressure level generated by sound calibrator under test shall be measured by standard microphone using an insert voltage technique.

This instrument has been calibrated against standards maintained at Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

Date of Receipt : 19 Feb. 2025

Date of Calibration : 21 Feb. 2025

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5

Head Office

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9036
Fax. (66) 0 2577 9009

Office/Laboratory

668 Mu 2 Tambon Bangpoomai, Amphoe Muang Samutprakan,
Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
(66) 08 3219 9440
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office

196 Phahonyothin Road, Ladyao, Chatuchak,
Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
(66) 08 1889 6827

THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-68/0220

MTC No. EEL. BP. 44/0268

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.

Nominal Output of Unit Under Test = 94 dB re 20 μ Pa at 1000 Hz

Acoustic Output in dB re 20 μ Pa, Corrected to Reference Conditions: 101.325 kPa, 23.0 °C and 50 %RH.

1. Sound Pressure Level

Standard Microphone Type	Measured Sound Pressure Level (dB)	Deviated value (dB)	Uncertainty (dB)	Tolerance limit IEC60942:2003 Class 1
1/2 inch Bruel&Kjaer 4180	93.81	-0.19	± 0.10	± 0.40 dB

2. Frequency

Standard Microphone Type	Measured Frequency (Hz)	Deviated value (Hz)	Uncertainty (Hz)	Tolerance limit IEC60942:2003 Class 1
1/2 inch Bruel&Kjaer 4180	999.9	-0.1	± 1.5	$\pm 1.0\%$

3. Total Distortion

Standard Microphone Type	Measured Total Distortion (%)	Uncertainty (%)	Tolerance limit IEC60942:2003 Class 1
1/2 inch Bruel&Kjaer 4180	0.95	± 0.50	$\pm 3.0\%$

Note : 1. No adjustment.

2. The calibrator pressure correction was not included.

3. The microphone volume correction was not included.

Calibrated by :



(Mr. Weerachai Deechaiyae)

Approved by :



(Mr. Prawate Kluaypa)

Director

Electrical and Electronic Standards Laboratory
Industrial Metrology and Testing Service Centre

Date of Calibration : 21 Feb. 2025

Date of Issue : 24 Feb. 2025

Ref : 2011268021900739001

End of Certificate

2 / 2

The results relate only to the items tested/calibrated or value assigned.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5

Head Office

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9036
Fax. (66) 0 2577 9009

Office/Laboratory

668 Mu 2 Tambon Bangpoornai, Amphoe Muang Samutprakan,
Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
(66) 08 3219 9440
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office

196 Phahonyothin Road, Ladyao, Chatuchak,
Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
(66) 08 1889 6827



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Noise B_422/25

Sound Level Meter Calibration Report

Acoustic Calibrator Data

Brand	ACO	Number	AC 03/56
Model	2127	Serial No.	130006
Calibration Range	94 dB, 1000 Hz	Last Calibration	21 February 2025
		Due Date	21 February 2026

Calibration Data

Sound Level Meter Data				Calibration Data		
SLM No.	Brand	Model	Serial No.	Date	Actual Reading [dB]	
					Before Adjustment	After Adjustment
ACO-B22	ACO	6236	00172060	01 September 2025	93.8	93.9
ACO-B33	ACO	6236	00182015	01 September 2025	93.9	93.9
ACO-R50	ACO	6236	00192062	01 September 2025	93.9	93.9
ACO-R52	ACO	6236	00192064	01 September 2025	93.9	93.9
Acoustic Certified Value : Thailand Institute of Scientific and Technological Research (TISTR)					93.81 ± 0.10 dB	

Calibrated by :



(Mr. Adul Dangklom)

Approved by :



(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com., www.spscon.com

Noise B_565/25

Sound Level Meter Calibration Report

Acoustic Calibrator Data

Brand	ACO	Number	AC 03/56
Model	2127	Serial No.	130006
Calibration Range	94 dB, 1000 Hz	Last Calibration	21 February 2025
		Due Date	21 February 2026

Calibration Data

Sound Level Meter Data				Calibration Data		
SLM No.	Brand	Model	Serial No.	Date	Actual Reading [dB]	
					Before Adjustment	After Adjustment
ACO-B30	ACO	6236	00182012	30 November 2025	93.7	93.9
ACO-R04	ACO	6236	00132030	30 November 2025	93.8	93.9
ACO-R11	ACO	6236	00192048	30 November 2025	93.8	93.9
ACO-R19	ACO	6236	00182001	30 November 2025	93.9	93.9
Acoustic Certified Value: Thailand Institute of Scientific and Technological Research (TISTR)					93.81 ± 0.10 dB	

Calibrated by:

(Mr. Adul Dangkom)

Approved by:

(Mr. Peera Detudom)

คุณภาพน้ำ



CALIBRATION LABORATORY Co., LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



CERTIFICATE OF CALIBRATION FOR

NOMENCLATURE : pH METER
MANUFACTURER : HANNA
MODEL / TYPE : HI3512/HI1332/HI7662-T
SERIAL NO. : 08685754/11250B7M/092806BN[PH04/56]
CLID. NO. : 272501562
JOB CONTROL NO. : 250617070523
CALIBRATION SERVICE : ☒ IN-LABORATORY ☐ ON-SITE

CUSTOMER : S.P.S. CONSULTING SERVICE CO., LTD.
7 SOI PHAHOLYOTHIN 24 ROAD, JOMPOL,
CHATUCHAK, BANGKOK 10900

DATE OF RECEIVED : 17 June 2025


DATE OF ISSUED : 20 June 2025

The report of calibration shall not be reproduced except in full without approval of the Calibration Laboratory Co., Ltd.

Calibrated By : Sukgasem Seehanart
Wenick Inchaisri
Calibration Engineer



Approved By : Mongkol Yotsoontorn
Authorized Signatory
20 June 2025



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)


Certificate No. Q25070523

F3-011-05/12-23

page 1 of 4



@clccalibration



REPORT OF CALIBRATION

FOR

NOMENCLATURE : **pH METER**
MANUFACTURER : **HANNA**
MODEL / TYPE : **HI3512/HI1332/HI7662-T**
SERIAL NO. : **08685754/11250B7M/092806BN[PH04/56]**
DATE OF CALIBRATION : **18 June 2025**

ENVIRONMENT CONDITIONS :

Temperature : $(25 \pm 2.5) ^\circ\text{C}$

Relative Humidity : $(50 \pm 15) \% \text{ RH}$

PROCEDURE USED :

This instrument was calibrated under procedure No. **CLC-CPCH-01** [pH Meter]. The calibration was performed by direct measurement with Certified Reference Material (CRM).

This instrument was calibrated under procedure No. **CLC-CPTH-04** [Temperature] based on **ASTM E 644-04** as calibration guidelines. The calibration was performed by using Calibration Bath, Precision Thermometer and IPRT which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

1. pH Standard Solution, NIMT TRM CODE TRM-S-2003, TRM CODE TRM-S-2007.
2. pH Standard Solution, Control Company Catalog Number 06664260,11754256, Lot Number CC787362.
3. Calibration Bath, Kambic Model OB-22/2 ULT S/N. 17115653.
4. Precision Thermometer, ASL Model F250 S/N. 1334023800.
5. IPRT, Wika Model CTP5000-250-D S/N. PO00043543-1-10-1.





CALIBRATION LABORATORY CO., LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



TRACEABILITY :

1. The measurements are traceable to International System of Units (SI) , through National Institute of Metrology (Thailand).
Lot Number. 080124 , 120124. Due Date 23 January 2026.
2. The measurements are traceable to International System of Units (SI) , through Control Company.
Certificate No. 4281-14495731 , Due Date 27 September 2025.
3. The measurements are traceable to International System of Units (SI) , through Calibration Laboratory Co., Ltd.
Certificate No. Q24120999, Due Date 26 November 2025.
4. The measurements are traceable to International System of Units (SI) , through Thailand Institute of Scientific and Technological Research (TISTR). Certificate No. PSL-T 1042/67, Due Date 16 October 2025.
5. The measurements are traceable to International System of Units (SI) , through National Institute of Metrology (Thailand).
Certificate No. TT-0146-24, Due Date 28 October 2025.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.

It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2022)"

Certificate No. Q25070523

F3-011-05/12-23

page 3 of 4



@clccalibration



CLC
Accredited
ISO/IEC 17025

CALIBRATION LABORATORY Co., LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Rd., Ladphrao, Bangkok 10230
Tel. 02-578-0353-4 Fax: 02-578-2672 www.cal-laboratory.com E-mail:sale@cal-laboratory.com



CONDITION OF CALIBRATION ITEM : RECEIVED IN GOOD OPERATIONAL CONDITION

MEASUREMENT RESULTS : (X) without adjustment () adjustment

The table in the following gives the calibration results and associated measurement uncertainties of pH meter.

CALIBRATION DATA

1. pH METER RESULT @ 25 °C

Standard pH Buffer Solution (pH)	pH Meter Reading (pH)	pH Meter Reading (mV)	Correction (pH)	Uncertainty of pH Measurement (\pm pH)	k Factor
4.003	4.005	168.2	-0.002	0.010	2,00
7.005	7.010	-8.1	-0.005	0.013	2,00
10.015	10.010	-177.7	+0.005	0.014	2,00

Technical Note. Setting function CAL 3 point (4,7,10).

Note. The Scope of Accredited ANAB Certificate No. ACDM-2814 Version 015 Page 4 of 68

2. TEMPERATURE RESULT

Immersion depth (mm)	Actual Temperature (°C)	DUC Reading (°C)	Correction (°C)	Uncertainty \pm (°C)
100	25.00	25.0	0.00	0.07

Technical Note. Type of sensor : Thermistor

Probe \varnothing 3 mm

Materials : Metal Sheath.

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor of $k = 2,00$.

Note. The Scope of Accredited ANAB Certificate No. ACDM-2814 Version 015 Page 56 of 68

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q25070523

F3-011-05/12-23

page 4 of 4



@clccalibration



CERTIFICATE No : 25M2256

REFERENCE No : 76365-3

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : DIGITAL BALANCE

MANUFACTURER : SARTORIUS

MODEL : BSA224S-CW

SERIAL No : 36591843

ID No : BA09/61

CONDITION AS RECEIVED : USED ITEM

SUBMITTED BY : S.P.S. CONSULTING SERVICE CO., LTD.
7 SOI PHAHOLYOTHIN 24, PHAHOLYOTHIN RD.,
JOMPOL, CHATUCHAK, BANGKOK 10900

CALIBRATED BY : ATSAWIN Y.

CALIBRATION DATE : 07-Mar-25

APPROVED BY : 
PONGSAK J.

ISSUED DATE : 13-Mar-25

RECEIVED DATE : 07-Mar-25

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.





CERTIFICATE No : 25M2256

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : DIGITAL BALANCE MODEL : BSA224S-CW
MANUFACTURER : SARTORIUS S/N : 36591843
ID No : BA09/61 RECEIVED DATE : 07-Mar-25
AIR PRESSURE : 1009mbar \pm 1mbar CALIBRATION DATE : 07-Mar-25
AMBIENT TEMPERATURE : 24° C \pm 1° C RELATIVE HUMIDITY : 52 %RH \pm 10 % RH

CONDITION OF THIS RESULTS OF CALIBRATION

1. THIS INSTRUMENT WAS CALIBRATED BY ACCORDING TO UKAS LAB 14 EDITION 6:2019 BY USING KNOWN WEIGHT STANDARD WEIGHT. THE BALANCE WAS NOT ADJUSTED BEFORE CALIBRATION. THE BALANCE HAS NO ZERO TRACKING FUNCTION. REPEATABILITY WAS MEASURED BY USING 10 REPEATED MEASUREMENTS. LINEARITY WAS MEASURED COVERING 10 POINTS, EVENLY SPREAD OVER THE RANGE. THE INSTRUMENT WAS SET ZERO BEFORE PERFORMING THE LINEARITY TEST. OFF-CENTER LOADING WAS MEASURED BY USING STANDARD WEIGHTS PLACED ON THE PAN AND MOVED TO VARIOUS POSITIONS ON THE PAN.

2. REFERENCE STANDARD INSTRUMENTS :-

INSTRUMENT	MODEL	SERIAL No	CERTIFICATE No	DUE DATE
1) STANDARD WEIGHT SET	E2	QK-I-151	C02250116	28-Jan-27
2) STANDARD WEIGHT	E2	15843	C02250117	29-Jan-27

3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.

4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.

5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-

- NATIONAL INSTITUTE OF METROLOGY (THAILAND)

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT

1. ZERO SETTING FUNCTION : NORMAL


2. TARE FUNCTION : NORMAL

3. REPEATABILITY OF READING AT 200 g WAS 0.000071 g

4. DEPARTURE FROM NOMINAL VALUE/ LINEARITY

NOMINAL VALUE (g)	BALANCE READING (g)	CORRECTION (g)	UNCERTAINTY (\pm g)
0.00	0.0000	0.0000	0.00012
0.10	0.1000	0.0000	0.00012
0.20	0.2000	0.0000	0.00012
0.50	0.5000	0.0000	0.00012
1.00	1.0000	0.0000	0.00012
2.00	2.0000	0.0000	0.00012
5.00	5.0000	0.0000	0.00012
10.00	10.0000	0.0000	0.00012
20.00	20.0001	-0.0001	0.00012
50.00	50.0000	0.0000	0.00014
100.00	100.0001	-0.0001	0.00019
200.00	200.0001	-0.0001	0.00032

5. OFF CENTER LOADING ERROR



POINT	READING (g)
1	100.0000
2	100.0000
3	100.0000
4	100.0000
5	100.0000
OFF-CENTER LOADING	0.0000

NOTE: THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA

THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR $k=2$, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.

END OF CALIBRATION REPORT



CERT.No.: HS-W015C

Calibration Date : 18 Mar 25
Submitted by : S.P.S CONSULTING SERVICE CO.,LTD
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol,
Chatuchak, Bangkok, Thailand 10900

Avg Room Temp : 20 °C
Avg Water Temp : 20 °C
Air Pressure : 760.00 mmHg
Salinity : 0 ppt

Model : YSI 5000
S/N : 15B100751
Probe : YSI 5010
S/N : 22D100097
ID NO. : -
Air Temp ref : S/N. F8065C26
Barometric ref : S/N. F8065C26
Water Temp ref : -
ID NO. HS001
Technician : Kittipong M.

Calibration Details

Calibration Point	100% air sat. (@20 °C, DO = 9.09 mg/l)	(status)	(status)
Measurement 1 (mg/l)	9.08	(PASS)	-
Measurement 2 (mg/l)	9.08	(PASS)	-
Measurement 3 (mg/l)	9.08	(PASS)	-
Measurement 4 (mg/l)	9.07	(PASS)	-
Measurement 5 (mg/l)	9.07	(PASS)	-
Measurement 6 (mg/l)	9.07	(PASS)	-
Measurement 7 (mg/l)	9.07	(PASS)	-
Measurement 8 (mg/l)	9.07	(PASS)	-
Measurement 9 (mg/l)	9.07	(PASS)	-
Measurement 10 (mg/l)	9.07	(PASS)	-

Mean Measurement	9.07	mg/l	-	-
Inaccuracy	0.02	mg/l	-	-

Overall Status (PASS)

Manufacturer Specification

Accuracy = +/- 0.02 mg/l

- 1) This certificate is issued based on the result that are found as shown on date and place of test only.
- 2) The calibration procedure followed in accordance with Harikul Science Co., Ltd.
- 3) This result shall not be used for advertising purpose.

Technician Signature
(Kittipong Maekwong)

Laboratory Manager
(Natenapha Pisatkunchon)



MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

Customer : <u>S.P.S.Consulting Service Co.,Ltd</u>	Date Tested: <u>December 18, 2025</u>	
	Recommendation Recertification	
Address : <u>7 Soi Phaholyothin 24</u>	Period <u>6</u> Months	
<u>Paholyothin Road</u>	Recertification Due: <u>June 28, 2026</u>	
<u>Jompol Chatuchak, Bangkok 10900</u>	Date Last Certified: <u>July 1, 2025</u>	
User Name: <u>K.Phenpha Viphasathawat</u>	Visit Number: <u>2 of 2</u>	
Phone: <u>083-9269252</u>	PerkinElmer Phone: <u>02-719-6420 ext 206</u>	
Fax: <u>02-513-4221</u>	PerkinElmer Fax: <u>02-318-5597</u>	

CONFIGURATION TESTED		ACCESSORIES/COMPONENT NOT INCLUDED
MODEL	SERIAL NUMBER	
<u>OPTIMA 5300DV</u>	<u>077C7042401</u>	
TESTED EQUIPMENT	CALIBRATION NUMBER	EXPIRATION
<u>IPV Methods</u>		
TEST STANDARD USED	PART NUMBER	EXPIRATION DATE
<u>Multielement Standard</u>	<u>N069-1579</u>	<u>November 30, 2026</u>
<u>Wavecal Solution</u>	<u>N058-2152</u>	<u>July 30, 2026</u>
<u>VIS Wavecal solution</u>	<u>N930-2946</u>	<u>August 30, 2026</u>
<u>Instrument Cal. STD4</u>	<u>N930-0221</u>	<u>November 30, 2026</u>
CUSTOMER SUPPLIED	COMMENTS	CUSTOMER INITIALS
<u>2 % HNO3</u>		
<u>10 % HNO3</u>		



MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

SERIAL NUMBER 077C7042401
DATE TESTED December 18, 2025
1. MECHANICAL CHECKS

A. Inspect and clean all fans and filters.

☐ OK

B. Inspect and replace as necessary, all torch components including the RF coil.

☐ OK

C. Inspect all tubing for sign of clacking or leaking.

☐ OK

D. Adjust water and gas pressure regulator settings.

☐ OK

E. Inspect and leak check pneumatics drawers.

☐ OK

F. Clean the exterior of the instrument.

☐ OK

2. OPTICAL CHECKS

A. Inspect and clean all optical components.

☐ OK

B. As required, check and replace all purgefilters.

☐ OK

C. Recheck optical alignment.

☐ OK

3. COOLING SYSTEM CHECKS

A. Perform preventive maintenance on chiller.

☐ OK

B. Flush out the chiller every year.

☐ N/A

4. PERFORMANCE CHECKS

A. Torch View Alignment.

☐ OK

B. Wavelength Calibration.

☐ OK



MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

SERIAL NUMBER : 077C7042401			DATE TESTED : December 18, 2025		
PARAMETER	SPECIFICATION			FINAL VALUE	
Spectral Resolution : UV	As	193.696 nm	≤ 0.007	0.00530	
	Ni	231.604 nm	≤ 0.008	0.00708	
	Ni	341.476 nm	≤ 0.012	0.00776	
Spectral Resolution : VIS	La	408.672 nm	≤ 0.020	0.01614	
	Ba	455.403 nm	≤ 0.025	0.02377	
Precision					
	As	193.656 nm	% RSD < 1.0	0.67	%
	Zn	213.856 nm	% RSD < 1.0	0.62	%
	Mn	257.610 nm	% RSD < 1.0	0.88	%
	La	379.478 nm	% RSD < 1.0	0.63	%
	Ba	455.403 nm	% RSD < 1.0	0.65	%
	Ba	493.408 nm	% RSD < 1.0	0.45	%
Detection Limits : Axial	Tl	190.080 nm	3(sd)	3.21	ppb
	As	193.696 nm	3(sd)	6.06	ppb
	Pb	220.353 nm	3(sd)	0.92	ppb
Detection Limits : Radial	As	193.696 nm	3(sd)	17.35	ppb
	Zn	213.856 nm	3(sd)	1.79	ppb
	Mn	257.610 nm	3(sd)	0.18	ppb
	La	379.478 nm	3(sd)	0.76	ppb
	Ba	455.403 nm	3(sd)	0.11	ppb
	Ba	493.408 nm	3(sd)	0.56	ppb
BEC : Axial (IB X 500)/(IS-IB)	Cd	226.502 nm	≤ 150 ppb	40.52	
BEC : Radial (IB X 1000)/(IS-IB)	Mn	257.610 nm	≤ 45 ppb	42.33	



MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

SERIAL NUMBER 077C7042401**DATE TESTED** December 18, 2025**Remarks :**

Commissioning follow as commissioning performance sheets.

This is to certify that the above tests have been performed and the configuration tested



meets




does not meet

the PerkinElmer Specifications listed on this certificate.

This certificate does not modify PerkinElmer's standard terms and condition of sale,
including warranty terms.

Service Department PerkinElmer Ltd.

Authorized Representative:

(Wiphan Promlumda)

Service Engineer

คุณภาพอากาศในสถานประกอบการ



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com., www.spscon.com

Personal Pump Calibration Report

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

Environmental Conditions

Temperature : 25 \pm 3 $^{\circ}$ C
Pressure : 1010 \pm 15 mmbar

Personal Pump Data					Calibration Data							
No.	Brand	Model	Serial No.	Date	Flow Rate (mL/min)						Value From Calibration Curve	
					Setting			Actual (Q std.)			y	R ²
					1	2	3	1	2	3		
B01	SKC	224-PCXR4	262101	03/10/2025	1,000	1,500	2,000	998	1,490	1,997	1.000x - 7.191	1.000
B02	SKC	224-PCXR4	626166	03/10/2025	1,000	1,500	2,000	1,007	1,500	2,008	0.999x + 2.537	1.000
B03	SKC	224-PCXR4	612968	03/10/2025	1,000	1,500	2,000	1,003	1,503	2,001	0.997x + 0.810	0.999
B04	SKC	224-PCXR4	602804	02/10/2025	1,000	1,500	2,000	998	1,494	1,993	1.001x - 6.035	1.000
B05	SKC	224-PCXR4	612693	02/10/2025	1,000	1,500	2,000	999	1,495	2,001	0.999x - 2.481	1.000
B06	SKC	224-PCXR4	262188	02/10/2025	1,000	1,500	2,000	997	1,510	2,000	0.998x + 0.064	0.999
B07	SKC	224-PCXR4	626262	01/10/2025	1,000	1,500	2,000	1,004	1,492	2,007	1.002x - 4.778	1.000
B08	SKC	224-PCXR4	626100	02/10/2025	1,000	1,500	2,000	1,005	1,500	2,005	1.004x - 7.223	1.000
B09	SKC	224-PCXR4	626479	01/10/2025	1,000	1,500	2,000	1,001	1,501	1,986	0.996x + 3.462	0.999
B10	SKC	224-PCXR4	091950	01/10/2025	1,000	1,500	2,000	997	1,504	2,000	1.003x - 8.822	1.000
B11	SKC	224-PCXR8	564315	03/10/2025	1,000	1,500	2,000	1,001	1,503	1,995	0.995x + 2.449	1.000
B12	SKC	224-PCXR4	034656	03/10/2025	1,000	1,500	2,000	997	1,506	2,003	1.003x - 9.062	0.999
B13	SKC	224-PCXR4	602073	03/10/2025	1,000	1,500	2,000	1,003	1,497	2,006	1.002x - 5.013	1.000
B14	SKC	224-PCXR4	626313	03/10/2025	1,000	1,500	2,000	998	1,501	1,992	1.005x - 11.702	0.999
B15	SKC	224-PCXR4	626474	03/10/2025	1,000	1,500	2,000	1,001	1,502	2,004	1.006x - 11.694	1.000
B16	SKC	224-PCXR4	626477	03/10/2025	1,000	1,500	2,000	996	1,498	1,992	1.007x - 16.329	0.999
B17	SKC	224-PCXR4	626860	02/10/2025	1,000	1,500	2,000	1,001	1,503	1,998	1.001x - 4.838	1.000
B18	SKC	224-PCXR4	691484	01/10/2025	1,000	1,500	2,000	997	1,514	1,996	0.996x + 5.360	1.000
B19	SKC	224-PCXR4	691599	01/10/2025	1,000	1,500	2,000	998	1,499	2,003	0.998x + 0.399	1.000
B20	SKC	224-PCXR4	691587	01/10/2025	1,000	1,500	2,000	1,001	1,501	1,999	0.995x + 1.520	0.999
B21	SKC	224-PCXR4	691531	03/10/2025	1,000	1,500	2,000	996	1,502	2,001	1.003x - 7.151	1.000
B22	SKC	224-PCXR4	691654	03/10/2025	1,000	1,500	2,000	1,001	1,500	1,998	0.997x - 0.666	1.000
B23	SKC	224-PCXR4	798393	03/10/2025	1,000	1,500	2,000	993	1,507	1,999	1.007x - 17.505	0.999
B24	SKC	224-PCXR4	626363	03/10/2025	1,000	1,500	2,000	994	1,498	1,995	1.000x - 3.941	1.000
B25	SKC	224-PCXR4	798489	01/10/2025	1,000	1,500	2,000	1,003	1,490	2,001	0.997x + 1.703	1.000
B26	SKC	224-PCXR4	798479	01/10/2025	1,000	1,500	2,000	1,001	1,509	1,995	1.002x - 8.057	0.999
B27	SKC	224-PCXR4	691673	01/10/2025	1,000	1,500	2,000	998	1,510	2,002	1.005x - 9.656	1.000
B28	SKC	224-PCXR4	691570	01/10/2025	1,000	1,500	2,000	1,011	1,508	2,009	0.999x + 3.729	0.999
B29	SKC	224-PCXR4	626472	01/10/2025	1,000	1,500	2,000	1,002	1,503	1,998	1.002x - 6.066	1.000
B30	SKC	224-PCXR4	691489	01/10/2025	1,000	1,500	2,000	997	1,506	2,001	1.004x - 8.049	1.000
B31	SKC	224-PCXR4	691509	02/10/2025	1,000	1,500	2,000	995	1,497	1,992	0.998x - 2.293	1.000
B32	SKC	224-PCXR4	091567	01/10/2025	1,000	1,500	2,000	1,002	1,500	2,003	1.008x - 15.778	0.999
B33	SKC	224-PCXR4	091756	02/10/2025	1,000	1,500	2,000	1,003	1,501	1,997	1.003x - 6.509	1.000
B34	SKC	224-PCXR4	612962	01/10/2025	1,000	1,500	2,000	996	1,512	1,996	1.001x - 5.867	0.999
B35	SKC	224-PCXR4	602682	01/10/2025	1,000	1,500	2,000	1,008	1,494	1,999	0.993x + 6.992	1.000
B36	SKC	224-PCXR4	626164	01/10/2025	1,000	1,500	2,000	997	1,502	1,992	0.999x - 3.235	1.000
B37	SKC	224-PCXR4	626256	01/10/2025	1,000	1,500	2,000	1,003	1,490	1,997	0.994x + 5.093	1.000
B38	SKC	224-PCXR4	626167	02/10/2025	1,000	1,500	2,000	998	1,513	1,995	1.000x - 5.277	0.999
B39	SKC	224-PCXR4	034637	03/10/2025	1,000	1,500	2,000	1,007	1,504	2,004	0.996x + 8.240	1.000
B40	SKC	224-PCXR4	798349	03/10/2025	1,000	1,500	2,000	998	1,510	2,002	0.998x + 3.905	1.000

Calibrated by :

(Mr. Abdul Dangklom)

Approved by :

(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด

S.P.S. CONSULTING SERVICE CO., LTD.

7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900

7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900

Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Personal Pump Calibration Report

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

Environmental Conditions

Temperature : 25 ± 3 °C
Pressure : 1010 ± 15 mmbar

Personal Pump Data				Calibration Data								
No.	Brand	Model	Serial No.	Date	Flow Rate (mL/min)						Value From Calibration Curve	
					Setting			Actual (Q std.)				
					1	2	3	1	2	3	y	R ²
B41	SKC	224-PCXR4	612669	01/10/2025	1,000	1,500	2,000	1,010	1,508	2,009	1.000x + 2.612	0.999
B42	SKC	224-PCXR4	626041	02/10/2025	1,000	1,500	2,000	1,004	1,494	1,994	0.997x + 1.344	1.000
B43	SKC	224-PCXR4	034636	01/10/2025	1,000	1,500	2,000	998	1,505	2,002	1.001x - 5.177	1.000
B44	SKC	224-PCXR8	529341	01/10/2025	1,000	1,500	2,000	999	1,496	1,998	0.996x + 0.909	1.000
B45	SKC	224-PCXR8	529594	01/10/2025	1,000	1,500	2,000	996	1,510	1,992	1.005x - 11.543	1.000
B46	SKC	224-PCXR8	566743	01/10/2025	1,000	1,500	2,000	1,003	1,488	1,997	0.994x + 3.717	1.000
B47	SKC	224-PCXR8	566747	01/10/2025	1,000	1,500	2,000	1,004	1,500	1,993	0.996x + 2.230	1.000
B48	SKC	224-PCXR8	566753	01/10/2025	1,000	1,500	2,000	1,002	1,501	1,991	1.000x - 4.116	0.999
B49	SKC	224-PCXR8	566780	01/10/2025	1,000	1,500	2,000	995	1,502	1,990	0.997x - 1.978	1.000
B50	SKC	224-PCXR8	500400	02/10/2025	1,000	1,500	2,000	997	1,503	2,001	1.004x - 10.178	1.000
B51	SKC	224-PCXR8	500363	01/10/2025	1,000	1,500	2,000	1,001	1,502	1,993	0.995x + 2.848	1.000
B52	SKC	224-PCXR8	093186	03/10/2025	1,000	1,500	2,000	996	1,510	1,999	1.005x - 12.252	0.999
B53	SKC	224-PCXR8	707670	03/10/2025	1,000	1,500	2,000	1,002	1,496	2,004	1.003x - 8.791	1.000
B54	SKC	224-PCXR3	509821	03/10/2025	1,000	1,500	2,000	999	1,501	1,995	0.999x - 2.090	1.000
B55	SKC	224-PCXR3	510710	02/10/2025	1,000	1,500	2,000	1,002	1,503	2,006	1.007x - 13.250	0.999
B56	SKC	224-PCXR3	511450	01/10/2025	1,000	1,500	2,000	995	1,505	1,997	1.002x - 7.594	1.000
B57	SKC	224-PCXR3	510798	02/10/2025	1,000	1,500	2,000	998	1,500	1,994	0.999x - 7.163	0.999
B58	SKC	224-PCXR3	509852	03/10/2025	1,000	1,500	2,000	1,002	1,494	1,996	0.993x + 6.485	1.000
B59	SKC	224-PCXR3	509862	01/10/2025	1,000	1,500	2,000	1,006	1,505	1,998	0.996x + 5.117	1.000
B60	SKC	224-PCXR3	512655	02/10/2025	1,000	1,500	2,000	1,004	1,501	2,003	1.010x - 14.223	0.999
B61	SKC	224-PCXR3	503915	03/10/2025	1,000	1,500	2,000	993	1,495	1,994	0.999x - 4.942	1.000
B62	SKC	224-PCXR3	505975	03/10/2025	1,000	1,500	2,000	995	1,500	2,005	1.009x - 16.396	1.000
B63	SKC	224-PCXR3	511432	03/10/2025	1,000	1,500	2,000	996	1,497	1,991	0.998x - 3.171	1.000
B64	SKC	224-PCXR3	508302	03/10/2025	1,000	1,500	2,000	1,008	1,506	1,998	0.992x + 8.667	0.999
B65	SKC	224-PCXR3	508310	03/10/2025	1,000	1,500	2,000	1,006	1,492	2,003	1.000x - 4.355	1.000
B66	SKC	224-PCXR3	509861	03/10/2025	1,000	1,500	2,000	994	1,496	1,994	0.997x - 0.275	1.000
B67	SKC	224-PCXR3	506295	01/10/2025	1,000	1,500	2,000	997	1,505	2,001	1.004x - 10.258	1.000
B68	SKC	224-PCXR3	505872	03/10/2025	1,000	1,500	2,000	998	1,512	1,992	0.999x - 3.554	0.999
B69	SKC	224-PCXR3	508375	01/10/2025	1,000	1,500	2,000	997	1,489	1,996	0.997x - 2.309	1.000
B70	SKC	224-PCXR3	510623	03/10/2025	1,000	1,500	2,000	1,001	1,496	1,991	0.992x + 7.131	1.000
B71	SKC	224-PCXR3	508367	03/10/2025	1,000	1,500	2,000	999	1,498	1,995	0.994x + 6.433	1.000
B72	SKC	224-PCXR3	505977	03/10/2025	1,000	1,500	2,000	996	1,507	1,999	1.003x - 7.490	1.000
B73	SKC	224-PCXR3	512606	03/10/2025	1,000	1,500	2,000	1,004	1,503	2,003	1.001x - 5.285	0.999
B74	SKC	224-PCXR3	505993	01/10/2025	1,000	1,500	2,000	1,006	1,501	1,997	0.993x + 8.232	1.000
B75	SKC	224-PCXR3	509820	02/10/2025	1,000	1,500	2,000	1,005	1,494	1,995	0.997x - 2.979	0.999
B76	SKC	224-PCXR3	509811	02/10/2025	1,000	1,500	2,000	996	1,503	1,999	1.005x - 10.613	1.000
B77	SKC	224-PCXR3	508301	01/10/2025	1,000	1,500	2,000	1,003	1,490	1,998	0.994x + 4.694	1.000
B78	SKC	224-PCXR3	510677	01/10/2025	1,000	1,500	2,000	1,004	1,492	1,997	0.995x + 4.036	1.000
B79	SKC	224-PCXR3	510920	01/10/2025	1,000	1,500	2,000	1,008	1,504	2,006	1.006x - 9.588	0.999

Calibrated by :

(Mr. Adul Dangklom)

Approved by :

(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด

S.P.S. CONSULTING SERVICE CO., LTD.

7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900

Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Personal Pump Calibration Report

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

Environmental Conditions

Temperature : 25 ± 3 °C
Pressure : 1010 ± 15 mmbar

Personal Pump Data				Calibration Data								
No.	Brand	Model	Serial No.	Date	Flow Rate (mL/min)						Value From Calibration Curve	
					Setting			Actual (Q std.)				
					1	2	3	1	2	3	y	R ²
B80	SKC	224-PCXR3	504569	01/10/2025	1,000	1,500	2,000	1,004	1,510	2,004	0.999x - 0.786	0.999
B81	SKC	224-PCXR3	503480	01/10/2025	1,000	1,500	2,000	1,004	1,509	2,002	1.000x - 1.340	0.999
B82	SKC	224-PCXR3	505673	01/10/2025	1,000	1,500	2,000	1,007	1,502	2,003	0.994x+ 9.425	1.000
B83	SKC	224-PCXR3	510785	01/10/2025	1,000	1,500	2,000	997	1,494	1,994	0.999x - 1.727	0.999
B84	SKC	224-PCXR3	508333	01/10/2025	1,000	1,500	2,000	1,000	1,505	2,005	1.002x - 5.217	1.000
B85	SKC	224-PCXR3	505757	03/10/2025	1,000	1,500	2,000	996	1,507	1,998	1.001x - 4.459	1.000
B86	SKC	224-PCXR3	512625	03/10/2025	1,000	1,500	2,000	998	1,506	1,993	0.997x - 4.180	0.999
B87	SKC	224-PCXR3	504324	03/10/2025	1,000	1,500	2,000	1,003	1,510	2,004	1.000x + 1.037	1.000
B88	SKC	224-PCXR3	508307	03/10/2025	1,000	1,500	2,000	995	1,505	1,999	1.007x - 17.485	0.999
B89	SKC	224-PCXR3	509860	03/10/2025	1,000	1,500	2,000	999	1,500	2,001	1.002x - 6.218	1.000
B90	SKC	224-PCXR3	508366	02/11/2025	1,000	1,500	2,000	996	1,505	2,004	1.008x - 15.379	1.000
B91	SKC	224-PCXR3	510919	01/10/2025	1,000	1,500	2,000	1,004	1,504	2,003	0.998x + 1.328	1.000
B92	SKC	224-PCXR3	510987	03/10/2025	1,000	1,500	2,000	997	1,505	2,005	1.009x - 17.601	0.999
B93	SKC	224-PCXR3	509845	03/10/2025	1,000	1,500	2,000	1,006	1,511	2,001	0.998x + 2.138	0.999

Calibrated by :

(Mr. Adul Dangklom)

Approved by :

(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com., www.spscon.com

Rotameter Calibration Report (For Personal Pump High Flow Adjust)

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

Calibration Data

Calibration Data											
Rotameter Data			Calibration Data								
No.	Brand	Model	Date	Flow Rate (ml/min)						Value From Calibration Curve	
				Flow Rate (Reading)			Actual (Q std.)				
				1	2	3	1	2	3	y	R ²
H-B01	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	502.3	998.1	1996.9	0.999x + 2.995	1.000
H-B02	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	501.2	998.9	1998.1	0.998x + 4.159	1.000
H-B03	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	500.4	999.5	2000.3	1.000x - 1.574	0.999
H-B04	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	499.5	1000.6	1999.5	0.999x + 0.880	1.000
H-B05	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	499.1	1001.3	2001.6	1.001x - 7.095	0.999
H-B06	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	499.4	999.8	1994.1	1.000x + 2.760	1.000
H-B07	Dwyer	VFB-65	02/10/2025	500	1,000	2,000	501.0	999.4	2000.7	0.997x + 1.623	1.000
H-B08	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	499.9	1001.7	2002.8	0.998x - 1.887	1.000
H-B09	Dwyer	VFB-65	02/10/2025	500	1,000	2,000	499.3	999.2	1996.6	0.999x + 1.428	0.999
H-B10	Dwyer	VFB-65	03/10/2025	500	1,000	2,000	498.7	1001.1	1997.8	0.996x + 6.669	1.000

Calibrated by :

(Mr. Abdul Dangklom)

Approved by :

(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com., www.spscon.com

Rotameter Calibration Report (For Personal Pump Low Flow Adjust)

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

Calibration Data

Rotameter Data			Calibration Data								
No.	Brand	Model	Date	Flow Rate (mL/min)						Value From Calibration Curve	
				Flow Rate (Reading)			Actual (Q std.)				
				1	2	3	1	2	3	y	R ²
L-B01	Dwyer	VFA-21	01/10/2025	50	100	200	50.7	99.8	199.6	1.000x + 0.067	1.000
L-B02	Dwyer	VFA-21	01/10/2025	50	100	200	50.2	99.4	198.7	1.001x - 1.181	0.999
L-B03	Dwyer	VFA-21	01/10/2025	50	100	200	50.8	99.2	201.5	1.002x + 0.053	1.000
L-B04	Dwyer	VFA-21	01/10/2025	50	100	200	49.7	101.6	200.8	1.001x + 0.344	1.000
L-B05	Dwyer	VFA-21	01/10/2025	50	100	200	50.5	100.4	201.6	0.998x + 0.225	0.999
L-B06	Dwyer	VFA-21	01/10/2025	50	100	200	50.1	100.5	201.8	1.003x - 0.103	1.000
L-B07	Dwyer	VFA-21	02/10/2025	50	100	200	50.6	100.8	201.3	0.998x + 0.877	1.000
L-B08	Dwyer	VFA-21	01/10/2025	50	100	200	49.8	101.3	198.9	1.000x - 0.165	1.000
L-B09	Dwyer	VFA-21	02/10/2025	50	100	200	50.1	99.6	200.7	1.002x - 0.766	0.999
L-B10	Dwyer	VFA-21	03/10/2025	50	100	200	50.9	100.8	201.2	1.003x + 0.694	1.000

Calibrated by :

(Mr. Adul Dangklom)

Approved by :

(Mr. Peera Detudom)



CERTIFICATE No : 25M2254
REFERENCE No : 76365-1

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : DIGITAL BALANCE
MANUFACTURER : METTLER TOLEDO
MODEL : XS105DU
SERIAL No : 1126422905
ID No : BA05/50
CONDITION AS RECEIVED : USED ITEM
SUBMITTED BY : S.P.S. CONSULTING SERVICE CO., LTD.
7 SOI PHAHOLYOTHIN 24, PHAHOLYOTHIN RD.,
JOMPOL, CHATUCHAK, BANGKOK 10900

CALIBRATED BY : ATSAWIN Y.

CALIBRATION DATE : 07-Mar-25

APPROVED BY : 
PONGSAK J.

ISSUED DATE : 13-Mar-25

RECEIVED DATE : 07-Mar-25

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.





CERTIFICATE No : 25M2254

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : DIGITAL BALANCE MODEL : XS105DU
MANUFACTURER : METTLER TOLEDO S/N : 1126422905
ID No : BA05/50 RECEIVED DATE : 07-Mar-25
AIR PRESSURE : 1009mbar \pm 1mbar CALIBRATION DATE : 07-Mar-25
AMBIENT TEMPERATURE : 24°C \pm 1°C RELATIVE HUMIDITY : 54 %RH \pm 10 % RH

CONDITION OF THIS RESULTS OF CALIBRATION

1. THIS INSTRUMENT WAS CALIBRATED BY ACCORDING TO UKAS LAB 14 EDITION 6:2019 BY USING KNOWN WEIGHT STANDARD WEIGHT. THE BALANCE WAS NOT ADJUSTED BEFORE CALIBRATION. THE BALANCE HAS NO ZERO TRACKING FUNCTION. REPEATABILITY WAS MEASURED BY USING 10 REPEATED MEASUREMENTS. LINEARITY WAS MEASURED COVERING 10 POINTS, EVENLY SPREAD OVER THE RANGE. THE INSTRUMENT WAS SET ZERO BEFORE PERFORMING THE LINEARITY TEST. OFF-CENTER LOADING WAS MEASURED BY USING STANDARD WEIGHTS PLACED ON THE PAN AND MOVED TO VARIOUS POSITIONS ON THE PAN.

2. REFERENCE STANDARD INSTRUMENTS :-

INSTRUMENT	MODEL	SERIAL No	CERTIFICATE No	DUE DATE
1) STANDARD WEIGHT SET	E2	QK-I-151	C02250116	28-Jan-27
2) STANDARD WEIGHT	E2	15843	C02250117	29-Jan-27

3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.

4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.

5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-
- NATIONAL INSTITUTE OF METROLOGY (THAILAND)

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT

1. ZERO SETTING FUNCTION : NORMAL


2. TARE FUNCTION : NORMAL

3. REPEATABILITY OF READING AT 120 g WAS 0.000055 g

4. DEPARTURE FROM NOMINAL VALUE/ LINEARITY

NOMINAL VALUE (g)	BALANCE READING (g)	CORRECTION (g)	UNCERTAINTY (\pm g)
0.00	0.00000	0.00000	0.000065
0.02	0.01999	0.00001	0.000065
0.10	0.10001	-0.00001	0.000066
0.20	0.20001	-0.00001	0.000066
0.50	0.50002	-0.00002	0.000065
1.00	1.00003	-0.00003	0.000066
2.00	2.00001	-0.00001	0.000067
5.00	5.00002	-0.00002	0.000068
10.00	10.00000	0.00000	0.000070
20.00	20.00004	-0.00004	0.000078
50.00	50.00000	0.00000	0.00013
100.00	100.0001	-0.0001	0.00019
120.00	120.0002	-0.0002	0.00022

5. OFF CENTER LOADING ERROR



POINT	READING (g)
1	50.0000
2	50.0000
3	50.0000
4	50.0000
5	50.0000
OFF-CENTER LOADING	0.0000

NOTE: THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA
THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR $k=2$, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.

END OF CALIBRATION REPORT





MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

Customer : <u>S.P.S.Consulting Service Co.,Ltd</u>	Date Tested: <u>December 18, 2025</u>	
	Recommendation Recertification	
Address : <u>7 Soi Phaholyothin 24</u>	Period <u>6</u> Months	
<u>Paholyothin Road</u>	Recertification Due: <u>June 28, 2026</u>	
<u>Jompol Chatuchak, Bangkok 10900</u>	Date Last Certified: <u>July 1, 2025</u>	
User Name: <u>K.Phenpha Viphasathawat</u>	Visit Number: <u>2 of 2</u>	
Phone: <u>083-9269252</u>	PerkinElmer Phone: <u>02-719-6420 ext 206</u>	
Fax: <u>02-513-4221</u>	PerkinElmer Fax: <u>02-318-5597</u>	

CONFIGURATION TESTED		ACCESSORIES/COMPONENT NOT INCLUDED
MODEL	SERIAL NUMBER	
<u>OPTIMA 5300DV</u>	<u>077C7042401</u>	
TESTED EQUIPMENT	CALIBRATION NUMBER	EXPIRATION
<u>IPV Methods</u>		
TEST STANDARD USED	PART NUMBER	EXPIRATION DATE
<u>Multielement Standard</u>	<u>N069-1579</u>	<u>November 30, 2026</u>
<u>Wavecal Solution</u>	<u>N058-2152</u>	<u>July 30, 2026</u>
<u>VIS Wavecal solution</u>	<u>N930-2946</u>	<u>August 30, 2026</u>
<u>Instrument Cal. STD4</u>	<u>N930-0221</u>	<u>November 30, 2026</u>
CUSTOMER SUPPLIED	COMMENTS	CUSTOMER INITIALS
<u>2 % HNO3</u>		
<u>10 % HNO3</u>		



MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

SERIAL NUMBER 077C7042401
DATE TESTED December 18, 2025
1. MECHANICAL CHECKS

A. Inspect and clean all fans and filters.

☐ OK

B. Inspect and replace as necessary, all torch components including the RF coil.

☐ OK

C. Inspect all tubing for sign of clacking or leaking.

☐ OK

D. Adjust water and gas pressure regulator settings.

☐ OK

E. Inspect and leak check pneumatics drawers.

☐ OK

F. Clean the exterior of the instrument.

☐ OK

2. OPTICAL CHECKS

A. Inspect and clean all optical components.

☐ OK

B. As required, check and replace all purgefilters.

☐ OK

C. Recheck optical alignment.

☐ OK

3. COOLING SYSTEM CHECKS

A. Perform preventive maintenance on chiller.

☐ OK

B. Flush out the chiller every year.

☐ N/A

4. PERFORMANCE CHECKS

A. Torch View Alignment.

☐ OK

B. Wavelength Calibration.

☐ OK



MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

SERIAL NUMBER : 077C7042401			DATE TESTED : December 18, 2025		
PARAMETER	SPECIFICATION			FINAL VALUE	
Spectral Resolution : UV	As	193.696 nm	≤ 0.007	0.00530	
	Ni	231.604 nm	≤ 0.008	0.00708	
	Ni	341.476 nm	≤ 0.012	0.00776	
Spectral Resolution : VIS	La	408.672 nm	≤ 0.020	0.01614	
	Ba	455.403 nm	≤ 0.025	0.02377	
Precision					
	As	193.656 nm	% RSD < 1.0	0.67	%
	Zn	213.856 nm	% RSD < 1.0	0.62	%
	Mn	257.610 nm	% RSD < 1.0	0.88	%
	La	379.478 nm	% RSD < 1.0	0.63	%
	Ba	455.403 nm	% RSD < 1.0	0.65	%
	Ba	493.408 nm	% RSD < 1.0	0.45	%
Detection Limits : Axial	Tl	190.080 nm	3(sd)	3.21	ppb
	As	193.696 nm	3(sd)	6.06	ppb
	Pb	220.353 nm	3(sd)	0.92	ppb
Detection Limits : Radial	As	193.696 nm	3(sd)	17.35	ppb
	Zn	213.856 nm	3(sd)	1.79	ppb
	Mn	257.610 nm	3(sd)	0.18	ppb
	La	379.478 nm	3(sd)	0.76	ppb
	Ba	455.403 nm	3(sd)	0.11	ppb
	Ba	493.408 nm	3(sd)	0.56	ppb
BEC : Axial (IB X 500)/(IS-IB)	Cd	226.502 nm	≤ 150 ppb	40.52	
BEC : Radial (IB X 1000)/(IS-IB)	Mn	257.610 nm	≤ 45 ppb	42.33	



MAINTENANCE AND TEST CERTIFICATE MODEL

OPTIMA 5300DV

SERIAL NUMBER 077C7042401**DATE TESTED** December 18, 2025**Remarks :**

Commissioning follow as commissioning performance sheets.

This is to certify that the above tests have been performed and the configuration tested



meets



does not meet

the PerkinElmer Specifications listed on this certificate.

This certificate does not modify PerkinElmer's standard terms and condition of sale,
including warranty terms.

Service Department PerkinElmer Ltd.

Authorized Representative:

(Wiphan Promlumda)

Service Engineer



Certificate of Calibration

Aquion : Anion (ID#894)

This certificate is to verify that instrument below are calibrated
by Archemica Lab Co.,Ltd.

AQUION S/N : 190840059

AS-DV S/N : 190915235

for

S.P.S. Consulting Service Co., Ltd.



Operator Signature : _____

Date : Jun 6, 2025

(Mr. Teerapat Boonla)

Application Chemist

ระดับเสียงในสถานประกอบการ



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-68/0220

MTC No. EEL. BP. 44/0268

CALIBRATION CERTIFICATE

Submitted by : S.P.S.Consulting Service Co.,Ltd.

Address : 7 Soi Phaholyothin 24, Phaholyothin Road, Jompol, Chatuchak, Bangkok 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

Instrument Calibrated :

Description : Sound Calibrator

Manufacturer : ACO

Model : 2127

Serial No. : 130006

Ambient Environment

Temperature : $(23 \pm 3) ^\circ\text{C}$

Relative Humidity : $(50 \pm 15) \%$

Ambient Pressure : $(101.325 \pm 1.500) \text{ kPa}$

- Standards used :
1. Digital Function Synthesizer NF Electronic DF-193A S/N 122037.
 2. Measuring Amplifier Bruel&Kjaer 2636 S/N 1537484.
 3. Programmable Attenuator Tamagawa TPA-303A S/N OF 2214.
 4. Digital Multimeter Agilent 34401A S/N MY44005560.
 5. Pressure Transmitter Vaisala PTB202AD S/N T0650001.
 6. Audio Analyzer Panasonic VP-7722A S/N 041477D122.
 7. Condenser Microphone B&K 4180 S/N 2889871.

Calibration Procedure: CP-102-04 based on IEC 60942-2003; The sound pressure level generated by sound calibrator under test shall be measured by standard microphone using an insert voltage technique.

This instrument has been calibrated against standards maintained at Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

Date of Receipt : 19 Feb. 2025

Date of Calibration : 21 Feb. 2025

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5

Head Office

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9036
Fax. (66) 0 2577 9009

Office/Laboratory

668 Mu 2 Tambon Bangpoomai, Amphoe Muang Samutprakan,
Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
(66) 08 3219 9440
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office

196 Phahonyothin Road, Ladyao, Chatuchak,
Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
(66) 08 1889 6827

THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-68/0220

MTC No. EEL. BP. 44/0268

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.

Nominal Output of Unit Under Test = 94 dB re 20 μ Pa at 1000 Hz

Acoustic Output in dB re 20 μ Pa, Corrected to Reference Conditions: 101.325 kPa, 23.0 °C and 50 %RH.

1. Sound Pressure Level

Standard Microphone Type	Measured Sound Pressure Level (dB)	Deviated value (dB)	Uncertainty (dB)	Tolerance limit IEC60942:2003 Class 1
1/2 inch Bruel&Kjaer 4180	93.81	-0.19	± 0.10	± 0.40 dB

2. Frequency

Standard Microphone Type	Measured Frequency (Hz)	Deviated value (Hz)	Uncertainty (Hz)	Tolerance limit IEC60942:2003 Class 1
1/2 inch Bruel&Kjaer 4180	999.9	-0.1	± 1.5	$\pm 1.0\%$

3. Total Distortion

Standard Microphone Type	Measured Total Distortion (%)	Uncertainty (%)	Tolerance limit IEC60942:2003 Class 1
1/2 inch Bruel&Kjaer 4180	0.95	± 0.50	$\pm 3.0\%$

Note : 1. No adjustment.

2. The calibrator pressure correction was not included.

3. The microphone volume correction was not included.

Calibrated by :


.....
(Mr. Weerachai Deechaiyae)

Approved by :


.....
(Mr. Prawate Kluaypa)
Director

Date of Calibration : 21 Feb. 2025

Date of Issue : 24 Feb. 2025

Electrical and Electronic Standards Laboratory
Industrial Metrology and Testing Service Centre

Ref : 2011268021900739001

End of Certificate

2 / 2

The results relate only to the items tested/calibrated or value assigned.
Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5

Head Office

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9036
Fax. (66) 0 2577 9009

Office/Laboratory

668 Mu 2 Tambon Bangpoornai, Amphoe Muang Samutprakan,
Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
(66) 08 3219 9440
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office

196 Phahonyothin Road, Ladyao, Chatuchak,
Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
(66) 08 1889 6827



Noise B_529/25

Sound Level Meter Calibration Report

Acoustic Calibrator Data

Brand	ACO	Number	AC 03/56
Model	2127	Serial No.	130006
Calibration Range	94 dB, 1000 Hz	Last Calibration	21 February 2025
		Due Date	21 February 2026

Calibration Data

Sound Level Meter Data				Calibration Data		
SLM No.	Brand	Model	Serial No.	Date	Actual Reading [dB]	
					Before Adjustment	After Adjustment
ACO-B41	ACO	6236	00192032	03 November 2025	93.9	93.9
ACO-B43	ACO	6236	00192034	03 November 2025	93.8	93.9
ACO-R51	ACO	6236	00192063	03 November 2025	93.9	93.9
ACO-R52	ACO	6236	00192064	03 November 2025	93.9	93.9
Acoustic Certified Value : Thailand Institute of Scientific and Technological Research (TISTR)					93.81 ± 0.10 dB	

Calibrated by :

(Mr. Adul Dangklom)

Approved by :

(Mr. Peera Detudom)

ปริมาณเสียงสะสมแบบติดตัวบุคคล



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-68/0514

MTC No. EEL. BP. 34/0868

CALIBRATION CERTIFICATE

Submitted by : S.P.S.Consulting Service Co.,Ltd.

Address : 7 Soi Phaholyothin 24, Phaholyothin Road, Jompol, Chatuchak, Bangkok 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre.
: Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

Instrument Calibrated :

Description : Sound Calibrator

Manufacturer : SVANTEK

Model : SV34

Serial No. : 83820

Ambient Environment

Temperature : $(23 \pm 3) ^\circ\text{C}$

Relative Humidity : $(50 \pm 15) \%$

Ambient Pressure : $(101.325 \pm 1.500) \text{ kPa}$

Standards used : 1. Digital Function Synthesizer NF Electronic DF-193A S/N 122037.
2. Measuring Amplifier Bruel&Kjaer 2636 S/N 1537484.
3. Programmable Attenuator Tamagawa TPA-303A S/N OF 2214.
4. Digital Multimeter Agilent 34401A S/N MY44005560.
5. Pressure Transmitter Vaisala PTB202AD S/N T0650001.
6. Audio Analyzer Panasonic VP-7722A S/N 041477D122.
7. Condenser Microphone B&K 4180 S/N 2633526.

Calibration Procedure: CP-102-04 based on IEC 60942-2003; The sound pressure level generated by sound calibrator under test shall be measured by standard microphone using an insert voltage technique.

This instrument has been calibrated against standards maintained at Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

Date of Receipt : 14 Aug. 2025

Date of Calibration : 22 Aug. 2025

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5

Head Office

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand

Tel. (66) 0 2577 9036

Fax. (66) 0 2577 9009

Office/Laboratory

668 Mu 2 Tambon Bangpoomai, Amphoe Muang Samutprakan,
Changwat Samutprakan 10280, Thailand

Tel. (66) 0 2323 1672-80 ext. 115, 116

(66) 08 3219 9440

E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office

196 Phahonyothin Road, Ladyao, Chatuchak,
Bangkok 10900, Thailand

Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217

(66) 08 1889 6827

THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-68/0514

MTC No. EEL. BP. 34/0868

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.

Nominal Output of Unit Under Test = 114 dB re 20 μ Pa at 1000 Hz

Acoustic Output in dB re 20 μ Pa, Corrected to Reference Conditions: 101.325 kPa, 23.0 °C and 50 %RH.

1. Sound Pressure Level

Standard Microphone Type	Measured Sound Pressure Level (dB)	Deviated value (dB)	Uncertainty (dB)	Tolerance limit IEC60942:2003 Class 2
1/2 inch Bruel&Kjaer 4180	114.02	0.02	± 0.10	± 0.75 dB

2. Frequency

Standard Microphone Type	Measured Frequency (Hz)	Deviated value (Hz)	Uncertainty (Hz)	Tolerance limit IEC60942:2003 Class 2
1/2 inch Bruel&Kjaer 4180	1000.0	0.0	± 1.5	$\pm 2.0\%$

3. Total Distortion

Standard Microphone Type	Measured Total Distortion (%)	Uncertainty (%)	Tolerance limit IEC60942:2003 Class 2
1/2 inch Bruel&Kjaer 4180	0.21	± 0.50	$\pm 4.0\%$

Note : 1. No adjustment.

2. The calibrator pressure correction was not included.

3. The microphone volume correction was not included.

Calibrated by :

.....

(Mr. Weerachai Deechaiyae)

Approved by :

.....

(Mr. Prawate Kluaypa)

Director

Electrical and Electronic Standards Laboratory

Industrial Metrology and Testing Service Centre

Date of Calibration : 22 Aug. 2025

Date of Issue : 25 Aug. 2025

Ref : 2011268081403169011

End of Certificate

2 / 2

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5

Head Office

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9036
Fax. (66) 0 2577 9009

Office/Laboratory

668 Mu 2 Tambon Bangpoomai, Amphoe Muang Samutprakan,
Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
(66) 08 3219 9440
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office

196 Phahonyothin Road, Ladyao, Chatuchak,
Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
(66) 08 1889 6827



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Noise Dose B_529_1/25

Noise Dose Meter Calibration Report

Acoustic Calibrator Data

Brand SVANTEK

Number SV 03/60

Model SV34

Serial No. 83820

Calibration Range 114 dB, 1000 Hz

Last Calibration 22 August 2025

Due Date 22 August 2026

Calibration Data

Sound Level Meter Data

Calibration Data

SLM No.	Brand	Model	Serial No.	Date	Actual Reading [dB]	
					Before Adjustment	After Adjustment
NMD-B11	SVANTEK	SV-104IS	80831	03 November 2025	114.0	114.0
NMD-B13	SVANTEK	SV-104IS	80834	03 November 2025	113.9	114.0
NMD-B15	SVANTEK	SV-104IS	80880	03 November 2025	114.0	114.0
NMD-B19	SVANTEK	SV-104IS	106124	03 November 2025	114.1	114.0

Acoustic Certified Value : Thailand Institute of Scientific and Technological Research (TISTR)

114.02 ± 0.10 dB

Calibrated by :

(Mr. Adul Dangklom)

Approved by :

(Mr. Peera Detudom)

ระดับความร้อนในสถานประกอบการ



ID LINE : IEC17025



Certificate of Calibration

Certificate Number : SPR24100363-5

Page : 1 of 3

Customer : S.P.S. CONSULTING SERVICE CO., LTD.

7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,
Bangkok 10900

Equipment Name : Area Heat Stress Monitor

Manufacturer : Quest Technologies

Model : QUESTemp 34

Serial Number : TEH060047

ID. Number : B05

Environmental Conditions

Ambient Temperature : $23^{\circ}\text{C} \pm 2^{\circ}\text{C}$

Received Date : 21 Oct 2024

Relative Humidity : $50\% \pm 15\%$

Calibration Date : 21 Oct 2024

Location of Calibration : In-Lab

Recommend Due Date : 21 Oct 2025

Calibration Procedure : SP-CPT-04-13

Date of Issue : 22 Oct 2024

Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs.

The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Surasak Ritthikaew

Calibration Officer

Approved by :

(Mr.Prayoon Topart)

Authorized Signatory



ID LINE : IEC17025



Calibration Report

Certificate Number : SPR24100363-5

Page : 2 of 3

Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Humidity Chamber	TH-80S	N/A	SPR24020149-7	23 Feb 2025
THERMO-HYGROMETER	5020A	A47046	QR24-0167	26 Jan 2025

Traceability

This certification is traceable to the International System of Unit maintained at :

SP Metrology - SP Metrology system (Thailand) Co.Ltd.

Quality Reborn Co., Ltd



ID LINE : IEC17025



Result of Calibration

Certificate Number : SPR24100363-5

Page : 3 of 3

Temperature Accuracy in the Measurement. (WET)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty (±)
30.0	30.014	30.1	0.086	0.20
35.0	35.012	35.1	0.088	0.20
40.0	40.017	40.1	0.083	0.20

Temperature Accuracy in the Measurement. (DRY)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty (±)
30.0	30.014	30.1	0.086	0.20
35.0	35.012	35.1	0.088	0.20
40.0	40.017	40.1	0.083	0.20

Temperature Accuracy in the Measurement. (GLOBE)

Unit : °C

Humidity Setting	Standard Reading	UUC Reading	Error	Uncertainty (±)
30.0	30.014	30.2	0.186	0.20
35.0	35.012	35.2	0.188	0.20
40.0	40.017	40.2	0.183	0.20

Note :

The result of calibration was found accurate as show on date and place of calibration only.
This Certificate is not certified for any commercial transaction.

Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor $k = 2$, providing a level of confidence approximately 95%.

- End of Certificate -



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72. Fax : (662) 513-4221. E-mail : sale@spscon.com., www.spscon.com

Heat B_289_1

Heat Stress WBGT Meter Verification Report			
Verification Data			
Heat Stress WBGT Meter No.	: B05	Verification Date	: 30 June 2025
Brand	: Quest Technologies	Ambient Temp.	: 24.5 °C
Model	: QUESTemp [®] 34	Barometric Pressure	: 1011 mmbar
Serial No.	: TEH060047	Relative Humidity	: 49 %
Verification Module (Electronic Sensor Check) :			
Verification Module No. : 21 WB = 12.5 °C, DB = 47.1 °C, G = 69.3 °C			
Result of Verification : Without Adjustment			
Wet Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
12.5	12.6	-0.1	± 0.5
Dry Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
47.1	47.0	0.1	± 0.5
Globe Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
69.3	69.3	0.0	± 0.5
UUC* = UNIT UNDER CALIBRATION			

Verified by :



(Mr. Adul Dangklom)

Approved by :



(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com., www.spscon.com

Heat B_470_1

Heat Stress WBGT Meter Verification Report			
Verification Data			
Heat Stress WBGT Meter No.	: B05	Verification Date	: 08 October 2025
Brand	: Quest Technologies	Ambient Temp.	: 24.5 °C
Model	: QUESTemp 34	Barometric Pressure	: 1011 mmbar
Serial No.	: TEH060047	Relative Humidity	: 49 %
Verification Module (Electronic Sensor Check) :			
Verification Module No. : 21 WB = 12.5 °C, DB = 47.1 °C, G = 69.3 °C			
Result of Verification : Without Adjustment			
Wet Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
12.5	12.7	-0.2	± 0.5
Dry Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
47.1	47.2	-0.1	± 0.5
Globe Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
69.3	69.2	0.1	± 0.5
UUC* = UNIT UNDER CALIBRATION			

Verified by :



(Mr. Adul Dangklom)

Approved by :



(Mr. Peera Detudom)



Certificate of Calibration

Certificate Number : SPR25030358-1

Page : 1 of 3

Customer : S.P.S. CONSULTING SERVICE CO., LTD.

7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,
Bangkok 10900

Equipment Name : Area Heat Stress Monitor

Manufacturer : Quest Technologies

Model : QUESTemp 34

Serial Number : TEG040059

ID. Number : B07

Environmental Conditions

Ambient Temperature : $23^{\circ}\text{C} \pm 2^{\circ}\text{C}$

Relative Humidity : $50\% \pm 15\%$

Location of Calibration : In-Lab

Calibration Procedure : SP-CPT-04-13

Received Date : 19 Mar 2025

Calibration Date : 22 Mar 2025

Recommend Due Date : 22 Mar 2026

Date of Issue : 23 Mar 2025

Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs.

The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr. Navaporn Uengseng

Calibration Officer

Approved



Calibration Report

Certificate Number : SPR25030358-1

Page : 2 of 3

Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Humidity Chamber	TH-80S	N/A	SPR25010173-14	30 Jan 2026
THERMO-HYGROMETER	5020A	A47046	TMU2500342	29 Jan 2026

Traceability

This certification is traceable to the International System of Unit maintained at :

SP Metrology - SP Metrology system (Thailand) Co.Ltd.

NA - NA Caltechnologies Co., Ltd.



Result of Calibration

Certificate Number : SPR25030358-1

Page : 3 of 3

Temperature Accuracy in the Measurement. (WET)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty (±)
30.0	30.015	29.9	-0.115	0.20
35.0	35.012	34.9	-0.112	0.20
40.0	40.016	39.9	-0.116	0.20

Temperature Accuracy in the Measurement. (DRY)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty (±)
30.0	30.015	29.8	-0.215	0.20
35.0	35.012	34.8	-0.212	0.20
40.0	40.016	39.8	-0.216	0.20

Temperature Accuracy in the Measurement. (GLOBE)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty (±)
30.0	30.015	29.9	-0.115	0.20
35.0	35.012	34.9	-0.112	0.20
40.0	40.016	39.9	-0.116	0.20

Note:

The result of calibration was found accurate as show on date and place of calibration only.
This Certificate is not certified for any commercial transaction.

Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor $k = 2$, providing a level of confidence approximately 95%.

- End of Certificate -



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72 Fax : (662) 513-4221 E-mail : sale@spscon.com., www.spscon.com

Heat B_289_2

Heat Stress WBGT Meter Verification Report			
Verification Data			
Heat Stress WBGT Meter No.	: B07	Verification Date	: 30 June 2025
Brand	: Quest Technologies	Ambient Temp.	: 24.5 °C
Model	: QUESTemp ^o 34	Barometric Pressure	: 1011 mmbar
Serial No.	: TEG040059	Relative Humidity	: 49 %
Verification Module (Electronic Sensor Check) :			
Verification Module No. : 21 WB = 12.5 °C, DB = 47.1 °C, G = 69.3 °C			
Result of Verification : Without Adjustment			
Wet Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
12.5	12.7	-0.2	± 0.5
Dry Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
47.1	47.1	0.0	± 0.5
Globe Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
69.3	69.4	-0.1	± 0.5
UUC* = UNIT UNDER CALIBRATION			

Verified by :

(Mr.Adul Dangklom)

Approved by :

(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72. Fax : (662) 513-4221. E-mail : sale@spscon.com., www.spscon.com

Heat B_470_2

Heat Stress WBGT Meter Verification Report			
Verification Data			
Heat Stress WBGT Meter No.	: 807	Verification Date	: 08 October 2025
Brand	: Quest Technologies	Ambient Temp.	: 24.5 °C
Model	: QUESTemp 34	Barometric Pressure	: 1011 mmbar
Serial No.	: TEG040059	Relative Humidity	: 49 %
Verification Module (Electronic Sensor Check) :			
Verification Module No. : 21 WB = 12.5 °C, DB = 47.1 °C, G = 69.3 °C			
Result of Verification : Without Adjustment			
Wet Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
12.5	12.6	-0.1	± 0.5
Dry Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
47.1	46.9	0.2	± 0.5
Globe Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
69.3	69.2	0.1	± 0.5
UUC* = UNIT UNDER CALIBRATION			

Verified by :

(Mr.Adul Dangklom)

Approved by :

(Mr. Peera Detudom)



ID LINE : IEC17025



Certificate of Calibration

Certificate Number : SPR24100363-3

Page : 1 of 3

Customer : S.P.S. CONSULTING SERVICE CO., LTD.

7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,
Bangkok 10900

Equipment Name : Area Heat Stress Monitor

Manufacturer : Quest Technologies

Model : QUESTemp 34

Serial Number : TEL080034

ID. Number : B11

Environmental Conditions

Ambient Temperature : $23\text{ }^{\circ}\text{C} \pm 2\text{ }^{\circ}\text{C}$

Received Date : 21 Oct 2024

Relative Humidity : $50\% \pm 15\%$

Calibration Date : 21 Oct 2024

Location of Calibration : In-Lab

Recommend Due Date : 21 Oct 2025

Calibration Procedure : SP-CPT-04-13

Date of Issue : 22 Oct 2024

Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs.

The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Chatchai Kittisopha

Calibration Officer

Approved by :

(Mr.Prayoon Topart)

Authorized Signatory



ID LINE : IEC17025



Calibration Report

Certificate Number : SPR24100363-3

Page : 2 of 3

Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Humidity Chamber	TH-80S	N/A	SPR24020149-7	23 Feb 2025
THERMO-HYGROMETER	5020A	A47046	QR24-0167	26 Jan 2025

Traceability

This certification is traceable to the International System of Unit maintained at :
SP Metrology - SP Metrology system (Thailand) Co.Ltd.

Quality Reborn Co., Ltd



ID LINE : IEC17025



Result of Calibration

Certificate Number : SPR24100363-3

Page : 3 of 3

Temperature Accuracy in the Measurement. (WET)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty (±)
30.0	30.021	30.3	0.279	0.20
35.0	35.018	35.3	0.282	0.20
40.0	40.019	40.3	0.281	0.20

Temperature Accuracy in the Measurement. (DRY)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty (±)
30.0	30.021	30.2	0.179	0.20
35.0	35.018	35.2	0.182	0.20
40.0	40.019	40.2	0.181	0.20

Temperature Accuracy in the Measurement. (GLOBE)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty (±)
30.0	30.021	30.2	0.179	0.20
35.0	35.018	35.2	0.182	0.20
40.0	40.019	40.2	0.181	0.20

Note :

The result of calibration was found accurate as show on date and place of calibration only.
This Certificate is not certified for any commercial transaction.

Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor $k = 2$, providing a level of confidence approximately 95%.

- End of Certificate -



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com., www.spscon.com

Heat B_289_3

Heat Stress WBGT Meter Verification Report			
Verification Data			
Heat Stress WBGT Meter No.	: B11	Verification Date	: 30 June 2025
Brand	: Quest Technologies	Ambient Temp.	: 24.5 °C
Model	: QUESTemp ^o 34	Barometric Pressure	: 1011 mmbar
Serial No.	: TEL080034	Relative Humidity	: 49 %
Verification Module (Electronic Sensor Check) :			
Verification Module No. : 21 WB = 12.5 °C, DB = 47.1 °C, G = 69.3 °C			
Result of Verification : Without Adjustment			
Wet Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
12.5	12.6	-0.1	± 0.5
Dry Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
47.1	46.9	0.2	± 0.5
Globe Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
69.3	69.2	0.1	± 0.5
UUC* = UNIT UNDER CALIBRATION			

Verified by :

(Mr.Adul Dangklom)

Approved by :

(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com., www.spscon.com

Heat B_470_3

Heat Stress WBGT Meter Verification Report			
Verification Data			
Heat Stress WBGT Meter No.	: B11	Verification Date	: 08 October 2025
Brand	: Quest Technologies	Ambient Temp.	: 24.5 °C
Model	: QUESTemp 34	Barometric Pressure	: 1011 mmbar
Serial No.	: TEL080034	Relative Humidity	: 49 %
Verification Module (Electronic Sensor Check) :			
Verification Module No. : 21 WB = 12.5 °C, DB = 47.1 °C, G = 69.3 °C			
Result of Verification : Without Adjustment			
Wet Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
12.5	12.6	-0.1	± 0.5
Dry Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
47.1	47.0	0.1	± 0.5
Globe Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
69.3	69.5	-0.2	± 0.5
UUC* = UNIT UNDER CALIBRATION			

Verified by :

(Mr. Adul Dangklom)

Approved by :

(Mr. Peera Detudom)



ID LINE : IEC17025



Certificate of Calibration

Certificate Number : SPR24100363-4

Page : 1 of 3

Customer : S.P.S. CONSULTING SERVICE CO., LTD.

7 Soi Phaholyothin 24 Phaholyothin Road., Jompol, Chatuchak,
Bangkok 10900

Equipment Name : Area Heat Stress Monitor

Manufacturer : Quest Technologies

Model : QUESTemp 32

Serial Number : TPA100010

ID. Number : B12

Environmental Conditions

Ambient Temperature : $23^{\circ}\text{C} \pm 2^{\circ}\text{C}$

Received Date : 21 Oct 2024

Relative Humidity : $50\% \pm 15\%$

Calibration Date : 21 Oct 2024

Location of Calibration : In-Lab

Recommend Due Date : 21 Oct 2025

Calibration Procedure : SP-CPT-04-13

Date of Issue : 22 Oct 2024

Method of Calibration

This certifies that the above instrument was calibrated in compliance with the calibration system requirement of ISO/IEC 17025:2017 in accordance with reference procedure. Standards used to perform this calibration are certified by to NIST or equivalent, National metrology institute, Natural physical constants, consensus standards. The result reported herein apply only to the calibration of the item described above as received. Our decision rule is to contact the customer if the item pass and fail calibration when the results include the uncertainties and the customer must determine if the results meets their needs.

The calibration certificate shall not be reproduced except in full, without written approval of SP Metrology System (Thailand).

Calibrated by : Mr.Surasak Ritthikaew

Calibration Officer

Approved by :

(Mr.Prayoon Topart)

Authorized Signatory



ID LINE : IEC17025



Calibration Report

Certificate Number : SPR24100363-4

Page : 2 of 3

Reference Standards

Equipment Name	Model	Serial No.	Certificate No.	Due. Date
Humidity Chamber	TH-80S	N/A	SPR24020149-7	23 Feb 2025
THERMO-HYGROMETER	5020A	A47046	QR24-0167	26 Jan 2025

Traceability

This certification is traceable to the International System of Unit maintained at :

SP Metrology - SP Metrology system (Thailand) Co.Ltd.

Quality Reborn Co., Ltd



ID LINE : IEC17025



Result of Calibration

Certificate Number : SPR24100363-4

Page : 3 of 3

Temperature Accuracy in the Measurement. (WET)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty (±)
30.0	30.019	30.2	0.181	0.20
35.0	35.017	35.2	0.183	0.20
40.0	40.019	40.2	0.181	0.20

Temperature Accuracy in the Measurement. (DRY)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty (±)
30.0	30.019	30.2	0.181	0.20
35.0	35.017	35.2	0.183	0.20
40.0	40.019	40.2	0.181	0.20

Temperature Accuracy in the Measurement. (GLOBE)

Unit : °C

Temperature Setting	Standard Reading	UUC Reading	Error	Uncertainty (±)
30.0	30.019	30.2	0.181	0.20
35.0	35.017	35.2	0.183	0.20
40.0	40.019	40.2	0.181	0.20

Note :

The result of calibration was found accurate as show on date and place of calibration only.
This Certificate is not certified for any commercial transaction.

Measurement Uncertainty

The reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor $k = 2$, providing a level of confidence approximately 95%.

- End of Certificate -



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com., www.spscon.com

Heat B_289_4


Heat Stress WBGT Meter Verification Report			
Verification Data			
Heat Stress WBGT Meter No.	: B12	Verification Date	: 30 June 2025
Brand	: Quest Technologies	Ambient Temp.	: 24.5 °C
Model	: QUESTemp ^o 32	Barometric Pressure	: 1011 mmbar
Serial No.	: TPA100010	Relative Humidity	: 49 %
Verification Module (Electronic Sensor Check) :			
Verification Module No. : 21 WB = 12.5 °C, DB = 47.1 °C, G = 69.3 °C			
Result of Verification : Without Adjustment			
Wet Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
12.5	12.5	0.0	± 0.5
Dry Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
47.1	47.2	-0.1	± 0.5
Globe Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
69.3	69.2	0.1	± 0.5
UUC* = UNIT UNDER CALIBRATION			

Verified by :



(Mr. Adul Dangklom)

Approved by :



(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221. E-mail : sale@spscon.com., www.spscon.com

Heat B_470_4

Heat Stress WBGT Meter Verification Report			
Verification Data			
Heat Stress WBGT Meter No.	: B12	Verification Date	: 08 October 2025
Brand	: Quest Technologies	Ambient Temp.	: 24.5 °C
Model	: QUESTemp 32	Barometric Pressure	: 1011 mmbar
Serial No.	: TPA100010	Relative Humidity	: 49 %
Verification Module (Electronic Sensor Check) :			
Verification Module No. : 21 WB = 12.5 °C, DB = 47.1 °C, G = 69.3 °C			
Result of Verification : Without Adjustment			
Wet Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
12.5	12.4	0.1	± 0.5
Dry Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
47.1	47.2	-0.1	± 0.5
Globe Probe Temperature Measurement			
Verification Module Reading (°C)	UUC* Reading (°C)	Correction (°C)	Tolerance Limit (°C)
69.3	69.1	0.2	± 0.5
UUC* = UNIT UNDER CALIBRATION			

Verified by :



(Mr.Adul Dangklom)

Approved by :



กลิ่นในสถานประกอบการ



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด

S.P.S. CONSULTING SERVICE CO., LTD.

7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900

7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900

Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Personal Pump Calibration Report

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

Environmental Conditions

Temperature : 25 ± 3 °C
Pressure : 1010 ± 15 mmbar

Personal Pump Data				Calibration Data								
No.	Brand	Model	Serial No.	Date	Flow Rate (mL/min)						Value From Calibration Curve	
					Setting			Actual (Q std.)				
					1	2	3	1	2	3	y	R ²
B80	SKC	224-PCXR3	504569	01/10/2025	1,000	1,500	2,000	1,004	1,510	2,004	0.999x - 0.786	0.999
B81	SKC	224-PCXR3	503480	01/10/2025	1,000	1,500	2,000	1,004	1,509	2,002	1.000x - 1.340	0.999
B82	SKC	224-PCXR3	505673	01/10/2025	1,000	1,500	2,000	1,007	1,502	2,003	0.994x+ 9.425	1.000
B83	SKC	224-PCXR3	510785	01/10/2025	1,000	1,500	2,000	997	1,494	1,994	0.999x - 1.727	0.999
B84	SKC	224-PCXR3	508333	01/10/2025	1,000	1,500	2,000	1,000	1,505	2,005	1.002x - 5.217	1.000
B85	SKC	224-PCXR3	505757	03/10/2025	1,000	1,500	2,000	996	1,507	1,998	1.001x - 4.459	1.000
B86	SKC	224-PCXR3	512625	03/10/2025	1,000	1,500	2,000	998	1,506	1,993	0.997x - 4.180	0.999
B87	SKC	224-PCXR3	504324	03/10/2025	1,000	1,500	2,000	1,003	1,510	2,004	1.000x + 1.037	1.000
B88	SKC	224-PCXR3	508307	03/10/2025	1,000	1,500	2,000	995	1,505	1,999	1.007x - 17.485	0.999
B89	SKC	224-PCXR3	509860	03/10/2025	1,000	1,500	2,000	999	1,500	2,001	1.002x - 6.218	1.000
B90	SKC	224-PCXR3	508366	02/11/2025	1,000	1,500	2,000	996	1,505	2,004	1.008x - 15.379	1.000
B91	SKC	224-PCXR3	510919	01/10/2025	1,000	1,500	2,000	1,004	1,504	2,003	0.998x + 1.328	1.000
B92	SKC	224-PCXR3	510987	03/10/2025	1,000	1,500	2,000	997	1,505	2,005	1.009x - 17.601	0.999
B93	SKC	224-PCXR3	509845	03/10/2025	1,000	1,500	2,000	1,006	1,511	2,001	0.998x + 2.138	0.999

Calibrated by :

(Mr. Abdul Dangklom)

Approved by :

(Mr. Peera Detudom)



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com., www.spscon.com

Rotameter Calibration Report (For Personal Pump High Flow Adjust)

Calibration Method : Dry Cal Primary Flowmeter

Model : Defender 510-H

S/N : 136164

Calibration Data

Calibration Data											
Rotameter Data			Calibration Data								
No.	Brand	Model	Date	Flow Rate (ml/min)						Value From Calibration Curve	
				Flow Rate (Reading)			Actual (Q std.)				
				1	2	3	1	2	3	y	R ²
H-B01	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	502.3	998.1	1996.9	0.999x + 2.995	1.000
H-B02	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	501.2	998.9	1998.1	0.998x + 4.159	1.000
H-B03	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	500.4	999.5	2000.3	1.000x - 1.574	0.999
H-B04	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	499.5	1000.6	1999.5	0.999x + 0.880	1.000
H-B05	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	499.1	1001.3	2001.6	1.001x - 7.095	0.999
H-B06	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	499.4	999.8	1994.1	1.000x + 2.760	1.000
H-B07	Dwyer	VFB-65	02/10/2025	500	1,000	2,000	501.0	999.4	2000.7	0.997x + 1.623	1.000
H-B08	Dwyer	VFB-65	01/10/2025	500	1,000	2,000	499.9	1001.7	2002.8	0.998x - 1.887	1.000
H-B09	Dwyer	VFB-65	02/10/2025	500	1,000	2,000	499.3	999.2	1996.6	0.999x + 1.428	0.999
H-B10	Dwyer	VFB-65	03/10/2025	500	1,000	2,000	498.7	1001.1	1997.8	0.996x + 6.669	1.000

Calibrated by :

(Mr. Abdul Dangklom)

Approved by :

(Mr. Peera Detudom)

Cert. No. : SP25026

Pages : 1 of 4

Calibration Certificate

Equipment :	UV-VIS SPECTROPHOTOMETER
Manufacturer :	PERKINELMER
Model :	LAMBDA 25
Serial No.:	501S14123010
ID No.:	SP03/58
Calibration Mode :	WAVELENGTH ACCURACY PHOTOMETRIC ACCURACY STRAY LIGHT
Condition As Found :	GOOD
Customer :	S.P.S CONSULTING SERVICE CO., LTD. 7 SOI PHAHOLYOTHIN 24, PHAHOLYOTHIN ROAD, CHOMPHON SUB-DISTRICT, CHATUCHAK DISTRICT, BANGKOK PROVINCE 10900 THAILAND.
Location :	ORGANIC LABORATORY IV
Ambient Temperature :	(22.9 ± 5) °C
Relative Humidity :	(53.7 ± 25) %
Received Date :	22 AUGUST 2025
Calibration Date :	22 AUGUST 2025
Date of Issue :	25 AUGUST 2025

Calibrated by :

Nitinun Srihawan

Approved by :

( Wichok Ekpongpradit)

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Cert. No. : SP25026

Job No. : VC68SP0019

Pages : 2 of 4

Calibration Method :

This instrument was calibrated by using on-site calibration procedure In-house method : CP-SP-01

The calibration procedure to direct measurement wavelength accuracy by using wavelength standard solution, Photometric accuracy by using absorbance standard filter and absorbance standard solution

The calibration procedure used was based on ASTM E275-01, ASTM E925-02

Condition of this result of calibration :

1. Certified reference materials

<u>Material</u>	<u>Ref. type</u>	<u>Cell serial No.</u>	<u>Cert. No.</u>	<u>Due Date</u>
Holmium liquid	RM-HL	29706	126461	24/10/2026
Didymium liquid	RM-DL	28912	126462	24/10/2026
Neutral density filter	RM-1N2N3N	13877	126457	24/10/2026
Potassium dichromate solutions	RM-0204060810	14204	126497	25/10/2026
Potassium Iodide solution	-	KI-0701-001	CI-0185-24	14/05/2026

2. This result of calibration was found accurate as shown on date and place of calibration only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 The UK National Physical Laboratory (NPL)

Result of calibration : Wavelength Accuracy

(Without adjustment)

Material	Certified Values of Reference Material (nm)	UUC* Reading (nm)	Error (nm)	Uncertainty ± (nm)	k Factor
RM-HL	278.13	278.21	0.08	0.16	2.00
	361.25	361.39	0.14	0.16	2.00
	467.82	467.71	-0.11	0.16	2.00
	536.56	536.50	-0.06	0.16	2.00
	640.50	640.36	-0.14	0.16	2.00
RM-DL	740.09	739.85	-0.24	0.16	2.00
	864.94	865.12	0.18	0.16	2.00

UUC* = Unit Under Calibration

Cert. No. : SP25026

Job No. : VC68SP0019

Pages : 3 of 4

Result of calibration : Photometric Accuracy

Material	Wavelength (nm)	Filter S/N	Nominal Absorbance (A)	Certified Absorbance (A)	UUC* Reading Absorbance (A)	Error (A)	Uncertainty ± (A)	k Factor
Neutral Density glass filter	440.0	29381	0.5	0.5443	0.5413	-0.0030	0.0043	2.00
		29914	0.7	0.7484	0.7455	-0.0029	0.0054	2.00
		29360	1.0	1.0527	1.0535	0.0008	0.0032	2.00
	465.0	29381	0.5	0.4948	0.4922	-0.0026	0.0041	2.00
		29914	0.7	0.6906	0.6877	-0.0029	0.0050	2.00
		29360	1.0	0.9695	0.9709	0.0014	0.0031	2.00
	546.1	29381	0.5	0.5090	0.5068	-0.0022	0.0036	2.00
		29914	0.7	0.6985	0.6960	-0.0025	0.0041	2.00
		29360	1.0	0.9814	0.9825	0.0011	0.0031	2.00
	590.0	29381	0.5	0.5375	0.5353	-0.0022	0.0034	2.00
		29914	0.7	0.7256	0.7231	-0.0025	0.0037	2.00
		29360	1.0	1.0213	1.0219	0.0006	0.0032	2.00
	635.0	29381	0.5	0.5223	0.5202	-0.0021	0.0033	2.00
		29914	0.7	0.6927	0.6901	-0.0026	0.0036	2.00
		29360	1.0	0.9744	0.9750	0.0006	0.0032	2.00

UUC* = Unit Under Calibration

Cert. No. : SP25026

Job No. : VC68SP0019

Pages : 4 of 4

Result of calibration : Photometric Accuracy

(Without adjustment)

Material	Wavelength (nm)	Solution (mg/l)	Certified Absorbance (A)	UUC* Reading Absorbance (A)	Error (A)	Uncertainty ± (A)	k Factor
Potassium dichromate solutions	235.0	20	0.2415	0.2443	0.0028	0.0101	2.00
		40	0.4866	0.4871	0.0005	0.0115	2.00
		60	0.7415	0.7295	-0.0120	0.0067	2.00
		80	0.9854	0.9844	-0.0010	0.0071	2.00
		100	1.2444	1.2425	-0.0019	0.0073	2.00

UUC* = Unit Under Calibration

Condition of this result of calibration : Spectrophotometer PERKINELMER Model LAMBDA 25 S/N 501S14123010

Resolution of Wavelength Mode 0.1 nm

Resolution of Photometric Mode 0.001 A

Parameter Setting

Measurement Mode Wavelength, Absorbance

Wavelength Scan 190 nm - 1100 nm

Scanning Speed 7.5 nm/min

Band width(Wavelength) 1.0

Band width(Vis) 1.0

Band width(Uv) 1.0

Stray Light** UUC* Reading at 220.0 nm	
Transimission T(%)	Absorbance(A)
0.020	3.7032

**Specific Acceptance :

Transmission ≤ 1.0 T(%), Absorbance ≥ 2.0 A

**Stray light not TISI Accredited

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95%

End of Calibration Certificate