

เอกสารแนบ 6

เอกสารสอบเทียบเครื่องมือที่ใช้ในการวิเคราะห์

Certificate No.: C0-1608001/24

Page 2 of total 4 pages

Reference Method:

- The calibration method used was CP-178 based on an in-house method.
- This certificate can be traceable to the national standards, which is realized the shown measurement units according to the International System of Units (SI Units).

Reference Standard:

Type	pH Value	Lot No.	Due Date	Traceability
pH Standard Solution	4.01	150823	Feb. 9, 2025	NIMT
	7.01	180723	Jan. 12, 2025	
	10.01	160823	Jan. 16, 2025	

Type	Serial No.	Certificate No.	Due Date	Traceability
Documenting Process Calibrator	2630521	10-2312001/23	Dec. 24, 2024	THC
Digital Thermometer with Sensor	1709138 / 4605984-005	10-0806001/24	Jun. 7, 2025	

Remark: This certificate is traceable to the International System of Unit (SI Unit) through:

- NIMT, National Institute of Metrology (Thailand).
- THC, Thai Heart Calibration Co., Ltd.

Measurement Results:

1. Function Simulated pH Meter

Standard Applied	Nominal Value	UUC Reading		Uncertainty
(mV)	(pH)	pH	mV	(± mV)
177.48	4.00	4.01	177.3	0.060
0.00	7.00	7.00	-0.1	0.060
-177.48	10.00	10.01	-177.4	0.060

UUC : Unit Under Calibration

Note : Adjust Curve to simulate pH (4,7,10)

CERTIFICATE OF CALIBRATION

Certificate No.: C0-1608001/24 Page 1 of total 4 pages

Customer WATER ANALYSIS CENTER CO., LTD.

1/94 Moo 5, T.Kanham,

A.U-thai, Ayuthaya 13210

Equipment pH Meter
Manufacturer METTLER TOLEDO Model SevenCompact S220
Serial No. B327527211 ID No. WWL 0068
Description Range : 0 - 14 pH, Resolution : 0.01 pH

Environment

Calibration

Received

Calibration

Date of Iss

Condition

Checked

() (K
() (Patiphan K.) (✓) (Onnapa P.)
() (Pongsak H.) () (Nitiiphong K.)
() (Kanung C.) () (Nonthachai K.)
() (Pramong P.) () (Noppol P.)

Certificate No.: CO-1608001/24

Page 4 of total 4 pages

Reference Method:

- The calibration method used was CP-096 based on an in-house method.
- The temperature scale used was an ITS-90.
- This certificate can be traceable to the national standards, which is realized the shown measurement units according to the International System of Units (SI Units).

Reference Standard Instruments:

Type	Serial No.	Cert. No.	Due Date	Traceability
Thermometer Readout	B7C853	10-0911001/23	Nov. 8, 2024	THC
Platinum Resistance Thermometer	4854	C0A30047	Oct. 22, 2025	FLUKE
Liquid Bath	XO111019	10-2405001/23	May 25, 2025	THC

Remark: This certificate is traceable to the International System of Unit (SI Unit) through:

- THC, Thai Heart Calibration Co., Ltd.
- FLUKE, Fluke Corporation, U.S.A.

Measurement Results:

(X) Without Adjustment

Dimension of probe : Diameter 4 mm. Sensor Type : RTD (PT100)

Immersion Depth (mm.)	Standard Reading (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
120	22.00	22.2	-0.20	0.065
120	25.00	25.2	-0.20	0.065
120	28.00	28.2	-0.20	0.065

UUC : Unit Under Calibration

The above reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor $k = 2.00$, providing a level of confidence approximately 95%.

- End of Certificate -

Calibrated by

FE-169

Calibrated by

FE-169

Certificate No.: CO-1608001/24

Page 3 of total 4 pages

Measurement Results (Cont.):

2. Calibration of pH Electrode (Serial No.: 3222623)

pH Standard Solution (pH)	Measured Value		Uncertainty (± pH)
	(pH)	(mV)	
4.01	4.01	186.1	0.013
7.01	7.01	9.3	0.013
10.01	10.00	-164.5	0.013

Note : Adjust Curve to Buffer Solution pH (4,7,10)
Temperature stability of micro bath : $25 \pm 0.2^\circ\text{C}$

The above reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor $k = 2.00$, providing a level of confidence approximately 95%.

Certificate No.: CO-1607004/24

Page 2 of total 2 pages

Reference Method:

- The calibration method used was CP-177 based on an in-house method.
- This certificate can be traceable to the national standards, which is realized the shown measurement units according to the International System of Units (SI Units).

Reference Standard :

Material	Batch Value	Lot Number	Due Date	Traceability
Conductivity Standard Solution	147.1 $\mu\text{S/cm}$	S230330005	Nov. 9, 2024	SCP Science
	1.423 mS/cm	S231129006	May 13, 2025	SCP Science

Remark: This certificate is traceable to the International System of Unit (SI Unit) through:

- SCP Science.

Measurement Results: (Probe Serial No. : 93X219065)

Conductivity Standard Solution	Measured Value	Correction	Uncertainty ($\pm$)
147.1 $\mu\text{S/cm}$	149.0 $\mu\text{S/cm}$	-1.9 $\mu\text{S/cm}$	2.5 $\mu\text{S/cm}$
1.423 mS/cm	1.425 mS/cm	-0.002 mS/cm	0.0052 mS/cm

Note : Adjustment points: 147.1 $\mu\text{S/cm}$ 1.423mS/cm

The above reported uncertainty of measurement is the expanded uncertainty obtained by multiplying the standard uncertainty with the coverage factor $k = 2.00$, providing a level of confidence approximately 95%.

- End of Certificate -

CERTIFICATE OF CALIBRATION

Certificate No.: CO-1607004/24 Page 1 of total 2 pages

Customer

WATER ANALYSIS CENTER CO., LTD.
1/94 Moo 5, T.Kanham,
A.U-thai, Ayuthaya 13210

Equipment

Conductivity Meter

Manufacturer

EUTECH

Model

CON 2700

Serial No.

2657889

ID No.

WWL-0136

Description

.

Environmental Conditions

Ambient Temperature: $(20 \pm 2) ^\circ\text{C}$

Relative Humidity: $(50 \pm 10) \%$

Calibration L

Received Date

Calibration D

Date of Issue

Condition of

Checked by

() (Krisy
() (Paiphon K.) (✓) (Onnapa P.)
() (Pongsak H.) () (Nitiphong K.)
() (Kanung C.) () (Nonthachai K.)
() (Pramong P.) () (Noppol P.)

This calibration certificate shall not be reproduced other than in full except with the prior written approval of the Thai Heart Calibration Co., Ltd.

FE-169

REV.02.02/24/21

Function

Calibration point

: Temperature measurement

: 20 °C

Result

Resolution

: Without adjustment


: 0.1 °C

Certificate No. : MT24-7016

Page : 2 of 2

Calibration point (°C)	Temperature of UUC* at each position (°C)									Uncertainty of measurement (+/- °C)
	Ch.1	Ch.2	Ch.3	Ch.4	Ch.5	Ch.6	Ch.7	Ch.8	Ch.9	
20	20.344	20.098	20.405	20.375	20.193	20.010	20.245	20.090	20.037	0.41

Setting temperature (°C)	Indicating Temperature (°C)	Measured stability (+/- °C)	Measured uniformity (°C)	Overall variation (°C)
20.0	20.0	0.30	0.68	0.86



UUC* = Unit under calibration

Uniformity = Maximum and Minimum difference of measured temperature at any probes and the measured temperature at the reference and same time.

Overall Variation = Difference of temperature value between the maximum and minimum any time.

Stability = One half of the maximum difference of measured temperatures at any one probe.

-000-

Certificate of Calibration

Certificate No. : MT24-7016
Page : 1 of 2

Customer

Address

: Water Analysis Center Co.,Ltd.

: 1/94 Moo 5 , Rojana Industrial Park , T.Kanham, A.U-Thai, Ayutthaya 13210

Description

Manufacturer

Model

Serial No.

Identification No.

Calibration Place

: Refrigerator

: B.T.Metrology Co.,Ltd.

: REF 940L

: BT-03-09-09

: WWL 0043

: Customer Laboratory

Order No.

Received date

Calibration date

Environment Condition :

Temperature

Humidity

: 2601/24

: Aug 02, 2024

: Aug 02, 2024

: (25+/-10) °C

: (50+/-30) %RH

Calibration Method : Calibration were conducted using In-house calibration procedure CP-MT-006 According to comparison with LXI Data Acquisition Switch Unit with sensor. The calibration methods based on Euramet Calibration Guide No.20 - guidelines on the Calibration of Temperature and/or Humidity Controlled Enclosures.

Reference Standard Instruments :

Instrument	Model	Serial No.	Certificate No.	Due Date
LXI Data Acquisition Switch Unit with Sensor	34972A	MY49020096	MT23-7163	Nov 30, 2024

The effect that the result relate only to the items calibrated. It was found accurate as shown on date and place of calibration only.

Traceability : This measurement are traceable to the International System of Unit (SI), through National Institute of Metrology Thailand (NIMT)

The reported expanded uncertainty of measurement was based on standard uncertainty multiplied by coverage factor 2, providing a level of confidence of not less than 95%

Calibrated by : Mr.Yuttakorn Jammeansri

Approved

Issue d

This calibration certificate shall not be reproduced other than in full except with the prior written approval of Incotech Metrological Center Co.,Ltd

CALIBRATION REPORT

Professional Calibration & Services Co.,Ltd.

Certificate Number : PL61070/24

Page 2 of 3

Equipment Standards Used

Description	Serial No.	Traceability to	Certificate No.	Cal. Due Date
Zero Oxygen Solution Set	-	NIST	S0060/23	01-May-28

Condition as received : Normal
Definitions :-

* NIST - National Institute of Standard and Technology



Professional Calibration & Services Co., Ltd.
50/888, 50/889 Moo 2, Rungsi-Nakornnayok Rd., Bungyeetho, Thunychuri,
Pothumthani 12130 Thailand
Tel : (+66)2150-6441 (Autoline)
Email : info@p-cal.com www.p-cal.com



Certificate of Calibration

Page 1 of 3



Certificate Number : PL61070/24
Control Number : PCAL174170
Customer Control : WWL 0073
Description : Dissolved Oxygen Meter
Manufacturer : YSI
Model : YSI 5000
Serial Number : 14C100917
Customer : Water Analysis Center Co.,Ltd
1/94 Moo 5 T.Kanham A.U-Thai Ayutthaya 13210 Thailand

Date of Receipt : 02-Dec-24
Date of Calibration : 02-Dec-24
Environment : Temperature 20 °C ± 2 °C
Relative Humidity 50 % ± 20 %
Calibration Method : Calibration Procedure Number CP-PL93
Calibration Results : See data attached

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%.

This certificate is issued in accordance with ISO/IEC17025 and the conditions of accreditation granted by the Accreditation Body which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. The results relate only to the item calibrated.

This certificate shall not be reproduced other than in full except without the prior written approval of the Head of Calibration Laboratory of Professional Calibration & Services Co., Ltd.

06-Dec-24

Issued Date



Inctech Metrological Center Co.Ltd.
39/1 Soi 82, Sukhapiban 5 Rd., O ngoen,
Salmat, Bangkok 10220, Thailand
Tel. (662) 909-8820 (Auto 10 lines) www.imcinstrument.com



Certificate of Calibration

Certificate No. : MT25-3161
Page : 1 of 2

Customer : Water Analysis Center Co.,Ltd.
Address : 1/94 Moo 5 , Rojana Industrial Park , T.Kanham, A.U-Thai, Ayutthaya 13210

Description : Hot Air Oven
Manufacturer : Memmert
Model : UF260
Serial No. : B620.0814
Identification No. : WWL 0212
Calibration Place : Customer Laboratory

Order No. : 1011/25
Received date : Mar 25, 2025
Calibration date : Mar 20, 2025
Environment Condition:
Temperature : (25 $\pm$ 10) °C
Humidity : (50 $\pm$ 30) %RH

Calibration Method : Calibration were conducted using In-house calibration procedure CP-MT-005 According to comparison with LXI Data Acquisition Switch Unit with sensor. The calibration methods based on Euramet Calibration Guide No.20 - guidelines on the Calibration of Temperature and/or Humidity Controlled Enclosures.

Reference Standard Instruments :

Instrument	Model	Serial No.	Certificate No.	Due Date
LXI Data Acquisition Switch Unit with Sensor	34972A	MY49028922	MT24-8770	Nov 22, 2025

The effect that the result relate only to the items calibrated. It was found accurate as shown on date and place of calibration only.

Traceability : This measurement are traceable to the International System of Unit (SI), through National Institute of Metrology Thailand (NIMT)

The reported expanded uncertainty of measurement was based on standard uncertainty multiplied by coverage factor of 2, providing a level of confidence of not less than 95%



approval of Inctech Metrological Center Co.,Ltd

CALIBRATION REPORT

Professional Calibration & Services Co., Ltd.

Certificate No.: PL61070/24

Page : 3 of 3

Calibration Results

Dissolved Oxygen Calibration

Description of Meter : Range : 0 to 60 mg/l
Resolution : 0.01 mg/l

Description of Electrode : Manufacturer : YSI
Model : 5010
Serial No. : 13C100067
Type : Electrochemical (Membrane)

Calibration Point	Standard Value	UUC Reading	UUC Error	Uncertainty ($\pm$)
0 mg/l	0.000 mg/l	0.00 mg/l	0.00 mg/l	0.03 mg/l
8 mg/l	8.454 mg/l	8.43 mg/l	-0.02 mg/l	0.05 mg/l
9 mg/l	9.020 mg/l	9.02 mg/l	0.00 mg/l	0.05 mg/l


Notes :

1). Calibration results that carry the double asterisk (**) are not accredited. Calibrations marked as such on this Certificate have been included for completeness.

...End...



Intech Metrological Center Co.,Ltd.
39/1 Soi 82, Sukhapiban 5 Rd., O ngoen,
Salmal, Bangkok 10220, Thailand
Tel. (662) 909-8820 (Auto 10 lines) www.imcinstrument.com



Certificate of Calibration



Equipment:

Balance
BL210S
Model: 06 December 2024
Serial No. (or ID.): WO-00053756
Manufacturer: Sartorius
Condition: 1 of 2
In condition

Customer:

Water Analysis Center Co., Ltd.
1/94 Moo 5, Rojana Industrial Park, Rojana Road,
Tambol Kanham, Amphur U-Thai, Ayutthaya 13210 Thailand

Environment Condition:

Temperature 24 °C ± 0.9 °C
Humidity 53 %RH ± 1.3 %RH

Calibration Place:

Water Analysis Center Co., Ltd. (วัดเครื่องวัด)
1/94 Moo 5, Rojana Industrial Park, Rojana Road,
Tambol Kanham, Amphur U-Thai, Ayutthaya 13210 Thailand

Calibration By:

Mr. Apiwit Chaosap
04 December 2024

Calibration Date:

In-house method, CAL-WI-47, based on UKAS Lab 14

The Method used:

This certificate is traceable to the SI Units maintained by National Institute of Metrology (NIMT), Thailand through DKSH Technology Co., Ltd. Certificate No. C02241786

Traceability:



This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.
The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).
These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of DKSH Technology Limited.

DKSH Technology Limited
2533 Sukhumvit Road, Bangkok, Thailand 10260
Phone: +66 2638 7000 Email: info.calibration@dksh.com Website: www.dksh.com/technical-thailand

Delivering Growth - in Asia and Beyond.

CAL-FM-C01-14; 12 Sep 2022

Function

Calibration point : Temperature measurement
Resolution : 104, 180 °C

Certificate No. : MT25-3161


Page : 2 of 2

Result : Without adjustment

Resolution : 0.1 °C

Calibration point (°C)	Temperature of UUC* at each position (°C)									Uncertainty of measurement (+/- °C)
	Ch.1	Ch.2	Ch.3	Ch.4	Ch.5	Ch.6	Ch.7	Ch.8	Ch.9	
104	103.767	103.648	104.174	103.965	104.090	104.047	104.160	103.891	104.284	0.32
180	179.673	179.787	179.762	179.908	179.891	179.615	179.920	179.806	179.752	0.50

Setting temperature (°C)	Indicating Temperature (°C)	Measured stability (+/- °C)	Measured uniformity (°C)	Overall variation (°C)
104.0	104.0 to 104.2	0.13	0.75	0.80
180.0	180.0 to 180.3	0.39	0.66	0.81



Front view

UUC* = Unit under calibration

Uniformity = Maximum and Minimum difference of measured temperature at any probes and the measured temperature at the reference and same time.

Overall Variation = Difference of temperature value between the maximum and minimum any time.

Stability = One half of the maximum difference of measured temperatures at any one probe.

-oOo-

PREVENTATIVE MAINTENANCE (PM) CHECK LIST
 FOR ATOMIC ABSORPTION SPECTROMETER
 Model & Serial Number: 240 FS AA & M418230004
 Customer : Water analysis center Co., Ltd.
 Date: 25 Apr 2024


Safety

- ☒ Flame, Inspect/replace o-ring nebulizer, spray chamber and burner
☒ Flame, Clean nebulizer, spray chamber and burner
☒ Flame, Check liquid trap interlock, burner interlock, pressure relief bung interlock and shield interlock
☐ Furnace, Clean work head , electrode and shroud N/A
☐ Furnace, Clean PSD and PSD tray N/A
☐ Furnace, Check water pressure N/A
☒ Check drain tube
☒ Check exhaust system
☒ Check gas pressure sensor interlock
☒ Check and all gas hoses for SpectrAA
☒ Clean computer control

- Optics

☒ Inspect/Replace that external optics surfaces
☒ Check Wavelength Accuracy the copper line at 323.0-326.0 nm = 324.7 nm
☒ Check that PMT % Gain the copper at 324.8 nm, 4 mA, 0.5 nm slit width, Gain = 29.1 (should be $\leq 64\%$ or $\leq 380V$)
☒ Flame, Check D2 lamp is work

Calibration Results:
Without Adjustment
 Eccentric Error: Weight to be 1/3 or 1/2 of Maximum capacity, taken from the center of the pan as a zero reference.



Repeatability: Determination of the standard deviation of weighing balance., Readability 0.0001 (g)

Nominal test value (g)	Standard Deviation
20	0.00005
200	0.00006

Error of indication from nominal or conventional mass value., Readability 0.0001 (g)			
Nominal Value (g)	Conventional Mass (g)	Displayed Value (g)	Error of indication (g)
1	1.00001	1.0000	0.0000
2	2.00001	2.0000	0.0000
5	5.00001	5.0000	0.0000
10	10.00001	10.0000	0.0000
20	20.00001	20.0000	0.0000
50	50.00000	50.0000	0.0000
70	70.00001	70.0001	0.0001
100	99.99996	100.0001	0.0001
120	119.99997	120.0001	0.0001
150	149.99996	150.0002	0.0002
200	199.99989	200.0007	0.0008
Uncertainty (g)			k
0.00011			2.04
0.00011			2.04
0.00011			2.04
0.00011			2.04
0.00012			2.03
0.00013			2.02
0.00016			2.01
0.00017			2.01
0.00021			2.00
0.00024			2.00
0.00030			2.00

The End of Certificate



บริษัท ไทยยูนิค จำกัด THAI UNIQUE CO., LTD.

80-82 ถนนประชาธิปไตย แขวงบางขุนพรหม เขตพระนคร กรุงเทพฯ 10200
80-82 Prachathipatai Rd., Bangkokphrom, Pranakorn, Bangkok 10200

Tel. 0-2629-0191-6, 0-2280-1787, Fax. 0-2280-1788, E-mail : thawat@thaiunique.com, Website : www.thaiunique.com

PREVENTATIVE MAINTENANCE (PM) CHECK LIST

FOR ATOMIC ABSORPTION SPECTROMETER

Model & Serial Number: 2402 AA & M418230004

Customer : Water Analysis Center Co., Ltd.

Date: 26 Apr 2024

Safety

- ☐ Flame, Inspect/replace o-ring nebulizer, spray chamber and burner N/A
- ☐ Flame, Clean nebulizer, spray chamber and burner N/A
- ☐ Flame, Check liquid trap interlock, burner interlock, pressure relief bung N/A
- ☒ interlock and shield interlock
- ☒ Furnace, Clean work head, electrode and shroud
- ☒ Furnace, Clean PSD and PSD tray
- ☒ Furnace, Check water pressure
- ☒ Check drain tube
- ☒ Check exhaust system
- ☒ Check gas pressure sensor interlock
- ☒ Check and all gas hoses for SpectraAA
- ☒ Clean computer control

Optics

- ☒ Inspect/Replace that external optics surfaces
- ☒ Check Wavelength Accuracy the copper line at 323.0-326.0 nm = 324.7 nm
- ☒ Check that PMT % Gain the copper at 324.8 nm, 4 mA, 0.5 nm slit width, Gain = 90.7% (should be $\leq 64\%$ or $\leq 380V$)
- ☐ Flame, Check D2 lamp is work N/A



บริษัท ไทยยูนิค จำกัด THAI UNIQUE CO., LTD.

80-82 ถนนประชาธิปไตย แขวงบางขุนพรหม เขตพระนคร กรุงเทพฯ 10200
80-82 Prachathipatai Rd., Bangkokphrom, Pranakorn, Bangkok 10200

Tel. 0-2629-0191-6, 0-2280-1787, Fax. 0-2280-1788, E-mail : thawat@thaiunique.com, Website : www.thaiunique.com

Electronics

- ☒ Check power supply voltage
- ☒ Check cables and connectors
- ☒ Check/Clean all boards in the instrument
- ☐ Furnace, Check camera and align** N/A

**Option for Graphite Zeeman only

Mechanisms

- ☒ Flame, Check the burner adjuster
- ☐ Furnace, Check PSD accessories N/A

Analytical performance

- ☒ Clear the sample compartment
- ☒ Flame, Check uptake rate form 7.2-10.6 mL per minute = 8.5 mL/min
- ☒ Test Photometric noise, STDV = 0.0001 Abs (should be ≤ 0.00050 Abs)
- ☒ Flame, Test high solids nebulizer setting use
- ☒ -Air/acet Cu 5 ppm = 0.79 Abs, and Precision (%RSD) = 0.4 % (should be > 0.55 Abs and $< 0.5\%$ RSD)
- or
- ☒ -N2O/Acet Cu 5 ppm = Abs, and Precision (%RSD) = % (should be > 0.3 Abs and $< 0.5\%$ RSD)
- ☐ Furnace, Characteristic mass and sensitivity Cu 25 ppb = Abs, and Precision (%RSD) = % (should be ≥ 0.15 Abs and $\leq 4.0\%$ RSD)

Signature
En



บริษัท ไทยยูนิค จำกัด THAI UNIQUE CO., LTD.

80-82 ถนนประชาธิปไตย แขวงบางขุนพรหม เขตพระนคร กรุงเทพฯ 10200
80-82 Prachathipatani Rd., Bangkokphrom, Pranakorn, Bangkok 10200
Tel. 0-2629-0191-6, 0-2280-1787, Fax. 0-2280-1788, E-mail : thawat@thaiunique.com, Website : www.thaiunique.com

PREVENTATIVE MAINTENANCE (PM) CHECK LIST

FOR ATOMIC ABSORPTION SPECTROMETER

Model & Serial Number: AA240FS & AA 0911M073
Customer : Water Analysis Center Co., Ltd.
Date: 11 Feb 2015

Safety

- ☒ Flame, Inspect/replace o-ring nebulizer, spray chamber and burner
- ☒ Flame, Clean nebulizer, spray chamber and burner
- ☒ Flame, Check liquid trap interlock, burner interlock, pressure relief/bung interlock and shield interlock
- ☐ Furnace, Clean work head, electrode and shroud N/A
- ☐ Furnace, Clean PSD and PSD tray N/A
- ☐ Furnace, Check water pressure N/A
- ☒ Check drain tube
- ☒ Check exhaust system
- ☒ Check gas pressure sensor interlock
- ☒ Check and all gas hoses for SpectraAA
- ☒ Clean computer control

Optics

- ☒ Inspect/Replace that external optics surfaces
- ☒ Check Wavelength Accuracy the copper line at 323.0-326.0 nm = 324.7 nm
- ☒ Check that PMT % Gain the copper at 324.8 nm, 4 mA, 0.5 nm slit width, Gain = 54% (should be $\leq 64\%$ or $\leq 380V$)
- ☒ Flame, Check D2 lamp is work



บริษัท ไทยยูนิค จำกัด THAI UNIQUE CO., LTD.

80-82 ถนนประชาธิปไตย แขวงบางขุนพรหม เขตพระนคร กรุงเทพฯ 10200
80-82 Prachathipatani Rd., Bangkokphrom, Pranakorn, Bangkok 10200
Tel. 0-2629-0191-6, 0-2280-1787, Fax. 0-2280-1788, E-mail : thawat@thaiunique.com, Website : www.thaiunique.com

Electronics

- ☒ Check power supply voltage
- ☒ Check cables and connectors
- ☒ Check/Clean all boards in the instrument
- ☒ Furnace, Check camera and align**

**Option for Graphite Zeeman only

Mechanisms

- ☐ Flame, Check the burner adjuster N/A
- ☒ Furnace, Check PSD accessories

Analytical performance

- ☒ Clear the sample compartment
- ☐ Flame, Check uptake rate form 7.2-10.6 mL per minute = _____ mL/min N/A
- ☒ Test Photometric noise, STDV = 0.0002 Abs (should be ≤ 0.00050 Abs)
- ☐ Flame, Test high solids nebulizer setting use N/A
- ☐ -Air/acet Cu 5 ppm = _____ Abs, and Precision (%RSD) = _____ % (should be > 0.55 Abs and $\leq 0.5\%$ RSD)
- or
- ☐ -N2O/Acet Cu 5 ppm = _____ Abs, and Precision (%RSD) = _____ % (should be > 0.3 Abs and $< 0.5\%$ RSD)
- ☒ Furnace, Characteristic mass and sensitivity Cu 25 ppb = 0.16 Abs, and Precision (%RSD) = 3 % (should be ≥ 0.15 Abs and $\leq 4.0\%$ RSD)

SIGN
Engineer

BSC Certification Test Report

Page 1 of 6

Certificate No. : M1439/24
Customer Name : LABORATORY WATER ANALYSIS CENTER COMPANY LIMITED
Customer Address : 1/94 Moo 5 Kham Ham Subdistrict,
Uthai District, Phra Nakhon Si Ayutthaya 13210

Equipment : Biological Safety Cabinet Class II Type A2
Manufacturer : Microtech
Model : V6-T
Serial No : 0972k097272
ID No. : WWL 0084

Were in accordance with ☒ EN 12469 ☐ NSF 49 ☐ Manufacturer's specification

Test Date : 15/10/2024
Due Date : 15/10/2025 or after HEPA filters are replaced or unit is moved
Test by : Mr. Pawut Wongnarakornkul

Approved by :

Issued Date : 16/10/2024

This calibration certificate documents the traceability to national standards, which realize the unit of measurement according to the International System of Units (SI).

This certificate may not be reproduced other than in full except with the prior written approval of the Megafil Company Limited.



บริษัท ไทยยูนิค จำกัด

THAI UNIQUE CO., LTD.

80-82 ถนนประชาธิปไตย แขวงบางขุนพรหม เขตพระนคร กรุงเทพฯ 10200
80-82 Prachathipatani Rd., Bangkhunphrom, Pranakorn, Bangkok 10200
Tel. 0-2629-0191-6, 0-2280-1787, Fax. 0-2280-1788, E-mail : thauunique@thauunique.com, Website : www.thauunique.com

Electronics

- ☒ Check power supply voltage
- ☒ Check cables and connectors
- ☒ Check/Clean all boards in the instrument
- ☐ Furnace, Check camera and align** **N/A**

**Option for Graphite Zeeman only

Mechanisms

- ☒ Flame, Check the burner adjuster
- ☐ Furnace, Check PSD accessories **N/A**

Analytical performance

- ☒ Clear the sample compartment
- ☒ Flame, Check uptake rate form 7.2-10.6 mL per minute = **9.5** mL/min
- ☒ Test Photometric noise, STDV = **0.0001** Abs (should be ≤ 0.00050 Abs)
- ☒ Flame, Test high solids nebulizer setting use
 - Air/acet Cu 5 ppm = **0.80** Abs, and Precision
 - (%RSD) = **0.2** % (should be ≥ 0.55 Abs and < 0.5 % RSD)
- or
- N2O/Acet Cu 5 ppm = _____ Abs, and Precision
- (%RSD) = _____ % (should be > 0.3 Abs and < 0.5 % RSD)
- ☐ Furnace, Characteristic mass and sensitivity Cu 25 ppb = _____ Abs, and **N/A**
- Precision (%RSD) = _____ % (should be ≥ 0.15 Abs and ≤ 4.0 % RSD)

SIGN
Engine

Certificate No. : M1439/24

2. Inflow velocity test.

Select method. : ☐ DIM ☒ Exhaust velocity.

☐ MFG's Specifications

MFG's Specifications method

(m/s.)

0.54	0.57	0.55	0.54	0.55
0.56	0.55	0.56	0.57	0.54
0.59	0.53	0.54	0.57	0.56
0.53	0.6	0.56	0.55	0.58
0.55	0.58	0.54	0.53	0.55

Average Inflow velocity 0.47 m/s (93 FPM.) Velocity range >0.40 m/s (>79 FPM.)

Inflow dimension 8 x 72 (inch x inch) Inflow area 4.00 SQ.FT

Inflow volume(Q) 372 CFM

Result Summary ☒ Pass ☐ Fail

Adjustments Required ☐ Fan Speed ☐ Damper

Equipment used : Thermo Anemometer Model 425 S/N : 02968605 Calibration date : 10/05/2024

3. HEPA filter leak test.

Measurement Data

HEPA Filter	PAO Upstream Conc.(calculated)	Specification	Measured leak penetration
Supply HEPA Filter	18 µg/L	<0.01%	<0.01%
Exhaust HEPA Filter	18 µg/L	<0.01%	<0.01%

Certificate No. : M1439/24

Procedure Used : European Standard EN12469 : 2000 has the status of British Standard,
Biotechnology Performance criteria for microbiological safety cabinets.
: NSF International Standard / American National Standard NSF / ANSI 49-2008
Biosafety Cabinet : Design, Construction, Performance and Field Certification.
: Australian Standard : AS 1807.23-2000 Determination of intensity of radiation from germicidal ultraviolet lamps.
: Manufacturer's specification.

1. Downflow velocity test.

Measurement Information

No. of Rows	No. of Readings	Grid Spacing Front-Back	Grid Spacing Side-Side	Probe height Above sash
2	8	1/4,3/4	1/8,3/8	100mm

Measurement Data. (m/s.)

0.37	0.43	0.41	0.39
0.36	0.35	0.32	0.34

Average velocity 0.37 m/s (73 FPM.) Velocity range 0.25-0.50 m/s (49-98 FPM.)

Uniformity(EN: +/-20%avg.) 0.30 - 0.44 m/s (58 - 88 FPM.)

Supply filter dimension 24 x 72 (inch x inch) Supply filter area 10.69 SQ.FT

Downflow volume (Q) 780 CFM.

Result Summary ☒ Pass ☐ Fail

Equipment used : Thermo Anemometer Model 425 S/N : 02968605 Calibration date : 10/05/2024

Certificate No. : M1439/24

Result Summary

Downflow Pattern test	☒ Accept	☐ Non-Conforming
View screen retention test	☒ Accept	☐ Non-Conforming
Work opening edge retention test	☒ Accept	☐ Non-Conforming
Sash/window seal test	☒ Accept	☐ Non-Conforming

5. Site installation

Sash Alarm.	☐ Pass	☐ Fail	☒ N/A
Interlock System.	☐ Pass	☐ Fail	☒ N/A
Exhaust System Performance	☐ Pass	☐ Fail	☒ N/A

Remark / Recommendation

ระบบ Site installation ไม่มีการตรวจสอบ เนื่องจากตู้ไม่ฟังก์ชันนี้

6. Illumination Test (Lighting) : Option

Lighting should be adequate for safe working within the cabinet. Illumination measured at the work surface.

Lux

585	936	917	514
849	1400	1465	755

Equipment used : Digital Light Meter Model Easy View 31 S/N : 160404993 Calibration date : 08/05/2024

Remark :

Certificate No. : M1439/24

Leak location

Supply HEPA Filter	Exhaust HEPA Filter
Back	Back
	

Result Summary ☒ Pass ☐ Fail

Equipment used : Aerosol Photometer Model TDA-2H S/N : 20138 Calibration date : 08/05/2024

Equipment used : Smoke Generator Model TDA-6C S/N : 20192

4. Airflow smoke patterns test

Measurement Information

- Downflow Pattern test : Smoke shall be passed from one end of the cabinet to the other, along the centerline of the work surface, at a height of 4 inch (10 cm) above the top of the access opening
- View screen retention test : Smoke shall be passed from one end of the cabinet to the other, 1.0 in (2.5 cm) behind the view screen, at a height 6.0 inch (15 cm) above the top of the access opening.
- Work opening edge retention test : Smoke shall be passed along the entire perimeter of the work opening. Particular attention should be paid to corners and vertical edges.
- Sash/window seal test : Smoke shall be passed up the inside of the window 2 in (5 cm) from the sides and along the top of the work area.

7. Ultraviolet Lamp Test (UV) : Option

Ultraviolet radiation where UV Lamp are fitted, the intensity of radiation at a wavelength of 254 nm. Shall be not less than 400 mW/m² when measures at work floor surface.

630	1450	1480	690
380	920	930	390

Equipment used : UVC LIGHT METER **Model** UVC-254SD **S/N :** Q879819 **Calibration date :** 08/05/2024

Remark:

-000-

Certificate of Calibration

LIQUID BATH



Page 1 of 3

Certificate No.: MC 2413808

Customer : Water Analysis Center Co., Ltd.
1/94 Moo 5, T.Kantham, A.U-Thai, Ayutthaya 13210.

Reference Job No. : 24-2841
Received Date : 16 December 2024

Description	: Water Bath	Resolution	: 0.1 °C
-------------	--------------	------------	----------

Manufacturer : ESSTELL Model : EWB-122D

Serial No. : 20180508122 ID. No.

Marking
; Additionally for the purpose of identification by this laboratory a label marked with this certificate number (MC 2413808) has been attached to the case.

Method : In-House calibration procedure MWI-T-029 this method is base on ASTM E 715-2007 "Liquid Bath".

Location of Calibration : Water Analysis Center Co., Ltd. ; Laboratory.

Environmental Conditions : Ambient Temperature : (25.2 to 25.6) °C

Relative Humidity : (49.0 to 51.0) %

Date of Calibration	: 16 December, 2024	Date of Issue	: 18 December 2024
---------------------	---------------------	---------------	--------------------

Checked by :

The uncertainties are for a confidence probability of approximately 95%

This certificate is issued in accordance with the conditions of accreditation granted by the National Standardization Council of Thailand-Office of the National Standardization Council that has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of Master Calibration Co., Ltd.

[MCF-Q-077 ; Rev.6 ; Date : 22/04/2021]

Certificate No.: MC 2413808

Page 3 of 3

2. Result of calibration :

Temperature Measurement Accuracy Test

Indicating Temperature (°C)	Measured Temperature (°C) at Spread Locations					Uncertainty of measurement (±°C)
	#1	#2	#3	#4	Ref. #5	
45.0	44.6	44.6	44.5	44.5	44.4	0.86

Chamber Characterization Result

Desired Temperature (°C)	Controller Temperature (°C)	Indicating Temperature (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
44.5	45.0	45.0	0.85	0.75	1.9

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2.0$, providing a level of confidence of approximately 95 %.

This certificate will certify of the calibrated equipment only.

End of Certificate

Checked by :

[MCF-Q-077 ; Rev.6 ; Date : 22/04/2021]

Certificate No.: MC 2413808

Page 2 of 3

Reference Standard Instrument :

Description	Certificate No.	Serial No.	Due date	Traceable thru
Data Acquisition/Switch Unit With Thermocouple Type "T" ID. No.27/1 to 27/5	MC 2403566	MY44020009	13 Mar 2025	MCAL

Traceability :

The measurement standard traceable to the international system of units (SI) through certificate as mentioned above


1. Calibration Procedure:

This Instrument was calibration according to ASTM E715 - 2007 by comparison with calibrated sensor under no load condition. The sensor were placed on five points and located one sensor in each of the eight corners of the chamber and was away from the each wall of 5 cm to 10 cm. And placed the five sensor within 2.5 cm of the geometric center of the chamber.

Temperature Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.

Temperature Stability - one-half of the greatest maximum difference of measured temperatures at any one sensor.

Overall Variation - The Difference of the maximum and minimum measured temperatures throughout observation.



- Overall Ambient Temperature around the Chamber variation : 1.1 °C
- Overall Line Voltage variator 0.0 V
- Chamber Size (W*H*D) : 50 cm x 12 cm x 30 cm
- Water Level : 7 cm

Checked by :

[MCF-Q-077 ; Rev.6 ; Date : 22/04/2021]

Certificate No.: MC 2413810

Page 2 of 3

Reference Standard Instrument:

Description	Certificate No.	Serial No.	Due date	Traceable thru
Data Acquisition/Switch Unit With RTD ID. No.10/1 to 10/9	MC 2400121	MY59002240	18 Mar 2025	MCAL

Traceability:

The measurement standard traceable to the international system of units (SI) through certificate as mentioned above

1. Calibration Procedure:

This Instrument was calibration according to TLAS G-20 by comparison with calibrated thermocouple type T under no load condition. The Thermocouples were placed on nine points and located one thermocouple in each of the eight corners of the chamber and was away from the each wall of 5 cm to 10 cm. And placed the ninth thermocouple within 2.5 cm of the geometric center of the chamber.

Temperature Uniformity - the maximum difference of measured temperatures at any sensors and the measured temperature at the reference location which are observed at the same time or at as close an observation time as possible to determine the temperature pattern or homogeneity within the chamber under steady state conditions. The reference sensor should preferably be located at the geometric center of the chamber.

Temperature Stability - one-half of the greatest maximum difference of measured temperatures at any one sensor.

Overall Variation - The Difference of the maximum and minimum measured temperatures throughout observation.




Figure 1: Sensor Installation Location

Checked by:

Certificate of Calibration

TEMPERATURE CONTROLLER ENCLOSURES



Page 1 of 3



Certificate No.: MC 2413810

Customer : Water Analysis Center Co., Ltd.
1/94 Moo 5, T.Kantham, A.U-Thai, Ayuthaya 13210.

Reference Job No.	: 24-2841	Received Date	: 16 December 2024
Description	: Incubator	Resolution	: 0.1 °C
Manufacturer	: Memmert	Model	: IN260
Serial No.	: D619.0170	ID. No.	: WWL 0192
Marking	: Additionally for the purpose of identification by this laboratory a label marked with this certificate number (MC 2413810) has been attached to the case.		
Method	: In-house calibration procedure MWL-T-033 this method Base on TLAS G-20-1/02-08 "Temperature Controlled Enclosures".		
Location of Calibration	: Water Analysis Center Co., Ltd.; Laboratory.		
Environmental Conditions	: Ambient Temperature : (23.3 to 24.1) °C		
	: Relative Humidity : (54.8 to 64.8) %		
Date of Calibration	: 16 December 2024	Date of Issue	: 18 December 2024

Checked by:

The uncertainties are for a confidence probability of approximately 95%

This certificate is issued in accordance with the conditions of accreditation granted by the National Standardization Council of Thailand-Office of the National Standardization Council that has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standards laboratory. This certificate may not be reproduced other than in full except with the prior written approval of Master Calibration Co.,Ltd.



Certificate No.: MC 2413810

Page 3 of 3

2. Result of calibration :

Temperature Measurement Accuracy Test

Indicating Temperature (°C)	Measured Temperature (°C) at Spread Locations									Uncertainty (±°C)	* Uncertainty does not include stability. (±°C)
	#1	#2	#3	#4	#5	#6	#7	#8	Ref. #9		
35.0	35.00	35.20	35.00	35.20	34.90	35.00	34.80	34.90	35.00	0.22	0.16

(*) : Non Accredited

Chamber Characterization Result

Desired Temperature (°C)	Controller Temperature (°C)	Indicating Temperature (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
35.0	35.0	35.0	0.08	0.25	0.50

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2.0$, providing a level of confidence of approximately 95 %.

This certificate will certify of the calibrated equipment only.

End of Certificate

Checked