

ภาคผนวก ญ

ผลตรวจวัดคุณภาพน้ำเสียก่อนการบำบัด น้ำทิ้งหลังการบำบัด
และเอกสารที่เกี่ยวข้อง

REF NO. : MGG0702025199

REPORT NO. : JOB01199

ต้นฉบับ

รายงานผลการวิเคราะห์คุณภาพน้ำ

เลขโครงการ : JB6807225 เลขที่รายงาน : WW6807225
บริษัท/โครงการ : โครงการ โรงแรม วิสา หัวหิน
ที่อยู่ของโครงการ : 222/222 ถนนเพชรเกษม ตำบลหัวหิน อำเภอหัวหิน จังหวัดประจวบคีรีขันธ์ 77110
วันที่เก็บตัวอย่าง : 7 กรกฎาคม 2568 วันที่รับตัวอย่าง : 7 กรกฎาคม 2568
วันที่วิเคราะห์ : 7-23 กรกฎาคม 2568 วันรายงานผล : 29 กรกฎาคม 2568
ผู้เก็บตัวอย่าง : นายมะห์เตร์ รอหมาน (ว-299-จ-0003) ประเภทตัวอย่าง : Wastewater

พารามิเตอร์	วิธีวิเคราะห์	ผลการวิเคราะห์		มาตรฐาน	หน่วย
		น้ำทิ้งก่อนบำบัด	บ่อกักน้ำทิ้งก่อนปล่อย		
pH	Electrometric Method	7.0	7.2	5.5-9.0	-
Total Dissolved Solids	Dried at 180 °C	266	256	≤1,300	mg/L
Total Suspended Solids	Dried at 103-105 °C	25	5	≤50	mg/L
Settleable Solids**	Settleable Solids	1.5	0.5	-	mL/L
Sulfide	Iodometric Method	0.3	0.4	≤1.0	mg/L
Oil and grease	Soxhlet Extraction Method	9.9	9.9	≤20	mg/L
Biochemical Oxygen Demand	Azide Modification & 5-Day BOD Test Method	135.4	4.1	≤40	mg/L

หมายเหตุ : มาตรฐาน : ประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุม การระบายน้ำทิ้งจากอาคารบางประเภทและบางขนาด พ.ศ.2567 (ประเภท ค)

Method : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23rd Edition, 2017.

Settleable Solids** : อยู่นอกเหนือขอบข่ายการขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม

ลักษณะตัวอย่าง : น้ำทิ้งก่อนบำบัด: เหลืองจาง มีตะกอน

บ่อกักน้ำทิ้งก่อนปล่อย: เหลืองจาง มีตะกอน

ห้องปฏิบัติการวิเคราะห์เอกชน

เลขทะเบียนที่ ว-299



ลงชื่อ..... Rinyarat sia.....รับรอง

(นางสาวริญญารัตน์ ศิริสุนทรพิงษ์)

ผู้ควบคุมห้องปฏิบัติการ

ว-299-ค-0001

ผลการวิเคราะห์นี้รับรองเฉพาะตัวอย่างที่ได้ทำการวิเคราะห์เท่านั้น

ห้ามคัดลอกหรือทำสำเนารายงานผลการวิเคราะห์เพียงบางส่วนโดยไม่ได้รับอนุญาตจากห้องปฏิบัติการเป็นลายลักษณ์อักษร

Analysis/Test Report

Customer Name : บริษัท เอ็ม กรีน กรุ๊ป จำกัด

Address : 188/46 ถนนประชาอุทิศ แขวงทุ่งครุ เขตทุ่งครุ กรุงเทพมหานคร 10140

Sampling Site : Job No.6807225

Sample Type : น้ำเสีย

Sampling by : ลูกค้า (บริษัท เอ็ม กรีน กรุ๊ป จำกัด)

Sampling Method : Grab

Sampling Date : 07/07/2568

Sampling Time : 10:13 น.

Received Date : 15/07/2568

Analytical Date : 15 - 18/07/2568

Report Date : 19/07/2568

Report No. : RS16207/68

Parameters	Unit	Method	TS17328 /68	มาตรฐาน ^a (อาคารประเภท ค)
			น้ำทิ้งก่อนบำบัด	
Total Kjeldahl Nitrogen	mg/L as N	SM 2023 (4500 N _{org} B)	16.1	≤ 40
Residual Chlorine	mg/L as Cl ₂	DPD Colorimetric	< 0.10	-
Fecal Coliform Bacteria	MPN/100 mL	SM 2023 (9221 E)	1.7 x 10 ⁴	-
Sample Condition		Observation	เหลืองจาง มีตะกอน	

Remark : 1. SM 2023 : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24th ed., 20232. a : อ้างอิงตามประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุมการระบายน้ำทิ้งจากอาคาร
บางประเภทและบางขนาด พ.ศ. 2567

K.Kornkanok

Miss KORNBANOK KHUNPITAK

Analyst

19/07/2568


Miss ORASA YUBUA
Technical Manager
19/07/2568

Reported results refer to the sample as received only.

Test report shall not be reproduced except in full, without written approved of the laboratory.

The laboratory has been accepted as an accredited laboratory complying with the ISO/IEC 17025

Analysis/Test Report

Customer Name : บริษัท เอ็ม กรีน กรุ๊ป จำกัด

Address : 188/46 ถนนประชาอุทิศ แขวงทุ่งครุ เขตทุ่งครุ กรุงเทพมหานคร 10140

Sampling Site : Job No.6807225

Sample Type : น้ำเสีย

Sampling by : ลูกค้า (บริษัท เอ็ม กรีน กรุ๊ป จำกัด)

Sampling Method : Grab

Sampling Date : 07/07/2568

Sampling Time : 10:15 น.

Received Date : 15/07/2568

Analytical Date : 15 - 18/07/2568

Report Date : 19/07/2568

Report No. : RS16208/68

Parameters	Unit	Method	TS17329 /68	มาตรฐาน ^ก (อาคารประเภท ค)
			บ่อน้ำทิ้งก่อนปล่อย	
Total Kjeldahl Nitrogen	mg/L as N	SM 2023 (4500 N _{org} B)	6.0	≤ 40
Residual Chlorine	mg/L as Cl ₂	DPD Colorimetric	< 0.10	-
Fecal Coliform Bacteria	MPN/100 mL	SM 2023 (9221 E)	9.2 x 10 ³	-
Sample Condition		Observation	เหลืองจาง มีตะกอน	

Remark : 1. SM 2023 : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24th ed., 20232. ก : อ้างอิงตามประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุมการระบายน้ำทิ้งจากอาคาร
บางประเภทและบางขนาด พ.ศ. 2567

K. Kornkanok

Miss KORNNANOK KHUNPITAK

Analyst

19/07/2568

Miss ORASA YUBUA
Technical Manager

19/07/2568

Reported results refer to the sample as received only.

Test report shall not be reproduced except in full, without written approved of the laboratory.

The laboratory has been accepted as an accredited laboratory complying with the ISO/IEC 17025

REF NO. : MGG082025225

REPORT NO. : JOB01225

ต้นฉบับ

รายงานผลการวิเคราะห์คุณภาพน้ำ

เลขโครงการ : JB6808241 เลขที่รายงาน : WW6808241
บริษัท/โครงการ : โครงการ โรงแรม วิสา หัวหิน
ที่อยู่ของโครงการ : 222/222 ถนนเพชรเกษม ตำบลหัวหิน อำเภอหัวหิน จังหวัดประจวบคีรีขันธ์ 77110
วันที่เก็บตัวอย่าง : 11 สิงหาคม 2568 วันที่รับตัวอย่าง : 11 สิงหาคม 2568
วันที่วิเคราะห์ : 11 สิงหาคม - 02 กันยายน 2568 วันรายงานผล : 3 กันยายน 2568
ผู้เก็บตัวอย่าง : นายมะห์เตร์ รอหมาน (ว-299-จ-0003) ประเภทตัวอย่าง : Wastewater

พารามิเตอร์	วิธีวิเคราะห์	ผลการวิเคราะห์		มาตรฐาน	หน่วย
		น้ำทิ้งก่อนบำบัด	บ่อกักน้ำทิ้งก่อนปล่อย		
pH	Electrometric Method	7.4	7.3	5.5-9.0	-
Total Dissolved Solids	Dried at 180 °C	385	175	≤1,300	mg/L
Total Suspended Solids	Dried at 103-105 °C	72	43	≤50	mg/L
Settleable Solids**	Settleable Solids	11.5	6.5	-	mL/L
Sulfide	Iodometric Method	0.8	0.5	≤1.0	mg/L
Oil and grease	Soxhlet Extraction Method	3.8	3.0	≤20	mg/L
Biochemical Oxygen Demand	Azide Modification & 5-Day BOD Test Method	122	3	≤40	mg/L

หมายเหตุ : มาตรฐาน : ประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุม การระบายน้ำทิ้งจากอาคารบางประเภทและบางขนาด พ.ศ.2567 (ประเภท ค)

Method : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23rd Edition, 2017.

Settleable Solids** : อยู่นอกเหนือขอบข่ายการขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม

ลักษณะตัวอย่าง : น้ำทิ้งก่อนบำบัด: เหลืองจาง ชุ่น มีตะกอน

บ่อกักน้ำทิ้งก่อนปล่อย: เหลืองจาง ชุ่น มีตะกอน

ห้องปฏิบัติการวิเคราะห์เอกชน

เลขทะเบียนที่ ว-299



ลงชื่อ.....*Rinyarat sua*.....รับรอง

(นางสาวริญญารัตน์ ศิริสุนทรพิงษ์)

ผู้ควบคุมห้องปฏิบัติการ

ว-299-ค-0001

ผลการวิเคราะห์นี้รับรองเฉพาะตัวอย่างที่ได้ทำการวิเคราะห์เท่านั้น

ห้ามคัดลอกหรือทำสำเนารายงานผลการวิเคราะห์เพียงบางส่วนโดยไม่ได้รับอนุญาตจากห้องปฏิบัติการเป็นลายลักษณ์อักษร

Analysis/Test Report

Customer Name : บริษัท เอ็ม กรีน กรุ๊ป จำกัด

Address : 188/46 ถนนประชาอุทิศ แขวงทุ่งครุ เขตทุ่งครุ กรุงเทพมหานคร 10140

Sampling Site : Job No.6808241

Sample Type : น้ำเสีย

Sampling by : ลูกค้า (บริษัท เอ็ม กรีน กรุ๊ป จำกัด)

Sampling Method : Grab

Sampling Date : 11/08/2568

Sampling Time : 11:08 น.

Received Date : 19/08/2568

Analytical Date : 19 - 22/08/2568

Report Date : 25/08/2568

Report No. : RS18998/68

Parameters	Unit	Method	TS20223 /68	มาตรฐาน ^a (อาคารประเภท ก)
			น้ำทิ้งก่อนบำบัด	
Total Kjeldahl Nitrogen	mg/L as N	SM 2023 (4500 N _{org} B)	35.0	≤ 40
Residual Chlorine	mg/L as Cl ₂	DPD Colorimetric	< 0.10	-
Fecal Coliform Bacteria	MPN/100 mL	SM 2023 (9221 E)	2.8 x 10 ⁴	-
Sample Condition		Observation	เหลืองขุ่น	

Remark : 1. SM 2023 : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24th ed., 20232. a : อ้างอิงตามประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุมการระบายน้ำทิ้งจากอาคาร
บางประเภทและบางขนาด พ.ศ. 2567

K. Kornkanok

Miss KORNKANOK KHUNPITAK

Analyst

25/08/2568


Miss ORASA YUBUA
TEST TECH CO., LTD.
Technical Manager

25/08/2568

Reported results refer to the sample as received only.

Test report shall not be reproduced except in full, without written approved of the laboratory.

The laboratory has been accepted as an accredited laboratory complying with the ISO/IEC 17025

Analysis/Test Report

Customer Name : บริษัท เอ็ม กรีน กรุ๊ป จำกัด

Address : 188/46 ถนนประชาอุทิศ แขวงทุ่งครุ เขตทุ่งครุ กรุงเทพมหานคร 10140

Sampling Site : Job No.6808241

Sample Type : น้ำเสีย

Sampling by : ลูกค้า (บริษัท เอ็ม กรีน กรุ๊ป จำกัด)

Sampling Method : Grab

Sampling Date : 11/08/2568

Sampling Time : 11:11 น.

Received Date : 19/08/2568

Analytical Date : 19 - 22/08/2568

Report Date : 25/08/2568

Report No. : RS18999/68

Parameters	Unit	Method	TS20224 /68	มาตรฐาน ^a (อาคารประเภท ค)
			บ่อพักน้ำทิ้งก่อนปล่อย	
Total Kjeldahl Nitrogen	mg/L as N	SM 2023 (4500 N _{org} B)	15.4	≤ 40
Residual Chlorine	mg/L as Cl ₂	DPD Colorimetric	< 0.10	-
Fecal Coliform Bacteria	MPN/100 mL	SM 2023 (9221 E)	1.3 x 10 ⁴	-
Sample Condition		Observation	เหลือกลิ่น	

Remark : 1. SM 2023 : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24th ed., 20232. a : อ้างอิงตามประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุมการระบายน้ำทิ้งจากอาคาร
บางประเภทและบางขนาด พ.ศ. 2567

K. Kornkanok

Miss KORNKANOK KHUNPITAK

Analyst

25/08/2568



Miss ORASA YUBUA

Technical Manager

25/08/2568

Reported results refer to the sample as received only.

Test report shall not be reproduced except in full, without written approved of the laboratory.

The laboratory has been accepted as an accredited laboratory complying with the ISO/IEC 17025

REF NO. : MGG092025253

REPORT NO. : JOB01253

ต้นฉบับ

รายงานผลการวิเคราะห์คุณภาพน้ำ

เลขโครงการ : JB6809259 เลขที่รายงาน : WW6809259
บริษัท/โครงการ : โครงการ โรงแรม วิสา หัวหิน
ที่อยู่ของโครงการ : 222/222 ถนนเพชรเกษม ตำบลหัวหิน อำเภอหัวหิน จังหวัดประจวบคีรีขันธ์ 77110
วันที่เก็บตัวอย่าง : 8 กันยายน 2568 วันที่รับตัวอย่าง : 8 กันยายน 2568
วันที่วิเคราะห์ : 8-17 กันยายน 2568 วันรายงานผล : 22 กันยายน 2568
ผู้เก็บตัวอย่าง : นายมะห์เตร์ รอหมาน (ว-299-จ-0003) ประเภทตัวอย่าง : Wastewater

พารามิเตอร์	วิธีวิเคราะห์	ผลการวิเคราะห์		มาตรฐาน	หน่วย
		น้ำทิ้งก่อนบำบัด	บ่อกักน้ำทิ้งก่อนปล่อย		
pH	Electrometric Method	7.2	7.5	5.5-9.0	-
Total Dissolved Solids	Dried at 180 °C	505	282	≤1,300	mg/L
Total Suspended Solids	Dried at 103-105 °C	103	4.3	≤50	mg/L
Settleable Solids**	Settleable Solids	6.1	<0.1	-	mL/L
Sulfide	Iodometric Method	1.5	0.6	≤1.0	mg/L
Oil and grease	Soxhlet Extraction Method	12.7	8.9	≤20	mg/L
Biochemical Oxygen Demand	Azide Modification & 5-Day BOD Test Method	83	4	≤40	mg/L

หมายเหตุ : มาตรฐาน : ประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุม การระบายน้ำทิ้งจากอาคารบางประเภทและบางขนาด พ.ศ.2567 (ประเภท ค)

Method : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23rd Edition, 2017.

Settleable Solids** : อยู่นอกเหนือขอบข่ายการขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม

ลักษณะตัวอย่าง : น้ำทิ้งก่อนบำบัด: เหลืองขุ่น มีตะกอน

บ่อกักน้ำทิ้งก่อนปล่อย: เหลืองจาง มีตะกอนละเอียด

ห้องปฏิบัติการวิเคราะห์เอกชน

เลขทะเบียนที่ ว-299



ลงชื่อ..... Rinyarat sia.....รับรอง

(นางสาวริญญารัตน์ ศิริสุนทรพิงษ์)

ผู้ควบคุมห้องปฏิบัติการ

ว-299-ค-0001

ผลการวิเคราะห์นี้รับรองเฉพาะตัวอย่างที่ได้ทำการวิเคราะห์เท่านั้น

ห้ามคัดลอกหรือทำสำเนารายงานผลการวิเคราะห์เพียงบางส่วนโดยไม่ได้รับอนุญาตจากห้องปฏิบัติการเป็นลายลักษณ์อักษร

Analysis/Test Report

Customer Name : บริษัท เอ็ม กรีน กรู๊ป จำกัด

Address : 188/46 ถนนประชาธิปไตย แขวงทุ่งครุ เขตทุ่งครุ กรุงเทพมหานคร 10140

Sampling Site : Job No.6809259

Sample Type : น้ำเสีย

Sampling by : ลูกค้า (บริษัท เอ็ม กรีน กรู๊ป จำกัด)

Sampling Method : Grab

Sampling Date : 08/09/2568

Sampling Time : 10:48 น.

Received Date : 13/09/2568

Analytical Date : 13 - 17/09/2568

Report Date : 18/09/2568

Report No. : RS21276/68

Parameters	Unit	Method	TS22727 /68	มาตรฐาน ^ก (อาคารประเภท ค)
			น้ำทิ้งก่อนบำบัด	
Total Kjeldahl Nitrogen	mg/L as N	SM 2023 (4500 N _{org} B)	28.0	≤ 40
Residual Chlorine	mg/L as Cl ₂	DPD Colorimetric	< 0.10	-
Fecal Coliform Bacteria	MPN/100 mL	SM 2023 (9221 E)	2.4 x 10 ³	-
Sample Condition		Observation	เหลืองขุ่น	

Remark : 1. SM 2023 : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24th ed., 2023

2. ก : อ้างอิงตามประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุมการระบายน้ำทิ้งจากอาคาร
บางประเภทและบางขนาด พ.ศ. 2567

H. Pornrawee

Miss PORNRWEE HIRANCHAROEN

Analyst

18/09/2568



Technical Manager

18/09/2568

Reported results refer to the sample as received only.

Test report shall not be reproduced except in full, without written approved of the laboratory.

The laboratory has been accepted as an accredited laboratory complying with the ISO/IEC 17025

Analysis/Test Report

Customer Name : บริษัท เอ็ม กรีน กรุ๊ป จำกัด

Address : 188/46 ถนนประชาธิปไตย แขวงทุ่งครุ เขตทุ่งครุ กรุงเทพมหานคร 10140

Sampling Site : Job No.6809259

Sample Type : น้ำเสีย

Sampling by : ลูกค้า (บริษัท เอ็ม กรีน กรุ๊ป จำกัด)

Sampling Method : Grab

Sampling Date : 08/09/2568

Sampling Time : 10:50 น.

Received Date : 13/09/2568

Analytical Date : 13 - 17/09/2568

Report Date : 18/09/2568

Report No. : RS21277/68

Parameters	Unit	Method	TS22728 /68	มาตรฐาน (อาคารประเภท ค)
			บ่อพักน้ำทิ้งก่อนปล่อย	
Total Kjeldahl Nitrogen	mg/L as N	SM 2023 (4500 N _{org} B)	8.0	≤ 40
Residual Chlorine	mg/L as Cl ₂	DPD Colorimetric	< 0.10	-
Fecal Coliform Bacteria	MPN/100 mL	SM 2023 (9221 E)	1.1 x 10 ³	-
Sample Condition		Observation	เหลืองขาง มีตะกอนละเอียด	

Remark : 1. SM 2023 : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24th ed., 20232. a : อ้างอิงตามประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุมการระบายน้ำทิ้งจากอาคาร
บางประเภทและบางขนาด พ.ศ. 2567

H. Pornraee

Miss PORNRAWEE HIRANCHAROEN

Analyst

18/09/2568



Miss ORASA LTD. YUBUA

Technical Manager

18/09/2568

Reported results refer to the sample as received only.

Test report shall not be reproduced except in full, without written approved of the laboratory.

The laboratory has been accepted as an accredited laboratory complying with the ISO/IEC 17025

REF NO. : MGG102025279

REPORT NO. : JOB01279

ต้นฉบับ

รายงานผลการวิเคราะห์คุณภาพน้ำ

เลขโครงการ : JB6810282 เลขที่รายงาน : WW6810282
บริษัท/โครงการ : โครงการ โรงแรม วิสา หัวหิน
ที่อยู่ของโครงการ : 222/222 ถนนเพชรเกษม ตำบลหัวหิน อำเภอหัวหิน จังหวัดประจวบคีรีขันธ์ 77110
วันที่เก็บตัวอย่าง : 6 ตุลาคม 2568 วันที่รับตัวอย่าง : 6 ตุลาคม 2568
วันที่วิเคราะห์ : 6-17 ตุลาคม 2568 วันรายงานผล : 17 ตุลาคม 2568
ผู้เก็บตัวอย่าง : นายมะห์เตร์ รอหมาน (ว-299-จ-0003) ประเภทตัวอย่าง : Wastewater

พารามิเตอร์	วิธีวิเคราะห์	ผลการวิเคราะห์		มาตรฐาน	หน่วย
		น้ำทิ้งก่อนบำบัด	บ่อกักน้ำทิ้งก่อนปล่อย		
pH	Electrometric Method	7.0	7.1	5.5-9.0	-
Total Dissolved Solids	Dried at 180 °C	531	277	≤1,300	mg/L
Total Suspended Solids	Dried at 103-105 °C	35.4	61.2	≤50	mg/L
Settleable Solids**	Settleable Solids	1.5	5.5	-	mL/L
Sulfide	Iodometric Method	0.8	0.5	≤1.0	mg/L
Oil and grease	Soxhlet Extraction Method	6.0	1.2	≤20	mg/L
Biochemical Oxygen Demand	Azide Modification & 5-Day BOD Test Method	20	12	≤40	mg/L

หมายเหตุ : มาตรฐาน : ประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุม การระบายน้ำทิ้งจากอาคารบางประเภทและบางขนาด พ.ศ.2567 (ประเภท ค)

Method : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23rd Edition, 2017.

Settleable Solids** : อยู่นอกเหนือขอบข่ายการขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม

ลักษณะตัวอย่าง : น้ำทิ้งก่อนบำบัด: เหลืองขุ่น มีตะกอน
บ่อกักน้ำทิ้งก่อนปล่อย: เหลืองขุ่น มีตะกอน

ห้องปฏิบัติการวิเคราะห์เอกชน

เลขทะเบียนที่ ว-299



ลงชื่อ...เพลินพรรณ บุญศักดิ์...รับรอง

(นางสาวเพลินพรรณ บุญศักดิ์)

ผู้ควบคุมห้องปฏิบัติการ

ว-299-ค-0003

ผลการวิเคราะห์นี้รับรองเฉพาะตัวอย่างที่ได้ทำการวิเคราะห์เท่านั้น

ห้ามคัดลอกหรือทำสำเนารายงานผลการวิเคราะห์เพียงบางส่วนโดยไม่ได้รับอนุญาตจากห้องปฏิบัติการเป็นลายลักษณ์อักษร

Analysis/Test Report

Customer Name : บริษัท เอ็ม กรีน กรุ๊ป จำกัด

Address : 188/46 ถนนประชาอุทิศ แขวงทุ่งครุ เขตทุ่งครุ กรุงเทพมหานคร 10140

Sampling Site : Job No.6810282

Sample Type : น้ำเสีย

Sampling by : ลูกค้า (บริษัท เอ็ม กรีน กรุ๊ป จำกัด)

Sampling Method : Grab

Sampling Date : 06/10/2568

Sampling Time : 11:24 น.

Received Date : 15/10/2568

Analytical Date : 15 - 18/10/2568

Report Date : 20/10/2568

Report No. : RS23936/68

Parameters	Unit	Method	TS25467 /68	มาตรฐาน ^a (อาคารประเภท ค)
			น้ำทิ้งก่อนบำบัด	
Total Kjeldahl Nitrogen	mg/L as N	SM 2023 (4500 N _{org} B)	14.7	≤ 40
Residual Chlorine	mg/L as Cl ₂	SM 2023 (4500-Cl G)	< 0.10	-
Fecal Coliform Bacteria	MPN/100 mL	SM 2023 (9221 E)	9.4 x 10 ³	-
Sample Condition		Observation	เหลือกลิ่น มีตะกอน	

Remark : 1. SM 2023 : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24th ed., 20232. a : อ้างอิงตามประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุมการระบายน้ำทิ้งจากอาคาร
บางประเภทและบางขนาด พ.ศ. 2567

H. Pornrawee

Miss PORNRAWEE HIRANCHAOEN

Analyst

20/10/2568


Miss ORASA YUBUA
TEST TECH CO., LTD.
Technical Manager

20/10/2568

Reported results refer to the sample as received only.

Test report shall not be reproduced except in full, without written approved of the laboratory.

The laboratory has been accepted as an accredited laboratory complying with the ISO/IEC 17025

Analysis/Test Report

Customer Name : บริษัท เอ็ม กรีน กรุ๊ป จำกัด

Address : 188/46 ถนนประชาอุทิศ แขวงทุ่งครุ เขตทุ่งครุ กรุงเทพมหานคร 10140

Sampling Site : Job No.6810282

Sampling by : ลูกค้า (บริษัท เอ็ม กรีน กรุ๊ป จำกัด)

Sampling Date : 06/10/2568

Received Date : 15/10/2568

Report Date : 20/10/2568

Sample Type : น้ำเสีย

Sampling Method : Grab

Sampling Time : 11:29 น.

Analytical Date : 15 - 18/10/2568

Report No. : RS23937/68

Parameters	Unit	Method	TS25468 /68	มาตรฐาน ^a (อาคารประเภท ค)
			ป้อนค่าที่ห้องก่อนปล่อย	
Total Kjeldahl Nitrogen	mg/L as N	SM 2023 (4500 N _{org} B)	10.2	≤ 40
Residual Chlorine	mg/L as Cl ₂	SM 2023 (4500-Cl G)	< 0.10	-
Fecal Coliform Bacteria	MPN/100 mL	SM 2023 (9221 E)	9.2 x 10 ³	-
Sample Condition		Observation	เหลืองขุ่น มีตะกอน	

Remark : 1. SM 2023 : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24th ed., 20232. a : อ้างอิงตามประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุมการระบายน้ำทิ้งจากอาคาร
บางประเภทและบางขนาด พ.ศ. 2567

H. Pornrawee

Miss PORNRAWEE HIRANCHAROEN

Analyst

20/10/2568


Miss ORISA YUBUA
TEST TECH CO., LTD.
Technical Manager
20/10/2568

Reported results refer to the sample as received only.

Test report shall not be reproduced except in full, without written approved of the laboratory.

The laboratory has been accepted as an accredited laboratory complying with the ISO/IEC 17025

REF NO. : MGG112025298

REPORT NO. : JOB01298

ต้นฉบับ

รายงานผลการวิเคราะห์คุณภาพน้ำ

เลขโครงการ : JB6811300 เลขที่รายงาน : WW6811300
บริษัท/โครงการ : โครงการ โรงแรม วิสา หัวหิน
ที่อยู่ของโครงการ : 222/222 ถนนเพชรเกษม ตำบลหัวหิน อำเภอหัวหิน จังหวัดประจวบคีรีขันธ์ 77110
วันที่เก็บตัวอย่าง : 10 พฤศจิกายน 2568 วันที่รับตัวอย่าง : 10 พฤศจิกายน 2568
วันที่วิเคราะห์ : 10-27 พฤศจิกายน 2568 วันรายงานผล : 1 ธันวาคม 2568
ผู้เก็บตัวอย่าง : นายมะห์เตร์ รอหมาน (ว-299-จ-0003) ประเภทตัวอย่าง : Wastewater

พารามิเตอร์	วิธีวิเคราะห์	ผลการวิเคราะห์		มาตรฐาน	หน่วย
		น้ำทิ้งก่อนบำบัด	บ่อกักน้ำทิ้งก่อนปล่อย		
pH	Electrometric Method	7.6	7.5	5.5-9.0	-
Total Dissolved Solids	Dried at 180 °C	284	180	≤1,300	mg/L
Total Suspended Solids	Dried at 103-105 °C	24.3	11	≤50	mg/L
Settleable Solids**	Settleable Solids	0.2	0.2	-	mL/L
Sulfide	Iodometric Method	0.7	0.5	≤1.0	mg/L
Oil and grease	Soxhlet Extraction Method	8.0	4.0	≤20	mg/L
Biochemical Oxygen Demand	Azide Modification & 5-Day BOD Test Method	20	15	≤40	mg/L

หมายเหตุ : มาตรฐาน : ประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุม การระบายน้ำทิ้งจากอาคารบางประเภทและบางขนาด พ.ศ.2567 (ประเภท ค)

Method : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23rd Edition, 2017.

Settleable Solids** : อยู่นอกเหนือขอบข่ายการขึ้นทะเบียนกับกรมโรงงานอุตสาหกรรม

ลักษณะตัวอย่าง : น้ำทิ้งก่อนบำบัด: เหลืองจางขุ่น มีตะกอน

บ่อกักน้ำทิ้งก่อนปล่อย: เหลืองจาง มีตะกอน

ห้องปฏิบัติการวิเคราะห์เอกชน

เลขทะเบียนที่ ว-299



ลงชื่อ.....เพลินพรรณ บุนยศักดิ์.....รับรอง

(นางสาวเพลินพรรณ บุนยศักดิ์)

ผู้ควบคุมห้องปฏิบัติการ

ว-299-ค-0003

ผลการวิเคราะห์นี้รับรองเฉพาะตัวอย่างที่ได้ทำการวิเคราะห์เท่านั้น

ห้ามคัดลอกหรือทำสำเนารายงานผลการวิเคราะห์เพียงบางส่วนโดยไม่ได้รับอนุญาตจากห้องปฏิบัติการเป็นลายลักษณ์อักษร

Analysis/Test Report

Customer Name : บริษัท เอ็ม กรีน กรุ๊ป จำกัด

Address : 188/46 ถนนประชาอุทิศ แขวงทุ่งครุ เขตทุ่งครุ กรุงเทพมหานคร 10140

Sampling Site : Job No.6811300

Sample Type : น้ำเสีย

Sampling by : ลูกค้า (บริษัท เอ็ม กรีน กรุ๊ป จำกัด)

Sampling Method : Grab

Sampling Date : 10/11/2568

Sampling Time : 11:06 น.

Received Date : 15/11/2568

Analytical Date : 15 - 20/11/2568

Report Date : 22/11/2568

Report No. : RS26955/68

Parameters	Unit	Method	TS28610 /68	มาตรฐาน ^a (อาคารประเภท ค)
			นำทิ้งก่อนบำบัด	
Total Kjeldahl Nitrogen	mg/L as N	SM 2023 (4500 N _{org} B)	42.7	≤ 40
Residual Chlorine	mg/L as Cl ₂	SM 2023 (4500-Cl G)	< 0.10	-
Fecal Coliform Bacteria	MPN/100 mL	SM 2023 (9221 E)	1.6 x 10 ⁴	-
Sample Condition		Observation	เหลือจางๆ มีตะกอน	

Remark : 1. SM 2023 : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24th ed., 20232. a : อ้างอิงตามประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุมการระบายน้ำทิ้งจากอาคาร
บางประเภทและบางขนาด พ.ศ. 2567

H. Pornrawee

Miss PORNRAWEE HIRANCHAROEN

Analyst

22/11/2568



22/11/2568

Reported results refer to the sample as received only.

Test report shall not be reproduced except in full, without written approved of the laboratory.

The laboratory has been accepted as an accredited laboratory complying with the ISO/IEC 17025

Analysis/Test Report

Customer Name : บริษัท เอ็ม กรีน กรุ๊ป จำกัด

Address : 188/46 ถนนประชาอุทิศ แขวงทุ่งครุ เขตทุ่งครุ กรุงเทพมหานคร 10140

Sampling Site : Job No.6811300

Sample Type : น้ำเสีย

Sampling by : ลูก้า (บริษัท เอ็ม กรีน กรุ๊ป จำกัด)

Sampling Method : Grab

Sampling Date : 10/11/2568

Sampling Time : 11:11 น.

Received Date : 15/11/2568

Analytical Date : 15 - 20/11/2568

Report Date : 22/11/2568

Report No. : RS26956/68

Parameters	Unit	Method	TS28611 /68	มาตรฐาน ^ก (อาคารประเภท ก)
			ข้อพิพาทก่อนปล่อย	
Total Kjeldahl Nitrogen	mg/L as N	SM 2023 (4500 N _{org} B)	18.2	≤ 40
Residual Chlorine	mg/L as Cl ₂	SM 2023 (4500-Cl G)	< 0.10	
Fecal Coliform Bacteria	MPN/100 mL	SM 2023 (9221 E)	7.0 x 10 ³	
Sample Condition		Observation	เหลืองจางๆ มีตะกอน	

Remark : 1. SM 2023 : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24th ed., 2023

2. ก : อ้างอิงตามประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุมการระบายน้ำทิ้งจากอาคารบางประเภทและบางขนาด พ.ศ. 2567

H. Pornrawee

Miss PORNRWEE HIRANCHAROEN

Analyst

22/11/2568



Miss ORASA YUBUA

Technical Manager

22/11/2568

Reported results refer to the sample as received only.

Test report shall not be reproduced except in full, without written approved of the laboratory.

The laboratory has been accepted as an accredited laboratory complying with the ISO/IEC 17025

REF NO. : MGG122025317

REPORT NO. : JOB01317

ต้นฉบับ

รายงานผลการวิเคราะห์คุณภาพน้ำ

เลขโครงการ : JB6812323 เลขที่รายงาน : WW6812323
บริษัท/โครงการ : โครงการ โรงแรม วิสา หัวหิน
ที่อยู่ของโครงการ : 222/222 ถนนเพชรเกษม ตำบลหัวหิน อำเภอหัวหิน จังหวัดประจวบคีรีขันธ์ 77110
วันที่เก็บตัวอย่าง : 15 ธันวาคม 2568 วันที่รับตัวอย่าง : 15 ธันวาคม 2568
วันที่วิเคราะห์ : 15-24 ธันวาคม 2568 วันรายงานผล : 25 ธันวาคม 2568
ผู้เก็บตัวอย่าง : นายมะห์เตร์ รอหมาน (ว-299-จ-0003) ประเภทตัวอย่าง : Wastewater

พารามิเตอร์	วิธีวิเคราะห์	ผลการวิเคราะห์		มาตรฐาน	หน่วย
		น้ำทิ้งก่อนบำบัด	บ่อกักน้ำทิ้งก่อนปล่อย		
pH	Electrometric Method	7.8	7.3	5.5-9.0	-
Total Dissolved Solids	Dried at 180 °C	254	173	≤1,300	mg/L
Total Suspended Solids	Dried at 103-105 °C	34	19	≤50	mg/L
Oil and grease	Soxhlet Extraction Method	5.1	3.1	≤20	mg/L

หมายเหตุ : มาตรฐาน : ประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุม การระบายน้ำทิ้งจากอาคารบางประเภทและบางขนาด พ.ศ.2567 (ประเภท ค)

Method : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23rd Edition, 2017.

ลักษณะตัวอย่าง : น้ำทิ้งก่อนบำบัด: เหลืองจางขุ่น มีตะกอน
บ่อกักน้ำทิ้งก่อนปล่อย: เหลืองจาง มีตะกอน

ห้องปฏิบัติการวิเคราะห์เอกชน

เลขทะเบียนที่ ว-299



ลงชื่อ...เพลินพรรณ บุนยศักดิ์...รับรอง

(นางสาวเพลินพรรณ บุนยศักดิ์)

ผู้ควบคุมห้องปฏิบัติการ

ว-299-ค-0003

ผลการวิเคราะห์รับรองเฉพาะตัวอย่างที่ได้ทำการวิเคราะห์เท่านั้น

ห้ามคัดลอกหรือทำสำเนารายงานผลการวิเคราะห์เพียงบางส่วนโดยไม่ได้รับอนุญาตจากห้องปฏิบัติการเป็นลายลักษณ์อักษร

REF NO. : MGG122025317

REPORT NO. : JOB01317

ต้นฉบับ

รายงานผลการวิเคราะห์คุณภาพน้ำ

เลขโครงการ : JB6812323 เลขที่รายงาน : WW6812323
บริษัท/โครงการ : โครงการ โรงแรม วิสา ห้วยหิน
ที่อยู่ของโครงการ : 222/222 ถนนเพชรเกษม ตำบลห้วยหิน อำเภอห้วยหิน จังหวัดประจวบคีรีขันธ์ 77110
วันที่เก็บตัวอย่าง : 15 ธันวาคม 2568 วันที่รับตัวอย่าง : 15 ธันวาคม 2568
วันที่วิเคราะห์ : 15 ธันวาคม 2568 วันรายงานผล : 25 ธันวาคม 2568
ผู้เก็บตัวอย่าง : นายมะห์เตร์ รอหมาน ประเภทตัวอย่าง : Wastewater
(บริษัท เอ็ม กรีน กรุ๊ป จำกัด)

พารามิเตอร์	วิธีวิเคราะห์	ผลการวิเคราะห์		มาตรฐาน	หน่วย
		น้ำทิ้งก่อนบำบัด	บ่อพักน้ำทิ้งก่อนปล่อย		
Settleable Solids	Settleable Solids	1.5	0.8	-	mL/L

หมายเหตุ : Method : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA, WEF, 23rd Edition, 2017.

ลักษณะตัวอย่าง : น้ำทิ้งก่อนบำบัด: เหลืองจางขุ่น มีตะกอน
บ่อพักน้ำทิ้งก่อนปล่อย: เหลืองจาง มีตะกอน

ห้องปฏิบัติการวิเคราะห์เอกชน

เลขทะเบียนที่ ว-299



ลงชื่อ.....เพลินพรรณ บุษยศักดิ์รับรอง

(นางสาวเพลินพรรณ บุษยศักดิ์)

ผู้ควบคุมห้องปฏิบัติการ

ผลการวิเคราะห์นี้รับรองเฉพาะตัวอย่างที่ได้ทำการวิเคราะห์เท่านั้น

ห้ามคัดลอกหรือทำสำเนารายงานผลการวิเคราะห์เพียงบางส่วนโดยไม่ได้รับอนุญาตจากห้องปฏิบัติการเป็นลายลักษณ์อักษร

Analysis/Test Report

Customer Name : บริษัท เอ็ม กรีน กรุ๊ป จำกัด

Address : 188/46 ถนนประชาอุทิศ แขวงทุ่งครุ เขตทุ่งครุ กรุงเทพมหานคร 10140

Sampling Site : Job No.6812323

Sampling by : ลูกค้า (บริษัท เอ็ม กรีน กรุ๊ป จำกัด)

Sampling Date : 15/12/2568

Received Date : 20/12/2568

Report Date : 05/01/2569

Sample Type : น้ำเสีย

Sampling Method : Grab

Sampling Time : 10:55 น.

Analytical Date : 20 - 27/12/2568

Report No. : RS00226/69

Parameters	Unit	Method	TS32358 /68	มาตรฐาน ^a (อาคารประเภท ค)
			น้ำทิ้งก่อนบำบัด	
BOD	mg/L	SM 2023 (5210 B, 4500-O G)	38	≤ 40
Total Kjeldahl Nitrogen	mg/L as N	SM 2023 (4500 N _{org} B)	40.6	≤ 40
Sulfide	mg/L as H ₂ S	SM 2023 (4500-S ²⁻ F)	< 0.30	≤ 1.0
Residual Chlorine	mg/L as Cl ₂	SM 2023 (4500-Cl G)	< 0.10	-
Fecal Coliform Bacteria	MPN/100 mL	SM 2023 (9221 E)	1.7 x 10 ⁴	-
Sample Condition		Observation	เหลืองจาง มีตะกอนน้ำตา	

Remark : 1. SM 2023 : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24th ed., 20232. a : อ้างอิงตามประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุมการระบายน้ำทิ้งจากอาคาร
บางประเภทและบางขนาด พ.ศ. 2567

H. Pornrawee

Miss PORNRAWEE HIRANCHAROEN

Analyst

05/01/2569


Miss ORASA YUBUA
TEST TECH CO., LTD.

Technical Manager

05/01/2569

Reported results refer to the sample as received only.

Test report shall not be reproduced except in full, without written approved of the laboratory.

The laboratory has been accepted as an accredited laboratory complying with the ISO/IEC 17025

Analysis/Test Report

Customer Name : บริษัท เอ็ม กรีน กรุ๊ป จำกัด

Address : 188/46 ถนนประชาอุทิศ แขวงทุ่งครุ เขตทุ่งครุ กรุงเทพมหานคร 10140

Sampling Site : Job No.6812323

Sample Type : น้ำเสีย

Sampling by : ลูกค้า (บริษัท เอ็ม กรีน กรุ๊ป จำกัด)

Sampling Method : Grab

Sampling Date : 15/12/2568

Sampling Time : 11:03 น.

Received Date : 20/12/2568

Analytical Date : 20 - 27/12/2568

Report Date : 08/01/2569

Report No. : RS00227-1/69

Parameters	Unit	Method	TS32359 /68	มาตรฐาน ^ก
			บ่อพักน้ำทิ้งก่อนปล่อย	(อาคารประเภท ค)
BOD	mg/L	SM 2023 (5210 B, 4500-O G)	16	≤ 40
Total Kjeldahl Nitrogen	mg/L as N	SM 2023 (4500 N _{org} B)	15.0	≤ 40
Sulfide	mg/L as H ₂ S	SM 2023 (4500-S ²⁻ F)	< 0.30	≤ 1.0
Residual Chlorine	mg/L as Cl ₂	SM 2023 (4500-Cl G)	< 0.10	-
Fecal Coliform Bacteria	MPN/100 mL	SM 2023 (9221 E)	2.4 x 10 ³	-
Sample Condition		Observation	เหลือจาก มีตะกอนเล็กน้อย	

Remark : 1. SM 2023 : Standard Methods for the Examination of Water and Wastewater, APHA, AWWA & WEF, 24th ed., 20232. a : อ้างอิงตามประกาศกระทรวงทรัพยากรธรรมชาติและสิ่งแวดล้อม เรื่อง กำหนดมาตรฐานควบคุมการระบายน้ำทิ้งจากอาคาร
บางประเภทและบางขนาด พ.ศ. 2567

H. Pornrawee

Miss PORNRAWEE HIRANCHAROEN

Analyst

08/01/2569



Miss ORASA YUBUA

Technical Manager

08/01/2569

Reported results refer to the sample as received only.

Test report shall not be reproduced except in full, without written approved of the laboratory.

The laboratory has been accepted as an accredited laboratory complying with the ISO/IEC 17025



ที่ อก ๐๓๑๐(๑)/ ๑๙ ๑ ๓ ๐

กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท
เขตราชเทวี กรุงเทพฯ ๑๐๕๐๐

๒๑ ธันวาคม ๒๕๖๕

เรื่อง ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

เรียน กรรมการผู้จัดการ บริษัท เอ็ม กรีน กรุ๊ป จำกัด

อ้างถึง คำขอขึ้นทะเบียน/เปลี่ยนแปลงบุคลากร และชนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน
ลงวันที่ ๖ พฤศจิกายน ๒๕๖๕

สิ่งที่ส่งมาด้วย เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
บริษัท เอ็ม กรีน กรุ๊ป จำกัด จำนวน ๑ แผ่น

ตามหนังสือที่อ้างถึง บริษัท เอ็ม กรีน กรุ๊ป จำกัด ขอต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการ
วิเคราะห์เอกชน เลขทะเบียน ว-๒๙๙ สถานที่เลขที่ ๑๘๘/๔๖ ถนนประชาอุทิศ แขวงทุ่งครุ เขตทุ่งครุ
กรุงเทพมหานคร ต่อกรมโรงงานอุตสาหกรรม นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว ให้บริษัท เอ็ม กรีน กรุ๊ป จำกัด ต่ออายุหนังสือรับขึ้นทะเบียน
ห้องปฏิบัติการวิเคราะห์เอกชน โดยมีองค์ประกอบดังนี้

ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์

๑) นางสาววิญญารัตน์ ศิริสุนทรพงษ์

ทะเบียนเลขที่ ว-๒๙๙-ค-๐๐๐๑

๒) นางสาวอารีรัตน์ ปิตถาภูมิพัชร

ทะเบียนเลขที่ ว-๒๙๙-ค-๐๐๐๒

ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์

๑) นางสาวพลีนพรรณ บุญยศักดิ์

ทะเบียนเลขที่ ว-๒๙๙-จ-๐๐๐๑

๒) นางสาวจินตนา มหาอ้น

ทะเบียนเลขที่ ว-๒๙๙-จ-๐๐๐๒

๓) นายมะห์เหม็ด รอดมาน

ทะเบียนเลขที่ ว-๒๙๙-จ-๐๐๐๓

๔) นางสาวตัสนัม ดอนิ

ทะเบียนเลขที่ ว-๒๙๙-จ-๐๐๐๔

๕) นายธีรทัศน์ ฉายศรีศิริ

ทะเบียนเลขที่ ว-๒๙๙-จ-๐๐๐๕

๖) นายธีรวิช เล่าปวีวรรณ

ทะเบียนเลขที่ ว-๒๙๙-จ-๐๐๐๖

ค. ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนให้วิเคราะห์ในน้ำเสีย ตามสิ่งที่ส่งมาด้วย

หนังสือฉบับนี้จะหมดอายุในวันที่ ๖ ธันวาคม ๒๕๖๘ หากประสงค์จะต่ออายุหนังสือ
รับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ให้ยื่นคำขอต่ออายุพร้อมเอกสารประกอบคำขอต่อ
กรมโรงงานอุตสาหกรรม ภายใน ๓๐ วัน ก่อนวันสิ้นอายุของหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
ทั้งนี้ สามารถยื่นคำขอผ่านระบบอิเล็กทรอนิกส์ได้ที่หน้าเว็บไซต์กรมโรงงานอุตสาหกรรม ตาม QR Code
ท้ายหนังสือฉบับนี้

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ

(นายประสม ดำรงพงษ์)

ผู้อำนวยการกองวิจัยและเฝ้าระวังมลพิษโรงงาน
ปฏิบัติราชการแทนอธิบดีกรมโรงงานอุตสาหกรรม

ยื่นคำขอผ่านระบบอิเล็กทรอนิกส์

กองวิจัยและเฝ้าระวังมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๕๓๐ ๖๓๑๒ ต่อ ๒๑๐๓-๕

โทรสาร ๐ ๒๕๓๐ ๖๓๑๒ ต่อ ๒๑๔๔

ไปรษณีย์อิเล็กทรอนิกส์ saraban@diw.mail.go.th



"อุตสาหกรรมก้าวไกล ประเทศไทยก้าวหน้า ร่วมกันพัฒนา อุตสาหกรรมสีเขียว"



เอกสารแนบท้ายหนังสือรับต่ออายุขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท เอ็ม กรีน กรุ๊ป จำกัด

เลขทะเบียน ว-๒๙๙

ที่ อก ๐๓๑๐(๑)/ ๑๙ ๑ ๓ ๐

ลงวันที่ ๒๑ ธันวาคม ๒๕๖๕

ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๘ รายการ

น้ำเสีย จำนวน 8 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Biochemical Oxygen Demand	5-Day BOD Test, Azide Modification Method
2	Chemical Oxygen Demand	Closed Reflux, Titrimetric Method
3	Oil & Grease	Soxhlet Extraction Method
4	pH	Electrometric Method
5	Sulfide	Iodometric Method
6	Temperature	Laboratory and Field Methods
7	Total Dissolved Solids	Dried at 180 °C
8	Total Suspended Solids	Dried at 103-105 °C

เอกสารอ้างอิง

APHA, AWWA, WEF. Standard Methods for the Examination of Water and Wastewater.

23rd ed. Washington, DC: APHA, 2017.

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ กองวิจัยและเฝ้าระวังมลพิษโรงงาน กรมโรงงานอุตสาหกรรม โทร. ๐ ๒๕๓๐ ๖๓๑๒ ต่อ ๒๑๐๓-๕

ที่ อก ๐๓๑๐(๑)/ ๑๗ ๑๑ ๖



กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ เขตราชเทวี
กรุงเทพมหานคร ๑๐๔๐๐

๑๖ ธันวาคม ๒๕๖๒

เรื่อง ขันทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

เรียน กรรมการผู้จัดการ บริษัท เอ็ม กรีน กรุป จำกัด

อ้างถึง คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และชนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน
ลงวันที่ ๒๓ กันยายน ๒๕๖๒

สิ่งที่ส่งมาด้วย เอกสารแนบท้ายหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
บริษัท เอ็ม กรีน กรุป จำกัด จำนวน ๑ แผ่น

ตามหนังสือที่อ้างถึง บริษัท เอ็ม กรีน กรุป จำกัด ขอขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
พร้อมรายชื่อผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ และรายการสาร
มลพิษที่จะทำการวิเคราะห์ ต่อกรมโรงงานอุตสาหกรรม นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว ให้บริษัท เอ็ม กรีน กรุป จำกัด ขึ้นทะเบียน
ห้องปฏิบัติการวิเคราะห์เอกชน มีเลขทะเบียน ว-๒๕๕๕ สถานที่ตั้งเลขที่ ๑๘๘/๔๖ ถนนประชาอุทิศ แขวงทุ่งครุ
เขตทุ่งครุ กรุงเทพมหานคร โดยมีองค์ประกอบดังนี้

ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์

นางสาววิญญารัตน์ ศิริสุนทรพงษ์

ทะเบียนเลขที่ ว-๒๕๕๕-ก-๘๔๒๔

ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์

๑) นางสาววิริยา สมตัว

ทะเบียนเลขที่ ว-๒๕๕๕-ก-๘๔๒๕

๒) นายทริท วรศัณห์

ทะเบียนเลขที่ ว-๒๕๕๕-ก-๘๔๒๖

ค. ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนให้วิเคราะห์ในน้ำเสีย จำนวน ๔ รายการ

ตามสิ่งที่ส่งมาด้วย

หนังสือฉบับนี้มีอายุ ๓ ปี นับจากวันที่กรมโรงงานอุตสาหกรรมออกหนังสือ
หากประสงค์จะต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ให้ยื่นคำขอต่ออายุพร้อม
เอกสารประกอบคำขอต่อกรมโรงงานอุตสาหกรรมภายใน ๓๐ วัน ก่อนวันสิ้นอายุของหนังสือรับ
ขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ซึ่งคำขอต่ออายุดังกล่าวขอรับได้ที่กรมโรงงานอุตสาหกรรม

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ

(นางจินดา เตชะศิริพันธ์)

ผู้อำนวยการกองวิจัยและเฝ้าระวังมลพิษโรงงาน

ปฏิบัติราชการแทนอธิบดีกรมโรงงานอุตสาหกรรม

กองวิจัยและเฝ้าระวังมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๒๐๒ ๔๑๔๖ ๐ ๒๒๐๒ ๔๐๐๒

โทรสาร ๐ ๒๓๕๔ ๓๒๐๘ ๐ ๒๓๕๔ ๓๔๑๕

เอกสารแนบท้ายหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท เอ็ม กรีน กรุป จำกัด

เลขทะเบียน ว-๒๕๕๕

ที่ อก ๐๓๑๐(๑)/ ๑๗ ๑๑ ๖

ลงวันที่ ๑๖ ธันวาคม ๒๕๖๒

ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๔ รายการ

น้ำเสีย จำนวน 4 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	pH	Electrometric Method
2	Temperature	Laboratory and Field Methods
3	Total Dissolved Solids	Dried at 180 °C
4	Total Suspended Solids	Dried at 103-105 °C

เอกสารอ้างอิง

APHA, AWWA, WEF. Standard Methods for the Examination of Water and
Wastewater. 23rd ed. Washington, DC: APHA, 2017.

(นางวิภาญ์ วัฒนสุกุล)

ผู้อำนวยการกลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ

และทะเบียนห้องปฏิบัติการ

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ กองวิจัยและเฝ้าระวังมลพิษโรงงาน กรมโรงงานอุตสาหกรรม โทร. ๐ ๒๒๐๒ ๔๑๔๖

สำเนา

ที่ กก ๐๓๑๐(๑)/

กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ เขตราชเทวี
กรุงเทพมหานคร ๑๐๔๐๐

เรื่อง ขันทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

เรียน กรรมการผู้จัดการ บริษัท เอ็ม กรีน กรุ๊ป จำกัด

อ้างถึง คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และชนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน
ลงวันที่ ๒๓ กันยายน ๒๕๖๒

สิ่งที่ส่งมาด้วย เอกสารแนบท้ายหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
บริษัท เอ็ม กรีน กรุ๊ป จำกัด จำนวน ๑ แผ่น

ตามหนังสือที่อ้างถึง บริษัท เอ็ม กรีน กรุ๊ป จำกัด ขอขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน
พร้อมรายชื่อผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ และรายการสาร
มลพิษที่จะทำการวิเคราะห์ ต่อกรมโรงงานอุตสาหกรรม นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว ให้บริษัท เอ็ม กรีน กรุ๊ป จำกัด ขึ้นทะเบียน
ห้องปฏิบัติการวิเคราะห์เอกชน มีเลขทะเบียน ว-๒๕๙๙ สถานที่ตั้งเลขที่ ๑๘๘/๔๖ ถนนประชาอุทิศ แขวงทุ่งครุ
เขตทุ่งครุ กรุงเทพมหานคร โดยมีองค์ประกอบดังนี้

ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์

นางสาววิญญารัตน์ ศิริสุนทรพงษ์ ทะเบียนเลขที่ ว-๒๕๙๙-ค-๘๔๒๔

ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์

๑) นางสาววิริยา สมตัว ทะเบียนเลขที่ ว-๒๕๙๙-จ-๘๔๒๕

๒) นายหริทร วงศ์ธานี ทะเบียนเลขที่ ว-๒๕๙๙-จ-๘๔๒๖

ค. ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนให้วิเคราะห์ในน้ำเสีย จำนวน ๔ รายการ

ตามสิ่งที่ส่งมาด้วย

หนังสือฉบับนี้มีอายุ ๓ ปี นับจากวันที่กรมโรงงานอุตสาหกรรมออกหนังสือ

หากประสงค์จะต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ให้ยื่นคำขอต่ออายุพร้อม

เอกสารประกอบคำขอต่อกรมโรงงานอุตสาหกรรมภายใน ๓๐ วัน ก่อนวันสิ้นอายุของหนังสือรับ

ขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ซึ่งคำขอต่ออายุดังกล่าวขอรับได้ที่กรมโรงงานอุตสาหกรรม

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ



(นางจินดา เตชะศรีนทร์)

ผู้อำนวยการกองวิจัยและเฝ้าระวังมลพิษโรงงาน

ปฏิบัติราชการแทนอธิบดีกรมโรงงานอุตสาหกรรม

กองวิจัยและเฝ้าระวังมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๒๐๒ ๔๑๔๖ ๐ ๒๒๐๒ ๔๐๐๒

โทรสาร ๐ ๒๒๕๔ ๓๒๐๘ ๐ ๒๒๕๔ ๓๔๑๕

เอกสารแนบท้ายหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท เอ็ม กรีน กรุ๊ป จำกัด

เลขทะเบียน ว-๒๕๙๙

ที่ กก ๐๓๑๐(๑)/

ลงวันที่ ๐๖ กันยายน ๒๕๖๒

ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๔ รายการ

น้ำเสีย จำนวน 4 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	pH	Electrometric Method
2	Temperature	Laboratory and Field Methods
3	Total Dissolved Solids	Dried at 180 °C
4	Total Suspended Solids	Dried at 103-105 °C

เอกสารอ้างอิง

APHA, AWWA, WEF, Standard Methods for the Examination of Water and
Wastewater, 23rd ed. Washington, DC: APHA, 2017.



(นางจินดา เตชะศรีนทร์)

ผู้อำนวยการกองมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษ
และทะเบียนห้องปฏิบัติการ

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ กองวิจัยและเฝ้าระวังมลพิษโรงงาน กรมโรงงานอุตสาหกรรม โทร. ๐ ๒๒๐๒ ๔๑๔๖

สำเนา

ที่ อก ๐๓๑๐(๑)/ ๓ ๒ ๐ ๗ ๕

กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ เขตราชเทวี
กรุงเทพมหานคร ๑๐๕๐๐

๒๘ ตุลาคม ๒๕๖๓

เรื่อง เปลี่ยนแปลงบุคลากรของห้องปฏิบัติการวิเคราะห์

เรียน กรรมการผู้จัดการ บริษัท เอ็ม กรีน กรุ๊ป จำกัด


อ้างถึง หนังสือบริษัท เอ็ม กรีน กรุ๊ป จำกัด ที่ MG ๐๐๓/๒๕๖๓ ลงวันที่ ๒ ตุลาคม ๒๕๖๓

ตามหนังสือที่อ้างถึง บริษัท เอ็ม กรีน กรุ๊ป จำกัด ห้องปฏิบัติการวิเคราะห์เอกชน เลขทะเบียน ๖-๒๕๙๙ สถานที่ตั้งเลขที่ ๑๘๘/๔๖ ถนนประชาอุทิศ แขวงทุ่งครุ เขตทุ่งครุ กรุงเทพมหานคร ของห้องปฏิบัติการวิเคราะห์ ความละเอียดแจ้งแล้ว นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว ให้ยกเลิกเจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๑ ราย ได้แก่ นางสาววิริยา สมด้ ทะเบียนเลขที่ ๖-๒๕๙๙-จ-๘๔๒๕

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ



(นางจินดา เดชะชินทร์)

ผู้อำนวยการกองวิจัยและเตือนภัยมลพิษโรงงาน
ปฏิบัติการแผนกป้องกันมลพิษโรงงานอุตสาหกรรม

กองวิจัยและเตือนภัยมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๒๐๒ ๔๐๐๒ ๐ ๒๒๐๒ ๔๑๔๖

โทรสาร ๐ ๒๓๕๔ ๓๒๐๘ ๐ ๒๓๕๔ ๓๔๑๕



ที่ อก ๐๓๑๐(๑)/ ๕ ๗ ๘๕

กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท
เขตราชเทวี กรุงเทพฯ ๑๐๕๐๐

๑๕ ตุลาคม ๒๕๖๔

เรื่อง เปลี่ยนแปลงบุคลากรของห้องปฏิบัติการวิเคราะห์

เรียน กรรมการผู้จัดการ บริษัท เอ็ม กรีน กรุ๊ป จำกัด

อ้างถึง คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และชนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน ลงวันที่ ๑ ตุลาคม ๒๕๖๔

ตามหนังสือที่อ้างถึง บริษัท เอ็ม กรีน กรุ๊ป จำกัด ห้องปฏิบัติการวิเคราะห์เอกชน เลขทะเบียน ๖-๒๕๙๙ สถานที่ตั้งเลขที่ ๑๘๘/๔๖ ถนนประชาอุทิศ แขวงทุ่งครุ เขตทุ่งครุ กรุงเทพมหานคร ขอเปลี่ยนแปลงบุคลากร ของห้องปฏิบัติการวิเคราะห์ ความละเอียดแจ้งแล้ว นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว มีความเห็นดังนี้

๑. ให้ยกเลิกเจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๑ ราย

นายหริทร วงศ์ธานี ทะเบียนเลขที่ ๖-๒๕๙๙-จ-๘๔๒๖

๒. ให้เพิ่มเจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๒ ราย

๑) นางสาวอารีรัตน์ ปิตถาภูมิพัชร ทะเบียนเลขที่ ๖-๒๕๙๙-จ-๙๖๔๙

๒) นางสาวพลินพรรณ บุญยศักดิ์ ทะเบียนเลขที่ ๖-๒๕๙๙-จ-๙๖๕๐

อนึ่ง หนังสือฉบับนี้จะหมดอายุพร้อมหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน ที่ อก ๐๓๑๐(๑)/๑๗๓๓๓ ลงวันที่ ๖ ธันวาคม ๒๕๖๒ คือในวันที่ ๖ ธันวาคม ๒๕๖๕

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ



(นางจินดา เดชะชินทร์)

ผู้อำนวยการกองวิจัยและเตือนภัยมลพิษโรงงาน
ปฏิบัติการแผนกป้องกันมลพิษโรงงานอุตสาหกรรม

กองวิจัยและเตือนภัยมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๒๐๒ ๔๐๐๒ ๐ ๒๒๐๒ ๔๑๔๖

โทรสาร ๐ ๒๓๕๔ ๓๒๐๘

ไปรษณีย์อิเล็กทรอนิกส์ saraban@diw.gmail.go.th



ที่ อก ๐๓๑๐(๑)/๗/๓๕

กรมโรงงานอุตสาหกรรม
ถนนพระรามที่ ๖ แขวงทุ่งพญาไท
เขตราชเทวี กรุงเทพฯ ๑๐๔๐๐

๒๖ มกราคม ๒๕๖๗

เรื่อง ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

เรียน กรรมการผู้จัดการ บริษัท เทสท์ เทค จำกัด

อ้างถึง คำขอขึ้นทะเบียน/ต่ออายุ/เปลี่ยนแปลงบุคลากร และชนิดสารมลพิษของห้องปฏิบัติการวิเคราะห์เอกชน
ลงวันที่ ๒๗ ธันวาคม ๒๕๖๖

- สิ่งที่ส่งมาด้วย ๑. รายชื่อผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ จำนวน ๑ แผ่น
๒. รายชื่อเจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๒ แผ่น
๓. ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๑๔ แผ่น

ตามหนังสือที่อ้างถึง บริษัท เทสท์ เทค จำกัด ขอต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน เลขทะเบียน ว-๒๔๕ สถานที่ตั้งเลขที่ ๓๐,๓๒ ซอยพระรามที่ ๒ ซอย ๖๓ แขวงแสมดำ เขตบางขุนเทียน กรุงเทพมหานครต่อกรมโรงงานอุตสาหกรรม นั้น

กรมโรงงานอุตสาหกรรมพิจารณาแล้ว ให้บริษัท เทสท์ เทค จำกัด ต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน โดยมีองค์ประกอบดังนี้

- ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ จำนวน ๕ ราย ตามสิ่งที่ส่งมาด้วย ๑
ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ ๔๔ ราย ตามสิ่งที่ส่งมาด้วย ๒
ค. ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนให้วิเคราะห์ในน้ำเสีย น้ำใต้ดิน สิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว และดิน ตามสิ่งที่ส่งมาด้วย ๓

หนังสือฉบับนี้จะหมดอายุในวันที่ ๒๘ มกราคม ๒๕๗๐ หากประสงค์จะต่ออายุหนังสือรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน สามารถยื่นคำขอผ่านระบบอิเล็กทรอนิกส์ได้ที่หน้าเว็บไซต์กรมโรงงานอุตสาหกรรม

จึงเรียนมาเพื่อทราบ

ขอแสดงความนับถือ

(นายประสม ดำรงพงษ์)

ผู้อำนวยการกองวิจัยและเฝ้าระวังมลพิษโรงงาน
ปฏิบัติราชการแทนอธิบดีกรมโรงงานอุตสาหกรรม


กองวิจัยและเฝ้าระวังมลพิษโรงงาน

กลุ่มมาตรฐานวิธีการวิเคราะห์ทดสอบมลพิษและทะเบียนห้องปฏิบัติการ

โทร. ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๐๓-๔

โทรสาร ๐ ๒๔๓๐ ๖๓๑๒ ต่อ ๒๑๔๔

ไปรษณีย์อิเล็กทรอนิกส์ saraban@diw.mail.go.th



“อุตสาหกรรมก้าวไกล ประเทศไทยก้าวหน้า ร่วมกันพัฒนา อุตสาหกรรมสีเขียว”



สิ่งที่ส่งมาด้วย ๑

เอกสารแนบท้ายหนังสือต่ออายุรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท เทสท์ เทค จำกัด

เลขทะเบียน ว-๒๔๕

ที่ อก ๐๓๑๐(๑)/ ๗ ๓ ๕

ลงวันที่ ๒๖ มกราคม ๒๕๖๗

ก. ผู้ควบคุมดูแลห้องปฏิบัติการวิเคราะห์ จำนวน ๕ ราย

๑) นางสาวเรวดี ศิริมงคล

ทะเบียนเลขที่ ว-๒๔๕-ค-๐๐๐๑

๒) นางสาวอรษา อยู่บัว

ทะเบียนเลขที่ ว-๒๔๕-ค-๐๐๐๒

๓) นางสาวณัฐวิภา อ่อนจัน

ทะเบียนเลขที่ ว-๒๔๕-ค-๐๐๐๓

๔) นางสาวดวงกมล บุญยั้ง

ทะเบียนเลขที่ ว-๒๔๕-ค-๐๐๐๔

๕) นางสาวจิตรา ลิ่มสืบพงษ์

ทะเบียนเลขที่ ว-๒๔๕-ค-๐๐๐๕

3/กนฟ

เอกสารแนบท้ายหนังสือต่ออาชญากรขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท เทสท์ เทค จำกัด

เลขทะเบียน ๖-๒๔๕

ที่ อก ๐๓๑๐(๑)/

ลงวันที่

ข. เจ้าหน้าที่ประจำห้องปฏิบัติการวิเคราะห์ จำนวน ๔๔ ราย

- ๑) นางสาวปรังคิทธิ์ รักษาสุข
- ๒) นางสาวบุษยา ศรีสว่าง
- ๓) นางสาวมินตรา ทีภูงา
- ๔) นายเฉลิมชัย เจริญยิ่ง
- ๕) นายอภิสิทธิ์ คุณมาศ
- ๖) นายชาญวิทย์ อุทัยเลี้ยง
- ๗) นางสาวบุญสิตา พราหมณ์นุช
- ๘) นางสาวณัฐนิชา ทองลอย
- ๙) นางสาวรุติพร เชื้อรพ
- ๑๐) นางสาวกนกพร มั่นนิม
- ๑๑) นางสาวนันนิภา สิมพริกซ์
- ๑๒) นางสาวเบญจมาศ เชื้อหงษ์
- ๑๓) นางสาวสุกิมน ดวงอินทร์
- ๑๔) นางสาวนุสริทธิ์ สุนธักดิ์
- ๑๕) นางสาวชนิทธิ์ทิพย์ สุขประทุมเนตร
- ๑๖) นางสาวพลอยรุ่ง สุท่ามา
- ๑๗) นางสาวธิดยา ดีมาก
- ๑๘) นางสาวรัชนิกา ลือเฟื่อง
- ๑๙) นางสาวศิริندا คำดี
- ๒๐) นางสาวลลิตา เสนาบุษ
- ๒๑) นางสาวไศยสุภา ใจดีเจย
- ๒๒) นายวัฒนา พันธุ์เดช
- ๒๓) นางสาวอ้อยใจ สระจันทร์
- ๒๔) นางสาวมาริสรา วิเศษสังข์
- ๒๕) นายณัฐวุฒิ ใจสุภาพ
- ๒๖) นายกิจดิพงษ์ เขื่อนงาม
- ๒๗) นายไกรทอง สีเขอน
- ๒๘) นายสุริยา ชื่นบาน
- ๒๙) นางสาวรัตนันท์ ก้องสุรินทร์
- ๓๐) นางสาวนุสรา สุระเวก
- ๓๑) นางสาวนริศรา สอนบุญชู
- ๓๒) นางสาวผ่องอำไพ ยางงาม
- ๓๓) นางสาวนิศาชล อึ้งเกลี้ยง
- ๓๔) นางสาวนริศรา ผงพิลา
- ๓๕) นางสาวศัลลียา ท้าวหาญ

- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๐๔
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๐๕
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๐๗
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๐๘
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๐๘
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๑๐
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๑๑
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๑๒
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๑๓
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๑๔
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๑๕
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๑๖
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๑๗
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๑๘
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๑๙
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๒๐
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๒๑
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๒๒
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๒๓
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๒๔
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๒๕
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๒๖
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๒๗
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๒๘
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๒๙
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๓๐
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๓๑
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๓๒
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๓๓
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๓๔
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๓๕
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๓๖
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๓๗
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๓๘
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๓๙

๓๖) นางสาวกรกนก..

- ๓๖) นางสาวกรกนก ขุนพิทักษ์
- ๓๗) นางสาวดวงหทัย เริ่มวานิชย์
- ๓๘) นางสาวจุไรรัตน์ จงประกอบกิจ
- ๓๙) นายกิตติพิชญ์ ไข่เกตุ
- ๔๐) นางสาวเจนจิรา พลดี
- ๔๑) นางสาวชลนิภาณต์ สิทธิธรรม
- ๔๒) นางสาวณัฐการณ์ ขวัญศรี
- ๔๓) นายดิษฐ์วัฒน์ นราวิชัยธำรงค์
- ๔๔) นายธนพล สะเอียบคง

- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๔๐
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๔๑
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๔๒
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๔๓
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๔๔
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๔๕
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๔๖
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๔๗
- ทะเบียนเลขที่ ๖-๒๔๕-จ-๐๐๔๘

รวม

รวม

เอกสารแนบท้ายหนังสือต่ออายุรับขึ้นทะเบียนห้องปฏิบัติการวิเคราะห์เอกชน

บริษัท เทสท์ เทค จำกัด

เลขทะเบียน ๖-๒๔๔

ที่ กก ๐๓๑๐(๑)/

ลงวันที่

ค. ขอบข่ายสารมลพิษที่ได้รับขึ้นทะเบียนจากกรมโรงงานอุตสาหกรรม จำนวน ๑๓๘ รายการ

น้ำเสีย จำนวน 45 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Aldrin	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
2	Arsenic	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
3	Barium	1) Digestion, Direct Nitrous Oxide-Acetylene Flame Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
4	α-BHC	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
5	β-BHC	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
6	δ-BHC	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
7	γ-BHC	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
8	Biochemical Oxygen Demand	1) 5-Day BOD Test, Azide Modification Method ^[3] 2) 5-Day BOD Test, Membrane Electrode Method ^[3]
9	Cadmium	1) Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
10	Chemical Oxygen Demand	1) Open Reflux, Titrimetric Method ^[3] 2) Closed Reflux, Colorimetric Method ^[3] 3) Closed Reflux, Titrimetric Method ^[3]
11	Chlordane	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[3]
12	Chromium	1) Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
13	Color	ADMI Weighted-Ordinate Spectrophotometric Method ^[3]
14	Copper	1) Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
15	Cyanide	Distillation, Colorimetric Method ^[3]
16	4,4'-DDD	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
17	4,4'-DDE	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
18	4,4'-DDT	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
19	Dieldrin	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
20	Endosulfan I	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
21	Endosulfan II	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
22	Endosulfan Sulfate	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
23	Endrin	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
24	Endrin Aldehyde	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
25	Formaldehyde	Distillation, Colorimetric Method ^[2]
26	Free Chlorine	1) Iodometric Method ^[3] 2) DPD Colorimetric Method ^[3]
27	Heptachlor	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
28	Heptachlor Epoxide	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
29	Hexavalent Chromium	Colorimetric Method ^[3]

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
30	Lead	1) Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
31	Manganese	1) Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
32	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[3]
33	Methoxychlor	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[3]
34	Nickel	1) Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
35	Oil & Grease	1) Liquid-Liquid, Partition-Gravimetric Method ^[3] 2) Soxhlet Extraction Method ^[3]
36	pH	Electrometric Method ^[3]
37	Phenols	Distillation, Direct Photometric Method ^[3]
38	Selenium	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
39	Sulfide	1) Iodometric Method ^[3] 2) Methylene blue Method ^[3]
40	Temperature	Laboratory and Field Methods ^[3]
41	Total Dissolved Solids	Dried at 180 °C ^[3]
42	Total Kjeldahl Nitrogen	Macro-Kjeldahl Method ^[3]
43	Total Suspended Solids	Dried from 103 to 105 °C ^[3]
44	Trivalent Chromium	1) Digestion, Direct Air-Acetylene Flame Method; Colorimetric Method; Calculation ^[3] 2) Digestion, Inductively Coupled Plasma Method; Colorimetric Method; Calculation ^[3]
45	Zinc	1) Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]

น้ำใต้ดิน...

น้ำใต้ดิน จำนวน 56 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Aldrin	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[3]
2	Antimony	Digestion, Inductively Coupled Plasma Method ^[3]
3	Arsenic	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
4	Barium	1) Digestion, Direct Nitrous Oxide-Acetylene Flame Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
5	Benzene	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[3]
6	Beryllium	1) Digestion, Direct Nitrous Oxide-Acetylene Flame Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
7	Bromodichloromethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[3]
8	Bromoform	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[3]
9	Cadmium	1) Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
10	Carbon tetrachloride	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[3]
11	Chlordane	Liquid-Liquid Extraction, Gas Chromatographic/Mass Spectrometric Method ^[3]
12	Chloroform	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[3]
13	Chlorodibromomethane	Purge and Trap, Gas Chromatographic/Mass Spectrometric Method ^[3]
14	Chromium	1) Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]

15 Chromium (III)...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
15	Chromium (III)	1) Digestion, Direct Air-Acetylene Flame Method; Colorimetric Method; Calculation ^[3] 2) Digestion, Inductively Coupled Plasma Method; Colorimetric Method; Calculation ^[3] .
16	Chromium (VI)	Colorimetric Method ^[3] .
17	Cyanide	Distillation, Colorimetric Method ^[3]
18	1,1-Dichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[3]
19	1,2-Dichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[3]
20	cis-1,2-Dichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[3]
21	trans-1,2-Dichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[3]
22	DDD	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
23	DDE	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
24	DDT	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
25	Dieldrin	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
26	Endosulfan	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
27	Endrin	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
28	Ethylbenzene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[3]
29	Heptachlor	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
30	Heptachlor epoxide	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]

31 Hexachlorobenzene...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
31	Hexachlorobenzene	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
32	α -HCH	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
33	β -HCH	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
34	γ -HCH	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
35	Lead	1) Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
36	Manganese	1) Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
37	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[3]
38	Methoxychlor	Liquid-Liquid Extraction, Gas Chromatographic/ Mass Spectrometric Method ^[3]
39	Methylene chloride	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[3]
40	Nickel	1) Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
41	pH	Electrometric Method ^[3]
42	Phenol	Distillation, Direct Photometric Method ^[3]
43	Selenium	1) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
44	Silver	1) Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
45	Styrene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[3]
46	Tetrachloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[3]

47 Toluene...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
47	Toluene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[3]
48	1,1,1-Trichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[3]
49	1,1,2-Trichloroethane	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[3]
50	Trichloroethylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[3]
51	Vanadium	1) Digestion, Direct Nitrous Oxide-Acetylene Flame Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]
52	Vinyl chloride	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[3]
53	m-Xylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[3]
54	o-Xylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[3]
55	p-Xylene	Purge and Trap, Gas Chromatographic/ Mass Spectrometric Method ^[3]
56	Zinc	1) Digestion, Direct Air-Acetylene Flame Method ^[3] 2) Digestion, Inductively Coupled Plasma Method ^[3]

สิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว จำนวน 20 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Antimony	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[1,4,7] 2) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[1,4,8] 3) Digestion, Inductively Coupled Plasma Method ^[5,7] 4) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]

2 Arsenic...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
2	Arsenic	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[1,4,7] 2) Waste Extraction, Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[1,4,9] 3) Digestion, Inductively Coupled Plasma Method ^[5,7] 4) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[5,9]
3	Barium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[1,4,7] 2) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[1,4,8] 3) Digestion, Inductively Coupled Plasma Method ^[5,7] 4) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
4	Beryllium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[1,4,7] 2) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[1,4,8] 3) Digestion, Inductively Coupled Plasma Method ^[5,7] 4) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
5	Cadmium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[1,4,7] 2) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[1,4,8] 3) Digestion, Inductively Coupled Plasma Method ^[5,7] 4) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
6	Chromium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[1,4,7] 2) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[1,4,8]

3) Digestion...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
7	Chromium (III)	3) Digestion, Inductively Coupled Plasma Method ^[5,7] 4) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8] 1) Waste Extraction, Digestion, Inductively Coupled Plasma Method; Waste Extraction, Colorimetric Method; Calculation Method ^[1,4,7,10] 2) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method; Waste Extraction, Colorimetric Method; Calculation Method ^[1,4,8,10] 3) Digestion, Inductively Coupled Plasma Method; Alkaline Digestion, Colorimetric Method; Calculation Method ^[5,6,7,10] 4) Digestion, Flame Atomic Absorption Spectrometric Method; Alkaline Digestion, Colorimetric Method; Calculation Method ^[5,6,8,10]
8	Chromium (VI)	1) Waste Extraction, Colorimetric Method ^[1,10] 2) Alkaline Digestion, Colorimetric Method ^[6,10]
9	Cobalt	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[1,4,7] 2) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[1,4,8] 3) Digestion, Inductively Coupled Plasma Method ^[5,7] 4) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
10	Copper	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[1,4,7] 2) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[1,4,8] 3) Digestion, Inductively Coupled Plasma Method ^[5,7] 4) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]

11 Lead...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
11	Lead	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[1,4,7] 2) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[1,4,8] 3) Digestion, Inductively Coupled Plasma Method ^[5,7] 4) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
12	Mercury	1) Waste Extraction, Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[1,11] 2) Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[12]
13	Molybdenum	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[1,4,7] 2) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[1,4,8] 3) Digestion, Inductively Coupled Plasma Method ^[5,7] 4) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
14	Nickel	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[1,4,7] 2) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[1,4,8] 3) Digestion, Inductively Coupled Plasma Method ^[5,7] 4) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
15	pH	Electrometric Method ^[17,18]
16	Selenium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[1,4,7] 2) Waste Extraction, Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[1,4,13] 3) Digestion, Inductively Coupled Plasma Method ^[5,7]

4) Digestion ...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
17	Silver	4) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[5,13] 1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[1,4,7] 2) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[1,4,8] 3) Digestion, Inductively Coupled Plasma Method ^[5,7] 4) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
18	Thallium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[1,4,7] 2) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[1,4,8] 3) Digestion, Inductively Coupled Plasma Method ^[5,7] 4) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
19	Vanadium	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[1,4,7] 2) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[1,4,8] 3) Digestion, Inductively Coupled Plasma Method ^[5,7] 4) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
20	Zinc	1) Waste Extraction, Digestion, Inductively Coupled Plasma Method ^[1,4,7] 2) Waste Extraction, Digestion, Flame Atomic Absorption Spectrometric Method ^[1,4,8] 3) Digestion, Inductively Coupled Plasma Method ^[5,7] 4) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]

รพช

ดิน...

ดิน จำนวน 17 รายการ

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
1	Antimony	1) Digestion, Inductively Coupled Plasma Method ^[5,7] 2) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
2	Arsenic	1) Digestion, Inductively Coupled Plasma Method ^[5,7] 2) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[5,9]
3	Barium	1) Digestion, Inductively Coupled Plasma Method ^[5,7] 2) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
4	Beryllium	1) Digestion, Inductively Coupled Plasma Method ^[5,7] 2) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
5	Cadmium	1) Digestion, Inductively Coupled Plasma Method ^[5,7] 2) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
6	Chromium	1) Digestion, Inductively Coupled Plasma Method ^[5,7] 2) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
7	Chromium (III)	1) Digestion, Inductively Coupled Plasma Method; Alkaline Digestion, Colorimetric Method; Calculation Method ^[5,6,7,10] 2) Digestion, Flame Atomic Absorption Spectrometric Method; Alkaline Digestion, Colorimetric Method; Calculation Method ^[5,6,8,10]
8	Chromium (VI)	Alkaline Digestion, Colorimetric Method ^[6,10]
9	Cyanide	Extraction, Distillation, Colorimetric Method ^[14,15,16]
10	Lead	1) Digestion, Inductively Coupled Plasma Method ^[5,7] 2) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
11	Manganese	1) Digestion, Inductively Coupled Plasma Method ^[5,7] 2) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]

รพช

12 Mercury ...

ลำดับที่	สารมลพิษ	วิธีวิเคราะห์
12	Mercury	Digestion, Cold-Vapor Atomic Absorption Spectrometric Method ^[5,7]
13	Nickel	1) Digestion, Inductively Coupled Plasma Method ^[5,7] 2) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
14	Selenium	1) Digestion, Inductively Coupled Plasma Method ^[5,7] 2) Digestion, Hydride Generation/Atomic Absorption Spectrometric Method ^[5,13]
15	Silver	1) Digestion, Inductively Coupled Plasma Method ^[5,7] 2) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
16	Vanadium	1) Digestion, Inductively Coupled Plasma Method ^[5,7] 2) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]
17	Zinc	1) Digestion, Inductively Coupled Plasma Method ^[5,7] 2) Digestion, Flame Atomic Absorption Spectrometric Method ^[5,8]

เอกสารอ้างอิง

- กระทรวงอุตสาหกรรม. ประกาศกระทรวงอุตสาหกรรม, พ.ศ. 2566. เรื่อง การจัดการสิ่งปฏิกูลหรือวัสดุที่ไม่ใช้แล้ว. ราชกิจจานุเบกษา. 31 พฤษภาคม 2566. เล่มที่ 140 ตอนพิเศษ 126 ง.
- สมาคมวิศวกรรมสิ่งแวดล้อมแห่งประเทศไทย. คู่มือวิเคราะห์น้ำเสีย. พิมพ์ครั้งที่ 4. กรุงเทพฯ: เรือนแก้วการพิมพ์, 2547.
- APHA, AWWA, WEF. Standard Methods for the Examination of Water and Wastewater. 24th ed. Washington, DC: APHA, 2023.
- United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. SW-846, 2014.
- United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Acid Digestion of Sediments, Sludges, and Soils. SW-846 Method 3050B, 2007.
- United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Alkaline Digestion for Hexavalent Chromium. SW-846 Method 3060A, 1996.

7. United States...

7. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Inductively Coupled Plasma-Optical Emission Spectrometry. SW-846 Method 6010C, 2000.

8. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Flame Atomic Absorption Spectrophotometry. SW-846 Method 7000B, 2007.

9. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Antimony and Arsenic (Atomic Absorption, Borohydride Reduction). SW-846 Method 7062, 1994.

10. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Chromium, Hexavalent (Colorimetric). SW-846 Method 7196A, 1992.

11. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Mercury in Liquid Waste (Manual Cold-Vapor Technique). SW-846 Method 7470A, 1994.

12. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Mercury in Solid or Semisolid Waste (Manual Cold-Vapor Technique). SW-846 Method 7471B, 2007.

13. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Selenium (Atomic Absorption, Borohydride Reduction). SW-846 Method 7742, 1994.

14. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Total and Amenable Cyanide: Distillation. SW-846 Method 9010C, 2004.

15. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Cyanide Extraction Procedure for Solids and Oil. SW-846 Method 9013A, 1996.

16. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Cyanide in Waters and Extracts Using Titrimetric and Manual Spectrophotometric Procedures. SW-846 Method 9014, 2014.

17. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. pH Electrometric Measurement. SW-846 Method 9040C, 2004.

18. United States Environmental Protection Agency. Test Methods for Evaluation Solid Waste Physical/Chemical Methods. Soil and Waste pH. SW-846 Method 9045D, 2004.

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com

NSC-TISI-TIS 17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 66-420087-1

Page : 1 of 2

Submitted by : M Green Group Co.,Ltd.

188/46 Wisetusakhon 25, Pracha-Uttd Rd., Thungkru Bangkok 10140 Thailand

Equipment : pH Meter with electrode

pH meter

Manufacturer : Eutech

Model : pH 700

Range : N/A pH

Resolution : 0.01 pH

Serial No. : 2884323

ID No. : N/A

Electrode

Model : N/A

Serial No. : 01X099320

Environment : On site calibration was carried out at the Laboratory, M Green Group Co.,Ltd.

Ambient Temperature : (25.0 to 25.5)°C

Relative Humidity : (45 to 50) %

Date of Received : 20 September 2023

Date of Calibration : 20 September 2023

Date of Issue : 25 September 2023

Calibrated by : Permpoon Chanpu

Calibration Method : In-house method CAL-M4201 direct measurement by using standard voltage calibrator and using certified reference material (CRM)

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Multiproduct Calibrator

ID No.	Cert. No.	Due Date	Traceability
400005	SG-E-00307/66	23 Aug 2025	National Institute of Metrology Thailand (NIMT)

2. Standard Buffer Solution

pH	Cert. No.	Lot No.	Exp. Date	Traceability
4.008	61270213	915161	19 Jul 2025	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025
6.985	61275614	898428	28 May 2024	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025
9.997	61281073	915163	19 Jul 2024	CPA Chem Ltd. Accredited to ISO 17034 and ISO/IEC 17025

Approved by :

(Surachai Promthong)

Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



CAL-F0031-01

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com

Certificate of Calibration

Certificate No. : 66-420087-1

Page : 2 of 2

Result of Calibration :

UUC Condition As-Received : Good

Function : Electrical measurement

pH meter

Performing standard curve by Multiproduct Calibrator at pH (4,7,10)

Adjustment Curve at nominal pH	Applied Voltage (mV)	Nominal Value (pH)	UUC Reading		Correction (mV)	Uncertainty (± mV)
			(pH)	(mV)		
4, 7, 10	177.4800	4	4.00	177.5	0.0	0.12
	0.0000	7	7.00	0.1	-0.1	0.086
	-177.4800	10	10.00	-177.4	-0.1	0.12

Function : pH meter with electrode

Performing a three - buffer standard curve using buffer nominal pH (4,7,10)

Adjustment Curve at nominal pH	Standard Buffer (pH)	UUC Reading (pH)	Correction (pH)	Uncertainty (± pH)
4, 7, 10	4.008	4.01	0.00	0.010
	6.985	7.00	-0.01	0.011
	9.997	10.01	-0.01	0.014

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2 , providing a level of confidence of approximately 95%

- oño -



CAL-F0031-03

ภาคผนวก ก/33

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com



NSC-TISI-TIS 17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 66-400519-1

Page : 1 of 2

Submitted by : M Green Group Co.,Ltd.
188/46 Wisatesukhakhon 25, Pracha-Uttd Rd., Thungkru Bangkok 10140 Thailand

Equipment : Digital Thermometer with Thermistor probe
Temperature Indicator

Manufacturer : Eutech **Model :** pH 700
Range : N/A °C **Resolution :** 0.1 °C
Serial No. : 2884323 **ID No. :** N/A

Thermistor probe
Model : N/A **Sheath Material :** Stainless
Diameter : 3.2 mm. **Length :** 100 mm.
Serial No. : PHSTEMB01P **ID No. :** N/A

Environment : On site calibration was carried out at the M Green Gpoup Co.,Ltd.

Ambient Temperature : (25.0 to 26.0) °C
Relative Humidity : (56 to 60) %
Line Voltage : (224.0 to 225.2) VAC

Date of Received : 20 September 2023

Date of Calibration : 20 September 2023

Date of Issue : 25 September 2023

Calibrated by : Permpon Chanpu

Calibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4003
by compared with PRT in the liquid bath at the constant controlled temperature.

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Platinum Resistance Thermometer (PRT)

ID No.	Cert. No.	Due Date	Traceability
400002	TT-0074-22	20 Jun 2024	National Institute of Metrology Thailand (NIMT)

2. Standard Digital Thermometer

ID No.	Cert. No.	Due Date	Traceability
400033	22E569	22 Feb 2024	National Institute of Metrology Thailand (NIMT)

Approved by : 
(Surachai Promthong)
Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel (02) 964-6211 Fax (02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com

Certificate of Calibration

Certificate No. : 66-400519

Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

Immersion Depth (mm.)	Standard Reading (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
100	25.006	24.9	0.1	0.19

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2,
providing a level of confidence of approximately 95%

- ๐0๐ -



CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel (02) 964-6211 Fax (02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com

NSC-TISI-TIS 17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 66-200300-1

Page : 1 of 2

Submitted by :

M Green Group Co., Ltd.

188/46 Wisatesuknakon 25, Pracha-Utid Rd., Thungkru, Bangkok 10140 Thailand

Equipment :

Electronic Balance

Manufacturer : SHIMADZU

Model : AP225WD

Serial No. : D316300690

Capacity : 220 g

Resolution : 0.00001g/102g, 0.0001g/220g

Environment :

On site calibration was carried out at the Laboratory, M Green Group Co., Ltd.

Ambient Temperature : (25.6 to 26.7) °C

Relative Humidity : (54.4 to 56.6) %

Air Pressure : 1010.0 mbar

Date of Received :

20 September 2023

Date of Calibration :

20 September 2023

Date of Issue :

22 September 2023

Calibrated by :

Akaradath Thippichai

Calibration Method :

In-house method CAL-M2001 based on UKAS Publication ref : LAB 14

Edition 7 - November 2022

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Weights

ID No.	Cert. No.	Due Date	Traceability
E261-E2624	C02222345	10 Nov 2023	National Institute of Metrology (Thailand), (NIMT)

Approved by :

(Surachai Promthong)

Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel (02) 964-6211 Fax (02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com

Certificate of Calibration

Certificate No. : 66-200300-1

Page : 2 of 2

Result of Calibration :

Without Adjustment

UUC Condition As-Received : Good

Departure of indication from nominal value


Nominal Value (g)	Correction (g)	Uncertainty ± (g)
0.001	0.00000	0.000012
0.01	0.00000	0.000013
0.1	0.00000	0.000015
1	0.00000	0.000026
10	0.00000	0.000053
20	-0.00003	0.000071
50	0.00004	0.00011
100	-0.00009	0.00020
150	0.0000	0.00038
200	-0.0001	0.00038

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2.15$, providing a level of confidence of approximately 95%

Eccentric error

Load test :	50 g
A	-0.00003
B	0.00000
C	0.00000
D	-0.00005
E	0.00000



Repeatability

Load test :	200 g
Stdev. :	0.000048 g

- o0o -



CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhaphrasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax (02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com



NSC-TISI-TIS17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 66-400531-1

Page : 1 of 2

Submitted by : M Green Group Co., Ltd.

188/46 Wisatesuknakhon 25, Pracha-Uttd Rd., Thungkru, Bangkok 10140 Thailand

Equipment : Liquid in Glass Thermometer

Manufacturer : N/A

Model : N/A

Range : 0 °C to 100 °C

Resolution : 1 °C

Serial No. : N/A

Immersion : Total

ID No. : 94-49747

Environment : Ambient Temperature : (23 ± 2) °C

Relative Humidity : (50 ± 15) %

Line Voltage : (220 ± 22) VAC

Date of Received : 21 September 2023

Date of Calibration : 23 September to 26 September 2023

Date of Issue : 26 September 2022

Calibrated by : Chortip Samchusri

Calibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4001 based on ASTM E77-07 by compared with PRT in the liquid bath at the constant controlled temperature.

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Platinum Resistance Thermometer (PRT)

ID No.	Cert. No.	Due Date	Traceability
400001	TT-0016-22	07 Feb 2024	National Institute of Metrology Thailand (NIMT)

2. Standard Digital Thermometer

ID No.	Cert. No.	Due Date	Traceability
400003	23E1866	01 Jun 2025	National Institute of Metrology Thailand (NIMT)
400004	23E1866	01 Jun 2025	National Institute of Metrology Thailand (NIMT)

Approved by :

(Surachai Promthong)
Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhaphrasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax (02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com

Certificate of Calibration

Certificate No. : 66-400531-1

Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

Ice point check : UUC* reading 0 °C Standard reading 0.0352 °C

Standard Reading (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
39.7228	40	-0.3	0.31

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%

- oOo -

181



CAL-F0031-03

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel (02) 964-6211 Fax (02) 964-5155, e-mail calibratech.cal@yahoo.com, calibratech.cal@hotmail.com



NSC-TISI-TIS17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 66-400520-1

Page : 1 of 2

Submitted by : M Green Group Co., Ltd.
188/46 Wisatesukhakhon 25, Pracha-Uttd Rd., Thungkru Bangkok 10140 Thailand

Equipment : Air Chamber (Refrigerator)
Manufacturer : Biobase Model : BXC-V250M (II)
Range : N/A °C Resolution : 0.1 °C
Serial No. : YC025025190108 ID No. : N/A

Environment : On site calibration was carried out at the Laboratory, M Green Group Co., Ltd.
Ambient Temperature : (25.0 to 26.0) °C
Relative Humidity : (40 to 50) %
Line Voltage : (226.0 to 230.0) V

Date of Received : 20 September 2023

Date of Calibration : 20 September 2023

Date of Issue : 25 September 2023

Calibrated by : Permpon Chanpu

Calibration Method : CAL-M4004, TLAS G-20

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Digital Thermometer with RTD Probe

ID No.	Cert. No.	Due Date	Traceability
400046 & 400042	66-400453-1	31 Jan 2024	National Institute of Metrology Thailand (NIMT)

Approved by : 
(Surachai Promthong)
Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com

Certificate of Calibration

Certificate No. : 66-400520-1


Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber
W = 0.50 m
D = 0.40 m
H = 1.20 m
Capacity = 0.24 m³

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (± °C)
			1	2	3	4	5	6	7	8	9	
4.0	2.0	2.0	4.05	4.04	4.27	4.89	4.10	4.05	4.92	4.37	4.43	0.46

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Uniformity (°C)	Measured Stability (°C)	Overall Variation (°C)
4.0	2.0	2.0	0.60	0.21	1.2

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

- o0o -



CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhaphrasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax (02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com



Certificate of Calibration

Certificate No. : 66-400520-2

Page : 1 of 2

Submitted by : M Green Group Co., Ltd.
188/46 Wisatesukhakhon 25, Pracha-Utid Rd., Thungkru Bangkok 10140 Thailand

Equipment : Air Chamber (Oven)
Manufacturer : Memmert Model : UF110
Range : N/A °C Resolution : 0.1 °C
Serial No. : B419.1092 ID No. : N/A

Environment : On site calibration was carried out at the Laboratory, M Green Group Co., Ltd.
Ambient Temperature : (25.0 to 26.0) °C
Relative Humidity : (40 to 50) %
Line Voltage : (226.0 to 230.0) V

Date of Received : 20 September 2023

Date of Calibration : 20 September 2023

Date of Issue : 25 September 2023

Calibrated by : Permpon Chanpu

Calibration Method : CAL-M4004, TLAS G-20

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units
Standard Digital Thermometer with Thermocouple probe

ID No.	Cert. No.	Due Date	Traceability
400046 & 400028	66-400184-3	04 Oct 2023	National Institute of Metrology Thailand (NIMT)

Approved by :

(Surachai Promthong)

Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhaphrasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com

Certificate of Calibration

Certificate No. : 66-400520-2


Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber
W = 0.56 m
D = 0.40 m
H = 0.48 m
Capacity = 0.11 m³

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (± °C)
			1	2	3	4	5	6	7	8	9	
103.0	103.0	103.0	103.3	103.0	103.7	103.3	103.1	103.0	103.8	102.7	103.3	0.69
105.0	105.0	105.0	105.3	105.0	105.7	105.3	105.2	105.0	105.8	104.6	105.3	0.71
180.0	180.0	180.0	180.4	180.1	181.2	180.4	180.3	180.0	181.4	179.0	180.5	0.95

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Uniformity (°C)	Measured Stability (°C)	Overall Variation (°C)
103.0	103.0	103.0	0.8	0.1	1.3
105.0	105.0	105.0	0.9	0.1	1.4
180.0	180.0	180.0	1.7	0.2	2.7

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

- o0o -

Signature



CAL-P0031-03

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com



NSC-TISI-TIS 17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 66-400520-3

Page : 1 of 2

Submitted by : M Green Group Co., Ltd.

188/46 Wisatesukhakhon 25, Pracha-Uttd Rd., Thungkru Bangkok 10140 Thailand

Equipment : Water Bath

Manufacturer : Memmert

Model : WNB29

Range : N/A °C

Resolution : 0.1 °C

Serial No. : L619.0037

ID No. : N/A

Environment : On site calibration was carried out at the Laboratory, M Green Group Co., Ltd.

Ambient Temperature : (25.0 to 26.0) °C

Relative Humidity : (40 to 50) %

Line Voltage : (226.0 to 230.0) V

Date of Received : 20 September 2023

Date of Calibration : 20 September 2023

Date of Issue : 25 September 2023

Calibrated by : Permpon Chanpu

Calibration Method : This instrument was calibrated by In-house method CAL-M4006 based on ASTM E715-80
The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units
Standard Digital Thermometer with RTD probe

ID No. Cert. No. Due Date Traceability

400046 & 400024 66-400184-2 06 Oct 2023 National Institute of Metrology Thailand (NIMT)

Approved by :

(Surachai Promthong)

Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com

Certificate of Calibration


Certificate No. : 66-400520-3

Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement



Front

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.					Uncertainty (± °C)	Measured Uniformity (°C)	Measured Stability (°C)
			1	2	3	4	5			
85.0	85.0	85.0	85.08	85.04	84.98	85.17	85.02	0.18	0.2	0.05

Remark The uncertainty is not combine uniformity of the water bath

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2 , providing a level of confidence of approximately 95%

- o0o -

APB



CAL-F0031-03

ภาคผนวก ก/39

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhaphrasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com



NSC-TISI-TIS17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 66-400520-4

Page : 1 of 2

Submitted by : M Green Group Co., Ltd.

188/46 Wisatesukhakhon 25, Pracha-Utid Rd., Thungkru Bangkok 10140 Thailand

Equipment : Air Chamber (Incubator)

Manufacturer : Biobase

Model : Biochemistry Incubator

Range : 0 °C to 65 °C

Resolution : 0.1 °C

Serial No. : KYP1502202003

ID No. : N/A

Environment : On site calibration was carried out at the Laboratory, M Green Group Co., Ltd.

Ambient Temperature : (25.0 to 25.5) °C

Relative Humidity : (45 to 50) %

Line Voltage : (226.0 to 230.0) V

Date of Received : 20 September 2023

Date of Calibration : 20 September 2023

Date of Issue : 25 September 2023

Calibrated by : Permpon Chanpu

Calibration Method : CAL-M4004, TLAS G-20

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

Standard Digital Thermometer with RTD Probe

ID No. Cert.No. Due Date Traceability

400029 & 400043 66-400226-1 27 Oct 2023 National Institute of Metrology Thailand (NIMT)

Approved by :

(Surachai Promthong)

Laboratory Manager

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd



CAL-F0031-03

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhaphrasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com

Certificate of Calibration

Certificate No. : 66-400520-4


Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

This instrument was setting air ventilation at position 0 (close)



Inside of Chamber

W = 0.45 m

D = 0.41 m

H = 0.85 m

Capacity = 0.16 m³

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Temperature (°C) @ Sensor No.									Uncertainty (± °C)
			1	2	3	4	5	6	7	8	9	
20.0	20.0	20.0	20.14	20.04	19.91	19.97	20.03	19.96	19.91	19.96	19.92	0.70

Test Point (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured	Measured	Overall
			Uniformity (°C)	Stability (°C)	
20.0	20.0	20.0	0.28	0.37	0.8

Remark The uncertainty is not combine uniformity of the air chamber

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2 , providing a level of confidence of approximately 95%

- 000 -

Signature



CAL-F0031-03

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com



Certificate of Calibration

Certificate No. : 66-300589-7

Page : 1 of 2

Submitted by : M Green Group Co., Ltd.

188/46 Wisatesuknakhon 25, Pracha-Utid Rd., Thungkru, Bangkok 10140 Thailand

Equipment : Burette

Manufacturer : GLASSCO

Class : A

Capacity : 10 ml

Graduation : 0.05 ml

ID No. : 2212-0344-1

Environment : Ambient Temperature : (20 ± 3) °C

Relative Humidity : (50 ± 10) %

Air Pressure : 1006.7 mbar.

Date of Received : 20 September 2023

Date of Calibration : 27 September 2023

Date of Issue : 27 September 2023

Calibrated by : Wipa Tovadee

Calibration Method : In-house method CAL-M3001 based on ASTM E 542-01

Reference Standard Instruments : This certification is traceable to the International System of Units

Electronic Balance

ID No.	Cert. No.	Due Date	Traceability
241003	66-200196-2	02 Dec 2023	National Institute of Metrology (Thailand) (NIMT)

Approved by :

(Wipa Tovadee)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



CAL-F0031-03

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com

Certificate of Calibration

Certificate No. : 66-300589-7

Page : 2 of 2

Result of Calibration : This result of true Volume is referred to standard temperature at 20 °C

UUC Condition As-Received : Good

Delivery Time : 21.33 sec.

Nominal Volume (ml)	Measuring Volume (ml)
10	9.9913

Uncertainty of measurement with in $\pm$ 0.0039 ml

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2.00$,

providing a level of confidence of approximately 95%

- o0o -



CAL-F0031-03

ภาคผนวก ญ/41

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhaphrachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com



NSC-TISI-TIS 17025
CALIBRATION 0030

Certificate of Calibration

Certificate No. : 66-300589-8

Page : 1 of 2

Submitted by : M Green Group Co., Ltd.

188/46 Wisatesuknakhon 25, Pracha-Utid Rd., Thungkru, Bangkok 10140 Thailand

Equipment : Burette

Manufacturer : GLASSCO

Class : A

Capacity : 25 ml

Graduation : 0.1 ml

ID No. : 2212-0344-2

Environment : Ambient Temperature : (20 ± 3) °C

Relative Humidity : (50 ± 10) %

Air Pressure : 1006.7 mbar.

Date of Received : 20 September 2023

Date of Calibration : 27 September 2023

Date of Issue : 27 September 2023

Calibrated by : Wipa Tovadee

Calibration Method : In-house method CAL-M3001 based on ASTM E 542-01

Reference Standard Instruments : This certification is traceable to the International System of Units

Electronic Balance

ID No.	Cert. No.	Due Date	Traceability
241003	66-200196-2	02 Dec 2023	National Institute of Metrology (Thailand) (NIMT)

Approved by :

(Wipa Tovadee)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd



CAL-F0031-03

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhaphrachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel (02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com

Certificate of Calibration

Certificate No. : 66-300589-8

Page : 2 of 2

Result of Calibration : This result of true Volume is referred to standard temperature at 20 °C

UUC Condition As-Received : Good

Delivery Time : 46.01 sec.

Nominal Volume (ml)	Measuring Volume (ml)
25	24.9741

Uncertainty of measurement with in $\pm$ 0.0066 ml

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2.00$,

providing a level of confidence of approximately 95%

- o0o -

(Signature)



CAL-F 0031-01

CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel (02) 964-6211 Fax (02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com

Certificate of Calibration

Certificate No. : 66-300590-1

Page : 1 of 2

Submitted by : M Green Group Co., Ltd.

188/46 Wisatosuknakhon 25, Pracha-Utid Rd., Thungkru, Bangkok 10140 Thailand

Equipment : Imhoff Cone

Manufacturer : VITLAB

Capacity : 1000 ml Graduation : 50 ml

ID No. : CY1000/01/22

Environment : Ambient Temperature : (20 ± 3) °C

Relative Humidity : (50 ± 10) %

Air Pressure : 1005.4 mbar.

Date of Received : 20 September 2023

Date of Calibration : 26 September 2023

Date of Issue : 26 September 2023

Calibrated by : Arcerat Sombun

Calibration Method : In-house method CAL-M3001 based on ASTM E 542-01

Reference Standard Instruments : This certification is traceable to the International System of Units

Electronic Balance

ID No.	Cert. No.	Due Date	Traceability
241002	66-200196-1	02 Dec 2023	National Institute of Metrology (Thailand) (NIMT)

Approved by :

(Wipa Tovadec)

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



CAL

Calibratech Co.,Ltd.

7/106-7 Moo 2, Sukhaprachasan 3 Rd., Bangpood, Pakkred, Nonthaburi 11120

Tel.(02) 964-6211 Fax.(02) 964-5155, e-mail : calibratech.cal@yahoo.com, calibratech.cal@hotmail.com

Certificate of Calibration

Certificate No. : 66-300590-1

Page : 2 of 2

Result of Calibration : This result of true Volume is referred to standard temperature at 20 °C

UUC Condition As-Received : Good

Nominal Volume (ml)	Measuring Volume (ml)
500	501.19
1000	1010.67

Uncertainty of measurement with in $\pm$ 0.17 ml

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor $k = 2.00$, providing a level of confidence of approximately 95%

- o0o -



CAL-F0031-03

การดูแลบำรุงรักษาเชิงป้องกัน

Preventive Maintenance



บริษัท ดีเคเอสเอช เทคโนโลยี จำกัด

ฝ่ายบริการหลังการขาย

โทร 0 2 639 7000 E-mail: service.tec.th@dksh.com

ฝ่ายขายและการตลาด

โทร 0 2 639 7000 E-Mail : marketing.tec.th@dksh.com

Website : www.dksh.co.th/technology/scientific-thailand

เงื่อนไขการให้บริการ Preventive Maintenance

Type text here

บริษัทฯ จะส่งวิศวกรผู้ชำนาญ เพื่อให้บริการตามขอบข่ายของการบริการ เฉพาะ ในวันและเวลา ราชการ หากมีความประสงค์ที่จะรับบริการนอกเหนือจากวัน เวลา ราชการ (วันหยุดเสาร์ – อาทิตย์ หรือวันหยุด นักชดถก) บริษัทฯ จะคิดค่าบริการเพิ่มเติมตามอัตราที่กฎหมายแรงงานกำหนดไว้

ขอบข่ายการบริการ

- ตรวจสอบสภาพการทำงานต่าง ๆ ของเครื่องมือ
- ทดสอบประสิทธิภาพการทำงานของเครื่องมือ
- รายการผลการตรวจสอบเครื่องมือ

หมายเหตุ

- ราคาไม่รวมถึงค่าบริการซ่อม หรือ เปลี่ยนอะไหล่ที่ชำรุดเสียหาย หรือหมดสภาพการใช้งาน
- ในกรณีที่ผู้รับบริการออกเขตพื้นที่ให้บริการ บริษัทฯ จำเป็นต้องคิดค่าใช้จ่ายเพิ่มเติม ได้แก่ ค่าเดินทาง เป็นต้น
- บริษัทฯ ขอสงวนสิทธิ์ในการเปลี่ยนแปลงราคา โดยไม่แจ้งให้ทราบล่วงหน้า

ช่องทางการติดต่อ



DKSH Technology Limited (บริษัท ดีเคเอสเอช เทคโนโลยี จำกัด)
เลขที่ 2533 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง กรุงเทพฯ 10260
เลขประจำตัวผู้เสียภาษี 010-555-001-4547 (สำนักงานใหญ่)



LINE: @dkshscientific



Call center 0 2 639 7000



DKSH Scientific



www.dksh.com/scientific-thailand



marketing_tec.th@dksh.com



@dkshscientific

Preventive Maintenance Contract

จำนวนในการทำสัญญาบริการ ...1...ครั้ง ต่อปี
ครั้งที่ 1 วันที่ 19/04/2024.....

รายละเอียดผู้รับบริการ

หน่วยงาน	บริษัท เทสท์ เทคโนโลยี จำกัด		
ที่อยู่	30,32 ซอยพระรามที่ 2 ซอย 63 ถนนพระรามที่ 2 แขวงสามยุค เขตบางขุนเทียน กรุงเทพมหานคร 10150		
โทรศัพท์	0-2893-4211-7	แฟกซ์	0-2893-4218

ผู้ติดต่อ

ชื่อ - นามสกุล	คุณกรรณก ขุนพิทักษ์		
ตำแหน่ง	หัวหน้าส่วน		
โทรศัพท์	087 398 9274	เบอร์ติดต่อ	-
E-mail	lab_center@testtech.co.th		

รายละเอียดผู้ให้บริการ

บริษัท ดีเคเอสเอช เทคโนโลยี จำกัด (ฝ่ายบริการหลังการขาย) (สำนักงานใหญ่)	
เลขที่ 2533 ถนนสุขุมวิท แขวงบางจาก เขตพระโขนง กรุงเทพฯ 10260	
โทรศัพท์ 0 2 693 7000 Email: sudarat.sk@dksh.com	
เจ้าหน้าที่ประสานงาน : คุณสุดาวิรัตน์ ศิริรัตน์ โทรศัพท์ 090 678 6925	
เจ้าหน้าที่ผู้ให้บริการ	นายจิรายุส สเลอาด
ตำแหน่ง	Specialist, Technical Service.
โทรศัพท์	0938138736
E-mail	Jirayut.js@dksh.com


ลงนามผู้รับบริการ	กนกกร ขุนพิทักษ์	ลงนามผู้ให้บริการ	
ตัวบรรจง	(นาย กนกกร ขุนพิทักษ์)	ตัวบรรจง	(นาย จิรายุส สเลอาด)
ตำแหน่ง	หัวหน้าส่วน	ตำแหน่ง	Specialist, Technical Service.
วันที่ / ประทับตราบริษัท	19/04/2024	วันที่ / ประทับตราบริษัท	19/04/2024

JOB No: LSPR2402440 MODEL: VAP300 S/N: GER5300210025

Operational Qualification (OQ)

ตรวจสอบสภาพเครื่อง


FRONT



No		PASS	FAIL	N/A
1	Quick clamping device with clamping block	☒	☐	☐
2	Digestion tube 250/300 ml	☒	☐	☐
3	PTFE steam inlet tubing	☒	☐	☐
4	Connection stopper, Viton	☒	☐	☐
5	Screw cap GL18	☒	☐	☐
6	PTFE-inlet tubing NaOH	☒	☐	☐
7	Distribution head made of glass	☒	☐	☐
8	Screw cap GL32	☒	☐	☐
9	Distillation condenser made of glass	☒	☐	☐
10	Screw cap GL14	☒	☐	☐
11	Ventilation valve	☒	☐	☐
12	Control panel	☒	☐	☐
13	Operating Button	☒	☐	☐
14	USB interface (with protective cap)	☒	☐	☐
15	Silicone tubing 8/10 for distillate discharge **	☐	☐	☒
16	Verprene tubing 4/8, receiver suction **	☐	☐	☒
17	Cable duct for electrode cable + titration tube**	☐	☐	☒
18	Silicone tubing 4/7, boric acid inlet**	☐	☐	☒
19	Sensor for level monitoring including connector**	☐	☐	☒
20	Agitator motor with propeller**	☐	☐	☒
21	Titration acid inlet tube **	☐	☐	☒
22	Receiver glass**	☐	☐	☒
23	Holder for pH electrode, removable**	☐	☐	☒
24	pH electrode (combined electrode)**	☐	☐	☒
25	Drip tray PP	☒	☐	☐


** only VAP 450

REAR




No		PASS	FAIL	N/A
1	Tube connection for sample H3BO3 supply	☐	☐	☒
2	Tube connection for sample H2O supply	☒	☐	☐
3	Tube connection for steam generator H2O supply	☒	☐	☐
4	Tube connection for NaOH supply	☒	☐	☐
5	Tube connection for receiver glass extraction	☒	☐	☐
6	Tube connection for sample waste extraction	☒	☐	☐
7	Tube connection, overpressure steam outlet	☒	☐	☐
8	Connection for cooling water supply (with cleaning sieve)	☒	☐	☐
9	Tube connection for cooling water outlet	☒	☐	☐
10	4 X USB interface	☒	☐	☐
11	1 X RS-232 Interface	☒	☐	☐
12	LAN Interface	☒	☐	☐
13	Screw cap for Perspex cover	☒	☐	☐
14	Connection socket for sample waste tank level monitoring	☒	☐	☐
15	Connection (not used)	☐	☐	☒
16	Connection socket for H2O tank level monitoring	☒	☐	☐
17	Connection socket for H3BO3 tank level monitoring	☒	☐	☐
18	Connection socket for NaOH tank level monitoring	☒	☐	☐
19	Overcurrent circuit breaker	☒	☐	☐
20	Apparatus socket (mains cable connection)	☒	☐	☐
21	Rating plate with serial number	☒	☐	☐
22	Exhaust air fan	☒	☐	☐
23	Excess temperature switch	☒	☐	☐

Inside Steam generator




No		PASS	FAIL	N/A
1	Steam generator	☒	☐	☐
2	Steam generator traverse	☒	☐	☐
3	Pinch valve	☒	☐	☐
4	Circuit board distributor	☒	☐	☐
5	Valve tubing connection	☒	☐	☐
6	Housing safety valve	☒	☐	☐
7	Safety valve SKT	☒	☐	☐
8	Excess temperature protection , steam generator	☒	☐	☐
9	Safety valve G 1/8 0,5 bar	☒	☐	☐
10	Ventilation glass pinch valve VAPODEST	☒	☐	☐
11	Hose clamp for ventilation clamp	☒	☐	☐
12	Distributor PP	☒	☐	☐
13	Angle connection PP	☒	☐	☐
14	Pressure transmitter	☒	☐	☐
15	Level switch	☒	☐	☐
16	Fixing bracket steam generator	☒	☐	☐
17	Relay HT+	☒	☐	☐
18	VA Hexagon nut 1/2"	☒	☐	☐
19	Angle connection 1/8"	☒	☐	☐
20	Bushing nipple 6-10-14	☒	☐	☐
21	VA Lens head screw M5 X 10	☒	☐	☐
22	Grounding connection , 2-pole	☒	☐	☐
23	VA Lens head screw M4 X 6	☒	☐	☐
24	Spacer bolt 5 mm	☒	☐	☐
25	VA Lens head screw M4 X 10	☒	☐	☐
26	Tubing connection	☒	☐	☐
27	Hose clamp 14,5 mm	☒	☐	☐
28	Module ball valve with nozzles	☒	☐	☐
29	Cross manifold with spout	☒	☐	☐
30	Seal copper G 1/8	☒	☐	☐
31	Locking screw 1/8"	☒	☐	☐
32	Pin strip	☒	☐	☐
33	Bundle clamp 12 H 4500	☒	☐	☐
34	Bundle clamp 12 H 4502	☒	☐	☐
35	Temperature switch 80°C	☒	☐	☐
36	VA Lens head screw M3 X 6	☒	☐	☐
37	VA Hexagon nut M4	☒	☐	☐
38	Lins head screw M4 X 8	☒	☐	☐
39	VA Spring washer	☒	☐	☐
40	Angle connection , reduced , 1/8" PP	☒	☐	☐

Module Pump holder VAP200 - 450 V3



No		PASS	FAIL	N/A
1	Peristaltic pump	☒	☐	☐
2	Diaphragm pump NaOH, with non return valve	☒	☐	☐
3	Circuit board	☒	☐	☐
4	Tubing connection module	☒	☐	☐
5	Flow controller	☒	☐	☐
6	Lens head screw M5 x 10	☒	☐	☐
7	Bushing nozzle	☒	☐	☐
8	Screw in socket	☒	☐	☐
9	Magnetic valve 2/2 way	☒	☐	☐
10	Circuit board distributor	☒	☐	☐
11	Bushing nozzle	☒	☐	☐
12	Screw 5 x 25	☒	☐	☐
13	Cylinder screw	☒	☐	☐
14	Screw 5 x 20	☒	☐	☐
15	Seal EPDM 15 x 4	☒	☐	☐
16	Tubing connection piece 51x10x6,5	☒	☐	☐
17	Tubing connection piece 51x10x10	☒	☐	☐
18	Screw M4x10	☒	☐	☐
19	Clamp	☒	☐	☐
20	Clamp	☒	☐	☐
21	Y-tube connector	☒	☐	☐
22	Spacer bolt 5 mm	☒	☐	☐
23	Bundle clamp	☒	☐	☐
24	Bundle clamp	☒	☐	☐
25	Retrofit earthing pump/v	☒	☐	☐
26	Snap ferrite	☒	☐	☐
27	Nut G 3/8"	☒	☐	☐
28	Pump holder plate	☒	☐	☐

Control panel



No		PASS	FAIL
1	Title bar	☒	☐
2	Status bar	☒	☐
3	Navigation button	☒	☐
4	Smart switch with multiple functions	☒	☐
5	USB interface	☒	☐

รายละเอียดการตรวจสอบ

ขั้นตอนการบริการ

ตรวจสอบระบบไฟฟ้า (Electrical Test)

- ความต้านทานทางไฟฟ้าของเครื่องกับกราวด์
- กระแสไฟฟ้าที่ใช้งาน

ตรวจสอบสภาพเครื่อง (Optical Test)

- Main cable
- Electric wiring
- Pumps
- Distribution Head
- Condensor
- Steam generator
- Tubing
- Viton cone

ตรวจสอบ Function การทำงาน (The Function Test)

- ระบบสร้างและควบคุมความดันของ Steam
- ระบบการเติมน้ำเข้า Sample Tube
- ระบบการเติม Na OH
- ระบบการเติม H3BO3

รายงานผลการให้บริวาร

1. TECHNICAL DATA

Main Supply 220 volt + 10% 50 Hz with ground
Nominal current

Pass	Fail	N/A	Remark
☒	☐	☐	
☒	☐	☐	

1.1 COOLING WATER BATH

Temperature 15-20 °C

Cooling Water Outlet

Control Temperature

Pass	Fail	N/A	Remark
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	

1.2 OPTICAL TESTYAP300

Screw cap GL14

Screw cap GL18

Screw cap GL32

Distillation Head

Condensor

Viton Cone

Ventilation Valve BV

Micro Switch Sample

Agitator motor for propeller

Pass	Fail	N/A	Remark
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☐	☐	☒	

2. SYSTEM COOLING WATER INLET

Cooling Water Inlet

Cooling Water Outlet

Flow control valve

Pass	Fail	N/A	Remark
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	

3. SYSTEM CONTROL

Display

Program

Adding NaOH

Adding H2O

Adding H3BO3

Suction Sample

Suction Receiver

Pass	Fail	N/A	Remark
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☐	☐	☒	
☒	☐	☐	
☐	☐	☒	

4. SYSTEM DISTILLATION

Boiler

Level Sensor

Novopren

Solenoid Valve Shut-Off

Solenoid Valve Steam

Solenoid Valve soft steam

Ventilation Valve Premount

Excess Pressure Detector

Heating Element

Pass	Fail	N/A	Remark
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	

5. PUMP

Pump H₂O Steam

- Non-Return Valve

Pump H₂O Sample

- Non-Return Valve

Pump NaOH

- Non-Return Valve

Pump H3BO3

- Non-Return Valve

Pump suction

Pump suction receiver

Pass	Fail	N/A	Remark
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☐	☐	☒	
☐	☐	☒	
☒	☐	☐	
☐	☐	☒	

6. The Following Program Run :

Addition H2O 0-999 mL

Addition NaOH 0-999 mL

Addition H3BO3 0-999 mL

Reaction Time 0-108 min

Distillation Time 0-108 min

Steam Capacity 10%-100%

Suction Sample

Suction Receiver

Pass	Fail	N/A	Remark
☒	☐	☐	
☒	☐	☐	
☐	☐	☒	
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☒	☐	☐	
☐	☐	☒	

7. Measured pumps

Pump NaOH

Volume :20.40.....ml

Pump H2O

Volume :10.00.....ml

Pump H3BO3

Volume :ml

Remark

Remark :

ข้อมูลสนับสนุนด้านเทคนิค (General Technical Support)

การบำรุงรักษาทั่วไป (Basic maintenance)

Cleaning program

Glass parts and tubes must be rinsed daily before starting analysis in order to prevent clogging by crystallizing chemicals.

The following settings are recommended for this:

parameter	Value
H ₂ O addition	120 ml
NaOH addition	0 ml
Dilution time	7 min
Steam power	100 %
Reaction time	0 s
Soak sample	30 s

Insert a digestion tube (without sample) and start the program.

All liquid carrying parts are cleaned. In the case of strong scaling, repeat. 10 ml of sulphuric acid can also be added to the digestion tube.

General error message

Error Message	Cause	Remedy
'Cooling water flow volume too low'	Cooling water pressure under 1 bar	- Open under tap. - Check cooled pressure. - Check coolant tube. Program continues automatically once error has been fixed.
'Sample tube missing'	Sample tube missing.	- Insert sample tube. - Continue program or restart.
'Distillation room protective door open'	Protection door not closed	Close protection door. Program continues automatically once error has been fixed.
'Reagent storage waste'	One or more reagent tanks are empty	- Fill storage tank. - Check correct seating of the universal sensors. The rearing program can be continued after rectification of the error.
	The sample waste tank is full.	- Empty sample waste tank. - Check correct seating of the universal sensors. The rearing program can be continued after rectification of the error.

Analytical errors

Error Message	Cause	Remedy
Analyte results too high	The chemicals used are contaminated with nitrogen compounds. Unrinsed residues in the digestion tube, sodium hydroxide drops get into the receiver. Glass bridge of the evaporator is broken or bent out, sodium hydroxide drops get into the receiver. Glass cleaning agents in the digestion tube. Enrichment of ammonia from the previous sample.	- Deflated checking of the chemicals. - Deterioration of a blank value. - Replace the chemicals if necessary. - Increase of the water addition amount. - Replacement of the glass container. - Clean digestion tube in advance with distilled water. - Increase dilution time. - Check whether the sample was previously sufficiently diluted.
Analyte results too low or no result	Insufficient distillation; distillation time too short. Ammonia escapes at testing phase.	- No quantitative oxidation of the ammonia content. - The distillation amount should be 100 ml. - Seal or defective Viton plug; clean or replace. - Check valve (if, screw connections) on the distribution funnel; replace if necessary. - Check valve of the condenser is gummed up; clean or replace. - Digestion tube is damaged at the neck; replace. - Distillation head glass lens; replace. - Check the standard flow rate of the NaOH pump (see Technical Data).
	Addition amount of the sodium hydroxide too little; no ammonia development. The low basic acid amount in the receiver; escaping ammonia is not completely reacted. Tubes not completely immersed in the acid receiver.	- Increase of the basic acid amount. - Increase of the acid amount.
	Formation of stable ammonia compounds which are not destroyed with sodium hydroxide.	This problem only occurs with catalysts containing mercury. Sodium sulphate solution destroys these components.



QUALITY CALIBRATION CO.,LTD.
235 Petchkasem 63/2 Road, Laksong, Bangkoe, Bangkok 10160
Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4584
www.qcalibration.com



CERTIFICATE No : 23T8796
REFERENCE No : 70515-4

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : WATER BATH
MANUFACTURER : MEMMERT
MODEL : WNE 45
SERIAL No : L720.0266
ID No : EQL-241
CONDITION AS RECEIVED : USED ITEM
SUBMITTED BY : TEST TECH CO., LTD.
30,32 RAMA II SOI 63, RAMA II RD., SAMAEDAM,
BANGKHUNTHIAN, BANGKOK 10150

CALIBRATED BY : CHAICHARN CH.

CALIBRATION DATE : 16-Aug-23

APPROVED BY : PONGSAK J.

ISSUED DATE : 16-Aug-23

RECEIVED DATE : 16-Aug-23

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.

F-G010 REV : 03



QUALITY CALIBRATION CO.,LTD.
235 Petchkasem 63/2 Road, Laksong, Bangkoe, Bangkok 10160
Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4584
www.qcalibration.com

CERTIFICATE No : 23T8796

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : WATER BATH
MANUFACTURER : MEMMERT
ID NUMBER : EQL-241
RECEIVED DATE : 16-Aug-23
AMBIENT TEMPERATURE : 25 °C ± 1 °C
MODEL : WNE 45
SERIAL NUMBER : L720.0266
CALIBRATION DATE : 16-Aug-23
RELATIVE HUMIDITY : 50 %RH ± 10 % RH

CONDITION OF THIS RESULTS OF CALIBRATION


1. THIS INSTRUMENT WAS CALIBRATED ACCORDING TO ASTM E715-80 (REAPPROVED 2001) BY COMPARISON WITH CALIBRATED RTD. THE PROBES WERE PLACED ON FIVE POINTS AND LOCATED ONE PROBE IN EACH OF THE FOUR CORNERS OF THE BATH AND PLACED THE FIFTH RTD WITHIN 2.5 cm. OF THE GEOMETRIC CENTER OF THE WATER VOLUME (REFERENCE LOCATION) UNDER NO LOAD CONDITION.

2. REFERENCE STANDARD INSTRUMENTS :

INSTRUMENT	MODEL	SERIAL No	CERTIFICATE No	DUE DATE
1) DATA LOGGER WITH RTD	2625A	6603614	23T6642	19-Jul-24

3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.
4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.
5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-
- NATIONAL INSTITUTE OF METROLOGY (THAILAND) THROUGH QUALITY CALIBRATION CO.,LTD.

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT



PROBE INSTALLATION
POSITION IN THE BATH

GENERAL INFORMATION

Overall Variation of Ambient Temperature around the Bath (°C) : 0.6
Overall Variation of Line Voltage (V) : 3
Instrument Condition : Normal
Bath Inner Size (W*L*H) : 59*35*20 cm

BATH PERFORMANCE

Calibrate Point (°C)	Average All Position Temp. (±°C)	Temperature Stability (±°C)	Radius Uniformity (°C)	Axial Uniformity (°C)	Overall Variation (°C)
83.0	83.09	0.05	0.07	0.05	0.16
92.0	92.13	0.11	0.06	0.06	0.28

TEMPERATURE MEASUREMENT ACCURACY TEST

Controller Temp (°C)	Indicating Temp (°C)	Measured Temperature (°C) at Spread Locations					Uncertainty (± °C)
		#1	#2	#3	#4	Ref. 5	
83.0	83.0	83.08	83.09	83.06	83.11	83.12	0.15
92.0	92.0	92.11	92.13	92.10	92.16	92.16	0.19

NOTE 1 : THE UNCERTAINTY OF MEASUREMENT EXCLUDED TEMPERATURE UNIFORMITY OF THE BATH.

NOTE 2 : THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA.

THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR k=2, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.

END OF CALIBRATION REPORT

F-G010 REV : 03



CERTIFICATE No : 24T1185
REFERENCE No : 72116-3

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : INCUBATOR
MANUFACTURER : MEMMERT
MODEL : IF 160
SERIAL No : D518.0082
ID No : EQL-205
CONDITION AS RECEIVED : USED ITEM
SUBMITTED BY : TEST TECH CO., LTD.
30,32 RAMA II SOI 63, RAMA II RD., SAMAEDAM,
BANGKHUNTHIAN, BANGKOK 10150

CALIBRATED BY : CHAICHARN CH.
CALIBRATION DATE : 09-Feb-24

APPROVED BY : PONGSAK J.

ISSUED DATE : 12-Feb-24

RECEIVED DATE : 09-Feb-24

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.

F-G010 REV : 03



CERTIFICATE No : 24T1185

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : INCUBATOR
MANUFACTURER : MEMMERT
MODEL : IF 160
ID No : EQL-205
RECEIVED DATE : 09-Feb-24
AMBIENT TEMPERATURE : 25 °C ± 1 °C
S/N : D518.0082
CALIBRATION DATE : 09-Feb-24
RELATIVE HUMIDITY : 53 %RH ± 10 %RH

CONDITION OF THIS RESULTS OF CALIBRATION


1. THIS INSTRUMENT WAS CALIBRATED ACCORDING TO TLAS G-20 BY COMPARISON WITH CALIBRATED RTD Pt100 UNDER NO LOAD CONDITION. THE TEMPERATURE PROBES WERE PLACED ON NINE POINTS AND LOCATED ONE THERMOMETER PROBE IN EACH OF THE EIGHT CORNERS OF THE CHAMBER AND WAS AWAY FROM THE EACH WALL OF 5 cm TO 10 cm. AND PLACED THE NINTH THERMOMETER PROBE WITHIN 2.5 cm. OF THE GEOMETRIC CENTER OF THE CHAMBER. THE UNIFORMITY WAS MEASURED BETWEEN REFERENCE PROBE AND OTHER PROBES AT THE SAME TIME.

2. REFERENCE STANDARD INSTRUMENTS :-

INSTRUMENT	MODEL	SERIAL No	CERTIFICATE No	DUE DATE
1) DATA LOGGER WITH RTD	HYDRA 2635A	7301307	23T6636	10-Jul-24

3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.
4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.
5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-
- NATIONAL INSTITUTE OF METROLOGY (THAILAND) THROUGH QUALITY CALIBRATION CO.,LTD.

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT



FRONT

GENERAL INFORMATION

Overall Ambient Temperature around the Chamber (°C) variation : 1
Overall Line Voltage (V) variation : 3
Instrument Condition : Normal
Chamber Size (W*L*H): 56*40*72 cm

CHAMBER PERFORMANCE

Calibrate Point (°C)	Average All Position Temp. (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
35.0	35.03	0.05	0.09	0.16
36.0	36.05	0.07	0.08	0.19
41.5	41.45	0.08	0.13	0.20

TEMPERATURE MEASUREMENT ACCURACY TEST

Controller Temp (°C)	Indicating Temp (°C)	Measured Temperature (°C) at Spread Locations									Uncertainty (± °C)
		#1	#2	#3	#4	Ref. 5	#6	#7	#8	#9	
35.0	35.0	34.98	35.01	35.00	35.00	35.02	35.08	35.07	35.04	35.10	0.25
36.0	36.0	36.00	36.03	36.03	36.02	36.04	36.09	36.10	36.04	36.12	0.25
41.5	41.5	41.45	41.45	41.39	41.46	41.46	41.47	41.43	41.44	41.49	0.36

NOTE 1 : THE UNCERTAINTY OF MEASUREMENT EXCLUDED TEMPERATURE UNIFORMITY OF THE CHAMBER.
NOTE 2 : LOCATION 5 WAS REFERENCE LOCATION.

NOTE 3 : THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA.
THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR k=2, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.
END OF CALIBRATION REPORT

F-G010 REV : 03



CERTIFICATE No : 24T1189
REFERENCE No : 72116-7

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : AUTOCLAVE
MANUFACTURER : HIRAYAMA
MODEL : HVE-50
SERIAL No : 30612085166
ID No : EQL-155
CONDITION AS RECEIVED : USED ITEM
SUBMITTED BY : TEST TECH CO., LTD.
30,32 RAMA II SOI 63, RAMA II RD.,
SAMAEDAM, BANGKHUNTHIAN, BANGKOK
10150

CALIBRATED BY : CHAICHARN CH.
CALIBRATION DATE : 09-Feb-24

APPROVED BY :
PONGSAK J.

ISSUED DATE : 12-Feb-24

RECEIVED DATE : 09-Feb-24

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.



CERTIFICATE No : 24T1189

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : AUTOCLAVE
MANUFACTURER : HIRAYAMA
ID NUMBER : EQL-155
RECEIVED DATE : 09-Feb-24
AMBIENT TEMPERATURE : 30° C ± 1° C
MODEL : HVE-50
SERIAL NUMBER : 30612085166
CALIBRATION DATE : 09-Feb-24
RELATIVE HUMIDITY : 53 %RH ± 10 % RH

CONDITION OF THIS RESULTS OF CALIBRATION


1. THIS INSTRUMENT WAS CALIBRATED BASED ON BS 2646-1:2021 BY COMPARISON WITH CALIBRATED RTD DATA LOGGERS UNDER NO LOAD CONDITION. THE SENSORS WERE PLACED ON FIVE LOCATIONS AS SHOWN IN THE PICTURE. THE SENSOR ON LOCATION 1 AND 2 WERE PLACED IN THE UPPER HALF AND LOWER HALF OF CHAMBER FREE SPACE RESPECTIVELY. THE THIRD SENSOR WAS PLACED WITHIN 2.5 cm. OF THE GEOMETRIC CENTER OF THE INSTRUMENT CHAMBER. SENSOR NUMBER 4 WAS ATTACHED TO THE LOAD TEMPERATURE PROBE, IF FITTED, WITHIN 15 mm OF ITS TIP. SENSOR NUMBER 5 WAS PLACED IN THE CHAMBER DRAIN OR VENT WITHIN 100 mm OF ITS CONNECTION TO THE CHAMBER.

2. REFERENCE STANDARD INSTRUMENTS :-

INSTRUMENT	MODEL	SERIAL No	CERTIFICATE No	DUE DATE
1) DATA LOGGER	VALPROBE	S350,S367,DV35,DN94	24T0890	26-Jan-25

3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.
4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.
5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-
- NATIONAL INSTITUTE OF METROLOGY (THAILAND) THROUGH QUALITY CALIBRATION CO.,LTD.

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT



FRONT

GENERAL INFORMATION

Overall Ambient Temperature around the Chamber variation : 0.5 °C
Autoclave Condition : Normal
Chamber Size (Diameter*H): 30 * 71 cm

CHAMBER PERFORMANCE

Calibrate Point (°C)	Average all Position (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)	Pressure (MPa)	Holding time (min)	Operating Cycle time (min)
115	115.74	0.09	0.11	0.25	0.090	20	60
121	121.59	0.06	0.21	0.28	0.125	20	60

TEMPERATURE MEASUREMENT ACCURACY TEST(° C)

Cont Temp	Ind Temp	Measured Temperature (°C) at Spread Locations					Uncertainty (± °C)
		#1	#2	#3	#4	#5	
115	115.	115.72	115.74	115.79	115.71	115.71	0.59
121	121	121.59	121.62	121.56	121.58	121.59	0.59

NOTE 1 : THE UNCERTAINTY OF MEASUREMENT OF TEMPERATURE MEASUREMENT ACCURACY TEST EXCLUDED TEMPERATURE UNIFORMITY OF THE CHAMBER.
NOTE 2 : THE STABILITY TERM IN THE UNCERTAINTY BUDGET WAS REPLACED BY THE STANDARD REPEATABILITY.
NOTE 3 : LOCATION 3 WAS REFERENCE LOCATION.
NOTE 4 : THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA.
THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR k=2, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.

END OF CALIBRATION REPORT



Bara Scientific Co., Ltd.
968 U Chu Liang Building Floor7 Rama4 Road
Silom Bangkok Bangkok Thailand 10500
Tel : 02-6324300 Fax : 02-6375496-7
www.barascientific.com



Certificate of Calibration

Number of Page(s) 1 of 3

Certificate No. BSCC-UV-166/24
Equipment UV/Vis Spectrophotometer
Model UV-1900i
Manufacturer Shimadzu
Serial No. A12535780311 ML
ID No. EQL-233
Date of receipt 26 April 2024
Date of calibration 26 April 2024
Date of issue 30 April 2024

Customer name Test Tech Co., Ltd.

Address 30,32 Rama II Soi 63, Rama II Road, Samae Dam, Bang Khun Thian, Bangkok 10150

Temperature (24.9 - 25.4) °C (On site)
Humidity (49.4 - 51.1) %RH (On site)

Equipment condition Good Operation

Calibration Location Water Room

Calibration Procedure In-house method WI-UV-702-01 based on ASTM E275-01

Traceability Wavelength Accuracy is traceable to certificate No. 106372 and 106371
Photometric Accuracy is traceable to certificate No. 106364 and 111398
Stray Light is traceable to certificate No. 106377
The above certificate are traceable to SI unit through Starna Scientific Ltd.
(UKAS accredited calibration laboratory NO. 0659)

Calibrated by Mr.Wanchana Janloey

Approved by

Mr.Sonthi Temboonsakdi
Service Manager

The above results are valid exclusively for the calibrated item(s) as mention in this report / certificate.
Advertising the report / Certificate and publicity of the results are prohibited and also shall not be reproduced
except in full, without written approval of the Bara Scientific Co., Ltd.



Bara Scientific Co., Ltd.
968 U Chu Liang Building Floor7 Rama4 Road
Silom Bangkok Bangkok Thailand 10500
Tel : 02-6324300 Fax : 02-6375496-7
www.barascientific.com



Certificate of Calibration

Certificate No. BSCC-UV-166/24

Number of Page(s) 2 of 3

Calibration Results:

1.Wavelength Accuracy

Certified Wavelength (nm)	UUC (nm)	Error (nm)	Uncertainty (±nm)
279.44	279.18	-0.26	0.18
418.53	418.46	-0.07	0.18
536.52	536.54	0.02	0.18
684.50	684.63	0.13	0.18
879.41	879.43	0.02	0.18

2.Photometric Accuracy (UV)

Wavelength (nm)	Certified Absorbance (A)	UUC (A)	Error (A)	Uncertainty (±A)
235	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
257	0.0000	0.0000	0.0000	0.0075
	0.8354	0.8333	-0.0021	0.0075
313	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
350	0.0000	-0.0001	-0.0001	0.0075
	0.6199	0.6190	-0.0009	0.0075

*CNR = Customer not request

The above results are valid exclusively for the calibrated item(s) as mention in this report / certificate.
Advertising the report / Certificate and publicity of the results are prohibited and also shall not be reproduced
except in full, without written approval of the Bara Scientific Co., Ltd.



Bara Scientific Co., Ltd.
968 U Chu Liang Building Floor7 Rama4 Road
Silom Bangrak Bangkok Thailand 10500
Tel : 02-6324300 Fax : 02-6375496-7
www.barascientific.com



Certificate of Calibration

Certificate No. BSCC-UV-166/24

Number of Page(s) 3 of 3

Calibration Results:

3. Photometric Accuracy (Visible)

Wavelength (nm)	Certified Absorbance (A)	UUC (A)	Error (A)	Uncertainty ($\pm$ A)
420.0	0.0000	0.0000	0.0000	0.0042
	0.5761	0.5791	0.0030	0.0042
	0.7119	0.7132	0.0013	0.0042
	1.0189	1.0221	0.0032	0.0042
440.0	0.0000	0.0000	0.0000	0.0042
	0.5610	0.5636	0.0026	0.0042
	0.7001	0.7012	0.0011	0.0042
	1.0026	1.0052	0.0026	0.0042
465.0	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
546.1	0.0000	0.0000	0.0000	0.0042
	0.5249	0.5260	0.0011	0.0042
	0.6975	0.6971	-0.0004	0.0042
	1.0009	1.0012	0.0003	0.0042
590.0	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
	CNR	CNR	CNR	CNR
635.0	0.0000	0.0000	0.0000	0.0042
	0.5666	0.5673	0.0007	0.0042
	0.7620	0.7611	-0.0009	0.0042
	1.0982	1.0976	-0.0006	0.0042

*CNR = Customer not request

4. Stray Light*

Standard cut-off wavelength (nm)	Unit Under Calibration(UUC)		
	Wavelength (nm)	Transmission (%T)	Absorbance (A)
200.85 $\pm$ 0.11nm	200.76	0.9795	2.0091

The Stray light transmission reference is less than 1.0%T and Stray light absorbance reference is greater than 2.00A

*Stray Light not NSC-ONSC Accredited.

The measurement uncertainty is base on a standard uncertainty multiplied by a coverage factor k=2 providing a level of confidence of approximately 95%.

End of Certificate

The above results are valid exclusively for the calibrated item(s) as mention in this report / certificate.
Advertising the report / Certificate and publicity of the results are prohibited and also shall not be reproduced except in full, without written approval of the Bara Scientific Co., Ltd.



Certificate of Calibration

Equipment: SPECTROPHOTOMETER

Model: DR6000

Serial No. (or ID.): 1693421 (EQL-197)

Manufacturer: HACH

Condition: In Condition

Certificate No.: C06240153

Issued Date: 18 April 2024

Job No.: WO-00024683

Page: 1 of 3

Customer: TEST TECH CO., LTD.

30,32 Rama II Soi 63, Rama II Rd.,
Samaedam, Bangkhuntien Bangkok 10150 Thailand

Environment Condition: Temperature 29.8 °C $\pm$ 0.1 °C
Humidity 45.7 %RH $\pm$ 6.9 %RH

Calibration Place: TEST TECH CO., LTD. (แผนกน้ำดี)

30,32 Rama II Soi 63, Rama II Rd.,
Samaedam, Bangkhuntien Bangkok 10150 Thailand

Calibration By: Miss.Kaewkan Suradech

Calibration Date: 18 April 2024

The Method used: In house method, CAL-WI-24, base on ASTM E 275-08 and ASTM E 387-04

Traceability: This certificate is traceable to the CRM maintained by National Institute of Standards and Technology (NIST) through Starna Scientific Limited.

The standard for Wavelength Certificate No. 118106 and 118118

The standard for Photometric Certificate No. 118123 and 118113

The standard for Stray light Certificate No. 118110 and 118112

The standard for Spectral resolution Certificate No. 118104

(Miss Kaewkan Suradech)

Person in charge

(Mr. Nitinun Srihawan)

Authorized signatory

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to international or national standard or other recognized national standard laboratories.

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor (k=2) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).

These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of DKSH Technology Limited.

บริษัท ดีเคเอสเอช เทคโนโลยี จำกัด
DKSH Technology Limited
2533 สุขุมวิท 70 ถนนสุขุมวิท แขวงคลองเตย เขตคลองเตย กรุงเทพมหานคร 10260
2533 Sukhumvit Road, Bangkok, Phra Khanong, Bangkok 10260
Phone: +66 2639 7000 Email: info.calibration@dksh.com Website: www.dksh.com/scientific-thailand

Delivering Growth - in Asia and Beyond.

Calibration Results:

Without Adjustment

Wavelength Accuracy (nm), The spectral bandwidth of Std at 2 nm and UUC at 2 nm

Standard Wavelength	Unit Under Calibration	Correction	Uncertainty
418.61	418.5	0.11	0.13
536.66	536.7	-0.04	0.13
637.98	637.9	0.08	0.13
748.48	748.6	-0.12	0.13
807.03	807.4	-0.37	0.13

Photometric Accuracy (Absorbance)

Wavelength	Standard absorbance	Unit Under Calibration	Correction	Uncertainty
420 nm	0.0000	0.000	0.0000	0.0045
	0.5772	0.576	0.0012	0.0045
	0.7198	0.719	0.0008	0.0045
	1.0394	1.039	0.0004	0.0045
440 nm	0.0000	0.000	0.0000	0.0045
	0.5608	0.560	0.0008	0.0045
	0.7062	0.705	0.0012	0.0045
	1.0189	1.018	0.0009	0.0045
465 nm	0.0000	0.000	0.0000	0.0045
	0.5214	0.521	0.0004	0.0045
	0.6652	0.664	0.0012	0.0045
	0.9577	0.957	0.0007	0.0045
546.1 nm	0.0000	0.000	0.0000	0.0045
	0.5192	0.518	0.0012	0.0045
	0.6907	0.689	0.0017	0.0045
	0.9949	0.993	0.0019	0.0045
590 nm	0.0000	0.000	0.0000	0.0045
	0.5530	0.551	0.0020	0.0045
	0.7555	0.753	0.0025	0.0045
	1.0761	1.073	0.0031	0.0045
635 nm	0.0000	0.000	0.0000	0.0045
	0.5604	0.559	0.0014	0.0045
	0.7418	0.739	0.0028	0.0045
	1.0467	1.044	0.0027	0.0045

บริษัท ดีเคเอสเอช เทคโนโลยี จำกัด
DKSH Technology Limited
2533 ถนนสุขุมวิท แขวงคลองเตย เขตคลองเตย กรุงเทพมหานคร 10260
2533 Sukhumvit Road, Bangchak, Phra Khanong, Bangkok 10260
Phone: +66 2639 7000 Email: info.calibration@dksh.com Website: www.dksh.com/scientific-thailand

Delivering Growth – In Asia and Beyond.

CAL-FM-C06-16: 11 Mar 2024

Calibration Results:

Without Adjustment

Photometric Accuracy (Absorbance)

Wavelength	Standard absorbance	Unit Under Calibration	Correction	Uncertainty
235 nm	0.0000	0.000	0.0000	0.0080
	0.7533	0.748	0.0053	0.0080
257 nm	0.0000	0.000	0.0000	0.0080
	0.8745	0.869	0.0055	0.0080
313 nm	0.0000	0.000	0.0000	0.0080
	0.2926	0.293	-0.0004	0.0080
350 nm	0.0000	0.000	0.0000	0.0080
	0.6486	0.644	0.0046	0.0080

Stray light *

Standard: cut-off	UUC: Wavelength (nm)	UUC: Transmission (%T)	Absorbance (A)
260.95 +/- 0.11 nm	261.0	0.9	2.046
392.04 +/- 0.11 nm	392.0	1.3	1.886

Spectral Resolution *

Nominal Concentration 0.02 % v/v	Peak	Trough	Ratio	SBW
Standard Wavelength (nm)	268.74	266.81	1.29	2.00
UUC: Wavelength (nm)	268.6	266.6		
Std Absorbance (A)	0.5137	0.3473		
UUC: Absorbance (A)	0.463	0.359		

* Calibration Marked " Not TISI Accredited " in this Certificate have been included for completeness.

The End of Certificate

บริษัท ดีเคเอสเอช เทคโนโลยี จำกัด
DKSH Technology Limited
2533 ถนนสุขุมวิท แขวงคลองเตย เขตคลองเตย กรุงเทพมหานคร 10260
2533 Sukhumvit Road, Bangchak, Phra Khanong, Bangkok 10260
Phone: +66 2639 7000 Email: info.calibration@dksh.com Website: www.dksh.com/scientific-thailand

Delivering Growth – In Asia and Beyond.

CAL-FM-C06-16: 11 Mar 2024



บริษัท ไทยยูนิค จำกัด

THAI UNIQUE CO., LTD.

80-82 ถนนประชาทิพย์ แขวงบางขุนพรหม เขตพระนคร กรุงเทพฯ 10200

80-82 Prachathipatai Rd., Bangkhunphrom, Pranakorn, Bangkok 10200

Tel. 0-2629-0191-6, 0-2280-1787, Fax. 0-2280-1788, E-mail : thawatt@thaiunique.com, Website : www.thaiunique.com

ATOMIC ABSORPTION SPECTROMETER TEST CERTIFICATE

Certificate No : SV2310/21414

Instrument Type : Atomic Absorption Spectrometer

Model : AA240FS

Serial Number : EL08043418

Organization : Test Tech Co., Ltd.

Address : 30.32 Soi 66 Rama II Rd., Sarnadarn Bangkhuntien, Bangkok 10150

Date : 25 Oct 2023

Hollow cathode lamps used

Element	Lamp number	Comments
Arsenic	56-101003-00	
Copper	56-101014-00	
Potassium	56-101042-00	
Iron	56-101027-00	
Manganese	56-101337-00	

Test description	Specification	Result	Comments
Light throughput (%Gain) or (EHT)			
Cu at 324.8 nm	≤ 64 % or 380 V	32 %	Pass
As at 193.7 nm	≤ 80 % or 540 V	55 %	Pass
K at 766.5 nm*	≤ 84 % or 540 V	64 %	Pass
Fe at 248.3 nm	≤ 80 % or 540 V	59 %	Pass
Mn at 279.5 nm	≤ 64 % or 380 V	46 %	Pass
Photometric noise Cu BGC off			
STDV @ 0 Abs	≤ 0.0001	0.0001	Pass



บริษัท ไทยยูนิค จำกัด

THAI UNIQUE CO., LTD.

80-82 ถนนประชาทิพย์ แขวงบางขุนพรหม เขตพระนคร กรุงเทพฯ 10200

80-82 Prachathipatai Rd., Bangkhunphrom, Pranakorn, Bangkok 10200

Tel. 0-2629-0191-6, 0-2280-1787, Fax. 0-2280-1788, E-mail : thawatt@thaiunique.com, Website : www.thaiunique.com

Wavelength accuracy			
Cu at 324.8 nm	323.0 nm – 326.0 nm	324.8 nm	Pass
As 193.7 nm	192.0 nm – 195.0 nm	193.7 nm	Pass
K at 766.5 nm*	765.0 nm – 768.0 nm	766.6 nm	Pass
Fe at 248.3 nm	246.8 nm - 249.8 nm	248.3 nm	Pass
Mn at 279.5 nm	278.0 nm - 281.0 nm	279.5 nm	Pass
High solids nebulizer setting**			
Uptake rate	7.2 – 10.6 ml / min	9.8 ml/min	Pass
Max Abs	≥ 0.75 Abs	0.81 Abs	Pass
Precision(%RSD)	≤ 0.5 %	0.2 %	Pass
Zeeman Background Correction Accuracy (%)***			
BCA @ Au 242.8 nm	< 3.7 %	***	***
Zeeman Magnetic Sensitivity Ratio (%)***			
MSR @ Cu 324.7 nm	> 70 %	***	***
Characteristic mass and sensitivity ****			
Sensitivity	≥ 0.21 Abs	****	****
Precision (%RSD)	≤ 4.0 %	****	****

* for Wideband PMT (Wavelength 190nm – 900nm)

** for Flame system

*** for Zeeman system

**** for Graphite furnace system

CALIBRATED BY :

Signature: [Signature]

Engineer : Suriya Nacharoen

Date : 25 / Oct / 2023



APPROVED BY :

Signature: [Signature]

Engineer : Suchai Sanguanklattichai

Date : 25 / Oct / 2023



ICP02329207

**MAINTENANCE AND TEST CERTIFICATE MODEL
OPTIMA 8000**

Customer : บริษัท เทสท์ เทคโนโลยี จำกัด
Address : 30, 32 ซอยพระรามที่ 2 ซอย 63
ถนนพระรามที่ 2 แขวงสามยุค
เขตบางขุนเทียน กรุงเทพฯ 10150
User Name: คุณเอื้อยใจ สรรค์จันทร์
Phone: 02-893-4211-17
Fax: lab_center@testtech.co.th

Date Tested: May 18, 2023
Recommendation Recertification
Period 12
Recertification Due: May 17, 2024
Date Last Certified: May 19, 2022
Visit Number: 1 of 1
PerkinElmer Phone: 02-719-6420 ext 206
PerkinElmer Fax: 02-318-5597

CONFIGURATION TESTED**MODEL SERIAL NUMBER**

OPTIMA 8000 078S1411171C

N0772045 2F1441085

EQL-180

TESTED EQUIPMENT CALIBRATION NUMBER

IPV Methods

TEST STANDARD USED PART NUMBER

Mixed standard 1/10 N069-1579

Mixed standard 1/100 N930-0221

CUSTOMER SUPPLIED COMMENTS

2 % HNO3

10 % HNO3

**ACCESSORIES/COMPONENT
NOT INCLUDED**

WinLab32 Version 5.5.0.0714

PN:6150T21E4Q1E

EXPIRATION**EXPIRATION DATE**

NOV 30, 2023

NOV 30, 2023

CUSTOMER INITIALS

Page 1 of 4



ICP02329207

**MAINTENANCE AND TEST CERTIFICATE MODEL
OPTIMA 8000****SERIAL NUMBER :** 078S1411171C**DATE TESTED :** May 18, 2023**1. MECHANICAL CHECKS**A. Inspect and clean all fans and filters. ☐B. Inspect and replace as necessary, all torch components including the RF coil. ☐C. Inspect all tubing for sign of clacking or leaking. ☐D. Adjust water and gas pressure regulator settings. ☐E. Inspect and leak check pneumatics drawers. ☐F. Clean the exterior of the instrument. ☐**2. OPTICAL CHECKS**A. Inspect and clean all optical components. ☐B. As required, check and replace all purge filters. ☐C. Recheck optical alignment. ☐**3. COOLING SYSTEM CHECKS**A. Perform preventive maintenance on chiller. ☐B. Flush out the chiller every six months. ☐**4. PERFORMANCE CHECKS**A. Torch View Alignment. ☐B. Wavelength Calibration. ☐

Page 2 of 4



ICP02329207

MAINTENANCE AND TEST CERTIFICATE MODEL
OPTIMA 8000

SERIAL NUMBER : 078S1411171C

DATE TESTED : May 18, 2023

PARAMETER	SPECIFICATION	FINAL VALUE
Spectral Resolution : UV	As 193.696 nm ≤ 0.009 nm	0.00720 nm
	Ni 231.604 nm ≤ 0.011 nm	0.00892 nm
	Ni 341.476 nm ≤ 0.015 nm	0.01343 nm
Spectral Resolution : VIS	Ba 455.403 nm ≤ 0.020 nm	0.01726 nm
Precision	Zn 206.200 nm % RSD $\leq 1.0 \%$	0.35 %
	Mg 280.271 nm % RSD $\leq 1.0 \%$	0.19 %
	Mg 285.213 nm % RSD $\leq 1.0 \%$	0.19 %
	Ba 455.403 nm % RSD $\leq 1.0 \%$	0.13 %
Detection Limits : Axial	Tl 190.801 nm 3(SD) ppb ≤ 10 ppb	1.54 ppb
	As 193.696 nm 3(SD) ppb ≤ 10 ppb	2.10 ppb
	Se 196.026 nm 3(SD) ppb ≤ 5.0 ppb	2.43 ppb
	Pb 220.353 nm 3(SD) ppb ≤ 3.0 ppb	1.40 ppb
Detection Limits : Radial	As 193.696 nm 3(SD) ppb ≤ 60 ppb	4.44 ppb
	Zn 213.857 nm 3(SD) ppb ≤ 2.0 ppb	0.12 ppb
	Mn 257.610 nm 3(SD) ppb ≤ 1.0 ppb	0.05 ppb
	La 379.478 nm 3(SD) ppb ≤ 3.0 ppb	0.21 ppb
	Ba 455.403 nm 3(SD) ppb ≤ 0.3 ppb	0.01 ppb
	Ba 493.408 nm 3(SD) ppb ≤ 0.6 ppb	0.01 ppb
BEC : Axial (IB X 1000)/(IS-IB)	Mn 257.610 nm ≤ 30 ppb	6.83 ppb
BEC : Radial (IB X 1000)/(IS-IB)	Mn 257.610 nm ≤ 30 ppb	9.29 ppb

Page 3 of 4



ICP02329207

MAINTENANCE REPORT AND IPV TEST CERTIFICATE
OPTIMA 8000

SERIAL NUMBER : 078S1411171C

DATE TESTED : May 18, 2023

Remarks :

Commissioning follow as commissioning performance sheets.

This is to certify that the above tests have been performed and the configuration tested

☒
☐

meets

does not meet

the PerkinElmer Specifications listed on this certificate.

This certificate does not modify PerkinElmer's standard terms and condition of sale, including warranty terms.

Service Department PerkinElmer Ltd.

Authorized Representative :

KLS

(Khwanchai Siangwong)

Customer Support Engineer

Page 4 of 4



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
53/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG, BANGKOK 10250
TEL. 0-2717-3000-24 FAX. 0-2719-9484



Certificate of Calibration

Certificate No. : 23H2216
Page : 1 of 2

Equipment : Dial Thermo-Hygrometer

Manufacturer: Barigo

Model : -

Serial No.: -

ID No.: EQL-064

Condition As-Received: Used Item

Received Date: 12 October 2023

Calibration Date: 17 October 2023
to 20 October 2023

Reference: 2310-0447DN

Submitted by: TEST TECH CO.,LTD. (HEAD Office)

Ambient Temperature: (25 ± 3) °C

30, 32 Rama II Soi 63, Rama II Rd.,

Relative Humidity: (50 ± 20) %

Samaedam, Bangkhunthian, Bangkok 10150

Procedure used: Calibration were conducted using in-house calibration procedure CP-H02 according to comparison with standard chilled mirror sensor for humidity measurement function and comparison with standard temperature probe for temperature measurement function into humidity / temperature chamber.

Condition of this result of calibration

1. Reference standards instruments :

Instrument	Model	Serial No.	Certificate No.	Due Date
1) Handheld Thermometer With Sensor	1523	3240076	231305	15 Mar 2024
2) Dew Point Hygrometer	Optidew 401	164756	TH-0158-22	13 Dec 2023

2. The certificate is valid only to the item calibrated on date and place of calibration.

3. This Certification is traceable to the International System of Unit maintained through:-

- Technology Promotion Association (Thailand-Japan), NSC-ONSC Accredited No. Calibration 0008
- National Institute of Metrology Thailand (NIMT)

Calibrated by : Surasit Phansudnoi
Issue Date : 26 October 2023

Approved Signatory :

[] Chakrit Waewwanjua
[] Ponthippa Tameyakul
[✓] Viporn Tantiyawutti

B 0327545



Cert. No.: 23H2216
Page.: 2 of 2

Result of Calibration:- Without Adjustment
Function: Humidity Measurement

Reference Temperature (°C)	Standard Humidity (%R.H.)	UUC* Reading (%R.H.)	Error (%R.H.)	Uncertainty of Measurement (±%R.H.)
25.0	30.1	30.0	-0.1	1.5
25.0	40.1	39.0	-1.1	1.5
25.0	50.1	49.0	-1.1	1.7
25.0	60.0	59.0	-1.0	1.7
25.0	75.2	75.5	0.3	1.8

Result of Calibration:- Without Adjustment
Function: Temperature Measurement

Standard Temperature (°C)	UUC* Reading (°C)	Error (°C)	Uncertainty of Measurement (±°C)
15.046	15.0	-0.046	0.72
19.975	20.0	0.025	0.72
25.022	25.0	-0.022	0.72
30.000	30.0	0.000	0.72

UUC* : Unit Under Calibration

The reported uncertainty of measurement was base on standard uncertainty multiplied by coverage factor k = 2.00, providing confidence level approximately 95%.

-o0o-

a 1185882



QUALITY CALIBRATION CO.,LTD.
235 Petchkasem 63/2 Road, Laksong, Bangkoe, Bangkok 10160
Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4584
www.qcalibration.com



CERTIFICATE No : 23M6754
REFERENCE No : 69854-1

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : DIGITAL BALANCE
MANUFACTURER : SARTORIUS
MODEL : BP210S
SERIAL No : S0736477
ID No : EQL-008
CONDITION AS RECEIVED : USED ITEM
SUBMITTED BY : TEST TECH CO., LTD.
30,32 RAMA II SOI 63, RAMA II RD.,
SAMAEDAM, BANGKHUNTHIAN, BANGKOK
10150

CALIBRATED BY : PRASERT D.
CALIBRATION DATE : 13-Jul-23

APPROVED BY : PONGSAK J.

ISSUED DATE : 17-Jul-23

RECEIVED DATE : 13-Jul-23

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.

F-G010 REV 03



QUALITY CALIBRATION CO.,LTD.
235 Petchkasem 63/2 Road, Laksong, Bangkoe, Bangkok 10160
Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4584
www.qcalibration.com

CERTIFICATE No : 23M6754

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : DIGITAL BALANCE MODEL : BP210S
MANUFACTURER : SARTORIUS S/N : S0736477
ID No : EQL-008 RECEIVED DATE : 13-Jul-23
AIR PRESSURE : 1011mbar $\pm$ 1mbar CALIBRATION DATE : 13-Jul-23
AMBIENT TEMPERATURE : 23°C $\pm$ 1°C RELATIVE HUMIDITY : 50%RH $\pm$ 10% RH

CONDITION OF THIS RESULTS OF CALIBRATION

1. THIS INSTRUMENT WAS CALIBRATED BY ACCORDING TO UKAS LAB 14 EDITION 6:2019 BY USING KNOWN WEIGHT STANDARD WEIGHT. THE BALANCE WAS ADJUSTED USING INTERNAL WEIGHT TO ADJUST. THE BALANCE HAS NO ZERO TRACKING FUNCTION. REPEATABILITY WAS MEASURED BY USING 10 REPEATED MEASUREMENTS. LINEARITY WAS MEASURED COVERING 10 POINTS, EVENLY SPREAD OVER THE RANGE. THE INSTRUMENT WAS SET ZERO BEFORE PERFORMING THE LINEARITY TEST. OFF-CENTER LOADING WAS MEASURED BY USING STANDARD WEIGHTS PLACED ON THE PAN AND MOVED TO VARIOUS POSITIONS ON THE PAN. THE INTERNAL WEIGHT WAS CHECKED BY USING

2. REFERENCE STANDARD INSTRUMENTS :-

INSTRUMENT	MODEL	SERIAL No	CERTIFICATE No	DUE DATE
1) STANDARD WEIGHT SET	E2	QK-I-151	M2302013S	02-Feb-25
2) STANDARD WEIGHT	E2	15843	M2302014S	02-Feb-25

3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.

4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.

5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-

- NATIONAL INSTITUTE OF METROLOGY (THAILAND) THROUGH CENTRAL BUREAU OF WEIGHTS&MEASURES

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT

1. ZERO SETTING FUNCTION : NORMAL


2. TARE FUNCTION : NORMAL

3. REPEATABILITY OF READING AT 200 g WAS 0 g

4. DEPARTURE FROM NOMINAL VALUE/ LINEARITY

NOMINAL VALUE (g)	BALANCE READING (g)	CORRECTION (g)	UNCERTAINTY ($\pm$ g)
0.0	0.0000	0.0000	0.000082
0.1	0.1000	0.0000	0.000083
0.2	0.2000	0.0000	0.000083
0.5	0.5000	0.0000	0.000083
1.0	1.0000	0.0000	0.000084
2.0	2.0000	0.0000	0.000084
5.0	5.0000	0.0000	0.000086
10.0	10.0000	0.0000	0.000089
20.0	20.0001	-0.0001	0.000094
50.0	49.9999	0.0001	0.00012
100.0	99.9999	0.0001	0.00019
200.0	199.9997	0.0003	0.00032

5. OFF CENTER LOADING ERROR



POINT	READING (g)
1	99.9998
2	99.9997
3	99.9998
4	99.9998
5	99.9998
OFF-CENTER LOADING	0.0001

6. INTERNAL WEIGHT ERROR : 0.00049999999988177 g

NOTE: THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA
THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A
COVERAGE FACTOR $k=2$, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.

END OF CALIBRATION REPORT

F-G010 REV 03

ภาคผนวก ก/61



CERTIFICATE No : 23T8798
REFERENCE No : 70515-6

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : HOT AIR OVEN
MANUFACTURER : MEMMERT
MODEL : UFE 500
SERIAL No : G508.0791
ID No : EQL-128
CONDITION AS RECEIVED : USED ITEM
SUBMITTED BY : TEST TECH CO., LTD.
30,32 RAMA II SOI 63, RAMA II RD., SAMAEDAM,
BANGKHUNTHIAN, BANGKOK 10150

CALIBRATED BY : CHAICHARN CH.
CALIBRATION DATE : 11-Sep-23

APPROVED BY :
PONGRAK J.
ISSUED DATE : 15-Sep-23
RECEIVED DATE : 11-Sep-23

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.

P-G010 REV : 03



CERTIFICATE No : 23T8798

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : HOT AIR OVEN
MANUFACTURER : MEMMERT
MODEL : UFE 500
ID No : EQL-128
RECEIVED DATE : 11-Sep-23
AMBIENT TEMPERATURE : 24 °C ± 1 °C
S/N : G508.0791
CALIBRATION DATE : 11-Sep-23
RELATIVE HUMIDITY : 51 %RH ± 10 %RH

CONDITION OF THIS RESULTS OF CALIBRATION


1. THIS INSTRUMENT WAS CALIBRATED ACCORDING TO TLAS G-20 BY COMPARISON WITH CALIBRATED RTD Pt100 UNDER NO LOAD CONDITION. THE TEMPERATURE PROBES WERE PLACED ON NINE POINTS AND LOCATED ONE THERMOMETER PROBE IN EACH OF THE EIGHT CORNERS OF THE CHAMBER AND WAS AWAY FROM THE EACH WALL OF 5 cm TO 10 cm. AND PLACED THE NINTH THERMOMETER PROBE WITHIN 2.5 cm. OF THE GEOMETRIC CENTER OF THE CHAMBER. THE UNIFORMITY WAS MEASURED BETWEEN REFERENCE PROBE AND OTHER PROBES AT THE SAME TIME.

2. REFERENCE STANDARD INSTRUMENTS :-

INSTRUMENT	MODEL	SERIAL No	CERTIFICATE No	DUPLICATE
1) DATA LOGGER WITH RTD	HYDRA 2635A	7301307	23T6636	10-Jul-24

3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.
4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.
5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-
- NATIONAL INSTITUTE OF METROLOGY (THAILAND) THROUGH QUALITY CALIBRATION CO.,LTD.

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT



FRONT

GENERAL INFORMATION

Overall Ambient Temperature around the Chamber (°C) variation : 1
Overall Line Voltage (V) variation : 10
Instrument Condition : Normal
Chamber Size (W*L*H): 56*40*48 cm

CHAMBER PERFORMANCE

Calibrate Point (°C)	Average All Position Temp. (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
104.0	104.49	0.28	0.66	0.93
180.0	180.25	0.32	0.62	1.11

TEMPERATURE MEASUREMENT ACCURACY TEST

Controller Temp (°C)	Indicating Temp (°C)	Measured Temperature (°C) at Spread Locations									Uncertainty (± °C)
		#1	#2	#3	#4	Ref. 5	#6	#7	#8	#9	
104.0	104.0	104.46	104.13	104.45	104.28	104.57	104.67	104.60	104.58	104.67	0.38
180.0	180.0	180.27	179.85	180.41	179.93	180.19	180.54	180.41	180.51	180.13	1.1

NOTE 1 : THE UNCERTAINTY OF MEASUREMENT EXCLUDED TEMPERATURE UNIFORMITY OF THE CHAMBER.

NOTE 2 : LOCATION 5 WAS REFERENCE LOCATION.

NOTE 3 : THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA.

THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR k=2, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.
END OF CALIBRATION REPORT

P-G010 REV : 03



QUALITY CALIBRATION CO.,LTD.
235 Petchkasem 63/2 Road, Laksong, Bangkac, Bangkok 10160
Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4584
www.qcalibration.com



CERTIFICATE No : 23T8799
REFERENCE No : 70515-7

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : HOT AIR OVEN
MANUFACTURER : MEMMERT
MODEL : UFE 500
SERIAL No : G512.2005
ID No : EQL-161
CONDITION AS RECEIVED : USED ITEM
SUBMITTED BY : TEST TECH CO., LTD.
30,32 RAMA II SOI 63, RAMA II RD., SAMAEDAM,
BANGKHUNTHIAN, BANGKOK 10150

CALIBRATED BY : CHAICHARN CH.
CALIBRATION DATE : 11-Sep-23

APPROVED BY :
PONGSAK J.
ISSUED DATE : 15-Sep-23
RECEIVED DATE : 11-Sep-23

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF
QUALITY CALIBRATION CO., LTD.

F-G010 REV : 03



QUALITY CALIBRATION CO.,LTD.
235 Petchkasem 63/2 Road, Laksong, Bangkac, Bangkok 10160
Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4584

CERTIFICATE No : 23T8799

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : HOT AIR OVEN
MANUFACTURER : MEMMERT
MODEL : UFE 500
ID No : EQL-161
RECEIVED DATE : 11-Sep-23
AMBIENT TEMPERATURE : 24 °C ± 1 °C
S/N : G512.2005
CALIBRATION DATE : 11-Sep-23
RELATIVE HUMIDITY : 51 %RH ± 10 %RH

CONDITION OF THIS RESULTS OF CALIBRATION


1. THIS INSTRUMENT WAS CALIBRATED ACCORDING TO TLAS G-20 BY COMPARISON WITH CALIBRATED RTD Pt100 UNDER NO LOAD CONDITION. THE TEMPERATURE PROBES WERE PLACED ON NINE POINTS AND LOCATED ONE THERMOMETER PROBE IN EACH OF THE EIGHT CORNERS OF THE CHAMBER AND WAS AWAY FROM THE EACH WALL OF 5 cm TO 10 cm. AND PLACED THE NINTH THERMOMETER PROBE WITHIN 2.5 cm. OF THE GEOMETRIC CENTER OF THE CHAMBER. THE UNIFORMITY WAS MEASURED BETWEEN REFERENCE PROBE AND OTHER PROBES AT THE SAME TIME.

2. REFERENCE STANDARD INSTRUMENTS :-

INSTRUMENT	MODEL	SERIAL No	CERTIFICATE No	DUE DATE
1) DATA LOGGER WITH RTD	HYDRA 2635A	7301307	23T6636	10-Jul-24

3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.
4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.
5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-
- NATIONAL INSTITUTE OF METROLOGY (THAILAND) THROUGH QUALITY CALIBRATION CO.,LTD.

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT



GENERAL INFORMATION

Overall Ambient Temperature around the Chamber (°C) variation : 1
Overall Line Voltage (V) variation : 10
Instrument Condition : Normal
Chamber Size (W*L*H): 56*40*48 cm

CHAMBER PERFORMANCE

Calibrate Piont (°C)	Average All Position Temp. (°C)	Temperature Stability (±°C)	Temperature Uniformity (°C)	Overall Variation (°C)
104.0	103.96	0.14	0.58	0.73
180.0	179.55	0.22	0.93	1.47

TEMPERATURE MEASUREMENT ACCURACY TEST

Controller Temp (°C)	Indicating Temp (°C)	Measured Temperature (°C) at Spread Locations									Uncertainty (± °C)
		#1	#2	#3	#4	Ref. 5	#6	#7	#8	#9	
104.0	104.0	104.16	104.13	104.20	103.98	103.76	103.76	104.06	103.71	103.93	0.38
180.0	180.0	179.73	179.89	180.04	179.54	179.30	178.98	179.75	178.97	179.77	1.1

NOTE 1 : THE UNCERTAINTY OF MEASUREMENT EXCLUDED TEMPERATURE UNIFORMITY OF THE CHAMBER.
NOTE 2 : LOCATION 5 WAS REFERENCE LOCATION.

NOTE 3 : THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA.
THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR k =2, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.
END OF CALIBRATION REPORT

F-G010 REV : 03

Certificate of Calibration

Equipment: TURBIDIMETER
Model: 2100N
Serial No. (or ID.): 970400003415 (EQL-024)
Manufacturer: HACH
Condition: In Condition

Certificate No.: C08230153
Issued Date: 15 September 2023
Job No.: WO-00005226
Page: 1 of 2

Customer: TEST TECH CO., LTD.
 30,32 Rama II Soi 63, Rama II Rd.,
 Samaedam, Bangkhuntien Bangkok 10150 Thailand

Environment Condition: Temperature 23 °C ± 2 °C
 Humidity 50 %RH ± 15 %RH


Calibration Place: Environment Laboratory, DKSH Technology Limited.
 2533 Sukhumvit Road, Bangchak,
 Phrakhanong, Bangkok 10260 Thailand

Calibration By: Miss.Orawan Khlaiphloi
Calibration Date: 14 September 2023
The Method used: In house method, CAL-WI-23, base on Hach Manufacturer Method 8195
Traceability: This certificate is traceable to Primary standard Fromazin and StablCal accepted by United States Environmental Protection Agency (EPA) through Hach Company Certificate No. A1075 , A1074 , A1091 , A1074 , A1074



(Miss Orawan Khlaiphloi)

Person in charge



(Mr. Nitinun Srihawan)

Authorized signatory

This certificate is issued the units of measurement according to the International System of Units (SI). It provides traceability of measurement to International or national standard or other recognized national standard laboratories.

The measurement uncertainty stated is the expanded uncertainty which is obtained from the standard uncertainty multiplied by the coverage factor ($k=2$) to provide a level of confidence of approximately 95%. It is determined in accordance with the Guide to Expression of Uncertainty in Measurement (GUM).

These results may be affected by deviations from specified conditions. The results relate only to the items tested, calibrated or sampled. The report shall not be reproduced except in full without approval of DKSH Technology Limited.

บริษัท ดีเคเอส อีเซีย จำกัด
 DKSH Technology Limited
 2533 ถนนสุขุมวิท แขวงคลองเตย เขตคลองเตย กรุงเทพมหานคร 10260
 2533 Sukhumvit Road, Bangchak, Phrakhanong, Bangkok 10260
 Phone: +66 2636 7000 Email: info.calibration@dksh.com Website: www.dksh.com/scientific-thailand

Delivering Growth - In Asia and Beyond.

CAL-FM-C08-08: 20 Jul 2022

Certificate No.: C08230153

Page 2 of 2

Calibration Results:

Before Adjustment

Std Turbidity (NTU)	UUC Reading	Correction	Deviation	Uncertainty
0.050	0.088	-0.038	0.0	0.070
20.40	19.1	1.30	0.0	1.0
205.0	195	10.0	0.5	10
1028.0	952	76.0	0.9	50
4068.0	3942	126.0	0.9	200

After Adjustment

Std Turbidity (NTU)	UUC Reading	Correction	Deviation	Uncertainty
0.050	0.084	-0.034	0.0	0.070
20.40	20.4	0.00	0.0	1.0
205.0	205	0.0	0.5	10
1028.0	1026	2.0	0.5	50
4068.0	4063	5.0	0.5	200

The End of Certificate

บริษัท ดีเคเอส อีเซีย จำกัด
 DKSH Technology Limited
 2533 ถนนสุขุมวิท แขวงคลองเตย เขตคลองเตย กรุงเทพมหานคร 10260
 2533 Sukhumvit Road, Bangchak, Phrakhanong, Bangkok 10260
 Phone: +66 2636 7000 Email: info.calibration@dksh.com Website: www.dksh.com/scientific-thailand

Delivering Growth - In Asia and Beyond.

CAL-FM-C08-08: 20 Jul 2022

๗๒

ใบรับรองการสอบเทียบ "เครื่องวัดความนำไฟฟ้า"
(Calibration Certificate of Conductivity Meter)



TECHNOLOGY PROMOTION ASSOCIATION (THAILAND-JAPAN)
CORPORATE SERVICES 3: EQUIPMENT CALIBRATION AND TESTING SERVICES
534/4 PATTANAKARN ROAD SOI 18, SUANLUANG, SUANLUANG BANGKOK 10250
TEL. 0-2717-3000-29 FAX. 0-2719-9484



Cert.No.: 24CH59
Page.: 1 of 3

Certificate of Calibration

Equipment :	Conductivity Meter
Manufacturer :	TOA
Model :	CM-41X
Serial No. :	842572
ID No. :	EQL-211
Condition As-Received:	Used Item
Received Date :	11 January 2024
Calibration Date :	15 January 2024
Reference :	2401-0300DN-1
Submitted by :	TEST TECH CO.,LTD. (HEAD Office) 30, 32 Rama II Soi 63, Rama II Rd., Samaedam, Bangkhunthian, Bangkok 10150
Ambient Temperature :	(25 ± 2.5) °C
Relative Humidity :	(50 ± 15) %
Calibration Procedure:	In -house method : - CP-CH6 by direct measurement with certified reference material (CRM) - CP-CH8 by comparison with standard thermometer

Calibrated by : Warakorn Lerngagtrakul

Approved by :

Saithip

Approved Signatory

(✓) Saithip Meangmai
() Warakorn Lerngagtrakul
() Ponpan Paipim

Issue Date : 17 January 2024

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full, except with the prior written
Approval of the head of Corporate Services 3, Equipment Calibration and Testing Services.

A 0062587



Cert.No.: 24CH59

Page.: 2 of 3

Condition of this result of calibration

1. Reference Standard Instrument :-

Instrument	Serial No.	ID No.	Certificate No.	Due date
1) Thermometer	1963878	130RC095	2311051	05 Sep 2024
2) Ref. Std. Thermometer	4982054	110RC044	231908	26 Jul 2024

- This Certification is traceable to SI Through Technology Promotion Association (Thailand - Japan)

2. Certified Reference Materials :-

- Conductivity calibration solution, CPA chem Ltd., The measurement results are traceable to SI through CPA chem Ltd., ANSI-ASQ National Accreditation Board, Accredited No. AR-1835

Conductivity Solution	Manufacturer	Lot No.	Exp. date
147.0 $\mu\text{S/cm}$	CPA Chem	913595	14 July 2024
1.413 mS/cm	CPA Chem	931955	30 Sep 2024
12.880 mS/cm	CPA Chem	913597	14 July 2024

- Control Conductivity calibration solution temperature by Water bath (25.0.1) °C

3. This certificate is valid only to the item calibrated on date and place of calibration.

Calibration resultsFunction : Conductivity Measurement(*) After Adjustment at 147.0, 1413.0, 12880 $\mu\text{S/cm}$

Conductivity Electrode Serial No.: 806F0005

Standard Conductivity Solution	After Adjustment UUC* Reading	Uncertainty of Measurement ($\pm$)	Coverage factor k
147.0 $\mu\text{S/cm}$	147.1 $\mu\text{S/cm}$	0.99 $\mu\text{S/cm}$	2.00
1.413 mS/cm	1.413 mS/cm	0.0092 mS/cm	2.00
12.880 mS/cm	12.88 mS/cm	0.086 mS/cm	2.00

Remark

- UUC* = Unit Under Calibration

- Adjustment Cell constant = 147.0 $\mu\text{S/cm}$ 96.8 m^{-1} , 1.413 mS/cm = 98.0 m^{-1} , 12.880 mS/cm = 99.4 m^{-1}

Santhip

a 1197672



Cert.No.: 24CH59

Page.: 3 of 3

Calibration ResultsFunction : Temperature Measurement

(*) Without adjustment

This equipment was connected with Temperature Probe;

- Model : CT-58101B

- Serial No. : 806F0005

Dimension of probe;

- Length : 114 mm

- Diameter : 12 mm

- Immersion Depth : 100 mm

Calibration Point (°C)	Standard Temperature (°C)	UUC* Reading (°C)	Error (°C)	Uncertainty of Measurement ($\pm$ °C)	Coverage factor k
25.0	25.003	25.0	-0.003	0.13	2.00

Remark : - UUC* = Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95 %.

-o0o-

Santhip

a 1197671

ใบรับรองการสอบเทียบ "ห้องเย็น"
(Calibration Certificate of Cool Room)

๗๑



Metrology

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhoi, Saraburi 18110, Thailand.

Saraburi Tel : +66 3627 3096 Fax : +66 3627 3100

Bangkok Tel : +668 9205 6851 , +669 8247 2360

Website : www.scieco.co.th E-Mail : calibrate@scg.com



NSC-TISI-TIS 17025
CALIBRATION 0244

Certificate No. T240070

Page 1 of 4

Certificate of Calibration

Equipment : Chamber (Cooling Room)

Manufacturer : -

Model : -

Serial No. : -

Customer Code : EQL-167

ID No. : T1447A1

Customer : Test Tech Co.,Ltd

30, 32 Rama II Soi 63, Rama II Rd., Samaedam,

Bangkhunthian Bangkok 10150

Customer Location : LABORATORY FLOOR 3

Date of Receipt : 12 January 2024

Calibrated By : Sujjar Naknakred (Site Calibration Manager)

Approved By :  / Boonchai Suriyawong (Site Calibration Manager)

Date of Issue : 24 JAN 2024

The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standard laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Metrology



Metrology

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhohi, Saraburi 18110, Thailand.



Certificate No. T240070

Page 2 of 4

Calibration Report

Equipment : Chamber (Cooling Room)
Date of Calibration : 16 January 2024
Environment : Temperature : 19.4-24.1 °C
Line Voltage : 221.3-226.1 V
Relative Humidity : 55 - 65 %RH

Condition of this results of calibration :

1. This equipment was calibrated by insert nine standard thermocouples type T into its chamber , the other one standard thermocouples type T use for ambient temperature measurement . The calibration was done in according to WI-T20 (based on ASTM E145-94 (Reapproved 2001) and AS2853-1986).
All data show below were final values and the initial data from customer request . The temperature scale used was based on ITS - 90 .

2. Reference Standard Instrument :

Instrument	Model	Instrument No.	Certificate No.	Due Date
TC	TYPE T	TN161-TN170	T230773	10 April 2024
TC	TYPE T	TN161-TN170	T230773	10 April 2024
DATA LOGGER	34970A	T149	T230773	10 April 2024

3. This certificate is traceable to :

National Institute of Metrology (Thailand) through Metrological Center (NSC-TISI-TIS 17025 CALIBRATION 0244.)

4. Condition of calibrated item : good

Equipment Description :

Time Constant - Hour 37 Minute At 3 °C
Fresh Air Damper ☐ Open ☐ Min ☐ Medium ☐ Max
☐ Close
☒ Not Available

5. Adjustment :

() without adjustment (X) after adjustment

Approved By. 

FM-L15 I18/18-08-66



Metrology

SCI ECO Services Company Limited


33/2 Moo 3, T.Banpa, A.Kaengkhohi, Saraburi 18110, Thailand.



Certificate No. T240070

Page 3 of 4


Calibration Report



C = Centre , F = Centre of Face , A = Corner , E = Centre of Edge

1C	=	TN161
2A	=	TN162
3A	=	TN163
4F	=	TN164
5A	=	TN165
6A	=	TN166
7F	=	TN167
8F	=	TN168
9A	=	TN169
10A	=	TN170

11F	=	TN161
12F	=	TN162
13A	=	TN163
14A	=	TN164
15F	=	TN165

Approved By. 

FM-L15 I18/18-08-66



Metrology

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhoi, Saraburi 18110, Thailand.



NSC-TISI-TIS 17025
CALIBRATION 0244

Certificate No. T240070

Page 4 of 4

Calibration Report

Measurement Results:

Calibration Point	Average Standard Reading at each position (°C)									
	TN161	TN162	TN163	TN164	TN165	TN166	TN167	TN168	TN169	TN170
3	3.17	3.11	3.11	3.33	2.94	3.06	2.95	3.17	2.86	2.59
	TN161	TN162	TN163	TN164	TN165					
	2.74	2.95	2.75	2.95	2.85					

Chamber (Cooling Room)			Temperature Distribution				
Setting (°C)	Reading (°C)		Average (°C)	Stability (± °C)	Uniformity (°C)	Uncertainty (± °C)	Coverage
	Min , Max	Average					Factor k
3.0	2.9 , 3.1	3.0	2.97	0.29	0.64	0.80	2.00

* The quoted uncertainty exclude " uniformity "

The calibration result apply only the above calibrated item.

The result of test was found accurate as shown on date and place of test only.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k which for a t-distribution, providing a level of confidence of approximately 95 % .

Approved By.

FM-L15 I18/18-08-66



Metrology

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhoi, Saraburi 18110, Thailand.

Saraburi Tel : +66 3627 3096 Fax : +66 3627 3100

Bangkok Tel : +668 9205 6851 , +669 8247 2360

Website : www.scieco.co.th E-Mail : calibrate@scg.com



NSC-TISI-TIS 17025
CALIBRATION 0244

Certificate No. T240161

Page 1 of 4

Certificate of Calibration

Equipment : Chamber (Cooling Room)

Manufacturer : -

Model : -

Serial No. : -

Customer Code : EQL-181

ID No. : T0399A5

Customer : Test Tech Co.,Ltd

30, 32 Rama II Soi 63, Rama II Rd., Samaedam,

Bangkhunthian Bangkok 10150

Customer Location : LABORATORY FLOOR 4

Date of Receipt : 24 January 2024

Calibrated By : Preecha Phisassutthikul (Temperature Calibration Manager)

Approved By : / Boonchai Suriyawong (Site Calibration Manager)

Date of Issue : 31 JAN 2024

The uncertainties are for a confidence probability of approximately 95%.

This Certificate is issued in accordance with the conditions of accreditation granted by the Thai Laboratory Accreditation Scheme which has assessed the measurement capability of the laboratory and its traceability to recognized national standards and to the units of measurement realized at the corresponding national standard laboratory. This certificate may not be reproduced other than in full except with the prior written approval of the Metrology.

FM-L14 I19/18-08-66



Metrology

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhoi, Saraburi 18110, Thailand.



Certificate No. T240161

Page 2 of 4

Calibration Report

Equipment : Chamber (Cooling Room)
Date of Calibration : 29 January 2024
Environment : Temperature : 25.4-27.9 °C
Line Voltage : 223.4-227.1 V
Relative Humidity : 45 - 49 %RH

Condition of this results of calibration :

1. This equipment was calibrated by insert 15 standard thermocouples type T into its chamber , the other one standard thermocouples type T use for ambient temperature measurement . The calibration was done in according to WI-T20 (based on ASTM E145-94 (Reapproved 2001) and AS2853-1986).

All data show below were final values and the initial data from customer request . The temperature scale used was based on ITS - 90 .

2. Reference Standard Instrument :

Instrument	Model	Instrument No.	Certificate No.	Due Date
TC	TYPE T	TN161-TN170	T230773	10 April 2024
TC	TYPE T	TN171-TN180	T230773	10 April 2024
DATA LOGGER	34970A	T149	T230773	10 April 2024

3. This certificate is traceable to :

National Institute of Metrology (Thailand) through Metrological Center (NSC-TISI-TIS 17025 CALIBRATION 0244.)

4. Condition of calibrated item : good

Equipment Description :

Time Constant 1 Hour 30 Minute At 3 °C
Fresh Air Damper ☐ Open ☐ Min ☐ Medium ☐ Max
☐ Close
☒ Not Available

5. Adjustment :

(X) without adjustment

() after adjustment

Approved By

FM-L15 I18/18-08-66



Metrology

SCI ECO Services Company Limited


33/2 Moo 3, T.Banpa, A.Kaengkhoi, Saraburi 18110, Thailand.



Certificate No. T240161

Page 3 of 4

Calibration Report



C = Centre , F = Centre of Face , A = Corner , E = Centre of Edge

1C = TN161	12F = TN172
2A = TN162	13A = TN173
3A = TN163	14A = TN174
4F = TN164	15F = TN175
5A = TN165	
6A = TN166	
7E = TN167	
8F = TN168	
9A = TN169	
10A = TN170	
11F = TN171	

Approved By

FM-L15 I18/18-08-66



Metrology

SCI ECO Services Company Limited

33/2 Moo 3, T.Banpa, A.Kaengkhoi, Saraburi 18110, Thailand.



Certificate No. T240161

Page 4 of 4

Calibration Report

Measurement Results

Calibration Point	Average Standard Reading at each position (°C)									
	TN161	TN162	TN163	TN164	TN165	TN166	TN167	TN168	TN169	TN170
3.0	2.81	3.01	2.99	2.87	2.92	3.08	3.04	2.93	3.31	3.10
	TN171	TN172	TN173	TN174	TN175					
	3.08	3.10	3.40	3.00	3.24					

Chamber (Cooling Room)			Temperature Distribution				
Setting (°C)	Reading (°C)		Average (°C)	Stability (± °C)	Uniformity (°C)	Uncertainty (± °C)	Coverage Factor <i>k</i>
	Min , Max	Average					
3.0	2.8 , 3.1	3.0	3.06	0.40	0.92	1.07	2.00

* The Acuoted uncertainty exclude "uniformity"

The calibration result apply only the above calibrated item.

The result of test was found accurate as shown on date and place of test only.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor *k* which for a t-distribution, providing a level of confidence of approximately 95 % .

Approved By _____

MAINTENANCE AND IPV TEST CERTIFICATE MODEL
OPTIMA 8000

Customer : บริษัท เทสท์ เทคโนโลยี จำกัด.
Address : 30,32 ซอยพรวรรณที่ 2 ซอย 63
ถนนพรวรรณที่ 2 แขวงสามยุค
เขตบางขุนเทียน กรุงเทพมหานคร 10150
User Name: คุณจุไรรัตน์ จงประกอบกิจ
Phone: 02-893-4211-17
Fax: lab_center@testtech.co.th

Date Tested: May 14, 2024
Recommendation Recertification
Period 12 **Months**
Recertification Due: May 14, 2025
Date Last Certified: May 18, 2023
Visit Number: 1 of 1
PerkinElmer Phone: 02-719-6420 ext 206
PerkinElmer Fax: 02-318-5597

CONFIGURATION TESTED

MODEL	SERIAL NUMBER	SOFTWARE
OPTIMA 8000 (EQL-180)	078S1411171C	ICP WinLab32 version 5

TESTED EQUIPMENT	CALIBRATION NUMBER	EXPIRATION
IPV Methods		

TEST STANDARD USED	PART NUMBER	EXPIRATION DATE
Multielement Standard	N069-1579	Dec 30, 2024
Instrument Cal. STD4	N930-0221	Nov 30, 2024

CUSTOMER SUPPLIED	COMMENTS	CUSTOMER INITIALS
2 % HNO3		
10 % HNO3		

MAINTENANCE AND IPV TEST CERTIFICATE MODEL
OPTIMA 8000

SERIAL NUMBER: 078S1411171C **DATE TESTED:** May 14, 2024

1. MECHANICAL CHECKS

A. Inspect and clean all fans and filters. ☐ OK

B. Inspect and replace as necessary, all torch components including the RF coil. ☐ OK

C. Inspect all tubing for sign of clacking or leaking. ☐ OK

D. Adjust water and gas pressure regulator settings. ☐ OK

E. Inspect and leak check pneumatics drawers. ☐ OK

F. Clean the exterior of the instrument. ☐ OK

2. OPTICAL CHECKS

A. Inspect and clean all optical components. ☐ OK

B. As required, check and replace all purgefilters. ☐ OK

C. Recheck optical alignment. ☐ OK

3. COOLING SYSTEM CHECKS

A. Perform preventive maintenance on chiller. ☐ OK

B. Flush out the chiller every six months. ☐ OK

4. PERFORMANCE CHECKS

A. Torch View Alignment. ☐ OK

B. Wavelength Calibration. ☐ OK

MAINTENANCE AND IPV TEST CERTIFICATE MODEL
OPTIMA 8000

SERIAL NUMBER: 078S1411171C		DATE TESTED: May 14, 2024	
PARAMETER	SPECIFICATION		FINAL VALUE
Spectral Resolution : UV			
As 193.696 nm	≤ 0.009 nm	0.00735	nm
NI 231.804 nm	≤ 0.011 nm	0.00913	nm
NI 341.476 nm	≤ 0.015 nm	0.01386	nm
Spectral Resolution : VIS			
Ba 455.403 nm	≤ 0.020 nm	0.01721	nm
Precision			
Zn 206.200 nm	% RSD ≤ 1.0 %	0.35	%
Mg 280.271 nm	% RSD ≤ 1.0 %	0.27	%
Mg 285.213 nm	% RSD ≤ 1.0 %	0.46	%
Ba 455.403 nm	% RSD ≤ 1.0 %	0.48	%
Detection Limits : Axial			
Ti 190.801 nm	3(sd) ≤ 10.0 ppb	1.00	ppb
As 193.696 nm	3(sd) ≤ 10.0 ppb	3.32	ppb
Se 196.026 nm	3(sd) ≤ 5.0 ppb	3.88	ppb
Pb 220.353 nm	3(sd) ≤ 3.0 ppb	1.45	ppb
Detection Limits : Radial			
As 193.696 nm	3(sd) ≤ 60.0 ppb	3.41	ppb
Zn 213.857 nm	3(sd) ≤ 2.0 ppb	0.30	ppb
Mn 257.610 nm	3(sd) ≤ 1.0 ppb	0.03	ppb
La 379.478 nm	3(sd) ≤ 3.0 ppb	0.27	ppb
Ba 455.403 nm	3(sd) ≤ 0.3 ppb	0.05	ppb
Ba 493.408 nm	3(sd) ≤ 0.6 ppb	0.06	ppb
BEC : Axial (IB X 1000)/(IS-IB)			
Mn 257.610 nm	≤ 30 ppb	10.70	ppb
BEC : Radial (IB X 1000)/(IS-IB)			
Mn 257.610 nm	≤ 30 ppb	21.54	ppb

MAINTENANCE AND IPV TEST CERTIFICATE MODEL
OPTIMA 8000

SERIAL NUMBER: 078S1411171C DATE TESTED: May 14, 2024

Remarks :

Commissioning follow as commissioning performance sheets.

This is to certify that the above tests have been performed and the configuration tested

☒ meets
☐ does not meet

the PerkinElmer Specifications listed on this certificate.

This certificate does not modify PerkinElmer's standard terms and condition of sale, including warranty terms.

Service Department PerkinElmer Scientific (Thailand)Co.,Ltd.

Customer Service Engineer:

()
Narong Watanakit
Service Engineer