

ภาคผนวกที่ 6

เอกสารสอบเทียบความถูกต้องของเครื่องมือ

เอกสาร 6-1	เอกสารสอบเทียบเครื่องมือการตรวจวัดคุณภาพอากาศ
เอกสาร 6-2	เอกสารการสอบเทียบเครื่องมือการตรวจวัดระดับเสียง
เอกสาร 6-3	เอกสารสอบเทียบเครื่องมือการตรวจวิเคราะห์คุณภาพน้ำ
เอกสาร 6-4	เอกสารสอบเทียบเครื่องมือการตรวจวิเคราะห์คุณภาพตะกอนดิน

**ตารางสรุปรายการเอกสารการสอบเทียบความถูกต้องของเครื่องมือเก็บตัวอย่าง
และเครื่องมือตรวจวิเคราะห์คุณภาพสิ่งแวดล้อม**

รายการตรวจวัด	เครื่องมือเก็บตัวอย่าง	เครื่องมือตรวจวิเคราะห์
	ชื่อเครื่องมือ	ชื่อเครื่องมือ
คุณภาพน้ำ		
- Temperature	-	- Thermometer
- Turbidity	-	- Turbidity Meter
- pH	-	- pH Meter
- Salinity	-	- Conductivity Meter
- Conductivity	-	- Conductivity Meter
- SS	-	- Digital Balance
- TDS	-	- Digital Balance
- DO	-	- DO Meter
- BOD ₅	-	- DO Meter
- Grease & Oil	-	- Digital Balance
คุณภาพอากาศ		
- TSP	- High Volume Air Sampler No. B32, B38, B40, B44	- Digital Balance
- PM-10	- PM10 Air Sampler No. B01, B02, B03, B04	- Digital Balance
- PM-2.5	- PM-2.5 Air Sampler No. DG-B09, B10, B11, B12	-
- SO ₂	- Gas Sampler Box No. B03, B05, B12, B14	- Spectrophotometer
- CO	- Personal Pump SKC No. B06, B07, B11, B14	- CO Analyzer No. B14
- NO ₂	- NO ₂ Analyzer No. B07, B09, B16, B19	- NO ₂ Analyzer No. B01, B02, B03, B04
ระดับเสียง		
- Leq 8 hr, Leq 24 hr, L90, Lmax, Ldn	- Acoustic Calibrator - Sound Level Meter ACO-No. B16, R42 ACO-C1-B01 ST-C1-B24, B27, B29	-

ตารางสรุปรายการเอกสารการสอบเทียบความถูกต้องของเครื่องมือเก็บตัวอย่าง
และเครื่องมือตรวจวิเคราะห์คุณภาพสิ่งแวดล้อม (ต่อ)

รายการตรวจวัด	เครื่องมือเก็บตัวอย่าง	เครื่องมือตรวจวิเคราะห์
	ชื่อเครื่องมือ	ชื่อเครื่องมือ
คุณภาพตะกอนดิน		
- Pb	-	- ICP
- Zn	-	- ICP
- Fe	-	- ICP

เอกสาร 6-1

เอกสารสอบเทียบเครื่องมือการตรวจวัดคุณภาพอากาศ



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจตุจักร เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72 Fax : (662) 513-4221 E-mail : sale@spscn.com, www.spscn.com

High Volume Air Sampler Calibration Report

Calibration Method : Multipoint Orifice Flow Transfer Standard Model : TE 5025A S/N : 3611

Calibration Data

High Volume Air Sampler Data		Calibration Data		
Recorder No.	Blower No.	Date	Actual Flowrate (ft ³ /min)	R ²
B01	B01	02/02/2026	y = 1.147x-3.983	0.998
B02	B02	02/02/2026	y = 1.158x-3.531	0.997
B03	B03	02/02/2026	y = 1.109x-2.595	0.999
B04	B04	02/02/2026	y = 1.123x-1.928	0.998
B05	B05	02/02/2026	y = 1.145x-3.659	0.999
B06	B06	02/02/2026	y = 1.100x-2.379	0.999
B07	B07	02/02/2026	y = 1.122x-2.377	0.999
B08	B08	02/02/2026	y = 1.102x-2.500	0.998
B09	B09	02/02/2026	y = 1.121x-3.479	0.999
B10	B10	02/02/2026	y = 1.143x-3.199	0.996
B11	B11	03/02/2026	y = 1.105x-1.251	0.999
B12	B12	03/02/2026	y = 1.136x-3.106	0.996
B13	B13	03/02/2026	y = 1.142x-2.536	0.998
B14	B14	03/02/2026	y = 1.135x-3.133	0.998
B15	B15	03/02/2026	y = 1.111x-2.932	0.998
B16	B16	04/02/2026	y = 1.104x-3.794	0.998
B17	B17	04/02/2026	y = 1.151x-2.773	0.998
B18	B18	04/02/2026	y = 1.140x-2.730	0.998
B19	B19	04/02/2026	y = 1.137x-2.624	0.998
B20	B20	04/02/2026	y = 1.139x-1.402	0.998
B21	B21	04/02/2026	y = 1.131x-1.657	0.999
B22	B22	04/02/2026	y = 1.166x-3.887	0.997
B23	B23	04/02/2026	y = 1.162x-4.278	0.998
B24	B24	04/02/2026	y = 1.128x-2.436	0.998
B25	B25	04/02/2026	y = 1.159x-4.833	0.998
B26	B26	03/02/2026	y = 1.103x-1.969	0.997
B27	B27	04/02/2026	y = 1.149x-3.761	0.997
B28	B28	04/02/2026	y = 1.134x-3.377	0.999
B29	B29	04/02/2026	y = 1.118x-2.903	0.997
B30	B30	04/02/2026	y = 1.157x-2.504	0.999
B31	B31	04/02/2026	y = 1.165x-3.108	0.999
B32	B32	04/02/2026	y = 1.113x-2.480	0.999
B33	B33	04/02/2026	y = 1.155x-2.774	0.997
B34	B34	04/02/2026	y = 1.125x-1.905	0.997

Calibrated by :



Approved by :



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจตุจักร เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72 Fax : (662) 513-4221 E-mail : sale@spscn.com, www.spscn.com

High Volume Air Sampler Calibration Report

Calibration Method : Multipoint Orifice Flow Transfer Standard Model : TE 5025A S/N : 3611

Calibration Data

High Volume Air Sampler Data		Calibration Data		
Recorder No.	Blower No.	Date	Actual Flowrate (ft ³ /min)	R ²
B35	B35	04/02/2026	y = 1.154x-3.696	0.996
B36	B36	04/02/2026	y = 1.119x-2.463	0.997
B37	B37	04/02/2026	y = 1.129x-3.721	0.998
B38	B38	02/02/2026	y = 1.182x-2.931	0.997
B39	B39	02/02/2026	y = 1.062x-1.512	0.999
B40	B40	02/02/2026	y = 1.163x-3.556	0.997
B41	B41	02/02/2026	y = 1.116x-1.975	0.996
B42	B42	02/02/2026	y = 1.152x-4.140	0.997
B43	B43	02/02/2026	y = 1.167x-3.646	0.996
B44	B44	02/02/2026	y = 1.178x-4.362	0.999
R01	R01	02/02/2026	y = 1.097x-2.630	0.997
R02	R02	02/02/2026	y = 1.138x-3.218	0.996
R03	R03	02/02/2026	y = 1.133x-3.084	0.998
R04	R04	02/02/2026	y = 1.121x-2.102	0.998
R05	R05	02/02/2026	y = 1.041x+0.894	0.997
R06	R06	02/02/2026	y = 1.075x+0.254	0.997
R07	R07	03/02/2026	y = 1.126x-2.841	0.996
R08	R08	03/02/2026	y = 1.091x-2.104	0.996
R09	R09	03/02/2026	y = 1.065x+0.557	0.996
R10	R10	03/02/2026	y = 1.093x-1.217	0.999
R11	R11	03/02/2026	y = 1.130x-3.938	0.999
R12	R12	05/02/2026	y = 1.164x-2.698	0.997
R13	R13	05/02/2026	y = 1.120x-3.336	0.998
R14	R14	05/02/2026	y = 1.034x+0.445	0.997
R15	R15	05/02/2026	y = 1.076x+0.429	0.998
R16	R16	05/02/2026	y = 1.107x-1.468	0.996
R17	R17	05/02/2026	y = 1.078x-1.668	0.998
R18	R18	05/02/2026	y = 1.160x-5.001	0.997
R19	R19	05/02/2026	y = 1.077x-2.680	0.997
R20	R20	05/02/2026	y = 1.185x-4.839	0.998

Calibrated by :



Approved by :





บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72 Fax : (662) 513-4221 E-mail : sale@spscon.com, www.spscon.com

High Volume PM-10 Air Sampler Calibration Report

Calibration Method : Multipoint Orifice Flow Transfer Standard Model : TE 5025A S/N : 3611

Calibration Data

High Volume PM-10 Data		Calibration Data		
Recorder No.	Blower No.	Date	Actual Flowrate (ft ³ /min)	R ²
B01	B01	02/02/2026	$y = 1.066x - 0.702$	0.998
B02	B02	02/02/2026	$y = 1.076x + 0.303$	0.999
B03	B03	02/02/2026	$y = 1.040x + 0.622$	0.997
B04	B04	02/02/2026	$y = 1.135x - 2.410$	0.998
B05	B05	02/02/2026	$y = 1.073x - 1.284$	0.997
B06	B06	02/02/2026	$y = 1.017x + 2.340$	0.998
B07	B07	02/02/2026	$y = 1.036x + 0.117$	0.998
B08	B08	02/02/2026	$y = 1.144x - 2.648$	0.997
B09	B09	02/02/2026	$y = 1.115x - 4.423$	0.999
B10	B10	02/02/2026	$y = 1.090x - 2.056$	0.999
B11	B11	03/02/2026	$y = 1.112x - 3.640$	0.997
B12	B12	03/02/2026	$y = 1.136x - 4.495$	0.997
B13	B13	03/02/2026	$y = 1.119x - 2.873$	0.999
B14	B14	03/02/2026	$y = 1.145x - 3.948$	0.996
B15	B15	03/02/2026	$y = 1.137x - 2.722$	0.999
B16	B16	03/02/2026	$y = 1.087x + 0.251$	0.997
B17	B17	03/02/2026	$y = 1.096x - 0.862$	0.998
B18	B18	03/02/2026	$y = 1.111x - 3.264$	0.996
B19	B19	03/02/2026	$y = 1.081x + 0.507$	0.998
B20	B20	03/02/2026	$y = 1.137x - 3.001$	0.999
B21	B21	03/02/2026	$y = 1.129x - 3.073$	0.999
B22	B22	03/02/2026	$y = 1.130x - 2.275$	0.998
B23	B23	03/02/2026	$y = 1.149x - 4.649$	0.997
B24	B24	03/02/2026	$y = 1.103x - 2.139$	0.999
B25	B25	03/02/2026	$y = 1.065x - 1.466$	0.999
B26	B26	03/02/2026	$y = 1.035x + 0.731$	0.998
B27	B27	03/02/2026	$y = 1.102x - 2.849$	0.998
B28	B28	03/02/2026	$y = 1.138x - 3.306$	0.996
B29	B29	03/02/2026	$y = 1.046x + 0.180$	0.997
B30	B30	03/02/2026	$y = 1.140x - 5.125$	0.999
B31	B31	03/02/2026	$y = 1.028x + 0.644$	0.998
B32	B32	04/02/2026	$y = 1.044x + 1.254$	0.999
B33	B33	04/02/2026	$y = 1.106x - 2.101$	0.999
B34	B34	04/02/2026	$y = 1.059x + 0.598$	0.998

Calibrated by :



Approved by :



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370 (Automatic 5 Lines) Fax : (662) 513-4221 E-mail : sale@spscon.com

CALIBRATION REPORT

PM2.5 AIR SAMPLER

DATE :	02 May 2026	SERIAL NO.	00402425091300145887		
NO.	PM2.5-DG-B09		AQMS-06		
CALIBRATING CONDITION					
Pressure	1011	mmbar	Temp. 24.5 °C % RH 50		
Calibration Method : Dry Cal Primary	Model : Defender 510 H	S/N : 136164			
CALIBRATION SETTING					
detaCal	PM2.5 AIR SAMPLER				
Flowrate Reading, L/min	Initial Flowrate Reading (Before Adj.), L/min	% Dif.	Final Flowrate Reading (After Adj.), L/min		
16.70	16.68	0.120	16.70		

Calibrated by :



Approved by :





บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel: (662) 939-4370 (Automatic 5 Lines) Fax: (662) 513-4221 E-mail: sale@spscon.com

CALIBRATION REPORT			
PM2.5 AIR SAMPLER			
DATE :	02 May 2026	SERIAL NO.	00402425091300025979
NO.	PM2.5-DG-B10		AQMS-06
CALIBRATING CONDITION			
Pressure	1011 mmbar	Temp.	24.5 °C % RH 48
Calibration Method : Dry Cal Primary		Model : Defender 510 H	S/N : 136164
CALIBRATION SETTING			
detaCal	PM2.5 AIR SAMPLER		
Flowrate Reading,L/min	Initial Flowrate Reading (Before Adj.),L/min	%Dif.	Final Flowrate Reading (After Adj.),L/min
16.70	16.65	0.299	16.70

Calibrated by : _____

Approved by : _____



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel: (662) 939-4370 (Automatic 5 Lines) Fax: (662) 513-4221 E-mail: sale@spscon.com

CALIBRATION REPORT			
PM2.5 AIR SAMPLER			
DATE :	02 May 2026	SERIAL NO.	00402425091300045188
NO.	PM2.5-DG-B11		AQMS-06
CALIBRATING CONDITION			
Pressure	1011 mmbar	Temp.	24.5 °C % RH 48
Calibration Method : Dry Cal Primary		Model : Defender 510 H	S/N : 136164
CALIBRATION SETTING			
detaCal	PM2.5 AIR SAMPLER		
Flowrate Reading,L/min	Initial Flowrate Reading (Before Adj.),L/min	%Dif.	Final Flowrate Reading (After Adj.),L/min
16.70	16.64	0.359	16.70

Calibrated by : _____

Approved by : _____



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel: (662) 939-4370 (Automatic 5 Lines) Fax: (662) 513-4221 E-mail: sale@spscon.com

CALIBRATION REPORT			
PM2.5 AIR SAMPLER			
DATE :	02 May 2026	SERIAL NO.	00402425072500032627
NO.	PM2.5-DG-B12		AQMS-06
CALIBRATING CONDITION			
Pressure	1011 mmbar	Temp.	24.5 °C % RH 50
Calibration Method : Dry Cal Primary		Model : Defender 510 H	S/N : 136164
CALIBRATION SETTING			
detaCal	PM2.5 AIR SAMPLER		
Flowrate Reading,L/min	Initial Flowrate Reading (Before Adj.),L/min	%Dif.	Final Flowrate Reading (After Adj.),L/min
16.70	16.66	0.240	16.70

Calibrated by : _____

Approved by : _____



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72 Fax : (662) 513-4221 E-mail : sale@spscon.com www.spscon.com

Gas Sampler Box Calibration Report							
Calibration Method : Dry Cal Primary Flowmeter				Model : Dry Cal DCL-ML		S/N : 136164	
Calibration Data							
Gas Sampler		Calibration Data					
No.	Rotameter	Date	Setting (Constant Flow) (ml/min)	Actual Flow Rate (ml/min)			
				Sampling Line A		Sampling Line B	
				Normal Condition	Standard Condition	Normal Condition	Standard Condition
B01	2 (A&B)	02/03/2026	200	200.6	200.7	200.3	200.4
B02	2 (A&B)	02/03/2026	200	200.4	200.5	200.6	200.7
B03	2 (A&B)	04/03/2026	200	200.3	200.4	200.1	200.2
B04	2 (A&B)	04/03/2026	200	200.4	200.5	200.7	200.8
B05	2 (A&B)	02/03/2026	200	200.2	200.3	200.5	200.6
B06	2 (A&B)	02/03/2026	200	200.4	200.5	200.1	200.2
B07	2 (A&B)	05/03/2026	200	199.8	199.9	200.3	200.4
B08	2 (A&B)	05/03/2026	200	200.2	200.3	200.6	200.7
B09	2 (A&B)	02/03/2026	200	200.6	200.7	200.4	200.5
B10	2 (A&B)	02/03/2026	200	200.5	200.6	200.7	200.8
B11	2 (A&B)	02/03/2026	200	200.3	200.4	199.8	199.9
B12	2 (A&B)	02/03/2026	200	200.4	200.5	200.1	200.2
B13	2 (A&B)	05/03/2026	200	200.2	200.3	200.5	200.6
B14	2 (A&B)	05/03/2026	200	200.4	200.5	200.7	200.8
B15	2 (A&B)	05/03/2026	200	200.5	200.6	200.2	200.3
B16	2 (A&B)	04/03/2026	200	200.2	200.3	200.6	200.7
B17	2 (A&B)	04/03/2026	200	200.1	200.2	200.4	200.5

Calibrated by : _____	Approved by : _____
-----------------------	---------------------



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72. Fax : (662) 513-4221. E-mail : sale@spscon.com. www.spscon.com

Personal Pump Calibration Report

Calibration Method : Dry Cal. Primary Flowmeter

Model : Defender 510-H

S/N : 136164

Environmental Conditions

Temperature : 25 ± 3 °C
Pressure : 1010 ± 15 mmbar

Personal Pump Data				Calibration Data									
No.	Brand	Model	Serial No.	Date	Flow Rate (ml/min)						Value From Calibration Curve		
					Setting			Actual (O std.)					
					1	2	3	1	2	3	y	R ²	
B01	SKC	224-PCXR4	262101	01/04/2026	1,000	1,500	2,000	1,002	1,503	2,004	1.002x - 1.364	1.000	
B02	SKC	224-PCXR4	626166	01/04/2026	1,000	1,500	2,000	1,004	1,501	2,003	1.001x - 0.594	1.000	
B03	SKC	224-PCXR4	612968	01/04/2026	1,000	1,500	2,000	1,003	1,502	1,994	1.003x - 0.098	0.999	
B04	SKC	224-PCXR4	602804	01/04/2026	1,000	1,500	2,000	1,000	1,502	1,998	0.998x + 0.642	1.000	
B05	SKC	224-PCXR4	612693	01/04/2026	1,000	1,500	2,000	1,001	1,504	1,999	1.000x - 1.767	1.000	
B06	SKC	224-PCXR4	262188	01/04/2026	1,000	1,500	2,000	999	1,500	2,002	1.001x - 3.590	1.000	
B07	SKC	224-PCXR4	626262	01/04/2026	1,000	1,500	2,000	997	1,506	2,003	1.004x - 9.859	0.999	
B08	SKC	224-PCXR4	626100	01/04/2026	1,000	1,500	2,000	1,003	1,503	2,005	1.003x - 4.160	1.000	
B09	SKC	224-PCXR4	626479	01/04/2026	1,000	1,500	2,000	998	1,497	1,997	1.000x - 3.382	1.000	
B10	SKC	224-PCXR4	091950	01/04/2026	1,000	1,500	2,000	1,004	1,506	1,996	1.005x - 7.323	0.999	
B11	SKC	224-PCXR4	564315	01/04/2026	1,000	1,500	2,000	1,000	1,493	2,003	1.005x - 10.601	1.000	
B12	SKC	224-PCXR4	034656	01/04/2026	1,000	1,500	2,000	997	1,495	1,997	0.998x - 1.755	1.000	
B13	SKC	224-PCXR4	602073	01/04/2026	1,000	1,500	2,000	1,001	1,501	1,996	0.999x + 2.894	0.999	
B14	SKC	224-PCXR4	626313	01/04/2026	1,000	1,500	2,000	999	1,494	1,997	0.999x - 1.898	0.999	
B15	SKC	224-PCXR4	626474	01/04/2026	1,000	1,500	2,000	996	1,498	1,993	1.007x - 17.126	0.999	
B16	SKC	224-PCXR4	626477	01/04/2026	1,000	1,500	2,000	1,002	1,499	1,997	0.997x + 1.352	1.000	
B17	SKC	224-PCXR4	626860	01/04/2026	1,000	1,500	2,000	996	1,510	1,999	1.005x - 12.252	0.999	
B18	SKC	224-PCXR4	691484	01/04/2026	1,000	1,500	2,000	995	1,499	1,997	1.006x - 12.384	1.000	
B19	SKC	224-PCXR4	691599	01/04/2026	1,000	1,500	2,000	998	1,501	1,991	1.004x - 10.506	0.999	
B20	SKC	224-PCXR4	691587	01/04/2026	1,000	1,500	2,000	997	1,495	1,998	1.003x - 7.247	1.000	
B21	SKC	224-PCXR4	691533	03/04/2026	1,000	1,500	2,000	998	1,496	1,996	1.000x - 3.454	1.000	
B22	SKC	224-PCXR4	691654	03/04/2026	1,000	1,500	2,000	996	1,494	1,997	0.998x - 2.784	1.000	
B23	SKC	224-PCXR4	798393	02/04/2026	1,000	1,500	2,000	1,000	1,503	1,999	0.999x - 1.308	1.000	
B24	SKC	224-PCXR4	626363	03/04/2026	1,000	1,500	2,000	1,001	1,509	1,995	1.002x - 6.940	0.999	
B25	SKC	224-PCXR4	798489	03/04/2026	1,000	1,500	2,000	1,002	1,502	2,004	1.004x - 7.267	1.000	
B26	SKC	224-PCXR4	798479	03/04/2026	1,000	1,500	2,000	996	1,503	2,003	1.006x - 12.675	1.000	
B27	SKC	224-PCXR4	691673	02/04/2026	1,000	1,500	2,000	1,000	1,504	2,000	0.999x + 0.483	1.000	
B28	SKC	224-PCXR4	691570	03/04/2026	1,000	1,500	2,000	997	1,510	2,001	0.998x - 0.435	0.999	
B29	SKC	224-PCXR4	626472	03/04/2026	1,000	1,500	2,000	998	1,494	1,992	1.000x - 5.221	1.000	
B30	SKC	224-PCXR4	691489	03/04/2026	1,000	1,500	2,000	1,008	1,504	2,006	1.003x - 2.114	1.000	
B31	SKC	224-PCXR4	691509	03/04/2026	1,000	1,500	2,000	1,003	1,503	2,002	0.997x + 0.610	0.999	
B32	SKC	224-PCXR4	091567	03/04/2026	1,000	1,500	2,000	1,002	1,502	1,996	0.996x + 7.008	1.000	
B33	SKC	224-PCXR4	091756	03/04/2026	1,000	1,500	2,000	998	1,501	1,992	1.006x - 12.101	0.999	
B34	SKC	224-PCXR4	612962	03/04/2026	1,000	1,500	2,000	1,001	1,494	2,004	1.004x - 8.196	1.000	
B35	SKC	224-PCXR4	602682	03/04/2026	1,000	1,500	2,000	997	1,504	1,998	1.000x - 2.804	1.000	
B36	SKC	224-PCXR4	626164	03/04/2026	1,000	1,500	2,000	996	1,512	1,996	1.001x - 5.867	0.999	
B37	SKC	224-PCXR4	626256	03/04/2026	1,000	1,500	2,000	999	1,496	2,005	1.007x - 11.487	1.000	
B38	SKC	224-PCXR4	626167	03/04/2026	1,000	1,500	2,000	1,002	1,500	2,003	1.008x - 16.377	0.999	
B39	SKC	224-PCXR4	034637	03/04/2026	1,000	1,500	2,000	998	1,490	1,994	0.996x - 0.044	1.000	
B40	SKC	224-PCXR4	798349	03/04/2026	1,000	1,500	2,000	1,001	1,501	2,000	0.995x + 0.921	0.999	

Calibrated by :

Approved by :



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72. Fax : (662) 513-4221. E-mail : sale@spscon.com. www.spscon.com

CALIBRATION REPORT

CHEMILUMINESCENT NO / NO₂ / NO_x ANALYZER

DATE : 02 May 2026

BRAND : API

MODEL : 200E

NO. NOX-807

SERIAL NO. 4338

Calibrator (Dilution System)

Brand : API

Model : 700

Last Cal. Date : 01 September 2025

Serial No. : 911

Reference Standard Gas

Standard Gas : Nitric Oxide (NO)

Cylinder No. : A006745V

Certified Date : 12 March 2025

Expired Date : 12 March 2028

Cylinder Conc. : 48.7 ppm

CALIBRATING CONDITION

Pressure : 1011 mmbar

Temp. : 24.5 °C

% RH : 48

CALIBRATION SETTING

Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB	
	Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response
Zero	0	0	0.10	-	0
NO Span	400	400	399.8	-0.050	400.0
NO _x Span	400	400	400.1	0.025	400.0

API Model 200E NO_x Analyzer Check List

Test Values	Observed Value	Units	Nominal Range
RANGE	500	PPB	500 standard
STABILITY (Zero Gas)	0.1	PPB	< 2 with zero air
SAMPLE FLOW	505	cc/min	500 ± 50
OZONE FLOW	78	cc/min	80 ± 15
PMT	103.1	mV	-20 - 150
AZERO	93.8	mV	-20 - 150
HVPS	675	V	420 - 900 constant
RCELL TEMP	50.1	°C	50 ± 1
BOX TEMP	29.5	°C	8 - 48
PMT TEMP	7.2	°C	7 ± 2
MOLY TEMP	315.3	°C	315 ± 5
RCELL PRESS	8.3	IN-Hg-A	2 - 10 constant
SAMPLE PRESS	28.6	IN-Hg-A	25 - 30 constant
NO Span Conc	400	PPB	20 - 20,000
NO _x Span Conc	400	PPB	20 - 20,000
NO Slope	1.007	-	1.0 ± 0.3
NO _x Slope	1.010	-	1.0 ± 0.3
NO Offset	1.2	mV	-20 to +150
NO _x Offset	0.8	mV	-20 to 150
Stability at Zero	0.1	PPB	< 0.2
Stability at Span	0.2	PPB	< 2 ppb @ 400 ppb span gas

Calibrated by :

Approved by :



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจตุจักร เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompoi, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

CALIBRATION REPORT					
CHEMILUMINESCENT NO / NO ₂ / NO _x ANALYZER					
DATE :	02 May 2026	BRAND :	API	MODEL :	200E
NO.	NOX-B09	SERIAL NO.	4412		
Calibrator (Dilution System)					
Brand :	API	Model :	700		
Last Cal. Date :	01 September 2025	Serial No. :	911		
Reference Standard Gas					
Standard Gas :	Nitric Oxide (NO)	Cylinder No. :	A006745V		
Certified Date :	12 March 2025	Expired Date :	12 March 2028	Cylinder Conc. :	48.7 ppm
CALIBRATING CONDITION					
Pressure	1011	mmbar	Temp.	24.5	°C
			% RH	48	
CALIBRATION SETTING					
Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response	Slope
Zero	0	-0.10	-	0	-
NO Span	400	399.9	-0.025	400.0	1.008
NO _x Span	400	400.2	0.050	400.0	1.011
API Model 200E NO _x Analyzer Check List					
Test Values	Observed Value	Units	Nominal Range		
RANGE	500	PPB	500 standard		
STABILITY (Zero Gas)	0.1	PPB	< 2 with zero air		
SAMPLE FLOW	508	cc/min	500 ± 50		
OZONE FLOW	78	cc/min	80 ± 15		
PMT	103.4	mV	-20 - 150		
AZERO	94.2	mV	-20 - 150		
HVPS	670	V	420 - 900 constant		
RCCELL TEMP	50.2	°C	50 ± 1		
BOX TEMP	29.4	°C	8 - 48		
PMT TEMP	7.3	°C	7 ± 2		
MOLY TEMP	314.9	°C	315 ± 5		
RCCELL PRESS	8.4	IN-Hg-A	2 - 10 constant		
SAMPLE PRESS	28.7	IN-Hg-A	25 - 30 constant		
NO Span Conc	400	PPB	20 - 20,000		
NO _x Span Conc	400	PPB	20 - 20,000		
NO Slope	1.008	-	1.0 ± 0.3		
NO _x Slope	1.011	-	1.0 ± 0.3		
NO Offset	1.4	mV	-20 to +150		
NO _x Offset	0.9	mV	-20 to 150		
Stability at Zero	0.1	PPB	< 0.2		
Stability at Span	0.2	PPB	< 2 ppb @ 400 ppb span gas		

Calibrated by :

Approved by :



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจตุจักร เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompoi, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

CALIBRATION REPORT					
CHEMILUMINESCENT NO / NO ₂ / NO _x ANALYZER					
DATE :	02 May 2026	BRAND :	API	MODEL :	200E
NO.	NOX-B16	SERIAL NO.	249		
Calibrator (Dilution System)					
Brand :	API	Model :	700		
Last Cal. Date :	01 September 2025	Serial No. :	911		
Reference Standard Gas					
Standard Gas :	Nitric Oxide (NO)	Cylinder No. :	A006745V		
Certified Date :	12 March 2025	Expired Date :	12 March 2028	Cylinder Conc. :	48.7 ppm
CALIBRATING CONDITION					
Pressure	1011	mmbar	Temp.	24.5	°C
			% RH	48	
CALIBRATION SETTING					
Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB	
Set Point	Expected Concentration	Analyzer Response	%Dif	Analyzer Response	Slope
Zero	0	0.10	-	0	-
NO Span	400	399.7	-0.075	400.0	1.005
NO _x Span	400	399.9	-0.025	400.0	1.009
API Model 200E NO _x Analyzer Check List					
Test Values	Observed Value	Units	Nominal Range		
RANGE	500	PPB	500 standard		
STABILITY (Zero Gas)	0.1	PPB	< 2 with zero air		
SAMPLE FLOW	507	cc/min	500 ± 50		
OZONE FLOW	78	cc/min	80 ± 15		
PMT	103.5	mV	-20 - 150		
AZERO	94.2	mV	-20 - 150		
HVPS	673	V	420 - 900 constant		
RCCELL TEMP	50.0	°C	50 ± 1		
BOX TEMP	28.9	°C	8 - 48		
PMT TEMP	7.1	°C	7 ± 2		
MOLY TEMP	314.7	°C	315 ± 5		
RCCELL PRESS	8.2	IN-Hg-A	2 - 10 constant		
SAMPLE PRESS	28.5	IN-Hg-A	25 - 30 constant		
NO Span Conc	400	PPB	20 - 20,000		
NO _x Span Conc	400	PPB	20 - 20,000		
NO Slope	1.005	-	1.0 ± 0.3		
NO _x Slope	1.009	-	1.0 ± 0.3		
NO Offset	1.1	mV	-20 to +150		
NO _x Offset	0.7	mV	-20 to 150		
Stability at Zero	0.1	PPB	< 0.2		
Stability at Span	0.2	PPB	< 2 ppb @ 400 ppb span gas		

Calibrated by :

Approved by :



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sales@spsc.com, www.spsc.com

CALIBRATION REPORT					
CHEMILUMINESCENT NO / NO ₂ / NO _x ANALYZER					
DATE :	02 May 2026	BRAND :	API	MODEL :	200E
NO :	NOX-B19	SERIAL NO. :	353		
Calibrator (Dilution System)					
Brand :	API	Model :	700		
Last Cal. Date :	01 September 2025	Serial No. :	911		
Reference Standard Gas					
Standard Gas :	Nitric Oxide (NO)	Cylinder No. :	A006745V		
Certified Date :	12 March 2025	Expired Date :	12 March 2028	Cylinder Conc. :	48.7 ppm
CALIBRATING CONDITION					
Pressure :	1011 mmbar	Temp. :	24.5 °C	% RH :	48
CALIBRATION SETTING					
Span	Initial Reading (Before Adj.),PPB			Final Reading (After Adj.),PPB	
Set Point	Expected Concentration	Analyzer Response	%Diff	Analyzer Response	Slope
Zero	0	0.11	-	0	-
NO Span	400	400.2	0.050	400.0	1.010
NO _x Span	400	400.3	0.075	400.0	1.013
API Model 200E NO _x Analyzer Check List					
Test Values	Observed Value	Units	Nominal Range		
RANGE	500	PPB	500 standard		
STABILITY (Zero Gas)	0.1	PPB	< 2 with zero air		
SAMPLE FLOW	512	cc/min	500 ± 50		
OZONE FLOW	79	cc/min	80 ± 15		
PMT	103.0	mV	-20 - 150		
AZERO	93.7	mV	-20 - 150		
HVPS	671	V	420 - 900 constant		
CELL TEMP	50.4	°C	50 ± 1		
BOX TEMP	29.0	°C	8 - 48		
PMT TEMP	7.5	°C	7 ± 2		
MOLY TEMP	315.4	°C	315 ± 5		
CELL PRESS	8.2	IN-Hg-A	2 - 10 constant		
SAMPLE PRESS	28.4	IN-Hg-A	25 - 30 constant		
NO Span Conc	400	PPB	20 - 20,000		
NO _x Span Conc	400	PPB	20 - 20,000		
NO Slope	1.010	-	1.0 ± 0.3		
NO _x Slope	1.013	-	1.0 ± 0.3		
NO Offset	1.7	mV	-20 to +150		
NO _x Offset	1.1	mV	-20 to 150		
Stability at Zero	0.1	PPB	< 0.2		
Stability at Span	0.2	PPB	< 2 ppb @ 400 ppb span gas		

Calibrated by :

Approved by :



QUALITY CALIBRATION CO.,LTD.

235 Petchkasem 63/2 Road, Laksong, Bangkac, Bangkok 10160

Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4584

www.qcalibration.com



CERTIFICATE No : 26M2266

REFERENCE No : 80204-1

PAGE : 1 OF 2

Certificate of Calibration

EQUIPMENT : DIGITAL BALANCE

MANUFACTURER : METTLER TOLEDO

MODEL : XS105DU

SERIAL No : 1126422905

ID No : BA 05/50

CONDITION AS RECEIVED : USED ITEM

SUBMITTED BY : S.P.S. CONSULTING SERVICE CO., LTD.
7 SOI PHAHOLYOTHIN 24, PHAHOLYOTHIN RD.,
JOMPOL, CHATUCHAK, BANGKOK 10900

CALIBRATED BY :

CALIBRATION DATE :

05-Mar-26

APPROVED BY :

ISSUED DATE :

06-Mar-26

RECEIVED DATE :

05-Mar-26

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF QUALITY CALIBRATION CO., LTD.





QUALITY CALIBRATION CO.,LTD.

235 Petchkasem 63/2 Road, Laksong, Bangkoe, Bangkok 10160

Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4584

www.qcalibration.com

CERTIFICATE No : 26M2266

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : DIGITAL BALANCE
MANUFACTURER : METTLER TOLEDO
ID No : BA 05/50
AIR PRESSURE : 1004mbar $\pm$ 1mbar
AMBIENT TEMPERATURE : 25° C $\pm$ 1° C
MODEL : XS105DU
S/N : 1126422905
RECEIVED DATE : 05-Mar-26
CALIBRATION DATE : 05-Mar-26
RELATIVE HUMIDITY : 50 %RH $\pm$ 10 % RH

CONDITION OF THIS RESULTS OF CALIBRATION

1. THIS INSTRUMENT WAS CALIBRATED BY ACCORDING TO UKAS LAB 14 EDITION 6:2019 BY USING KNOWN WEIGHT STANDARD WEIGHT. THE BALANCE WAS NOT ADJUSTED BEFORE CALIBRATION. THE BALANCE HAS NO ZERO TRACKING FUNCTION. REPEATABILITY WAS MEASURED BY USING 10 REPEATED MEASUREMENTS. LINEARITY WAS MEASURED COVERING 10 POINTS, EVENLY SPREAD OVER THE RANGE. THE INSTRUMENT WAS SET ZERO BEFORE PERFORMING THE LINEARITY TEST. OFF-CENTER LOADING WAS MEASURED BY USING STANDARD WEIGHTS PLACED ON THE PAN AND MOVED TO VARIOUS POSITIONS ON THE PAN.

2. REFERENCE STANDARD INSTRUMENTS :-

INSTRUMENT	MODEL	SERIAL No	CERTIFICATE No	DUE DATE
1) STANDARD WEIGHT SET	E2	QK-1-151	C02250116	28-Jan-27
2) STANDARD WEIGHT	E2	15843	C02250117	29-Jan-27

3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.

4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.

5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-
- NATIONAL INSTITUTE OF METROLOGY (THAILAND)

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT

1. ZERO SETTING FUNCTION : NORMAL
2. TARE FUNCTION : NORMAL
3. REPEATABILITY OF READING AT 120 g WAS 0.000055 g
4. DEPARTURE FROM NOMINAL VALUE/ LINEARITY

NOMINAL VALUE (g)	BALANCE READING (g)	CORRECTION (g)	UNCERTAINTY ($\pm$ g)
0.00	0.00000	0.00000	0.000063
0.02	0.02000	0.00000	0.000065
0.10	0.10000	0.00000	0.000066
0.20	0.20000	0.00000	0.000066
0.50	0.50001	-0.00001	0.000065
1.00	1.00001	-0.00001	0.000066
2.00	2.00000	0.00000	0.000067
5.00	5.00001	-0.00001	0.000068
10.00	10.00004	-0.00004	0.000070
20.00	20.00006	-0.00006	0.000078
50.00	50.00000	0.00000	0.00012
100.00	100.00002	-0.00002	0.00019
120.00	120.00002	-0.00002	0.00022

5. OFF CENTER LOADING ERROR



POINT	READING (g)
1	50.0000
2	50.0000
3	50.0000
4	50.0000
5	50.0000
OFF-CENTER LOADING	0.0000

NOTE: THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA
THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR $k=2$, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.

END OF CALIBRATION REPORT



SITHIPORN ASSOCIATES CO., LTD. CALIBRATION LABORATORY

451-451/1 Sirinthorn Road, Bangburmu, Bangplud, Bangkok, 10700 Thailand
Tel. +66 2433 8331 Email : calibration@sithiporn.com

SITHIPORN
associates



Cert. No. : SP25026

Pages : 1 of 4

Calibration Certificate

Equipment : UV-VIS SPECTROPHOTOMETER
Manufacturer : PERKINELMER
Model : LAMBDA 25
Serial No.: 501S14123010
ID No.: SP03/58
Calibration Mode : WAVELENGTH ACCURACY
PHOTOMETRIC ACCURACY
STRAY LIGHT

Condition As Found : GOOD

Customer : S.P.S CONSULTING SERVICE CO., LTD.
7 SOI PHAHOLYOTHIN 24, PHAHOLYOTHIN ROAD,
CHOMPHON SUB-DISTRICT, CHATUCHAK DISTRICT,
BANGKOK PROVINCE 10900 THAILAND.

Location : ORGANIC LABORATORY IV

Ambient Temperature : (22.9 $\pm$ 5) °C

Relative Humidity : (53.7 $\pm$ 25) %

Received Date : 22 AUGUST 2025

Calibration Date : 22 AUGUST 2025

Date of Issue : 25 AUGUST 2025

Calibrated by :

Approved by :

This certificate is issued in accordance with the requirements of ISO/IEC 17025 standard, may not be reproduced other than in full, except with the prior written approval of the head of Calibration Laboratory.

Cert. No. : SP25026
Job No. : VC68SP0019
Pages : 2 of 4

Calibration Method :

This instrument was calibrated by using on-site calibration procedure In-house method : CP-SP-01
The calibration procedure to direct measurement wavelength accuracy by using wavelength standard solution, Photometric accuracy by using absorbance standard filter and absorbance standard solution
The calibration procedure used was based on ASTM E275-01, ASTM E925-02

Condition of this result of calibration :

1. Certified reference materials

Material	Ref. type	Cell serial No.	Cert. No.	Due Date
Holmium liquid	RM-HL	29706	126461	24/10/2026
Didymium liquid	RM-DL	28912	126462	24/10/2026
Neutral density filter	RM-1N2N3N	13877	126457	24/10/2026
Potassium dichromate solutions	RM-0204060810	14204	126497	25/10/2026
Potassium Iodide solution	-	KI-0701-001	CI-0185-24	14/05/2026

2. This result of calibration was found accurate as shown on date and place of calibration only.

3. This certificate is traceable to the international system of unit maintained at :

3.1 The UK National Physical Laboratory (NPL)

Result of calibration : Wavelength Accuracy

(Without adjustment)

Material	Certified Values of Reference Material (nm)	UUC* Reading (nm)	Error (nm)	Uncertainty ± (nm)	k Factor
RM-HL	278.13	278.21	0.08	0.16	2.00
	361.25	361.39	0.14	0.16	2.00
	467.82	467.71	-0.11	0.16	2.00
	536.56	536.50	-0.06	0.16	2.00
	640.50	640.36	-0.14	0.16	2.00
RM-DL	740.09	739.85	-0.24	0.16	2.00
	864.94	865.12	0.18	0.16	2.00

UUC* = Unit Under Calibration

Michael P.

Cert. No. : SP25026
Job No. : VC68SP0019
Pages : 3 of 4

Result of calibration : Photometric Accuracy

Material	Wavelength (nm)	Filter S/N	Nominal Absorbance (A)	Certified Absorbance (A)	UUC* Reading Absorbance (A)	Error (A)	Uncertainty ± (A)	k Factor
Neutral Density glass filter	440.0	29381	0.5	0.5443	0.5413	-0.0030	0.0043	2.00
		29914	0.7	0.7484	0.7455	-0.0029	0.0054	2.00
		29360	1.0	1.0527	1.0535	0.0008	0.0032	2.00
	465.0	29381	0.5	0.4948	0.4922	-0.0026	0.0041	2.00
		29914	0.7	0.6906	0.6877	-0.0029	0.0050	2.00
		29360	1.0	0.9695	0.9709	0.0014	0.0031	2.00
	546.1	29381	0.5	0.5090	0.5068	-0.0022	0.0036	2.00
		29914	0.7	0.6985	0.6960	-0.0025	0.0041	2.00
		29360	1.0	0.9814	0.9825	0.0011	0.0031	2.00
	590.0	29381	0.5	0.5375	0.5353	-0.0022	0.0034	2.00
		29914	0.7	0.7256	0.7231	-0.0025	0.0037	2.00
		29360	1.0	1.0213	1.0219	0.0006	0.0032	2.00
	635.0	29381	0.5	0.5223	0.5202	-0.0021	0.0033	2.00
		29914	0.7	0.6927	0.6901	-0.0026	0.0036	2.00
		29360	1.0	0.9744	0.9750	0.0006	0.0032	2.00

UUC* = Unit Under Calibration

Michael B.

Cert. No. : SP25026
Job No. : VC68SP0019
Pages : 4 of 4

Result of calibration : Photometric Accuracy

(Without adjustment)

Material	Wavelength (nm)	Solution (mg/l)	Certified Absorbance (A)	UUC* Reading Absorbance (A)	Error (A)	Uncertainty $\pm$ (A)	k Factor
Potassium dichromate solutions	235.0	20	0.2415	0.2443	0.0028	0.0101	2.00
		40	0.4866	0.4871	0.0005	0.0115	2.00
		60	0.7415	0.7295	-0.0120	0.0067	2.00
		80	0.9854	0.9844	-0.0010	0.0071	2.00
		100	1.2444	1.2425	-0.0019	0.0073	2.00

UUC* = Unit Under Calibration

Condition of this result of calibration : Spectrophotometer PERKINELMER Model LAMBDA 25 S/N 501S14123010

Resolution of Wavelength Mode 0.1 nm

Resolution of Photometric Mode 0.001 A

Parameter Setting

Measurement Mode Wavelength, Absorbance

Wavelength Scan 190 nm - 1100 nm

Scanning Speed 7.5 nm/min

Band width(Wavelength) 1.0

Band width(Vis) 1.0

Band width(Uv) 1.0

Stray Light** UUC* Reading at 220.0 nm	
Transmission T(%)	Absorbance(A)
0.020	3.7032

**Specific Acceptance :

Transmission $\leq$ 1.0 T(%), Absorbance $\geq$ 2.0 A

**Stray light not TISI Accredited

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k , providing a level of confidence of approximately 95%

End of Calibration Certificate



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24 ถนนพหลโยธิน แขวงจอมพล เขตจตุจักร กรุงเทพฯ 10900
7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Calibration Report			
Non-Dispersive Infrared CO Analyzer			
Date :	01 April 2026	Brand :	API
No.	CO-B14	Model :	300EU
		Serial No.	131
Calibrator (Dilution System)			
Brand :	API	Model :	700
Last Cal. Date :	01 September 2025	Serial No. :	911
Reference Standard Gas			
Standard Gas :	Carbon Monoxide (CO)	Cylinder No. :	D711839
Certified Date :	14 March 2024	Expired Date :	14 March 2032
		Cylinder Conc. :	4,580 ppm
Calibrating Condition			
Pressure	1011 mmbar	Temp.	24.5 °C
		% RH	48
Calibration Setting			
Span	Initial Reading (Before Adj.), PPM		Final Reading (After Adj.), PPM
Set Point	Expected Concentration	Analyzer Response	%Diff
Zero	0	0.10	-
CO Span	40.00	39.94	-0.150
			40.00
API Model 300EU CO Analyzer Check List			
Parameter	Observed Value	Units	Nominal Range
Range	50	PPM	0-1000 ppm
Stability	0.10	PPM	+ 1 ppm With Zero Air
CO Measure	4016.4	mV	2500-4800 mV
CO Reference	3948.2	mV	2500-4800 mV
Measure/Reference Ratio	1.180	-	1.1-1.3 W/Zero Air
Sample Pressure	28.5	In-Hg-A	~2"± Ambient Absolute Pressure
Sample Flow	808	CC/Min	800 ± 10%
Sample Temperature	48.3	°C	48 ± 4
Bench Temperature	48.1	°C	48 ± 2
Wheel Temperature	68.2	°C	68 ± 2
Box Temperature	30.8	°C	Ambient Temp + 7 ± 10
Photo-Drive	3036.1	mV	250 mV to 4750 mV
Slope	1.017	-	1.0 ± 0.3
Offset	0.2	-	0 ± 0.3

Calibrated by :

Approved by :

เอกสาร 6-2

เอกสารการสอบเทียบเครื่องมือการตรวจวัดระดับเสียง



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-69/0204

MTC No. EEL. BP. 152/0169

CALIBRATION CERTIFICATE

Submitted by : S.P.S.CONSULTING SERVICE CO.,LTD.

Address : 7 Soi Phaholyothin 24, Phaholyothin Road, Jompol, Chatuchak, Bangkok 10900.

Calibrated at : Electrical and Electronic Standards Laboratory, Industrial Metrology and Testing Service Centre,
Soi 1C, Bangpoo Industrial Estate, Sukhumvit Rd., Muang, Samutprakan 10280.

Instrument Calibrated :

Description : Sound Calibrator

Manufacturer : ACO

Model : 2127

Serial No. : 130006

Ambient Environment

Temperature : (23 + 3) °C

Relative Humidity : (50 ± 15) %

Ambient Pressure : (101.325 ± 1.500) kPa

- Standards used :
1. Digital Function Synthesizer NF Electronic DF-193A S/N 122037.
 2. Measuring Amplifier Bruel&Kjaer 2636 S/N 1537484.
 3. Programmable Attenuator Tamagawa TPA-303A S/N OF 2214.
 4. Digital Multimeter Agilent 34401A S/N MY44005560.
 5. Pressure Transmitter Vaisala PTB202AD S/N T0650001.
 6. Audio Analyzer Keithley 2015 P S/N 4106495.
 7. Condenser Microphone Bruel&Kjaer 4180 S/N 2889871.

Calibration Procedure: CP-102-04 based on IEC 60942-2003; The sound pressure level generated by sound calibrator under test shall be measured by standard microphone using an insert voltage technique.

This instrument has been calibrated against standards maintained at Electrical and Electronic Standards Laboratory (EEL), which are traceable to the International System of Units through the National Institute of Metrology (Thailand).

The information on actual reading is attached herewith and the uncertainty limits quoted refer to the measured values only.

Date of Receipt : 29 Jan. 2026

Date of Calibration : 11 Feb. 2026

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5



THAILAND INSTITUTE OF SCIENTIFIC AND TECHNOLOGICAL RESEARCH (TISTR)

Request No. 21-69/0204

MTC No. EEL. BP. 152/0169

The reported expanded uncertainty is based upon a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.

Nominal Output of Unit Under Test = 94 dB re 20μPa at 1000 Hz

Acoustic Output in dB re 20μPa, Corrected to Reference Conditions: 101.325 kPa, 23.0 °C and 50 %RH.

1. Sound Pressure Level

Standard Microphone Type	Measured Sound Pressure Level (dB)	Deviated value (dB)	Uncertainty (dB)	Tolerance limit IEC60942:2003 Class 1
1/2 inch Bruel&Kjaer 4180	93.89	-0.11	± 0.10	±0.40 dB

2. Frequency

Standard Microphone Type	Measured Frequency (Hz)	Deviated value (Hz)	Uncertainty (Hz)	Tolerance limit IEC60942:2003 Class 1
1/2 inch Bruel&Kjaer 4180	999.9	-0.1	± 1.5	±1.0%

3. Total Distortion

Standard Microphone Type	Measured Total Distortion (%)	Uncertainty (%)	Tolerance limit IEC60942:2003 Class 1
1/2 inch Bruel&Kjaer 4180	1.90	± 0.50	±3.0%

Note : 1. No adjustment.

2. The calibrator pressure correction was not included.

3. The microphone volume correction was not included.

Calibrated by :

Approved by :

Date of Calibration : 11 Feb. 2026

Date of Issue : 12 Feb. 2026

Electrical and Electronic Standards Laboratory
Industrial Metrology and Testing Service Centre

Ref : 2011269012900497001

End of Certificate

2 / 2

The results relate only to the items tested/calibrated or value assigned.

Advertising the Report/Certificate and publicity of the results except in full are prohibited unless written permission is obtained from the governor of TISTR.

FM.BL.MTC.002 Rev.5

Head Office

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9036
Fax. (66) 0 2577 9009

Office/Laboratory

668 Mu 2 Tambon Bangpoo Mai, Amphoe Muang Samutprakan,
Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
(66) 08 3219 9440
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office

196 Phaholyothin Road, Ladyao, Chatuchak,
Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
(66) 08 1889 6827

Head Office

35 Mu 3 Tambon Khlong Ha, Amphoe Khlong Luang,
Changwat Pathumthani 12120, Thailand
Tel. (66) 0 2577 9036
Fax. (66) 0 2577 9009

Office/Laboratory

668 Mu 2 Tambon Bangpoo Mai, Amphoe Muang Samutprakan,
Changwat Samutprakan 10280, Thailand
Tel. (66) 0 2323 1672-80 ext. 115, 116
(66) 08 3219 9440
E-mail : mtc@tistr.or.th Website : www.tistr.or.th

Office

196 Phaholyothin Road, Ladyao, Chatuchak,
Bangkok 10900, Thailand
Tel. (66) 0 2579 1121-30 ext. 5219, 5225, 5217
(66) 08 1889 6827



บริษัท เอส.พี.เอส. คอนซัลติ้ง เซอร์วิส จำกัด
S.P.S. CONSULTING SERVICE CO., LTD.
7 ซอยพหลโยธิน 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900
Tel : (662) 939-4370-72, Fax : (662) 513-4221, E-mail : sale@spscon.com, www.spscon.com

Noise B_250/26

Sound Level Meter Calibration Report

Acoustic Calibrator Data

Brand	ACO	Number	AC 03/56
Model	2127	Serial No.	130006
Calibration Range	94 dB, 1000 Hz	Last Calibration	11 February 2026
		Due Date	11 February 2027

Calibration Data

Sound Level Meter Data				Calibration Data		
SLM No.	Brand	Model	Serial No.	Date	Actual Reading [dB]	
					Before Adjustment	After Adjustment
ACO-B16	ACO	6236	00172039	02 May 2026	93.9	93.9
ACO-R42	ACO	6236	00192054	02 May 2026	93.8	93.9
ACO-C1-B01	ACO	6238	00223038	02 May 2026	93.9	93.9
ST-C1-B24	SCARLET	ST-12D	10352330	02 May 2026	93.8	93.9
ST-C1-B27	SCARLET	ST-12D	10352334	02 May 2026	93.7	93.9
ST-C1-B29	SCARLET	ST-12D	10352342	02 May 2026	93.9	93.9
Acoustic Certified Value : Thailand Institute of Scientific and Technological Research (TISTR)					93.89 ± 0.10 dB	

Calibrated by :



Approved by :



เอกสาร 6-3

เอกสารสอบเทียบเครื่องมือการตรวจวิเคราะห์คุณภาพน้ำ

Certificate of Calibration

Certificate No. : 69-400023-2

Page : 1 of 2

Submitted by : S. P. S Consulting Service Co., Ltd.

7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol, Chatuchak, Bangkok 10900

Equipment : Liquid in Glass Thermometer

Manufacturer : SK

Model : N/A

Range : 0 °C to 100 °C

Resolution : 1 °C

Serial No. : N/A

Immersion : Total

ID No. : TM21/59

Environment : Ambient Temperature : (23 ± 2) °C

Relative Humidity : (50 ± 15) %

Line Voltage : (220 ± 22) VAC

Date of Received : 13 January 2026

Date of Calibration : 16 January 2026

Date of Issue : 16 January 2026

Calibrated by : Chortip Samchusri

Calibration Method : This instrument was calibrated by In-house method comparison technique CAL-M4001 based on ASTM E77-07 by compared with PRT in the liquid bath at the constant controlled temperature.

The temperature scale used was based on ITS-90

Reference Standard Instruments : This certification is traceable to the International System of Units

1. Platinum Resistance Thermometer (PRT)

ID No.	Cert. No.	Due Date	Traceability
400001	TT-0023-24	16 Feb 2026	National Institute of Metrology Thailand (NIMT)

2. Standard Digital Thermometer

ID No.	Cert. No.	Due Date	Traceability
400003	25E1656	22 May 2027	National Institute of Metrology Thailand (NIMT)
400004	25E1656	22 May 2027	National Institute of Metrology Thailand (NIMT)

Approved by :

Supervisor

The Uncertainties are for a confidence probability of approximately 95%

This certificate may not be reproduced other than in full except with the prior written approval of the Calibratech Co.,Ltd.



Certificate of Calibration

Certificate No. : 69-400023-2

Page : 2 of 2

Result of Calibration : Without Adjustment

UUC Condition As-Received : Good

Function : Temperature measurement

Ice point check : UUC* reading 0 °C Standard reading 0.4773 °C

Standard Reading (°C)	UUC Reading (°C)	Correction (°C)	Uncertainty (± °C)
20.3922	20	0.4	0.31

Remark

UUC : Unit Under Calibration

This result of calibration was found accurate as shown on date and place of calibration only.

This reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2 , providing a level of confidence of approximately 95%

- o O o -





CERTIFICATE No : 26E3363
REFERENCE No : 80514-1

PAGE : 1 OF 3

Certificate of Calibration

EQUIPMENT : pH METER
MANUFACTURER : HANNA
MODEL : HI3512
SERIAL No : TH118035
ID No : PH 04/56
CONDITION AS RECEIVED : USED ITEM
SUBMITTED BY : S.P.S. CONSULTING SERVICE CO., LTD.
7 SOI PHAHOLYOTHIN 24, PHAHOLYOTHIN RD.,
JOMPOL, CHATUCHAK, BANGKOK 10900

CALIBRATED BY : 
CALIBRATION DATE : 02-Apr-26

APPROVED BY : 

ISSUED DATE : 02-Apr-26

RECEIVED DATE : 01-Apr-26



CERTIFICATE No : 26E3363

PAGE : 2 OF 3

Calibration Report

EQUIPMENT : pH METER
MANUFACTURER : HANNA
ID No : PH 04/56
RECEIVED DATE : 01-Apr-26
AMBIENT TEMPERATURE : 23 ° C ± 3 ° C
MODEL : HI3512
SERIAL NUMBER : TH118035
CALIBRATION DATE : 02-Apr-26
RELATIVE HUMIDITY : 50 % RH ± 10% RH

CONDITION OF THIS RESULTS OF CALIBRATION

1. THIS INSTRUMENT WAS CALIBRATED BY DIRECT MEASUREMENT METHOD BASED ON WI-TQ-062 AND WI-TQ-063. THE DISPLAY UNIT WAS TESTED BY GENERATING STANDARD VOLTAGE TO THE UNIT AND READING THE VALUE COMPARED WITH THE CALCULATED VALUE. THE DISPLAY AND ELECTROD WAS CALIBRATED BY USING STANDARD pH BUFFER
2. REFERENCE STANDARD INSTRUMENTS :-

INSTRUMENT	MODEL	SERIAL No/ LOT No	CERTIFICATE No	DUE DATE
1) pH STANDARD SOLUTION	00651-36	CC829914	4956-16178366	11-Sep-27
2) pH STANDARD SOLUTION	00651-38	CC831093	4957-16224157	01-Oct-27
3) pH STANDARD SOLUTION	00651-40	CC826321	4958-16046326	10-Jul-27
4) PROCESS CALIBRATOR	CA150	9186079	26E936	17-Mar-27
5) BATH	260014	1247 48074	25T9431	12-Sep-26
6) THERMOMETER WITH PROBE	421504	55000379	25T9466	12-Sep-26

3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.
4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.
5. THIS CERTIFICATE IS TRACEABLE TO SI UNIT MAINTAINED AT :-
 - NATIONAL INSTITUTE OF STANDARD AND TECHNOLOGY, USA.
 - NATIONAL INSTITUTE OF METROLOGY (THAILAND)

RESULT OF CALIBRATION : ADJUSTMENT

1. DISPLAY UNIT ONLY

SLOPE FACTOR k = 2.303 RT/F = 59 mV/pH

mV APPLIED	UUC READING (mV)	CORRECTION (mV)	UUC READING (pH)	UNCERTAINTY OF MEASUREMENT (± mV)	COVERAGE FACTOR k
414.11	414.3	-0.19	-0.256	0.15	2.00
354.95	355.1	-0.15	0.759	0.15	2.00
295.80	296.0	-0.20	1.773	0.15	2.00
236.64	236.8	-0.16	2.787	0.15	2.00
177.48	177.7	-0.22	3.801	0.15	2.00
118.32	118.5	-0.18	4.816	0.15	2.00
59.16	59.4	-0.24	5.831	0.15	2.00
0.00	0.2	-0.20	6.845	0.15	2.00
-59.16	-58.9	-0.26	7.878	0.15	2.00
-118.32	-118.1	-0.22	8.915	0.15	2.00
-177.48	-177.2	-0.28	9.952	0.15	2.00
-236.64	-236.4	-0.24	10.990	0.15	2.00
-295.80	-295.5	-0.30	12.027	0.15	2.00
-354.95	-354.7	-0.25	13.064	0.15	2.00
-414.11	-413.8	-0.31	14.101	0.15	2.00

END OF CALIBRATION REPORT PAGE 2 OF 3





QUALITY CALIBRATION CO., LTD.
235 Petchkasem 63/2 Road, Laksong, Bangkok, Bangkok 10160
Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4584

CERTIFICATE No : 26E3363

PAGE : 3 OF 3

Calibration Report

RESULT OF CALIBRATION (CONTINUE) :

2. DISPLAY UNIT WITH pH ELECTRODE S/N: 11340D0M

STANDARD pH BUFFER SOLUTION (pH)	UUC READING (pH)	CORRECTION (pH)	VALUE BEFORE ADJUSTMENT	UNCERTAINTY OF MEASUREMENT ($\pm$ pH)	COVERAGE FACTOR k
4.008	4.008	0.000	4.142	0.012	2.00
6.998	7.000	-0.002	7.068	0.012	2.00
10.007	10.009	-0.002	10.226	0.014	2.00

3. DISPLAY UNIT WITH TEMPERATURE

STANDARD READING ($^{\circ}$ C)	UUC READING ($^{\circ}$ C)	CORRECTION ($^{\circ}$ C)	VALUE BEFORE ADJUSTMENT	UNCERTAINTY OF MEASUREMENT ($\pm$ $^{\circ}$ C)	COVERAGE FACTOR k
25.004	25.0	0.004	---	0.0085	2.00

4. PERCENT SLOPE 91%

UUC : UNIT UNDER CALIBRATION

THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR k , PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.

END OF CALIBRATION REPORT



CALIBRATION LABORATORY Co., LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Road, Ladphrao, Bangkok 10230
Tel. 02-578-0353-4 Fax. 02-578-2672 www.cal-laboratory.com E-mail: sale@cal-laboratory.com



CERTIFICATE OF CALIBRATION

FOR

NOMENCLATURE : CONDUCTIVITY METER
MANUFACTURER : METTLER TOLEDO
MODEL / TYPE : SEVEN COMPACT S230
SERIAL NO. : C141708983/5821320179[CD 05/65]
CLID. NO. : 272300452
JOB CONTROL NO. : 260202014122
CALIBRATION SERVICE : ☒ IN-LABORATORY ☐ ON-SITE
CUSTOMER : S.P.S. CONSULTING SERVICE CO., LTD.
7 SOI PHAHOLYOTHIN 24 ROAD., JOMPOL,
CHATUCHAK, BANGKOK 10900

DATE OF RECEIVED : 02 February 2026

DATE OF ISSUED : 04 February 2026

The report of calibration shall not be reproduced except in full without approval of the calibration Laboratory Co., Ltd.

Calibrated By :



Calibration Engineer

Approved By :



Authorized Signatory

04 February 2026



This Calibration Certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI)

Certificate No. Q26014122

F3-011-05/12-23

page 1 of 4



REPORT OF CALIBRATION

FOR

NOMENCLATURE : CONDUCTIVITY METER
MANUFACTURER : METTLER TOLEDO
MODEL / TYPE : SEVEN COMPACT S230
SERIAL NO. : C141708983/5821320179[CD 05/65]
DATE OF CALIBRATION : 03 February 2026

ENVIRONMENT CONDITIONS :

Temperature : $(25 \pm 2.5) ^\circ\text{C}$ Relative Humidity : $(50 \pm 15) \% \text{ RH}$

PROCEDURE USED :

This instrument [Conductivity Meter] was calibrated under procedure No. WI-305-130.

The calibration was performed by direct measurement with Certified Reference Material (CRM) and Reference Material (RM) .

This instrument [Temperature] was calibrated by comparison with Temperature Bath, Precision Thermometer and IPRT which maintained by the Calibration Laboratory Co., Ltd.

REFERENCE STANDARD USED :

1. Conductivity Standard , Hanna Product Code HI7033L Lot Number 0101.
2. Potassium Chloride Solution, Sigma-Aldrich Canada Co Product No. 1.01203.0500.
3. Potassium Chloride Solution, Sigma-Aldrich Canada Co Product No. 1.01254.0500.
4. Temperature Bath, Jingmi Model RTS-35S S/N. 16339.
5. Precision Thermometer, ASL Model F200-A-8 S/N. 014433/03.
6. IPRT, ASL Model T100-250-1D S/N. PO106346-1-13.

Certificate No. Q26014122

F3-011-05/12-23

page 2 of 4



TRACEABILITY :

1. The measurements are traceable to International System of Units (SI) , through Hanna instruments.
Certificate No. 07D45, Due Date April 2027.
2. The measurements are traceable to International System of Units (SI) , through Sigma-Aldrich Canada Co.
Lot Number. HC56384703, Due Date 31 January 2028.
3. The measurements are traceable to International System of Units (SI) , through Sigma-Aldrich Canada Co.
Lot Number. HC33528554, Due Date 30 November 2026.
4. The measurements are traceable to International System of Units (SI) , through Calibration Laboratory Co., Ltd.
Certificate No. Q25097314, Due Date 29 August 2026.
5. The measurements are traceable to International System of Units (SI) , through Thailand Institute of Scientific and Technological Research (TISTR). Certificate No. PSL-T 0177/68, Due Date 10 February 2026.
6. The measurements are traceable to International System of Units (SI) , through National Institute of Metrology (Thailand).
Certificate No. TT-1008-25, Due Date 04 March 2026.

UNCERTAINTY :

The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor complies with the table which for a normal distribution corresponds to a coverage probability of approximately 95 %.

It has been evaluated according to the "Evaluation of the Uncertainty of Measurement in Calibration (EA-4/02 M:2022)"

Certificate No. Q26014122

F3-011-05/12-23

page 3 of 4





CALIBRATION LABORATORY Co.,LTD.

2/10-11,14,55 Soi Prasert Manukit 29 Yaek 4, Prasert Manukit Road, Ladphrao, Bangkok 10230
Tel. 02-578-0353-4 Fax. 02-578-2672 www.cal-laboratory.com E-mail: sale@cal-laboratory.com



CONDITION OF CALIBRATION ITEM : RECEIVED IN GOOD OPERATIONAL CONDITION

MEASUREMENT RESULTS : (X) without adjustment () adjustment

The table in the following gives the calibration results and associated measurement uncertainties of Conductivity Meter.

CALIBRATION DATA

1. Conductivity Solution Test @ 25°C

Standard Conductivity Solution	DUC Reading	Uncertainty of Measurement	k Factor
*84 µS/cm	83.95 µS/cm [Cell Constant 0.470]	± 1.15 µS/cm	2,00
1414 µS/cm	1414 µS/cm [Cell Constant 0.510]	± 21 µS/cm	2,00
12.83 mS/cm	12.83 mS/cm [Cell Constant 0.570]	± 0.19 mS/cm	2,00

Note. The Scope of Accredited TISI Certificate No. 23-LB0092 Issue 02 Page 91 of 138

* means Calibrations marked "Not TISI Accredited" in this Certificate have been included for completeness.

*2. TEMPERATURE RESULT

Immersion depth (mm)	Actual Temperature (°C)	DUC Reading (°C)	Correction (°C)	Uncertainty ± (°C)
100	25.00	25.1	-0.10	0.08

Technical Note. Type of sensor : Conductivity Probe

Probe Ø 12 mm

The reported uncertainty is based on a standard uncertainty multiplied by coverage factor of $k = 2$.

Note. * means Calibrations marked "Not TISI Accredited" in this Certificate have been included for completeness.

This report is valid for the above stated instrument/s only.

End of Certificate

Certificate No. Q26014122

F3-011-05/12-23

page 4 of 4



QUALITY CALIBRATION CO.,LTD.

235 Petchkasem 63/2 Road, Laksong, Bangkac, Bangkok 10160

Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4584

www.qcalibration.com



PAGE : 1 OF 2

Certificate of Calibration

CERTIFICATE No : 26M2268

REFERENCE No : 80204-3

EQUIPMENT : DIGITAL BALANCE

MANUFACTURER : SARTORIUS

MODEL : BSA224S-CW

SERIAL No : 36591843

ID No : BA 09/61

CONDITION AS RECEIVED : USED ITEM

SUBMITTED BY : S.P.S. CONSULTING SERVICE CO., LTD.
7 SOI PHAHOLYOTHIN 24, PHAHOLYOTHIN RD.,
JOMPOL, CHATUCHAK, BANGKOK 10900

CALIBRATED BY : 

CALIBRATION DATE : 05-Mar-26

APPROVED BY : 

ISSUED DATE : 06-Mar-26

RECEIVED DATE : 05-Mar-26

THIS CERTIFICATE MAY NOT BE REPRODUCED OTHER THAN IN FULL EXCEPT WITH THE PRIOR WRITTEN APPROVAL OF QUALITY CALIBRATION CO., LTD.



F-G010 REV 03

**QUALITY CALIBRATION CO.,LTD.**

235 Petchkasem 63/2 Road, Laksong, Bangkok, Bangkok 10160

Tel (662) 421-5402, (662) 444-0152-3, Fax (662) 809-4584

www.qcalibration.com

CERTIFICATE No : 26M2268

PAGE : 2 OF 2

Calibration Report

EQUIPMENT : DIGITAL BALANCE MODEL : BSA224S-CW
 MANUFACTURER : SARTORIUS S/N : 36591843
 ID No : BA 09/61 RECEIVED DATE : 05-Mar-26
 AIR PRESSURE : 1004mbar ± 1mbar CALIBRATION DATE : 05-Mar-26
 AMBIENT TEMPERATURE : 25° C ± 1° C RELATIVE HUMIDITY : 50 %RH ± 10 % RH

CONDITION OF THIS RESULTS OF CALIBRATION

1. THIS INSTRUMENT WAS CALIBRATED BY ACCORDING TO UKAS LAB 14 EDITION 6:2019 BY USING KNOWN WEIGHT STANDARD WEIGHT. THE BALANCE WAS NOT ADJUSTED BEFORE CALIBRATION. THE BALANCE HAS NO ZERO TRACKING FUNCTION. REPEATABILITY WAS MEASURED BY USING 10 REPEATED MEASUREMENTS. LINEARITY WAS MEASURED COVERING 10 POINTS, EVENLY SPREAD OVER THE RANGE. THE INSTRUMENT WAS SET ZERO BEFORE PERFORMING THE LINEARITY TEST. OFF-CENTER LOADING WAS MEASURED BY USING STANDARD WEIGHTS PLACED ON THE PAN AND MOVED TO VARIOUS POSITIONS ON THE PAN.

2. REFERENCE STANDARD INSTRUMENTS :-

INSTRUMENT	MODEL	SERIAL No	CERTIFICATE No	DUE DATE
1) STANDARD WEIGHT SET	E2	QK-1-151	C02250116	28-Jan-27
2) STANDARD WEIGHT	E2	15843	C02250117	29-Jan-27

3. THE CERTIFICATE IS VALID FOR THE ITEM CALIBRATED AS SHOWN ON THE DATE AND PLACE OF CALIBRATION ONLY.

4. THIS RESULT EXCLUDE LONG TERM STABILITY OF THE UNIT UNDER CALIBRATION.

5. THIS CERTIFICATE IS TRACEABLE TO THE INTERNATIONAL SYSTEM OF UNIT MAINTAINED AT:-

- NATIONAL INSTITUTE OF METROLOGY (THAILAND)

RESULT OF CALIBRATION :- WITHOUT ADJUSTMENT


1. ZERO SETTING FUNCTION : NORMAL

2. TARE FUNCTION : NORMAL

3. REPEATABILITY OF READING AT 200 g WAS 0.000055 g

4. DEPARTURE FROM NOMINAL VALUE/ LINEARITY

NOMINAL VALUE (g)	BALANCE READING (g)	CORRECTION (g)	UNCERTAINTY (± g)
0.0	0.0000	0.0000	0.00010
0.1	0.1000	0.0000	0.00011
0.2	0.2000	0.0000	0.00010
0.5	0.5000	0.0000	0.00011
1.0	1.0000	0.0000	0.00011
2.0	2.0000	0.0000	0.00011
5.0	5.0000	0.0000	0.00011
10.0	10.0000	0.0000	0.00011
20.0	20.0001	-0.0001	0.00011
50.0	50.0000	0.0000	0.00013
100.0	100.0001	-0.0001	0.00019
200.0	200.0001	-0.0001	0.00032

5. OFF CENTER LOADING ERROR

POINT	READING (g)
1	100.0000
2	100.0000
3	100.0000
4	100.0000
5	100.0000
OFF-CENTER LOADING	0.0000

NOTE: THIS CALIBRATION WAS CARRIED OUT AT THE CUSTOMER'S PLACE AT LABORATORY AREA THE REPORTED UNCERTAINTY OF MEASUREMENT WAS BASED ON A STANDARD UNCERTAINTY MULTIPLIED BY A COVERAGE FACTOR $k=2$, PROVIDING A LEVEL OF CONFIDENCE APPROXIMATELY 95%.

END OF CALIBRATION REPORT



CERT.No.: HS-X014C

Calibration Date : 16 March 2026

Submitted by : S.P.S CONSULTING SERVICE CO.,LTD

7 Soi Phaholyothin 24, Phaholyothin Rd., Jompol,

Chatuchak, Bangkok, Thailand 10900

Avg Room Temp : 20 °C

Avg Water Temp : 20 °C

Air Pressure : 760.00 mmHg

Salinity : 0 ppt

Model : YSI 5000

S/N : 15B100751

Probe : YSI 5010

S/N : 22D100097

ID NO. :

Air Temp ref : S/N. F8065C26

Barometric ref : S/N. F8065C26

Water Temp ref : -

ID NO. HS001

Technician : Kittipong M.

Calibration Details

Calibration Point	100% air sat. (@20 °C, DO = 9.09 mg/l)	(status)	(status)
Measurement 1 (mg/l)	9.09	(PASS)	-
Measurement 2 (mg/l)	9.09	(PASS)	-
Measurement 3 (mg/l)	9.09	(PASS)	-
Measurement 4 (mg/l)	9.09	(PASS)	-
Measurement 5 (mg/l)	9.09	(PASS)	-
Measurement 6 (mg/l)	9.08	(PASS)	-
Measurement 7 (mg/l)	9.08	(PASS)	-
Measurement 8 (mg/l)	9.08	(PASS)	-
Measurement 9 (mg/l)	9.08	(PASS)	-
Measurement 10 (mg/l)	9.08	(PASS)	-

Mean Measurement	9.09	mg/l	-	-
Inaccuracy	0.00	mg/l	-	-

Overall Status (PASS)

Manufacturer Specification

Accuracy = +/- 0.02 mg/l

- 1) This certificate is issued based on the result that are found as shown on date and place of test only.
- 2) The calibration procedure followed in accordance with Harikul Science Co., Ltd.
- 3) This result shall not be used for advertising purpose.

Technician Signature

Laboratory Manager

เอกสาร 6-4

เอกสารสอบเทียบเครื่องมือการตรวจวิเคราะห์คุณภาพตะกอนดิน



WO-11962532

MAINTENANCE AND IPV TEST CERTIFICATE MODEL

Optima 5300

Customer :	S.P.S. Consulting Service Co.,Ltd.	Date Tested:	June 18, 2026
Address :	7, 24 Soi Phaholyothin, Ladyao, Khet, Bangkok 10900	Recommendation Recertification	
User Name		Period	6 Months
Phone:		Recertification Due:	June 18, 2027
E - Mail :		Date Last Certified:	N/A
		Visit Number:	1 of 2
		PerkinElmer Phone:	02-719-6420 ext 206
		PerkinElmer Fax:	02-318-5597

CONFIGURATION TESTED

MODEL	SERIAL NUMBER	SOFTWARE
Optima5300	077C7042401	Winlab32
TESTED EQUIPMENT	CALIBRATION NUMBER	EXPIRATION
IPV Method	N/A	N/A
TEST STANDARD USED	PART NUMBER	EXPIRATION DATE
Multielement Standard	N069-1579	Dec-26
Instrument Cal. STD4	N930-0221	Dec-26
CUSTOMER SUPPLIED	COMMENTS	CUSTOMER INITIALS
2 % HNO3	N/A	N/A
10 % HNO3	N/A	N/A



WO-11962532

MAINTENANCE AND IPV TEST CERTIFICATE MODEL

Optima 5300

SERIAL NUMBER:	077C7042401	DATE TESTED:	June 18, 2026
-----------------------	-------------	---------------------	---------------

1. MECHANICAL CHECKS

A. Inspect and clean all fans and filters. ☐

B. Inspect and replace as necessary, all torch components including the RF coil. ☐

C. Inspect all tubing for sign of clacking or leaking. ☐

D. Adjust water and gas pressure regulator settings. ☐

E. Inspect and leak check pneumatics drawers. ☐

F. Clean the exterior of the instrument. ☐

2. OPTICAL CHECKS

A. Inspect and clean all optical components. ☐

B. As required, check and replace all purgefilters. ☐

C. Recheck optical alignment. ☐

3. COOLING SYSTEM CHECKS

A. Perform preventive maintenance on chiller. ☐

B. Flush out the chiller every year. ☐

4. PERFORMANCE CHECKS

A. Torch View Alignment. ☐

B. Wavelength Calibration. ☐

MAINTENANCE AND IPV TEST CERTIFICATE MODEL

Optima 5300

SERIAL NUMBER: 077C7042401		DATE TESTED:		June 18, 2026	
PARAMETER		SPECIFICATION		FINAL VALUE	
Spectral Resolution : UV					
	As 193.696 nm	≤ 0.009	nm	0.006	nm
	Ni 231.604 nm	≤ 0.011	nm	0.008	nm
	Ni 341.476 nm	≤ 0.015	nm	0.011	nm
Spectral Resolution : VIS					
	Ba 455.403 nm	≤ 0.020	nm	0.015	nm
Precision					
	Zn 206.200 nm	% RSD	≤ 1.0 %	0.39	%
	Mg 280.271 nm	% RSD	≤ 1.0 %	0.41	%
	Mg 285.213 nm	% RSD	≤ 1.0 %	0.31	%
	Ba 455.403 nm	% RSD	≤ 1.0 %	0.68	%
Detection Limits : Axial					
	Tl 190.801 nm	3(sd)	<10	9.23	ppb
	As 193.696 nm	3(sd)	<10	6.78	ppb
	Se 196.026 nm	3(sd)	<5	4.12	
	Pb 220.353 nm	3(sd)	<3	2.34	ppb
Detection Limits : Radial					
				1.23	
	As 193.696 nm	3(sd)	<60	55.23	ppb
	Zn 213.857 nm	3(sd)	<2	1.03	ppb
	Mn 257.610 nm	3(sd)	<0.75	0.65	ppb
	La 379.478 nm	3(sd)	<3	3.15	ppb
	Ba 455.403 nm	3(sd)	<0.3	0.23	ppb
	Ba 493.408 nm	3(sd)	<0.6	1.11	ppb
BEC : Axial (IB X 1000)/(IS-IB)					
	Mn 257.610 nm	≤ 30	ppb	15.87	ppb
BEC : Radial (IB X 1000)/(IS-IB)					
	Mn 257.610 nm	≤ 30	ppb	10.35	ppb

MAINTENANCE AND IPV TEST CERTIFICATE MODEL

Optima 5300

SERIAL NUMBER: 077C7042401 DATE TESTED: June 18, 2026

Remarks :

Test all pass

This is to certify that the above tests have been performed and the configuration tested



meets



does not meet

the PerkinElmer Specifications listed on this certificate.

This certificate does not modify PerkinElmer's standard terms and condition of sale, including warranty terms.

Service Department PerkinElmer Ltd.

Customer Service Engineer:

()

Sr. Customer Support Engineer